Longitudinal Awake Imaging of Mouse Deep Brain Microvasculature with Super-resolution Ultrasound Localization Microscopy

Yike Wang[1,2,6], Matthew R. Lowerison[1,2,6], Zhe Huang[1,6], Qi You[1,3], Bing-Ze Lin[1,2,6], Daniel A. Llano[1,4,5], Pengfei Song[1,2,3,5,6*]

[1] Beckman Institute for Advanced Science and Technology, University of Illinois Urbana-Champaign, Urbana, IL 61801

[2] Department of Electrical and Computer Engineering, University of Illinois Urbana-Champaign, Urbana, IL 61801

[3] Department of Bioengineering, University of Illinois Urbana-Champaign, Urbana, IL 61801

[4] Department of Molecular and Integrative Physiology, University of Illinois Urbana-Champaign, Urbana, IL 61801

[5] Neuroscience Program, University of Illinois Urbana-Champaign, Urbana, IL 61801

[6] Department of Biomedical Engineering, Duke University, Durham, NC, 27708

*Corresponding Author: Dr. Pengfei Song (pengfei.song@duke.edu)

Abstract

Ultrasound localization microscopy (ULM) is an emerging imaging modality that resolves microvasculature in deep tissues with high spatial resolution. However, existing preclinical ULM applications are largely constrained to anesthetized animals, introducing confounding vascular effects such as vasodilation and altered hemodynamics. As such, ULM quantifications (e.g., vessel diameter, density, and flow velocity) may be confounded by the use of anesthesia, undermining the usefulness of ULM in practice. Here we introduce a method to address this limitation and achieve ULM imaging in awake mouse brain. Pupillary monitoring was used to support the presence of the awake state during ULM imaging. Vasodilation induced by isoflurane was observed by ULM. Upon recovery to the awake state, reductions in vessel density and flow velocity were observed across different brain regions. In the cortex, the effects induced by isoflurane are more pronounced on venous flow than on arterial flow. In addition, serial *in vivo* imaging of the same animal brain at weekly intervals demonstrated the highly robust longitudinal imaging capability of the proposed technique. The consistency was further verified through quantitative analysis on individual vessels, cortical regions of arteries and veins, and subcortical regions. This study demonstrates longitudinal ULM imaging in the awake mouse brain, which is crucial for many ULM brain applications that require awake and behaving animals.

Introduction

Sensitive imaging of correlates of activity in the awake brain is fundamental for advancing our understanding of neural function and neurological diseases. In the field of neuroscience, there is a growing interest in utilizing imaging techniques to study the rodent brain, which serves as a valuable model for investigating brain function[1,2]. Brain imaging modalities for rodents encompass a wide range of techniques, including but not limited to functional magnetic resonance imaging (fMRI)[3,4], positron emission tomography (PET)[5,6], one- and two-photon imaging[7,8], photoacoustic imaging[9,10], and more recently ultrasound localization microscopy (ULM)[11]. ULM is uniquely capable of imaging microvasculature situated in deep tissue (e.g., at a depth of several centimeters). ULM can also be combined with the principles of functional ultrasound (fUS)[12–15] to image whole-brain neural activity at a microscopic scale[16]. The structural and functional imaging capabilities of ULM have opened new doors for numerous basic research and clinical applications that involve cerebral microvasculature[17–19].

At present, a key limitation associated with existing ULM brain imaging studies is the use of anesthesia, which induces profound alterations to cerebral blood flow (CBF), including changes in vessel size (e.g., diameter) and flow velocity[20]. As such, CBF measurements under anesthesia do not reflect the blood flow under the normal physiological state of the brain. In addition, anesthetics also have a significant attenuating effect on neural responses to sensory inputs, thereby impacting the neurovascular coupling process[21–26]. Therefore, in neuroscience research, brain imaging in the awake state is often preferred over imaging under anesthesia (e.g., fluorescence imaging[27], photoacoustic imaging[28,29], fMRI[4], PET[6], and fUS[30–33]). As ULM is gaining traction in many preclinical brain imaging applications, enabling ULM for awake animals has become crucial to eliminate the confounding vascular effects of anesthetics and obtain accurate structural and functional cerebrovascular measurements.

Another challenge associated with preclinical ULM brain imaging is to conduct long-term, longitudinal studies, which are essential for tracking disease progression or therapeutic impacts for many neurological disease applications[34–36]. The key technical challenge for longitudinal ULM brain imaging is to find the same imaging plane and reconstruct consistent ULM images across different imaging sessions. Misalignment of imaging planes or tissue movement will undermine ULM imaging quality and result in inconsistent cerebrovascular quantifications. Although intact skull imaging has been shown to be feasible for short-term studies (e.g., over a few days)[18], long-term monitoring can be challenging because changes in skull properties over time (e.g., thickness and composition) could negatively impact ULM imaging quality. Currently there is a strong need for methodological developments to enable longitudinal brain imaging with ULM.

In this study, we developed a method for awake and longitudinal ULM brain imaging in a mouse model to eliminate the confounding vascular effects of anesthesia and enable monitoring of cerebral vasculature over a three-week period in the same animal. We constructed a head-fixed awake imaging platform and established a ULM image reconstruction metric to allow comparisons of ULM images acquired at different states of wakefulness (e.g., awake vs. anesthesia). Our method allowed detailed comparisons of local and global variations of the cerebral vasculature and blood flow under awake and anesthesia conditions. Detailed quantitative analysis of vessel diameter and blood flow velocity was performed. We also demonstrated robust longitudinal ULM brain imaging on same animals with high repeatability over multiple weeks.

Results

ULM brain imaging can be performed in head-fixed awake mice.

To achieve consistent ULM brain imaging while allowing limited movement in awake animals, a head-fixed imaging platform with a chronic cranial window was used in this study. Pre-surgery handling and post-surgery habituation were employed to alleviate animal stress and facilitate awake imaging[33]. To further minimize tissue motion, a 3D-printed body tube was adopted[33] for animal immobilization (Fig. 1a). A headpost was implanted during cranial window surgery to fix the animal's head to the body tube. In this procedure, the skull was removed, and the brain tissue was protected using a polymethyl pentene (PMP) membrane[37], which was further protected by a layer of silicone rubber (Fig. 1b). Animals were imaged after one week of surgical recovery and head-fixation habituation. Prior to each imaging session, the silicone rubber was gently removed with forceps, and the transparent PMP membrane allowed detailed examination of the brain surface to ensure absence of tissue damage (Fig. 1c).

Fig.1 | Experimental setup for awake mice ULM brain imaging. a, 3D model of the body tube enabling rapid fixation of the headpost onto the tube. **b,** Top view photograph of the mouse after cranial window surgery, with the headpost protected by silicone rubber. **c,** Photograph taken after the post-surgery recovery, with the removal of the silicone rubber protection. **d,** Image captured by the camera positioned in front of the mouse during the imaging session. **e,** Change of the pupillary area over time, after microbubble injection (T = 0s), along with magnified portion of the pupil photo in d at six time points (T=0, 200, 400, 600, 800, and 1000 second). Pupil is outlined with red dashed circle. **f,** Awake ULM image obtained from data collected during a fully awake state, as indicated by the gray shaded area in **e**. (scale bar: 1 mm)

Contrast enhancing microbubbles (MBs) were administered intravenously via tail vein in this study (details
provided in Methods). Isoflurane anesthesia was terminated right after microbubble (MB) injection,
marking the beginning of the data acquisition (i.e., T = 0s). Mice were allowed to gradually regain
consciousness. Throughout the awakening process, an infrared camera was used to monitor the pupil to
provide a reference for the state of arousal[38] (Fig. 1d). Fig. 1e shows the change in pupillary area over time,
which reveals a gradual enlargement of the pupil after the cessation of anesthesia. Due to differences in
tail vein injection timing and anesthesia depth, the time required for each mouse to fully awaken varied.
Although it was not feasible to get pupil size stabilized just after 500 seconds for each animal, ULM
reconstruction only used the data that acquired after the animal reached full pupillary dilation, to ensure
that ULM accurately captures the cerebrovascular characteristics in the awake state. Ultrafast plane wave
imaging data were collected for ULM reconstruction (details provided in Methods) and a representative
awake ULM brain image is presented in Fig. 1f.

Microbubble count serves as a quantitative metric for awake ULM image reconstruction.

Due to the stochastic nature of MB localization and fluctuations of MB concentration in the blood stream,
it is challenging to compare ULM images acquired at different states of wakefulness with different cerebral
blood flow conditions. Therefore, to obtain complete, fully saturated ULM images under different
physiological conditions and MB concentrations, we used MB count as a measure of vessel saturation to
determine the completeness of ULM reconstructions. To ensure high quality ULM imaging, only MBs that
were successfully tracked by the uTrack algorithm (details provided in Methods) were considered as
effective MB signals that were utilized for ULM image quantification.

To facilitate equitable comparison of brain perfusion at different states, a practical saturation point
enabling stable quantification of most vessels needs to be established. Our observations indicated that
when the cumulative MB count reached 5 million, ULM images achieved a relatively stable state.
Accordingly, in this study, the saturation point was defined as a cumulative MB count of 5 million. There
are also possible alternatives for ULM image normalization. For example, different ULM images can be
normalized to have the same saturation rate. However, the proposed method of using the same number
of cumulative MB count for normalization enables the analysis of blood flow distribution across different
brain regions from a probabilistic perspective. The following analysis substantiates this criterion.

Fig. 2a compares ULM directional vessel density maps and flow speed maps generated with 1, 3, 5, and 6
million MBs, using the same animal as shown in Fig. 1. To quantitatively confirm saturation, multiple vessel
segments were selected for further analysis. Fig. 2b presents the measured vessel diameter for a specific
segment at various MB counts. After binarizing the ULM map, the vessel diameter was measured by
calculating the distance from the vessel centerline to the edge. Each point along the centerline of the
segment provided a diameter measurement, enabling calculation of the mean and standard error. At low
MB counts, vessels appeared incompletely filled, leading to inaccurate estimation of vessel diameter due
to incomplete profiles. For example, at 1–2 million MBs, the binarized ULM map displayed a width of only
one or two pixels along the segment. As a result, the measurements always yielded the same diameter
values (two pixels, ~10um) with a consistently low standard error of the mean across the entire segment.
With increased MB counts, the measured vessel diameter gradually rose, ultimately reaching saturation.
The plots in Fig. 2b show that vessel diameter stabilized at 5 million MB count. Additionally, Fig. 2c

illustrates the changes in flow velocity measured at different cumulative MB counts. The violin plots display
the distribution of flow speed estimates for all valid centerline pixels within the selected segment. At low
MB counts (1–3 million), flow velocity estimates fluctuated, but they stabilized as the MB count increased
(4–6 million MBs). At 5 million MBs, flow velocity estimates were nearly identical to those at 6 million MBs,
corroborating previous findings that vessel velocity measurements stabilize as MB count grows[39]. To assess
the generalizability of the 5 million MB saturation condition, vessel segments from three different mice
across various brain regions were examined. The results, shown in Supplementary Fig. 2, confirm that this
saturation criterion applies broadly. Although the 5 million MB threshold may not ensure absolute
saturation for all vessels, it is generally effective for vessels larger than 15 μm. This MB count threshold
was therefore adopted as a practical criterion.

To further verify that the proposed MB bolus injection method can help to achieve ULM image saturation
shortly after mice awaken from anesthesia, an analysis on the change in MB concentration over time was
conducted once pupil size had stabilized (T = 500s). Fig. 2d shows a clear trend of decreasing MB
concentration in the blood stream and increasing cumulative MB count with time. Fig. 2e demonstrates a
flattened vessel saturation curve and a rapidly reducing vessel filling rate, which is typical for ULM
reconstruction[39]. The pixel filling rate dropped below 5% of the initial rate after 300 seconds of ULM data
accumulation (T = 800s), indicating ULM image saturation. Fig. 2f further examines the relationship
between the pixel filling rate and the cumulative MB count, which is independent of data acquisition time.
The pixel filling rate at 5 million cumulative MBs is always below 5% of the initial rate for each experiment,
ensuring image saturation. In summary, all the quantitative measurements indicate that ULM images
obtained using the proposed metric (i.e., 5 million MBs) were complete and can be used to consistently
measure cerebral vascular properties such as vessel diameter and blood flow velocity. All subsequent ULM
images in this study were produced using this criterion.

Fig.2 | Data processing standards for awake mice ULM imaging. a, ULM directional vessel density maps and flow velocity maps at cumulative MB count of 1, 3, 5, and 6 million. **b,** Vessel diameter measurements along the selected segment at varying cumulative MB counts. Each point along the vessel centerline provides a measurement, with the mean and standard error of the mean (SEM) plotted to show average diameter and variability. **c,** Flow velocity measurements along the selected segment at varying cumulative MB counts. Each non-zero centerline pixel provides a velocity value, with the distribution across positions shown as a violin plot. **d,** Time courses of MB count in each second (blue curve) and the cumulative MB count (orange curve) starting from T = 500s. The vertical gray dashed lines indicate the time points when the cumulated MB count reaches 1, 2, 3, 4, 5, and 6 million. **e,** Time courses of saturation level of the ULM image (blue curve) and the filling rate of pixels (orange curve). The filling rate is calculated by taking the derivative of the filled pixel count, and then normalized to the initial filling rate at the beginning of ULM reconstruction (T = 500s). **f,** Relationship between pixel filling rate and cumulative MB count, eliminating the time axis by plotting the orange curves from d and e together. This figure presents a case study based on the same mouse shown in Fig 1. The x-axis for **d-f** begins at 500 seconds because, at this point, the mouse's pupil size stabilized, indicating it had recovered to an awake state. Consequently, ULM images were accumulated starting from this time. It is important to note that not every mouse requires 500 seconds to fully awaken; the time to reach a stable awake state varies across individual mice.

Fig.3 | Comparison of ULM images in isoflurane anesthetized and awake states. The ULM images of three different mice are shown. The upper panel shows the comparison of directional vessel density maps, while the lower panel shows the comparison of flow velocity. Four regions of interest (ROIs) are selected within each coronal plane (indicated by white dashed boxes in the whole-plane view of the vessel density map) for zoom-in comparison.

ULM reveals increase in blood flow induced by isoflurane anesthesia.

Fig. 3 presents a comparison of ULM directional vessel density maps and flow velocity maps in awake and anesthetized states for three different coronal planes from three animals. Four regions of interest (ROIs) were selected within each imaging plane to provide detailed comparisons. When comparing vessel density maps, ULM images that were acquired in the awake state demonstrate a global reduction of vascularity, which refers to percentage of pixels that occupied by blood vessels. The reduction is also clearly observed in magnified local regions especially for regions 5, which encompasses the pretectal region. In addition to reduced vascularity, ULM also reveals decreased vessel diameter which reflects vasoconstriction after the animals woke up (e.g., white arrows in regions 1, 3, and 6, corresponding to the thalamus, cortex and midbrain/cortex overlap region). Focusing on ULM flow speed maps, a global reduction in flow speed can be clearly observed when transitioning from anesthetized to awake. Regional maps further revealed the significant flow speed reduction for most of the vessels throughout the brain (e.g., blue arrows).

To quantitatively compare the results from Fig. 3, we performed various analyses in Fig. 4. Since previous studies have shown that isoflurane induces different effects on arteries and veins[20,40,41], we first selected a cortical artery and vein for comparison. In Fig. 4a, a cortical region near Region 3 from Fig. 3 was selected, and one artery and one vein in this region were analyzed. Fig. 4b compares the vessel diameters of these segments in anesthetized and awake states. For the selected artery, no significant change in diameter was observed, with both anesthetized and awake states showing an average diameter of 22 μm. In contrast, the selected vein displayed a notable response to isoflurane, with its diameter decreasing significantly from 54 μm under isoflurane anesthesia to 41 μm in the awake state. Flow velocity analysis of the same vessels further suggests that isoflurane may have a greater impact on venous blood flow velocity: arterial flow velocity increased slightly from 7.33 mm/s under anesthesia to 8.04 mm/s in the awake state, whereas venous flow velocity decreased from 10.48 mm/s to 8.11 mm/s, with both the magnitude and statistical significance of the decrease greater in the vein.

While single-vessel analysis provides valuable insights, its generalizability is limited. To validate the broader applicability of our findings, we conducted ROI-based analyses of cortical arteries and veins, using vascularity and mean velocity as primary metrics. The relationship between arteriovenous classification and blood flow direction is more straightforward to infer on cortex, with downward flow representing cortical arteries and upward flow representing cortical veins[16]. Fig. 4d shows, for Mouse 1, the comparison of cortical upward (venous) and downward (arterial) flow density maps under different states. Supplementary Fig. 3 also provides a detailed comparison of upward and downward flow for all three mice in both anesthetized and awake conditions. Fig. 4e demonstrates that transitioning from isoflurane anesthesia to the awake state led to a reduction in vascularity for both arteries and veins within the selected cortical ROI across all three mice. However, the decrease in venous vascularity (averaging 55% across mice) was greater than that of arterial vascularity (averaging 35%). Fig. 4f further illustrates the change in mean velocity within the selected ROI, where both arteries and veins showed a decreasing trend; however, venous flow velocity reduced by an average of 38%, compared to a 19% reduction in arterial flow velocity across the cortical ROI. These results underscore the different responses of arteries and veins to isoflurane anesthesia, which is also indicated by previous studies[20,40,41].

To visualize flow velocity changes more clearly, we highlighted the flow velocity comparison for the same region in Fig. 4g and provided a violin plot of velocity estimates within the ROI in Fig. 4h. The violin plots

reveal a consistently greater reduction in venous flow velocity from anesthesia to the awake state across
all three mice.

For deeper brain regions, flow analyses were conducted without separating arteries and veins, as this
physiological separation is not established for ULM in these regions. Fig. 4i displays the selected ROIs for
each mouse, while Fig. 4j and 4k present the trends in vascularity and flow velocity changes across
different brain regions. Results indicate that vascularity and flow velocity decreased across the brain as
mice transitioned from anesthetized to awake states, with subcortical ROIs showing more pronounced
changes than cortical ROIs. For example, in the thalamus and midbrain regions, average vascularity
decreased by 37% and average flow velocity by 17%, whereas cortical ROIs exhibited smaller decreases of
4% in vascularity and 11% in flow velocity. This work is consistent with a previous fMRI study, where it was
also observed that isoflurane-induced cerebral hyperemia was not most pronounced in the cerebral cortex
compared with other deeper brain regions[42].

In summary, statistical analysis revealed a decrease in individual vessel diameter, vascularity, and blood
flow velocity (particularly in venules) after awakening. These findings align with existing research,
indicating higher blood perfusion during isoflurane anesthesia[20,40,41,43,44].

Fig.4 | Quantitative comparison of the ULM images in isoflurane anesthetized and awake states. a, Example cortical region near Region 3 in Mouse 1 (from Fig. 3), with vessel density maps displayed separately by flow direction. In this cortical region, upward flow corresponds to venous blood, while downward flow corresponds to arterial blood. One artery and one vein segment were selected within the boxed areas for analysis of flow velocity and diameter. **b,c,** Comparisons of vessel diameter **(b)** and flow velocity **(c)** for the selected arterial and venous segments. Statistical analysis was conducted using t-test at each measurement point along the segments. **d,** Whole cortical region in Mouse 1, showing separate blood flow density maps for upward (venous) and downward (arterial) flow, with a defined ROI (white circled region) used for quantitative analysis. **e,f,** Quantitative assessment of vascularity **(e)** and mean flow velocity **(f)** within the cortical ROI for Mouse 1–3, presented as individual values in each state (left y-axis) and the relative reduction from the anesthetized baseline to the awake state (right y-axis). **g,** Flow velocity map for the same cortical region shown in **d**, using Mouse 1 as an example. **h,** Flow velocity distributions within the cortical ROI for Mouse 1–3, comparing anesthetized and awake states. **i,** Regional analysis across different brain areas using bidirectional ULM maps, with selected ROIs from the cortex (CTX), hippocampal formation (HPF), thalamus (Thal), and midbrain (MBr) for each mouse. **j,k,** Quantitative analysis of vascularity **(j)** and mean blood flow velocity **(k)** for the ROIs in each brain region in Mouse 1–3. (*: p<0.05, **: p<0.01, ***: p<0.001)

Awake ULM imaging demonstrates high consistency in longitudinal imaging across different weeks.

Longitudinal awake ULM brain imaging was feasible using the surgical and imaging techniques presented in this study. Fig. 5 presents the results of awake brain imaging performed on the same brain region over three consecutive weeks. Two ROIs at different depths were selected for each mouse to compare micro vessel reconstruction across different time points. Fig. 5 demonstrates a high level of consistency in both directional vessel density maps and flow speed maps obtained from the three imaging sessions, although some minor discrepancies were observed. The inconsistency could potentially be attributed to physiological variation and/or slight misalignment.

To quantitatively evaluate the consistency of ULM imaging in longitudinal studies, we conducted a similar analysis to that shown in Fig. 4. First, a case study was performed on individual vessels by selecting an artery and a vein in the cortex of Mouse 4 (Fig. 6a) and measuring vessel diameter (Fig. 6b) and flow velocity (Fig. 6c) along the selected segments. The results show small variation in measurements across weeks. For vessel diameter, both arterial and venous segments demonstrated statistical equivalence across the three weekly measurements. Statistical testing of equivalence was conducted using the two one-sided test (TOST) procedure, which evaluates the null hypothesis that the difference between the two weeks is larger than three times the standard deviation of one week. Large p-values from the TOST analyses indicate rejection of the null hypothesis, supporting that the measurements are equivalent. For flow velocity, apart from a slight decrease in the venous flow during the second week (4.49 mm/s, compared to 6.03 mm/s in the first week and 6.00 mm/s in the third week), no other significant differences were observed. Overall, quantitative measurements maintained high consistency, though slight variations were expected due to differences in probe positioning and physiological variations among the mice.

To further generalize these findings and examine longitudinal variation in ROI-based analysis, we used Mouse 4 as an example to show the consistency of blood flow density across different flow directions in the cortex (Fig. 6d) and extended the quantitative analysis to all three mice (Fig. 6e). In Supplementary Fig. 4, we present a comparison of upward and downward flow images for all three mice over the three-week longitudinal study period. In the comparative study shown in Fig. 4, clear differences were seen between arterial and venous responses under various conditions, while in the longitudinal study, both arteries and veins exhibited similar levels of variation. For instance, in Mouse 6, both arterial and venous vascularity showed a notable decrease from around 60% in the first week to around 40% in the second week, which is a relatively large variation compared with other measurements. This quantitative analysis aims to inform readers of the expected range of variation of this technique. Fig. 6f displays the flow velocity map for the same region as in Fig. 6d, with flow velocity distributions inside the selected ROI for all three mice shown in Fig. 6g. While arterial and venous flow velocity distributions exhibit clear distinctions, their variations over the three weeks remained acceptable. Even in the previously mentioned case of Mouse 6 during Week 1 and Week 2, arterial mean flow velocity only shifted from 6.31 mm/s in the first week to 7.00 mm/s in the second week.

We also conducted ROI analyses without distinguishing between arteries and veins for subcortical regions (Fig. 6h–k), using the same ROIs as those shown in Fig. 5. Although subcortical regions show larger difference to anesthesia-induced changes, they maintained relatively stable values in the longitudinal study. In terms of vascularity (Fig. 6i), Mouse 5 exhibited the greatest variation among the three mice, with vascularity increasing from 55% in the first week to 67% in the third week. Compared to the maximum 40%

257 reduction seen from anesthesia to awake states in Fig. 4j, vascularity measurement in subcortical regions
258 remained consistent. This observation also applied to mean velocity measurements.

Fig.5 | Longitudinal awake ULM imaging results on the same coronal plane for three consecutive weeks. The ULM images of three different mice are shown. The upper panel shows the comparison of directional vessel density maps, while the lower panel shows the comparison of flow velocity. Two regions of interest (a cortical ROI and a sub cortical ROI) are selected within each coronal plane (indicated by white dashed boxes in the whole-plane view of the vessel density map) for zoom-in comparison.

Fig.6 | Analysis of ULM images acquired in the three consecutive weeks. a, Two example cortical regions in Mouse 4 (from Fig. 5), with one for artery selection and the other for vein. The vessel segments were selected within the boxed areas for analysis of flow velocity and diameter. **b,c,** Comparisons of vessel diameter **(b)** and flow velocity **(c)** for the selected arterial and venous segments. Statistical analysis was conducted using the two one-sided test (TOST) procedure, which evaluates the null hypothesis that the difference between the two weeks is larger than three times the standard deviation of one week. **d,** Whole cortical region in Mouse 4, showing separate blood flow density maps for upward (venous) and downward (arterial) flow, with a defined ROI (white circled region) used for quantitative analysis. **e,** Bar plot for the vascularity comparison of the cortical artery and vein among Mouse 4-6. **f,** Flow velocity map for the same cortical region shown in **d**, using Mouse 4 as an example. **g,** Flow velocity distributions within the cortical ROI for Mouse 4-6, comparing arterial and venous flow across three weeks. **h,** Bidirectional ULM vessel density map of subcortical ROIs for the three different mice. **i,** Bar plot for the vascularity comparison of the subcortical ROIs in **h. j,** ULM flow velocity map of the same subcortical ROIs for the three different mice. **k,** Bar plot for the mean velocity comparison of the subcortical ROIs. (*: p<0.05, **: p<0.01, ***: p<0.001)

Discussion

In this study, we introduced a method for performing ULM brain imaging in awake mice under a longitudinal study setting. Our method enabled high-resolution imaging of deep cerebral micro-vasculature with the animal under the awake state. We translated the awake imaging techniques previously described in fUS[33] to our study to enable awake ULM, and established a quantitative metric for ULM image reconstruction. Based on the setup above, we studied CBF changes induced by anesthesia, which aligned well with literature. Isoflurane has been shown to increase vascular diameter and CBF in mice, as validated by multiple imaging modalities including optical coherence tomography[40], photoacoustic microscopy[41], two-photon microscopy[43], and laser speckle imaging[20,44]. These effects have also been validated in larger animal models such as rats[42], dogs[45], and marmosets[46]. In human, vasodilation and increased CBF caused by volatile anesthetics such as isoflurane have also been reported[22].

Statistical analysis from Fig. 4 shows that certain vessels exhibit a larger diameter under isoflurane anesthesia, and the vascularity, calculated as the percentage of vascular area within selected brain region ROIs, is also higher in the anesthetized state. These findings suggest a vasodilation effect induced by isoflurane, consistent with existing research[20,40,41,43,44]. It is worth noting that although our data indicate a global elevation of CBF under isoflurane anesthesia, individual vessels exhibit large discrepancies in behavior. For example, the vessel at the left lower corner from Mouse 3 in the entorhinal cortex (Fig. 3) shows almost no blood flow during anesthesia but then exhibits high vessel pixel density after awakening. The wide range of vessel behaviors were also previously reported in literature[20,46]. Our results indicate that awake ULM imaging has ample spatial resolution and imaging depth of penetration to resolve individual vessel variations down to micron-sized vessels deep into the brain. This is a unique capability that is not available from other biomedical imaging modalities.

Increased blood flow velocity induced by isoflurane has also been reported by other studies[40,41]. However, previous research presented different speculations on the predominant factor contributing to the increase in CBF induced by anesthesia, specifically whether the increase is attributed to vasodilation or increase in blood flow velocity[20]. One study found significant changes in both blood flow and vessel diameter but minor changes in flow velocity, suggesting that the increase in blood flow was largely driven by vasodilation[40]. Conversely, another study drew the opposite conclusion[20]. Benefitting from the large field of view (FOV) of ULM and its capability to directly quantify microvascular blood flow velocity, we can make a more comprehensive inference regarding the relationship among the changes in vessel diameter, flow speed, and flow volume from anesthetized to awake states. For arteries, the change in blood flow velocity is not significant, indicating that the alteration in blood flow may be primarily due to vasodilation instead of velocity change. Isoflurane causes vasodilation by acting on the ion channels (e.g., potassium channel) of smooth muscle[45], which is more abundantly found in arteries than in veins. In the case of veins, which do not actively dilate or constrict, their vessel diameter and blood flow variations are more likely controlled by passive mechanisms. Fig. 5 reveals significant differences in flow velocity of veins between anesthesia and awake state, suggesting that the changes in flow velocity may have a greater impact on venous blood flow volume compared with arterial volume.

The differences in cerebral vasculature between anesthetized and awake states observed using ULM are also in agreement with other studies[20,40,41,44]. However, previous studies mostly used optical imaging techniques, which have limited penetration depth and can only observe surface pial vessels in the cortex. Some other studies using fMRI can detect deeper CBF changes in the whole brain, but they do not provide

insights about small vessel blood flow variations due to insufficient spatial resolution[42]. As a bridging
imaging modality between MRI and optical techniques, awake ULM enables observations of detailed
microvascular variations induced by anesthesia across the whole depth of the brain, which provides
complementary information to existing biomedical imaging modalities.

Although isoflurane is widely used in ultrasound imaging because it provides long-lasting and stable
anesthetic effects, it is important to note that the vasodilation observed with isoflurane is not
representative of all anesthetics. Some anesthesia protocols, such as ketamine combined with
medetomidine, do not produce significant vasodilation and are therefore preferred in experiments where
vascular stability is essential, such as functional ultrasound imaging[47]. Therefore, in future studies, it would
be valuable to design more rigorous control experiments with larger sample sizes to systematically
compare the effects of isoflurane anesthesia, awake states, and other anesthetics that do not induce
vasodilation on cerebral blood flow.

Our proposed method enabled repeatable longitudinal brain imaging over a three-week period,
addressing a key limitation of conventional ULM imaging and offering potential for various preclinical
applications. However, there are still some limitations in this study.

One of the limitations is the lack of objective measures to assess the effectiveness of head-fix habituation
in reducing anxiety. This may introduce variability in stress levels among mice. Recent studies suggest that
tracking physiological parameters such as heart rate, respiratory rate, and corticosterone levels during
habituation can confirm that mice reach a low stress state prior to imaging[48]. This approach would be
highly beneficial for future awake imaging studies. Furthermore, alternative head-fixation setups, such as
air-floated balls or treadmills, which allow the free movement of limbs, have been shown to reduce anxiety
and facilitate natural behaviors during imaging[30]. Adopting these approaches in future studies could
enhance the reliability of awake imaging data by minimizing stress-related confounds.

Another limitation of this study is the potential residual vasodilatory effect of isoflurane anesthesia on
awake imaging sessions. The awake imaging sessions were conducted shortly after the mice had emerged
from isoflurane anesthesia, required for the MB bolus injections. The lasting vasodilatory effects of
isoflurane may have influenced vascular responses, potentially contributing to an underestimation of
differences in vascular dynamics between anesthetized and awake state. Future applications of awake
ULM in functional imaging using an indwelling jugular vein catheter presents a promising alternative to
enable more accurate functional imaging in awake animals, addressing current limitations associated with
anesthesia-induced vascular effects.

In our longitudinal study, consistent imaging results were obtained over a three-week period,
demonstrating the feasibility of awake ULM imaging for this duration. However, for certain research
applications, a monitoring period of several months would be valuable. Extending the duration of
longitudinal awake ULM imaging to enable such long-term studies is a potential direction for future
development.

Tissue motion is also a critical concern of ULM imaging. Benefiting from the robust head-fixing protocols
and successful habituation, we did not observe significant motion artifacts in the ULM images.
Nevertheless, we still employed the correlation method to register the ULM images[49] in order to remove
subtle tissue movement and improve the ULM image quality. This method preserves the maximum fidelity
for the MB tracks, but may not be adequate for non-rigid motion. While rigid motion correction is often

effective in anesthetized animals, awake animal imaging presents greater challenges due to the more
prominent non-rigid motion, particularly in deeper brain regions. This is evidenced in Supplementary Fig.
1 (Mouse 7), where cortical vessels remain relatively stable, but regions around the colliculi and
mesencephalon exhibit more noticeable motion artifacts, indicating that displacement is more
pronounced in deeper areas. To address these deeper, non-rigid motions, recent studies suggest
estimating non-rigid transformations from unfiltered tissue signals before applying corrections to ULM
vascular images[16,50]. Such advanced motion correction strategies may be more effective for awake ULM
imaging, which experiences higher motion variability. The development of more robust and effective
motion correction techniques will be crucial to reduce motion artifacts in future awake ULM applications.
Meanwhile, with 2D imaging, we cannot correct for out-of-plane motion, which necessitates 3D imaging.
In the future, 3D motion correction techniques that account for complex tissue motions and are
computationally efficient need to be developed for awake and longitudinal ULM imaging.

Based on the maximum linking distance and gap closing parameters outlined in the Methods section,
blood flow with velocities below 50 mm/s can be detected. However, the use of a directional filter to
estimate flow direction may introduce aliasing. MBs moving at higher velocities may be subject to incorrect
flow direction estimation due to aliasing effects. Given that the compounded frame rate is 1000 Hz, with
an ultrasound center frequency of 20 MHz and a sound speed of 1540 m/s, the relationship between
Doppler frequency and the axial blood flow velocity[12] indicates that aliasing will not occur for axial flow
velocities below 19.25 mm/s. In all flow velocity maps presented in this study, the range is limited to a
maximum of 15 mm/s, remaining below the critical threshold for aliasing. Additionally, all vessels analyzed
in the violin plots for arteriovenous flow comparisons fall within this range. While cortical arterioles and
venules generally exhibit moderate flow speeds, aliasing remains a factor to consider when combining
directional filtering with velocity analysis.

Advances in ULM imaging methods can benefit longitudinal awake imaging. For instance, dynamic ULM
can differentiate between arteries and veins by leveraging pulsatility features[51]. 3D ULM, with volumetric
imaging array[52,53], enables the reconstruction of whole-brain vascular network, providing a more
comprehensive understanding of vessel branching patterns. Meanwhile, 3D ULM also helps to mitigate
the challenge of aligning the identical coronal plane for longitudinal imaging, a process that requires
precise manual alignment in 2D ULM to ensure consistency. Additionally, this alignment issue can also be
alleviated in 2D imaging using backscattering amplitude method, which may assist in estimating out-of-
plane positioning during longitudinal imaging[54].

Longitudinal brain imaging in the awake state offers a promising tool for neuroscience research as it not
only avoids the confounding effects of anesthesia on cerebral vasculature, but also enables observations
of intrinsic dynamics of the vasculature within the same subject, minimizing potential sources of bias
associated with inter-subject variability. In the future, this technique is expected to be further integrated
with disease models to study the changes in cerebral vasculature during the development of diseases.
Also, this technique can be further combined with the latest functional ULM (fULM) studies[16] to allow
awake fULM imaging. Our study laid the foundation for these studies with awake fULM, which is expected
to improve the sensitivity of conventional fULM techniques because hemodynamic responses are much
stronger in the awake state than in anesthesia[24–26].

Methods

Animal preparation

Six healthy female C57 mice (8-12 weeks) were used for this study. All experimental procedures were conducted in accordance with the guidelines set by the University of Illinois Institutional Animal Care and Use Committee (IACUC Protocol number #22165). The animals were housed in an environment with a 12-hour light/dark cycle and had free access to food and water. Prior to the cranial window surgery, the mice were kept in group housing, and they were individually housed post-surgery.

Pre-surgery handling

A week before performing the cranial window surgery, the animals underwent tunnel handling, a procedure shown to significantly reduce the levels of anxiety in mice[55,56]. Specifically, a commercially available polycarbonate mouse transfer tube (TRANS-TUBE 130X50MM, Braintree Scientific, Inc.) was used for the initial three days. The mice were encouraged to enter the tube from their cage, after which the tube was lifted, allowing the mouse to remain inside for 30 seconds. The mouse was then returned to its cage for free movement for one minute, before a second identical tunnel handling was carried out. This procedure was done twice daily. Approximately three days into this routine, the mice were accustomed to the tunnel handling method. The animals were then picked up using the commercial tunnel, and then a 3D-printed body tube was attached to the other end of the tunnel. The mice were allowed to enter the body tube voluntarily. Subsequently, the body tube was utilized for handling the mice in the following 5 days as a replacement for the commercial tunnel.

Cranial window surgery

The entire procedure closely followed a previously published protocol[33]. To minimize brain swelling, the animal received an intraperitoneal injection of Dexamethasone (0.5 mg/kg body weight). The mouse was anesthetized by inhalation of isoflurane (3% for induction and 1%-1.5% for maintenance). After confirming that the mouse was anesthetized, the head of the animal was fixed in the stereotaxic frame. Afterward, the scalp of the mouse was incised, and the temporalis muscle was dissected from the skull using forceps. Tissue adhesive (3M, cat. no. 70200742529) was applied to secure the retracted muscles and incision edges, ensuring proper closure of the skin. Subsequently, a headpost was fixed to the skull using tissue adhesive and dental cement (Sun Medical Co., Ltd, cat. no. Super-Bond C&B). The skull was then carefully thinned using a high-speed rotary micromotor (Foredom, cat. no. K.1070) to create a cranial window with a width of approximately 8 mm laterally and which extending from bregma to lambda in the anterior-posterior direction. The skull was continuously thinned along this outline until it became loose, allowing for separation of the bone from the dura mater using forceps. After removing the skull, a polymethyl pentene plastic sheet (Goodfellow, cat. no. ME311051) was placed over the cranial window and fixed to the bone using tissue adhesive and dental cement. To provide protection, the window was covered with biocompatible silicone rubber (Smooth-on, cat. no. Body Double-Fast Set). Once the silicone rubber was secure, the anesthesia was discontinued, and the mouse was allowed to regain consciousness.

Post-surgical care and head-fix habituation

Following the surgery, subcutaneous administration of Carprofen (5-10 mg/kg body weight) was provided to the animals for immediate post-operative pain management. The animals were allowed to recover from anesthesia in their individual cages, with a heating pad provided to maintain optimal body temperature during the recovery period. Observations were made every 15 minutes until the animals reached sternal recumbency. To aid in their recovery, dry food soaked in water was provided, along with the use of recovery gel (ClearH2O, cat. no. 72-06-5022) to facilitate chewing and hydration. In the case that any signs of pain were observed, additional subcutaneous doses of 5-10 mg/kg Carprofen were administered every 12 to 24 hours to alleviate surgical discomfort for 3 days. Post-operative monitoring of the animals was performed daily for a duration of 14 days following the surgery. Once the animals had fully recovered from the surgical procedure, the animals were habituated daily to head fixation after walking through the body tube. The duration of the head-fixation periods was gradually increased over time, starting at approximately 10 minutes on day 1 and extending to up to 1 hour after 3 days[33]. If any signs of discomfort, such as excessive movement or vocalization, were observed during the head-fixation sessions then the sessions would be immediately discontinued.

The use of 3D-printed body tube

The body tube was secured to the table using four screws on the outer side, while two small screws on the top were used to clamp the headpost. Two semicircular grooves that fit the size of the small screws were made on the headpost (Fig. 1a). During the process of head-fixing the animal, the experimenter guided the animal to enter the body tube from the rear. Once the animal protruded its head from the front end, the experimenter manually grasped the headpost, gently restrained the head of the animal, and subsequently aligned the semicircular grooves with the screws to immobilize the animal.

Experimental procedure of imaging sessions

On the day of imaging, the animal was guided to walk into the body tube, and the headpost was firmly secured to the tube. The protective silicone rubber and headpost were cleaned using 70% ethanol, followed by rinsing with sterile saline. Then, forceps were inserted between the silicone rubber and the cement to detach the silicone rubber from the headpost. Due to the challenges of conducting continuous infusion in awake animals, contrast microbubbles were administered via tail vein bolus injection in this study. Animals were anesthetized with isoflurane (3% for induction and 1% for maintenance) before tail vein catheterization to alleviate pain and stress. Ultrasound coupling gel was applied, and the transducer was positioned to find the imaging plane. When the imaging plane was identified, the relative position of the probe and the headpost was recorded, serving as a reference point for subsequent longitudinal studies. B-mode and power Doppler images were also saved as references to facilitate the identification of the same imaging plane in the following weeks.

Once the imaging plane was confirmed, the isoflurane anesthesia was terminated, and 0.1 ml of DEFINITY® (Lantheus, North Billerica, MA) microbubble was administered via the tail vein catheter. The completion of the injection was considered as the starting point, denoted as T = 0. From T = 0 onwards, continuous ultrasound data was saved. After completion of the imaging session, the cranial window was filled with the same biocompatible silicone rubber described above, allowing approximately 10 minutes for

solidification. The headpost was then removed from the body tube, and the animal was returned to its
home cage. Imaging sessions were conducted once per week over a three-week period.

For the experiment of comparing differences between anesthesia and awake state (Fig. 3-5), once the mice
were fixed to the body tube, they were continuously anesthetized with 2% isoflurane for more than 15
minutes. During this process, the tail vein catheterization was completed, and the imaging plane was
confirmed. Following the bolus injection of 0.1 ml MB, ULM data under anesthesia was acquired for 1000
seconds. The tail vein catheter was remained in place. After the ULM acquisition under anesthesia, it was
essential to wait until no bubbles were circulating within the blood stream before terminating anesthesia.
Once the animal approached full arousal, another bolus injection of 0.1 ml MB was administered. The
catheter was then removed, and ULM acquisition started (marked as T = 0). Other procedures were the
same as those described in the previous two paragraphs. For the ULM data collected under anesthesia,
since there was no specific time point at which mice achieved full pupillary dilatation, we consistently
employed the RF dataset starting from T = 500s until the cumulative MB count reached 5 million for ULM
reconstruction to make the MB concentration similar as in the awake cases.

Pupillary recording and measurement

Monitoring the pupillary area of rodents is commonly used to determine their level of arousal[57]. The pupils
constrict under isoflurane anesthesia and enlarge upon awakening[38,58]. In this study, video recording of
pupillary area performed using a camera (ace acA800-510um, Basler Inc., Exton, PA) placed in front of the
mouse, with timestamps synchronized to the ultrasound acquisition. The ULM imaging session was carried
out in a dark room. An infrared (IR) flashlight (EVOLVA Future Technology T20) was positioned above the
camera. Since mice are functionally blind to IR wavelengths[59,60], the use of IR illumination can minimize
the effect of light exposure on pupil changes so that the pupillary area is mostly influenced by anesthesia.
The videos recorded by the camera were analyzed using ImageJ software, and the contour of the pupil
was manually circled. The equivalent area of a single pixel was calibrated, which allowed for the
quantification of the pupillary area according to the number of pixels occupied by the pupil.

Ultrasound imaging sequence

A Vantage 256 system (Verasonics Inc., Kirkland, WA) was used for this experiment. The ultrasound system
was connected to a linear-array transducer (L35-16vX, Verasonics Inc.), which was fixed onto a 3D-printed
holder attached to a translation motor (VT-80, Physik Instrumente, Auburn, MA). The motor allowed
precise control of transducer movement in the elevational direction, enabling adjustment of the imaged
coronal plane. Ultrasound was transmitted at a center frequency of 20 MHz. A 9-angle compounded plane
wave technique was used (angles: -8°, -6°, -4°, -2°, 0°, 2°, 4°, 6°, and 8°), with a post-compounded frame
rate of 1000 Hz. Interleaved sampling is employed to capture high-frequency echoes more effectively. With
the system's sampling rate limited to 62.5 MHz, the upper limit of the center frequency of the transducer
passband is 15.625 MHz. To mitigate aliasing, two transmissions are sent per angle, staggered in time. This
approach effectively doubles the sampling rate, ensuring more accurate image reconstruction. The
ensemble size of each dataset was 800 frames, and the acquired radiofrequency (RF) data was saved for
offline reconstruction. Beamforming was conducted using the Verasonics built-in program to reconstruct
the in-phase and quadrature (IQ) data. The IQ data had a pixel size of half the wavelength in the axial
direction and one wavelength in the lateral direction.

ULM image processing

To ensure consistent localization of bubbles at different depths, a noise equalization method was employed to adjust signal intensity[61]. High pass filtering with a cutoff frequency of 30Hz was applied to the IQ data to enhance sensitivity for MB extraction. MBs moving toward and away from the transducer were separated into two distinct datasets (upward flow and downward flow) based on the positive and negative Doppler shifts[62]. Subsequently, singular value decomposition (SVD) filtering was applied to the IQ data to further eliminate clutter signals from tissue and extract MB signals[63,64]. The singular value cutoff was adaptively determined to achieve objective and stable filtering results[65]. The filtered MB data was then spline interpolated to a resolution with a lateral and axial pixel size of one-tenth of the wavelength (4.928 µm). Each interpolated frame was subjected to 2D normalized cross-correlation with the point spread function of the imaging system, which was empirically determined. Regional maxima in the cross-correlation results indicated the center position of MBs[49]. Thresholding of image intensity was applied to reject low-intensity values and prevent noise from being erroneously identified as bubbles. The coordinates of the microbubble center points obtained from each frame were tracked using the uTrack algorithm[66,67]. To ensure the detection of reliable MB tracks, only tracks with a minimum length of 10 consecutive frames were considered valid. The choice of 10 consecutive frames (10 ms) was based on established practice[11] but can be adjusted as needed[68]. For the uTrack algorithm, two additional key parameters were specified: the maximum linking distance and the gap-filling distance, both set to 10 pixels (~50 microns). This configuration means that only bubble centroids within 10 pixels of each other across consecutive frames are considered part of the same bubble trajectory. Additionally, when the start and end points of two tracks fall within this threshold, the gap-filling parameter merges them into a single, continuous track. It is important to select these parameters carefully, as overly large values could lead to an overestimation of flow velocity. By setting the maximum linking distance to 10 pixels, we effectively limited the measurable velocity to 50 mm/s, under the assumption that no bubble would exceed a 50-micron displacement within the 1 ms interval between frames. Considering the velocity distribution across the mouse brain[69], this 50 mm/s limit ensures that most of the blood flow is captured accurately. After determining bubble tracks with the specified parameters for uTrack algorithm, accumulating the MB tracks resulted in the flow intensity map. The final values were square-root transformed to compress the dynamic range of the image for display. Furthermore, the distance traveled by MB between adjacent frames was calculated during the tracking process to determine blood flow velocity, which was then assigned to each individual bubble track. The average velocity computed from multiple tracks was used to generate a comprehensive flow speed map.

Quantitative analysis of ULM images

For single-vessel analysis, the vessel density map (either bidirectional or unidirectional) was binarized, and the center line of the vessel was extracted using MATLAB "bwskel" function. The radius of each vessel was defined as the distance from the central line to the nearest vessel edge on the binarized map, with diameter calculated by doubling this radius. Given a pixel size of approximately 4.928 µm, the minimum detectable vessel diameter using this approach is 9.856 µm, explaining why the smallest measured radius in unsaturated conditions, as shown in Fig. 2b, always falls into this value. To analyze a specific vessel, we manually selected a vessel segment. Each point on the central line provided a diameter measurement. For flow velocity, each non-zero pixel within the chosen segment was used to estimate effective blood flow velocity in the vessel.

For ROI-based analysis, ROIs were manually defined based on a brain atlas to ensure consistency across both anesthetized and awake states, as well as across weekly imaging sessions, enhancing comparability of results. In ROI-based analysis, we focused on two primary parameters: vascularity and mean velocity. Vascularity was defined as the proportion of the pixel count occupied by blood vessels within each ROI, obtained by binarizing the ULM vessel density maps and calculating the percentage of the pixels with MB signal. Mean velocity was calculated by averaging all non-zero pixel of velocity estimates within the ROI. The velocity distribution within each ROI was also visualized using violin plots, as shown in Fig. 2, 4 and 6, to illustrate the range and density of flow velocity estimates across different acquisition.

Acknowledgements

This study was partially supported by the National Institute on Aging, the National Institute of Biomedical Imaging and Bioengineering, and the National Institute of Neurological Disorders and Stroke of the National Institutes of Health under grant numbers R21EB030072, R21EB030072-01S1, R21AG077173, R56NS131516, and by the National Science Foundation CAREER Award 2237166. The content is solely the responsibility of the authors and does not necessarily represent the official views of the NIH and NSF. MRL was supported by a Beckman Institute Postdoctoral Fellowship. We thank Dr. Baher Ibrahim and Mr. Gang Xiao from Dr. Daniel Llano's lab for their assistance with the awake imaging setup. We thank Dr. Danqing Hu from Emory University for her insights regarding the impact of anesthetics on cerebral blood flow.

Reference

1. Misgeld, T. & Kerschensteiner, M. In vivo imaging of the diseased nervous system. *Nat. Rev. Neurosci.* **7**, 449–463 (2006).
2. Ji, N., Freeman, J. & Smith, S. L. Technologies for imaging neural activity in large volumes. *Nat. Neurosci.* **19**, 1154–1164 (2016).
3. Grandjean, J. *et al.* Common functional networks in the mouse brain revealed by multi-centre resting-state fMRI analysis. *NeuroImage* **205**, 116278 (2020).
4. Gutierrez-Barragan, D. *et al.* Unique spatiotemporal fMRI dynamics in the awake mouse brain. *Curr. Biol.* **32**, 631-644.e6 (2022).
5. Toyama, H. *et al.* PET imaging of brain with the β-amyloid probe, [11C]6-OH-BTA-1, in a transgenic mouse model of Alzheimer's disease. *Eur. J. Nucl. Med. Mol. Imaging* **32**, 593–600 (2005).
6. Miranda, A. *et al.* Awake [18] F-FDG PET Imaging of Memantine-Induced Brain Activation and Test–Retest in Freely Running Mice. *J. Nucl. Med.* **60**, 844–850 (2019).
7. Ma, Y. *et al.* Wide-field optical mapping of neural activity and brain haemodynamics: considerations and novel approaches. *Philos. Trans. R. Soc. B Biol. Sci.* **371**, 20150360 (2016).
8. Zong, W. *et al.* Miniature two-photon microscopy for enlarged field-of-view, multi-plane and long-term brain imaging. *Nat. Methods* **18**, 46–49 (2021).
9. Yao, J. *et al.* High-speed label-free functional photoacoustic microscopy of mouse brain in action. *Nat. Methods* **12**, 407–410 (2015).
10. Pang, Z. *et al.* Multi-modality photoacoustic/ultrasound imaging based on a commercial ultrasound platform. *Opt. Lett.* **46**, 4382 (2021).
11. Errico, C. *et al.* Ultrafast ultrasound localization microscopy for deep super-resolution vascular imaging. *Nature* **527**, 499–502 (2015).
12. Macé, E. *et al.* Functional ultrasound imaging of the brain. *Nat. Methods* **8**, 662–664 (2011).

13. Mace, E. *et al.* Functional ultrasound imaging of the brain: theory and basic principles. *IEEE Trans.*
Ultrason. Ferroelectr. Freq. Control **60**, 492–506 (2013).

14. Deffieux, T., Demené, C. & Tanter, M. Functional Ultrasound Imaging: A New Imaging Modality for
Neuroscience. *Neuroscience* **474**, 110–121 (2021).

15. Edelman, B. J. & Macé, E. Functional ultrasound brain imaging: Bridging networks, neurons, and
behavior. *Curr. Opin. Biomed. Eng.* **18**, 100286 (2021).

16. Renaudin, N. *et al.* Functional ultrasound localization microscopy reveals brain-wide neurovascular
activity on a microscopic scale. *Nat. Methods* **19**, 1004–1012 (2022).

17. Lowerison, M. R. *et al.* Aging-related cerebral microvascular changes visualized using ultrasound
localization microscopy in the living mouse. *Sci. Rep.* **12**, 619 (2022).

18. Chavignon, A., Hingot, V., Orset, C., Vivien, D. & Couture, O. 3D transcranial ultrasound localization
microscopy for discrimination between ischemic and hemorrhagic stroke in early phase. *Sci. Rep.*
12, 14607 (2022).

19. Demené, C. *et al.* Transcranial ultrafast ultrasound localization microscopy of brain vasculature in
patients. *Nat. Biomed. Eng.* **5**, 219–228 (2021).

20. Sullender, C. T., Richards, L. M., He, F., Luan, L. & Dunn, A. K. Dynamics of isoflurane-induced
vasodilation and blood flow of cerebral vasculature revealed by multi-exposure speckle imaging. *J.*
Neurosci. Methods **366**, 109434 (2022).

21. Masamoto, K. & Kanno, I. Anesthesia and the Quantitative Evaluation of Neurovascular Coupling. *J.*
Cereb. Blood Flow Metab. **32**, 1233–1247 (2012).

22. Slupe, A. M. & Kirsch, J. R. Effects of anesthesia on cerebral blood flow, metabolism, and
neuroprotection. *J. Cereb. Blood Flow Metab.* **38**, 2192–2208 (2018).

23. Reimann, H. M. & Niendorf, T. The (Un)Conscious Mouse as a Model for Human Brain Functions: Key
Principles of Anesthesia and Their Impact on Translational Neuroimaging. *Front. Syst. Neurosci.* **14**, 8
(2020).

24. Aksenov, D. P., Li, L., Miller, M. J., Iordanescu, G. & Wyrwicz, A. M. Effects of Anesthesia on BOLD
Signal and Neuronal Activity in the Somatosensory Cortex. *J. Cereb. Blood Flow Metab.* **35**, 1819–
1826 (2015).

25. Pisauro, M. A., Dhruv, N. T., Carandini, M. & Benucci, A. Fast Hemodynamic Responses in the Visual
Cortex of the Awake Mouse. *J. Neurosci.* **33**, 18343–18351 (2013).

26. Desai, M. *et al.* Mapping brain networks in awake mice using combined optical neural control and
fMRI. *J. Neurophysiol.* **105**, 1393–1405 (2011).

27. Flusberg, B. A. *et al.* High-speed, miniaturized fluorescence microscopy in freely moving mice. *Nat.*
Methods **5**, 935–938 (2008).

28. Guo, H., Chen, Q., Qin, W., Qi, W. & Xi, L. Detachable head-mounted photoacoustic microscope in
freely moving mice. *Opt. Lett.* **46**, 6055 (2021).

29. Tang, J. *et al.* Noninvasive High-Speed Photoacoustic Tomography of Cerebral Hemodynamics in
Awake-Moving Rats. *J. Cereb. Blood Flow Metab.* **35**, 1224–1232 (2015).

30. Bertolo, A. *et al.* Whole-Brain 3D Activation and Functional Connectivity Mapping in Mice using
Transcranial Functional Ultrasound Imaging. *J. Vis. Exp.* 62267 (2021) doi:10.3791/62267.

31. Macé, É. *et al.* Whole-Brain Functional Ultrasound Imaging Reveals Brain Modules for Visuomotor
Integration. *Neuron* **100**, 1241-1251.e7 (2018).

32. Sans-Dublanc, A. *et al.* Optogenetic fUSI for brain-wide mapping of neural activity mediating
collicular-dependent behaviors. *Neuron* **109**, 1888-1905.e10 (2021).

33. Brunner, C. *et al.* Whole-brain functional ultrasound imaging in awake head-fixed mice. *Nat. Protoc.*
16, 3547–3571 (2021).

34. Beckmann, L. *et al.* Longitudinal deep-brain imaging in mouse using visible-light optical coherence
tomography through chronic microprism cranial window. *Biomed. Opt. Express* **10**, 5235 (2019).

35. Wang, Y., Liang, G., Liu, F., Chen, Q. & Xi, L. A Long-Term Cranial Window for High-Resolution Photoacoustic Imaging. *IEEE Trans. Biomed. Eng.* **68**, 706–711 (2021).

36. Tournissac, M. *et al.* Cranial window for longitudinal and multimodal imaging of the whole mouse cortex. *Neurophotonics* **9**, (2022).

37. Sieu, L.-A. *et al.* EEG and functional ultrasound imaging in mobile rats. *Nat. Methods* **12**, 831–834 (2015).

38. Turner, K. L., Gheres, K. W. & Drew, P. J. Relating Pupil Diameter and Blinking to Cortical Activity and Hemodynamics across Arousal States. *J. Neurosci.* **43**, 949–964 (2023).

39. Hingot, V. *et al.* Microvascular flow dictates the compromise between spatial resolution and acquisition time in Ultrasound Localization Microscopy. *Sci. Rep.* **9**, 2456 (2019).

40. Rakymzhan, A., Li, Y., Tang, P. & Wang, R. K. Differences in cerebral blood vasculature and flow in awake and anesthetized mouse cortex revealed by quantitative optical coherence tomography angiography. *J. Neurosci. Methods* **353**, 109094 (2021).

41. Cao, R. *et al.* Functional and oxygen-metabolic photoacoustic microscopy of the awake mouse brain. *NeuroImage* **150**, 77–87 (2017).

42. Sicard, K. *et al.* Regional Cerebral Blood Flow and BOLD Responses in Conscious and Anesthetized Rats under Basal and Hypercapnic Conditions: Implications for Functional MRI Studies. *J. Cereb. Blood Flow Metab.* **23**, 472–481 (2003).

43. Lyons, D. G., Parpaleix, A., Roche, M. & Charpak, S. Mapping oxygen concentration in the awake mouse brain. *eLife* **5**, e12024 (2016).

44. Takuwa, H. *et al.* Hemodynamic changes during somatosensory stimulation in awake and isoflurane-anesthetized mice measured by laser-Doppler flowmetry. *Brain Res.* **1472**, 107–112 (2012).

45. Iida, H., Ohata, H., Iida, M., Watanabe, Y. & Dohi, S. Isoflurane and Sevoflurane Induce Vasodilation of Cerebral Vessels via ATP-sensitive K+Channel Activation. *Anesthesiology* **89**, 954–960 (1998).

46. Santisakultarm, T. P. *et al.* Two-photon imaging of cerebral hemodynamics and neural activity in awake and anesthetized marmosets. *J. Neurosci. Methods* **271**, 55–64 (2016).

47. Réaux-Le-Goazigo, A. *et al.* Ultrasound localization microscopy and functional ultrasound imaging reveal atypical features of the trigeminal ganglion vasculature. *Commun. Biol.* **5**, 330 (2022).

48. Chabouh, G. *et al.* An open-source platform for 3D transcranial Ultrasound Localization Microscopy in awake mice. Preprint at https://doi.org/10.21203/rs.3.rs-4925712/v1 (2024).

49. Song, P. *et al.* Improved Super-Resolution Ultrasound Microvessel Imaging With Spatiotemporal Nonlocal Means Filtering and Bipartite Graph-Based Microbubble Tracking. *IEEE Trans. Ultrason. Ferroelectr. Freq. Control* **65**, 149–167 (2018).

50. Hingot, V., Errico, C., Tanter, M. & Couture, O. Subwavelength motion-correction for ultrafast ultrasound localization microscopy. *Ultrasonics* **77**, 17–21 (2017).

51. Bourquin, C., Poree, J., Lesage, F. & Provost, J. In Vivo Pulsatility Measurement of Cerebral Microcirculation in Rodents Using Dynamic Ultrasound Localization Microscopy. *IEEE Trans. Med. Imaging* **41**, 782–792 (2022).

52. McCall, J. R., Santibanez, F., Belgharbi, H., Pinton, G. F. & Dayton, P. A. Non-invasive transcranial volumetric ultrasound localization microscopy of the rat brain with continuous, high volume-rate acquisition. *Theranostics* **13**, 1235–1246 (2023).

53. Heiles, B. *et al.* Ultrafast 3D Ultrasound Localization Microscopy Using a 32 × 32 Matrix Array. *IEEE Trans. Med. Imaging* **38**, 2005–2015 (2019).

54. Renaudin, N., Pezet, S., Ialy-Radio, N., Demene, C. & Tanter, M. Backscattering amplitude in ultrasound localization microscopy. *Sci. Rep.* **13**, 11477 (2023).

55. Hurst, J. L. & West, R. S. Taming anxiety in laboratory mice. *Nat. Methods* **7**, 825–826 (2010).

56. Gouveia, K. & Hurst, J. L. Optimising reliability of mouse performance in behavioural testing: the major role of non-aversive handling. *Sci. Rep.* **7**, 44999 (2017).

57. Privitera, M. *et al.* A complete pupillometry toolbox for real-time monitoring of locus coeruleus
activity in rodents. *Nat. Protoc.* **15**, 2301–2320 (2020).

58. Kum, J. E., Han, H.-B. & Choi, J. H. Pupil Size in Relation to Cortical States during Isoflurane
Anesthesia. *Exp. Neurobiol.* **25**, 86–92 (2016).

59. Breuninger, T., Puller, C., Haverkamp, S. & Euler, T. Chromatic Bipolar Cell Pathways in the Mouse
Retina. *J. Neurosci.* **31**, 6504–6517 (2011).

60. Chang, L., Breuninger, T. & Euler, T. Chromatic Coding from Cone-type Unselective Circuits in the
Mouse Retina. *Neuron* **77**, 559–571 (2013).

61. Song, P., Manduca, A., Trzasko, J. D. & Chen, S. Noise Equalization for Ultrafast Plane Wave
Microvessel Imaging. *IEEE Trans. Ultrason. Ferroelectr. Freq. Control* **64**, 1776–1781 (2017).

62. Huang, C. *et al.* Short Acquisition Time Super-Resolution Ultrasound Microvessel Imaging via
Microbubble Separation. *Sci. Rep.* **10**, 6007 (2020).

63. Demene, C. *et al.* Spatiotemporal Clutter Filtering of Ultrafast Ultrasound Data Highly Increases
Doppler and fUltrasound Sensitivity. *IEEE Trans. Med. Imaging* **34**, 2271–2285 (2015).

64. Baranger, J. *et al.* Adaptive Spatiotemporal SVD Clutter Filtering for Ultrafast Doppler Imaging Using
Similarity of Spatial Singular Vectors. *IEEE Trans. Med. Imaging* **37**, 1574–1586 (2018).

65. Song, P., Manduca, A., Trzasko, J. D. & Chen, S. Ultrasound Small Vessel Imaging With Block-Wise
Adaptive Local Clutter Filtering. *IEEE Trans. Med. Imaging* **36**, 251–262 (2017).

66. Jaqaman, K. *et al.* Robust single-particle tracking in live-cell time-lapse sequences. *Nat. Methods* **5**,
695–702 (2008).

67. Stein, S. C. & Thiart, J. TrackNTrace: A simple and extendable open-source framework for developing
single-molecule localization and tracking algorithms. *Sci. Rep.* **6**, 37947 (2016).

68. Couture, O., Hingot, V., Heiles, B., Muleki-Seya, P. & Tanter, M. Ultrasound Localization Microscopy
and Super-Resolution: A State of the Art. *IEEE Trans. Ultrason. Ferroelectr. Freq. Control* **65**, 1304–
1320 (2018).

69. Demeulenaere, O. *et al.* In vivo whole brain microvascular imaging in mice using transcranial 3D
Ultrasound Localization Microscopy. *eBioMedicine* **79**, 103995 (2022).

708
709

Animal ID	Use of the animal	Corresponding figures
Mouse1	• Comparison between anesthetized and awake imaging	Figure 3 Figure 4
Mouse2	• Comparison between anesthetized and awake imaging • Showcase of surgical defect or problematic ULM image	Figure 3 Figure 4 Supplementary figure 1
Mouse3	• Comparison between anesthetized and awake imaging	Figure 3 Figure 4
Mouse4	• Longitudinal imaging across 3 weeks • Example of demonstrating system schematic and image saturation criteria	Figure 1 Figure 2 Figure 5 Figure 6 Supplementary figure 2
Mouse5	• Longitudinal imaging across 3 weeks • Example of demonstrating image saturation criteria	Figure 5 Figure 6
Mouse6	• Longitudinal imaging across 3 weeks • Example of demonstrating image saturation criteria • Showcase of surgical defect or problematic ULM image	Figure 5 Figure 6 Supplementary figure 1 Supplementary figure 2
Mouse7	• Showcase of surgical defect or problematic ULM image	Supplementary figure 1
Mouse8	• Showcase of surgical defect or problematic ULM image	Supplementary figure 1

Mouse 7: Cortical bulge after surgical damage   **Mouse 8: Loss of perfusion after cortical damage**   **Mouse 2: Inflammation on the surface of the cortex**   **Mouse 6: Signal attenuation from air bubble**

Supplementary Fig. 1 | Overview of all experimental mice and examples of poor imaging outcomes. The table details the usage and locations where each experimental mouse is referenced in this study. Not all mice yielded optimal imaging results. Mice with poor imaging outcomes are included as examples of failed cases.

Supplementary Fig. 2 | Perfusion under the 5 million saturation standard for various small vessels. To ensure the findings in Figure 2b are not specific to a single vessel, vascular segments were selected from three different brain regions in three different mice. Comparisons of vessel diameter and blood flow velocity

713
714

Supplementary Fig 3 | Comparison of ULM images in isoflurane anesthetized and awake states (Mouse 1 to Mouse 3), with distinctions made between upward and downward flow.

Supplementary Fig 4 | Longitudinal comparison of Mouse 4 to Mouse 6 with distinctions made between upward and downward flow.

718