Motion Blur Microscopy

Utku Goreke[1], **Ayesha Gonzales**[2], **Brandon Shipley**[2], **Yuncheng Man**[1], **William Wulftange**[1], **Ran An**[1], **Michael Hinczewski**[2,*], **and Umut A. Gurkan**[1,3,†]

[1]Department of Mechanical and Aerospace Engineering, Case Western Reserve University, Cleveland, OH
[2]Department of Physics, Case Western Reserve University, Cleveland, OH
[3]Department of Biomedical Engineering, Case Western Reserve University, Cleveland, OH
[*]michael.hinczewski@case.edu
[†]umut@case.edu

ABSTRACT

Imaging and characterizing the dynamics of cellular adhesion in blood samples is of fundamental importance in understanding biological function. *In vitro* microscopy methods are widely used for this task, but typically require diluting the blood with a buffer to allow for transmission of light. However whole blood provides crucial mechanical and chemical signaling cues that influence adhesion dynamics, which means that conventional approaches lack the full physiological complexity of living microvasculature. We propose to overcome this challenge by a new *in vitro* imaging method which we call motion blur microscopy (MBM). By decreasing the source light intensity and increasing the integration time during imaging, flowing cells are blurred, allowing us to identify adhered cells. Combined with an automated analysis using machine learning, we can for the first time reliably image cell interactions in microfluidic channels during whole blood flow. MBM provides a low cost, easy to implement alternative to intravital microscopy, the *in vivo* approach for studying how the whole blood environment shapes adhesion dynamics. We demonstrate the method's reproducibility and accuracy in two example systems where understanding cell interactions, adhesion, and motility is crucial—sickle red blood cells adhering to laminin, and CAR-T cells adhering to E-selectin. We illustrate the wide range of data types that can be extracted from this approach, including distributions of cell size and eccentricity, adhesion times, trajectories and velocities of adhered cells moving on a functionalized surface, as well as correlations among these different features at the single cell level. In all cases MBM allows for rapid collection and processing of large data sets, ranging from thousands to hundreds of thousands of individual adhesion events. The method is generalizable to study adhesion mechanisms in a variety of diseases, including cancer, blood disorders, thrombosis, inflammatory and autoimmune diseases, as well as providing rich datasets for theoretical modeling of adhesion dynamics.

Main

Cellular interactions, including cell adhesion, migration, and chemotaxis, are important to investigate the mechanisms of diseases including cancer, thrombosis, inflammatory diseases, anemia, and vasculopathy. *In vitro* cellular imaging techniques for hematology generally require the use of an aqueous buffer [1–3], which dilutes the sample and allows transmission of light for imaging.[4] However, buffer solutions replace the original whole blood medium, potentially affecting the biological mechanisms under investigation. For instance, plasma proteins facilitate the interaction of red blood cells with endothelial cells, and red blood cells induce margination of leukocytes and platelets to the vascular wall.[5,6] Comprehensively understanding these phenomena requires a physiologically realistic approach that includes the presence of whole blood. Therefore, intravital methods remain the gold standard for studies of important dynamic processes associated with cellular interactions.[7]

Intravital methods include multi- and single-photon microscopy, confocal microscopy, Brillouin spectroscopy combined with light microscopy, lightsheet microscopy, and endomicroscopy.[8–13] However, intravital microscopy methods are highly costly, and require intensive effort in both setup and analysis, and therefore have limited applicability for the broader research community.[14,15] Total internal reflection fluorescence microscopy is a potential *in vitro* alternative for visualizing cellular interactions under whole blood flow (and examples with buffer flow already exist[16,17]). But fluorophore labeling in whole blood can be challenging. Finally, laser optical imaging has been used for obtaining the number of platelet interactions that occur with a protein substrate in a microfluidic channel,[18] but this method is limited to a single numeric output (intensity of light scattered over time), and has additional experimental setup complexity.

Here, we describe a practical, accessible, and easily adaptable microscopy method that enables real-time

imaging of dynamics of cellular interactions under whole blood flow *in vitro* by eliminating the need for blood sample dilution completely. We call our approach *motion blur microscopy* (MBM). MBM leverages blurring to make the cellular interactions that take place at slower velocity scales discernable (Fig. 1). Individual cells with a velocity substantially less than the bulk flow (i.e. immobile, adhered cells, or those that are rolling/migrating while in contact with the surface) can be visualized within whole blood flow. MBM allows *in vitro* analysis of various static and dynamic properties of cellular interactions, all while mimicking key *in vivo* conditions.

MBM works by reducing the light source and increasing the exposure time, resulting in streaks of flowing cells that generate noisy images. Thus, to identify and analyze adhered cells, an experimenter must spend a considerable amount of time and effort to distinguish cells. While this may not be a difficult task when analyzing individual MBM images, manually analyzing dynamic interactions from MBM videos consisting of hundreds or thousands of frames can be impractical and error-prone. Therefore, we developed an automated machine-learning-based analysis, which can efficiently characterize the dynamics of cellular interactions in MBM videos. The automated analysis is a two-phase system, where phase one identifies groups of pixels in an MBM image corresponding to adhered cells, and phase two classifies these groups by cell type (Fig 2). The phase one task is completed using a machine learning segmentation network. The phase two task is completed by classifying cells by their size, or by using a machine learning classification network, depending on the complexity of the system under study.

With our automated analysis, we can process MBM images in an accurate and high throughput manner. Adhered cells can be identified, and morphological features, such as the size and eccentricity of each individual cell can be extracted. MBM videos enable studying the dynamics of cells on the individual level, and determine kinetic properties like adhesion times, and average velocities (for the case of cells that roll or migrate while on the surface). The individual cell data can be aggregated to produce group statistics, such as distributions of morphological features and dynamic quantities, or mean squared displacements. Importantly, the method allows to identify and analyze the properties of hundreds of thousands of cells. These data can be used to help us understand cellular adhesion dynamics, or be used in clinical studies. Since our approach relies on a basic experimental microscopy setup, we anticipate that it will be highly accessible to the broader research community.

Characterizing fundamental cellular interactions on the single-cell level can reveal small sub-populations of cells that initiate pathogenesis. Here, we focus on two diseases where the cellular interactions in the vascular space carry central importance. Analyzing these sub-populations typically requires intravital microscopy, and thus, serves as a means to demonstrate the effectiveness of MBM in challenging applications. The first of these diseases is sickle cell disease, where hemoglobin, the fundamental protein underlying red blood cells' oxygen transport, polymerizes in a deoxygenated environment. When the hemoglobin polymerizes, a cascade of events can occur, leading to a debilitating disease complication known as a vaso-occlusive crisis. Red blood cells containing polymerized sickle hemoglobin experience physical and chemical changes, resulting in abnormally stiff, dense, and adhesive cells[19–21]. The second disease in focus is malignant solid tumors and T-cell migration is crucial for immunotherapy of this type of malignancy. Many chimeric antigen receptor T-cell (CAR-T) cell therapies are under development for solid tumors, and they hold great promise for treating refractory cancers. Understanding the migration of T cells, specifically CAR-T cells, would lead to better design of future cell therapies when diapedesis/migration is an important aspect of the therapy, as in the case of solid tumors. An accessible, reliable, single-cell level *in vitro* method for analyzing CAR-T cell behavior, like MBM, would accelerate overcoming this bottleneck.

To demonstrate MBM and put its validity to the test, three distinct experimental setups were used. One experimental setup used a laminin functionalized channel to observe sickle red blood cells under flow. A second experimental set-up used an E-selectin functionalized channel to observe CAR-T cells under flow. The final experimental set-up used a P-selectin functionalized channel to observe a combination of red blood cells and CAR-T cells under flow. We include this final set-up as an additional demonstration of the flexibility of MBM, and the details of its analysis are in the Supplementary Information (SI). In this work, we aim to accomplish three main goals. One, we show that MBM with automated analysis is accurate, insofar as it can correctly classify cells for a variety of inputs in a reproducible manner. Two, we show that relevant physical properties of identified cells can be found. Three, we show that the identified cells and their properties constitute a data set with practical uses, ranging from studying cellular

a Moving objects obstruct the background

b Long Exposure

c Motion Blur Microscopy

Flowing cells become blurry
Adhered cells become discernable

25 μm

(adjusted contrast)

| 0.8 ms Exposure | 30 ms Exposure | 1200 ms Exposure |
| 12V (100%) Bulb Voltage | 9V Bulb Voltage | 2.7V Bulb Voltage |

d Reduced light

long exposure

Figure 1. Motion blur microscopy (MBM). a) Moving objects in the foreground obstruct the stationary objects in the background. **b)** By adjusting the exposure, moving objects are blurred to obtain a clear view of the background. **c)** The same principle is applied to microscale whole blood flow. Shown are three views of whole blood flow in the same microscopic field with different microscope settings. Increasing the camera exposure (integration time) helps blur the foreground, which consists of non-adhered (flowing) cells. Long exposure results in excess brightness which is then compensated by a reduction of the light source voltage to obtain a clear view of the adhered cells in the background. Yellow arrows show the flow direction. The shear rate of the flow is 50 s^{-1} which is enough to induce the blur at 1200 ms integration time. **d)** Schematic illustration of the microfluidic channel, with clearly visible adhered cells and blurry flowing cells.

adhesion mechanics to aiding in clinical studies.

Overview of MBM with automated analysis

Here we summarize the complete process for analyzing blood samples using MBM (full details can be found in the Methods and SI). Given a cell type (or types) of interest, we functionalize a microchannel with suitable adhesion proteins. Blood samples sent through the microchannels are then imaged to produce either individual MBM images or videos. Cells of interest adhere to the functionalized protein surfaces, and become visible against the blurred foreground of non-adhered flowing cells. The MBM images are analyzed via a two-stage automated process schematically illustrated in Fig. 2. The first phase consists of a segmentation neural network, which labels pixels from MBM images corresponding to adhered objects, and groups together neighboring labeled pixels. In phase two of the analysis, groups of labeled pixels are then classified by cell type using either a size threshold (in applications where size is a sufficient criterion) or a specially trained classification

neural network (in more complicated scenarios). After all of the cells in an MBM image or MBM video have been identified, we then generate data for each identified cell. Morphological properties like the size and eccentricity of each cell can be extracted, as well as dynamical properties such as adhesion time or mean velocity. Using the automated analysis, we were able to extract properties of individual cells for 175 sickle red blood cell / laminin images, one CAR-T / P-selectin video with 999 frames, and one CAR-T / E-selectin video with 499 frames, all together containing hundreds of thousands of adhesion events.

MBM-based cell classification is reproducible and accurate

In cases where only a single cell type is expected to be visible in an MBM image, we demonstrate that we can distinguish groups of adhered pixels identified by the phase one segmentation network from debris and other artifacts using morphological properties of the group.

In Fig 3, we show joint probability densities of the

Figure 2. A cartoon diagram of the automated analysis pipeline. The pipeline is conducted in two distinct phases. In the first phase, groups of adhered pixels that may correspond to an adhered cell are identified. In phase two, these groups are classified by cell type.

size and eccentricity of all groups of adhered pixels identified by the segmentation network for two collections of MBM images: sickle red blood cell adhesion to laminin (Fig 3a) and CAR-T adhesion to E-selectin (Fig 3b). The histograms on the top and right edges of the figures are marginal distributions of either the size or eccentricity alone. The joint probabilities reveal three distinct classes of objects: small size / high eccentricity, small size / zero eccentricity, and large size / high eccentricity. The last category corresponds to sickle red blood cell or CAR-T cells, and hence we can use a size threshold (indicated by dashed vertical lines) to distinguish cells from other adhered objects (i.e. debris). Each group of pixels above the threshold is classified as a cell.

To validate the inter-experimenter reproducibility of this classification scheme, two researchers replicated five consecutive experiments each. The experimenters used a single tube of blood collected from a homozygous sickle-cell disease subject and analyzed the number of sickle cells using the MBM approach five times each over a fifteen minute period per experiment. The results did not show a significant difference between the experimenters (Fig 4a, p=0.696, two-way ANOVA with replication). Furthermore, the coefficient of variation within the replications of each experimenter demonstrated a decreasing trend with increasing experiment duration, (Fig 4b) and reached <25%, an important precision benchmark for bioanalytical method validation[22]. Collectively, these results provide reasonable assurance for the acquisition of meaningful results by showing that our experimental procedure and analysis pipeline are precise and independent of the experimenter.

Finally, we tested the accuracy of the procedure in counting cells, by comparing manual human and automated counts of adhered cells in various MBM images

for both sickle red blood cell adhesion to laminin (Fig 4c, $N = 175$) and CAR-T adhesion to E-selectin (Fig 4d, $N = 172$). The pipeline performed very well, with an R^2 value of 0.99 for both the sRBC and CAR-T cases. Importantly, the counts remained accurate even as the number of adhered cells becomes large in an individual image.

MBM can provide high-throughput single-cell dynamic data

The effectiveness of MBM at analyzing adhered cells in static images generalizes to videos, allowing us to characterize the dynamics of adhered cells on protein-functionalized surfaces. To facilitate this, we combined the classification procedure described above (applied to each frame of the video) with a cell tracking algorithm, described in detail in the SI. The tracking analysis distinguishes the motion of adhered cells between sequential frames from new adhesion events, and quantifies the total time spent by a cell on the surface before detachment.

We show eight representative cell trajectories in Fig 5a&b, four for sickle red blood cell adhesion to laminin, and four for CAR-T cell adhesion to E-selectin. We highlight three distinct types of adhesion events: adhesion events with large displacements, adhesion events with small displacements in the same direction as the flow, and adhesion events with small displacements in the opposite direction of the flow. These trajectories are generated for each adhered cell in an MBM video, giving us a high-throughput approach to collect dynamical information about large numbers of cells.

The adhesion time distributions in Fig 6 are one example of the dynamical data that can be compiled through our method, corresponding to thousands of individual trajectories. Fig 6a shows sickle red blood cell adhesion to laminin, and Fig 6b shows CAR-T cell adhesion to

Figure 3. Size and eccentricity distributions under whole blood flow. Joint probability distributions for the size and eccentricity of **a)** objects adhered to laminin functionalized channels from 176 images ($N = 211,879$) and **b)** objects adhered to E-selectin functionalized channels from 485 consecutive video frames ($N = 66,399$). Marginal distributions of size and eccentricity are shown on the top and right axes respectively. Dotted vertical lines indicate the size threshold used for classification, with objects above the threshold corresponding to sickle red blood cells (in panel a) or CAR-T cells (in panel b).

E-selectin. In both cases there is a peak in the distribution at low adhesion times, corresponding to a large number of rapid attachment/detachment events, but also a non-trivial proportion of long-lived trajectories. The tail of the distribution in the sickle red blood cell case is approximately power-law, while for CAR-T cells there is a plateau at intermediate times followed by a more rapid decay. The longest adhesion times (hundreds of seconds) occur infrequently, but the number of events collected by MBM is sufficiently large to resolve these rare cases. Adhesion time distributions are essential raw data for biophysical modeling of bond dynamics between cells and surface proteins under flow conditions, providing a valuable starting point for future studies.

But MBM also has the potential of revealing more subtle dynamical relationships that have not been systematically explored, for example how morphological features of cells are correlated with dynamical behaviors at the single cell level. For example, a longstanding question about sickle red blood cells is whether elongated, irreversibly sickled cells are more adhesive to endothelium than sickle discocytes[23]. In Fig 7, we show MBM data for the relationship between average sickle red blood cell eccentricity (calculated over the entire cell trajectory) and adhesion time to laminin, revealing an

unexpected negative trend ($p = 0.005$): cells with longer adhesion times are slightly less elongated on average.

We can also investigate how the orientation of the cell motion relative to the flow direction influences the dynamics. For each cell trajectory, the initial and final position (before detachment) defines a net displacement, which divided by adhesion time gives us an average velocity vector. The left column of Fig 8 shows distributions of the components of this vector parallel (with flow) and perpendicular (cross flow) to the flow direction. Moreover, we can correlate the average velocity components with adhesion time (right column of Fig 8).

This analysis reveals a variety of interesting features. The left panel of Fig 8a shows sickle red blood cell velocities on laminin, projected parallel to the flow direction. The distribution is peaked at zero velocity, but has contributions from both positive (with flow) and negative (against flow) velocities. As expected, the distribution is distinctly asymmetric: cells are less likely to move against the flow. The right panel of Fig 8a bins the sickle red blood cells by adhesion time, and depicts the average parallel velocity for each bin. The shortest adhesion times have the largest velocities, in agreement with the sample trajectories of Fig 5a. For velocities perpendicular to the flow direction (Fig 8b, left) the asymmetry

Figure 4. Validation and benchmarking of MBM. a) We establish inter-experimenter reproducibility for adhered sickle red blood cell (sRBC) counts for different experimental durations, carried out by two different researchers. Shaded regions around each line are one standard deviation. The cell adhesion results show no significant difference when two different experimenters perform MBM using the aliquots of the same patient sample **b)** We establish the sensitivity of MBM for adhered sickle red blood cell counts. After seven minutes of whole blood flow, the coefficient of variation drops below 25%, an important benchmark for the precision of bioanalytical methods. Finally, we establish the accuracy of MBM in counting **c)** sickle red blood cells adhered to laminin and **d)** CAR-T cells adhered to E-selectin respectively. There is a strong agreement between automated and human counts in both cases, as indicated by the R^2 value being close to one.

a

sRBC Adhesion To Laminin

b

CAR-T Cell Adhesion To E-selectin Surface

Figure 5. Example adhered cell trajectories. Shown are **a)** sickle red blood cell (sRBC) motion on laminin and **b)** CAR-T cell motion on E-selectin. Generally speaking, the longer a cell is adhered, the smaller the displacement of the cell in the direction of the flow. CAR-T cells tend to have higher motility relative to sickle red blood cells.

a

sRBC Adhesion Time Probabilities

b

CAR-T Cell Adhesion Time Probabilities

Figure 6. Probability distributions of adhesion times. Shown are **a)** sickle red blood cell (sRBC) adhesion to laminin ($N = 5666$) and **b)** CAR-T cell adhesion to E-selectin ($N = 2566$).

in the distribution vanishes: average velocities in either the up or down perpendicular direction are equally likely. The large tail at positive velocities that was visible in the parallel distribution also disappears. In order to achieve velocities with magnitudes much greater than 2 μm/s one clearly requires the assistance of flow in the same direction as the cell motion.

Fig 8c analyzes the motility of CAR-T cells along E-selectin, parallel to the flow direction. We noted in Fig 5 that CAR-T cells appeared to be more mobile than sickle red blood cells, and this characteristic is validated in the distribution of average velocities (Fig 8c, left). Relative to the velocity distribution of sickle red blood cells (Fig 8a, left), we see about an order of magnitude higher probabilities at large positive velocities. In fact

there is a subsidiary peak in the CAR-T distribution around 2 μs/s, in addition to the zero-velocity peak. These two subpopulations (≈ 0 and ≈ 2 μm/s velocity cells) are also evident when looking at the correlation of velocity with adhesion time (Fig. 8c, right). On the top of the overall trend of shorter adhesion times associated with higher velocities, we see the 2 μm/s group cluster around 75 s adhesion time and the 0 μm/s group cluster around 350 s adhesion time.

Fig. 8 is just one illustration of the versatility of MBM as an *in vitro* platform for blood cell motility analysis, which to date has been exclusively performed with intravital microscopy. Another promising area for exploration is the impact of environmental signals on motility, for example, CD19 activation of CAR-T cells. The majority of CAR-T cells evaluated for B-cell malignancies target

Figure 7. Average eccentricity vs. adhesion time for sickle red blood cell (sRBC) adhesion to laminin under normoxic conditions. Dotted lines show 95% confidence intervals for the line of best fit. The blue histogram is the distribution of average eccentricities (N = 5666). There is a negative relationship P = (0.005) between average eccentricity and adhesion time. Error lines for each scatter point are one standard deviation.

CD19. T cells express programmed cell death protein 1 (PD-1) during activation and You et al. used intravital microscopy to show that T cell motility was proportional to PD-1 expression[24,25]. As an in vitro alternative, we compare CAR-T velocities on E-selectin (parallel to the flow) with and without CD19 activation by using MBM Fig. 9. Though the E-selectin-functionalized substrate is used here for simplicity, in the future this same kind of analysis can be done on endothelialized substrates for added physiological realism. Fig. 9a shows the velocity distribution for unactivated cells as a control reference, the same distribution as in Fig. 8c but plotted on a linear scale. The velocity bins are color-coded by different velocity regimes, and these same colors are used to label the trajectories in Fig. 9b by their respective velocity regimes. In the absence of CD19 activation, adhered cells fall mainly into relatively immobile populations (the peak around zero velocity, highlighted in purple) and those moving at below 2 μ/s (light blue). As expected, the small velocity population consists mainly of cells that adhere at one location, while the mobile cells show a range of trajectory lengths. Activation significantly enhances velocities above 2 μm/s (Fig. 9c), and we see

a similar distribution of trajectory lengths across the different velocity regimes (Fig. 9d) for mobile cells.

Microfluidics holds great promise for research into the cellular interactions that take place in blood, and as a result various microscopy approaches have been developed which complement microfluidic systems. Yet faithfully recreating *in vivo* conditions in microfluidics is a challenging task, and a bottleneck for its wider adoption. The further a microfluidic study veers from true-to-life conditions, the less physiologically relevant the results become. Therefore, if we are to generate truly meaningful results under *in vitro* conditions, our methods should strive to reproduce the environment of the microvasculature as much as possible. Researchers have developed a variety of approaches to mimic the microvasculature[26,27], but the blood—the main constituent with its full complexity—has always been missing in these microfluidic systems. MBM enables researchers to investigate the dynamics of cellular interactions in the presence of blood flow. This means that many of the relevant physical forces and chemical signals which modulate cellular interactions will be present during an experiment. Environmental completeness would be especially crucial when the functional pathways of cells are actively regulated by their environment. One example of such regulation is the influence of the complement and renin-angiotensin systems on leukocyte function[28]. Therefore, we anticipate that MBM will have a formative effect on microfluidic studies. Even so, a new method is only as good as the available tools for analysis. Our machine learning pipeline for MBM can automatically analyze hundreds of thousands of observed adhesion events. Typically, the task of completing this analysis manually would be impossible, as the amount of time required, or the number of people required, would be far too large. Our automated analysis approach is flexible and mitigates errors due to individual human biases in cell counting and classification.

The applications of MBM are wide-ranging for both fundamental investigations into biophysical mechanisms and clinical studies. For example, adhesion duration distributions are crucial for studies of force-dependent binding/unbinding of protein complexes and cell-cell interactions[29]. MBM can complement existing approaches in this area like numerical simulations or force spectroscopy experiments (i.e. atomic force microscopy)[30–36]. MBM could provide a valuable comparison point for bond lifetime simulations, adding extra physiological realism due to the blood flow. Because of its ease of

Figure 8. Velocities of cells under whole blood flow. Each row shows a velocity probability distribution (left) and a corresponding average velocity versus adhesion time plot (right). Three cases are depicted: **a)** With flow (parallel to the flow direction) for sickle red blood cell adhesion (sRBC) to laminin ($N = 5666$). The probability distribution shows that a large majority of adhesion events have near-zero velocities. **b)** Cross flow (perpendicular to the flow direction) for sickle red blood cell (sRBC) adhesion to laminin ($N = 5666$). Because this is the cross-flow direction, we expect the distribution to approach that of a random walk, and the average velocities be near zero. **c)** With flow for CAR-T cell adhesion to E-selectin ($N = 2566$). When comparing this row to **a**, we see that CAR-T cells adhering to E-selectin tend to have larger with flow velocities. Cross flow analysis of CAR-T cells is not shown, but the results are similar to those of **b**.

Figure 9. The effect of CD19 activation on the motility of CAR-T cells adhered to E-selectin. a,c) Velocity distribution of CAR-T cells for control (a) and CD19 activated (c) systems, with color coding highlighting different velocity ranges. **b,d)** Trajectories of adhered cells for the control (b) and activated (d) cases, labeled by their respective velocity ranges according to the same color scheme as in the distribution panels. CD19 activation leads to a large increase in trajectories with velocities above 2 μm/s.

implementation, it is also accessible to a broader research community than atomic force microscopy techniques. On the clinical front, there are a variety of promising applications for specific diseases. In the case of cancer, MBM could help visualize crucial interactions of circulating tumor and endothelial cells with improved physiological accuracy[37]. It can also be utilized to distinguish and isolate rare circulating tumor cells from billions of other cells flowing through a microfluidic channel. For sickle cell disease, MBM could serve as a means to understand how sickle cells behave under changes in blood pressure,

flow rates, and viscosity, as well as concentrations of relevant constituents in whole blood. Furthermore, the concentration of adhered sickle red blood cells, as well as the morphological characteristics of the cells may be used to monitor if a sickle cell disease patient is undergoing a flare-up, or if they are in an asymptomatic diseased state. Other possible disease contexts where MBM may be useful are solid tumor malignancies. MBM may help us characterize the motility of immune cells including T-cells and macrophages with great physiological relevance. As a final example, MBM could be used to

further our understanding of autoimmune diseases, such as rheumatoid arthritis, where leukocyte recruitment is an important part of the pathology. For example, MBM can be used in tandem with "joint on a chip" systems, where complex interactions among blood cells, endothelium and skeletal components are studied with *in vitro* experiments[38].

Aside from clinical studies, MBM with automated analysis has the ability to efficiently gather large quantities of data that may provide an ideal basis for theoretical modeling of cellular adhesion mechanics. For example, the velocity distributions along the flow direction from Fig. 8 qualitatively resemble those of molecular motors that are capable of forward and backward stepping along a cytoskeletal track, which can be described via coarse-grained kinetic models of the underlying biochemical cycle[39]. Similar mathematical approaches could be brought to bear on the cell velocity data, giving us a more complete picture of the cell interactions with the endothelium that give rise to these kinds of dynamics. The excellent statistics of the data set is crucial in this regard, since in principle it enables us to distinguish between competing models.

As with any experimental approach, there are also limitations. In sufficiently complex scenarios where large densities and/or multiple cell types adhere to the surface, extracting single-cell dynamical information may become more difficult. Large densities, with many overlapping adhered cells, could impede the video analysis in determining if two cells in consecutive frames are the same or not. Overlaps were rare in the examples we investigated, allowing for straightforward extraction of cell trajectories. In physical systems where this would be an issue, however, more sophisticated methods could be incorporated into MBM analysis workflow for tracking cells[40,41]. Similarly, both the sickle red blood cell and CAR-T datasets we focused on in the main text involve a single adhered cell type, which could be identified reliably using a size threshold. In cases with multiple cell types of similar sizes, we would have to employ a more involved classification procedure, using other morphological characteristics of the cells. The SI shows results from one such example, where a convolutional neural network was trained to reliably distinguish CAR-T cell adhesion to P-selectin from red blood cells, which also adhere to the surface. Thus we believe all these limitations may be substantially addressed with additions to the data analysis approach.

In summary, MBM is a robust, easy-to-implement,

high-throughput method to study cell adhesion dynamics in the presence of blood flow. Its flexibility allows for broad deployment in both clinical contexts and studies of fundamental biophysical mechanisms underlying cell adhesion. Combined with automated analysis for cell classification and tracking, it promises to be a general platform for elucidating how interactions with the complex whole blood environment influence and regulate cellular adhesion.

Methods

Blood sample collection and microfluidic adhesion assays

Surplus EDTA-anticoagulated whole blood samples were collected under an Institutional Review Board approved protocol (registered at www.clinicaltrials.gov as NCT02824471, "Sickle Cell Disease Biofluid Chip Technology"). The microfluidic platforms were fabricated and functionalized as described previously[42]. For studying the adhesion of sickle red blood cells, unprocessed whole blood samples from homozygous sickle cell disease subjects were used and microchannels were coated with laminin (murine, laminin-1). For demonstrating CAR-T motility, the microchannels were coated with either P- or E-selectin (human, CD62P and CD62E). Whole blood samples from subjects with no hemoglobinopathies were first leukodepleted and then premixed with CAR-T cell populations. To prevent non-specific cellular interactions, microchannels were incubated with bovine serum albumin at 4°C overnight, and additionally, 250 μl of bovine serum albumin was injected into the microchannels at room temperature with a 5 μl/min flow rate 2 hours prior to the experiments.

Motion blur microscopy

Microchannels were visualized with Olympus CellSens software using an Olympus IX83 inverted microscope and QImaging EXi Blue CCD camera with 10x objective (numerical aperture 0.3, pixel area 6.5 μm^2). To induce motion blur, camera exposure was set to 1.2 seconds. Images and videos of the microchannel surface were saved uncompressed to reduce noise. Frame rates of the videos were kept at $(1/1.2)$ s^{-1}, which is the maximum frame rate for a 1.2 s exposure time. High integration time was compensated for by adjusting the voltage of the light source to 2.7V (maximum 12V). It should be noted that it is key to adjust the focus map for the surface of interest before the microchannels are injected with the sample. Unprocessed or leukodepleted whole blood was

withdrawn into a syringe which then was loaded into a constant displacement syringe pump (NE-1000, New Era Pump Systems Inc.). MBM with automated analysis requires a minimum flow velocity of approximately 150 μm/s in the background to create the minimum particle streak for MBM. Higher background flow velocity yields better distinction of cellular interactions. The flow velocity was kept at approximately 500 μm/s for 20 minutes for demonstrating sickle red blood cell adhesion to laminin. For the demonstration of CAR-T cell adhesion to E-selectin, the flow velocity was swept linearly from 500 μm/s to 3500 μm/s in 10 minutes.

Automated image analysis

To analyze MBM images, we developed an automated analysis procedure with two distinct phases. Phase 1 performs segmentation on MBM images at the pixel level and phase 2 classifies groups of pixels by cell type. After classification, we can perform further analysis on the classified cells (i.e. morphology characterization and trajectory tracking). While we focus on three experimental examples in this paper, our approach can easily be generalized to other systems.

Phase 1: pixel segmentation

Phase 1 uses a machine learning network to segment every pixel in an MBM image into one of two categories: adhered or background. Adhered pixels correspond to a pixel of the image belonging to an adhered object, with the remaining pixels falling into the background category.

The segmentation network is based on a modified U-Net architecture from our previous work[43] (further details are provided in the SI). To train the network, one or more MBM images is chosen as training/validation images. A labeled mask of the training/validation images is created, by manually labelling each pixel as adhered or background. Each labeled training/validation mask, as well as each training/validation image, is then be split into tiles of size 150x150 pixels, which are subsequently resized to 128x128 pixels, as that is the required input of the segmentation network. Once all of the training/validation tiles are compiled, we can then train the segmentation network for our particular task.

For the purposes of this paper, we have trained and used two distinct segmentation networks, one for analysis of sickle red blood cell adhesion to laminin, and one for analysis of CAR-T cell adhesion to E-selectin. The first network was trained on one full MBM channel image yielding 1515 tiles, while the second network was trained on thirteen frames of an MBM video, yielding 702 tiles. In both cases 80% of the tiles were used for training, and 20% for validation.

Phase 2: cell classification

The output of the segmentation network is a fully segmented image, where each pixel belongs to one of two classes. In the second phase of the analysis, groups of adhered pixels are extracted and classified. Generally speaking, the cell identification task falls into one of two cases. In the first case, only one adhered cell type is expected to be distinguishable in the MBM image. In the second case, two or more cell types are expected to be distinguishable in the MBM image. For brevity, we will only highlight the single cell type case in the main text, and interested readers can refer to the SI for information on the multiple cell type case.

When only a single cell type of interest is expected to adhere to the functionalized channel, we can classify groups of adhered pixels using a physical property of the group. For example, when we look at the attachment of sickle red blood cells to laminin or CAR-T cells to E-selectin, the vast majority of groups of adhered pixels correspond to one cell type. In this case, we can use a size threshold to distinguish adhered cells from groups of adhered pixels falsely classified by the segmentation network as adhered.

When training both segmentation networks, the size threshold for the classification of groups of adhered pixels was optimized with the objective to produce the closest agreement between human and machine counts of relevant cells. To find the optimum threshold, an array of threshold values can be used to count cells for an array of inputs. The automated cell counts can then be compared to manual cell counts, and the threshold which produces the greatest agreement can be chosen. For single-cell type analysis, two optimum thresholds were found, one for sickle red blood cell adhesion to laminin and one for CAR-T cell adhesion to E-selectin.

In the laminin case, 175 MBM images were analyzed using a trained segmentation network and size thresholds in the range 0-250 px (corresponding to the area of the adhered pixel group), in increments of ten, were used to produce sickle red blood cell counts. Agreement between human and machine counts was checked for each threshold value, and the optimum threshold was found to be 90 pixels (38 μm^2), where groups of adhered pixels with a size greater than 90 pixels were classified

as sickle red blood cells, and groups of adhered pixels with size less than 90 pixels were classified as other. For the CAR-T case, 172 frames from an MBM video were analyzed using a trained segmentation network and size thresholds in the range 0-150 px, in increments of ten, were used to produce CAR-T cell counts. The optimum threshold was found to be 40 pixels (17 μm^2).

Post-classification analysis

After phase 1 and phase 2 are completed, the groups of adhered pixels that are positively classified as the cell type(s) of interest can be further analyzed. Properties like the size, eccentricity and position of each individual cell are easily obtainable. These static data can be analyzed individually or aggregated together to produce group statistics. When MBM video is available, we can analyze the dynamics of cells under flow. We define a parameter for each adhesion event called the survival time—the time from when a cell attaches to the surface until it detaches or leaves the frame. With the survival time in hand, as well as the static data for each frame, we can generate dynamic data for each cell. Individual cell dynamics like average velocity, average size, and average eccentricity can be found, or, we can aggregate cells and produce group statistics like mean-squared displacement, or distributions of dynamic quantities.

Identification rules for tracking moving cells in MBM videos

MBM video analysis brings forth additional challenges beyond those of individual images. In particular, we need to track cells over time, requiring us to determine if cells in consecutive frames are the same cell or not. Two distinct issues arose when making this determination: duplicate counts, and tracking motile cells.

Duplicate counts occur when the segmentation network erroneously segments a single group of adhered pixels as two distinct groups of adhered pixels, whose edges are close, but not touching. To minimize the duplicate counts, we checked to see if the centroids of any two groups of adhered pixels were within 50 pixels (1 px = 0.322 μm) of one another. Also, the shortest distance between the edges of all such pairs was computed, and if this distance was 5 pixels or fewer for any pair, the two groups of adhered pixels were considered to be one. In this case, the two groups of adhered pixels were replaced by the smallest polygon that contains them, known as the *convex hull*. One can imagine this shape as the result of stretching a rubber band over the two groups of adhered pixels and letting it rest taut.

Another case of duplicate counting occurs if phase 2 misclassified an adhered cell as "other" for any amount of frames between two positive classifications. This error would cause the analysis to detect a "new" group of adhered pixels, despite it being the same cell. We avoid this by giving each adhesion event a disappearance allowance. Each cell for an adhesion event was allowed to "disappear" for a number of frames before considering the adhesion event completed. If a cell "disappeared" for a number of frames below the allowance, and then "reappeared", the adhesion event would continue as if the misclassification never occurred. For sickle red blood cells, we allowed cells to go undetected for up to two frames and still count as the same cell. For CAR-T cells, they could go undetected for up to nine frames. These frame thresholds helped us to not preemptively end adhesion events.

Interacting motile cells are constantly adhering, detaching, and persisting in their adhesion to the surface, and as such, the results of analyzing many frames present a complex picture of many adhesion events happening at various times. Thus, we devised rules by which to determine if a cell in consecutive frames was a new adhesion event or a continuation of a prior adhesion event. If a positively identified group of adhered pixels was within 30 pixels parallel to the flow direction, or within 5 pixels normal to the flow direction of a positively identified group of adhered pixels in the previous frame, the group of adhered pixels was classified as a continuation of the adhesion event. Otherwise, the positively identified group of adhered pixels was classified as the beginning of a new adhesion event. These movement rules plus the frame thresholds protected us from splitting a singular adhesion event into two due to any misclassifications by the automated analysis.

Statistical methods

Statistical analyses were performed using the Microsoft Excel Analysis ToolPak add-in. $p < 0.05$ was chosen to indicate a significant difference.

Code and data availability

All the code associated with the MBM analysis, as well as a link to the Open Science Framework repository containing the datasets, can be found at: https://github.com/hincz-lab/motion-blur-microscopy.

Acknowledgments

We express deep gratitude to Paolo F. Caimi at the Case Comprehensive Cancer Center for sharing CAR-T cells, to Paul Tesar at the Case Western Reserve University School of Medicine for cell culture resources, and to Jane Reese Koc at the Hematopoietic Biorepository and Cellular Therapy Core. This work was supported by the National Heart Lung and Blood Institute R01HL133574, OT2HL152643, T32HL134622, K25HL159358, and National Science Foundation CAREER awards 1552782 and 1651560. This work was partially supported by the National Science Foundation CAREER award 2023525. The authors acknowledge with gratitude the contributions of research participants and clinicians at Seidman Cancer Center (University Hospitals, Cleveland).

Authorship Contributions

U.G. and U.A.G developed the idea. U.G., B.S., A.G., M.H. and U.A.G. designed the study. U.G., W.W., and Y.M. performed the experiments. U.G., B.S., A.G. and M.H. developed the automated algorithm. U.G., B.S., A.G., Y.M, W.W, R.A, M.H. and U.A.G. analyzed the data. U.G., B.S., A.G., W.W, M.H. and U.A.G. discussed and interpreted the data. U.G., B.S., A.G., wrote the manuscript, U.G., B.S., A.G., prepared the figures. M.H. and U.A.G. edited the manuscript.

Conflict-of-Interest Disclosure

R.A., U.A.G., and Case Western Reserve University have financial interests in Hemex Health Inc. U.A.G., and Case Western Reserve University have financial interests in BioChip Labs Inc. U.A.G. and Case Western Reserve University have financial interests in Xatek Inc. U.A.G. has financial interests in DxNow Inc. Financial interests include licensed intellectual property, stock ownership, research funding, employment, and consulting. Hemex Health Inc. offers point-of-care diagnostics for hemoglobin disorders, anemia, and malaria. BioChip Labs Inc. offers commercial clinical microfluidic biomarker assays for inherited or acquired blood disorders. Xatek Inc. offers point-of-care global assays to evaluate the hemostatic process. DxNow Inc. offers microfluidic and bio-imaging technologies for *in vitro* fertilization, forensics, and diagnostics. The competing interests of Case Western Reserve University employees are overseen and managed by the Conflict of Interests Committee according to a Conflict-of-Interest Management Plan.

References

1. Sai, J., Rogers, M., Hockemeyer, K., Wikswo, J. P. & Richmond, A. Study of chemotaxis and cell–cell interactions in cancer with microfluidic devices. In *Methods in Enzymology*, vol. 570, 19–45 (Elsevier, 2016).

2. Hines, P. C. *et al.* Flow adhesion of whole blood to P-selectin: a prognostic biomarker for vaso-occlusive crisis in sickle cell disease. *Br. J. Haematol.* **194**, 1074–1082 (2021).

3. Man, Y. *et al.* Leukocyte adhesion to P-selectin and the inhibitory role of crizanlizumab in sickle cell disease: A standardized microfluidic assessment. *Blood Cells, Mol. Dis.* **83**, 102424 (2020).

4. Hernandez, A. A. *et al.* An allosteric shift in CD11c affinity activates a proatherogenic state in arrested intermediate monocytes. *J. Immunol.* **205**, 2806–2820 (2020).

5. Barshtein, G., Zelig, O., Gural, A., Arbell, D. & Yedgar, S. Determination of red blood cell adhesion to vascular endothelial cells: A critical role for blood plasma. *Colloids Surfaces B: Biointerfaces* 112226 (2021).

6. Gutierrez, M., Fish, M. B., Golinski, A. W. & Eniola-Adefeso, O. Presence of rigid red blood cells in blood flow interferes with the vascular wall adhesion of leukocytes. *Langmuir* **34**, 2363–2372 (2018).

7. Buscher, K., Marcovecchio, P., Hedrick, C. C. & Ley, K. Patrolling mechanics of non-classical monocytes in vascular inflammation. *Front. Cardiovasc. Medicine* **4**, 80 (2017).

8. Cabrales, P. *et al.* The macrophage stimulating anti-cancer agent, RRx-001, protects against ischemia-reperfusion injury. *Expert. Rev. Hematol.* **10**, 575–582 (2017).

9. Cheburkanov, V., Keene, E., Pipal, J. & Yakovlev, V. V. Towards in vivo larynx imaging: assessing mechanical properties of larynx with Brillouin microscopy. In *Optical Biopsy XX: Toward Real-Time Spectroscopic Imaging and Diagnosis*, vol. 11954, 22–28 (SPIE, 2022).

10. Hwang, K. *et al.* Handheld endomicroscope using a fiber-optic harmonograph enables real-time and in vivo confocal imaging of living cell morphology and capillary perfusion. *Microsystems & Nanoeng.* **6**, 1–11 (2020).

11. Wang, F. *et al.* In vivo NIR-II structured-illumination light-sheet microscopy. *Proc. Natl. Acad. Sci.* **118**, e2023888118 (2021).

12. Condeelis, J. & Weissleder, R. In vivo imaging in cancer. *Cold Spring Harb. Perspectives Biol.* **2**, a003848 (2010).

13. Sundd, P. *et al.* 'Slings' enable neutrophil rolling at high shear. *Nature* **488**, 399–403 (2012).

14. da Silva, B. C. G., Tam, R. & Ferrari, R. J. Detecting cells in intravital video microscopy using a deep convolutional neural network. *Comput. Biol. Medicine* **129**, 104133 (2021).

15. Gregorio da Silva, B. C., Carvalho-Tavares, J. & Ferrari, R. J. Detecting and tracking leukocytes in intravital video microscopy using a Hessian-based spatiotemporal approach. *Multidimens. Syst. Signal Process.* **30**, 815–839 (2019).

16. Jennissen, H., Sanders, A., Schnittler, H. & Hlady, V. TIRF-rheometer for measuring protein adsorption under high shear rates: constructional and fluid dynamic aspects. *Materialwissenschaft und Werkstofftechnik: Entwicklung, Fertigung, Prüfung, Eigenschaften und Anwendungen technischer Werkstoffe* **30**, 850–861 (1999).

17. Pisapati, A. V. *et al.* Characterizing single-molecule conformational changes under shear flow with fluorescence microscopy. *J. Vis. Exp.* e60784 (2020).

18. Avtaeva, Y. N., Mel'nikov, I., Okhota, S., Zozulya, N. & Gabbasov, Z. Kinetics of platelet adhesion to protein-coated surface in whole blood samples at high flow rates. *Bull. Exp. Biol. Medicine* **169**, 229–232 (2020).

19. Goreke, U., Bode, A., Yaman, S., Gurkan, U. A. & Durmus, N. G. Size and density measurements of single sickle red blood cells using microfluidic magnetic levitation. *Lab on a Chip* **22**, 683–696 (2022).

20. Man, Y. *et al.* Occlusionchip: a functional microcapillary occlusion assay complementary to ektacytometry for detection of small-fraction red blood cells with abnormal deformability. *Front. Physiol.* 1526.

21. An, R. *et al.* Sickle red blood cell-derived extracellular vesicles activate endothelial cells and enhance sickle red cell adhesion mediated by von Willebrand factor. *Br. J. Haematol.* **201**, 552–563 (2023).

22. Guidance for Industry, Bioanalytical Method Validation. *U.S. Department of Health and Human Services, Food and Drug Administration, Center for Drug Evaluation and Research (CDER), Center for Veterinary Medicine (CMV)* (2018).

23. Goreke, U. *et al.* Membrane bending and sphingomyelinase-associated, sulfatide-dependent hypoxic adhesion of sickle mature erythrocytes. *Blood Adv.* **7**, 2094–2104 (2023).

24. Sharpe, A. H. & Pauken, K. E. The diverse functions of the PD1 inhibitory pathway. *Nat. Rev. Immunol.* **18**, 153–167 (2018).

25. You, R. *et al.* Active surveillance characterizes human intratumoral T cell exhaustion. *J. Clin. Investig,* **131** (2021).

26. Wong, K. H., Chan, J. M., Kamm, R. D. & Tien, J. Microfluidic models of vascular functions. *Annu. Rev. Biomed. Eng.* **14**, 205–230 (2012).

27. Goreke, U. *et al.* Catch bonds in sickle cell disease: shear-enhanced adhesion of red blood cells to laminin. *Biophys. J.* (2023).

28. Bekassy, Z., Lopatko Fagerström, I., Bader, M. & Karpman, D. Crosstalk between the renin–angiotensin, complement and kallikrein–kinin systems in inflammation. *Nat. Rev. Immunol.* 1–18 (2021).

29. Bell, G. Models for the specific adhesion of cells to cells. *Science* **200**, 618–27 (1978).

30. Hillringhaus, S., Dasanna, A. K., Gompper, G. & Fedosov, D. A. Stochastic bond dynamics facilitates alignment of malaria parasite at erythrocyte membrane upon invasion. *Elife* **9**, e56500 (2020).

31. Anvari, S., Osei, E. & Maftoon, N. Interactions of platelets with circulating tumor cells contribute to cancer metastasis. *Sci. Reports* **11**, 1–16 (2021).

32. Lansche, C. *et al.* The sickle cell trait affects contact dynamics and endothelial cell activation in Plasmodium falciparum-infected erythrocytes. *Commun. Biol.* **1**, 1–14 (2018).

33. Strnad, M. *et al.* Nanomechanical mechanisms of Lyme disease spirochete motility enhancement in extracellular matrix. *Commun. Biol.* **4**, 1–9 (2021).

34. Arora, N., Hazra, J. P. & Rakshit, S. Anisotropy in mechanical unfolding of protein upon partner-assisted pulling and handle-assisted pulling. *Commun. Biol.* **4**, 1–10 (2021).

35. Yang, J. *et al.* Molecular interaction and inhibition of SARS-CoV-2 binding to the ACE2 receptor. *Nat. Commun.* **11**, 1–10 (2020).

36. Li, N. *et al.* Ligand-specific binding forces of LFA-1 and Mac-1 in neutrophil adhesion and crawling. *Mol. biology cell* **29**, 408–418 (2018).

37. Priyadarshani, J., Roy, T., Das, S. & Chakraborty, S. Frugal approach toward developing a biomimetic, microfluidic network-on-a-chip for in vitro analysis of microvascular physiology. *ACS Biomater. Sci. & Eng.* **7**, 1263–1277 (2021).

38. Paggi, C. A., Teixeira, L. M., Le Gac, S. & Karperien, M. Joint-on-chip platforms: entering a new era of in vitro models for arthritis. *Nat. Rev. Rheumatol.* **18**, 217–231 (2022).

39. Vu, H. T., Chakrabarti, S., Hinczewski, M. & Thirumalai, D. Discrete step sizes of molecular motors lead to bimodal non-gaussian velocity distributions under force. *Phys. Rev. Lett.* **117**, 078101 (2016).

40. Kok, R. *et al.* OrganoidTracker: Efficient cell tracking using machine learning and manual error correction. *PLOS One* **15**, e0240802 (2020).

41. Aragaki, H., Ogoh, K., Kondo, Y. & Aoki, K. LIM tracker: a software package for cell tracking and analysis with advanced interactivity. *Sci. Reports* **12** (2022).

42. Alapan, Y. *et al.* Sickle cell disease biochip: a functional red blood cell adhesion assay for monitoring sickle cell disease. *Transl. Res.* **173**, 74–91 (2016).

43. Praljak, N. *et al.* Integrating deep learning with microfluidics for biophysical classification of sickle red blood cells adhered to laminin. *PLoS Comput. Biol.* **17**, e1008946 (2021).

44. Simonyan, K. & Zisserman, A. Very deep convolutional networks for large-scale image recognition. *arXiv:1409.1556* (2014).

45. Russakovsky, O. *et al.* ImageNet large scale visual recognition challenge. *arXiv:1409.0575* (2015).

Supplementary Information: Motion Blur Microscopy

Phase 1 training details

Ideally, the chosen training/validation images are dense in the objects of interest, which will combat any class imbalance during training. Labeled masks for the training/validation images were created by coloring each pixel of the images one of two colors: one color for adhered pixels, and one color for background pixels. This process was completed by first coloring all adhered pixels one color using the software GIMP. Any non-colored pixels were then filled in automatically with a Python script. When resizing tiles from 150x150 pixels to 128x128 pixels, the cubic interpolation method of the Python library OpenCV was used.

To train the segmentation networks (with architecture summarized in Fig. SI1), a custom data generator was created, which loaded tiles into the network in batches of 32 and applied data augmentation. Data augmentations included are rotations of 90, 180, 270, and 360 degrees, horizontal flips, and vertical flips. In total, the augmentations can produce 8 unique orientations of each tile. A stop function was included in the training, which automatically stopped the training after the validation loss had not decreased for five straight epochs. The weights of the network after the five straight epochs were used as the trained network weights. The loss function used for optimization was a categorical cross-entropy function from the Keras library. The training was completed using an Adam optimizer with a learning rate of 0.001, and carried out in a Jupyter notebook using an NVIDIA GeForce RTX 2060 SUPER GPU with 8GB of memory and an AMD Ryzen 7 3700X 8-Core CPU.

The first segmentation network was trained on one full MBM image taken from a channel functionalized with laminin for sickle red blood cell adhesion. The chosen image originally had dimensions of 15171x2392 pixels. When creating the mask for this segmentation network, pixels corresponding to the adhered category were manually colored white. After the manual labeling was completed, the mask was sent through a python code which labeled all background pixels as black. The number of training/validation tiles generated for this segmentation network was 1515 tiles, where 80 percent were used for training, and 20 percent for validation. Augmentations were applied to the training tiles, resulting in 9696 tiles for training, and 303 tiles for validation.

The second segmentation network was trained on 13 frames of an MBM video taken from a channel functionalized with E-selectin for CAR-T cell adhesion. Each frame had dimensions of 1392x1040 pixels. When creating the masks for this segmentation network, pixels corresponding to the adhered category were manually colored green. After the manual labeling was completed, the masks were sent through a python code which labeled all background pixels as black. These frames generated a total of 702 training/validation tiles, where 80 percent were used for training, and 20 percent for validation. Augmentations were applied to the training tiles, resulting in 4496 tiles for training, and 140 tiles for validation.

MBM with automated analysis with multiple cell types

When multiple cell types are expected to be visible in an MBM image, phase 2 classification becomes more complicated. In general for this case using a simple physical parameter (like cell size / eccentricity) for groups of adhered pixels will not be sufficient for accurate classification. The reason for this is because the distribution of physical parameters can overlap between the different types. Thus, trying to classify with only a physical parameter may lead to errors. One solution to this problem is using a machine learning classification network for phase 2, which we demonstrate here as a proof of concept to classify CAR-T cell adhesion to P-selectin in the presence of red blood cells (where both CAR-T and red blood cells are expected to adhere).

Results

For simplicity, we re-used the phase 1 segmentation network from analysis of sickle red blood cell adhesion to laminin for analysis of CAR-T cell adhesion to P-selectin. To demonstrate the necessity of a machine learning approach for phase 2, we show simple physical characteristics (eccentricity versus size) of the adhered pixel groups in Fig. SI2. As in main text Fig. 3, there are only two distinct peaks: one at small areas (misclassified pixels / detritus) and one at large areas (all adhered cells). The two types of adhered cells do not form distinguishable peaks.

To get reliable CAR-T cell classification, we trained a VGG16 convolutional neural network[44] as decscribed in the next section. The goal was to identify adhered pixel groups as either CAR-T cells or as non-CAR-T ("other"), starting with pixel groups that had sufficiently large areas to be cells (i.e. after passing the size threshold). The output of this binary classification network is a

Fig. SI1. The architecture of the phase 1 segmentation network. The architecture is a modified U-Net, based on Ref. 43. We show an example input and output for sickle red blood cell adhesion to laminin, as well as manual classification for comparison. The groups of adhered pixels found by the segmentation network appear to match manual classification well. Conv = 2D Convolutional. BN = Batch Normalization. MP = Max Pooling. ConvT = 2D Convolutional Transpose. Concat = Concatenation.

scalar between 0-1, which can be roughly thought of as the probability of the group being a CAR-T cell. To positively assign a CAR-T classification, we have a choice of confidence threshold, with all network outputs above the threshold labeled as CAR-T. The default threshold would be 0.5, but higher thresholds can also be used to get more conservative assignments. The effectiveness of the binary classification can be measured via an F_1 score, which is harmonic mean of precision and recall (with values in the range 0-1, where 1 corresponds to perfect classification). As shown in Fig. SI3a, most confidence thresholds in the range 0.5-0.99 (except the highest ones) give excellent F_1 scores $\gtrsim$ 90% for classifying a set of 115/770 CAR-T/other cells. We chose a threshold of 0.8, and for this value Fig. SI3b shows the corresponding confusion matrix and Fig. SI3c shows a comparison to manual counting of CAR-T cells by human experts. The excellent agreement ($R^2 = 0.95$) shows that is possible to have accurate phase 2 classification of MBM images via neural networks even in the presence of multiple cell types.

Phase 2 training details

The classification network is a VGG16 architecture[44] pre-trained on ImageNet[45] with additional trainable layers added at the end (Fig. SI4). Note there is a special pre-processing applied to MBM images for CAR-T cell adhesion to P-selectin. Each pixel from each MBM image is color adjusted according to the equation:

$$c_i' = \frac{c_i - (\mu - 4\sigma)}{(\mu + 4\sigma) - (\mu - 4\sigma)}, \tag{1}$$

where c_i' is the color of a pixel after adjustment, c_i is the original color of the pixel, μ is the mean color of the MBM image, and σ is the standard deviation of the color of the MBM image. This color adjustment is not strictly necessary, but it helps the phase 1 segmentation network generalize to a new experimental set-up (the alternative would have been to re-train the phase 1 network for this experimental case).

To train the classification network, 40x40 pixel regions are extracted from MBM images. These regions are centered on groups of adhered pixels classified by the phase 1 segmentation network which have areas larger than the size threshold of 180 pixels. To create a training/validation dataset these regions are manually classified as CAR-T or other. We extracted 8830 regions, and classified 1143 CAR-T cells, and 7687 other, where 70% of regions were used for training, 20% of images were used for validation, and 10% of regions were used for generating F_1 scores and confusion matrices. Data augmentations such as horizontal and vertical flips were applied to the training/validation data set, resulting in four unique orientations for each image, and giving 3200/21523 CAR-T/other images for training and 914/6149 CAR-T/other images for validation. A stop

Fig. SI2. CAR-T cell/sickle red blood cell adhesion to P-selectin size vs. eccentricity distribution. Hexplot showing the joint distribution of size and eccentricity for CAR-T/red blood cell adhesion to P-selectin for 947 consecutive video frames ($N = 103894$). Projections on each axis show the marginal distributions of size and eccentricity alone. Note that the distribution peak at large areas corresponds to both CAR-T and red blood cells, and so the two types cannot be distinguished via size/eccentricity characteristics alone.

function was included in the training, which automatically stopped the training after the validation loss had not decreased for 10 straight epochs. The best weights of the network after the 10 straight epochs were used as the trained network weights. The training was carried out via an Adam optimizer with a learning rate of 0.001 and batch size of 32. The loss function used for optimization was a categorical cross-entropy function from the Keras library. The training was completed on a Jupyter notebook using an NVIDIA GeForce RTX 2060 SUPER GPU with 8GB of memory and an AMD Ryzen 7 3700X 8-Core CPU.

Fig. SI3. Validation of MBM with automated analysis with more than one cell type a) F_1 score vs. confidence threshold plot. There are a range of confidence thresholds which produce acceptable F_1 scores. **b)** Confusion matrix for the classification network on a subset of groups of adhered pixels ($N = 885$), for a confidence threshold of 0.8. The true positive rate is large for quantified cells. **c)** Comparison between automated and human counts of CAR-T cell adhesion to P-selectin. The R^2 value indicates good agreement between the two counts.

Fig. SI4. Phase 2 classification network architecture a) Schematic of the architecture for the phase 2 classification network. Only the final three layers are trainable. Conv = 2D Convolutional. MP = Max Pooling. DwD = Dense with Dropout. **b)** Cartoon representation of MBM with automated analysis for analysis of multiple cell types. Classification now requires a neural network.