`

Deep learning-based aberration compensation improves contrast and resolution in fluorescence
microscopy

Min Guo[1,2, *], Yicong Wu[2, 3, ^], Yijun Su[2, 3, %], Shuhao Qian[1], Eric Krueger[2, %], Ryan Christensen[2, %], Grant
Kroeschell[2, %], Johnny Bui[2, %], Matthew Chaw[2, %], Lixia Zhang[3], Jiamin Liu[3], Xuekai Hou[1], Xiaofei Han[2],
Xuefei Ma[4], Alexander Zhovmer[4], Christian Combs[5], Mark Moyle[6], Eviatar Yemini[7], Huafeng Liu[1], Zhiyi
Liu[1], Patrick La Riviere[8, 9], Daniel Colón-Ramos[9, 10], Hari Shroff[2, 3, 9, %]

1.  Current address: State Key Laboratory of Modern Optical Instrumentation, College of Optical
Science and Engineering, Zhejiang University, Hangzhou, China
2.  Laboratory of High Resolution Optical Imaging, National Institute of Biomedical Imaging and
Bioengineering, National Institutes of Health, Bethesda, Maryland, USA
3.  Advanced Imaging and Microscopy Resource, National Institutes of Health, Bethesda, Maryland,
USA
4.  Center for Biologics Evaluation and Research, U.S. Food and Drug Administration, Silver Spring,
MD, USA
5.  NHLBI Light Microscopy Facility, National Institutes of Health, Bethesda, MD, USA
6.  Department of Biology, Brigham Young University-Idaho, Rexburg, ID, USA
7.  Department of Neurobiology, UMass Chan Medical School, Worcester, MA
8.  Department of Radiology, University of Chicago, Chicago, IL, USA
9.  MBL Fellows Program, Marine Biological Laboratory, Woods Hole, MA, USA
10. Wu Tsai Institute, Department of Neuroscience and Department of Cell Biology, Yale University
School of Medicine, New Haven, CT, USA

^ Current address: Nanodelivery Systems and Devices Branch, Cancer Imaging Program, Division of
Cancer Treatment and Diagnosis, National Cancer Institute, National Institutes of Health, Rockville,
MD, USA

% Current address: Janelia Research Campus, Howard Hughes Medical Institute (HHMI), Ashburn,
VA, USA

* Correspondence to guom@zju.edu.cn

**Abstract**

Optical aberrations hinder fluorescence microscopy of thick samples, reducing image signal, contrast,
and resolution. Here we introduce a deep learning-based strategy for aberration compensation,
improving image quality without slowing image acquisition, applying additional dose, or introducing
more optics into the imaging path. Our method (i) introduces synthetic aberrations to images acquired
on the shallow side of image stacks, making them resemble those acquired deeper into the volume and
(ii) trains neural networks to reverse the effect of these aberrations. We use simulations to show that
applying the trained 'de-aberration' networks outperforms alternative methods, and subsequently apply
the networks to diverse datasets captured with confocal, light-sheet, multi-photon, and super-resolution
microscopy. In all cases, the improved quality of the restored data facilitates qualitative image
inspection and improves downstream image quantitation, including orientational analysis of blood
vessels in mouse tissue and improved membrane and nuclear segmentation in *C. elegans* embryos.

`

## Introduction

Fluorescence microscopes offer diffraction-limited imaging only when optical aberrations are absent. Such aberrations can arise due to optical path length differences introduced anywhere in the imaging path, including from instrument misalignment, optical imperfections, or differences in refractive index between the heterogenous and refractile sample, immersion media, or objective immersion oil. Sample-induced optical aberrations usually dominate and are often the reason that three-dimensional (3D) fluorescence image volumes show obvious deterioration in image signal-to-noise ratio (SNR), contrast, and resolution deeper into the image volume.

One method of compensating for these aberrations is via adaptive optics (AO[1,2]), a broad class of techniques that measure the aberrated wavefront and subsequently apply an equal and opposite 'corrective' wavefront, restoring diffraction-limited[3] or even super-resolution[4] imaging throughout the image volume. Once the aberrated wavefront is determined, an adaptive element such as a deformable mirror or spatial light modulator is used to apply the correction. Although these methods are effective, the process of determining the wavefront typically slows acquisition and/or applies more illumination dose than imaging without AO. From a practical perspective, implementing AO is nontrivial and adds considerable expense to the underlying microscope. Thus, AO remains the province of relatively few labs, and there is a need for new methods that can reverse the effects of optical aberrations without sacrificing temporal resolution, imparting more dose to the sample, or adding additional hardware to the microscope.

Deep learning approaches can computationally reverse image degradation, and have been used successfully in denoising[5,6], deconvolution[7,8], and super-resolution applications[9,10]. By incorporating information about the underlying object, such methods can also learn to predict the wavefront associated with aberrated images[11-13]. With sufficient training data (matched pairs of diffraction-limited and aberrated data), we reasoned that a neural network ought to be able to directly predict the diffraction-limited image from the aberrated image. The challenge then becomes accumulating appropriate training data, which would ideally be obtained without relying on AO.

Here we address this problem by (i) introducing synthetic aberrations to easily obtained near-diffraction limited data so that they resemble aberrated data and (ii) training neural networks to reverse the effect of these aberrations. We use simulations to show that application of our 'content-aware' approach outperforms other image restoration methods, including deconvolution with the known aberrated point spread function (PSF). We then apply our techniques to diverse volumetric data captured with confocal, light-sheet, multi-photon, and super-resolution microscopes, finding that in all cases, resolution and contrast are substantially improved over the raw data. In addition to facilitating biological inspection, the restored data also enhanced quantitative investigation, including orientational analysis of blood vessels in mouse tissue and improved accuracy of membrane and nuclear segmentation in *C. elegans* embryos.

## Results

*Compensating for aberrations with deep learning*

First, we intentionally synthetically aberrate the images acquired by fluorescence microscopes given knowledge of the physics of image formation[14,15] (**Fig. 1, Methods, Supplementary Note 1**). Aberrations are chosen so that the aberrated images resemble those acquired deeper into the sample,

`

where aberrations are more pronounced. The key insight of our approach is that the 'shallow' images on
the 'near side' of the three-dimensional fluorescence volume are usually near-diffraction-limited and
thus provide ground truth data that can be used to train a network to reverse the effect of the
synthetically introduced aberrations. The trained neural network model (termed 'DeAbe') can then be
used to reverse depth-dependent blurring on data unseen by the network, effectively mitigating the
effect of aberrations without recourse to AO.
To benchmark our method, we began by simulating 3D phantoms consisting of randomly
oriented and positioned dots, lines, spheres, circles, and spherical shells.  We then degraded these
structures by adding random aberrations and noise and evaluated the extent to which DeAbe could
reverse the degradation (**Fig. 1b, Supplementary Figs. 1-5**). Visual assessments in lateral (**Fig. 1c, d,**
**Supplementary Video 1**) and axial (**Fig. 1e, Supplementary Video 2**) views, as well as quantitative
comparisons (**Fig. 1f**) demonstrated that the DeAbe model outperformed blind deconvolution[16],
Richardson-Lucy deconvolution with an ideal point spread function (PSF), and even Richardson-Lucy
deconvolution with the aberrated PSF (known in these simulations, but unknown in general). We
attribute the superior performance of DeAbe to its ability to learn a sample-specific prior, thereby better
conditioning its solution relative to Richardson-Lucy deconvolution.
Importantly, simulations allowed us to further characterize DeAbe, offering insight into the
regimes in which the method excels and where performance suffers. First, we found optimal
performance when aberration magnitudes in the training data match the aberration magnitude in the
test data (**Supplementary Fig. 1**). Over the conditions we tested, the model improved images
contaminated with root mean square (RMS) wavefront distortion exceeding four radians (the highest
value we tested), although performance degrades as wavefront distortion increases. Second, although
we performed tests with training data containing up to the 7$^{th}$ Zernike order, the improvement offered
past order four (the value used in this work) is negligible (**Supplementary Fig. 2**). Third, DeAbe trained
on a mixture of Zernike basis functions also provides notable improvement on images corrupted solely
by individual Zernike functions (**Supplementary Fig. 3**), although dedicated models trained to correct
specific Zernike modes are better if these modes are known in advance (**Supplementary Fig. 4**). Finally,
although DeAbe's performance suffers in the presence of noise, it still offers noticeable visual and
quantitative improvements in image quality for SNR above ~5 (**Supplementary Fig. 5**).
114
115 *Computational aberration compensation improves image quality on diverse volumetric data*
116
We subsequently applied DeAbe to experimental data acquired with different microscope
modalities, in each case training models on images derived from the shallow side of image volumes (**Fig.**
**2, Supplementary Fig. 6, Supplementary Table 1**). First, we imaged live *C. elegans* embryos expressing a
pan-nuclear GFP-histone marker with inverted selective plane illumination microscopy (iSPIM)[17,18],
finding that the raw image data displayed progressive loss of contrast and resolution as a function of
increasing depth, making it difficult or impossible to discern subnuclear structure (or even individual
nuclei) at deeper imaging planes (**Fig. 2a, i, Supplementary Video 3**). By contrast, the DeAbe prediction
restored these structures, also improving axial views (**Fig. 2a, iii**). Richardson-Lucy deconvolution also
offered some improvement in image quality, albeit not to the extent of the DeAbe prediction, while also
undesirably amplifying noise (**Fig. 2a, ii**). Second, we used spinning-disk confocal microscopy to image
thicker adult *C. elegans* expressing the multicolor NeuroPAL transgene[19], used for resolving neuronal
identities. Depth-dependent image degradation produced raw images with dim or diffuse nuclear signal

`

in each color channel. The DeAbe prediction improved SNR dramatically (**Supplementary Fig. 7,**
**Supplementary Video 4**), which we suspect may prove useful in improving the accuracy of neuronal
identification. Third, we applied DeAbe to images of NK-92 cells stained with Alexa Fluor 555 wheat
germ agglutinin and embedded in collagen matrices, acquired with instant SIM[20], a super-resolution
imaging technique (**Fig. 2b-d**, **Supplementary Fig. 8, Supplementary Video 5**). Post deconvolution, the
DeAbe prediction better resolved clusters of membrane-bound glycoproteins, intracellular vesicles, and
membranes ('DeAbe+', **Fig. 2c, d**) than the raw (or deconvolved raw, **Supplementary Fig. 8**) data,
especially near the limits of the 45 μm thick imaging volume. Fourth, we used two-photon microscopy to
image live murine cardiac tissue expressing Tomm20-GFP, marking the outer mitochondrial membrane
(**Fig. 2e**). Although mitochondrial boundaries were evident in the raw data 20 μm into the volume,
aberrations caused a progressive loss in resolution that hindered visualization of subcellular structure at
greater depths (**Fig. 2e, f**). The DeAbe prediction restored resolution throughout the 150 μm thick
volume (**Fig. 2f, Supplementary Fig. 9, Supplementary Video 6**), unlike Richardson-Lucy deconvolution
(**Fig. 2f**) which amplified noise without restoring the mitochondria. The DeAbe prediction similarly
improved contrast and resolution when applied to volumes of fixed mouse liver stained with membrane
labeled tdTomato, imaged with two-photon microscopy (**Supplementary Video 7**).
Next, we applied DeAbe to samples ~10,000-fold larger in volumetric extent (**Fig. 3a,**
**Supplementary Video 8**).  We fixed and CLARITY[21]-cleared E11.5 mouse embryos immunostained for
neurons (Alexa Fluor TuJ1) and blood vessels (Alexa Fluor 594) and imaged them with low magnification
confocal microscopy. Although tissue clearing nominally produces a sample with the same refractive
index everywhere, we still observed pronounced depth-dependent degradation from the 'near' to 'far'
side of the embryo, including in intensity (likely due to photobleaching during the acquisition) and
resolution. We were able to largely reverse this deterioration by digitally compensating for
photobleaching[22] (**Methods**), applying DeAbe, and finally deconvolving the data (**Fig. 3b, Supplementary**
**Fig. 10**). While the improvement in image quality was particularly striking in axial views (**Fig. 3b**),
restorations also improved the appearance of fibrillar structures in lateral views, in both channels,
throughout the volume (e.g., the vicinity of the vagus nerve and its associated nerve roots, **Fig. 3c, d**).
We further investigated this qualitative impression by using automated tools[23,24] to
quantitatively assess the mean 3D orientation and directional variance (a measure of the spread in
angular orientation) at each voxel in the blood vessel channel (**Fig. 3e-g**, **Supplementary Figs. 11, 12,**
**Supplementary Video 9**). The DeAbe restoration resulted in cleaner separation between vessels, which
aided voxel-wise quantification of these metrics even in dense regions containing many crisscrossing
vessels (**Fig. 3e, Supplementary Video 9**). In deeper regions of the volume (**Fig. 3f**), the DeAbe results
produced narrower angular histogram distributions of vessels than the noisy raw data (**Fig. 3f**). The
improvement in quantification was also reflected in directional variance analysis. For example, when
visually inspecting different regions of interest (ROI) with differential vessel alignment (**Fig. 3g**,
comparing vicinity of aortic arches, (ROI 1), to diencephalon, (ROI 2)) we observed a greater difference
in mean directional variance when using the DeAbe reconstruction vs. the raw data (**Supplementary Fig.**
**12**).

*Incorporating DeAbe in multi-step restoration further enhances resolution and contrast in 4D imaging*
*applications*

`

Given the performance of DeAbe thus far, we wondered if we could further boost image quality
by combining DeAbe with additional networks designed to enhance spatial resolution. To test this
possibility, we acquired dual-view light sheet microscopy (diSPIM[25,26]) volumetric time-lapse ('4D')
recordings of *C. elegans* embryos expressing labels marking cell membranes and nuclei, and then passed
the raw single-view data through three networks designed to sequentially compensate for aberrations
(i.e., DeAbe), deconvolve the resulting predictions ('DL Decon'), and improve resolution isotropy[5] ('DL
Iso', **Fig. 4a-d, Supplementary Figs. 13-16**). As expected, (**Fig. 4a**), the raw data showed increasing
depth-dependent degradation in resolution and contrast, which confounded our ability to discern
distinct nuclei or cell boundaries on the 'far' side of the volume. In comparison, the multi-step
procedure offered striking improvements in resolution and contrast in both nuclear and membrane
channels, largely alleviating the degradation (**Fig. 4a, b, Supplementary Figs. 14, 15, Supplementary
Video 10**). Ablation experiments in which one or more of the networks were removed produced inferior
results, further substantiating our hypothesis that the gains in image quality benefited from applying all
three networks (**Supplementary Fig. 17**). In the membrane channel, the multi-step restoration enabled
us to automatically segment cell boundaries more accurately than in the raw data and further refine the
segmentations manually up to 421 cells (**Fig. 4c, Supplementary Fig. 16, Supplementary Video 11**),
exceeding previous efforts limited to the 350-cell stage[27]. The automatic segmentation additionally
provided a cell count closer to manual ground truth[28] than either the raw data or a subset of the
networks (**Fig. 4d**).
Next, we explored replacing the final network (DL Iso) with a network designed to further
enhance resolution based on ground truth acquired with expansion microscopy[9,29] ('DL Expan',
**Supplementary Fig. 13b**). After verifying that DL Expan improved resolution more than 2-fold on data
unseen by the model (**Supplementary Fig. 18**), we applied the new multi-step restoration method to *C.
elegans* embryos expressing a GFP-membrane marker labeling head neurons and gut cells (**Fig. 4e**).
Compared to the raw data, the enhanced resolution offered by the deep learning prediction better
resolved closely spaced membranes within and between cells (**Fig 4f-h, Supplementary Figs. 19, 20**).
This capability proved especially useful when tracking the development of neurites projecting in the
nerve ring, a neuropil that constitutes the brain of the animal, and which is composed of hundreds of
tightly packed interwoven neurites. While the position of the neurites within the neuropil determines
circuit identity and connectivity, the sequence of events leading to its innervation has not been
described because of limits in resolving these structures. We focused our analyses on the closely
positioned neurons AIY and SMDD, which we identified based on morphology by comparison to labeled
images in ref.[30] and ref.[31].  SMDD is a central pioneering neuron in the nematode brain[31-33], while its
sister cell AIY[30] is a first layer interneuron[34] involved in thermotaxis and locomotion[35]. Observing both
neurons over our 120-minute recording, we found that SMDD's neurites grew out first, followed by AIY's
neurite. AIY's neurite entered the nerve ring after SMDD, consistent with the SMDD's role as a pioneer
neuron (**Fig 4i, Supplementary Video 12**). Such developmental dynamics were difficult or impossible to
observe in the raw data (**Supplementary Fig. 21**), or joint deconvolutions of the dual-view data due to
artifacts resulting from motion between the two views (**Supplementary Fig. 22**). Finally, to illustrate that
these gains in image quality can be extended to a different label imaged in a different microscope, we
also restored images of nuclei labeled with a GFP histone marker and acquired with high NA diSPIM[18],
finding similarly dramatic improvements in contrast and resolution (**Supplementary Fig. 23,
Supplementary Videos 13, 14)**.
215

`

**Discussion**

As we show on diverse microscopes and samples, DeAbe can compensate for optical aberrations without recourse to AO, improving SNR, contrast, and resolution in fluorescence microscopy volumes. We suspect this capability will be useful for most labs, which lack access to sophisticated AO setups but still wish to improve the quality of imaging volumes acquired using existing hardware. Besides improving the qualitative appearance of images (**Fig. 1-4**), which facilitates inspection of biological features deep within imaging volumes, DeAbe also quantitatively improves downstream image analysis. We highlight this capability by refining vessel segmentation in large, cleared tissue samples (**Fig. 3e-g**) and in enhancing the segmentation of densely packed nuclei and membranes in *C. elegans* embryos (**Fig. 4**). The latter capability may prove particularly useful in the creation or extension of 4D morphological atlases[27], which depend on high quality image data.

Several caveats are worth noting in the context of current limitations and with an eye towards future applications. First, the performance of DeAbe depends critically on the quality of the training data, and specifically on the assumption that fluorescently labeled structures are similar throughout the image volume. While this assumption was met for the samples in this work, we encourage caution when applying DeAbe on highly heterogenous specimens. Second, although here we mainly trained on semi-synthetic data (**Fig. 2-4**), it would also be worth investigating how well the training derived from fully synthetic data[7] (**Fig. 1**) generalizes to experimental data. Such an approach might prove useful in ameliorating system aberrations introduced by microscope hardware. Third, we focused here on correcting depth-dependent aberrations, in which the training data was corrupted by a constant aberration in each image plane. A useful future direction would be to extend our approach to explicitly account for laterally varying aberrations, as such aberrations are problematic particularly for large specimens. Finally, although we used a mixture of random low-order aberrations to train our model, enhanced performance is likely if aberrations specific to the sample (or instrument) can be inferred and used in the training procedure (**Supplementary Fig. 4**).

**Author Contributions**

Conceived project and directed research: H.S. Implemented DeAbe framework: M.G. Designed simulations: M.G., Y.W., H.S. Wrote software: M.G., Y.W., J.L., X.Han, S.Q., Z.L. Designed experiments: M.G., Y.S., R.C., X.M, A.Z., C.C., M.M., E.Y., D.C-R., H.S. Prepared samples: M.G., Y.S., E.K., R.C., G.K., J.B., M.C., L.Z., X.M., A.Z., C.C., M.M., E.Y. Performed experiments: M.G., Y.S., E.K., R.C., G.K., J.B., M.C., X.Han, X.M., A.Z., C.C., M.M., E.Y. Designed and performed segmentation analysis: M.G., Y.S., J.L., X. Hou, S.Q., Z.L. All authors examined and analyzed data. Wrote manuscript: M.G. and H.S., with advice from all authors. Provided biological insight and advice: R.C., X.M., A.Z., C.C., M.M., E.Y., D.C-R. Supervised research: E.Y., H.L, Z.L., P.L-R., D.C-R., H.S.

**Acknowledgments**

We thank O.H. for supporting the NeuroPAL work in his lab and allowing us to use the data generated therein for this paper, Robert Weigert (NCI) for the gift of the fixed mouse liver sample, Nikolaj Reiser for helpful discussions, and Courtney Johnson and Xuesong Li for their comments on the manuscript. This research was supported by the intramural research programs of the National Institute of Biomedical Imaging and Bioengineering and the National Heart, Lung, and Blood Institute within the National

`

Institutes of Health (NIH). This work was supported by the Howard Hughes Medical Institute (HHMI).
This article is subject to HHMI's Open Access to Publications policy. HHMI laboratory heads have
previously granted a non-exclusive CC BY 4.0 license to the public and a sub-licensable license to HHMI
in their research articles. Pursuant to those licenses, the author-accepted manuscript of this article can
be made freely available under a CC BY 4.0 license immediately upon publication. This research is
funded in part by the Gordon and Betty Moore Foundation. We thank the Office of Data Science
Strategy, NIH, for providing a seed grant enabling us to test and validate the initial deep learning
frameworks using cloud-based computational resources. H.S., P.L.R. and D.C.-R. acknowledge the
Whitman and Fellows program at MBL for providing funding and space for discussions valuable to this
work. Z.L. acknowledges the funding support from Natural Science Foundation of Zhejiang Province
(LR20F050001). H.L. acknowledges the funding support from the National Key Research and
Development Program of China (2020AAA0109502), the National Natural Science Foundation of China
(U1809204) and the Talent Program of Zhejiang Province (2021R51004). E.Y. acknowledges support
from the Esther A. and Joseph Klingenstein Fund, the Simons Foundation, and the Hypothesis Fund. X.M.
and A.Z. were supported by Intramural FDA funding.

**Methods**

278    *Deep learning-based de-aberration model*

Building a de-aberration model (DeAbe) requires appropriate training data and the use of a
neural network. First, based on the physics of image formation, we derived forward imaging models that
allowed us to synthetically aberrate the data produced for multiple systems, including wide field, light
sheet, confocal, two photon, and super-resolution structured illumination microscopes (**Supplementary
Note 1**). Second, we extracted subvolumes from the shallow side of the experimentally acquired image
stacks, using these data as ground truth. Third, based on the forward imaging models, we synthetically
added aberrations to the ground truth images so that they resembled aberrated data present deeper
within the image stacks. Together, the paired ground truth data and associated synthetically degraded
data constitute training pairs. Fourth, we used these training pairs in conjunction with our 3D RCAN
network[9] to train a DeAbe model to reverse the effect of synthetic aberration. Finally, we applied the
trained network to reduce the effects of aberrations in experimentally acquired image volumes unseen
by the network.

We define the 'shallow side' of an image stack by the planes nearest to the detection objective,
which are typically contaminated with least aberration and thus offer the best image quality. We then
selected subvolumes on the shallow side ('shallow subvolumes') by visually inspecting image quality in
real and Fourier space (**Supplementary Fig. 6, Supplementary Table 1**). We extracted shallow
subvolumes from image stacks by manually cropping with ImageJ when image size and content differed
substantially across a given specimen type, or automatically with customized ImageJ macros when
considering specimens with more stereotyped image size and content (e.g., as for time-lapse image
volumes). For the cleared mouse embryo images (**Fig. 3**), the shallow subvolumes were further divided
into smaller subvolumes (~80 MB/volume) due to their large volume size in raw data (**Supplementary
Table 1**).

`

301       As described in **Supplementary Note 1**, we expressed the aberrated wavefront $\phi(r, \theta)$ at the
302    back focal plane of the objective using Zernike basis functions $\phi_m(r, \theta)$ and associated Zernike
303    coefficients $c_m$

$$\phi(r, \theta) = \sum_{m=0}^{M} c_m \phi_m(r, \theta), \tag{1}$$

with $M$ the maximum Zernike index chosen in our aberration.
We generated synthetic aberrations by using semi-randomly generated Zernike coefficients (**Fig.**
**1a**). We used the ANSI convention[36] when indexing the Zernike coefficients, customizing aberrations by
using different Zernike coefficients for different datasets acquired from different microscopes. For all
experimental datasets, we added aberrations up to the 4[th] Zernike order (i.e., $M$ = 14), except for piston
and tilt components (Z = 0, 1, 2). The amplitudes of the Zernike coefficients were randomly generated,
but subject to pre-defined bounds. We initially set an upper bound of 0.5 rad for all Zernike coefficients,
then added an additional 1 rad for defocus (Z = 4) and spherical (Z = 12) components to mimic the more
severe contamination caused by defocus and spherical aberrations commonly encountered in
experimental datasets, i.e:

$$\begin{cases} c_z = 0, & for\ Z = 0, 1, 2 \\ |c_z| \leq 1.5, & for\ Z = 4, 12 \\ |c_z| \leq 0.5, & otherwise\ for\ Z \leq M, \end{cases} \tag{2}$$

with $M$ = 14 for all experimental datasets.
For each shallow side subvolume, 10 independent sets of aberrations were generated and used
for synthetic degradation, thereby augmenting the data 10-fold. Processing was performed with custom
MATLAB code (MathWorks, R2022b), with further details provided in the *Code availability* section.
We employed 3D RCAN (https://github.com/AiviaCommunity/3D-RCAN), appropriate for 3D
image volumes, for generating the DeAbe model based on the training data pairs. We trained individual
DeAbe models for each microscope and each sample type. For training, we set the number of epochs to
200; the number of steps per epoch to 400; the training patch size to 64 × 64 × 64; the number of
residual blocks to 5; the number of residual groups to 5; and the number of channels to 32. The training
was performed within Python 3.7.0 on a Windows 10 workstation (CPU: Intel Xeon, Platinum 8369B, two
processors; RAM: 256 GB; GPU: NVIDIA GeForce RTX 3090 with 24 GB memory). More details on
datasets and training parameters are listed in **Supplementary Table 1**.

***Multi-step image restoration with deep learning***
The multi-step image restoration pipeline combines the DeAbe model with two additional
networks to progressively improve image resolution and contrast: (1) the DeAbe model to reverse
degradation from aberrations ("DL DeAbe"); (2) a deconvolution network designed to mimic the image
quality improvement afforded by multiview imaging ("DL Decon", see the section *Deep learning-based*
*deconvolution*); (3) an axial resolution enhancement network to improve resolution isotropy ("DL Iso",
see the section *Deep learning-based axial resolution enhancement*); or a network designed to predict
the improved resolution provided by expanded samples ("DL Expan", see the section *Deep learning-*
*based expansion*).

***Deep learning-based deconvolution***

`

As for our previous attempts at deep-learning based multiview deconvolution[8], we used a
single-view image volume as input, and attempted to restore image resolution and contrast that
approximated the result from multiview joint deconvolution. The training data were acquired by dual-
view light sheet microscopy[25], either a 'symmetric' diSPIM equipped with 0.8/0.8 NA objectives[26] (**Fig 4e-**
**i, Supplementary Figs. 17-22**) or a higher NA 'asymmetric' diSPIM equipped with 1.1 / 0.67 NA
objectives[18] (**Fig 4a-d, Supplementary Figs. 14-16, 23**). First, raw images were de-aberrated with the
DeAbe model. Then de-aberrated images from the two views were jointly deconvolved to achieve
reconstructions with near isotropic spatial resolution and good image quality throughout the
reconstruction. With training data consisting of the single-view de-aberrated images as input and the
jointly deconvolved images as ground truth, we then used another 3D RCAN for the deconvolution
model (DL Decon). For all datasets, the number of epochs for training was 200; the number of steps per
epoch was 400; the training patch size was $64 \times 64 \times 64$; the number of residual blocks was 5; the
number of residual groups was 5; and the number of channels was 32. The training was performed
within Python 3.7.0 on a Windows 10 workstation (CPU: Intel Xeon, Platinum 8168, two processors;
RAM: 512 GB; GPU: Nvidia Quadro RTX6000 with 24 GB memory). We note that although training DL
Decon required dual-view image volumes, applying DL Decon needs only single-view image volumes
acquired from single-view light sheet microscopy (iSPIM).
358
***Deep learning-based axial resolution enhancement***
The images predicted by the DL Decon model were not perfectly isotropic, i.e., the axial
resolution (although improved over the raw input images) is worse than the lateral resolution. Thus, for
some experiments we used an additional network to enhance axial resolution (DL Iso, **Fig. 4a, b,**
**Supplementary Figs. 14-17, Supplementary Videos 10, 11**). CARE[5] software
(https://github.com/CSBDeep/CSBDeep) was employed to train the a 'DL Iso' model based on the
predictions derived from serially applying the DeAbe and Decon models to raw input images. We used
100 3D volumes, each spanning $360 \times 480 \times 310$ voxels, for training data. Training was performed on the
xy planes (lateral views), using a 2D PSF (consisting of a point blurred with a 1D Gaussian function, sigma
= 2.5 pixels along the y dimension) an axial downsampling factor of 6, and a patch size of $64 \times 64$ to
create training pairs. The training was performed within Python 3.7.0 on a Windows 10 workstation
(CPU: Intel Xeon, Platinum 8168, two processors; RAM: 512 GB; GPU: Nvidia Quadro RTX6000 with 24
GB memory).
372
***Deep learning-based expansion***
As an alternative to DL Iso, we also trained a model to improve the resolution based on data
acquired with expansion microscopy (DL Expan). First, physically expanded samples (**Supplementary Fig.**
**18**) were imaged on the symmetric 0.8 NA diSPIM. Second, dual-view raw images were jointly
deconvolved and used as ground truth images. Third, the ground truth images were synthetically
degraded to resemble low-resolution conventional images acquired on the diSPIM, following our
previous procedure[9]. Last, the 3D RCAN network was employed to train the DL Expan model based on
the training data (i.e., synthetically degraded and ground truth pairs).
For the worm embryo data with DAPI labeled nuclei (**Supplementary Fig. 23**), dual-view raw
image volumes from 15 expanded worm embryos were acquired and jointly deconvolved to produce 15
high-resolution image volumes. These 15 volumes were then synthetically degraded to generate low-
resolution images. For the worm embryo data with TTX3B neurites labeled (**Fig 4e-i**, **Supplementary**

**Figs. 18-21**), dual view image volumes from 71 expanded worm embryos were acquired and manually
cropped to select regions containing TTX3B neurites (this was necessary given the sparsely labeled
neurites present in the raw images). Cropped images were jointly deconvolved to produce 71 high-
resolution image volumes. These 71 volumes were then synthetically degraded to generate synthetic
low-resolution image data. For each dataset, the low-resolution and high-resolution paired volumes
were then used to train the 3D RCAN based DL Expan model. The number of epochs for training was set
to 300; the number of steps per epoch to 400; the training patch size to $64 \times 64 \times 64$; the number of
residual blocks to 5; the number of residual groups to 5; and the number of channels to 32. The training
was performed within Python 3.7.0 on a Windows 10 workstation (CPU: Intel Xeon, Platinum 8369B, two
processors; RAM: 256 GB; GPU: NVIDIA GeForce RTX 3090 with 24 GB memory).
395

396    ***Simulations on phantom objects***
397            To evaluate the quality and performance of our DeAbe model, we generated 3D phantom
objects consisting of five types of structures in MATLAB (Mathworks, R2022b, with the Image Processing
Toolbox): dots, lines, circles, spheres, and spherical shells[22]. Phantoms were randomly oriented and
located in a volume of $256 \times 256 \times 256$ voxels, with voxel size $0.13 \times 0.13 \times 0.13$ μm³. We simulated the
blurring introduced by light sheet microscopy (**Supplementary Note 1**) by convolving the phantom with
an ideal, noise-free PSF resembling that of our light sheet system (with 1.1 NA water dipping objective,
detection wavelength of 0.532 μm and an illumination light sheet thickness of 2 μm). Aberrated data
was generated by altering the ideal PSF according to the synthetic aberration procedure described
above.
406            To create synthetic aberrations, we adopted Equation (1) and generated Zernike coefficients
407    semi-randomly in MATLAB, with each Zernike coefficient $c_m$ subject to a pre-defined upper bound $T_m$:

$$|c_m| \leq T_m, \ for \ m \leq M, \tag{3}$$

409    with $m$ the Zernike index following the ANSI convention and $M$ the maximum Zernike index.
410    We omitted piston and tilt components ($m$ = 0, 1, 2) and weighted lower order Zernike components
411    (Defocus $m$ =4, astigmatism $m$=3,5, and spherical $m$=12) more as these aberrations are commonly
412    observed in real samples:

$$T_m = \begin{cases} 0, & for \ m = 0, 1, 2 \\ 1.5, & for \ m = 3, 4, 5, 12 \\ 0.5, & otherwise \ for \ m \leq M, \end{cases} \tag{4}$$

414    with $M$ defined based on the desired Zernike order:

$$M = \begin{cases} 9, & for \ Zernike \ order \ of \ 3 \\ 14, & for \ Zernike \ order \ of \ 4 \\ 20, & for \ Zernike \ order \ of \ 5 \\ 27, & for \ Zernike \ order \ of \ 6 \\ 35, & for \ Zernike \ order \ of \ 7 \ . \end{cases} \tag{5}$$

416    For **Supplementary Fig. 2**, we varied $M$ to explore the effect of different Zernike orders on de-aberration
417    performance by setting $M = 9, 14, 20, 27,$ and $35$ corresponding to Zernike orders 3-7. For all other
418    simulations, we set $M = 14$.
419    The Root Mean Square (RMS) wavefront distortion of an aberration with Zernike coefficients $c_m$ ($m =$
420    $3, 4, 5, \dots, M$) is:

`

$$421 \qquad RMS_c = \sqrt{\sum_{m=3}^{M} c_m{}^2}. \qquad (6)$$

422    The RMS wavefront distortion for aberrations defined by upper bounds $T_m$ $(m = 3, 4, 5, …, M)$ is:

$$423 \qquad RMS_T = \sqrt{\sum_{m=3}^{M} T_m{}^2}. \qquad (7)$$

424    To create training data, we synthetically aberrated phantoms with two types of aberrations:

425    1) a random mixture of aberrations containing different Zernike components, with the
426    amplitude of the aberrations subject to upper bounds. This type of aberrations was first generated with
427    a set of initial Zernike coefficients $c_m$ based on Equations (3-5), and then rescaled to a maximum RMS of
428    $\Omega$ wavefront distortion (e.g., $\Omega = 1, 2,$ or 4 rad) to obtain the final Zernike coefficients $c_{m-final}$:

$$429 \qquad c_{m-final} = \frac{\Omega}{RMS_T} c_m, \; for \; m \le M. \qquad (8)$$

430    These aberrated training data were used to train the general DeAbe models (i.e., all but the model
431    trained to counter the defocus mode specifically) used in all figures and videos showing simulated
432    phantoms.

433    2) a single aberration mode of defocus with amplitude subject to upper bounds, i.e., the upper
434    bounds of each Zernike coefficient were zeros except for the defocus mode ($m$ =4):

$$435 \qquad T_m = \begin{cases} 1.5, & for \; m = 4 \\ 0, & otherwise \; for \; m \le M. \end{cases} \qquad (9)$$

By replacing Equation (4) with Equation (9), we could generate the defocus aberration the same way as
for the first aberration type (1). These training data were only used to train the specific defocus DeAbe
model used in **Supplementary Fig 4**.

For each training session, we created 50 phantoms, each consisting of different random objects.
For each phantom, we generated 10 independent aberrated images with each image containing random
mixtures of aberrations (**Fig 1, Supplementary Figs 1-5, Supplementary Videos 1-2**) or only defocus
aberrations (**Supplementary Fig 4**), for a total of 500 training data pairs per session. We also added
Poisson noise to the aberrated images by defining the SNR as

$$444 \qquad SNR = \sqrt{S}, \qquad (10)$$

where $S$ is the signal defined by the average of all pixels with intensity above a threshold (here set as 1%
of the maximum intensity of the blurred objects in the noise-free image).

We employed 3D RCAN to train the DeAbe model based on simulated training data. We set the
number of epochs to 200; the number of steps per epoch to 400; the training patch size to 64 × 64 × 64;
the number of residual blocks to 5; the number of residual groups to 5; and the number of channels to
32. Training was performed with Python 3.7.0 on a Windows 10 workstation (CPU: Intel Xeon, Platinum
8369B, two processors; RAM: 256 GB; GPU: NVIDIA GeForce RTX 3090 with 24 GB memory).

To benchmark the performance of the DeAbe model, we created synthetic phantoms with three
types of aberrations:

1) a random mixture of aberrations containing different Zernike components, with the
amplitude of the aberrations subject to upper bounds. This type of aberration is the same used for
training the general DeAbe models and was generated following Equations (3-5) and (8). This aberration
mixture was used in **Fig 1, Supplementary Fig 2, and Supplementary Videos 1-2**.

2) a random mixture of aberrations containing different Zernike components, with the
amplitude of the aberrations fixed at a certain RMS value. This aberration mixture was first generated
with a set of initial Zernike coefficients $c_z$ based on Equations (3-5), and then rescaled to a fixed
amplitude with RMS $\Upsilon$ (e.g., $\Upsilon = 1, 2,$ or 4 rad) wavefront distortion to obtain the final Zernike
coefficients $c_{m-final}$:

$$c_{m-final} = \frac{\Upsilon}{RMS_c} c_m, \quad for \ m \leq M. \tag{11}$$

This aberration mixture was used for **Supplementary Figs 1,3,5**.
3) single aberration modes with a fixed RMS value, i.e., Zernike coefficients were set to zero
except for the desired aberration mode. The single aberration modes tested in the paper include
defocus ($m$=4), astigmatism ($m$=3,5), coma ($m$=7,8), trefoil ($m$=6,9), and spherical ($m$=12). If the RMS
wavefront distortion is defined as $\Upsilon$ (e.g., $\Upsilon = 1, 2,$ or 4 rad), each single aberration mode's Zernike
coefficients are:
Defocus: $c_4 = \Upsilon, \ otherwise \ c_m = 0 \ for \ m \leq M$
Astigmatism: $\sqrt{c_3{}^2 + c_5{}^2} = \Upsilon, \ otherwise \ c_m = 0 \ for \ m \leq M$
Coma: $\sqrt{c_7{}^2 + c_8{}^2} = \Upsilon, \ otherwise \ c_m = 0 \ for \ m \leq M$
Trefoil: $\sqrt{c_6{}^2 + c_9{}^2} = \Upsilon, \ otherwise \ c_m = 0 \ for \ m \leq M$
Spherical: $c_{12} = \Upsilon, \ otherwise \ c_m = 0 \ for \ m \leq M$
These aberrations were used to test the DeAbe performance on single aberration modes
(**Supplementary Figs. 3,4**).
477
478    For quantitative analysis, we used structural similarity index (SSIM) and peak signal-to-noise
479 ratio (PSNR) to evaluate the restored images provided by deep learning as well as by traditional
480 deconvolution. The SSIM and PSNR were calculated based on image volumes with MATLAB (Mathworks,
481 R2022b). Their mean value and standard deviation were computed from 100 simulations, each with
482 random object structures and input aberrations.
483

### Preprocessing, attenuation correction, traditional deconvolution, and multiview fusion

Raw images acquired with iSIM and light sheet imaging were preprocessed by subtracting a
uniform background with intensity equivalent to the average of 100 dark (no excitation light)
background images. When diSPIM was operated in stage scan mode, the images were also deskewed to
correct the distortion induced by stage-scan acquisition before further processing.
For the cleared mouse embryos imaged with confocal microscopy (**Fig 3, Supplementary Fig. 10,**
**Supplementary Video 8**), raw data was additionally preprocessed with intensity attenuation correction.
The attenuation correction was performed by multiplying the raw intensity values with an exponential
compensation factor:

$$I(z) = I_0(z)e^{\alpha z} \tag{11}$$

with $I_0(z)$ the raw intensity, $z$ the depth and $\alpha$ the attenuation factor. Here we set $\alpha = 0.01$.
For the comparison of DeAbe with traditional deconvolution, we implemented both Richardson-
Lucy (RL) deconvolution[37,38] (**Fig. 1c-f, Fig. 2 and Supplementary Figs. 8-10**) and blind deconvolution[16]
(**Fig. 1c-f**) on the raw aberrated images. For blind deconvolution, we used the MATLAB function
*deconvblind* with default settings (https://www.mathworks.com/help/images/ref/deconvblind.html).
For RL deconvolution, we adopted our previously developed deconvolution package[8]
(https://github.com/eguom/regDeconProject). In one synthetic dataset ('RL Decon 2', **Fig. 1c-f**), we used

an aberrated PSF that was generated as described in **Supplementary Note 1** and matched the
aberrations in the synthetic dataset; otherwise, we used an aberration-free ideal PSF for all other
datasets (**Fig. 1c-f, Fig. 2 and Supplementary Figs. 8-10**). Additionally, we also performed RL
deconvolution on several datasets after DeAbe processing (**Fig. 2b-d, Supplementary Fig. 8, 10**), setting
the number of iterations to 20 unless specified otherwise. All deconvolution was performed in MATLAB
(MathWorks, R2022b) on a Windows 10 workstation (CPU: Intel Xeon, Platinum 8369B, two processors;
RAM: 256 GB; GPU: NVIDIA GeForce RTX 3090 with 24 GB memory).
For data acquired by diSPIM, we performed multiview fusion on several datasets either for
generating DL training data (**Fig. 4**, **Supplementary Figs. 14, 15, 19-21, 23**) or for comparisons to the DL
Decon model (**Supplementary Figs. 17, 22**). The diSPIM data typically contain two view volumes,
referred to as View A and View B volumes. The multiview fusion process involves registration and joint
deconvolution to combine two views into a single volumetric image stack with improved resolution. The
registration first rotates View B by 90 degrees along the Y-axis to align View B's orientation with View A
and then maximizes the cross-correlation function between View A and View B with affine
transformations. After registration, View A and registered View B were deconvolved jointly using a
modified Richardson–Lucy deconvolution algorithm as previously described[25]. Multiview fusion was
achieved using custom software (https://github.com/eguom/diSPIMFusion) on a Windows 10
workstation (CPU: Intel Xeon, Platinum 8369B, two processors; RAM: 256 GB; GPU: NVIDIA GeForce RTX
3090 with 24 GB memory).
520
**Sample preparation and imaging**
*Live nematode embryos imaged with light sheet microscopy*
Nematode strains were kept at 20°C, and grown on NGM media plates seeded with *E. coli* OP50.
Strains used in this paper included BV514 (ujIS113 [*pie-1p::mCherry::H2B + unc-119(+); Pnhr-*
*82::mCherry::histone + unc-119(+)*]), OD58 (ltIs38 [*pie1p::GFP::PH(PLC1delta1) + unc-119(+)*]), DCR6268
(olaEx3632 [*pttx-3b::SL2::PHD::GFP:: unc-54 3' UTR + pelt-7::mCh::NLS::unc-54 3' UTR*]), and SLS164
( ltIS138[*pie-1p::GFP::PH(PLC1delta1) + unc-119(+)*]; ujIS113 [*pie-1p::mCherry::H2B + unc-119(+); Pnhr-*
*82::mCherry::histone + unc-119(+)*]). SLS164 was made by crossing together strains BV514 and OD58 and
may have unc-119(ed3) III in the background. Strains BV514 and OD58 were gifts from Dr. Zhirong Bao.
Nematode samples were prepared for diSPIM imaging as previously described[17,26,39]: gravid adult
hermaphrodites were picked into a watch glass with M9 buffer, adults were cut in half to liberate
embryos, and embryos were transferred onto a poly-L-lysine coated coverslip in a diSPIM imaging
chamber. For strain DCR6268 ((olaEx3632 [*pttx-3b::SL2::PHD::GFP:: unc-54 3' UTR + pelt-*
*7::mCh::NLS::unc-54 3' UTR*]), labeling neuron and gut cells), embryos were imaged once they reached
the bean stage of development using a fiber-coupled symmetric diSPIM (with 0.8NA/0.8NA
objectives)[26]. Volumes were captured once per minute over two hours in light sheet scan mode.  Each
volume comprised 50 slices, with a 1 μm step size and a total acquisition time per volume of ~1 second.
For strain SLS164 (labeling cell membrane and nuclei), embryos were imaged from the 2- or 4-cell stage
using a fiber-coupled asymmetric diSPIM (with 1.1NA/0.67NA objectives)[18]. Volumes were captured
once every 3 minutes over 450-minute duration in stage scan mode. Each volume comprised 70 slices,
with a 1.1 μm stage step size and a total acquisition time of ~1.4 s per volume. For strain BV514 (labeling
cell nuclei), embryos were imaged from the bean stage to hatching using the asymmetric diSPIM.
Volumes were captured every 5 minutes in stage scan mode. Each volume comprised 60 slices, with a
1.4 μm stage step size and a total acquisition time per volume of ~1.2 seconds. For strain OD58 (labeling

`

545 cell membranes), embryos were imaged from the 4- or 8-cell stage using a symmetric diSPIM. Volumes
546 were captured once every 3 minutes over a 450-minute period in light sheet scan mode. Each volume
547 comprised 45 slices, with a 1 μm step size and a total acquisition time per volume of ~0.9 seconds. For
548 all imaging, images were acquired using 488 nm excitation (for GFP labels) or 561 nm excitation (for
549 mCherry labels).
550
*Expanded nematode embryos*
*C. elegans* embryos from strain DCR6268 (labeling neurites and gut cells) were immobilized on
Poly-L-Lysine (PLL) coated glass bottom dishes, bleached, digested by yatalase, fixed, and expanded. The
procedure takes approximately 2 days, and is adapted from our published method[22].
First, glass bottom dishes were coated with PLL. PLL (Sigma, Cat# P5899) powder was
reconstituted in distilled water to 1mg/mL, aliquoted, and stored at -20°C. Prior to experiments, 30-50
μL of PLL was placed on the glass bottom dish (MatTek, Cat# P35G-1.5-14-C) and air dried at room
temperature (RT). Coated coverslips were usually prepared up to 1 day before pre-treatment of *C.
elegans* for expansion microscopy.
Second, embryos were digested, fixed, and stained with DAPI. Gravid adult *C. elegans* worms
were deposited in a petri dish in PBS buffer and cut with a surgical blade to release eggs. Eggs were
immobilized on a PLL coated glass bottom dish in PBS and could be processed immediately or stored at
25°C in M9 buffer until the embryos developed to the desired stage. Embryos were treated with a
bleaching mixture containing 1% sodium hypochlorite (Sigma, Cat# 425044) in 0.1M NaOH/water for 2-3
minutes, rinsed 3 times in PBS, digested in 50 mg/mL Yatalase in PBS (Takara Bio, Cat# T017) for 40
minutes at RT and rinsed 3 times with PBS. It was important to treat eggs with bleach only after
immobilization on the PLL surface, otherwise embryos tended to detach from the glass at later steps.
Digested embryos were fixed in 4% paraformaldehyde/PBS (Electron Microscopy Sciences, Cat#
RT15710) for 1 hour, then rinsed 3 times with PBS to remove fixative. Fixed embryos were
permeabilized in 0.1% Triton X-100/PBS (Sigma, Cat# 93443) for 1 hour at RT with 1 μL/mL of DAPI
(Thermo Fisher Scientific, Cat# D1306).
Optionally, GFP signal can be boosted by immunolabeling. Yatalase digested embryos were
permeabilized with staining buffer (0.1% Triton X-100/PBS) for 1 hour before immunolabeling. Embryos
were stained by an anti-GFP primary antibody (Abcam, Cat# ab290) in the staining buffer at 4°C
overnight at 1 μg/mL. After primary antibody labeling, embryos were washed 3 times (30 min intervals
between washes) in the staining buffer and labeled using donkey-anti-rabbit-biotin secondary antibody
(Jackson ImmunoResearch, Cat# 711-067-003) in the staining buffer at 4°C overnight at 1 μg/mL. After
secondary antibody labeling, the embryos were washed 3 times in the staining buffer (30 mins intervals
between washes) and labeled with Alexa Fluor 488 Streptavidin in the staining buffer at 4°C overnight at
2 μg/mL (Jackson ImmunoResearch, Cat# 016-540-084). Labeled embryos were washed 3 times in the
staining buffer (30 minutes between washes) before being processed for expansion microscopy.
Immunolabeling was only performed on the data shown in **Supplementary Fig 18a**.
Finally, embryos were expanded. Embryos were treated with 1 mM MA-NHS (Sigma, Cat#
730300) in PBS for 1 hour at RT. Samples were rinsed 3 times in PBS, and treated with monomer
solution, which was made up of acrylamide (Sigma, Cat# A9099), sodium acrylate (Santa Cruz
Biotechnology, Cat# 7446-81-3), N, N'-methylenebis(acrylamide) (Sigma, Cat# 146072) and 4-Hydroxy-
TEMPO (Sigma, Cat# 176141), diluted with PBS, with a final concentration of 10%, 19%, 0.1%, and 0.01%,
respectively. After the treatment for 1 hour at RT, the monomer solution was replaced by gelation

`

solution. The gelation solution shared the same reagents and concentrations as monomer solution, with
the addition of tetramethylethylenediamine (TEMED, Thermo Fisher Scientific, Cat# 17919, reaching a
final concentration of 0.2%) and ammonium persulfate (APS, Thermo Fisher Scientific, Cat# 17874,
reaching a final concentration of 0.2%). APS was added at last, and the fresh gelation solution was
immediately applied to the embryos sandwiched between the glass bottom dish and another coverslip
surface for 2 hours at RT. It was important to control the gelation speed with 4-hydroxy-TEMPO as
premature gelation can distort embryos and result in poor expansion quality. The polymerized embryo-
hydrogel hybrid was cut out by a razor blade and digested with 0.2 mg/mL Proteinase K (Thermo Fisher
Scientific, Cat# AM2548) in digestion buffer (0.5 M sodium chloride (Quality Biological, Cat # 351-036-
101); 0.8 M guanidine hydrochloride (Sigma, Cat# G9284); and 0.5% Triton X-100) at 45°C overnight.
Digested embryos were expanded ~3.3-3.7 fold in distilled water, exchanging the water every 30 min
until expansion was complete. Expanded samples were flipped over so that embryos were 'on top'
(suitable for diSPIM imaging), mounted on PLL coated #1.5 coverslips (VWR, Cat# 48393-241) and
secured in an imaging chamber filled with distilled water. Finally, samples were imaged using the
symmetric 0.8/0.8 NA diSPIM in stage scan mode. Depending on the orientation of embryos, ~200-300
planes were acquired for each embryo, with 1.414 μm stage step size and 20 ms per-plane exposure
time.
606
*Live nematode adults imaged with spinning disk confocal microscopy*
*C. elegans* strain OH15500 (*otIs669[NeuroPAL]; otIs672[panneuronal::GCaMP6s]*) were raised at
20°C and grown on NGM media plates seeded with OP50 *E. coli* . Young adult worms (with 2 or less
visible eggs in their uterus) were picked and immobilized inside a microfluidic chip as previously
described[19]. Worms were imaged by a spinning disk confocal microscope (Nikon, Ti-e) equipped with a
60×/1.2 NA water objective (Nikon, CFI Plan Apochromat VC 60XC WI), a confocal scan unit (Yokagawa,
CSU-X1) and an electron multiplying CCD (EM-CCD, Andor, iXon Ultra 897). Four excitation lasers (405
nm, 488 nm, 561 nm, and 640 nm) were used for illumination, in conjunction with emission filters
spanning 420-470 nm, 500-545 nm, 570-650 nm, and 660-800 nm bandwidths, respectively. The pixel
size was 0.27 μm in the XY dimension and each Z-stack volume comprised 21 slices for each color, with
1.5 μm step size. Each multicolor Z-stack volume was captured at a rate of just over 1 minute.
618
*Fixed WGA-labeled NK-92 samples*
NK-92 cells (ATCC®, CRL-2407™) were rinsed with 1× PBS, and fixed with 1 ml of 4%
paraformaldehyde in 1× PBS for 30 min at room temperature, rinsed in 1 ml of 1x PBS, and
permeabilized in 0.1% Triton X-100 in 1× PBS for 15 min. Next, samples were rinsed with 1× PBS, and
blocked with buffer containing 1% BSA (Fisher, Cat# BP9700100) in 1× PBS for 1 hour. Blocking buffer
was removed, and the samples were stained with 500 μl of 1x PBS with a 1:100 dilution of Alexa Fluor
555 labelled WGA (Invitrogen, Cat# W32464), 10 U/mL phalloidin-ATTO 647N conjugate (Millipore-
Sigma, Cat #65906), and 1:1000 dilution of Hoechst solution (Tocris, Cat#5117) for 1 h. Cells were
washed in 1× PBS three times. We mounted samples using 90% Glycerol (Sigma, Cat# G5516) in 1x PBS.
In preparation for imaging, cells were cultured in collagen-I gels in the ImmunoCult-XF T Cell
Expansion Medium (STEMCELL Technologies, Cat# 10981) with the addition of Human Recombinant
Interleukin 2 (STEMCELL Technologies, Cat# 78036.3). To prepare 3 mg/ml collagen-I gel, we assembled
a gel premix on ice in a prechilled Eppendorf tube. Briefly, to 1 volume of CellAdhere™ type I bovine
(STEMCELL Technologies, Cat# 07001) we added 8/10 volume of DMEM, 1/10 volume of 10x PBS, 1/20

`

volume of 1M HEPES, and 1/20 volume of 1M (in DMSO) Alexa Fluor 488 ester (Molecular Probes, Cat#
A20000). A drop of premixed gel (~50 µL) was spread immediately on a glass surface of a plasma-
treated glass-bottom 35 mm Petri dish (MatTek Corp., Cat# P35G-1.5-14-C) with a pipette tip. During
polymerization (room temperature, for overnight), gels were covered with 1 mL of mineral oil (Sigma-
Aldrich, Cat# M8410) to prevent evaporation of water. Before adding NK-92 cells, polymerized gels were
rinsed with PBS to remove the unpolymerized gel components.
Instant structured illumination microscopy (iSIM) was performed using the commercial instant
structured illumination microscope system (VisiTech Intl, Sunderland, UK) equipped with an Olympus
UPlanSAapo 60×/1.3NA Sil objective, two Flash-4 scientific CMOS cameras (Hamamatsu, Corp., Tokyo,
Japan), an iSIM scan head (VisiTech Intl, Sunderland, UK), and a Nano-Drive piezo Z stage (Mad City
Laboratories, Madison, WI). The iSIM scan head included the VT-Ingwaz optical destriping unit. The
exposure time was set to 250 ms per image frame. The voxel size was 64 x 64 x 250 nm, in x, y, and z,
respectively.
646

*Two-photon microscopy on live and fixed mouse tissue*
Fixed mouse liver samples and fresh ex-vivo mouse heart muscle strips were imaged with two-
photon microscopy using a Leica SP8 two photon DIVE upright microscope (Mannheim, Germany), a
pulsed dual beam Insight X3 Ti-Sapphire laser (MKS Spectra-Physics, Milpitas CA), a Leica 25x 1.0 NA (HC
PL IRAPO) water dipping lens, and emission bandwidth tunable Leica HyD detectors in the non-
descanned emission pathway. Liver samples were prepared from freshly excised liver from a 10 week-
old mouse expressing a membrane-targeted peptide fused with tdTomato[40]. After excision, the mouse
liver was washed in cold saline three times, fixed with 4% formaldehyde in PBS for 2 hours, and stored in
PBS.  Tissue harvesting procedures were approved by the NCI (for mouse liver) and NHLBI (for mouse
heart) Animal Care User Committees (ACUC) respectively. Freshly excised heart muscle strips from
transgenic mice expressing mitochondrial TOMM20-mNeonGreen were prepared for imaging as
described[41].  tdTomato and mNeonGreen were excited using 1045 nm and 960 nm excitation with
emission bandwidths of 550-700 nm and 500-600 nm, respectively. Laser excitation (ramped as a
function of depth in some experiments and optimized by adjusting the objective motorized correction
collar) were in the range of 1% for tdTomato and less than 20% for mNeonGreen. HyD detector gains
were kept at 100% for tdTomato and 150% for mNeonGreen. Tiled images volumes of liver membrane
expressing tdTomato were collected with voxels sizes set to 400 nm in the XY dimension and 500 nm in
the z dimension. Z-stack volumes of mNeonGreen expressing heart strip were collected with voxels sizes
set to 120 nm in the XY dimension and 500 nm in the z dimension. All imaging was conducted at an
imaging speed of 600 Hz with a pinhole size of 1 A.U.

*Cleared mouse embryos imaged with confocal microscopy*
E11.5-day mouse embryos were collected in phosphate-buffered saline (PBS) and directly
immersed in 4% paraformaldehyde (PFA) in PBS (pH 7.4) at 4°C overnight. Following fixation, the
samples were washed with PBS and stored in PBS at 4°C for further analyses. Wholemount
immunofluorescence staining was performed at 4°C. The mouse embryos were permeabilized with 0.2%
Triton/PBS overnight and blocked with 10% normal goat serum and 1% BSA in 0.2% Triton/PBS
overnight.  The embryos were then stained with monoclonal antibody against PECAM1 (CD31, clone
MEC 13.3, Cat# 553700, BD Pharmingen, 1:200 dilution) and monoclonal anti-β-tubulin III (TuJ1))
antibody (clone 2G10, Cat# T8578, Sigma-Aldrich, 1:500 dilution) in blocking buffer overnight. After
washing with 0.2% Triton/PBS, the embryos were stained with secondary antibodies with Alexa 488 goat

anti-rat IgG and Alexa 594 goat-anti-mouse IgG (1:250, Invitrogen, Carlsbad, CA) in blocking buffer
overnight. The embryos were cleared with CLARITY and imaged using a Zeiss LSM 880 Confocal
microscope with a 10X, 0.5NA air objective. To compensate for focal shift effects due to the refractive
index difference between air and CLARITY we scaled the axial voxel size of images by 1.45 before
processing for DeAbe.
682

683     *Calculation of vessel orientation and alignment*
684           Orientations were estimated in 3D using a weighted vector summation algorithm[23], adapting it
685     for the volumetric images of fiber-like structures corresponding to the CD31 channel (i.e., blood vessel
686     images) in CLARITY-cleared mouse embryos (**Fig. 3**).
687           For a given voxel within the 3D image, an $n \times n \times n$ voxel window was generated surrounding the
voxel under assessment. To segment the effective voxels, six-level Otsu intensity thresholding was applied
to the image, with five thresholds dividing the intensity into six levels. The lowest level was designated as
background noise, and regions assigned to the upper five levels defined the vessel signals. The window
size $n$ was typically set as two to three times the vessel thickness. All vectors passing through the center
voxel were defined and weighted by their length and intensity variations, and the direction of the sum of
all the weighted vectors was designated as the orientation of the center voxel[23], with associated azimuthal
angle $\theta$ (ranging from 0° to 180°) and polar angle $\varphi$ (ranging from 0° to 180°). However, since the
calculation of the polar angle $\varphi$ was not straightforward, we defined two additional azimuthal angles, $\beta$
and $\gamma$ (**Supplementary Fig. 11a**), which were symmetrical to the azimuthal angle $\theta$. $\beta$ was defined as the
angle between the projection of the vessel in the $zx$ plane and the $x$ axis, and $\gamma$ was the angle between
the projection in the $yz$ plane and the $-y$ axis. These two angles were related to the polar angle $\varphi$ via:
$\tan^2 \varphi = 1/\tan^2 \beta + 1/\tan^2 \gamma$.

700           We also derived the 3D directional variance (DV) metric, quantifying the spread in orientations[24,42].
The value of DV ranges from 0 to 1, with 0 corresponding to perfectly parallel alignment, and 1
corresponding to complete disorder (**Supplementary Fig. 11b**). The directional variance $\bar{D}_{3D}$ was defined
as:

$\bar{D}_{3D} = 1 - (\bar{C}_{3D}^2 + \bar{S}_{3D}^2 + \bar{Z}_{3D}^2)^{1/2}$,

where:

$\bar{C}_{3D} = (1/k)\sum_{j=1}^{k}(f_j/\sqrt{1+f_j^2})\cos(2\theta_j)$,

$\bar{S}_{3D} = (1/k)\sum_{j=1}^{k}(f_j/\sqrt{1+f_j^2})\sin(2\theta_j)$,

$\bar{Z}_{3D} = (1/k)\sum_{j=1}^{k}(SI/\sqrt{1+f_j^2})$,

with $f_j = \sqrt{1/\tan^2(2\beta_j) + 1/\tan^2(2\gamma_j)}$, and $SI = (-1)\cdot(\varphi-90)/|\varphi-90|$, where $\varphi$ was acquired from the
determination of $\beta$ and $\gamma$ as described above, $k$ was the number of fiber voxels in the region, and $\theta$, $\beta$
and $\gamma$ were calculated azimuthal angles as described above.

`

*Membrane segmentation*

For the images of live worm embryos dual-labeled with nuclear and membrane markers (**Fig. 4c,**
**d, Supplementary Fig. 16**), raw data was restored using our multiple-step deep learning pipeline (Steps 1-
3 in **Supplementary Fig. 13a**) prior to cell membrane segmentation. We performed automatic membrane
segmentation using segmented nuclei as seeds:

First we used the Keras and Tensorflow-based implementations of Mask RCNN[43]
(https://github.com/matterport/Mask_RCNN) to perform nuclear segmentation (**Supplementary Fig.**
**16d**). We then manually segmented 8 volumes (3 acquired with diSPIM, 3 with iSPIM, and 2 from
multiview confocal microscopy[22] for a total of 1963 nuclei) for training. Of these 8 volumes, 6 volumes
with a total of 1688 nuclei were used for training a segmentation network and 2 volumes with a total of
275 nuclei were used for validation. We used a ResNet-50 model as the backbone for our network,
initialized the model using weights obtained from pretraining on the MS COCO dataset[44], and proceeded
to train all layers in three stages. Training took ~10 hours and applying the model took ~ 3 minutes per
volume on a Windows workstation equipped with an Intel(R) Xeon(R) W-2145 CPU operating at 3.70 GHz,
an Nvidia Quadro P6000 GPU, and 128 GB of RAM. After Mask RCNN segmentation, we applied a marker-
controlled watershed operation (https://www.mathworks.com/help/images/marker-controlled-
watershed-segmentation.html) to the nuclear segmentations to separate touching nuclei.

Second, we applied the vascular structure enhancement filter[45]
(https://github.com/timjerman/JermanEnhancementFilter) to the membrane data to enhance
boundaries (**Supplementary Fig. 16c**). Scales were set to [2.0, 2.25, 2.5] and all other parameters were set
to the default.

Third, the centroids of segmented nuclei were used as seeds, and we used the seeded watershed
algorithm (https://github.com/danielsnider/Simple-Matlab-Watershed-Cell-Segmentation) for
membrane segmentation (**Supplementary Fig. 16f**).

This workflow was applied both to the raw image data and restored images after each step in our
multi-step pipeline to demonstrate the benefit of segmentation enhancement from DL processing.

For selected volumes (**Fig. 4c, Supplementary Video 11**), we also performed manual editing on
the automatic segmentations produced by the multi-step deep learning pipeline. Manual editing was
performed within the ImageJ plugin Labkit (https://imagej.net/plugins/labkit/). After automatic
segmentations were imported to Labkit, segmentation labels were manually edited interactively in lateral
views (XY planes), and then were edited in axial views (YZ planes). Since the manual editing was conducted
in 2D views and initial editing in either view was not sufficient to ensure smoothness in 3D, we iterated
twice to further improve our results.

*Code availability*

Training and applying deep learning models were achieved using Python 3.7.0. Generation of synthetic
aberrated data and quantitative image analysis was performed in MATLAB (Mathworks, R2022b).
Customized code and software are available at https://github.com/eguomin/DeAbePlus/. RCAN and
CARE software were installed from https://github.com/AiviaCommunity/3D-RCAN and
https://github.com/CSBDeep/CSBDeep, and code for RL deconvolution and multiview fusion is available
at https://github.com/eguomin/diSPIMFusion/.

*Data availability*

The data that support the findings of this study are included in **Supplementary Figs. 1–23** and
**Supplementary Videos 1–14**. Some representative data from the figures (**Supplementary Figs. 6, 17**) are
publicly available at https://zenodo.org/record/8424246. Other datasets (training data and intermediate
data for deep learning) are available from the corresponding author upon reasonable request due to
their large file size.

`

**Fig. 1, Concept and simulations illustrating deep learning-based aberration compensation**. **a)**
Schematic. *Left*: Fluorescence microscopy volumes are collected and near-diffraction-limited images
from the shallow side of each stack are synthetically degraded to resemble aberrated images deeper
into the stack. A neural network (e.g., a three-dimensional residual channel attention network, 3D
RCAN) is trained to reverse this degradation given the ground truth on the shallow side of the stack, and
the trained neural network (DeAbe model) subsequently applied to images throughout the stack,

`

improving contrast and resolution. *Right*: More detailed view of synthetic degradation process. Zernike
basis functions and associated coefficients (coeffs) are used to add random aberrations by modifying the
ideal point spread function (iPSF) to generate an aberrated PSF (aPSF). Ground truth images (GT) are
Fourier transformed (FT) and multiplied by the ratio of the Fourier transformed aberrated and ideal PSFs
(essentially a modified optical transfer function, mOTF). Inverse Fourier transforming (IFT) the result and
adding noise generates the synthetically aberrated images. See **Methods** for further detail on this
procedure. OBJ: objective lens used to collect the stack. **b)** Simulated three-dimensional phantoms
consisting of randomly oriented and positioned dots, lines, spheres, spherical shells, and circles
comparing maximum intensity projections of aberrated input image (left, random aberration with root
mean square (RMS) wavefront distortion < 2 radians and Poisson noise added for an SNR of ~16,
corresponding PSF in inset), network prediction (DeAbe) given aberrated input (middle), and ground
truth (GT, right). Higher magnification views of dashed rectangular region are shown in **c)** (maximum
intensity projection) and **d)** (single plane), additionally showing restoration given blind deconvolution
(Blind Decon), Richardson-Lucy deconvolution with diffraction-limited PSF (RL Decon 1), Richardson-Lucy
deconvolution with aberrated PSF (RL Decon 2). Yellow arrows indicate a reference structure for visual
comparison. Twenty iterations were used for RL deconvolution and ten for blind deconvolution. **e)** As in
**c, d)** but showing axial plane along dashed blue line in **b)**. **f)** Quantitative comparisons for the
restorations shown in **b-e)** using structural similarity index (SSIM, top) and peak signal-to-noise ratio
(PSNR, bottom) with ground truth reference. Means and standard deviations are shown for 100
simulations (10 independent phantom volumes, each aberrated with 10 randomly chosen aberrations).
Scale bars: 5 μm **b)** and 2.5 μm **c-e)**. See also **Supplementary Figs. 1-5**.

**Figure 2, Computational aberration compensation on fluorescence microscopy image volumes. a)** Live
*C. elegans* embryos expressing a pan-nuclear GFP histone marker were imaged with light sheet
microscopy (**i**, left column) and the raw data processed with Richardson-Lucy deconvolution (**ii**, 10
iterations, middle column) or with a trained DeAbe model (**iii**, right column). First two rows show single

planes 20 and 28 μm into the sample, third row shows axial view. Comparative line profiles through

dashed blue, yellow, and green lines are shown in insets, comparing ability to discriminate nuclei. Red

arrow highlights nuclei for visual comparison. See also **Supplementary Video 3**. **b)** NK-92 cells stained

with Alexa Fluor 555 wheat germ agglutinin and embedded in collagen matrices were fixed and imaged

with instant SIM, a super-resolution imaging technique. Left: raw data, right: after application of DeAbe

and deconvolution (DeAbe+, 20 iterations Richardson-Lucy). Lateral maximum intensity projections

(MIP, top) or single axial planes (bottom) are shown in **b)**, and **c, d** show higher magnification views

corresponding to green **c)** or blue **d)** dashed rectangular regions in **b)**. Colored arrows in **c, d** highlight

fine features obscured in the raw data and better revealed in the DeAbe+ reconstructions. See also

**Supplementary Video 5**, **Supplementary Fig. 8. e)** Live cardiac tissue containing cardiomyocytes

expressing Tomm20-GFP was imaged with two photon microscopy. Raw data (left) are compared with

DeAbe prediction (right) at indicated depths, with insets showing corresponding Fourier transform

magnitudes. Blue circles in Fourier insets in **e)** indicate $1/300$ nm$^{-1}$ spatial frequency just beyond

resolution limit. See also **Supplementary Video 6**. **f)** Higher magnification views of white dashed

rectangular region in **e)**, emphasizing recovery of mitochondrial boundaries by DeAbe model. See also

**Supplementary Fig. 9, Supplementary Video 7.** Scale bars: 10 μm **a, e)**; 5 μm **b, f)**; 2 μm **c, d)**; **e)**

diameter of Fourier circle: $300$ nm$^{-1}$.

**Fig. 3, Computational aberration compensation on mm-scale cleared mouse embryo volumes. a)** Fixed
and CLARITY-cleared E11.5-day mouse embryos were immunostained for neurons (TuJ1, cyan) and
blood vessels (CD31, magenta), imaged with confocal microscopy and processed with a trained DeAbe
model. See also **Supplementary Video 8**. **b)** Axial view corresponding to dotted rectangular region in **a)**,
comparing raw data and depth-compensated, de-aberrated, and deconvolved data (DeAbe+). See also
**Supplementary Fig. 10**. **c)** Higher magnification lateral view at axial depth of 1570 μm indicated by the
orange double headed arrowheads in **b)**. **d)** Higher magnification views of white dotted region in **c)**,
comparing raw (left) and DeAbe+ processing (right) for neuronal (top) and blood vessel (bottom) stains.
**e)** Orientation (θ, transverse angle) analysis on blood vessel channel of DeAbe+ data, here shown on
single lateral plane at indicated axial depth. See also **Supplementary Fig. 11, Supplementary Video 9**. **f)**
Higher magnification lateral view of white dotted region in **e)** (note that axial plane is different),
comparing intensity (left) and orientation (right) views between raw (top row) and DeAbe+ prediction

(middle row). Righthand insets show higher magnification views of vessel and surrounding region
highlighted by dotted lines. Bottom row indicates histogram of all orientations in the vessel highlighted
with dotted ellipse, full-width-at-half maximum (FWHM) in peak region of histogram is also shown. **g)**
Directional variance of blood vessel stain within the indicated plane, with higher magnification region of
interest (ROI) views at right. Histogram of directional variance in both regions also shown. See also
**Supplementary Fig. 12**. Scale bars: 500 $\mu$m **a, b, c, e)**; 100 $\mu$m **d)**, 50 $\mu$m inset; 300 $\mu$m **f)**, 50 $\mu$m inset;
300 $\mu$m **g)**, 50 $\mu$m inset.

**Fig. 4, Incorporating aberration compensation into multi-step restoration dramatically improves image quality in volumetric time-lapse imaging. a)** *C. elegans* embryos expressing GFP-labeled membrane marker (green) and mCherry-labeled nuclear marker (magenta) were imaged with dual-view

light-sheet microscopy (diSPIM) and the raw data (left) from single-view recordings processed through

neural networks that progressively de-aberrated, deconvolved, and isotropized spatial resolution (3-step

DL, right). Single planes from lateral (top) and axial (bottom) perspectives are shown, with arrow in

lower panel indicating direction of increasing depth. See also **Supplementary Video 10**, **Supplementary**

**Figs. 14, 15**. **b)** Higher magnification axial views of membranes (top) and nuclei (bottom) deep into

embryo, corresponding to dashed rectangle in **a)**. **c)** Examples of automatic segmentation on raw (left,

319 cells), 3-step DL prediction (middle, 421 cells), and manually corrected segmentation based on DL

prediction (right, 421 cells). Single planes corresponding to the upper planes in **a)** are shown. Red and

blue dashed ellipses highlight regions for visual comparison. See also **Supplementary Video 11**. **d)**

Number of cells detected by automatic segmentation of membrane marker vs. time for raw data

(purple), and after successively applying each step in the multistep restoration (Steps 1-3, blue, green,

and red curves). Ground truth from manual expert (black curve) is also shown for comparison. Inset

(ellipse with dotted blue lines) highlights number count at early timepoints. **e)** Maximum intensity

projection (MIP) images of *C. elegans* embryos expressing membrane-localized GFP under control of the

*ttx3-3b* promoter, imaged with diSPIM, comparing raw single-view recordings (left) and multi-step

restoration that progressively de-aberrated, deconvolved, and super-resolved the data (right, 3-step DL).

Boundary of the embryo has been outlined in light blue for clarity. See also **Supplementary Figs. 19, 20,**

**Supplementary Video 12.** Higher magnification MIP (**f**) or single lateral (**g**) or axial (**h**) plane comparisons

corresponding to dashed lines or rectangle in **e)** are also shown. **i)** Time series based on 3-step DL MIP

predictions highlight developmental progression of AIY (blue) and SMDD (magenta) neurites as they

enter the nerve ring region. Top and bottom parts of each panel at each time point show MIP (neurites

highlighted as dotted lines) vs. model of the neurites, respectively. See also **Supplementary Fig. 21**.

Scale bars: 5 $\mu$m **a, c, e, f, h**); 2 $\mu$m **b, d, g**).

Ji, N. Adaptive optical fluorescence microscopy. *Nature Methods* **14**, 374 (2017).

Hampson, K. M. *et al.* Adaptive optics for high-resolution imaging. *Nat Rev Methods Primers* **1**, 1-26 (2021).

Wang, K. *et al.* Rapid adaptive optical recovery of optimal resolution over large volumes. *Nat Methods* **11**, 625-628 (2014).

Zheng, W. *et al.* Adaptive optics improves multiphoton super-resolution imaging. *Nat Methods* **14**, 869-872 (2017).

Weigert, M. *et al.* Content-aware image restoration: pushing the limits of fluorescence microscopy. *Nat Methods* **15**, 1090-1097 (2018).

Qiao, C. *et al.* Rationalized deep learning super-resolution microscopy for sustained live imaging of rapid subcellular processes. *Nat Biotechnol.* (2022). https://doi.org:doi: 10.1038/s41587-022-01471-3

Li, Y. *et al.* Incorporating the image formation into deep learning improves network performance. *Nature Methods* (2022). https://doi.org:doi: 10.1038/s41592-022-01652-7

Guo, M. *et al.* Rapid image deconvolution and multiview fusion for optical microscopy. *Nature Biotechnol.* **38**, 1337-1346 (2020). https://doi.org:https://doi.org/10.1038/s41587-020-0560-x

Chen, J. *et al.* Three-dimensional residual channel attention networks denoise and sharpen fluorescence microscopy image volumes. *Nature Methods* **18**, 678-687 (2021). https://doi.org:https://doi.org/10.1101/2020.08.27.270439

Wang, H. *et al.* Deep learning enables cross-modality super-resolution in fluorescence microscopy. *Nat Methods* **16**, 103-110 (2019).

Tian, Q. *et al.* DNN-based aberration correction in a wavefront sensorless adaptive optics system. *Optics Express* **27**, 10765-10776 (2019).

Saha, D. *et al.* Practical sensorless aberration estimation for 3D microscopy with deep learning. *Opt Express* **28**, 29044-29053 (2020).

Vinogradova, K. & Myers, E. W. Estimation of Optical Aberrations in 3D Microscopic Bioimages. *2022 7th International Conference on Frontiers of Signal Processing (ICFSP)*, 97-103 (2022). https://doi.org:doi: 10.1109/ICFSP55781.2022.9924879

Hanser, B. M., Gustafsson, M. G. L., Agard, D. A. & Sedat, J. W. Phase retrieval for high-numerical-aperture optical systems. *Optics Letters* **28**, 801-803 (2003).

Hanser, B. M., Gustafsson, M. G. L., Agard, D. A. & Sedat, J. W. Phase-retrieved pupil functions in wide-field fluorescence microscopy. *J Microsc.* **216(Pt 1)**, 32-48 (2004).

Holmes, T. J. *et al.* in *Handbook of Biological Confocal Microscopy*   (ed James B. Pawley)  389-402 (Springer, 1995).

Wu, Y. *et al.* Inverted selective plane illumination microscopy (iSPIM) enables coupled cell identity lineaging and neurodevelopmental imaging in Caenorhabditis elegans. *Proc. Natl. Acad. Sci. USA* **108**, 17708-17713 (2011).

Wu, Y. *et al.* Reflective imaging improves spatiotemporal resolution and collection efficiency in light sheet microscopy. *Nat Commun.* **8**, 1452 (2017).

Yemini, E. *et al.* NeuroPAL: A Multicolor Atlas for Whole-Brain Neuronal Identification in *C. elegans*. *Cell* **184**, 272-288 (2021).

York, A. G. *et al.* Instant super-resolution imaging in live cells and embryos via analog image processing. *Nat Methods* **10**, 1122-1126 (2013).

Chung, K. *et al.* Structural and molecular interrogation of intact biological systems. *Nature* **497**, 332-337 (2013).

22    Wu, Y. *et al.* Multiview confocal super-resolution microscopy. *Nature* **600**, 279-284 (2021).
https://doi.org:https://doi.org/10.1101/2021.05.21.445200
23    Liu, Z. *et al.* Rapid three-dimensional quantification of voxel-wise collagen fiber orientation.
*Biomedical Optics Express* **6**, 2294-2310 (2015).
24    Liu, Z. *et al.* Automated quantification of three-dimensional organization of fiber-like structures
in biological tissues. *Biomaterials* **116**, 34-47 (2017).
25    Wu, Y. *et al.* Spatially isotropic four-dimensional imaging with dual-view plane illumination
microscopy. *Nat Biotechnol.* **31**, 1032-1038 (2013).
26    Kumar, A. *et al.* Dual-view plane illumination microscopy for rapid and spatially isotropic
imaging. *Nature Protocols* **9**, 2555-2573 (2014).
27    Cao, J. *et al.* Establishment of a morphological atlas of the *Caenorhabditis elegans* embryo using
deep-learning-based 4D segmentation. *Nature Communications* **11**, 6254 (2020).
28    Sulston, J. E., Schierenberg, E., White, J. G. & Thomson, J. N. The embryonic cell lineage of the
nematode Caenorhabditis elegans. *Dev. Biol.* **100**, 64-119 (1983).
29    Chen, F., Tillberg, P. & Boyden, E. S. Expansion microscopy. *Science* **347**, 543-548 (2015).
30    Bertrand, V. & Hobert, O. Linking Asymmetric Cell Division to the Terminal Differentiation
Program of Postmitotic Neurons in *C. elegans*. *Developmental Cell* **16**, 563-575 (2009).
31    Rapti, G., Li, C., Shan, A., Lu, Y. & Shaham, S. Glia initiate brain assembly through noncanonical
Chimaerin-Furin axon guidance in *C. elegans*. *Nature Neuroscience* **20**, 1350-1360 (2017).
32    Moyle, M. W. *et al.* Structural and developmental principles of neuropil assembly in *C. elegans*.
*Nature* **591**, 99-104 (2021).
33    Brittin, C. A. *et al.* Multicellular rosettes organize neuropil formation. *bioRxiv* (2020).
34    White, J. G., Southgate, E., Thomson, J. N. & Brenner, S. The Structure of the Nervous System of
the Nematode Caenorhabditis elegans. *Phil. Trans. R. Soc. London Ser. B* **314**, 1-340 (1986).
35    Mori, I. & Ohshima, Y. Neural regulation of thermotaxis in *Caenorhabditis elegans*. *Nature* **376**,
344-348 (1995).
36    Thibos, L. N., Applegate, R. A., Schwiegerling, J. T., Webb, R. & Members, V. S. T. Standards for
Reporting the Optical Aberrations of Eyes. *OSA Technical Digest*, SuC1 (2000).
37    Richardson, W. H. Bayesian-Based Iterative Method of Image Restoration. *JOSA* **62**, 55-59
(1972).
38    Lucy, L. B. An iterative technique for the rectification of observed distributions. *Astronomical
Journal* **79**, 745-754 (1974).
39    Bao, Z. & Murray, J. I. in *Imaging in Developmental Biology: A Laboratory Manual*  (eds James
Sharpe, Rachel Wong, & Raphael Yuste)  (Cold Spring Harbor Laboratory Press, 2010).
40    Muzumdar, M. D., Tasic, B., Miyamichi, K., Li, L. & Luo, L. A global double-fluorescent Cre
reporter mouse. *Genesis* **45**, 593-605 (2007).
41    Kosmach, A. *et al.* Monitoring mitochondrial calcium and metabolism in the beating MCU-KO
heart. *Cell Reports* **37**, 109846 (2021).
42    Liu, Z. *et al.* Label-free, multi-parametric assessments of cell metabolism and matrix remodeling
within human and early-stage murine osteoarthritic articular cartilage. *Communications Biology*
**6**, 405 (2023).
43    He, K., Gkioxari, G., Dollár, P. & Girshick, R. Mask R-CNN. *2017 IEEE conference on Computer
Vision (ICCV)*, 2980-2988 (2017).
44    Lin, T.-Y. *et al.* in *ECCV 2014* Vol. 8693  (eds D. Fleet, T. Pajdla, B. Schiele, & T. Tuytelaars)
(Springer, 2014).
45    Jerman, T., Pernuš, F., Likar, B. & Špiclin, Ž. Enhancement of Vascular Structures in 3D and 2D
Angiographic Images. *IEEE Trans Med Imaging* **35**, 2107-2118 (2016).