Deep learning-based aberration compensation improves contrast and resolution in fluorescence
microscopy

Min Guo[1,2, *], Yicong Wu[2, 3, ^], Chad M. Hobson[4], Yijun Su[2, 3, 4], Shuhao Qian[1], Eric Krueger[2, 4], Ryan
Christensen[2, 4], Grant Kroeschell[2, 4], Johnny Bui[2, 4], Matthew Chaw[2, 4], Lixia Zhang[3], Jiamin Liu[3], Xuekai
Hou[1], Xiaofei Han[2], Zhiye Lu[5], Xuefei Ma[5], Alexander Zhovmer[6], Christian Combs[7], Mark Moyle[8], Eviatar
Yemini[9], Huafeng Liu[1], Zhiyi Liu[1], Alexandre Benedetto[10], Patrick La Riviere[11, 12], Daniel Colón-Ramos[12, 13],
Hari Shroff[2, 3, 4, 12]

1. State Key Laboratory of Modern Optical Instrumentation, College of Optical Science and
Engineering, Zhejiang University, Hangzhou, China
2. Laboratory of High Resolution Optical Imaging, National Institute of Biomedical Imaging and
Bioengineering, National Institutes of Health, Bethesda, Maryland, USA
3. Advanced Imaging and Microscopy Resource, National Institutes of Health, Bethesda, Maryland,
USA
4. Janelia Research Campus, Howard Hughes Medical Institute (HHMI), Ashburn, VA, USA
5. Laboratory of Molecular Cardiology, National Heart, Lung, and Blood Institute, National
Institutes of Health, Bethesda, Maryland, USA
6. Center for Biologics Evaluation and Research, U.S. Food and Drug Administration, Silver Spring,
MD, USA
7. NHLBI Light Microscopy Facility, National Institutes of Health, Bethesda, MD, USA
8. Department of Biology, Brigham Young University-Idaho, Rexburg, ID, USA
9. Department of Neurobiology, UMass Chan Medical School, Worcester, MA
10. Faculty of Health and Medicine, Division of Biomedical and Life Sciences, Lancaster University, UK
11. Department of Radiology, University of Chicago, Chicago, IL, USA
12. MBL Fellows Program, Marine Biological Laboratory, Woods Hole, MA, USA
13. Wu Tsai Institute, Department of Neuroscience and Department of Cell Biology, Yale University
School of Medicine, New Haven, CT, USA

^ Current address: Nanodelivery Systems and Devices Branch, Cancer Imaging Program, Division of
Cancer Treatment and Diagnosis, National Cancer Institute, National Institutes of Health, Rockville,
MD, USA

* Correspondence to guom@zju.edu.cn

Abstract

Optical aberrations hinder fluorescence microscopy of thick samples, reducing image signal, contrast,
and resolution. Here we introduce a deep learning-based strategy for aberration compensation,
improving image quality without slowing image acquisition, applying additional dose, or introducing
more optics into the imaging path. Our method (i) introduces synthetic aberrations to images acquired
on the shallow side of image stacks, making them resemble those acquired deeper into the volume and
(ii) trains neural networks to reverse the effect of these aberrations. We use simulations and
experiments to show that applying the trained 'de-aberration' networks outperforms alternative
methods, providing restoration on par with adaptive optics techniques; and subsequently apply the
networks to diverse datasets captured with confocal, light-sheet, multi-photon, and super-resolution
microscopy. In all cases, the improved quality of the restored data facilitates qualitative image

inspection and improves downstream image quantitation, including orientational analysis of blood
vessels in mouse tissue and improved membrane and nuclear segmentation in *C. elegans* embryos.

Introduction

Fluorescence microscopes offer diffraction-limited imaging only when optical aberrations are
absent. Such aberrations can arise due to optical path length differences introduced anywhere in the
imaging path, including from instrument misalignment, optical imperfections, or differences in refractive
index between the heterogenous and refractile sample, immersion media, or objective immersion oil.
Sample-induced optical aberrations usually dominate and are often the reason that three-dimensional
(3D) fluorescence image volumes show obvious deterioration in image signal-to-noise ratio (SNR),
contrast, and resolution deeper into the image volume.
One method of compensating for these aberrations is via adaptive optics (AO[1,2]), a broad class of
techniques that measure the aberrated wavefront and subsequently apply an equal and opposite
'corrective' wavefront, restoring diffraction-limited[3] or even super-resolution[4] imaging throughout the
image volume. Once the aberrated wavefront is determined, an adaptive element such as a deformable
mirror or spatial light modulator is used to apply the correction. Although these methods are effective,
the process of determining the wavefront typically slows acquisition and/or applies more illumination
dose than imaging without AO. From a practical perspective, implementing AO is nontrivial and adds
considerable expense to the underlying microscope. Thus, AO remains the province of relatively few
labs, and there is a need for new methods that can reverse the effects of optical aberrations without
sacrificing temporal resolution, imparting more dose to the sample, or adding additional hardware to
the microscope.
Deep learning approaches can computationally reverse image degradation, and have been used
successfully in denoising[5,6], deconvolution[7,8], and super-resolution applications[9,10]. By incorporating
information about the underlying object, such methods can also learn to predict the wavefront
associated with aberrated images[11-13]. With sufficient training data (matched pairs of diffraction-limited
and aberrated data), we reasoned that a neural network ought to be able to directly predict the
diffraction-limited image from the aberrated image. The challenge then becomes accumulating
appropriate training data, which would ideally be obtained without relying on AO.
Here we address this problem by (i) introducing synthetic aberrations to easily obtained near-
diffraction limited data so that they resemble aberrated data and (ii) training neural networks to reverse
the effect of these aberrations. We use simulations to show that application of our 'content-aware'
approach outperforms other image restoration methods, including deconvolution with the known
aberrated point spread function (PSF). We also show that our method provides performance on par with
direct wavefront sensing-based AO[3], by comparing its output to experimental ground truth. We then
apply our techniques to diverse volumetric data captured with confocal, light-sheet, multi-photon, and
super-resolution microscopes, finding that in all cases, resolution and contrast are substantially
improved over the raw data. In addition to facilitating biological inspection, the restored data also
enhanced quantitative investigation, including orientational analysis of blood vessels in mouse tissue
and improved accuracy of membrane and nuclear segmentation in *C. elegans* embryos.
82
83 **Results**
84

Compensating for aberrations with deep learning

First, we intentionally synthetically aberrate the images acquired by fluorescence microscopes given knowledge of the physics of image formation[14,15] (**Fig. 1, Methods, Supplementary Note 1**). Aberrations are chosen so that the aberrated images resemble those acquired deeper into the sample, where aberrations are more pronounced. The key insight of our approach is that the 'shallow' images on the 'near side' of the three-dimensional fluorescence volume are usually near-diffraction-limited and thus provide ground truth data that can be used to train a network to reverse the effect of the synthetically introduced aberrations. The trained neural network model (termed 'DeAbe') can then be used to reverse depth-dependent blurring on data unseen by the network, effectively mitigating the effect of aberrations without recourse to AO.

To benchmark our method, we began by simulating 3D phantoms consisting of randomly oriented and positioned dots, lines, spheres, circles, and spherical shells. We then degraded these structures by adding random aberrations and noise and evaluated the extent to which DeAbe could reverse the degradation (**Fig. 1b, Supplementary Figs. 1-7**). Visual assessments in lateral (**Fig. 1c, d, Supplementary Video 1**) and axial (**Fig. 1e, Supplementary Video 2**) views, as well as quantitative comparisons (**Fig. 1f**) demonstrated that the DeAbe model outperformed blind deconvolution[16], Richardson-Lucy deconvolution with an ideal point spread function (PSF), Richardson-Lucy deconvolution with the aberrated PSF (known in these simulations, but unknown in general), and denoising methods (**Supplementary Figs. 6, 7**). We attribute the superior performance of DeAbe to its ability to learn a sample-specific prior, thereby better conditioning its solution relative to Richardson-Lucy deconvolution.

Importantly, simulations allowed us to further characterize DeAbe, offering insight into the regimes in which the method excels and where performance suffers. First, we found optimal performance when aberration magnitudes in the training data match the aberration magnitude in the test data (**Supplementary Fig. 1**). Over the conditions we tested, the model improved images contaminated with root mean square (RMS) wavefront distortion exceeding four radians (the highest value we tested), although performance degrades as wavefront distortion increases. Second, although we performed tests with training data containing up to the 7th Zernike order, the improvement offered past order four (the value used in this work) is negligible (**Supplementary Fig. 2**). Third, DeAbe trained on a mixture of Zernike basis functions also provides notable improvement on images corrupted solely by individual Zernike functions (**Supplementary Fig. 3**), although dedicated models trained to correct specific Zernike modes are better if these modes are known in advance (**Supplementary Fig. 4**). Fourth, although DeAbe's performance suffers in the presence of noise, it still offers noticeable visual and quantitative improvements in image quality for SNR above ~5 (**Supplementary Fig. 5**). Finally, we explored different networks for implementing DeAbe, finding that our previous 3D RCAN[9] offered better performance than CARE[5], RLN[7], or BasicVSR++[17] architectures (**Supplementary Figs. 8, 9**).

Comparing DeAbe predictions to experimental ground truth

We next benchmarked DeAbe against experimental datasets acquired with a lattice light sheet microscope[18] equipped with adaptive optics for inducing and correcting aberrations (AO-LLSM[19], **Fig. 2, Supplementary Table 1**). When imaging phalloidin-stained PtK2 cells (**Fig. 2a-f**), we induced aberrations

that obscured the fine actin mesh at the cell periphery, filamentous actin, and stress fibers (**Fig. 2b-d**).
Training a DeAbe model with a mixture of random aberrations restored these structures, improving
contrast and resolution to a level approaching the aberration-free ground truth (**Fig. 2e, f**) or AO result
(**Supplementary Fig. 10**). As for the simulations (**Supplementary Fig. 4, 6, 7**), we confirmed that training
DeAbe with Zernike modes matching the underlying aberration enhanced performance compared to a
random mixture of modes (**Supplementary Fig. 11**) and that DeAbe outperformed deconvolution
(**Supplementary Fig. 12**) and denoising (**Supplementary Fig. 13**).
We compared the performance of DeAbe to AO on a more challenging sample, fixed 5dpf
zebrafish embryos expressing a GFP membrane marker labeling glutamatergic neurons (**Fig. 2g-m**),
When acquiring image volumes 40-140 μm from the surface of the fish, AO correction and the DeAbe
prediction improved lateral (**Fig. 2i-k**) and axial (**Fig. 2l**) views of the raw data, enhancing spatial
resolution (**Fig. 2m**). Intriguingly, we also found examples in which the visual clarity of the DeAbe
prediction appeared better than the AO correction (**Supplementary Fig. 14**), perhaps reflecting
imperfect AO correction. The cell and fish samples also allowed us to investigate whether models
trained on one sample type generalized to the other. As we[20] and others[21] have reported, we obtained
superior results when training models specific to each sample type (**Supplementary Fig. 15**).
145
146   *Computational aberration compensation improves image quality on diverse volumetric data*
147
We subsequently applied DeAbe to diverse datasets acquired with different microscope
modalities, in each case training models on images derived from the shallow side of image volumes (**Fig.
3, Supplementary Fig. 16-17, Supplementary Table 1**). First, we imaged live *C. elegans* embryos
expressing a pan-nuclear GFP-histone marker with inverted selective plane illumination microscopy
(iSPIM)[22,23], finding that the raw image data displayed progressive loss of contrast and resolution as a
function of increasing depth, making it difficult or impossible to discern subnuclear structure (or even
individual nuclei) at deeper imaging planes (**Fig. 3a, i, Supplementary Video 3**). By contrast, the DeAbe
prediction restored these structures, also improving axial views (**Fig. 3a, iii**). Richardson-Lucy
deconvolution also offered some improvement in image quality, albeit not to the extent of the DeAbe
prediction, while also undesirably amplifying noise (**Fig. 3a, ii**). Second, we used spinning-disk confocal
microscopy to image thicker adult *C. elegans* expressing the multicolor NeuroPAL transgene[24], used for
resolving neuronal identities. Depth-dependent image degradation produced raw images with dim or
diffuse nuclear signal in each color channel. The DeAbe prediction improved SNR dramatically
(**Supplementary Fig. 18, Supplementary Video 4**), which we suspect may prove useful in improving the
accuracy of neuronal identification. Third, we applied DeAbe to images of NK-92 cells stained with Alexa
Fluor 555 wheat germ agglutinin and embedded in collagen matrices, acquired with instant SIM[25], a
super-resolution imaging technique (**Fig. 3b-d**, **Supplementary Fig. 19, Supplementary Video 5**). Post
deconvolution, the DeAbe prediction better resolved clusters of membrane-bound glycoproteins,
intracellular vesicles, and membranes ('DeAbe+', **Fig. 3c, d**) than the raw (or deconvolved raw,
Supplementary Fig. 19) data, especially near the limits of the 45 μm thick imaging volume. Fifth, we
verified that the DeAbe prediction restored the shapes of neuronal nuclei located on the 'far side' of
anesthetized adult *C. elegans* imaged with instant SIM, matching ground truth experiments in which we
flipped the worm over (**Supplementary Fig. 20**). Sixth, we used two-photon microscopy to image live
murine cardiac tissue expressing Tomm20-GFP, marking the outer mitochondrial membrane (**Fig. 3e**).
Although mitochondrial boundaries were evident in the raw data 20 μm into the volume, aberrations

caused a progressive loss in resolution that hindered visualization of subcellular structure at greater
depths (**Fig. 3e, f**). The DeAbe prediction restored resolution throughout the 150 µm thick volume (**Fig.**
3f, Supplementary Fig. 21, Supplementary Video 6), unlike Richardson-Lucy deconvolution (**Fig. 3f**)
which amplified noise without restoring the mitochondria. The DeAbe prediction similarly improved
contrast and resolution when applied to volumes of fixed mouse liver stained with membrane labeled
tdTomato, imaged with two-photon microscopy (**Supplementary Video 7)**. Quantitative contrast metrics
(**Methods, Supplementary Fig. 22**) confirmed our visual impressions of contrast improvement provided
by DeAbe.
Next, we applied DeAbe to samples ~10,000-fold larger in volumetric extent (**Fig. 4a,**
Supplementary Video 8). We fixed and iDISCO[26]-cleared E11.5 mouse embryos immunostained for
neurons (Alexa Fluor TuJ1) and blood vessels (Alexa Fluor 594) and imaged them with low magnification
confocal microscopy. Although tissue clearing nominally produces a sample with the same refractive
index everywhere, we still observed pronounced depth-dependent degradation from the 'near' to 'far'
side of the embryo, including in intensity (likely due to photobleaching during the acquisition) and
resolution. We were able to largely reverse this deterioration by digitally compensating for
photobleaching[27] (**Methods**), applying DeAbe, and finally deconvolving the data (**Fig. 4b, Supplementary**
Fig. 23). While the improvement in image quality was particularly striking in axial views (**Fig. 4b**),
restorations also improved the appearance of fibrillar structures in lateral views, in both channels,
throughout the volume (e.g., the vicinity of the vagus nerve and its associated nerve roots, **Fig. 4c, d**).
We further investigated this qualitative impression by using automated tools[28,29] to
quantitatively assess the mean 3D orientation and directional variance (a measure of the spread in
angular orientation) at each voxel in the blood vessel channel (**Fig. 4e-g**, **Supplementary Figs. 24, 25,**
Supplementary Video 9). The DeAbe restoration resulted in cleaner separation between vessels, which
aided voxel-wise quantification of these metrics even in dense regions containing many crisscrossing
vessels (**Fig. 4e, Supplementary Video 9**). In deeper regions of the volume (**Fig. 4f**), the DeAbe results
produced narrower angular histogram distributions of vessels than the noisy raw data (**Fig. 4f**). The
improvement in quantification was also reflected in directional variance analysis. For example, when
visually inspecting different regions of interest (ROI) with differential vessel alignment (**Fig. 4g**,
comparing vicinity of aortic arches, (ROI 1), to diencephalon, (ROI 2)) we observed a greater difference
in mean directional variance when using the DeAbe reconstruction vs. the raw data (**Supplementary Fig.**
25).
204
205   *Incorporating DeAbe in multi-step restoration further enhances resolution and contrast in 4D imaging*
206   *applications*
207
Given the performance of DeAbe thus far, we wondered if we could further boost image quality
by combining DeAbe with additional networks designed to enhance spatial resolution. To test this
possibility, we acquired dual-view light sheet microscopy (diSPIM[30,31]) volumetric time-lapse ('4D')
recordings of *C. elegans* embryos expressing labels marking cell membranes and nuclei, and then passed
the raw single-view data through three networks designed to sequentially compensate for aberrations
(i.e., DeAbe), deconvolve the resulting predictions ('DL Decon'), and improve resolution isotropy[5] ('DL
Iso', **Fig. 5a-d, Supplementary Figs. 26-29**). As expected, (**Fig. 5a**), the raw data showed increasing
depth-dependent degradation in resolution and contrast, which confounded our ability to discern
distinct nuclei or cell boundaries on the 'far' side of the volume. In comparison, the multi-step

procedure offered striking improvements in resolution and contrast in both nuclear and membrane
channels, largely alleviating the degradation (**Fig. 5a, b, Supplementary Figs. 27, 28, Supplementary**
Video 10). Ablation experiments in which one or more of the networks were removed produced inferior
results, further substantiating our hypothesis that the gains in image quality benefited from applying
DeAbe (**Supplementary Fig. 30**). In the membrane channel, the multi-step restoration enabled us to
automatically segment cell boundaries more accurately than in the raw data and further refine the
segmentations manually up to 421 cells (**Fig. 5c, Supplementary Fig. 29, Supplementary Video 11**),
exceeding previous efforts limited to the 350-cell stage[32]. Automated segmentation by successively
applying DeAbe and DL Decon additionally provided a cell count closer to manual ground truth[33] than
the raw data (**Fig. 5d**) or DL Decon alone, with DL Iso providing no benefit to automated segmentation
(**Supplementary Fig. 31**).
Next, we explored replacing the final network (DL Iso) with a network designed to further
enhance resolution based on ground truth acquired with expansion microscopy[9,34] ('DL Expan',
Supplementary Fig. 26b). After verifying that DL Expan improved resolution more than 2-fold on data
unseen by the model (**Supplementary Fig. 32**), we applied the new multi-step restoration method to *C.*
elegans embryos expressing a GFP-membrane marker labeling head neurons and gut cells (**Fig. 5e**).
Compared to the raw data, the enhanced resolution offered by the deep learning prediction better
resolved closely spaced membranes within and between cells (**Fig 5f-h, Supplementary Figs. 33, 34**).
This capability proved especially useful when tracking the development of neurites projecting in the
nerve ring, a neuropil that constitutes the brain of the animal, and which is composed of hundreds of
tightly packed interwoven neurites. While the position of the neurites within the neuropil determines
circuit identity and connectivity, the sequence of events leading to its innervation has not been
described because of limitations in resolving these structures. We focused our analyses on the closely
positioned neurons AIY and SMDD, which we identified based on morphology by comparison to labeled
images in ref.[35] and ref.[36]. SMDD is a central pioneering neuron in the nematode brain[36-38], while its
sister cell AIY[35] is a first layer interneuron[39] involved in thermotaxis and locomotion[40]. Observing both
neurons over our 120-minute recording, we found that SMDD's neurites grew out first, followed by AIY's
neurite. AIY's neurite entered the nerve ring after SMDD, consistent with the SMDD's role as a pioneer
neuron (**Fig 5i, Supplementary Video 12**). Such developmental dynamics were difficult or impossible to
observe in the raw data (**Supplementary Fig. 35**), or joint deconvolutions of the dual-view data due to
artifacts resulting from motion between the two views (**Supplementary Fig. 36**). To illustrate that these
gains in image quality can be extended to a different label imaged in a different microscope, we also
restored images of nuclei labeled with a GFP histone marker and acquired with high NA diSPIM[23], finding
similarly dramatic improvements in contrast and resolution (**Supplementary Fig. 37, Supplementary**
Videos 13, 14).
In these neuronal (**Fig. 5e-i, Supplementary Videos 12**) and nuclear (**Supplementary Fig. 37**,
Supplementary Videos 13, 14) recordings, although the inter-volume recording time spanned several
minutes, the volume acquisition time was 1 s and 1.2 s (10 ms and 20 ms per plane, respectively),
necessary to ameliorate motion blur in these rapidly repositioning[30] embryo samples. As DeAbe is
applied after data acquisition, there is no loss in temporal resolution relative to raw image capture. This
capability is advantageous over AO, which always entails additional temporal cost due to the need for
wavefront sensing and correction (e.g., several seconds for a single loop of correcting aberrations in the
AO-LLSM experiments presented in **Fig. 2**). While this cost may be acceptable for correcting aberrations
in static or slowly moving samples prior to image acquisition (by far the most common use case in AO

enabled microscopy), it is too slow for the highly dynamic embryos imaged here, which would ideally benefit from rapid AO correction at each plane, at each time point.

To further underscore this point, we used the iSIM to image adult worms with a GCaMP6 marker targeted to neurons. On anesthetized (**Supplementary Fig. 38**) or partially immobilized (**Supplementary Fig. 39**) worms, DeAbe restored fine structure otherwise masked by aberrations. When performing continuous volumetric recordings at 1.5 Hz, necessary to follow calcium transients in the moving worm head and pharynx, DeAbe improved quality sufficiently that we could resolve structural details in the nerve ring that was obscured in the raw data (**Supplementary Videos 15-17, Supplementary Fig. 39**). As for the embryos, such restoration is currently infeasible with AO, due to its slow speed.

Discussion

As we show on diverse microscopes and samples spanning multiple spatial and temporal scales, DeAbe can compensate for optical aberrations without recourse to AO: improving SNR, contrast, and resolution in fluorescence microscopy volumes without compromising the temporal resolution of data acquisition. We anticipate this capability will be useful for most labs, which lack access to sophisticated AO setups but still need to improve the quality of imaging volumes acquired using existing hardware. Besides improving the qualitative appearance of images (**Fig. 1-5**), which facilitates inspection of biological features deep within imaging volumes, DeAbe also quantitatively improves downstream image analysis. We highlight this capability by refining vessel segmentation in large, cleared tissue samples (**Fig. 4e-g**) and in enhancing the segmentation of densely packed nuclei and membranes in *C. elegans* embryos (**Fig. 5**). The latter capability may prove particularly useful in the creation or extension of 4D morphological atlases[32], which depend on high quality image data.

Several caveats are worth noting in the context of current limitations and with an eye towards future applications. First, as for any deep learning method, DeAbe provides a *prediction* at best and cannot fully recover lost information that is not present in the raw data. Second, the performance of DeAbe depends critically on the quality of the training data, and specifically on the assumption that fluorescently labeled structures are similar throughout the image volume. While this assumption was met for the samples in this work, we encourage caution when applying DeAbe on highly heterogenous specimens (or when applying DeAbe trained on one sample type to another, **Supplementary Fig. 15**), lest hallucinations arise. Third, although here we mainly trained on semi-synthetic data (**Fig. 2-5**), it would also be worth investigating how well the training derived from fully synthetic data[7] (**Fig. 1**) generalizes to experimental data. Such an approach might prove useful in ameliorating system aberrations introduced by microscope hardware. Fourth, we focused here on correcting depth-dependent aberrations, in which the training data was corrupted by a constant aberration in each image plane. A useful future direction would be to extend our approach to explicitly account for laterally varying aberrations, as such aberrations are problematic particularly for large specimens. Finally, although we used a mixture of random low-order aberrations to train our model, enhanced performance is likely if aberrations specific to the sample (or instrument) can be inferred and used in the training procedure (**Supplementary Fig. 4, 11, 15**).

Author Contributions

Conceived project and directed research: H.S. Implemented DeAbe framework: M.G. Designed simulations: M.G., Y.W., H.S. Wrote software: M.G., Y.W., J.L., X.Han, S.Q., Z.L. Designed experiments:

M.G., C.M.H., Y.S., R.C., X.M, A.Z., C.C., M.M., E.Y., A.B., D.C-R., H.S. Prepared samples: M.G., C.M.H., Y.S., E.K., R.C., G.K., J.B., M.C., L.Z., Z.Lu, X.M., A.Z., C.C., M.M., E.Y., A.B. Performed experiments: M.G., C.M.H., Y.S., E.K., R.C., G.K., J.B., M.C., X.Han, X.M., A.Z., C.C., M.M., E.Y., A.B. Designed and performed segmentation analysis: M.G., Y.S., J.L., X. Hou, S.Q., Z.Liu. All authors examined and analyzed data. Wrote manuscript: M.G. and H.S., with advice from all authors. Provided biological insight and advice: R.C., X.M., A.Z., C.C., M.M., E.Y., A.B., D.C-R. Supervised research: E.Y., H.L, Z.Liu, A.B., P.L-R., D.C-R., H.S.

Acknowledgments

We thank Oliver Hobert for supporting the NeuroPAL work in his lab and allowing us to use the data generated therein for this paper; Robert Weigert for the gift of the fixed mouse liver sample; Manuel Zimmer for providing strain ZIM1997; Daniela Malide for helping to prepare and image the cleared mouse tissue datasets; Leanna Eisenman for the PtK2 cell sample preparation; Emmanuel Marquez-Legorreta for helping us to prepare the zebrafish samples; Teng-Leong Chew, the Advanced Imaging Center, and the Light Microscopy facility at HHMI Janelia Research Campus for supporting experiments with the AO-LLSM system; Steve Coleman for assisting us with the Visitech iSIM; Dan Milkie for his assistance with generating and interpreting wavefront images from the modified AO-LLSM system; Nikolaj Reiser for helpful discussions; and Courtney Johnson and Xuesong Li for their comments on the manuscript. This research was supported by the intramural research programs of the National Institute of Biomedical Imaging and Bioengineering and the National Heart, Lung, and Blood Institute within the National Institutes of Health (NIH). This work was supported by the Howard Hughes Medical Institute (HHMI). This article is subject to HHMI's Open Access to Publications policy. HHMI laboratory heads have previously granted a non-exclusive CC BY 4.0 license to the public and a sub-licensable license to HHMI in their research articles. Pursuant to those licenses, the author-accepted manuscript of this article can be made freely available under a CC BY 4.0 license immediately upon publication. This research is funded in part by the Gordon and Betty Moore Foundation. We thank the Office of Data Science Strategy, NIH, for providing a seed grant enabling us to test and validate the initial deep learning frameworks using cloud-based computational resources. H.S., P.L.R. and D.C.-R. acknowledge the Whitman and Fellows program at MBL for providing funding and space for discussions valuable to this work. M.G. acknowledges the funding support from the Hundred Talents Program of Zhejiang University. We thank the Janelia Visiting Scientist program for supporting M.G. and A.B. Z.L. acknowledges the funding support from Natural Science Foundation of Zhejiang Province (LR20F050001). H.L. acknowledges the funding support from the National Key Research and Development Program of China (2020AAA0109502), the National Natural Science Foundation of China (U1809204) and the Talent Program of Zhejiang Province (2021R51004). E.Y. acknowledges support from the Esther A. and Joseph Klingenstein Fund, the Simons Foundation, and the Hypothesis Fund. X.M. and A.Z. were supported by Intramural FDA funding. A.B. acknowledges UKRI BBSRC (project grant BB/S017127/1).

Methods

Deep learning-based de-aberration model

Building a de-aberration model (DeAbe) requires appropriate training data and the use of a
neural network. First, based on the physics of image formation, we derived forward imaging models that
allowed us to synthetically aberrate the data produced for multiple systems, including wide field, light
sheet, confocal, two photon, and super-resolution structured illumination microscopes (**Supplementary**
Note 1). Second, we extracted subvolumes from the shallow side of the experimentally acquired image
stacks, using these data as ground truth; alternatively, when we could obtain whole aberration-free
volumes, we used them as ground truth (e.g., aberration-free images of synthetic phantoms in **Fig. 1b,**
Supplementary Figs. 1-9, and stacks of cells in **Fig. 2a and Supplementary Figs. 10-13,** where the
aberrations are negligible due to the thickness of the sample). Third, based on the forward imaging
models, we synthetically added aberrations to the ground truth images so that they resembled
aberrated data present deeper within the image stacks. Together, the paired ground truth data and
associated synthetically degraded data constitute training pairs. Fourth, we used these training pairs in
conjunction with our 3D RCAN network[9] to train a DeAbe model to reverse the effect of synthetic
aberration. Finally, we applied the trained network to reduce the effects of aberrations in
experimentally acquired image volumes unseen by the network.
We define the 'shallow side' of an image stack by the planes nearest to the detection objective,
which are typically contaminated with least aberration and thus offer the best image quality. We then
selected subvolumes on the shallow side ('shallow subvolumes') by visually inspecting image quality in
real and Fourier space. We also examined quantitative metrics for this choice, finding that our visual
impression usually coincided with a resolution degradation of ~20% (**Supplementary Fig. 16, 17,**
Supplementary Table 1). We extracted shallow subvolumes from image stacks by manually cropping
with ImageJ when image size and content differed substantially across a given specimen type, or
automatically with customized ImageJ macros when considering specimens with more stereotyped
image size and content (e.g., as for time-lapse image volumes). For the cleared mouse embryo images
(**Fig. 4**), the shallow subvolumes were further divided into smaller subvolumes (~80 MB/volume) due to
their large volume size in raw data (**Supplementary Table 1**).
As described in **Supplementary Note 1**, we expressed the aberrated wavefront $\phi(r,\theta)$ at the
back focal plane of the objective using Zernike basis functions $\phi_m(r,\theta)$ and associated Zernike
coefficients c_m

$$\phi(r,\theta) = \sum_{m=0}^{M} c_m \phi_m(r,\theta), \qquad\qquad (1)$$

with M the maximum Zernike index chosen in our aberration.
We generated synthetic aberrations by using semi-randomly generated Zernike coefficients (**Fig.**
1a). We used the ANSI convention[41] when indexing the Zernike coefficients, customizing aberrations by
using different Zernike coefficients for different datasets acquired from different microscopes. For all
experimental datasets, we added aberrations up to the 4th Zernike order (i.e., M = 14), except for piston
and tilt components (Z = 0, 1, 2). The amplitudes of the Zernike coefficients were randomly generated,
but subject to pre-defined bounds. We initially set an upper bound of 0.5 rad for all Zernike coefficients,
then added an additional 1 rad for defocus (Z = 4) and spherical (Z = 12) components to mimic the more
severe contamination caused by defocus and spherical aberrations commonly encountered in
experimental datasets, i.e:

$$\begin{cases} c_z = 0, & for\ Z = 0,1,2 \\ |c_z| \leq 1.5, & for\ Z = 4,12 \\ |c_z| \leq 0.5, & otherwise\ for\ Z \leq M, \end{cases} \quad (2)$$

with M = 14 for all experimental datasets.

For each shallow side subvolume, 10 independent sets of aberrations were generated and used
for synthetic degradation, thereby augmenting the data 10-fold. Processing was performed with custom
MATLAB code (MathWorks, R2022b), with further details provided in the *Code availability* section.

We employed 3D RCAN (https://github.com/AiviaCommunity/3D-RCAN), appropriate for 3D
image volumes, for generating the DeAbe model based on the training data pairs. We trained individual
DeAbe models for each microscope and each sample type. For training, we set the number of epochs to
200; the number of steps per epoch to 400; the training patch size to 64 × 64 × 64; the number of
residual blocks to 5; the number of residual groups to 5; and the number of channels to 32. When
applying the model, the patch size was set to 256 × 256 × 256. Image volumes larger than this patch size
were divided into patches, the network applied to each patch, and the patches stitched together via
linear blinding to minimize boundary artifacts[8] (unless specified otherwise, we used this setting for
applications of 3D RCAN). Training and model application was performed within Python 3.7.0 on a
Windows 10 workstation (CPU: Intel Xeon, Platinum 8369B, two processors; RAM: 256 GB; GPU: NVIDIA
GeForce RTX 3090 with 24 GB memory). More details on datasets and training parameters are listed in
Supplementary Table 1.

Multi-step image restoration with deep learning

The multi-step image restoration pipeline combines the DeAbe model with two additional
networks to progressively improve image resolution and contrast: (1) the DeAbe model to reverse
degradation from aberrations ("DL DeAbe"); (2) a deconvolution network designed to mimic the image
quality improvement afforded by multiview imaging ("DL Decon", see the section *Deep learning-based*
deconvolution); (3) an axial resolution enhancement network to improve resolution isotropy ("DL Iso",
see the section *Deep learning-based axial resolution enhancement*); or a network designed to predict
the improved resolution provided by expanded samples ("DL Expan", see the section *Deep learning-*
based expansion).

Deep learning-based deconvolution

As for our previous attempts at deep-learning based multiview deconvolution[8], we used a
single-view image volume as input, and attempted to restore image resolution and contrast that
approximated the result from multiview joint deconvolution. The training data were acquired by dual-
view light sheet microscopy[30], either a 'symmetric' diSPIM equipped with 0.8/0.8 NA objectives[31] (**Fig 5e-**
i, Supplementary Figs. 30, 32-36) or a higher NA 'asymmetric' diSPIM equipped with 1.1 / 0.67 NA
objectives[23] (**Fig 5a-d, Supplementary Figs. 27-29, 37**). First, raw images were de-aberrated with the
DeAbe model. Then de-aberrated images from the two views were jointly deconvolved to achieve
reconstructions with near isotropic spatial resolution and good image quality throughout the
reconstruction. With training data consisting of the single-view de-aberrated images as input and the
jointly deconvolved images as ground truth, we then used another 3D RCAN for the deconvolution
model (DL Decon). For all datasets, the number of epochs for training was 200; the number of steps per
epoch was 400; the training patch size was 64 × 64 × 64; the number of residual blocks was 5; the

number of residual groups was 5; and the number of channels was 32. Training and model application
was performed within Python 3.7.0 on a Windows 10 workstation (CPU: Intel Xeon, Platinum 8168, two
processors; RAM: 512 GB; GPU: Nvidia Quadro RTX6000 with 24 GB memory). We note that although
training DL Decon required dual-view image volumes, applying DL Decon needs only single-view image
volumes acquired from single-view light sheet microscopy (iSPIM)[22].

Deep learning-based axial resolution enhancement
The images predicted by the DL Decon model were not perfectly isotropic, i.e., the axial
resolution (although improved over the raw input images) is worse than the lateral resolution. Thus, for
some experiments we used an additional network to enhance axial resolution (DL Iso, **Fig. 5a, b,**
Supplementary Figs. 27-30, Supplementary Videos 10, 11). CARE[5] software
(https://github.com/CSBDeep/CSBDeep) was employed to train the a 'DL Iso' model based on the
predictions derived from serially applying the DeAbe and Decon models to raw input images. We used
100 3D volumes, each spanning 360 × 480 × 310 voxels, for training data. Training was performed on the
xy planes (lateral views), using a 2D PSF (consisting of a point blurred with a 1D Gaussian function, sigma
= 2.5 pixels along the y dimension) an axial downsampling factor of 6, and a patch size of 64 × 64 to
create training pairs. The training was performed within Python 3.7.0 on a Windows 10 workstation
(CPU: Intel Xeon, Platinum 8168, two processors; RAM: 512 GB; GPU: Nvidia Quadro RTX6000 with 24
GB memory).

Deep learning-based expansion
As an alternative to DL Iso, we also trained a model to improve the resolution based on data
acquired with expansion microscopy (DL Expan). First, physically expanded samples (**Supplementary Fig.**
32) were imaged on the symmetric 0.8 NA diSPIM. Second, dual-view raw images were jointly
deconvolved and used as ground truth images. Third, the ground truth images were synthetically
degraded to resemble low-resolution conventional images acquired on the diSPIM, following our
previous procedure[9]. Last, the 3D RCAN network was employed to train the DL Expan model based on
the training data (i.e., synthetically degraded and ground truth pairs).
For the worm embryo data with DAPI labeled nuclei (**Supplementary Fig. 37**), dual-view raw
image volumes from 15 expanded worm embryos were acquired and jointly deconvolved to produce 15
high-resolution image volumes. These 15 volumes were then synthetically degraded to generate low-
resolution images. For the worm embryo data with TTX3B neurites labeled (**Fig 5e-i**, **Supplementary**
Figs. 32-35), dual view image volumes from 71 expanded worm embryos were acquired and manually
cropped to select regions containing TTX3B neurites (this was necessary given the sparsely labeled
neurites present in the raw images). Cropped images were jointly deconvolved to produce 71 high-
resolution image volumes. These 71 volumes were then synthetically degraded to generate synthetic
low-resolution image data. For each dataset, the low-resolution and high-resolution paired volumes
were then used to train the 3D RCAN based DL Expan model. The number of epochs for training was set
to 300; the number of steps per epoch to 400; the training patch size to 64 × 64 × 64; the number of
residual blocks to 5; the number of residual groups to 5; and the number of channels to 32. The training
was performed within Python 3.7.0 on a Windows 10 workstation (CPU: Intel Xeon, Platinum 8369B, two
processors; RAM: 256 GB; GPU: NVIDIA GeForce RTX 3090 with 24 GB memory).

Simulations on phantom objects

To evaluate the quality and performance of our DeAbe model, we generated 3D phantom
objects consisting of five types of structures in MATLAB (Mathworks, R2022b, with the Image Processing
Toolbox): dots, lines, circles, spheres, and spherical shells[27]. Phantoms were randomly oriented and
located in a volume of 256 × 256 × 256 voxels, with voxel size 0.13 × 0.13 × 0.13 μm³. We simulated the
blurring introduced by light sheet microscopy (**Supplementary Note 1**) by convolving the phantom with
an ideal, noise-free PSF resembling that of our light sheet system (with 1.1 NA water dipping objective,
detection wavelength of 0.532 μm and an illumination light sheet thickness of 2 μm). Aberrated data
was generated by altering the ideal PSF according to the synthetic aberration procedure described
above.
To create synthetic aberrations, we adopted Equation (1) and generated Zernike coefficients
semi-randomly in MATLAB, with each Zernike coefficient c_m subject to a pre-defined upper bound T_m:

$$|c_m| \leq T_m, \ for \ m \leq M, \tag{3}$$

485   with m the Zernike index following the ANSI convention and M the maximum Zernike index.
486   We omitted piston and tilt components (m = 0, 1, 2) and weighted lower order Zernike components
487   (Defocus m =4, astigmatism m=3,5, and spherical m=12) more as these aberrations are commonly
488   observed in real samples:

$$T_m = \begin{cases} 0, & for \ m = 0,1,2 \\ 1.5, & for \ m = 3,4,5,12 \\ 0.5, & otherwise \ for \ m \leq M, \end{cases} \tag{4}$$

490   with M defined based on the desired Zernike order:

$$M = \begin{cases} 9, & for \ Zernike \ order \ of \ 3 \\ 14, & for \ Zernike \ order \ of \ 4 \\ 20, & for \ Zernike \ order \ of \ 5 \\ 27, & for \ Zernike \ order \ of \ 6 \\ 35, & for \ Zernike \ order \ of \ 7 \end{cases} \tag{5}$$

492   For **Supplementary Fig. 2**, we varied M to explore the effect of different Zernike orders on de-aberration
493   performance by setting $M = 9, 14, 20, 27,$ and 35 corresponding to Zernike orders 3-7. For all other
494   simulations, we set $M = 14$.
495   The Root Mean Square (RMS) wavefront distortion of an aberration with Zernike coefficients c_m ($m =$
496    $3, 4,5, \dots, M$) is:

$$RMS_c = \sqrt{\sum_{m=3}^{M} c_m{}^2} . \tag{6}$$

498   The RMS wavefront distortion for aberrations defined by upper bounds T_m ($m = 3, 4,5, \dots, M$) is:

$$RMS_T = \sqrt{\sum_{m=3}^{M} T_m{}^2} . \tag{7}$$

500   To create training data, we synthetically aberrated phantoms with two types of aberrations:
501   1) a random mixture of aberrations containing different Zernike components, with the
502   amplitude of the aberrations subject to upper bounds. This type of aberrations was first generated with
503   a set of initial Zernike coefficients c_m based on Equations (3-5), and then rescaled to a maximum RMS of
504   Ω wavefront distortion (e.g., $\Omega = 1,2,$ or 4 rad) to obtain the final Zernike coefficients $c_{m-final}$:

$$c_{m-final} = \frac{\Omega}{RMS_T} c_m, \quad for \; m \leq M. \tag{8}$$

506 These aberrated training data were used to train the general DeAbe models (i.e., all but the model
507 trained to counter the defocus mode specifically) used in all figures and videos showing simulated
508 phantoms.
509   2) a single aberration mode of defocus with amplitude subject to upper bounds, i.e., the upper
510 bounds of each Zernike coefficient were zeros except for the defocus mode (m =4):

$$T_m = \begin{cases} 1.5, & for \; m = 4 \\ 0, & otherwise \; for \; m \leq M. \end{cases} \tag{9}$$

By replacing Equation (4) with Equation (9), we could generate the defocus aberration the same way as
for the first aberration type (1). These training data were only used to train the specific defocus DeAbe
model used in **Supplementary Fig. 4**.
For each training session, we created 50 phantoms, each consisting of different random objects.
For each phantom, we generated 10 independent aberrated images with each image containing random
mixtures of aberrations (**Fig 1, Supplementary Figs. 1-9, Supplementary Videos 1-2**) or only defocus
aberrations (**Supplementary Fig. 4**), for a total of 500 training data pairs per session. We also added
Poisson noise to the aberrated images by defining the SNR as

$$SNR = \sqrt{S}, \tag{10}$$

where S is the signal defined by the average of all pixels with intensity above a threshold (here set as 1%
of the maximum intensity of the blurred objects in the noise-free image).
We employed 3D RCAN to train the DeAbe model based on simulated training data. We set the
number of epochs to 200; the number of steps per epoch to 400; the training patch size to 64 × 64 × 64;
the number of residual blocks to 5; the number of residual groups to 5; and the number of channels to
32. Training was performed with Python 3.7.0 on a Windows 10 workstation (CPU: Intel Xeon, Platinum
8369B, two processors; RAM: 256 GB; GPU: NVIDIA GeForce RTX 3090 with 24 GB memory).
To benchmark the performance of the DeAbe model, we created synthetic phantoms with three
types of aberrations:
1) a random mixture of aberrations containing different Zernike components, with the
amplitude of the aberrations subject to upper bounds. This type of aberration is the same used for
training the general DeAbe models and was generated following Equations (3-5) and (8). This aberration
mixture was used in **Supplementary Fig 2**.
2) a random mixture of aberrations containing different Zernike components, with the
amplitude of the aberrations fixed at a certain RMS value. This aberration mixture was first generated
with a set of initial Zernike coefficients c_z based on Equations (3-5), and then rescaled to a fixed
amplitude with RMS Υ (e.g., $\Upsilon = 1, 2, or \; 4$ rad) wavefront distortion to obtain the final Zernike
coefficients $c_{m-final}$:

$$c_{m-final} = \frac{\Upsilon}{RMS_c} c_m, \quad for \; m \leq M. \tag{11}$$

This aberration mixture was used for **Fig 1, Supplementary Figs 1,3,5, and Supplementary Videos 1-2**.
3) single aberration modes with a fixed RMS value, i.e., Zernike coefficients were set to zero
except for the desired aberration mode. The single aberration modes tested in the paper include
defocus (m=4), astigmatism (m=3,5), coma (m=7,8), trefoil (m=6,9), and spherical (m=12). If the RMS
wavefront distortion is defined as Υ (e.g., $\Upsilon = 1, 2, or \; 4$ rad), each single aberration mode's Zernike
coefficients are:

Defocus: $c_4 = \Upsilon$, $otherwise\ c_m = 0\ for\ m \leq M$

Astigmatism: $\sqrt{c_3{}^2 + c_5{}^2} = \Upsilon$, $otherwise\ c_m = 0\ for\ m \leq M$

Coma: $\sqrt{c_7{}^2 + c_8{}^2} = \Upsilon$, $otherwise\ c_m = 0\ for\ m \leq M$

Trefoil: $\sqrt{c_6{}^2 + c_9{}^2} = \Upsilon$, $otherwise\ c_m = 0\ for\ m \leq M$

Spherical: $c_{12} = \Upsilon$, $otherwise\ c_m = 0\ for\ m \leq M$

These aberrations were used to test the DeAbe performance on single aberration modes

(**Supplementary Figs. 3,4**).

For quantitative analysis, we used structural similarity index (SSIM) and peak signal-to-noise

ratio (PSNR) to evaluate the restored images provided by deep learning as well as by traditional

deconvolution. The SSIM and PSNR were calculated based on image volumes with MATLAB (Mathworks,

R2022b). Their mean value and standard deviation were computed from 100 simulations, each with

random object structures and input aberrations.

To benchmark the performance of DeAbe using different neuronal networks, we compared our

default 3D-RCAN choice with three other state-of-the-art 3D networks including CARE, RLN, and

BasicVSR++. For a fair comparison, the training data pairs of phantom objects for **Fig. 1** (generated with

random mixtures of aberrations) were used to train the CARE, RLN, and BasicVSR++ in addition to 3D-

RCAN. Then models trained using different networks were applied to aberrated images and the

prediction results compared in **Supplementary Figs. 8, 9**. 1) The CARE package was downloaded from

https://github.com/CSBDeep/CSBDeep. The patch size was set to a 3D shape of 64 × 64 × 64 and the

patch number was set to 32; the training epoch was 50 and the training steps per epoch was 30; and all

other parameters were set to default values. 2) The RLN package was downloaded from

https://github.com/MeatyPlus/Richardson-Lucy-Net. The training files and folders were reorganized to

fit the input format as required by RLN. All training parameters were set to default values. 3) The

BasicVSR++ package was downloaded from https://github.com/XPixelGroup/BasicSR. The batch size was

set as 2 and the patch size of the 3D shape was 10 × 256 × 256; the learning rate for all modules was set

to 1×10^{-4}; and all other parameters were set at default values.

To distinguish de-aberration from denoising (**Supplementary Figs. 6, 7, 13**), we compared

DeAbe performance with nonlocal means (NLM) and an unsupervised deep learning network,

Noise2Void (N2V). The NLM denoising algorithm was implemented using the OpenCV library

(https://docs.opencv.org/3.4/d5/d69/tutorial_py_non_local_means.html). We used the function

fastNlMeansDenoising with the parameters h as 5, templateWindowSize as 7, and searchWindowSize as

21. The N2V package was downloaded from https://github.com/hanyoseob/pytorch-noise2void. The

training files and folders were reorganized to fit the input format as required by N2V. The training epoch

was 5000 and the batch size was 4; and all other parameters were set to default values.

Preprocessing, attenuation correction, traditional deconvolution, and multiview fusion

Raw images acquired with iSIM and light sheet imaging were preprocessed by subtracting a

uniform background with intensity equivalent to the average of 100 dark (no excitation light)

background images. When diSPIM was operated in stage scan mode, the images were also deskewed to

correct the distortion induced by stage-scan acquisition before further processing.

For the cleared mouse embryos imaged with confocal microscopy (**Fig 4, Supplementary Fig. 23,**
Supplementary Video 8) and nematodes imaged with iSIM (**Supplementary Fig. 20**), raw data was
additionally preprocessed with intensity attenuation correction. The attenuation correction was
performed by multiplying the raw intensity values with an exponential compensation factor:
$$I(z) = I_0(z)e^{\alpha z} \tag{11}$$
with $I_0(z)$ the raw intensity, z the depth and α the attenuation factor. We set $\alpha = 0.01$ for all datasets.
For the comparison of DeAbe with traditional deconvolution, we implemented both Richardson-
Lucy (RL) deconvolution[42,43] (**Fig. 1c-f, Fig. 3, Supplementary Figs. 12, 19, 21, 23**) and blind
deconvolution[16] (**Fig. 1c-f, Supplementary Fig. 12**) on the raw aberrated images. For blind
deconvolution, we used the MATLAB function $deconvblind$ with default settings
(https://www.mathworks.com/help/images/ref/deconvblind.html). For RL deconvolution, we adopted
our previously developed deconvolution package[8] (https://github.com/eguom/regDeconProject). In one
synthetic dataset ('RL Decon 2', **Fig. 1c-f**), we used an aberrated PSF that was generated as described in
Supplementary Note 1 and matched the aberrations in the synthetic dataset; otherwise, we used an
aberration-free ideal PSF for all other datasets (**Fig. 1c-f, Fig. 3 and Supplementary Figs. 19, 21, 23**).
Additionally, we also performed RL deconvolution on several datasets after DeAbe processing (**Fig. 3b-d,**
Supplementary Fig. 19, 23), setting the number of iterations to 20 unless specified otherwise. All
deconvolution was performed in MATLAB (MathWorks, R2022b) on a Windows 10 workstation (CPU:
Intel Xeon, Platinum 8369B, two processors; RAM: 256 GB; GPU: NVIDIA GeForce RTX 3090 with 24 GB
memory).
For data acquired by diSPIM, we performed multiview fusion on several datasets either for
generating DL training data (**Fig. 5**, **Supplementary Figs. 27, 28, 33-35, 37**) or for comparisons to the DL
Decon model (**Supplementary Figs. 30, 36**). The diSPIM data typically contain two view volumes,
referred to as View A and View B volumes. The multiview fusion process involves registration and joint
deconvolution to combine two views into a single volumetric image stack with improved resolution. The
registration first rotates View B by 90 degrees along the Y-axis to align View B's orientation with View A
and then maximizes the cross-correlation function between View A and View B with affine
transformations. After registration, View A and registered View B were deconvolved jointly using a
modified Richardson–Lucy deconvolution algorithm as previously described[30]. Multiview fusion was
achieved using custom software (https://github.com/eguom/diSPIMFusion) on a Windows 10
workstation (CPU: Intel Xeon, Platinum 8369B, two processors; RAM: 256 GB; GPU: NVIDIA GeForce RTX
3090 with 24 GB memory).
621
Sample preparation and imaging
Live nematode embryos imaged with light sheet microscopy
Nematode strains were kept at 20°C, and grown on NGM media plates seeded with *E. coli* OP50.
Strains used in this paper included BV514 (ujIS113 [*pie-1p::mCherry::H2B + unc-119(+); Pnhr-*
82::mCherry::histone + unc-119(+)]), OD58 (ltIs38 [*pie1p::GFP::PH(PLC1delta1) + unc-119(+)*]), DCR6268
(olaEx3632 [*pttx-3b::SL2::PHD::GFP:: unc-54 3' UTR + pelt-7::mCh::NLS::unc-54 3' UTR*]), and SLS164
(ltIS138[*pie-1p::GFP::PH(PLC1delta1) + unc-119(+)*]; ujIS113 [*pie-1p::mCherry::H2B + unc-119(+); Pnhr-*
82::mCherry::histone + unc-119(+)]). SLS164 was made by crossing together strains BV514 and OD58 and
may have unc-119(ed3) III in the background. Strains BV514 and OD58 were gifts from Dr. Zhirong Bao.
Nematode samples were prepared for diSPIM imaging as previously described[22,31,44]: gravid adult
hermaphrodites were picked into a watch glass with M9 buffer, adults were cut in half to liberate

embryos, and embryos were transferred onto a poly-L-lysine coated coverslip in a diSPIM imaging
chamber. For strain DCR6268 ((olaEx3632 [*pttx-3b::SL2::PHD::GFP:: unc-54 3' UTR + pelt-*
7::mCh::NLS::unc-54 3' UTR]), labeling neuron and gut cells), embryos were imaged once they reached
the bean stage of development using a fiber-coupled symmetric diSPIM (with 0.8NA/0.8NA
objectives)[31]. Volumes were captured once per minute over two hours in light sheet scan mode. Each
volume comprised 50 slices, with a 1 μm step size and a total acquisition time per volume of ~1 second.
For strain SLS164 (labeling cell membrane and nuclei), embryos were imaged from the 2- or 4-cell stage
using a fiber-coupled asymmetric diSPIM (with 1.1NA/0.67NA objectives)[23]. Volumes were captured
once every 3 minutes over 450-minute duration in stage scan mode. Each volume comprised 70 slices,
with a 1.1 μm stage step size and a total acquisition time of ~1.4 s per volume. For strain BV514 (labeling
cell nuclei), embryos were imaged from the bean stage to hatching using the asymmetric diSPIM.
Volumes were captured every 5 minutes in stage scan mode. Each volume comprised 60 slices, with a
1.4 μm stage step size and a total acquisition time per volume of ~1.2 seconds. For strain OD58 (labeling
cell membranes), embryos were imaged from the 4- or 8-cell stage using a symmetric diSPIM. Volumes
were captured once every 3 minutes over a 450-minute period in light sheet scan mode. Each volume
comprised 45 slices, with a 1 μm step size and a total acquisition time per volume of ~0.9 seconds. For
all imaging, images were acquired using 488 nm excitation (for GFP labels) or 561 nm excitation (for
mCherry labels).
651
Expanded nematode embryos
C. elegans embryos from strain DCR6268 (labeling neurites and gut cells) were immobilized on
Poly-L-Lysine (PLL) coated glass bottom dishes, bleached, digested by yatalase, fixed, and expanded. The
procedure takes approximately 2 days, and is adapted from our published method[27].
First, glass bottom dishes were coated with PLL. PLL (Sigma, Cat# P5899) powder was
reconstituted in distilled water to 1mg/mL, aliquoted, and stored at -20°C. Prior to experiments, 30-50
μL of PLL was placed on the glass bottom dish (MatTek, Cat# P35G-1.5-14-C) and air dried at room
temperature (RT). Coated coverslips were usually prepared up to 1 day before pre-treatment of *C.*
elegans for expansion microscopy.
Second, embryos were digested, fixed, and stained with DAPI. Gravid adult *C. elegans* worms
were deposited in a petri dish in PBS buffer and cut with a surgical blade to release eggs. Eggs were
immobilized on a PLL coated glass bottom dish in PBS and could be processed immediately or stored at
25°C in M9 buffer until the embryos developed to the desired stage. Embryos were treated with a
bleaching mixture containing 1% sodium hypochlorite (Sigma, Cat# 425044) in 0.1M NaOH/water for 2-3
minutes, rinsed 3 times in PBS, digested in 50 mg/mL Yatalase in PBS (Takara Bio, Cat# T017) for 40
minutes at RT and rinsed 3 times with PBS. It was important to treat eggs with bleach only after
immobilization on the PLL surface, otherwise embryos tended to detach from the glass at later steps.
Digested embryos were fixed in 4% paraformaldehyde/PBS (Electron Microscopy Sciences, Cat#
RT15710) for 1 hour, then rinsed 3 times with PBS to remove fixative. Fixed embryos were
permeabilized in 0.1% Triton X-100/PBS (Sigma, Cat# 93443) for 1 hour at RT with 1 μL/mL of DAPI
(Thermo Fisher Scientific, Cat# D1306).
Optionally, GFP signal can be boosted by immunolabeling. Yatalase digested embryos were
permeabilized with staining buffer (0.1% Triton X-100/PBS) for 1 hour before immunolabeling. Embryos
were stained by an anti-GFP primary antibody (Abcam, Cat# ab290) in the staining buffer at 4°C
overnight at 1 μg/mL. After primary antibody labeling, embryos were washed 3 times (30 min intervals

between washes) in the staining buffer and labeled using donkey-anti-rabbit-biotin secondary antibody
(Jackson ImmunoResearch, Cat# 711-067-003) in the staining buffer at 4°C overnight at 1 µg/mL. After
secondary antibody labeling, the embryos were washed 3 times in the staining buffer (30 mins intervals
between washes) and labeled with Alexa Fluor 488 Streptavidin in the staining buffer at 4°C overnight at
2 µg/mL (Jackson ImmunoResearch, Cat# 016-540-084). Labeled embryos were washed 3 times in the
staining buffer (30 minutes between washes) before being processed for expansion microscopy.
Immunolabeling was only performed on the data shown in **Supplementary Fig 32a**.
684   Finally, embryos were expanded. Embryos were treated with 1 mM MA-NHS (Sigma, Cat#
730300) in PBS for 1 hour at RT. Samples were rinsed 3 times in PBS, and treated with monomer
solution, which was made up of acrylamide (Sigma, Cat# A9099), sodium acrylate (Santa Cruz
Biotechnology, Cat# 7446-81-3), N, N'-methylenebis(acrylamide) (Sigma, Cat# 146072) and 4-Hydroxy-
TEMPO (Sigma, Cat# 176141), diluted with PBS, with a final concentration of 10%, 19%, 0.1%, and 0.01%,
respectively. After the treatment for 1 hour at RT, the monomer solution was replaced by gelation
solution. The gelation solution shared the same reagents and concentrations as monomer solution, with
the addition of tetramethylethylenediamine (TEMED, Thermo Fisher Scientific, Cat# 17919, reaching a
final concentration of 0.2%) and ammonium persulfate (APS, Thermo Fisher Scientific, Cat# 17874,
reaching a final concentration of 0.2%). APS was added at last, and the fresh gelation solution was
immediately applied to the embryos sandwiched between the glass bottom dish and another coverslip
surface for 2 hours at RT. It was important to control the gelation speed with 4-hydroxy-TEMPO as
premature gelation can distort embryos and result in poor expansion quality. The polymerized embryo-
hydrogel hybrid was cut out by a razor blade and digested with 0.2 mg/mL Proteinase K (Thermo Fisher
Scientific, Cat# AM2548) in digestion buffer (0.5 M sodium chloride (Quality Biological, Cat # 351-036-
101); 0.8 M guanidine hydrochloride (Sigma, Cat# G9284); and 0.5% Triton X-100) at 45°C overnight.
Digested embryos were expanded ~3.3-3.7 fold in distilled water, exchanging the water every 30 min
until expansion was complete. Expanded samples were flipped over so that embryos were 'on top'
(suitable for diSPIM imaging), mounted on PLL coated #1.5 coverslips (VWR, Cat# 48393-241) and
secured in an imaging chamber filled with distilled water. Finally, samples were imaged using the
symmetric 0.8/0.8 NA diSPIM in stage scan mode. Depending on the orientation of embryos, ~200-300
planes were acquired for each embryo, with 1.414 µm stage step size and 20 ms per-plane exposure
time.
707
708   *PtK2 cells imaged with adaptive optical lattice light-sheet microscopy (AO-LLSM)*
709   PtK2 cell samples were prepared by placing one 25 mm round coverslip (Warner Instruments,
CS-25R17) into a 35 mm culture dish (Corning, 430165) and seeding cells at 100k cells per dish the day
before fixation. Cells were washed quickly 3 times with pre-warmed PBS before fixing in 4%
formaldehyde for 5 minutes at room temperature. 3 additional PBS washes were performed, and cells
were permeabilized in 0.1% IGEPAL (Sigma-Aldrich, I8896) for 5 minutes at room temperature. Cells
were washed with PBS 3 times, after which 250 µl of a primary antibody solution of 0.1% iGf-free BSA
(Jackson ImmunoResearch, 001-000-162) and 1:400 Phalloidin Alexa Fluor 488 (ThermoFisher Scientific,
A12379) in PBS was added to each coverslip. Cells were incubated at 37C for 1 hour, and a final wash of
PBS with 0.05% Tween-20 (Sigma-Aldrich, P1379) and 2 additional PBS washes were performed.
718   Cells were imaged in PBS on a modified adaptive optical lattice light-sheet microscope[18,19]. First,
a system correction was performed as previously described[19]. Lattice light sheet excitation was
performed using a 488 nm laser line, a Thorlabs TL20x-MPS 0.6 NA objective lens, and a square lattice

pattern (Outer NA: 0.4, Inner NA: 0.3, Cropping: 10, Envelope: 5). Image stacks (256x1500 pixel field of
view (FOV) with 401 z steps) were acquired by scanning the sample stage horizontally at an angle of
32.45° relative to the optical axis of the detection objective (Zeiss Plan-Apo 20x, NA 1.0 DIC M27 75 mm)
with a step size of 0.4 μm and an exposure time of 20 ms. Emission light was filtered through a Semrock
BrightLine 523/40-25 emission filter and reflected onto a Hamamatsu Orca Flash 4.0 sCMOS camera via
a Semrock Di03-R561-t3-32x40 dichroic. After data collection, images were deskewed using a custom
analysis pipeline (https://github.com/aicjanelia/LLSM). The final voxel size after deskewing was 0.108 x
0.108 x 0.215 nm.
For training data, 40 random FOVs were selected and imaged as described above. For aberration
experiments (**Fig. 2a-f, Supplementary Figs. 10-12**), a random FOV was selected and a ground truth data
set was acquired. Next, an aberration was applied to the deformable mirror (DM; ALPAO DM69). These
aberrations were either random, wherein each actuator on the mirror was pushed or pulled by a
random amount with a fixed maximum amplitude, or a predefined Zernike mode (astigmatism, coma, or
spherical). For each type of aberration, 3 different magnitudes were used, and for each magnitude 3
different FOVs were selected, yielding a total of 36 experiments. After the aberration was applied to the
DM, a stack was collected. The microscope configuration was then changed to the adaptive optics (AO)
configuration.
The methods for AO correction have been described previously[19]. A focused two photon
(Coherent 1335240 Chameleon) spot was directed through the detection objective and scanned through
the same FOV to be imaged. The collected emission was passed through a microlens array and imaged
to the same camera used for image collection to function as a Shack-Hartmann (SH) wavefront sensor.
The distance each spot in the SH image moves is calculated relative to a reference image, after which
the DM is updated to correct the measured aberration. This process is repeated 2 additional times as
the AO correction will iteratively improve until it converges. The microscope is then switched back to
LLSM mode, and a final stack is acquired.
For comparative denoising experiments (**Supplementary Fig. 13**), a random FOV was selected
and a ground truth stack was acquired. Then, aberrations (random, astigmatism, and coma) were
applied to the DM at a single magnitude; 3 separate FOVs were examined per aberration. Once the
aberration was applied to the DM, stacks were acquired with the original laser power (high SNR,
Supplementary Fig. 13c) as well as 1/5 laser power (low SNR, **Supplementary Fig. 13b**).
751
Zebrafish embryos imaged with adaptive optical lattice light-sheet microscopy
Transgenic Zebrafish Tg(vGlut2a:Gal4); (UAS:CoChR-eGFP), featuring eGFP localized in the
membrane of glutamatergic neurons, were fixed overnight at 5 dpf in 4% PFA at 4C and subsequently
washed with and stored in PBS. A total of n=6 fish were used for experiments. To mount the fish onto 25
mm round coverslips, the coverslips were first treated with Poly-l-lysine, after which a thin layer of 1.5%
agarose (ThermoFisher Scientific, 16520050) was cured onto the coverslip. A small channel was carved
into the center of the agarose, and the fish was placed ventral side down into the channel. Finally, a
small drop of 1.5% agarose was placed on top of the fish.
Fish were imaged in milliQ water on the modified AO-LLSM described above. In this case, a
square lattice pattern (Outer NA: 0.4, Inner NA: 0.34, Cropping: 10, Envelope: 10) was used for
excitation. Image stacks (256x512 pixel FOV with 101 z steps) were acquired by scanning the sample
stages horizontally and vertically simultaneously such that the sample moved directly along the optical
axis of the detection objective with a step size of 0.2 μm and an exposure time of 100 ms. Emission light

was captured as described above. In this instance, deskewing of the data was not necessary and the final
voxel size was 108 x 108 x 200 nm.
For training data, 42 FOVs were selected near the surface (~0-20 µm) of the fish and imaged as
described above. Next, 15 FOVs deeper within the fish (~40 – 120 µm) were imaged first without AO
correction, and next with an identical AO correction procedure as described above (**Fig. 2g-m**,
Supplementary Fig. 14).
771
Live nematode adults imaged with spinning disk confocal microscopy
C. elegans strain OH15500 (*otIs669[NeuroPAL]; otIs672[panneuronal::GCaMP6s]*) were raised at
20°C and grown on NGM media plates seeded with OP50 *E. coli* . Young adult worms (with 2 or less
visible eggs in their uterus) were picked and immobilized inside a microfluidic chip as previously
described[24]. Worms were imaged by a spinning disk confocal microscope (Nikon, Ti-e) equipped with a
60×/1.2 NA water objective (Nikon, CFI Plan Apochromat VC 60XC WI), a confocal scan unit (Yokagawa,
CSU-X1) and an electron multiplying CCD (EM-CCD, Andor, iXon Ultra 897). Four excitation lasers (405
nm, 488 nm, 561 nm, and 640 nm) were used for illumination, in conjunction with emission filters
spanning 420-470 nm, 500-545 nm, 570-650 nm, and 660-800 nm bandwidths, respectively. The pixel
size was 0.27 µm in the XY dimension and each Z-stack volume comprised 21 slices for each color, with
1.5 µm step size. Each multicolor Z-stack volume was captured at a rate of just over 1 minute.
783
Fixed WGA-labeled NK-92 samples imaged with instant structured illumination microscopy
NK-92 cells (ATCC®, CRL-2407™) were rinsed with 1× PBS, and fixed with 1 ml of 4%
paraformaldehyde in 1× PBS for 30 min at room temperature, rinsed in 1 ml of 1x PBS, and
permeabilized in 0.1% Triton X-100 in 1× PBS for 15 min. Next, samples were rinsed with 1× PBS, and
blocked with buffer containing 1% BSA (Fisher, Cat# BP9700100) in 1× PBS for 1 hour. Blocking buffer
was removed, and the samples were stained with 500 µl of 1x PBS with a 1:100 dilution of Alexa Fluor
555 labelled WGA (Invitrogen, Cat# W32464), 10 U/mL phalloidin-ATTO 647N conjugate (Millipore-
Sigma, Cat #65906), and 1:1000 dilution of Hoechst solution (Tocris, Cat#5117) for 1 h. Cells were
washed in 1× PBS three times. We mounted samples using 90% Glycerol (Sigma, Cat# G5516) in 1x PBS.
In preparation for imaging, cells were cultured in collagen-I gels in the ImmunoCult-XF T Cell
Expansion Medium (STEMCELL Technologies, Cat# 10981) with the addition of Human Recombinant
Interleukin 2 (STEMCELL Technologies, Cat# 78036.3). To prepare 3 mg/ml collagen-I gel, we assembled
a gel premix on ice in a prechilled Eppendorf tube. Briefly, to 1 volume of CellAdhere™ type I bovine
(STEMCELL Technologies, Cat# 07001) we added 8/10 volume of DMEM, 1/10 volume of 10x PBS, 1/20
volume of 1M HEPES, and 1/20 volume of 1M (in DMSO) Alexa Fluor 488 ester (Molecular Probes, Cat#
A20000). A drop of premixed gel (~50 µL) was spread immediately on a glass surface of a plasma-
treated glass-bottom 35 mm Petri dish (MatTek Corp., Cat# P35G-1.5-14-C) with a pipette tip. During
polymerization (room temperature, for overnight), gels were covered with 1 mL of mineral oil (Sigma-
Aldrich, Cat# M8410) to prevent evaporation of water. Before adding NK-92 cells, polymerized gels were
rinsed with PBS to remove the unpolymerized gel components.
Instant structured illumination microscopy (iSIM) was performed using the commercial instant
structured illumination microscope system (VisiTech Intl, Sunderland, UK) equipped with an Olympus
UPlanSAapo 60×/1.3NA Sil objective, two Flash-4 scientific CMOS cameras (Hamamatsu, Corp., Tokyo,
Japan), an iSIM scan head (VisiTech Intl, Sunderland, UK), and a Nano-Drive piezo Z stage (Mad City
Laboratories, Madison, WI). The iSIM scan head included the VT-Ingwaz optical destriping unit. The

exposure time was set to 250 ms per image frame. The voxel size was 64 x 64 x 250 nm, in x, y, and z,
respectively.
811
Nerve ring calcium imaging of trapped C. elegans with instant structured illumination microscopy
Strain ABA0001 ((*lite-1(xu7); goeIs247 [ceh-24p::GCaMP6s::mKate2::unc-54 3'UTR + unc-119(+)]*) was
generated by crossing TQ1101[45] and HBR1077[46]. Adult day 1 (24 hours after late L4 stage) ABA0001
worms were raised at 20 C on standard 6 cm-diameter NGM plates seeded with *E. coli* OP50 bacteria[47].
Individual worms were picked for imaging using BIO-133 (MY Polymers) as sticky glue (in lieu of bacterial
paste) into another drop of BIO-133[48] set on a high-precision 50 x 24 mm^2 #1.5 glass coverslip (Thorlabs,
CG15KH1) between two 18 x 18 mm^2 #1 glass coverslips (Brand, 470045) used as spacers. Another high
precision 50x24 mm^2 #1.5 glass coverslip was carefully laid on top and gently pressed downwards. The
assembly was cooled on ice to ensure minimal worm movement, then flood-exposed on an aluminum
sheet to 365 nm light dispensed by a LED array for 1-2 min until BIO-133 had cured[48]. The "coverslip-
sandwiched" worms were then imaged with a qCMOS Orca Quest (Hamamatsu, C15550-22UP) through
a 40x/1.15NA water objective (Olympus, UAPON-340) on a VisiTech iSIM imaging platform driven by
Micro-Manager 2.0[49], equipped with a 300 μm-range Z-piezo (ASI, PZ-2300FT) and 405 nm, 488 nm, and
561 nm lasers.
Image volumes of *Pceh-24::GCaMP6s* expression in the worm head were then acquired using
the single-channel fast-sequence mode, with 1.2 μm axial spacing, yielding a volume acquisition rate of
~1.5 Hz (voxel dimensions: 0.115 x 0.115 x 1.2 μm^3). The exposure time was 14 ms. GCaMP6s
fluorescence was filtered through a ET525/50m emission filter (Chroma).
830
Imaging anesthetized adult C. elegans with instant structured illumination microscopy
Adult day 1 ZIM1997 (*mzmIs52; lite-1(ce314);otIs670*)[50] or ABA001 worms were raised at 20 C
on standard 6 cm-diameter NGM plates seeded with *E. coli* OP50 bacteria[47] and subsequently exposed
to unseeded NMG plates containing 0.02% levamisole prepared in M9 buffer for 10 min. Worms were
next mounted in BIO-133 as previously described, and imaged with the aforementioned VisiTech iSIM
imaging platform. 3D volumes were acquired with 300 nm Z-steps at full XY-resolution (voxel
dimensions: 0.115 x 0.115 x 0.300 μm^3) sequentially (XY-Z-C) for each channel (starting with the longest
excitation wavelength).
For ZIM1997, imaging was performed twice per worm (before and after flipping) so that both
sides of the worm were imaged with the more favorable 'near-side' configuration (**Supplementary Fig.
20**). The imaging parameters for each label were as follows: 1) mTagBFP2 with 405 nm excitation, 40 ms
exposure time, and an ET460/50m emission filter; 2) GCaMP6f with 488 nm excitation, 20 ms exposure
time, and an emission filter of ET525/50m; 3) CYO1FP with 488 nm excitation, 30 ms exposure time, and
an emission filter of ET600/50m; 4) TagRFP-T with 561 nm excitation, 40 ms exposure time, and an
emission filter of ET600/50m; 5) mNeptune2.5 with 561 nm excitation, 60 ms exposure time, and an
emission filter of ET690/50m. All emission filters were purchased from Chroma.
For ABA001, imaging parameters were: 1) GCaMP6s with 488 nm excitation, 30 ms exposure
time, and an emission filter of ET525/50m; 2) mKate2 with 561 nm excitation, 30 ms exposure time, and
an emission filter of ET600/50m. All emission filters were purchased from Chroma.
850
Two-photon microscopy on live and fixed mouse tissue

Fixed mouse liver samples and fresh ex-vivo mouse heart muscle strips were imaged with two-
photon microscopy using a Leica SP8 two photon DIVE upright microscope (Mannheim, Germany), a
pulsed dual beam Insight X3 Ti-Sapphire laser (MKS Spectra-Physics, Milpitas CA), a Leica 25x 1.0 NA (HC
PL IRAPO) water dipping lens, and emission bandwidth tunable Leica HyD detectors in the non-
descanned emission pathway. Liver samples were prepared from freshly excised liver from a 10 week-
old mouse expressing a membrane-targeted peptide fused with tdTomato[51]. After excision, the mouse
liver was washed in cold saline three times, fixed with 4% formaldehyde in PBS for 2 hours, and stored in
PBS. Tissue harvesting procedures were approved by the NCI (for mouse liver) and NHLBI (for mouse
heart) Animal Care User Committees (ACUC) respectively. Freshly excised heart muscle strips from
transgenic mice expressing mitochondrial TOMM20-mNeonGreen were prepared for imaging as
described[52]. tdTomato and mNeonGreen were excited using 1045 nm and 960 nm excitation with
emission bandwidths of 550-700 nm and 500-600 nm, respectively. Laser excitation (ramped as a
function of depth in some experiments and optimized by adjusting the objective motorized correction
collar) were in the range of 1% for tdTomato and less than 20% for mNeonGreen. HyD detector gains
were kept at 100% for tdTomato and 150% for mNeonGreen. Tiled images volumes of liver membrane
expressing tdTomato were collected with voxels sizes set to 400 nm in the XY dimension and 500 nm in
the z dimension. Z-stack volumes of mNeonGreen expressing heart strip were collected with voxels sizes
set to 120 nm in the XY dimension and 500 nm in the z dimension. All imaging was conducted at an
imaging speed of 600 Hz with a pinhole size of 1 A.U.

Cleared mouse embryos imaged with confocal microscopy
E11.5-day mouse embryos were collected in phosphate-buffered saline (PBS) and directly
immersed in 4% paraformaldehyde (PFA) in PBS (pH 7.4) at 4°C overnight. Following fixation, the
samples were washed with PBS and stored in PBS at 4°C for further analyses. Wholemount
immunofluorescence staining was performed at 4°C. The mouse embryos were permeabilized with 0.2%
Triton/PBS overnight and blocked with 10% normal goat serum and 1% BSA in 0.2% Triton/PBS
overnight. The embryos were then stained with monoclonal antibody against PECAM1 (CD31, clone
MEC 13.3, Cat# 553700, BD Pharmingen, 1:200 dilution) and monoclonal anti-β-tubulin III (TuJ1))
antibody (clone 2G10, Cat# T8578, Sigma-Aldrich, 1:500 dilution) in blocking buffer overnight. After
washing with 0.2% Triton/PBS, the embryos were stained with secondary antibodies with Alexa 488 goat
anti-rat IgG and Alexa 594 goat-anti-mouse IgG (1:250, Invitrogen, Carlsbad, CA) in blocking buffer
overnight. The embryos were cleared with iDISCO[26] and imaged using a Zeiss LSM 880 Confocal
microscope with a 10X, 0.5NA air objective. To compensate for focal shift effects due to the refractive
index difference between air and iDISCO we scaled the axial voxel size of images by 1.56 before
processing for DeAbe.
886
887 **Quantitative image quality analysis**
888
Decorrelation resolution metric
Decorrelation analysis[53] was used to estimate image resolution (**Fig. 2f, m, Supplementary Fig.
13f, Supplementary Fig. 16**). Code was downloaded from https://github.com/Ades91/ImDecorr, and the
MATLAB version of the code was used. For statistical analysis, the resolution of each image was
estimated first, then means and standard deviations were calculated from N=12 (**Fig. 2f**) or N= 15 (**Fig.
2m**) images.
895

896 *Normalized Discrete Cosine Transform Shannon Entropy.*
897   The Normalized Discrete Cosine Transform Shannon Entropy (DCTS) a helpful metric for
898 quantifying image sharpness in the frequency domain. We used it to analyze image quality degradation
899 vs. imaging depth (**Supplementary Fig. 17**). The definition of DCTS has been described in ref[54], and we
900 implemented it via customized MATLAB code.
901
902 *Image contrast metric*
903   We adopted a commonly used contrast metric – the root mean square (RMS) contrast (RMSC[55])
904 to quantify image contrast (**Supplementary Fig. 22**). The RMSC of an image is defined as:

905
$$RMSC = \sqrt{\frac{1}{N-1}\sum_{i=1}^{N}(x_i - \bar{x})^2}$$

Where x_i is the intensity of each pixel i, $\bar{x}$ is the mean intensity of the image, and N is the total pixel
number. To compare across different images, we first divide each image by its mean intensity and bin 3-
fold to reduce noise before computing the RMSC.

Image intensity correction for time-lapse images
When applying the DeAbe model to predict images, the 3D-RCAN network automatically
normalizes the input raw images to an intensity range of 0-1 by default. For time lapse images, this
normalization was performed independently at each time point, resulting in additional intensity
fluctuations. To compensate for these fluctuations, we applied corrections to the DeAbe predictions for
the GCaMP calcium signal in the live worm experiments (**Supplementary Fig. 39, Supplementary Videos**
15-17).

We first calculated the normalization ratio at each time point:

917
$$r_k = \frac{Raw_{k,pre-norm}}{Raw_{k,post-norm}}$$

918 where r_k is the normalization ratio of time point k; $Raw_{k,pre-norm}$ is the average intensity of the raw
919 image volume before normalization and $Raw_{k,post-norm}$ is the average intensity of the raw image
920 volume after normalization.
921   Next, we rescaled the image intensity of the DeAbe images based on the normalization ratios by
922 matching all time points to the first time point:

923
$$Prediction_{k,final} = \frac{r_k}{r_1}Prediction_k$$

924 where r_1 is the normalization ratio of the first time point; $Prediction_k$ is the images predicted by DeAbe
925 model; $Prediction_{k,final}$ is the final images with intensity fluctuation compensation for quantitative
926 analysis.
927
Calculation of vessel orientation and alignment
Orientations were estimated in 3D using a weighted vector summation algorithm[28], adapting it
for the volumetric images of fiber-like structures corresponding to the CD31 channel (i.e., blood vessel
images) in iDISCO-cleared mouse embryos (**Fig. 4**).

For a given voxel within the 3D image, an $n \times n \times n$ voxel window was generated surrounding the
voxel under assessment. To segment the effective voxels, six-level Otsu intensity thresholding was applied
to the image, with five thresholds dividing the intensity into six levels. The lowest level was designated as
background noise, and regions assigned to the upper five levels defined the vessel signals. The window
size n was typically set as two to three times the vessel thickness. All vectors passing through the center
voxel were defined and weighted by their length and intensity variations, and the direction of the sum of
all the weighted vectors was designated as the orientation of the center voxel[28], with associated azimuthal
angle θ (ranging from 0° to 180°) and polar angle φ (ranging from 0° to 180°). However, since the
calculation of the polar angle φ was not straightforward, we defined two additional azimuthal angles, β
and γ (**Supplementary Fig. 24a**), which were symmetrical to the azimuthal angle θ. β was defined as
the angle between the projection of the vessel in the zx plane and the x axis, and γ was the angle
between the projection in the yz plane and the $-y$ axis. These two angles were related to the polar
angle φ via:

$\tan^2 \varphi = 1/\tan^2 \beta + 1/\tan^2 \gamma$.

We also derived the 3D directional variance (DV) metric, quantifying the spread in orientations[29,56].
The value of DV ranges from 0 to 1, with 0 corresponding to perfectly parallel alignment, and 1
corresponding to complete disorder (**Supplementary Fig. 24b**). The directional variance $\bar{D}_{3D}$ was defined
as:

$\bar{D}_{3D} = 1 - (\bar{C}_{3D}{}^2 + \bar{S}_{3D}{}^2 + \bar{Z}_{3D}{}^2)^{1/2}$,

where:

$\bar{C}_{3D} = (1/k)\sum_{j=1}^{k}(f_j / \sqrt{1+f_j^2})\cos(2\theta_j)$,

$\bar{S}_{3D} = (1/k)\sum_{j=1}^{k}(f_j / \sqrt{1+f_j^2})\sin(2\theta_j)$,

$\bar{Z}_{3D} = (1/k)\sum_{j=1}^{k}(SI / \sqrt{1+f_j^2})$,

with $f_j = \sqrt{1/\tan^2(2\beta_j)+1/\tan^2(2\gamma_j)}$, and $SI = (-1)\cdot(\varphi-90)/|\varphi-90|$, where φ was acquired from the
determination of β and γ as described above, k was the number of fiber voxels in the region, and θ ,
β and γ were calculated azimuthal angles as described above.

Membrane segmentation

For the images of live worm embryos dual-labeled with nuclear and membrane markers (**Fig. 5c,**
d, Supplementary Fig. 29), raw data was restored using our multiple-step deep learning pipeline (Steps 1-
3 in **Supplementary Fig. 26a**) prior to cell membrane segmentation. We performed automatic membrane
segmentation using segmented nuclei as seeds:
First we used the Keras and Tensorflow-based implementations of Mask RCNN[57]
(https://github.com/matterport/Mask_RCNN) to perform nuclear segmentation (**Supplementary Fig.**

29d). We then manually segmented 8 volumes (3 acquired with diSPIM, 3 with iSPIM, and 2 from
multiview confocal microscopy[27] for a total of 1963 nuclei) for training. Of these 8 volumes, 6 volumes
with a total of 1688 nuclei were used for training a segmentation network and 2 volumes with a total of
275 nuclei were used for validation. We used a ResNet-50 model as the backbone for our network,
initialized the model using weights obtained from pretraining on the MS COCO dataset[58], and proceeded
to train all layers in three stages. Training took ~10 hours and applying the model took ~ 3 minutes per
volume on a Windows workstation equipped with an Intel(R) Xeon(R) W-2145 CPU operating at 3.70 GHz,
an Nvidia Quadro P6000 GPU, and 128 GB of RAM. After Mask RCNN segmentation, we applied a marker-
controlled watershed operation (https://www.mathworks.com/help/images/marker-controlled-
watershed-segmentation.html) to the nuclear segmentations to separate touching nuclei.
975   Second, we applied the vascular structure enhancement filter[59]
976   (https://github.com/timjerman/JermanEnhancementFilter) to the membrane data to enhance
977   boundaries (**Supplementary Fig. 29c**). Scales were set to [2.0, 2.25, 2.5] and all other parameters were set
978   to the default.
979   Third, the centroids of segmented nuclei were used as seeds, and we used the seeded watershed
980   algorithm (https://github.com/danielsnider/Simple-Matlab-Watershed-Cell-Segmentation) for
981   membrane segmentation (**Supplementary Fig. 29f**).
982   This workflow was applied both to the raw image data and restored images after each step in our
983   multi-step pipeline to demonstrate the benefit of segmentation enhancement from DL processing.
984   For selected volumes (**Fig. 5c, Supplementary Video 11**), we also performed manual editing on
the automatic segmentations produced by the multi-step deep learning pipeline. Manual editing was
performed within the ImageJ plugin Labkit (https://imagej.net/plugins/labkit/). After automatic
segmentations were imported to Labkit, segmentation labels were manually edited interactively in lateral
views (XY planes), and then were edited in axial views (YZ planes). Since the manual editing was conducted
in 2D views and initial editing in either view was not sufficient to ensure smoothness in 3D, we iterated
twice to further improve our results.
991

Code availability

Training and applying deep learning models were achieved using Python 3.7.0. Generation of synthetic
aberrated data and quantitative image analysis was performed in MATLAB (Mathworks, R2022b).
Customized code and software are available at https://github.com/eguomin/DeAbePlus/. RCAN and
CARE software were installed from https://github.com/AiviaCommunity/3D-RCAN and
https://github.com/CSBDeep/CSBDeep, and code for RL deconvolution and multiview fusion is available
at https://github.com/eguomin/diSPIMFusion/.

Data availability

The data that support the findings of this study are included in **Figs. 1-5, Supplementary Figs. 1–39** and
Supplementary Videos 1–17. Some representative data from the figures (**Fig. 2a, Supplementary Figs.**
16, 30) are publicly available at https://doi.org/10.5281/zenodo.8424245. Other datasets (training data
and intermediate data for deep learning) are available from the corresponding author upon reasonable
request due to their large file size.

1005

Fig. 1, Concept and simulations illustrating deep learning-based aberration compensation. **a)**
Schematic. *Left*: Fluorescence microscopy volumes are collected and near-diffraction-limited images
from the shallow side of each stack are synthetically degraded to resemble aberrated images deeper
into the stack. A neural network (e.g., a three-dimensional residual channel attention network, 3D
RCAN) is trained to reverse this degradation given the ground truth on the shallow side of the stack, and
the trained neural network (DeAbe model) subsequently applied to images throughout the stack,

improving contrast and resolution. *Right*: More detailed view of synthetic degradation process. Zernike
basis functions and associated coefficients (coeffs) are used to add random aberrations by modifying the
ideal point spread function (iPSF) to generate an aberrated PSF (aPSF). Ground truth images (GT) are
Fourier transformed (FT) and multiplied by the ratio of the Fourier transformed aberrated and ideal PSFs
(essentially a modified optical transfer function, mOTF). Inverse Fourier transforming (IFT) the result and
adding noise generates the synthetically aberrated images. See **Methods** for further detail on this
procedure. OBJ: objective lens used to collect the stack. **b)** Simulated three-dimensional phantoms
consisting of randomly oriented and positioned dots, lines, spheres, spherical shells, and circles
comparing maximum intensity projections of aberrated input image (left, random aberration with root
mean square (RMS) wavefront distortion of 2 radians and Poisson noise added for an SNR of ~16,
corresponding PSF in inset), network prediction (DeAbe) given aberrated input (middle), and ground
truth (GT, right). Higher magnification views of dashed rectangular region are shown in **c)** (maximum
intensity projection) and **d)** (single plane), additionally showing restoration given blind deconvolution
(Blind Decon), Richardson-Lucy deconvolution with diffraction-limited PSF (RL Decon 1), Richardson-Lucy
deconvolution with aberrated PSF (RL Decon 2). Yellow arrows indicate a reference structure for visual
comparison. Twenty iterations were used for RL deconvolution and ten for blind deconvolution. **e)** As in
c, d) but showing axial plane along dashed blue line in **b)**. **f)** Quantitative comparisons for the
restorations shown in **b-e**) using structural similarity index (SSIM, top) and peak signal-to-noise ratio
(PSNR, bottom) with ground truth reference. Means and standard deviations are shown for 100
simulations (10 independent phantom volumes, each aberrated with 10 randomly chosen aberrations).
Scale bars: 5 μm **b)** and 2.5 μm **c-e**). See also **Supplementary Figs. 1-5**.

1033

1034

Fig. 2, Benchmarking DeAbe against experimental ground truth and adaptive optics (AO) correction. a)
Fixed Ptk2 cells were stained for actin using Phalloidin Alexa Fluor 488 and imaged with an AO- lattice
light sheet microscope. Aberrated (i), DeAbe prediction using a model trained on random aberrations
(ii), and ground truth (GT, iii) are shown. Inset in (i) shows applied aberration; right hand insets in i)-iii)
show Fourier transforms, blue ellipse with 1/500 nm^{-1} horizontal extent and 1/400 nm^{-1} vertical extent.
Note images have been rotated so viewing is normal to the coverslip surface, which results in

anisotropic resolution in the lateral plane. **b)** Higher magnification insets of green rectangular region in
a). **c)** Higher magnification views of the yellow rectangular region in **b**). **d)** Higher magnification view of
blue rectangular region in **a**). **e)** Line profiles along red arrowheads in **d)** comparing aberrated image
(blue), DeAbe prediction (red), and ground truth (GT, black). **f)** Decorrelation resolution analysis of
images in **a)**. Means, standard deviations and individual data points from 12 images are shown. Green
arrows in **b)** and blue arrows in **d)** highlight features improved in DeAbe or GT relative to aberrated
image. XY: lateral views of sample (single planes). See also **Supplementary Figs. 10-13.** 5 dpf zebrafish
embryos expressing a GFP membrane marker labeling glutamatergic neurons were fixed and imaged in
an AO-lattice light sheet microscope. Image volumes were collected 40-140 μm from the surface of the
fish and passed through DeAbe or corrected via AO. **g)** Depth coded lateral (XY) maximum intensity
projection of volume after DeAbe compensation. Volume spans 20 μm. **h)** Single lateral plane 13 μm
into imaging volume. DeAbe prediction is shown. Note images are displayed in the native view so axial
direction is along optical axis of detection objective, resulting in isotropic resolution in the lateral plane.
i-k) Higher magnification views of green, orange, and blue rectangular regions in **h)**, comparing raw (iv),
DeAbe prediction (v), or AO correction (vi). **l)** Axial cross section along dashed white line in **g)**. Arrows in
i-l) highlight membrane regions for comparisons. **m)** Lateral resolution estimates from decorrelation
analysis. Means, standard deviations, and individual data points derived from 15 volumes are shown.
See also **Supplementary Fig. 14.** Scale bars: 10 μm and 0.4 μm^{-1} vertical/ 0.5 μm^{-1} horizontal (insets) **a**); 5
μm **b, d, g, h**); 2 μm **c, i, j, k, l**). Data shown are representative samples from N = 12 experiments for **a-d)**
and N=15 for **g-l)**.

1061

1062

Fig. 3, Computational aberration compensation on variety of fluorescence microscopy image volumes.

a) Live *C. elegans* embryos expressing a pan-nuclear GFP histone marker were imaged with light sheet

microscopy (**i**, left column) and the raw data processed with Richardson-Lucy deconvolution (**ii**, 10

iterations, middle column) or with a trained DeAbe model (**iii**, right column). First two rows show single

planes 20.0 and 27.7 μm into the sample, third row shows axial view. Comparative line profiles through

blue, yellow, and green lines are shown in insets, comparing ability to discriminate nuclei. Red arrow

highlights nuclei for visual comparison. See also **Supplementary Video 3**. **b)** NK-92 cells stained with

Alexa Fluor 555 wheat germ agglutinin and embedded in collagen matrices were fixed and imaged with

instant SIM, a super-resolution imaging technique. Left: raw data, right: after application of DeAbe and

deconvolution (DeAbe+, 20 iterations Richardson-Lucy). Lateral maximum intensity projections (MIP,

top) or single axial planes (bottom) are shown in **b)**, and **c, d** show higher magnification views

corresponding to green **c)** or blue **d)** dashed rectangular regions in **b)**. Colored arrows in **c, d** highlight

fine features obscured in the raw data and better revealed in the DeAbe+ reconstructions. See also

Supplementary Video 5, **Supplementary Fig. 19. e)** Live cardiac tissue containing cardiomyocytes

expressing Tomm20-GFP was imaged with two photon microscopy. Raw data (left) are compared with

DeAbe prediction (right) at indicated depths, with insets showing corresponding Fourier transform

magnitudes. Blue circles in Fourier insets in **e)** indicate $1/300$ nm^{-1} spatial frequency just beyond

resolution limit. See also **Supplementary Video 6**. **f)** Higher magnification views of white dashed

rectangular region in **e)**, emphasizing recovery of mitochondrial boundaries by DeAbe model. See also

Supplementary Fig. 21, Supplementary Video 7. Scale bars: 10 μm **a, e)**; 5 μm **b, f)**; 2 μm **c, d)**; **e)**

diameter of Fourier circle: 300 nm^{-1}. Data shown are representative samples from N = 3 experiments.

1084

1085

1086

Fig. 4, Computational aberration compensation on mm-scale cleared mouse embryo volumes. a) Fixed
and iDISCO-cleared E11.5-day mouse embryos were immunostained for neurons (TuJ1, cyan) and blood
vessels (CD31, magenta), imaged with confocal microscopy and processed with a trained DeAbe model.
See also **Supplementary Video 8**. **b)** Axial view corresponding to dotted rectangular region in **a)**,
comparing raw data and depth-compensated, de-aberrated, and deconvolved data (DeAbe+). See also
Supplementary Fig. 23. **c)** Higher magnification lateral view at axial depth of 1689 μm indicated by the
orange double headed arrowheads in **b)**. **d)** Higher magnification views of white dotted region in **c)**,
comparing raw (left) and DeAbe+ processing (right) for neuronal (top) and blood vessel (bottom) stains.
e) Orientation (θ, transverse angle) analysis on blood vessel channel of DeAbe+ data, here shown on
single lateral plane at indicated axial depth. See also **Supplementary Fig. 24, Supplementary Video 9**. **f)**
Higher magnification lateral view of white dotted region in **e)** (note that axial plane is different),
comparing intensity (left) and orientation (right) views between raw (top row) and DeAbe+ prediction
(middle row). Righthand insets show higher magnification views of vessel and surrounding region
highlighted by dotted lines. Bottom row indicates histogram of all orientations in the vessel highlighted

with dotted ellipse, full-width-at-half maximum (FWHM) in peak region of histogram is also shown. **g)**

Directional variance of blood vessel stain within the indicated plane, with higher magnification region of

interest (ROI) views at right. Histogram of directional variance in both regions also shown. See also

Supplementary Fig. 25. Scale bars: 500 μm **a, b, c, e)**; 100 μm **d)**, 50 μm inset; 300 μm **f)**, 50 μm inset;

300 μm **g)**, 50 μm inset. Data shown are representative samples from N = 3 experiments for **a-d)** and

N=1 for **e-g)**.

1109

1110

1111

Fig. 5, Incorporating aberration compensation into multi-step restoration dramatically improves image quality in volumetric time-lapse imaging. a) *C. elegans* embryos expressing GFP-labeled membrane marker (green) and mCherry-labeled nuclear marker (magenta) were imaged with dual-view

light-sheet microscopy (diSPIM) and the raw data (left) from single-view recordings processed through
neural networks that progressively de-aberrated, deconvolved, and isotropized spatial resolution (3-step
DL, right). Single planes from lateral (top) and axial (bottom) perspectives are shown, with arrow in
lower panel indicating direction of increasing depth. See also **Supplementary Video 10**, **Supplementary**
Figs. 27, 28. **b)** Higher magnification axial views of membranes (top) and nuclei (bottom) deep into
embryo, corresponding to dashed rectangle in **a)**. **c)** Examples of automatic segmentation on raw (left,
319 cells), 3-step DL prediction (middle, 421 cells), and manually corrected segmentation based on DL
prediction (right, 421 cells). Single planes corresponding to the upper planes in **a)** are shown. Red and
blue dashed ellipses highlight regions for visual comparison. See also **Supplementary Video 11**. **d)**
Number of cells detected by automatic segmentation of membrane marker vs. time for raw data
(purple), and after successively applying the first two steps in the multistep restoration (Steps 1-2, blue
and green curves), with means and standard deviations statistically derived from 3 different embryos.
Ground truth from manual expert (black curve) is also shown for comparison. Inset (ellipse with dotted
blue lines) highlights number count at early timepoints. See also **Supplementary Fig. 31. e)** Maximum
intensity projection (MIP) images of *C. elegans* embryos expressing membrane-localized GFP under
control of the *ttx3-3b* promoter, imaged with diSPIM, comparing raw single-view recordings (left) and
multi-step restoration that progressively de-aberrated, deconvolved, and super-resolved the data (right,
3-step DL). Boundary of the embryo has been outlined in light blue for clarity. See also **Supplementary**
Figs. 33, 34, Supplementary Video 12. Higher magnification MIP (**f**) or single lateral (**g**) or axial (**h**) plane
comparisons corresponding to dashed lines or rectangle in **e**) are also shown. **i)** Time series based on 3-
step DL MIP predictions highlight developmental progression of AIY (blue) and SMDD (magenta) neurites
as they enter the nerve ring region. Top and bottom parts of each panel at each time point show MIP
(neurites highlighted as dotted lines) vs. model of the neurites, respectively. See also **Supplementary**
Fig. 35. Scale bars: 5 μm **a, c, e, f, h**); 2 μm **b, d, g**). Data shown are representative samples from N = 3
experiments.

1140

1141

1　Ji, N. Adaptive optical fluorescence microscopy. *Nature Methods* **14**, 374 (2017).

2　Hampson, K. M. *et al.* Adaptive optics for high-resolution imaging. *Nat Rev Methods Primers* **1**,
1-26 (2021).

3　Wang, K. *et al.* Rapid adaptive optical recovery of optimal resolution over large volumes. *Nat*
Methods **11**, 625-628 (2014).

4　Zheng, W. *et al.* Adaptive optics improves multiphoton super-resolution imaging. *Nat Methods*
14, 869-872 (2017).

5　Weigert, M. *et al.* Content-aware image restoration: pushing the limits of fluorescence
microscopy. *Nat Methods* **15**, 1090-1097 (2018).

6　Qiao, C. *et al.* Rationalized deep learning super-resolution microscopy for sustained live imaging
of rapid subcellular processes. *Nat Biotechnol.* (2022). https://doi.org:doi: 10.1038/s41587-022-
01471-3

7　Li, Y. *et al.* Incorporating the image formation into deep learning improves network
performance. *Nature Methods* (2022). https://doi.org:doi: 10.1038/s41592-022-01652-7

8　Guo, M. *et al.* Rapid image deconvolution and multiview fusion for optical microscopy. *Nature*
Biotechnol. **38**, 1337-1346 (2020). https://doi.org:https://doi.org/10.1038/s41587-020-0560-x

9　Chen, J. *et al.* Three-dimensional residual channel attention networks denoise and sharpen
fluorescence microscopy image volumes. *Nature Methods* **18**, 678-687 (2021).
https://doi.org:https://doi.org/10.1101/2020.08.27.270439

10　Wang, H. *et al.* Deep learning enables cross-modality super-resolution in fluorescence
microscopy. *Nat Methods* **16**, 103-110 (2019).

11　Tian, Q. *et al.* DNN-based aberration correction in a wavefront sensorless adaptive optics
system. *Optics Express* **27**, 10765-10776 (2019).

12　Saha, D. *et al.* Practical sensorless aberration estimation for 3D microscopy with deep learning.
Opt Express **28**, 29044-29053 (2020).

13　Vinogradova, K. & Myers, E. W. Estimation of Optical Aberrations in 3D Microscopic Bioimages.
2022 7th International Conference on Frontiers of Signal Processing (ICFSP), 97-103 (2022).
https://doi.org:doi: 10.1109/ICFSP55781.2022.9924879

14　Hanser, B. M., Gustafsson, M. G. L., Agard, D. A. & Sedat, J. W. Phase retrieval for high-
numerical-aperture optical systems. *Optics Letters* **28**, 801-803 (2003).

15　Hanser, B. M., Gustafsson, M. G. L., Agard, D. A. & Sedat, J. W. Phase-retrieved pupil functions in
wide-field fluorescence microscopy. *J Microsc.* **216(Pt 1)**, 32-48 (2004).

16　Holmes, T. J. *et al.* in *Handbook of Biological Confocal Microscopy* (ed James B. Pawley) 389-
402 (Springer, 1995).

17　Chan, K. C. K., Zhou, S., Xu, X. & Loy, C. C. Basic VR++: Improving Video Super-Resolution with
Enhanced Propagation and Alignment. *2022 IEEE/CVF Conference on Computer Vision and*
Pattern Recognition (CVPR), 5962-5971 (2022).

18　Chen, B. C. *et al.* Lattice light-sheet microscopy: imaging molecules to embryos at high
spatiotemporal resolution. *Science* **346**, 1257998 (2014).

19　Liu, T.-L. *et al.* Observing the cell in its native state: Imaging subcellular dynamics in multicellular
organisms. *Science* **360**, eaaq1392 (2018).

20　Li, X. *et al.* Three-dimensional structured illumination microscopy with enhanced axial
resolution. *Nature Biotechnology* **41**, 1307-1319 (2023).

21　Shroff, H., Testa, I., Jug, F. & Manley, S. Live-cell imaging powered by computation. *Nat Rev Mol*
Cell Biol **online ahead of print** (2024).

22   Wu, Y. *et al.* Inverted selective plane illumination microscopy (iSPIM) enables coupled cell
1188   identity lineaging and neurodevelopmental imaging in Caenorhabditis elegans. *Proc. Natl. Acad.*
1189   *Sci. USA* **108**, 17708-17713 (2011).

1190   23   Wu, Y. *et al.* Reflective imaging improves spatiotemporal resolution and collection efficiency in
1191   light sheet microscopy. *Nat Commun.* **8**, 1452 (2017).

1192   24   Yemini, E. *et al.* NeuroPAL: A Multicolor Atlas for Whole-Brain Neuronal Identification in *C.*
1193   *elegans*. *Cell* **184**, 272-288 (2021).

1194   25   York, A. G. *et al.* Instant super-resolution imaging in live cells and embryos via analog image
1195   processing. *Nat Methods* **10**, 1122-1126 (2013).

1196   26   Renier, N. *et al.* iDISCO: a simple, rapid method to immunolabel large tissue samples for volume
1197   imaging. *Cell* **159**, 896-910 (2014).

1198   27   Wu, Y. *et al.* Multiview confocal super-resolution microscopy. *Nature* **600**, 279-284 (2021).
1199   https://doi.org:https://doi.org/10.1101/2021.05.21.445200

1200   28   Liu, Z. *et al.* Rapid three-dimensional quantification of voxel-wise collagen fiber orientation.
1201   *Biomedical Optics Express* **6**, 2294-2310 (2015).

1202   29   Liu, Z. *et al.* Automated quantification of three-dimensional organization of fiber-like structures
1203   in biological tissues. *Biomaterials* **116**, 34-47 (2017).

1204   30   Wu, Y. *et al.* Spatially isotropic four-dimensional imaging with dual-view plane illumination
1205   microscopy. *Nat Biotechnol.* **31**, 1032-1038 (2013).

1206   31   Kumar, A. *et al.* Dual-view plane illumination microscopy for rapid and spatially isotropic
1207   imaging. *Nature Protocols* **9**, 2555-2573 (2014).

1208   32   Cao, J. *et al.* Establishment of a morphological atlas of the *Caenorhabditis elegans* embryo using
1209   deep-learning-based 4D segmentation. *Nature Communications* **11**, 6254 (2020).

1210   33   Sulston, J. E., Schierenberg, E., White, J. G. & Thomson, J. N. The embryonic cell lineage of the
1211   nematode Caenorhabditis elegans. *Dev. Biol.* **100**, 64-119 (1983).

1212   34   Chen, F., Tillberg, P. & Boyden, E. S. Expansion microscopy. *Science* **347**, 543-548 (2015).

1213   35   Bertrand, V. & Hobert, O. Linking Asymmetric Cell Division to the Terminal Differentiation
1214   Program of Postmitotic Neurons in *C. elegans*. *Developmental Cell* **16**, 563-575 (2009).

1215   36   Rapti, G., Li, C., Shan, A., Lu, Y. & Shaham, S. Glia initiate brain assembly through noncanonical
1216   Chimaerin-Furin axon guidance in *C. elegans*. *Nature Neuroscience* **20**, 1350-1360 (2017).

1217   37   Moyle, M. W. *et al.* Structural and developmental principles of neuropil assembly in *C. elegans*.
1218   *Nature* **591**, 99-104 (2021).

1219   38   Brittin, C. A. *et al.* Multicellular rosettes organize neuropil formation. *bioRxiv* (2020).

1220   39   White, J. G., Southgate, E., Thomson, J. N. & Brenner, S. The Structure of the Nervous System of
1221   the Nematode Caenorhabditis elegans. *Phil. Trans. R. Soc. London Ser. B* **314**, 1-340 (1986).

1222   40   Mori, I. & Ohshima, Y. Neural regulation of thermotaxis in *Caenorhabditis elegans*. *Nature* **376**,
1223   344-348 (1995).

1224   41   Thibos, L. N., Applegate, R. A., Schwiegerling, J. T., Webb, R. & Members, V. S. T. Standards for
1225   Reporting the Optical Aberrations of Eyes. *OSA Technical Digest*, SuC1 (2000).

1226   42   Richardson, W. H. Bayesian-Based Iterative Method of Image Restoration. *JOSA* **62**, 55-59
1227   (1972).

1228   43   Lucy, L. B. An iterative technique for the rectification of observed distributions. *Astronomical*
1229   *Journal* **79**, 745-754 (1974).

1230   44   Bao, Z. & Murray, J. I. in *Imaging in Developmental Biology: A Laboratory Manual* (eds James
1231   Sharpe, Rachel Wong, & Raphael Yuste) (Cold Spring Harbor Laboratory Press, 2010).

1232   45   Liu, J. *et al. C. elegans* phototransduction requires a G protein-dependent cGMP pathway and a
1233   taste receptor homolog. *Nature Neuroscience* **13**, 715-722 (2010).

46   Schwarz, J. & Bringmann, H. Analysis of the NK2 homeobox gene *ceh-24* reveals sublateral
motor neuron control of left-right turning during sleep. *Elife* **28**, e24846 (2017).
47   Stiernagle, T. Maintenance of *C. elegans*. *WormBook* **11**, 1-11 (2006).
48   Han, X. *et al.* A polymer index-matched to water enables diverse applications in fluorescence
microscopy. *Lab Chip* **21**, 1549-1562 (2021).
49   Edelstein, A. D. *et al.* Advanced methods of microscope control using μManager software.
Journal of Biological Methods **1**, e11 (2014).
50   Uzel, K., Kato, S. & Zimmer, M. A set of hub neurons and non-local connectivity features support
global brain dynamics in *C. elegans*. *Curr Biol.* **32**, 3443-3459 (2022).
51   Muzumdar, M. D., Tasic, B., Miyamichi, K., Li, L. & Luo, L. A global double-fluorescent Cre
reporter mouse. *Genesis* **45**, 593-605 (2007).
52   Kosmach, A. *et al.* Monitoring mitochondrial calcium and metabolism in the beating MCU-KO
heart. *Cell Reports* **37**, 109846 (2021).
53   Descloux, A., Grußmayer, K. S. & Radenovic, A. Parameter-free image resolution estimation
based on decorrelation analysis. *Nature Methods* **16**, 918-924 (2019).
54   Royer, L. A. *et al.* Adaptive light-sheet microscopy for long-term, high-resolution imaging in live
organisms. *Nat Biotechnol.* **34**, 1267-1278 (2016).
55   Pavel, M., Sperling, G., Riedl, T. & Vanderbeek, A. Limits of visual communication: the effect of
signal-to-noise ratio on the intelligibility of American Sign Language. *Journal of the Optical
Society of America A* **4**, 2355-2365 (1987).
56   Liu, Z. *et al.* Label-free, multi-parametric assessments of cell metabolism and matrix remodeling
within human and early-stage murine osteoarthritic articular cartilage. *Communications Biology*
6, 405 (2023).
57   He, K., Gkioxari, G., Dollár, P. & Girshick, R. Mask R-CNN. *2017 IEEE conference on Computer
Vision (ICCV)*, 2980-2988 (2017).
58   Lin, T.-Y. *et al.* in *ECCV 2014* Vol. 8693 (eds D. Fleet, T. Pajdla, B. Schiele, & T. Tuytelaars)
(Springer, 2014).
59   Jerman, T., Pernuš, F., Likar, B. & Špiclin, Ž. Enhancement of Vascular Structures in 3D and 2D
Angiographic Images. *IEEE Trans Med Imaging* **35**, 2107-2118 (2016).

1263