Open-top Bessel beam two-photon light sheet microscopy for three-dimensional pathology

Running title: OT-TP-LSM for 3D pathology

Won Yeong Park[1], Jieun Yun[1], Jinho Shin[3], Byung Ho Oh[4], Gilsuk Yoon[5], Seung-Mo Hong[6], and Ki Hean Kim[1, 2, *]

[1] Department of Mechanical Engineering, Pohang University of Science and Technology, Pohang, Gyeongsangbuk, 37673, Republic of Korea.

[2] Department of I-Bio, Pohang University of Science and Technology, Pohang, Gyeongsangbuk, 37673, Republic of Korea.

[3] Department of Medicine, University of Ulsan College of Medicine, Seoul, 05505, Republic of Korea.

[4] Department of Dermatology, College of Medicine, Yonsei University, Seoul, 03722, Republic of Korea.

[5] Department of Pathology, School of Medicine, Kyungpook National University, Daegu, 41944, Republic of Korea.

[6] Department of Pathology, Asan Medical Center, University of Ulsan College of Medicine, Seoul, 05505, Republic of Korea.

*Corresponding authors

Ki Hean Kim

Department of Mechanical Engineering and

Division of Integrative Biosciences and Biotechnology

Pohang University of Science and Technology (POSTECH)

Mechanical Laboratory Building 212, 77 Cheongam-ro, Nam-gu,

Pohang, Gyeongbuk 37673, Republic of Korea

Phone No: 82-54-279-2190

Email Address: kiheankim@postech.ac.kr

Keywords: nondestructive pathology, 3D pathology, two-photon light sheet microscopy,

Bessel beam illumination, 3D imaging, unsupervised deep learning

Abstract

Nondestructive pathology based on three-dimensional (3D) optical microscopy holds promise as a complement to traditional destructive hematoxylin and eosin (H&E) stained slide-based pathology by providing cellular information in high throughput manner. However, conventional techniques provided superficial information only due to shallow imaging depth. Herein, we developed open-top two-photon light sheet microscopy (OT-TP-LSM) for intraoperative 3D pathology. An extended depth of field two-photon excitation light sheet was generated by scanning a nondiffractive Bessel beam, and selective planar imaging was conducted with cameras at 400 frames/s max during the lateral translation of tissue specimens. Intrinsic second harmonic generation was collected for additional extracellular matrix (ECM) visualization. OT-TP-LSM was tested in various human cancer specimens including skin, pancreas, and prostate. High imaging depths were achieved owing to long excitation wavelengths and long wavelength fluorophores. 3D visualization of both cells and ECM enhanced the ability of cancer detection. Furthermore, an unsupervised deep learning network was employed for the style transfer of OT-TP-LSM images to virtual H&E images. The virtual H&E images exhibited comparable histological characteristics to real ones. OT-TP-LSM may have the potential for histopathological examination in surgical and biopsy applications by rapidly providing 3D information.

Introduction

Conventional histopathological examination using thin sectioning and hematoxylin and eosin (H&E) staining has inherent limitations, including long processing time, tissue loss, sampling error, limited two-dimensional (2D) information, and artifacts. High-speed three-dimensional (3D) optical microscopy, which can visualize cellular structures without thin sectioning, holds promise for nondestructive 3D pathological examination [1-4]. Various high-speed 3D microscopy techniques have been developed [5-12], and open-top light sheet microscopy (OT-LSM) is a promising technique that enables high-throughput imaging with sheet illumination and planar imaging on uneven tissue surfaces [13-16]. OT-LSM is a specialized form of LSM for imaging large tissue specimens such as biopsy specimens by placing all the optical components below the sample holder. However, fluorescence 3D microscopy techniques, including OT-LSM, have shallow imaging depths in turbid tissue owing to high excitation light scattering and absorption, which limits precise 3D histopathology. As alternative method for high depth imaging, two-photon microscopy (TPM) has been used [17-19]. TPM is based on nonlinear excitation of fluorophores and achieves high imaging depths up to a few hundred micrometers by using long excitation wavelengths, which reduce light scattering. Furthermore, TPM provides additional intrinsic second harmonic generation (SHG) contrast for imaging collagen in the extracellular matrix (ECM) [20-22]. Two-photon LSM (TP-LSM) techniques have been developed for high-speed and high-depth imaging of live small animals and biological studies [23-27], but they were not in an open-top configuration.

In this study, we present the development of an open-top TP-LSM (OT-TP-LSM) with a Bessel beam for high-throughput and high-depth 3D pathological examination of large biopsied tissue specimens. A nondiffractive Bessel beam was used for the generation of an extended depth of field (EDOF) excitation light sheet with minimal side lobe effects [28, 29].

High imaging depths were achieved by using long excitation wavelengths and fluorescent

nuclear probes emitting relatively long wavelengths. After development and characterization,

the system was applied to high-throughput 3D imaging of various human cancer specimens,

including skin, pancreas, and prostate in comparison with conventional H&E stained images

for verification. Moreover, OT-TP-LSM images were converted to virtual H&E images using

deep learning style transfer. We demonstrate that OT-TP-LSM imaging is comparable to that

obtained from H&E stained slides for several human cancer types and has the potential for

rapid and accurate nondestructive 3D pathology.

Results

Open-top two-photon light sheet microscopy (OT-TP-LSM) for 3D pathology

Schematics and characterization results of OT-TP-LSM are presented in Figure 1. OT-TP-

LSM is an orthogonally arranged dual objective LSM with a liquid prism interface (Figure

1A). All optical components of OT-TP-LSM were positioned underneath a sample holder to

accommodate large size specimens without physical interference with the illumination and

imaging optics. Two 20x air objective lenses with a moderate numerical aperture (NA) of

0.42 were used to deliver a thin excitation light sheet and collect emission light efficiently.

The custom liquid prism was used to transmit excitation light into the specimen and to collect

emission light out with minimal aberration by refractive index (RI) matching with different

mediums where the sample was immersed (Figure 1B). The liquid prism was filled with

diluted glycerol in distilled water, whose RI was matched to that of a quartz (RI: 1.45)

window in a sample holder. A Ti-Sapphire laser beam was expanded and converted to a

Bessel beam using an axicon lens. The Bessel beam was reflected on a galvanometer scanner

and relayed to the sample through the liquid prism by a combination of a tube lens and an

objective lens. The scanner scanned the Bessel beam to form an EDOF excitation light sheet.

Emission light generated by the sample was collected by the imaging objective lens through

the liquid prism, split into two channels, one for fluorescence and the other for SHG, using a

dichroic mirror, and imaged simultaneously by two scientific complementary metal-oxide–

semiconductor (sCMOS) cameras with individual gain control.

The excitation light sheet was designed to have a DOF of 180 μm with a thickness of

0.9 μm (Figure 1 - figure supplement 1). The imaging field of view was 600 μm × 170 μm

with 2,048 × 512 pixels and was limited by the DOF of the Bessel beam. Emission light

scattering in the specimen further limited the imaging depth. Selective planar images along

the oblique excitation light sheet were acquired sequentially with a stepwise X-axis

translation of the specimen by the motorized stage (Figure 1C). A graphical user interface

based on LabVIEW was used to generate trigger signals for both the cameras and the

motorized stage. OT-TP-LSM was characterized by imaging fluorescent microspheres (0.5

μm in diameter) embedded in agarose gel, whose RI was matched to that of the solution in

the liquid prism (Figure 1D). The image throughput was 0.24 mm^2/s at an acquisition rate of

400 fps and step size of 1 μm. The image resolution was measured to be 0.9 μm in all

dimensions. The lateral resolution in the Y-Z' plane was approximately 128% of the

theoretical value due to spherical aberration induced by the RI mismatch between air and the

immersion medium, but this was sufficient for detecting cellular features.

The imaging depth of OT-TP-LSM was evaluated using freshly excised human skin

specimens. Proflavine was topically instilled onto the specimen for nuclear labeling before

imaging. A representative Y-Z' plane image is shown in Figure 1E. The imaging depth in the

skin was approximately 90 μm in the 45°-tilted image plane, and the actual depth was 64 μm

from the surface. To compare the imaging depth, the same skin specimen was imaged using

conventional one-photon excitation-based LSM (1PLSM). The imaging depth was

approximately 28 μm in the 45°-tilted image plane and 20 μm from the surface, which was

approximately three times less than that of OT-TP-LSM. While the imaging depth of OT-TP-

LSM was sufficient for visualizing 3D cell structures in the skin, it was relatively low

compared to that of conventional point-scanning based TPM due to the effect of emission

photon scattering.

3D OT-TP-LSM imaging of human skin cancers

After development and characterization, the performance of OT-TP-LSM was assessed in

human skin cancer specimens. Representative OT-TP-LSM images of two basal cell

carcinoma (BCC) specimens are presented along with H&E stained histology images in

Figure 2. The H&E images were obtained from the same specimens at slightly different

depths. The OT-TP-LSM images visualized both cell structures and ECM in the cross-

sectioned skin specimens, where proflavine-labeled cell nuclei and SHG-emitting collagen

were depicted in green and blue, respectively. A mosaic image of the first BCC specimen

visualized the epidermis as a thin layer on the top and the underlying dermis, most of which

was occupied by BCC nests (Figure 2A). The epidermis and dermis were easily identifiable

with SHG contrasts. Two regions of interest (ROI) were selected in the epidermis and BCC,

and magnified images of these ROIs are presented to show the detailed cell structures.

Magnified image in the epidermis (ROI 1) showed layered squamous epithelial cells (Figure

2B). Squamous cells in the basal layer were relatively large and individually resolved, while

those in the upper layers were unresolved with relatively small cytoplasm and appeared as a

band. Magnified images of BCC (ROI 2) showed many irregularly shaped tumor nests

composed of densely packed monomorphic tumor cells of the same size (Figure 2C).

The mosaic image of the second BCC specimen mainly showed two different cell

structures in the dermis (Figure 2D). The structures on the right side were BCCs, while those

on the left side had intact ECM composition and differed from BCCs. The corresponding

H&E stained histology image confirmed these as BCCs and normal gland structures in the

dermis, respectively. Two ROIs were selected in the nontumorous skin region, and magnified

images at two different depths are presented along with their corresponding H&E stained

images. Magnified images of ROI 3 visualized clusters of small round tubular structures

(Figure 2E). The image at depth 2 showed a duct structure composed of single-layered

cuboidal epithelial cells, which is typical of eccrine glands. Magnified images of ROI 4 at

two different depths showed pear-shaped clusters consisting of cells with clear and bulging

cytoplasm, which is typical of sebaceous glands (Figure 2F). Using 3D visualization, normal

glandular structures in the dermis were distinguished from BCC tumor nests (Video 1). Both

eccrine and sebaceous glands could appear similar to BCC nests in 2D images at certain

depths. Hence, 3D visualization would be important for distinguishing them.

3D OT-TP-LSM imaging of human pancreatic cancers

We then used OT-TP-LSM to visualize human pancreas specimens including those with

pancreatic cancer. Instead of proflavine, propidium iodide (PI) was used as a nuclear labeling

agent to increase the imaging depth (Figure 3 - figure supplement 1). The proflavine-labeled

pancreas specimens had a relatively shallow imaging depth (up to 30 µm approximately)

owing to the dense desmoplastic stromal cell composition. PI labeling allowed for deeper

imaging up to approximately 55 µm from the surface due to the longer peak emission

wavelength of 640 nm compared to proflavine. PI cytoplasmic labeling also occurred in

addition to nuclear labeling of both acinar and neoplastic cells, probably owing to PI labeling

of both DNA and RNA (Figure 3 - figure supplement 2) [30, 31]. Cell nuclei were identified

with relatively strong fluorescence in the membrane and weaker fluorescence in the core.

Representative OT-TP-LSM images and corresponding H&E stained slide images of normal

pancreas and premalignant lesion specimens, and a pancreatic cancer (also known as

pancreatic ductal adenocarcinoma, PDAC) specimen are shown in Figures 3 and 4,

respectively. The mosaic image of the normal pancreas visualized a dense distribution of

acini composed of exocrine acinar cells in good correlation with the corresponding H&E

image (Figure 3A). Two ROIs were selected, and magnified images are presented to show

detailed cytologic features of the normal pancreas. Magnified images of ROI 1 visualized the

cellular structure of acini and islets of Langerhans expressing different fluorescence

intensities (Figure 3B). Acini were composed of lobules of acinar cells, and the boundaries of

the acini were identified with relatively low PI fluorescence. The islets of Langerhans

expressed relatively low PI fluorescence, probably due to the high composition of crystallized

insulin in beta cells [32]. Magnified images of ROI 2 visualized a duct in the normal pancreas

(Figure 3C). The duct was composed of single-layered cuboidal epithelial cells with round

nuclei and was surrounded by dense fibrous tissue expressing SHG. The mosaic image of the

pancreatic premalignant lesion exhibited different and irregular structures, including dilated

ducts with intraductal proliferation of papillary structures (Figure 3D). Two different types of

precursor lesions, intraductal papillary mucinous neoplasm (IPMN) and pancreatic

intraepithelial neoplasia (PanIN), were observed [33, 34], and tumor cells were tall columnar

containing intracytoplasmic mucin [35]. A corresponding H&E histology image confirmed

the precursor lesions. Magnified images of PanIN (ROI 3) at two different depths showed

proliferation of papillary epithelial cells with intracytoplasmic mucin, and 3D morphology of

the PanIN (Figure 3E). OT-TP-LSM images were very similar to the corresponding H&E

images, indicating OT-TP-LSM provided comparable histopathological information to that of

H&E stained slides without thin sectioning.

A mosaic image of the PDAC arising from the IPMN specimen showed various

abnormal ductal structures including cancer glands, IPMN, and surrounding pancreatitis,

which were confirmed by the corresponding H&E stained slide image (Figure 4A). More

desmoplastic fibrous stroma and a few abnormal cancer glands were observed compared to

normal pancreas tissue [36, 37]. Two ROIs were selected in the PDAC region and the

boundary region with nonneoplastic pancreatitis, and magnified images are presented to show

detailed cellular structures. Magnified images of ROI 1 (PDAC) at two different depths

showed irregularly shaped glands with sharp angles (Figure 4B). The glands exhibited

disorganized cancer cell arrangement with nuclear membrane distortion. Magnified images of

ROI 2 showed both nonneoplastic ducts and cancer glands in different cell arrangements

(Figure 4C). The nonneoplastic ducts had small, evenly distributed cells with relatively high

nuclear fluorescence. Cancer glands, on the other hand, had disorganized and multilayered

nuclei.

3D OT-TP-LSM imaging of human prostatic cancers

OT-TP-LSM was used to visualize human prostatectomy specimens, and two representative

OT-TP-LSM images of benign and adenocarcinoma specimens are presented in Figure 5. The

mosaic image of the histologically confirmed benign prostate showed benign glands

dispersed throughout the specimen with relatively weak SHG composition between the

glands (Figure 5A). An ROI was selected on benign glands, and magnified images are

presented to show cytologic features alongside the corresponding H&E stained slide image

(Figure 5B). Multiple cell layers, consisting of secretory cells and basal cells, lined the

central lumen in the ducts of benign glands. Although images at two different depths are

presented, the cell distribution was similar because of the relatively small depth difference

compared to the gland size. The mosaic image of the prostate adenocarcinoma in Figure 5C

showed both the benign and cancer regions, whose boundary was marked with a white solid

line, and they were confirmed by the corresponding H&E stained image. The cancer region

was full of crowded small gland structures, whereas the benign region consisted of relatively

large glands surrounded by fibrous stroma. Magnified images of ROI 2 in the cancer region

showed small glands composed of monolayered cell walls (Figure 5D). The cancer glands

had single-layered malignant secretory cells only due to the loss of basal cell layers. The

corresponding H&E histology was diagnosed as Gleason score 3+3 prostatic adenocarcinoma.

OT-TP-LSM provided histological information equivalent to that of the H&E stained image

without the need for sectioning.

Deep learning-based style transfer for virtual H&E

OT-TP-LSM images visualized histological features with high similarity to those in H&E

images, but they may be unfamiliar to pathologists and clinicians accustomed to traditional

H&E stained slide histology. PI-labeled OT-TP-LSM images were converted to virtual H&E

images using Cycle-consistent generative adversarial network (CycleGAN) [38].

Representative virtual H&E images of pancreatic premalignant (PanIN) and PDAC are

presented along with the corresponding real H&E stained images for comparison in Figure 6.

Mosaic virtual H&E images of both PanIN and PDAC showed similar patterns to the

corresponding real H&E stained histology images (Figure 6A, C). Although the normal

region in the lower left corner of the virtual PanIN image was not transformed properly

possibly due to insufficient image data of normal acini for training (Figure 6A), the

magnified virtual H&E images clearly displayed distinct cell distributions of PanIN and

PDAC by labeling cell cytoplasm and nuclei in pink and purple, respectively (Figure 6B, D).

PanINs were visualized as slightly dilated, irregularly clustered ducts with preserved

structural integrity (Figure 6B), whereas PDAC had disorganized cell arrangement with

haphazard patterns, infiltrating into the surrounding tissue with desmoplastic stroma (Figure

6D). Histological features found in the virtual H&E images were diagnostic factors for the

discrimination of PDAC and PanIN. Virtual H&E staining of OT-TP-LSM via CycleGAN

could provide comparable cellular information to conventional H&E stained histology.

Discussion

OT-TP-LSM was developed to enable rapid and precise 3D pathological examination of

excised tissue specimens during both biopsy and surgery, allowing visualization of 3D cell

structures. OT-TP-LSM generated a two-photon excitation light sheet using one-dimensional

Bessel beam scanning, and planar images were acquired by collecting emission light from the

excitation light sheet with sCMOS cameras at 400 fps maximum. Proflavine and PI were

used for nuclear labeling and cell structure visualization, and intrinsic SHG contrast was used

for ECM visualization. The performance of OT-TP-LSM was characterized in comparison

with histopathologic images of H&E stained slides in various human cancer specimens

including skin, pancreas, and prostate. 3D OT-TP-LSM images were well matched with the

corresponding 2D H&E histological images with all the human tissue specimens used. The

relatively high imaging depths of OT-TP-LSM enable the visualization of detailed 3D cell

structures with high contrast and without distortion. The high-quality images of OT-TP-LSM

may be partially attributed to the use of the nondiffractive Bessel beam for the generation of

an EDOF excitation light sheet [39].

OT-TP-LSM could visualize 3D cell structures in the skin up to a depth of 60 μm by

using long wavelength excitation light and proflavine labeling. Although the imaging depth

was not high compared with conventional TPM, it was sufficient to visualize 3D cell

structures in the skin without sectioning artifacts. The imaging depth varied depending on the

tissue type, and the imaging depths of both pancreas and prostate were relatively shallow in

comparison to that of the skin. This low imaging depth could be attributed to the dense cell

composition and higher RI than the current immersion medium. To increase the imaging

depth, PI with a relatively long emission wavelength was used. Additionally, an immersion

medium with 1.5 RI could improve the imaging depth [40], and a quick optical clearing

method could also be used to enhance the imaging depth [41].

We applied the unsupervised deep learning network of CycleGAN to PI-labeled OT-

TP-LSM images of pancreas specimens to generate virtual H&E images, and the cytoplasm

and cell nuclei, which exhibited intensity differences in greyscale, were transformed to H&E-

like pseudo colors. Virtual H&E images in pancreatic cancer provided histological features

equivalent to the conventional H&E. Although the CycleGAN network showed the potential

for virtual H&E staining, further development is necessary to ensure robust performance and

universal application. Therefore, CycleGAN will be trained using more OT-TP-LSM image

data from different tissue types such as skin and prostate.

The current OT-TP-LSM has several limitations that require improvement. One issue

is spherical aberration induced by the RI mismatch between the liquid prism and air when

using air objective lenses in both illumination and imaging arms. Spherical aberration

degraded the image resolution, particularly in the imaging arm. An immersion imaging

objective lens with a high NA would be desirable to improve both image resolution and

emission light collection efficiency. The excitation light sheet was generated through the

rapid scanning of a single Bessel beam, resulting in a short pixel dwell time compared to that

of a conventional single photon 2D light sheet. The generation of excitation light sheet with

the scanning of multiple Bessel beams would increase the pixel dwell time and image

contrast [42, 43]. The relatively short excitation wavelength of 770 nm was used for all the

imaging to efficiently collect additional SHG light [20], but the imaging depth was limited.

Longer excitation wavelengths with long emission wavelength probes would increase the

imaging depth beyond those presented here.

In conclusion, OT-TP-LSM was developed for high-throughput and high-depth tissue

examination. OT-TP-LSM was tested with various human cancers including skin, pancreas,

and prostate by visualizing 3D cell structures. Intrinsic SHG contrast was used for visualizing

ECM and detecting fibrosis. Normal glands and BCC nests were clearly distinguished, and

both pancreatic and prostate cancers were clearly identified using 3D visualization of cell

structures with OT-TP-LSM. Furthermore, virtual H&E staining of OT-TP-LSM using

CycleGAN provided histopathologic information comparable to that from H&E stained slides.

Therefore, OT-TP-LSM could be useful for rapid and precise nondestructive 3D pathology.

Materials and methods

OT-TP-LSM setup

The schematics of OT-TP-LSM are shown in Figure 1A. As the light source, OT-TP-LSM

used a titanium–sapphire pulse laser (Chameleon II, Coherent) with 140-fs pulse width and

80-MHz pulse repetition rate. The excitation wavelength was set at 770 nm for imaging

human tissue specimens. In the illumination arm, the laser beam was expanded from 1.4 mm

to 1.86 mm and then converted to a Bessel beam by using an axicon lens (131-1270 +

ARB825, Eksma). The Bessel beam was relayed onto the specimen through two lens pairs of

L1 (AC254-045-B, Thorlabs) and L2 (AC254-080-B, Thorlabs), and L3 (AC254-150-B,

Thorlabs) and a 20× air illumination objective lens with 0.42 NA (MY20X-804, Mitutoyo).

The Bessel beam diameter at the back aperture of the objective lens was approximately 6.8

mm to prevent beam clipping, and the effective illumination NA was 0.34. The excitation

light sheet was generated by Y-axis scanning of the Bessel beam with a galvanometric

scanner (GVS102, Thorlabs), which was positioned between L1 and L2. The liquid prism

transmitted the excitation light normally through a coverslip (HSU-0101122, Marienfeld

Superior), RI-matching solution, and a 0.25 mm thick round quartz window (26043, Ted

Pella) at the bottom of the sample holder, and then to the sample (Fig. 1B). Diluted 85%

glycerol (Samchun) in distilled water was used for RI-matching with the quartz window (RI:

1.45). The sample holder was connected to an x-y motorized translation stage (XY MS-2000,

Applied Scientific Instrumentation) for sample translation during imaging. To remove air

gaps between the quartz window of the sample holder and irregular-shaped tissue specimen,

the specimen was partially immersed in the RI-matching solution and light pressure was

applied on the top of the specimen. In the imaging arm, emission light was collected by the

imaging objective lens (MY20X-804, Mitutoyo) through the liquid prism, spectrally

separated into two channels by a dichroic mirror (T425lpxr, Chroma), and collected at two

sCMOS cameras (pco-edge 4.2, PCO) after passing through a tube lens (AC254-200-A,

Thorlabs), respectively. SHG light was collected in the short pass channel with an emission

filter (ET376/30x, Chroma), and fluorescence was collected in the long pass channel with an

emission filter (ET680SP, Chroma). Both fluorescence and SHG images were simultaneously

acquired with individual gain control.

Human skin cancer specimen collection and imaging protocol

Fresh skin cancer specimens were collected under the Institutional Review Board (IRB) of

Yonsei University, Severance Hospital, approved protocol (approval # 4-2021-0128). The

study was conducted according to the Declaration of Helsinki Principles. OT-TP-LSM

imaging was conducted within 12 h post-excision. Patient information was deidentified.

Before imaging, fresh skin specimens were cut cross-sectionally to approximately 4 mm

thickness and topically instilled with 400 μg/mL proflavine solution (Sigma) for 5 min, and

then rinsed in phosphate-buffered saline. OT-TP-LSM imaging was conducted in 3D on the

sectioned surface. After imaging, the skin specimens were fixed and processed for

histopathological evaluation with H&E stained slides. Microscopic images were obtained by

using a wide-field microscope (Leica Z16 APO, Leica Microsystems).

Human pancreatic and prostate cancer specimen collection and imaging protocol

Formalin-fixed, paraffin-embedded (FFPE) tissue blocks of pancreatic and prostate cancers

were collected under approval (approval #, 2021-0393) from the IRB of Asan Medical Center,

University of Ulsan College of Medicine. 4 μm thickness sections were made for H&E

stained slides, and the remaining blocks were used for OT-TP-LSM imaging. The FFPE

tissue blocks were deparaffinized with xylene (99%, Daejung) solution overnight, 1:1 xylene-

100% methanol (Optima grade, Fisher scientific) mixture solution for 1 h with shaking,

followed by sequentially decreasing the methanol concentration in the xylene and methanol

mixture to 90%, 80%, and 50% for 1 h each. After deparaffinization, specimens were stained

with 80 µg/mL PI solution (Sigma) for 5 min, and then OT-TP-LSM imaging was conducted

on the sectioned surface.

Virtual H&E staining of OT-TP-LSM via deep learning network

CycleGAN is an unsupervised deep learning network without the need for well-aligned image

pairs [38] and was used for style transfer of OT-TP-LSM images to virtual H&E images [44-

46]. CycleGAN model consisted of two generators (G: real OT-TP-LSM $\rightarrow$ fake H&E and

F: real H&E $\rightarrow$ fake OT-TP-LSM), and two corresponding discriminators ($D_{OT-TP-LSM}$ and

$D_{H\&E}$) to identify each fake image (Figure 6 - figure supplement 1). Cycle-consistent loss was

combined with adversarial loss to prevent the well-known problem of mode collapse, where

all input images mapped to the same output images [38]. For CycleGAN training, PI-labeled

OT-TP-LSM and real H&E images were collected from pancreas specimens, including each

of normal, premalignant lesions, and cancer. Collected images were cropped into 512×512

pixels patch images, and augmented by applying rotation (90°, 180°, 270°) and flip (vertical,

horizontal). The total training dataset was 14,322 and 22,644 patches for OT-TP-LSM and

H&E, respectively. Adam solver was used as an optimizer [47]. Hyper-parameters were set

with a batch size of 1 and an initial learning rate of 0.0002 and 100 epochs. Once the training

was completed, generator G was used to transform OT-TP-LSM patches into virtual H&E

ones.

Acknowledgments

This work was supported in part by the National Research Foundation (NRF) grant funded by

the Korea government (NRF-2020R1A2C3009309); and the Korea Medical Device

Development Fund (KMDF) grant funded by the Korean government (Project Number:

9991006749, KMDF_PR_20200901_0076).

Conflict of interest statements

No conflicts of interest were declared.

References

1.   Liu JTC, Glaser AK, Bera K, *et al*. Harnessing non-destructive 3D pathology. *Nat*

Biomed Eng 2021; **5:** 203-218.

2.   Almagro J, Messal HA, Thin MZ, *et al*. Tissue clearing to examine tumour

complexity in three dimensions. *Nat Rev Cancer* 2021; **21:** 718-730.

3.   Bishop KW, Maitland KC, Rajadhyaksha M, *et al*. In vivo microscopy as an

adjunctive tool to guide detection, diagnosis, and treatment. *J Biomed Opt* 2022; **27:** 040601.

4.   Serafin R, Koyuncu C, Xie W, *et al*. Nondestructive 3D pathology with analysis of

nuclear features for prostate cancer risk assessment. *J Pathol* 2023; **260:** 390-401.

5.   Filler TJ, Peuker ET. Reflection contrast microscopy (RCM): a forgotten technique?

J Pathol 2000; **190:** 635-638.

6.   van Royen ME, Verhoef EI, Kweldam CF, *et al*. Three-dimensional microscopic

analysis of clinical prostate specimens. *Histopathology* 2016; **69:** 985-992.

397 7.  Verhoef EI, van Cappellen WA, Slotman JA, *et al*. Three-dimensional analysis

398 reveals two major architectural subgroups of prostate cancer growth patterns. *Modern Pathol*

399 2019; **32:** 1032-1041.

400 8.  Yoshitake T, Giacomelli MG, Cahill LC, *et al*. Direct comparison between confocal

401 and multiphoton microscopy for rapid histopathological evaluation of unfixed human breast

402 tissue. *J Biomed Opt* 2016; **21:** 126021.

403 9.  Tao YK, Shen DJ, Sheikine Y, *et al*. Assessment of breast pathologies using

404 nonlinear microscopy. *P Natl Acad Sci USA* 2014; **111:** 15304-15309.

405 10.  Baugh LM, Liu ZY, Quinn KP, *et al*. Non-destructive two-photon excited

406 fluorescence imaging identifies early nodules in calcific aortic-valve disease. *Nature*

407 *Biomedical Engineering* 2017; **1:** 914-924.

408 11.  Patel KB, Liang WX, Casper MJ, *et al*. High-speed light-sheet microscopy for the in-

409 situ acquisition of volumetric histological images of living tissue. *Nature Biomedical*

410 *Engineering* 2022; **6:** 569-583.

411 12.  Voleti V, Patel KB, Li WZ, *et al*. Real-time volumetric microscopy of in vivo

412 dynamics and large-scale samples with SCAPE 2.0. *Nat Methods* 2019; **16:** 1054-1062.

413 13.  McGorty R, Liu H, Kamiyama D, *et al*. Open-top selective plane illumination

414 microscope for conventionally mounted specimens. *Opt Express* 2015; **23:** 16142-16153.

415 14.  Glaser AK, Reder NP, Chen Y, *et al*. Light-sheet microscopy for slide-free non-

416 destructive pathology of large clinical specimens. *Nature Biomedical Engineering* 2017; **1:**

417 0084.

418 15.  Glaser AK, Bishop KW, Barner LA, *et al*. A hybrid open-top light-sheet microscope

419 for versatile multi-scale imaging of cleared tissues. *Nat Methods* 2022; **19:** 613-619.

16.   Gao G, Miyasato D, Barner LA, *et al*. Comprehensive Surface Histology of Fresh

Resection Margins With Rapid Open-Top Light-Sheet (OTLS) Microscopy. *Ieee T Bio-Med*

Eng 2023; **70:** 2160-2171.

17.   Denk W, Strickler JH, Webb WW. Two-photon laser scanning fluorescence

microscopy. *Science* 1990; **248:** 73-76.

18.   So PTC, Dong CY, Masters BR, *et al*. Two-photon excitation fluorescence

microscopy. *Annu Rev Biomed Eng* 2000; **2:** 399-429.

19.   Helmchen F, Denk W. Deep tissue two-photon microscopy. *Nat Methods* 2005; **2:**

932-940.

20.   Zipfel WR, Williams RM, Christie R, *et al*. Live tissue intrinsic emission microscopy

using multiphoton-excited native fluorescence and second harmonic generation. *Proc Natl*

Acad Sci U S A 2003; **100:** 7075-7080.

21.   Fast A, Lal A, Durkin AF, *et al*. Fast, large area multiphoton exoscope (FLAME) for

macroscopic imaging with microscopic resolution of human skin. *Sci Rep* 2020; **10:** 18093.

22.   Park WY, Kim B, Chun JH, *et al*. High-contrast visualization of human skin cancers

with combined reflectance confocal and moxifloxacin-based two-photon microscopy: An ex

vivo study. *Lasers Surg Med* 2022; **54:** 1226-1237.

23.   Fahrbach FO, Gurchenkov V, Alessandri K, *et al*. Light-sheet microscopy in thick

media using scanned Bessel beams and two-photon fluorescence excitation. *Opt Express*

2013; **21:** 13824-13839.

24.   Mahou P, Vermot J, Beaurepaire E, *et al*. Multicolor two-photon light-sheet

microscopy. *Nat Methods* 2014; **11:** 600-601.

25.   Wolf S, Supatto W, Debregeas G, *et al*. Whole-brain functional imaging with two-

photon light-sheet microscopy. *Nat Methods* 2015; **12:** 379-380.

26.   Maioli V, Boniface A, Mahou P, *et al*. Fast in vivo multiphoton light-sheet

microscopy with optimal pulse frequency. *Biomed Opt Express* 2020; **11:** 6012-6026.

27.   de Vito G, Turrini L, Mullenbroich C, *et al*. Fast whole-brain imaging of seizures in

zebrafish larvae by two-photon light-sheet microscopy. *Biomed Opt Express* 2022; **13:** 1516-

1536.

28.   Planchon TA, Gao L, Milkie DE, *et al*. Rapid three-dimensional isotropic imaging of

living cells using Bessel beam plane illumination. *Nat Methods* 2011; **8:** 417-423.

29.   Fan JL, Rivera JA, Sun W, *et al*. High-speed volumetric two-photon fluorescence

imaging of neurovascular dynamics. *Nat Commun* 2020; **11:** 6020.

30.   Rieger AM, Hall BE, Le TL, *et al*. Conventional apoptosis assays using propidium

iodide generate a significant number of false positives that prevent accurate assessment of

cell death. *J Immunol Methods* 2010; **358:** 81-92.

31.   Niu JF, Li CM, Wu HH, *et al*. Propidium iodide (PI) stains Nissl bodies and may

serve as a quick marker for total neuronal cell count. *Acta Histochem* 2015; **117:** 182-187.

32.   Lemaire K, Chimienti F, Schuit F. Zinc transporters and their role in the pancreatic

β - cell. *Journal of Diabetes Investigation* 2012; **3:** 202-211.

33.   Hruban RH, Adsay NV, Albores-Saavedra J, *et al*. Pancreatic intraepithelial

neoplasia: a new nomenclature and classification system for pancreatic duct lesions. *Am J*

Surg Pathol 2001; **25:** 579-586.

34.   McGinnis T, Bantis LE, Madan R, *et al*. Survival Outcomes of Pancreatic

Intraepithelial Neoplasm (PanIN) versus Intraductal Papillary Mucinous Neoplasm (IPMN)

Associated Pancreatic Adenocarcinoma. *J Clin Med* 2020; **9:** 3102.

35.   Adsay NV, Bandyopadhyay S, Basturk O, *et al*. Chronic pancreatitis or pancreatic

ductal adenocarcinoma? *Semin Diagn Pathol* 2004; **21:** 268-276.

36.   Drifka CR, Tod J, Loeffler AG, *et al*. Periductal stromal collagen topology of

pancreatic ductal adenocarcinoma differs from that of normal and chronic pancreatitis.

Modern Pathol 2015; **28:** 1470-1480.

37.   Ray A, Callaway MK, Rodriguez-Merced NJ, *et al*. Stromal architecture directs early

dissemination in pancreatic ductal adenocarcinoma. *JCI Insight* 2022; **7:** e150330.

38.   Zhu JY, Park T, Isola P, *et al*. Unpaired Image-to-Image Translation using Cycle-

Consistent Adversarial Networks. *Ieee I Conf Comp Vis* 2017; 2242-2251.

39.   Olarte OE, Licea-Rodriguez J, Palero JA, *et al*. Image formation by linear and

nonlinear digital scanned light-sheet fluorescence microscopy with Gaussian and Bessel

beam profiles. *Biomed Opt Express* 2012; **3:** 1492-1505.

40.   Young PA, Clendenon SG, Byars JM, *et al*. The effects of refractive index

heterogeneity within kidney tissue on multiphoton fluorescence excitation microscopy. *J*

Microsc-Oxford 2011; **242:** 148-156.

41.   Aoyagi Y, Kawakami R, Osanai H, *et al*. A Rapid Optical Clearing Protocol Using

2,2'-Thiodiethanol for Microscopic Observation of Fixed Mouse Brain. *Plos One* 2015; **10:**

e0116280.

42.   Minamikawa T, Hashimoto M, Fujita K, *et al*. Multi-focus excitation coherent anti-

Stokes Raman scattering (CARS) microscopy and its applications for real-time imaging. *Opt*

Express 2009; **17:** 9526-9536.

43.   Chen X, Zong W, Li R, *et al*. Two-photon light-sheet nanoscopy by fluorescence

fluctuation correlation analysis. *Nanoscale* 2016; **8:** 9982-9987.

44.   Chen Z, Yu W, Wong IHM, *et al*. Deep-learning-assisted microscopy with ultraviolet

surface excitation for rapid slide-free histological imaging. *Biomed Opt Express* 2021; **12:**

5920-5938.

45.   Cao R, Nelson SD, Davis S, *et al*. Label-free intraoperative histology of bone tissue

via deep-learning-assisted ultraviolet photoacoustic microscopy. *Nat Biomed Eng* 2023; **7:**

124-134.

46.   Abraham T, Shaw A, O'Connor D, *et al*. Slide-free MUSE microscopy to H&E

histology modality conversion via unpaired image-to-image translation GAN Models. v1.

arXiv.org 2020. Available from: https://arxiv.org/abs/2008.08579. Accessed: [14 August

2023]. Not peer-reviewed.

47.   Kingma DP, Ba J. Adam: A method for stochastic optimization. v9. arXiv.org 2014.

Available from: https://arxiv.org/abs/1412.6980. Accessed: [14 August 2023]. Not peer-

reviewed.

Figure legends

Figure 1. Schematics and characteristics of open-top two-photon light sheet microscopy

(OT-TP-LSM). (A) Overall design of OP-TP-LSM. L: lens; M: mirror; DM: dichroic mirror;

SPF: short pass filter; BPF: band pass filter; OBJ: objective lens. (B) Detailed schematic of

the sample interface, including the liquid prism filled with refractive index (RI)-matching

solution. (C) Illustration of sequential light sheet imaging with stepwise lateral translation.

(D) Image resolution characterized by three-dimensional (3D) imaging of 0.5 μm fluorescent

microspheres. (E) Imaging depth characterized by imaging a proflavine-labeled fresh human

skin specimen in comparison with that of one-photon light sheet microscopy (1PLSM) using

Gaussian excitation sheet.

Figure 2. 3D OT-TP-LSM images of fresh human skin cancer (basal cell carcinoma,

BCC) specimens. (A) Large-sectional OT-TP-LSM and hematoxylin and eosin (H&E)

stained slide images of a BCC case 1. (B and C) Two magnified images of regions of interest

(ROIs) with corresponding H&E images. ROI 1 and 2 were in the epidermis and superficial

dermis and the BCC region below, respectively. (D) Large-sectional OT-TP-LSM and H&E

images of a BCC case 2. (E and F) Two magnified images of ROIs with corresponding H&E

images. ROI 3 and 4 were in eccrine and sebaceous glands, respectively. Superficial cell

layers of both glands are marked with a white arrow. In the OT-TP-LSM images, proflavine

fluorescence and second harmonic generation (SHG) are displayed in green and blue,

respectively.

Figure 3. 3D OT-TP-LSM images of two human pancreas specimens: normal and

premalignant lesion. (A) Large-area OT-TP-LSM and H&E stained images of normal

human pancreas. (B and C) Magnified two-depth OT-TP-LSM images of ROI 1 and 2 in the

normal pancreas and the corresponding H&E stained slide images. ROI 1 and 2 were on the

acini and islets of Langerhans and the pancreatic duct, respectively. 3D morphological

change of the duct was marked with a white arrow. (D) Large-area OT-TP-LSM and H&E

stained images of a pancreatic premalignant lesion. Red box indicates the intraductal

papillary mucinous neoplasm (IPMN). (E) Magnified two-depth OT-TP-LSM images of ROI

3 and corresponding H&E image of pancreatic intraepithelial neoplasia (PanIN). 3D

morphological change of papillary in the PanIN was marked with a cyan arrow. In OT-TP-

LSM images, propidium iodide (PI)-labeled cells and SHG-emitting collagen are displayed in

green and blue, respectively.

Figure 4. 3D OT-TP-LSM images of human pancreatic ductal adenocarcinoma (PDAC).

(A) Large-area OT-TP-LSM and H&E stained slide images of PDAC arising from IPMN

specimen. Pink solid line indicates the IPMN area. (B and C) Magnified two-depth OT-TP-

LSM images of ROI 1 and 2 in the PDAC region and the corresponding H&E stained images.

ROI 1 was on the PDAC, and ROI 2 was on the boundary region between nonneoplastic

pancreatitis and PDAC, respectively. A nonneoplastic duct and a PDAC were marked with

green and red arrows, respectively. A white dashed line indicated the boundary between

nonneoplastic pancreatitis and PDAC. In the OT-TP-LSM images, PI-labeled cells and SHG-

emitting collagen are displayed in green and blue, respectively.

Figure 5. 3D OT-TP-LSM images of human prostate specimens: benign and

adenocarcinoma. (A) Large-area OT-TP-LSM and H&E stained slide images of a benign

prostate specimen. (B) Magnified two-depth OT-TP-LSM images of ROI 1 in the benign

gland and a corresponding H&E image. (C) Large-area OT-TP-LSM and H&E images of a

prostatic cancer specimen. A white solid line indicated the boundary between

adenocarcinoma and benign regions. (D) Magnified two-depth OT-TP-LSM images of ROI 2

in adenocarcinoma glands and a corresponding H&E image. In OT-TP-LSM images, PI-

labeled cells and SHG-emitting collagen were visualized in green and blue, respectively.

Figure 6. Style transfer of OT-TP-LSM images for virtual H&E by cycle-consistent generational adversarial networks (CycleGAN). (A and C) Large-area virtual H&E images and corresponding H&E stained slide images of PanIN and PDAC specimens, respectively. (B and D) Magnified virtual H&E images and corresponding H&E stained images of PanIN and PDAC, respectively.

Supplementary figure legends

Figure 1 - figure supplement 1. Detail optical configuration of illumination arm in open-top two-photon light sheet microscopy (OT-TP-LSM). (A) Illustration of beam path at Y-Z' plan. L: lens, BF: barrier filter, AL: axicon lens, SM: scanning mirror, OBJ: objective lens. Two-photon light sheet with extended depth of field (DOF) was generated by scanning a Bessel beam in the Y-axis direction rapidly. (B) Illustration of beam path at Z-X plan. (C) Magnified image and designed dimension of the Bessel beam based light sheet. The excitation light sheet was designed to have a DOF of 180 μm with a thickness of 0.9 μm.

Figure 3 - figure supplement 1. Comparison of imaging depth between propidium iodide (PI) and proflavine in human pancreas. (A) 45° oblique OT-TP-LSM image of pancreas specimens after PI cell labeling. (B) 45° oblique OT-TP-LSM image of pancreas specimens after proflavine cell labeling. PI labeling enabled the deeper imaging up to approximately 55 μm from the surface owing to its relatively long peak emission wavelength compared to proflavine. Both pancreas specimens were incubated with PI and proflavine for 5 mins, respectively.

Figure 3 - figure supplement 2. PI cell staining characteristics in human pancreas. (A) OT-TP-LSM image of acinar cell structures in normal pancreas specimens after PI cell labeling and corresponding H&E image. (B) OT-TP-LSM image of pancreatic intraepithelial neoplasia (PanIN) structures in premalignant pancreas after PI cell labeling and

corresponding H&E image. In OT-TP-LSM images, PI fluorescence was displayed in green

color. PI showed nuclear and cytoplasmic labeling properties of cells in both acini and

neoplastic structures. Cell nuclei were identified with relatively strong and weak fluorescence

in the membrane and core, respectively.

Figure 6 - figure supplement 1. Cycle-consistent generational adversarial networks

(CycleGAN) workflow for style transfer of OT-TP-LSM image to virtual hematoxylin

and eosin (H&E) image. Real OT-TP-LSM image and real H&E image were assigned as x

and y, and the corresponding fake image is additionally marked with a prime. The generator

G transformed the real OT-TP-LSM image into a fake virtual H&E-stained image. Generator

F transformed the real H&E image into a fake OT-TP-LSM image. The discriminator Dx

classifies real OT-TP-LSM image x and fake OT-TP-LSM image x'. The discriminator Dy

classifies real H&E image y and fake virtual H&E image y'. Both Cycle-consistent loss and

adversarial loss were used as the loss function, which was to guarantee the performance of

inverse transform from G and F by preserving the morphological features of real images.

Video legend

Video 1. Three-dimensional (3D) visualization of basal cell carcinoma (BCC) structures and

normal cell structures.

A

BF   AL   L1   SM(Y)   L2   L3
      α: 5°   F:45   F:80   F:150

Beam width: 0.9mm   Ring D: 3.6mm   Beam width: 1.7mm   Ring D: 6.8mm
β: 2.3 deg   β: 1.3 deg   Effective NA 0.34

B

C

OBJ lens 1 Ill.

Sheet area
0.6 mm x 0.18 mm

Sheet thickness	Depth of field
0.90 μm	180 μm

A

B

A
B