Widefield *in vivo* imaging system with two fluorescence and two reflectance channels, a single sCMOS detector, and shielded illumination

Abbreviated title: Four-color mesoscope

Patrick R. Doran[1], Natalie Fomin-Thunemann[2], Rockwell P. Tang[1], Bernhard Zimmerman[2],

Kıvılcım Kılıç[2], Emily A. Martin[2], Sreekanth Kura[2], Harrison P. Fisher[1], Grace Chabbott[3], Joel

Herbert[4], John X. Jiang[2], Sava Sakadzic[5], David A. Boas[2], Anna Devor[2,5], Ichun Anderson Chen[2],

Martin Thunemann[2]

[1]Graduate Program in Biomedical Engineering, Boston University, Boston, MA 02215, USA

[2]Department of Biomedical Engineering, Boston University, Boston, MA 02215, USA

[3]Undergraduate Program in Biomedical Engineering, Boston University, Boston, MA 02215, USA

[4]Undergraduate Program in Neuroscience, Boston University, Boston, MA 02215, USA

[5]Athinoula A. Martinos Center for Biomedical Imaging, Department of Radiology, Harvard Medical

School, Massachusetts General Hospital, Charlestown, MA 02129, USA

Key words: mesoscope; mesoscale; cerebral cortex; fluorescence; hemoglobin absorption

§**Corresponding authors:**

Anna Devor and Martin Thunemann

610 Commonwealth Ave

Boston, MA 02215

USA

Email: adevor@bu.edu; martinth@bu.edu

Acknowledgements: We thank Ahmed Abdelfattah for generously providing purified mApple

fluorescent protein for calibration of our imaging system. We gratefully acknowledge support from

the NIH (BRAIN Initiative R01MH111359, BRAIN Initiative U19NS123717, R01DA050159,

R01NS108472). Patrick Doran was supported by the Ruth L. Kirschstein Predoctoral Fellowship

F31NS118949. Rockwell P. Tang was supported by NIGMS T32GM145455. The authors

acknowledge that data analysis was performed on the Shared Computing Cluster which is

administered by Boston University's Research Computing Services. We thank the members of

the Neurovascular Imaging Laboratory for helpful discussions.

Author contributions: Patrick R. Doran: Investigation, Formal Analysis, Software, Visualization,

Writing – original draft, Writing – review & editing; Natalie Fomin-Thunemann: Investigation,

Writing – review & editing; Rockwell P. Tang: Investigation, Visualization, Writing – original draft,

Writing – review & editing; Bernhard Zimmermann, Kıvılcım Kılıç, Joel D. Herbert, Grace

Chabbott, John X. Jiang: Resources, Writing – review & editing; Sreekanth Kura, Harrison P.

Fisher: Software, Writing – review & editing; Sava Sakadzic, David A. Boas: Conceptualization,

Writing – review & editing; Anna Devor: Conceptualization, Funding acquisition, Supervision,

Writing – original draft, Writing – review & editing; Ichun Anderson Chen: Conceptualization,

Methodology, Supervision, Writing – review & editing; Martin Thunemann: Conceptualization,

Software, Supervision, Visualization, Writing – original draft, Writing – review & editing.

Data availability: Design blueprints, parts inventory, data acquisition and analysis software, and

data presented in this article are publicly available on github (https://github.com/NIL-

NeuroScience/WidefieldImagingAnalysis).

Abstract

Significance: Widefield microscopy of the entire dorsal part of mouse cerebral cortex enables large-scale ("mesoscopic") imaging of neuronal activity with fluorescent indicators as well as hemodynamics via oxy- and deoxyhemoglobin absorption. Versatile and cost-effective imaging systems are needed for large-scale, color-multiplexed imaging of multiple fluorescent and intrinsic contrasts.

Aim: Develop a system for mesoscopic imaging of two fluorescent and two reflectance channels.

Approach: Excitation of red and green fluorescence is achieved through epi-illumination. Hemoglobin absorption imaging is achieved using 525- and 625-nm LEDs positioned around the objective lens. An aluminum hemisphere placed between objective and cranial window provides diffuse illumination of the brain. Signals are recorded sequentially by a single sCMOS detector.

Results: We demonstrate performance of our imaging system by recording large-scale spontaneous and stimulus-evoked neuronal, cholinergic, and hemodynamic activity in awake head-fixed mice with a curved "crystal skull" window expressing the red calcium indicator jRGECO1a and the green acetylcholine sensor GRAB$_{ACh3.0}$. Shielding of illumination light through the aluminum hemisphere enables concurrent recording of pupil diameter changes.

Conclusions: Our widefield microscope design with single camera can be used to acquire multiple aspects of brain physiology and is compatible with behavioral readouts of pupil diameter.

Introduction

With the rapidly expanding list of fluorescent probes for neuronal activity and the development of large-scale measurements (Abdelfattah et al. 2022), widefield microscopy has become the measurement modality of choice for mesoscale studies of neuronal circuits and underlying behavior when cellular resolution is not required (Bouchard et al. 2009, Mohajerani et al. 2013, Ma et al. 2016b, Vanni et al. 2017, Wright et al. 2017, Couto et al. 2021). Probes available in distinct spectral variants (Zhao et al. 2011) can be spectrally multiplexed for simultaneous imaging of different aspects of neuronal activity (Lin and Schnitzer 2016). Using multiplexed probes for in vivo widefield imaging of head-fixed mice with "crystal skull" cranial windows (Kim et al. 2016) provides a view of the entire dorsal part of the cerebral cortex including the primary sensory, motor and higher association areas (Higley and Cardin 2022).

In the healthy brain, increases in neuronal activity are often accompanied by increases in blood flow, volume, and oxygenation in the active region. The increase in blood flow and volume occurs, in large part, due to dilation of cerebral arterioles responding to vasoactive messengers released by active neurons (Uhlirova et al. 2016b, Echagarruga et al. 2020). This hemodynamic response increases oxygen supply beyond the actual demand leading to a net increase in blood oxygenation. Both oxy- and deoxyhemoglobin (HbO and HbR, respectively) strongly absorb light in the blue-green spectrum. Therefore, the hemodynamic response results in time-variant modulation of fluorescence signals within that spectral range. The methods developed to correct fluorescent signals for this hemodynamic artifact fall into two categories. For the first correction method, enhanced green fluorescent protein (EGFP)-based indicators, including GCaMP, iGluSnFR, and GRAB indicators, are excited off-peak at their isosbestic point in the 390-420 nm range (Lerner et al. 2015, Allen et al. 2017, Musall et al. 2019, Lake et al. 2020, Couto et al. 2021). When excited at these wavelengths, fluorescence emission is independent from the analyte concentration (i.e., $[Ca^{2+}]$, [Glutamate], etc.), but affected by hemodynamic changes (Dana et al.

2019, Jing et al. 2020). This signal can then be regressed out of the analyte-dependent signal to

remove hemodynamic artifacts. Another correction method requires simultaneous acquisition of

light absorption of (at least two) wavelengths to quantify changes in HbO and HbR concentrations

(Δ[HbO] and Δ[HbR])(Ma et al. 2016b, Mitra et al. 2018). Furthermore, quantitative estimates of

Δ[HbO] and Δ[HbR] are important measurables on their own; they aid in understanding

cerebrovascular physiology and interpretation of non-invasive imaging signals generated through,

for example, functional magnetic resonance imaging (fMRI) (Uhlirova et al. 2016a).

Here, we describe a versatile widefield imaging system that enables two-color fluorescence

imaging in addition to concurrent acquisition of two absorption channels for quantitative estimation

of Δ[HbO] and Δ[HbR] with a single sCMOS detector. We shield the illumination light to prevent it

from reaching the animal's eyes using an aluminum light-blocking hemisphere to avoid unwanted

visual stimulation of the animal. Here, we provide optical design blueprints, a wiring diagram, a

timing chart, data acquisition software, and an inventory of all parts. We demonstrate the

performance of our new widefield microscope in awake mice expressing the mApple-based

calcium indicator jRGECO1a (Dana et al. 2016) and the EGFP-based acetylcholine (ACh) probe

GRAB$_{ACh3.0}$ (Jing et al. 2020).

Results

We designed and manufactured a widefield imaging system with a field of view of 10×10 mm^2 covering the dorsal cortex of a mouse. We use a single sCMOS detector for sequential recording of two fluorescent indicators (here: EGFP-derived GRAB$_{ACh3.0}$ and mApple-derived jRGECO1a) as well as hemoglobin absorption at 525 and 625 nm at an effective frame rate of 10 Hz or better without changes in the configuration of the emission path. **Figure 1** and **Supplementary Figure 1** provide an overview of the system layout and illustrate the spectral properties of the light sources and filter elements together with HbO and HbR absorption spectra as well as jRGECO1a and GRAB$_{ACh3.0}$ excitation and emission spectra. For fluorescence excitation, we use epi-illumination through the objective lens. A 470-nm LED with a maximal power of 3.7 W provides light for GRAB$_{ACh3.0}$ excitation, which is filtered through a 466/40 nm excitation bandpass filter, resulting in a maximal power of 403 mW at the sample surface. For jRGECO1a excitation, we use light from a 565-nm LED with a maximal power of 8.6 W which is filtered through a 560/14 nm excitation bandpass filter reaching a maximal power of 75 mW at the sample surface. A secondary peak in the 565-nm LED spectrum at ~425 nm is removed by an additional 450-nm longpass filter in the excitation path (**Figure 1A**). The collimated and filtered light is coupled into the infinity space using a multiband dichroic mirror (488/561 nm, 'main dichroic mirror'). This, in combination with a multi-band 'detection' filter (523/610 nm), allows GRAB$_{ACh3.0}$ and jRGECO1a fluorescence at 500-540 nm and 580-640 nm, respectively, to reach the sCMOS detector while blocking excitation light at 440-480 nm and 550-570 nm from reaching the camera (**Figure 1C-D**). Typical powers at the sample surface during image acquisition were 90-177 mW and 68-75 mW for 470 nm and 565 nm, respectively. We observed no or minimal bleaching over continuous acquisition periods of up to 30 minutes. The wavelengths for hemoglobin absorption imaging are within the respective emission bands of the green (525 nm) and red (625 nm) fluorophore (**Figure 1E**). The LEDs used for hemoglobin absorption imaging are placed in a custom-designed reflectance illumination ring

(RIR) located between the objective lens and the sample surface (**Figure 1B** and **Supplementary**

Figure 1D). The RIR contains six 525-nm LEDs and six 625-nm LEDs positioned at equal

distances with maximal powers of 170 mW and 371 mW per LED, respectively. To enable a

simple procedure for estimation of Δ[HbR] and Δ[HbO], we placed 10-nm bandpass filters in front

of each LED. This leads to a maximal power of 39.5 mW (525 nm) and 51 mW (625 nm) at the

sample surface. We placed a hollow aluminum hemisphere between the RIR and the sample

surface internally reflecting the light of the 525-nm and 625-nm LEDs. For imaging, the opening

of the hemisphere is brought close to the sample surface achieving diffuse and homogenous

illumination and efficiently eliminating most reflections from the curved glass window covering the

brain surface. Typical powers at the sample surface during image acquisition were 10.5 mW and

4.4 mW for the 525-nm and 625-nm LEDs, respectively.

To acquire all parts of the image simultaneously with a sCMOS camera with rolling shutter, we

ensure that the LEDs emit light only when all rows of the sCMOS chip are exposed. Therefore,

we control the LED drivers with two digital trigger signals combined through a logical AND gate

integrated circuit (IC) (**Figure 2**). The first input to the AND gate IC is generated by the camera

and is in 'high'-state when all rows are being exposed; the second input is generated by the

microscope control computer and selects the respective LED wavelength to be switched on

(**Figure 2**). Only when both signals are in 'high' state, the LED power supply receives a trigger

signal to switch the respective LED on. Timing delays of the AND gate IC and the LED driver are

<9 ns and <100 μs, respectively, enabling virtually delay-free LED power switches for typical

exposure times of 2-8 ms per individual LED.

The imaging setup is controlled through a DAQ interface programmed via MATLAB. The main

acquisition script generates digital trigger signals controlling the imaging setup based on exposure

time, which can be different for every wavelength, acquisition rate, and total acquisition time.

Additional triggers are generated to synchronize or control external devices such as stimulus

apparatuses, behavioral cameras, electrophysiology recording equipment, and behavioral control

devices. The currents of the individual LEDs are set in MATLAB and sent via a USB-based

remote-control interface to the LED drivers. The sCMOS camera is controlled through the

manufacturer's software where parameters such as field of view and on-camera pixel binning are

set. This software also drives image acquisition; however, the camera is set to respond to

exposure triggers generated by the DAQ interface, with the (variable) trigger length determining

the exposure time. During acquisition, the DAQ interface records feedback signals from camera,

microscope control, and external devices allowing synchronization of data streams during off-line

post-processing and analysis.

While our single-camera approach reduces complexity and cost, we use multiband filters to image

two fluorophores present in the sample at the same time. The lack of spectral separation on the

emission side can potentially create crosstalk when fluorophores are excited off-resonance. Note

that both fluorophores undergo dynamic changes in fluorescence intensity, so the degree of

crosstalk will vary in time and space. Published spectra show that $GRAB_{ACh3.0}$ (or EGFP) is

essentially not excited at 565 nm, while jRGECO1a (or mApple) fluorescence is excited at

470 nm. To quantify the amount of crosstalk under well-controlled conditions, we collected images

of purified EGFP and mApple proteins with various combinations of illumination power and

exposure time for both the 470-nm and 565-nm LEDs. In line with the published spectra, we

observed that the 565-nm LED practically does not excite purified EGFP while mApple shows

fluorescence when excited with the 470 nm LED (**Supplementary Figure 2A**). Based on this *in*

vitro calibration with EGFP and mApple and the typical acquisition parameters we used for *in vivo*

imaging in $GRAB_{ACh3.0}$ and jRGECO1a expressing animals, we determined that the level of

crosstalk of $GRAB_{ACh3.0}$ to fluorescence measured upon 565-nm excitation is <1%, while

fluorescence from jRGECO1a contributes about 3-9% to the fluorescence intensity measured

upon 470-nm excitation (**Supplementary Figure 2**; see **Appendix B** for a detailed discussion of

the crosstalk analysis). The in-vitro calibration allows to correct for crosstalk; however, given the

relatively small degree of crosstalk, we did not apply the correction term as it potentially adds

additional noise without sufficient benefit (see **Appendix B** for a detailed discussion).

To demonstrate the capabilities of our new imaging system design, we recorded spontaneous

and stimulus-induced changes in neuronal calcium, acetylcholine, and hemodynamic signals in

awake head-fixed mice expressing jRGECO1a and $GRAB_{ACh3.0}$ (**Figure 3**). Mice were implanted

with a curved glass cranial window exposing both hemispheres of the dorsal cortex while leaving

the bone in-between the two hemispheres above the superior sagittal sinus intact (Kilic et al.

2020) (**Figure 3A**). Images were registered to the Allen Mouse Common Coordinate Framework

(Wang et al. 2020) to generate time courses from pre-defined regions of the dorsal cortex.

Measurements of spontaneous activity show a high correlation between neuronal calcium and

ACh time courses (**Figure 3B** shows right secondary motor cortex), which is consistent with

published results (Lohani et al. 2022). We then recorded the response to a sensory stimulus

consisting of a 2-s, 3-Hz air puff sequence (20 repetitions, 25 s inter-stimulus interval) delivered

to the right whisker pad (**Figure 3C-E**). Ratio maps show neuronal activation and functional

hyperemia in the contralateral barrel cortex (green ROI in calcium map, **Figure 3C**). The stimulus-

averaged time course calculated from the contralateral barrel cortex shows that the imaging

system operates at sufficient signal to noise ratio to record the neuronal calcium response to the

six individual air puffs comprising one stimulus train (**Figure 3D**). Despite high trial-to-trial

variability (**Figure 3E**), which is consistent with published widefield imaging studies in awake mice

(Musall et al. 2019), responses to individual air puffs are still discernible in individual trial

recordings. The time course of the hemodynamic response shows a broad peak and post-stimulus

undershoot. The hemodynamic correction we apply to the green fluorophore compensates for

darkening that occurs when hemoglobin levels increase. **Supplementary Figure 3** compares the

trial averaged ACh time courses with and without correction for hemodynamic artifacts.

The aluminum hemisphere in-between the cranial window and the imaging system reflects 525-

and 625-nm light to improve the even distribution of illumination light for hemoglobin absorption

imaging. However, we observe some degree of inhomogeneity due to the curvature of glass

window (**Figure 3A**). At the same time, the hemisphere acts as a shield preventing the strobing

illumination light from reaching the mouse's eyes. For example, four-color imaging with an

effective acquisition rate of 10 Hz would produce an undesired visual stimulus at 40 Hz. To

illustrate the shielding effect, we recorded changes of the animal's pupil diameter during

mesoscopic imaging in presence or absence of the aluminum hemisphere (**Figure 4**). For this

experiment, a behavioral camera recorded the pupil diameter continuously while widefield

illumination across the four wavelengths at 40 Hz was switched on and off at 60-s intervals; we

recorded 20 off-to-on and on-to-off transitions per test. The images shown in **Figure 4A** show the

pupil before and after turning illumination on; **Supplementary Video 1** shows a recording of the

off-to-on transition with and without hemisphere. **Figure 4B** shows representative 2-min time

series covering the off-to-on and on-to-off transitions in presence or absence of the hemisphere.

Without the hemisphere, pupil size measurements are largely dominated by the state of widefield

illumination. In the presence of the hemisphere, pupil size changes are largely independent from

widefield illumination (**Figure 4C**). Then, we compared the relative change in pupil size from 1.5-

0.5 s before versus 2-3 s after start of widefield illumination averaged over 20 transitions and

observe a significantly larger drop in pupil size without the hemisphere ($p<0.001$; t-test; **Figure**

4D). Pupil tracking is often used as an important variable to define behavioral states (Vinck et al.

2015, Reimer et al. 2016) and it is therefore desirable to combine pupil recordings and mesoscale

imaging (Shahsavarani et al. 2023). In our system, the hemisphere eliminates interference of

widefield illumination on behavioral recordings and enables high-quality measurements of pupil

dynamics during mesoscale imaging.

Discussion

Here, we describe a new widefield microscope design for mesoscopic imaging of dorsal mouse

cortex. Several publications describe widefield calcium imaging with a single fluorescent indicator

and concurrent hemodynamic imaging (Bouchard et al. 2009, Ma et al. 2016b, Wright et al. 2017,

Mitra et al. 2018, Shahsavarani et al. 2023). Others describe two-color fluorescence widefield

imaging without hemodynamic imaging (Feng et al. 2019, Lohani et al. 2022). Compared to

previous publications, our system has several unique features:

First, the system enables quasi-simultaneous imaging of two spectrally distinct fluorophores

(here: EGFP-based GRAB$_{ACh3.0}$ and mApple-based jRGECO1a) together with quantitative

estimates of Δ[HbO] and Δ[HbR] through two-color absorption imaging. Changes in [HbO] and

[HbR] are used to correct hemodynamic artifacts of the recorded fluorescence signals. Further,

we can use them to study the relationship between cortical hemodynamics and different aspects

of brain activity, such as neuronal firing and neuromodulation thereby gaining further insights into

brain state-dependent regulation of cerebral blood flow and metabolism (Machler et al. 2021).

Second, we use a single detector. Using one camera drastically reduces the overall cost to build

the system but requires sequential image acquisition for every wavelength (here: four

wavelengths). We achieve sufficient SNR with negligible spectral crosstalk at a 10 Hz acquisition

rate, which is within the range for widefield imaging studies in awake animals (Lake et al. 2020,

Lohani et al. 2022).

Finally, the illumination light is contained within an aluminum hemisphere. Besides improving the

uniformity of illumination with the 525- and 625-nm LEDs, the hemisphere blocks most of the

illumination light from the animal's view, thereby preventing interference with behavioral

measurements of pupil diameter as well as unwanted stimulation of the visual system. Pupil size

measurements are often used as a non-invasive, indirect readout of alertness (Joshi et al. 2016,

Reimer et al. 2016). We expect our shielded illumination design will enable seamless integration

of mesoscale imaging with behavioral experiments based on visual stimulation (Vinck, Batista-

Brito et al. 2015, Peters, Fabre et al. 2021) and virtual reality (Pinto, Rajan et al. 2019).

The overall cost to build our widefield imaging system with the components used is around

$45,000, with the camera comprising the bulk of the cost (~$25000). While this is a lower price

than for typical commercially available widefield imaging systems, modifications could further

reduce the overall cost of the system. Depending on the brightness of the used fluorophores, their

expression level, as well as the experimental demands, it is possible that sufficient imaging

performance could be achieved with a less expensive camera. Furthermore, we used

commercially available LED drivers as well as commercially available LEDs for fluorescence

excitation. Using a custom design or already available open-source components with similar

performance characteristics, especially regarding their response time to digital triggers, could

further reduce the overall cost.

Our imaging system can easily be modified or expanded to suit the requirements of different

experimental studies; changing a LED and the corresponding optical filters will allow combining

other fluorescent indicators if they are spectrally separated. This would, for example, allow

integration of calcium indicators excited with near-infrared light (Qian et al. 2020). As the spectral

diversity of genetically encoded molecular probes expands, widefield imaging with three spectrally

separated fluorescent indicators might soon become a feasible approach. Furthermore, we and

others combined mesoscale imaging with extracellular electrophysiology, providing simultaneous

access to cortical network activity as well as single- or multi-unit activity in targeted cortical and

subcortical areas (Allen et al. 2017, Thunemann et al. 2018, Lu et al. 2023).

Our current system design has the following limitations. First, uneven illumination causes different

levels of absolute image brightness across the field of view. In practice, however, varying absolute

intensities will not affect the readout of relative (i.e., intensity-normalized) changes in fluorescence

or absorption, except for potentially reduced signal-to-noise levels in darker regions of the image.

A possible solution to overcome uneven illumination is the incorporation of an integration device

with uniformity correction, such as EUCLID (Celebi et al. 2023) into our imaging system. Further,

uneven illumination in the medial-to-lateral direction is also caused by the curvature of the glass

window covering the cranial exposure (see **Figure 3A**). Use of curved glass also causes different

cortical depths to match the focal plane of the imaging system, which potentially introduces

spatially heterogenous sampling bias. Here, use of adaptive lens optics, such as meta-lenses

(Yang et al. 2023) could improve uniformity of illumination and could be used to match the focal

plane across the field of view. Finally, in our experimental animals, we removed the dorsal cranium

and replaced it with a curved glass window. In some preparations for mesoscopic imaging, the

skull is thinned, polished, and covered with a glass coverslip, leaving the interface between brain

and bone intact (Silasi et al. 2016). While we do not expect performance or quality issues to arise

when imaging transcranial preparations with our setup, we consider a careful side-by-side

comparison of these two window preparations a valuable subject of future experiments. Designing

an imaging system with a single camera reduces overall system cost. However, the use of two

cameras would allow for higher acquisition rates through spectral multiplexing. Further, the dual-

band dichroic mirror and emission filter cut the available spectral bandwidth of the

EGFP/GRAB$_{ACh3.0}$ emission band (500-540 nm) and the mApple/jRGECO1a excitation band (550-

570 nm). This potentially enables crosstalk due to off-peak excitation of the red fluorophore.

However, we achieve sufficient SNR with negligible spectral crosstalk at a 10 Hz acquisition rate,

which is within the range of appropriate temporal resolution for widefield imaging studies in awake

animals (Lake, Ge et al. 2020, Lohani, Moberly et al. 2022). Furthermore, the extension of the

imaging system by a second camera requires minimal modifications to the overall layout, allowing

for higher acquisition speeds, if desired. Another consideration during the design of our system

was the use of lasers instead of LEDs for excitation of the green and red fluorescent proteins.

While less expensive, LEDs have a broad emission spectrum that requires clean-up through

excitation filters. This limits the emitted light to the usable spectral range (see, e.g., **Figure 1D**,

565-nm LED) leading to an overall reduction in the available power compared to the nominal

output (75 mW versus 8.6 W for the 565-nm LED), potentially requiring longer integration times

to achieve sufficient SNRs. Here, use of a laser would enable excitation at higher power levels, if

required. A drawback of using lasers instead of LEDs would be a substantial increase in the cost

of the system.

In summary, given the versatility of our system and the comparatively low cost in contrast to

commercial intravital imaging setups, we expect our widefield imaging system design to become

a useful platform in the neuroscience and neurophotonics community.

Methods

Experimental animals

All procedures were conducted with approval from Boston University Institutional Animal Care and Use Committee and in accordance with the National Institutes of Health Guide for the Care and Use of Laboratory Animals. Standard rodent chow and water were provided ad libitum. Mice were housed in a 12-hour light cycle with lights turning on at 7:30 AM. We used three female mice of the Thy1-jRGECO1a GP8.20 (Dana et al 2016, Dana et al 2019, Lin et al 2016) line; cortex-wide GRAB$_{ACh3.0}$ expression was induced by injection of 1.5 µL AAV9-hSyn-GRAB$_{ACh3.0}$ (WZ Biosciences, 3.06×10^{13} GC/mL) into each transverse sinus (3 µL per animal) of 1-day old neonates according to procedures described in Lohani et al. (2022).

Headpost implantation and craniotomy in adult (8-16 week old) animals follows procedures described in Kilic et al. (2020). Briefly, a custom-designed headpost machined from titanium was attached to the cranium and a modified crystal skull curved glass window was used to replace the dorsal cranium on both hemispheres (Kim et al. 2016). Prior to surgery, the original curved glass (12 mm width, labmaker.org) was cut in half to obtain separate glass pieces for each hemisphere. Following the implantation procedure, a silicone plug was placed on top of the glass window and a 3D-printed cap was fixed to the headpost to prevent heat loss. Mice were allowed to recover for at least one week following surgery before behavioral training commenced.

To adapt mice to be comfortable under head fixation, animals were handled until they remained calm in the gloved hand of the investigator. Mice were then head-fixed for increasing periods while they received a reward of sweetened condensed milk. Data acquisition was started when animals tolerated one hour of head fixation.

Imaging system optical components

The imaging system is assembled from standard optomechanical components compatible with 2-

inch lens tubes. The imaging system is placed on an elevated platform (61×61 cm, Thorlabs,

MB2424) on a vibration-isolating table.

A 470-nm LED (Thorlabs, SOLIS-470C) is used to excite $GRAB_{ACh3.0}$ and a 565-nm LED

(Thorlabs, SOLIS-565C) is used to excite jRGECO1a. A 466/40 nm excitation filter (Semrock,

FF01-466/40, diameter 50 mm) is placed in front of the 470-nm LED and a 560/14 nm excitation

filter (Semrock, FF01-560/14, diameter 50 mm) is placed in front of the 565-nm LED. A dichroic

mirror combines excitation light from 565-nm and 470-nm LEDs (Semrock, FF520-Di02, 50 x 72

mm). The output of these LEDs is filtered through a 450-nm longpass filter (Chroma, AT450lp) to

remove low wavelength light from the 565 nm LED that leaks through the primary excitation filter.

A 45-degree mirror (Thorlabs, PF20-03-P01, diameter 50.8 mm) reflects excitation light down

towards the mouse and reflects emitted light into the plane of the camera. A dual-band 'main'

dichroic beam splitter (Semrock, Di03-R488/461, 50 x 72 mm) reflects excitation light towards the

mirror and transmits emitted light towards the camera. A dual-band emission filter (Semrock,

FF01-523/610, diameter 50 mm) is placed between a dichroic mirror and camera. A tube lens

(Olympus, MVX-TLU-MVX10) projects the image onto the sCMOS camera (Andor, Zyla 4.2P-

CL10 or Sona 4.2B-6).

An objective lens (Olympus, ZDM-1-MVX063) focuses the fluorescence excitation light onto the

brain for epi-illumination and projects the emitted light towards the camera. The objective lens is

surrounded by the reflective illumination ring (RIR) assembly, which contains six equally spaced

525 -nm LEDs (CREE LED, XBDGRN-00-0000-000000C02) and six equally spaced 625-nm

LEDs (CREE LED, XBDRED-00-0000-000000701) for reflectance imaging. Bandpass filters were

cut to 9.4-mm diameter and placed in front of the reflectance LEDs. These filters are 532/10 nm

bandpass filters (Edmund Optics, 65094) for the 525-nm LEDs and 632/10 nm bandpass filters

(Edmund Optics, 65105) for the 625-nm LEDs. Six Peltier elements (Ferro Tec, 81044) controlled

by a feedback-controlled temperature controller (Thorlabs, TED200C) actively cool the LEDs in

the RIR to maintain the LEDs at room temperature during data acquisition. Temperature is

measured on the LED ring with a 10-kΩ thermistor (TDK, NTCG163JF103FT1). Heat from the

Peltier elements is transferred to a custom heat sink made of aluminum. The RIR allows fixation

of custom hemisphere made of aluminum (bead-blasted) with a 15-mm opening located directly

above the cranial window.

Control of imaging acquisition

We use a computer-controlled DAQ device (National Instruments, PCIe-6363) to operate the

imaging system. The DAQ device is connected to two I/O connector boards (National Instruments,

CB-68LPR) inside a control interface box. The control interface relays a DAQ-generated exposure

trigger to the camera. An integrated circuit with an OR gate (Texas Instruments, SN74AHC32N)

within the control interface receives the "Fire ALL" trigger from the camera as an input. The "Fire

ALL" trigger is positive when all camera rows are exposed by the rolling shutter. The output of the

OR gate is the common input to an integrated circuit with four AND gates (Texas Instruments,

SN74AC08N). The second – wavelength-specific – input to each individual AND gate is the LED

selection trigger generated by the DAQ system. Each output of the AND gate connects to one

LED driver (Thorlabs, DC2200) controlling one type of LED. During normal operation of the

imaging system, the "Fire ALL" trigger controls when the LED turns on. The other input to the OR

gate is a camera bypass trigger that allows the LEDs to be triggered while the camera is off. We

use a custom MATLAB script to control the DAQ system. We used image capture software (Andor,

SOLIS) to control camera and image acquisition.

The animal was positioned under 525-nm illumination and the platform carrying the animal was

adjusted to bring the brain surface into focus. Then, the aluminum dome is lowered to minimize

the distance between its opening and the brain surface. Illumination intensities (i.e., LED currents)

were adjusted to bring image intensities into the dynamic range of the camera at 16-bit image

depth. The acquisition window was cropped to the size of the field of view; image height (number

of rows) was a variable entered into the MATLAB script controlling image acquisition: the length

of the period when all rows are exposed ("FIRE all") depends on image height and requires

adjustment of the camera exposure trigger generated by the DAQ system. Typical image sizes

were 460×360 pixels with a 9.2x7.2 mm field of view and 20 μm resolution at 4×4 binning. Based

on selected exposure times and frame rate, the MATLAB script calculates the total number of

frames per acquisition run; this number is entered into the image capture software. The camera

is set to "External Exposure" mode and image acquisition is started. Then, execution of the

MATLAB script is started causing the DAQ interface to send LED selection and camera triggers

to the control interface.

Sensory stimulation and behavioral recordings

The MATLAB acquisition script generates additional triggers for synchronization with behavioral

apparatuses, such as to deliver stimuli by puffing compressed air to the animals' whiskers. Stimuli

are defined by duration, frequency, repetitions, interstimulus interval for individual stimuli and

sequences of pre-stimulus baseline, stimulus, and post-stimulus observation period. Based on

these parameters, the MATLAB script calculates the required imaging time and number of

images. For air puff stimuli, the 5-V trigger is sent to a picopump (WPI, PV830) connected to a

2 -mm diameter plastic tube that terminates in a glass capillary to deliver pressurized air to the

mouse whiskers.

The animal's face and whiskers are recorded during imaging with a CCD camera (Basler,

acA1920-150uc) through a variable zoom lens (Edmund Optics, 67715). A 940-nm LED

(Thorlabs, M940L3) continuously illuminates the animal, and a 920-nm long pass filter prevents

light from the microscope from contaminating the camera recording. The mesoscope control

system triggers the CCD camera to acquire a single frame during each acquisition cycle.

An accelerometer (Analog Devices, ADXL335) is placed underneath the mouse bed to record

animal's motion. The accelerometer generates analog signals for motion in X, Y and Z direction

which are recorded by a separate DAQ system (NI, USB-6363) receiving a synchronization signal

from the mesoscope control system.

Image processing

The camera control software (Solis, Andor) stores the acquired images as raw data ('.dat')

together with metadata files ('.ini' and '.sifx'). For all subsequent import and processing steps, we

use MATLAB. Raw image data is loaded into MATLAB, bypassing conversion into image (.tif) files

by the camera software and enabling multi-stream data import from high-performance solid-state

drives with MATLAB's multi-processor environment. For all channels, a baseline image

representing the temporal average of all images for that channel is generated. For fluorescence

channels, ΔF/F images are computed on a pixel-by-pixel basis by dividing each image by the

baseline image. After calculating ΔF/F, linear trends are removed from the time course of each

pixel using the MATLAB function "detrend". The modified Beer-Lambert law is used to calculate

changes in HbO and HbR concentrations from 525 nm and 625 nm reflectance (see

implementation for MATLAB's multi-processor environment in Appendix A). Estimated Δ[HbO]

and Δ [HbR] are used to correct green fluorescence measurements for artificial intensity changes

caused by dynamic changes in excitation and emission light absorption by hemoglobin. The

correction method has been described in Ma et al. (2016a) and its implementation is further

detailed in **Appendix A**. In this correction approach, the measured fluorescence is equal to the

artifact-free change in fluorescence multiplied by a correction factor containing the pathlength of

in tissue and the tissue absorption coefficient. Pathlengths have been previously estimated based

on Monte Carlo simulations (Ma et al 2016). Since hemoglobin accounts for most of the light

absorption in brain tissue, light absorption depends largely on Δ[HbO] and Δ[HbR] and the known

HbO and HbR extinction coefficients at the respective wavelengths of fluorescence excitation and

emission. We calculate the light absorption coefficient for every pixel at every time point based

on our estimations of Δ[HbO] and Δ[HbR] and apply it to correct for hemodynamic darkening of

ΔF/F in the green (EGFP/GRAB$_{ACh3.0}$) channel.

Measurements with several consecutive trials are interpolated at 100-ms intervals using a time

vector relative to the stimulus onset. Trial-averaged ratio maps are calculated by averaging

individual time courses across trials.

Behavioral video recordings are stored as image sequences and imported into MATLAB. To

calculate pupil size, an elliptical region is drawn defining the eye and an intensity threshold is

manually defined. The pupil size is calculated as the number of pixels below the threshold. The

resulting time series is normalized to its maximal and minimal pupil size within the recording.

***In Vitro* Crosstalk calibration**

Solutions of purified 0.03 mg/mL EGFP (Abcam, ab84191) and 0.4 mg/mL mApple (kindly

provided by Ahmed Abdelfattah) in phosphate-buffered saline (PBS) were prepared prior to

imaging. 30-μL samples of EGFP, mApple, or PBS-only control solution were each injected into

one channel of separate 6-channel slides (ibidi, μ-Slide VI – Flat). The center of the fluorophore-

filled channel was focused under the widefield imaging system before starting image acquisition

for each slide. Illumination during an imaging sequence alternated between the 470-nm and

565-nm LEDs while stepping through a series of camera exposure times and increasing LED

currents. Three trials were conducted for each sample, with the position of the slide under the

objective being shifted between trials to account for intensity variations. To calculate the

fluorescence intensity, pixel values were averaged over a region of interest covering the center

of the fluorophore-containing channel. Then, we took the mean of the three trials per condition

and subtracted the fluorescence intensity of the PBS control. Further analysis of crosstalk is

detailed in **Appendix B**.

Figure Legends

Figure 1. Schematics of the wide field imaging system and fluorescence and reflectance spectra. A. A single sCMOS camera is used to measure emission of green and red fluorophores and two hemoglobin absorption wavelengths. A 565-nm LED excites mApple-based jRGECO1a and a 470-nm LED excites EGFP-based GRAB$_{ACh3.0}$; both LEDs are filtered through excitation filters. The main dichroic mirror reflects excitation light towards a 45°-mirror and then a 0.63× objective lens to focus the excitation light onto the sample. Below the objective lens is the reflectance illumination ring (RIR) further detailed in panel B. An aluminum hemisphere between the RIR and the sample surface internally reflects the 525- and 625-nm LEDs for homogenous illumination and prevents illumination light from reaching the animal's eyes. Fluorescence emission and reflectance pass the main dichroic and the detection filter before being focused by a tube lens to the camera. **B.** Cross-section of the reflectance illumination ring (RIR) setup. Each of the six 525- and 625-nm LEDs are arranged on a ring-shaped board at equal distances. Peltier elements actively cool the LEDs to prevent temperature changes resulting in unstable illumination intensities. 10 nm wide excitation filters are mounted in front of each LED. **C.** Excitation (solid blue line) and emission (solid green line) of EGFP-derived GRAB$_{ACh3.0}$ (Jing et al. 2020). The light-blue and dark-blue shaded areas represent the 470-nm LED spectrum before and after passing the excitation filter (BP466/40, blue dotted line). **D.** Excitation (solid yellow line) and emission (solid red line) spectrum of mApple-derived jRGECO1a (Dana et al. 2016). The light-orange and dark-orange areas represent the 565-nm LED spectrum before and after passing the excitation filter (BP560/14, orange dotted line). **E.** Absorption spectra of oxyhemoglobin (HbO, red line) and deoxyhemoglobin (HbR, blue line) from Prahl (1999). LED spectra are confined by excitation filters (525 nm in light and dark green; 625 nm in light and dark red before and after passing through the respective excitation filters).

Figure 2. Timing diagram to trigger LEDs and camera acquisition. MATLAB controls the DAQ

system via digital triggers controlling camera and LED drivers. A different LED illuminates the

sample during each frame for near-simultaneous acquisition. The effective frame rate is

determined by the length of the four-frame cycle (typically 100 ms). Grey lines next to the camera

Fire ALL trigger represent exposure of pixel rows in rolling shutter mode. The Fire ALL trigger is

sent when all rows are exposed. Each LED driver (bottom left) is controlled by the output of a

separate AND gate. The Fire ALL trigger is the common input to every AND gate while the

individual LED selection triggers are the second input for each AND gate.

Figure 3: Spontaneous and stimulus-evoked activity in the somatosensory cortex. A.

Average of the first ten images collected during a ten-minute acquisition for each illumination

wavelength. The two hemispheres were manually masked around the cranial window. A green

polygon was drawn around the right secondary motor cortex, which was determined by

registration to the Allen atlas of the mouse brain. **B.** Spontaneous activity in the right secondary

motor cortex (green polygon in panel A) for calcium, acetylcholine (ACh), and oxygenated,

deoxygenated, and total hemoglobin (HbO, HbR, HbT). **C.** Ratio maps averaged across 20 trials

showing the response to a 2-s, 3-Hz train of air puffs to the right whisker pad. Maps show data

from 1.7 s after stimulus onset except for the calcium map, which shows data from 0.1 s after

stimulus onset. **D.** Average time course of the stimulus-evoked response in the contralateral barrel

cortex (green region in panel C). The pink dashed line indicates when the calcium ratio map in

panel C is shown whereas the black dashed line indicates when ACh, HbT and HbO ratio maps

in panel C are shown. The grey-shaded area indicates the duration of the stimulus. **E.** Response

of the contralateral barrel cortex to individual stimulus trains. The trials are sorted according to

the magnitude of the calcium response.

Figure 4: The aluminum hemisphere shields illumination and reduces pupil constriction

during mesoscale imaging. A. Images of the mouse face 0.5 s before and 5 s after widefield

imaging acquisition begins. Green circles highlight the pupils. **B.** Example two-minute pupil

diameter time courses showing the off-to-on and on-to-off transition. The grey-shaded area

indicates the period when widefield acquisition was performed **C.** Average pupil diameter changes

upon beginning of widefield imaging (20 trials; traces show average ± standard deviation). Pupil

diameter 1.5-0.5 s before imaging was defined as baseline for each trial. The green-dashed line

indicates the start of widefield imaging, the black dashed lines indicate time points of the images

shown in panel A. **D.** Comparison of the ratio of pupil size 2-3 s after start of imaging to pupil size

1.5-0.5 s before imaging. Significance was tested with a student's t test.

References

Abdelfattah, A. S., S. Ahuja, T. Akkin, S. R. Allu, J. Brake, D. A. Boas, E. M. Buckley, R. E. Campbell, A. I. Chen, X. Cheng, T. Cizmar, I. Costantini, M. De Vittorio, A. Devor, P. R. Doran, M. El Khatib, V. Emiliani, N. Fomin-Thunemann, Y. Fainman, T. Fernandez-Alfonso, C. G. L. Ferri, A. Gilad, X. Han, A. Harris, E. M. C. Hillman, U. Hochgeschwender, M. G. Holt, N. Ji, K. Kilic, E. M. R. Lake, L. Li, T. Li, P. Machler, E. W. Miller, R. C. Mesquita, K. Nadella, U. V. Nagerl, Y. Nasu, A. Nimmerjahn, P. Ondrackova, F. S. Pavone, C. Perez Campos, D. S. Peterka, F. Pisano, F. Pisanello, F. Puppo, B. L. Sabatini, S. Sadegh, S. Sakadzic, S. Shoham, S. N. Shroff, R. A. Silver, R. R. Sims, S. L. Smith, V. J. Srinivasan, M. Thunemann, L. Tian, L. Tian, T. Troxler, A. Valera, A. Vaziri, S. A. Vinogradov, F. Vitale, L. V. Wang, H. Uhlirova, C. Xu, C. Yang, M. H. Yang, G. Yellen, O. Yizhar and Y. Zhao (2022). "Neurophotonic tools for microscopic measurements and manipulation: status report." Neurophotonics **9**(Suppl 1): 013001.

Allen, W. E., I. V. Kauvar, M. Z. Chen, E. B. Richman, S. J. Yang, K. Chan, V. Gradinaru, B. E. Deverman, L. Luo and K. Deisseroth (2017). "Global Representations of Goal-Directed Behavior in Distinct Cell Types of Mouse Neocortex." Neuron **94**(4): 891-907 e896.

Bouchard, M. B., B. R. Chen, S. A. Burgess and E. M. Hillman (2009). "Ultra-fast multispectral optical imaging of cortical oxygenation, blood flow, and intracellular calcium dynamics." Opt Express **17**(18): 15670-15678.

Celebi, I., M. Aslan and M. S. Unlu (2023). "A spatially uniform illumination source for widefield multi-spectral optical microscopy." PLoS One **18**(10): e0286988.

Couto, J., S. Musall, X. R. Sun, A. Khanal, S. Gluf, S. Saxena, I. Kinsella, T. Abe, J. P. Cunningham, L. Paninski and A. K. Churchland (2021). "Chronic, cortex-wide imaging of specific cell populations during behavior." Nat Protoc **16**(7): 3241-3263.

Dana, H., B. Mohar, Y. Sun, S. Narayan, A. Gordus, J. P. Hasseman, G. Tsegaye, G. T. Holt, A. Hu, D. Walpita, R. Patel, J. J. Macklin, C. I. Bargmann, M. B. Ahrens, E. R. Schreiter, V. Jayaraman, L. L. Looger, K. Svoboda and D. S. Kim (2016). "Sensitive red protein calcium indicators for imaging neural activity." Elife **5**.

Dana, H., Y. Sun, B. Mohar, B. K. Hulse, A. M. Kerlin, J. P. Hasseman, G. Tsegaye, A. Tsang, A. Wong, R. Patel, J. J. Macklin, Y. Chen, A. Konnerth, V. Jayaraman, L. L. Looger, E. R. Schreiter, K. Svoboda and D. S. Kim (2019). "High-performance calcium sensors for imaging activity in neuronal populations and microcompartments." Nat Methods **16**(7): 649-657.

Echagarruga, C. T., K. W. Gheres, J. N. Norwood and P. J. Drew (2020). "nNOS-expressing interneurons control basal and behaviorally evoked arterial dilation in somatosensory cortex of mice." Elife **9**.

Feng, J., C. Zhang, J. E. Lischinsky, M. Jing, J. Zhou, H. Wang, Y. Zhang, A. Dong, Z. Wu, H. Wu, W. Chen, P. Zhang, J. Zou, S. A. Hires, J. J. Zhu, G. Cui, D. Lin, J. Du and Y. Li (2019). "A

Genetically Encoded Fluorescent Sensor for Rapid and Specific In Vivo Detection of Norepinephrine." Neuron **102**(4): 745-761 e748.

Higley, M. J. and J. A. Cardin (2022). "Spatiotemporal dynamics in large-scale cortical networks." Curr Opin Neurobiol **77**: 102627.

Jing, M., Y. Li, J. Zeng, P. Huang, M. Skirzewski, O. Kljakic, W. Peng, T. Qian, K. Tan, J. Zou, S. Trinh, R. Wu, S. Zhang, S. Pan, S. A. Hires, M. Xu, H. Li, L. M. Saksida, V. F. Prado, T. J. Bussey, M. A. M. Prado, L. Chen, H. Cheng and Y. Li (2020). "An optimized acetylcholine sensor for monitoring in vivo cholinergic activity." Nat Methods **17**(11): 1139-1146.

Joshi, S., Y. Li, R. M. Kalwani and J. I. Gold (2016). "Relationships between Pupil Diameter and Neuronal Activity in the Locus Coeruleus, Colliculi, and Cingulate Cortex." Neuron **89**(1): 221-234.

Kilic, K., M. Desjardins, J. Tang, M. Thunemann, S. Sunil, S. E. Erdener, D. D. Postnov, D. A. Boas and A. Devor (2020). "Chronic Cranial Windows for Long Term Multimodal Neurovascular Imaging in Mice." Front Physiol **11**: 612678.

Kim, T. H., Y. Zhang, J. Lecoq, J. C. Jung, J. Li, H. Zeng, C. M. Niell and M. J. Schnitzer (2016). "Long-Term Optical Access to an Estimated One Million Neurons in the Live Mouse Cortex." Cell Rep **17**(12): 3385-3394.

Lake, E. M. R., X. Ge, X. Shen, P. Herman, F. Hyder, J. A. Cardin, M. J. Higley, D. Scheinost, X. Papademetris, M. C. Crair and R. T. Constable (2020). "Simultaneous cortex-wide fluorescence Ca(2+) imaging and whole-brain fMRI." Nat Methods **17**(12): 1262-1271.

Lerner, T. N., C. Shilyansky, T. J. Davidson, K. E. Evans, K. T. Beier, K. A. Zalocusky, A. K. Crow, R. C. Malenka, L. Luo, R. Tomer and K. Deisseroth (2015). "Intact-Brain Analyses Reveal Distinct Information Carried by SNc Dopamine Subcircuits." Cell **162**(3): 635-647.

Lin, M. Z. and M. J. Schnitzer (2016). "Genetically encoded indicators of neuronal activity." Nat Neurosci **19**(9): 1142-1153.

Lohani, S., A. H. Moberly, H. Benisty, B. Landa, M. Jing, Y. Li, M. J. Higley and J. A. Cardin (2022). "Spatiotemporally heterogeneous coordination of cholinergic and neocortical activity." Nat Neurosci **25**(12): 1706-1713.

Lu, X., Y. Wang, Z. Liu, Y. Gou, D. Jaeger and F. St-Pierre (2023). "Widefield imaging of rapid pan-cortical voltage dynamics with an indicator evolved for one-photon microscopy." Nature Communications **14**(1): 6423.

Ma, Y., M. A. Shaik, S. H. Kim, M. G. Kozberg, D. N. Thibodeaux, H. T. Zhao, H. Yu and E. M. Hillman (2016a). "Wide-field optical mapping of neural activity and brain haemodynamics: considerations and novel approaches." Philos Trans R Soc Lond B Biol Sci **371**(1705): 20150360.

Ma, Y., M. A. Shaik, M. G. Kozberg, S. H. Kim, J. P. Portes, D. Timerman and E. M. Hillman
(2016b). "Resting-state hemodynamics are spatiotemporally coupled to synchronized and
symmetric neural activity in excitatory neurons." Proc Natl Acad Sci U S A **113**(52): E8463-E8471.

Machler, P., T. Broggini, C. Mateo, M. Thunemann, N. Fomin-Thunemann, P. R. Doran, I. Sencan,
K. Kilic, M. Desjardins, H. Uhlirova, M. A. Yaseen, D. A. Boas, A. A. Linninger, M. Vergassola, X.
Yu, L. D. Lewis, J. R. Polimeni, B. R. Rosen, S. Sakadzic, R. B. Buxton, M. Lauritzen, D. Kleinfeld
and A. Devor (2021). "A Suite of Neurophotonic Tools to Underpin the Contribution of Internal
Brain States in fMRI." Curr Opin Biomed Eng **18**: 100273.

Mitra, A., A. Kraft, P. Wright, B. Acland, A. Z. Snyder, Z. Rosenthal, L. Czerniewski, A. Bauer, L.
Snyder, J. Culver, J. M. Lee and M. E. Raichle (2018). "Spontaneous Infra-slow Brain Activity Has
Unique Spatiotemporal Dynamics and Laminar Structure." Neuron **98**(2): 297-305 e296.

Mohajerani, M. H., A. W. Chan, M. Mohsenvand, J. LeDue, R. Liu, D. A. McVea, J. D. Boyd, Y.
T. Wang, M. Reimers and T. H. Murphy (2013). "Spontaneous cortical activity alternates between
motifs defined by regional axonal projections." Nat Neurosci **16**(10): 1426-1435.

Musall, S., M. T. Kaufman, A. L. Juavinett, S. Gluf and A. K. Churchland (2019). "Single-trial
neural dynamics are dominated by richly varied movements." Nat Neurosci **22**(10): 1677-1686.

Prahl, S. (1999). "Optical Absorption of Hemoglobin."   Retrieved 11/3/2023, 2023, from
https://omlc.org/spectra/hemoglobin/.

Qian, Y., D. M. O. Cosio, K. D. Piatkevich, S. Aufmkolk, W. C. Su, O. T. Celiker, A. Schohl, M. H.
Murdock, A. Aggarwal, Y. F. Chang, P. W. Wiseman, E. S. Ruthazer, E. S. Boyden and R. E.
Campbell (2020). "Improved genetically encoded near-infrared fluorescent calcium ion indicators
for in vivo imaging." PLoS Biol **18**(11): e3000965.

Reimer, J., M. J. McGinley, Y. Liu, C. Rodenkirch, Q. Wang, D. A. McCormick and A. S. Tolias
(2016). "Pupil fluctuations track rapid changes in adrenergic and cholinergic activity in cortex."
Nat Commun **7**: 13289.

Shahsavarani, S., D. N. Thibodeaux, W. Xu, S. H. Kim, F. Lodgher, C. Nwokeabia, M. Cambareri,
A. J. Yagielski, H. T. Zhao, D. A. Handwerker, J. Gonzalez-Castillo, P. A. Bandettini and E. M. C.
Hillman (2023). "Cortex-wide neural dynamics predict behavioral states and provide a neural
basis for resting-state dynamic functional connectivity." Cell Rep **42**(6): 112527.

Silasi, G., D. Xiao, M. P. Vanni, A. C. Chen and T. H. Murphy (2016). "Intact skull chronic windows
for mesoscopic wide-field imaging in awake mice." J Neurosci Methods **267**: 141-149.

Thunemann, M., Y. Lu, X. Liu, K. Kilic, M. Desjardins, M. Vandenberghe, S. Sadegh, P. A. Saisan,
Q. Cheng, K. L. Weldy, H. Lyu, S. Djurovic, O. A. Andreassen, A. M. Dale, A. Devor and D. Kuzum
(2018). "Deep 2-photon imaging and artifact-free optogenetics through transparent graphene
microelectrode arrays." Nat Commun **9**(1): 2035.

Uhlirova, H., K. Kilic, P. Tian, S. Sakadzic, L. Gagnon, M. Thunemann, M. Desjardins, P. A.
Saisan, K. Nizar, M. A. Yaseen, D. J. Hagler, Jr., M. Vandenberghe, S. Djurovic, O. A.
Andreassen, G. A. Silva, E. Masliah, D. Kleinfeld, S. Vinogradov, R. B. Buxton, G. T. Einevoll, D.
A. Boas, A. M. Dale and A. Devor (2016a). "The roadmap for estimation of cell-type-specific
neuronal activity from non-invasive measurements." Philos Trans R Soc Lond B Biol Sci
371(1705).

Uhlirova, H., K. Kilic, P. Tian, M. Thunemann, M. Desjardins, P. A. Saisan, S. Sakadzic, T. V.
Ness, C. Mateo, Q. Cheng, K. L. Weldy, F. Razoux, M. Vandenberghe, J. A. Cremonesi, C. G.
Ferri, K. Nizar, V. B. Sridhar, T. C. Steed, M. Abashin, Y. Fainman, E. Masliah, S. Djurovic, O. A.
Andreassen, G. A. Silva, D. A. Boas, D. Kleinfeld, R. B. Buxton, G. T. Einevoll, A. M. Dale and A.
Devor (2016b). "Cell type specificity of neurovascular coupling in cerebral cortex." Elife **5**.

Vanni, M. P., A. W. Chan, M. Balbi, G. Silasi and T. H. Murphy (2017). "Mesoscale Mapping of
Mouse Cortex Reveals Frequency-Dependent Cycling between Distinct Macroscale Functional
Modules." J Neurosci **37**(31): 7513-7533.

Vinck, M., R. Batista-Brito, U. Knoblich and J. A. Cardin (2015). "Arousal and locomotion make
distinct contributions to cortical activity patterns and visual encoding." Neuron **86**(3): 740-754.

Wang, Q., S. L. Ding, Y. Li, J. Royall, D. Feng, P. Lesnar, N. Graddis, M. Naeemi, B. Facer, A.
Ho, T. Dolbeare, B. Blanchard, N. Dee, W. Wakeman, K. E. Hirokawa, A. Szafer, S. M. Sunkin,
S. W. Oh, A. Bernard, J. W. Phillips, M. Hawrylycz, C. Koch, H. Zeng, J. A. Harris and L. Ng
(2020). "The Allen Mouse Brain Common Coordinate Framework: A 3D Reference Atlas." Cell
181(4): 936-953 e920.

Wright, P. W., L. M. Brier, A. Q. Bauer, G. A. Baxter, A. W. Kraft, M. D. Reisman, A. R. Bice, A.
Z. Snyder, J. M. Lee and J. P. Culver (2017). "Functional connectivity structure of cortical calcium
dynamics in anesthetized and awake mice." PLoS One **12**(10): e0185759.

Yang, F., M. Y. Shalaginov, H.-I. Lin, S. An, A. Agarwal, H. Zhang, C. Rivero-Baleine, T. Gu and
J. Hu (2023). "Wide field-of-view metalens: a tutorial." Advanced Photonics **5**(03): 033001.

Zhao, Y., S. Araki, J. Wu, T. Teramoto, Y. F. Chang, M. Nakano, A. S. Abdelfattah, M. Fujiwara,
T. Ishihara, T. Nagai and R. E. Campbell (2011). "An expanded palette of genetically encoded
Ca(2)(+) indicators." Science **333**(6051): 1888-1891.

Figure 1

A 565 nm LED
560/14 nm Filter
Combining dichroic
470 nm LED
450 nm Longpass filter
466/40 nm Filter
Mirror
Main dichroic
Detection filter
Lens
Camera
Lens
RIR
Hemi-sphere

······ Reflectance
—— Excitation
—— Emission

B <u>R</u>eflectance <u>I</u>llumination <u>R</u>ing
Mounting adapter
Mount / heatsink
Peltier circuit board
Peltier element
Copper ring
LED circuit board
LED and filter mount
Bandpass filter

C Fluorescence GRAB$_{ACh3.0}$ (EGFP-derived)
470 nm LED
Excitation
Emission
466/40

D Fluorescence jRGECO1a (mApple-derived)
565 nm LED
Excitation
Emission
560/14

E Reflectance (Hemoglobin absorption)
525 nm LED
625 nm LED
HbO
HbR

Wavelength (nm)

650   --- Main dichroic   ---Detection filter   ---Combining dichroic

651 **Figure 2**

Figure 3

655   **Figure 4**

31