1

An enclosed live-imaging platform to reveal ultrasound-inflicted mechanolysis and wound healing response

Ziyue Bai[1], Zaimeng Li[2], and Yue Shao[1,*]

[1] Institute of Biomechanics and Medical Engineering, Department of Engineering Mechanics, School of Aerospace Engineering, Tsinghua University, Beijing 100084, China; [2] Institute of Fluid Mechanics, Department of Engineering Mechanics, School of Aerospace Engineering, Tsinghua University, Beijing 100084, China

* Correspondence should be addressed to Y. Shao (yshao@tsinghua.edu.cn)

ABSTRACT

Focused ultrasound (FUS) technology as a non-invasive therapy has been widely adopted in medical and life science research. Although various physical and biological effects of FUS have been well-documented, there is still a lack of understanding and direct evidence on the biological mechanism of therapeutic cell ablation caused by high-intensity ultrasound (HIFU) and the subsequent wound healing responses. Herein, this study designed and built an enclosed cell culture device that can synergistically combine non-invasive FUS stimulation and long-term live-cell imaging, serving as an in vitro platform to explore both short and long-term biological effects of ultrasound. Further, the process, mechanism and wound healing response of cell ablation induced by HIFU were studied, revealing a unique mechanism, termed ultrasound-inflicted mechanolysis, that is mediated by growing cavitation air bubbles under confined contact with cells. This discovery provides a previously unappreciated mechanism for understanding the biomechanical principles of ultrasound-based ablative therapy. We also found a post-ablation phantom layer (PAPL) remaining attached to neighboring live cell monolayer after HIFU stimulation. Such PAPL serves as a guiding cue for collective cell migration during post-ablation wound healing process, thereby providing a biomimetic model for studying wound healing after HIFU-inflicted damage. Together, the findings of this study might provide theoretical and technological basis for advancing our understanding of the biological effects of ultrasound-based ablative therapy and inspiring clinically relevant applications in the future.

INTRODUCTION

In recent years, focused ultrasound (FUS) technology has attracted extensive attention as a non-invasive therapeutic modality and has rapidly advanced in various research areas[1-3]. Currently, the most common application of therapeutic FUS is in the ablation of tumors[4, 5]. Furthermore, FUS exhibits deep penetration and high spatial precision in comparison to other therapeutic physical stimuli. Low-frequency focused ultrasound (frequency < 1MHz) can non-invasively focus within the cranial region[6] and it can be combined with other therapeutic modalities such as microbubbles and phase-change liquid droplets[7]. By promoting transcellular endocytosis, inducing endothelial cell membrane opening, forming cavities or channels within the cytoplasm, and opening tight junctions between endothelial cells, FUS facilitates the delivery of therapeutic agents, *e.g.*, cisplatin[8], doxorubicin[9, 10], mistletoe lectin-1 (ML1)[11], trastuzumab[12], and genetic materials[13] across the blood-brain barrier (BBB)[14, 15] or blood-tumor barrier (BTB)[16-18]. On neuroregulation, FUS can either inhibit or activate neural activity[19-21], showing significant potential in analgesia[22]. Its stimulation of the hepatic portal venous plexus holds promise as a non-pharmacological treatment for glucose homeostasis dysregulation[23]. Moreover, FUS has been applied in thrombolysis[24], as well as in therapeutic research related to inflammation[25, 26]. Overall, FUS-based therapeutic techniques offer vast avenues for clinically relevant applications.

The therapeutic principle of FUS technology is based on the energy accumulation generated when ultrasound waves are focused on specific areas in tissues, leading to thermal or mechanical effects. Subsequently, a series of biological effects are induced to achieve therapeutic objectives. The thermal effect of focused ultrasound refers to the higher energy accumulation at the focal point, causing a significant temperature increase in the targeted tissue compared to non-focal tissue. Different temperature elevation can achieve varying regulatory effects[27, 28]. For instance, using low-intensity focused ultrasound (LIFU) can mildly increase the temperature (*e.g.*, a rise of 2-6℃) of the targeted

tissue as a form of local hyperthermia[29]. On the other hand, employing high-intensity focused ultrasound

(HIFU) for a short duration can heat the targeted tissue to a high temperature (above 55°C), leading to

coagulative necrosis and "thermal ablation" of tissue. The mechanical effects of FUS primarily include

cavitation effects, which are divided into stable cavitation and inertial cavitation. Stable cavitation refers

to the stable oscillation of microbubbles in the focal region under the effect of pressure waves. Inertial

cavitation involves the drastic volume changes of microbubbles under the influence of strong pressure

waves, ultimately leading to their rupture and generating shock waves[30]. Despite considerable research

on the physical effects of FUS, the investigation into the biological effects mechanism such as cell

ablation induced by HIFU, and the subsequent wound healing response, remains limited. This hinders a

comprehensive understanding of the mechanisms underlying HIFU therapy and limits improvements

and optimizations in ultrasound therapeutic methods. Therefore, it is critically needed to explore the

mechanism of HIFU-induced cell ablation and associated wound healing responses.

To investigate the mechanisms of ultrasound therapy, *e.g.*, HIFU-induced cell ablation, it is

essential to develop experimental models that can simulate multicellular damage and repair processes

under ultrasound treatment in vitro. Traditionally, for in vitro studies on the biological effects of

ultrasound, researchers would typically expose cells to FUS treatment and then observe them under a

microscope[31-33]. This method is relatively simple and convenient but lacks real-time observation of cell

responses during FUS stimulation, which is necessary for studying short-term cell responses. Currently,

real-time monitoring devices for studying ultrasound-induced responses mainly rely on either an upright

or an inverted microscope. For example, Zhihai *et al.*[34] integrated a FUS device with an upright

microscope, placing the focused ultrasound transducer beneath the culture dish at a 45-degree angle.

When cells were stimulated with FUS, the incident sound waves were focused to the bottom of the

culture chamber through a waveguide material, then reflected at a right angle to reach an absorber.

Simultaneously, real-time imaging was performed by immersing the objective into the upright-placed

cell culture dish during stimulation. Recently, Yao-Shen et al.[35] combined a FUS device with an

inverted microscope, with the FUS transducer was immersed vertically into the cell culture medium.

However, these setups required immersing the objective lens or ultrasound transducer into an open cell

culture environment, leading to a significant risk of contamination. Despite that they might be suitable

for short-term experiments (0-30 minutes), they are insufficient to support continuous studies on the cell

ablation mechanisms and subsequent wound healing responses induced by HIFU. This hinders the

exploration and understanding of relevant mechanisms. Therefore, it is imperative to develop new

technologies that integrate non-invasive ultrasound stimulation, real-time live-cell imaging, and long-

term cell culture into a single platform. This is a key technical challenge that this study aims to address.

Herein, this study has designed and constructed a fully enclosed cell culture device (ECCD)

capable of simultaneously achieving non-invasive ultrasound stimulation and long-duration live-cell

imaging. This device serves as an experimental platform for exploring the biological effects of FUS.

Unlike previously reported open culture systems, this platform allows for real-time and programmable

FUS stimulation of cells in the target region, while also enabling long-term, online monitoring of

cellular responses using an inverted microscope. We applied this platform to obtain a stereotypic focal-

area geometry under FUS, and analyze the biophysical effects involved in different sub-regions of the

focal-area, including short and long-term cellular responses during and after HIFU-induced cell ablation.

This study not only provides direct experimental evidence and insights into classical thermal and

mechanical effects of FUS, but also reveals previously unappreciated biomechanical mechanisms

underlying the cell damage and wound healing involved in FUS-based ablative therapy.

RESULTS

Design of an enclosed cell culture device compatible with FUS stimulation and long-term live-cell imaging

In order to simultaneously achieve non-invasive FUS stimulation and long-term live-cell imaging, this study established an enclosed cell culture device (ECCD). Through optimized design of the acoustic path, acoustic windows, and sound-absorbing materials, the device enables FUS to act on a focal cell culture area at the bottom of the cell culture device (**Fig. 1a**). Meanwhile, real-time imaging captures changes and responses of the cultured cells in this region. Specifically, this design allows the FUS to be emitted from the transducer, reach the outside of the cell culture device via the coupling device, pass through the acoustic window, and further propagate in the cell culture medium. Ultrasonic waves then focus at the bottom of the ECCD, where a glass coverslip for cell culture is horizontally positioned. The axis of the ultrasound cone forms a 45-degree angle with coverslip, and are subsequently reflected at a 45-degree angle. Such reflected waves pass through the exit acoustic window and are absorbed by a sound-absorbing material attached to the window, thereby avoiding additional reflections that may affect the focused acoustic field within the cell culture device.

The ECCD is assembled from multiple components. To meet requirements for FUS stimulation and routine cell culture, we adopted a four-sided pyramid structure (**Fig. 1b**). The left and right sides of the device are inclined at 45-degree against the bottom surface, with circular and rectangular windows are opened on each, henceforth referred to as "entrance acoustic window (EnAW)" and "exit acoustic window (ExAW)", respectively. Both acoustic windows are sealed with 1 mm thick sheet of acoustically transparent material. This allows the FUS beam to transmit with extremely low loss along the expected acoustic path while avoiding direct contact between the ultrasound transducer and the liquid inside the culture device. A circle of grooves is extended from the inner edge of the EnAW and ExAW for positioning and stabilizing the acoustically transparent sheet. To ensure correct positioning and light contact of the ultrasound transducer coupler and the EnAW, a set of protruding keys were manufactured

at the inner rim of the EnAW to restrict the translational movement of the coupler. A circular opening

and recess structure are made at the bottom of the ECCD to allow the attachment of a coverslip for cell

culture (**Fig. 1b**).

Considering the relatively large size of the FUS transducer, a coupler is required to integrate it

with the ECCD. To match the shape of the ultrasound transducer and the ultrasound cone, the coupler is

designed to contain a conical head (**Fig. 1c&d**), which has a circular opening covered with a 1 mm thick

acoustically transparent sheet. The coupler also features a set of protruding keys complementary to those

on the rim of the EnAW of ECCD, ensuring the centered assembly and tight fit between them (**Fig. 1e**).

During operation, the coupler is filled with degassed water and connected through inlet/outlet ports to a

circulation system for cooling. The rear part of the coupler is a cylindrical cavity for placing the

transducer, before the coupler is sealed by a lid and a silicone ring upon mechanical tightening.

To examine whether the focused sound field in the device meets the design expectations, we

conducted multiphysics simulations using COMSOL. For simplicity, a two-dimensional model was used

(see **Methods** for details). The calculated sound pressure field, sound pressure level field

(**Supplementary Fig. 1c**), and sound intensity field (**Fig. 1f**) were normalized and plotted as heat maps.

These results showed that the ultrasound waves converged along the expected two-dimensional cone to

reach the focal point. Subsequently, they reached the ExAW along the ideal reflection angle and were

absorbed by the sound-absorbing material (modeled as a perfect matching layer in the simulation).

These results clearly indicate a focal ultrasound stimulation at the target region of the cell culture area,

with significantly lower level of ultrasound outside of the focal region. Similar results can be observed

in the sound intensity field image as well. Such focused stimulation performance meets the goal of our

device design. To evaluate the importance of sound-absorbing materials, we attempted to remove it in

the simulation (**Supplementary Fig. 1d**) and then observed highly disordered, non-focal sound pressure

field and sound pressure level field in the device (**Supplementary Fig. 1e**), as illustrated by sound waves undergoing multiple reflections. Under such condition, the energy cannot be accumulated at the focal point, and a stable energy distribution cannot be calculated in the steady-state sound intensity field. Therefore, the sound-absorbing material in the design of ECCD is crucial for ensuring a stable and predictable focused sound field. In addition, we compared the acoustic field of ECCD with that generated in a conventional upright imaging system, which was previously used to stimulate cells with ultrasound (**Supplementary Fig. 2**). Our simulation results showed that in contrast to the focal field formed in the ECCD, the conventional system only generated disordered acoustic field despite the use of a focused ultrasound transducer, due to the lack of design to optimize the acoustic path. Therefore, our results implicate the importance of ECCD device design to assure desirable combination of focal stimulation and unobstructed live-imaging.

An experimental platform for FUS stimulation and live-cell imaging based on ECCD

In order to integrate ECCD with laser confocal microscopy, a carrier system was designed. This system allows the transducer, coupler, and culture device to move freely in three dimensions within the microscope's field of view, facilitating optical focusing and field-of-view selection. To this end, a three-axis displacement platform, referred to as the primary displacement system, was employed (**Fig. 2a&b**). This system also mechanically isolates the weight of the ultrasound transducer, coupler, and culture device, from the microscope platform, thereby ensuring minimal disturbance to the optical systems as well as universal applicability to different kinds of optical setups. Furthermore, ensuring alignment and close contact between the ultrasound transducer coupler and the culture device is crucial for maintaining the stability and predictability of the acoustic pathway during experiments. Therefore, a secondary displacement system was mounted on the primary platform to control the relative motion between the coupler and the culture device. This system, with only one degree of freedom in the vertical direction

(Z-axis), independently controls the vertical motion and applies additional load to achieve precise

alignment between the coupler and the EnAW. With such a two-level carrier system, independent

adjustments can be made to the acoustic pathway as well as the optical path relative to the microscope,

decoupling the control of ultrasound stimulation and that of microscopic imaging.

High-precision 3D printing was chosen to manufacture the coupler and ECCD. The ECCD was

made from MED610 photopolymer resin, which is commonly used in orthopedic and oral surgery[36] and

expected to exhibit good biocompatibility, thermal stability, and chemical stability under cell culture and

ultrasound stimulation conditions. It also demonstrates mechanical stability to prevent deformation,

breakage, fatigue, and creep during assembly and experiments. The coupler was made from ABS

photopolymer resin (acrylonitrile-butadiene-styrene copolymer) using high-precision 3D printing (**Fig.**

2c&d), due to the robustness, density, and stability of the materials.

For the acoustic windows mounted on the ECCD and the coupler, 1mm-thick acoustically

transparent material was used to both ensure seamless coupling and avoid window bending due to liquid

pressure under working conditions. These acoustic windows should be made from materials that do not

significantly reflect sound waves at its interface with the culture medium (considered as water). Firstly,

several common laboratory materials were tested for acoustic impedance, from which

polydimethylsiloxane (PDMS) was found to have an impedance value close to that of water (**Fig. 2e&f**).

Subsequently, different PDMS were tested, and the acoustic impedance of PDMS slab at 4:1 mixing

ratio was found closest to that of water at the same temperature. Therefore, PDMS with base: curing

agent ratios at 4:1 was used for making acoustic windows. Before conducting experiments, the primary

displacement platform carrying the ultrasound stimulation system and the secondary displacement

platform carrying the ECCD (**Fig. 2g**) are assembled and mounted on the optical table of the microscope

(**Fig. 2h**).

In addition, we calibrated the acoustic field generated by the FUS transducer used in this study.

In particular, the calibration was conducted using a hydrophone to measure the sound intensity field near

the focal plane of the transducer. The results showed that when the input voltage to the transducer was

20 V, the frequency 1 MHz, the repetition frequency 10 kHz, and the duty cycle 30% for burst waves,

the -3 dB focal region (where the sound intensity dropped to 50% of the maximum intensity) on the X-Y

plane (perpendicular to the axis of the focusing acoustic cone) was approximately a circular area with a

diameter of 3 mm; while on the X-Z plane, the -3 dB focal region was an elliptical shape with a major

axis of about 18 mm and a minor axis of about 3 mm (**Supplementary Fig. 3a**).

Given that thermal effect due to HIFU is inevitable during experiments, we studied the dynamics

of average temperature change at the focal region. Specifically, we used a high-speed thermistor

measurement method. Specifically, a T25 culture bottle with a thermistor attached to the bottom and

filled with water was placed on the path of the FUS (**Supplementary Fig. 3b**). To ensure high thermal

sensitivity and minimal scattering of the acoustic waves, the thermistor was made from a platinum wire

(50 μm in diameter). To measure the average temperature over the focal region, the platinum wire was

bent and attached to a 3×3 mm^2 square area matching the size of the focal region (**Supplementary Fig.**

3c-e). Our results showed that under FUS stimulation, the temperature at the focal region could rise and

reach a steady state within about 5 s, and drop to the environmental temperature within 3 s after the FUS

was stopped, thus demonstrating the rapid temperature response at the focal region. Moreover, the target

temperature of the focal region can also be modulated by varying the ultrasound intensity and the duty

cycle of burst waves (**Supplementary Fig. 3f&g**).

HIFU-induced cell ablation does not show significant correlation with ultrasound-caused

thermal effect

To mimic the interaction between FUS and tissues (especially epithelial tissues, which are

commonly involved in ultrasound therapy), this study selected epithelial monolayer of Madin-Daby

canine kidney cells (MDCK cells) as an in vitro model. Firstly, to validate the compatibility of ECCD

with cell culture, MDCK cells were plated in ECCD and regular 60 mm culture dish and cultured for 2

days. At 50% and 100% confluence, no significant difference was observed between above two groups

in cell proliferation and epithelial colony morphology (**Fig. 3a**). In order to observe the shape and size of

focal region produced by FUS on MDCK monolayer in ECCD, we next set to simultaneously investigate

the temperature change of focal region and cell ablation under FUS stimulation. Although

abovementioned results from thermistor experiment showed the dynamics of average temperature

change in the focal region, it was not suitable for measuring the temperature field distribution of the

focal area. To characterize the temperature field, custom-made thermochromic tapes were used. These

tapes contained thermochromic inks with pre-calibrated color-changing temperatures of 31, 38, 45, and

65°C in red, yellow, green, and purple, respectively, and were mixed at volume ratio of 4:2:2:1. These

yinks changed color from brick red at room temperature to green, blue, purple, and milky white at

temperatures of 34, 38, 43, and 62°C, respectively (**Fig. 3b**). These tapes were applied to the bottom of

the culture device before FUS experiments, and the color changes were recorded to reconstruct the

temperature field changes during ultrasound stimulation. Additionally, to detect cell ablation after

ultrasound stimulation, Live dye was used to label the entire monolayer of MDCK cells at 100%

confluence before the experiment. Of note, while Live dye emits a strong green fluorescence (FITC)

inside cells with intact plasma membrane, it produces little or no fluorescence signal outside the cell or

once cell membrane was ruptured, thereby serving as an indicator of cell ablation. It is also important to

note that the disappearance of FITC signal only reflect cell ablation, not necessarily cell detachment.

Given that HIFU-based thermal effects are widely believed to be associated with ultrasound-

induced cell ablation in previous studies[4, 27], we next compared the changes in the focal region

temperature with the distribution of cell ablation. The "temperature field" was henceforth defined based

on the images of the thermochromic tape, with particular attention to the region where color change (*i.e.*,

temperature change) occurred. This temperature field could clearly visualize the focal area under FUS

treatment. Correspondingly, the images of cells captured under FITC fluorescence were henceforth

defined as the "cell field". Specifically, we define areas where cells were ablated (shown by the

disappearance of fluorescence during FUS treatment) as the "cell ablation area", and an elliptical-like

enveloping area that encloses all cell ablation area was defined as the focal area of cell field. Our results

indicated that under FUS stimulation, the focal area of the temperature field reached a steady state

within a short time (1-10 s). After stopping ultrasound, the focal area of the temperature field also

returned to ambient temperature within a relatively short time (**Supplementary Fig. 4a**). This was

consistent with the temperature change dynamics measured by thermistor-based method

(**Supplementary Fig. 3f&g**).

Of note, our results showed that the contours of the temperature field and the cell field were

roughly elliptical, with the major axis direction along the direction of ultrasound propagation (**Fig.**

3c&d). By measuring the lengths of the major and minor axes of the elliptical outline of the focal region

of the temperature field and cell field (**Fig. 3e**), we observed a positive correlation between FUS

intensity and the size of focal region of the temperature field, as well as that of the cell field. As

expected, similar correlation was also observed between FUS treatment time and above focal regions

(**Fig. 3c&d**).

However, by comparing the cell ablation area with the focal area of the temperature field, it is

clear that the focal area of the temperature field showed continuously decreasing temperature from the

center to the periphery, while the distribution of the cell ablation area was not necessarily continuous or

following the temperature gradient. Notably, the cell ablation area was a hollow elliptical shape under

conditions of high ultrasound intensity (transducer voltage 50 V, exposure time 3 min), and it became a

circular elliptical band under conditions of lower ultrasound energy (transducer voltage 20 V, exposure

time 3min), with non-uniform cell ablation within it (transducer voltage 40 V). Further quantitative

comparative analysis revealed no significant correlation in terms of shape and size between the cell

ablation area and the focal region of the temperature field (**Fig. 3f&g**). These observations suggested that

although changes in the temperature field occurred as expected within the elliptical focal region, the

continuous changes in the temperature field and the non-continuous, non-uniform changes in the cell

ablation area were inconsistent with the conventional understanding of ultrasound-based thermal

ablation theory. This implies that mechanisms other than thermal effects might exist for cell ablation

under FUS.

To further demonstrate that the thermal effects within the FUS focal area were insufficient to

ablate cells within this region in the short experimental observation time of this study, we placed the

ECCD (with a monolayer of MDCK cells at over 90% confluence) into water baths at 65°C and 85°C,

respectively, for 5 minutes. Compared with the untreated group, the proportion of cell ablation caused

by short-term treatment was less than 10% in both 65°C and 85°C groups (**Supplementary Fig. 4b&c**).

This is significantly different from the proportion of the cell ablation area within the focal region of cell

field under FUS stimulation. These results indicated that cell ablation produced under FUS stimulation

in a short time and the associated thermal effects in the focal area were not significantly correlated in a

global manner. Therefore, it is necessary to explore cell ablation mechanisms beyond thermal effects.

HIFU-induced cell ablation is associated with cavitation bubbles

To explore cell ablation mechanisms under FUS stimulation beyond global thermal effects, we

performed experiments using MDCK cells expressing red fluorescent protein in cell nuclei (H2B-RFP).

Firstly, we applied Live dye to visualize cell ablation region. **Fig. 4a** shows a typical structure of the

cell-ablation area after HIFU treatment, which features an elliptical ablation band surrounding a central

area with streak-like ablation domains and a small, central, high-temperature domains. Specifically,

when the FUS was activated for a few seconds, the streak-like ablation pattern appears near the

nucleation-like high-temperature domain (**Fig. 4a**), and expands from the side proximal to the incident

FUS waves, as shown in **Supplementary Fig. 5**. Of note, there were no bubbles in the streak-like

ablation domains during FUS treatment, suggesting it is perhaps starts from a combination of high-

temperature and shear force and then continued to expand due to acoustic streaming driven by the

nonlinear effect of ultrasound[37].In the meantime, when the FUS was activated for a few seconds, a large

number of bubbles gradually appeared near the side proximal to the incident FUS waves, causing cell

ablation at this site, which evolved into a crescent-like shape and eventually a complete elliptical

ablation band. Even after the cessation of FUS, some bubbles remained in the crescent-like region (**Fig.**

4b). These bubbles are opaque under the bright field, but almost transparent in the fluorescent channel,

with only a halo around the edge. These data suggest that above-observed cell ablation is closely

associated with bubble formation under FUS.

To further investigate the relationship between FUS-induced bubbles and cell ablation, we used

Dead dye to monitor cells that died during FUS treatment in real time. A band of bubbles (black arc

shown in **Fig. 4c&d**) first appeared, wherein bubbles were constantly generated, expanded and finally

burst under FUS. Due such mechanical effect of this band of bubbles, cells in the corresponding region

were ablated rapidly, as reflected by the presence of the Dead dye signal. Such ablated cell monolayer

eventually broke down and fractured and left a narrower void band. Meanwhile, we did not observe

notable temperature rise in this region, ruling out that the bubbles in this region are caused by the

heating and boiling of the liquid in this region. Therefore, we reason that this phenomenon might be

attributed to the inertial cavitation phenomenon.

Strikingly, we observed a large number of steadily expanding bubbles in the crescent-like region

proximal to the incident FUS waves. These bubbles, in contrast to above-mentioned inertial cavitation,

remain stable and expanding without bursting, in the absence of insignificant temperature increase. This

suggested that these bubbles and corresponding cell ablation in this particular region might reflect a

mechanism different from canonical effects such as thermal-ablation or cavitation bubble collapse.

To further illustrate the relationship between stable FUS-induced bubbles and cell ablation

observed in the proximal region of the crescent-like region, we treated the cells with a pulsed FUS of

varying intensity with a duty cycle of 10% and a pulse repetition rate of 1 kHz (**Supplementary Fig. 6**).

Compared with abovementioned continuous FUS waves, pulsed FUS with the same transducer input

voltage showed greatly reduced temperature within the focal region. Under pulsed FUS, the streak-like

ablation domains in the central region of the focal plane are still evident, further indicating it is likely

due to non-cavitation mechanical effect such as shear stream instead of thermal effect. Importantly, with

pulsed FUS, there was no notable formation of stable bubbles, nor formation of elliptical ablation band

with stable bubbles or cell ablation associated with it. These results indicate that FUS-induced cell

ablation in the unique crescent-like domain requires the formation of stable bubbles.

Notably, after cell ablation within the elliptical band, a layer of dead cells remained attached (**Fig.
4e**). Bright-field imaging results showed that while this layer of dead cells still appear to connect with

the rest of the cell monolayer, its morphology differed significantly from live cells. This suggests that

cell ablation under ultrasound may not result in the complete disappearance of cells and their substances

but leaves behind a post-ablation phantom layer (PAPL) composed of cellular remnants. Importantly,

the connections formed by the cytoskeleton in the phantom layer appear intact, maintaining connections

to adjacent live cell monolayer. Although such connections might hold the phantom layer in its original

place, it was also observed that after long-term FUS treatment, or at the location where large bubble

clusters form, the phantom layer could break away from the substrate (**Fig. 4f**). In this process, even part

of the live cell monolayer connected to the phantom layer could be peeled off the substrate. Therefore,

abovementioned results indicate a significant correlation between FUS-induced air bubbles and the cell

ablation potentially via mechanical interactions between them. This suggests that it is critical to

investigate deeper into the mechanical interactions between FUS-induced bubbles and cells.

Additionally, the formation of the PAPL in above FUS-induce tissue damage model might represent a

previously unappreciated opportunity to recapitulate the wound created by HIFU-based cell ablation.

Mechanolysis induced by subcellular bubble formation in the ablation band

Above experiments provided macroscopic level clues regarding a mechanical mechanism

underlying HIFU-induced cell ablation mediated by bubbles. To further explore the correlation between

bubble formation and cell ablation in the elliptical ablation band at the microscopic level, we observed

cellular responses during FUS treatment at single-cell resolution. Live dye was used to track cell

viability in real-time. Our results showed that, as ultrasound stimulation proceeded, a subcellular air

bubble suddenly appeared where cells are attached. As the bubble grew, the fluorescence intensity of the

cell associated with the bubble started to decrease until almost no fluorescence signal could be detected

in the whole cell (**Fig. 5a**). Quantitative analysis of the average fluorescence intensity after the

appearance of the bubble revealed a rapid decay within a few seconds, followed by a minimally

detectable plateau (**Fig. 5b**). These results suggest that as the bubble grows, the cell membrane was

broken, leading to the outflow of cellular content and eventually cell death. Statistical analysis indicated

that the fluorescence signal of the cell decreased from the level of live cells to the background level in

an average time of about 10 seconds after the appearance of the bubble, (**Fig. 5c**). Compared to classical

processes like apoptosis or necrosis, this is a remarkably rapid process, implying a different mechanism

for cell ablation through mechanical interaction between cells and FUS-induced subcellular bubbles.

To investigate the mechanism by which bubbles induce cell ablation, we first needed to

determine the relative positions of bubbles and cells. Thus, we performed optical sections and 3D

reconstruction of GFP-expressing MDCK cells during FUS treatment. Specifically, there were 10 optical

sections (X-Y planes) were acquired along the z-axis, with section spacing 11.29 μm, and the total

scanning height 101.62 μm. The lowest optical section was lower than the bottom surface of the cell

(where there is no fluorescence), and the top optical section was higher than the top surface of the cell

(where there is no fluorescence) to ensure that the 3D reconstruction included the signals across entire

thickness of the cell. 3D reconstructed images (**Fig. 5d**) showed that, before bubble formation, the green

fluorescence was uniformly distributed within the cell. As the bubble formed, a non-fluorescent circular

area appeared inside the cell (from X-Y cross-section). X-Z and Y-Z sections further revealed that the

bubble's formation created a low-fluorescence area between the cell's original position and the substrate

(coverslip), surrounded by the GFP fluorescence signal of the cell itself, reflecting the bubble might

push part of the cell away from the substrate upon its formation. This suggests that under ultrasound

treatment, subcellular air bubbles may originate between the cell and the substrate, exerting a repulsive

force on the cell.

To further confirm that abovementioned bubbles, whose appearance resulted in a low-

fluorescence region inside the cell, were indeed formed between the cell and the substrate undeath it, we

compared them with another type of bubbles that are known to be positioned above, instead of

underneath, the cell, and thus do not generate a repulsive effect. Specifically, our quantitative analysis

showed that bubbles that eventually caused cell ablation significantly reduced the fluorescence signal at

where they appear, even before the cells died (**Fig. 5e**). Continuing to track these bubbles, we observed

that, after ablating the cell at its original location, they continued to attach to the substrate, grow, and

reach the culture space above surrounding live cells. In this case, these bubbles were simply positioned

above, instead of intruding into, the cells below, and such bubbles did not significantly affect the

fluorescence signal from cells underneath (**Fig. 5e; Supplementary Fig. 7a**). The comparison of above

two situations suggests that bubbles causing cell ablation and those floating on the cell surface exhibit

significant differences in optical effects, with the latter lacking a repulsive effect on the cell. This further

supports our hypothesis that bubbles causing cell ablation originate between the cell and the substrate,

creating repulsive forces that might eventually rupture the cell.

To further substantiate above hypothesis, we investigated the relationship between subcellular

bubble formation and cell membrane deformation, we performed Z-stack imaging on MDCK cells that

only expressed membrane-GFP during FUS treatment (**Fig. 5f**). To confirm that bubbles only appeared

between the cell and the substrate, we scanned only the bottom half of the cell close to the substrate. The

scanning height was 16.60 μm, consisting of 5 layers with a spacing of 4.15 μm. Under this setting, the

uppermost optical section was at the middle of the cell, and the lowest was below the cell membrane,

thus could capture the formation and growth of the bubble along with cell membrane deformation once

the bubble formed between the cell and the substrate. Our result showed that a bubble appeared at 7.6 s

after FUS treatment, and the membrane-GFP signal enveloping the bubble was also clearly observed

(**Fig. 5f**), suggesting that the bubble indeed formed between the cell and the substrate and caused close

contact with cell membrane. As the bubble gradually expanded, the membrane-GFP signal enveloping it

also expanded, reflecting the direct mechanical repulsive effect of the bubble on the cell membrane.

Importantly, at 38 s, the membrane-GFP signal enveloping the bubble no longer exhibited a continuous

curved shape but instead showed structures resembling bifurcated cracks, followed by cell rupture. This

provided direct experimental evidence that bubble formation and growth caused the expansion and

ultimate rupture of the cell membrane (**Fig. 5f**). Additionally, we observed that some bubbles

disappeared on their own after formation, and with time, the fluorescence signal of the live cells in their

occupied area gradually recovered. This further supports the repulsive effect caused by bubble to the cell

(**Fig. 5g**).

Altogether, while we do not rule out the possibility that thermal effects by FUS may contribute

to the ablation of some cells within the focal zone, the above results suggest that the formation of

subcellular bubbles between the cell and the substrate and their repulsive effect on the cell impose

significant mechanical loads on the cell membrane, leading to continuous expansion and eventual

rupture. This process is distinct from classical mechanisms of ultrasound-induced cell ablation and

represents a previously unappreciated mechanism, which we termed ultrasound-inflicted "mechanolysis".

Furthermore, whether bubbles exert mechanical effects on the cell nucleus is another important

question. Therefore, we conducted FUS stimulation experiments on MDCK cells expressing H2B-RFP,

simultaneously using Live dye to monitor cell viability. The results showed that bubbles generally did

not form directly beneath the cell nucleus (**Supplementary Fig. 7b**), and the morphology of the cell

nuclei in the PAPL did not show significant distortions after cell ablation, suggesting less significant

effect of subcellular bubbles on cell nucleus than that on the cell membrane. In some cells, it was indeed

observed that when the formation of a bubble occurred near the cell nucleus, it could cause mechanical

compression and deformation of the cell nucleus, but not to a degree significant enough to induce

notable translocation or rupture of the cell nucleus (**Supplementary Fig. 7c**). This suggests that the

mechanical effect of bubbles on the cell membrane may be a critical factor in FUS-inflicted

mechanolysis

Therefore, based on above findings, the interactions between HIFU-induced bubbles and the cell

could be largely grouped into three different types (**Fig. 5h**). When bubbles form between the cell and

the substrate, if they may persist for a short time before disappearing, the cell could gradually refill the

void space created by the bubble. For bubbles that form and consistently grow between the substate and

the cell, they can cause significant deformation of the cell membrane, ultimately leading to

mechanolysis of the cell. Finally, as these bubbles further enlarge, they may eventually penetrate

through the PAPL and detach from the substrate.

To investigate whether this ultrasound-induced bubble generation can be achieved in thicker

tissues, we performed FUS treatment (continuous wave) to 3D extracellular matrix made of Matrigel

domes attached to the ECCD. As shown in **Supplementary Fig. 8a**, bubbles are continuously generated

and expanded in the 3D matrix during the FUS treatment. Z-stack scanning images shown in

Supplementary Fig. 8b indicated that bubbles are distributed throughout the 3D matrix, indicating that

this FUS-induced bubble phenomenon is not limited to monolayer cultured cells in vitro, but also occurs

in more physiological-like 3D matrix.

Collective cell migration and wound healing after HIFU-induced cell ablation

The wound healing response after ultrasound-induced cell ablation holds significant practical

implications, yet there is a lack of exploration in previous studies. Using only mechanical scratch or

place-holders to create empty space that cells could migrate into does not recapitulate key features of

ultrasound-inflicted wounding, such as the residue of cell/tissue debris left at the original place after

ultrasound ablation, thereby limiting their applications in understanding subsequent wound healing

response. In this study, the HIFU-induced ablation successfully created a unique type of wound, termed

the PAPL, within a continuous epithelial monolayer, providing a more biomimetic model of HIFU-

inflicted wounding. Herein we apply this model to study the collective cell migration and wound healing

after HIFU-induced ablation.

To this end, we continued to culture cells in ECCD after HIFU ablation and tracked the

migration of the cells close to the wounded PAPL area (**Fig. 6**). Under experimental conditions that

allow the PAPL to remain entirely or partially attached to the substrate after ablation, we explored how

it might regulate the collective migration of the neighboring live cells and the closure of the wound.

Using H2B-mCherry MDCK cell monolayers, we found that live cells actively and directionally

extended and migrated towards the area where the PAPL remained after ablation until the ultrasound-

inflicted wound was completely closed. During this process, there remained a connection between the

PAPL and live cells, and live cells also reconstructed the PAPL during the closure of the wound.

Additionally, under conditions where ultrasound was strong enough to cause the detachment of the

PAPL, we observed that live cells around the wound area could still migrate into the clear area left at the

original place of the PAPL. This indicates that cells are intrinsically able to migrate towards the wound

area regardless of the presence of the PAPL. Importantly, during the process of collective cell migration

and wound healing, we observed a phenomenon of "fracture" between some "leading cells" at the

migration front and the cells behind them, forming new tissue gaps. This implies a process of secondary

damage that is generated spontaneously during the repair of the primary wound inflicted by ultrasound

ablation) (**Fig. 6b**). This may be due to the different directions and speeds of movement of cells within

the cell monolayer nearby the wound. Such secondary damages were often repaired again after the

closure of the primary ultrasound-induced wound.

To further investigate the influence of the PAPL on the dynamics of collective cell migration and

wound healing response after ultrasound ablation, we selected cases where part of the PAPL remained

attached to the substrate, and tracked and analyzed the cell migration nearby. As shown in **Fig. 7a**, the

live cell monolayer in this case could migrate into both the PAPL and the free area (where the PAPL had

detached but still available for live cells to migrate into), providing a basis for analyzing and comparing

the differential guiding effect imposed by the PAPL and the free area on cell migration. As shown in **Fig.**

7b, by tracking cell migration trajectories, we found that although the initial migration direction was

relatively uniform, it was significantly influenced by the PAPL and changed direction gradually,

preferentially migrating into the area where the PAPL remained. By calculating the angle θn

(counterclockwise) between the migration direction of each cell and the horizontal direction of the

viewport, and comparing it with the orientation angle of the interface between the PAPL and the free

area (θ_1 and θ_2 in **Fig. 7b**), we found that as the wound healing proceeded, the difference between θ_n and

θ_1 or θ_2 decreased (**Fig. 7c**). In the meantime, analysis of cell movement revealed an average speed of

18-43.2 μm/h during the collective cell migration and wound healing process (**Fig. 7d**). Above results

indicates that the cell migration direction gradually turned towards the PAPL, and away from the clear

area, in a manner approximately aligned with the interface between the PAPL and the free area. These

results suggest that the PAPL, although no longer possessing live-cell activity, still exhibits biological

activity and can interact with neighboring cells to regulate their movement. This is different from the

traditional free-boundary wound healing process and indicates a previously underexplored mechanism

that might play an important role in dictating collective cell migration and wound healing after HIFU-

induced ablation. Such mechanism poses an interesting question worthy of future studies.

DISCUSSION

In this study, we designed and built an enclosed cell culture device capable of simultaneously

achieving non-invasive ultrasound stimulation and long-term online live-cell imaging. Additionally, we

established an in vitro platform to explore the biological effects of ultrasound. In comparison to previous

experimental methods, this platform offers several advantages. Firstly, compared to conventional upright

FUS-imaging systems, the design of 45°-tilt angle of incidence built into the ECCD allows a clear and

uninterrupted acoustic path, thus it can achieve controllable FUS stimulation and ablation of cells within

a pre-defined target region, while cells outside the focal area could remain live, thus providing a

technical basis for investigating changes between cells subject to different levels of ultrasonic

stimulation in the same culture environment. Secondly, this platform is compatible with long-term cell

culture and real-time microscopic imaging, wherein multiple serial excitation and continuous imaging

experiments can be performed simultaneously. Thirdly, conventional upright imaging systems expose

the culture system to open environment and significantly increase the risk of contamination of the

culture. Even for short-term ultrasound stimulation, using upright imaging system could lead to

contamination observable within 24-48 hours. By avoiding the cell contamination risks associated with

previous methods, the ECCD platform allows the study of both the passive response of cells to short-

term FUS stimulation and the continuous observation of the active post-stimulation response of cells

over an extended period. Compared to conventional systems wherein FUS treatment and real-time

imaging are carried out using decoupled, open-environment devices, which often result in discontinuous

and unstable experimental conditions, the ECCD developed in this study provides an integrative

technical platform for on-demand spatiotemporal modulation of FUS treatment and continuous live-

imaging in a single, enclosed, stable environment, thereby facilitating the study of long-term biological

effects and responses after ultrasound therapy.

In previous studies, cavitation effect, which describes the process of bubble generation,

oscillation and collapse in target tissues under ultrasonic treatment, including stable cavitation and

inertial cavitation[38], is the most well-accepted effect of FUS. Stable cavitation is generally triggered by

low intensity ultrasound, which is characterized by alternating positive and negative pressures that cause

stable volume oscillations around the equilibrium position of the bubble. This effect can put cyclic stress

and shear forces on the cell, resulting in repairable, reversible perforation[30,39,40]. When the ultrasonic

intensity exceeds a certain threshold, inertial cavitation can cause violent volume oscillation and rapid

volume increase of bubbles, and finally lead to bubble rupture and collapse, producing high temperature

and high pressure, strong shock wave[41] or microjet[42], and reactive oxygen (ROS)[43], and lead to

reversible or irreversible perforation of cell membrane. Severe cases can lead to membrane tearing and

mechanical cell lysis[44, 45]. Nevertheless, direct experimental observation of cavitation bubbles' effect on

cell ablation are still limited except for a study by Wang *et al.* [33, 41, 44, 46]. In this study, such significant

cavitation occurs at the "initiation band of bubbles" at the periphery of the focal area in the ECCD,

which leads to rapid cell death and even the breakdown of the PAPL, thereby providing further

processive understanding of classical effects of FUS.

Despite various types of FUS-related effects reported in previous studies, they were studied separately

in different systems that vary in the spatial and temporal scales, as well as in the absence of a full view

of the integration of these effects throughout the focal region of FUS. By far, a single experimental

system to study and reconcile such diverse FUS-induced phenomena remains elusive. In this work, our

ECCD platform realizes real-time visual recording of the complete focal region of FUS, which serves as

an integrative system for studying the location-specific, mechanistically distinct effects throughout the

focal region. Indeed, we have observed a variety of biological effects in the focal region, including

acoustic streaming, cavitation effect, fine packet intense cracking and thermal effect, which includes and

extends beyond classical understanding of FUS effect. This platform not only establishes a convenient

system for the study of these effects, but also provides useful information for dissecting and future

rational optimization of FUS devices. Particularly, in addition to classical phenomena associated with

FUS, our results provide direct experimental evidence for the mechanolysis caused by HIFU, a

previously unreported biomechanical mechanism of ultrasound-induced cell ablation that occurs within

a specific area of cell ablation. In our study, we found that cavitation bubbles created under HIFU can

also cause mechanolysis of cells without collapse. In this case, because the bubble does not collapse

during the process of cell ablation, but has obvious volume growth, the mechanism of stable cavitation

and inertial cavitation proposed above is obviously different from that of previous studies. Therefore, the

mechanism of mechanolysis mediated by ultrasonic cavitation bubble formation and growth identified in

this study provides a previously unrecognized mechanism that may provide theoretical guidance for the

innovative development of future FUS applications.

It should be noted that the bubbles produced by the cavitation effect in this study are largely

derived from the dissolved gas in the culture system under high energy input from continuous FUS

waves. However, under the pulse wave with the same amplitude of 10% duty cycle, the transducer is

reduced to 1/10 of the power of the continuous wave (belonging to the intensity of LIFU), and cannot

input enough energy for the focal region, so the experimental results show that there is no significant

bubble formation under the pulse FUS. Nevertheless, our findings about the mechanolysis under HIFU

in this work can also extend to other contexts of applications wherein bubbles can be elicited through

assistive mechanisms under LIFU. At present, in the field of LIFU combined with microbubbles,

injection microbubbles are mostly used to provide cavitation nuclei[47], which has the advantage that the

addition of injection microbubbles can significantly reduce the cavitation threshold (about 100 times)[48],

and reduce the additional damage caused by high-intensity stimulation to tissues (LIFU intensity is 0.1-

1kW/cm^2 [49]). Therefore, by combining the mechanolysis principle found in this work with LIFU and

exogenous bubbles, it could inspire new strategies for future ultrasound-mediated cell ablation therapy.

The experimental platform developed in this study also provides a foundation for exploring long-

term biological responses following ultrasound stimulation, such as wound healing responses after

HIFU-induced ablation. Wound healing is a significant research topic in the field of medicine. For

studying the wound healing process, in vitro models, due to their more controllable environment

compared to in vivo models, are widely adopted for investigating wound healing mechanisms. Currently,

commonly used in vitro wound healing models are based on mechanical scratches or pre-occupied void

area[50, 51] to create free spaces for cells to migrate into. In addition to scratching, various methods such as

stamping, thermal stimulation, electric shocks, and laser illumination have been employed in monolayer

cell cultures to create "wounds"[52-55]. However, these models remain difficult to recapitulate the tissue

damage inflicted by ultrasound ablation, thereby hindering the understanding of the wound healing

process post-ultrasound treatment and limiting insights into the factors regulating the collective cell

migration during such wound healing. In this study, we achieved FUS ablation in a monolayer of cells in

vitro and observed a residual layer of cell remnants (the PAPL) left at the site of the ablation, providing

a unique model and method for understanding and studying the collective cell migration and wound

healing response after ultrasound treatment. Compared to traditional models of free-edge cell migration,

our model provides a more biomimetic representation of the ultrasound-inflicted wound. The

interactions between the PAPL and neighboring live cells further demonstrate its biological activity and

effects that can guide the direction of collective cell migration, a phenomenon not reported in previous

studies. Regarding such correlation between the PAPL and collective cell migration in wound healing,

we speculate that it might be related to intercellular connections between live cells and the PAPL.

Additionally, the release of lysophosphatidic acid during cell lysis and its activation of various

downstream signals might also contribute to this phenomenon[56, 57]. The specific mechanisms behind this

phenomenon await further exploration. In the future, this wound healing model might be used to explore

the therapeutic effects of FUS-based tumor ablation and the subsequent wound healing process in

adjacent normal tissues. Based on findings reported in this study, there might be more biomechanical

regulatory mechanisms and related application to be explored regarding the post-ablation wound healing process.

In summary, the ECCD developed in this study integrates non-invasive ultrasound stimulation and long-term live-cell imaging, providing a unique technological platform for in-depth exploration of the mechanisms underlying ultrasound-induced biological effects. Based on this, we discovered mechanism of ultrasound ablation via cavitation bubble formation and growth-mediated mechanolysis of cells, adding another layer of our understanding of the biomechanical principles of ultrasound therapy. The post-ablation wound healing model feature a PAPL layer and its interaction with collective cell migration also offer new perspectives and methods for investigating tissue wound healing mechanisms. The findings of this study might contribute to broaden our understanding of the biological effects of ultrasound, and inspire advancements in the applications for FUS-based therapy.

ACKNOWLEDGEMENTS

We thank Dr. Haidong Wang for assistance in thermistor measurement. We thank Dr. Jiabin Zhang and Dr. Feng Feng for assistance in ultrasound transducer calibration. We thank Dr. Yan Yi for the gift of the ultrasound-absorbent rubber. We thank Dr. Jianbo Bai for the gift of the H2B-RFP MDCK cell line. Y.S.'s work is supported by the National Natural Science Foundation of China (U21A20203, 12102229), the National Key Research and Development Program of China (2021YFA0719301), the Oversea High-level Scholar Introduction Program, Tsinghua University Dushi Program, and the Tsinghua University Startup Funding.

AUTHOR CONTRIBUTIONS

616   Z.B. and Y.S. conceived the project; Z.B. designed and performed experiments; Z.B., Z.L., and Y.S.

617   analyzed data. Y.S. supervised the project. All authors contributed to the manuscript.

619   **COMPETING FINANCIAL INTERESTS**

620   The authors declare no competing financial interest.

REFERENCES

1. Krishna V, Sammartino F, Rezai A. A Review of the Current Therapies, Challenges, and Future Directions of Transcranial Focused Ultrasound Technology: Advances in Diagnosis and Treatment. *JAMA Neurol* **75**, 246-254 (2018).

2. Izadifar, Izadifar, Chapman, et al. An Introduction to High Intensity Focused Ultrasound: Systematic Review on Principles, Devices, and Clinical Applications. *J. Clin. Med.* **9**, 460 (2020).

3. Meng, Y., Hynynen, et al. Applications of focused ultrasound in the brain: from thermoablation to drug delivery. *Nat. Rev. Neurol.* **17**, 7-22 (2021).

4. Recker, F., Thudium, et al. Multidisciplinary management to optimize outcome of ultrasound-guided high-intensity focused ultrasound (HIFU) in patients with uterine fibroids. *Sci. Rep.* **11**, 22768 (2021).

5. Yu Tang, Chun Chen, Binglei Jiang, et al. Bifidobacterium bifidum-Mediated Specific Delivery of Nanoparticles for Tumor Therapy. *Int. J. Nanomed.* **16**, 4643-4659 (2021).

6. Lipsman, N., Meng, et al. Blood–brain barrier opening in Alzheimer's disease using MR-guided focused ultrasound. *Nat. Commun.* **9**, 2336 (2018).

7. Shashank R. Sirsi, Mark A. Borden. State-of-the-art materials for ultrasound-triggered drug delivery. *Adv. Drug Delivery Rev.* **72**, 3-14 (2014).

8. Kelsie F Timbie, Umara Afzal, Abhijit Date, et al. MR image-guided delivery of cisplatin-loaded brain-penetrating nanoparticles to invasive glioma with focused ultrasound. *J. Controlled Release* **263**, 120-131 (2017).

9. Saira Alli, Carlyn A. Figueiredo, Brian Golbourn, et al. Brainstem blood brain barrier disruption using focused ultrasound: A demonstration of feasibility and enhanced doxorubicin delivery. *J. Controlled Release* **281**, 29-41 (2018).

645   10.   Cao Y, Chen Y, Yu T, et al. Drug Release from Phase-Changeable Nanodroplets Triggered by

646   Low-Intensity Focused Ultrasound. *Theranostics* **8**, 1327-1339 (2018).

647   11.   Matos, Maria B. C. De, et al. Ultrasound-Sensitive Liposomes for Triggered Macromolecular

648   Drug Delivery: Formulation and In Vitro Characterization. *Frontiers in Pharmacology*, 10

649   (2019).

650   12.   Ying Meng et al. MR-guided focused ultrasound enhances delivery of trastuzumab to Her2-

651   positive brain metastases. *Sci. Transl. Med.* **13**, 4011 (2021).

652   13.   Stavarache, M. A., Petersen, N., et al. Safe and stable noninvasive focal gene delivery to the

653   mammalian brain following focused ultrasound. *J. Neurosurg.* **130**, 989-998 (2019).

654   14.   Nickolai Sheikov, Nathan McDannold, Natalia Vykhodtseva, et al. Cellular mechanisms of the

655   blood-brain barrier opening induced by ultrasound in presence of microbubbles. *Ultrasound in*

656   *Medicine & Biology* **30**, 979-989 (2004).

657   15.   Chen H, Konofagou EE. The Size of Blood–Brain Barrier Opening Induced by Focused

658   Ultrasound is Dictated by the Acoustic Pressure. *J. Cereb. Blood Flow Metab.* **34**, 1197-1204

659   (2014).

660   16.   Colleen T. Curley et al. Augmentation of brain tumor interstitial flow via focused ultrasound

661   promotes brain-penetrating nanoparticle dispersion and transfection. *Sci. Adv.* **6**, 1344 (2020).

662   17.   Miguel N. Centelles, Michael Wright, Po-Wah So, et al. Image-guided thermosensitive

663   liposomes for focused ultrasound drug delivery: Using NIRF-labelled lipids and topotecan to

664   visualise the effects of hyperthermia in tumours. *J. Controlled Release* **280**, 87-98 (2018).

665   18.   Paul Cressey, Maral Amrahli, Po-Wah So, et al. Image-guided thermosensitive liposomes for

666   focused ultrasound enhanced co-delivery of carboplatin and SN-38 against triple negative breast

667   cancer in mice. *Biomaterials* **271**, 120758(2021).

19.   Patricia C. Rinaldi, Joie P. Jones, Frederick Reines, et al. Modification by focused ultrasound

pulses of electrically evoked responses from an in vitro hippocampal preparation. *Brain Res.* **558**,

36-42 (1991).

20.   Po-Hsiang Tsui, Shyh-Hau Wang, Chih-Chung Huang. In vitro effects of ultrasound with

different energies on the conduction properties of neural tissue. *Ultrasonics* **43**, 560-565 (2005).

21.   David P. Darrow, Parker O'Brien, Thomas J. Richner, et al. Reversible neuroinhibition by

focused ultrasound is mediated by a thermal mechanism. *Brain Res.* **12**, 1439-1447 (2019).

22.   T. Zhang et al. Transcranial Focused Ultrasound Stimulation of Periaqueductal Gray for

Analgesia. *IEEE Trans. Biomed. Eng.* **69**, 3155-3162 (2022).

23.   Cotero, V., Graf, et al. Stimulation of the hepatoportal nerve plexus with focused ultrasound

restores glucose homoeostasis in diabetic mice, rats and swine. *Nat. Biomed. Eng.* **6**, 683–705

(2022).

24.   Yixin Zhong, Yu Zhang, Jie Xu, et al. Low-Intensity Focused Ultrasound-Responsive Phase-

Transitional Nanoparticles for Thrombolysis without Vascular Damage: A Synergistic

Nonpharmaceutical Strategy. *ACS Nano* **13**, 3387-3403 (2019).

25.   Zachs, D.P., Offutt, et al. Noninvasive ultrasound stimulation of the spleen to treat inflammatory

arthritis. *Nat. Commun.* **10**, 951 (2019).

26.   Huerta, T.S., Devarajan, et al. Targeted peripheral focused ultrasound stimulation attenuates

obesity-induced metabolic and inflammatory dysfunctions. *Sci. Rep.* **11**, 5083 (2021).

27.   Kennedy, J. High-intensity focused ultrasound in the treatment of solid tumours. *Nat. Rev.*

Cancer **5**, 321–327 (2005).

28.   Kok H P, Cressman E N K, Ceelen W, et al. Heating technology for malignant tumors: A review.

Int. J. Hyperthermia **37**, 711-741 (2020).

29.   Hynynen, K., et al. A scanned, focused, multiple transducer ultrasonic system for localized

692   hyperthermia treatments. *Int. J. Hyperthermia* **3**, 21-35 (1987).

693   30.   Junru Wu, Wesley L. Nyborg. Ultrasound, cavitation bubbles and their interaction with cells.

694   *Adv. Drug Delivery Rev.* **60**, 1103-1116 (2008).

695   31.   Hu, Zhang, Unger, et al. Focused Ultrasound-Induced Cavitation Sensitizes Cancer Cells to

696   Radiation Therapy and Hyperthermia. *Cells* **9**, 2595 (2020).

697   32.   Manabu Kinoshita, Kullervo Hynynen. A novel method for the intracellular delivery of siRNA

698   using microbubble-enhanced focused ultrasound. *Biochem. Biophys. Res. Commun.* **335**, 393-

699   399 (2005).

700   33.   Robyn K. Schlicher, Harish Radhakrishna, Timothy P. Tolentino, et al. Apkarian, Vladimir

701   Zarnitsyn, Mark R. Prausnitz, Mechanism of intracellular delivery by acoustic cavitation.

702   *Ultrasound in Medicine & Biology* **32**, 915-924 (2006).

703   34.   Qiu, Zhihai, et al. The Mechanosensitive Ion Channel Piezo1 Significantly Mediates In Vitro

704   Ultrasonic Stimulation of Neurons. *iScience* **21**, (2019).

705   35.   Yao-Shen Huang, Ching-Hsiang Fan, Ning Hsu, et al. Sonogenetic Modulation of Cellular

706   Activities Using an Engineered Auditory-Sensing Protein. *Nano Lett.* **20**, 1089-1100 (2020).

707   36.   Hideki Kitamori, Iori Sumida, Tomomi Tsujimoto, et al. Evaluation of mouthpiece fixation

708   devices for head and neck radiotherapy patients fabricated in PolyJet photopolymer by a 3D

709   printer. *Physica Medica* **58**, 90-98 (2019).

710   37.   Yuanru Zhao, Jingwen Xu, Yuanyuan Zhang, et al. Biomimetic redox-responsive prodrug

711   micelles with diselenide linkage for platinum nanozymes augmented sonodynamic/chemo

712   combined therapy of colon cancer. *Chem. Eng. J.* **472**, 144911 (2023).

38.   Xuancheng Fu, Xiaoran Hu. Ultrasound-Controlled Prodrug Activation: Emerging Strategies in

Polymer Mechanochemistry and Sonodynamic Therapy. *ACS Appl. Bio Mater.*, 2024.

39.   Hidemi Honda, Qing-Li Zhao, Takashi Kondo. Effects of dissolved gases and an echo contrast

agent on apoptosis induced by ultrasound and its mechanism via the mitochondria-caspase

pathway. *Ultrasound in Medicine & Biology* **28**, 673-682 (2002).

40.   Alexander A. Doinikov, Ayache Bouakaz. Theoretical investigation of shear stress generated by

a contrast microbubble on the cell membrane as a mechanism for sonoporation. *J. Acoust. Soc.*

Am. **128**, 11–19 (2010).

41.   Lokhandwalla, M., Sturtevant, et al. Mechanical haemolysis in shock wave lithotripsy (swl): i.

analysis of cell deformation due to swl flow-fields. *Phys. Med. Biol.* **46**, 413-413 (2001).

42.   E.A Brujan. The role of cavitation microjets in the therapeutic applications of ultrasound.

Ultrasound in Medicine & Biology **30**, 381-387 (2004).

43.   Zeyu Jiang, Wenjing Xiao, Qinrui Fu. Stimuli responsive nanosonosensitizers for sonodynamic

therapy. *J. Controlled Release* **361**, 547-567 (2023).

44.   Boris Krasovitski, Victor Frenkel, Shy Shoham et al. Intramembrane cavitation as a unifying

mechanism for ultrasound-induced bioeffects. *Proc. Natl. Acad. Sci.* **108**, 3258-3263 (2011).

45.   Liu Rengeng, Zhang Qianyu, Lang Yuehong et al. Sonodynamic therapy, a treatment developing

from photodynamic therapy. *Photodiagn. Photodyn. Ther.* **19**, 159-166 (2019).

46.   Maochen Wang, Yi Zhang, Chenliang Cai et al. Acoustic and microbubble-cell parameters

Sonoporation-induced cell membrane permeabilization and cytoskeleton disassembly at varied.

Sci. Rep. **8**, 3885 (2018).

47.   J. Deprez, G. Lajoinie, Y. Engelen et al. Opening doors with ultrasound and microbubbles:

Beating biological barriers to promote drug delivery. *Adv. Drug Delivery Rev.* **172**, 9-36 (2021).

48.   Ambre Dauba, Anthony Delalande, Hermes A. S. Kamimura et al. Recent Advances on

737   Ultrasound Contrast Agents for Blood-Brain Barrier Opening with Focused Ultrasound.

738   *Pharmaceutics* **12**,1125 (2020).

739   49.   Martijn Hoogenboom, Dylan Eikelenboom, Martijn H. den Brok et al. Mechanical High-

740   Intensity Focused Ultrasound Destruction of Soft Tissue: Working Mechanisms and Physiologic

741   Effects. *Ultrasound in Medicine & Biology* **41**, 1500-1517 (2015).

742   50.   Tao, H., Berno, et al. In vitro human keratinocyte migration rates are associated with SNPs in the

743   KRT1 interval. *PLoS One* **2**, 697 (2007).

744   51.   Nadine B. Wolf, Sarah Küchler, Michal R. Radowski, et al. Influences of opioids and

745   nanoparticles on in vitro wound healing models. *Eur. J. Pharm. Biopharm.* **73**, 34-42 (2009).

746   52.   Stamm, Anne, Reimers, et al. In vitro wound healing assays – state of the art. *BioNanoMaterials*

747   **17**, 79-87 (2016).

748   53.   Kai Safferling, Thomas Sütterlin, Kathi Westphal, et al. Wound healing revised: A novel

749   reepithelialization mechanism revealed by in vitro and in silico models. *J. Cell Biol.* **25**, 691–709

750   (2013).

751   54.   Van Kilsdonk, Jeroen WJ, et al. An in vitro wound healing model for evaluation of dermal

752   substitutes. *Wound Repair and Regeneration* **21**, 890-896 (2013).

753   55.   Ud‑Din, Sara, Ardeshir Bayat. Non‑animal models of wound healing in cutaneous repair: In

754   silico, in vitro, ex vivo, and in vivo models of wounds and scars in human skin. *Wound Repair*

755   *and Regeneration* **25**, 164-176 (2017).

756   56.   George Broughton, I. I., Jeffrey E. Janis, et al. The basic science of wound healing. *Plast.*

757   *Reconstr. Surg.* **117**, 12S-34S (2006).

758   57.   Schreml, Stephan, et al. Wound healing in the 21st century. *J. Am. Acad. Dermatol.* **63**, 866-881

759   (2010).

766
Figure 1. Principle and design of the enclosed cell culture device compatible with focused

ultrasound stimulation and long-term microscopy imaging. **(a)** Schematic diagram of the device

design and the acoustic pathway for focused ultrasound (FUS). **(b)** Structural design of the ECCD. **(c)**

Structural design of the FUS transducer coupler. **(d)** Schematic showing the assembly of the FUS

transducer and cell culture device with the acoustic pathway (schematic was drawn real component

proportions). **(e)** Design of the transducer coupler and positioning structure on the cell culture device. **(f)**

Computational results of the FUS acoustic field. Zoom-in displays the intensity field in the focal region.

The curve on the right shows the variation of intensity along the line drawn across the focal region, with

intensity represented by I_{sppa} (spatial peak pulsed average intensity, the maximum pulsed average

intensity on the acoustic field or a specified plane, in units of W/cm²).

Figure 2

ECCD   Carrier system

Coupler   Sealed to coupler

e

Materials	Sound velocity (m·s⁻¹·10³)	Density (Kg·m⁻³·10³)	Acoustic impedance (Kg·m⁻²·s⁻¹·10⁶)
Plastic (T25 flask)	2.766	1.035	2.863
Plastic (35mm dish)	3.020	1.046	3.159
Silica gel	1.003	1.049	1.052
Acrylic plate	4.023	1.189	4.783
Plastic (culture plate)	2.753	1.044	2.874
MED610 light-cured resin	2.711	1.168	3.166
PDMS (4:1)	1.203	1.030	1.239
Pure water	1.500	1.000	1.500

f

Ratio of base to curing agent	Sound velocity (m·s⁻¹·10³)	Density (Kg·m⁻³·10³)	Acoustic impedance (Kg·m⁻²·s⁻¹·10⁶)
10:1	1.058	1.034	1.094
5:1	1.123	1.031	1.157
4:1	1.203	1.030	1.239
3:1	1.115	1.029	1.148
2:1	1.059	1.025	1.085
1:1	1.052	1.024	1.077

Coupler

ECCD

Second-level displacement system

Frist-level displacement system   Microscope

Figure 2. Experimental setup for ultrasound stimulation and live-cell imaging based on the ECCD. **(a)** Flowchart showing the experimental setup and its operational procedure. **(b)** Design of the primary and secondary displacement devices. **(c)** Photos of the ECCD and its tray system. The tray system is connected to the displacement systems mentioned above. Scale bar: 10 mm. **(d)** Photos of the FUS transducer coupler. Scale bar: 10 mm. **(e)** Selected acoustically transparent materials and their measured acoustic parameters. Testing temperature: 25℃. **(f)** Table of acoustic parameters for PDMS slab formed by mixing base liquid and curing agent in different proportions. Testing temperature: 25℃. **(g)** Assembly of the displacement system, support system, ultrasound stimulation system, and cell culture system. Scale bar: 10 cm. **(h)** Photos of the assembly of the ultrasound cell stimulation experiment system and the microscope system.

Figure 3

Figure 3. HIFU-induced cell ablation does not show significant correlation with ultrasound-caused thermal effect. (a) Bright-field images of MDCK cells cultured in regular 35 mm culture dishes and ECCD at confluence of 50% and 100%, respectively. Scale bar: 100 μm. **(b)** Table showing the temperature-dependent coloration of the four-color thermochromic ink at different temperatures. The mixing ratio is red: yellow: green: purple = 4:2:2:1. **(c)** The top row shows images of the thermosensitive color-changing sticker attached to the ECCD bottom changing color at the focal point under the influence of FUS. The FUS transducer emits continuous waves at a frequency of 1MHz with indicated voltages for 3 min. The second row shows zoom-in images of the focal region. The third row shows FITC fluorescence images at the focal point after the termination of FUS treatment. Live dye is used on cells, and non-fluorescent regions reflect cell ablation areas. The fourth row shows zoom-in images boxed in the third row. Scale bar: 5 mm. **(d)** The top row shows images of the thermosensitive color-changing sticker attached to the ECCD bottom changing color at the focal point under indicated duration of FUS treatment. The FUS transducer emits continuous waves at a voltage of 40 V. The second row shows zoom-in images of the focal region. The third row shows FITC fluorescence images at the focal point after the termination of FUS treatment. Live dye is used on cells, and non-fluorescent regions reflect cell ablation areas. The fourth row shows zoom-in images boxed in the third row. Scale bar: 5 mm. **(e)** The second image from the left shows the temperature field and cell field at the focal point under FUS exposure with a transducer voltage of 30 V and an exposure time of 3 min. The arrows indicate the size of the focal region along the major and minor axis. The third image is an ellipse fitted to the focal region. **(f)** Line charts showing the lengths of the major and minor axis of the focal region in the temperature field and cell field under different ultrasound intensities (representing different excitation voltages) and different exposure times. **(g)** Scatter plots of the aspect ratio of the focal region in the temperature field and cell field under different ultrasound intensities (*left*), with a line fitted through the least squares method ($R^2=0.1753$), and scatter plots of the aspect ratio of the focal region in

817   the temperature field and cell field under different FUS treatment duration (*right*), with a line fitted

818   through the least squares method (R^2=0.1322). The cells used in the above experiments were all wild-

819   type MDCK cell lines.

820
821

Figure 4

Figure 4. Significant association between cell ablation zone and bubble generation zone under

FUS. (a) FITC fluorescence images of cells taken after FUS treatment. The FUS transducer is powered

at 40 V, emits continuous waves at a frequency of 1 MHz, and has an exposure time of 1 min. The

direction of ultrasound incidence is indicated by black arrows. Live dye is used to label living cells, and

non-fluorescent regions reflect the cell ablation zone. Scale bar: 1 mm. The red box shows the streak-

like ablation domains in the focal area. The area surrounded by the two blue ovals is elliptical ablation

region where bubbles appear. The area enclosed by the white irregular lines is the area with a

temperature greater than 62 ℃ during FUS treatment. **(b)** Bright field, FITC and RFP fluorescence

images captured near the focal region after FUS treatment. The FUS transducer is powered at 40 V,

emits continuous waves at a frequency of 1 MHz, and has an exposure time of 3 min. Bubbles are

opaque and spherical under bright field. H2B-RFP MDCK cells are used. Live dye is used on cells, and

non-fluorescent regions reflect the cell ablation zone. Scale bar: 100 μm. **(c)** Time-lapse micrographs

captured under bright field and RFP fluorescence during FUS treatment. Yellow arrows indicate bands

of bubbles. The FUS transducer is powered at 40 V, emits continuous waves at a frequency of 1 MHz.

Cells are stained with red fluorescent dead cell dye, and fluorescent regions reflect the cell ablation zone.

Scale bar: 100 μm. (The time-stamp starts from the first image shown here) **(d)** Images captured at the

edge of the focal region under bright field during FUS exposure. Yellow lines represent the edges of

stable bubble bands. Red and green arrows indicate the expansion and retreat directions of the

movement of bubble clusters. The FUS transducer is powered at 40 V, emits continuous waves at a

frequency of 1 MHz. Scale bar: 100 μm. (The time-stamp starts from the first image shown here) **(e)** The

top-left image shows the bright field microscopic image after ultrasound exposure. The FUS transducer

is powered at 40 V, emits continuous waves at a frequency of 1 MHz, and has an exposure time of 3 min.

The other three images are zoom-in views of the areas boxed in by different color squares in the top-left

image. The bubble zone is spatially correlated with the cell ablation zone, and the morphology of the

post-ablation phantom layer (PAPL) after cell ablation is significantly different from the live cell

monolayer. Scale bar: 100 μm. **(f)** The top two rows show bright field and FITC fluorescence images

after FUS exposure. The FUS transducer is powered at 40 V, emits continuous waves at a frequency of 1

MHz, and has an exposure time of 3 min. Live dye is used on cells, and non-fluorescent regions reflect

the cell ablation zone. The bottom three rows show 3D reconstructed microscopic images of the PAPL

and its adjacent live cell monolayer. Point 1 shows the PAPL that remained in its original place; Points

2-3 show PAPL rupture and peeled off the substrate by bubbles, along with some live cell monolayer

connected to it; Point 4 shows the PAPL forming a "dome" shape, with its central part detached but

peripheral part connect to the live cell monolayer on the substrate. Scale bar: 100 μm. The cells used in

Fig. 4a&c-f were wild-type MDCK cells.

Figure 5

Figure 5. Subcellular bubble formation and FUS-induced mechanolysis of the cell. (a) Time-

lapse FITC fluorescence micrographs of MDCK cells under FUS treatment. The three rows represent

images where bubbles appear in a single cell, three different cells, and multiple cells, respectively. Live

dye is used on cells, and the disappearance of the fluorescent signal reflects cell ablation. The FUS

transducer is powered at 40 V, emits continuous waves at a frequency of 1 MHz. Scale bar: 50 μm. **(b)**

Scatter plot and interpolated curve showing the average fluorescence intensity of the cell cytoplasm

(excluding bubble area) over time for the cells shown in the first row of images in **(a)**. **(c)** Scatter plot of

the time taken from bubble appearance to the average fluorescence intensity of the cell dropping to the

background fluorescence intensity. (box: 25% - 75%, bar-in-box: median, and whiskers: min and max).

$n = 3$ independent experiments; $n_{cell} = 350$. The FUS transducer is powered at 40 V, emits continuous

waves at a frequency of 1 MHz. **(d)** FITC fluorescence images showing bubbles generated at cells

during FUS exposure. The X-Y, X-Z, and Y-Z images were shown for the cell. The green area on the

image represents the cell region, and the black area appearing in the cell region represents the bubble

region. The intersection points correspond to the center points of the bubbles. Cells are MDCK cells

expressing GFP on the cell membrane. The FUS transducer is powered at 40 V, emits continuous waves

at a frequency of 1 MHz. Scale bar: 10 μm. **(e)** Analysis of fluorescence intensity at the location of the

bubble. The fluorescence intensity at the center of the bubble causing cell ablation is denoted as h_1 when

it first appears and h_2 after it caused cell ablation, and the fluorescence intensity at that position before

bubble generation is denoted as h_0. The scatter plot shows h_1/h_0 and h_2/h_0. Data were plotted as mean $\pm$

s.d. $n = 3$ independent experiments; $n_{samples} = 59$. **(f)** Bright field and FITC fluorescence micrographs

showing bubbles generated underneath the cell and eventually causing cell membrane rupture during

FUS exposure. GFP-cytomembrane MDCK cells were used. Scale bar: 10 μm. **(g)** Time-lapse FITC

fluorescence micrographs showing bubbles appearing and disappearing underneath the cell during FUS

treatment. Live dye is added to the cells. The FUS transducer is powered at 40 V, emits continuous

waves at a frequency of 1 MHz. Scale bar: 10 μm. **(h)** Schematic of different types of interactions

between HIFU-induced bubbles and cells observed in this study. The cells used in above experiments

were wild-type MDCK cells.

Figure 6

Figure 6. Observation of collective cell migration and wound healing after FUS ablation. (a, b)

The first three columns depict time-lapse bright-field and RFP fluorescence micrographs showing the

collective migration of live cells towards the PAPL or clear area after HIFU-induced ablation. In the

third column, the white area represents the PAPL (with remaining RFP-labeled cell nuclei but no

movement) or the clear area (where PAPL detached), with white arrows indicating the direction of

collective cell migration. The yellow area represents secondary tissue gaps that appear during the

primary wound healing process, largely due to the different directions and speeds of cell migration

within the live cell monolayer, with yellow arrows indicating the cell migration direction. The fourth

column shows the trajectories of individual cell nuclei during the collective migration (Colored line).

The fifth to sixth columns in **(a)** and the fifth to seventh columns in **(b)** provide zoom-in views of the

boxed regions in the fourth column. Scale bar: 100 μm. White arrows indicate the location of bubbles.

The cells used in above experiments were H2B-RFP MDCK cells

Figure 7

Figure 7. Effect of the PAPL on the collective migration and wound healing after HIFU-induced

ablation. (a) The first three rows show time-lapse bright-field and RFP fluorescence micrographs after

focused exposure, illustrating the collective migration of live cells toward the PAPL or the free area

(where the PAPL has detached). In the third row, the white area represents the PAPL or the free area,

with white arrows indicating cell migration direction. The yellow area represents secondary gaps formed

in the collectively migration cell sheet nearby the wound, largely due to different directions and speeds

of cells within the monolayer. Yellow arrows indicate the cell migration direction at the edge of the free

area. The fourth row displays the trajectories of individual cell nuclei during the wound healing process.

The fifth to sixth rows provide zoom-in views of the boxed regions in the fourth row. Scale bar: 100 μm.

(b) Trajectories of individual cells during collective cell migration. Scale bar: 100 μm. **(c)** Angles

between the migration direction of individual cells and the orientation of the interface between the

PAPL and the free area at different time points as indicated. (T_1: 0.00-0.67 h; T_2: 2.33-3.00 h; T_3: 4.67-

5.33 h; T_4: 7.00-7.67 h; T_5: 8.50-9.17; T_6: 10.17-10.83h; T_7: 11.83-12.50 h). Data were plotted as mean

± s.e.m. * $P < 0.05$, ** $P < 0.01$, *** $P < 0.001$, **** $P < 0.0001$. **(d)** Line graphs depicting the

migration speed of individual cells (left) and the average speed of the cell population (right) during

collective migration.