Development and Characterization of a low intensity vibrational system for microgravity

studies

Omor M. Khan[1], Will Gasperini[1], Chess Necessary[2], Zach Jacobs[2], Sam Perry[2], Paul Kuehl[2],

Kendall Nelson[2], Paul Gamble[2], Maximilien DeLeon[3,4,5], Sean Howard[1], Anamaria Zavala[1], Mary

C Farach-Carson[3,4,5], Elizabeth Blaber[6], Danielle Wu[3,4,5], Aykut Satici[1*], Gunes Uzer[1*]

[1]Department of Mechanical and Biomedical Engineering, Boise State University, [2]Space Tango,

Inc., 611 Winchester Lexington KY 40505, United States, [3] Department of Diagnostic and

Biomedical Sciences, UTHealth Houston School of Dentistry, [4] Department of Bioengineering,

Rice University,[5]Department of Biosciences , Rice University, [6]Department of Biomedical

Engineering, Rensselaer Polytechnic Institute

*Corresponding authors:

Aykut Satici PhD, Gunes Uzer PhD

Boise State University

Department of Mechanical & Biomedical Engineering

1910 University Drive, MS-2085

Boise, ID 83725-2085

Ph. (208) 426-4461

Email: aykutsatici@boisestate.edu, gunesuzer@boisestate.edu

Keywords: vibrational bioreactor, low intensity vibration, microgravity, scaffolds

Funding support: NSF 2025505, 1929188 and NIH AG059923.

Abstract

The advent of extended-duration human spaceflight demands a better comprehension of the physiological impacts of microgravity. One primary concern is the adverse impact on the musculoskeletal system, including muscle atrophy and bone density reduction. Ground-based microgravity simulations have provided insights, with vibrational bioreactors emerging as potential mitigators of these negative effects. Despite the potential they have, the adaptation of vibrational bioreactors for space remains unfulfilled, resulting in a significant gap in microgravity research. This paper introduces the first automated low-intensity vibrational (LIV) bioreactor designed specifically for the International Space Station (ISS) environment. Our research covers the bioreactor's design and characterization, the selection of an optimal linear guide for consistent 1-axis acceleration, a thorough analysis of its thermal and diffusion dynamics, and the pioneering use of BioMed Clear resin for enhanced scaffold design. This advancement sets the stage for more authentic space-based biological studies, vital for ensuring the safety of future space explorations.

Introduction

The 21st century has witnessed notable advancements in the field of space exploration, as both private and public sectors have made substantial efforts to expand the boundaries of human presence beyond Earth[1–9]. In order to ensure the continued viability of long-term spaceflight, it is critical to address the physiological challenges that arise from microgravity as mission duration and reach increase[2,4,9,10]. Microgravity poses a unique challenge to astronaut health, particularly affecting the musculoskeletal system. Muscle atrophy and bone density reduction are among the most pressing health issues faced during extended missions, necessitating innovative solutions to mitigate these effects[10–19].

Ground-based studies have provided valuable insights into the challenges of microgravity.

Simulated microgravity experiments, particularly those employing low- and high-intensity

vibrational bioreactors, have shown promise in mitigating some of the adverse physiological

effects of simulated microgravity[16,17,19–27]. In addition, pharmacological treatments for bone loss

resulting from exposure to microgravity are also being tackled and researched in both real and

simulated microgravity[10,11,28].

However, the full mechanism and impact of microgravity on the human body, especially in real

space conditions, are still not entirely understood[12]. Astronauts aboard the International Space

Station (ISS) undergo rigorous exercise regimens to combat the adverse effects of living in

space[10–12,17–19,28]. While these exercises help, they don't completely negate the physiological

changes induced by microgravity[12,19].

This highlights the crucial necessity of doing direct cellular studies in a microgravity setting and

evaluating the possible advantages of interventions such as vibrational bioreactors in the context

of space exploration. While vibrational bioreactors have shown promise on Earth[14–17,21–26], there

is a conspicuous absence of such technology tailored for use in space[19], especially one that meets

the stringent standards of the ISS.

Our research is motivated by the current limitations in understanding the effects of microgravity

on cellular level. The primary aim of our study is to design and characterize an automated low-

intensity vibrational (LIV) bioreactor for microgravity. This bioreactor is tailored to meet the

stringent requirements of the International Space Station (ISS). Until now, there have been no

vibrational bioreactors in space. With the introduction of our LIV bioreactor, we aim to bridge this

gap, enabling a direct comparison between experiments conducted in space and those on Earth.

This endeavor is crucial for advancing our understanding and ensuring the safety and
sustainability of future human space exploration.

Experimental Design

The objective of our research was to engineer a low-intensity vibrational (LIV) bioreactor suitable

for microgravity conditions aboard the International Space Station (ISS). Our design had to

conform to the CubeLab Interface Control Document (ICD)[29] standards, ensuring compatibility

with the ISS's operational environment. We utilized the CubeLab 9U module (**Supplementary**

Figure 2), provided by Space Tango[30], which offers a standardized platform for microgravity

experiments[2,31], allowing for a secure and controlled setting for our bioreactor.

In a precursor study within our laboratory, we developed a 3D bone analog scaffold utilizing gyroid

shapes to mimic the complex architecture of bone tissue[32]. For the current bioreactor design,

these scaffolds are integral to our assessment, serving as a platform to implement and evaluate

the bioreactor's capacity to support biological studies. The focus of our investigation is the

application of vibrational stimuli to bone cells, addressing the challenge of bone density

reduction—a significant concern for astronauts in microgravity conditions. This approach aims to

leverage the mechanical signals provided by our bioreactor to enhance bone cell function and

mitigate the effects of space-induced osteopenia.

For the delivery of precise vibrational stimuli, we selected the piezoelectric actuator (P-841.30,

Physik Instrumente (PI), Karlsruhe, Germany) and its closed-loop controller (E-610.S0, Physik

Instrumente (PI), Karlsruhe, Germany), chosen for their compliance with the CubeLab's size,

weight, and power constraints. These components were tasked with providing a consistent

vibration regime at 90Hz with a 0.7g peak to peak (1g=9.81 m/s^2) acceleration, ensuring the

application of low-intensity vibrations to the cells.

Our methodology began with the selection of an appropriate linear guide to facilitate precise one-

axis acceleration, crucial for the bioreactor's performance. This selection process was informed

by evaluating multiple linear guides against our design criteria.

Subsequently, we conducted a thermal and vibrational analysis of the bioreactor's behavior using

both COMSOL simulations and empirical testing to ensure that the applied vibrations were within

the operational parameters and did not induce too much heat and resonant frequencies that could

interfere with the bioreactor's function or cell viability.

To validate the bioreactor's compatibility with cellular studies, we performed XTT and Live/Dead

assays to confirm that the bioreactor design did not adversely affect cell viability. These assays

were critical in demonstrating the biocompatibility of our system.

Lastly, we addressed the challenge of media exchange within the CubeWells. A diffusion study[33],

supported by COMSOL simulations and visual assessments using a real-time camera setup, was

conducted to ascertain the efficiency of media introduction and distribution within the wells,

ensuring that all cells received adequate nutrients without the risk of shear stress-induced

damage.

Methods and Data Collection

A custom amplifier circuit was designed with a Sallen-Key[34] second-order lowpass filter topology

(**Supplementary Figure 1**). The filter was tuned to isolate a 90 Hz sine wave from a modulated

PWM signal, removing the extraneous frequencies. This process produced a clean sine wave to

drive the piezoelectric actuator, providing the precise vibrational stimulus required for the

biological experiments.

Linear guide selection

To select a linear guide capable of delivering precise one-axis acceleration, we conducted

performance evaluations on three different models. Using LabVIEW software, we measured the

acceleration profiles of each guide with a MEMSIC CXL04GP3 accelerometer (**Supplementary**

Figure 3). The guides were tested for their ability to sustain a 0.7g peak-to-peak acceleration at

a frequency of 90 Hz over a 20-minute duration. The objective was to identify a guide that would

restrict vibrations to the intended axis, ensuring accurate delivery of mechanical stimuli to the

samples. The tested guides (**Figure 1**) included: a custom-built roller-based guide (SS2EB,

Misumi, Schaumburg, IL), known for smooth motion; a low-maintenance plastic linear guide (N

Prism Preloaded, Igus Inc., Rumford, RI); and a Schneeberger roller-based linear guide (NKL 2-

95, Schneeberger Inc., Woburn, MA).

Thermal Characterization

The thermal performance of the LIV bioreactor was evaluated to determine its suitability for

sustaining biological experiments within the CubeLab module. The fully assembled bioreactor,

including the piezoelectric actuator, closed-loop controller, and the selected linear guide, was

installed inside the CubeLab. Thermocouples (Type K) were positioned: directly inside the

controller housing, above the actuator, adjacent to the biological sample area, and within the

external ambient environment of the CubeLab (**Figure 2**).

LabVIEW software with the NI 9211 DAQ card (National Instruments, Austin, TX, USA) was

utilized for both control of the vibrational input and data acquisition of the thermal output. The

system was tested with two different payload conditions, 400 grams and 1100 grams, to simulate

varying experimental loads. The operational test consisted of a 25-minute vibration cycle at 0.7g

peak-to-peak acceleration and a frequency of 90Hz.

Thermal characterization was conducted under two distinct CubeLab configurations: an 'Open'

setup with the top cover removed and a 'Closed' setup where the module was entirely sealed.

These tests were performed without the integration of active cooling systems to assess the

bioreactor's inherent thermal management capabilities under operational conditions. Data

collected during these tests were analyzed to understand the thermal dynamics of the system,

with particular attention to the temperature stability and heat dissipation efficiency in both 'Open'

and 'Closed' states.

Vibration Characterization

The vibrational characteristics of our bioreactor were analyzed through both simulation and

experimental testing. In the simulation phase, COMSOL was employed to study the system's

dynamics. A detailed CAD model of the bioreactor, including the base, actuator, linear guide, and

payload holder, was integrated into the simulation environment and set the material to standard

Al 6061. The setup mirrored real-world constraints: roller constraints for the linear guide, in-plane

constraints at the base's bottom, and attachment constraints at the actuator-bioreactor interface

and the connection between the payload holder and the linear guide. This simulation, executed

with a standard mesh configuration, focused on primary vibration modes within the solid

mechanics physics domain.

Concurrently, experimental tests were conducted to validate the simulation results and further

understand the bioreactor's vibrational behavior. The system was subjected to vibrations at a

targeted 0.4g peak-to-peak acceleration across a frequency range of 50 Hz to 500 Hz, with finer

increments around the critical 90 Hz range. Subsequent tests aimed for a 0.7g peak-to-peak

vibration; however, due to the actuator's constraints, only at frequencies from 80 Hz to 150 Hz.

Scaffold Viability

Cell viability on SLA-printed scaffolds made from BioMed Clear resin was assessed using XTT

assays for quantitative analysis and Live/Dead imaging for qualitative evaluation. The resin was

selected for its precision in creating complex structures. Scaffolds were integrated with a hydrogel

to improve the biomimetic quality for cell culture studies.

Cell Culture and Seeding:

Murine mesenchymal stem cells (MSCs) were employed for this study. These cells were

encapsulated within a hydrogel[35(p3)] composed of Thiolated Hyaluronic Acid (HA) (Advanced

Biomatrix, #GS22F), functionalized with specific amino sequences: PQ (GGGPQ↓IWGQGK

concentrated at 44.973 mgs/mL) and RGD (GRGDS concentrated at 73.7mgs/mL) with a volume

ratio of 4 HA:1 RGD: 1PQ. The encapsulated cells were then seeded into gyroid-shaped scaffolds,

designed to mimic the bone marrow microenvironment. These scaffolds were fabricated using the

BioMed Clear resin from Formlabs and were SLA printed using a Form 2 printer.

Osteogenic Media Preparation:

The osteogenic media used for cell culture was prepared using α-MEM, supplemented with 10%

FBS, 100 U/mL penicillin, 100 µg/mL streptomycin, Vitamin C at 50ug/ml, and beta-

glycerophosphate ranging from 10mM-20mM.

Vibration Regime:

Post-seeding, the scaffolds were transferred to a 96-well plate containing the osteogenic media.

The entire setup was then placed in an incubator. A distinct vibration regime was implemented,

wherein the samples were subjected to a consistent vibration of 0.7g peak-to-peak at 90 Hz for

20 minutes. This vibration was administered twice daily, with a 2-hour interval between each

session.

XTT Assay:

On the first and fourth day, an XTT assay was performed to assess cell viability. The assay utilized

two primary reagents: an electron mediator solution and the XTT developer reagent (Cayman

Chemical, #10010200). Following the manufacturer's protocol, the samples were incubated for

approximately 2 hours at 37°C. The absorbance of the assay wells was then measured using a

microtiter plate reader, capturing values between 400-500 nm wavelengths.

Live/Dead Imaging:

On the fifth day, live/dead imaging was conducted. The cells were stained using Calcein AM

(green) to identify live cells, Propidium Iodide (red) to mark dead cells, and Hoechst 33342 (blue)

for nuclear staining according to the manufacturer's specifications (EMD Millipore #CBA415). The

stained samples were then visualized under a fluorescence microscope. While no quantitative

data was extracted from these images, they provided a visual representation of the ratio between

live and dead cells, offering insights into cell health and viability.

Controls and Replicates:

The study employed both vibrated and non-vibrated samples as experimental and control groups

(+LIV/-LIV), respectively. Each week's samples (n=4/week for 2 weeks) were biological replicates,

with fresh samples prepared for the subsequent week following the identical protocol as the initial

week.

Diffusion dynamics within the CubeWells

To ensure consistent cell culture conditions within the custom-designed CubeWells by Space

Tango, a diffusion study was conducted. Initial tests verified the CubeWells' structural integrity

through a visual leak assessment. The primary objectives of the diffusion study[33] were to evaluate

the efficiency of a miniaturized pump in introducing new media into the CubeWell and to confirm

uniform media distribution across the well. This was achieved through both COMSOL simulations

and real-time camera observations.

<u>COMSOL Simulation:</u>

To predict the diffusion dynamics within the custom-designed CubeWell system, a multiphysics

simulation was conducted using COMSOL. The study combined two primary modules: Laminar

Flow and Transport of Diluted Species. The Laminar Flow module was employed to simulate the

fluid flow into the CubeWell, while the Transport of Diluted Species module was used to visualize

the influx of new media, set at a concentration of 1 mol/m^3. The simulation was constrained by

defining the inlet at the top and outlet at the bottom of the CubeWell (**Figure 5b**), along with the

boundary walls, ensuring a realistic representation of the media exchange process.

<u>Experimental Setup:</u>

For the in-vitro experiment, the CubeWells were pre-filled with 8 mL of water. Within this

environment, SLA-printed scaffolds made from BioMed Clear resin were introduced. These

scaffolds, measuring 5x5x5mm, were encapsulated in hydrogels (**Figure 4b**), mimicking the

cellular constructs intended for subsequent cell culture experiments.

<u>Media Exchange and Observation:</u>

To simulate the introduction of fresh media, a mixture of water and Royal Blue Icing color dye

(Wilton, #17111150) was prepared. This simulated medium was then pumped into the CubeWell

at a consistent rate of 1 mL/min. The entire process of media exchange, including the diffusion of

the dye and fluid movement within the CubeWells, was captured using a high-speed camera

(Photron FASTCAM MINI UX50) set at 1000 fps. This recording spanned a duration of 5 minutes,

providing a detailed visual account of the diffusion dynamics.

<u>Data Analysis:</u>

Post-recording, the captured footage was analyzed to quantify diffusion within the CubeWell

system. Five distinct locations within the CubeWell were selected for this analysis. At these points,

the intensity values of the dye were measured over time, offering a quantitative perspective on

the diffusion rates and patterns.

Results

<u>Linear guide selection</u>

The vibration experiment results (**Figure 1)** confirmed the Schneeberger NKL 2-95 linear guide's

superior performance, with a 95.47% and 52.60% more consistent y-axis acceleration than the N

preloaded prism carriage and the custom-built laboratory guide, respectively. The Schneeberger

guide maintained a stable acceleration profile with lower mean values (X: 0.031g, Y: 0.704g, Z:

0.018g) and standard deviations (X: ±0.000g, Y: ±0.005g, Z: ±0.006g), indicating its precise

control over vibrational input. In comparison, the N preloaded prism carriage showed higher

variability, particularly in the x-axis (0.473g±0.024) and z-axis (0.133g±0.008), while the

laboratory guide using Misumi roller guides presented intermediate stability (X: 0.089g±0.007, Y:

0.684g±0.011, Z: 0.071g±0.000). All guides met the experimental requirement of 0.7g peak to

peak acceleration in the y-axis, demonstrating their adequacy for the intended vibrational studies.

<u>Thermal profile in the CubeLab environment</u>

Thermal profile (**Figure 2**) of each component measured in the CubeLab maintained a steady

temperature in different conditions (Open and Closed), with variations within ±0.80°C over the 25-

minute LIV period. In the 'Open' CubeLab configuration, where the top cover was removed, the

temperature readings indicated a gradual increase over time in the controller and the ambient air

temperature 1 and 2 in the CubeLab. Although the ambient air temperature increased, the

temperatures for ambient 1 (23±0.72°C) and ambient 2 (24±0.60°C) still stabilized below the

critical threshold for cell viability. In the 'Closed' configuration, temperatures were consistently

higher, yet they remained within acceptable limits, suggesting that passive heat dissipation

mechanisms were sufficient to maintain a safe operational environment for the duration of the

experiments.

<u>Vibration Characterization</u>

Experimental characterization of the bioreactor's vibrational dynamics, supplemented by

COMSOL simulations, confirmed operational stability across the frequency spectrum. The

simulations indicated no resonant frequencies within the operational range, with the first mode of

vibration occurring at 929 Hz (**Figure 3**), thus minimizing the risk of resonance interference.

Empirical tests validated the simulation, demonstrating effective vibrational isolation at

operational frequencies, with X/Y and Z/Y acceleration ratios remaining below 0.13 at lower

frequencies. As frequency increased, a corresponding rise in these ratios was observed,

suggesting higher transference of vibration to non-targeted axes. **Table 1** and **2** encapsulates

these findings, showing the bioreactor's performance at various frequencies and accelerations,

with a clear delineation between low and high-frequency behavior.

<u>Scaffold viability</u>

The XTT assay results indicated no significant difference in metabolic activity between the

vibrated (+LIV) (p=0.9973) and non-vibrated (-LIV) (p>0.9999) samples from day 1 (D1) to day 4

(D4). Also, from D1 to D4, the metabolic activity of cells on the new scaffold material slightly

increased in both vibrated (D1 mean: 1.4584 to D4 mean: 1.5453) and non-vibrated (D1 mean:

1.4584 to D4 mean: 1.4986) conditions. This showed that the BioMed Clear resin used in the

scaffolds was biocompatible.

For a more direct observation of cell viability, Live/Dead imaging was employed. The imaging

results, presented in **Figure 4c-f** and (**Supplementary Figure 4**), visually demonstrate the

proportion of live (green) to dead (red) cells, providing a qualitative assessment of cell health. The

pictures show that there are mostly living cells in both +LIV and -LIV conditions. There doesn't

seem to be an increase in cell death after LIV was added or when the new material for the scaffold

was added. The comparative analysis of the live/dead ratio further supports the conclusion that

the new scaffold material, in conjunction with the LIV bioreactor, maintains a conducive

environment for cell growth and viability.

<u>Diffusion Study</u>

The diffusion characteristics within the CubeWells were initially modeled using COMSOL

simulations to establish an efficient media exchange protocol. According to the simulations

illustrated in **Figure 5b-c**, it was seen that by employing a pumping rate of 1mL/min,

approximately 80% of the volume of CubeWell would be occupied by fresh media during a span

of 3 minutes. This rate of media replacement would result in total replacement of the media by

the 5-minute mark. These findings informed the operational parameters for the media pump in

subsequent in-vitro validation experiments.

In practice, the diffusion regime for the in-vitro experiment, visualized in **Figure 5a**, involved a 5-

minute media pumping phase at the same flow rate, followed by a 15-minute recirculation period.

The high-speed camera monitoring revealed a consistent media distribution pattern (**Figure 5d-**

e), with a notable 55-second delay in reaching the center, bottom-left, and bottom-right locations

compared to the simulation. Despite this delay, both the simulation and the experimental

observations confirmed that 100% media exchange was accomplished within the 5-minute

duration. This parallel between the predicted and actual diffusion patterns validates the

simulation's utility in setting practical experimental conditions and underscores the system's

capability to maintain cell viability by preventing prolonged exposure to stagnant media.

Conclusions and Discussion

This study presented a wide-ranging characterization of a LIV bioreactor designed for microgravity
conditions, such as those on the ISS. The research focused on validating the bioreactor's
mechanical and biological performance, ensuring its compatibility with the stringent requirements
of space missions.

The selection of the linear guide was critical for the bioreactor's performance. Our results indicated
that the Schneeberger NKL 2-95 linear guide provided consistent one-axis acceleration, crucial
for minimizing disturbances during vibration. Despite its larger size, it demonstrated superior
performance in maintaining the desired vibrational profile with minimal cross-axis acceleration, as
compared to the Igus N Prism guide and the Misumi linear guide. The Misumi linear guide,
previously used in the laboratory, served as an effective control, confirming the reliability of the
Schneeberger guide under the test conditions.

Thermal characterization within the CubeLab environment demonstrated that the system could
sustain a stable temperature profile, with variations within ±0.80°C over a 25-minute operational
period. Notably, while the controller exhibited an increase in temperature, the duration of
operation and the volume of air within the CubeLab were insufficient to cause a significant rise in
the overall ambient temperature around the samples. The CubeLab's exterior design proved
effective in dissipating heat from the controller, thereby minimizing thermal transfer to the
biological samples. This indicates that the bioreactor's thermal management is adequate for the
tested operational conditions without the need for active cooling systems. Nonetheless, the design
accommodates the integration of minimal cooling solutions for experiments with cooler ambient
requirements, ensuring adaptability for a broad spectrum of research needs.

The vibrational characterization of the bioreactor was approached with a strategy that combined predictive COMSOL simulations with empirical testing. The simulations provided theoretical assurance that the bioreactor's design avoids resonant frequencies within the operational range, with the first resonant mode predicted at 929 Hz. While empirical testing did not directly confirm this resonant frequency, it did reveal an increase in the ratio of non-targeted to targeted axis accelerations at higher frequencies. This trend suggests that the bioreactor's actual vibrational isolation performance aligns with the simulation outcomes, especially since no resonance was observed during the tested frequency range.

The increase in acceleration ratios at higher frequencies highlights the need for careful monitoring of vibrational behavior as operational frequencies approach the upper limit. Still, the bioreactor met the initial goals for vibration isolation, showing that it could keep accelerations under control on the main axis and prevent transference to the non-vibrated axes. This finding is crucial for microgravity research applications, where the fidelity of mechanical stimuli delivery to biological samples is of utmost importance.

The simulation's finding that there is no resonance during routine operations, along with empirical evidence of vibration isolation, point to the bioreactor as a promising tool for spaceflight experiments. It is prepared to provide a stable environment for the study of microgravity's effects on biological systems. Future design enhancements may focus on further reducing vibrational transfers at higher frequencies, leveraging the insights gained from this study to refine the bioreactor's performance.

The evaluation of the scaffold's biological performance within the LIV bioreactor demonstrated a slight increase in metabolic activity by day four, with a marginally higher increase observed in the vibrated (+LIV) samples. This increment, although subtle, hints at a potential positive interaction between mechanical stimulation and cellular activity over time with the use of the new scaffold

material. The use of SLA-printed scaffolds across all samples ensures a consistent comparison

base, highlighting the specific effects of LIV on cell viability.

The BioMed Clear resin scaffolds, employed in both vibrated and non-vibrated conditions, have

shown promising compatibility for extended cell culture applications. This compatibility is crucial

for future microgravity studies, where material behavior under prolonged exposure to LIV can

significantly influence experimental outcomes. The data suggest that these scaffolds can support

cell growth and maintain viability, which is essential for the success of long-duration biological

experiments in space. Further research, with an extended timeline, will be instrumental in

understanding the full scope of LIV's effects on cellular systems in microgravity. The robustness

of the SLA-printed scaffolds for intricate designs also opens avenues for more complex tissue

engineering applications in space, where precise geometric control is crucial. The findings thus

far provide a strong foundation for the extended use of these materials and methods in

microgravity research.

Diffusion studies, supported by COMSOL simulations and in-vitro experiments, verified the

CubeWell system's media exchange efficiency. The congruence of simulation and experimental

outcomes validated the media pump's operational parameters, ensuring timely and uniform cell

nourishment without prolonged exposure.

In conclusion, the contributions of this research are versatile, addressing previous limitations in

space research and providing a comprehensive solution that enhances the scope of biological

experimentation in microgravity. The insights gained from this work are expected to have a

profound impact on the health and safety of astronauts by informing the design of future space

missions and the development of countermeasures against the adverse effects of living in space.

Acknowledgements

This study was supported by, AG059923, NSF1929188 and, NSF 2025505

Data Availability

The datasets used and/or analyzed during the current study are available from the corresponding

author on reasonable request.

Competing interests

The author(s) declare no competing interests financial or otherwise.

Contributions

Omor Khan: concept/design, data analysis/interpretation, manuscript writing

Will Gasperini: concept/design, data analysis/interpretation

Chess Necessary: concept/design

Zach Jacobs: concept/design

Sam Perry: concept/design

Paul Kuehl: concept/design

Maximilien DeLeon: concept/design, data analysis/interpretation, manuscript writing

Kendall Nelson: concept/design, data analysis/interpretation

Paul Gamble: concept/design, final approval of manuscript

Anamaria Zavala: data analysis, concept/design, final approval of manuscript

Sean Howard: data analysis, concept/design, final approval of manuscript

Mary Farach-Carson: data analysis, concept/design

Elizabeth Blaber: data analysis, concept/design

Danielle Wu: data analysis, concept/design

Aykut Satici: concept/design, data analysis/interpretation, financial support, manuscript writing,

final approval of manuscript

Gunes Uzer: concept/design, data analysis/interpretation, financial support, manuscript writing,

final approval of manuscript

Figure 1: Comparison of peak-to-peak accelerations for three different linear guides under

90Hz, 0.7g peak-to-peak vibration. The guides tested were an N-Prism linear guide, a

Schneeberger linear guide, and a Misumi linear guide. Box and violin plots represent the

distribution of peak-to-peak accelerations in the X, Y, and Z axes over a 20-minute test period.

The Schneeberger guide demonstrated the most consistent performance, with a peak-to-peak

acceleration of 0.7g in the Y-axis and minimal vibration translation to the X and Z axes.

Figure 2: Differential Heating Effects on the Vibrational Bioreactor and Surrounding

Environment. This figure illustrates the varying temperature profiles observed during a 25-minute

vibrational experiment conducted within a closed and open-lid system. Internal and external

thermocouples recorded the temperature of the vibrational controller (rising to 59°C) and the

ambient air within and outside the enclosure (increasing from 21°C to 25°C). The results confirm

that despite the significant heat generated by the vibrational actuator, the ambient conditions

remained within a safe range for cell viability.

Figure 3: Vibration profile characterization of the vibrational bioreactor. First mode shape

of the vibrational bioreactor observed at 929 Hz in COMSOL simulation.

Figure 4: Analysis of Stem Cell Viability and Metabolic Activity in SLA-Printed Scaffolds.

This figure presents the results of a 4-day experiment investigating the survival and metabolic

activity of stem cells grown in SLA-printed scaffolds (blue arrows). **(a)** Metabolic activity was

measured using an XTT assay at day 1 (mean = 1.4, SD = 0.23) and at day 4 (mean = 1.5, SD =

0.19), showing a slight increase over time despite high standard deviations due to a small sample

size (n = 4 per group). Analysis showed no significance between groups and days. Concurrently,

(c-f) Live/Dead staining and imaging highlighted a predominance of viable (green) cells over dead

ones, reinforcing the potential for successful cell growth within these scaffolds.

Figure 5: Diffusion study of custom well plate for cell culture. **(a)** For most efficient volume

change, a 5-minute pump followed by a 15 min. recirculation is implemented in the media change

regimen for the real experiment **(b)** COMSOL simulation depicting media concentration profile

and its **(c)** quantitative analysis. **(d)** Real experiment replicating the COMSOL simulation using

dye as new media for visualization and **(e)** intensity values of the real experiment showcasing the

concentration gradient within the 5 min. pump and 15 min. recirculation in highlighted region

Supplementary Figure 1: Custom circuit diagram (top) to generate sine wave signal from Teensy

4.0 microcontroller. PWM signal (square wave) coming from the microcontroller is filtered by the

demodulating circuit to get a sine wave output that will be used as an input signal for the LIV.

Supplementary Figure 2: CubeLab (top and bottom-left) facility that will house the vibrational

bioreactor and the biological samples. The PAUL (top and bottom-right) facility designed to house

and provide mechanical, electrical, and network interface from the outside.

Supplementary Figure 3: Experiment schematic of the LIV regimen from signal generation to

data acquisition and visualization in LabVIEW.

Supplementary Figure 4: Additional Live/Dead assay images across various scaffold locations

and samples, demonstrating a predominance of live cells over dead ones. These images further

validate the biocompatibility and cell-supportive nature of the scaffold materials used in the study.

Table 1: Acceleration ratio in each non-targeted axes to the acceleration of target direction,

highlighting the relative distribution of vibrational energy across all axes. The bioreactor

was subjected to vibration of 0.3g peak-to-peak across the entire 60 Hz to 500 Hz frequency

range.

Frequency (Hz)	X/Y	Z/Y
60	0.06	0.06
70	0.06	0.13
80	0.06	0.13
85	0.06	0.13
90	0.06	0.13
95	0.06	0.10
110	0.06	0.13
130	0.06	0.13
150	0.06	0.10
200	0.13	0.16
250	0.13	0.16
300	0.13	0.19
500	0.14	0.19

497 **Table 2: Acceleration ratio in each non-targeted axes to the acceleration of target direction,**

498 **highlighting the relative distribution of vibrational energy across all axes**. The bioreactor

499 was subjected to vibration of 0.7g peak-to-peak limited to frequencies between 80 Hz and 150 Hz

500 due to system constraints.

Frequency (Hz)	X/Y	Z/Y
80	0.06	0.09
85	0.06	0.09
90	0.06	0.09
95	0.055	0.09
110	0.07	0.09
130	0.09	0.09
150	0.06	0.10

Abbreviations: LIV means low intensity vibration, MSCs means Mesenchymal Stem Cells

References

1. Shiba D, Mizuno H, Yumoto A, et al. Development of new experimental platform 'MARS'—

Multiple Artificial-gravity Research System—to elucidate the impacts of micro/partial gravity

on mice. *Sci Rep*. 2017;7(1):10837. doi:10.1038/s41598-017-10998-4

2. Poghosyan A, Golkar A. CubeSat evolution: Analyzing CubeSat capabilities for conducting

science missions. *Prog Aerosp Sci*. 2017;88:59-83. doi:10.1016/j.paerosci.2016.11.002

3. Lumpp JE, Erb DM, Clements TS, Rexroat JT, Johnson MD. The CubeLab Standard for

improved access to the International Space Station. In: *2011 Aerospace Conference*. ;

2011:1-6. doi:10.1109/AERO.2011.5747232

4. Garzaniti N, Tekic Z, Kukolj D, Golkar A. Review of technology trends in new space

missions using a patent analytics approach. *Prog Aerosp Sci*. 2021;125:100727.

doi:10.1016/j.paerosci.2021.100727

5. Gao Z. Development and the State-of-Art Facility for Aerospace Vehicle. *Highlights Sci Eng*

Technol. 2023;38:251-257. doi:10.54097/hset.v38i.5814

6. Ficca A, Marulo F, Sollo A. An open thinking for a vision on sustainable green aviation. *Prog*

Aerosp Sci. 2023;141:100928. doi:10.1016/j.paerosci.2023.100928

7. Bozdoğan R, Kotil T, Öztürk F. Technological Development of Turkish Aerospace. In: Kacır

MF, ed. *National Technology Initiative: Social Reflections and Türkiye's Future*. 1. Türkiye

Bilimler Akademisi Yayınları; 2022:475-494. doi:10.53478/TUBA.978-625-8352-17-7.ch25

8. Adler EJ, Martins JRRA. Hydrogen-powered aircraft: Fundamental concepts, key

technologies, and environmental impacts. *Prog Aerosp Sci*. 2023;141:100922.

doi:10.1016/j.paerosci.2023.100922

9. Gibney E. European space windfall will fast-track science missions. *Nature*. Published

online November 29, 2019. doi:10.1038/d41586-019-03707-w

10. Stavnichuk M, Mikolajewicz N, Corlett T, Morris M, Komarova SV. A systematic review and

meta-analysis of bone loss in space travelers. *Npj Microgravity*. 2020;6(1):13.

doi:10.1038/s41526-020-0103-2

11. Smith SM, Heer M, Shackelford LC, et al. Bone metabolism and renal stone risk during

International Space Station missions. *Bone*. 2015;81:712-720.

doi:10.1016/j.bone.2015.10.002

12. Scott JM, Feiveson AH, English KL, et al. Effects of exercise countermeasures on

multisystem function in long duration spaceflight astronauts. *Npj Microgravity*. 2023;9(1):11.

doi:10.1038/s41526-023-00256-5

13. The impact of microgravity on bone in humans. *Bone*. 2016;87:44-56.

doi:10.1016/j.bone.2015.12.057

14. Zhang C, Li J, Zhang L, et al. Effects of mechanical vibration on proliferation and osteogenic

differentiation of human periodontal ligament stem cells. *Arch Oral Biol*. 2012;57(10):1395-

1407. doi:10.1016/j.archoralbio.2012.04.010

15. Tirkkonen L, Halonen H, Hyttinen J, et al. The effects of vibration loading on adipose stem

cell number, viability and differentiation towards bone-forming cells. *J R Soc Interface*.

Published online December 7, 2011. doi:10.1098/rsif.2011.0211

16. Shikata T, Shiraishi T, Morishita S, Takeuchi R, Saito T. Effects of Amplitude and Frequency

of Mechanical Vibration Stimulation on Cultured Osteoblasts. *J Syst Des Dyn*.

2008;2(1):382-388. doi:10.1299/jsdd.2.382

17. Liu Y, Fan Y, Chen X. Effects of whole-body vibration training in static and dynamic semi-

squat patterns on the lower limb muscle activity. *Sci Rep*. 2023;13(1):14432.

doi:10.1038/s41598-023-40985-x

18. Juhl OJ, Buettmann EG, Friedman MA, DeNapoli RC, Hoppock GA, Donahue HJ. Update

on the effects of microgravity on the musculoskeletal system. *Npj Microgravity*. 2021;7(1):1-

15. doi:10.1038/s41526-021-00158-4

19. English KL, Downs M, Goetchius E, et al. High intensity training during spaceflight: results

from the NASA Sprint Study. *Npj Microgravity*. 2020;6(1):1-9. doi:10.1038/s41526-020-

00111-x

20. Mochi F, Scatena E, Rodriguez D, Ginebra MP, Del Gaudio C. Scaffold-based bone tissue

engineering in microgravity: potential, concerns and implications. *Npj Microgravity*.

2022;8(1):1-13. doi:10.1038/s41526-022-00236-1

21. Touchstone H, Bryd R, Loisate S, et al. Recovery of stem cell proliferation by low intensity

vibration under simulated microgravity requires LINC complex. *Npj Microgravity*.

2019;5(1):1-9. doi:10.1038/s41526-019-0072-5

22. Halonen HT, Hyttinen JAK, Ihalainen TO. Mechanical impact stimulation platform tailored for

high-resolution light microscopy. *Health Technol*. 2020;10(1):87-99. doi:10.1007/s12553-

019-00382-9

23. Sharma S, Gold GE. Low-intensity Vibration Therapy for Bone Health in Renal

Osteodystrophy. *Acad Radiol*. 2017;24(11):1329-1331. doi:10.1016/j.acra.2017.08.001

24. Campsie P, Childs PG, Robertson SN, et al. Design, construction and characterisation of a

novel nanovibrational bioreactor and cultureware for osteogenesis. *Sci Rep*.

2019;9(1):12944. doi:10.1038/s41598-019-49422-4

25. Tian JQ, Wei TF, Wei YR, et al. Effect of whole body vibration therapy in the rat model of

steroid-induced osteonecrosis of the femoral head. *Front Cell Dev Biol*. 2023;11:1251634.

doi:10.3389/fcell.2023.1251634

26. Thompson M, Woods K, Newberg J, Oxford JT, Uzer G. Low-intensity vibration restores

nuclear YAP levels and acute YAP nuclear shuttling in mesenchymal stem cells subjected

to simulated microgravity. *Npj Microgravity*. 2020;6(1):1-11. doi:10.1038/s41526-020-00125-

5

27. Chaudhuri J, ed. *Bioreactors for Tissue Engineering: Principles, Design and Operation*.

Springer; 2005.

28. LeBlanc A, Matsumoto T, Jones J, et al. Bisphosphonates as a supplement to exercise to

protect bone during long-duration spaceflight. *Osteoporos Int*. 2013;24(7):2105-2114.

doi:10.1007/s00198-012-2243-z

29. CubeLab-Interface-Control_Document-Rev2-1.pdf. Accessed November 7, 2023.

https://ueberflieger.space/wp-content/uploads/2021/08/CubeLab-Interface-

Control_Document-Rev2-1.pdf

30. Space Tango [@space_tango]. Today, NASA Astronaut @Astro_Jessica successfully

installed our P.A.U.L. facility on the #ISS. This certified powered locker facility enables

complex microgravity research that requires power during all phases of flight. #OpenOrbit

https://t.co/BTGEbZwvSc. Twitter. Published December 9, 2019. Accessed November 12,

2023. https://twitter.com/space_tango/status/1204141228339675141

31. Jacobs ZA. PROVIDING A PERSISTENT SPACE PLUG-AND-PLAY AVIONICS

NETWORK ON THE INTERNATIONAL SPACE STATION.

32. Regner AM, DeLeon M, Gibbons KD, et al. Increased deformations are dispensable for cell

mechanoresponse in engineered bone analogs mimicking aging bone marrow. Published

online October 26, 2023:2023.09.24.559187. doi:10.1101/2023.09.24.559187

33. Hubka KM, Carson DD, Harrington DA, Farach-Carson MC. Perlecan domain I gradients

establish stable biomimetic heparin binding growth factor gradients for cell migration in

hydrogels. *Acta Biomater.* 2019;97:385-398. doi:10.1016/j.actbio.2019.07.040

34. Zumbahlen H. CHAPTER 8 - Analog Filters. In: Zumbahlen H, ed. *Linear Circuit Design*

Handbook. Newnes; 2008:581-679. doi:10.1016/B978-0-7506-8703-4.00008-0

35. Sablatura LK, Tellman TV, Kim A, Farach-Carson MC. Bone Marrow Endothelial Cells

Increase Prostate Cancer Cell Apoptosis in 3D Triculture Model of Reactive Stroma.

Biology. 2022;11(9):1271. doi:10.3390/biology11091271

Figure 1: Comparison of peak-to-peak accelerations for three different linear guides under 90Hz, 0.7g peak-to-peak vibration. The guides tested were an N-Prism linear guide, a Schneeberger linear guide, and a Misumi linear guide. Box and violin plots repraesent the distribution of peak-to-peak accelerations in the X, Y, and Z axes over a 20-minute test period. The Schneeberger guide demonstrated the most consistent performance, with a peak-to-peak acceleration of 0.7g in the Y-axis and minimal vibration translation to the X and Z axes.

Figure 2: Differential Heating Effects on the Vibrational Bioreactor and Surrounding Environment. This figure illustrates the varying temperature profiles observed during a 25-minute vibrational experiment conducted within a closed and open-lid system. Internal and external thermocouples recorded the temperature of the vibrational controller (rising to 59°C) and the ambient air within and outside the enclosure (increasing from 21°C to 25°C). The results confirm that despite the significant heat generated by the vibrational actuator, the ambient conditions remained within a safe range for cell viability.

Figure 3: Vibration profile of the vibrational bioreactor. First mode shape of the vibrational bioreactor observed at 929 Hz in COMSOL simulation.

Figure 4: Analysis of Stem Cell Viability and Metabolic Activity in SLA-Printed Scaffolds. This figure presents the results of a 4-day experiment investigating the survival and metabolic activity of stem cells grown in SLA-printed scaffolds (blue arrows). **(a-b)** Metabolic activity was measured using an XTT assay at day 1 (mean = 1.4, SD = 0.23) and at day 4 (mean = 1.5, SD = 0.19), showing a slight increase over time despite high standard deviations due to a small sample size (n = 4 per group). Analysis showed no significance between groups and days. Concurrently, **(c-f)** Live/Dead staining and imaging highlighted a predominance of viable (green) cells over dead ones, reinforcing the potential for successful cell growth within these scaffolds.

Figure 5: Diffusion study of custom well plate for cell culture. **(a)** For most efficient volume change, a 5-minute pump followed by a 15 min. recirculation is implemented in the media change regimen for the real experiment **(b)** COMSOL simulation depicting media concentration profile and its **(c)** quantitative analysis. **(d)** Real experiment replicating the COMSOL simulation using dye as new media for visualization and **(e)** intensity values of the real experiment showcasing the concentration gradient within the 5 min. pump and 15 min. recirculation in highlighted region

Supplementary Figures:

Development and Characterization of a low intensity vibrational system for microgravity studies

Omor M. Khan[1], Will Gasperini[1], Chess Necessary[2], Zach Jacobs[2], Sam Perry[2], Paul Kuehl[2], Kendall Nelson[2], Paul Gamble[2], Maximilien DeLeon[3,4,5], Sean Howard[1], Anamaria Zavala[1], Mary C Farach-Carson[3,4,5], Elizabeth Blaber[6], Danielle Wu[3,4,5], Aykut Satici[1*], Gunes Uzer[1*]

[1]Department of Mechanical and Biomedical Engineering, Boise State University, [2]Space Tango, Inc., 611 Winchester Lexington KY 40505, United States, [3]Department of Diagnostic and Biomedical Sciences, UTHealth Houston School of Dentistry, [4]Department of Bioengineering, Rice University,[5]Department of Biosciences , Rice University,

[6]Department of Biomedical Engineering, Rensselaer Polytechnic Institute

Supplementary Figure 1: Custom circuit diagram (top) to generate sine wave signal from Teensy 4.0 microcontroller. PWM signal (square wave) coming from the microcontroller is filtered by the demodulating circuit to get a sine wave output that will be used as an input signal for the LIV.

Supplementary Figure 2: CubeLab (top and bottom-left) facility that will house the vibrational bioreactor and the biological samples. The PAUL (top and bottom-right) facility designed to house and provide mechanical, electrical, and network interface from the outside

Supplementary Figure 3: Experiment schematic of the LIV regimen from signal generation to data acquisition and visualization in labVIEW

Supplementary Figure 4: Additional Live/Dead assay images across various scaffold locations and samples, demonstrating a predominance of live cells over dead ones. These images further validate the biocompatibility and cell-supportive nature of the scaffold materials used in the study.