Revealing 3D cancer tissue structures using holotomography and virtual hematoxylin and eosin staining via deep learning

Authors: Juyeon Park[1,2+], Su-Jin Shin[3+], Minji Kim[4+], Geon Kim[1,2], Hyungjoo Cho[5], Dongmin Ryu[5], Daewoong Ahn[5], Ji Eun Heo[6], Inyeop Jang[4], Hyun-seok Min[5], Kwang Suk Lee[7], Tae Hyun Hwang[4,8,9*] and YongKeun Park[1,2,5*]

[1]*Department of Physics, Korea Advanced Institute of Science and Technology (KAIST), Daejeon, 34141, Republic of Korea*

[2]*KAIST Institute for Health Science and Technology, KAIST, Daejeon 34141, Republic of Korea*

[3]*Department of Pathology, Gangnam Severance Hospital, Yonsei University College of Medicine, Seoul 06273, Korea*

[4]*Department of Artificial Intelligence and Informatics, Mayo Clinic, Jacksonville, Florida*

[5]*Tomocube Inc., Daejeon 34109, Republic of Korea*

[6]*Department of Urology, Yonsei University College of Medicine, Seoul 06273, Korea*

[7]*Department of Urology, Gangnam Severance Hospital, Yonsei University College of Medicine, Seoul 06273, Korea*

[8]*Department of Cancer Biology, Mayo Clinic, Jacksonville, Florida*

[9]*Department of Immunology, Mayo Clinic, Jacksonville, Florida*

[+]*Contributed equally*

[*]*yk.park@kaist.ac.kr and Hwang.TaeHyun@mayo.edu*

Abstract

In standard histopathology, hematoxylin and eosin (H&E) staining stands as a pivotal tool for cancer tissue analysis. However, this method is limited to two-dimensional (2D) analysis or requires labor-intensive preparation for three-dimensional (3D) inspection of cancer tissues. In this study, we present a method for 3D virtual H&E staining of label-free cancer tissues, employing holotomography and deep learning. Holotomography is used to measure the 3D refractive index (RI) distribution of the label-free cancer slides. A deep learning-based image-to-image translation framework is integrated into the resulting 3D RI distribution, enabling virtual H&E staining in 3D. Our method has been applied to colon cancer tissue slides with thicknesses up to 20 μm, with conventional chemical H&E staining providing a direct validation for the method. This framework not only bypasses the conventional staining process but also provides 3D structures of glands, lumens, and individual nuclei. The results demonstrate enhancement in histopathological efficiency and the extension of the standard histopathology into the 3D realm. To validate the repeatability and scalability of the approach, we applied the framework to the gastric cancer slides obtained from different institute and imaging devices.

Introduction

Histopathology is a procedure involving the microscopic examination of tissue samples to study the
structural cellular alterations in diseased tissues. Standard histopathological procedures meticulously stain
tissue slides with hematoxylin and eosin (H&E), a widely used technique for enhancing the visibility of the
cellular and tissue structures when examined under bright-field (BF) microscopes. The standard procedures,
employed for centuries, provide valuable insights into microscopic structural changes in tissues, facilitating
diagnosis, grading, and classification of diseases (Fig. 1a).

However, the standard histopathology procedure is limited to two-dimensional (2D) analysis. Although
biological tissues inherently possess three-dimensional (3D) structures, standard histopathology has
limitations in providing the 3D histopathological structures because of the challenges posed by BF
microscopy and H&E staining when dealing with thick 3D tissue samples. Recent studies have highlighted
the critical need for 3D histopathological analysis in histopathology, offering insights into 3D features such
as vascular and neural structures[1], tissue heterogeneity[2,3], and cancer grading[4,5]. However, the current
implementation of 3D histopathological analysis relies on heavy sample preparation steps, including
hundreds of sectioning and H&E staining or optical clearing.

To mitigate the resource-intensive demands of 3D histopathology, numerous efforts have been made to
eliminate staining procedures. Notably, quantitative phase imaging (QPI) techniques are one endeavor to
investigate biomedical specimens in a label-free manner, which quantitatively measures the optical phase
delay of the specimen without staining[6]. With the advantages of label-free imaging capabilities and
quantitative measurements, QPI has become an invaluable tool for investigating histopathological
specimens, including structures of diverse cancers[7-16], Alzheimer's disease[17], and Crohn's disease[18]. While
the studies predominantly focused on 2D specimens, recent advancements in 3D QPI techniques are now
being applied to various histopathological specimens[19-22].

While using label-free imaging techniques, QPI images may lack subcellular specificity. To address this
limitation, deep learning techniques have been actively integrated with QPI images to enhance subcellular
specificity[23]. Among the various deep learning techniques for interpreting QPI images, the concept of
virtual staining is gaining attention. Virtual staining involves an image-to-image translation framework,
using deep learning to convert label-free images to the desired stained images. This approach has been
applied to diverse histopathological slides and target stains. For instance, unsupervised deep learning has
successfully generated immunohistochemistry-stained images from phase images of kidney tissue slides[24].
Additionally, supervised deep learning enabled virtual staining of nuclei and F-actin using the label-free
phase images of kidney tissue slides[25]. Furthermore, Phasestain has generated H&E, John's, and Masson's
trichrome stained images from label-free phase images of skin, kidney, and liver tissue, respectively[26].
These studies have demonstrated the potential of virtual staining to make standard histopathology more
efficient by eliminating the staining procedures and replacing them with computational staining. However,
it is important to note that these research efforts have primarily focused on 2D applications.

In this study, we aim to address two key challenges–confinement in two-dimension and resource-intensive
H&E staining procedures–by employing holotomography and deep learning. Our approach involves the
elimination of traditional H&E staining procedures and the extension of H&E stained images into three-
dimensional, providing 3D H&E stained structures of colon cancer tissue slides without the need for

staining (Fig. 1b). Holotomography, a 3D QPI technique, measures the label-free 3D refractive index (RI)
distribution of the sample with high spatial resolution and optical sectioning capabilities. To realize the
holotomography, we utilized the deconvolution phase microscopy which involves the deconvolution of
intensity images using optical transfer functions to capture the 3D RI distribution of specimen without
external staining[27] (Fig. 1c). A deep learning technique for image-to-image translation is integrated with
the holotomography to generate virtual H&E stained images from label-free RI images in 3D (Fig. 1d). The
framework includes training and testing phase. During the training phase, the neural network learns to map
between the RI images and H&E stained images obtained from a standard histopathological slide. Once the
training is completed, the optimized neural network is directly applied to the 3D RI images acquired from
10 and 20 μm-thick, label-free colon cancer tissue slides, resulting in the 3D virtual H&E stained images.
94

95
96
Fig.1 | **Overview of the proposed framework a.** Procedures for standard histopathology including H&E
staining and imaging under BF microscopy. **b**. The proposed framework of 3D virtual H&E staining.
Integration of holotomography and deep learning enabled label-free 3D virtual H&E staining of thick colon
cancer tissue slides. **c**. Optical setup for holotomography and reconstruction procedures. Raw intensity
images are deconvolved with the optical transfer functions, reconstructing 3D RI images. **d**. 3D Virtual
H&E staining framework. The neural network is trained to map between RI and BF images using the

conventional thin, H&E-stained colon and tissue slides during the training phase. The trained neural
network is directly applied to the 3D RI images obtained from label-free, 10 and 20 μm-thick colon cancer
tissue slides.
106
107 **Results**
108
109 **Network training**
110
To train the network, we prepared the paired training dataset from a 4 μm-thick, H&E-stained colon cancer
tissue slide (Supplementary Fig. 1a). We first measured the slide using holotomography to obtain RI images.
It is important to note that the resulting 3D RI images were all-in-focused to ensure that the cellular details
from slightly different axial positions could be integrated into a single focal plane[28] (See Methods).
Subsequently, the same slide was imaged using a whole slide scanner (WSS) to obtain the H&E-stained BF
images. Image registration was performed between the BF images obtained from WSS and RI images
measured using holotomography based on the spatial transform network[29], and we cropped the images into
1,024 × 1,024-pixel patches, resulting in the paired dataset (See Methods). A total of 2,538 patches are
prepared and the dataset is divided into 1,996 and 542 patches for training and validation, respectively.
120
For the architecture of the neural network, we employed a conditional generative adversarial network to
train the model (Supplementary Fig. 1b). The network consists of a generator and discriminator, which are
adversarially learning to map RI images to BF images. Scalable neural architecture search (SCNAS) has
been used as the generator, an optimized architecture for 3D medical image datasets[30] (Supplementary Fig.
2a). The discriminator consists of five convolutional layers (Supplementary Fig. 2b). While the generator
creates the virtual H&E stained images from input RI images, the discriminator's role is to distinguish the
created image from the ground truth images when provided with input images.
128
To evaluate the performance of the trained networks, we compared virtual H&E images and ground truth
images. First, we obtained a wide-field, all-in-focused RI image from a different region of the same slide
used for training (Fig. 2a). The wide-field RI image revealed various histological features, including glands
(Figs. 2a(i–v)) and stroma (Fig. 2a(vi)). Subsequently, the wide-field RI image is cropped into 1024 ×
1024-pixel patches with a 50% overlap and we fed the patches into the trained network. The generated
H&E images are stitched back to reconstruct the wide-field image using the ImageJ plugin[31] (Fig. 2b). To
directly compare the virtual H&E images with ground truth images, we obtained the WSS image of the
corresponding region (Fig. 2c). The results demonstrate that the trained network successfully predicted
most of the histological and anatomical features including glands (Figs. 2b (i–v) and 2c (i–v)) and stroma
(Figs. 2b (vi) and 2c (vi)). For the quantitative assessment, the structural similarity index measure (SSIM)
values were calculated for each selected region of interest (See Methods).
140

Fig. 2 | Validations of the trained network with a 4 μm-thick, H&E-stained colon cancer tissue slide.
a, Wide-field RI image obtained from a 4 μm-thick, H&E-stained colon cancer slide and its detailed images.
b, Wide-field H&E stained images generated by the trained neural network and its detailed images. **c,**
Ground truth H&E images obtained using a WSS and its detailed images.

3D label-free RI distribution of thick colon tissue slides

To apply the trained neural network to label-free thick colon cancer tissue slides, we acquired 3D RI images
from two colon cancer slides obtained from different patients. The slides were prepared with thicknesses
of 10 and 20 μm, which are significantly thicker than the 4 μm-thick conventional histopathology slides,
without any staining. We acquired approximately 1.2 × 1.2 mm regions of 3D RI images from each slide

using holotomography (Fig 3). The wide-field RI images effectively depicted the anatomical structures,
including glands and lumens, which are commonly found in colon cancer slides (Figs. 3a and 3c). Focusing
more on the anatomical structures, we zoomed into the 3D structures of glands and lumens (Figs. 3b and
3d). Within these RI images, the lumen structures with vacant regions are distinctly clarified, as well as
visualizing the surrounding glandular structures. Notably, the 3D structures exhibit the warping of lumens
and surrounding glands along the axial axis. This dynamic presentation provides clear evidence of 3D
reconstruction of the anatomical structures.
161

162
163
164 **Fig. 3 | Label-free 3D RI images of thick colon cancer tissue slides. a, c.** Wide-field RI images were
165 obtained from the label-free 10 (a) and 20 μm-thick colon cancer tissue slides (c) using holotomography.
166 A single section of 3D RI image is presented. **b, d.** Detailed 3D RI images of glands and lumens obtained
167 from 10 (b) and 20 μm-thick colon cancer tissue slides (d).
168

3D virtual H&E images predicted by the trained network from label-free RI images

To generate the 3D H&E images from the label-free thick tissue slides, we employed the 3D RI images acquired from 10 μm and 20 μm-thick colon cancer tissue slides as input for our trained network. Specifically, we cropped each wide-field RI image from every axial position into 1024 × 1024-pixel patches with 50% overlap. These patches were used as input to the trained neural network, generating H&E images for each patch. The predicted H&E patches were stitched back to form a wide-field H&E images (Figs. 4a and 4c). By performing this prediction for all axial positions, we successfully generated 3D H&E images from the label-free thick colon cancer tissue slides.

For both slides, anatomical structures including glands and lumens are successfully predicted as H&E stained images from the label-free RI images as shown in detailed images (Figs. 4b and 4d). Lumen structures with void areas as well as the surrounding glands are clearly identified. Additionally, the necrotic regions within the gland are clarified as indicated by cyan arrows. Building on the patterns observed in the RI images, the warping of the glands and lumens along the axial axis is consistently demonstrated in the predicted H&E images.

Fig. 4 | 3D virtual H&E images predicted using trained neural network a, c. Wide-field H&E images predicted from the label-free 10 (a) and 20 μm-thick colon cancer tissue slides (c). A single section of the 3D H&E image is presented. **b, d.** Detailed images of 3D H&E images predicted from 10 (b) and 20 μm-thick colon cancer tissue slides (d). Cyan arrows indicate the necrotic structures.

Analyzing 3D histological features from the 3D virtual H&E images

To explore the detailed 3D outcomes of our trained neural network, we examined histological features in 3D (Fig. 5). In the 10 μm-thick sample, cross-sectional views revealed 3D gland and lumen structures (Fig. 5a). Void structures of lumens were continuously constructed along the z-axis, with width alterations in cross sections indicating the 3D construction of lumens. Notably, these cross sections also displayed the 3D structures of surrounding glands and the distribution of nuclei within the glands.

Focusing on the scale of individual nuclei, we zoomed into 3D H&E images and compared them with input RI images (Fig. 5b). In our direct comparison, we observed circular structures of the nucleus in the input RI images, which is consistent with the predicted H&E images of blue circular morphologies. As the z-

204 position changes, nuclei dynamically alter size and shape–contracting or expanding–indicating 3D
205 construction of the nucleus. Furthermore, we examined the relative 3D distribution among nuclei by
206 tracking the emergence and disappearance of some nuclei along the *z*-axis. This dynamic presentation
207 provides a comprehensive understanding of the 3D arrangements of nuclei within the glands.
208
209 From the 20 μm-thick tissue, we similarly observed the 3D structures of glands and lumens (Fig. 5c). The
210 cross sections depicted the 3D architecture of void structures of the lumen, exhibiting structural variations
211 of widening, shrinking, or warping along the z-axis. Additionally, the cross sections of the surrounding
212 glands revealed the 3D structures of the glands, along with the 3D distribution of nuclei within the glandular
213 structures.
214
215 A detailed comparison between the input RI images and the predicted 3D H&E images provided 3D
216 analyses of individual nuclei (Fig 5d). Specifically, as the axial position varied, some nuclei became visible,
217 others disappeared, and certain nuclei underwent changes in both shape and size. This dynamic presentation
218 unveils both 3D relative distribution among nuclei and 3D structures of individual nuclei within the glands.
219
220

221
222
223 **Fig. 5 | 3D histological features of colon cancer tissue a, c.** The *x-y*, *y-z*, and *x-z* cross sections of predicted
224 3D H&E images of 10 (a) and 20 μm-thick colon cancer tissue slides (c). **b, d.** Detailed images to visualize
225 individual nuclei in input RI images and corresponding predicted H&E images from 10 (b) and 20 μm-
226 thick tissue slides (d).

227
Validation of the virtual H&E images with the chemically stained slides
229

We further validated the capabilities of our method in the 3D histopathology of thick colon cancer tissue
slides. To achieve this, we directly compared the 3D virtual H&E images to the chemically stained images
of thick tissue slides (Fig. 6). After measuring label-free RI images of the two thick colon cancer slides, we
chemically stained the same slide and obtained images using WSS. Because the WSS images are only
available in 2D, we applied minimum intensity projection to the 3D virtual H&E images along the z-axis
to ensure a fair comparison.
236

The comparison between virtual and chemical H&E images for the 10 μm-thick tissue slide highlighted the
reliability and fidelity of the proposed virtual staining framework. The analysis of the two wide-field images
exhibited the accurate reproduction of anatomical features by the trained network (Figs. 6a and 6b). The
anatomical structures including glands and stroma are visible in the virtual H&E stained images. To provide
a detailed examination, we zoomed the glandular structures, revealing comparable nucleus distributions to
the chemically stained slides (Figs. 6a (i, ii) and 6b (i, ii)).
243

Comparing virtual and chemical H&E for the 20 μm-thick tissue slides demonstrated the robustness of the
virtual staining framework, even for sample thickness where the chemical staining becomes abnormal. The
colorimetric characteristics of virtual H&E remain consistent with minimal deviation, while chemical
staining resulted in a distinctly different color distribution. Nevertheless, we could compare the anatomical
features between virtual and chemical H&E images (Figs. 6c and 6d). The structures of glands and stroma
were successfully reproduced by the trained network. For a detailed investigation of glandular structures,
we zoomed into the selected glands, clearly illustrating the consistency of lumens and surrounding glandular
structures (Figs. 6c (i, ii) and 6d (i, ii)). All regions indicated with arrows and boxes are confirmed by the
experienced pathologist (SJS).
253

a Wide-field virtual H&E images - 10 μm-thick tissue slide   **b** Wide-field chemical H&E images - 10 μm-thick tissue slide

c Wide-field virtual H&E images - 20 μm-thick tissue slide   **d** Wide-field chemical H&E images - 20 μm-thick tissue slide

Fig. 6 | **Validations using the standard histopathology procedures a, c,** Minimum intensity projection of predicted 3D H&E images from label-free 10 (a) and 20 μm-thick colon cancer tissue slide (c). **b, d,** Same 10 (b) and 20 μm-thick tissue slides (d) were stained with H&E and imaged using WSS. Detailed images of selected glandular structures are presented (i, ii).

Inter-organ, inter-institutional, and inter-device validation of 3D virtual H&E staining

To ensure the scalability and repeatability of our framework, it is crucial to demonstrate the direct application of our virtual staining technique across diverse sample origins, institutional settings, and imaging devices. To achieve this, we conducted virtual staining procedures on gastric cancer slides obtained from the Mayo Clinic. The procedures for generating the training dataset remained consistent with those used for the colon cancer dataset, differing only in the number of slides and imaging devices used (See Methods). This resulted in a dataset consisting of 9,231 patches, with 7,002 patches allocated for training and 2,231 patches for validation. Following training, SSIM values were computed for each selected window to quantitatively evaluate the network's performance (Supplementary Fig. 3). Additionally, we utilized identical architecture and hyperparameters of the neural network as those employed for training on colon cancer data, ensuring consistency and robustness of our neural network architecture.

Following training, we prepared a 20 μm-thick gastric cancer tissue slide without any staining. Then, we acquired 3D RI images spanning approximately 1.92 × 1.22 mm regions using holotomography (Fig. 7a). These wide-field RI images clearly depicted anatomical structures such as smooth muscle and blood vessels, with vascular structures displaying individual cells within circular surrounding tissue structures, alongside wavy textures in the smooth muscle (Fig. 7b).

Applying the same cropping and stitching procedures used for colon cancer data, we inputted the 3D RI

images into our trained network to generate virtually stained 3D H&E images of the gastric cancer slides

(Fig. 7c). Notably, the network effectively predicted most anatomical structures, including smooth muscle

and vascular structures, as evidenced in wide-field images. Detailed examination revealed a successful

depiction of vascular structures as red blood cells with surrounding circular connective tissue structures and

wavy textures of smooth muscle (Fig. 7d). Leveraging the 3D nature of the data, we observed changes in

the cross-sectional morphology of vascular structures along the axial axis, as well as alterations in the

distribution of red blood cells within it. Moreover, variations in the wavy texture of smooth muscle along

the axial axis were visualized, highlighting the ability of virtual 3D H&E staining to reveal intricate 3D

tissue textures.

To further elucidate structural features in 3D, we visualized cross sections of detailed images depicting

vascular and muscular structures (Figs. 7e and 7f). Notably, these cross sections delineated continuous

structural changes of vessels and muscles, exhibiting curved propagation along the axial axis. Additionally,

void structures are clearly distinguished from the tissue regions, enhancing the clarity of tissue structures

in three dimensions.

Fig. 7 | Validations of the virtual staining across different organ type, institute, and imaging device a.
A wide-field RI image was obtained from the label-free 20 μm-thick gastric cancer tissue slide using holotomography. A single section of 3D RI image is presented. **b.** Detailed 3D RI images of muscular and vascular structures obtained from the 20 μm-thick gastric cancer tissue slide. **c.** A wide-field H&E image predicted from the label-free 20 μm-thick gastric cancer tissue slide. A single section of the 3D H&E image is presented. **d.** Detailed images of 3D H&E images predicted from the 20 μm-thick gastric cancer tissue slide. **e, f.** The *x-y*, *y-z*, and *x-z* cross sections of predicted 3D H&E images of the 20 μm-thick gastric cancer tissue slide.

Discussion

The proposed virtual H&E staining framework effectively addresses two limitations–labor-intensive
sample staining procedures and 2D limited analysis–simultaneously. By utilizing a single routinely
generated histopathological slide, we trained the neural network that translates RI images into H&E stained
images. This trained network, when applied to 3D RI images of label-free thick tissue, successfully
generated 3D H&E images. This approach not only eliminates the need for staining procedures and
significantly reduces the number of slicing processes but also provides comprehensive 3D glandular
structures for thick colon cancer tissues.

Moreover, the versatility of this method is noteworthy, as it transcends specific organ types. Its versatility
stems from the fact that sample preparation, imaging, and network training do not involve organ-specific
steps, allowing scalability to tissue slides from various organs. Similarly, our approach can extend to
cytology slides as well, which also exhibit 3D structures but have conventionally been studied in a 2D
context using BF microscopy.

Likewise, our framework demonstrates remarkable repeatability across different imaging devices. These
results highlight the robustness of the proposed method, encompassing both holotomography and neural
network architecture. This consistency emphasizes the potential for widespread adoption and utilization in
diverse research and clinical settings.

While our proposed virtual staining method has been successfully applied to tissues with a thickness of up
to 20 μm, the applicable thickness can be pushed further through technical advances of holotomography.
Currently, holotomography faces challenges related to image degradation as the sample becomes thicker.
The degradation in image quality can be attributed to the multiple scattering of the light within the thick
sample, which hinders the accurate reconstruction of RI values through holotomography. Nonetheless,
several strategies are being employed to address this challenge, including enhancements to optical
conditions, the implementation of reconstruction algorithms tailored for thicker samples[32], and the
application of aberration correction algorithms[33]. With these strategies aimed at improving image quality
in holotomography, we believe in a synergistic growth between virtual staining and holotomography.

Moreover, holotomography quantifies RI values, determining the protein concentration in specific regions.
This empowers the proposed framework to conduct further quantitative analysis of subcellular structures.
Specifically, segmentation or classification of cell types within the generated 3D H&E stained images can
be performed using open-source, pre-trained resources[34], with results directly applicable to the RI images
for calculating 3D morphological and biophysical parameters. With these capabilities, we believe that the
proposed framework can effectively overcome current limitations in histopathology, and broaden its utility
across various histopathological studies, aiding in diagnosis and precision medicine.

Methods

Colon cancer slides preparation from Gangnam Severance Hospital

To create the training datasets, we utilized a 4 μm-thick colon cancer tissue slide prepared as tumor
microarrays. The slide was newly diagnosed and reviewed by a pathologist (SJS) at Gangnam Severance
Hospital in January 2023. Among 26 cases, a colon cancer tissue slide with grade two was randomly

selected. From the chosen case, a formalin-fixed paraffin-embedded tissue block was obtained. The tissue
block was sliced to a 4 μm thickness using a microtome. The slice was carefully placed on a slide glass and
deparaffinized using xylene. Subsequently, the slide was then stained with H&E, mounted with a mounting
medium, which has an RI value of 1.495 measured by a refractometer (R-5000, Atago), and finally covered
with coverslips. Annotations of glands and stroma in the H&E stained images were carried out by an
experienced pathologist (SJS), specializing in uropathology. This study was approved by the institutional
review board (IRB) of Gangnam Severance Hospital (IRB No. 3-2022-0083) and conducted in accordance
with the Declaration of Helsinki.
360
361 For the label-free thick tissue slides used in the testing phase, we randomly selected two cases of colon
362 cancer. Formalin-fixed paraffin-embedded blocks from the selected cases were sliced into 10 and 20 μm-
363 thick sections using a microtome, respectively. The subsequent procedures were consistent with those of
364 training datasets, except that we did not stain the slides and mounted the slides immediately after the
365 deparaffinization.
366
367 **Gastric cancer slides preparation from Mayo Clinic**
368
369 For the training datasets, we obtained four gastric cancer tissue slides, each with a 5 μm thickness and H&E
370 stain. The slide was newly diagnosed and reviewed by pathologists at the Mayo Clinic in January 2023.
371 The slide preparation procedures were consistent with those conducted at Gangnam Severance Hospital.
372 This study was approved by the institutional review board (IRB) of Mayo Clinic (IRB No. 3-2022-0083)
373 and was conducted in accordance with the Declaration of Helsinki.
374
375 For testing purposes, a label-free, 20 μm-thick tissue slide was randomly selected from the gastric cancer
376 cases. The slide was prepared using the same procedures as those performed at Gangnam Severance
377 Hospital.
378
379 **Image acquisition**
380
To measure the RI distributions of tissue slides, we employed a low-coherence holotomography setup (HT-
X1, Tomocube). This system measures transmitted intensity images of a sample using four optimized
Köhler illumination patterns[27], from which the 3D RI distribution was reconstructed by deconvolution of
the intensity images with theoretically estimated point spread functions. In the experimental setup, a light-
emitting diode (LED) with a center wavelength of 449 nm within a digital micromirror device (DMD)
module (DLP4500, Texas Instrument) was used as the illumination source. The blue wavelength was
chosen to avoid overlap with the peak absorption spectrum of H&E staining. The illumination intensity was
controlled using a DMD located in the Fourier plane, which was then projected onto the sample using a
condenser lens (f = 180 mm, numerical aperture (NA) = 0.75). The diffracted light from the specimen was
collected using an objective lens (numerical aperture (NA) = 0.95) and measured using a CMOS camera
(FS-U3-28S5, FLIR) located in the image plane. The RI images offer a lateral and axial resolution of 155
nm and 1.07 μm, respectively. In addition to RI images, this system offers single-channel BF (scBF) images,
which are single-intensity images taken under uniform circular pattern illumination. Wide-field RI images
were obtained by stitching the tiled RI images of each field of view with the size of 227 × 227 μm. For the

stitching, the raw intensity images were captured with an overlap of 23 μm for both RI and scBF images.
To obtain the ground truth images, the H&E stained slides were scanned using the whole slide scanner
manufactured by 3D HISTECH for colon cancer slides and by Leica for gastric cancer slides.
398
399   **All-in-focus**
400
We applied an all-in-focus algorithm only to the RI images obtained from a slide used for network training[28].
This algorithm ensures that the single focal plane includes most of the cellular structures. As demonstrated
in the previous study, we used a vignette size of 100 pixels with a step size of 10 pixels. Normalized variance
was calculated in each window for all axial sections, and the window from the section with the maximum
normalized variance value was selected as the best focal plane. Repeating the procedures for the whole 3D
RI images resulted in a single all-in-focused RI image.
407
408   **Registration**
409
Image registration between the RI and WSS images involved three steps. First, we manually cropped the
regions from the WSS images where they were imaged using holotomography. Subsequently, we applied a
spatial transform network[29] to register the manually cropped WSS images with the scBF images acquired
using holotomography. This step resulted in paired images of scBF and WSS at a wide-field scale. Then,
we cropped the paired datasets into $1{,}024 \times 1{,}024$-pixel patches and repeated the registration using a spatial
transform network between these patches to ensure precise alignment at each patch scale. Note that aligning
the WSS images with scBF images directly leads to the alignment with RI images, as scBF and RI images
are captured at the same position.
418
419   All the registration steps, except for the manual cropping, were conducted using spatial transform
420   networks[29]. Specifically, EfficientNet[35] was trained to discover the optimal affine transform matrix for the
421   input WSS images, which will be used to transform the WSS images to be aligned with the target scBF
422   images. The training was achieved by minimizing the loss defined using Pearson's correlation coefficient
423   (PCC):
424

$$\mathcal{L}_{pcc}(x,y) = 1 - \frac{\mathbb{E}\big[(x - \mu_x)(y - \mu_y)\big]}{\sigma_x \sigma_y},$$

426
427   where x and y refer to the input WSS and target scBF images, respectively. μ and σ denote the mean and
428   the standard deviation of each image, respectively. Once the training was completed, we employed the
429   trained network to predict the optimized affine transform matrix, which was then applied to register the
430   WSS images.
431
432   **Training details**
433
434   A conditional GAN was used to train the network for generating 3D H&E images, primarily based on a
435   previous study[36]. In this framework, the generator's objective is to transfer the RI images into the feasible
436   H&E stained BF images by minimizing the $\mathcal{L}_{pcc}$ between the images generated by the network and their

corresponding ground truth images. Simultaneously, the discriminator aims to distinguish the images from the network and genuine images when provided with the corresponding input images. The final objective of this framework becomes

$$G^* = \arg \min_G \max_D \mathcal{L}_{cGAN}(G, D) + \lambda \mathcal{L}_{pcc}(G(x), y),$$

where $\mathcal{L}_{cGAN}$ is defined as:

$$\mathcal{L}_{cGAN}(G, D) = \mathbb{E}[\log D(x, y)] + \mathbb{E}[\log(1 - D(x, G(x)))],$$

where $G(\cdot)$ and $D(\cdot)$ refer to the generator and discriminator network operators, respectively. x denotes an input image and y indicates a ground truth image.

For the architecture of the generator, we used SCNAS[30], an optimized network specifically tailored for 3D medical image segmentation (Supplementary Fig. 2a). SCNAS employs a stochastic sampling algorithm within a gradient-based bi-level optimization framework to simultaneously search for the optimal network parameters at multiple levels using generic 3D medical imaging datasets. This search resulted in the discovery of a U-Net-like encoder-decoder structure with skip connections. In each micro-level architecture, we added a squeeze excitation block[37]. The key network parameters were as follows: activation function, leaky ReLU; normalization function, instance normalization; the size of the initial feature map, 12; the number of layers, 8; feature map multiplier, 2.

The discriminator consists of five convolutional layers with a kernel size of 4 and a stride of 2 (Supplementary Fig. 2b). The first four layers use leaky ReLU with a slope of 0.2, while the last layer employs ReLU as an activation function. Following the last layer, the result is activated by a Sigmoid function. The first and last layers use batch normalization, while the rest of the layers use instance normalization.

Both the generator and discriminator were trained using an adaptive moment estimation optimizer (ADAM) to update the learnable parameters[38]. Image augmentation techniques, such as adding blur, adjusting brightness, and introducing Gaussian noise, were randomly added to the training patches. A learning rate of 1×10^{-4} was maintained throughout the training process. Training was conducted for 40 epochs with a batch size of 1, and early stopping was employed to prevent overfitting. The network was implemented using Python version 3.8.0 and PyTorch[39] version 1.13.1, and the procedures were executed on NVIDIA GeForce 4090 GPU.

The training results are compared with the ground truth images using SSIM, which is defined as below:

$$SSIM(x, y) = \frac{1}{3} \sum_{i=1,2,3} \frac{(2\mu_{x,i}\mu_{y,i} + C_1)(2\sigma_{xy,i} + C_2)}{(\mu_{x,i}^2 + \mu_{y,i}^2 + C_1)(\sigma_{x,i}^2 + \sigma_{y,i}^2 + C_2)},$$

where x and y are the network-generated output and corresponding ground truth image, respectively. μ and σ refer to the mean and the standard deviation of each image, respectively, where an index i refers to the

479 RGB channels of each image. $\sigma_{xy,i}$ represents the covariance of both images calculated from i-th image
480 channels. C_1 and C_2 are regularization constants, set as 6.5025 and 58.5225, respectively, by default. The
481 SSIM values were calculated for the selected windows. To make the spatial frequency range identical
482 between the images, we performed low-pass filtering with the same kernel size for both images before
483 calculating SSIM.
484
485 **Supplementary information**
486

487
488
Supplementary Fig. 1| Dataset preparation and conditional generative adversarial network. a,
Workflow of dataset preparation for the training phase. A conventional 4 μm-thick, H&E stained colon
cancer tissue slides are imaged under holotomography and a whole slide scanner. Then, the images are
registered using a spatial transform network[29], resulting in a paired dataset for supervised learning. **b,**
Overview of the conditional generative adversarial network[36]. The generator creates H&E stained images
from input RI images, while the discriminator is trained to distinguish between the generated images and
ground truth images.
496

497
498
499 **Supplementary Fig. 2| Network architecture. a,** Architecture of the SCNAS used for generator[30]. The
500 structure includes eight encoding micro-level architectures, a bridge, and another decoding eight micro-
501 level architectures, resulting in a U-Net-like structure. **b,** Architecture of the discriminator. The structure is
502 benchmarked from the previous work[36], which includes five sequential convolutional layers.
503

504
505
506 **Supplementary Fig. 3| Validations of the trained network with a 5 μm-thick, H&E-stained gastric**
507 **cancer tissue slide. a,** Wide-field RI image obtained from a 5 μm-thick, H&E-stained gastric cancer slide
508 and its detailed images. **b,** Wide-field H&E stained images generated by the trained neural network and its
509 detailed images. **c,** Ground truth H&E images obtained using a WSS and its detailed images.
510
511 **Acknowledgement**
512

This work was supported by the National Research Foundation of Korea (2015R1A3A2066550,
2022M3H4A1A02074314, RS-2023-00241278), an Institute of Information & Communications
Technology Planning & Evaluation (IITP; 2021-0-00745) grant funded by the Korea government (MSIT),
a KAIST Institute of Technology Value Creation, Industry Liaison Center (G-CORE Project) grant
funded by MSIT (N10240002), the Korea Health Technology R&D Project through the Korea Health
Industry Development Institute (KHIDI), funded by the Ministry of Health & Welfare, Korea
(HI21C0977, HR22C1605).

Competing interests

D.R., D.A., H.C., H.-s.M. and Y.K.P. have financial interests in Tomocube, a company that commercializes
holotomography and quantitative phase imaging instruments. All other authors declare no competing
interests.

References

1   Kiemen, A. L. *et al.* CODA: quantitative 3D reconstruction of large tissues at cellular
resolution. *Nature Methods* **19**, 1490-1499 (2022).
2   Olson, E., Levene, M. J. & Torres, R. Multiphoton microscopy with clearing for three
dimensional histology of kidney biopsies. *Biomed. Opt. Express* **7**, 3089-3096,
doi:10.1364/BOE.7.003089 (2016).
3   Paul, D., Cowan, A. E., Ge, S. & Pachter, J. S. Novel 3D analysis of Claudin-5 reveals
significant endothelial heterogeneity among CNS microvessels. *Microvascular Research*
86, 1-10, doi:https://doi.org/10.1016/j.mvr.2012.12.001 (2013).
4   Glaser, A. K. *et al.* Light-sheet microscopy for slide-free non-destructive pathology of large
clinical specimens. *Nature Biomedical Engineering* **1**, 0084, doi:10.1038/s41551-017-
0084 (2017).
5   Liu, J. T. C. *et al.* Harnessing non-destructive 3D pathology. *Nature Biomedical
Engineering* **5**, 203-218, doi:10.1038/s41551-020-00681-x (2021).
6   Park, Y., Depeursinge, C. & Popescu, G. Quantitative phase imaging in biomedicine.
Nature Photonics **12**, 578-589, doi:10.1038/s41566-018-0253-x (2018).
7   Park, J. *et al.* Quantification of structural heterogeneity in H&E stained clear cell renal cell
carcinoma using refractive index tomography. *Biomed. Opt. Express* **14**, 1071-1081,
doi:10.1364/BOE.484092 (2023).
8   Zhang, J. K., He, Y. R., Sobh, N. & Popescu, G. Label-free colorectal cancer screening
using deep learning and spatial light interference microscopy (SLIM). *APL Photonics* **5**,
doi:10.1063/5.0004723 (2020).
9   Majeed, H., Nguyen, T. H., Kandel, M. E., Kajdacsy-Balla, A. & Popescu, G. Label-free
quantitative evaluation of breast tissue using Spatial Light Interference Microscopy (SLIM).
Scientific Reports **8**, 6875, doi:10.1038/s41598-018-25261-7 (2018).
10   Gladstein, S. *et al.* Correlating colorectal cancer risk with field carcinogenesis progression
using partial wave spectroscopic microscopy. *Cancer Medicine* **7**, 2109-2120,
doi:https://doi.org/10.1002/cam4.1357 (2018).
11   Takabayashi, M., Majeed, H., Kajdacsy-Balla, A. & Popescu, G. Disorder strength
measured by quantitative phase imaging as intrinsic cancer marker in fixed tissue biopsies.
PLOS ONE **13**, e0194320, doi:10.1371/journal.pone.0194320 (2018).
12   Sridharan, S., Macias, V., Tangella, K., Kajdacsy-Balla, A. & Popescu, G. Prediction of
Prostate Cancer Recurrence Using Quantitative Phase Imaging. *Scientific Reports* **5**,
9976, doi:10.1038/srep09976 (2015).
13   Uttam, S. *et al.* Early Prediction of Cancer Progression by Depth-Resolved Nanoscale
Mapping of Nuclear Architecture from Unstained Tissue Specimens. *Cancer Research* **75**,
4718-4727, doi:10.1158/0008-5472.Can-15-1274 (2015).
14   Wang, P. *et al.* Nanoscale nuclear architecture for cancer diagnosis beyond pathology via
spatial-domain low-coherence quantitative phase microscopy. *Journal of Biomedical
Optics* **15**, 066028 (2010).

Wang, P., Bista, R., Bhargava, R., Brand, R. E. & Liu, Y. Spatial-domain low-coherence quantitative phase microscopy for cancer diagnosis. *Opt. Lett.* **35**, 2840-2842, doi:10.1364/OL.35.002840 (2010).

Subramanian, H. *et al.* Nanoscale Cellular Changes in Field Carcinogenesis Detected by Partial Wave Spectroscopy. *Cancer Research* **69**, 5357-5363, doi:10.1158/0008-5472.Can-08-3895 (2009).

Lee, M. *et al.* Label-free optical quantification of structural alterations in Alzheimer's disease. *Scientific Reports* **6**, 31034, doi:10.1038/srep31034 (2016).

Bokemeyer, A. *et al.* Quantitative Phase Imaging Using Digital Holographic Microscopy Reliably Assesses Morphology and Reflects Elastic Properties of Fibrotic Intestinal Tissue. *Scientific Reports* **9**, 19388, doi:10.1038/s41598-019-56045-2 (2019).

Lee, A. J. *et al.* Volumetric Refractive Index Measurement and Quantitative Density Analysis of Mouse Brain Tissue with Sub-Micrometer Spatial Resolution. *Advanced Photonics Research* **4**, 2300112, doi:https://doi.org/10.1002/adpr.202300112 (2023).

Hugonnet, H. *et al.* Multiscale label-free volumetric holographic histopathology of thick-tissue slides with subcellular resolution. *Advanced Photonics* **3**, 026004 (2021).

Filan, C., Song, H., Platt, M. & Robles, F. Analysis of structural effects of sickle cell disease on brain vasculature of mice using three-dimensional quantitative phase imaging. *Journal of Biomedical Optics* **28**, 096501 (2023).

Abraham, T. M. *et al.* Label- and slide-free tissue histology using 3D epi-mode quantitative phase imaging and virtual hematoxylin and eosin staining. *Optica* **10**, 1605-1618, doi:10.1364/OPTICA.502859 (2023).

Park, J. *et al.* Artificial intelligence-enabled quantitative phase imaging methods for life sciences. *Nature Methods*, doi:10.1038/s41592-023-02041-4 (2023).

Wang, R. *et al.* Virtual brightfield and fluorescence staining for Fourier ptychography via unsupervised deep learning. *Opt. Lett.* **45**, 5405-5408 (2020).

Guo, S.-M. *et al.* Revealing architectural order with quantitative label-free imaging and deep learning. *elife* **9**, e55502 (2020).

Rivenson, Y. *et al.* PhaseStain: the digital staining of label-free quantitative phase microscopy images using deep learning. *Light: Science & Applications* **8**, 23 (2019).

Hugonnet, H., Lee, M. & Park, Y. Optimizing illumination in three-dimensional deconvolution microscopy for accurate refractive index tomography. *Opt. Express* **29**, 6293-6301, doi:10.1364/OE.412510 (2021).

Liang, M. *et al.* All-in-focus fine needle aspiration biopsy imaging based on Fourier ptychographic microscopy. *Journal of Pathology Informatics* **13**, 100119, doi:https://doi.org/10.1016/j.jpi.2022.100119 (2022).

Jaderberg, M., Simonyan, K. & Zisserman, A. Spatial transformer networks. *Advances in neural information processing systems* **28** (2015).

Kim, S. *et al.* in *Medical Image Computing and Computer Assisted Intervention–MICCAI 2019: 22nd International Conference, Shenzhen, China, October 13–17, 2019, Proceedings, Part III 22.* 220-228 (Springer).

Preibisch, S., Saalfeld, S. & Tomancak, P. Globally optimal stitching of tiled 3D microscopic image acquisitions. *Bioinformatics* **25**, 1463-1465, doi:10.1093/bioinformatics/btp184 (2009).

Lee, M., Hugonnet, H. & Park, Y. Inverse problem solver for multiple light scattering using modified Born series. *Optica* **9**, 177-182, doi:10.1364/OPTICA.446511 (2022).

Chen, M., Phillips, Z. F. & Waller, L. Quantitative differential phase contrast (DPC) microscopy with computational aberration correction. *Opt. Express* **26**, 32888-32899, doi:10.1364/OE.26.032888 (2018).

Graham, S. *et al.* Hover-net: Simultaneous segmentation and classification of nuclei in multi-tissue histology images. *Medical image analysis* **58**, 101563 (2019).

35   Tan, M. & Le, Q. in *International conference on machine learning.* 6105-6114 (PMLR).
36   Isola, P., Zhu, J.-Y., Zhou, T. & Efros, A. A. in *Proceedings of the IEEE conference on*
computer vision and pattern recognition. 1125-1134.
37   Hu, J., Shen, L. & Sun, G. in *Proceedings of the IEEE conference on computer vision and*
pattern recognition. 7132-7141.
38   Kingma, D. P. & Ba, J. Adam: A method for stochastic optimization. *arXiv preprint*
arXiv:1412.6980 (2014).
39   Paszke, A. *et al.* Pytorch: An imperative style, high-performance deep learning library.
Advances in neural information processing systems **32** (2019).
628