behaviorMate: An *Intranet* of Things Approach for Adaptable Control of Behavioral and Navigation-Based Experiments

John C. Bowler[1,3,4+,†], George Zakka[1,3+,†], Hyun Choong Yong[1,3], Wenke Li[6], Bovey Rao[1,2,3], Zhenrui Liao[1,3], James B. Priestley[5], Attila Losonczy[1,3,†]

[1] Department of Neuroscience
[2] Doctoral Program in Neurobiology and Behavior
[3] Mortimer B. Zuckerman Mind Brain Behavior Institute,
Columbia University, New York, NY 10027 USA
[4] Department of Neurobiology University of Utah, Salt Lake City, UT 84112, USA
[5] Brain Mind Institute, École polytechnique fédérale de Lausanne
[6] Aquabyte, San Francisco, CA 94111
[+] indicates equal contribution

[†] Correspondence should be addressed to:
A.L.: al2856@columbia.edu
J.C.B.: jack.bowler@utah.edu
G.Z.: gz2333@columbia.edu

1 Abstract

Investigators conducting behavioral experiments often need precise control over the timing of the delivery of stimuli to subjects and to collect the precise times of the subsequent behavioral responses. Furthermore, investigators want fine-tuned control over how various multi-modal cues are presented. behaviorMate takes an "Intranet of Things" approach, using a networked system of hardware and software components for achieving these goals. The system outputs a file with integrated timestamp-event pairs that investigators can then format and process using their own analysis pipelines. We present an overview of the electronic components and GUI application that make up behaviorMate as well as mechanical designs for compatible experimental rigs to provide the reader with the ability to set up their own system. A wide variety of paradigms are supported, including goal-oriented learning, random foraging, and context switching. We demonstrate behaviorMate's utility and reliability with a range of use cases from several published studies and benchmark tests. Finally, we present experimental validation demonstrating different modalities of hippocampal place field studies. Both treadmill with burlap belt and virtual reality with running wheel paradigms were performed to confirm the efficacy and flexibility of the approach. Previous solutions rely on proprietary systems that may have large upfront costs or present frameworks that require customized software to be developed. behaviorMate uses open-source software and a flexible configuration system to mitigate both concerns. behaviorMate has a proven record for head-fixed imaging experiments and could be easily adopted for task control in a variety of experimental situations.

2 Introduction

In this work, we present *behaviorMate*, an open-source system of modular components that communicate with each other over a local area network (LAN), orchestrated by a custom Java application. We validate our approach through benchmarking and experimental results, demonstrating the ability of behaviorMate to control head-fixed navigational experiments both in a self-powered treadmill as well as in a fully visual virtual reality (VR) setup with a running wheel. Depending on the nature of the investigation, this solution permits researchers to select the configuration of cues they wish to

present at run-time, resulting in a system that is capable of adapting quickly to incorporate changes to experimental protocols. This design makes it easy to adapt to shifting demands of research on the same behavioral setups while producing a consistent, easy-to-parse output file describing the experiment.

Proprietary systems such as National Instrument's Data Acquisition Systems (DAQ) and their accompanying software controller, LabView, typically have large upfront and licensing costs for each experimental rig their system is deployed on. behaviorMate exclusively uses open-source software and is simple to construct with parts that are significantly cheaper. The customized electronics may be ordered as printed circuit boards (PCB) and assembled by hand or purchased preassembled. The latter approach saves assembly time while adding significant expense; however, this can be mitigated by bulk ordering. The total cost of one of our VR rigs including 5 computer monitors, Android computers to render the visual scene, a frame for holding the monitors, and a light-weight running wheel was approximately $3700. The total cost for a belt-based treadmill system is around $1700. For two-photon imaging, additional components may be needed, such as a light-tight and laser resistant enclosure, which which were not factored into the calculation. While other open-source and modular behavioral control systems have been developed (Akam et al. 2022), behaviorMate does not require the user to write software code and uses compiled Arduino programs that do not incur the overhead of code interpreters to maximize performance.

Many experiments rely on using head-restrained animals to study the neuronal circuits underlying complex behaviors; such immobilization poses a fundamental challenge to behavioral research as it stymies animals' natural movement through their environments. In order to address this issue, a plethora of "treadmill" systems have been developed relying on belt-based systems (Royer et al. 2012, Lovett-Barron et al. 2014, Jordan et al. 2021), VR (Dombeck et al. 2010), or augmented reality (AR), where experimenters move "spatial" cues to create the illusion of locomotion without actual movement in physical space (Jayakumar et al. 2019). Belt-based systems involve placing an animal on a physical belt treadmill where, as the animal runs, physical cues affixed to the belt provide spatial information. VR- and AR-systems operate under closed loop control, where the environment responds to the animal's behavior (Dombeck & Reiser 2012). Normally, these cues are visually displayed on one or more computer monitors, but olfactory and auditory stimuli

have also been utilized at regular virtual distance intervals to enrich the experience with more salient cues. Furthermore, belt-based and VR/AR-based elements have been combined to enrich experiences and enable more complex behavioral tasks. At times, traditional "open-loop" stimulus may be required, such as timed cue presentations, allowing pre- and post- event neuronal activity to be examined. Increasingly, studies combine open- and closed-loop behavioral paradigms to identify stimulus-response profiles or behavioral state driven changes to stimulus responses. The proliferation of various experimental paradigms demands a high level of flexibility in experimental setups, so that methods can be integrated within single experiments or across lines of research to interrogate neuronal function.

Many of the spatial representations observed in freely moving animals are conserved in head-restrained animal setups (Aronov & Tank 2014, Dombeck et al. 2010). Observations within the medial temporal lobe, focused on all subregions of the hippocampus as well as the entorhinal cortex, have shown that prominent neural correlates of navigation behaviors exist in both the treadmill belt style systems as well as in the visual only VR systems (Aronov & Tank 2014, Dombeck et al. 2010, Royer et al. 2012, Danielson et al. 2016). In CA1 (Danielson et al. 2017, Zaremba et al. 2017) and CA3 (Terada et al. 2022), place cells, goal-related tuning, and other feature-selective tuning have been observed. Additionally, in visual VR systems, grid cell-like tuning has been observed in the medial entorhinal cortex (MEC), as was distance tuning (Aronov & Tank 2014, Heys et al. 2014). The increased control over reward distribution and cue manipulation that VR systems afford has been pivotal for recent findings relating to how cells in CA1 form place fields, uncovering novel mechanisms underlying synaptic plasticity (Gonzalez et al. 2023, Bittner et al. 2017, Priestley et al. 2022), demonstrating the power of these systems to test specific theoretical predictions.

3 Overview

behaviorMate is an integrated system composed of several sub-components that communicate on LAN using JSON-formatted packets transmitted via a standard Datagram packet protocol (UDP)

(Fig. 1). This approach allows for modularity and encapsulation of different concerns. We dub the 84

resulting system as an "*Intranet* of Things" approach, mimicking the Internet of Things systems 85

commonly found in modern "smart homes," which communicate exclusively on a dedicated LAN 86

using statically assigned internet protocol (IP) addresses. This allows for a central user interface 87

(UI) or "real-time" controller to seamlessly send and receive messages from various accessory 88

downstream devices that have dedicated simple functions. Importantly, the techniques we describe 89

are possible implementations of the behaviorMate system. In our experimental setups, we use 90

behaviorMate to combine *in vivo* head-fixed mouse navigation experiments with two-photon (2p) 91

microscopy, focusing on one-dimensional (1D) spatial navigation tasks (Fig. 3E). Due to its modular 92

design, however, behaviorMate can be easily reconfigured to provide closed-loop or open-loop 93

control during a variety of behavioral paradigms (and has been used for non-navigation based 94

studies; such as the head-fixed trace fear conditioning task described in Ahmed et al. has been 95

implemented on this system). 96

In addition to running a PC-based UI, we combine Arduino open-source micro controllers with 98

custom designed printed circuit boards (PCBs) to run the experiments (Fig. 1, 2D, E *also see* Sup- 99

plementary Materials). The circuits include Ethernet adapters allowing the Arduinos to communicate 100

on the network as well as connectors that interface them to sensors or actuators for controlling 101

the experiments or reporting the behavioral state back to the UI. We outline a general-purpose 102

input/output (GPIO) circuit that can connect to a variety of sensors and reward valves generally 103

using a transistor-transistor logic (TTL) pulse/square wave as well as I^2C and serial peripheral 104

interface (SPI) protocols (Fig. 2D), a position tracking circuit (Fig. 2E) that is dedicated mainly to 105

reading updates of the animal's location from a rotary encoder, and a VR controller setup which can 106

be used to provide a more immersive visual display to animals during the behaviors. Importantly 107

these individual components can be swapped out or added to with increased experimental demands. 108

For additional integration of software components or testing, the UI can be configured to send 109

messages to the PC's `localhost` address (i.e. 127.0.0.1 on most PCs). To speed up development, 110

we created virtual implementations of the electronics in Python code to test UI changes before 111

testing them on a physical system. Most modern programming languages, including Python, have 112

built-in support for JSON processing and UDP networking making component and new feature 113
testing easy with the help of utility scripts that can be written in users' preferred environments (*see* 114
Supplementary Materials for references to example Python code). 115

Lastly, since our main focus is on 1D navigational tasks for two-photon imaging, we provide two 117
designs for physical rigs, which hold animals fixed in place under a microscope. However, they 118
could also be adapted to any other methods for *in-vivo* recording. One system is a self-powered, 119
belt-based treadmill system, wherein animals are placed on a fabric belt that they pull beneath 120
them to advance their virtual position. Locations and contexts can be distinguished in this setup 121
by having various fabrics and physical cues. Alternatively, the other system provides navigational 122
contexts through a series of computer displays for a purely visual experience, but allowing for 123
additional flexibility of within-task manipulations to context and cues. We describe the details of both 124
of these systems and, additionally, note that with the following methods, it is possible to combine 125
aspects from both belt-based and VR systems to meet novel experimental designs. To clarify, for 126
the remainder of the paper, "treadmill" will refer to these belt-based systems while "VR" will refer to 127
the visual-based system in which animals moving a plastic running wheel. 128

4   The User Interface 130

The central hub for running experiments is the behaviorMate user interface (UI) (Fig. 2A). The UI 131
collects input from various sensors, assembling information about virtual position and behavior state 132
data asynchronously. The UI performs three main functions: (1) accepting user input and displaying 133
the current state of experiments to ensure proper functioning, (2) writing hardware information in a 134
time-stamped streaming text file which can be processed later for analysis of the experiment, and 135
(3) serving as the central communications hub for sending and receiving UDP-formatted messages 136
to the downstream components. The system is event-driven such that the hardware components 137
send messages to the UI when the state changes rather than the UI continuously polling. There 138
are currently two versions of the UI. The first is a long-term support (LTS) release that evolved 139

alongside the experimental systems and has been used for data collection for numerous published 140
studies; the second is a streamlined beta version that has been rewritten to take advantage of 141
the JavaFX visual development toolbox. Java was chosen for the UI because it is cross-platform 142
in that it can be run on Windows, Linux, and Mac, and supports standard networking protocols. 143
For consistency, interchangeability, and space optimization, we run the UI on an Intel(TM) NUC 144
mini PC with a 2.4GHz base-frequency (up to 4.2GHz based on processor load) 4-core i5-1135G7 145
processor, 16GB DDR4-3200MHz RAM, and a 512GB NVMe M.2 SSD hard drive running Window 146
10 Professional Edition, although many other configurations are supported. Notably, due to the sys- 147
tem's modular architecture, different UIs could be implemented in any programming language and 148
swapped in without impacting the rest of the system. The only requirement for interchangeability of 149
the UI is to implement sending and receiving of the same JSON formatted UDP messages packets 150
to the downstream hardware components. 151

The UI executes in three distinct phases during run-time. First a setup routine is executed, then 153
the program enters the main event loop, and finally a termination routine is executed when an 154
experimental trial is completed (Figure 2B). The main event loop begins executing as soon as the 155
UI is loaded, but additional steps for checking and logging the contexts' state are added during the 156
execution of a trial. Devices still receive input and deliver output to the UI even when a trial is not 157
currently running, allowing experimenters to confirm that everything is working before starting the 158
trial. Experiments using behaviorMate are configured by user-generated JSON files with key-value 159
pairs specified in the documentation. These settings files can be manually generated in a text editor 160
or programmatically generated using a script (*see also* Supplementary Materials). 161

The main structure of the application logic is an event loop that iterates over enabled contexts. 163
This will define presentation schemes for both simple stimuli such as an LED lighting up and water 164
reward delivery, as well as more complex stimuli like flashing LEDs, groups of stimuli, or even visual 165
VR scenes. A context can be applied globally or to one or several locations; a list of locations and 166
rules governing an individual context's activation is referred to as a Context List. When a Context 167
List is active, the devices assigned to it will be triggered. For example, a user could configure an 168

LED which blinks at several designated locations along the track. Context Lists are implemented in the UI using an Object-Oriented approach that streamlines implementing logic for governing the presentation and control of novel cue types. The main rules governing how a Context List works are implemented in a `check` method which is passed the current state of the experiment as input arguments and returns a Boolean variable as `true` or `false` corresponding to whether a Context List should be active or not. In this way novel Context List types may be added to the program easily through inheritance of a base class and overriding this one method. Additionally, Context Lists are instantiated using the "Factory Method" pattern (Gamma et al. 1994) and appended to a dynamic list. This pattern allows for adding complex functionality while minimizing the impact on performance and increasing code reuse and modularity. Existing Context List types can be modified by applying "decorators" (Gamma et al. 1994), resulting in *composable* behavioral paradigms, meaning that novel experiments can be implemented based on wrapping Context Lists with on or several of the existing decorators, without requiring any software modification. Settings files following JSON syntax specify the configuration as well as the context lists and decorators. For example, a particular decorator can be configured to cause a context to only be active on even laps while a second only permits the contexts to be active after a certain amount of time has passed. Composing these two rules means it is possible to trigger the contexts on even laps, only after 5 have passed since starting the experiment.

5 Behavior Tracking & Control

5.1 GPIO/Behavior Controller

The primary GPIO circuit we implement, referred to as the Behavior Controller, is composed of an Arduino attached to an integrated circuit (Fig. 2D). The custom PCB handles voltage regulation and provides the necessary connectors to operate downstream devices. The Arduino program distributed with behaviorMate controls all connected devices and communication between them and the UI. The program wraps "action" messages with "routes" (Fig. 2C). This pattern of routing messages simplifies debugging since it is clear which classes of the Arduino code received the

messages and where replies were generated. The messages are JSON formatted text, so they are human-readable and can be simulated within the behaviorMate PC for testing and debugging purposes. The PCB has various connectors and other components that makes it easier to connect sensors and actuators to the Arduino and provides an Ethernet adapter for the Arduino to permit LAN communication. The Behavior Controller has several important functions; it receives signals from the UI to activate rewards and cues such as odor valves, LEDs, and tone generators and passes input from sensors to the PC. In our setup for 2p imaging, the Behavior Controller also sends a synchronizing signal to the microscope to initiate imaging after a trial is started through the behaviorMate UI. The timestamp of this signal is used to align information about the animals' behavior with recorded microscopy data. By default, this signal is a TTL pulse (3.3 V in our implementation, Fig. 2D *also see* Supplementary Materials) that can be configured to activate other types of recording devices such as video cameras or electrophysiology interfaces.

5.2 Position Controller

The position controller's function is to detect animal movement and report it to behaviorMate. It is composed of an Arduino Nano attached to a custom circuit which has an onboard quadrature decoder, 16 bit counter, and connectors for attaching the rotary encoder and lap reset circuit. For either treadmill or running wheel based setups (i.e. for a VR setup), a quadrature rotary encoder device is coupled to the shaft of a wheel that turns as the animal runs. The turning of the rotary encoder generates necessary signals, or "ticks", which are passed to the quadrature decoder to calculate the instantaneous velocity of the animals. These quadrature ticks are decoded and counted in the counter. The custom circuit (Fig. 2E) uses a 10MHz oscillator which is capable of counting high resolution encoders (>4096 ticks/turn) at greater than 1K RPM. The onboard counter enables the tick counting to be decoupled from the speed of the on-board Arduino. Consequently, the Arduino can check the counter at a much slower rate (100Hz in our setup, but can be as slow as 5Hz) without losing count of the ticks. The Arduino is only responsible for checking the counter and generating a JSON formatted text string to transmit to the behavior PC via Ethernet, and only transmits when there is movement. Significantly, this eliminates polling from the behavior PC and

allows the communication between the Arduino and the PC to be one way instead of bidirectional. 224
The behavior PC then translates the counts into a meaningful measurement of the linear distance 225
traveled by the animal on the belt or running wheel. 226

6   Behavioral Apparatus 228

6.1   Treadmill system 229

Self-powered treadmill systems have been extensively used in the study of navigational behaviors 230
(Royer et al. 2012, Tuncdemir et al. 2022, Lovett-Barron et al. 2014, Grienberger & Magee 2022, 231
Geiller et al. 2017). The following section outlines a protocol for the construction of a behaviorMate 232
compatible mouse treadmill setup that has been used for head-fixed imaging (Kaufman et al. 233
2020a, Terada et al. 2022, Rolotti et al. 2022b). In combination with the software and circuits de- 234
scribed above, the treadmill system provides a flexible platform for implementing novel experimental 235
paradigms. 236

6.1.1   Frame 238

The treadmill system requires a frame to support it and maintain wheel alignment, so we provide 239
designs for a frame made of extruded aluminum rails (Fig. 3A). The rails support wheels on either 240
end for a fabric "belt" that can be moved by a mouse. In addition to the wheels, we added two 241
wheel-rollers that support the belt near the animal's fixed position (Fig. 3E). This design does not 242
have a platform directly beneath the mouse, allowing the belt to absorb vibrations, which reduces 243
motion artifacts and allows for a stable field of view during *in vivo* imaging experiments. Treadmill 244
belts are constructed from 2-inch wide fabric ribbon and may be embellished by linking multiple 245
fabrics or adding various physical cues such as foam balls, sequins, or hook-and-loop fasteners for 246
texture (Fig. 3F). These cues help the animal to orient itself and are useful for information encoding. 247
In this setup, the animal is head-fixed beneath a microscope, but is able to move the belt and their 248
attached physical cues which mimics navigation while still allowing investigators to perform imaging 249

of cellular and even sub-cellular regions of interest. To further optimize the treadmill for imaging, 250

a two goniometer device placed underneath the belt is used to adjust the angle of the animal's 251

head along the roll (side-to-side) and pitch (forward and backward) axes. This setup allows the 252

investigator to level the field of view, ensuring the neuronal plane of interest is perpendicular to 253

the optical axis of the microscope, thus increasing the number of neurons that can be imaged and 254

improves image clarity. Finally, the frame has a sliding mechanism to adjust the distance between 255

the wheels and accommodate various sized belts to ensure proper tension. Proper belt tension 256

ensures the animal can move easily and reduces the chance of belt slippage along the wheels. 257

6.1.2 Electronics and Peripherals 259

Interfacing between the treadmill and behaviorMate requires the previously described electronics. 260

Any number of optional peripheral devices such as LEDs, odor release systems, etc., may be added. 261

For position tracking, a rotary encoder is attached to one of the two wheel shafts and connected 262

to the position tracking circuit to track the animal's position. The behaviorMate UI handles the 263

conversion of the rotary encoder's spinning to the linear distance traveled. It also maintains the 264

animal's current lap number. An additional "lap-reset" feature prevents drift from accumulating 265

between the digital tracking and physical treadmill belt. We provide designs for an optical lap-reset 266

circuit for determining when a complete revolution of the belt has been made (Fig. 3A). Setting up 267

a lap-reset circuit additionally permits a calibration function that can be temporarily enabled prior to 268

running an experiment to adjust the scaling between rotary encoder "ticks" and distance moved 269

along the treadmill belt. 270

Generally position tracking drift is small (Fig. 3C) but can occur (especially with more "slippery 272

fabrics" or if belt tension is set too low). It is important to accurately specify the track length and 273

calibrate the position scale factor to tightly couple the PC and the electronics. If the system is not 274

set up properly or the belt is not sufficiently tight, the belt may slip, leading to inaccurate position 275

readings. Therefore, we provide a benchmark of the position tracking performance under normal 276

conditions. To compute the linear distance traveled, a small LED was attached to the top of the belt 277

and the entire treadmill system was moved to a completely dark room with an IR camera pointed 278
directly at the belt. The belt was moved by hand and upon each complete revolution, the LED 279
would pass within the field of view of the camera. OpenCV (Bradski 2000) was used to extract 280
the lap completion times from the recorded video. These were taken to be the "ground truth" lap 281
completion times. The quantities of interest were the time and position differences between the 282
LED and those tracked by the behaviorMate UI. The mean difference between lap times was 0.102 283
seconds (Fig. 3B, D), while the mean difference between position upon lap completion was 13 mm 284
and did not exceed 23 mm (Fig. 3C). Given that the mean and max differences were a fraction of 285
the average size of a subject mouse and significantly smaller than the typical size of reward zones, 286
these metrics were deemed more than acceptable. Enabling the lap-reset feature on self-powered 287
treadmill prevents accumulation of these errors, ensuring that position, as tracked by the UI, is 288
aligned with the fabric belt and cue presentations are aligned with the positions tracked by the UI. 289

The UI synchronizes any number of additional sensory cues such as LEDs, tones, and odor 291
release systems, which can be present at pre-defined times or locations along the belt (Fig. 3G). 292
For example, a liquid reward can be coupled to a specific location on the treadmill belt that is 293
either *operant* (delivered only if the animal licks within a "reward zone") or *non-operant* (delivered 294
immediately when the animal reaches a "reward zone"). These "reward zones" are defined by a 295
location and radius, so there is user-specified tolerance for the areas animals can receive operant 296
rewards. The rules governing the rewards are fully customizable and specified in the settings file 297
for the UI. 298

6.1.3 Use in Past Studies   300

The physical design of the treadmill was inspired by Royer et al. and influenced by several other 301
hippocampus studies focusing on behavioral navigation tasks (Kaifosh et al. 2013, Lovett-Barron 302
et al. 2014, Zaremba et al. 2017). Our design is flexible and has been validated by the myriad 303
experimental paradigms that have been successfully implemented in other labs as well (Tuncdemir 304
et al. 2023, Vancura et al. 2023, O'Hare et al. 2022, Rolotti et al. 2022a, Tuncdemir et al. 2022, 305

Rolotti et al. 2022b, Terada et al. 2022, Geiller et al. 2022, Blockus et al. 2021, Grosmark et al. 2021, Dudok et al. 2021a,b, Geiller et al. 2020, Kaufman et al. 2020*a*, Turi et al. 2019, Zaremba et al. 2017, Danielson et al. 2017, 2016). Multi-modal cues can be incorporated across several days of experiments to probe different aspects of navigation and associative learning (Fig. 3G). For example, the treadmill has been used to test goal-oriented learning, where goal locations are "hidden" and animals must report their understanding of the correct location by licking at the reward spout in order to trigger water delivery (Zaremba et al. 2017, Danielson et al. 2016). During these tasks, novel cues can be presented by introducing new treadmill belts as well as novel non-spatial cues such as background tones, blinking LEDs, or odor stimuli. Moving the reward location forces animals to learn updates to task rules mid-experiment and has been used to investigate deficits in mouse models of psychiatric disorder (Zaremba et al. 2017) as well as neural circuit analysis of the mechanisms behind reward learning (Kaufman et al. 2020*a*) and the development of signals underpinning spatial navigation (Rolotti et al. 2022b, Terada et al. 2022, Danielson et al. 2016). Cues can be tied to time, location, or both. Switching between a time-dependent presentation and a location-dependent one provides a powerful tool for assessing how the same cue may be processed differently based on the current context, location, or task (Terada et al. 2022, Tuncdemir et al. 2022). In addition to cue presentation, the setup has also been used to trigger optogenetic stimulation (Kaufman et al. 2020*b*, Geiller et al. 2022, O'Hare et al. 2022, Rolotti et al. 2022b); this coupling of navigational cues provided by the treadmill belt with stimulation in hippocampal region CA1 was used for the first successful optical induction of place fields (Rolotti et al. 2022b).

6.2 VR System

The use of Virtual Reality (VR) cues can add significant flexibility to behavioral studies and has thus become increasingly popular in neuroscience. Although different VR implementations have been described, for this manuscript, VR is defined as closed-loop control from the behaviorMate UI that integrates position updates virtually and advances cues past the animal to create the illusion of movement, unlike a belt which physically moves past the animal. Various systems for accomplishing this have been proposed and implemented previously (Dombeck et al. 2010, Harvey et al. 2012,

Sheffield et al. 2017, Hainmueller & Bartos 2018, Campbell et al. 2018, Arriaga & Han 2019, 2017, ₃₃₄
Dudok et al. 2021). Mice can learn to use VR cues as landmarks for traversing environments to ₃₃₅
specific locations. Neuronal circuits thought to be involved in navigation in freely moving mice ₃₃₆
have been shown to be similarly engaged during the described VR tasks (Dombeck & Reiser ₃₃₇
2012, Dombeck et al. 2010, Aronov & Tank 2014). behaviorMate represents a novel contribution ₃₃₈
in that it incorporates visual VR displays into its architecture for 1D navigation and is designed to ₃₃₉
adapt seamlessly to novel arrangements of VR displays with minimal computational load for each ₃₄₀
additional display (Priestley et al. 2022, Bowler & Losonczy 2023). This is achieved by rendering ₃₄₁
the virtual environments on separate android devices (one per display) with a separate application ₃₄₂
(VRMate). Since the same behaviorMate UI is used for treadmill and VR experiments, all of the ₃₄₃
experimental paradigms described in the treadmill section can be implemented in the VR system ₃₄₄
with minimal changes to the settings file. Additionally, VR permits rapid visual context switches ₃₄₅
and the addition of novel virtual objects in a familiar scene. Furthermore, "hybrid" setups are also ₃₄₆
supported in which visual displays can be added to physical treadmills in order to combine both ₃₄₇
tactile and visual cues. ₃₄₈

₃₄₉

6.2.1 VRMate ₃₅₀

VRMate is the companion program to behaviorMate that runs the visual VR interface to display cues ₃₅₁
to animals as they move in virtual space. VRMate is a program developed using the Unity(TM) game ₃₅₂
engine and written in C#; it listens for incoming UDP packets carrying position updates from the ₃₅₃
behaviorMate UI. In our implementation, each visual display is connected to an ODROID C4 which ₃₅₄
is a single-board computer equipped with an Amlogic S905X3 Quad-Core Cortex-A55 (2.016GHz) ₃₅₅
ARMv8-A Processor, 4 GiB of DDR4 RAM, Mali-G31 MP2 GPU, and 1 LAN port, all mounted ₃₅₆
on a single PCB and running the Android operating system. Each display-ODROID pair runs an ₃₅₇
instance of the VRMate program, which means scene rendering is done independently by dedicated ₃₅₈
hardware for each monitor. This allows for scalability to any number of additional displays without ₃₅₉
compromising frame rate or resolution. Prior to starting an experiment, behaviorMate supplies each ₃₆₀
instance of VRMate with all the information it needs to correctly display the virtual environment ₃₆₁

from a particular point of view (Fig. 4F). Notably, Unity is platform independent, so VRMate can 362

be built for Windows, OSX, and other systems. Therefore, it is also not strictly necessary to use 363

dedicated Android devices, and it is possible to run behaviorMate and VRMate on the same PC. 364

Various setups can be configured to meet individual researchers' needs. 365

Specifically for researchers conducting imaging experiments, Unity supports the implementation 367

of custom "shader programs" which can be enabled through the UI. Implementing custom shaders 368

allows you to edit color-pixel values in the visual scenes just prior to when they are rendered. For 369

example, this means it is possible to remove a particular color emission from the scenes or display 370

with just a single color (Fig. 4F). This is particularly useful for experiments where green or red 371

fluorescent proteins are used to indicate neuronal activity. In this case, rendering scenes with 372

the green or red component of each pixel set to 0 minimizes the risk of artifacts in the recorded data. 373

6.2.2 Frame and Display Setup 375

Our standard setup for immersive mouse VR experiments contains 5 monitors surrounding the 376

animal. We supply plans for a VR frame to construct this (Fig. 4A, *also see* Supplementary 377

Materials). The VR frame is designed with extruded aluminum rails and holds the displays at the 378

proper angle and distance from the animal. The goal is to completely encompass the animal's 379

visual field with the virtual scene, so the displays are placed in an octagonal setup with the animal 380

at the center. The angle between each display is 135 degrees and the distance between the mouse 381

and the center display is 15 inches. Since mice primarily have a visual field oriented overhead 382

(Dräger & Olsen 1980), the mouse is positioned 5 inches above the bottom edge of each monitor in 383

order to align the visual stimulus to the animals field of view. 384

In cases when the standard 5 monitor setup will not fit due to space constraints, we use a 386

wireless, tablet-based setup. VRMate is built to run on the Android OS and can be installed on 387

any Android tablet. Our tablet system consists of 5 "Fire HD 10" (2019, 9th Gen), 10.1 inch tablets 388

with a screen resolution of 1920 x 1200 IPS while running Android 9. Additional specs include 2 389

GB RAM, 4xARM Cortex-A73 (2.0 GHZ) and 4xARM Cortex-A53 (2.0 GHz) processors, and a   390
ARM Mali-G72 MP3 GPU. Tablets may be connected to the behaviorMate intranet using a wireless   391
router. The PC running behaviorMate connects through this router to interface with the displays.   392
Each device on the network must be assigned a unique static IP address. This is achieved by   393
configuring the IP tables on the Wi-Fi-enabled router to bind each tablet's IP address to its MAC   394
address. A low-latency router with multiple antennae is recommended and should be placed as   395
close as possible to the tablet displays.   396

Typical VR setups position animals on top of a running wheel which is less bulky and me-   398
chanically simpler than the fabric belt treadmill system. The running wheel design used in our   399
setup, as described previously by Warren et al., is lightweight and optimized to have low rotational   400
momentum. This provides animals with a natural feeling of running while minimizing the effort   401
required to traverse the virtual environments. Mice are head-fixed on top of the wheel using custom   402
head restraining bars and a matching holder using a dove-tail and clamping system. The implant   403
holders (Fig. S1B) are two L-shaped components machined from solid steel. They have two M6   404
holes for being attached to pillar posts of aluminum rails. In addition, the implant holders each   405
have a channel which allows the clamps to slide and accommodate head-bars of different lengths.   406
Both the clamps and clamp holders have thumb screws not shown in the figure. The screw for the   407
clamps tightly grips the head-bar and thus minimizes motion artifacts in imaging data. The screw   408
for the clamp holders fixes the clamps' position and reduces motion artifacts. The head restraining   409
bars for mice presented with visual cues (Fig. 4G) are designed to assist with blocking light that   410
may otherwise be detected during imaging. These can be 3D printed from titanium and reused   411
between animals if properly cleaned and sterilized.   412

In some cases, a modified running wheel design with goniometers is required to tilt both the im-   414
plant holders and the entire wheel along the roll and pitch axes (Fig. 4B). For head-fixed two-photon   415
imaging, this is used to align the imaging plane to maximize the size of the field of view or, as   416
needed, level the glass coverslip placed over the brain (which minimizes the amount of the material   417
the beam must travel through and maximizes image clarity). For this modified design, aluminum   418

rails are arranged in a "U" shape and then placed on top of two goniometers. The implant holders 419

sit on top of these rails, and the wheel is suspended below the rails, offset from the goniometers. 420

The entire assembly sits on sliding mounts connected to pillar posts so the elevation of the animal's 421

head can be adjusted. 422

6.2.3   Example setup 424

A simple VR system could have 2 displays in an arrow configuration directly in front of a mouse 425

which is head fixed and mounted on top of a running wheel. Attached to the back of each monitor 426

would be a small Android computer (ODROID) connected via an HDMI cable. The left screen could 427

be given an IP address of 192.168.1.141. The right machine will be assigned 192.168.1.142. Static 428

IP assignment is done through the Android devices' settings menu. Arduinos running the Behavior 429

and Position Controllers are given the IPs of 192.168.1.101 and 192.168.1.102 respectively (which 430

is the default in our supplied code (*see Supplementary Materials*). Finally, a PC running the UI is 431

assigned an IP of 192.168.1.100. We use standard IP address and ports for each of the displays 432

and electronics; however, the IPs that behaviorMate will use are specified in the settings file and 433

determined at run-time (the Arduino IP assignments are made when the devices are loaded). Each 434

device is also assigned a port which can be selected arbitrarily and can be changed as needed. 435

The only requirement is that all devices are on the same subnet. In this example setup, all devices 436

share the same network prefix of 192.168.1.x so they are part of the same subnet and will be able 437

to communicate. Given that all components are connected to the same local network, usually via 438

an unmanaged switch, packets will arrive to their intended recipient. 439

When the animal runs, turning the wheel and rotary encoder, behaviorMate will receive and 441

integrate movement updates from the Position Controller. Once the experimenter begins a trial 442

from the UI, the PC will evaluate which contexts are active and send position updates to both 443

screens simultaneously. The VRMate instances will receive these packets and render their view 444

of the same scene independently. Latency between the UI and the computers running VRMate 445

is fairly low (Fig. 4C) so this does not present an issue. As the animal runs, it enters a reward 446

zone. behaviorMate will send a message to the Behavior Controller triggering the "reward context", [447] causing a reward valve to release a drop of water or sucrose reward. As the mouse runs, the visual [448] scene updates accordingly. Once the mouse travels a linear distance equal to the "track_length" [449] parameter specified in the settings file, behaviorMate advances the lap count and resets the position [450] to 0. Between-lap "time outs" or dark periods may also be configured. Furthermore, the user may [451] choose to run a "context-switch" experiment: this type of experiment might involve one scene being [452] displayed on odd laps and another on even laps. If a scene change is necessary, between laps, [453] behaviorMate will send a message to both running instances of VRMate that specify the switch to [454] the alternate scene. [455]

[456]

6.2.4 VR System Performance [457]

One key concern with closed-loop control VR experimental setups is the latency between when the [458] animal moves the physical wheel and when the displays update position in the virtual environment. [459] We therefore performed a benchmarking test on our standard 5 monitors VR/running wheel setup [460] (with scenes rendered by ODROIDs connected to the UI via an unmanaged Ethernet switch). An im- [461] pulse was applied to the running wheel, and the time taken for the screens to update was observed. [462] The screen delay was measured using a high-speed 150 fps camera. Latency was sampled 20 times [463] and found to have a median value of 0.067 seconds, with no sample exceeding 0.087 seconds (Fig. [464] 4C). Since mouse reaction times are approximately 0.1 seconds (Dombeck & Reiser 2012, Mauk [465] & Buonomano 2004), VR position updates using this system are nearly instantaneous to the mouse. [466]

[467]

As with the fabric treadmill, another characteristic that should be considered for the running [468] wheel is the accuracy of the position tracking system. Accuracy refers to how closely the mouse's [469] virtual position recorded by behaviorMate matches the true linear distance it has run on the wheel. [470] To find the "true" linear distance traveled, a small dim LED was attached to the side of the running [471] wheel, and the whole setup was placed in a dark enclosure, so the only significant source of light [472] was the LED. The wheel was advanced by hand, and using OpenCV (Bradski 2000), the total [473] distance traversed by the LED was extracted. Concurrently, behaviorMate was detecting the current [474]

position using the attached rotary encoder. Graphs of position vs. time from the data collected using computer-vision and the data recorded by behaviorMate were overlaid. From 4D, it is clear that there is little discrepancy between the two positions. Across all time points, the mean difference between the true and recorded position was 4.1 mm.

6.2.5   Use in Past Experiments

While the the introduction of a purely visual VR option for behaviorMate represents one of the newest updates to the setup, studies have already been published that specifically leverage this capability, primarily to examine the effect of rapid within-session context changes (Priestley et al. 2022, Bowler & Losonczy 2023). VR experiments have the ability to transport animals to completely novel scenes which has been useful for investigating the mechanisms of synaptic plasticity, since it is possible to capture the animals first exposure to an environment in a controlled way (Priestley et al. 2022). Since the visual VR setup is integrated into behaviorMate, it is also possible to test how animals respond to changes in the rules governing reward delivery and visual scene presentation. The system further allows for simultaneous delivery of multi-modal stimuli simultaneously such as audio and visual. Thus, our system comprises an adaptable framework for probing the function of navigational circuits in the brain and the relationship between goal and context representations (Bowler & Losonczy 2023).

7   Experimental Validation

To investigate if spatially tuned 'place cells' (O'Keefe & Dostrovsky 1971) with similar properties emerge in both the belt treadmill system (TM) and the VR running wheel system, we performed 2p imaging in CA1 of the hippocampus while animals explored in either VR or TM. Mice were injected with GCaMP8m (AAV1-syn-GCaMP8m-WPRE, Addgene) and implanted with a glass window to have optical access to CA1 (Fig. 5). Once mice recovered from surgeries, they were trained to do a spatial navigation task in either VR or TM. After the mice were fully trained, 2p imaging was performed during navigation. Calcium signals were processed with Suite2p for motion correction,

cell segmentation, and fluorescent signal and neuropil extraction (Pachitariu et al. 2017). Extracted 502
raw fluorescent signals were corrected for neuropil contamination, and resulting ΔF/F was used to 503
estimate spike events using OASIS (Friedrich et al. 2017). A median filter was applied to binarize 504
spike amplitudes into bins of position, and all subsequent analyses used this binarized train as an 505
estimation of each ROIs activity. We recorded on average 1533$\pm$481 and 1057$\pm$169 CA1 neurons 506
in VR and TM, respectively. Although VR had significantly fewer CA1 neurons being classified as 507
place cells compared to TM (VR: 34.9% (3253/9195), TM: 44.9% (6101/13454), two sample z-test, 508
z=14.96, p=1.27e-50), place cells tiled the entire track in both VR and TM (*see* Supplementary 509
Materials). In all subsequent analyses, we used CA1 neurons identified as a place cell unless 510
otherwise stated. To assess differences in the calcium signal of place cells found in context- 511
environment pairs, we first compared the ΔF/F amplitude and frequency of individual ROIs identified 512
as place cells. Only ΔF/F events that were 3 standard deviations above the mean ΔF/F were 513
included in these analyses. To examine the effect of environment (VR vs. TM) and context (familiar 514
(F) vs. novel (N)) on ΔF/F amplitude and frequency, we conducted two-way repeated measures of 515
ANOVA with both environment and context as a within-subject factors. There was no significant 516
effect of environment or context on mean ΔF/F amplitude (environment: $F(1,5)$=1.9387, p=0.222, 517
context: $F(1,5)$=0.0037, p=0.9537) or on mean ΔF/F frequency (environment: $F(1,5)$=4.37, p=0.091, 518
context: $F(1,5)$=0.0789, p=0.7901). These results suggest that calcium transient properties of place 519
cells observed in both environments and contexts were similar. Next, we quantified individual place 520
cells' specificity, sensitivity, and spatial information to identify any differences in place cell firing 521
properties between context-environment pairs. As described above, two-way repeated measures 522
of ANOVA were used with both context and environment as within-subject factors. The two-way 523
ANOVA revealed a significant effect of environment on sensitivity (environment: $F(1,5)$=9.85, 524
p=0.0257, context: $F(1,5)$=3.34, p=0.13) and specificity (environment: $F(1,5)$=53.32, p=0.0008, 525
context: $F(1,5)$=2.55, p=0.17), suggesting that place cells in TM have higher trial-by-trial in-field 526
firing compared to those from VR. However, we did not observe any significant effect of environment 527
or context on spatial information (Skaggs & Mcnaughton 1998) (environment: $F(1,5)$=0.74, p=0.43, 528
context: $F(1,5)$=3.79, p=0.11). These results are consistent with previous findings that place cells 529
in VR have reduced trial-by-trial reliability (Ravassard et al. 2013, Aghajan et al. 2015), but still 530
encode spatial information (Dombeck et al. 2010, Aronov & Tank 2014, Hainmueller & Bartos 2018, 531

Priestley et al. 2022). The observed differences in the properties of VR and TM place cells may be ⁵³² the result of the absence of proximal visual cues and tactile cues (which are speculated to be more ⁵³³ salient than visual cues) during VR tasks. Next, we wanted to investigate if place cells remap in ⁵³⁴ both VR and TM environments when mice are exposed to a novel context (Muller & Kubie 1987, ⁵³⁵ Leutgeb et al. 2005). As expected, a subset of place cells changed their place tuning properties ⁵³⁶ when exposed to novel contexts (Fig. 6A,B). Moreover, we did not observe differences in the ⁵³⁷ between context rate map correlation between VR and TM, suggesting that remapping happens in ⁵³⁸ both environment types. To further quantify the extent of remapping in each environment, we built a ⁵³⁹ naive Bayes decoder to decode position based on population activity of CA1 neurons. When the ⁵⁴⁰ decoder was trained with population activity from familiar trials and tested with population activity ⁵⁴¹ from a held-out familiar trial, the absolute decoding error was significantly below chance (p<0.05) ⁵⁴² in both VR and TM. However, when the decoder trained with familiar trials was tested to decode ⁵⁴³ position in a novel trial, the absolute decoding error was at the chance level (Fig. 6C,D) in both VR ⁵⁴⁴ and TM. These suggest that place cells remap in both VR and TM when exposed to novel context. ⁵⁴⁵

8 Discussion ⁵⁴⁶

behaviorMate is, most generally, a solution for controlling and collecting feedback from devices ⁵⁴⁷ in behavioral experiments. The behaviorMate methodology is built around an "Intranet of Things" ⁵⁴⁸ approach, expanding to incorporate additional components to meet experimental demands but also ⁵⁴⁹ compatible with sparser setups to ease complexity when not needed. It has typically been used ⁵⁵⁰ concurrently with neuronal imaging, since the time-stamped output of behavior events is relatively ⁵⁵¹ simple to align with imaging data, providing experimenters with the ability to correlate neuronal ⁵⁵² events with behavior. In addition, behaviorMate is designed to be flexible, expanding to incorporate ⁵⁵³ additional components as specified in the settings while making minimal assumptions. Experi- ⁵⁵⁴ menters may incorporate other techniques such as imaging and electrophysiology recordings or just ⁵⁵⁵ run behavior. For experimental paradigms involving any combination of displays, one-dimensional ⁵⁵⁶ movement (wheels, treadmills, etc.), sensors (LEDs, Capacitance lick sensors, etc.), and actuators ⁵⁵⁷ (odor release systems, solenoid valves, etc.), this is a cost-effective and reliable solution. ⁵⁵⁸

⁵⁵⁹

The main modules of behaviorMate are the Java GUI application, referred to as the UI, Behavior 560
Controller, and Position Controller. The UI coordinates most communication between devices, 561
although sometimes devices directly communicate with each other for performance reasons. The 562
Behavior Controller sends messages to and receives data from nearly any kind of peripheral 563
devices such as 2-photon microscopes, LEDs, speakers, computer displays, electrophysiology 564
systems, etc. The Position Controller is solely responsible for receiving position updates from the 565
rotary encoder. Any number of additional devices may be used by connecting them to the behavior 566
controller using standard connectors. Furthermore, the companion software of VRMate allows for 567
low-cost integration of virtual reality experiments that are already compatible with behaviorMate. 568
The appeal of behaviorMate is its modularity, reliability, and open source code base. 569

Finally, many published studies have relied on this system and the resulting data collected 571
(Bowler & Losonczy 2023, Tuncdemir et al. 2023, Vancura et al. 2023, Priestley et al. 2022, O'Hare 572
et al. 2022, Rolotti et al. 2022a, Tuncdemir et al. 2022, Rolotti et al. 2022b, Terada et al. 2022, Geiller 573
et al. 2022, Blockus et al. 2021, Grosmark et al. 2021, Dudok et al. 2021a,b, Geiller et al. 2020, 574
Kaufman et al. 2020a, Turi et al. 2019, Zaremba et al. 2017, Danielson et al. 2017, 2016). These 575
paradigms varied greatly and included random foraging, goal-oriented learning, multi-modal cue 576
presentations, and context switches (both in VR environments and physical treadmills). Instructions 577
for downloading the UI software and the companion VRMate software, and assembling a Behavior 578
Controller, Position Controller, running wheel, VR monitor frame, and treadmill, can be found on the 579
Losonczy Lab website (www.losonczylab.org/software). 580

Figure 1. Overview of the behaviorMate system with optional components *Diagram shows the modular components behaviorMate can interact with. The colored arrows show the direction of information flow. For example, the Position Tracking module only receives position updates. The Visual Display module sends data to the Display Controllers to render the scene. Devices using GPIO may both send and receive data to and and from behaviorMate.*

Figure 2. Details of UI function. (**A**) Screenshot of the UI. The interface provides a snapshot of the animal's current status (*center*) as well as cumulative information about the current experiment (*upper-left*). The control panel along the left side provide: *Top.* an input for a project name and mouse id, which control where the resulting behavior file will be saved (these boxes are disabled when an experiment is currently running), *Middle.* controls to trigger the GPIO to send a TTL pulse i.e. to issue a water reward or turn on an LED, and *Bottom.* controls to start/stop experiments as well as to load settings files. (**B**) Details of UI event loop. The UI is continuously executing an event loop on every update step which checks for messages, writes the behavior file and updates the display. (**C**) JSON formatted message passing. The PC sends and receives JSON formatted messages via UDP to active components of the system. Messages have a nested structure to allow for internal routing and subsequent actions to be taken. (**D**) Rendering of the GPIO circuit used. *Left.* JST headers connected to I/O pins and pull-down resisters to allow for easy connections with peripheral devices that can be activated by TTL pulses (both at 3.3V and 12V, depending on the pins used). Additionally, power connections and a BNC output trigger are provided. *Right.* This board attaches to an Arduino Due microcontroller. (**E**) Updates to the position have a dedicated circuit for decoding quadrature information and sending updates to the PC. Connections for a quadrature based rotary encoder, power, and Ethernet are provided. Additionally, an input for a "Lap Reset" sensor is provided and headers to interface with an Arduni Mini.

Figure 3. Details of Treadmill System. (**A**) CAD model of voluntary treadmill that can expand or contract to accommodate different sized running belts. Implant holders sit on goniometers to allow mouse head angle to be adjusted. (**B**) Lap time error is defined for each lap (full turn of the treadmill belt) as the difference between the time recorded by behaviorMate and the computer vision benchmark in seconds. (**C**) Absolute error is defined for each lap as the difference between the position in behaviorMate and the benchmark position in millimeters. (**D**) The time to complete a lap according to behaviorMate and the benchmark were nearly identical. (**E**) Schematic of head-fixed behavioral task. Two-photon objective. (**F**) Cue-rich belts for treadmill behavioral tasks. Belts are easily interchangeable. (**G**) Example test condition spanning 3 days and involving 5 sensory cues. Each row represents a sensory modality (visual, auditory, etc.). Cues are tied to locations on the belt or wheel as indicated by the colored rectangles. Locations and durations of cue presentations can be changed across trials.

Figure 4. Details of VR System. (**A**) The complete assembly. 5 monitors are suspended in an octagonal pattern in front of the subject and display the virtual scenes. (**B**) A running wheel attached to a goniometer to allow for angle of subject's head to be adjusted. Placed directly underneath the objective. (**C**) Box plot describing the delay between when the running wheel is moved and when the virtual scene is updated to reflect the new position. (**D**) Plot showing the difference between the position of the animal computed by behaviorMate and the ground-truth position tracked using computer vision. (**E**) Box plot describing the absolute difference between the current position of the mouse according to behaviorMate and the computer vision benchmark at each point in time. The median error was found to be 3.3 mm. (**F**) A 2-D projection of the displays which will resemble what the test animal will see. When viewed on the actual monitors, the scene will appear more immersive. (**I**) A VR scene as the subject moves down the virtual track. Modifying a few numerical parameters in the settings files allows one to change the appearance and view angle of a scene. Bottom: A scene with all red color shaders removed. (**G**) Left: Top view of implant. This side will make contact with microscope objective. Right: Bottom view. This side will make contact with the mouse's brain. (**H**) Sketch of how the implant will appear after being surgically attached to skull.

Figure 5. 2-photon imaging of CA1 population for experimental validation (A) Field of view obtained from virtual reality (VR) and treadmill (TM) **(B)** Examples of place cells recorded using VR or TM. In both familiar and novel contexts, place cells encoding different locations on the track are found. Note that they are different place cells from individual animals. **(C)** Mean Δf/f amplitude of place cells recorded each context-environment pair (F = familiar context, N = novel context). **(D)** Mean Δf/f frequency of place cells. For (C) and (D), only significant Δf/f events were used. See methods for details. **(E)** Place sensitivity in each context-environment pair. Significant main effect of environment (VR vs. TM, p = 0.0257). **(F)** Place cell specificity in each context-environment pair. Significant main effect of environment (VR vs. TM, p = 0.008) **(G)** Spatial information (bits/event) in each context-environment pair. Two-way repeated measures of ANOVA was used unless otherwise stated. Refer to Table ST1 for ANOVA table. For all panels, n.s. = non-significant, *P<0.05, ***P<0.001

Figure 6. Place cells remap in both environments (A) Place cells recorded in VR or TM sorted by their familiar peak (top) or novel peak (bottom) locations. A subset of CA1 neurons remaps when exposed to novel context in both VR and TM. **(B)** Rate map correlation of individual place cells between familiar and novel contexts. No difference was observed between VR and TM (Wilcoxon rank-sum test, F=1.38, p=0.17). **(C)** Confusion matrices of the Bayesian decoder for VR and TM. F-F indicates a model trained and tested with trials from the familiar context. F-N is a model trained with trials from familiar context, and tested with a novel trial. **(D)** Absolute decoding error for the maximum likelihood estimator. A dashed line indicates a chance level (88±1.96cm for VR/ 50cm for TM, see methods). Note that when the decoder is trained and tested with familiar context trials, absolute decoding accuracy is below chance in both VR and TM (VR: 27.62±1.44cm, p<0.01, TM: 17.78±0.44cm, p<0.01, permutation test). However, when the decoder is trained with familiar context trials, but tested with novel context trials, decoding accuracy was increased to the chance level (VR: 99.80±3.95cm, p=1.00, TM: 56.82±2.46cm, p=1.00).

Acknowledgments 581

J.C.B. is supported by NIMH F31NS110316. J.B.P. is supported by the EPFL ELISIR fellowship and 582
the Fondation Marina Cuennet-Mauvernay. A.L. is NIMH R01MH124047, NIMH R01MH124867, 583
NINDS R01NS121106,NINDS U01NS115530, NINDS R01NS133381, NINDS R01NS131728, NIA 584
RF1AG080818. We thank Drs. Ali Kaufman and Ally Lowell for troubleshooting and debugging the 585
initial iterations of behaviorMate as well as Dr. Andres Grosmark for continued feedback on the 586
project. Additionally, we thank Dr. Stephanie Herrlinger, Abhishek Shah, and all other members of 587
the Losonczy lab for comments and discussion on the manuscript. 588

Author Contributions 589

J.C.B. and A.L. conceived of the original method and experimental requirements. Funding support 590
from A.L. J.C.B. developed software with assistance from W.L. and Z.L. W.L developed custom 591
PCBs. G.Z. developed a JavaFX version of behaviorMate, performed benchmarking and mechanical 592
design. J.C.B., G.Z. and H.C.Y. wrote the original draft with input from all authors. H.C.Y. performed 593
the experimental validation with assistance from B.R. J.B.P. developed the VR layout. J.C.B., G.Z., 594
B.R., Z.L. and J.B.P. performed testing and debugging. All authors contributed to review and editing. 595

Declaration of Interests 596

The authors declare no competing interests. 597

9 Methods

9.1 Lead Contact and Materials Availability

Further information and requests for resources and reagents should be directed to the Lead Contact
Attila Losonczy (al2856@columbia.edu). All unique resources generated in this study are available
from the Lead Contact with a completed Materials Transfer Agreement.

9.2 Experimental Model and Subject Details

9.2.1 Animals

All experiments were conducted in accordance with the NIH guidelines and with the approval of the
Columbia University Institutional Animal Care and Use Committee. Experiments were performed
with six adult mice between 8-16 weeks. 5 mice were Unc5b crossed with C57Bl/6, and 1 mouse
was VGAT-cre crossed with Ai9 (Jackson Laboratory). 3 female and 3 male mice were used in this
experiment to balance for sex.

9.2.2 Surgical procedures

Surgeries were done as previously described (Lovett-Barron et al. 2014). Briefly, animals were
anesthetized with isoflurane and injected with jGCaMP8m (rAAV1-syn-GCaMP8m-WPRE, Addgene:
162375-AAV1) in dorsal CA1 (AP: 2.1, ML: 1.4, DV: 1.3, 1,2, 1,1, 100uL per site) using Nanoject
III (Drummond). Animals were allowed to recover from the injection surgery for 7 days, and the
cortex lying above CA1 was aspirated to implant a glass cannula to enable optical access to CA1.
Finally, a titanium headpost was cemented to the skull using dental cement to head-fix animals for
2-photon functional imaging. Analgesics were provided as per Columbia University Institutional
Animal Care and Use Committee prior and after the surgeries for up to 3 days.

9.2.3 2-photon imaging and behavior

After recovering from surgeries, animals were habituated to head fixation in a VR environment for
15 minutes. Subsequently, animals were trained to run on a lightweight wheel, and 5% sucrose
was given at random locations to motivate animals' to run in the VR or TM. When animals reliably

ran at least 60 laps in VR, or 30 laps in TM, the reward was fixed to either 1 or 2 locations. A 623

3m VR track was used for all animals except jy030, which had 4m VR track instead, and 2m TM 624

belt was used in all animals. For VR experiments, animals were imaged for one session, and the 625

context switch happened after animals ran 30 laps in familiar context. Animals were required to 626

run at least another 30 laps in the novel context. In TM, imaging was done over two days (first day: 627

familiar belt, second day: novel belt). 2p imaging was performed using a 8kHz resonant scanner 628

(Bruker) and a 16x Nikon water immersion objective (0.8 NA, 3mm working distance). In order to 629

image GCaMP8m signals, 920nm excitation wavelength was used (Coherent), and the power from 630

the tip of the objective never exceeded 100mW. GCaMP signals were collected using a GaAsP 631

photomultiplier tube detector (Hamamatsu, 7422P-40) following amplification via a custom dual 632

stage preamplifier (1.4 x 105 dB, Bruker). All imaging was performed using 512 x 512 pixels with 633

digital zoom between 1 and 1.4. 634

9.2.4   Calcium signal processing 635

Imaging data was processed as previously described (Priestley et al. 2022). In short, the SIMA 636

software package (Kaifosh et al. 2014) was used to organize the imaging data set, and the imaging 637

data was processed with Suite2p (Pachitariu et al. 2017) for motion correction, signal/neuropil 638

extraction, and ROI detection. Following ROI detection, individual ROIs were manually curated 639

using the Suite2p GUI to exclude non-somatic ROIs. ΔF/F signals were calculated after subtracting 640

the neuropil and correcting for baseline. 641

9.2.5   Data analysis 642

As previously described, all analyses were done using binarized event signals (Ahmed et al. 2020, 643

Priestley et al. 2022). Briefly, ΔF/F signals were deconvolved using OASIS (Friedrich et al. 2017), 644

and the resulting estimated spike amplitude train was binarized by selecting events whose ampli- 645

tudes were above 4 median absolute deviations of the raw trace. We do not claim that this reflects 646

the true spiking activities of individual ROIs. Binarized event trains were binned into 4cm bins 647

and smoothed with a Gaussian filter (S.D. = 2) to calculate tuning curves. Individual ROIs were 648

classified as a place cell by identifying bins that have activities greater than the 99th percentile 649

of surrogate tuning curves (Priestley et al. 2022). In short, surrogate average tuning curves were 650
calculated by performing 1,000 circular shuffles of their occupancy of individual trials. Bins that had 651
activity greater than 99th percentile of the tuning curves were identified as potentially significant 652
bins. In order to be classified as a place cell, ROIs needed to have at least 3 consecutive significant 653
bins (12cm), but less than 25 consecutive bins (100cm). Moreover, in order to avoid spurious 654
detection of significant bins, binarized events must have been detected in at least 50% of trials 655
within those significant bins. 656

For mean ΔF/F amplitude and frequency calculations, only bins with 3 standard deviations 658
above the mean were used. For place cell sensitivity, representing the proportion of laps that had 659
active events within the significant field, the number of laps that had at least 1 binarized event within 660
the significantly detected fields was divided by the total number of laps. For place cell specificity, 661
total number of events within the significant field was divided by the total number of events observed 662
within the lap. Then, it was averaged across all laps to have a single value for each ROI. If multiple 663
fields were detected, they were computed separately and averaged across fields to have a single 664
value for each ROI. For spatial information, it was calculated as described previously (Skaggs & 665
Mcnaughton 1998). For remapping analysis, one animal was used for the analysis (jy065). Given 666
that TM imaging happened across two days, FOVs recorded from TM were matched using CellReg 667
(Sheintuch et al. 2017), and spatial mask correlation was used to find the same ROIs. Results were 668
manually curated within the Suite2p GUI. This was not necessary for VR since the context switch 669
happened within the same session. To calculate rate map correlation, the Pearson correlation 670
coefficient was computed between average tuning curves from the famliar and novel contexts. For 671
TM, given that place cells may anchor on seams of the belt, the shift that generated the maximum 672
population vector correlation between familiar and novel contexts was calculated. All ROIs recorded 673
in the novel context were circularly shifted by the same amount, and the rate map correlation was 674
calculated as described above. A naive Bayesian classifier was used to decode animals' position 675
on the track (Zhang et al. 1998). For each frame and spatial bin: 676

$$P(pos|a_{all}) = C \left(\prod_{i=1}^{N} f_i(pos)^{a_i} \right) e^{-\tau \sum_{i=1}^{N} f_i(pos)}$$

where N is the total number of cells, $f_i(pos)$ is the average binned activity of a cell, τ is the bin

size, and a_i is the activity on the frame. C is the normalization constant. The position bin with the

highest probability was selected as a decoded position. Chance level for VR was calculated by

randomly shuffling cell labels of the testing set. For TM, a chance level was determined as track

length / 4, which is a chance level of a circular environment (50cm). The decoder was trained with

n-1 trials and tested on the left out trial. To match ROI numbers recorded between VR and TM,

500 ROIs were randomly selected, and the decoding accuracy was calculated 50 times. Average

absolute decoding accuracy was compared to the chance level calculated from each iteration, and

the number of times that it was below the chance level was considered to be its p value.

References

Aghajan, Z. M., Acharya, L., Moore, J. J., Cushman, J. D., Vuong, C. & Mehta, M. R. (2015), 'Impaired spatial selectivity and intact phase precession in two-dimensional virtual reality.', *Nature neuroscience* **18**, 121–8.

Ahmed, M. S., Priestley, J. B., Castro, A., Stefanini, F., Canales, A. S. S., Balough, E. M., Lavoie, E., Mazzucato, L., Fusi, S. & Losonczy, A. (2020), 'Hippocampal network reorganization underlies the formation of a temporal association memory.', *Neuron* **107**, 283–291.e6.

Akam, T., Lustig, A., Rowland, J. M., Kapanaiah, S. K., Esteve-Agraz, J., Panniello, M., Márquez, C., Kohl, M. M., Kätzel, D., Costa, R. M. & Walton, M. E. (2022), 'Open-source, python-based, hardware and software for controlling behavioural neuroscience experiments', *eLife* **11**, e67846.
URL: *https://doi.org/10.7554/eLife.67846*

Aronov, D. & Tank, D. W. (2014), 'Engagement of neural circuits underlying 2d spatial navigation in a rodent virtual reality system', *Neuron* **84**, 442–456.
URL: *http://dx.doi.org/10.1016/j.neuron.2014.08.042*

Arriaga, M. & Han, E. B. (2017), 'Dedicated hippocampal inhibitory networks for locomotion and immobility', *Journal of Neuroscience* **37**(38), 9222–9238.
URL: *https://www.jneurosci.org/content/37/38/9222*

Arriaga, M. & Han, E. B. (2019), 'Structured inhibitory activity dynamics in new virtual environments', *eLife* **8**, e47611.
URL: *https://doi.org/10.7554/eLife.47611*

Bittner, K. C., Milstein, A. D., Grienberger, C., Romani, S. & Magee, J. C. (2017), 'Behavioral time scale synaptic plasticity underlies ca1 place fields.', *Science (New York, N.Y.)* **357**, 1033–1036.
URL: *http://www.ncbi.nlm.nih.gov/pubmed/28883072*

Blockus, H., Rolotti, S. V., Szoboszlay, M., Peze-Heidsieck, E., Ming, T., Schroeder, A., Apostolo, N., Vennekens, K. M., Katsamba, P. S., Bahna, F., Mannepalli, S., Ahlsen, G., Honig, B., Shapiro, L., de Wit, J., Losonczy, A. & Polleux, F. (2021), 'Synaptogenic activity of the axon guidance molecule robo2 underlies hippocampal circuit function.', *Cell reports* **37**, 109828.

Bowler, J. C. & Losonczy, A. (2023), 'Direct cortical inputs to hippocampal area ca1 transmit complementary signals for goal-directed navigation', *Neuron* .

URL: *https://linkinghub.elsevier.com/retrieve/pii/S0896627323007018*

Bradski, G. (2000), 'The OpenCV Library', *Dr. Dobb's Journal of Software Tools* .

Campbell, M. G., Ocko, S. A., Mallory, C. S., Low, I. I. C., Ganguli, S. & Giocomo, L. M. (2018), 'Principles governing the integration of landmark and self-motion cues in entorhinal cortical codes for navigation.', *Nature neuroscience* **21**, 1096–1106.

Danielson, N. B., Kaifosh, P., Zaremba, J. D., Lovett-Barron, M., Tsai, J., Denny, C. A., Balough, E. M., Goldberg, A. R., Drew, L. J., Hen, R., Losonczy, A. & Kheirbek, M. A. (2016), 'Distinct contribution of Adult-Born hippocampal granule cells to context encoding', *Neuron* **90**(1), 101–112.

URL: *https://doi.org/10.1016/j.neuron.2016.02.019*

Danielson, N. B., Turi, G. F., Ladow, M., Chavlis, S., Petrantonakis, P. C., Poirazi, P. & Losonczy, A. (2017), 'In vivo imaging of dentate gyrus mossy cells in behaving mice.', *Neuron* **93**, 552–559.e4.

Dombeck, D. A., Harvey, C. D., Tian, L., Looger, L. L. & Tank, D. W. (2010), 'Functional imaging of hippocampal place cells at cellular resolution during virtual navigation', *Nature Neuroscience* **13**(11), 1433–1440.

URL: *https://doi.org/10.1038/nn.2648*

Dombeck, D. A. & Reiser, M. B. (2012), 'Real neuroscience in virtual worlds', *Current Opinion in Neurobiology* **22**(1), 3–10. Neurotechnology.

URL: *https://www.sciencedirect.com/science/article/pii/S0959438811001814*

Dräger, U. C. & Olsen, J. F. (1980), 'Origins of crossed and uncrossed retinal projections in pigmented and albino mice.', *The Journal of comparative neurology* **191**, 383–412.

URL: *http://www.ncbi.nlm.nih.gov/pubmed/7410600*

Dudok, B., Klein, P. M., Hwaun, E., Lee, B. R., Yao, Z., Fong, O., Bowler, J. C., Terada, S., Sparks, F. T., Szabo, G. G., Farrell, J. S., Berg, J., Daigle, T. L., Tasic, B., Dimidschstein, J., Fishell, G., Losonczy, A., Zeng, H. & Soltesz, I. (2021), 'Alternating sources of perisomatic inhibition during

behavior', *Neuron* **109**(6), 997–1012.e9.

URL: *https://doi.org/10.1016/j.neuron.2021.01.003*

Dudok, B., Klein, P. M., Hwaun, E., Lee, B. R., Yao, Z., Fong, O., Bowler, J. C., Terada, S., Sparks, F. T., Szabo, G. G., Farrell, J. S., Berg, J., Daigle, T. L., Tasic, B., Dimidschstein, J., Fishell, G., Losonczy, A., Zeng, H. & Soltesz, I. (2021b), 'Alternating sources of perisomatic inhibition during behavior.', *Neuron* **109**, 997–1012.e9.

Dudok, B., Szoboszlay, M., Paul, A., Klein, P. M., Liao, Z., Hwaun, E., Szabo, G. G., Geiller, T., Vancura, B., Wang, B.-S., McKenzie, S., Homidan, J., Klaver, L. M. F., English, D. F., Huang, Z. J., Buzsáki, G., Losonczy, A. & Soltesz, I. (2021a), 'Recruitment and inhibitory action of hippocampal axo-axonic cells during behavior.', *Neuron* **109**, 3838–3850.e8.

Friedrich, J., Zhou, P., Paninski, L., Staneva, V., Chklovskii, D. & Pnevmatikakis, E. (2017), 'Fast online deconvolution of calcium imaging data', *PLOS Computational Biology* **13**, e1005423.

URL: *http://dx.plos.org/10.1371/journal.pcbi.1005423*

Gamma, E., Helm, R., Johnson, R. & Vlissides, J. M. (1994), *Design Patterns: Elements of Reusable Object-Oriented Software*, 1 edn, Addison-Wesley Professional.

Geiller, T., Fattahi, M., Choi, J.-S. & Royer, S. (2017), 'Place cells are more strongly tied to landmarks in deep than in superficial ca1', *Nature Communications* **8**, 14531.

URL: *http://www.nature.com/articles/ncomms14531*

Geiller, T., Sadeh, S., Rolotti, S. V., Blockus, H., Vancura, B., Negrean, A., Murray, A. J., Rózsa, B., Polleux, F., Clopath, C. & Losonczy, A. (2022), 'Local circuit amplification of spatial selectivity in the hippocampus.', *Nature* **601**, 105–109.

Geiller, T., Vancura, B., Terada, S., Troullinou, E., Chavlis, S., Tsagkatakis, G., Tsakalides, P., Ócsai, K., Poirazi, P., Rózsa, B. J. & Losonczy, A. (2020), 'Large-scale 3d two-photon imaging of molecularly identified ca1 interneuron dynamics in behaving mice.', *Neuron* **108**, 968–983.e9.

Gonzalez, K. C., Negrean, A., Liao, Z., Polleux, F. & Losonczy, A. (2023), 'Synaptic basis of behavioral timescale plasticity 1 2 3 4 5', *bioRxiv* .

URL: *https://doi.org/10.1101/2023.10.04.560848*

Grienberger, C. & Magee, J. C. (2022), 'Entorhinal cortex directs learning-related changes in ca1 representations.', *Nature* .

URL: *http://www.ncbi.nlm.nih.gov/pubmed/36323779*

Grosmark, A. D., Sparks, F. T., Davis, M. J. & Losonczy, A. (2021), 'Reactivation predicts the consolidation of unbiased long-term cognitive maps', *Nature Neuroscience* **24**, 1574–1585.

URL: *https://www.nature.com/articles/s41593-021-00920-7*

Hainmueller, T. & Bartos, M. (2018), 'Parallel emergence of stable and dynamic memory engrams in the hippocampus', *Nature* **558**(7709), 292–296.

URL: *https://doi.org/10.1038/s41586-018-0191-2*

Harvey, C. D., Coen, P. & Tank, D. W. (2012), 'Choice-specific sequences in parietal cortex during a virtual-navigation decision task', *Nature* **484**(7392), 62–68.

URL: *https://doi.org/10.1038/nature10918*

Heys, J. G., Rangarajan, K. V. & Dombeck, D. A. (2014), 'The functional micro-organization of grid cells revealed by cellular-resolution imaging', *Neuron* **84**, 1079–1090.

URL: *https://linkinghub.elsevier.com/retrieve/pii/S0896627314009684*

Jayakumar, R. P., Madhav, M. S., Savelli, F., Blair, H. T., Cowan, N. J. & Knierim, J. J. (2019), 'Recalibration of path integration in hippocampal place cells', *Nature* **566**, 533–537.

URL: *http://dx.doi.org/10.1038/s41586-019-0939-3 http://www.nature.com/articles/s41586-019-0939-3*

Jordan, J. T., McDermott, K. D., Frechou, M. A., Shtrahman, M. & Gonçalves, J. T. (2021), 'Treadmill-based task for assessing spatial memory in head-fixed mice.', *STAR protocols* **2**, 100770.

Kaifosh, P., Lovett-Barron, M., Turi, G. F., Reardon, T. R. & Losonczy, A. (2013), 'Septo-hippocampal GABAergic signaling across multiple modalities in awake mice', *Nature Neuroscience* **16**(9), 1182–1184.

URL: *https://doi.org/10.1038/nn.3482*

Kaifosh, P., Zaremba, J. D., Danielson, N. B. & Losonczy, A. (2014), 'Sima: Python software for

analysis of dynamic fluorescence imaging data', *Frontiers in Neuroinformatics* **8**. 793

    URL: *http://journal.frontiersin.org/article/10.3389/fninf.2014.00080/abstract* 794

Kaufman, A. M., Geiller, T. & Losonczy, A. (2020*a*), 'A role for the locus coeruleus in hippocampal 795
ca1 place cell reorganization during spatial reward learning', *Neuron* **105**(6), 1018–1026.e4. 796
    URL: *https://doi.org/10.1016/j.neuron.2019.12.029* 797

Kaufman, A. M., Geiller, T. & Losonczy, A. (2020*b*), 'A role for the locus coeruleus in hippocampal 798
CA1 place cell reorganization during spatial reward learning', *Neuron* **105**(6), 1018–1026.e4. 799
    URL: *https://doi.org/10.1016/j.neuron.2019.12.029* 800

Leutgeb, S., Leutgeb, J. K., Barnes, C. A., Moser, E. I., McNaughton, B. L. & Moser, M.-B. (2005), 801
'Independent codes for spatial and episodic memory in hippocampal neuronal ensembles', 802
Science **309**, 619–623. 803
    URL: *https://www.science.org/doi/10.1126/science.1114037* 804

Lovett-Barron, M., Kaifosh, P., Kheirbek, M. A., Danielson, N., Zaremba, J. D., Reardon, T. R., Turi, 805
G. F., Hen, R., Zemelman, B. V. & Losonczy, A. (2014), 'Dendritic inhibition in the hippocampus 806
supports fear learning', *Science* **343**(6173), 857–863. 807
    URL: *https://www.science.org/doi/abs/10.1126/science.1247485* 808

Mauk, M. D. & Buonomano, D. V. (2004), 'The neural basis of temporal processing', *Annual Review* 809
of Neuroscience **27**(1), 307–340. PMID: 15217335. 810
    URL: *https://doi.org/10.1146/annurev.neuro.27.070203.144247* 811

Muller, R. & Kubie, J. (1987), 'The effects of changes in the environment on the spatial firing of 812
hippocampal complex-spike cells', *The Journal of Neuroscience* **7**, 1951–1968. 813
    URL: *https://www.jneurosci.org/lookup/doi/10.1523/JNEUROSCI.07-07-01951.1987* 814

O'Hare, J. K., Gonzalez, K. C., Herrlinger, S. A., Hirabayashi, Y., Hewitt, V. L., Blockus, H., 815
Szoboszlay, M., Rolotti, S. V., Geiller, T. C., Negrean, A., Chelur, V., Polleux, F. & Losonczy, A. 816
(2022), 'Compartment-specific tuning of dendritic feature selectivity by intracellular ca2+ release.', 817
Science (New York, N.Y.) **375**, eabm1670. 818

O'Keefe, J. & Dostrovsky, J. (1971), 'The hippocampus as a spatial map. preliminary evidence from unit activity in the freely-moving rat', *Brain Research* **34**, 171–175.
URL: *https://linkinghub.elsevier.com/retrieve/pii/0006899371903581*

Pachitariu, M., Stringer, C., Dipoppa, M., Schröder, S., Rossi, L. F., Dalgleish, H., Carandini, M. & Harris, K. D. (2017), 'Suite2p: beyond 10,000 neurons with standard two-photon microscopy', *bioRxiv* .
URL: *https://doi.org/10.1101/061507*

Priestley, J. B., Bowler, J. C., Rolotti, S. V., Fusi, S. & Losonczy, A. (2022), 'Signatures of rapid plasticity in hippocampal ca1 representations during novel experiences', *Neuron* **110**(12), 1978–1992.e6.
URL: *https://doi.org/10.1016/j.neuron.2022.03.026*

Ravassard, P., Kees, A., Willers, B., Ho, D., Aharoni, D. A., Cushman, J., Aghajan, Z. M. & Mehta, M. R. (2013), 'Multisensory control of hippocampal spatiotemporal selectivity.', *Science (New York, N.Y.)* **340**, 1342–1346.

Rolotti, S. V., Ahmed, M. S., Szoboszlay, M., Geiller, T., Negrean, A., Blockus, H., Gonzalez, K. C., Sparks, F. T., Canales, A. S. S., Tuttman, A. L., Peterka, D. S., Zemelman, B. V., Polleux, F. & Losonczy, A. (2022b), 'Local feedback inhibition tightly controls rapid formation of hippocampal place fields', *Neuron* **110**, 783–794.e6.
URL: *https://linkinghub.elsevier.com/retrieve/pii/S089662732100996X*

Rolotti, S. V., Blockus, H., Sparks, F. T., Priestley, J. B. & Losonczy, A. (2022a), 'Reorganization of ca1 dendritic dynamics by hippocampal sharp-wave ripples during learning.', *Neuron* **110**, 977–991.e4.

Royer, S., Zemelman, B. V., Losonczy, A., Kim, J., Chance, F., Magee, J. C. & Buzsáki, G. (2012), 'Control of timing, rate and bursts of hippocampal place cells by dendritic and somatic inhibition', *Nature Neuroscience* **15**(5), 769–775.
URL: *https://doi.org/10.1038/nn.3077*

Sheffield, M. E., Adoff, M. D. & Dombeck, D. A. (2017), 'Increased prevalence of calcium transients

across the dendritic arbor during place field formation', *Neuron* **96**(2), 490–504.e5.

　URL: *https://www.sciencedirect.com/science/article/pii/S0896627317308735*

Sheintuch, L., Rubin, A., Brande-Eilat, N., Geva, N., Sadeh, N., Pinchasof, O. & Ziv, Y. (2017), 'Tracking the same neurons across multiple days in ca2+ imaging data.', *Cell reports* **21**, 1102–1115.

Skaggs, W. E. & Mcnaughton, B. L. (1998), 'Spatial firing properties of hippocampal ca1 populations in an environment containing two visually identical regions', *The Journal of neuroscience* **18**, 8455–66.

Terada, S., Geiller, T., Liao, Z., O'Hare, J., Vancura, B. & Losonczy, A. (2022), 'Adaptive stimulus selection for consolidation in the hippocampus', *Nature* **601**(7892), 240–244.

　URL: *https://doi.org/10.1038/s41586-021-04118-6*

Tuncdemir, S. N., Grosmark, A. D., Chung, H., Luna, V. M., Lacefield, C. O., Losonczy, A. & Hen, R. (2023), 'Adult-born granule cells facilitate remapping of spatial and non-spatial representations in the dentate gyrus.', *Neuron* .

Tuncdemir, S. N., Grosmark, A. D., Turi, G. F., Shank, A., Bowler, J. C., Ordek, G., Losonczy, A., Hen, R. & Lacefield, C. O. (2022), 'Parallel processing of sensory cue and spatial information in the dentate gyrus', *Cell Reports* **38**.

Turi, G. F., Li, W. K., Chavlis, S., Pandi, I., O'Hare, J., Priestley, J. B., Grosmark, A. D., Liao, Z., Ladow, M., Zhang, J. F., Zemelman, B. V., Poirazi, P. & Losonczy, A. (2019), 'Vasoactive intestinal polypeptide-expressing interneurons in the hippocampus support goal-oriented spatial learning', *Neuron* **101**, 1150–1165.e8.

Vancura, B., Geiller, T., Grosmark, A., Zhao, V. & Losonczy, A. (2023), 'Inhibitory control of sharp-wave ripple duration during learning in hippocampal recurrent networks.', *Nature neuroscience* **26**, 788–797.

Warren, R. A., Zhang, Q., Hoffman, J. R., Li, E. Y., Hong, Y. K., Bruno, R. M. & Sawtell, N. B. (2021), 'A rapid whisker-based decision underlying skilled locomotion in mice', *eLife* **10**, e63596.

　URL: *https://doi.org/10.7554/eLife.63596*

Zaremba, J. D., Diamantopoulou, A., Danielson, N. B., Grosmark, A. D., Kaifosh, P. W., Bowler, [873]
J. C., Liao, Z., Sparks, F. T., Gogos, J. A. & Losonczy, A. (2017), 'Impaired hippocampal place cell [874]
dynamics in a mouse model of the 22q11.2 deletion', *Nature Neuroscience* **20**(11), 1612–1623. [875]
URL: *https://doi.org/10.1038/nn.4634* [876]

Zhang, K., Ginzburg, I., McNaughton, B. L. & Sejnowski, T. J. (1998), 'Interpreting neuronal [877]
population activity by reconstruction: unified framework with application to hippocampal place [878]
cells.', *Journal of neurophysiology* **79**, 1017–44. [879]

10 Supplementary Materials

10.1 Table of Contents

- **Link to website for quickstart guides and file downloads:**

  https://www.losonczylab.org/behaviormate

- **Link to example settings files:**

  https://github.com/losonczylab/behaviorMate/tree/main/example_settings_files

- **Link to source code for the UI:**

  https://github.com/losonczylab/behaviorMate

- **Link to online java documentation for the UI:**

  https://www.losonczylab.org/behaviorMate-1.0.0

- **Link to electronics schematics:**

  https://github.com/losonczylab/Hardware/tree/main/Electronics

- **Link to treadmill and VR CAD files:**

  https://github.com/losonczylab/Hardware/tree/main

- **Figure S1: Additional details on VR Frame**

- **Figure S2: Experimental Validation**

- **Table ST1: ANOVA results**

Figure S1. Alternative VR system (**A**) A version of the VR system that can better fit in tight spaces. Can be made to be self-contained in the event it needs to be moved or swapped with other systems. (**B**) Fixes the position of the mouse's head by tightly clamping onto each side of the bar implanted into its skull. The screws used to close the two clamps are not shown.

Figure S2. Additional place cell properties (A) Proportion of place cells in VR and TM. VR had significantly less ROIs that were classified as a place cell compared to TM (two sample z-test, F=14.96 p=1.27e^{-50}). **(B)** Proportion of place cells in each context-environment pair. **(C)** Mean velocity per lap in each context-environment pair. Significant main effect of environment (F(1,5)=35.95, p=0.0019). **(D)** Peak location of place cells recorded in each context-environment pair. Note that place cells tile entire track in all pairs.

Table ST1. ANOVA Results

△ **F/F Amplitude** (Fig. 5C)	**F value**	**DOF**	**PR > F**
Context (familiar vs. novel)	0.0037	(1, 5)	0.9537
Environment (VR vs Treadmill)	1.9387	(1, 5)	0.2226
Interaction	1.4364	(1, 5)	0.2844

△ **F/F Frequency** (Fig. 5D)	**F value**	**DOF**	**PR > F**
Context (familiar vs. novel)	0.0789	(1, 5)	0.7901
Environment (VR vs Treadmill)	4.3798	(1, 5)	0.0906
Interaction	0.2100	(1, 5)	0.6660

Sensitivity (Fig. 5E)	**F value**	**DOF**	**PR > F**
Context (familiar vs. novel)	3.3375	(1, 5)	0.1273
Environment (VR vs Treadmill)	9.8469	(1, 5)	0.0257
Interaction	4.8578	(1, 5)	0.0787

Specificity (Fig. 5F)	**F value**	**DOF**	**PR > F**
Context (familiar vs. novel)	2.5521	(1, 5)	0.1710
Environment (VR vs Treadmill)	53.3198	(1, 5)	0.0008
Interaction	3.7118	(1, 5)	0.1120

Spatial Information (Fig. 5G)	**F value**	**DOF**	**PR > F**
Context (familiar vs. novel)	3.7945	(1, 5)	0.1090
Environment (VR vs Treadmill)	0.7352	(1, 5)	0.4304
Interaction	2.3396	(1, 5)	0.1867

Velocity (Fig. S2C)	**F value**	**DOF**	**PR > F**
Context (familiar vs. novel)	0.1262	(1, 5)	0.7369
Environment (VR vs Treadmill)	35.9548	(1, 5)	0.0019
Interaction	2.7198	(1, 5)	0.1600