Systematic benchmarking of imaging spatial transcriptomics platforms in FFPE tissues

Huan Wang[1,*], Ruixu Huang[2,*], Jack Nelson[1,*], Ce Gao[3], Miles Trans[3], Anna Yeaton[4], Kristen

Felt[5], Kathleen L. Pfaff[6], Teri Bowman[7], Scott J. Rodig[6,7], Kevin Wei[,3,7], Brittany A. Goods[2,**],

Samouil L. Farhi[1,**]

Affiliations

[1]Spatial Technology Platform, Broad Institute of MIT and Harvard, Cambridge, MA 02142

USA.

[2]Thayer School of Engineering, Molecular and Systems Biology, and Program in Quantitative

Biomedical Sciences at Dartmouth College, Hanover, NH 03755, USA.

[3]Division of Rheumatology, Inflammation, and Immunity, Brigham and Women's Hospital and

Harvard Medical School, Boston, MA 02215, USA.

[4]Present affiliation: Immunai, New York, NY 10016, USA.

[5]ImmunoProfile, Brigham & Women's Hospital and Dana-Farber Cancer Institute, Boston, MA

02215, USA.

[6]Center for Immuno-Oncology, Tissue Biomarker Laboratory, Dana-Farber Cancer Institute,

Boston, MA 02215, USA.

[7]Department of Pathology, Brigham and Women's Hospital, Boston, MA 02215, USA.

[*] These authors contributed equally to this work

[**] These authors jointly supervised this work

Corresponding author: Samouil L. Farhi: sfarhi@broadinstitute.org

Abstract

Emerging imaging spatial transcriptomics (iST) platforms and coupled analytical

methods can recover cell-to-cell interactions, groups of spatially covarying genes, and gene

signatures associated with pathological features, and are thus particularly well-suited for

applications in formalin fixed paraffin embedded (FFPE) tissues. Here, we benchmarked the

performance of three commercial iST platforms on serial sections from tissue microarrays

(TMAs) containing 23 tumor and normal tissue types for both relative technical and biological

performance. On matched genes, we found that 10X Xenium shows higher transcript counts per

gene without sacrificing specificity, but that all three platforms concord to orthogonal RNA-seq

datasets and can perform spatially resolved cell typing, albeit with different false discovery rates,

cell segmentation error frequencies, and with varying degrees of sub-clustering for downstream

biological analyses. Taken together, our analyses provide a comprehensive benchmark to guide

the choice of iST method as researchers design studies with precious samples in this rapidly

evolving field.

MAIN

Spatial transcriptomics (ST) tools measure the gene expression profiles of tissues or cells *in situ*. These approaches overcome the limitations of single-cell RNA-sequencing (scRNA-seq) methods by negating the need for cellularization and maintaining both local and global spatial relationships between cells within a tissue. ST can thus recover cell-cell interactions with high confidence, groups of spatially covarying genes, groups of cells predictive of cancer survival, and gene signatures associated with pathological features [1, 2]. These advantages, coupled with rapidly emerging computational and analytical methods, have led to substantial excitement about deploying these platforms in fundamental biology studies, and in the clinic for research and diagnostic purposes [3, 4, 5].

ST tools can be split into two broad categories: sequencing (sST) and imaging (iST) based modalities. sST methods tag transcripts with an oligonucleotide address indicating spatial location, most commonly by placing tissue slices on a barcoded substrate; isolating tagged mRNA for next-generation sequencing; and computationally mapping transcript identities to locations [6]. In contrast, iST methods most commonly use variations of fluorescence *in situ* hybridization (FISH) where mRNA molecules are tagged with hybridization probes which are detected in a combinatorial manner over multiple rounds of staining with fluorescent reporters, imaging, and de-staining **(Fig. 1a)** [7]**.** Computational reconstruction then yields maps of transcript identity with single-molecule resolution. Compared to sST methods, iST methods are targeted to subsets of the transcriptome due to their reliance on pre-defined gene panels and they adopt the higher spatial resolution and sensitivity of FISH, yielding single-cell resolution data [8].

While the iST methods share some similarities, significant differences arise in primary signal detection and amplification, sample processing, and the subsequent fluorescent cycling

chemistry (**Fig. 1b**) [9,10,11]. The need for amplification of signal is coupled to the sample processing,

namely whether the sample is cleared, gel-embedded, or photobleached to quench

autofluorescence. There are tradeoffs due to differences in sample processing for each iST method.

For example, clearing of the sample increases signal quality but can prevent follow-up H&E

staining and complicate immunostaining, which, in turn, can make cell segmentation more

challenging. Finally, there are tradeoffs between imaging time, molecular plex, and imaging area

covered, which result from the particular combination of the molecular protocol and the imaging

hardware implementation [12].

A key historic limitation in the widespread use of iST methods with human clinical samples

was the incompatibility of most methods with formalin-fixed, paraffin-embedded (FFPE) tissue

samples [13, 14]. FFPE is the standard format for clinical sample preservation for pathology due to

its ability to maintain tissue morphology and sample stability at room temperature for decades [15].

The ability to process FFPE samples with iST would enable the use of archival tissue banks for

studies and obviate the need for specialized sample harvesting workflows. However, FFPE

samples tend to suffer from decreased RNA integrity, particularly after having been stored in

archives for extended periods of time [16].

Three companies recently released the first FFPE compatible commercial iST platforms:

10X's Xenium, Nanostring's CosMx, and Vizgen's MERSCOPE [9,10,11,17]. These three platforms

each use different protocols, probe designs, signal amplification strategies, and computational

processing methods, and therefore may potentially yield different sensitivities and downstream

results. The main chemistry difference lies in transcript amplification: 10X Xenium uses a small

number of padlock probes with rolling circle amplification; CosMx uses a low number of probes

amplified with branch chain hybridization; and MERSCOPE uses direct probe hybridization but

amplifies by tiling the transcript with many probes (**Fig. 1b**). However, no head-to-head

performance comparisons on matched samples have been published. Understanding the key

differences across platforms will allow users to make better-informed decisions regarding panel

design, method choice, and sample selection as they design costly experiments, often on precious

samples that have been bio-banked for years [18].

In this study, we compared currently available FFPE-compatible iST platforms on matched

tissue samples. We prepared a set of samples representative of typical archival FFPE tissues,

comprised of 23 different tissue types, and acquired matched data from sequential sections

according to the manufacturer's best practices at the time of writing, generating a dataset of >3.3M

cells. We analyzed the relative sensitivity and specificity of each method on shared transcripts,

and further quantified the concordance of the iST data across each platform with orthogonal data

sets from The Cancer Genome Atlas (TCGA) program and Genotype-Tissue Expression (GTEx)

databases [19,20]. Then we focused on cell-level comparisons, evaluating the out-of-the-box

segmentation for each platform based on detected genes and transcripts and coexpression patterns

of known disjoint markers. Finally, we cross-compared the ability of each platform to identify cell

type clusters with breast and breast cancer tissues as an example use case. Taken together, our

work provides the first head-to-head comparison of these platforms across multiple archival

healthy and cancerous FFPE tissue types.

RESULTS

Collection of matched iST data across 23 FFPE tissue types reveals high transcript counts obtained by Xenium and CosMx.

To test the performance of the latest generation of FFPE-compatible iST tools, we measured the spatial expression of the same genes on the same samples as much as possible given current panel configurations. To accomplish this, we used two previously generated multi-tissue tissue microarrays (TMAs) from clinical discarded tissue (**see Methods**). We focused on FFPE tissues as the standard method for sample processing and archival in pathology. One TMA consisted of one hundred and seventy-three 0.6 mm diameter cores (i.e. sampled regions) from seven different cancer types, with 3-6 patients per cancer type, and 3-6 cores per patient (**Fig. 1c,d, Supplementary Table 1**). A separate TMA consisted of forty-eight 1.2 mm diameter cores spanning sixteen normal tissue types isolated with each tissue type coming from one patient and represented in 2-3 cores (**Fig. 1c,d, Supplementary Table 2**). CosMx and Xenium suggest pre-screening samples based on H&E, while MERSCOPE recommends a $DV_{200} > 60\%$. Since our goal was to determine the compatibility of iST platforms under typical workflows for biobanked FFPE tissues, and since TMAs are challenging to assay by DV_{200}, samples were not prescreened based on RNA integrity. Samples were screened by H&E in the process of TMA assembly. Both TMAs were sliced into serial sections for processing by 10X Xenium, Vizgen MERSCOPE, and Nanostring CosMx, following manufacturer instructions (**see Methods**).

The three different iST platforms offer different degrees of customizability and panel compositions. In terms of panel design, MERSCOPE and Xenium offer either fully customizable panels or standard panels with optional add-on genes, while CosMx offers a standard 1K (substantially larger plex) panel with optional add-on genes. We opted to run the CosMx 1K panel

as available commercially, as well as the Xenium human breast, lung, and multi-tissue off-the-

shelf panels. We then designed two MERSCOPE panels to match the pre-made Xenium breast and

lung panels, by filtering out any genes which could potentially lead to high expression flags in any

tissue in the Vizgen online portal. This resulted in a total of six panels, with each panel overlapping

the others on >94 genes (**Fig. 1e, Supplementary Table 3**). Samples were run following

manufacturer instructions over the course of 74 days after slicing (**Fig. 1f, Supplementary Table**

4), with efforts made to ensure that head-to-head comparisons were available at similar time points

for each pair of platforms. In data review, we noticed that MERSCOPE breast and lung panel were

originally acquired with a 5 µm imaging depth, which was unintentionally thinner than the

manufacturer recommendation of 10 µm, and could thus lead to aberrantly low counts. Thus, a

second round of breast panel acquisition was performed with a 10 µm imaging depth

(**Supplementary Table 1a**), resulting in a median 3.0-fold increase in expression across all

transcripts. We excluded the 5 µm MERSCOPE breast panel data from all further comparisons but

left the lung panel data in as an illustrative example of an unsuccessful run. However, we

emphasize that MERSCOPE performance should be judged based on the rerun breast panel.

Each data set was processed according to the standard base-calling and segmentation

pipeline provided by each manufacturer. The resulting count matrices and detected transcripts were

then subsampled and aggregated to individual cores of the TMA (**Methods**). Across all datasets

we generated >190 million transcripts, >3.3 million cells, across 7 tumor types, and 16 normal

tissue types. Overall, we found that the cores from each TMA were generally well adhered to the

tissue and detected transcripts, and we were able to collect data from all three modalities for 217

cores (**Supplementary Table 4**). The total number of transcripts recovered for each run was

highest for Xenium, followed closely by CosMx, and then MERSCOPE (**Supplementary Table**

4). The total number of cells initially reported was highest for Xenium followed by MERSCOPE

and CosMx (**Supplementary Table 4**). Based on the initially reported number of transcripts, the

tumor TMA appeared to provide more counts than the normal tissue TMA, which we ascribed to

a higher tissue quality in the tumor samples (**Supplementary Table 4**). We note that the total

number of transcripts from the MERSCOPE normal TMA run was below what would be typically

thought of as a successful run, even when rerunning with the breast panel at 10 μm imaging depth.

Such a sample would normally be excluded from analysis, but we continued the data through to

illustrate how low transcript capture affects downstream results.

Figure 1: Experimental design and iST platforms. (a) Overall approach for generating iST data.
(b) Different amplification approaches for Xenium, MERSCOPE, and CosMx. (c) Overview of

the tissue types and numbers of cores used in this study. BlC = bladder cancer, BrC = breast cancer,
CRC = colorectal cancer, HNSCC = head and neck squamous cell carcinoma, Mel = Melanoma,
NSCLC = non-small cell lung cancer, OvC = ovarian cancer. (d) DAPI images from the Xenium
run of each TMA, including tumors (top) and normal tissues (bottom) (e) The number of common
target genes in each panel used in this study. (f) Overall timeline of the imaging days for each
study. Day = 0 corresponds to the day of slicing. † denotes the MERSCOPE breast and lung panels
acquired with a 5 µm imaging thickness, thinner than manufacturer instructions.
166

10X Xenium shows higher transcript counts per gene without sacrificing specificity

We next sought to directly compare the performance of each iST platform on matched

genes. We began with a pseudo-bulk-based approach at the core level since this would not depend

on differences in cell segmentation performance (see **Methods**) [21].

First, we examined the run-to-run reproducibility within a single platform for Xenium and

MERSCOPE, finding that the total transcript count of all shared genes was highly correlated across

data sets acquired with different panels, regardless of the tissue of origin (**Supplementary Fig.**

1a). We also examined the pseudo-bulk gene expression correlation for cores from the same patient

in the same dataset and found that correlation was high (Pearson's r => 0.7) in almost all cases

(**Supplementary Fig. 1b-c**), indicating good sample-to-sample reproducibility within a given

platform.

To evaluate the relative sensitivity of each platform, we plotted the total transcript counts

of every shared gene between all combinations of platform and panel, summed across all matched

cores. We found generally linear relationships between all pairs of platforms (**Fig. 2a-c,**

Supplementary Fig. 2). Xenium consistently showed higher expression levels on the same genes

than CosMx in the tumor TMA, with the Xenium breast having 14.6-fold more counts than the

CosMx multi-tissue data sets (**Fig. 2a**). The Xenium multi-tissue panel data showed a slightly

smaller difference, with 12.3-fold higher expression on the same genes (**Fig. 2a**), while the lung

panel, which was acquired closest in time following slicing, also displayed a median of 14.0-fold

higher expression **(Fig. 2a)**. MERSCOPE showed higher expression levels than CosMx when

using the breast (10 µm) panel (median of 5.4-fold higher), and comparable expression levels even

when using the lung (5 µm) panel (median of 1.1-fold) **(Fig. 2b)**. Finally, Xenium showed 2.6-

fold higher median expression with the breast panels (10 µm) than MERSCOPE, and 13.6-fold

higher median expression with the lung panels (5 µm) **(Fig. 2c)**. In the normal tissue TMA, we

found that results were generally consistent, except that the MERSCOPE breast panel showed

decreased transcript counts relative to the same panel in the tumor TMA (**Supplementary Fig. 2b-**

c), which is consistent with this TMA being unsuccessful for MERSCOPE. Considering the overall

higher transcripts per cell across platforms for the tumor TMA **(Supplementary Table 4)**, this

suggests that the ability to detect transcripts falls off more strongly with sample quality with

MERSCOPE, altering the performance relative to CosMx but not Xenium. Examining the CosMx

as compared to Xenium data also revealed an upward curve in the lower expression regime

indicative of higher-than-expected calls associated with the low expression regime by CosMx

(**Supplementary Fig. 2a**).

We next wanted to assess the specificity of each platform. Each of the three platforms

includes negative controls which are used to evaluate sample quality [22, 23]. Xenium and CosMx

include both negative probes (e.g. real probes targeting nucleic acids that are not present in human

tissue) and negative barcodes (e.g. algorithmically allowable barcodes that are not associated with

any probe in the experimental panel). MERSCOPE includes only negative barcodes by default. To

determine specificity, we first calculated the fraction of negative barcodes and probes relative to

the number of transcripts for each tissue type **(Fig. 2e).** We found that MERSCOPE and Xenium

consistently showed the highest on-target fraction, while CosMx was lower across each tissue type

(**Fig. 2e**). However, this measurement is biased because of the relative numbers of controls and

target barcodes. We therefore also adopted a false discovery rate (FDR) calculation which

normalizes for these differences and is calculated against both the negative probes and negative

barcodes (see **Methods, Fig. 2f-g**). We found that Xenium consistently showed the lowest FDR

while CosMx showed the highest FDR regardless of whether we standardized to negative control

barcodes or probes. This finding is consistent with the upswing in the gene-gene expression plots

in **Fig 2a** and **Supplementary Fig. 2a**—as both indicate a higher FDR at the low end of the gene

expression range. These results are consistent when visualized across panel **(Fig. 2e-g)**.

Finally, we used the negative control barcodes to evaluate the number of genes reliably

detected by each platform in each tissue type. For each core, we calculated the number of genes

that were detected two standard deviations above the average expression of the negative control

probes. These numbers were then averaged for cores of the same tissue type. Because the CosMx

panel was almost three times larger, it yielded a larger absolute number of detected genes in 14

out of 20 tissue types while the Xenium breast panel was higher in the remaining 6 tissue types

(**Fig. 2h**, **Supplementary Table 5**). However, Xenium consistently detected the highest fraction

of genes in a panel, followed by MERSCOPE and CosMx (**Fig. 2i**).

Figure 2: Technical performance comparison of iST platforms grouped by tissue types. (a) Scatter plots of summed gene expression levels (natural log transformed) of every shared gene between Xenium (breast/lung) and CosMx (1k) data, captured from matched tumor TMA cores. Each data point corresponds to a gene. **(b)** Same as (a) but between MERSCOPE (breast/lung) and CosMx(1k). **(c)** Same as (a) but between Xenium(breast/lung) and MERSCOPE(breast/lung). **(d)** Same as (a) but between Xenium(multi-tissue) and CosMx(1k). **(e)** Bar plot of percentage of all transcripts corresponding to genes relative to the total number of calls (including negative control probes and unused barcodes) averaged across cores of the same tissue type. Results are presented by panel including breast, lung, and multi-tissue panels from Xenium; breast and lung panels from MERSCOPE; and multi-tissue 1k panel from CosMx. **(f)** Bar plot of false discovery rate (FDR) where FDR(%) = (blank barcode calls / total transcript calls) x (Number of panel genes /Number of blank barcode) x 100. FDR values were log_{10} transformed to better show the differences between panels. **(g)** Same as (f) but using negative control probes to replace blank barcodes. MERSCOPE is missing in this bar plot as it does not have negative control probe by design. **(h)** Bar plot of number of genes detected above noise, estimated as two standard deviations above average of the negative control probes. **(i)** Same as (h) but normalized to the number of genes in a panel. † denotes the MERSCOPE lung panel acquired with a 5 μm imaging thickness.

iST platforms are all concordant with orthogonal RNA-seq data sets

In the absence of ground truth, it is difficult to evaluate whether a higher number of expressed genes is representative of increased sensitivity to real biology or increased false positive rates. We thus evaluated the correlation of iST data to reference RNA-seq data. We first aggregated pseudo-bulk normal tissue TMA results from all panels of the three platforms and compared them to data from the TCGA program **(see Methods)** [11]. We observed similar correlation coefficients across all gene panels relative to pseudo-bulk RNA-seq expression data **(Fig. 3a-b, Supplementary Table 8)**. However, notably, the CosMx data showed a characteristic upswing in the low expression regime, similar to that observed when plotting gene-by-gene expression against MERSCOPE and Xenium **(Fig. 2, Supplementary Fig. 2)**.

We also compared the pseudo-bulk results from the normal tissue TMA with bulk RNA-seq data obtained from GTEx[12] The Xenium breast, Xenium multi-tissue, and CosMx data sets showed similar correlations to breast data obtained from GTEx, while the MERSCOPE had

significantly lower correlation, consistent with a run which doesn't pass QC (Pearson's r of 0.36

vs 0.71, 0.68, and 0.69, respectively, **Fig. 3c**). Similar trends were observed in the lung data, with

MERSCOPE lung (5 µm) showing the lowest correlation while the other three data sets showed

higher correlations to GTEx data (**Fig. 3d**). These relative trends remained true across most normal

tissue types, though we found that thyroid, pancreas, and lymph nodes showed the lowest

correlations across all panels while prostate, tonsil, and liver showed the highest correlations

(**Supplementary Table 9)**. Overall, our comparison to TCGA and GTEx data suggests that while

some platforms may be more highly correlated to reference datasets in some cases, all are within

a similar correlation regime regardless of tissue type.

We next wanted to determine how the expression of tissue-specific transcript markers

varied across each platform. To accomplish this, we curated tissue markers that are unique to each

tissue type by selecting genes whose expression in a single tissue exceeds 20 times the sum of

other tissues from the GTEx database (see **Methods**). We found tissue-specific expression patterns

of several of these markers across all selected panels when visualized across each healthy tissue

type (**Fig. 3e**). MERSCOPE consistently showed expression of tissue-specific markers in multiple

tissue types, consistent with unsuccessful normal TMA data acquisition. However, some canonical

markers were not enriched in certain tissues. For example, we found that ABCC11 was highly

expressed in the breast tissue in the Xenium and MERSCOPE datasets, but not in CosMx. In fact,

although CosMx showed satisfying expression patterns for some tissue markers, many canonical

markers are not enriched in the expected tissues, possibly due to the high false discovery rate **(Fig.**

3f-g). Across marker genes, Xenium data had a distinct expression pattern in all tissues, whereas

CosMx and MERSCOPE showed a less distinct pattern in many tissue types.

Figure 3: Concordance of iST data with reference RNA-seq datasets. (a) Scatter plots of overlapping genes, showing the averaged expression of a gene across breast cancer cores profiled by the indicated panel, normalized to 100,000 vs the average FPKM from TCGA for all samples

of a matched tissue type (BRCA). **(b)** Same as (a) but for lung cancer cores plotted vs averaged
LUAD and LUSC samples from TCGA. **(c)** Same as (a) but showing breast cores vs averaged
nTPM values from GTEx breast samples. **(d)** Same as (a) but for lung cores and samples. † denotes
the MERSCOPE lung data acquired with a 5-µm imaging depth on FFPE sample. ‡ denotes the
normal tissue TMA data of MERSCOPE which failed initial QC. **(e)** Heatmap of Z-scored average
gene expression for several canonical marker genes in the indicated tissue cores for the Xenium
multi-tissue panel (left) and CosMx 1K panel (right).
289
Out of the box segmentation and filtration can yield cells with comparable numbers of

detected transcripts and genes from each platform.

Next, we compared the performance of each iST method on a single-cell level. The three

platforms generate cell boundaries based on a DAPI image alone (Xenium) or a DAPI image

combined with a membrane marker (CosMx and MERSCOPE). When we visually examined the

segmentation outputs, Xenium data showed cell boundaries that appeared to include large regions

of non-cellular space, in contrast to MERSCOPE and CosMx which tightly followed the visualized

cell nucleus (**Fig. 4a**). When transcripts were overlaid with these segmentation boundaries,

Xenium cell boundaries fell between regions of transcripts and thus most transcripts were assigned

to cells. MERSCOPE and CosMx's tighter nuclei removed more transcripts, though those that

remain appeared more confidently assigned to cells. Overall, when normalized to the imaged tissue

area, Xenium and CosMx identified the most putative cells, followed by MERSCOPE (**Fig. 4b,**

Supplementary Table 6). In line with the visual inspection, Xenium cells were consistently larger,

regardless of data set or panel, followed by CosMx and finally MERSCOPE (**Fig. 4c,**

Supplementary Table 7).

We filtered out empty regions of space and cells without any transcripts for downstream

examination and quantified the fraction of cells containing differing numbers of transcripts per cell

(Fig. 4d). We chose a permissive threshold of removing cells with fewer than 10 transcripts for

Xenium and MERSCOPE, and 20 transcripts for CosMx from downstream analysis. [11, 24, 25]. The

tumor TMA consistently had a greater fraction of cells passing filtration, with Xenium retaining

the most cells (97.43% breast, 97.10% multi-tissue, 95.08% lung) followed by CosMx (83.41%)

and MERSCOPE (68.46% breast (10 µm), 25.77% lung (5 µm) (**Supplementary Table 3**). The

normal tissue TMA had overall lower cell retention performance, but the relative performance of

the platforms based on the fractions of cells remained the same. Notably, while CosMx and

Xenium still retained > 77% of the cells, MERSCOPE data of the normal tissue TMA had <3% of

cells retained and was thus not used in downstream analysis. Unsurprisingly, filtration decreased

the number of retained cells per unit area for all platforms, with the smallest decrease coming for

CosMx. The cells retained from CosMx had similar areas, while filtration of the Xenium and

MERSCOPE data sets resulted in a higher average cell area (**Fig. 4c**).

After filtration, we compared the number of transcripts and the number of unique genes

per retained cell across all tissues and all panels, focusing only on cores that were sampled by all

three platforms (**Fig. 4e-f**). Xenium breast panel gave the highest numbers of transcripts per cell

in most tissue types, 17 out of 22. The CosMx data showed the highest numbers of transcripts in

heart, lymph node, spleen, thyroid, and tonsil; and comparable transcript counts in breast cancer,

ovary, and ovarian cancer to the Xenium breast panel. The MERSCOPE data generally had the

lowest number of transcripts per cell, though bladder cancer and breast cancer measured with the

MERSCOPE breast panels approached the results from Xenium, and the bladder cancer and skin

data sets had higher transcripts per cell than CosMx. As expected given its larger panel size,

CosMx found many unique genes per cell, showing the largest numbers in 9 tissue types: breast

cancer, colon, heart, lymph node, ovary, ovarian cancer, spleen, thyroid, and tonsil; while Xenium

breast panel found the most unique genes per cell in 12 tissue types: bladder cancer, bladder, breast,

CRC, HNSCC, kidney, lung, melanoma, NSCLC, pancreas, prostate, and skin (**Fig. 4f**).   If these

analyses were restricted to only the shared genes across all panels, numbers were much lower, but

Xenium showed higher expression levels and unique numbers of genes than either CosMx or

MERSCOPE (**Supplementary Fig 3c-d**).

We then wanted to determine how effectively different iST platforms' segmentation

algorithms perform. We examined the co-expression of *CD19*, a canonical B-cell marker, and

CD3e, a canonical T-cell marker across all filtered cells; the co-expression of *CD8* and *CD4*,

markers of T-cell subsets; and the co-expression of *CD3e* and *EPCAM*, a marker for epithelial

cancer cells [26, 27]. All these marker gene pairs are disjointly expressed, and a well-performing

segmentation algorithm should yield few cells expressing both markers. We pooled all the filtered

cells from matched cores and all available panels of each platform and plotted the expression of

one gene against the other and converted the scatter plot to a heatmap to show cell fractions. We

found that Xenium—despite its less visually accurate cell boundaries—and MERSCOPE, showed

clear patterns of disjoint expression, separating cells from different lineages, while CosMx showed

such a pattern for *EPCAM* vs *CD3e* but not for the other two pairs (**Fig. 4g**). Given the low counts

of the immune genes, it was difficult to determine if these were false positive calls or segmentation

errors. Nevertheless, since the CosMx panel is much higher plex, and retained similar numbers of

transcripts and genes to Xenium, we wondered how these two methods performed in terms of cell

type recovery.

Figure 4: Comparison of cell segmentation results from each iST platform. (a) Top row: DAPI image overlaid with cell segmentation boundaries (subset). Middle row: all the transcripts in green dots, white lines for the cell boundaries, and *EPCAM* in blue dots. Bottom row: segmented cell boundaries before and after filtration. **(b)** Violin plot of segmented cells per unit area before (left half) and after filtration (right half) grouped by panel with tumor and normal TMA data combined. **(c)** Same as (b) but showing cell areas before and after filtration. **(d)** Line plot showing remaining cells in percentage after filtering with various thresholds (transcripts per cell). Dotted lines indicate selected thresholds: 10 transcripts or above for Xenium and MERSCOPE and 20 for CosMx. **(e)** Heatmap of transcripts per cell after filtration. All available genes are considered here for each

360 panel. MERSCOPE lung panel (5 μm) excluded from this heatmap. **(f)** Same as (e) but showing
361 unique genes per cell. **(g)** Co-expression density map for three pairs of disjoint genes (rows) from
362 all three platforms (columns). All cells across all tissues which include at least one detected
363 transcript of either of the indicated genes are plotted together, with color indicating the number of
364 cells at the indicated expression levels of each gene.
365

Clustering analyses reveal differences in cell type recovery across platforms

In a typical iST workflow, a key step is reducing the dimensionality of the data by

identifying cell types, their unique states, and their expression patterns for further analysis

leveraging spatial information[28]. To compare across platforms, we clustered the data from the

filtered cells from all the cores for each TMA with a focus on breast tissues. The initial clustering

of TMAs from datasets (except MERSCOPE normal tissue) showed expected batch effects caused

by patients and tissue types with broadly similar cluster arrangements around morphological tissue

features (**Supplementary Fig. 4a-d**). We removed batch effects (**see Methods**) and then

performed targeted clustering and cell type annotation for breast samples from the CosMx and

Xenium breast datasets; lung samples from the CosMx and Xenium lung datasets; and breast

cancer from the CosMx, MERSCOPE breast, and Xenium breast datasets.

In breast samples, we were able to identify nine cell types, including all known major cell

types, (adipocytes, alveolar cells, B cells, basal cells, fibroblast cells, hormone-sensing cells,

myeloid, T cells, vascular & lymphatic cells) from the Xenium data, using previously established

markers (**Fig. 5a, Supplementary Fig. 5**) [29,30,31]. In the CosMx data, we were only able to identify

six cell types, including several major cell types, but failed to recognize cell subtypes (B cells,

basal cells, fibroblast, hormone-sensing cells, immune cells, and vascular & lymphatic cells) (**Fig.**

5a, Supplementary Fig. 5). A high gene-to-gene correlation was found between all overlapping

cell types between Xenium and CosMx (**Fig. 5a**). Similarly, in the lung samples, we were able to

identify nine cell types (alveolar epithelial type 1 cell, alveolar epithelial type 2 cell, endothelial

capillary cells, endothelial cells, fibroblasts, immune cells, macrophages, mast cells, and stroma

cells) in the Xenium lung panel, successfully covering all known major cell types (**Fig. 5b,**

Supplementary Fig. 5) [32,33]. Four of the six major clusters were identified and annotated in the

lung samples from the CosMx data (endothelial cells, epithelial cells, immune cells, and

macrophages), while the other two clusters remained difficult to annotate due to the non-traditional

enriched gene markers (**Fig. 5b, Supplementary Fig. 5**). Correlation heatmaps show a strong

correlation between the two macrophage clusters identified in Xenium and CosMx (**Fig. 5b**). The

epithelial cell cluster from CosMx correlates strongly with alveolar epithelial type 1 cell from

Xenium and the endothelial cell cluster from CosMx correlates with endothelial capillary cell

cluster from Xenium (**Fig. 5b**).

Finally, in breast cancer, after batch effect removal (**Supplementary Fig. 5d-f**), Xenium

resulted in nine cell types (alveolar cells, B cells, basal cells, fibroblast, hormone-sensing cells,

immune cells, myeloid, T cells, and vascular & lymphatic cells) (**Fig. 5c, Supplementary Fig. 5**)

[34,35,36]. On the other hand, CosMx resulted in eight cell types (alveolar cells, basal cells, epithelial

cells, fibroblast cells, hormone-sensing cells, immune cells, myeloid, and vascular & lymphatic

cells). MERSCOPE resulted in six cell types, including alveolar cells, fibroblast cells, hormone-

sensing cells, myeloid cells, T cells, and vascular & lymphatic cells. The cell type annotation of

Xenium and CosMx is comparable in terms of both transcriptomic profile and subtype depth, with

CosMx only unable to annotate immune cell subtypes (B cell and T cell). Gene expression of the

same cell type from both platforms correlated well (**Fig. 5c, Supplementary 5**). The cell type

annotation of CosMx, however, was especially difficult compared to Xenium because of its

atypical gene markers shown for each cluster in the heatmaps (**Supplementary Fig. 5**) and low

expression of transcripts from canonical markers (**Supplementary Fig. 5g-h**). MERSCOPE, on

the other hand, identified most, but not all, the cell types recognized by Xenium and CosMx,

including alveolar cells, fibroblast cells, hormone-sensing cells, myeloid, T cells, and vascular &

lymphatic cells. MERSCOPE and Xenium showed a high correlation for almost all matching

clusters. The correlation map shows a clearer one-to-one mapping between MERSCOPE and

Xenium clusters than Xenium and CosMx clusters.

415
416

Figure 5: Cell type recovery performance across technology. (a) Clustering results of breast samples in normal TMA from Xenium breast (bottom left) and CosMx multi-tissue 1k (top). Correlation plot showing the correlation between cell types identified (bottom right). **(b)** Clustering results of lung samples in normal TMA from Xenium lung (bottom left) and CosMx multi-tissue 1k (top). Correlation plot showing the correlation between cell types identified (bottom right). **(c)** Clustering results of breast cancer samples in tumor TMA from Xenium breast (bottom left), MERFISH (bottom right) and CosMx multi-tissue 1k (top). Correlation plot showing the correlation between cell types identified in CosMx and Xenium (left), MERFISH and Xenium (right).

DISCUSSION

In this study, we compared data obtained with three commercially available iST platforms with archival FFPE tissues to assess overall technical performance and help guide experimental design with human samples that represent an important use case of these platforms. We focused our analyses on technical performance as a function of tissue type, including 7 different tumor types and 16 normal tissue types. Overall, we found that each iST platform presented various tradeoffs in terms of implementation, panel design and panel options, and resulting total transcript quantification and downstream analyses, including cell segmentation, cell quality, and biological interpretation. All these factors must be considered when designing iST experiments.

There are significant workflow differences between the different platforms which factor into the choice of method. Cutting samples onto MERSCOPE coverslip is more difficult than on standard microscope slides. The total hands-on time for running a slide on Xenium is 2-3 days compared to 5-7 days for MERSCOPE and 2 days on CosMx. We found that MERSCOPE and CosMx are well set up for batch processing in the wet lab, either due to built-in pause points or the instrument's ability to run multiple samples. Xenium is limited for batch processing by a need for a separate thermocycler for each slide pair processed in parallel. After staining, selecting regions of interest (ROIs) presented a surprising challenge for some systems: the Xenium platform could readily image the entire slide as a single ROI which easily covered entire TMAs, but the MERSCOPE ran into a 1cm^2 imaging area limit which meant cores in the addressable region were left unimaged, while the CosMx workflow required a demanding manual selection of ROIs for each core. These factors are likely to change as each company updates its protocol, but currently, Xenium offers the shortest and least hands-on workflow.

From a technical perspective, we analyzed each resulting dataset with a combination of

manufacturer recommended processes for each platform and computational tools that can be

implemented by the user downstream. These pipelines each result in count matrices and detected

transcripts that can be analyzed using a whole suite of emerging tools. For our purposes, when

analyzed at a core level to abrogate the effects of individual cell-segmentation performance, we

found that the total number of transcripts varied substantially across iST platform, with Xenium

yielding the highest number of transcripts captured followed by CosMx. Indeed, this trend held

when normalized for the number of cores imaged and on a per cell basis per area. When this

analysis was also restricted to shared genes, we also found that Xenium consistently had higher

expression levels across each tissue type, with no clear differences between performance on either

tumor or normal tissue.

Using a pseudo-bulk approach, again at the core level, we assessed overall correlation,

reproducibility, and sensitivity of each platform. We found high correlation between replicates of

the same patient, suggesting that there is high reproducibility across technical replicates on each

platform. This is important to consider since cost or input material availability can be prohibitive

to implementing experimental designs that leverage technical replicates—though additional tissue

may still be valuable for powering cell-cell interaction analysis. We additionally found high

correlation on a gene-by-gene basis between MERSCOPE and Xenium platforms. Xenium and

MERSCOPE also showed consistently high specificity across tissue types. CosMx displayed a

characteristic upward curve when compared to MERSCOPE or Xenium on a gene-by-gene basis,

indicating more frequent calls in the lower expression regime. This, coupled with the lower

specificity across several tissues for CosMx and the high false discovery rate, suggest that CosMx

is prone to errors in calling lowly expressed genes. Finally, Xenium had the highest sensitivity

across tissues. CosMx and MERSCOPE both detected fewer transcripts than Xenium.

MERSCOPE outperformed CosMx in the higher quality (as judged by relative performance across

all platforms) tumor TMA but underperformed it in lower quality normal tissue TMA. In general,

our analyses also suggest similar performance within a given platform across a vast array of tissue

types assayed here. We note that given the small number of replicates from each tissue, particularly

in the normal tissue, we stop short of making blanket statements about relative performance across

a particular tissue type. The data suggests, instead, that Xenium and MERSCOPE provide more

reliable true-positive signals of lowly expressed genes and that Xenium's overall performance is

less dependent on sample input quality than the other two platforms. MERSCOPE, especially,

appears to be particularly sensitive to sample input, highlighting the importance of prescreening

RNA integrity according to manufacturer instructions.

When we compared each dataset to existing RNA-seq datasets, we found comparable

correlation of pseudo-bulk data to RNA-seq data from GTEx or the TCGA across each panel and

platform. However, the presence of a characteristic upswing for CosMx, even when comparing to

orthogonal data, further shows that there is a higher false positive rate for lower expression level

genes in CosMx data. This upswing could be explained by the absence of probing genes in a

particular tissue in a larger panel. However, the Xenium multi tissue panel also includes genes not

expressed in breast and lung but does not show a similar upswing. Thus, a more likely

interpretation is that the CosMx is prone to a high FDR at the low expression regime. This could

also suggest that the CosMx transcript counts and detected gene numbers may be slightly inflated

by false discoveries.

From a tissue-specific expression perspective, Xenium showed a distinct expression

pattern of key tissue-markers, whereas CosMx and MERSCOPE did not. Additionally,

MERSCOPE and CosMx consistently showed expression of known tissue markers in unexpected

tissue types. This could be partly explained by an overall low performance on this particular normal

tissue TMA for MERSCOPE due to RNA quality. This performance difference could be

problematic for studies that are designed to compare tissue-specific factors. For studies whose

main biological variable of interest are within the same tissue, factors like sensitivity, specificity,

and panel availability may be a more important guide for iST experimental design.

A significant advantage of spatial transcriptomics data is the ability to map expression in

single cells. We compared each platform on a cell-level basis by assessing cell identification and

cell clustering. Overall, it appears that the out of the box segmentation from Xenium performs

poorly in terms of drawing cell boundaries specific to a single cell, while MERSCOPE and Xenium

much more closely match cell boundaries. This did not appear to differ on a tissue-by-tissue basis,

thus, is likely inherent to the overall approach used by each platform. After applying an expression

level filter, Xenium overall retained the highest number of cells across various filtering

stringencies. Despite Xenium's cell boundaries not clearly matching nuclei, both it and

MERSCOPE were able to effectively separate cells from different lineage markers, as judged by

finding minimal coexpression of disjoint markers, while CosMx showed more double positive

cells (out of, it should be noted, fewer cells expressing the target genes overall).

To determine whether clearer identification of lineage markers resulted in improved ability

to identify cell types, we performed clustering analyses specifically in the breast tissue and breast

cancer samples. We note that we used the full panel, not only the shared genes, when performing

these clustering analyses. Xenium allowed for identification of all major cell lineages in the breast

when compared to several reference breast atlases. Both the global and tissue clustering results

show that CosMx is also able to recognize the major cell types, but cannot identify cell subtypes.

Additionally, since the cluster-enriched genes do not correspond to well-known markers, probably

due to the low expression caused by low sensitivity and specificity, cell type annotation was

particularly difficult. Lastly, despite lower transcript counts and fewer cells, MERFISH still

successfully identified cell groups, capturing the patterns seen in other platforms. These

differences in cell typing, can also be attributed to the differential performance of cell segmentation

pipelines [37,38 39,40]. Since all platforms provide the underlying DAPI stain and morphology images

(in the case of CosMx and MERSCOPE) it is likely that segmentation performance could be

improved on a sample-by-sample or tissue-type-by tissue-type basis. Future work should seek to

assess cell segmentation tools and their performance across data from each platform to help inform

the choice of analytical method where needed.

Plex is an important factor in ST experiments which we have not explicitly considered. The

kinds of questions that may be answered by a 1,000-plex panel are clearly different than those

answered by a 300-plex panel, offering more opportunities to explore intra- and intercellular

signaling interactions. Thus, we note that for the right question, the higher false positive rates and

lower sensitivities of CosMx relative to Xenium could be tolerated for a broader coverage of the

biology. On the other hand, the fully configurable nature of Xenium and MERSCOPE panels could

be better suited for branches of biology not well sampled by the 1,000 plex CosMx panel. We

recommend subsampling existing atlas data to determine whether the gene set which can be studied

will be sufficient to cluster the cell types of interest and identify the necessary biological programs.

We note that each of the manufacturers has publicly stated plans to grow their product offerings

to increasing panel sizes.

There are several limitations of our study. While we attempted to match time post slicing,

the unintentional acquisition of MERSCOPE tissues at thinner thicknesses meant that the rerun

MERSCOPE data had a longer time on slide than other panels. However, the fact that the increase

in counts in the rerun matched the increase in imaging volume, and the fact that Xenium runs

showed stable expression levels over time suggests that this contribution was minimal.

Additionally, our panel design for MERSCOPE required removal of genes so the panel was

compatible with all tissues, lowering the plexity slightly. This could have compromised

MERSCOPE's ability to identify cell types relative to Xenium.

Most importantly, we only attempted to compare the performance of iST platforms under

typical use cases for clinical samples obtained from archival biobanks. Our results don't

necessarily extend to non-human samples, frozen samples, and even FFPE samples which have

been extensively validated for high RNA integrity. Indeed, there have been reports that

MERSCOPE, in previous studies of the mouse brain, shows comparable or even superior results

to those reported by 10X Xenium[41]. Given the large change in data quality between the normal

and tumor TMA, we cannot exclude the possibility that in the highest quality samples MERSCOPE

would provide higher transcript numbers, with the associated downstream benefits relative to

Xenium and CosMx. However, the current guidance of $DV_{200} > 0.6$ restricts studies to the upper

regime sample quality and limits archival investigations.

Despite these limitations, our overall interpretation of these results is that amplification of

RNA signal is especially important for recovery of transcript counts by iST in low-quality samples

where RNA may be highly degraded and fewer landing sites are available for probes. Platforms

(such as Xenium) which rely on small numbers of landing sites and are subsequently heavily

amplified are robust to RNA degradation and are thus more broadly compatible with a broad range

of samples. On the other hand, when sample quality is high (as in some of our tumor samples) the

gap between amplified and unamplified platforms' performance closes and most platforms can

yield useful data for subsequent downstream spatial analysis.

Methods

Sample choice and TMA construction

Two TMAs were constructed using FFPE clinical discards at Brigham and Women's

Hospital Pathology Core and were acquired with a waiver of consent for non-sequencing based

readouts under IRB 2014 P 001721. The samples included:

1. A tumor TMA of 170 cores, 0.6 mm in diameter, including a variety of cancer samples

571   and healthy lymphoid tissue as a positive staining control. The TMA samples were

572   selected from samples previously characterized by ImmunoProfile and were selected

573   to encompass both high and low levels of the biomarkers in the ImmunoProfile panel

574   [*CD8*, *PD-1*, *PD-L1*, *Foxp3*, tumor marker (*Cytokeratin*, *Sox10*, or *PAX8*)].

575   Annotations were performed by KF and SR based on H&E and immunofluorescence

576   staining. Cores included both tumor and healthy control annotation, though for the

577   purpose of this study, all were combined under their tumor label. Tumors were also

578   chosen to be a mixture of *PD-L1* high and *PD-L1* low, a parameter to be analyzed at a

579   future date. This TMA had previously been studied by both H&E, and several highly

580   multiplexed immunostaining approaches, and was known to be of high morphological

581   integrity.

582   2. A normal TMA of 45 cores 1.2 mm in diameter representing a broad range of normal

583   tissues. Samples were sourced from the same patient in either duplicate or triplicate.

584   This TMA was chosen for the breadth of tissue lineages included and the relatively

585    large core size.

586 All samples were fully de-identified before assembly into TMAs. The breakdown of the number

587 of samples per tissue and the number of cores per tissue is included in **Supplementary Table 1-**

588 **2**.

589 **Preparation of sequential sections**

590   Sequential sections were prepared according to manufacturer instructions ("Tissue

591 Preparation Guide Demonstrated Protocol CG000578" for Xenium, "91600112 MERSCOPE User

592 Guide Formalin-Fixed Paraffin-Embedded Tissue Sample Preparation RevB" for Vizgen, and

593 "MAN-10159-01CosMx SMI Manual Slide Preparation Manual" for CosMx) at the Brigham and

594 Women's Hospital Pathology Core. Prior to collecting samples, ~50 μm of each TMA were faced

595 off to reach deeper into the sample where RNA integrity was likely higher. 5 μm sequential

596 sections were then collected, floated in a 37˚C water bath, and adhered to Xenium slides (10X, PN

597 1000460), Vizgen FFPE coverslips (Vizgen, PN 10500102), or standard Superfrost+ slides for

598 CosMx (Leica BOND PLUS slides, Leica Biosystems S21.2113.A). TMAs were sliced as close to

599 the center of the active area as possible for each platform. Samples were baked at 42˚C for 3 hours

600 for Xenium, 55 ˚C for 15 minutes for MERSCOPE, and 60˚C for 16 hours for CosMx. Sections

601 were stored according to manufacturer instructions prior to processing, with 10X Xenium stored

602 in a desicator at room temperature, Vizgen MERSCOPE coverslips stored at -20˚C, and Nanostring

603 CosMx slides stored at 4˚C. Samples for 10X Xenium and Vizgen MERSCOPE were brought to

604 the Spatial Technology Platform at the Broad Institute for processing, while samples for

605 Nanostring CosMx were processed at the Wei lab at Brigham and Women's Hospital.

Vizgen MERSCOPE probe selection

Pre-designed probe panels from Vizgen were not available at the time of the experiment. Therefore, we ordered custom gene panels to match the pre-released gene panels from 10X for the human breast and human lung panels. Gene lists were uploaded to the Vizgen panel design portal and were checked against all profiled tissues, removing genes that were overexpressed in any individual tissue based on Vizgen's design guidelines (FPKM > 900), and ensuring that the total panel FPKM did not exceed the allowed limit in any individual sample type. Panels were manufactured at the 300 gene scale as custom panels BP0892 and BP0893. The final gene lists, for all three iST modalities are available in **Supplementary Table 3.**

Vizgen MERSCOPE data acquisition

MERSCOPE samples were imaged according to manufacturer protocol "9160001 MERSCOPE Instrument User Guide RevF". Samples were processed in two batches, the first of four samples, two of each TMA and with each library prepped in parallel; and a follow up sample of each TMA re-run with the breast panel. Samples were first hybridized with anchoring probes overnight before being embedded in a polyacrylamide gel. Samples were incubated for two hours with a digestion solution at 37°C and then overnight at 47°C overnight in a detergent clearing solution and proteinase K to remove native proteins while the anchoring probes kept nucleic acids bound to the gel. After clearing, samples were additionally photobleached using Vizgen's MERSCOPE Photobleacher for three hours at room temperature in the clearing solution. Samples were hybridized with encoding probes and a cell boundary stain (PN 10400118) and then imaged with imaging kits (PN 10400005). Samples were stored at 37°C in clearing solution after hybridization and before final imaging. After an initial examination of the data, a second batch of

both TMAs was run a second time with the human breast panel, increasing the set imaging capture

thickness from 5 µm to 10 µm to capture more tissue from cores that had lifted during the gel

embedding process. Data was processed on premises through the standard Vizgen workflow to

generate cell by gene and transcript by location matrices. We segmented the data with a built-in

Cellpose method on the most accurate looking cell boundary stain.

10X Xenium data acquisition

10X Xenium samples were processed in three batches according to manufacturer protocols

"Probe Hybridization, Ligation & Amplification, User Guide CG0000582" and "Decoding &

Imaging, User Guide CG000584". Samples were stained utilizing 10X's predesigned Human

Breast (10X, PN 1000463), Human Multi-Tissue and Cancer (10X, PN 1000626), and Human

Lung panels (10X, PN 1000601), as they became available from the manufacturer. Slides for both

TMAs were processed in pairs according to which probe library they were receiving. Slides were

stained with a Xenium imaging kit according to manufacturer instructions (10X, PN 1000460).

Briefly, padlock probes were incubated overnight before rolling circle amplification and native

protein autofluorescence was reduced with a chemical autofluorescence quencher. Slides were

processed on a 10 Xenium Analyzer, with ROIs selected to cover the entire TMA region. Data was

processed on premises through the standard 10X workflow to generate cell by gene and transcript

by location matrices.

Nanostring CosMx

Nanostring CosMx samples were prepared with one 1000 plex panel. Samples were

hybridized with probes and stained with cell markers. Samples were loaded onto the CosMx SMI

at the same time for imaging, during which branched fluorescent probes were hybridized onto the

samples to amplify the signal above the background.

Nanostring CosMx samples were prepared with Human Universal Cell Characterization

1000 Plex Panel (part number 122000157) according to manufacturer protocol "MAN-10159-01

CosMx SMI Manual Slide Preparation Manual". Firstly, slides were baked at 60℃ overnight for

better tissue adherence. After baking, slides were treated sequentially with deparaffinization, target

retrieval (15 min at 100℃), permeabilization (3μg/mL proteinase K, 15 min at 40℃), fiducials

application, post-fixation, NHS-acetate application and then hybridized with denatured probes

from universal panel and default add-on panel. After *in situ* hybridization (18 hours at 37℃), slides

were washed and incubated with DAPI (15 min at RT) and marker stain mix (with *PanCK, CD45,*

CD68 and cell segmentation marker *CD298/B2M*). Slides were washed and loaded onto the

CosMx SMI for UV bleaching, imaging, cycling and scanning. Raw images were decoded by

default pipeline on Atomx SIP (cloud-based service). Machine: CosMx_0020. Serial Number:

INS2301H0020

Data preprocessing

After data acquisition, the resulting outputs were uploaded to a Google bucket associated

with a terra.bio Workspace for distribution and follow on analysis.

To facilitate standardized data formatting and subsequent analytical processes, we built a

data ingestion pipeline with the following objectives: a) to grab cell-level and transcript-level

data from diverse platforms and normalize the data structure; b) to tag each cell and transcript

with essential metadata including tissue type, tumor status, *PD-L1* status, among others

(**Supplementary Fig. 6**); and c) to transform the data into various formats tailored to the

requirements of particularized analyses. Specifically, to tag the data, core centers in the TMA

were manually identified using DAPI images (Xenium) or cell metadata that contains global

coordinates (MERSCOPE and CosMx) using QGIS(version:3.16.10-Hannover). Cells or

transcripts within a specified radius were then labeled with core metadata via spatial joining

(implemented by GeoPandas, version:0.13.0). In instances where cores are in close proximity or

when a uniform radius cannot be applied effectively, we manually generated the core boundary

masks.

Reproducibility

To evaluate panel to panel reproducibility we summed the expression level of shared

genes between indicated panels (breast vs. multi-tissue and breast vs. lung panels from Xenium

and breast vs. lung panels from MERCOPE) over an individual core and plotted all cores present

in each panel, before calculating a Pearson's correlation. The format of the data used is shown in

Supplementary Table 10. To evaluate core to core reproducibility, the individual gene counts of

core 1 were plotted against those of core 2 and a Pearson's r correlation was calculated.

On target rates and false-discovery measurements

To compare across panels and platforms, we subset all datasets to include only cores

assayed in all runs. The fraction of on-target barcodes was calculated as a percentage of all

transcripts corresponding to genes relative to the total number of calls (including negative

control probes and unused barcodes or blank barcodes). These measurements were performed on

individual cores and averaged across all cores of the same tissue type.

Because the difference in relative numbers of controls and target barcodes across

different platforms, we adopted the false discovery rate (FDR) calculation to evaluate the

specificity in a more normalized way (**Fig. 2f-g**). We calculated the FDR of platform p panel m

data in tissue t using the following equation and cell level data (**see example in Supplementary**

Table 11):

$$FDR(negative\ control\ probes)_{p,m}^{t} = mean\left(\frac{\sum_j neg_{jn}}{\sum_i g_{in} + \sum_j neg_{jn}} \times \frac{I}{J} \times \%\right), \quad n = \{1, ..., N\}$$

Where N is the total number of cores that belong to tissue type t, I is the total number of unique

genes, J is the total number of negative control probes, g_{in} is the gene expression of gene i in

core n, neg_{jn} is the total calls negative control probe j in core.

Since MERSCOPE does not include negative control probes, FDR was recalculated by

substituting negative control with blank barcodes (**Fig. 2f**) using the following equation:

$$FDR(blank\ barcodes)_{p,m}^{t} = mean\left(\frac{\sum_l blank_{ln}}{\sum_i g_{in} + \sum_l blank_{ln}} \times \frac{I}{L} \times \%\right), \quad n = \{1, ..., N\}$$

Where N is the total number of cores that belong to tissue type t, I is the total number of unique

genes, L is the total number of unused barcodes or blank barcodes, g_{in} is the gene expression of

gene i in core n, $blank_{ln}$ is the total calls of unused barcode or blank barcode l in core n,

specifically, we used BLANK for Xenium data, Blank for MERSCOPE data, and SystemControl

for CosMx data. We only used the data from matched cores, so N is same for different platform

p.

Sensitivity comparison

716   Sensitivity was measured by the percentage of the total number of unique genes detected

717   above noise level, where the noise was estimated as two standard deviation above average

718   expression of the negative control probes.

Orthogonal RNA-Seq concordance analysis

RNA TCGA cancer sample gene data summarizes 7,932 samples from 17 different cancer

types, and it provides FPKM for each gene documented. We used all samples which were

annotated as BRCA (Breast cancer), BLCA (Bladder cancer), COAD and READ (colorectal

cancer), HNSC (head and neck squamous cell carcinoma), LUAD and LUSC (non-small cell lung

cancer), SKCM (melanoma), and OV (ovarian cancer). For GTEx, we selected the tissue types

matching the annotation in our normal tissue TMA. For each panel, the genes probed by iST were

averaged across all patients with the matching tissue label from the RNA-seq database.

To get pseudo-bulked iST values, the expression level of each gene in each core was normalized

to the sum of all genes in that core and scaled by 100,000. We then averaged these scaled pseudo-

bulk expression values across cores and plotted them against the averaged FPKMs from reference

RNA-seq data sets.

Tissue marker enrichment analysis

To determine the assay's ability to specifically identify known lineage markers, we focused

on the normal tissue TMA profiled with multi-tissue panel of Xenium, breast panel of MERSCOPE,

and 1K panel of CosMx. We selected genes with known canonical expression patterns using based

on transcriptomics data from GTEx. If a gene had 20-fold higher expression in a specific tissue

than every other tissue combined, this gene was considered to be a tissue marker and was used for

assessing specificity for each platform. Counts for each gene were normalized to the total counts

within the core, and then the Z-score of this gene across tissue types was plotted in a heatmap **Fig.**

3e. We calculated average expression of a gene across cores of the same tissue type and normalized

to the total averaged expression of all genes. Z-scores were calculated with the mean and standard

deviation across all averaged genes.

Evaluation of cell segmentation performance

In the absence of ground truth data, we conducted a comparative analysis of cell counts,

cell areas, coexpression across various platforms and panels, utilizing the segmentation results

supplied by each respective company. To facilitate comparison, cell counts were normalized to a

consistent area of 1000 μm². Both cell count and cell area were then delineated at two distinct

levels of detail: a consolidated assessment encompassing all tissue types (see **Fig. 4b-c**), as well

as a segregated evaluation by individual tissue types (refer to **Supplementary Table 6-7**). To

evaluate the biological performance of the segmentation, we plotted coexpression plots of gene

pairs that are mutually exclusive including *CD3e* vs. *CD19*, *CD4* vs. *CD8*, and *CD3e* vs.

EPCAM. We pooled all the filtered cells from matched cores and all available panels of each

platform, dropped cells which do not express either gene, plotted the expression of one gene

against the other, and converted the scatter plot to a 2D histogram showing cell numbers in each

co-expression bin (**Fig. 4g**).

Cells per area quantification

Segmented cells were aggregated by TMA cores. For Xenium and MERSCOPE data, the

estimation of tissue area was performed by calculating the area of a discernible circle, utilizing

respective radius of 0.3 µm and 0.6 µm for tumor and normal TMA samples. Conversely, for the

CosMx dataset, the tissue area estimation was approached differently due to its square-like data

presentation, a result of the FOV selection process. Here, the tissue area was deduced by

multiplying the number of FOVs covered by each core with the area of a single FOV.

Clustering

For cell filtering, cells with less than 10 transcript counts in MERFISH and Xenium

datasets were removed, and cells with less than 20 transcript counts in CosMx datasets were

removed. We followed standard processes to then cluster and annotate cell types across each

dataset using Scanpy[42]. Briefly, data was normalized and scaled, dimensionality reduction was

performed and cell clusters were identified[43, 44]. To identify the cell type for each cluster, we used

a t-test to find the markers for each Leiden cluster and annotated them according to previous

literature[29-36]. These are some of the example markers used for cell type annotation: in breast

samples, PIGR and KIT for alveolar cells, for B cells, KRT5, DST, and MYLK for basal cells,

LUM, MMP2 and CXCL12 for fibroblast, etc. Heatmaps of the top 3 markers for each cluster are

drawn for each dataset from all three panels (refer to **Supplementary Figure 5a-c**). For datasets

that showed batch effect with patients, Harmony was used to remove this variance[45]. Correlation

heatmaps were generated over overlapping genes that exist in both datasets, and the Pearson

correlation coefficient was calculated.

Acknowledgment

We thank Nir Hacohen, Ilya Korsunsky, Roopa Madhu, and Kseniia Anufrieva for helpful

discussions, as well as Patricia Rogers and Natan Pierete for assistance with the 10X Xenium data

acquisition. We appreciate 10X Genomics, Vizgen, and NanoString Technologies for reviewing

data and analyses for quality. This work is supported by a Broad Institute SPARC grant and by a

Brigham and Women's Hospital Department of Medicine - Broad Institution collaborative

research Award. K.W. is supported by a NIH-NIAMS K08AR077037, a Burroughs Wellcome

Fund Career Awards for Medical Scientists, and a Doris Duke Charitable Foundation Clinical

Scientist Development Award. B.A.G is supported in part through the Geisel School of Medicine

at Dartmouth's Center for Quantitative Biology through a grant from the National Institute of

General Medical Sciences (NIGMS, P20GM130454) of the NIH.

Author contributions

Conceptualization: A.Y., S.F., tissue-microarray construction: K.F., K.P., T.B., S.R.,

pathological annotation: K.P., S.R., gene selection: A.Y., S.F., Xenium and MERSCOPE data

acquisition: J.N., CosMx data acquisition: C.G., M.T.,K.W., analysis: H.W., R.H., B.G.,S.F.,

figure generation: H.W., R.H., B.G.,S.F., writing original draft: H.W., R.H., J.N., B.G., S.F., draft

reviewing and editing: H.W., R.H., B.G., S.F., supervision: B.G., S.F., funding acquisition: S.F.

K.W.

Data and Code Availability

Cell level data and gene level data will be made available at the time of final publication.

All code used in this manuscript for data processing and analysis will be made available on

github prior to final publication.

Declaration of interests

All authors declare that they have no conflicts of interest.
807

References

1. Chen W.T., Lu A. Croissants K. et al. Spatial Transcriptomics and In Situ Sequencing to

Study Alzheimer's Disease. Cell. 182 (4), 976–991 (2020).

2. Hunter, M.V., Moncada, R., Weiss, J.M. et al. Spatially resolved transcriptomics reveals

the architecture of the tumor-microenvironment interface. Nat Commun 12, 6278 (2021).

3. Lein E, Borm LE, Linnarsson S. The promise of spatial transcriptomics for neuroscience

in the era of molecular cell typing. Science. 2017 Oct 6;358(6359):64-69.

4. Lewis SM, Asselin-Labat ML, Nguyen Q, et al. Spatial omics and multiplexed imaging to

explore cancer biology. Nat Methods. 2021 Sep;18(9):997-1012.

5. Rao, A., Barkley, D., França, G.S. et al. Exploring tissue architecture using spatial

transcriptomics. Nature 596, 211–220 (2021).

6. Haque, A., Engel, J., Teichmann, S.A. et al. A practical guide to single-cell RNA-

sequencing for biomedical research and clinical applications. Genome Med 9, 75 (2017).

7. Williams, C.G., Lee, H.J., Asatsuma, T. et al. An introduction to spatial transcriptomics for

biomedical research. Genome Med 14, 68 (2022).

8. Tian, L., Chen, F. & Macosko, E.Z. The expanding vistas of spatial transcriptomics. Nat
Biotechnol 41, 773–782 (2023).

9. Janesick A., Shelansky R., Andrew D.G., et al. High resolution mapping of the breast
cancer tumor microenvironment using integrated single cell, spatial and insitu analysis of
FFPE tissue. bioRxiv 2022.10.06.510405

10. He S., Bhatt R., Birditt B., et al. High-plex multiomic analysis in FFPE tissue at single-
cellular and subcellular resolution by spatial molecular imaging. bioRxiv
2021.11.03.467020

11. Moffitt, J. R. et al. Molecular, spatial, and functional single-cell profiling of the
hypothalamic preoptic region. Science 362, eaau5324 (2018).

12. Du J., Yang Y.C., An Z.J., et al. Advances in spatial trasnciptomics and related data
analysis strategies. J Transl Med. 2023; 21: 330.

13. Phan, H.V., van Gent, M., Drayman, N. et al. High-throughput RNA sequencing of
paraformaldehyde-fixed single cells. Nat Commun 12, 5636 (2021).

14. Xu, Z., Zhang, T., Chen, H. et al. High-throughput single nucleus total RNA sequencing
of formalin-fixed paraffin-embedded tissues by snRandom-seq. Nat Commun 14, 2734
(2023).

15. Gnanapragasam VJ. Unlocking the molecular archive: the emerging use of formalin-fixed
paraffin-embedded tissue for biomarker research in urological cancer. BJU Int. 2010
Jan;105(2):274-8.

16. Matsunaga H., Arikawa K., Yamazaki M., et al. Reproducible and sensitive micro-tissue
RNA sequencing from formalin-fixed paraffin-embedded tissues for spatial gene
expression analysis. Sci Rep 12, 19511 (2022).

17. Kok Hao Chen et al. Spatially resolved, highly multiplexed RNA profiling in single

cells.Science 348, aaa 6090 (2015).

18. Walsh, L.A., Quail, D.F. Decoding the tumor microenvironment with spatial technologies.

Nat Immunol 24, 1982–1993 (2023).

19. Tomczak K., Czerwinska P., Wiznerowicz M. The cancer genome atlas (TCGA): an

immeasuable source of knowledge

20. GTEx Consortium. Genetic effects on gene expression across human tissues. Nature 550,

204–213 (2017).

21. Junttila S., Smolander J., Elo L., et al. Benchmarking methods for detecting differential

states between conditions from multi-subject single-cell RNA-seq data. Briefings in

Bioinformatics, Volume 23, Issue 5, Sep. 2022.

22. Zimmerman S. M., Fropf R., Kulasekara B. R., et al. Spatially resolved whole

transcriptome profiling in human and mouse tissue using digital spatial profiling. Genome

Res. 2022 Oct;32(10):1892-1905.

23. Trevethan. R. Sensitivity, specificity and predictive values: foundations, liabilities, and

pitfalls in research and practice. Front Public Health. 2017 Nov 20;5:307.

24. He S, Bhatt R, Brown C. et al. High-plex imaging of RNA and proteins at subcellular

resolution in fixed tissue by spatial molecular imaging. Nat Biotechnol. 2022

Dec;40(12):1794-1806.

25. Garrido-Trigo A, Corraliza AM, Veny M et al. Macrophage and neutrophil heterogeneity

at single-cell spatial resolution in human inflammatory bowel disease. Nat Commun. 2023

Jul 26;14(1):4506. doi: 10.1038/s41467-023-40156-6. PMID: 37495570; PMCID:

PMC10372067.

26. Nieto P., Elosua-Nayes M., Trincado J. L., et al. A single-cell tumor immune atlas for precision oncology. Genome Res. 2021 Oct; 31(10): 1913–1926.

27. Elmentaite R., Conde C. D., Yang L., et al. Single-cell atlases: shared and tissue-specific cell types across human organs.

28. Heumos L., Schaar A. C., Lance C., et al. Best practices for single-cell analysis across modalities. Nat Rev Genet 24, 550–572 (2023).

29. Kumar, T., Nee, K., Wei, R. et al. A spatially resolved single-cell genomic atlas of the adult human breast. Nature 620, 181–191 (2023).

30. Gray G.K., Li C.M., Rosenbluth J.M. et al. A human breast atlas integrating single-cell proteomics and transcriptomics. Developmental Cell, Volume 57, Issue 11, 2022.

31. Reed A.D., Pensa S., Steif A. et al. A human breast cell atlas mapping the homeostatic cellular shifts in the adult breast. bioRxiv 2023.04.21.537845

32. Travaglini K.J., Nabhan A.N., Penland L. et al. A molecular cell atlas of the human lung from single-cell RNA sequencing. Nature 587, 619–625 (2020).

33. Sikkema, L., Ramírez-Suástegui, C., Strobl, D.C. et al. An integrated cell atlas of the lung in health and disease. Nat Med 29, 1563–1577 (2023).

34. Wu, S.Z., Al-Eryani, G., Roden, D.L. et al. A single-cell and spatially resolved atlas of human breast cancers. Nat Genet 53, 1334–1347 (2021).

35. Dave A., Charytonowicz D., Francoeur N. et al. The breast cancer single-cell atlas: defining cellular heterogeneity within model cell lines and primary tumors to inform disease subtype, stemness, and treatment options. Cell Oncol (Dordr). 2023 Jun;46(3):603-628.

36. Yeo S.K., Zhu X., Okamoto T. et al. Single-cell RNA sequencing reveals distinct patterns of cell state heterogeneity in mouse models of breast cancer. eLife 9:e58810.

37. Greenwald, N.F., Miller, G., Moen, E. et al. Whole-cell segmentation of tissue images with

human-level performance using large-scale data annotation and deep learning. Nat

Biotechnol 40, 555–565 (2022).

38. Stringer, C., Wang, T., Michaelos, M. et al. Cellpose: a generalist algorithm for cellular

segmentation. Nat Methods 18, 100–106 (2021).

39. Petukhov, V., Xu, R.J., Soldatov, R.A. et al. Cell segmentation in imaging-based spatial

transcriptomics. Nat Biotechnol 40, 345–354 (2022).

40. Chen, H., Li, D. & Bar-Joseph, Z. SCS: cell segmentation for high-resolution spatial

transcriptomics. Nat Methods 20, 1237–1243 (2023).

41. Marco Salas, S., Czarnewski, P., Kuemmerle, L. B., Helgadottir, S., Mattsson Langseth,

C., Tiesmeyer, S., ... & Nilsson, M. (2023). Optimizing Xenium In Situ data utility by

quality assessment and best practice analysis workflows. bioRxiv, 2023-02.

42. Wolf, F., Angerer, P. & Theis, F. SCANPY: large-scale single-cell gene expression data

analysis. Genome Biol 19, 15 (2018).

43. Dries R., Chen J., Del Rossi N., et al. Advances in spatial transcriptomic data analysis.

Genome Res. 2021 Oct;31(10):1706-1718.

44. Traag, V.A., Waltman, L. & van Eck, N.J. From Louvain to Leiden: guaranteeing well-

connected communities. Sci Rep 9, 5233 (2019).

45. Korsunsky, I., Millard, N., Fan, J. et al. Fast, sensitive and accurate integration of single-

cell data with Harmony. Nat Methods 16, 1289–1296 (2019).

Supplementary Figures and Supplementary Tables: [link]

Supplementary Figure 1: Reproducibility across different panels and cores from same patient.
(a) Scatter plots of cumulative gene expression levels (natural log transformed) of shared genes between two panels within each platforms, captured from matched tissue cores. Column 1: Xenium breast vs. Xenium lung; Column 2: Xenium breast vs. Xenium multi-tissue; Column 3: MERSCOPE breast round 1(5 μm) vs. MERSCOPE breast round 2(10 μm). Each data point corresponds to a TMA core. **(b)** Scatter plots of gene expression levels (natural log transformed) of every shared gene between two cores of the same tissue type from the same patient. In this example, cores are from breast cancer tissue. Each data point corresponds to a gene. **(c)** Heatmap of correlation coefficient expressed as Pearson's r values, indicating good core-to-core or sample-to-sample reproducibility. Core pairs are selected from same tissue/tumor type from the same patients.

Supplementary Figure 2: Gene by gene plots of iST results by panel and by tissue microarray.
(a) Scatter plots of summed gene expression levels (natural log transformed) of every shared gene between Xenium (breast/lung) and CosMx (1k) data, captured from matched normal tissue TMA cores. Each data point corresponds to a gene. **(b)** Same as (a) but between MERSCOPE (breast/lung) and CosMx(1k). **(c)** Same as (a) but between Xenium(breast/lung) and MERSCOPE(breast/lung). **(d)** Same as (a) but between Xenium(multi-tissue) and CosMx(1k). † denotes the MERSCOPE lung panel acquired with a 5 μm imaging thickness.

Supplementary Figure 3: Tissue marker analyses and cell level measurements.
(a) Heatmap of Z-scored gene expression showing CosMx's ability to specifically identify known lineage markers. We focused on the normal tissue TMA profiled with multi-tissue panel and selected genes with canonical expression patterns for this analysis. **(b)** Same as (a) but for MERSCOPE (breast panel). **(c)** Heatmap of transcripts per cell after filtration. Only shared genes (40) are considered here for each panel. **(d)** Same as (c) but showing unique transcripts from the same gene set.

Supplementary Fig: 4 Global clustering analyses
(a) Global Clustering results of tumor TMA from Xenium breast panel (top), Xenium lung panel (middle), and Xenium panhuman panel (bottom). **(b)** Global Clustering results of tumor TMA from MERFISH breast panel (top), MERSCOPE lung panel (middle), and CosMx multitissue panel (bottom).

Supplementary Fig: 4 Global clustering analyses
(c) Global Clustering results of normal TMA from Xenium breast panel (top), Xenium lung panel (middle), and Xenium panhuman panel (bottom). **(d)** Global Clustering results of normal TMA from MERSCOPE breast panel (top), MERSCOPE lung panel (middle), and CosMx multitissue panel (bottom).

Supplementary Fig: 5. Cell type recovery and UMAPs

(a) Heatmap showing the top gene markers for cell types annotated in breast samples of normal TMA from Xenium breast (left) and Cosmx multitissue (right). **(b)** Heatmap showing the top gene markers for cell types annotated in lung samples of normal TMA from Xenium lung (left) and Cosmx multitissue (right). **(c)** Heatmap showing the top gene markers for cell types annotated in breast cancer samples of tumor TMA from Xenium breast (left), Cosmx multitissue (middle), NERSCOPE breast (right).

Supplementary Fig: 5. Cell type recovery and UMAPs
(d) UMAP of breast cancer samples of tumor TMA from Xenium breast panel pre (left) and post (right) batch effect removal. **(e)** UMAP of breast cancer samples of tumor TMA from Cosmx multitissue panel pre (left) and post (right) batch effect removal. **(f)** UMAP of breast cancer samples of tumor TMA from MERFISH breast panel pre (left) and post (right) batch effect removal

Supplementary Fig: 5. Cell type recovery and UMAPs
(g) UMAP plot of well-known gene markers for BrC, in breast cancer samples of tumor TMA from CosMx multitissue panel. **(h)** UMAP plot of well-known gene markers for BrC, in breast cancer samples of tumor TMA from Xenium multitissue panel.

	cell_id	fov	x_centroid	y_centroid	transcript_counts	geometry
	247310500000100001	1	12200.436214	359.843544	0	POINT (12200.43621 359.84354)
	247310500000100002	1	12187.780341	368.994318	9	POINT (12187.78034 368.99432)
	247310500000100003	1	12209.730058	361.731687	0	POINT (12209.73006 361.73169)
	247310500000100005	1	12222.551206	359.055678	0	POINT (12222.55121 359.05568)
	247310500000100006	1	12234.644291	358.311934	0	POINT (12234.64429 358.31193)

Supplementary Figure 6: Workflow for tagging imaging spatial transcriptomics data
(a) To facilitate standardized data formatting and subsequent analytical processes, we built this data ingestion pipeline with the following objectives: 1) to grab cell-level and transcript-level data from diverse platforms and normalize the data structure; 2) to tag each cell and transcript with essential metadata including tissue type, tumor status, PD-L1 status, among others; and 3) to transform the data into various formats tailored to the requirements of particularized analyses. Specifically, to tag the data, core centers in the TMA were pinpointed using DAPI images (Xenium) or cell metadata that contains global coordinates (MERSCOPE and CosMx) using QGIS(version:3.16.10-Hannover). Cells or transcripts within a specified radius were then labeled with core metadata via spatial joining (implemented by GeoPandas, version:0.13.0). In instances where the cores are in close proximity or when a uniform radius cannot be applied effectively, we manually generated the core boundary masks.