Single-cell measurement of microbial growth rate with Raman microspectroscopy

Running title: Microbial growth rate measured with Raman spectroscopy.

Tristan A. Caro [a#], Srishti Kashyap [a], George Brown [b], Claudia Chen [b], Sebastian H. Kopf [a], Alexis S. Templeton [a]

a. Department of Geological Sciences, University of Colorado Boulder, Colorado, United States
b. Department of Applied Mathematics, University of Colorado Boulder, Boulder, Colorado, United States

 # Corresponding Author
Tristan Caro
tristan.caro@colorado.edu
2200 Colorado Ave
Benson Earth Sciences Building
Rm. 285. UCB 399
Boulder, CO 80309

Competing interest statement: The authors declare no conflicts of interest.

This file includes:
    Main Text
    Figures 1 - 4
    Tables 1 - 2
    References

Key words: Stable isotope probing, Raman microspectroscopy, deuterium, single-cell analysis, growth rate, nanoSIMS

Abstract

Rates of microbial activity and growth are fundamental to understanding environmental geochemistry and ecology. However, measuring the heterogeneity of microbial activity at the single-cell level, especially within complex populations and environmental matrices, remains a forefront challenge. Stable Isotope Probing (SIP) is a standard method for assessing microbial activity and involves measuring the incorporation of an isotopically labeled compound into microbial biomass. Here, we assess the utility of Raman microspectroscopy as a SIP technique, specifically focusing on the measurement of deuterium (^{2}H), a tracer of microbial biomass production. We generate calibrations of microbial biomass ^{2}H values and find that Raman microspectroscopy reliably quantifies ^{2}H incorporation ranging between 0 and 40 at. %. Applying the results of this calibration to a SIP model, we explicitly parameterize the factors controlling microbial growth quantification, demonstrating how Raman-SIP can measure the growth of microorganisms with doubling times ranging from hours to years. Furthermore, we correlatively compare our Raman-derived measurements with those of nanoscale secondary ion mass spectrometry (nanoSIMS) to compare the relative strengths of nanoSIMS- and Raman-based SIP approaches. We find that Raman microspectroscopy is a robust, accessible methodology that can readily differentiate and quantify the growth of individual microbial cells in complex samples.

Importance

Growth rate, the rate at which organisms grow and reproduce, is a key metric with which to evaluate microbial physiology and contributions to system-level processes. The heterogeneity of microbial growth across space, time, and populations is often difficult to capture with bulk-scale techniques. Single-cell methods hold promise for measuring the heterogeneity of microbial growth rates and responses to changing conditions *in situ*, without the need for cultivation of microbial isolates. In this study, we evaluated the ability of Raman microspectroscopy, a non-destructive and rapid technique, to measure the assimilation of isotopically labeled water into individual microbial cells and thereby calculate their rates of growth. We explicitly parameterize the factors controlling the quantification of microbial growth rate and compare this technique to standard methods. The framework we report allows researchers to couple single-cell and aggregate rate measurements to functional or system-level properties, a forefront challenge in microbiology.

Introduction

Microbial growth rate is a critical parameter for assessing biogeochemical cycling, environmental habitability, and microbial fitness. Not only does the rate of microbial growth represent a geochemical transformation, but it is also a parameter that microorganisms modulate in response to changing conditions: microorganisms may more readily replicate when conditions are favorable to them and adopt alternative physiologic states under adverse conditions.

Methods for measuring microbial growth often rely upon the addition of an isotopically labeled substrate to a natural sample and measuring its incorporation into microbial biomass, a method generally termed stable isotope probing (SIP) (1–3). SIP has been applied with the stable isotopes of various elements including H, C, N, O, S, etc., through measurement of single-cells (4–8), membrane lipids (3,9,10), and nucleic acids (2,11–13). One approach to probe cellular biosynthesis is to amend an environmental sample with deuterated water (^2H_2O) (3–6,8,10,14–18). Water is incorporated into microbial biomass during the synthesis of lipids, nucleic acids, and proteins, and so probing a sample with deuterated water allows the quantification of microbial biomass growth. ^2H_2O has been termed a "passive" tracer for microbial anabolism, as it is nutritionally neutral and taxonomically agnostic.

In microbial communities, spatial relationships dictate ecological and evolutionary dynamics, as well as community structure and function (26–30). However, microbial growth is known to be extremely heterogenous in environments such as soil matrices (30), marine particles, (31–33), the rock-hosted subsurface (34,35) and others. Spatially-resolved measurements of microbial growth can identify the habitability and biological activity of environmental samples at the microscale. Furthermore, single-cell measurements of microbial

growth are especially useful for microbial ecologists as they provide information on anabolic

heterogeneity across and within populations. However, describing the spatial distributions of

microorganisms in the environment, in addition to their degrees of activity, remains one of the

primary challenges in microbial ecology.

Nanoscale secondary ion mass spectrometry (nanoSIMS) is a standard method for

conducting spatially-resolved SIP measurements of individual cells (4,6,7,10,15,18,36–39).

NanoSIMS relies on the measurement of secondary ions released via ablation by a primary ion

beam and measures isotopes of multiple elements at high spatial resolution. However, isotope

ratios measured via nanoSIMS may become depressed due to the inclusion of stains, nucleic acid

probes, and sample washing (40–42). Additionally, nanoSIMS remains a costly and time

intensive method that is inaccessible to many researchers. We therefore set out to assess

alternative methodologies that may circumvent these difficulties.

In 2015, foundational work by Berry and colleagues established Raman

microspectroscopy as a viable tool for detecting deuterium enrichment of microbial biomass at

the single-cell level (8). Raman-based measurements of deuterium in microbial biomass rely on

the principle that there exists a strong band centered at 2800 cm^{-1}, corresponding to C-H bonding

environments associated with lipids, proteins, and nucleic acids of cells. When deuterium is

incorporated into microbial biomass, this band, now representing C-^{2}H ("C-D") bonds, becomes

red-shifted and centers at 2200 cm^{-1} **(Fig. 1B)**. In previous work, the "CD% metric", defined as

the fractional abundance of CD and CH peak areas, (CD% = CD / (CD + CH) × 100%) was

found to correlate with the deuterium fractional abundance (^2F) of the growth medium provided

to the organisms (8). Berry et al. illustrated applications of Raman-SIP in combination with

fluorescence in situ hybridization (FISH) and cell-sorting methodologies to capture the identity

of active microbial community members.

For our study, we built upon this prior work in four key aspects. First, we generated a

robust hydrogen isotopic calibration for Raman spectroscopy that specifically targets two

environmentally-relevant taxa **(Fig. 2)**. Second, we combined the results of our isotopic

calibration with a numerical model of microbial growth to determine the feasibility and

sensitivity of Raman-SIP for quantitative inference of microbial biomass growth rate **(Fig. 3)**.

The ability to translate Raman-derived measurements of deuterium enrichment into biomass

synthesis/growth estimates would pave the way for future investigations to not only identify

microbial activity *in situ*, but also to quantify biosynthetic growth rate. Third, the cells used to

generate our Raman-based isotopic calibration were correlatively measured by nanoSIMS to

determine the extent to which nanoSIMS and Raman-based measurements agree. Finally, we

provide a computational framework for Raman-SIP that allows users to optimize SIP incubations

for their specific study system.

Results and Discussion

Single-cell isotopic measurements

To assess the sensitivity and precision of Raman microspectroscopy as a reporter of cellular ^{2}H

incorporation, we generated a large isotopic dataset of bacterial and archaeal cells grown to

equilibrium with deuterated (^{2}H$_2$O) media of varying hydrogen isotope compositions (0 – 50 at.

% ^{2}H). We focused our study on two anaerobic organisms isolated from the subsurface

biosphere: a sulfate reducing bacterium (SRB) *Thermodesulfovibrio hydrogeniphilus* (43)*,* and a

methanogenic archaeon *Methanobacterium NSHQ04* (44). We acquired individual measurements

of cellular deuterium enrichment (n = 351) with Raman spectroscopy and nanoSIMS (**Fig. 1**).

We refer to hydrogen isotope content using both Raman and nanoSIMS as the fractional

abundance of deuterium, ^2F, reported in atom percent (at. %) (see *Materials and Methods*),

where $^2F = {}^2H / (H + {}^2H) \times 100 \%$. To generate hydrogen isotope calibrations, we fit simple

linear models relating the hydrogen isotopic composition of microbial biomass ($^2F_{biomass}$) to the

organism's growth water ($^2F_{gw}$). (**Fig. 2**).

With Raman microspectroscopy, biomass ^2H enrichment ($^2F_{biomass}$) increased linearly and

exhibited strong positive correlation (R^2 = 0.94, 0.95, for *T. hydrogeniphilus and M. NSHQ04,*

respectively) with the deuterium content of the growth water ($^2F_{gw}$) (**Fig. 2, Table 1B**). Raman

measurements of biomass deuterium content exhibited an average standard deviation of σ_{2F} =

2.89 at. %. The variability of Raman-derived $^2F_{biomass}$ increased with the isotopic composition of

the growth water: a maximum standard deviation was observed at the $^2F_{gw}$ = 50% ^2H_2O

condition where σ_{2F} = 3.77 at. %. Corresponding measurements by nanoSIMS of the same cells

followed a similar trend: $^2F_{Biomass}$ increased linearly with the growth water applied (R^2 = 0.84,

0.96 for *T. hydrogeniphilus* and *M. NSHQ04*, respectively) and the average standard deviation of

nanoSIMS ^2F was σ_{2F} = 1.86 at. %. Similar to Raman, the standard deviation of nanoSIMS

measurements increased along with the isotopic composition of growth water, reaching a

maximum at the 50% label where σ_{2F} = 3.20 at. % (**Table 1A**). With both analytical techniques

and across both organisms, we observed depressed biomass enrichments at $^2F_{gw}$ = 50 at. %,

likely resulting from deuterium toxicity effects (see section *Applying Raman microspectroscopy*

to SIP growth rate measurements). For analytical uncertainty, we report the average standard

deviation observed in Raman and nanoSIMS measurements across growth water conditions: 2.89

and 1.86 at. %, respectively. We emphasize that these are conservative over-estimates of

analytical error due to heterogeneity in the hydrogen isotopic enrichment of individual cells. Our

data support previous observations by Berry et al. that point to high reproducibility of single-cell

measurements (8).

Applying Raman microspectroscopy to SIP growth rate measurements

Having generated a Raman-based hydrogen isotopic calibration of microbial biomass, we sought

to test the applicability of this analytical method to measurements of cell-specific growth rate.

Microbial growth rate can be inferred from SIP incubations because the degree of an organism's

biomass isotopic enrichment over time depends on the isotopic composition of a given substrate

and its rate of biomass synthesis (14). Therefore, if measurements of ^{2}H incorporation are

reliable, then these values should translate to cell-specific growth rates. To this end, we sought to

parameterize sources of uncertainty and propagation to downstream calculations to define

acceptable experimental conditions and analytical uncertainties for quantitative measurement of

microbial growth rate.

The growth rate of an organism in the presence of a stable isotope tracer is calculated

with **Eq. 1** (14). In brief, microbial growth rate is a logarithmic relationship between an

organism's isotopic composition at the start of an incubation (F_0) and its isotopic composition at

the time (t) of sampling (F_T). Cellular isotopic enrichment is compared to the enrichment of the

isotopic label (F_L) offset by a water hydrogen assimilation efficiency constant (a_w) – a value

representing the fraction of biomass hydrogen that a cell acquires from its growth water, as

opposed to other metabolic sources and associated fractionation effects.

174
$$\mu = -\frac{1}{t} \cdot ln\left(\frac{F_T - a_w \cdot F_L}{F_0 - a_w \cdot F_L}\right)$$
(Eq. 1)

176   We modeled growth rates (μ) (**Eq. 1**) across a wide range of SIP experimental conditions

177   (varying F_L, t, and a_w,) and applied standard propagation of uncertainty to determine the

178   associated errors (σ_μ) (**Eq. 2**) that result from compounding uncertainty of Raman-SIP

179   experimental aspects (3,14):

181
$$\sigma_\mu = \frac{\sqrt{(a_w F_L - F_T)^2 \sigma_{F_0}^2 + (a_w F_L - F_0)^2 \sigma_{F_T}^2 + a_w^{\,2} \cdot (F_0 - F_T)^2 \sigma_{F_L}^2 + F_L^2 \cdot (F_0 - F_T)^2 \sigma_{a_w}^2}}{t \cdot (a_w F_L - F_0) \cdot (a_w F_L - F_T)}$$
(Eq. 2)

183   Where sigma (σ) terms represent uncertainties of each subscripted parameter (e.g., σ_{FT}

184   represents uncertainty in biomass isotopic composition at time of sampling T). We convert

185   growth rate to biomass generation time using the relationship:

187
$$T_G = \frac{ln(2)}{\mu}$$
(Eq. 3)

189   where μ is growth rate in units of t^{-1} and T_G is generation time in units of t.

190   To a first order, the range of growth rates (μ) that can be measured with SIP depends on

191 the isotopic composition of the label and the incubation duration. Uncertainty in growth rate (σ_μ)

192 depends on analytical uncertainty in label strength (σ_{FL}), analytical uncertainty of biomass

193 isotopic enrichment (σ_{FT}), and microbial water hydrogen assimilation efficiency (σ_{aw}). Our model

194 constrains the region of quantification for both isotopic enrichment (**Fig. 3C**) and corresponding

195 growth rate (**Fig. 3D-E**) in Raman SIP experiments. We propagate the uncertainty estimates of

196 $^2F_{biomass}$ estimated from our hydrogen isotope calibration as σ_{FT} to estimate associated uncertainty

197 in growth rate (σ_μ) that results from a given SIP experiment. We report uncertainty in growth rate

198 relative to the growth rate itself as *Relative Error = (σ_μ / μ) × 100%*. We set our limit for relative

199 error to be 50% such that errors below this cutoff result in growth rate differences that are

200 distinguishable with $> 2\sigma$ confidence.

201   In **Fig. 3** we display example outputs from our model using incubation times of *t = 5* and

202 *t = 30 days* across a range of ^2H_2O label strengths. For example, the range of quantification for

203 an experiment applying a 35% label is 6.9 - 32.6 at. % deuterium enrichment. Ranges of

204 quantification and error optima for additional isotopic label strengths are reported in **Table 2.**

205 For a 5-day incubation, this range of ^2F quantification corresponds to growth rates between 0.53

206 – 0.04 d^{-1}, or generation times of 1.3 and 17 days. For a 30-day incubation, this range of ^2F

207 quantification corresponds to growth rates between 0.09 - 0.007 d^{-1} or generation times between

208 7.7 and 101.9 days (**Fig. 3E**).

209   Asymptotic rises in uncertainty at high and low biomass isotopic enrichment,

210 respectively, correspond to (i) uncertainty in growth rate as cell isotopic enrichment approaches

211 that of the label solution and (ii) uncertainty resulting from biomass isotopic signal not

significantly exceeding that of the noise inherent to the analytical method. The first condition (i)

represents cells that grow faster than can be discriminated using the set incubation time and label

strength, while (ii) represents cells that grow too slow to be distinguished from noise. It should

be noted that our ranges of quantification should be viewed conservative estimates, as the source

of uncertainty we apply corresponds to heterogeneity of ^{2}H-incorporation across homogenous

cell populations, as opposed to the error across individual Raman acquisitions. This latter source

of error is far smaller both in our dataset and in previously reported datasets (8).

Incubation time (t) shifts the dynamic range of SIP (**Supplementary Fig. S8**): shorter

incubations can resolve faster-growing organisms from each other but may not capture slower-

growing organisms due to insufficient label incorporation. Conversely, longer incubations can

capture slower-growing organisms, but faster-growing organisms may become saturated with the

label and will no longer be resolvable from each other (**Fig. 3**). Growth rates of fast-growing

organisms, those that are readily saturated with deuterium, could lead to underestimates of

aggregate microbial turnover.

Increasing the strength of the label (F_L) expands the dynamic range of the SIP method

i.e., both slower and faster growth rates can be captured as there is more isotopic "space" that

can be measured (**Table 2, Fig. 3**). However, increasing F_L must be weighed against the risk that

higher deuterium concentrations can impose toxicity or unintended physiological effects.

Deuterated water at high levels of isotopic labeling have been shown to have no effect, inhibition

and even stimulation among different microorganisms (8,10,45). In our experiments, we

observed depressed isotopic enrichment in growth water containing *50 at. %* ^{2}H$_2$O. These effects

were noted for both organisms and both the Raman and nanoSIMS data (grey point intervals in

Fig. 2). We conjecture that this result stems from isotopic toxicity affecting microbial growth or

metabolic switching at high isotopic enrichment, but this mechanism cannot be explained by our

study. Therefore, we caution against applying ^2H_2O label strengths at or above 50 at. % due to

varying physiological effects that may arise at this threshold, especially when applying this

approach to natural microbial communities of unknown and mixed composition. Uncertainties in

isotopic label strength (σ_{FL}), especially at high isotopic enrichment, due to pipetting or

gravimetric dilution contribute relatively minor components to total uncertainty.

The water hydrogen assimilation efficiency constant (a_w) is essential for estimation of

microbial growth rate, as it sets the upper limit of biomass deuterium enrichment. This limit is

represented by the $a_w \cdot F_L$ terms in **Eq. 1 (Fig 3A)**. Therefore, similar to increasing the strength

of the label, an organism with a greater a_w has more isotopic "space" in which to become

enriched. In our study, we use Raman-derived ^2F measurements to empirically determine the a_w

of *T. hydrogeniphilus* and *M. NSHQ04* to be $a_w = 0.76 \pm 0.02$ and $a_w = 0.88 \pm 0.02$, respectively

(**Table 1C**, **Fig. 2**). The higher a_w observed with the methanogen *M. NSHQ04* is in line with

primarily autotrophic growth (44) with an offset from $a_w = 1$ due to hydrogen isotopic

fractionation. The slightly lower a_w for the sulfate reducing *T. hydrogeniphilus* is indicative of a

hydrogen contribution from its acetotrophic metabolism or from greater fractionation effects. We

note that these values represent a_w of whole-cell biomass, which may be distinct from those

calculated via compound-specific analysis (3,21,25). It was previously noted that uncertainty in

a_w is a significant driver of uncertainty in hydrogen SIP-derived growth rate estimates (3). For

mixed communities, this uncertainty must be propagated through growth rate calculations. For

organisms with known metabolisms, this term can be estimated as common metabolic modes

often constrain a_w factor to a degree (3,21,25).

Raman and nanoSIMS measure different pools of cellular hydrogen.

Currently, nanoscale secondary ion mass spectrometry (nanoSIMS) serves as a standard method

for reporting single-cell isotope ratios. A NanoSIMS instrument applies a primary ion beam to

ionize and ablate a sample, inducing the release of secondary ions whose isotopic composition is

measured by magnetic sector isotope ratio mass spectrometry. NanoSIMS has many benefits as

an analytical technique: it can measure multiple isotopes simultaneously and allows

spatial/elemental imaging at the nanoscale. However, nanoSIMS destroys the sample and is time

and cost-intensive. Furthermore, depression of isotopic enrichment measured by nanoSIMS due

to sample preparation has been widely reported for multiple stable isotope tracers (^{13}C, ^{18}O, ^{15}N,

^{2}H, and ^{34}S) under various fixation and staining conditions (8,10,15,36,40–42,46). Given these

limitations, it would be useful to have additional methods that can complement the strengths of

nanoSIMS while negating some of its drawbacks.

Deuterium enrichments measured by Raman were greater than those measured by

nanoSIMS, when compared to the culture growth water (**Fig. 2**). When we correlated cell-

specific $^{2}F_{biomass}$ measurements we observed that, cell-to-cell, nanoSIMS-derived hydrogen

isotope values were severely depressed compared to corresponding Raman-derived values

(**Supplemental Fig. 2**). From this correlative dataset, we calculated an average hydrogen isotope

dilution factor of 58.7 ± 6.0 % across all growth water labeling conditions (*Supplementary Text*).

Such a significant dilution of hydrogen isotopes speaks to a fundamental difference in the

nature of nanoSIMS- and Raman-based measurements. Owing to how the primary ion beam

ionizes and ablates a cell, nanoSIMS is mostly agnostic to the sources of $^2H/^1H$ it measures,

setting aside potential differences in hydrogen ionization efficiency across different classes of

molecules. Therefore, $^1H^-$ and $^2H^-$ ions reaching the instrument detectors can derive from a

variety of cellular hydrogen sources including remnant intracellular water, adsorbed extracellular

water, and biomass hydrogen in both non-exchangeable sites and exchangeable sites that are

readily mixed with natural-abundance washing buffers during sample preparation. Protonated

sites in biomolecules (e.g., O-H, N-H, S-H, etc.) experience rapid exchange with aqueous wash

solutions on the order of picoseconds to minutes (47,48), meaning that anabolically produced O-

^2H, N-^2H, S-^2H, etc. bonds would be rapidly overprinted by natural-abundance hydrogen during

standard washing procedures (**Fig. 4**). As described in *Materials and Methods*, no stains or

probes were applied to the cells in this study. Therefore, the hydrogen dilution factors reported

here result solely from sample washing. This significant and persistent isotopic dilution may

confound efforts to use deuterium tracers as measures of biosynthesis in nanoSIMS-based studies

(10). This problem is compounded at lower-isotopic enrichment, where substantial variability in

$^2H/^1H$ values results from low ^2H ion counts as well as random variation in which cell

subcomponents (e.g., highly exchangeable vs. non-exchangeable) are ablated (10). Our

observation that the magnitude of isotopic dilution appears variable across samples (**Fig. 2**)

suggests that dilution factors could contribute substantial uncertainty to corrected values. Future

work should apply caution when applying dilution factors to back-calculate isotopic abundance,

as these can vary widely between the sample preparation and organism (40), an effect that may

be pronounced in environmental systems with mixed microbial communities.

A key advantage of Raman-SIP is that the C-H (2800 - 3100 cm^{-1}) and C-D (2040 - 2300

cm^{-1}) bands report solely non-exchangeable carbon-hydrogen and carbon-deuterium stretching

modes associated with mostly lipids, proteins and nucleic acids (8,49). These spectral regions are

oblivious to exchangeable bonding environments including O-H, N-H, S-H, etc., as well as

remnant water. Raman spectroscopy therefore focuses on biosynthetically produced organic

bonds without the confounding effects of remnant water or exchanged biomolecular hydrogen.

Deuterium measurements by Raman-SIP appear to exhibit a greater reliability in tracking

biosynthetic incorporation of deuterium into cellular biomass.

Future Directions for Raman-SIP

Raman spectroscopy has several advantages for measuring microbial growth rates. First, it is a

non-destructive, rapid method that is readily coupled with other techniques such as fluorescence

microscopy, fluorescence in situ hybridization (FISH), scanning electron microscopy (SEM), or

cell sorting (8,15). A single Raman spectrum takes between seconds to one minute to acquire,

and many Raman instruments are equipped with fluorescence modules that can enable

visualization of the cells in addition to transmitted or reflected light microscopy. Second, Raman

is comparatively accessible relative to nanoSIMS and allows analysis with minimal preparation,

making it capable of being deployed directly on many forms of microbial systems including

filters from aquatic environments, rock faces, plant material, etc. Third, Raman

microspectroscopy has microscale spatial resolution and provides diagnostic information about

cellular organic composition. The fingerprint region (200 – 1800 cm^{-1}) is routinely used to

identify biomolecules (49) as well as differentiate taxa, down to the species/strain level and

growth stage (50–52). Fourth, the Raman fingerprint region enables identification of biologically

precipitated minerals and the mineralogical/material context of a cell (37,53). Raman SIP with

deuterium, specifically, makes identification of individual cell and mineral components in

organic-mineral assemblages feasible because the C-D band occupies a typically silent region in

inorganic Raman spectra. Co-registered mineral and quantitative cell-activity measurements

open the door for novel investigations into the cell-mineral and cell-cell relationships that drive

microbial activity in the rock-hosted biosphere (35). Furthermore, measurement of ^{13}C and ^{15}N

enrichment in microbial cells has been demonstrated with Raman (55–59) and the potential

exists for the co-registration of multiple stable isotopes at the single-cell level.

The main disadvantage of deuterium Raman-SIP is its detection limit. Classical or

spontaneous Raman microspectroscopy of microbial cells is limited by low signal intensities,

particularly in samples with high autofluorescence (54). A key outcome of our study is the

definition of Raman-SIP ranges of quantification. Our study defines the minimum cutoff for

quantitative measurement as ~ 7 at. % (**Table 2**), depending on the label strength applied. In the

coming years, advancements in Raman technology hold potential for increasing its sensitivity

and subsequently improving isotopic measurements. Specifically, surface- or tip-enhanced

Raman spectroscopy (SERS or TERS) (50,60–63), stimulated Raman scattering (SRS),

resonance Raman spectroscopy, or coherent anti-Stokes Raman spectroscopy (CARS), which are

shown to dramatically improve acquisition times, and enhance signal, could be applied to

determine isotopic enrichments (55,56,60,64–69). However, at present, robust validation of these

methods for deuterium SIP is required (**Supplemental Text, Supplemental Fig. S4 – S5**).

Our results point to Raman microspectroscopy being a powerful tool for conducting SIP,

especially where complex environmental matrices may complicate bulk-omics or other single-

cell approaches. For the benefit of future researchers, we designed an interactive graphical user

interface (GUI), *Shiny R-SIP*, that allows users to implement our SIP model and optimize their

experiments by considering how the relevant sources of uncertainty impact the design of a

Raman-SIP experiment (*Supplementary Data* or online: https://apps.kopflab.org/login with login

credentials username: *shiny-rSIP*, password: *public-access*). The GUI allows users to adjust

parameters of label strength, incubation time, and water hydrogen assimilation efficiency, as well

associated uncertainties in these terms, in order to design effective SIP experiments. The

parameters we define in our microbial growth model are specific to the study organisms used and

our Raman instrument: they are intended to provide reasonable estimates of uncertainty that may

be encountered. Users of this GUI are encouraged to define their own uncertainty terms by

constraining the technical variation in individual CD% measurements inherent to their Raman

instrument. For samples where microbial growth may exhibit large variation, we suggest users

consider multiple incubation times (or sub-sampling efforts) to effectively capture a wide range

of biomass growth rates in experiments targeting natural microbial communities, while also

quantifying their threshold of tracer saturation as an upper-limit of growth-rate quantification.

Conclusions

In this work, we provide the basis for the quantitative measurement of single-cell microbial growth rates via Raman-SIP. We develop hydrogen isotope calibrations for two organisms, *T. hydrogeniphilus* (sulfate-reducing bacterium) and *M. NSHQ04* (methanogenic archaeon). These calibrations validate the utility of Raman microspectroscopy as a method for measuring microbial ^{2}H/H enrichment. Applying our Raman-derived isotopic calibration to a model of microbial growth, we define ranges of quantification for Raman SIP experiments where single-cell growth rates can be sensitively distinguished. We find that with reasonable (20 - 40 at. % ^{2}H$_2$O) isotopic label strengths and incubation times, Raman-SIP can capture a wide array of microbial generation times, with this range being defined by the parameters of the SIP experiment. These ranges of quantification can guide the design and interpretation of SIP experiments where Raman is used to track cellular ^{2}H incorporation. Finally, we observe that hydrogen isotopic values derived from nanoSIMS suffer from severe dilution that may complicate the use of nanoSIMS for quantifying cellular ^{2}H uptake. In conclusion, this work provides a robust framework for applying deuterium Raman-SIP to spatially resolved, quantitative investigations of microbial activity in environmental and model systems.

Materials and Methods

Microorganisms and Growth Conditions

We cultivated *Thermodesulfovibrio hydrogeniphilus* (43) and *Methanobacterium NSHQ04* (44), in media containing 0%, 10%, 20%, 30%, 40%, or 50% ^{2}H$_2$O (0% = local ultra-pure water which contains ~140ppm ^{2}H$_2$O naturally). Base media (described below) were modified with different

percentages (vol/vol) of ^2H_2O [Cambridge Isotope Laboratories], depending on the final isotopic

enrichment desired for the experiment. All cultures were grown in the dark in anaerobic 60 ml

serum vials containing 25 ml of growth medium.

Thermodesulfovibrio hydrogeniphilus HBr5^T (DSM 18151) (43), a sulfate reducing

bacterium, was obtained from the Deutsche Sammlung von Mikroorganismen und Zellkulturen

(DSMZ). It was grown in a modified DSM 641 medium that contained (per liter): 1.0g NH_4Cl,

2.0g Na_2SO_4, 1.0g $Na_2S_2O_3x5H_2O$, 1.0g $MgSO_4x7H_2O$, 0.1g $CaCl_2x2H_2O$, 0.5g KH_2PO_4, 2.0g

$NaHCO_3$, 1ml trace element solution SL-10, 1ml selenite-tungstate solution, 1g yeast extract,

0.5ml sodium resazurin (0.1% wt/vol), 0.2g sodium acetate, 10ml 141 vitamin solution, and 0.1g

Na_2S x $9H_2O$. The headspace was flushed and over-pressurized with 1 bar $H_2:CO_2$ (80:20).

Methanobacterium NSHQ04 (44) is a methanogenic archaeon enriched from groundwater

isolated from the Samail ophiolite in Oman. The culture was grown in a synthetic site water

(NSHQ04) medium (44) containing (per liter): 420mg NaCl, 683mg $CaCl_2x2H_2O$, 1.2mg

H_4SiO_4, 1.7mg NaBr, 60mg $MgCl_2x6H_2O$, 100mg NH_4Cl, 10 ml 141 trace elements, 10ml 141

vitamins, 0.5ml sodium resazurin (0.1 wt/vol), 20ml antibiotic cocktail [containing per 100ml

MilliQ water, 500mg penicillin G, 500mg streptomycin, 500ul ampicillin (10mg/ml stock)],

10ml KH_2PO_4 (1.1 wt/vol), 4ml yeast extract (5% wt/vol), 2ml $Fe(NH_4)_2(SO_4)_2$ (0.2% w/v),

10ml sodium formate (100mM), 10ml cysteine-HCl (2.5% w/v), and 10ml Na_2S (2.5% w/v). The

headspace was flushed and over-pressurized with 2 bar $H_2:N_2$ (5:95).

Cultures of *T. hydrogeniphilus* and *M. NSHQ04* were incubated in forced-air incubators

at 65°C and 40°C, respectively. Cultures were transferred three times into media of identical

deuterium enrichment. Before each transfer, cultures were grown to stationary phase, which was

reached in 3-4 days for *T. hydrogeniphilus* and 12-14 days for *Methanobacterium NSHQ04*. This

ensured that the cells measured were at or near isotopic equilibrium with their respective growth

water, offset by a_w. After the third transfer, 1ml of each culture was harvested and fixed by

addition of paraformaldehyde to a final concentration of 2% (vol/vol). Fixed cells were washed

by centrifugation in successively dilute phosphate-buffered saline (PBS) of rinses [1X, 1%,

0.1%, and 0.01% (vol/vol) PBS) to ensure the removal of excess salts and fixative. PBS was

composed of Dulbecco's formula: KCl 200 mg/L, KH_2PO_4 200 mg/L, NaCl 8000 mg/L,

Na_2HPO_4 1150 mg/L. After the final wash, cells were resuspended in 50μl of 0.01% (vol/vol)

PBS.

Sample Preparation

Sample coupons were prepared by raster engraving of aluminum-coated glass slides (Deposition

Research Lab Inc.) with an Epilog Mini 24 CO_2 laser machine operating at 80 W laser power,

10% etching speed, to create a 3 x 6 grid. These etchings were used to manually break the glass

slide, producing 7mm^2 coupons that are compatible with both Raman and nanoSIMS sample

mounts. Aluminum coupons were used instead of silicon wafers (typically used for nanoSIMS

studies) because silicon exhibits a large peak at 520 cm^{-1} in the Raman spectrum. All coupons

were sterilized and rendered organic-clean by combustion for 8 hours at 450°C in a muffle

furnace. 5μl of fixed and washed cells was spotted on individual coupons and allowed to air-dry.

Raman Spectroscopy, Fitting, and SIP Calculations

Raman spectroscopy was conducted at the Raman Microspectroscopy Laboratory, Department of

Geological Sciences, University of Colorado-Boulder (RRID:SCR_019305) on a Horiba

LabRAM HR Evolution Raman spectrometer equipped with a 100mW 532 nm excitation laser.

The laser was focused with a 100X (NA = 0.90) air objective lens, resulting in a spot size of

~1µm. Single-cell spectra were captured in the 200-3150 cm^{-1} range using 100% laser power

(2.55 mW) over 2 acquisitions of 45 seconds. A 100 µm confocal pinhole and 600 lines/mm

diffraction grating were used, resulting in a spectral resolution of ~ 4.5 cm^{-1}. Duplicate

acquisitions were averaged to remove noise and cosmic ray spikes. Spectra were baseline-

subtracted using a polynomial fit in LabSpec 6 (Horiba Scientific). Fitting of bands in the 1800 -

3150 cm^{-1} region was performed using Gaussian-Lorentzian curve fitting algorithms to integrate

C-D (2040 – 2300 cm^{-1}) and C-H (2800 – 3100 cm^{-1}) bands. Raman-derived ^{2}F was determined

by calculating the CD% value of each spectrum: $^{2}F = CD\% = CD / (CH + CD) \times 100\%$ where

CD and CH are the areas of the C-D and C-H bands, respectively.

To correlate Raman spectra with nanoSIMS data, extensive context maps of the sample

coupons were acquired using reflected light microscopy on the Raman instrument. Before cell

spotting, fiducial markings were etched into the aluminum coupon surface with a Leica

LMD7000 microdissection system.

As described in *Results and Discussion*, modeling of growth rate was carried out with

Eq. 1 – 3. (3,14). Using these formulae, we modeled μ across a range of F_L, t, and a_w. We define

an acceptable cutoff using a relative error term defined as the fraction of uncertainty in

calculated growth rate over the growth rate itself (σ_μ / μ) where quantifiable growth rate is where

Relative Error = (σ_μ / μ) × 100 % < 50% **(Fig. 3)**.

The value of the water hydrogen assimilation constant (a_w) was determined for both

organisms from the slopes of the linear regression relating $^2F_{biomass}$ vs. $^2F_{water}$. This value, $a_w =$

$X_w \times \alpha_{biomass/w}$, corresponds to the isotopic offset between microbial biomass relative to its

growth water. This term includes both X_w, the mole fraction of lipid H sourced from growth

water, and $\alpha_{biomass/w}$, the isotopic fractionation between whole-cell biomass and growth water

(21).

Nanoscale Secondary Ion Mass Spectrometry

Samples were analyzed with a CAMECA NanoSIMS 50L (CAMECA, Gennevilliers, France)

housed in the Caltech Microanalysis Center at the California Institute of Technology. Before,

analysis, cells were sputter-coated with 25nm of Au. Cells were analyzed using a 2.5 pA primary

Cs+ beam current, and a presputter time of 6 – 20 minutes depending on the size of the raster

area. Two masses were collected in parallel ($^1H^-$, $^2H^-$) using electron multipliers. Individual

samples were identified using the NanoSIMS CCD camera and correlated with Raman-

measurements using previously generated image maps. For all analyses, at least two frames were

collected. All ion images were recorded at 512 x 512-pixel area. Single-cell isotope values were

quantified using a custom script (see *Data Availability*) relying on the *sims* Python library

(https://github.com/zanpeeters/sims). Cell areas were manually defined using the GNU Image

Manipulation Program (GIMP). Isotope ratios (^2R) were converted to fractional abundances (^2F)

with the relations: $^2F = {}^2R / (1 + {}^2R)$ and $^2R = {}^2H / {}^1H$, where ^1H and ^2H represent total ion

counts in detectors EM1 and EM2, respectively, averaged across a cell area. Additionally, a

deadtime correction (70) was applied:

$$N = \frac{N_m}{1 - N_m \cdot \tau / T}$$ (Eq. 4)

where N is the real number of secondary ions, N_m is the number of secondary ions measured, τ is

deadtime (in seconds) and T is dwell time (seconds per pixel). This instrument's detector

deadtime was 44 ns. Deadtime correction had a negligible effect on isotopic values

(**Supplementary Text, Supplementary Fig. S7**).

Author Contributions

Tristan Caro and Srishti Kashyap designed the study and performed the experiments. Tristan

Caro, Srishti Kashyap, Claudia Chen, and George Brown analyzed the data. Alexis Templeton

and Sebastian Kopf provided financial support for the study. Tristan Caro and Sebastian Kopf

designed the GUI. Alexis Templeton and Sebastian Kopf contributed to the design of the study

and interpretation of its findings. Tristan Caro wrote the manuscript with input from all authors.

Data Availability

Raw data and analysis scripts used to process data and generate visualizations are available at

https://github.com/tacaro/Caro-et-al-Raman-SIP. The interactive GUI referenced in this

manuscript is available at https://apps.kopflab.org/login with login username *shiny-rSIP* and

password *public-access.* In addition, a copy can be downloaded for offline use from the
manuscript GitHub repository.

Acknowledgements

The authors would like to thank the University of Colorado Boulder, Department of Geological

Sciences for support of this study. This work was also supported by a NASA Exobiology

Program grant to Alexis Templeton (80NSSC21K0489) and an Army Research Office grant to

Sebastian Kopf (W911NF2120119 / 78484-LS). The authors sincerely thank Yunbin Guan of the

Caltech Microanalysis Center for expertise with nanoSIMS measurements. We thank Eric

Ellison of the CU Boulder Raman Microspectroscopy Laboratory (RRID:SCR_019305) for

assistance with Raman measurements and data analysis. We extend sincere thanks to John

Magyar for helpful discussions during the design of this study. We credit Dan Utter for

assistance with nanoSIMS data processing and initial drafting of the nanoSIMS data extraction

script. We thank Rebecca Wipfler for assistance with laser microdissection microscopy and

coupon etching. Tristan Caro was supported by a National Science Foundation Graduate

Research Fellowship and through the BioFrontiers Institute of the University of Colorado

Boulder. Claudia Chen and George Brown were supported by the Laboratory for

Interdisciplinary Statistical Analysis at the University of Colorado Boulder. The Laboratory for

Interdisciplinary Statistical Analysis at the University of Colorado Boulder is supported by

National Science Foundation Grant No. 1955109.

Figures

Figure 1: Representative Raman and nanoSIMS data collected in this study.

Across all panels data from two representative cells, "1" and "2", are noted. Representative nanoSIMS

isotopic images of are shown in **Panel A**, with pixel counts labeled on the x and y axes. From left to right,

the pixel intensity in each panel corresponds to direct ion counts for $^1H^-$, $^2H^-$, and the fractional abundance

of deuterium (^2F), respectively. The top row shows cells grown in 0% ^2H_2O (natural abundance); the

bottom row displays cells grown in 50 at. % ^2H_2O. Representative single-cell fitted Raman spectra are

shown in **Panel B**, with the characteristic C-H band between 2800 and 3000 cm^{-1} clearly visible. For cells

grown in deuterated media (Cell 2), the C-D band emerges between 2040 and 2300 cm^{-1}. **Panel C**

displays a typical micrograph captured through a confocal Raman microspectroscope. The scale bar

represents 2µm. Cells in all panels represent *T. hydrogeniphilus*. The yellow boxes in the top-left facet of

Panel A, and Panel C, represent a correlated region imaged with nanoSIMS and Raman (reflected light),

respectively.

Figure 2: Hydrogen isotope calibrations relating growth water and biomass ^{2}H content.

Individual measurements of single-cell biomass deuterium enrichment ($^2F_{biomass}$) conducted with Raman

spectroscopy (top panels) or nanoSIMS (bottom panels) are plotted against the corresponding deuterium

content of the microbial growth media ($^2F_{gw}$). The dotted line indicates the 1:1 line; the solid blue line

represents the line of best fit whose equation is noted in each panel. Points and error bars displayed are

528   mean and standard deviation of biomass deuterium enrichment, respectively. Note that the linear models

529   displayed in each panel are calculated only with data between deuterium enrichments of 0 – 40 at. % to

530   ensure that these models are not affected by the toxicity effects observed at 50 at. % label isotopic

531   enrichment (point intervals in gray).

Figure 3: Modeling of microbial growth in the presence of enriched isotopic label (A, B) and associated error in biomass ^{2}F quantification (C) and growth rate/generation time (D, E).

A schematic figure (**Panel A**) visualizes how microbial growth rate is calculated as a function of biomass isotopic enrichment (F_T), label strength (F_L), incubation time (t), and water assimilation efficiency

constant (a_w) (**Eq. 1**). Similarly, **Panel B** displays example output of the microbial growth model, where

$a_w = 0.7$. Isotopic enrichment increases as a function of time but depends upon isotopic label strength (F_L)

and growth rate (μ, noted above each plot). In **Panels C-E**, we display quantification ranges where

relative error in the growth rate measurement is < 50% of the growth rate itself (Relative Error = σ_μ / μ ×

100% where μ is apparent growth rate). In **Panel C,** relative error is plotted against biomass deuterium

enrichment ($^2F_{biomass}$) for label strengths ^2F_L ranging from 20-50 at. %. Plotted points indicate the of

minimum error for each isotopic label. Along the top axis of **Panel C**, the lower limit, error optimum, and

upper limit of quantification, are noted, for the isotopic label strength of $^2F_L = 35$ at. %. The key results

from **Panel C** are summarized in **Table 2.** In **Panels D** and **E,** relative error is plotted against biomass

generation times in the context of a simulated SIP incubation. In these examples, the incubation times of 5

days (**Panel D**) and 30 days (**Panel E**) are chosen. Along the top axes of **Panels D** and **E**, the minimum,

optimum, and maximum quantifiable generation times are noted for the isotopic tracer strength of $^2F_L =$

35 at. %. These generation times correspond to the limits of quantification and optima noted in Panel A.

Figure 4. Schematic diagram showing the dilution of ^{2}H/H isotopic signal due to sample preparation, and how Raman and nanoSIMS measure different pools of biomass ^{2}H/H.

(**A**) The original pool of deuterated cell mass is biosynthesized, sourcing ^{2}H from the growth medium. During sample washing, hydrogen in exchangeable, or "protic" bonding environments rapidly exchange with wash solutions, leaving a diluted residual pool of ^{2}H/H. Carbon-hydrogen and carbon-deuterium bonds are not affected by this dilution effect, as ^{2}H/H in these bonds do not exchange during washing. (**B**) Within the C-H and C-D (C-^{2}H) wavenumber regions, Raman is only sensitive to carbon-hydrogen and

561 carbon-deuterium bonds, which are not affected by dilution due to washing, whereas **(C)** nanoSIMS

562 ostensibly measures all pools of ^{2}H/H, including those affected by dilution.

Tables

Table 1. Results and statistical summary of the isotopic calibration.
Here we summarize key outputs of our analyses: (A) analytical uncertainty in biomass measurements across growth water isotopic compositions ($^2F_{gw}$), (B) Linear models relating the isotopic composition of biomass ($^2F_{biomass}$) to that of growth water ($^2F_{gw}$), (C) Water hydrogen assimilation efficiency factors (a_w) derived from the hydrogen isotopic calibration, with associated standard error, (D) Assessment of correlation between biomass isotopic values estimated with Raman spectroscopy ($^2F_{Raman}$)and nanoSIMS ($^2F_{nanoSIMS}$) (**Supplementary Fig. S2, S3**).

	Raman	**nanoSIMS**
(A) Analytical Uncertainty in Cellular ^2F (SD) (at. %)		
$^2F_{gw}$= 0% (SD of the method blank)	1.13	0.01
$^2F_{gw}$ = 10%	2.56	1.24
$^2F_{gw}$ = 20%	3.34	1.04
$^2F_{gw}$ = 30%	3.07	2.95
$^2F_{gw}$ = 40%	3.46	2.74
$^2F_{gw}$ = 50%	3.78	3.20
Mean SD across all conditions	2.89	1.86
(B) Hydrogen Isotope Calibration Models		
where: $y = {}^2F_{Biomass},\ x = {}^2F_{gw}$		

T. hydrogeniphilus	y = 0.761x + 3.05 $R^2 = 0.94$	y = 0.29x + 1.11 $R^2 = 0.84$
M. NSHQ04	y = 0.878x + 1.89 $R^2 = 0.95$	y = 0.512x – 0.40 $R^2 = 0.96$
(C) Water hydrogen assimilation efficiency $a_w \pm SE$		
T. hydrogeniphilus	0.761 ± 0.014	0.290 ± 0.009
M. NSHQ04	0.878 ± 0.020	0.512 ± 0.010
(D) Correlative Statistics		
Linear model equation *where:* $y = {}^2F_{nanoSIMS}$ $x = {}^2F_{Raman}$ (Supplementary Text, Supplementary Fig. S2-S3)	y = 0.439x – 0.3617 Slope SE = 0.01 Intercept SE = 0.24	
RMSE (at. %)	2.52	

Table 2: Range of Quantification for Raman-derived (CD%-based) measurements of single-cell deuterium content (Fig. 3C).

The optimum value represents the $^2F_{biomass}$ value at which relative error in growth rate calculation is minimized. The lower and upper limits of quantification represent where relative error of growth rate calculation exceeds 50%.

Isotopic label strength (^2F_L at. %)	Lower limit of quantification ($^2F_{biomass}$ at. %)	Optimum ($^2F_{biomass}$ at. %)	Upper limit of quantification ($^2F_{biomass}$ at. %)
20	7.4	12.6	16.7
25	6.8	15.8	22.4
30	6.6	18.9	27.7
35	6.4	21.8	32.6
40	6.4	25.0	37.7
45	6.3	28.1	42.8
50	6.2	31.3	48.0

References

1. Dumont MG, Murrell JC. Stable isotope probing — linking microbial identity to function. Nat Rev Microbiol. 2005 Jun;3(6):499–504.

2. Hungate BA, Mau RL, Schwartz E, Caporaso JG, Dijkstra P, van Gestel N, et al. Quantitative Microbial Ecology through Stable Isotope Probing. Appl Environ Microbiol. 2015 Nov 1;81(21):7570–81.

3. Caro TA, McFarlin J, Jech S, Fierer N, Kopf S. Hydrogen stable isotope probing of lipids demonstrates slow rates of microbial growth in soil. Proc Natl Acad Sci. 2023 Apr 18;120(16):e2211625120.

4. Trembath-Reichert E, Morono Y, Ijiri A, Hoshino T, Dawson KS, Inagaki F, et al. Methyl-compound use and slow growth characterize microbial life in 2-km-deep subseafloor coal and shale beds. Proc Natl Acad Sci. 2017 Oct 31;114(44):E9206–15.

5. Cui L, Yang K, Zhu YG. Stable Isotope-Labeled Single-Cell Raman Spectroscopy Revealing Function and Activity of Environmental Microbes. In: Dumont MG, Hernández García M, editors. Stable Isotope Probing: Methods and Protocols [Internet]. New York, NY: Springer; 2019 [cited 2022 Mar 28]. p. 95–107. (Methods in Molecular Biology). Available from: https://doi.org/10.1007/978-1-4939-9721-3_8

6. Eichorst SA, Strasser F, Woyke T, Schintlmeister A, Wagner M, Woebken D. Advancements in the application of NanoSIMS and Raman microspectroscopy to investigate the activity of

microbial cells in soils. FEMS Microbiol Ecol [Internet]. 2015 Oct 1 [cited 2021 Feb 2];91(fiv106). Available from: https://doi.org/10.1093/femsec/fiv106

7. Trembath-Reichert E, Shah Walter SR, Ortiz MAF, Carter PD, Girguis PR, Huber JA. Multiple carbon incorporation strategies support microbial survival in cold subseafloor crustal fluids. Sci Adv. 2021 Apr;7(18):eabg0153.

8. Berry D, Mader E, Lee TK, Woebken D, Wang Y, Zhu D, et al. Tracking heavy water (D2O) incorporation for identifying and sorting active microbial cells. Proc Natl Acad Sci. 2015 Jan 13;112(2):E194–203.

9. Wegener G, Kellermann MY, Elvert M. Tracking activity and function of microorganisms by stable isotope probing of membrane lipids. Curr Opin Biotechnol. 2016 Oct 1;41:43–52.

10. Kopf SH, McGlynn SE, Green-Saxena A, Guan Y, Newman DK, Orphan VJ. Heavy water and 15N labelling with NanoSIMS analysis reveals growth rate-dependent metabolic heterogeneity in chemostats. Environ Microbiol. 2015;17(7):2542–56.

11. Koch BJ, McHugh TA, Hayer M, Schwartz E, Blazewicz SJ, Dijkstra P, et al. Estimating taxon-specific population dynamics in diverse microbial communities. Ecosphere. 2018;9(1):e02090.

12. Li J, Mau RL, Dijkstra P, Koch BJ, Schwartz E, Liu XJA, et al. Predictive genomic traits for bacterial growth in culture versus actual growth in soil. ISME J. 2019 Sep;13(9):2162–72.

13. Coskun ÖK, Özen V, Wankel SD, Orsi WD. Quantifying population-specific growth in

benthic bacterial communities under low oxygen using H 2 18 O. ISME J. 2019

Jun;13(6):1546–59.

14. Kopf SH, Sessions AL, Cowley ES, Reyes C, Sambeek LV, Hu Y, et al. Trace incorporation

of heavy water reveals slow and heterogeneous pathogen growth rates in cystic fibrosis

sputum. Proc Natl Acad Sci. 2016 Jan 12;113(2):E110–6.

15. Schaible GA, Kohtz AJ, Cliff J, Hatzenpichler R. Correlative SIP-FISH-Raman-SEM-

NanoSIMS links identity, morphology, biochemistry, and physiology of environmental

microbes. ISME Commun. 2022 Jun 30;2(1):1–10.

16. Wegener G, Bausch M, Holler T, Thang NM, Mollar XP, Kellermann MY, et al. Assessing

sub-seafloor microbial activity by combined stable isotope probing with deuterated water

and 13C-bicarbonate. Environ Microbiol. 2012;14(6):1517–27.

17. Kellermann MY, Wegener G, Elvert M, Yoshinaga MY, Lin YS, Holler T, et al. Autotrophy

as a predominant mode of carbon fixation in anaerobic methane-oxidizing microbial

communities. Proc Natl Acad Sci. 2012 Nov 20;109(47):19321–6.

18. Morono Y, Terada T, Nishizawa M, Ito M, Hillion F, Takahata N, et al. Carbon and nitrogen

assimilation in deep subseafloor microbial cells. Proc Natl Acad Sci. 2011 Nov

8;108(45):18295–300.

19. Damoglou AP, Dawes EA. Studies on the lipid content and phosphate requirement of

glucose- and acetate-grown Escherichia coli. Biochem J. 1968 Dec 1;110(4):775–81.

20. Neidhardt FC, editor. Escherichia coli and Salmonella typhimurium: cellular and molecular

biology. Washington, D.C: American Society for Microbiology; 1987. 2 p.

21. Zhang X, Gillespie AL, Sessions AL. Large D/H variations in bacterial lipids reflect central

metabolic pathways. Proc Natl Acad Sci. 2009 Aug 4;106(31):12580–6.

22. Sessions AL, Jahnke LL, Schimmelmann A, Hayes JM. Hydrogen isotope fractionation in

lipids of the methane-oxidizing bacterium Methylococcus capsulatus. Geochim Cosmochim

Acta. 2002 Nov 15;66(22):3955–69.

23. Valentine DL, Sessions AL, Tyler SC, Chidthaisong A. Hydrogen isotope fractionation

during H2/CO2 acetogenesis: hydrogen utilization efficiency and the origin of lipid-bound

hydrogen. Geobiology. 2004;2(3):179–88.

24. Leavitt WD, Kopf SH, Weber Y, Chiu B, McFarlin JM, Elling FJ, et al. Controls on the

hydrogen isotope composition of tetraether lipids in an autotrophic ammonia-oxidizing

marine archaeon. Geochim Cosmochim Acta. 2023 Jul 1;352:194–210.

25. Wijker RS, Sessions AL, Fuhrer T, Phan M. 2H/1H variation in microbial lipids is controlled

by NADPH metabolism. Proc Natl Acad Sci. 2019 Jun 18;116(25):12173–82.

26. Boetius A, Ravenschlag K, Schubert CJ, Rickert D, Widdel F, Gieseke A, et al. A marine

microbial consortium apparently mediating anaerobic oxidation of methane. Nature. 2000

Oct;407(6804):623–6.

27. Cordero OX, Ventouras LA, DeLong EF, Polz MF. Public good dynamics drive evolution of

iron acquisition strategies in natural bacterioplankton populations. Proc Natl Acad Sci. 2012

Dec 4;109(49):20059–64.

28. Marlow J, Spietz R, Kim KY, Ellisman M, Girguis P, Hatzenpichler R. Spatially resolved

correlative microscopy and microbial identification reveal dynamic depth- and mineral-

dependent anabolic activity in salt marsh sediment. Environ Microbiol. 2021;23(8):4756–77.

29. Védère C, Vieublé Gonod L, Nunan N, Chenu C. Opportunities and limits in imaging

microorganisms and their activities in soil microhabitats. Soil Biol Biochem. 2022 Nov

1;174:108807.

30. Sokol NW, Slessarev E, Marschmann GL, Nicolas A, Blazewicz SJ, Brodie EL, et al. Life

and death in the soil microbiome: how ecological processes influence biogeochemistry. Nat

Rev Microbiol. 2022 Feb 28;1–16.

31. Ebrahimi A, Schwartzman J, Cordero OX. Cooperation and spatial self-organization

determine rate and efficiency of particulate organic matter degradation in marine bacteria.

Proc Natl Acad Sci. 2019 Nov 12;116(46):23309–16.

32. Kirchman DL. Growth Rates of Microbes in the Oceans. Annu Rev Mar Sci. 2016;8(1):285–

309.

33. Grossart HP, Kiørboe T, Tang K, Ploug H. Bacterial Colonization of Particles: Growth and

Interactions. Appl Environ Microbiol. 2003 Jun;69(6):3500–9.

34. Casar CP, Kruger BR, Flynn TM, Masterson AL, Momper LM, Osburn MR. Mineral-hosted

biofilm communities in the continental deep subsurface, Deep Mine Microbial Observatory,

SD, USA. Geobiology. 2020;18(4):508–22.

35. Templeton AS, Caro TA. The Rock-Hosted Biosphere. Annu Rev Earth Planet Sci.

2023;51(1):493–519.

36. Woebken D, Burow LC, Behnam F, Mayali X, Schintlmeister A, Fleming ED, et al.

Revisiting N2 fixation in Guerrero Negro intertidal microbial mats with a functional single-

cell approach. ISME J. 2015 Feb;9(2):485–96.

37. Suzuki Y, Yamashita S, Kouduka M, Ao Y, Mukai H, Mitsunobu S, et al. Deep microbial

proliferation at the basalt interface in 33.5–104 million-year-old oceanic crust. Commun

Biol. 2020 Apr 2;3(1):1–9.

38. Dekas AE, Parada AE, Mayali X, Fuhrman JA, Wollard J, Weber PK, et al. Characterizing

Chemoautotrophy and Heterotrophy in Marine Archaea and Bacteria With Single-Cell Multi-

isotope NanoSIP. Front Microbiol [Internet]. 2019 [cited 2023 Jun 7];10. Available from:

https://www.frontiersin.org/articles/10.3389/fmicb.2019.02682

39. Arandia-Gorostidi N, Parada AE, Dekas AE. Single-cell view of deep-sea microbial activity

and intracommunity heterogeneity. ISME J. 2022 Oct 6;1–11.

40. Meyer NR, Fortney JL, Dekas AE. NanoSIMS sample preparation decreases isotope

enrichment: magnitude, variability and implications for single-cell rates of microbial activity.

Environ Microbiol. 2021;23(1):81–98.

41. Stryhanyuk H, Calabrese F, Kümmel S, Musat F, Richnow HH, Musat N. Calculation of

Single Cell Assimilation Rates From SIP-NanoSIMS-Derived Isotope Ratios: A

Comprehensive Approach. Front Microbiol [Internet]. 2018 [cited 2023 Mar 31];9. Available

from: https://www.frontiersin.org/articles/10.3389/fmicb.2018.02342

42. Musat N, Stryhanyuk H, Bombach P, Adrian L, Audinot JN, Richnow HH. The effect of

FISH and CARD-FISH on the isotopic composition of 13C- and 15N-labeled Pseudomonas

putida cells measured by nanoSIMS. Syst Appl Microbiol. 2014 Jun 1;37(4):267–76.

43. Haouari O, Fardeau ML, Cayol JL, Fauque G, Casiot C, Elbaz-Poulichet F, et al.

Thermodesulfovibrio hydrogeniphilus sp. nov., a new thermophilic sulphate-reducing

bacterium isolated from a Tunisian hot spring. Syst Appl Microbiol. 2008 Mar 19;31(1):38–

42.

44. Miller HM, Chaudhry N, Conrad ME, Bill M, Kopf SH, Templeton AS. Large carbon

isotope variability during methanogenesis under alkaline conditions. Geochim Cosmochim

Acta. 2018 Sep 15;237:18–31.

45. Lester Jr. W, Sun SH, Seber A. Observations on the Influence of Deuterium on Bacterial

Growth*. Ann N Y Acad Sci. 1960;84(16):667–77.

46. Pernice M, Dunn SR, Tonk L, Dove S, Domart-Coulon I, Hoppe P, et al. A nanoscale

secondary ion mass spectrometry study of dinoflagellate functional diversity in reef-building

corals. Environ Microbiol. 2015;17(10):3570–80.

47. Englander SW, Sosnick TR, Englander JJ, Mayne L. Mechanisms and uses of hydrogen

exchange. Curr Opin Struct Biol. 1996 Feb;6(1):18–23.

48. Bai Y, Milne JS, Mayne L, Englander SW. Primary structure effects on peptide group

hydrogen exchange. Proteins Struct Funct Bioinforma. 1993;17(1):75–86.

49. Cui D, Kong L, Wang Y, Zhu Y, Zhang C. In situ identification of environmental

microorganisms with Raman spectroscopy. Environ Sci Ecotechnology. 2022 Jul

1;11:100187.

50. Mosier-Boss PA. Review on SERS of Bacteria. Biosensors. 2017 Dec;7(4):51.

51. Wang K, Chen L, Ma X, Ma L, Chou KC, Cao Y, et al. Arcobacter Identification and

Species Determination Using Raman Spectroscopy Combined with Neural Networks. Appl

Environ Microbiol. 2020 Oct;86(20):e00924-20.

52. Novelli-Rousseau A, Espagnon I, Filiputti D, Gal O, Douet A, Mallard F, et al. Culture-free

Antibiotic-susceptibility Determination From Single-bacterium Raman Spectra. Sci Rep.

2018 Mar 2;8(1):3957.

53. Klein F, Humphris SE, Guo W, Schubotz F, Schwarzenbach EM, Orsi WD. Fluid mixing

and the deep biosphere of a fossil Lost City-type hydrothermal system at the Iberia Margin.

Proc Natl Acad Sci. 2015 Sep 29;112(39):12036–41.

54. Hatzenpichler R, Krukenberg V, Spietz RL, Jay ZJ. Next-generation physiology approaches

to study microbiome function at single cell level. Nat Rev Microbiol. 2020 Apr;18(4):241–

56.

55. Cui L, Yang K, Zhou G, Huang WE, Zhu YG. Surface-Enhanced Raman Spectroscopy

Combined with Stable Isotope Probing to Monitor Nitrogen Assimilation at Both Bulk and

Single-Cell Level. Anal Chem. 2017 Jun 6;89(11):5793–800.

56. Ivleva NP, Kubryk P, Niessner R. Raman microspectroscopy, surface-enhanced Raman

scattering microspectroscopy, and stable-isotope Raman microspectroscopy for biofilm

characterization. Anal Bioanal Chem. 2017 Jul 1;409(18):4353–75.

57. Wang Y, Huang WE, Cui L, Wagner M. Single cell stable isotope probing in microbiology

using Raman microspectroscopy. Curr Opin Biotechnol. 2016 Oct 1;41:34–42.

58. Weber F, Zaliznyak T, Edgcomb VP, Taylor GT. Using Stable Isotope Probing and Raman

Microspectroscopy To Measure Growth Rates of Heterotrophic Bacteria. Appl Environ

Microbiol. 2021 Oct 28;87(22):e01460-21.

59. Wang Y, Xu J, Kong L, Liu T, Yi L, Wang H, et al. Raman–deuterium isotope probing to

study metabolic activities of single bacterial cells in human intestinal microbiota. Microb

Biotechnol. 2020;13(2):572–83.

60. Chisanga M, Muhamadali H, Ellis DI, Goodacre R. Surface-Enhanced Raman Scattering

(SERS) in Microbiology: Illumination and Enhancement of the Microbial World. Appl

Spectrosc. 2018 Jul 1;72(7):987–1000.

61. Chisanga M, Muhamadali H, Kimber R, Goodacre R. Quantitative detection of isotopically

enriched E. coli cells by SERS. Faraday Discuss. 2017 Nov 30;205(0):331–43.

62. Efrima S, Zeiri L. Understanding SERS of bacteria. J Raman Spectrosc. 2009;40(3):277–88.

63. Mosier-Boss PA, Sorensen KC, George RD, Obraztsova A. SERS substrates fabricated using

ceramic filters for the detection of bacteria. Spectrochim Acta A Mol Biomol Spectrosc.

2016 Jan 15;153:591–8.

64. Kubryk P, Kölschbach JS, Marozava S, Lueders T, Meckenstock RU, Niessner R, et al.

Exploring the Potential of Stable Isotope (Resonance) Raman Microspectroscopy and

Surface-Enhanced Raman Scattering for the Analysis of Microorganisms at Single Cell

Level. Anal Chem. 2015 Jul 7;87(13):6622–30.

65. Hu F, Shi L, Min W. Biological imaging of chemical bonds by stimulated Raman scattering

microscopy. Nat Methods. 2019 Sep;16(9):830–42.

66. Cicerone M. Molecular imaging with CARS micro-spectroscopy. Curr Opin Chem Biol.

2016 Aug 1;33:179–85.

67. Camp Jr CH, Cicerone MT. Chemically sensitive bioimaging with coherent Raman

scattering. Nat Photonics. 2015 May;9(5):295–305.

68. Ivleva NP, Wagner M, Horn H, Niessner R, Haisch C. Raman microscopy and surface-

enhanced Raman scattering (SERS) for in situ analysis of biofilms. J Biophotonics.

2010;3(8–9):548–56.

69. Weiss R, Palatinszky M, Wagner M, Niessner R, Elsner M, Seidel M, et al. Surface-

enhanced Raman spectroscopy of microorganisms: limitations and applicability on the

single-cell level. Analyst. 2019;144(3):943–53.

70. Kilburn MR, Wacey D. CHAPTER 1 Nanoscale Secondary Ion Mass Spectrometry

(NanoSIMS) as an Analytical Tool in the Geosciences. 2014;1–34.