Using splines for point spread function calibration at non-uniform depths in localization microscopy

Serafim Korovin[1,*]**, Dylan Kalisvaart**[1]**, Shih-Te Hung**[1]**, Jelmer Cnossen**[1]**, Coen de Visser**[2]**, and Carlas S. Smith**[1,*]

[1]Delft University of Technology, Delft Center for Systems and Control, 2628 CN Delft, the Netherlands
[2]Delft University of Technology, Control and Simulation Section, 2629 HS Delft, the Netherlands
[*]Correspondence: s.korovin@tudelft.nl or c.s.smith@tudelft.nl

ABSTRACT

Single-molecule localization microscopy methods extensively leverage the microscope point spread function (PSF) for fitting the molecules. Calibrating an accurate PSF model is especially difficult in the presence of depth-dependent aberrations which alter the PSF shape depending on the imaging depth. The aberrations at depths of a few micrometers become substantial enough to considerably impoverish the conventional calibration methods' performance. In our work, we propose a novel spline model which enables the depth-dependent PSF model calibration by interpolating between the beads at arbitrary depths. We show that diffspline reduces the PSF intensity overestimation by 67.8 percentage points and underestimation by 21.8 percentage points. Moreover, it eliminates the depth-dependent bias and improves the localization precision two-fold compared to previous approaches.

Introduction

Recent years have seen the emergence of single-molecule localization microscopy (SMLM)[1–3] as a powerful tool for resolving three-dimensional (3D) biological structures[4–6] beyond the diffraction limit. This technique relies on the sparse excitation of photoswitchable fluorophores to create isolated molecule images. The images contain molecule shapes blurred with the microscope emission pattern, or the point spread function (PSF), which can be experimentally designed and modelled to encode additional positional information in its shape[5,7,8]. The molecule positions are determined through maximum likelihood estimation (MLE)[9] or maximum a posteriori (MAP) estimation[10] of the position parameters within the point spread function (PSF) model. Therefore, creating a flexible PSF model that accurately represents the true experimental PSF is crucial for the ultimate localization accuracy and precision[11].

To accurately model the PSF, piece-wise polynomials called spline functions[12] were employed. Specifically, cubic splines (csplines)[13] have been shown to enable the estimation precision that achieves the Cramèr-Rao lower bound (CRLB)[14,15], the theoretical bound on the highest possible localization precision of unbiased estimators. By deriving the PSF model directly from the experimental data, splines provide an accurate PSF representation for varying PSF shapes[16].

Spline models are conventionally calibrated on image frames obtained by scanning through the bead sample along the Z-axis. The resulting Z-stack of images contains fluorescent beads located on the coverslip. However, when imaging deeper into the sample, depth-induced spherical aberrations emerge and become exacerbated in the presence of the refractive index mismatch between the objective lens immersion oil and sample medium[17]. These aberrations lead to considerable errors in PSF shapes, even for depths of only a few micrometers[17].

Several methods have been proposed for addressing the challenges posed by depth-induced aberrations. Active aberration correction methods[18,19] aim to estimate the aberrations directly from the wavefront, but these methods require complex adaptive optics hardware and are difficult to operate for non-experts. Other approaches use phase-retrieved PSF methods[20–24] to construct a pupil function that describes the magnitude and the phase of the emission light wavefront. However, such methods require specific imaging system parameters (e.g., refractive indices and emission wavelengths) to build the pupil function, and thus are susceptible to parameter deviations in the form of noise or optical imperfections.

In turn, a cspline-based approach[25] was devised to correct depth-induced aberrations without any requirements for optical parameters or hardware. It estimates the cspline bias for every depth and subtracts the bias from the localization estimates. However, since the correction is only applied to the final estimates directly, the PSF model remains incorrect. This means that the localization precision is unaffected by the correction. Therefore, depth information must be incorporated into the spline PSF model for localization to be accurate and precise.

An alternative strategy is to extend the cspline model from three-dimensional (3D) to four-dimensional (4D) by interpolating

between the point spread functions (PSFs) at various depths. However, this approach is computationally expensive, as the number of coefficients for which the spline has to be calibrated increases exponentially with the spline dimensionality. This issue was alleviated with the multivariate B-spline[26] which employs basis functions to obtain the interpolation, thus reducing the coefficient complexity to depend only on the number of voxels present in the PSF image. However, applying B-splines to depth-dependent PSF calibration is challenging, since B-splines assume uniformly spaced calibration points in all dimensions, while the actual bead depths in the calibration Z-stack are not uniformly spread.

To this end, we propose a depth-dependent spline PSF model formulated as a non-uniform Catmull-Rom spline. Similar to the B-spline, the Catmull-Rom spline[27] reduces the coefficient complexity as it builds a spline function based on pixel values. Furthermore, Barry and Goldman[28] proposed a recursive algorithm that allowed the Catmull-Rom spline to be built on arbitrarily spaced points. By applying Barry and Goldman's definition of the Catmull-Rom spline, we were able to build a spline PSF model capable of interpolating between beads at non-equidistant depths.

The resulting non-uniform Catmull-Rom spline, which we name Depth-Interpolating Fitting Function spline (diffspline), is evaluated against the cspline approach on simulated beads at various depths. We compared the diffspline with the cspline in terms of the PSF shapes and localization errors (bias, precision, and CRLB). We demonstrated that the diffspline PSF produces a more accurate PSF shape, with the intensity overestimation reduced by 67.8 percent points and with the underestimation reduced by 21.8 percent points. The effectiveness of the diffspline for deep sample localization was demonstrated by achieving an approximately two-fold improvement in localization precision compared to the coverslip-calibrated cspline. Moreover, the precision of our model closely followed its CRLB, suggesting that the diffspline delivers the correct CRLB for the evaluated depth.

Figure 1. Schematic overview of this study. (a) Depending on the depth of the imaged beads, their 3D PSFs change. On the right side, the PSF shapes are illustrated as intensity profiles in *ZX* with a fixed cutoff for schematic purposes. (b) 3D cspline approach creates a smooth PSF model by interpolating between the given intensity values. The 3D cspline model is calibrated on the bead data from the coverslip. (c) 4D spline can also interpolate between the bead data from different arbitrary depths. This way, we can obtain 3D PSF models from the depths that are not present in the data. (d) Depth-dependent 4D spline interpolation produces more accurate PSF models, and thus results in more accurate z-localizations.

Simulated pipeline

Our approach was verified on simulated data. We devised the following pipeline to generate PSF Z-stacks for any given depth.

Figure 2. Overview of the processing pipeline that we simulate in our work. (a) The simulated sample contains beads at multiple depths. The sample stage movement along the X and Y axis is included in the simulation. (b) The simulated bead sample is moved in both X and Y planes and scanned at various positions. In this manner, we can acquire different realizations of the same bead across the field of view. (c) The obtained realizations are aligned and averaged into one PSF.

First, we simulate a 3D bead sample by generating PSF Z-stacks at the given depths using a vectorial model[29]. The bead PSFs at different depths are simulated by adding first- and second-order spherical aberrations, as proposed by Booth et al.[17]. Each Z-stack contains one bead, which is assigned a random intensity between 5000 and 15000 photons, and a constant background of 50 photons/pixel. Moreover, we simulate the shot noise in the Z-stack with a Poisson distribution[9]. To improve the PSF calibration fidelity from the noisy Z-stacks, we obtain several realizations of the same bead. As depicted in Figure 2 (a, b), we achieve this by changing the field of view (FOV) of the images by moving the bead sample along the X and Y axes. This effect is simulated by adding a random position phase in the pupil phase function of the vectorial model. In our simulations, we use the emission wavelength $\lambda = 680$ nm, an astigmatic PSF with vertical astigmatism $Z_2^2 = 147m\lambda$, the refractive index for the sample medium is 1.33 and for the immersion oil is 1.52. The simulated voxel size is 100 nm in X and Y axes, and the Z-range spans from -500 nm to 500 nm with 10 nm step.

Subsequently, we apply the alignment and processing methods proposed by Li et al.[13]. First, all PSFs are gathered and aligned around a reference bead[13]. Next, the PSFs are divided into groups containing realizations of the same bead. Then, for each group, we apply the processing by Li et al.[13] wherein we average the bead PSFs, eliminate the background by Z-stack minimum subtraction, normalize the PSF amplitude by the cumulative intensity of the central axial slice, and smooth the PSF shape in Z-direction with a smoothing B-spline. As an outcome of the processing, which is depicted in Figure 2 (c), we obtain a processed PSF Z-stack for each simulated bead. This Z-stack is the output of the simulation pipeline.

Spline calibration

We calibrate spline models on the processed PSF Z-stacks. The cspline is calibrated on the simulated PSF Z-stack on the coverslip (0 nm depth)[13].

For the diffspline calibration, we first generate simulated beads at depths d_i, which are specified separately for each experiment. To calibrate the diffspline at a specific depth d_{eval}, we interpolate between four consecutive PSFs with depths d_i closest to d_{eval}, i.e. we find i s.t. $d_i : d_{i-1} < d_i \leq d_{eval} < d_{i+1} < d_{i+2}$). The interpolation is performed with the non-uniform Catmull-Rom spline, as described in Supplementary Note S1. We further discuss the effects of the Poisson noise on the spline calibration in Supplementary Note S1. Thus, we obtain an interpolated 3D PSF model for the required depth. Since csplines have an established localization methodology[16], we convert the interpolated 3D Catmull-Rom spline into a cspline form for a direct comparison between the two spline types. The conversion is performed by computing the spline coefficients with Catmull-Rom and saving them in a cspline model.

Then, we evaluate the converted model by using it to localize on the simulated data. To do so, we first use the vectorial model to generate PSFs at the image center in XY for a given depth with an intensity of 10000 photons and a constant background of

50 photons/pixel. Next, we create 100 realizations of each Z-position by applying the shot noise to the PSF. Following that, we
obtain the maximum likelihood estimates of the emitter Z-positions from the generated frames with the Levenberg-Marquardt
algorithm[30] using the damping factor = 10 and 100 iterations. Finally, we calculate the localization bias and precision from the
estimates given the known ground truth. The bias is calculated as the mean absolute error, and the precision is calculated as the
standard deviation of the estimates over 100 frames. For the PSF shape evaluation, we employ the pixel-wise relative difference
between the target PSF and the ground truth PSF, which is the normalized version of the pixel-wise difference used by Babcock
et al.[16]. It is calculated as the difference between the target and ground truth, then divided by the ground truth pixel values.
This metric, multiplied by 100, describes the percentage of deviation from the ground-truth reference.

Results

Several experiments were conducted to quantify the diffspline performance. First, we compared it with the cspline approach
by evaluating the splines on the beads that were present in the z-stack ("on-bead evaluation"). This tested the capabilities of
the splines as PSF models without interpolation effects. We evaluated the spline PSF shapes and localization metrics, such as
bias and precision. Furthermore, we tested the interpolation capabilities of the diffspline model. We calculated the diffspline
localization bias and precision for depths between uniformly and nonuniformly spaced beads. Moreover, we investigated the
influence of the distance between beads on the resulting localization errors.

On-bead evaluation

96 First, we compared the cspline approach with our method by evaluating them on PSFs at 6000 nm depth. The simulation
97 pipeline was leveraged to generate PSFs at 0 nm depth (coverslip), which was used for the cspline model calibration, and
98 5900, 6000, 6100, 6200 nm, which were used for the diffspline calibration. Both models were evaluated at depth $d = 6000$ nm.

Figure 3. PSF visualization - comparing cspline and diffspline PSFs to the simulated ground truth (GT) at the 6000 nm depth
for the bead intensity = 10000 photons and the background of 50 photons/pixel. The first three columns show the PSF intensity
in X and Y, and the last two columns show the respective pixel-wise difference from the ground truth in percentages. Since the
cspline is calibrated on the coverslip, it fails to capture the depth-dependent aberrations at larger depths and therefore
considerable shape differences with the simulated ground truth occur.

Figure 4. Diffspline localization performance quantification. (a, b) We compare the coverslip-calibrated cspline and diffspline approaches at 6000 nm depth by measuring their localization biases (average absolute errors) and precisions (standard deviations of errors) with CRLBs (theoretical lower bound of the standard deviations). (c) Estimations produced by the cspline model calibrated on the coverslip. (d) Estimations by the diffspline calibrated on the simulated deep bead sample. (e) Diffspline and cspline precisions were compared over varying intensities at different depths. (f) Diffspline localization error distributions are plotted for different distances between the calibration beads at 1000 nm and 4000 nm depths. (g, h) Evaluating the diffspline interpolation between the uniformly (g) and non-uniformly (h) distributed beads against the cspline approach.

Figure 3 shows that the 4D diffspline was able to capture the PSF shape more accurately than the cspline, which had significant shape deformations when compared to the ground truth (GT). The largest negative relative difference, or under-estimation, in the pixel-wise relative intensity difference between the cspline and the GT being -63.5% at $Z = -500$, and the largest positive difference, or overestimation, is 76.3% at $Z = 500$ (see Supplementary Note S1). In contrast, the largest under- and overestimation for the diffspline were -41.7% at $Z = 500$ and 8.5% at $Z = -500$. Therefore, we observed that the largest intensity overestimation was reduced by 67.8 percent points, whereas the largest underestimations were reduced by 21.8 percent points. Moreover, as can be observed in Figure 3, the diffspline underestimations followed the general PSF shape, and thus could be attributed to an insufficient shape amplitude rather than an incorrect lateral shape profile.

For the estimation results, the coverslip-calibrated cspline at the depth of 6000 nm displayed a large bias, as can be observed in Figure 4(a). Although aberration-correction[25] can be employed to correct the depth-induced bias in the cspline approach, the precision remains unaddressed, which results in highly imprecise localizations, as shown in Figure 4(b). In contrast, the diffspline approach displays a consistently low bias of approximately 10 nm on average. Moreover, Figure 4(b) demonstrates that the diffspline precision remains consistently low at approximately 12 nm on average, while it stays close to the CRLB with the difference not exceeding 4 nm.

Furthermore, we demonstrate the precision advantages of our method by comparing cspline and diffspline precisions. We evaluated the cspline and diffspline models on frames with bead intensities varying from 1000 to 5000 photons with a constant background of 50 photons/pixel. The evaluation was performed by calculating the precision averaged over the Z-range. Figure 4(e) shows that the diffspline displays lower precision for all depths and intensities than the cspline. The diffspline precision is lower by 60 nm on average when evaluated at a 6000 nm depth with a 1000 photon intensity and by approximately 45 nm on average with 5000 photon intensity. This precision difference is lower when evaluated on the coverslip, where the difference of 10 nm at 1000 photon intensity is almost eliminated with an increase in intensity to 5000 photons. The lack of a similar decrease for the 6000 nm depth suggests that the model mismatch between the real PSF and the cspline prevents the latter from localizing more precisely.

To provide a more detailed look into the distribution of localization estimates over depth, we evaluated cspline and diffspline models at every depth from 0 to 6000 nm with a 100 nm step. For every depth, we estimated Z-positions for Z-range from -500 to 500 nm with a 10 nm step for 100 simulated frames, and averaged the predictions over the frames. The resulting Z-position estimates were plotted against the ground truth depths, as depicted in Figures 4(c, d). Each color represents a separate evaluation depth, e.g. the rightmost green line represents the estimates for the evaluation at a 6000 nm depth, thus showing the estimates from which the bias and precision in Figures 4 (a, b) were calculated. We can observe that the cspline estimates tend to collapse to the $Z = 0$ or $Z = -500$ regions at larger depths, which results in the bias increase outside those regions, whereas the diffspline estimations remain unaltered throughout all depths.

Interpolation evaluation

Next, we evaluated the effects of the distance between calibration beads on the resulting diffspline localization performance. For each "reference" depth $\rho \in \{1000, 4000\}$ nm and for each bead distance $\delta \in \{100, 300, 500, 700, 1000\}$ nm, we generated PSFs at depths $\rho - \delta, \rho, \rho + \delta, \rho + 2 \cdot \delta$, thus obtaining the four required calibration PSFs. The evaluation was performed at 10 evaluation depths uniformly spaced between ρ and $\rho + \delta$. For each evaluation depth, the errors between the Z-estimates and the ground truth were calculated. The errors were combined and depicted in the boxplot shown in Figure 4(f). We observed an increase in the mean error by 5 nm with an increase in the bead distance from 100 to 1000 nm, as well as a slight increase in the standard deviation by 4 nm.

Finally, we compared the localization bias of the cspline and diffspline calibrated on uniformly and nonuniformly spaced beads. For the uniform case, we generated calibration PSFs at depths 4000, 5000, 6000, 7000, 8000 nm, and for the non-uniform case - at depths 4500, 5600, 6366, 7277, 7877, 8400 nm. We then evaluated both splines in the depth range from the second to penultimate depth of the respective calibration PSFs. We chose five equidistant interpolation points between each consecutive pair of depths and evaluated the splines at these depths. Figures 4(g, h) show that the diffspline PSF shows a consistently low bias of approximately 12 nm and a precision of around 11 nm on average in both the uniform and non-uniform cases. The cspline estimation bias and precision both rise with an increase in the estimation depth, with the bias ranging from 150 to 250 nm and the precision rising from 15 to 100 nm.

Conclusion

In our work, we propose the diffspline model that enables PSF calibration from bead data at nonuniformly spaced depths. The model reduced the intensity overestimation by up to 67.8 percent points and the underestimation by up to 21.8 percent points at larger depths. This, in turn, led to a consistently low bias of 12 nm for the diffspline, and its precision closely followed the CRLB, with the differences not exceeding 4 nm across the entire Z-range.

The diffspline-interpolated models were shown to produce estimation results on par with the on-bead evaluation, thus enabling accurate and precise PSF calibration at arbitrary depths between those present in the bead stack. We also showed that an increase in the distance between the calibration beads influences the diffspline localization errors. The increase in the bead distance from 100 to 1000 nm induces an estimation bias of approximately 5 nm owing to interpolation errors.

Moreover, our approach reduced the computational complexity of the spline calibration from $O(4^d V)$ to $O(V)$ as we demonstrated in Supplementary Note S1, which suggests potential applicability for higher-dimensional splines. However, one of the limitations of our current approach is the reduction in the continuity requirement from C^2 in cspline to C^1 in diffspline. Future research could be directed at including the second-order continuity into the Catmull-Rom spline by expanding the spline order to 5. This way, a quintic Catmull-Rom spline would include extra smoothness requirements in its definition and thus have extended interpolation capacity which could alleviate the interpolation-induced errors.

References

1. Betzig, E. *et al.* Imaging intracellular fluorescent proteins at nanometer resolution. *Science* **313**, 1642–1645, DOI: 10.1126/science.1127344 (2006). https://www.science.org/doi/pdf/10.1126/science.1127344.

2. Rust, M., Bates, M. & Zhuang, X. Sub-diffraction-limit imaging by stochastic optical reconstruction microscopy (storm). *Nat. methods* **3**, 793–5, DOI: 10.1038/nmeth929 (2006).

3. Huang, B., Bates, M. & Zhuang, X. Super-resolution fluorescence microscopy. *Annu. Rev. Biochem.* **78**, 993–1016, DOI: 10.1146/annurev.biochem.77.061906.092014 (2009). PMID: 19489737, https://doi.org/10.1146/annurev.biochem.77.061906.092014.

4. von Diezmann, L., Shechtman, Y. & Moerner, W. E. Three-dimensional localization of single molecules for super-resolution imaging and single-particle tracking. *Chem. Rev.* **117**, 7244–7275, DOI: 10.1021/acs.chemrev.6b00629 (2017). PMID: 28151646, https://doi.org/10.1021/acs.chemrev.6b00629.

5. Pavani, S. R. P. *et al.* Three-dimensional, single-molecule fluorescence imaging beyond the diffraction limit by using a double-helix point spread function. *Proc. Natl. Acad. Sci.* **106**, 2995–2999, DOI: 10.1073/pnas.0900245106 (2009). https://www.pnas.org/doi/pdf/10.1073/pnas.0900245106.

6. Huang, F. *et al.* Ultra-high resolution 3d imaging of whole cells. *Cell* **166**, 1028–1040, DOI: https://doi.org/10.1016/j.cell.2016.06.016 (2016).

7. Huang, B., Wang, W., Bates, M. & Zhuang, X. Three-dimensional super-resolution imaging by stochastic optical reconstruction microscopy. *Science* **319**, 810–813, DOI: 10.1126/science.1153529 (2008). https://www.science.org/doi/pdf/10.1126/science.1153529.

8. Shechtman, Y., Sahl, S. J., Backer, A. S. & Moerner, W. E. Optimal point spread function design for 3d imaging. *Phys. Rev. Lett.* **113**, 133902, DOI: 10.1103/PhysRevLett.113.133902 (2014).

9. Smith, C., Joseph, N., Rieger, B. & Lidke, K. Fast, single-molecule localization that achieves theoretically minimum uncertainty. *Nat. methods* **7**, 373–5, DOI: 10.1038/nmeth.1449 (2010).

10. Kalisvaart, D. *et al.* Precision in iterative modulation enhanced single-molecule localization microscopy. *Biophys. J.* **121**, 2279–2289, DOI: 10.1016/j.bpj.2022.05.027 (2022).

11. van Dijk, R., Kalisvaart, D., Cnossen, J. & Smith, C. S. Bayesian posterior density estimation reveals degeneracy in three-dimensional multiple emitter localization. *Sci. Reports* **13**, DOI: 10.1038/s41598-023-49101-5 (2023).

12. Holden, S. J., Uphoff, S. & Kapanidis, A. N. Daostorm: An algorithm for high- density super-resolution microscopy. *Nat. Methods* **8**, 279–280, DOI: 10.1038/nmeth0411-279 (2011).

13. Li, Y. *et al.* Real-time 3d single-molecule localization using experimental point spread functions. *Nat. Methods* **15**, 367–369, DOI: 10.1038/nmeth.4661 (2018).

14. Kay, S. M. *Fundamentals of statistical signal processing: estimation theory* (Prentice-Hall, Inc., 1993).

15. Ober, R., Ram, S. & Ward, E. Localization accuracy in single-molecule microscopy. *Biophys. journal* **86**, 1185–200, DOI: 10.1016/S0006-3495(04)74193-4 (2004).

16. Babcock, H. P. & Zhuang, X. Analyzing single molecule localization microscopy data using cubic splines. *Sci. Reports* **7**, 552, DOI: 10.1038/s41598-017-00622-w (2017).

17. Booth, M., Neil, M. & Wilson, T. Aberration correction for confocal imaging in refractive-index-mismatched media. *J. Microsc.* **192**, 90–98, DOI: https://doi.org/10.1111/j.1365-2818.1998.99999.x (1998). https://onlinelibrary.wiley.com/doi/pdf/10.1111/j.1365-2818.1998.99999.x.

18. Burke, D., Patton, B., Huang, F., Bewersdorf, J. & Booth, M. J. Adaptive optics correction of specimen-induced aberrations in single-molecule switching microscopy. *Optica* **2**, 177–185, DOI: 10.1364/OPTICA.2.000177 (2015).

19. Mlodzianoski, M. *et al.* Active psf shaping and adaptive optics enable volumetric localization microscopy through brain sections. *Nat. Methods* **15**, DOI: 10.1038/s41592-018-0053-8 (2018).

20. Hanser, B. M., Gustafsson, M. G. L., Agard, D. A. & Sedat, J. W. Phase-retrieved pupil functions in wide-field fluorescence microscopy. *J. Microsc.* **216**, 32–48, DOI: https://doi.org/10.1111/j.0022-2720.2004.01393.x (2004). https://onlinelibrary.wiley.com/doi/pdf/10.1111/j.0022-2720.2004.01393.x.

21. McGorty, R., Schnitzbauer, J., Zhang, W. & Huang, B. Correction of depth-dependent aberrations in 3d single-molecule localization and super-resolution microscopy. *Opt. Lett.* **39**, 275–278, DOI: 10.1364/OL.39.000275 (2014).

22. Liu, S., Kromann, E. B., Krueger, W. D., Bewersdorf, J. & Lidke, K. A. Three dimensional single molecule localization using a phase retrieved pupil function. *Opt. Express* **21**, 29462–29487, DOI: 10.1364/OE.21.029462 (2013).

23. Aristov, A., Lelandais, B., Rensen, E. & Zimmer, C. Zola-3d allows flexible 3d localization microscopy over an adjustable axial range. *Nat. Commun.* **9**, DOI: 10.1038/s41467-018-04709-4 (2018).

24. Xu, F. *et al.* Three-dimensional nanoscopy of whole cells and tissues with in situ point spread function retrieval. *Nat. Methods* **17**, 531–540, DOI: 10.1038/s41592-020-0816-x (2020).

25. Li, Y., Wu, Y.-L., Hoess, P., Mund, M. & Ries, J. Depth-dependent psf calibration and aberration correction for 3d single-molecule localization. *Biomed. Opt. Express* **10**, 2708–2718, DOI: 10.1364/BOE.10.002708 (2019).

26. Li, M. *et al.* Fast and universal single molecule localization using multi-dimensional point spread functions. *bioRxiv* DOI: 10.1101/2023.10.17.562517 (2023). https://www.biorxiv.org/content/early/2023/10/20/2023.10.17.562517.full.pdf.

27. Catmull, E. & Rom, R. A class of local interpolating splines. In BARNHILL, R. E. & RIESENFELD, R. F. (eds.) *Computer Aided Geometric Design*, 317–326, DOI: https://doi.org/10.1016/B978-0-12-079050-0.50020-5 (Academic Press, 1974).

28. Barry, P. J. & Goldman, R. N. A recursive evaluation algorithm for a class of catmull-rom splines. *SIGGRAPH Comput. Graph.* **22**, 199–204, DOI: 10.1145/378456.378511 (1988).

29. Siemons, M., Hulleman, C. N., Thorsen, R. Ø., Smith, C. S. & Stallinga, S. High precision wavefront control in point spread function engineering for single emitter localization. *Opt. Express* **26**, 8397–8416, DOI: 10.1364/OE.26.008397 (2018).

30. Marquardt, D. W. An algorithm for least-squares estimation of nonlinear parameters. *J. Soc. for Ind. Appl. Math.* **11**, 431–441 (1963).

31. Lay, D., Lay, S. & McDonald, J. *Linear Algebra and Its Applications* (Pearson Education, 2015).

32. Verhaegen, M. & Verdult, V. *Filtering and System Identification: A Least Squares Approach* (Cambridge University Press, 2007).

33. Korovin, S. *et al.* Data underlying the publication: Using splines for point spread function calibration at non-uniform depths in localization microscopy, DOI: 10.4121/5f75fadc-8f80-4e27-a049-1137031b19c0 (2024).

34. Korovin, S. *et al.* Software underlying the publication: Using splines for point spread function calibration at non-uniform depths in localization microscopy (2024).

Author contributions statement

S. K. wrote the code, conducted experiments, and wrote the manuscript. D.K. and S.H. provided feedback on the results and writing. J.C. conducted the derivations and developed the code for the 3D Catmull-Rom spline model. C.V. provided feedback on the writing. C.S.S. supervised the project.

Additional information

Acknowledgements   S.K., D.K., S.H., J.C., and C.S.S. were supported by the Netherlands Organisation for Scientific Research (NWO), under NWO START-UP project no. 740.018.015.

Competing interests   The authors declare no competing interests.

Supporting citations   References[31,32] appear in the supporting information.

Data and code availability   The data that support the findings of this study are openly available in 4TU.ResearchData[33] at doi.org/10.4121/5f75fadc-8f80-4e27-a049-1137031b19c0. The code that supports the findings of this study is openly available on GitHub[34] at https://github.com/qnano/diffspline.