Depth-dependent scaling of axial distances in light microscopy

S.V. LOGINOV,[1†] D.B. BOLTJE,[1,2,†,*] M.N.F. HENSGENS,[1] J.P. HOOGENBOOM[1] AND E.B. VAN DER WEE[1,‡,**]

[1]*Department of Imaging Physics, Delft University of Technology, Delft, The Netherlands*
[2]*Delmic B.V., Delft, The Netherlands*
[†]*These authors contributed equally to this work.*
[‡]*Current address: Cell Biology, Neurobiology and Biophysics, Department of Biology, Utrecht University, Utrecht, The Netherlands*
[*]*d.b.boltje@tudelft.nl,* [**]*e.b.vanderwee@uu.nl*

Abstract: In volume fluorescence microscopy, refractive index matching is essential to minimize aberrations. There are however, common imaging scenarios, where a refractive index mismatch (RIM) between immersion and sample medium cannot be avoided. This RIM leads to an axial deformation in the acquired image data. Over the years, different axial scaling factors have been proposed to correct for this deformation. While some reports have suggested a *depth-dependent* axial deformation, so far none of the scaling theories has accounted for a depth-dependent, non-linear scaling. Here, we derive an analytical theory based on determining the leading constructive interference band in the objective lens pupil under RIM. We then use this to calculate a depth-dependent re-scaling factor as a function of the numerical aperture (NA), the refractive indices n_1 and n_2, and the wavelength λ. We compare our theoretical results with wave-optics calculations and experimental results obtained using a novel measurement scheme for different values of NA and RIM. As a benchmark, we recorded multiple datasets in different RIM conditions, and corrected these using our depth-dependent axial scaling theory. Finally, we present an online web applet that visualizes the depth-dependent axial re-scaling for specific optical setups. In addition, we provide software which will help microscopists to correctly re-scale the axial dimension in their imaging data when working under RIM.

1. Introduction

Optical sectioning has enabled imaging of large volumes by fluorescence microscope (FM), as realized in confocal, two-photon microscopy [1], structured illumination microscopy (SIM) [2] and light sheet microscopy [3]. For optimal imaging, aberrations need to be minimized by avoiding a RIM between the detection microscope objective immersion medium n_1 and specimen n_2 [1]. Failing to do so results in the blurring of the point spread function (PSF) of the microscope and therefore a loss in resolving power, as well as in a deformation of the recorded volume along the optical axis.

This axial deformation arises from the refraction of the peripheral rays on the RIM interface, causing an axial shift of their focal point with respect to the focal point of the paraxial rays [4]. This effect can be characterized using focal shift $\Delta f = \text{AFP} - \text{NFP}$, where AFP is the actual focal position (the real position of the object) and NFP is the apparent or nominal focal position (the microscope z-position where the object is found in focus) [1]. Just as the true lateral distances are recalculated back from an image using the magnification of the objective M, so ought the AFPs (i.e. the true axial distances) to be re-scaled using the re-scaling factor $\zeta = \text{AFP}/\text{NFP}$. The accurate knowledge of the re-scaling factor enables reliable quantitative volumetric microscopy.

While it is generally unfavorable to have a refractive index mismatch, for e.g. illumination intensities in confocal microscopy, there are common imaging scenarios in which a mismatch is still present in the most optimal configuration [1, 4]. For instance, as the resolving power of the microscope depends on the NA of the detection objective, high-NA oil immersion objectives

($n_1 = 1.52$) are used to image water-like specimens ($n_1 \approx 1.33$). In addition, embedding and fixation media rarely match exactly, in terms of refractive index, the immersion media of air, water, silicone, glycerol or oil objectives, leading to the axial deformation of the imaged volumes.

Refractive index mismatches are often more pronounced in integrated cryogenic fluorescence and/or correlative microscopy [5–8]. In such systems, light and electron microscopy are combined in a single setup, and air objectives are often used for imaging specimens with a higher refractive index. The use of air objectives is a straight-forward choice, as the specimen resides in a vacuum chamber for electron microscopy. In cryogenic fluorescence microscopy the specimen is cooled to temperatures below 120 K, which makes the use of a non-touching air objectives favorable, as any other (touching) immersion objectives are challenging in terms of engineering [9–12].

The development of the confocal microscope, along with its optical sectioning capability, allowed for 3D imaging, and hence the need for axial scaling theories arose. In the paraxial approximation, the axial distances are simply re-scaled using the ratio of the two refractive indices [13], which works well for low NA objectives. Visser *et al.* presented a scaling theory which is based on the contribution of the high-angle (or marginal) rays to the axial scaling [14]. More recently, several scaling theories have been presented, which result in re-scaling factors somewhere in between the paraxial and high-angle approximations [4, 15, 16]. The most accurate method to determine axial scaling are full wave-optics calculations of the microscope's point spread function under RIM [17–19]. As these calculations are computationally expensive and complex, they are hardly used by microscopists to calculate the axial re-scaling factor.

The re-scaling factor can also be measured experimentally by observing the NFP of a fluorescent bead or interface through a medium with refractive index n_2, whilst knowing the AFP of said bead or interface. This can be done by constructing a sample cell where the fluorescence is present far from the coverslip, which is later filled with a liquid with n_2 [4, 20, 21]. The AFP can be obtained through imaging with an objective where $n_1 = n_2$ [4, 20, 21] or by measuring Fabry Perrot fringes in the transmission spectrum of the cell [21]. Alternatively, the apparent axial deformation of a spherical object, larger than the PSF of the microscope, can be used to measure the axial scaling [14, 22]. Recently, a different approach was presented where a coverslip was step-wise coated with a low-index polymer ($n_2 \approx 1.33$) and the AFP was measured using stylus profilometry [23]. With this method, the axial scaling can be measured in the range of a few microns from the coverslip, while the former methods are used to measure tens of microns away from the coverslip.

While the explicit axial scaling theories in literature are all depth-independent, there exist some reports in literature that this factor is actually depth-dependent. In 1993, Hell *et al.* wrote *"...it can be expected that the regions close to the cover glass are slightly more scaled than those in deeper regions of the specimen."* [17]. Later, wave-optics calculations showed a non-linear dependence of the focal shift Δf on the imaging depth for high-NA objectives and large RIMs [18, 19]. For instance, Sheppard and Török reported a non-linear dependence of the focal shift Δf at a depth <30 µm from the coverslip using wave-optics calculations (NA = 1.3, n_1 = 1.52, n_2 = 1.33) [18], where the re-scaling factor was 5% larger close to the coverslip than at large distances. More recently, the measurements by Petrov *et al.* showed significant non-linear axial scaling for imaging depths < 4 µm [23].

There is, up to now, no straightforward equation which can be used to explicitly calculate the re-scaling factor as a function of depth. Moreover, a depth-dependence of the re-scaling factor has not been measured experimentally for large depths, and for several NAs and multiple RIM conditions.

Here, we present an analytical theory which calculates the depth-dependent re-scaling factor as a function of the NA, the refractive indices n_1 and n_2, and the wavelength λ. The gist of the theory is in the determination of the leading constructive interference band in the objective lens' pupil under RIM. We compare the theory to both full wave-optics calculations and experiments.

In the experiments, we have imaged the gap between two substrates that were brought closer to each other step by step. By filling the space between the substrates with a liquid with index n_2 we were able to measure the NFP, while the AFP was determined independently from the microscope by the piezo-stage holding the top substrate. We have measured the re-scaling factor ζ for several objectives with various NAs, immersion refractive index n_1, and sample refractive index n_2, with both RIMs where $n_1 < n_2$ and $n_1 > n_2$ in a wide range of depths and compared them to the analytical theory and the wave-optics calculations. We demonstrate that the axial re-scaling of 3D microscopy data, recorded with a refractive index mismatch, using the depth-dependent re-scaling factor outperforms the re-scaling using existing linear re-scaling theories from literature. Finally, we provide the reader with an online web applet where one can visualize the depth-dependent axial re-scaling factor for their specific optical setup and PYTHON software to re-scale data acquired under RIM.

2. Scaling of axial distances due to refractive index mismatch

2.1. Geometrical optics

Given a fluorescent object emitting light in an ideal, spherical fashion, a flat interface with a refractive index mismatch will cause disturbance to the fluorescent light propagation – in the form of refraction. When an objective lens is collecting light under RIM, this effect will be more pronounced with increasing NA of the objective lens (OL) (and therefore collection angle), see Figure 1a for the overall geometry.

This is further illustrated using geometrical optics in Figure 1b through d, where a light source is located $10\,\mu m$ away from the interface between n_1 and n_2. In Figure 1b no RIM is present ($n_1 = n_2 = 1.0$), where the undisturbed wave-fronts and rays are shown in solid blue and black respectively. The dotted horizontal line depicts the interface with the RIM. Figure 1c and d show two cases where RIMs are present respectively for $n_1 < n_2$ and $n_1 > n_2$. The black rays follow Snell's law when crossing the interface, resulting in the wave-fronts (solid blue) deviating from the ideal spherical shape (dashed blue). The caustic surface (purple) indicates where geometrical optics breaks down. From this simple illustration, we see that the effect of RIM becomes more pronounced with increasing collection angles (numerical aperture) and increasing RIM contrast $\Delta n = |n2 - n1|$.

When imaging with a microscope, there are two (independent) axial coordinates to be considered: the actual depth of the fluorescent emitter z_A, as measured from the RIM interface ($z_A = 10\,\mu m$ in Figure 1) and the nominal depth of the focal plane of the microscope into the sample z_N. A 3D image stack recorded with a microscope under RIM has z_N as an axial coordinate. Similar to using the magnification M to re-calculate the true lateral dimensions, the axial coordinates should be re-scaled using the re-scaling factor $\zeta \equiv z_A/z_N$ to obtain a 3D stack with the actual depth z_A as an axial coordinate.

Geometrical optics can produce several estimates for the re-scaling factor ζ. The paraxial rays are nicely focused, even under RIM, producing the estimate [13]:

$$\zeta_{\text{paraxial}} = n_2/n_1. \tag{1}$$

Another geometrical optics estimate comes from Visser *et al.*, obtained from the marginal rays still fitting into the NA [14]:

$$\zeta_{\text{marginal}} = \sqrt{\frac{(n_2^2 - \text{NA}^2)}{(n_1^2 - \text{NA}^2)}}. \tag{2}$$

The two re-scaling factor values ζ_{paraxial} and ζ_{marginal} are irreconcilable in a modern high-performance objective lenses (with $\text{NA} \rightarrow 0.95 \times n_1$). For the example shown in Figure 1c,

Fig. 1. The effects of RIM demonstrated with geometrical optics. **(a)** An OL imaging a fluorescent object (green star) with- and without refractive index mismatch, left and right respectively. The dotted boxes are detailed out in panels b through d. **(b)** A fluorescent object (green star) sits $z_A = 10\,\mu m$ deep from the refractive index interface (dotted line) between n_1 and n_2. With $n_1 = n_2 = 1.0$, rays (black) and wave-fronts (blue) are depicted. **(c)** In the case where $n_1 = 1.0, n_2 = 1.33$ the ideal spherical wave-fronts (dashed blue) are deformed (solid blue) after crossing the refractive index (RI) interface. Rays are depicted in red to indicate RIM and the caustical surface (purple) shows where geometrical optics breaks down. **(d)** In the opposite RIM case, where $n_1 = 1.52, n_2 = 1.33$, again the wave-fronts are deformed under RIM and total internal reflection occurs. The NA is 0.95 **(b, c)** and 1.4 **(d)**.

$\zeta_{paraxial} = 1.33$ and $\zeta_{marginal} = 2.98$, whilst for Figure 1d $\zeta_{paraxial} = 0.88$ and $\zeta_{marginal} \rightarrow 0$. Although these estimates are derived from geometrical optics, $\zeta_{paraxial}$ and $\zeta_{marginal}$ do provide lower and upper bounds when $n_1 < n_2$ (reverse when $n_1 > n_2$) and are still useful checks in the wave-optics treatment.

2.2. Analytical expression describing depth-dependent axial scaling

When considering the PSF of a wide-field microscope under RIM two intrinsic length scales need to be considered: (i) several strictly geometrical parameters n_1, n_2, z_A, z_N, & NA, and (ii) the physical parameter of the (vacuum) wavelength of light used $\lambda = 2\pi/k$ which is independent from the exact RIM geometry, where k is the wave number. With changing imaging depth, the geometrical parameters z_A, z_N do change, while the wavelength of light λ does not. Following the derivation outlined below, we find it is exactly the interplay between the two length scales that yields a depth-dependent re-scaling factor:

$$\zeta = \begin{cases} \min[\zeta_{univ}, \zeta_{crit}], & \text{if } n_2 \geq n_1 \\ \max[\zeta_{univ}, \zeta_{crit}], & \text{otherwise} \end{cases} \tag{3}$$

151 with

$$\zeta_{\text{univ}} = \frac{n_2}{n_1} \times \frac{(1 - \epsilon) + \frac{m}{;} n_1 \sqrt{\epsilon(2 - \epsilon)}}{1 - \left(\frac{n_2}{n_1}\right)^2 \epsilon(2 - \epsilon)},$$

$$\zeta_{\text{crit}} = \text{Re}\left(\frac{n_1 - \sqrt{n_1^2 - \text{NA}^2}}{n_2 - \sqrt{n_2^2 - \text{NA}^2}}\right),$$

$$m^2 = n_2^2 - n_1^2,$$

$$\epsilon = \frac{\lambda/4}{z_A n_2} \times \begin{cases} +1, & \text{if } n_2 \geq n_1 \\ -1, & \text{otherwise} \end{cases}$$

152 where ζ is the axial re-scaling factor ($\zeta \equiv z_A/z_N$), n_1 is the immersion refractive index, n_2 the
153 sample refractive index, λ the wavelength (in vacuum), z_A the actual depth of the imaged object
154 and NA the numerical aperture.

155 2.3. Derivation of analytical theory through wave-optics

With geometrical optics ambiguous in defining the re-scaling factor ζ we have to turn to wave-
optics. The most complete wave-optics theory describing aberrations occurring under RIM was
developed by Hell *et al.* [17]. Their calculation scheme takes into account the vectorial nature of
light and, is in essence, equivalent to the 3-fold application of the scalar model described in [24].
The scalar model in turn is based on the general Diffraction Integral applied to the ray fans in
RIM [25–27]. A scalar (radially symmetric) PSF of a wide-field microscope under RIM is an
integral over Bessel-beams:

$$\text{PSF}(x, y, z_A, z_N) = \left| \int A(\theta) \times J_0(r k_r) \times e^{i\Phi(\theta)} \sin\theta \, d\theta \right|^2, \tag{4}$$

where the PSF is dependent on both the actual depth z_A and on the nominal depth z_N, as an
image with intensities $I(x, y, z_N)$ is a convolution of fluorophore distribution $f(x, y, z_A)$ and
the PSF(x, y, z_A, z_N). Furthermore, $r = \sqrt{x^2 + y^2}$ as the PSF is cylindrically symmetric, the
amplitude factor $A(\theta) = 1/\sqrt{\cos\theta}$ is due to the Abbe-sine condition, J_0 is the Bessel function,
$i^2 = -1$ and the integral is taken over the NA of the objective lens. The effects of RIM are
engraved in the phase function $\Phi(\theta)$:

$$\Phi(\theta) = k_z'(\theta) \times z_A - k_z(\theta) \times z_N, \tag{5}$$

where k_z and k_z' are the axial wave-numbers of the Bessel beams, respectively for a matched
case and one undergoing RIM. In the full vectorial treatment, there are three integrals of the
form of Equation 4; I_0, I_1 and I_2 (one for each of the E-vector directions, with different orders of
Bessel functions). The first integral I_0 has a Bessel function of the order zero and is, in essence,
the Diffraction Integral used in the scalar-light theory. The other two integrals I_1 and I_2 of the
full vectorial theory contain the same phase function $\Phi(\theta)$ as in I_1, and, moreover, they vanish
at $r = 0$. Thus, we will consider only the scalar-light version of the wave-optics Equation 4 for
finding an explicit analytical form of the re-scaling factor ζ in the following paragraphs.

Compared to the ambiguous behavior of geometrical optics, Equation 4 provides a clear
prediction of the behavior of a microscope imaging under RIM. For a given depth z_A of a
fluorescent object there is a focal plane depth $z_N(z_A)$ where the PSF is maximal (i.e. the object is
in focus), and $\zeta(z_A) \equiv z_A/z_N(z_A)$. Under RIM, two intrinsic length scales need to be considered:
(i) several strictly geometrical parameters n_1, n_2, z_A, z_N, and NA. (ii) the (vacuum) wavelength

Fig. 2. The structure of $\Phi(\theta)$ for $n_1 < n_2$, where the diffraction integral in k-space is illustrated with the arcs of Ewalds' spheres. k'_z and k_z are the axial wave numbers with and without RIM. **(a)** After transition to a new medium the NA is preserved (due to the Snell's law) while a different Ewalds' sphere is used, going from black (no RIM) to red (under RIM). **(b)** The RIM phase-function $\Phi(\theta)$ re-scales (multiplies) the spheres in the axial directions in z_A (red) and z_N (black), and the phase derivative can be set to zero only at one point on the arc depending on the exact values of z_N and z_A (for non-zero θ). The Ewald sphere curvature under RIM changes with changing depth (red translucent, Δz_N) and with it the stationary point $d\Phi/d\theta = 0$.

$\lambda = 2\pi/k$ which is totally independent from the RIM geometry. These two length scales are
fused together in Equation 4 via the phase function $\Phi(\theta)$. With changing imaging depth, the
geometrical parameters also change (e.g. the path length difference between the paraxial and
the marginal rays). Reversely, the wavelength of light does not depend on the depth. So it is
the interplay between geometrical and physical length scales that results in a depth-dependent
re-scaling factor $\zeta(z_A) \neq \text{const}$.
To find ζ without calculating the integrals like Equation 4 we estimate the maximum of such
an integral by finding the leading terms of a simpler integral:

$$I_{\text{approx}} = \left| \int e^{i\Phi(\theta)} \, d\theta \right|^2. \tag{6}$$

To estimate the maxima positions of I_{approx} and assuming that $e^{i\Phi}$ is oscillating fast (for example at
large depth), we try to find the stationary points θ^* where $d\Phi/d\theta = 0$ [28]. The structure of $\Phi(\theta)$
in case of $n_1 < n_2$ is shown in Figure 2. The axial wave numbers under RIM and without RIM, k'_z
and k_z belong to the Ewalds' spheres (Figure 2a, red and black), with $k'_r(\theta) \equiv k_r(\theta) = k_1 \times \sin\theta$
conserved under RIM due to Snell's law. Radial wave number conservation allows the use of
$\sin\theta \, d\theta = dk_r/k_2$, making Equation 4 more similar to I_{approx}. The two integrals (Equation 4 and
Equation 6) are taken over spherical segments having the same width, but different radii due to
RIM, where the larger sphere is flattened compared to the smaller one. The RIM phase-function
$\Phi(\theta)$ re-scales these spheres in the axial directions in z_A and z_N, see Figure 2b. The stationary
points θ^* are found where the two re-scaled spheres are running in parallel to each other (dashed

lines, $d\Phi/d\theta = 0$). There is always a stationary point at $\theta = 0$, and a stationary point might appear for a single non-zero θ. For a given z_A and z_N, there is an interference between the contributions from the two critical points ($e^{i\Phi(0)}$ and $e^{i\Phi(\theta^*)}$) which controls the shape of the PSF. And when the interference is the most constructive – then the object appears in focus.

To find the non-zero stationary point θ^* we rewrite Φ using Ewalds' spheres and Snell's law:

$$k_z'(\theta)^2 = k'^2 - k_r(\theta)^2,$$
$$k_z(\theta)^2 = k^2 - k_r(\theta)^2,$$

so that when differentiating Φ by k_r:

$$\frac{d\Phi}{dk_r} = \frac{k_r}{k_z'}z_A - \frac{k_r}{k_z}z_N, \tag{7}$$

and at a stationary point $\theta^* : \frac{d\Phi}{d\theta} = 0$ ($k_r^* : \frac{d\Phi}{dk_r} = 0$):

$$\left.\frac{k_z'}{k_z}\right|_{\theta^*} = \frac{z_A}{z_N}, \tag{8}$$

that is the smaller sphere was multiplied by the larger depth z_A and the larger sphere by the smaller z_N as sketched in the Figure 2b. The stationary phase points of the integral in Equation 6 correspond to the rays of geometrical optics [26–28]. Hence, the re-scaling factor must fall within the bounds set by ζ_{paraxial} and ζ_{marginal}. Contrary to the purely geometrical optics approach, we have now two contributions to balance: the paraxial one $e^{i\Phi(0)}$ and the intermediate one $e^{i\Phi(\theta^*)}$.

We estimate for which z_A/z_N the two stationary point contributions $e^{i\Phi(0)}$ and $e^{i\Phi(\theta^*)}$, build the largest constructive interference. The relative strengths of the two leading contributions in the integral I_{approx} (Equation 6) depend on the width of the stationary points of Φ, and hence on the second derivative of Φ. The contributions of stationary-points in Equation 4 are also weighted by pupil-amplitude factor $A(\theta)$. Instead of finding the exact weight of each contribution, we study how the phase function $\Phi(\theta)$ behaves in several focal positions $z_N(z_A)$, which can be determined by full vectorial PSF calculations. This is shown in Figure 3a and Figure 3b for NA = 0.95, $n_1 < n_2$ and NA = 1.4, $n_1 > n_2$ respectively. The phase function at the second stationary point lies between $\frac{\pi}{3}$ and $\frac{2\pi}{3}$ from the phase function at the first stationary point in both imaging scenarios. Therefore, we select the following condition when deriving an explicit formula for ζ:

$$\Phi(\theta^*) - \Phi(0) \approx \frac{\pi}{2}, \tag{9}$$

where a fixed $\pi/2$ separation is assumed, as if the two stationary points constitute a single and uninterrupted constructive interference area when the microscope is in focus. With this approximate condition set, Snell's law, and the Ewalds' spheres scaling Equation 8 established, we solve a quadratic equation to derive an explicit formula for ζ. From Equation 9 and the definition of ζ we have:

$$\Phi(0) = n_2 k_0 z_A - n_1 k_0 z_N,$$
$$\Phi(\theta^*) = k_z'(\theta^*)z_A - k_z(\theta^*)^2 z_N,$$
$$k_z'(\theta^*) - \frac{k_z(\theta^*)}{\zeta} = n_2 k_0 - \frac{n_1 k_0}{\zeta} + \frac{\pi}{2z_A},$$
$$\zeta\frac{k_z'(\theta^*)}{k_0} - \frac{k_z(\theta^*)}{k_0} = n_2\zeta \times (1 - \frac{\lambda/4}{n_2 z_A}) - n_1,$$

Fig. 3. The RIM phase function difference $\Phi - \Phi(0)$ taken in the focal position $z_N(z_A)$, determined from the wave-optics calculations for fluorophore depths z_A set to $10\,\mu m$ (red), $20\,\mu m$ (green), $50\,\mu m$ (blue) and $100\,\mu m$ (black). **(a)** Objective NA of 0.95, $n_1 = 1.0$, $n_2 = 1.33$ (solid lines) and $n_2 = 1.52$ (dashed lines). **(a)** Objective NA of 1.4, $n_1 = 1.52$, $n_2 = 1.33$ (solid lines) and $n_2 = 1.0$ (dashed lines). For both cases, the second stationary point is somewhere between $\pi/3$ and $2\pi/3$ from the paraxial stationary point at $k_r/k_0 = 0$.

where the dimensionless parameter $\epsilon \equiv \frac{\lambda/4}{n_2 z_A}$ appears. This parameter embodies the before-mentioned interplay between the purely geometrical and purely physical length scales of the RIM. Applying Equation 8:

$$\frac{k_z'(\theta^*)^2 - k_z(\theta^*)^2}{k_z(\theta^*)k_0} = n_2\zeta(1 - \epsilon) - n_1. \tag{10}$$

Using Ewalds' spheres, Snell's law, and Equation 8 yields:

$$k_z'(\theta)^2 + k_r(\theta)^2 = n_2^2 k_0^2,$$
$$k_z(\theta)^2 + k_r(\theta)^2 = n_1^2 k_0^2,$$
$$k_z'(\theta)^2 - k_z(\theta)^2 = (n_2^2 - n_1^2)k_0^2 \equiv m^2 k_0^2,$$
$$k_z(\theta^*)^2(\zeta^2 - 1) = m^2 k_0^2.$$

Going back to the left hand-side of Equation 10:

$$\frac{k_z'(\theta^*)^2 - k_z(\theta^*)^2}{k_z(\theta^*)k_0} = \frac{m^2 k_0}{k_z(\theta^*)} = m\sqrt{\zeta^2 - 1},$$

and hence for the equation:

$$m\sqrt{\zeta^2 - 1} = n_2\zeta(1 - \epsilon) - n_1, \tag{11}$$

$$\zeta^2\left((1-\epsilon)^2 - \left(\frac{m}{n_2}\right)^2\right) - 2\frac{n_1}{n_2}\zeta + 1 = 0. \tag{12}$$

Which, when solved for ζ leads to:

$$\zeta(z_A) = \frac{n_2}{n_1} \times \frac{(1 - \epsilon) + \frac{m}{n_1}\sqrt{\epsilon(2 - \epsilon)}}{1 - \left(\frac{n_2}{n_1}\right)^2 \epsilon(2 - \epsilon)}. \tag{13}$$

2.4. Critical value at shallows depths

At very large depths, the re-scaling factor ζ approaches the paraxial regime (Equation 1), as $\epsilon \to 0$. At very shallow depths (i.e. large ϵ) Equation 13 becomes problematic: $\zeta(z_A = 0) = \infty$ when $n_2 > n_1$ and $\zeta(z_A = 0) = 0$ for $n_2 < n_1$, going beyond the bounds of the geometrical optics ($\zeta_{\mathrm{marginal}}$ in Equation 2). This is related to the fact that the actual value of the NA of the objective lens was omitted in the derivation so far.

At ever decreasing depths, the second critical point θ^* moves to higher wave numbers k_r and higher angles θ (peak of the phase difference in Figure 3) until it falls outside the NA of the optical system (Figure 2a). Hence, there is some (NA dependent) depth below which the assumptions we used to derive Equation 13 break down. We will hence derive a NA dependent critical value of re-scaling ζ_{crit} to be used at small depths.

With the whole range of phase function values $\Phi(\theta)$ becoming smaller than $\pi/2$ at small depths, the earlier condition on the separation of the critical points (Equation 9) can not be satisfied anymore within the confines of the NA. At such small depth, the entire pupil participates in constructive interference and the breakdown condition can be written as:

$$\Phi(0) = \Phi(\theta^*) - \frac{\pi}{2} = \Phi(\theta_{max}). \tag{14}$$

From the definition of $\Phi(\theta)$ and $k_r(\theta_{max}) = k_0 \times \mathrm{NA}$:

$$n_2 k_0 z_A - n_1 k_0 z_N = \sqrt{n_2^2 - \mathrm{NA}^2}k_0 z_A - \sqrt{n_1^2 - \mathrm{NA}^2}k_0 z_N,$$

And as $\zeta \equiv \frac{z_A}{z_N}$ one gets:

$$\zeta_{\mathrm{crit}} = \frac{n_1 - \sqrt{n_1^2 - \mathrm{NA}^2}}{n_2 - \sqrt{n_2^2 - \mathrm{NA}^2}}. \tag{15}$$

Equation 15, together with the universal NA-independent expression for ζ Equation 13 form Equation 3 and describe the depth-dependent axial scaling when imaging under RIM.

2.5. Re-scaling microscopy data

When re-scaling microscopy data from the native z_N coordinate into the true depth coordinate z_A, it is convenient to use the re-scaling factor as a function of the *nominal depth z_N* instead of the actual depth z_A utilized in Equation 13. In this case, a different dimensionless parameter δ can be used for the derivation:

$$\delta \equiv \frac{\lambda/4}{n_1 z_N}, \tag{16}$$

and Equation 12 can be re-written as:

$$m\sqrt{\zeta^2 - 1} = n_2 \zeta - n_1(1 + \delta),$$

$$\zeta^2 - 2\frac{n_2}{n_1}(1 + \delta)\zeta + (\frac{m}{n_1})^2 + (1 + \delta)^2 = 0,$$

yielding the re-scaling factor as a function of z_N:

$$\zeta(z_N) = \frac{n_2}{n_1} \times (1 + \delta) + \frac{m}{n_1}\sqrt{\delta(2 + \delta)}. \tag{17}$$

 3. Methods

 3.1. Wave-optics calculations

 For the wave-optics calculations of the PSFs we adapted the most complete description from Hell
 et al. [17]. The radially symmetric fields (of both excitation and emission light) are represented
 as integrals over component Bessel-beams. The wave-optics is linked with geometrical optics
 through an elegant mathematical correspondence between Bessel beams and light-cone sections
 of the ray manifold [25, 26, 29]. The effects of RIM are taken into account as (RIM *dependent*)
 phase shifts of the component Bessel-beams, resulting in a new interference pattern of the field
 and thus in a new PSF shape. The (dimensionless) intensity of a focal spot can be found by
 computing three integrals I_0, I_1 and I_2:

$$\text{PSF} = |I_0|^2 + 4\,|I_1|^2 + |I_2|^2 \tag{18}$$

 Integrals I_i each have a Bessel function of the first kind $J_i(rk_r(\theta))$ weighted by some coefficients
 of geometric origin. The common factors are $\sin\theta$ (from the Jacobian), $\sqrt{\cos\theta}$ (from Abbe's sine
 condition) and $e^{i\Phi(\theta)}$ (the RIM induced phase shifts). The argument of the Bessel functions is
 $rk_r(\theta) = r \times n_2 k_0 \sin\theta$. Taking into account the light polarization, the following factors appear:
 I_0 contains a factor of $(1 + \cos\theta)$, I_1 a factor of $\sin\theta$ and I_2 a factor of $1 - \cos\theta$. For each
 polarization orientation s and p, the partial transmission is taken into account as $t_s(\theta)$ and $t_p(\theta)$
 coefficients. Thus, the polarization factors become: $t_p(\theta) + t_s(\theta)\cos\theta$ in I_0, $t_s(\theta)\sin\theta$ in I_1 and
 $t_p(\theta) - t_s(\theta)\cos\theta$ in I_2. The integrals then become:

$$I_0 = \int_0^{\theta_{max}} J_0(rk_r(\theta))e^{i\Phi(\theta)}\sin\theta(t_p(\theta) + t_s(\theta)\cos\theta)\sqrt{\cos\theta}\sin\theta d\theta,$$

$$I_1 = \int_0^{\theta_{max}} J_1(rk_r(\theta))e^{i\Phi(\theta)}t_s(\theta)\sin\theta\sqrt{\cos\theta}\sin\theta d\theta,$$

$$I_2 = \int_0^{\theta_{max}} J_2(rk_r(\theta))e^{i\Phi(\theta)}(t_p(\theta) - t_s(\theta)\cos\theta)\sqrt{\cos\theta}\sin\theta d\theta,$$

 with $\theta_{max} = \text{Re}(\arcsin \text{NA}/n_2)$ the maximum angle of the light cone, and the phase function:

$$\Phi(\theta) = k_0 \times (n_2 z_A \cos\theta - n_1 z_N \sqrt{1 - n_2^2/n_1^2 \sin^2\theta}),$$

 and the transmission coefficients $t_s(\theta)$ and $t_p(\theta)$:

$$t_s(\theta) = \frac{2\cos\theta}{\cos\theta + \sqrt{1 - n_2^2/n_1^2 \sin^2\theta}},$$

$$t_p(\theta) = \frac{n_1}{n_{n2}}\frac{2\cos\theta}{\cos\theta + \sqrt{1 - n_2^2/n_1^2 \sin^2\theta}}.$$

 The computations of the integrals were performed through a MATLAB script producing 2D
 PSF cross-sections ($r \times z_N$, $10 \times 20\,\mu\text{m}$) on a $N^2 = 2048 \times 2048$ pixel grid from $M = 2000$
 Bessel-beam components. To minimize the number of Bessel function invocations, each of the
 integrals I_i was represented as a double matrix product:

$$I_i = J_i^{N \times N} \times A_i^{M \times M} \times E^{M \times N},$$

where $J_i^{N \times N}$ is the matrix containing the Bessel function, $A_i^{M \times M}$ is the diagonal matrix created from all the position independent factors and $E^{M \times N}$ is the matrix containing the z_N dependent phase change of M Bessel-beams.

For every RIM condition, a series of 2D PSF cross-sections (lateral × axial, XZ planes) was generated with the actual depth z_A changing from 0.2 to 5 µm with a step of 0.2 µm. Step sizes of 0.5 µm and 1.0 µm were used respectively from 5.5 to 25 µm and from 26 to 150 µm. The axial center of the calculated PSF frame was shifted with the actual depth z_A using the paraxial scaling factor $z_N^0 = z_A(n_1/n_2)$. In the lateral center of each frame, the Z profile was extracted and passed through a Butterworth low-pass filter (5th order, sampling and critical frequencies $f_s = 0.1\,\text{nm}^{-1}$, $f_c = 0.003\,\text{nm}^{-1}$) to remove high frequency ringing oscillations. The axial focal position z_N^* was determined by taking the maximum of the filtered profile and the re-scaling factor was computed as $\zeta = z_A/z_N^*$.

3.2. Axial scaling measurements

Table 1. Overview of microscope objectives and refractive indices in the measurements.

NA	Nikon #	M	Pixel size [nm]	FOV [µm]	Correction collar	n_1	n_2
0.70	MRH08630	60 ×	108.3	110.9	Yes	1.0	1.0
|	|	|	|	|	|	1.0	1.34
|	|	|	|	|	|	1.0	1.52
0.85	MUE35900	100 ×	65	133.12	Yes	1.0	1.0
|	|	|	|	|	|	1.0	1.34
|	|	|	|	|	|	1.0	1.52
0.95	MRD00605	60 ×	108.3	110.9	Yes	1.0	1.0
|	|	|	|	|	|	1.0	1.34
|	|	|	|	|	|	1.0	1.52
1.25	MRD77400	40 ×	162.5	166.4	No	1.34	1.34
1.40	MRD01901	100 ×	65	133.12	No	1.52	1.52
|	|	|	|	|	|	1.52	1.34
|	|	|	|	|	|	1.52	1.0

We measured the depth-dependent axial scaling using a fluorescence microscope setup described earlier [9], for different refractive index mismatches and numerical apertures, as listed in Table 1. A schematic of the measurement scheme is shown in Figure 4a, where a sapphire ball (BA, Ceratec, 2 mm diameter, grade 10) is separated from a coverslip (CS, Thorlabs #CG15CH2, 22 × 22 mm, #1.5H thickness). The sapphire ball was glued to a glass strip (GS) with a small droplet of UV curable glue (Norland Optical Adhesive 63). Fluorescent beads (FB, diameter 190 nm, excitation 480 nm, emission 520 nm, Bangs Labs #FSDG002) were dropcasted on the coverslip using a micropipette. Both surfaces were cleaned with acetone, isopropyl alcohol and dried with nitrogen. The glued sapphire ball was dipped in the solution and dried three times to apply the beads to the surface. Beads on both surfaces were imaged by acquiring a z-stack with the fluorescence microscope using an excitation wavelength of 485 nm (Lumencor SPECTRA X), dichroic mirror (Semrock #FF410/504/582/669-Di01-25x36) and an emission filter (Semrock #FF01-525/30-25). The objective lens was mounted upright and was moved by

(a)   (b)   (c)

Fig. 4. The experimental setup used to measure depth-dependent axial scaling. **(a)** Schematic illustration showing the movable optical objective lens (OL) and sapphire ball (BA) at opposite sides of the static coverslip (CS). The green spheres represent deposited fluorescent beads (FB) on the surface. **(b)** Detail photograph showing the objective lens (OL), coverslip (CS) and sapphire ball (BA). The ball is glued to a glass strip (GS), which is mounted to the sample shuttle holder (SSH). Scale bar 3 mm. **(c)** Maximum intensity projection along the Y axis of a recorded z-stack where $n_1 = n_2 = 1.0$. The bottom fluorescent signal originates from fluorescent beads residing on the coverslip. The signal at the top comes from the beads on the sapphire ball, shown by the visible curvature.

stick-slip piezo positioners (SmarAct GmbH) in X, Y and Z. The correction collar setting of the OL was optimized to minimize spherical aberrations when imaging beads on the coverslip. The glass strip with sapphire ball was mounted to the sample shuttle holder (SSH), Figure 4b. This holder was moved by piezo positioners, and hence the distance between the coverslip and the ball was controlled. Details on the positioning systems can be found in [9].

We used different objective lenses, but all were compatible with the Nikon CFI60 optical system, and were used together with a 200 mm tube lens (Nikon #MXA20696). Care was taken to center the pupil of the objective lens to the center of the tube lens, after which the objective lens was solely moved along the optical axis. With the distance between ball and coverslip set, a z-stack was acquired by moving the objective lens. In Figure 4c the maximum intensity projection along Y is shown for such a z-stack, where the fluorescent beads were imaged using a 100×, 0.85 NA objective (Nikon #MUE35900) in the absence of a refractive index mismatch ($n_1 = n_2 = 1.0$). The bottom fluorescent signal originates from fluorescent beads residing on the coverslip. The signal at the top comes from the beads on the sapphire ball, as recognized by the visible curvature.

The AFP was varied by changing the distance between ball and coverslip, and the contact point between them was first approximately set by having both the fluorescent beads on the coverslip and bottom side of the ball in focus of the microscope. Next, the starting AFP (z_A, typically around 100 µm) was set by moving the ball away from the coverslip and the AFP was reduced step by step (see Figure 5a, I through III). At each set depth z_A, a z-stack is recorded around the coverslip and ball separately (by moving the objective lens, −4 to 8 µm and −8 to 8 µm respectively, both with a step size of 0.25 µm). To ensure the second z-stack was acquired around the beads on the surface of the ball, we used ζ_marginal and ζ_paraxial estimates for < 20 µm and, > 20 µm respectively. At the end of the measurement sequence the ball was forced into contact with the coverslip, slightly deforming it by setting negative values of AFP on purpose (to

Fig. 5. Experimental procedure and data analysis. **(a)** Schematic showing three different distances between ball and coverslip, going from clear separation (I) to contact (II) and deformation of the coverslip (III). The dashed lines indicate the position of both ball (red) and coverslip (black). The green dots indicate deposited fluorescent beads. **(b)** The positions for the coverslip and ball as a function of the set ball-coverslip distance z_S (top) and the same data near the coverslip (bottom). **(c)** The coverslip position (z_{CS}) is subtracted from the lowest point of the ball (z_B) yielding the landing curve in z_N coordinates, plotted against the set depth z_S (top). The resulting values from a piecewise linear fit are used to determine when contact between ball and coverslip takes place, and are used to obtain $z_A = z_S - z_{S0}$. This finally yields the observed focal depth z_N versus the actual depth z_A from which the re-scaling factor $\zeta = z_A/z_N$ was calculated (bottom).

about $-2\,\mu m$), see Figure 4a, III.

The 3D positions of all fluorescent beads in the recorded z-stack were acquired using the PSF Extractor software [9, 30] and the X, Y and Z locations of all fluorescent beads were extracted. From these, the position of the coverslip z_{CS} was determined by taking the median of the z positions of the beads on the cover slip and a least squares paraboloid fit on the coordinates of the beads residing on the ball surface was used to determine the lowest point of the ball z_B. These z positions are plotted against the set depth z_S in Figure 5b for the ball (black) and coverslip (red). Around $5.5\,\mu m$ the ball and coverslip make contact, effectively deforming the coverslip at lower set depths. The coverslip position was subtracted from the lowest point of the ball $(z_N = z_B - z_{CS})$ and plotted against (z_S), see Figure 5c. A piece wise linear function of the form

$$z_N(z_S) = \begin{cases} z_{N0}, & \text{if } z_S \le z_{S0} \\ z_{N0} + (z_S - z_{S0})/m, & \text{otherwise} \end{cases}$$

is used to determine the offset z_{N0} between coverslip and ball, the inflection point z_{S0} and the slope of the curve m. The fit is performed using a Levenberg-Marquardt non-linear least squares algorithm. We fit the data from -2 to $20\,\mu m$ around the inflection point, and pass the measured values to the uncertainty parameter *sigma* in SCIPY.OPTIMIZE.CURVE_FIT [31]. This artificially increases the uncertainty with increasing depth, and ensures that the inflection point can be found reproducibly for all datasets. The fit values are used to calculate the AFP for each set depth $z_A = z_S - z_{N0} + mz_{S0}$. The result is shown in Figure 5c (bottom), where the observed nominal depth z_N versus the actual depth z_A is plotted, from which the re-scaling factor $(\zeta = z_A/z_N)$ was

353 computed.

3.3. Re-scaling of 3D microscopy data

To demonstrate and test the re-scaling of 3D microscopy data, we recorded z-stacks of fluorescent
beads embedded in an agarose hydrogel for $n_1 < n_2$, $n_1 = n_2$ and $n_1 > n_2$, where the $n_1 = n_2$
case serves as a ground truth for the re-scaling of the mismatched cases. To prepare the sample,
we first coated a cover slip with fluorescent beads (190 nm, excitation 480 nm, emission 520 nm,
Bangs Labs #FSDG002) by dropcasting from isopropyl alcohol (IPA). Next, an agarose gel was
prepared (2w%) by mixing Milli-Q water and agarose (Low EEO, Fisher Scientific) and heating
to boil in a microwave. 25 µL of the same beads in IPA were added to 225 mL agarose gel while
warm. The mixture was pipetted into a glass sample cell composed of a microscope slide to
which two small pieces of microscope slide as spacers were glued (using Norland 81 optical
adhesive) roughly 0.5 cm apart, on top of which a #1.5 cover slip (180 µm thickness) was glued.
After filling the cell with the agarose gel containing the beads, the cell was sealed with optical
adhesive. After each gluing step, the sample was cured by exposure to UV light (~350 nm) for
90 s, and during the last curing step the beads were protected using a piece of aluminum foil.
Confocal z-stacks of the same volume were repeatedly recorded with different objectives
using a Nikon C2-SHS C2si confocal on a Nikon Eclipse Ti inverted microscope with a 488 nm
excitation laser. If the microscope objective had a correction collar, we optimized its position to
minimize spherical aberrations while inspecting beads on the cover slip before the recording of
the z-stack. We used the following procedure to record the z-stacks (with a 500 nm z step size).
On the back of the sample (i.e. the side of the microscope slide) a spot was drawn using a felt
marker. We then imaged this spot in bright field using a 10×/0.4 NA air objective and chose a
reference feature in this spot to find back the same position in the sample. Next, we switched to
the 100×/1.4 NA oil objective, and recorded a confocal z-stack including the cover slip interface.
We then removed the immersion liquid from the sample, and found the same position again
using the 10×/0.4 air objective. At this position, we again recorded a z-stack, but now with the
100×/0.85 NA air objective. Finally, we recorded at the same position a (ground truth) z-stack
using a 40×/1.25 NA water objective. For single bead comparison, we cropped volumes in the
three stacks comprising the same beads in the ground truth (water immersion) stack and the
axially deformed stacks (air, oil immersion). We determined the cover-slip-to-sample interface
as the intensity peak of the beads deposited on the cover slip and cropped the z-stack above the
cover slip.
Re-scaling of the data was done either using the linear theories [4, 13, 15, 16] where the voxel
size in z was simply corrected, or using the depth-dependent re-scaling. For the latter, the AFP
of each slide in the stack was calculated using equation 17, with the Lyakin scaling factor as
the critical value instead of equation 15. Next, a linear z-stack was generated with a z step size
corresponding to the original stack and a range corresponding to the final AFP. The intensities in
this z-stack were interpolated (using inverse distance weighting) from the two nearest slices in
the (non-linear) AFP z-stack. A Jupyter Notebook to perform the linear and non-linear re-scaling
is available at [32].

3.4. Refractive index measurements

394 The refractive indices of the immersion oils used in this study were measured using an Abbe
395 refractometer (Atago 3T). The indices were measured at $T = 25$ °C and at a wavelength
396 $\lambda_D = 589.3$ nm. Using a measured dispersion value, the indices were converted using Cauchy's
397 relation to $\lambda = 520$ nm [33]. This resulted in the following refractive indices: $n^{25}_{520} = 1.5197$ (Type
398 DF immersion oil, Cargille) and $n^{25}_{520} = 1.3421$ (Immersol^{TM} W 2010 immersion oil, Zeiss).

Fig. 6. Axial re-scaling factor ζ versus AFP for different NAs and RIM contrasts, obtained through analytical expression Equation 3 (colored solid) and as calculated using wave-optics (colored dashed). For each NA the re-scaling factor as calculated by Lyakin *et al.* is included as a black solid line [16]. **(a)** ζ as a function of AFP for NAs of 0.95, 0.85 and 0.70 at $n_1 = 1.0$ and $n_2 = 1.33$. **(b)** ζ as a function of AFP for an NA of 1.4 at $n_1 = 1.52$ and $n_2 = 1.33$.

4. Results

4.1. Validation through wave-optics calculations

We first validate the derived analytical expression, by comparing it to wave-optics calculations, performed as described in subsection 3.1. This is shown in Figure 6a and b respectively for RIM contrasts of $n_1 = 1.0 \rightarrow n_2 = 1.33$ and $n_1 = 1.52 \rightarrow n_2 = 1.33$. The axial re-scaling factor ζ is plotted versus depth, as obtained through analytical expression Equation 3 (colored solid) and as calculated using wave-optics (colored dashed).

For $n_1 < n_2$ (Figure 6a) the re-scaling factor increases at shallow depths, where the re-scaling (constant) maximum is determined by the numerical aperture of the optical system (i.e. the z_A-independent value of ζ_{crit}). At increasing depths it levels towards the n_2/n_1 limit, following the universal (i.e. NA-independent) curve ζ_{univ}. The wave-optics calculations deviate from the analytical expression due to some oscillations remaining after data processing of the computed 2D PSFs. Despite this, they follow the general trend in the NA-independent regime (ζ_{univ} in Equation 3) for the different NAs plotted. At shallow depths, the plateau from the analytical expression (ζ_{crit} in Equation 3) is reproduced by the wave-optics calculations, though the exact re-scaling factor value fits between the analytical expression and the re-scaling factor as calculated by Lyakin *et al.* [16].

In the case of $n_1 > n_2$ (Figure 6b) the re-scaling factor as given by Equation 3, decreases at shallow depths, also reaching a plateau value determined by the numerical aperture. For this RIM contrast the plateau is very small, yielding a fast changing re-scaling factor in the first $10\,\mu\text{m}$ after the refractive index interface. Imaging deeper, the re-scaling factor again levels towards the n_2/n_1 limit. The wave-optics calculations show less extreme axial scaling at shallow depths, while we find good agreement for depths > $15\,\mu\text{m}$.

4.2. Validation through experiments

Five different objective lenses are used to measure depth-dependent axial re-scaling factors as explained in subsection 3.2, varying in numerical aperture, immersion refractive index n_1 and magnification. Immersol^{TM} W 2010 immersion oil (Zeiss, $n_{520}^{25} = 1.3421$) and Type DF immersion oil (Cargille, $n_{520}^{25} = 1.5197$) were used in experiments. We will first discuss two

Fig. 7. Axial re-scaling factor ζ versus AFP for an NA of 0.85, imaging under a RIM contrast of $n_1 = 1.0 \rightarrow n_2 = 1.34$. Measurement data (solid blue dots) are plotted alongside the wave-optics calculations (dashed blue lines), analytical solution (solid black lines, Equation 3) and depth-independent theories (solid lines). Two sets of measurement data are plotted individually and from this data the mean (solid black dots) is computed by binning along z_A, plotted in the center of each bin (bin sizes of 1 and 10 µm respectively for 0 to 10 µm and 10 to 100 µm). The measurement error (shaded blue area) is the sum of the measurements standard deviation and an estimated upper error limit.

cases in detail below, $n_1 = 1.0 \rightarrow n_2 = 1.34$ for a NA of 0.85 and $n_1 = 1.52 \rightarrow n_2 = 1.34$ for a NA of 1.4. For the other cases listed in Table 1, figures are presented in Supplement 1 and will be discussed afterwards.

4.2.1. NA=0.85, $n_1 = 1.0$, $n_2 = 1.34$

In Figure 7, the axial re-scaling factor ζ is plotted against the actual depth z_A. The measurements (solid blue dots) are plotted alongside the wave-optics calculations (dashed blue lines), analytical solution (solid black lines) and literature theories (solid colored lines). From our individual sets of measurement data, the mean (solid black dots) is computed by binning along z_A, plotted in the center of each bin (bin sizes of 1 and 10 µm respectively for 0 to 10 µm and 10 to 100 µm). We estimate an upper error limit of $(100\,\text{nm}/z_N + 100\,\text{nm}/z_A)\,\zeta$, in determining the various z positions, composed of errors from: (i) positioning of the piezo stages, (ii) determining the position of ball and coverslip by fluorescent bead localization, and (iii) fitting uncertainties in the landing curve. The total measured error for the re-scaling factor (shaded blue area) is the sum of the rolling standard deviation of the measurements and the upper error limit. It increases significantly at shallow depths, which is confirmed by acquiring data without RIM present, and

applying the same data analysis as shown in Figures S4 and S10.

Results are shown for a RIM contrast of $n_1 = 1.0 \rightarrow n_2 = 1.34$ in Figure 7, where the uncertainty in the measurement data increases at shallow depths, while at larger depths the measurements capture the decay of the re-scaling factor but are slightly higher than the wave-optics calculations and analytical theory. Although we have taken care to align the optical system, we do expect some misalignment and/or residual optical aberrations to be the cause of this deviation. Especially, residual spherical aberration induced by a non-perfect correction collar setting could influence the measurements.

Looking at the exact value of the plateau at shallow depths, the analytical expression overestimates the re-scaling factor compared with the measurements and wave-optics calculations. In fact, both re-scaling factors from Lyakin and Stallinga give better agreement with the measured data [15, 16] and in practical terms seems to provide a better value for ζ_{crit}. At extreme depths of $z_A = 10$ mm the depth-dependent re-scaling factor approaches the $n_2/n_1 = 1.34$ limit set by Carlsson *et al.* (not shown in Figure 7) [13]. At more realistic imaging depths of $100\,\mu$m, the median from Diel coincides with the depth-dependent re-scaling factor and their mean gives a good agreement at $z_A = 40\,\mu$m.

4.2.2. NA=1.4, $n_1 = 1.52$, $n_2 = 1.34$

In Figure 8, the axial re-scaling factor ζ is plotted against the actual depth z_A. We find good agreement between measurement data and the analytical theory when imaging under a RIM contrast of $n_1 = 1.52 \rightarrow n_2 = 1.34$, as show in Figure 8. The measurements presented agree with the analytical theory and only start to drop more at shallow depths ($< 4\,\mu$m). The relative error when approaching $z_A = 0$ increases for z_A and z_N as the error in both is affected by determining the point at which the sapphire ball touches the coverslip. We have also included re-scaling factors as measured by Petrov *et al.* where they use a method with a much smaller measurement errors (see also inset) [23]. Combining the measurement data from Petrov at shallow depths, and our own measurements data at larger depths, we find very good agreement with the analytical theory presented. Although the analytical expression provides a plateau for the re-scaling factor $\zeta_{\text{crit}} = 0.64$, this is not reproduced in the measurement data from Petrov.

Our wave-optics calculations also match the analytical theory at larger depths, but deviate below $z_A = 4\,\mu$m. At the same time, they do overlap with wave-optics calculations from Egner & Hell [19]. At these shallow depths, an offset of a few tens of nanometers in the NFP has a dramatic effect on the re-scaling factor and hence super-critical angle fluorescence (SAF) effects are significant [34]. Although SAF is not included in the analytical derivation, it is considered in the wave-optics calculations through the Fresnel transmission coefficients when using the vector PSF model. The axial deformation is, however, enhanced by SAF, as the NFP shifts <40 nm away from the objective [34]. Therefore, the difference in the re-scaling factor between the wave-optics calculations and the analytical theory has already been reduced by the incorporation of SAF in the calculations.

Up to depths of $z_A = 1\,\mu$m, there is a ~250 nm difference in the NFP found between $\zeta = 0.64$ (analytical theory) and $\zeta = 0.78$ (wave-optics calculations). This is too large to be caused by the chosen method of determining the focal position (i.e. maximum axial intensity [19] vs. minimal lateral width [23]). As such, we do not have a fitting explanation for the discrepancy between the wave-optics calculations and the analytical theory along with measurements from Petrov *et al.*.

4.2.3. Other imaging scenarios

The measurement data for the remaining cases listed in Table 1, are shown in Figures S1 through S11, where the re-scaling factor ζ is plotted against z_A, alongside the wave-optics calculations, analytical solution and depth-independent theories. For each objective lens used we acquired data without a refractive index mismatch present and applied the same data analysis procedures

Fig. 8. Axial re-scaling factor ζ versus AFP for an NA of 1.4, imaging under a RIM contrast of $n_1 = 1.52 \rightarrow n_2 = 1.34$. Measurement data (solid blue dots) are plotted alongside the wave-optics calculations (dashed blue lines), analytical solution (solid black lines, Equation 3) and depth-independent theories (solid lines). Three sets of measurement data are plotted individually and from this data the mean (solid black dots) is computed by binning along z_A, plotted in the center of each bin (bin sizes of 1 and 10 μm respectively for 0 to 10 μm and 10 to 100 μm). The measurement error (shaded blue area) is the sum of the measurements standard deviation and an estimated upper error limit. Measured data and theory as presented by Petrov *et al.* and wave-optics calculations from Egner & Hell are included [19, 23].

as for axially scaled measurements (Figures S1, S4, S6, S9, S10). In all cases, the spread in the measurement data increases below $z_A = 10\,\mu$m as uncertainties in determining both z_A and z_N increase. At higher depths, no reproducible deviations away from $\zeta = 1.0$ are found in these data, and we note a maximum error of about 2 % in the measurements. The data shown in Figure S9 is of special interest where we have used a water immersion objective to verify our measurement method and analysis when having immersion oil ($n = 1.34$) present between the coverslip and sapphire ball. We find a mean re-scaling factor of 0.99 ± 0.05 (standard deviation) over the entire depth range, and apart from the increased scatter near $z_A = 0$, the re-scaling factor remains approximately constant ($\zeta = 1.0$) as a function of depth for each individual dataset.

The objective lens with the lowest numerical aperture which we used in our measurements had a NA of 0.7, of which the results are presented in Figure S2 and S3 ($n_1 = 1.0 \rightarrow n_2 = 1.34$ & $n_1 = 1.0 \rightarrow n_2 = 1.52$). In both cases, the wave-optics calculations reproduce the analytical expression, while the measurements only reproduce the general trend for higher RIM contrast of $n_1 = 1.0 \rightarrow n_2 = 1.52$. In the case of $n_1 = 1.0 \rightarrow n_2 = 1.34$, the measurement uncertainties obscure a clear trend in the measurement data.

The RIM of $n_1 = 1.0 \rightarrow n_2 = 1.52$ for the 0.85 NA objective lens is shown in Figure S5. Both the wave-optics calculations and measurement data reproduce the behavior of the re-scaling factor as predicted by the analytical expression. Comparing against the measurement data, Lyakin and Stallinga again provide the best critical value ζ_{crit}.

In the case of the 0.95 NA air objective (Figures S7 and S8), again the wave-optics calculations reproduce the analytical expression over the entire depth range. For the measured data however, the trend of the NA independent regime $\zeta_{\mathrm{universal}}$ is reproduced, but the measurements fail to replicate the strong increase to the plateau given by ζ_{crit} below $z_A = 20\,\mu$m. Along with depth-dependent re-scaling being more prominently present with increasing NA, so is the sensitivity to any remaining (i.e. spherical or tilt) aberrations in the optical setup. We have taken care to minimize these, but we estimate that residual aberrations lower the effective NA, obscuring the rise to the plateau in the measurements.

For the 1.4 NA oil immersion objective, we present the RIM case of $n_1 = 1.52 \rightarrow n_2 = 1.0$ in Figure S11. Again the general depth-dependent behavior is reproduced in the measurements, but the re-scaling factor is underestimated at the coverslip while it is overestimated when imaging deeper into the sample.

4.3. *Axial re-scaling of 3D microscopy data*

To test axial re-scaling of 3D microscopy data using our depth-dependent theory, we recorded confocal z-stacks of beads suspended in an agarose hydrogel, from the cover slip to $100\,\mu$m depth (z_N). These z-stacks were recorded using the $100\times$/0.85 NA air, $40\times$/1.25 NA water and $100\times$/1.4 oil objectives (see Table 1). As the refractive index of the agarose gel ($n_2 = 1.3356$ [35]) is very close to the refractive index of the immersion medium of the water objective, we can use these z-stacks as a ground truth for the axial re-scaling of the data. We recorded the z-stacks at the exact same location in the sample to allow for direct comparison of the same beads in different imaging scenarios.

4.3.1. $n_1 < n_2$

The case $n_1 < n_2$ is particularly relevant for cryo-fluorescence microscopy where air immersion is used to observe a frozen or vitrified sample. Thus, we evaluate a typical situation with NA = 0.85, $n_1 = 1.0$ and $n_2 = 1.336$ [9]. Before re-scaling the data, we quantitatively compare the re-scaling factors of the depth-dependent and linear theories, as plotted in Figure 9. As we have found that the theories of Lyakin and Stallinga result in approximately equal re-scaling factors, we choose to plot only the Lyakin re-scaling factor. As the critical value of the depth-dependent re-scaling factor, we use the Lyakin re-scaling factor. Therefore, the depth-dependent re-scaling

Fig. 9. Quantitative comparison of the linear scaling theories from literature to the depth-dependent analytical theory for NA=0.85, $n_1 = 1.0$ and $n_2 = 1.336$. **(a)** The re-scaling factor ζ vs AFP for the depth-dependent theory (solid), Lyakin *et al.* (dashed) [16], Diel *et al.* mean (dash-dot) and median (dot-dot) [4]. **(b)** The focal shift vs AFP of the aforementioned theories. **(d)** Axial difference vs AFP for the theories. **(d)** Relative difference between the theories.

factor is 1.532 close to the cover slip and falls off to ~1.4 deeper into the sample (see Figure 9a). When plotting the focal shift Δf as a function of depth (AFP) we see that depth-dependent focal shift equals the Lyakin focal shift close to the cover slip and gradually transitions to the Diel (median) focal shift at larger depth (see Figure 9b).

To see how significant the difference between the theories is for this imaging scenario, we plot the absolute and relative difference between the depth-dependent and linear theories as function of depth (AFP). Figure 9c shows that for Lyakin the difference is obviously zero near the cover slip, but quickly increases to a large difference (>1 µm) after a depth of >25 µm. On the other hand, Diel (median) differs significantly (~1 µm) at a depth of 20 µm, but is the linear theory with the smallest relative difference at >50 µm depth. Finally, Diel (mean) is in between the other two linear theories, having a smaller difference close to the cover slip than Diel (median), and a smaller difference at large depth than Lyakin. Where the relative difference between the depth-dependent theory and the theories of Lyakin and Diel median can exceed 5 %, the relative difference for the Diel mean theory is below 3 % (see Figure 9d).

Figure 10 shows the re-scaling of the 3D microscopy data of beads in an agarose hydrogel with NA = 0.85, $n_1 = 1.0$ and $n_2 = 1.336$. Overlays are plotted of the re-scaled z-stacks (grays)

Fig. 10. Axial re-scaling of 3D microscopy data with NA = 0.85, $n_1 = 1.0$, $n_2 = 1.336$ using the depth-dependent and linear theories. Maximum intensity projections along Y of the (re-scaled) confocal z-stack of beads embedded in an agarose hydrogel. The stacks have been re-scaled using the depth-dependent (left), Diel (mean) (middle) and Diel (median) (right) re-scaling factors. Overlays are plotted of the re-scaled z-stacks (grays) and the ground truth z-stack recorded with the water immersion objective (reds). The two bottom rows are cut-outs of the upper row, where the cut-outs at larger depths appear more noisy due to the intensity re-scaling due to fluorescence intensity loss at large depths. The beads are slightly displaced in X due to imperfections in the manual overlay of the z-stacks.

and the ground truth z-stack recorded with the water immersion objective (reds). As $n_1 < n_2$, the recorded z-stack was stretched in the axial direction to return to the ground truth axial distances. Deep into the sample ($z_A \approx 90 \, \mu m$), Diel (mean) shows the largest re-scaling error where the z-stack is overstretched, whereas the depth-dependent and Diel (median) re-scaling collapses onto the ground truth data, although the PSF is elongated in the axial direction due to spherical aberrations induced by the refractive index mismatch. Closer to the cover slip ($z_A \approx 20 \, \mu m$), both the depth-dependent and Diel (mean) re-scaling overlap with the ground truth z-stack, whereas Diel (median) has a slight (absolute) error as the axial distances have not been stretched enough. This agrees with the re-scaling factors plotted in Figure 9. To quantify this effect, we localize all individual fluorescent beads present in the data recorded using the PSF-EXTRACTOR

Fig. 11. Relative error of the re-scaled z position (NA = 0.85, n_1 = 1.0, n_2 = 1.336) compared to the ground truth z_A as recorded using a water immersion objective. The relative error is binned along z_A in bins of 10 µm and the mean of each bin is plotted in the center of each bin. The standard deviation is plotted as an error bar. The dashed gray line indicates 0 and acts as a guide to the eye. We measure an error in the axial re-scaling using the depth-dependent theory below 2 %, whereas the linear theories result in errors close to 5 %.

software [30]. We compare the z position found in the re-scaled data to the ground truth value z_A and divide this difference over z_A. This is shown in Figure 11, where we bin the data (bin size 10 µm) and plot the mean of each bin along with the standard deviation as error bar. The data is plotted centered with respect to the bin, and we omit the first bin (0 to 10 µm) as it only contains 3 data points. The depth-dependent theory outperforms both Diel theories in the axial re-scaling of this data, where we note a maximum error of 2 % when using depth-dependent axial re-scaling, whereas the errors for the linear theories approaches 5 %.

4.3.2. $n_1 > n_2$

The case of $n_1 > n_2$ is relevant for oil immersion observation of samples in water. Thus, we here evaluate a typical situation with NA = 1.4, n_1 = 1.52 and n_2 = 1.336. We quantitatively compare the re-scaling factors of the depth-dependent and linear theories in Figure 12. As NA > n_2, we cannot use Diel (mean). The depth-dependent re-scaling factor equals the Lyakin re-scaling factor close to the cover slip, but surpasses the Diel (median) at $z_A \approx 30$ µm. The same trend is seen in the focal shift Δf. In this imaging scenario, the axial positional difference between Lyakin and the depth-dependent re-scaling theory increases after a few micrometers in depth, resulting in a significant difference >6 %. While the absolute difference between Diel (median) and the depth-dependent theory is small close to the cover slip (<1 µm) and ~2 µm um at large depths, the relative error is large (~20 %) close to the cover slip and small at large depths (~2 %).

In Figure 13 the re-scaling of the 3D microscopy data is plotted. As $n_1 > n_2$, the recorded z-stack was compressed in the axial direction to return to the ground truth axial distances. At larger depth ($z_A \approx 65$ µm), the depth-dependent re-scaling nicely coincides with the ground truth, whereas the axial distances in the Diel (median) re-scaling has been compressed too much (see second row). Closer to the cover slip ($z_A \approx 15$ µm) the depth-dependent re-scaling again nicely coincides with the ground truth, but the Diel (median) re-scaling has been compressed too little. To quantify the error in re-scaling using the Diel (median) and depth-dependent theory, we

Fig. 12. Quantitative comparison of scaling theories in case of NA = 1.4, n_1 = 1.52, n_2 = 1.336. **(a)** The re-scaling factor ζ vs AFP for the depth-dependent theory (solid), Lyakin *et al.* (dashed) [16] and median (dot-dot) [4]. **(b)** The focal shift vs AFP of the aforementioned theories. **(d)** Axial difference vs AFP for the theories. **(d)** Relative difference between the theories.

again localized all fluorescent beads and compared the rescaled z position of the beads to the ground truth z_A, as shown in Figure 14. There is not one theory that outperforms the others over the depths measured here. Still, the depth-dependent theory results in a smaller error closer to the cover slip, as is to be expected following Figure 8. In addition, this measurement did not include beads near the cover glass (z_A < 20 µm) where a relative difference up to 10 % between the two theories is expected (see Figure 12). We should note that the objective lens used in this experiment did not have a correction collar. This means that we could not completely get rid of spherical aberration (SA) when imaging close to the cover slip. The presence of SA near the cover slip contradicts the optical conditions in the derivation of the analytical theory and therefore affects the legitimacy of the depth-dependent re-scaling factor, resulting in a higher re-scaling error. This shows that proper re-scaling also requires an optimization of the correction collar before the acquisition of the 3D microscopy data.

4.4. Software for depth-dependent re-scaling

4.4.1. Re-scaling acquired z-stacks using Python

To re-scale z-stack acquired under RIM, we have written PYTHON software which is available at https://github.com/hoogenboom-group/SF [32]. Jupyter Notebooks are used to

Fig. 13. Axial re-scaling of 3D microscopy data with NA = 1.4, n_1 = 1.52, n_2 = 1.336. Maximum intensity projections along Y of the (re-scaled) confocal z-stack of beads embedded in an agarose hydrogel. The stacks have been re-scaled using the depth-dependent (left) and Diel (median) (right) re-scaling factors. Overlays are plotted of the re-scaled z-stacks (grays) and the ground truth z-stack recorded with the water immersion objective (reds). The two bottom rows are cut-outs of the upper row. The beads are slightly displaced in X due to imperfections in the manual overlay of the z-stacks.

read image data and, based on the specific imaging conditions, use inverse distance weighting to rescale the intensities along the axial coordinate correctly. The same axial pixel size from the original data is used, but depending on the RIM mismatch, the re-scaled z-stack will map a shorter or larger axial range.

4.4.2. Interactive online tool for plotting the re-scaling factor versus imaging depth

We have made an online interactive tool where the depth-dependent re-scaling factor can be plotted and compared to existing depth independent scaling theories using PLOTLY DASH [36], the source code of which is available at [32]. The code is accessible via the URL https://axialscaling.pythonanywhere.com/. The refractive indices n_1 and n_2, the NA and the wavelength λ can be varied. In addition, the focal shift Δf = can be plotted

Fig. 14. Relative error of the re-scaled z position (NA = 1.4, n_1 = 1.52, n_2 = 1.336) compared to the ground truth z_A. The relative error is binned along z_A in bins of 10 µm and the mean of each bin is plotted in the center of each bin. The standard deviation is plotted as an errorbar. The dashed gray line indicates 0 and acts as a guide to the eye. For this measurement, we find that the maximum error in the axial re-scaling using the depth-dependent and Diel (median) linear theory is 5 %.

for the depth-dependent re-scaling factor, as well as for two depth independent scaling theories. Both the re-scaling factors and the focal shift resulting from the depth-dependent theory can be exported to file. All the measurement data and wave optics calculation results found in this manuscript and the accompanying supplemental materials are also included in the interactive plot.

5.   Discussion

The various re-scaling theories found in the literature can be understood as a result of different (overt or covert) assumptions on which maximum constructive interference contribution is leading. For example, Lyakin *et al.* used an analysis similar to ours, but explicitly sets the critical point θ^* such that $k_z'(\theta^*) = \frac{1}{2} \times (k_z'(0) + k_z'(\theta_{\max}))$, and then from Equation 8 follows [16]:

$$\zeta_{\text{Lyakin}} = \frac{n_1 + \sqrt{n_1^2 - \text{NA}^2}}{\sqrt{4\left(n_2^2 - n_1^2\right) + \left(n_1 + \sqrt{n_1^2 - \text{NA}^2}\right)^2}} \tag{19}$$

The same value was proposed earlier by Stallinga while posing different criteria of minimal variation of the phase function: $\theta^* : \min \left\langle (\Phi - \Phi(\theta^*))^2 \right\rangle$ [15]. For Diel *et al.* the ζ_{median} corresponds to setting $\theta^* : k_r(\theta^*) = 1/2k_r(\theta_{\max}) = \frac{1}{2}k_0 \times \text{NA}$ [4], such that

$$\zeta_{\text{median}} = \sqrt{\frac{n_2^2 - \frac{\text{NA}^2}{4}}{n_1^2 - \frac{\text{NA}^2}{4}}} \tag{20}$$

These estimates for ζ are geometrical in nature, where the actual wavelength is not taken into account – with the radii of the Ewalds' spheres being cancelled out of the fraction Equation 8.

This results in an implicit omission of the interference nature of the PSF under RIM. Somewhat less obviously, the criterion of Stallinga is also geometrical [15]. While it does demand minimal variation in the value of Φ in order to achieve the highest possible constructive interference, it does not take into account that the well minimized phase function can become several 2π in range (when both z_A and z_N are large enough) and thus lead to destructive interference anyway.

In order to be able to correct for depth-dependent axial scaling in experiments where imaging under RIM is unavoidable, it is crucial that (i) the position of the coverslip (or more general, the RIM interface) is determined precisely and, (ii) spherical aberration at the interface is reduced to an absolute minimum by selecting the cover glass the objective is optimized for, or by optimizing the position of the correction collar. In practical terms, this would require that the objective lens used is fitted with a correction collar, which is often the case for high-NA objective lenses.

One such case would be cryo-fluorescence microscopy used in the cryo-electron tomography workflow, where a RIM is generally present due to the large temperature difference between the optical objective and the sample. With fluorescence microscopy, targets can be identified in frozen hydrated cells and consequently sufficiently thin sections can be prepared by ablating the excess cellular material surrounding the target with a focused ion beam [37–39]. The aimed thickness of this frozen section is approximately 100 to 200 nm, which makes it crucial to precisely determine the target position with respect to the RIM interface (outer cell surface) [40]. If axially scaled distances measured with the light microscope are not corrected, targeting errors in the range of 300 to 1200 nm will occur with cell thicknesses ranging up to micrometers. Moreover, the *depth-dependence* of these errors will be important when fabricating sections out of thicker ice whilst using fluorescence microscopy to find targets in for instance organoids, as can be done in high pressure frozen samples [41, 42].

With the provided software [32], one can easily plot the depth-dependent re-scaling factor for the relevant imaging scenario, but also re-scale their 3D microscopy data, without having to judge which linear theory will hold in this scenario or perform full wave-optics calculations. We should note, however, that although the data will be axially re-scaled, our software does not provide any correction of the shape of the point spread function (PSF) due to spherical aberrations. To correct for this, one would require deconvolution with a depth-dependent PSF as demonstrated in [43].

Our measurements are limited by the relatively high uncertainty in determining the re-scaling factor near $z_A = 0$. It would therefore be interesting, for future research, to use the method from Petrov *et al.* to measure the re-scaling factor for the scenario with a 0.85 NA air objective when imaging $n_1 = 1.0 \rightarrow n_2 = 1.33$.

6. Conclusion

We have presented an analytical theory to correct for the depth-dependent axial deformation when imaging with light microscopy in the presence of a refractive index mismatch between the sample and the microscope objective immersion medium. Using the resulting equation, a re-scaling factor as a function of depth can be calculated from imaging parameters numerical aperture (NA), the refractive indices of the objective (n_1) and sample (n_2), and the wavelength (λ), for a RIM with both $n_1 < n_2$ and $n_1 > n_2$. We performed wave-optics calculations to verify the theory and find a very good agreement between the two. In addition, we performed experiments to measure the axial scaling, where we find a good agreement at larger depths, whereas closer to the cover slip, the measurements suffer from large uncertainties. We do find good agreement with the accurate measurements done by Petrov *et al.* [23] in the case when imaging with a high-NA oil immersion objective into a water sample.

Next, we tested the depth-dependent theory versus existing linear theories in the literature on 3D microscopy data with a known ground truth. We find that for NA = 0.85, $n_1 = 1$, $n_2 = 1.33$ the depth-dependent theory outperforms existing linear theories for depths up to 80 µm with a maximum relative error of 2 %. For NA = 1.4, $n_1 = 1.52$, $n_2 = 1.33$, the depth-dependent theory

performs as good as the best linear theory in literature. However, we think its performance was compromised by the presence of a small spherical aberration near the cover slip, which could not be corrected for as the objective did not have a correction collar. Moreover, we could not compare close to the cover glass ($< 10\,\mu m$), where the linear theory used in the comparison was expected to break down.

Finally, we have presented a web applet to be used by microscope users to calculate the re-scaling factor for their imaging parameters. In addition, we have shared software to re-scale 3D data sets using the depth-dependent scaling factor.

Our re-scaling theory is the first to include the depth dependence of axial scaling due to a refractive index mismatch. It will be of use in imaging scenarios where the refractive indices of the sample and objective cannot be matched, such as in integrated, cryogenic, and correlative light and electron microscopy setups, or in the imaging of water-like samples using high-NA oil immersion objectives.

Funding. This work is part of the Cryo3Beams project (project N⁰ 17152) financed via the High Tech Systems and Materials programme of the Applied and Engineering Sciences domain of the Dutch Research Council (NWO).

Acknowledgments. We thank Hans C. Gerritsen and Gerhard A. Blab for the useful discussions, Bernd Rieger for feedback on our results and Alfons van Blaaderen for the use of the refractometer.

Disclosures. SVL, MNFH and EBW declare no conflicts of interest. DBB is employed by Delmic BV, JPH has a financial interest in Delmic B.V.

Data availability. Data underlying the results presented in this paper are available in Ref. [44].

Supplemental document. See Supplement 1 for supporting content.

References

1. J. Pawley, *Handbook of biological confocal microscopy*, vol. 236 (Springer Science & Business Media, 2006).
2. M. Saxena, G. Eluru, and S. S. Gorthi, "Structured illumination microscopy," Adv. Opt. Photonics **7**, 241–275 (2015).
3. E. H. Stelzer, F. Strobl, B.-J. Chang, *et al.*, "Light sheet fluorescence microscopy," Nat. Rev. Methods Primers **1**, 73 (2021).
4. E. E. Diel, J. W. Lichtman, and D. S. Richardson, "Tutorial: avoiding and correcting sample-induced spherical aberration artifacts in 3d fluorescence microscopy," Nat. protocols **15**, 2773–2784 (2020).
5. P. De Boer, J. P. Hoogenboom, and B. N. Giepmans, "Correlated light and electron microscopy: ultrastructure lights up!" Nat. methods **12**, 503–513 (2015).
6. T. Ando, S. P. Bhamidimarri, N. Brending, *et al.*, "The 2018 correlative microscopy techniques roadmap," J. physics D: Appl. physics **51**, 443001 (2018).
7. G. Wolff, C. Hagen, K. Grünewald, and R. Kaufmann, "Towards correlative super-resolution fluorescence and electron cryo-microscopy," Biol. Cell **108**, 245–258 (2016).
8. R. Kaufmann, C. Hagen, and K. Grünewald, "Fluorescence cryo-microscopy: current challenges and prospects," Curr. opinion chemical biology **20**, 86–91 (2014).
9. D. B. Boltje, J. P. Hoogenboom, A. J. Jakobi, *et al.*, "A cryogenic, coincident fluorescence, electron, and ion beam microscope," Elife **11**, e82891 (2022).
10. S. Gorelick, G. Buckley, G. Gervinskas, *et al.*, "Pie-scope, integrated cryo-correlative light and fib/sem microscopy," Elife **8**, e45919 (2019).
11. W. Li, J. Lu, K. Xiao, *et al.*, "Integrated multimodality microscope for accurate and efficient target-guided cryo-lamellae preparation," Nat. Methods pp. 1–8 (2023).
12. S. Li, Z. Wang, X. Jia, *et al.*, "Eli trifocal microscope: A precise cryogenic fabrication system to prepare target cryo-lamellae of cells for in situ cryo-et study," bioRxiv pp. 2022–08 (2022).
13. K. Carlsson, "The influence of specimen refractive index, detector signal integration, and non-uniform scan speed on the imaging properties in confocal microscopy," J. Microsc. **163**, 167–178 (1991).
14. T. Visser, J. Oud, and G. Brakenhoff, "Refractive index and axial distance measurements in 3-d microscopy," SPIE MILESTONE SERIES MS **131**, 286–289 (1996).
15. S. Stallinga, "Finite conjugate spherical aberration compensation in high numerical-aperture optical disc readout," Appl. optics **44**, 7307–7312 (2005).
16. D. Lyakin, L. Maksimova, A. Y. Sdobnov, and V. Ryabukho, "The influence of the numerical aperture of a beam probing an object on the determination of the thickness of a layered object in confocal microscopy," Opt. Spectrosc. **123**, 487–494 (2017).

17. S. Hell, G. Reiner, C. Cremer, and E. H. Stelzer, "Aberrations in confocal fluorescence microscopy induced by mismatches in refractive index," J. microscopy **169**, 391–405 (1993).

18. C. Sheppard and P. To¨ ro¨ k, "Effects of specimen refractive index on confocal imaging," J. microscopy **185**, 366–374 (1997).

19. A. Egner and S. W. Hell, "Aberrations in confocal and multi-photon fluorescence microscopy induced by refractive index mismatch," Handb. biological confocal microscopy pp. 404–413 (2006).

20. C. J. de Grauw, J. M. Vroom, H. T. van der Voort, and H. C. Gerritsen, "Imaging properties in two-photon excitation microscopy and effects of refractive-index mismatch in thick specimens," Appl. optics **38**, 5995–6003 (1999).

21. T. H. Besseling, J. Jose, and A. V. Blaaderen, "Methods to calibrate and scale axial distances in confocal microscopy as a function of refractive index," J. microscopy **257**, 142–150 (2015).

22. A. Diaspro, F. Federici, and M. Robello, "Influence of refractive-index mismatch in high-resolution three-dimensional confocal microscopy," Appl. optics **41**, 685–690 (2002).

23. P. N. Petrov and W. Moerner, "Addressing systematic errors in axial distance measurements in single-emitter localization microscopy," Opt. express **28**, 18616–18632 (2020).

24. S. F. Gibson and F. Lanni, "Experimental test of an analytical model of aberration in an oil-immersion objective lens used in three-dimensional light microscopy," JOSA A **8**, 1601–1613 (1991).

25. C. J. Bouwkamp, "Diffraction theory," Reports on progress physics **17**, 35 (1954).

26. J. B. Keller, "Geometrical theory of diffraction," Josa **52**, 116–130 (1962).

27. E. Wolf, "A scalar representation of electromagnetic fields: Ii," Proc. Phys. Soc. **74**, 269 (1959).

28. J. J. Stamnes, "Waves, rays, and the method of stationary phase," Opt. express **10**, 740–751 (2002).

29. H. Green and E. Wolf, "A scalar representation of electromagnetic fields," Proc. Phys. Soc. Sect. A **66**, 1129 (1953).

30. D. B. Boltje, E. B. van der Wee, and R. Lane, *PSF-Extractor* (2022).

31. P. Virtanen, R. Gommers, T. E. Oliphant, *et al.*, "Scipy 1.0: fundamental algorithms for scientific computing in python," Nat. methods **17**, 261–272 (2020).

32. E. B. van der Wee, D. B. Boltje, and S. Loginov, "Re-scaling factor online plotting tool," (2023).

33. F. A. Jenkins, "Fundamentals of optics: By francis a. jenkins and harvey e. white," Tech. rep., McGraw-Hill (1976).

34. M. E. Siemons, L. C. Kapitein, and S. Stallinga, "Axial accuracy in localization microscopy with 3d point spread function engineering," Opt. Express **30**, 28290–28300 (2022).

35. E. Fujiwara, T. D. Cabral, M. Sato, *et al.*, "Agarose-based structured optical fibre," Sci. reports **10**, 7035 (2020).

36. Plotly, *Dash* (2023).

37. E. Y. Chiang, A. Hidalgo, J. Chang, and P. S. Frenette, "Imaging receptor microdomains on leukocyte subsets in live mice," Nat. methods **4**, 219–222 (2007).

38. E. Villa, M. Schaffer, J. M. Plitzko, and W. Baumeister, "Opening windows into the cell: focused-ion-beam milling for cryo-electron tomography," Curr. opinion structural biology **23**, 771–777 (2013).

39. R. K. Hylton and M. T. Swulius, "Challenges and triumphs in cryo-electron tomography," Iscience **24**, 102959 (2021).

40. M. Vulović, R. B. Ravelli, L. J. van Vliet, *et al.*, "Image formation modeling in cryo-electron microscopy," J. structural biology **183**, 19–32 (2013).

41. O. H. Schiøtz, C. J. Kaiser, S. Klumpe, *et al.*, "Serial lift-out: sampling the molecular anatomy of whole organisms," Nat. Methods pp. 1–9 (2023).

42. K. Kelley, A. M. Raczkowski, O. Klykov, *et al.*, "Waffle method: A general and flexible approach for improving throughput in fib-milling," Nat. Commun. **13**, 1857 (2022).

43. B. Kim and T. Naemura, "Blind deconvolution of 3d fluorescence microscopy using depth-variant asymmetric psf," Microsc. Res. Tech. **79**, 480–494 (2016).

44. D. Boltje, S. V. Loginov, M. N. Hensgens, *et al.*, "Data underlying the publication: Depth-dependent scaling of axial distances in light microscopy," (2024).