Soft Tissue Mechanics in Hip Distraction after Total Knee Arthroplasty: A Finite Element Analysis

Sophia Soehnlen, MS[1,2,] Sara Sadeqi, PhD[2,3], Yogesh Kumaran, MS[2], Sudharshan Tripathi, MS[2],

Ryan K. Jones, BS[2], David H. Sohn, MD, JD[2], Vijay K. Goel, PhD[2]

[1]Ohio State University Wexner Medical Center, Department and Orthopaedics, Columbus, OH, USA

[2]Engineering Center for Orthopedic Research (E-CORE), Departments of Bioengineering and Orthopaedic Surgery, University of Toledo, Toledo, OH, USA

[3]Food and Drug Administration, Washington DC, USA

Address Correspondence to:

Sophia Soehnlen

Sophia.soehnlen@osumc.edu

Abstract: INTRODUCTION: Improvement in diagnostic and surgical techniques in hip arthroscopy have led to a surge in hip distraction procedures over the recent years with the predicted annual frequency being four out of every 10,000 orthopedic procedures in 2017. Due to the large traction force required to achieve the appropriate joint spacing intra-operatively, an emergence of traction-related neurological and soft tissue injuries have surfaced. Pre-existing hip joint pathologies and surgical procedures disrupt the biomechanical stability of the joint and significantly increase the risk of iatrogenic damage. Furthermore, patients with total knee arthroplasties are often subject to intra-articular ligament releases, leading to reduced stability; however, it is not well understood how this may impact their outcomes of hip arthroscopic procedures. The current study aims to investigate the biomechanical behavior of various instrumented knee joints subjected to traction forces to aid clinical understanding and advancements of hip arthroscopy techniques. METHODS: A validated finite element (FE) model of the pelvis and lower extremity was developed from computed tomography (CT) scans of a healthy 45-year-old female. Three different models were assembled according to different TKA techniques performed: Bi-Cruciate Retaining (BCR) model, Posterior-Cruciate Retaining (PCR) model, and Posterior Stabilized (PS) model. The BCR model is noted by retaining all native ligaments of the knee joint (ACL, PCL, MCL, and LCL), whereas the PCR model was subject to ACL removal and the PS model required ACL and PCL removal (Figure 1). The pelvis was encastered to prevent translation under the traction forces as motion of the patient's trunk is restrained, intraoperatively. To simulate the loading condition of hip distraction, an axial force was coupled to the distal fibula and tibia and incrementally increased from 100N to 500N. Joint spacing and ligament strain in the hip and knee joint were analyzed to assess the effects of traction forces.

RESULTS: The medial and lateral compartment stiffness of the knee joint was analyzed under hip distraction for the three different TKA scenarios. The BCR model displayed the greatest average knee complex stiffness. Release of the ACL resulted in a larger decrease of stiffness compared to release of the PCL. There was no change in forces required for hip distraction as result of changes in the knee joint stiffness (Figure 3). The PCR and PS models were subject to excess knee joint distraction that exceeded 12 mm and ligament strain greater than 20% before adequate hip joint distraction of 10 mm was achieved. The BCR model remained below 10 mm of knee distraction and 15% ligament strain at 10 mm of hip joint distraction. DISCUSSION: Our study reveals patients undergoing hip distraction with a prior TKA may experience increased soft tissue damage or iatrogenic dislocation due to reduced knee joint stability. The PCR and PS models outline a trend suggesting patients who have undergone ligament sacrificing TKAs experience large reductions in knee joint stability, causing strain levels that are indicative of soft tissue injury. The BCR TKA was indicated to be the safest under the distraction conditions as joint spacing and strain levels were largely reduced comparatively; however, when surpassing 10 mm of knee joint distraction at forces greater than 350 N, the strain levels in the ACL suggest minor injury may occur.

 Introduction:

Hip distraction is a versatile surgical technique employed in minimally invasive hip

arthroscopies for treating intra- and extraarticular pathologies and in a wide range of other

interventions including joint preservation surgeries, management of hip dysplasia, treatment of

avascular necrosis, fracture reduction and fixation, external fixation for pelvic injuries, and

conservative management of hip conditions [add cite]. Recent advancements in hip arthroscopy,

including novel instrumentation and refined surgical techniques, have not only increased the

frequency of these procedures to four per 10,000 orthopedic interventions in 2017 [1], but also

significantly reduced surgical complications and expedited patient recovery, presenting a marked

improvement over traditional open surgical methods [2, 3]. During the procedure, achieving

adequate joint visualization necessitates mild dislocation, achieved through the application of

traction. Distraction devices employ a consistent biomechanical method, wherein the patient's foot

is secured in a boot connected to a gliding mechanism. This setup facilitates the application of

traction forces, pulling the boot—and consequently the leg—away from the patient, generating an

axial force through the lower extremity and joints [4, 5].

Due to the large traction force provided to achieve the appropriate joint spacing, an

emergence of distraction-related injuries has surfaced that lead to neurologic and soft tissue

damage [3, 6, 7]. Factors such as body weight, sex differences, and bone morphology play a crucial

role in determining the necessary traction force for ideal joint spacing [8-10]. Moreover, pre-

existing conditions and previous surgeries of the knee and hip joints can compromise the

biomechanical stability of the joint, elevating the risk of iatrogenic damage during hip distraction

[11]. A particular concern arises in patients with prior total knee arthroplasty (TKA), where the

removal of the cruciate ligaments diminishes knee joint stability post-operatively [12, 13]. In

response, the bicruciate sparing TKA technique has been introduced to preserve the biomechanical

characteristics of a healthy knee, enhancing post-operative outcomes [14]. Nevertheless, the

biomechanical integrity of the knee joint complex may still be at risk during hip distraction,

regardless of the surgical technique employed.

To date, the mechanical behavior of the knee joint subjected to varying traction forces

during hip arthroscopy in patients with a history of TKA remains unexplored. This gap in

knowledge represents a critical barrier to optimizing surgical outcomes and minimizing the risk of

iatrogenic injury. Our study seeks to bridge this gap by quantifying the displacement and strain

exerted on the knee joint and its ligaments during hip distraction, specifically after different TKA

techniques. Through a comprehensive parametrized finite element (FE) analysis, we aim to

provide surgeons with detailed insights into the biomechanics of hip distraction, thereby informing

safer surgical practices and enhancing patient care.

Methods:

Model Development

A previously validated finite element (FE) model of the lower extremity was utilized in this study

[15]. The bony and main ligament geometries of a 45-year-old female were obtained from

computed tomography and magnetic resonance imaging, respectively. Mimics v15 (Materialise,

Leuven, Belgium) was used for image reconstruction and imported into Geomagic Studio

(3DSystems, Morrisville, North Carolina, USA) to smooth the geometries. Three-dimensional

geometries for TKA instrumentation were created using Solidworks computer aided design

(Dassault Systèmes, Vélizy-Villacoublay, France). Blender (Blender foundation) was utilized to

perform Boolean functions to remove the bony features of the 3D geometry where the implant was

placed. Following a previous study [15], the TKA was implanted under guidance of an orthopedic

surgeon.

The hip capsule (Figure 2) was created in SolidWorks to fit the geometry of the model based on

capsule attachment points noted by Stewart et al [16]. The capsule thickness fell within the

ranges listed in literature [17].

A mesh convergence study was performed on the components of the model to ensure proper mesh

density and quality. Meshlab [18] was used to apply surface mesh to each component which was

converted into hybrid tetrahedral elements in Abaqus 2019 (Simulia, Providence, Rhode Island,

USA). Mesh size was reduced to approximately half the total number of elements in varying

iterations. The parts were subjected to the same loading and boundary conditions and the maximum

von Mises stress was recorded. The mesh density was chosen by selecting the element size for

each category that fell within 5% of the most densely meshed trial, as shown in Supplemental

Table A. The meshed components were assembled in Abaqus 2019.

Model Validation

Both linear elastic and hyperelastic material properties were used for the ligament and other soft

tissue structures, while only linear elastic material properties were used to model the implant

components and bony features as outlined in Supplemental Table 4.1 and 4.2 [15, 19]. In the knee

joint, the anterior cruciate ligament (ACL), posterior cruciate ligament (PCL), medial collateral

ligament (MCL), and lateral collateral ligament (LCL) were modeled with hyperelastic material

models [20]. Peters et al [21] reported experimental stress-strain data of the knee ligaments for a

range of ages in males and females. To age-match the material properties closely to the current

model, the experimental data from a healthy 43-year-old female was utilized. A hyperelastic, neo-

Hookean material model was applied to the ACL, PCL, MCL, and LCL [22]. The other soft tissue

structures of the knee complex were modeled using truss elements with linear elastic material

properties based on previous studies [23].

To validate the hip joint under traction forces, hip joint dislocation was compared against cadaveric

hip load-displacement data [24]. The pelvis was fixed, and an axial load was applied distally on

the femur, at the intercondylar fossa. Frictionless surface to surface interactions with exponential

pressure-overclosure behavior were used to model the hip joint based on a previous FE study [25].

The hyperelastic, Yeoh material model was used to model the hip capsule as suggested by Stewart

at el [16]. Yeoh material coefficients (Table 1) were experimentally determined to fit the curve of

the reported cadaveric load-displacement data (Figure 4) [16, 24].

The stiffness of the knee ligament complex was validated against clinical data reported for

ligament balancing using the Depuy Synthes Knee Balancer (Depuy Synthes, Warsaw, IN, USA)

in PCR TKAs. For the validation study, the distal portion of tibia and fibula was encastered to

restrict translation and rotation. Using SolidWorks, metal paddles resembling the Knee Balancer

device were created and placed on the femoral condyles as directed by device protocol [26]. An

axial force was applied on each paddle similar to the device mechanism described by Völlner et al

[26]. Force displacement curves were constructed and compared against previously collected

clinical data that used the same device [26, 27].

Simulations

Three different models were assembled according to different TKA techniques performed: Bi-

Cruciate Retaining (BCR) model, Posterior-Cruciate Retaining (PCR) model, and the Posterior

Stabilized (PS) model (Figure 3A-C). The BCR model is noted by retaining all native ligaments

of the knee joint (ACL, PCL, MCL, and LCL), whereas the PCR model was subject to ACL

removal and the PS model required ACL and PCL removal [28]. The TKA instrumentation was

modelled with a rigid "tie" interaction at the bone interface to limit motion between implant and

bone. Frictionless, hard contact surface to surface interactions were applied at the patellofemoral

and tibiofemoral joints. The femur was angled at 10° relative to the pelvis under the guidance of

an orthopedic surgeon to fall within the range of previously reported positioning of the lower

extremity [4, 29]. The pelvis was encastered to prevent its motion under the traction forces as

motion of the patient's trunk is restrained, intraoperatively [4]. To simulate the loading condition

of hip distraction, an axial force was coupled to the distal fibula and tibia [29, 30]. Clinical studies

have indicated a wide range of required traction forces (100N-600N) and a recommended

minimum of 10 mm of hip joint distraction [8, 31-35]; thus, the present model was subjected to

traction forces ranging between 100N to 500N to analyze joint behavior.

Data Analysis

Joint spacing and ligament strain were analyzed to assess the effects of traction forces. Relative

strain (i.e. strain relative to pre-traction length) was calculated following methodology previously

reported [36]. To measure the compartmental knee joint spacing, two points approximately vertical

to each other were selected in the center region of the tray component and femoral component of

the TKA system for the medial and lateral compartments. The difference before and after traction

was measured to calculate the knee joint spacing. For hip joint distraction, a point selected on the

femoral head near the centroid was used to calculate the displacement of the femur respective to

the hip joint as suggested by a previous study [37].

Results

Validation

The FEA simulation of hip dislocation fell within an agreeable range of previously measured

cadaveric data (Figure 4). Moreover, the joint stiffness for the PCR model fell within the ranges

of clinical data shown in Table 2. The trend data for ligament balancing during a PCR TKA in

Figure 4 fell within range of *in-vivo* data.

Hip Joint Distraction

Figure 6 displays the hip joint distraction for each of the implant scenarios. The implant scenario

did not alter the force needed to distract the hip joint. Based on the clinical suggestion of 10 mm

of hip joint spacing, adequate joint visualization for the present model was achieved approximately

between 250 N to 350 N [37].

Knee Complex

The lateral and medial compartmental stiffness of the knee joint was analyzed under hip distraction

for the three different TKA scenarios. The BCR model displayed the greatest average knee

complex stiffness of 39.65 N/mm for the medial compartment and 37.48 N/mm for the lateral

compartment compared to the PCR and PS implants (Figure 7). With release of the ACL for the

PCR implant, there was a 31% and 34% decrease in the medial and lateral compartment stiffness,

respectively, compared to the intact BCR model. Release of the PCL and ACL for the PS implant

resulted in a stiffness decrease of 42% medially and 43% laterally compared to the BCR, but only

a 16% and 15% decrease in medial and lateral compartment stiffness compared to the PCR case.

The knee joint spacing during distraction was recorded. The PS implant resulted in the greatest

knee joint spacing at each force, whereas the BCR model had the least amount of joint spacing for

both compartments (Figure 8). On average, the PCR model's medial compartment resulted in an

average of 45% greater joint spacing compared to the BCR model, and the PS model resulted in a

72% increase in medial compartment spacing.

Knee Ligament Strain

The relative strain of the four main ligaments were measured under traction forces. The MCL ligament displayed the highest levels of strain in the PS model and the lowest strains in the BCR model. Comparing the PCR and PS cases, the presence of the PCL played a larger role in reducing MCL strain compared to the LCL, especially at higher loading conditions (Figure 9). The ACL was noted to have the lowest strain out of the ligaments present in the BCR model. The presence of the ACL played a greater role in reducing MCL and LCL strain compared to the PCL. The percent increase in strain of the MCL and LCL for the PCR model compared to the BCR model was 54% and 67%, respectively. The percent increase in strain for the MCL and LCL for the PS model compared to the PCR model was 18% and 15%, respectively. However, the strain for the MCL and LCL for the PS model compared to the BCR model increased by 87% and 95%, respectively.

Discussion

Our findings indicate that knee joints post-surgery, particularly those with ligament release, may face a heightened risk of further ligamentous damage due to excessive distraction. The PCR knee and PS model were subject to excess knee joint distraction and ligament strain before adequate hip joint distraction was achieved. The bi-cruciate retaining TKA was indicated to be the safest under the distraction conditions, however strain levels were indicative of potential soft tissue injury, which supports the common clinical reports of pain and instability in the knee joint, post-operatively. This suggests a need for the development of safer distraction techniques to minimize knee joint loading during hip distraction in vulnerable knee joints.

The reported complication rates for hip arthroscopies vary from 1.4% to 7%, encompassing issues such as transient neurapraxia and iatrogenic soft tissue injury [38-41] largely attributed to the

substantial forces required for hip joint distraction [7, 40]. Despite awareness of these risks,

research into the biomechanical impact of distraction forces on the knee joint during surgery

remains scarce [6]. Beutel et al [42] investigated hip arthroscopy complications in a small cohort

of 9 patients with previous lower-extremity arthroplasty, finding no adverse effects. However, with

only five of these patients having undergone knee arthroplasty (either total or uni-compartmental),

the findings offer limited generalizability. This study, therefore, seeks to fill a critical gap by

examining the biomechanical responses of knee joints with prostheses when subjected to traction

forces, aiming to enhance the safety and efficacy of hip arthroscopy techniques.

The results of the validation simulations agreed with previous literature findings for knee ligament

complex stiffness in patients with posterior cruciate retaining TKRs. The hip joint dislocation

under axial traction forces were similar to cadaveric load displacement curves (Figure 4).

Recognizing the variability in patient-specific traction forces [31], the current study utilized a wide

range of forces to account for these differences. Recent research has identified a positive

correlation between traction forces and factors such as body weight and the presence of hip

osteoarthritis [31]. However, Yin at el suggests lower extremity muscle volume is a greater

indicator of higher traction forces than BMI [43]. Thus, these patient factors must be considered

when determining risk for patients with prior TKAs.

The stiffness of the hip joint (Figure 4) was shown to be greater than that of the knee complex

subjected to PCR and PS TKAs (Figure 7). Previous literature supports these findings with the hip

capsule stiffness ranging between 12.71 N/mm to 60.20 N/mm based on section [24]. However,

there is limited knowledge of the intact stiffness of the knee complex under tensile forces [44].

Our results from the BCR model with intact knee ligaments report a stiffness of 39.65 N/mm and

37.78 N/mm for the medial and lateral compartments, respectively. Based on the previous

literature findings for hip joint stiffness and our model predictions, the lower stiffness of the knee

joint may predispose it to negative effects from hip distraction forces. Moreover, we observed a

large reduction in knee complex stiffness following ligament releases for PCR and PS TKAs

(Figure 7). Ligament balancing, aimed at correcting varus and valgus deformities in TKR patients,

often involves sequential ligament releases in PCR TKAs. Yet, literature advises against complete

MCL releases due to its critical role in passive knee stabilization [44] with stiffness potentially

decreasing by up to 40% post-release [44]. Based on the trends outlined in the present study,

patients undergoing additional ligament releases may be at an increased risk for injury during hip

distraction. The contribution of the PCL to knee stability under tensile forces remains a topic of

debate, with no consensus on its biomechanical role when the knee is fully extended [44, 45].

Figure 8 suggests a relationship between the PCL's contribution to stability and the applied traction

force. Our results indicate that the PCL's role in passive stabilization is comparatively minor, with

only a 15% and 16% stiffness difference observed between the PCR and PS models, respectively.

This underscores the predominant stabilizing functions of the ACL and MCL.

Knee complex stiffness was observed to vary with changes in intraoperative knee joint spacing

and ligament strain, showing an average increase in joint spacing of 45% for PCR and 72% for PS

models relative to the intact BCR model (Figure 8). While injury can occur within a wide range of

strain values depending on patient specific factors, there is limited data quantifying ligament

strains in intact knee kinematics [46, 47]. In the present study, within the traction force ranges of

250N to 350N needed for adequate joint visualization, the strain for the MCL and LCL was greater

than 20% for the PCR and PS simulations which may suggest potential for complications [46].

The BCR model resulted in strain levels that remained below 20%. Moreover, when knee joint

spacing exceeds 10 mm of distraction there may be risk for ligament failures and iatrogenic

dislocation due to high strain values noted in Figure 9. BCR implants have been debated due to

poor surgical outcomes with unclear indications for use, making it a less widely used implant

compared to the PCR [48]. In one study of a small cohort, only 16% of patients received a BCR

with the main indication of use being an intact ACL [49]. The present study suggests BCR implants

may reduce the chance of traction related complications because of retaining the native ligaments.

The strain values noted in the BCR model were within reasonable range compared to strain values

reported during native knee motion, suggesting a minimal risk of injury [50, 51]. However, when

surpassing the recommended 10 mm of distraction, the strain levels in the ACL suggest minor

injury may occur. Our FE data findings, which show large changes in knee complex stiffness and

joint spacing, align numerically with clinical observations reported in the literature, where

approximately 50% of patients experience sensations of instability and pain in the knee following

procedures involving hip distraction [6]. This numerical agreement suggests a potential

biomechanical underpinning for the clinical symptoms observed post-procedure, warranting

further investigation to explore the direct impact of biomechanical alterations on patient outcomes.

The high risk of traction related problems has been recognized, and in response, various surgical

procedures have emerged to reduce traction force and time [52, 53]. Lateral or supine patient

positioning are used during hip arthroscopy, but the superiority of one over the other is debated

[54]. Despite the approach used, abduction and flexion of the hip are used to relax the hip capsule

to allow for easier distraction [10, 55, 56]. Additionally, various surgical interventions have been

proposed to alter the resistive forces of the hip capsule. Distension combined with traction

increases joint distraction upwards of 2.25 times compared to traction alone [57, 58]. Furthermore,

joint venting prior to the application of traction significantly reduces the amount of traction needed,

however, the required force still fell within the range of 100 N to 450N [34, 59]. Based on the

trend outlined in the present study, the force required remains in a harmful range for patients with

PCR and PS TKRs. Moreover, capsulotomies have been a debated procedure to reduce traction

forces [60, 61]. The size of the capsulotomy is known to reduce the traction force; however,

capsule repair is essential to return the capsule to its native functionality [62, 63]. The combination

of joint venting and capsulotomies further reduces the traction force safely, but the lower extremity

joints remain exposed to excessive traction [8, 34]. Further research is needed to study and combat

hip distraction forces for vulnerable patient populations.

Limitations

While our findings underscore the pivotal role of the ACL in knee stabilization during distraction

forces, it's important to note that the material properties used in our analysis were derived from a

healthy, non-arthritic individual. Given the known susceptibility of intraarticular ligaments to

degradation from osteoarthritis (OA) and aging [21, 64] our study's application of semi-patient-

specific material properties may not fully encapsulate the biomechanical realities of younger

patients or those with advanced OA. Despite this limitation, our results indicate a discernible

pattern: individuals with a history of TKA, potentially exhibiting compromised ACL and PCL

properties due to OA, could be at an elevated risk of injury during surgery. This insight, while

derived from a constrained model, highlights the necessity for further research into the

biomechanical implications of ligament degradation on surgical outcomes.

Our approach to modeling muscle forces as uniaxial truss-connector elements with linear, elastic

material properties, while simplified [19], closely mirrors the clinical scenario during hip

distraction procedures where patients are under general anesthesia. In this state, muscles indeed

function as passive stabilizers, akin to the passive mechanical behavior assumed in our simulations.

This aspect of our model is particularly relevant given that, under anesthesia, the dynamic and

active properties of muscle forces are substantially reduced, aligning with the conditions we have

simulated. A previous study highlighted increased hip traction forces in awake patients [32], a

situation divergent from standard practice where anesthesia is employed, thus reinforcing the

clinical applicability of our modeling assumptions. Therefore, rather than detracting from the

study's validity, this simplification may, in fact, enhance the relevance of our findings to the typical

clinical environment, where the procedure is performed under conditions that render muscle forces

passive. This alignment with clinical practice supports the legitimacy of our study and highlights

the potential applicability of our results in understanding the biomechanical implications of hip

distraction procedures.

Lastly, the PS model was simplified by using the PCR model with a PCL release. While simplified

in the present study, a PS implant commonly has a uniquely designed cam-post that limits anterior-

posterior translation. However, this would not alter the mechanism under traction forces [65].

Conclusions

While traction forces are a significant concern for patient outcomes in hip distraction procedures,

the detailed biomechanical impact on the lower extremity, particularly following TKA, has not

been fully explored. Our biomechanical models suggest that individuals who have undergone hip

arthroscopy post-TKA with ligament release might be at a heightened risk for subsequent soft

tissue damage, highlighting the importance of carefully evaluating a patient's TKA history when

assessing the risks associated with hip distraction. Because of the inherent instability present in

TKAs, post-operatively, surgeons should exercise increased caution and incorporate techniques to

lower traction forces for patients with prior knee ligament releases. The insights from our study

point towards the necessity of further research into the biomechanical consequences of hip

distraction in patients with previous TKAs, aiming to refine safety protocols and enhance patient

care.

References

1. Marin-Pena, O., et al., *The current situation in hip arthroscopy.* EFORT Open Rev, 2017. **2**(3): p. 58-65.
2. Chung, A.S., et al., *Is Less Really More? Economic Evaluation of Minimally Invasive Surgery.* Global Spine Journal, 2020. **11**(1_suppl): p. 30S-36S.
3. Arriaza, C.R., et al., *Traction-related complications in hip arthroscopy for 26 years. A systematic review.* Journal of Hip Preservation Surgery, 2023.
4. Byrd, J.W., *Hip arthroscopy utilizing the supine position.* Arthroscopy, 1994. **10**(3): p. 275-80.
5. Mei-Dan, O., et al., *Hip Distraction Without a Perineal Post: A Prospective Study of 1000 Hip Arthroscopy Cases.* Am J Sports Med, 2018. **46**(3): p. 632-641.
6. Frandsen, L., et al., *Traction-related problems after hip arthroscopy.* J Hip Preserv Surg, 2017. **4**(1): p. 54-59.
7. Papavasiliou, A.V. and N.V. Bardakos, *Complications of arthroscopic surgery of the hip.* Bone Joint Res, 2012. **1**(7): p. 131-44.
8. Roling, M.A., et al., *Traction force for peroperative hip dislocation in hip arthroscopy.* Hip Int, 2020. **30**(3): p. 333-338.
9. Arvidsson, I., *The hip joint: forces needed for distraction and appearance of the vacuum phenomenon.* Scand J Rehabil Med, 1990. **22**(3): p. 157-61.
10. Dienst, M., et al., *Effects of traction, distension, and joint position on distraction of the hip joint: an experimental study in cadavers.* Arthroscopy, 2002. **18**(8): p. 865-71.
11. Truntzer, J.N., et al., *Hip arthroscopy in the United States: an update following coding changes in 2011.* J Hip Preserv Surg, 2017. **4**(3): p. 250-257.
12. Cromie, M.J., et al., *Posterior cruciate ligament removal contributes to abnormal knee motion during posterior stabilized total knee arthroplasty.* Journal of Orthopaedic Research, 2008. **26**(11): p. 1494-1499.
13. Verra, W.C., et al., *Retention versus sacrifice of the posterior cruciate ligament in total knee arthroplasty for treating osteoarthritis.* Cochrane Database Syst Rev, 2013. **2013**(10): p. Cd004803.
14. Sabatini, L., et al., *Bicruciate-retaining total knee arthroplasty: What's new?* World J Orthop, 2021. **12**(10): p. 732-742.
15. Ingels, M.L., *Mechanics of Patellofemoral Maltracking in Total Knee Arthroplasty: A Finite Element Analysis*. 2016, University of Toledo.
16. Stewart, K.J., et al., *Implementing capsule representation in a total hip dislocation finite element model.* Iowa Orthop J, 2004. **24**: p. 1-8.
17. Kay, J., et al., *The dimensions of the hip capsule can be measured using magnetic resonance imaging and may have a role in arthroscopic planning.* Knee Surg Sports Traumatol Arthrosc, 2020. **28**(4): p. 1246-1261.
18. Cignoni, P.C., Massimiliano; Ranzuglia, Guido, *MeshLab: an Open-Source 3D Mesh Processing System.* ERCIM News, 2008.
19. Kiapour, A., et al., *Finite element model of the knee for investigation of injury mechanisms: development and validation.* J Biomech Eng, 2014. **136**(1): p. 011002.
20. Sadeqi, S.S., Rodney; Erbulut, Deniz U.; Goel, Vijay K., *Optimization of Material Coefficients in the Holzapfel-Gasser-Ogden Material Model for the Main Four Ligaments of the Knee Joint-A Finite Element Study.* Applied Mathematics, 2021. **Vol. 12**(No.12).
21. Peters, A.E., et al., *Ligament mechanics of ageing and osteoarthritic human knees.* Front Bioeng Biotechnol, 2022. **10**: p. 954837.

22.   Mootanah, R., et al., *Development and validation of a computational model of the knee joint for*
the evaluation of surgical treatments for osteoarthritis. Computer Methods in Biomechanics and
Biomedical Engineering, 2014. **17**(13): p. 1502-1517.
23.   Erbulut, D.U., et al., *Tibiofemoral Cartilage Contact Pressures in Athletes During Landing: A*
Dynamic Finite Element Study. J Biomech Eng, 2021. **143**(10).
24.   Stewart, K.J., et al., *Spatial distribution of hip capsule structural and material properties.* J
Biomech, 2002. **35**(11): p. 1491-8.
25.   Joukar, A., et al., *Effects on hip stress following sacroiliac joint fixation: A finite element study.*
JOR Spine, 2019. **2**(4): p. e1067.
26.   Vollner, F., et al., *A simple method for determining ligament stiffness during total knee*
arthroplasty in vivo. Sci Rep, 2019. **9**(1): p. 5261.
27.   Vollner, F., et al., *Stability of knee ligament complex of Thiel-embalmed cadaver compared to in*
vivo knee. J Mech Behav Biomed Mater, 2017. **71**: p. 392-396.
28.   D'Anchise, R., et al., *Posterior cruciate ligament-retaining and posterior-stabilized total knee*
arthroplasty: differences in surgical technique. Joints, 2013. **1**(1): p. 5-9.
29.   Mannava, S., et al., *Basic Hip Arthroscopy: Supine Patient Positioning and Dynamic Fluoroscopic*
Evaluation. Arthrosc Tech, 2015. **4**(4): p. e391-6.
30.   Metz, A.K., et al., *Female Patients and Decreased Hip Capsular Thickness on Magnetic Resonance*
Imaging Associated With Increased Axial Distraction Distance on Examination Under Anesthesia:
An In Vivo Study. Arthroscopy, 2022. **38**(12): p. 3133-3140.
31.   Ellenrieder, M., et al., *Patient-specific factors influencing the traction forces in hip arthroscopy.*
Arch Orthop Trauma Surg, 2017. **137**(1): p. 81-87.
32.   Eriksson, E., I. Arvidsson, and H. Arvidsson, *Diagnostic and operative arthroscopy of the hip.*
Orthopedics, 1986. **9**(2): p. 169-76.
33.   Griffiths, E.J. and V. Khanduja, *Hip arthroscopy: evolution, current practice and future*
developments. Int Orthop, 2012. **36**(6): p. 1115-21.
34.   O'Neill, D.C., et al., *Joint Venting Prior to Hip Distraction Minimizes Traction Forces During Hip*
Arthroscopy. Arthroscopy, 2021. **37**(7): p. 2164-2170.
35.   Grontvedt, T. and L. Engebretsen, *Arthroscopy of the hip.* Scand J Med Sci Sports, 1995. **5**(1): p.
7-9.
36.   Sadeqi, S., et al., *Effect of Whole Body Parameters on Knee Joint Biomechanics: Implications for*
ACL Injury Prevention During Single-Leg Landings. Am J Sports Med, 2023. **51**(8): p. 2098-2109.
37.   Kapron, A.L., et al., *Patient-Specific Parameters Associated With Traction in Primary and Revision*
Hip Arthroscopic Surgery. Orthop J Sports Med, 2018. **6**(11): p. 2325967118807707.
38.   Casp, A. and F.W. Gwathmey, *Hip Arthroscopy: Common Problems and Solutions.* Clin Sports
Med, 2018. **37**(2): p. 245-263.
39.   Jamil, M., et al., *Hip arthroscopy: Indications, outcomes and complications.* Int J Surg, 2018.
54(Pt B): p. 341-344.
40.   Simpson, J., H. Sadri, and R. Villar, *Hip arthroscopy technique and complications.* Orthop
Traumatol Surg Res, 2010. **96**(8 Suppl): p. S68-76.
41.   Kowalczuk, M., et al., *Complications following hip arthroscopy: a systematic review and meta-*
analysis. Knee Surg Sports Traumatol Arthrosc, 2013. **21**(7): p. 1669-75.
42.   Beutel, B.G., et al., *Hip arthroscopy outcomes, complications, and traction safety in patients with*
prior lower-extremity arthroplasty. Int Orthop, 2015. **39**(1): p. 13-8.
43.   Yin, Y., et al., *The traction force of the pulled limb in hip arthroscopic surgery is determined by*
stiffness coefficient which is significantly related to muscle volume. Knee Surg Sports Traumatol
Arthrosc, 2023. **31**(7): p. 2708-2715.

44.   Vollner, F., et al., *Weakening of the knee ligament complex due to sequential medial release in*
total knee arthroplasty. Arch Orthop Trauma Surg, 2019. **139**(7): p. 999-1006.

45.   Matsueda, M., et al., *Soft tissue release in total knee arthroplasty. Cadaver study using knees*
without deformities. Clin Orthop Relat Res, 1999(366): p. 264-73.

46.   Butler, D.L., M.D. Kay, and D.C. Stouffer, *Comparison of material properties in fascicle-bone units*
from human patellar tendon and knee ligaments. J Biomech, 1986. **19**(6): p. 425-32.

47.   Quapp, K.M. and J.A. Weiss, *Material characterization of human medial collateral ligament.* J
Biomech Eng, 1998. **120**(6): p. 757-63.

48.   Trecci, A., *Bicruciate retaining.* Ann Transl Med, 2016. **4**(7): p. 128.

49.   Christensen, J.C., et al., *Higher Frequency of Reoperation With a New Bicruciate-retaining Total*
Knee Arthroplasty. Clin Orthop Relat Res, 2017. **475**(1): p. 62-69.

50.   Taylor, K.A., et al., *In vivo measurement of ACL length and relative strain during walking.* J
Biomech, 2013. **46**(3): p. 478-83.

51.   Hosseini Nasab, S.H., et al., *Loading Patterns of the Posterior Cruciate Ligament in the Healthy*
Knee: A Systematic Review. PLoS One, 2016. **11**(11): p. e0167106.

52.   O'Neill, D.C., et al., *Postless Hip Distraction Systems Decrease the Amount of Traction Force*
Needed to Obtain Adequate Hip Distraction Versus a Conventional Post Hip Distractor.
Arthroscopy, 2023. **39**(3): p. 740-747.

53.   Gupta, A., et al., *Safety measures in hip arthroscopy and their efficacy in minimizing*
complications: a systematic review of the evidence. Arthroscopy, 2014. **30**(10): p. 1342-8.

54.   de Sa, D., et al., *A Comparison of Supine and Lateral Decubitus Positions for Hip Arthroscopy: A*
Systematic Review of Outcomes and Complications. Arthroscopy, 2016. **32**(4): p. 716-25 e8.

55.   Glick, J.M., et al., *Hip arthroscopy by the lateral approach.* Arthroscopy, 1987. **3**(1): p. 4-12.

56.   Lall, A.C., et al., *Perineal Pressure During Hip Arthroscopy Is Reduced by Use of Trendelenburg: A*
Prospective Study With Randomized Order of Positioning. Clin Orthop Relat Res, 2019. **477**(8): p.
1851-1857.

57.   Dienst, M., et al., *Effects of traction, distension, and joint position on distraction of the hip joint:*
an experimental study in cadavers. Arthroscopy, 2002. **18**(8): p. 865-71.

58.   Byrd, J.W. and K.Y. Chern, *Traction versus distension for distraction of the joint during hip*
arthroscopy. Arthroscopy, 1997. **13**(3): p. 346-9.

59.   Mortensen, A.J., et al., *Hip Joint Venting Decreases the Traction Force Required to Access the*
Central Compartment During Hip Arthroscopy. Arthrosc Sports Med Rehabil, 2023. **5**(3): p. e589-
e596.

60.   Khair, M.M., et al., *The Effect of Capsulotomy and Capsular Repair on Hip Distraction: A*
Cadaveric Investigation. Arthroscopy, 2017. **33**(3): p. 559-565.

61.   Perry, A.K., et al., *Top Ten Pearls for Successful Hip Arthroscopy for Femoroacetabular*
Impingement. Arthrosc Tech, 2021. **10**(8): p. e2033-e2042.

62.   Weber, A.E., et al., *Vertical Extension of the T-Capsulotomy Incision in Hip Arthroscopic Surgery*
Does Not Affect the Force Required for Hip Distraction: Effect of Capsulotomy Size, Type, and
Subsequent Repair. Am J Sports Med, 2018. **46**(13): p. 3127-3133.

63.   Hartwell, M.J., S.G. Moulton, and A.L. Zhang, *Capsular Management During Hip Arthroscopy.*
Curr Rev Musculoskelet Med, 2023.

64.   Schulze-Tanzil, G., *Intraarticular Ligament Degeneration Is Interrelated with Cartilage and Bone*
Destruction in Osteoarthritis. Cells, 2019. **8**(9).

65.   Murakami, K., et al., *In vivo kinematics of gait in posterior-stabilized and bicruciate-stabilized*
total knee arthroplasties using image-matching techniques. Int Orthop, 2018. **42**(11): p. 2573-
2581.

Figures

Figure 1. Meshed parts of the A.) Posterior Cruciate retaining total knee replacement system and
517   the B.) Bi-cruciate retaining total knee replacement system.

Figure 2. Hip capsule geometry and anatomical attachment sites based on Stewart et al [16]

Figure 3. FE models of instrumented TKA. A) PCR TKA with simulated ACL release. B) PS
TKA with simulated ACL and PCL releases. C) BCR TKA with intact ligaments.

Figure 4. The load-displacement curve for the hip joint under various distraction forces. The gray lines represent experimental cadaveric data measuring hip joint dislocation [24]. The red line represents the FE model data with experimental curve fitted Yeoh hyperelastic material coefficients for the hip capsule.

Figure 5. Knee ligament balancing for the A) medial and B) lateral compartment for PCR TKAs. The gray points represent cadaveric and *in*-vivo reported load-displacement values [26, 27]. The orange points represent the load-displacement data for the PCR model under knee ligament balancing.

Figure 6. Hip joint distraction under axial traction force for the varying TKA implant scenarios.

602 **Figure 7**. Averaged stiffness for the knee complex during the range of distraction forces.

Figure 8. Knee joint distraction under the axially applied traction force for the A) medial and B) lateral compartment.

605
606
607

Figure 9. Measured strain in the A) MCL, B) LCL, C) PCL, and D) ACL under the traction
forces.

610
611

623   Tables
624   Table 1. Hyperelastic Yeoh material model coefficients for the hip capsule.

Yeoh Coefficients	C10	C20	C30	D1	D2	D3
	0.65	0.75	6.85	0.000887	0.000487	0.000487

651 **Table 2**. Knee compartmental stiffness for PCR TKAs during knee ligament balancing. The
652 clinical measurements were taken with the DePuy Knee Balancer System (DePuy Synthes,
653 Johnson and Johnson, Raynham, MA). The compartmental stiffness in the model was averaged
654 across a range of simulated forces used in knee ligament balancing.

Compartmental Stiffness		
	lateral compartment (N/mm)	Medial compartment (N/mm)
Cadaveric [27]	31.6	26.6
In vivo **[26, 27]**	25.3	25.7
	19.9± 0.9	28.4 ± 1.2
FE Model	25.7	27.8

Supplemental Data

Mesh Convergence

A component from each of the bones, implant, and ligaments geometries, were utilized in the mesh
convergence study to appropriately assign mesh density. The maximum Von Mises stress was recorded
for varying mesh sizes. For the bone components (Supplemental Table 1), a mesh size of 3 was selected.
A mesh size of 1 was selected for the 3D soft tissue components (ligaments and hip capsule)
(Supplemental Table 2). Moreover, a mesh size of 2 was selected for each of the TKA implant
components (Supplemental Table 3).

Supplemental Table 1. Mesh convergence study of the femur under tensile load of 50N.

# of Elements	Mesh Size	Max Von Mises (MPa)
297541	2	1.851
225953	2.25	1.827
173626	2.5	1.824
125185	2.75	1.8
100724	3	1.782

687 **Supplemental Table 2**. Mesh convergence study of the LCL ligament under tensile load of 50N.

# of Elements	Mesh Size	Max Von Mises (MPa)
17285	0.75	3.703
7495	1	3.695
4034	1.25	3.683

692 **Supplemental Table 3**. Mesh convergence study of the tibial component under tensile load of 50N.

# of Elements	Mesh Size	Max Von Mises (MPa)
126727	1	2.888
77747	1.25	3.013
46128	1.5	3.072
29369	1.75	3.099
20793	2	3.034

698 Material Properties

699 **Supplemental Table 4.1**. Implant material properties [15]

Part	Material	Young's Modulus (MPa)	Poisson's Ratio
Femoral Component	Titanium	120,000	0.3
Tibial Tray	Polyethylene	690	0.46
Tibial Component	Titanium	120,000	0.3
Patellar Dome	Polyethylene	690	0.46

701 **Supplemental Table 4.2**. Hard tissue material properties [15]

Material	Young's Modulus (MPa)	Poisson's Ratio
Cortical Bone	15,000	0.3
Cancellous Bone	400	0.3