Evaluation of phantom doping materials in quantitative susceptibility mapping

Padriac Hooper[1,2], Jin Jin[1,2,3], Kieran O'Brien[1,2,3], Monique Tourell[1,2,6], Simon Daniel Robinson[2,4,5] and Markus Barth[1,2,6]

[1] ARC Training Centre for Innovation in Biomedical Imaging Technology, Brisbane, Australia,

[2] Centre for Advanced Imaging, The University of Queensland, Brisbane, Australia,

[3] Siemens Healthcare Pty Ltd, Brisbane, Australia,

[4] High Field MR Center, Department of Biomedical Imaging and Image-Guided Therapy, Medical University of Vienna, Austria,

[5] Christian Doppler Laboratory for MR Imaging Biomarkers, Department for Biomedical Imaging and Image-Guided Therapy, Medical University of Vienna, Austria,

[6] School of Information Technology and Electrical Engineering, The University of Queensland, Brisbane, Australia

Abstract

Purpose: To measure magnetic susceptibility (χ) with Quantitative Susceptibility Mapping (QSM) and evaluate its repeatability using four phantom doping materials relevant to QSM applications.

Methods: A cylindrical phantom was constructed containing vials of agarose gel doped with two paramagnetic materials (ferritin, USPIO) and two diamagnetic materials ($CaCl_2$, $CaCO_3$) at five concentrations each. Single orientation QSM measurements (MEDI+0) were carried out on the phantom at 3T and 7T. We measured molar susceptibility (χ_{mol}) from QSM and evaluated the test-retest repeatability of χ using the standard error of the measurement (SEM). We evaluated material lifespan by conducting a t-test of χ_{mol} at various timepoints.

Results: χ_{mol} (ppm·L·mmol^{-1}) were measured as 1.67 ± 0.24 and 0.74 ± 0.09 (USPIO: 3T and 7T, respectively), $10^{-2} \times (8.13 \pm 1.35; 8.13 \pm 1.19)$ (Ferritin: 3T; 7T), $10^{-4} \times (-2.68 \pm 0.24; -2.71 \pm 0.37)$ ($CaCl_2$: 3T; 7T), and $10^{-5} \times (-9.52 \pm 1.44; -9.53 \pm 1.18)$ ($CaCO_3$: 3T; 7T). The USPIO SEM (1.5 ± 2.0; 5.1 ± 2.0 ppb at 3T; 7T) was greater than the ferritin SEM (1.2 ± 1.0; 2.2 ± 1.3 ppb at 3T; 7T). The $CaCl_2$ SEM (7.5 ± 5.5; 1.2 ± 0.6 ppb at 3T; 7T) was greater than the $CaCO_3$ SEM (1.2 ± 0.6; 0.9 ± 0.7 ppb at 3T; 7T). We observed no significant changes in molar susceptibility for ferritin and $CaCO_3$ over the measured timeframes (24 months and 15 months, respectively).

Conclusion: We recommend using ferritin and $CaCO_3$ in the construction of susceptibility phantoms, removing later echo times for $CaCO_3$ QSM reconstructions.

Word count: 5,034 words

Corresponding author: Markus Barth; m.barth@uq.edu.au

1 Introduction

MRI employs electromagnetic fields to excite and detect nuclear spin resonance. These spins act as probes of the local microenvironment, and with the appropriate model-based analysis, provide a means for quantifying physical properties related to tissue structure and composition. Quantitative parameters (e.g., T_1, T_2, diffusion coefficients) can be derived for every voxel in an MRI image and can serve as biomarkers of disease profiles[1,2]. The magnetic susceptibility of tissue, χ, is sensitive to both tissue structure and composition and can be measured using the phase component of the MRI signal acquired from T_2^*-weighted images in Quantitative Susceptibility Mapping (QSM). Clinically, QSM has produced a groundswell of interest, finding applications in mapping calcifications[3], venous oxygenation[4], and iron content[5,6]. However, several challenges remain in QSM imaging. Firstly, local χ variations (e.g., air-tissue interfaces) are a source of field distortions and produce regions of low SNR[7], which lead to artifacts in the susceptibility map[8]. Secondly, the dipole kernel contains zeros in k-space at spatial frequencies corresponding to that of a double cone[9,10], making dipole inversion an ill-posed problem necessitating regularization. Because of the absence of spatial frequencies at the center of k-space, QSM requires referencing to a known susceptibility value, which is difficult to define in vivo[11]. Additionally, the susceptibility of some tissues are not scalar but tensor, and susceptibility is affected by tissue microstructure, e.g., the radial anisotropy of the myelin sheath[12].

Susceptibility imaging phantoms that contain uniform regions of known susceptibility are not affected by patient/biological factors[13–16]. As a result, they provide reliable reference values, assess errors in MR acquisition and QSM reconstructions, and enable calibration across scanners and imaging sites[17,18]. In vivo, the predominant cause of tissue susceptibility are iron- or calcium-containing materials that produce paramagnetic (positive) or diamagnetic (negative) susceptibility contrast, motivating the use of iron-based or calcium-based materials, respectively, as χ sources in phantom inclusions. Most phantom inclusions are composed of gel mixtures in place of aqueous mixtures, since gels mimic the in vivo relaxation properties of soft tissues[19] and embed particles in a fixed position after solidification. Mimicking the signal relaxation and susceptibility of iron- and calcium-based materials provides a realistic evaluation of QSM reconstructions. The protein ferritin, is biologically relevant as a dominant form of iron stored within deep grey matter[20]. To mimic ferritin for QSM applications, Cuna et al. synthesized an iron-filled hydrogel phantom material with variable cluster size, and a comparable molar susceptibility to ferritin *in vivo*[18]. The same research group evaluated QSM susceptibility measurements with a SQUID magnetometer, showing the iron-filled hydrogel phantom material had comparable measurements to those acquired with ultra-high-field scanners[21]. An iron-filled hydrogel phantom material bears a resemblance to iron clustering systems, which is observed Alzheimer's beta-amyloid plaques[5,6]. Another paramagnetic material of interest is ultrasmall superparamagnetic iron oxide (USPIO), which is applied to QSM as a blood-pool contrast agent[4,22,23] and in magnetic fluid hyperthermia[24]. USPIO has been used to validate R_2^* mapping at

across various field strengths[25,4]. An important material observed *in vivo* are the insoluble polyhedral salt crystals, calcium carbonate ($CaCO_3$), which are present in biochemical analyses within bone mineralization and calcifications[26–28]. Emmerich et al. evaluated $CaCO_3$ particles using clinical and ultra-high-field scanners[29] and used $CaCO_3$ particles to study the separation of χ sources in QSM[30], with *in vivo* comparison to multiple sclerosis lesions[31]. An alternative diamagnetic χ source is calcium chloride ($CaCl_2$), which has been used previously by Hopkins et al. to match the susceptibility of bone[32], and is of practical utility in a phantom due to its high diamagnetism, water solubility and inert chemical properties[33].

This study aimed to estimate χ and evaluate the repeatability of MR-based χ measurements of two paramagnetic and two diamagnetic materials: USPIO, ferritin (paramagnetic), $CaCl_2$ and, $CaCO_3$ (diamagnetic). To do this, a cylindrical phantom was constructed containing vials of doped agarose gel. Susceptibility measurements were carried out on the phantom at 3T and 7T, the test-retest repeatability of susceptibility measurements were assessed and evaluated over a 9- to 24-month period. Based on our results we draw conclusions on the suitability of doping materials in QSM phantom studies.

2 Methods

2.1 Phantom design

Relevant design factors for producing the QSM phantom: it should (i) fit within most RF head coils, (ii) reduce B_0 and B_1 inhomogeneities, (iii) provide a means to trap air bubbles, and (iv) be stable over the long-term ($\geq$ 12 months).

The phantom schematics are shown in Figure 1. To ensure fitting in what is probably the most widely used head coil at ultra-high-field currently, the 7 Tesla Nova Medical 1Tx/32Rx head coil (Nova Medical, Wilmington, MA), the outer diameter of the phantom was 150 mm. As most other head coils – including other field strengths – are larger, they would be able to accommodate a phantom of this size. The phantom contains 20 NMR vials (10 mm outer diameter, N-51A, Kimble Glass) of doped agarose gel (details below) surrounded by a solution of ultrapure water (Milli-Q IQ 7000, Merck & Co.), 44.3 weight percentage (wt.%) PVP-40 (PVP-40, Merck & Co.) and 1.7 wt.% NaCl (746398, Merck & Co.) to mimic the electric permittivity and conductivity of white matter at 7T. To reduce B_1 artifacts and errors during background field correction, vials were positioned at least 10 mm away from the phantom periphery. The phantom was designed to accommodate a bubble trapping compartment outside the Field of View (FOV), connected to the main compartment by a small hole (see Figure 1 a). The tight-fitting non-magnetic closure and an NMR tube cap prevented microbes from entering the phantom and samples, respectively. The modification of electric permittivity in the phantom was important to prevent standing wave artifacts at 7T that would lead to image

inhomogeneities.

Figure 1: Side (A) and top (B) views of phantom. (C) CAD drawing of the phantom within an

approximation of the 7T Nova Medical 32 Rx coil. (D) CAD drawing indicating relative positioning of

vials.

The phantom design features four quadrants, each containing five vials. The samples were

prepared by doping hot agarose gel with one of the four dopants (USPIO, ferritin, $CaCl_2$ and $CaCO_3$)

at five equally spaced concentrations. Other relevant dopants, Gd-DTPA[34], hydroxyapatite[35], and

tungsten carbide[30] were not covered in this study. The concentrations were chosen to more than cover

the magnetic susceptibility range expected for in vivo human brain scans, with USPIO ranging from

0.22 to 0.67 mmol/L, ferritin from 3.76 to 10.21 mmol/L, $CaCl_2$ from 0.9 to 4.5 mol/L and, $CaCO_3$

from 1.0 to 5.0 mol/L. Ultrapure water (Type 1) was used as a solution for the PVP-NaCl mixture to

limit the likelihood of microorganisms or magnetic ions entering the samples. The phantom was given

at least 48 hours for the PVP-NaCl mixture to settle before scanning[36].

2.2   MR acquisition

15   The phantom was scanned at two different field strengths of 3T and 7T (all Siemens Healthineers,

16   Erlangen, Germany) with a 20-channel and 32-channel head coil, respectively. Phase and magnitude

17   images were acquired with a multi-echo 3D GRE pulse sequence using the parameters shown in Table

18   1. We used bipolar acquisitions with shortest possible echo spacings as the results of phase

unwrapping are generally more reliable when the shortest possible echo spacings are used[37,38] because SNR is high and wraps fewer in number[39]. B_0 shimming was performed using Siemens' GRE brain sequence and Prescan normalize correction[40] was used with the 20-channel head coil at 3 T. Images were reconstructed using ASPIRE[41] for phase and root sum-of-squares for magnitude and converted to NIFTI[42] format for QSM processing. The phantom was scanned twice at baseline (t_0), again after 9 months (t_1), then again twice at 24 months (t_2). The test-retest scans were acquired within 15 minutes of one another. Some materials were included and removed from the phantom at different timepoints; USPIO and ferritin included at t_0, t_1 and t_2; $CaCl_2$ included at t_0, t_1; $CaCO_3$ included at t_1, t_2.

	3T	7T
Model	Magnetom Prisma	Magnetom 7T Plus
Gradient readout	Bipolar	
$TE_1:\Delta TE:TE_{max}$; TR (ms)	1.87:1.87:22.44; 26	3.15:3.15:28.35; 32
Coil	20-ch Rx head coil	32-ch Rx head coil
Voxel size	1.0 mm^3 isotropic	0.7 mm^3 isotropic
Acquisition matrix	$192\times192\times112$	$272\times272\times160$
Flip angle; Bandwidth; Averages	15°; 1000 Hz/pix; 1	15°; 340 Hz/pix; 1
Acceleration? type and factor	✗	GRAPPA 2
Partial Fourier? factor; Elliptical scanning?	✗;✗	6/8 PFA; ✓
Prescan normalize correction?	✓	✗

Table 1: MR acquisition parameters used in this study.

2.3 Image pre-processing and corrections

Noise was measured from unfiltered magnitude images as the standard deviation of the intensity in air regions (SD_{air}), outside the phantom. Regions of interest (ROIs) $20 \times 20 \times 160$ mm^3 were drawn manually at the image corners, ensuring ROIs were free of artifacts and matrix borders. SD_{air} has a Rayleigh distribution, which can be corrected to give the expected noise level; $SD_{air}' = SD_{air}/0.66$[43]. The signal was measured using cylindrical ROIs (7 mm diameter) drawn manually within each vial, excluding partial volume and/or signal loss at the vial edge, and excluding aliasing artifacts at the 15 distal slices on either end.

The criterion for exclusion data at specific echo time for QSM and R_2^* was set to SNR less than or equal to 10:1. No echo times were excluded for USPIO, ferritin, $CaCl_2$; however, for $CaCO_3$ echo times longer than 11 ms at 3T and 3 ms at 7T (see Table S1) were excluded. For R_2^* mapping with $CaCO_3$ at 7T, echo times longer than 9 ms were excluded. For the statistical analysis of R_2^* mapping, we removed $CaCO_3$ concentrations ≥ 4.0 mol·L^{-1} since the SNR was less than or equal to 10:1 at echo times 6 to 9 ms at 7T.

To correct for slight differences in the phantom orientation and positioning between longitudinal

acquisitions, the real and imaginary images were manually co-registered (ITK-SNAP v4.0[44]) using the 1st echo magnitude as a reference, interpolating using B-splines. Phase offsets associated with opposite polarity of odd and even echoes due to a bipolar acquisition[45] were corrected using MCPC-3D-S[37]. χ-induced geometric distortions occur in the readout direction of bipolar GRE acquisitions due to the opposing direction of odd and even readout gradients[46]. Note that geometric distortions were not observed at 3T due to the high readout pixel bandwidth (see Table 1). Field maps of odd and even echoes were created, combining the phase using nonlinear complex fitting[47], then unwrapped using SPURS[48]. The voxel displacement map was calculated by dividing the field map by the readout pixel bandwidth[49]. The warped field was generated by applying smoothing to the voxel displacement map[50]. The real and imaginary images were unwarped (SPM12)[51] with tri-linear interpolation. Gibbs artifacts associated with sharp transitions in signal intensity were observed at the vial boundary. A Gibbs ringing correction was applied to the real and imaginary images using sub-voxel shifts in all three spatial dimensions[52,53].

Zero-padding of the complex GRE data matrix by a factor of 1.5 was used prior to non-linear field map estimation as it is known to result in a more accurate field map[54]. Moreover, it is well established that zero-padding reduces aliasing associated with the Fourier Transform, which affects both background field correction and dipole inversion stages of QSM processing.

2.4 R_2^* mapping

R_2^* maps were generated by voxel-wise nonlinear fitting of the multi-echo magnitude image using the Levenberg-Marquardt algorithm. Clump/agglomerate masks were produced as follows: (1) determining local R_2^* outliers (median + 3*IQR within each vial), then (2) finding continuous (6-connectivity) regions of R_2^* outliers larger than 1 mm^3. From the clump/agglomerate masks, we report (i) the percentage of local R_2^* outliers per vial, and (ii) the range of clump/agglomerate sizes in mm^3.

2.5   QSM

The phase across echo times was fitted using nonlinear complex fitting[47], then unwrapped using SPURS[48]. Background fields were then corrected using V-SHARP[55], SMV-radii = 1:1:10 mm (rounding to the nearest integer towards infinity). An 'initial mask' was generated by thresholding values within the 1st-echo magnitude image greater than 2 % of its maximum intensity. A morphological opening operation, which removes disconnected voxels from the mask, was applied using a structural spherical element of 5 mm radius. The maximum SMV radius was set to 10 mm since the vials positioned close to the perimeter were 10 mm from the mask edge, which is a limitation of V-SHARP[56]. The mask imposed on the field map and on χ during dipole inversion was set as the entire matrix as including the spatial distribution of all frequencies and χ sources improves background

field correction[57]. For 7T data, residual RF transmit coil fields (B_1^+) were corrected by fitting and subtracting a fourth-order 3D polynomial[58,59].

A 'reference mask' ($M_{reference}$) was derived analogous to the so-called 'CSF mask' for zero-referencing in MEDI+0 [60,61] in 3 steps: (1) R_2^* thresholding at 5 s^{-1}, (2) morphological closing (3 mm radius), then (3) morphological erosion (10 mm radius), see Figure S1. The 'weighting map' (used to weigh the data consistency term) was computed as described in SEPIA documentation (https://sepia-documentation.readthedocs.io/en/latest/method/weightings.html)[62]. The field noise map was inverted, normalized using the median and upper IQR, re-centered to 1, then global outliers (defined as median + 3*IQR) were replaced with a $3 \times 3 \times 3$ voxel box filtered copy. The relative residual (Equation 3) of $S_{measured}$ (Equation 2) and $S_{simulated}$ (Equation 1) was computed as follows,

$$S_{measured} = S(TE) \cdot e^{-j \cdot \phi_{S(TE_1)}} \quad (1)$$
$$S_{simulated} = S_0 \cdot e^{-R_2^* \cdot TE + j \cdot \omega \cdot TE - j \cdot \omega \cdot TE_1} \quad (2)$$
$$relative residual = \frac{\sum^{TE}|S_{simulated}(TE) - S_{measured}(TE)|^2}{\sum^{TE}|S_{measured}(TE)|^2} \quad (3)$$

In these equations, ω was the angular frequency determined during field mapping, S_0 was the extrapolated signal magnitude at TE=0, $S_{measured}$ was the measured data with the phase subtracted from the 1st echo, and $S_{simulated}$ was the simulated mono-exponential model signal with the phase subtracted from the 1st echo. The relative residual map was brought into a weighting component using a threshold of 0.3, which was then used to modulate the weighting map[62].

For dipole inversion, we used MEDI+0[60,61] using the default regularization parameters (λ_1=1000, λ_2=100). The MERIT parameter was set to false. MEDI with SMV-filtering was not used since it may emphasize the high frequency components more than low frequency components within a susceptibility distribution[8]. The percentage threshold parameter (c_∇) was optimized by minimizing the streaking artifact with MEDI+0 turned 'off', which was quantified as the standard deviation within the reference mask (SD_{Ref})[60,63]. The optimal c_∇ was found to be 0.5 for both 3T and 7T scans (see Figure S2) and was used for subsequent QSM reconstructions.

The phantom design and vial positioning lends itself to more direct quantification of susceptibility without performing dipole inversion. We used the following analytical model[7] (referred to as the 'infinite cylinder model') to generate a simulated susceptibility map,

$$\chi_{analytical} = \frac{6\delta}{3 \cos^2 \theta - 1} \quad (4)$$

In Equation (4), $\delta = \Delta f_{local}/f_0$ is the local frequency in Hz (Δf_{local}) normalized to the Larmor frequency (f_0) in parts-per-million (ppm). The angle relative to the applied field vector (θ) is assumed

to equal $0°$, therefore making $\chi_{analytical} = 3\delta$.

2.6   Automatic segmentation

We performed automatic segmentation instead of manual segmentation to obtain clear and consistent cylindrical ROIs for statistical analysis. The 'initial mask' (refer to section 2.5) was eroded by 10 mm, which is the distance between the distal vials and the mask edge. The complement of the non-eroded reference mask $M_{reference}$ (refer to Figure S1) was multiplied by the eroded initial mask, obtaining a mask of the 10 mm vials. To exclude erroneous voxels at the vial edge, the obtained mask was eroded by 3 mm. We also excluded the 15 distal slices on either end which were prone to aliasing artifacts. Segmentation was performed on the obtained mask using the cluster function within FSL[64].

2.7   Statistical analysis

The molar concentration (c_{mol}) was fitted against the mean ROI measurement for χ and R_2^*, respectively, with a least-squares regression to determine the linear fit:

$$R_2^* = R_{2\,0}^* + c_{mol} \cdot R_{2\,mol}^* \quad (5)$$

$$\chi = \chi_0 + c_{mol} \cdot \chi_{mol} \quad (6)$$

We used robust regression with a bisquare weighting function, which reduces the weight of independent variables with a high least-squares residual. $R_{2\,0}^*$; χ_0 are equal to the agarose R_2^* value; χ value, respectively. $R_{2\,mol}^*$; χ_{mol} are equal to the R_2^* relaxivity; molar susceptibility, respectively.

To assess test-retest repeatability of susceptibility measurements, a single-score coefficient of reliability (ICC) was calculated using a two-way ANOVA model with absolute agreement ('A-1')[66,67]. We also computed standard error of the measurement (SEM), which is given in terms of the standard deviation of test-retest measurements (SD) and the ICC[17],

$$SEM = SD \cdot \sqrt{(1 - ICC)} \quad (7)$$

Smaller values of SEM represent greater test-retest precision. To determine material lifespan, we performed a t-test to detect a significant difference between baseline χ_{mol} values to timepoint χ_{mol} value; the first timepoint at which there was a significant change was determined to be the material lifespan. We also performed a t-test to detect a possible correlation between time (t=0, 9, 24 month) and timepoint χ_{mol} values.

Bland–Altman (BA) analysis for repeated measurements per subject (phantom) was applied to evaluate agreement between 3T and 7T scanners[68]. We performed the following BA analyses: (1) to

determine the agreement in χ with field strength; ferritin, $CaCl_2$ and $CaCO_3$), (2) to determine the agreement in the product of the magnetization ($M=\chi \cdot B_0$) with field strength; USPIO[4], (3) to determine the agreement in R_2^*/B_0 with field strength; $CaCO_3$ [31], (4) to determine the agreement in R_2^* with field strength; USPIO, ferritin and $CaCl_2$. We also performed a t-test to detect linear correlation between cross-field measurements; if detected, then we performed linear regression to quantify the trend between cross-field measurements. To validate the dipole inversion step, Bland–Altman (BA) analysis for repeated measurements per subject (phantom)[66] was applied to evaluate agreement between the simulated analytical susceptibility map ($\chi_{analytical}$ from Equation (4)) and the measured susceptibility map (χ_{MEDI+0}).

3 Results

Maps and plots of susceptibility for the four used materials and five concentrations are shown in Figures 2 to 5. On the scatter plots, the fitted regression was dotted, and, where given, the 95% confidence intervals were solid. The vertical error bars were the standard deviation of the ROI measurement. The Bland-Altman mean difference (bias) line was solid, and the limits-of-agreement were dotted. On the correlation plots, the identity line (y = x) was dotted; the fitted regression and 95% confidence interval lines were solid.

Figure 2: USPIO χ maps (A) and R_2^* maps (B). The windowing of the χ map at 7T was set to half that of 3T. The yellow arrowhead indicated the presence of an agglomerate at USPIO 0.36 mmol/L, manifesting as bright on both the χ and R_2^* maps. (C) Scatter plot of χ as a function of USPIO concentration. (D) Bland-Altman and (E) correlation plots of $(\chi \cdot B_0)_{3T}$ versus $(\chi \cdot B_0)_{7T}$. (F) Scatter plot of R_2^* as a function of USPIO concentration. (G) Bland-Altman and (H) correlation plots of $(R_2^*)_{3T}$ versus $(R_2^*)_{7T}$.

Figure 3: Ferritin χ maps (A) and R_2^* maps (B). At 10.2 mmol/L, ferritin clumps were marked with a blue arrowhead, manifesting as dark on χ (diamagnetic) and bright on R_2^*. (C) Scatter plot of χ as a function of ferritin concentration. (D) Bland-Altman and (E) correlation plots of χ_{3T} versus χ_{7T}. (F) Scatter plot of R_2^* as a function of ferritin concentration. (G) Bland-Altman and (H) plots of $(R_2^*)_{3T}$ versus $(R_2^*)_{7T}$.

Figure 4: Calcium chloride ($CaCl_2$) χ maps (A) and R_2^* maps (B). The voxels corresponding to $CaCl_2$ 4.5 mol/L were heterogeneous, as indicated with a red arrowhead on the 7T χ map. (C) Scatter plot of χ as a function of concentration of $CaCl_2$. (D) Bland-Altman plot of χ_{3T} versus χ_{7T}. (E) Scatter plot of R_2^* as a function of concentration of $CaCl_2$. (F) Bland-Altman plot of $(R_2^*)_{3T}$ versus $(R_2^*)_{7T}$.

Figure 5: Calcium carbonate ($CaCO_3$) χ maps (A) and R_2^* maps (B) of. At 7T, the R_2^* at concentrations $\geq$ 4.0 mol/L were heterogeneous, indicated with a purple arrowhead. (C) Scatter plot of χ as a function of concentration of $CaCO_3$. (D) Bland-Altman and (E) correlation plots of χ_{3T} versus χ_{7T}. (F) Scatter plot of R_2^* as a function of concentration of $CaCO_3$. (G) Bland-Altman plot of $(R_2^*/B_0)_{3T}$ versus $(R_2^*/B_0)_{7T}$.

3.1 Analysis of clumps/agglomerates and other confounds

We termed "agglomerates" as bright clusters on a susceptibility map and "clumps" as dark clusters on a susceptibility map; clumps/agglomerates each were bright on the R_2^* maps. At 3T, the mean clump/agglomerate volumes of ferritin (3.49 mm^3) were roughly 6% larger than those of USPIO (3.27 mm^3). At 7T, the mean clump/agglomerate volumes for ferritin (13.01 mm^3) were roughly twice those of USPIO (6.40 mm^3). A single agglomerate was observed in USPIO 0.45 mmol/L (see Figure 2), and several clumps were observed in ferritin 10.2 mmol/L (see Figure 3).

For CaCl$_2$ 4.5 mol/L, we observed heterogeneity, which was attributed to it reaching the water solubility limit[33]; thus, we excluded CaCl$_2$ 4.5 mol/L from all statistical analysis. CaCO$_3$ concentrations $\geq$ 4.0 mol/L were heterogeneous on the R_2^* map (see Figure 5) due to low SNR; thus, these two concentrations were excluded from the R_2^* statistical analysis. At 3T, the mean clump/agglomerate volumes for CaCl$_2$ (1.46 mm^3) were roughly 15% larger than CaCO$_3$ (1.27 mm^3). At 7T, the mean clump/agglomerate volumes for CaCl$_2$ (5.59 mm^3) were roughly half that of CaCO$_3$ (10.40 mm^3). The mean number of R_2^* map outlier voxels per vial was less than 3.0 % for each material group at each field strength (see Table S2).

3.2 Linear regression (c_{mol} versus χ)

The linearity was excellent, $R^2 > 0.9$, for all 4 materials and field strengths. For USPIO, χ_{mol} at 3T (1.67 ppm·mmol^{-1}·L) was roughly 2.3 times greater than its χ_{mol} at 7T (0.74 ppm·mmol^{-1}·L). For ferritin, χ_{mol} at 3T and 7T were equal (8.13×10^{-2} ppm·mmol^{-1}·L). For CaCl$_2$, χ_{mol} at 3T (2.68 ×10^{-4} ppm·mmol^{-1}·L) was roughly 1% less than its χ_{mol} at 7T (2.71 ×10^{-4} ppm·mmol^{-1}·L). For CaCO$_3$, χ_{mol} at 3T (9. 52 ×10^{-5} ppm·mmol^{-1}·L) was also roughly 1% less than its χ_{mol} at 7T (9.53 ×10^{-5} ppm·mmol^{-1}·L).

		USPIO	Ferritin	CaCl$_2$	CaCO$_3$
χ_{mol} (ppm·L·mmol^{-1})	3T	1.67 ± 0.24	(8.13 ± 1.35) ×10^{-2}	(-2.68 ± 0.24) ×10^{-4}	(-9.52 ± 1.44) ×10^{-5}
	7T	0.74 ± 0.09	(8.13 ± 1.19) ×10^{-2}	(-2.71 ± 0.37) ×10^{-4}	(-9.53 ± 1.18) ×10^{-5}
χ_0 (ppb)	3T	253 ± 61	80 ± 81	25 ± 80	190 ± 43
	7T	188 ± 28	53 ± 91	19 ± 91	165 ± 39
R^2 (c_{mol} versus χ)	3T	0.95	0.95	0.98	0.94
	7T	0.95	0.94	0.97	0.96

Table 2: Linear fitting coefficients (c_{mol} versus χ) and linearity (R^2) at 3T and 7T, respectively, from MEDI+0 reconstructed susceptibility maps. Coefficients are reported with standard error.

3.3 Linear regression (c_{mol} versus R_2^*)

The linearity was poor ($R^2 < 0.5$) for $CaCl_2$ at 3T, good ($0.75 < R^2 < 0.9$) for $CaCl_2$ at 7T and $CaCO_3$ at 7T, and excellent ($R^2 > 0.9$) for $CaCO_3$ at 3T, USPIO and ferritin at both field strengths (see Table S3). For USPIO, $R_{2\ mol}^*$ at 7T (77.5 $s^{-1} \cdot mmol^{-1} \cdot L$) was roughly 6% less than $R_{2\ mol}^*$ at 3T (82.3 $s^{-1} \cdot mmol^{-1} \cdot L$). For ferritin, $R_{2\ mol}^*$ at 7T (0.77 $s^{-1} \cdot mmol^{-1} \cdot L$) was roughly 3.6 times less than $R_{2\ mol}^*$ at 7T (2.78 $s^{-1} \cdot mmol^{-1} \cdot L$). The linearity of $CaCl_2$ was poor at 3T, we report $R_{2\ mol}^*$ mean ± standard deviation as 13.1 ± 0.6 s^{-1}; 14.0 ± 1.1 s^{-1} at 3T; 7T, respectively. For $CaCO_3$, $R_{2\ mol}^*$ at 7T (0.869 $s^{-1} \cdot mmol^{-1} \cdot L$) was roughly 2.5 times greater than $R_{2\ mol}^*$ at 7T (0.352 $s^{-1} \cdot mmol^{-1} \cdot L$).

3.4   Test-retest repeatability

Coefficient of reliability (ICC) and standard error of the measurement (SEM) of test-retest measurements are shown in table 3. The test-retest repeatability was 'excellent' (ICC > 0.90) for each material at each field strength. For ferritin and $CaCO_3$, the SEM each were within 5 ppb at both field strengths. At 3T, the USPIO SEM (1.5 ± 2.0 ppb) was 25% larger than the ferritin SEM (1.2 ± 1.0 ppb). At 7T, the USPIO SEM (5.1 ± 2.0 ppb) was 1.8 times larger than the ferritin SEM (2.2 ± 1.3 ppb). At 3T, the $CaCl_2$ SEM (7.5 ± 5.5 ppb) was 625 % larger than the $CaCO_3$ SEM (1.2 ± 0.6 ppb). At 7T, the $CaCl_2$ SEM (1.2 ± 0.6 ppb) was 33% larger than the $CaCO_3$ SEM (0.9 ± 0.7 ppb).

3.5   Material lifespan

Comparing the molar susceptibility measurements from t_0 to t_1 showed no significant change for USPIO (p=0.32; p=0.06 at 3T; 7T), ferritin (p=0.55; p=0.11 at 3T; 7T), $CaCl_2$ at 3T (p=0.71), but there was a significant change for $CaCl_2$ at 7T (p=0.04), see table 3. Comparing the molar susceptibility measurements from t_0 to t_2 showed a significant change for USPIO at 3T (p=0.05) but not at 7T (p=0.10), and no significant change for ferritin (p=0.24; p=0.32 at 3T; 7T). Comparing the molar susceptibility measurements from t_1 to t_2 showed no significant change for $CaCO_3$ (p=0.38; p=0.12 at 3T; 7T). There was no significant linear correlation between time ($t_{month}=[0,9,24]$) and molar susceptibility (χ_{mol}) for USPIO (p=0.38; p=0.82 at 3T; 7T) and ferritin (p=0.28; p=0.70 at 3T; 7T).

	USPIO	Ferritin	$CaCl_2$	$CaCO_3$
ICC, test-retest	0.99-1.00; 0.99-1.00	0.99-1.00; 1.00-1.00	0.99-1.00; 1.00-1.00	1.00-1.00; 1.00-1.00
SEM, test-retest (ppb)	1.5 ± 2.0; 5.1 ± 2.0	1.2 ± 1.0; 2.2 ± 1.3	7.5 ± 5.5; 1.2 ± 0.6	0.8 ± 0.6; 0.9 ± 0.7
Material lifespan	≤ 24 months (p=0.05 at 3T)	> 24 months (no significant change observed)	≤ 9 months (p=0.04 at 7T)	> 15 months (no significant change observed)

Table 3: Test-retest repeatability and lifespan of χ measurements for concentrations 1 to 5 of each material group. Besides material lifespan, values given at 3T; 7T, respectively. Coefficient of reliability (ICC) reported as: ICC lower interval to ICC upper interval. Standard error of the measurement (SEM) of test-retest measurements reported as mean ± standard deviation. Material lifespan was based on the t-tests comparing baseline χ_{mol} values to χ_{mol} value at various timepoints.

3.6 Cross-field strength agreement

For both ferritin and $CaCO_3$, the χ_{3T} versus χ_{7T} relationship showed a significant bias; -0.03 ppm (p=0.00) for ferritin and -0.02 ppm (p=0.01) for $CaCO_3$. For both ferritin and $CaCO_3$, the correlation between χ_{3T} versus χ_{7T} was significant (p=0.02), and the correlation line was y=1.00x-0.03 for ferritin, and y=0.99x-0.02 for $CaCO_3$. For $CaCl_2$, the χ_{3T} versus χ_{7T} relationship showed no significant bias (p=0.50), and no significant correlation (p=0.08). The product of USPIO susceptibility and field strength, $\chi \cdot B_0$, showed a significant bias of 2.0 ppm·T (p=0.00). The correlation between $(\chi \cdot B_0)_{3T}$ and $(\chi \cdot B_0)_{7T}$ was significant (p=0.02), and the correlation line was y=1.69x+0.29.

The USPIO $(R_2^*)_{3T}$ and $(R_2^*)_{7T}$ showed significant bias of -2.6 s^{-1} (p=0.01). The correlation between $(R_2^*)_{3T}$ and $(R_2^*)_{7T}$ was significant (p=0.02), and the correlation line was y=0.93x+0.55. The ferritin $(R_2^*)_{3T}$ and $(R_2^*)_{7T}$ showed significant bias 5.7 s^{-1} (p=0.00). The correlation between $(R_2^*)_{3T}$ and $(R_2^*)_{7T}$ was significant (p=0.02), and the correlation line was y=1.87x-9.48. The $CaCl_2$ $(R_2^*)_{3T}$ and $(R_2^*)_{7T}$ showed no significant bias (p=0.32), and no significant correlation (p=1.00). The quotient of $CaCO_3$ R_2^* and field strength, R_2^*/B_0, showed a significant bias of -1.4 $s^{-1} \cdot T^{-1}$ (p=0.01). The correlation between $(R_2^*/B_0)_{3T}$ and $(R_2^*/B_0)_{7T}$ was not significant (p=0.33).

3.7 Comparison to analytical model

$\chi_{analytical}$ and χ_{MEDI+0} showed no significant bias for each material (USPIO, ferritin, $CaCl_2$, $CaCO_3$) and field strength (3T, 7T) as indicated in Table S4 (p>0.05). The correlation between $\chi_{analytical}$ and χ_{MEDI+0} was significant for each material (USPIO, ferritin, $CaCl_2$, $CaCO_3$) and field strength (3T, 7T) as indicated in Table S4 (p=0.02). At 3T, the correlation lines were y=1.00x+0.00 (USPIO), y=0.99x+0.01 (ferritin), y=0.99x+0.00 ($CaCl_2$ and $CaCO_3$). At 3T, the correlation lines were y=0.97x+0.01 (USPIO), y=0.98x+0.01 (ferritin), y=0.97x-0.02 ($CaCl_2$) and y=0.98x+0.02 ($CaCO_3$).

4 Discussion

4.1 Acquisition and processing

The field-to-susceptibility dipole inversion is formulated as a L_1-/L_2-norm regularization; requiring careful tuning of the regularization terms, trading off excessive noise propagation (streaking artifacts) versus the suppression of image features within the noise level[69]. We used MEDI-based QSM reconstruction as the information from magnitude images helped to accurately regularize susceptibility gradients reduceing error propagation during dipole inversion. Moreover, the MEDI algorithm uses a non-linear data fidelity term, which better handles dipole incompatible fields than the linear data fidelity variant[8].

We observed minor deviations in SNR and local field surrounding the clumps/agglomerates. The mean number of R_2^* map outlier voxels per vial was small (less than 3.0 % for each material group, as indicated in Table S2). The SNR within phase images are higher than SNR within magnitude images[70,71]; which implies that SNR within susceptibility maps are higher than SNR within R_2^* maps. This could explain why the minor deviations in SNR surrounding clumps/agglomerates led to larger deviations in the R_2^* maps than in the susceptibility maps. Errors associated with clumps/agglomerates were mitigated by modulating the weighting map using information from a simulated monoexponential signal (as described in section 2.5).

4.2 Test-retest repeatability and material lifespan

A key aim within the quantitative MRI community is the development of phantoms that can assist system stability[72,73]. This requires materials that are stable within minutes of repeated measurements (test-retest measurements), as well as within months (longitudinal measurements). At both field strengths, our test-retest measurements indicated that the ferritin SEM was less than the USPIO SEM; and the $CaCO_3$ SEM was less than the $CaCl_2$ SEM. Dielectric artifacts are an intrinsic limitation that should be considered when using $CaCl_2$ in phantoms. The dielectric artifacts lead to the elevated $CaCl_2$ SEM in the test-retest measurement. At both 3T and 7T, shading about the $CaCl_2$ vials was observed (see Figure 4B), and was attributed to the interaction between $CaCl_2$ conductivity and the B_1 field, which leads to signal attenuations that are unevenly distributed throughout the imaging FOV[74]. For USPIO, the high R_2^* dephasing and the high local fields contributed to the elevated USPIO SEM. For ferritin and $CaCO_3$, the SEM each were within 5 ppb at both field strengths. In terms of material lifespan, ferritin and $CaCO_3$ each were stable over the measured timeframes (9 & 24 months for ferritin; 15 months for $CaCO_3$). Ferritin and $CaCO_3$ could therefore be used within susceptibility phantoms to support 3T/7T harmonization studies, using our measurement data as a reference when constructing such a phantom.

4.3   Cross-field strength agreement

A key motivation of ultra-high-field MRI is the comparability of quantitative measurements of phase and/or susceptibility to clinical MRI[75]. Optimizing and standardizing QSM acquisition and reconstruction protocols across field strength is essential to make susceptibility a robust biomarker. Initiatives such as the German Ultrahigh Field Imaging network traveling heads study[76,77] and the United Kingdom 7T study[78] are steps in this direction. Cross-field strength agreement was conducted in the current work using Bland-Altman analysis. For both ferritin and $CaCO_3$, the χ_{3T} versus χ_{7T} relationship showed a significant bias; -0.03 ppm (p=0.00) for ferritin and -0.02 ppm (p=0.01) for $CaCO_3$. Curiously, the agreement between χ_{3T} and χ_{7T} was best at concentrations 1 and 5 (vials positioned closest to the edge). This could be due to the closer proximity to the receive elements, therefore, resulting in a higher SNR than concentrations 2 to 4 (vials positioned closest to the center). For both ferritin and $CaCO_3$, the correlation between χ_{3T} versus χ_{7T} was significant (p=0.02), and the correlation line was y=1.00x-0.03 for ferritin, and y=0.99x-0.02 for $CaCO_3$. The consistent cross-field strength bias might provide insight into future 3T/7T harmonization studies. An important property of $CaCO_3$ is that the quotient of R_2^* and field strength, R_2^*/B_0, is invariant with field strength[31]. The correlation between $(R_2^*/B_0)_{3T}$ and $(R_2^*/B_0)_{7T}$ was y=0.98x-0.81 (R^2=1.00). The findings reinforce the use of $CaCO_3$ in phantoms for susceptibility source differentiation which require diamagnetic material.

4.4   Comparison to reference values

In a USPIO phantom study by Liu et al., it is described that USPIO's magnetization ($\chi \cdot B_0$) reaches a saturation point prior to 3T, becoming invariant between 3T and 7T[4]. Liu et al. reported that the USPIO χ_{mol} at 3T; 7T is approximately 1.79; 0.74 ppm·L·mmol^{-1} [4]. These values fall within the standard errors of χ_{mol} measured in our study, 1.67 ± 0.24; 0.74 ± 0.09 ppm·L·mmol^{-1} at 3T; 7T. Liu et al. reported that the USPIO $R_2^*_{mol}$ at 3T; 7T is 87.2; 106.5 s^{-1}·L·mmol^{-1} [4]. Their data fall within the standard errors of $R_2^*_{mol}$ at 3T (84.2 ± 7.9 s^{-1}·L·mmol^{-1}), but not at 7T (79.5 ± 4.4 s^{-1}·L·mmol^{-1}). A possible explanation could be that the USPIO used in this study, Molday Ion, differs from the USPIO used by Liu et al., Ferumoxytol[4]. Previous studies have reported the ferritin susceptibility at 293 K is approximately 0.077 ppm·L·mmol^{-1} [79] and 0.080 ppm·L·mmol^{-1} [80]. These values fall within the range of ferritin χ_{mol} measured in our study, (8.13 ± 1.35)×10^{-2}; (8.13 ± 1.19)×10^{-2} ppm·L·mmol^{-1} at 3T; 7T. The susceptibility value of any paramagnetic material may deviate with temperature (by Curie's Law)[79], and the internal temperature of the phantom bore frequently deviates from the ambient temperature of the scanner room[72]. Future studies would benefit from using integrated temperature monitoring and/or control[81,82].

4.5   Recommendations

In future studies, we recommend susceptibility matching the undoped agarose to the surrounding fluid by doping each agarose mixture with the same concentration of conductivity modifier used to dope the surrounding fluid. By Equation (6), $\chi_{undoped\ agarose} = \chi_{surrounding\ fluid}$ would ensure that $\chi'_0 = 0$. We recommend using ferritin and $CaCO_3$ as paramagnetic and diamagnetic susceptibility sources for the validation of QSM imaging

Conclusion

This research performed a range of quantitative analysis of materials for QSM phantom construction: signal-to-noise ratio, χ and R_2^* maps, outliers, test-retest repeatability, cross-field strength agreement, and material lifespan. Based on the results, we recommend using ferritin and $CaCO_3$ as paramagnetic and diamagnetic susceptibility sources for the validation of QSM imaging.

Acknowledgements

We are grateful to Carl Dixon from the CAI workshop for engineering design input and for machining the phantom. We thank Martijn Cloos for providing a model of the 7T Nova Medical 32 RX-element coil. This work was published in preliminary form[65]. This research was funded by the Australian Government through the Australian Research Council (project number IC170100035). The authors acknowledge the facilities and scientific and technical assistance of National Imaging Facility, a National Collaborative Research Infrastructure Strategy (NCRIS) capability, at the University of Queensland.

Data availability statement

We facilitate the reproducibility of this study and the phantom by providing scripts (https://github.com/paddyhooper93/Thesis, commit hash 5ea6db1).

References

1. Russek SE, Stupic KF, Biller JR, Boss MA, Keenan KE, Mirowski E. Electromagnetics for Quantitative Magnetic Resonance Imaging. In: *Compendium on Electromagnetic Analysis*. World Scientific; 2020:95-147.

2. Hall MG, Cashmore MTD, McGrath C, McCann A, Tofts PS. The perfect diagnostic imaging machine and what it means for quantitative MRI reproducibility. *IPEM-Translation*. 2023;6-8:100019.

3. Deistung A, Schweser F, Wiestler B, et al. Quantitative Susceptibility Mapping Differentiates between Blood Depositions and Calcifications in Patients with Glioblastoma. Kleinschnitz C, ed. *PLoS ONE*. 2013;8(3):e57924.

4. Liu S, Brisset JC, Hu J, Haacke EM, Ge Y. Susceptibility weighted imaging and quantitative susceptibility mapping of the cerebral vasculature using ferumoxytol. *Journal of Magnetic Resonance Imaging*. 2018;47(3):621-633.

5. Ravanfar P, Loi SM, Syeda WT, et al. Systematic Review: Quantitative Susceptibility Mapping (QSM) of Brain Iron Profile in Neurodegenerative Diseases. *Frontiers in Neuroscience*. 2021;15:618435.

6. Wang Y, Spincemaille P, Liu Z, et al. Clinical quantitative susceptibility mapping (QSM): Biometal imaging and its emerging roles in patient care. *Journal of Magnetic Resonance Imaging*. 2017;46(4):951-971.

7. Neelavalli J, Cheng YCN. Magnetic Susceptibility. In: Haacke EM, Reichenbach JR, eds. *Susceptibility Weighted Imaging in MRI : Basic Concepts and Clinical Applications*. 1st ed. Hoboken, N.J.; 2011:17-31.

8. Kee Y, Liu Z, Zhou L, et al. Quantitative Susceptibility Mapping (QSM) Algorithms: Mathematical Rationale and Computational Implementations. *IEEE Transactions on Biomedical Engineering*. 2017;64(11):2531-2545.

9. Marques J, Bowtell R. Application of a Fourier-based method for rapid calculation of field inhomogeneity due to spatial variation of magnetic susceptibility. *Concepts in Magnetic Resonance Part B: Magnetic Resonance Engineering: An Educational Journal*. 2005;25(1):65-78.

10. Salomir R, de Senneville BD, Moonen CT. A fast calculation method for magnetic field inhomogeneity due to an arbitrary distribution of bulk susceptibility. *Concepts in Magnetic Resonance*. 2003;19B(1):26-34.

11. Straub S, Schneider TM, Emmerich J, et al. Suitable reference tissues for quantitative susceptibility mapping of the brain. *Magnetic Resonance in Medicine*. 2017;78(1):204-214.

12. Liu C. Susceptibility tensor imaging. *Magnetic Resonance in Medicine*. 2010;63(6):1471-1477.

13. Sullivan DC, Obuchowski NA, Kessler LG, et al. Metrology Standards for Quantitative Imaging Biomarkers. *Radiology*. August 2015.

14. Raunig DL, McShane LM, Pennello G, et al. Quantitative imaging biomarkers: A review of statistical methods for technical performance assessment. *Stat Methods Med Res*. 2015;24(1):27-67.

15. Obuchowski NA, Mozley PD, Matthews D, Buckler A, Bullen J, Jackson E. Statistical Considerations for Planning Clinical Trials with Quantitative Imaging Biomarkers. *Journal of the National Cancer Institute*. 2019;111(1):19-26.

16. Ye S, Lim JY, Huang W. Statistical considerations for repeatability and reproducibility of quantitative imaging biomarkers. *BJR|Open*. 2022;4(1):20210083.

17. Deh K, Kawaji K, Bulk M, et al. Multicenter reproducibility of quantitative susceptibility mapping in a gadolinium phantom using MEDI+0 automatic zero referencing. *Magnetic Resonance in Medicine*. 2019;81(2):1229-1236.

18. Cuña EG, Schulz H, Tuzzi E, et al. Simulated and experimental phantom data for multi-center quality assurance of quantitative susceptibility maps at 3 T, 7 T and 9.4 T. *Physica Medica: European Journal of Medical Physics*. 2023;110.

19. Antoniou A, Georgiou L, Christodoulou T, et al. MR relaxation times of agar-based tissue-mimicking phantoms. *Journal of Applied Clinical Medical Physics*. 2022;23(5):e13533.

20. Duyn JH, Schenck J. Contributions to magnetic susceptibility of brain tissue. *NMR in Biomedicine*. 2017;30(4):e3546.

21. Hagberg GE, Engelmann J, Göring E, Cuña EG, Scheffler K. Magnetic properties of iron-filled hydrogel clusters: a model system for quantitative susceptibility mapping with MRI. *Frontiers in Physics*. 2023;11. https://www.frontiersin.org/articles/10.3389/fphy.2023.1209505. Accessed August 21, 2023.

22. Shen Y, Hu J, Eteer K, et al. Detecting sub-voxel microvasculature with USPIO-enhanced susceptibility-weighted MRI at 7 T. *Magnetic Resonance Imaging*. 2020;67:90-100.

23. Han SH, Cho JH, Jung HS, et al. Robust MR assessment of cerebral blood volume and mean vessel size using SPION-enhanced ultrashort echo acquisition. *NeuroImage*. 2015;112:382-389.

24. Deh K, Zaman M, Vedvyas Y, et al. Validation of MRI quantitative susceptibility mapping of superparamagnetic iron oxide nanoparticles for hyperthermia applications in live subjects. *Scientific Reports*. 2020;10(1):1171.

25. Brown GC, Cowin GJ, Galloway GJ. A USPIO doped gel phantom for R2* relaxometry. *Magn Reson Mater Phy*. 2017;30(1):15-27. doi:10.1007/s10334-016-0576-x

26. Macpherson P, Matheson MS. Comparison of calcification of pineal, habenular commissure and choroid plexus on plain films and computed tomography. *Neuroradiology*. 1979;18(2):67-72.

27. Du Boulay GH. *Principles of X-Ray Diagnosis of the Skull*. 2nd ed. Butterworth-Heinemann; 2016.

28. Baconnier S, Lang SB. Calcite microcrystals in the pineal gland of the human brain: second harmonic generators and possible piezoelectric transducers. *IEEE Transactions on Dielectrics and Electrical Insulation*. 2004;11(2):203-209. doi:10.1109/TDEI.2004.1285888

29. Emmerich J, Bachert P, Ladd ME, Straub S. On the influence of two coexisting species of susceptibility-producing structures on the R2∗ relaxation rate. *Magnetic Resonance Imaging*. 2020;71:170-177.

30. Emmerich J, Bachert P, Ladd ME, Straub S. A novel phantom with dia- and paramagnetic substructure for quantitative susceptibility mapping and relaxometry. *Physica Medica*. 2021;88:278-284.

31. Emmerich J, Bachert P, Ladd ME, Straub S. On the separation of susceptibility sources in quantitative susceptibility mapping: Theory and phantom validation with an in vivo application to multiple sclerosis lesions of different age. *Journal of Magnetic Resonance*. 2021;330:107033. doi:10.1016/j.jmr.2021.107033

32. Hopkins JA, Wehrli FW. Magnetic susceptibility measurement of insoluble solids by NMR: Magnetic susceptibility of bone. *Magnetic Resonance in Medicine*. 1997;37(4):494-500.

33. Rumble J R. Physical Constants of Organic Compounds. In: *CRC Handbook of Chemistry and Physics*. 104th Edition (Internet Version 2023). Boca Raton, FL.: CRC Press/Taylor & Francis; 2023.

34. Weis J, Nilsson S, Ericsson A, Wikström M, Sperber GO, Hemmingsson A. Measurement of magnetic susceptibility and MR contrast agent concentration. *Magnetic Resonance Imaging*. 1994;12(6):859-864.

35. Kanazawa Y, Matsumoto Y, Harada M, Hayashi H, Matsuda T, Otsuka H. Appropriate echo time selection for quantitative susceptibility mapping. *Radiological Physics and Technology*. 2019;12(2):185-193.

36. Ianniello C, de Zwart JA, Duan Q, et al. Synthesized tissue-equivalent dielectric phantoms using salt and polyvinylpyrrolidone solutions: Synthesized Tissue-Equivalent Dielectric Properties Using a Water-Soluble Polymer. *Magnetic Resonance in Medicine*. 2018;80(1):413-419.

37. Eckstein K, Dymerska B, Bachrata B, et al. Computationally Efficient Combination of Multi-channel Phase Data From Multi-echo Acquisitions (ASPIRE). *Magnetic Resonance in Medicine*. 2018;79(6):2996-3006.

38. Dymerska B, Eckstein K, Bachrata B, et al. Phase unwrapping with a rapid opensource minimum spanning tree algorithm (ROMEO). *Magnetic Resonance in Medicine*. 2021;85(4):2294-2308.

39. Jutras JD, Wachowicz K, Gilbert G, De Zanche N. SNR efficiency of combined bipolar gradient echoes: Comparison of three-dimensional FLASH, MPRAGE, and multiparameter mapping with VFA-FLASH and MP2RAGE. *Magnetic Resonance in Medicine*. 2017;77(6):2186-2202.

40. Schmitt T, Rieger JW. Recommendations of Choice of Head Coil and Prescan Normalize Filter Depend on Region of Interest and Task. *Front Neurosci*. 2021;15:735290. doi:10.3389/fnins.2021.735290

41. Eckstein K, Dymerska B, Bachrata B, et al. Computationally Efficient Combination of Multi-channel Phase Data From Multi-echo Acquisitions (ASPIRE): Combination of Multi-Channel Phase Data from Multi-Echo Acquisitions (ASPIRE). *Magn Reson Med*. 2018;79(6):2996-3006. doi:10.1002/mrm.26963

42. Li X, Morgan PS, Ashburner J, Smith J, Rorden C. The first step for neuroimaging data analysis: DICOM to NIfTI conversion. *Journal of Neuroscience Methods*. 2016;264:47-56.

43. Dietrich O, Raya JG, Reeder SB, Reiser MF, Schoenberg SO. Measurement of signal-to-noise ratios in MR images: Influence of multichannel coils, parallel imaging, and reconstruction filters. *Journal of Magnetic Resonance Imaging*. 2007;26(2):375-385.

44. Yushkevich PA, Piven J, Hazlett HC, et al. User-guided 3D active contour segmentation of anatomical structures: Significantly improved efficiency and reliability. *NeuroImage*. 2006;31(3):1116-1128.

45. Li J, Chang S, Liu T, et al. Phase-corrected bipolar gradients in multi-echo gradient-echo sequences for quantitative susceptibility mapping. *Magn Reson Mater Phy*. 2015;28(4):347-355.

46. Jezzard P, Balaban RS. Correction for geometric distortion in echo planar images from B0 field variations. *Magnetic Resonance in Medicine*. 1995;34(1):65-73.

47. Liu T, Wisnieff C, Lou M, Chen W, Spincemaille P, Wang Y. Nonlinear formulation of the magnetic field to source relationship for robust quantitative susceptibility mapping: Robust QSM With Nonlinear Data Fidelity Constraint. *Magnetic Resonance in Medicine*. 2013;69(2):467-476.

48. Dong J, Liu T, Chen F, et al. Simultaneous Phase Unwrapping and Removal of Chemical Shift (SPURS) Using Graph Cuts: Application in Quantitative Susceptibility Mapping. *IEEE Transactions on Medical Imaging*. 2015;34(2):531-540.

49. Dymerska B, Josephs O, Callaghan M. Characterising the impact of distortion correction in high-resolution 3D GRE R_2^* mapping at 7T. In: *Proceedings of International Society for Magnetic Resonance in Medicine*. London, UK; 2022.

50. Garcia D. Robust smoothing of gridded data in one and higher dimensions with missing values. *Computational Statistics & Data Analysis*. 2010;54(4):1167-1178.

51. SPM 12 Software. Wellcome Trust Centre for Neuroimaging. https://www.fil.ion.ucl.ac.uk/spm/software/spm12/. Published 2014. Accessed October 24, 2024.

52. Kellner E, Dhital B, Kiselev VG, Reisert M. Gibbs-ringing artifact removal based on local subvoxel-shifts. *Magnetic Resonance in Medicine*. 2016;76(5):1574-1581.

53. Tournier JD, Smith R, Raffelt D, et al. MRtrix3: A fast, flexible and open software framework for medical image processing and visualisation. *NeuroImage*. 2019;202:116137.

54. Eskreis-Winkler S, Zhou D, Liu T, et al. On the influence of zero-padding on the nonlinear operations in Quantitative Susceptibility Mapping. *Magn Reson Imaging*. 2017;35:154-159.

55. Li W, Wu B, Liu C. Quantitative susceptibility mapping of human brain reflects spatial variation in tissue composition. *NeuroImage*. 2011;55(4):1645-1656.

56. Schweser F, Robinson SD, de Rochefort L, Li W, Bredies K. An illustrated comparison of processing methods for phase MRI and QSM: removal of background field contributions from sources outside the region of interest. *NMR in Biomedicine*. 2017;30(4):e3604.

57. Sandgaard AD, Shemesh N, Jespersen SN, Kiselev VG. To mask or not to mask? Investigating the impact of accounting for spatial frequency distributions and susceptibility sources on QSM quality. *Magnetic Resonance in Medicine*. 2023;90(1):353-362.

58. Langkammer C, Schweser F, Shmueli K, et al. Quantitative susceptibility mapping: Report from the 2016 reconstruction challenge. *Magnetic Resonance in Medicine*. 2018;79(3):1661-1673.

59. Bilgic B, Ye H, Wald LL, Setsompop K. Simultaneous Time Interleaved MultiSlice (STIMS) for Rapid Susceptibility Weighted acquisition. *NeuroImage*. 2017;155:577-586.

60. Liu T, Xu W, Spincemaille P, Avestimehr AS, Wang Y. Accuracy of the Morphology Enabled Dipole Inversion (MEDI) Algorithm for Quantitative Susceptibility Mapping in MRI. *IEEE Transactions on Medical Imaging*. 2012;31(3):816-824.

61. Liu Z, Spincemaille P, Yao Y, Zhang Y, Wang Y. MEDI+0: Morphology enabled dipole inversion with automatic uniform cerebrospinal fluid zero reference for quantitative susceptibility mapping: QSM With Automatic Uniform CSF Zero Reference. *Magnetic Resonance in Medicine*. 2018;79(5):2795-2803.

62. Chan KS, Marques JP. SEPIA—Susceptibility mapping pipeline tool for phase images. *NeuroImage*. 2021;227:117611.

63. Shmueli K, de Zwart JA, van Gelderen P, Li TQ, Dodd SJ, Duyn JH. Magnetic susceptibility mapping of brain tissue in vivo using MRI phase data. *Magnetic Resonance in Medicine*. 2009;62(6):1510-1522.

64. Jenkinson M, Beckmann CF, Behrens TEJ, Woolrich MW, Smith SM. FSL. *NeuroImage*. 2012;62(2):782-790.

65. Hooper P, Tourell M, O'Brien K, Jin J, Robinson SD, Barth M. Development of a Quantitative Susceptibility Mapping (QSM) phantom for validation of acquisition strategies and post-processing tools at 3 T. In: *Proceedings of the International Society for Magnetic Resonance in Medicine*. London, UK; 2022.

66. McGraw KO, Wong SP. Forming inferences about some intraclass correlation coefficients. *Psychological Methods*. 1996;1(1):30-46.

67. Koo TK, Li MY. A Guideline of Selecting and Reporting Intraclass Correlation Coefficients for Reliability Research. *Journal of Chiropractic Medicine*. 2016;15(2):155-163.

68. Bland JM, Altman DG. Agreement Between Methods of Measurement with Multiple Observations Per Individual. *Journal of Biopharmaceutical Statistics*. 2007;17(4):571-582.

69. QSM Consensus Organization Committee, Bilgic B, Costagli M, et al. Recommended implementation of quantitative susceptibility mapping for clinical research in the brain: A consensus of the ISMRM electro-magnetic tissue properties study group. *Magnetic Resonance in Medicine*. 2024:1-29.

70. Gudbjartsson H, Patz S. The Rician Distribution of Noisy MRI Data. *Magnetic Resonance in Medicine*. 1995;34(6):910-914.

71. Conturo TE, Smith GD. Signal-to-noise in phase angle reconstruction: Dynamic range extension using phase reference offsets. *Magnetic Resonance in Medicine*. 1990;15(3):420-437.

72. Stupic KF, Ainslie M, Boss MA, et al. A standard system phantom for magnetic resonance imaging. *Magnetic Resonance in Medicine*. 2021;86(3):1194-1211.

73. Keenan KE, Ainslie M, Barker AJ, et al. Quantitative magnetic resonance imaging phantoms: A review and the need for a system phantom. *Magnetic Resonance in Medicine*. 2018;79(1):48-61.

74. Tropp J. Image brightening in samples of high dielectric constant. *Journal of Magnetic Resonance*. 2004;167(1):12-24.

75. Möller HE, Bossoni L, Connor JR, et al. Iron, Myelin, and the Brain: Neuroimaging Meets Neurobiology. *Trends in Neurosciences*. 2019;42(6):384-401.

76. Voelker MN, Kraff O, Brenner D, et al. The traveling heads: multicenter brain imaging at 7 Tesla. *Magnetic Resonance Materials in Physics, Biology and Medicine*. 2016;29(3):399-415.

77. Voelker MN, Kraff O, Goerke S, et al. The traveling heads 2.0: Multicenter reproducibility of quantitative imaging methods at 7 Tesla. *NeuroImage*. 2021;232:117910.

78. Rua C, Clarke WT, Driver ID, et al. Multi-centre, multi-vendor reproducibility of 7T QSM and R2* in the human brain: Results from the UK7T study. *NeuroImage*. 2020;223:117358.

79. Schenck JF. Health and Physiological Effects of Human Exposure to Whole-Body Four-Tesla Magnetic Fields during MRI. *Annals of the New York Academy of Sciences*. 1992;649(1):285-301.

80. Xie H, Cheng YCN, Kokeny P, et al. A quantitative study of susceptibility and additional frequency shift of three common materials in MRI. *Magnetic Resonance in Medicine*. 2016;76(4):1263-1269.

1   81.   Keenan KE, Stupic KF, Russek SE, Mirowski E. MRI-visible liquid crystal thermometer. *Magnetic Resonance*
2   *in Medicine*. 2020;84(3):1552-1563.

3   82.   Kan H, Uchida Y, Arai N, et al. Decreasing iron susceptibility with temperature in quantitative susceptibility
4   mapping: A phantom study. *Magnetic Resonance Imaging*. 2020;73:55-61.

(A) Physical Isocenter

175 mm

220 mm

10 mm-diameter vials (× 20)
Imaging FOV: [190 190 100] mm
Air bubble trapping zone

(D) Vial arrangement

15 mm

10 mm

10 mm

(B) Small hole

Small hole to *remove bubbles* from imaging FOV to trapping zone.

(C) 7 T Head coil

(A) USPIO χ-map

3 T   7 T

1.4 ppm
0

0.7 ppm
0

(B) USPIO R₂*-map

3 T   7 T

75 s⁻¹
0

(C)

χ (1)

3 T
7 T

0.09 0.18 0.27 0.36 0.45 0.54
USPIO cFe (mmol/L)

(D)

(χ · B₀)₇ₜ − (χ · B₀)₃ₜ (ppm · T)

3.0 (+1.96SD)

2.0 (p=0.00)

2.0 (−1.96SD)

2   3   4   5
Mean (χ · B₀)₃ₜ & (χ · B₀)₇ₜ (ppm · T)

(E)

(χ · B₀)₇ₜ (ppm · T)

(χ · B₀)₃ₜ (ppm · T)

(F)

R₂* (s⁻¹)

3 T
7 T

0.09 0.18 0.27 0.36 0.45 0.54
USPIO cFe (mmol/L)

(G)

(R₂*)₇ₜ − (R₂*)₃ₜ (s⁻¹)

-0.08 (+1.96SD)

2.6 [p=0.01]

-5.2 (−1.96SD)

20   30   40   50   60
Mean (R₂*)₃ₜ & (R₂*)₇ₜ (s⁻¹)

(H)

(R₂*)₇ₜ (s⁻¹)

(R₂*)₃ₜ (s⁻¹)