1   **Title: Multiplexed dynamic control of temperature to probe and observe mammalian cells.**
2

3   **Authors:** William Benman[1], Pavan Iyengar[2#], Thomas Mumford[1#], Lukasz J. Bugaj[1,3,4]*
4
5   **Affiliations:**
6   [1]Department of Bioengineering, University of Pennsylvania, Philadelphia, PA, 19104, USA
7   [2]Department of Chemistry, University of Pennsylvania, Philadelphia, PA, 19104, USA
8   [3]Abramson Cancer Center, University of Pennsylvania, Philadelphia, PA, 19104, USA
9   [4]Institute of Regenerative Medicine, University of Pennsylvania, Philadelphia, PA, 19104, USA
10
11   [#]equal contribution
12   *corresponding author
13
14   **Correspondence:** bugaj@seas.upenn.edu

15
16
Abstract

Temperature is a critical parameter for biological function, yet there is a lack of approaches to
modulate the temperature of biological specimens in a dynamic and high-throughput manner.
We present the thermoPlate, a device for programmable control of temperature in each well of a
96-well plate in a manner compatible with mammalian cell culture and live cell imaging. The
thermoPlate maintains precise feedback control of temperature patterns independently in each
well, with minutes-scale heating and cooling through ΔT ~15°C. We built a computational model
that predicts thermal diffusion through sample plates to guide optimal design of heating
protocols. We used the thermoPlate in conjunction with live-cell confocal microscopy to
measure the dynamics of stress granule formation in response to time-varying temperature
stimuli, uncovering robust adaptation and biochemical memory formation that varied non-linearly
as a function of both time and temperature of heat shock. The capabilities and open-source
nature of the thermoPlate enable the study and engineering of a wide range of
thermoresponsive phenomena.

Introduction

Temperature plays important roles in the function and control of biological systems. In humans, core temperature is narrowly maintained around 37°C and, although temperature can rise during infection, the febrile response is also tightly controlled to remain below ~41°C[1]. In the biosciences, temperature can be harnessed as a non-invasive perturbation, for example with temperature-sensitive protein mutants that provide conditional knockdown of proteins[2–4]. Separately, heat-shock promoters can allow expression of custom transgenes in response to a brief increase in temperature. While such technologies can control genes in model organisms[5], they also offer spatiotemporally-precise control of engineered cells deployed within mammals because heat can effectively penetrate opaque tissues[6–8]. Exposing mammalian cells to temperatures above ~42°C induces heat shock, which can be toxic over extended periods. This response has been harnessed for cancer therapy, where thermal ablation of tumors is a common, minimally-invasive treatment strategy [9].

Despite its importance, there are relatively few experimental methods to systematically manipulate temperature. Incubators can be set to desired temperatures, but changes in temperatures are slow, challenging studies of phenomena that occur over fast (~10s of min) timescales. Incubators are also expensive, have a large footprint, and allow only one temperature to be tested at a time. Commercial devices such as the CherryTemp or VAHEAT are similarly limited in throughput. Thermocyclers can be used to specify multiple temperatures simultaneously[10], but these are not compatible with long term cell survival or simultaneous imaging. The current lack of methods for microscope-compatible temperature modulation in microwell plates limits our understanding of how biological systems respond to temperature fluctuations and hinders the engineering of thermally-controllable biological systems.

Here, we report the thermoPlate, a device for programmable heating and thermometry of samples in 96-well plates. The thermoPlate allows heating of individual wells with arbitrary temperature profiles and is compatible with both live-cell microscopy and standard cell culture incubators. We characterize thermoPlate performance, finding accurate heating across the plate to < 0.1°C of the desired set point. We also present a model that accurately predicts heat diffusion across the sample plate, allowing the user to rapidly test the feasibility of arbitrary desired heating programs *in silico*. We harness the utility of the thermoPlate to characterize the dynamics of heat stress in mammalian cells, finding that both the timing and intensity of heat contribute to the formation of cellular memories of stress.

Results

thermoPlate Design, Calibration, and Dynamics

The thermoPlate maintains a desired temperature by continuous deposition and measurement of heat in each well of a 96-well plate (**Figure 1A**). This is achieved with a custom-designed circuit board that accommodates a pair of thermistors in each of 96 positions in the format of a 96-well plate. A thermistor is a resistor whose resistance changes as a function of temperature, thus enabling its common use as a thermometer ("reader"). However, thermistors can also deposit heat when sufficient current is passed through them ("heater"). By pairing one reader and one heater thermistor in each well, the thermoPlate achieves

independent measurement and heating of all 96 wells (**Figure 1A, B**). To operate, the user
uploads the desired heating protocols to the on-board Arduino Micro. The Arduino regulates the
current of each heater through one of 12 shift registers linked to 96 transistors, and it receives
the temperature readings from each reader connected to a voltage divider through one of 6
multiplexers (**Supplemental Figure 1**). With the set point and temperature of each well, the
Arduino implements proportional-integral-derivative (PID) feedback control for each well,
adjusting the current of each heater to reach and maintain the sample at the desired set point
for every well (**Figure 1C**). For more details on the feedback implementation, see **Methods**.
The thermoPlate sits atop a 96-well culture plate to which it is mated with a 3D printed adapter
(**Figure 1D**). The thermoPlate can operate in standard cell culture incubators. When the culture
plate has a transparent bottom, this assembly is also compatible with live cell imaging using an
inverted microscope. The thermoPlate can run in a standalone format but can also be
connected to a computer to view live readouts of the current temperature of each well.
88   With all components and 3D printed parts in hand, the thermoPlate can be assembled in
89   8 hours, for a total cost of ~$300. Design files, a parts list, assembly instructions, and operation
90   protocols can be found in the **Methods** section and our online repository (link in **Methods**).
91   Accurate multiplexed temperature control requires consistent readings between each of
the 96 readers. We characterized the variability in temperature readings in a fully-assembled
thermoPlate in a cell culture incubator that was set to one of 5 temperatures between 26-45°C,
with the true temperature verified by 2 independent digital thermometers. While the average
error between thermoPlate and true temperature increased slightly at higher temperatures, we
observed a consistent pattern of variability between wells (SD ~ 0.26°C) (**Figure 1E**). From
these results, we derived adjustment factors for each well that corrected readings to within
0.1°C of the true temperature across all temperatures, with a standard deviation of 0.06°C
between wells in subsequent experiments (**Figure 1F**). Once calculated, calibration values
maintained accurate readings when tested after weeks of use (**Supplemental Figure 2**).
101   We next tested the thermoPlate's ability to control prespecified temperatures, with
temperature measured by the calibrated reader thermistors. 15 wells were set to simultaneously
maintain temperatures ranging from 28°C to 42°C (3-17°C above ambient) in 1 degree
increments (**Figure 1G**). After a rapid approach and equilibration at the designated set point,
each well maintained its temperature with high accuracy (average error of 0.04°C from set point)
and precision (average SD of 0.05°C within individual trace) over 15 hrs of operation (**Figure
1G**). Moreover, temperatures could be dynamically toggled between multiple setpoints,
returning to the desired temperature with high fidelity (at steady state: average error of 0.07°C,
average within-trace SD = 0.06°C) (**Figure 1H**).
110
111   *Characterizing speed and accuracy of thermoPlate heating*
112   We next characterized key operational features of thermoPlate heating, including its
range, accuracy, heating/cooling speed, and overshoot. We devised an experiment where the
thermoPlate would sequentially heat each well from 3°C above ambient temperature to a
prespecified higher temperature for 20 minutes, followed by cooling back to 3°C above ambient.
Note that because the thermoPlate has no active cooling, in most experiments the baseline
(lowest) experimental temperature should be set to higher than ambient to allow rapid cooling
to—and maintenance of—this desired low temperature. **Figure 2A** demonstrates the above

described experiment for heating through ΔT = 10°C. All wells experienced this heat pulse with
high consistency between wells (**Figures 2A,B**). From this type of experiment, we could
systematically quantify error (difference between the temperature and the set point after 20
minutes), rise/fall time (the time until temperature reached 99% of the target temperature), and
overshoot/undershoot (difference between the maximum/minimum and setpoint).
To quantify performance as a function of heating magnitude, we chose 11 wells
distributed across the plate and repeated the above experiment with heating of either 5, 10, 15,
or 20°C above ambient, with each well heated one after the other (**Figure 2C**). While all wells
reached their target temperature for up to ΔT = 15°C, only some wells could reach ΔT = 20°C
within 20 minutes. The rate at which a well approached ΔT = 20°C was highly dependent on its
position on the plate, with corner wells approaching 20°C more rapidly than center wells
(Supplemental Figure 3). Thus the variability in heating speed likely arises because of
differences in heat diffusion in different parts of the plate, with higher numbers of neighbor wells
acting as a larger heatsink that limits heating. Rise/fall times were 2 min for ΔT = 5°C and 4 min
for ΔT = 10°C, with increasing variability between wells with increasing ΔT (**Figure 2D**).
However, error (0.05-0.15°C) and overshoot/undershoot (0.1-0.3°C) were minimal and did not
correlate with ΔT.
Although heat diffusion limited the heating range of individual wells, it could also be
harnessed to boost the heating range within groups of wells. We tested this concept by heating
either individual wells or groups of 4 or 9 wells to ΔT = 20°C. Although, as before, individual
wells struggled to reach this level of heating, wells in a 3x3 matrix rapidly equilibrated to this
high set point (**Figure 2E**). Grouped wells achieved a faster approach to high temperatures at
the expense of small increases in error (0.15-0.2°C) and overshoot (~1°C). In sum, the
thermoPlate can reliably and rapidly achieve heating of up to ΔT = 15°C in individual wells, and
can also be programmed to achieve ΔT >20°C with appropriate arrangement of the heated
wells.
145
Modeling heat diffusion in the thermoPlate
Heat diffusion constrains the heating protocols that can be run simultaneously because it
places limits on the temperature differential between two neighboring wells. For example, a well
heated to 15°C above ambient will diffusively heat its neighboring wells by ~10°C, placing a
lower limit on the possible temperatures of those neighbors. As the number of wells in use
increases, it becomes increasingly challenging to predict whether a desired protocol is
achievable, particularly when dynamic heating protocols are desired.
We thus created a model of heat diffusion in 96-well plates to allow users to rapidly
assess the feasibility of particular heating programs. We implemented an explicit finite
difference approximation of a modified heat equation. This model simulates dynamic heat
stimuli and calculates the rate of heat spread through each well of the plate over time. Unlike
the classical heat equation, our model has 6 (vs 1) parameters that describe the rate of heat
spread. These additional parameters account for observed differences in heat spread across
different contexts, for example in edge wells vs center wells (**Figure 3A**). These parameters
also accounted for small asymmetries around heated wells (left vs right, top vs bottom) likely
due to asymmetric positioning of the electrical components.

To fit the model, we generated eight sets of training data where different sets of wells (1-
12 wells) underwent dynamic heating/cooling programs, with temperature changes of 1-15°C
over durations of up to 90 min (**Figure 3B**). These training experiments were designed with
increasing levels of complexity, from simple heating of one well (Dataset 1) to more complex,
dynamic heating of 12 wells arranged both individually and in groups (Dataset
8)(**Supplementary Figure 4**). The model was parameterized for each training set by
minimizing the root mean squared error (RMSE) of the model relative to the training data. These
8 parameter sets were then averaged to create a generic set of parameters that simulate
thermoPlate behavior. We validated the model using Leave-One-Out-Cross-Validation, wherein
fitted parameters from 7 training sets were averaged and used to predict the 8th dataset,
resulting in a mean average error of 0.23°C across all wells and times. Comparing the
predictions of the full model vs the data across all data sets, the average error over an
experimental run was 0.22°C (**Figure 3C-E**), with error falling within 1°C for > 99% of timepoints
and within 0.5°C for > 75% of timepoints (**Figure 3E**), showing close agreement between model
and data. Transient spikes in error (1-2°C) could on occasion be observed but only near wells
that experienced a sudden large temperature change (**Figure 3D**).
The predictive power of this model allows users to rapidly test whether particular heating
patterns and spatial arrangements are permissible. For this purpose, we created an online
interface where users can upload the desired heating protocol and easily simulate and visualize
heat generation and spread (**Figure 3F**). The visualization also highlights impermissible
temperature set points, allowing the user to modify their experimental design accordingly.
183
Characterizing dynamics and memory of heat stress in mammalian cells.
The thermoPlate was designed to probe biological systems with dynamic temperature
inputs in a manner compatible with live-cell imaging. To highlight this capability, we examined
the dynamics of stress granule formation in response to heat shock in mammalian cells. Stress
granules (SGs) are condensates of protein and RNA that form in the cytosol in response to
stressors including heat, hypoxia, and oxidation[11]. While the composition and function of SGs in
response to heat stress have been studied extensively[12–17], their dynamics are relatively less
explored.
We first confirmed that the thermoPlate was fully compatible with mammalian cell
culture, with no observed impact on proliferation or survival of HEK 293T cells after even 48 hr
of direct contact with the thermoPlate (**Supplemental Figure 5**). We next characterized media
evaporation during thermoPlate use and found that, while significant evaporation could be
observed in edge wells under prolonged heating of >10°C, evaporation could be prevented by
adding a thin layer of paraffin oil over each well (**Supplemental Figure 6**). To observe stress
granules, we stably expressed a fluorescent fusion of the SG scaffold G3BP1-mCherry in HeLa
cells, and we imaged G3BP1 condensation as a function of time and temperature.
We first asked at what temperature SGs could be observed by exposing cells to constant
heat between 41 to 47°C (**Figure 4A**). We found a highly non-linear relationship between
temperature and stress-granule formation. Although no SGs were observed at 41°C, SG rapidly
formed at 42°C and increased in magnitude at 43°C and 44°C. Surprisingly, maximal SG
strength decreased with progressively further increases in temperature (**Figure 4B-D**).
Although cells remained morphologically unperturbed with SGs at 42°C and 43°C over at least 8

hrs, cells heated with 44°C retracted their cytoplasms and became immobile, with higher temperatures leading to faster onset of this phenotype (**Figure 4C, Movie S1).** In all heating conditions**,** SGs formed rapidly within the first 5 min. Notably, SG formation was transient across all conditions despite constant heating, and in most cells G3BP1 returned to a diffuse state within several hours of the initial heat shock (**Figure 4B**). SG adaptation is consistent with known negative feedback of SGs, including transcription of heat shock proteins (HSPs)[18–22] and of suppressors of the integrated stress response including GADD34[23–25].

We next asked if SG adaptation required constant heat stress or alternatively could form in response to short heat pulses. We examined SG formation in response to periodic 1-hr pulses of heat stress (43.5°C), a pulse duration shorter than the time scale of SG adaptation (**Figure 4B**). Although the SGs formed and dissolved in sync with the first two heating and cooling pulses, SGs largely did not form during the 3rd pulse, suggesting that cells continued adapting to heat stress despite heat not being applied for the entire duration of adaptation (**Figure 4E,F, Movie S2**). Indeed, even though the second heat pulse yielded strong SG formation, SG magnitude began to decrease during this second window, indicating that the first heat pulse influenced SG behavior during the second heat pulse (**Figure 4E**). Collectively, these results show that cells store memories of prior heat exposure that shape future responses to repeated stress.

We next characterized the timescales of cellular memory formation. We programmed the thermoPlate to apply two pulses of heat stress (43.5°C) separated by varying periods of relaxation at 37°C (**Figure 4G**), and we quantified SG formation during the second pulse. SG formation depended on the interval between pulses in a biphasic manner. Although SGs were prominent during the second pulse after a 1 or 2 hr interval, they were not observed with intervals of 3 and 4.5 hrs. At longer intervals, SG formation began to reappear on the second pulse. With a 9 hr interval, SG formation during the 2nd pulse closely resembled that of the first pulse (**Figure 4G,H**). Thus, cellular memory of prior stress was maximal for 3-5 hrs after which the memory began to fade and cells again responded to heat shock as on the first exposure.

Finally, we wondered whether memories of prior heat shock depended on the duration or the temperature of heat shock, or both. We designed a "pump-probe experiment" where we delivered a "pump" of heat with variable duration (15, 30, 60 min) and temperature (43-44.5°C) and we later probed SG formation with a 1 hr of heat shock (43.5°C) after a fixed 3 hr interval (**Figure 4I**). Indeed, memory (suppression of SG formation on the second pulse) depended on both time and temperature of the pump pulse. Suppression of SGs was stronger with increasing temperatures, and the magnitude of suppression increased with increasing pump duration. Notably, even a short 15 min pump was sufficient for measurable SG suppression at higher temperatures (**Figure 4I,J**). Additionally, the rate of dissolution of SGs after removal of the pump pulse depended on the integrated heat stress, with slower dissociation in response to longer and more intense heating (**Figure 4I**), indicating a second form of stress memory. This effect could only be observed at or above 44°C heating, though at these temperatures the effect could be observed even after a brief 15 min pump duration (**Figure 4K**). In sum, our results indicate multiple mechanisms by which cells can store memories of prior heat stress and demonstrate the utility of the thermoPlate to explore these stress responses efficiently with high precision, throughput, and live-cell readouts.

Discussion

250   The thermoPlate introduces the ability to modulate the temperature of biological samples in a multiplexed and small format that is compatible with both long term cell culture and live-cell microscopy. After calibration, the thermoPlate achieved its desired set point temperatures rapidly (minutes) and accurately (<0.2°C error) with minimal variation between wells (between-well SD = 0.06°C), allowing profiling of the response to thermal dynamics with high resolution in temperature and in time. The device is open source and can be assembled by non-specialists in several hours by following the assembly protocols associated with this manuscript, for a total cost of ~$300 in parts. The design files can also be modified to adapt the device for specific applications as needed.

          Heat diffusion is a central challenge in designing multi-well thermal control of microwell plates. Although diffusion constrains the types of possible heating protocols, these constraints need not limit throughput. For example, if a sustained temperature sweep is desired, a temperature gradient can be specified across the plate, and all 96-wells can be used simultaneously. For more complex and dynamic patterns, samples must be placed more strategically to allow the samples to heat and cool appropriately. While this is straightforward for a few samples, it becomes more challenging with each additional heated well. For this reason, we developed a model of thermal diffusion that accurately predicts heating dynamics of each well for arbitrary protocols. The user can rapidly simulate a desired protocol *in silico* and observe whether the intended thermal patterns are permitted within the constraints of thermal diffusion. If they are not, the user adjusts the inputs, simulates, and iterates until a suitable arrangement is reached. We also introduce a web-based GUI for easy access to this model.

          Although thermal diffusion presents a constraint, we also harness it to passively cool the samples. The simplicity of passive cooling is balanced by two additional constraints: 1) potentially slow cooling kinetics and 2) inability to maintain certain wells at ambient temperatures while heating other wells, since over time all wells will increase temperature by a few degrees as heat spreads throughout the plate faster than it diffuses out to the environment. To address both concerns, we maintained the environment at least ~3°C cooler than the lowest experimental temperature. This ensures that low experimental temperatures can be maintained and also results in faster cooling to this low level. With this approach, heating and cooling occur on comparable ~minutes timescales (**Figure 2D**).

          We harnessed the capabilities of the thermoPlate to explore the dynamics of the heat-shock response by observing stress granule (SG) formation in response to time-varying temperatures. We found a rich set of dynamic features of SG formation that respond to both the time and temperature of heat stress in a non-linear manner. Rather than a binary response, SGs formed in a measurably different manner with each additional degree of heating above 41°C, with a maximal peak amplitude at 44°C. Notably, SGs showed adaptation to heat stress, such that G3BP1 puncta spontaneously dissolved despite persistent heating. Such adaptation of SGs is consistent with known mechanisms of negative feedback [19,22–25]. Nevertheless, to our knowledge the adaptation of SGs to chronic heat stress has not been previously reported, demonstrating the utility of simultaneously modulating temperature and outcomes in cellular systems.

          Heat stress resulted in the formation of at least two types of biochemical memories. One hour pulses of stress were sufficient to suppress SG formation upon subsequent stress, with

maximal suppression observed 3-6 hrs after the initial stress pulse. These kinetics are in line with a recent finding that 1 hr of heat stress induced expression of GADD34 after 3-5 hrs, which suppressed ISR stimulation in response to a second heat pulse[23]. However, this prior work did not observe the status of SG formation during this second pulse. Notably, our measurements also revealed SG suppression after only 15 and 30 min of initial heat shock. With these shorter durations, the strength of SG suppression increased with increasing temperature of prior heat shock. A second form of biochemical memory appeared in the kinetics of SG dissolution after stress removal. This memory was observed immediately after removal of heat stress but only above 44°C, consistent with fast post-translational mechanisms like protein clustering distinct from SG formation[26]. While further work will be required to identify the specific mechanisms that underlie our observations, our results highlight the capabilities for thermal modulation with the thermoPlate, as well as the discovery potential of instrumentation for easy, precise, and dynamic control of temperature under a microscope.

We anticipate that the thermoPlate will find use across the biosciences wherever precise thermal control is desired. One emerging area is in engineering thermogenetic control of cells, where temperature acts as an inducer for genetic or protein activity[6–8]. The thermoPlate would allow rapid characterization of these systems over a range of timescales. The thermoPlate will also have impact in understanding the *in vitro* and *in cellulo* behavior of thermally-switchable proteins, which have been identified with increasing frequency over the past several years[27–31]. Although designed for control of 96-well plates, the thermoPlate could be readily adapted to larger well formats where multiple reader/heater pairs would regulate a single well. Because the thermoPlate occupies only the top side of a well plate, it is also compatible with other instruments that stimulate or measure samples from the bottom of the plate[32–34], opening the door for integration of additional inputs and outputs for thermoPlate experiments.

Acknowledgements:
The authors thank Dr. James Shorter for helpful discussions. This work was supported by the National Science Foundation (CAREER CBET 2145699 to L.J.B., GRFP to W.B.), the National Institutes of Health (R35GM138211 to L.J.B), and the Penn Center for Precision Engineering for Health (CPE4H). Cell sorting was performed on a BD FACSAria Fusion that was obtained through an NIH S10 grant (S10OD026986).

Author Contributions:
W.B. and L.J.B conceived study. W.B. P.I., L.J.B. designed hardware structure and component layout. W.B. and P.I. constructed hardware, performed experiments, and analyzed data. T.M. and P.I. developed the model of thermal diffusion. L.J.B. supervised the work. W.B., T.M., and L.J.B. drafted manuscript and made figures. All authors contributed to editing.

METHODS

PID Implementation
PID (Proportional, Integral, Derivative) control was implemented using a discrete form of the
standard PID algorithm. PID constants were generated empirically by testing several
combinations. In brief, the Arduino collects the temperature of every well every ~2 seconds.
These values are used to calculate the current error (proportional), change in error from the
previous reading per unit time (derivative), and integral (time weighted sum of all past error).
The algorithm applies the PID constants to each error value and calculates a new duty ratio to
apply to the corresponding heater. For more detailed information on programming architecture,
see the **Code** section of the **Repository**.

thermoPlate Calibration
Calibration was performed by placing a thermoPlate on a 96 well plate containing PBS. This
assembly was equilibrated in a standard mammalian cell culture incubator at ~25°C. Two digital
thermometers (RC-4 Elitech Digital Temperature Data Logger) were placed inside two 15mL
conical tubes filled with water and sealed with parafilm (to prevent evaporation). Both conical
tube/thermometer were placed next to the thermoPlate, and a plastic box was placed over the
thermoPlate and thermometers to avoid variation due to airflow. The average of the two digital
thermometers was recorded as the true temperature. 5 temperature readings were recorded
from each well. The average of those 5 readings was assigned as the final temperature reading
of that well. The incubator temperature was then increased to 30°C and allowed to equilibrate
for 2 hours. Readings from the thermoPlate and the digital thermometers were recorded, and
this procedure was repeated in 5°C steps until 45°C. For each well, the temperature readings of
the thermoPlate vs digital thermometers was then plotted, and the slope and y-intercept of the
linear fit were extracted. These values were used to calibrate each raw temperature reading
during thermoPlate operation. See **Manual** for more details on thermoPlate operation.

thermoPlate Modeling
A model of heat transfer within a thermoPlate-heated 96-well plate was made in MATLAB to
simulate the temperature of each well given a desired heating input. Well temperatures were
simulated through time. At each timestep, the instantaneous heat derivative of each well was
calculated and multiplied by the simulation timestep to obtain the well's temperature at the next
timestep. The model uses 8 parameters. H_p describes heat transfer between neighboring
unheated wells. Around heated wells, H_{left}, H_{right}, H_{top}, and H_{bottom} describe heat transfer
between the heated well and its neighbors to the left, right, top, and bottom, respectively. This
was done to capture the consistent asymmetries seen in temperature around wells that are
actively heated. H_e describes heat transfer from an edge well to the environment, *loss*
describes a consistent loss of heat from each well to the environment. Finally, *heat* is a constant
temperature increment that is added to each heated well during a simulation step where the well
is below its set temperature. When a heated well is at or above its setpoint, this *heat* is not
added for that step. Example calculations of the change in temperature of wells over a single
time step are shown in **Figure S7**. Parameters were obtained by fitting the model to eight real
datasets and averaging resultant 8 parameter sets to obtain generalizable parameters. Fitting
was done by minimizing an error function, defined as the difference between the simulation and

the real data summed across every well and timepoint. Minimization was done using the
MATLAB function fminsearch.

Cell Culture
HeLa cells were maintained in 10% fetal bovine serum (FBS) and 1% penicillin/streptomycin
(P/S) in DMEM. Cell lines were not verified after purchase. Cells were not cultured in proximity
to commonly misidentified cell lines.

Plasmid design and assembly
Constructs for stable transduction and transient transfection were cloned into the pHR lentiviral
backbone with an SFFV promoter driving the gene of interest. The pHR backbone was
linearized using MluI and NotI restriction sites. G3BP1 (Ophir Shalem Lab) and mCherry inserts
were generated via PCR and cloned into the linearized viral backbone as a single fusion protein
using HiFi cloning mix. H2B-iRFP nuclear marker was (Addgene Plasmid #90237).

Plasmid transfection.
HEK 293T cells were transfected using the calcium phosphate method, as follows: Per 1 mL of
media of the cell culture to be transfected, 50 µL of 2x HeBS buffer, 1 µg of each DNA
construct, and H_2O up to 94 µL was mixed. 6 µL of 2.5mM $CaCl_2$ was added after mixing of
initial components, incubated for 1:45 minutes at room temperature, and added directly to cell
culture.

Lentiviral packaging and cell line generation
Lentivirus was packaged by cotransfecting the pHR transfer vector, pCMV-dR8.91 (Addgene,
catalog number 12263), and pMD2.G (Addgene, catalog number 12259) into Lenti-X HEK293T.
Briefly, cells were seeded one day prior to transfection at a concentration of 350,000 cells/mL in
a 6-well plate. Plasmids were transfected using the calcium phosphate method. Media was
removed one day post-transfection and replaced with fresh media. Two days post-transfection,
media containing virus was collected and centrifuged at 800 x g for 3 minutes. The supernatant
was passed through a 0.45 µm filter. 500 µL of filtered virus solution was added to 700,000
HeLa cells seeded in a 6-well plate. Cells were expanded over multiple passages, and
successfully transduced cells were enriched through fluorescence activated cell sorting (BD
FACS Aria Fusion).

Preparation of cells for plate-based experiments
All experiments were carried out in Falcon™ 353219 Black 96 well plates. Briefly, wells were
coated with 50 uL of MilliporeSigma™ Chemicon™ Human Plasma Fibronectin Purified Protein
fibronectin solution diluted 100x in PBS and were incubated at 37 °C for 30 min. HeLa cells
were seeded in wells at a density of 60,000 cells/well in 100 µL and were spun down at 20 x g
for 1 minute to ensure an even distribution of cells across the wells.

Imaging
Live-cell imaging. Live-cell imaging was performed using a Nikon Ti2-E microscope equipped
with a Yokagawa CSU-W1 spinning disk, 405/488/561/640 nm laser lines, an sCMOS camera
(Photometrics), a motorized stage, and an environmental chamber (Okolabs). HeLa cells

expressing the construct of interest were imaged with a 20X air objective at variable
temperatures and 5% CO_2.

Imaging quantification. Images were processed using Cell Profiler [35]. Cells were segmented
using the H2B-iRFP imaging channel, and cytoplasm was identified using a 20 pixel ring around
the nucleus. Nuclear and cytoplasmic fluorescence values were then exported and analyzed
using R (https://cran.r-project.org/) and R-Studio (https://rstudio.com/). Data was processed
and visualized using the tidyR [46] and ggplot2 [47] packages. Stress granules were quantified by
averaging the maximum pixel intensity of each cell within a frame and then averaging those
values across multiple replicate frames.

Figures

Figure 1. Design and use of the thermoPlate for simultaneous, independent temperature
control in 96 well plates. a) Image of a fully assembled thermoPlate device. b) Enlarged image
of a single heater/reader pair, which regulates the temperature of a single well of a 96 well plate.
c) An Arduino takes readings from a reader and dynamically adjusts the duty ratio of the heater
using PID feedback control. e) Raw readings taken from the reader of each well when placed in
a cell culture incubator set to various temperatures before and after calibration. Color and
dashed lines represent the incubator temperature. See Methods for details on calibration. f)
Traces showing 16 wells that were set to 16 different temperatures in 1°C increments (28-43°C)
with an ambient temperature of 25°C for 15 hrs. h) Dynamic temperature control in multiple
wells over multiple temperatures. 7 wells were heated to different temperatures for repeated
cycles of 30 minutes of heating followed by 10 minutes of cooling to 28°C, with an ambient
temperature of 25°C.

438
Figure 2. Characterization of thermoPlate heating and cooling a) Temperature profile of
each well heated with a 20 min pulse of 10°C. Wells were heated sequentially. b) The mean of
all 96 traces from (a) labeled with key metrics of thermoPlate performance. Heating error =
difference between the measured and set point temperature at 20 minutes. Rise time = duration
at which temperature reached 0.5°C of the desired temperature. Overshoot = difference
between maximum measured and set point temperature during heating. Cooling error, fall time,
and undershoot represent the same parameters during the cooling phase. Error ribbons = 1 SD.
s.s. = steady state. c) Heating each of 12 wells through 5, 10, 15, or 20°C for 20 minutes and
then cooling to 28°C (ambient = 25°C). d) Mean of each performance metric from (c). Error bars
represent = SD. e) Heating amplitude and speed can be increased by heating groups of wells.
Wells were programmed to heat by ΔT = 20°C as either single wells or a group of 9 wells. f)
Performance metrics of heating as a function of well grouping. Wells were heated by ΔT = 10°C.
Data represents mean +/- SD from 3-9 wells.

Figure 3. Mathematical modeling predicts the temperature of all wells in a sample plate.
a) Diagram of parameters used to model heat diffusion in a 96-well plate that is heated by a thermoPlate. b) Strategy for parameter fitting. 8 parameter sets were fit to 8 separate experiments that spanned a range of wells, temperature, and heating dynamics. Parameters were then averaged across the parameter sets. c) Comparisons of set points, model predictions, actual temperatures, and error between predictions and model for the 60 min time points of two heating datasets. d) Traces of error (predicted-measured) within each well over the entire duration of two datasets. Each trace represents the error within a single well. e) Summaries of error for each well and time point of each data set. f) A Shiny App GUI allows users to upload heating profiles, review the profiles visually, and simulate heating and heat diffusion during a run. The simulation shows the predicted temperature of all wells and indicates whether any prespecified temperature profiles are unachievable due to thermal diffusion from neighboring wells (red circles in right panel). The user can then adjust and simulate the adjusted protocol until no errors are found.

Figure 4. The thermoPlate allows visualization of mammalian heat shock response with high resolution in temperature and time. a) Heat induced stress granule (SG) formation was observed by measuring condensation of G3BP1-mCh in response to time varying temperature changes in HeLa cells. b) SG formation in response to constant heating. Each trace represents mean SG formation from two wells. c) Representative images of experiment in (b). d) Maxima of SG formation across all temperatures tested. e) SG response dynamics in response to pulsatile stimuli. Red bars indicate heating intervals (1 hr at 43.5°C), which were separated by 1 hr intervals at 37°C. f) Interrogating temporal properties of heat shock memories that suppress stress granule formation. Cells were heated for 1 hr intervals (43.5°C) separated by variable intervals at 37°C, and SG formation on the second pulse was measured. h) The degree of

suppression of SGs on the second pulse from (g) is quantified as a function of the interval
between heat pulses. i) A "pump-probe" experiment to interrogate the effects of temperature
and duration of heat shock on SG memory. The temperature and duration of the "pump" heat
pulse were varied, and the magnitude of SG formation on a second heat pulse (1 hr, 43.5°C)
was measured. j) Quantification of SG memory (suppression during the probe pulse) as a
function of pump temperature and duration from experiment in (i). k) Quantification of kinetics of
SG dissolution after the pump pulse. Dissociation time was calculated as the time at which SG
magnitude dropped 90% from its peak value before adaptation. Data represents mean +/- SD of
four replicates.

Supplementary Figures

Figure S1. Wiring diagram of a single Heater/Reader thermistor pair. Each Heater/Reader
pair is controlled via an Arduino microcontroller which coordinates the following actions: 1) The
temperature of a well is requested by the Arduino from the multiplexer. 2) The multiplexer
retrieves and returns the temperature of the well via the Reader. The Arduino calculates the
necessary heating intensity (in the form of a duty cycle) via a PID algorithm. 3) The Arduino
requests that the shift register accordingly adjusts the on/off state of the Heater. 4) The shift
register will adjust the Heater accordingly. This process is repeated 96 times for each well of a
96-well plate.

Figure S2. Calibration stability after 1 month of use. a) The same procedure performed in Figure 1e (raising the ambient temperature of an incubator in 5°C increments and recording thermoPlate readings at those temperatures) was repeated on the same thermoPlate 1 month after initial calibration, with temperature readings adjusted using the same calibration values produced in Figure 1e. Average reading error remained unchanged compared to prior calibration (<0.1°C) while average SD between wells increased marginally to .086°C, indicating calibration values are stable over time. Color and dashed lines represent the incubator temperature. See Methods for details on calibration.

Figure S3. Edge wells heat faster and reach higher steady states. a) Traces are taken from
Figure 2c, showing kinetics and steady state temperature from wells undergoing constant
heating. Color represents the position on the plate (center wells labeled in red and edge wells
labeled in blue) showing that edge wells are able to achieve faster and higher steady state
heating than center wells. b) Diagram of the position on the plate of each trace found in (a).

Figure S4. Visualization of heating profiles used to fit the thermoPlate model. a) Each
diagram corresponds to one dataset used for model fitting with color representing the maximum
set temperature of each well throughout the experiment. Traces represent the times at which
specific wells either raise or lower their setpoint, normalized between max and min for that well.

Figure S5. Cell growth quantification with and without exposure to the thermoPlate. a)
Every well of a 96 well plate was seeded with HeLa cells and kept for two days in a standard
cell culture incubator, with half of the wells exposed to thermoPlate heater/readers. Cells were
then fixed, stained with DAPI, imaged, and nuclei were counted to quantify population size in
each well. No significant difference in cell numbers was observed between the two conditions,

indicating that cell growth and survival are not significantly altered by thermoPlate exposure.
Each bar represents the mean +/- SD of N = 48 wells.

523
Figure S6. Evaporation during thermoPlate usage. a) A 96 well plate filled with 150uL of
PBS was placed in a cell culture incubator and wells were set to heat by either 5 or 10°C for 24
hours. The volume of PBS remaining in the well was measured via a micropipette. Color
represents the percent evaporation in each well. b) The same procedure as (a) was repeated
heating wells by either 15 or 20°C. Major evaporation (>50%) was limited to wells set to 20°C
above ambient. c) The same procedure as (b) was performed with a thin layer of 50uL of
paraffin oil coating each well to prevent evaporation. Evaporation in all wells was <10%
regardless of heating demonstrating that paraffin oil can be used to extend the duration of long
term heating.

Figure S7. Diagram of all parameters used for model fitting. To fit heat spread across a 96 well plate during thermoPlate operation, wells were categorized into 1 of 3 categories: a) a central well that transfers heat only to neighbor wells and vertically to the environment, b) edge wells that transfer heat to neighbors, vertically to the environment, and horizontally to the environment, and c) heated wells that can transfer heat in any of the ways shown in (a, b) but also receive heat deposition from a thermoPlate heater. For all scenarios, the temperature difference between the well of interest and its neighbors/environment are weighted by fitted parameters to calculate heat transfer.

Supplementary Movie Captions

Movie S1. Stress granule formation during constant heat stress. HeLa cells expressing G3BP1-mCherry were exposed to various temperatures (41°C-47°C) and imaged for ~8hrs. Stress granules form in response to 42°C but are transient in all cases, decreasing over time after initial activation. Time is hh:mm. Scale bar = 30 µM

Movie S2. Stress granule adaptation during chronic heat stress. HeLa cells expressing G3BP1-mCherry were exposed to continuous rounds of 1 hr at 43.5°C followed by 1hr at 37°C. Stress granules form in response to the first two pulses of heat but are significantly attenuated upon the third pulse, indicating a memory of previous stress. Time is hh:mm. Scale bar = 30 µM.

References

1. Mackowiak, P. A. & Boulant, J. A. Fever's glass ceiling. *Clin. Infect. Dis.* **22**, 525–536
(1996).
2. Horowitz, N. H. Biochemical Genetics of Neurospora. in *Advances in Genetics* (ed.
Demerec, M.) vol. 3 33–71 (Academic Press, 1950).
3. Talavera, A. & Basilico, C. Temperature sensitive mutants of BHK cells affected in cell
cycle progression. *J. Cell. Physiol.* **92**, 425–436 (1977).
4. Varadarajan, R., Nagarajaram, H. A. & Ramakrishnan, C. A procedure for the prediction of
temperature-sensitive mutants of a globular protein based solely on the amino acid
sequence. *Proc. Natl. Acad. Sci. U. S. A.* **93**, 13908–13913 (1996).
5. Halloran, M. C. *et al.* Laser-induced gene expression in specific cells of transgenic
zebrafish. *Development* **127**, 1953–1960 (2000).
6. Abedi, M. H., Lee, J., Piraner, D. I. & Shapiro, M. G. Thermal Control of Engineered T-cells.
ACS Synth. Biol. **9**, 1941–1950 (2020).
7. Miller, I. C. *et al.* Enhanced intratumoural activity of CAR T cells engineered to produce
immunomodulators under photothermal control. *Nat Biomed Eng* **5**, 1348–1359 (2021).
8. Wu, Y. *et al.* Control of the activity of CAR-T cells within tumours via focused ultrasound.
Nat Biomed Eng **5**, 1336–1347 (2021).
9. Chu, K. F. & Dupuy, D. E. Thermal ablation of tumours: biological mechanisms and
advances in therapy. *Nat. Rev. Cancer* **14**, 199–208 (2014).
10. Piraner, D. I., Wu, Y. & Shapiro, M. G. Modular Thermal Control of Protein Dimerization.
ACS Synth. Biol. **8**, 2256–2262 (2019).
11. Protter, D. S. W. & Parker, R. Principles and Properties of Stress Granules. *Trends Cell*
Biol. **26**, 668–679 (2016).
12. Khong, A. *et al.* The Stress Granule Transcriptome Reveals Principles of mRNA
Accumulation in Stress Granules. *Mol. Cell* **68**, 808–820.e5 (2017).
13. Jain, S. *et al.* ATPase-Modulated Stress Granules Contain a Diverse Proteome and
Substructure. *Cell* **164**, 487–498 (2016).
14. Nunes, C. *et al.* MSGP: the first database of the protein components of the mammalian
stress granules. *Database* **2019**, (2019).
15. Markmiller, S. *et al.* Context-Dependent and Disease-Specific Diversity in Protein
Interactions within Stress Granules. *Cell* **172**, 590–604.e13 (2018).
16. Youn, J.-Y. *et al.* High-Density Proximity Mapping Reveals the Subcellular Organization of
mRNA-Associated Granules and Bodies. *Mol. Cell* **69**, 517–532.e11 (2018).
17. Qin, W. *et al.* Dynamic mapping of proteome trafficking within and between living cells by
TransitID. *Cell* **186**, 3307–3324.e30 (2023).
18. Solís, E. J. *et al.* Defining the Essential Function of Yeast Hsf1 Reveals a Compact
Transcriptional Program for Maintaining Eukaryotic Proteostasis. *Mol. Cell* **63**, 60–71
(2016).
19. Yoo, H., Bard, J. A. M., Pilipenko, E. V. & Drummond, D. A. Chaperones directly and
efficiently disperse stress-triggered biomolecular condensates. *Mol. Cell* **82**, 741–755.e11
(2022).
20. McMillan, D. R., Xiao, X., Shao, L., Graves, K. & Benjamin, I. J. Targeted Disruption of Heat

Shock Transcription Factor 1 Abolishes Thermotolerance and Protection against Heat-
inducible Apoptosis*. *J. Biol. Chem.* **273**, 7523–7528 (1998).

21. Pirkkala, L., Alastalo, T. P., Zuo, X., Benjamin, I. J. & Sistonen, L. Disruption of heat shock
factor 1 reveals an essential role in the ubiquitin proteolytic pathway. *Mol. Cell. Biol.* **20**,
2670–2675 (2000).

22. Krakowiak, J. *et al.* Hsf1 and Hsp70 constitute a two-component feedback loop that
regulates the yeast heat shock response. *Elife* **7**, (2018).

23. Klein, P. *et al.* Temporal control of the integrated stress response by a stochastic molecular
switch. *Sci Adv* **8**, eabk2022 (2022).

24. Batjargal, T. *et al.* Optogenetic control of the integrated stress response reveals
proportional encoding and the stress memory landscape. *Cell Syst* **14**, 551–562.e5 (2023).

25. Kojima, E. *et al.* The function of GADD34 is a recovery from a shutoff of protein synthesis
induced by ER stress: elucidation by GADD34-deficient mice. *FASEB J.* **17**, 1573–1575
(2003).

26. Riback, J. A. *et al.* Stress-Triggered Phase Separation Is an Adaptive, Evolutionarily Tuned
Response. *Cell* **168**, 1028–1040.e19 (2017).

27. Dietler, J. *et al.* A Light-Oxygen-Voltage Receptor Integrates Light and Temperature. *J. Mol.*
Biol. **433**, 167107 (2021).

28. Chen, D. *et al.* Integration of light and temperature sensing by liquid-liquid phase separation
of phytochrome B. *Mol. Cell* **82**, 3015–3029.e6 (2022).

29. Benman, W. *et al.* Temperature-responsive optogenetic probes of cell signaling. *Nat.*
Chem. Biol. **18**, 152–160 (2022).

30. Huang, Z., Benman, W., Dong, L. & Bugaj, L. J. Rapid optogenetic clustering in the
cytoplasm with BcLOVclust. *J. Mol. Biol.* 168452 (2024).

31. Almblad, H. *et al.* Bacterial cyclic diguanylate signaling networks sense temperature. *Nat.*
Commun. **12**, 1986 (2021).

32. Bugaj, L. J. & Lim, W. A. High-throughput multicolor optogenetics in microwell plates. *Nat.*
Protoc. **14**, 2205–2228 (2019).

33. Repina, N. A., McClave, T., Bao, X., Kane, R. S. & Schaffer, D. V. Engineered illumination
devices for optogenetic control of cellular signaling dynamics. *bioRxiv* 675892 (2019)
doi:10.1101/675892.

34. Datta, S. *et al.* High-throughput feedback-enabled optogenetic stimulation and
spectroscopy in microwell plates. *bioRxiv* 2022.07.13.499906 (2022)
doi:10.1101/2022.07.13.499906.

35. Lamprecht, M. R., Sabatini, D. M. & Carpenter, A. E. CellProfiler: free, versatile software for
automated biological image analysis. *Biotechniques* **42**, 71–75 (2007).