# Quantitating SARS-CoV-2 Neutralizing Antibodies from Human Dried Blood Spots

Katherine Berman, Greta Van Slyke, Hayley Novak, Jean M. Rock, Rachel Bievenue, Amanda Damjanovic, Kate DeRosa, Gianna Mirabile, Roxie C. Girardin, Alan P. Dupuis, Kathleen A. McDonough, Monica M. Parker, Linda M. Styer[*], and Nicholas J. Mantis[*]

Division of Infectious Diseases, Wadsworth Center, New York State Department of Health, Albany, NY 12208

**Running title:** Neutralizing antibodies in dried blood spots

*co-corresponding authors: Linda Styer

Nicholas J Mantis, Division of Infectious Diseases, Wadsworth Center, Albany, NY 12208,
email: Linda.Styer@health.ny.gov
email: nicholas.mantis@health.ny.gov

30                          **ABSTRACT**

**Background:** In the earliest days of COVID-19 pandemic, the collection of dried blood

spots (DBS) enabled public health laboratories to undertake population-scale

seroprevalence studies to estimate rates of SARS-CoV-2 exposure. With SARS-CoV-2

seropositivity levels now estimated to exceed 94% in the United States, attention has

turned to using DBS to assess functional (neutralizing) antibodies within cohorts of

interest.

**Methods:** Contrived DBS eluates from convalescent, fully vaccinated and pre-COVID-

19 serum samples were evaluated in SARS-CoV-2 plaque reduction neutralization titer

(PRNT) assays, a SARS-CoV-2 specific 8-plex microsphere immunoassay, a cell-based

pseudovirus assay, and two different spike-ACE2 inhibition assays.

**Results:** DBS eluates from convalescent individuals were compatible with RBD-ACE2

inhibition assays, an in-house Luminex-based RBD-ACE2 inhibition assay, and

commercial real-time PCR-based neutralization assay (NAB-Sure™) but not cell-based

pseudovirus assays or PRNT. The insensitivity of cell-based pseudovirus assays was

overcome with DBS eluates from vaccinated individuals with high SARS-CoV-2

antibody titers.

**Conclusion:** SARS-CoV-2 neutralizing titers can be derived with confidence from DBS

eluates, thereby opening the door to the use of these biospecimens for the analysis of

vulnerable populations and normally hard to reach communities.

55              **Keywords:** Serology; Neutralizing; COVID-19; Antibody; Human

**1. INTRODUCTION**

The collection and serologic analysis of dried blood spots (DBS) from population-based serosurveys proved integral to understanding the dynamics of SARS-CoV-2 infection in the first months of the COVID-19 pandemic. In New York State, for example, a statewide seroprevalence study among ~15,000 individuals in public settings (e.g., grocery store fronts) conducted in April 2020 afforded some of the first insights into overall disease incidence and demographics of virus exposure in urban and rural communities (1). Similarly, a cross-sectional serosurvey of ~400,000 newborns in New York State that employed heel stick DBS collected between November 2019 and November 2021 afforded unprecedent insight into COVID-19 exposures and vaccination status among a diverse population of reproductive-aged females (2). Other groups have leveraged clinician-assisted and self-collected DBS to conduct large scale serosurveys of vulnerable cohorts, school age children, and otherwise hard to reach communities (3-10). In addition to ease of collection, once collected onto Whatman filter paper, DBS samples are stable for long periods of time under refrigeration (4-8$^\circ$C) or frozen (up to 20 $^\circ$C) in the presence of desiccants (11, 12).

Within the COVID-19 pandemic, DBS have largely been employed in serological studies aimed at determining SARS-CoV-2 S and N reactivity as measures of infection and vaccination. However, with SARS-CoV-2 seropositivity levels now estimated to exceed 94% in the United States, the value of such serosurveys has diminished (13). Furthermore, the emergence of omicron subvariants with highly mutated spike proteins confounds one's ability to extrapolate neutralizing antibody titers from binding antibody units (BAUs) (14, 15). For those and other reasons, there is considerable interest in developing a method to assess SARS-CoV-2 neutralizing titers directly from DBS eluates.

Indeed, several reports have already described the use of DBS eluates to assess functional activity of SARS-COV-2 antibodies. For example, Itell and colleagues examined paired plasma and fingerstick (FS) DBS from convalescent individuals in SARS-COV-2 spike-pseudotyped lentiviral particle assay in 96- and 384-well format (16). In another report, Sancilio and colleagues successfully used an ACE2-RBD

binding assay on the Mesoscale Diagnostics (MSD) platform to estimate neutralizing

titers in contrived DBS (cDBS) eluates (17) and in self-collected DBS from a community-

based serology survey (18). However, neither study benchmarked their results against

plaque reduction neutralization titer (PRNT) values, nor did they perform cross

comparisons of the different *in vitro* assays (e.g., pseudovirus versus ACE2-RBD).

Therefore, to further advance the utility and validity of DBS for the purpose of

functional SARS-COV-2 serology, we investigated the compatibility of contrived DBS

eluates with different SARS-COV-2 neutralization assays and compared their outputs to

PRNT values derived from matched serum. While DBS eluates proved incompatible

with cell-based SARS-COV-2 pseudovirus neutralization assays, we found that two

different ACE2-RBD inhibition assays, an in-house Luminex-based assay and a

commercial real-time PCR-based assay (NAB-Sure™), yielded neutralization values

that strongly correlated with PRNT. An 8-plex MIA based solely on binding antibodies to

spike antigens also served as a proxy for PRNT. As the two RBD-ACE2 inhibition

assays employed in this study are amenable to high-throughput testing, we propose to

integrate one or both into our DBS workflow for analysis of cohort specific studies going

forward.

**2. METHODS**

**2.1 SARS-CoV-2 reagents**. The source and descriptions of SARS-CoV-2 reagents

used in this study is provided in **Table 1**.

**2.2 Human serum samples.** A set of serum samples from 30 PCR positive donors was

purchased from Access Biologicals, LLC, (Vista, CA), herein referred to as "Panel

D". Positive test results were reported in the initial U.S. wave of infections between

3/15/2020 and 4/25/2020; and samples were drawn between 6/3/2020-6/4/2020,

indicating that all donors were infected with the Washington variant of SARS-CoV-

2. The panel consists of 21 female and 9 male donors; ages ranged from 25- to 66-

years-old; all donors were symptomatic. A set of 87 COVID-19 negative serum samples

collected between 3/24/2017 and 11/9/2018 was also purchased from Access

Biologicals, LLC. This set is referred to as "Panel E", which consists of pre-pandemic

SARS-CoV-2 sero-negative donors (73 females and 14 males between 18 and 64-

years-old). Upon receipt, Panel D and E serum samples were stored at -80ºC.

**2.3 Contrived DBS.** Contrived DBS were created by mixing an equal volume of Panel

D or E serum with washed human red blood cells (RBC).  Fifty microliters (50 μL) of the

RBC-serum mixture were spotted onto Whatman 903 filter paper (SigmaAldrich, St.

Louis, MO) and allowed to air dry at room temperature for 4 h before being packaged in

semi-permeable bags with desiccant and stored at -20ºC. Red blood cells (RBC) were

collected by centrifugation (3,000 RPM at 4ºC) from commercially obtained fresh human

whole blood [Type O EDTA ] (ZenBio, Durham, NC).

**2.4 Plaque reduction neutralization titer (PRNT).** PRNT was determined as described

(19, 20). Briefly, 2-fold serially diluted serum (100 μl) was mixed with an equal volume of

150-200 plaque forming units (PFUs) of SARS-CoV-2 (isolate USA-WA1/2020, BEI

Resources, NR-52281) and incubated for 1 h at 37°C, 5% $CO_2$. The virus:serum mixture

(100 μl) was applied to VeroE6 cells grown to 95%–100% confluency in 6 well plates.

Adsorption of the virus:serum mixture was allowed to proceed for 1 h at 37°C, 5% $CO_2$.

Following the adsorption period, the wells were overlaid with 0.6% agar prepared in cell

culture medium (Eagle's Minimal Essential Medium, 2% heat inactivated FBS,

100 μg/ml penicillin G, 100 U/ml streptomycin). Two days post-infection, a second agar

overlay containing 0.2% neutral red prepared in cell culture medium was applied.

Plaque determinations were conducted 1-2 days later. Neutralizing titers were defined

as the inverse of the highest dilution of serum providing 50% (PRNT50) or 90%

(PRNT90) viral plaque reduction relative to a virus only control.

**2.5 Microsphere Immunoassay (MIA)**. The development and validation of an 8-plex

MIA consisting of three different nucleocapsid N antigens, RBD, Spike S1 (S1), full

length Spike (FLS), and full length Trimer (FLT) is described elsewhere (J. Rock,

L.Styer, *manuscript in preparation*). Magplex-C microspheres (Luminex Corp) were

coupled with 7 different SARS-CoV-2 antigens and 1 internal control antibody. Each

antigen/antibody was coupled to a different bead region allowing the beads to be
multiplexed in the same well. A 3.3 mm DBS punch from each sample was combined
with 250 µl elution buffer (Tris-buffered saline, 1% casein blocker) (Bio-Rad
Laboratories) in a 96 well plate. The DBS punches were eluted for 1 hour at room
temperature. Antigen-coated beads (diluted to 1,250 beads/bead set/well with assay
buffer (PBS, 2% BSA, pH 7.4)) and 25 µl DBS eluate were combined in a 384 well plate
and incubated at 37°C for 30 minutes while shaking (300 RPM) in the dark. Plates were
washed for two cycles using the BioTek 405TS microplate washer and 50 µl wash buffer
(PBS, 2% BSA, 0.02% Tween-20, 0.05% sodium azide, pH 7.4). 50 µl phycoerythrin-
tagged goat anti-human IgG was added to each well. Plates were incubated at 37°C for
30 minutes while shaking (300 RPM) in the dark and washed as described above.
Antibody-bead complexes were resuspended with 90 µl xMap sheath fluid (Luminex),
plates were incubated for 1 minute at room temperature while shaking (300 RPM), and
70 µl from each well was read on a Luminex FlexMAP 3D instrument. Results were
presented in median fluorescence intensity (MFI) for each well. Positive, negative, and
background controls were included with every plate.
167
**2.6 Reporter virus particle (RVP) neutralization assays.** Pseudo-typed lentiviral
reporter virus particles (RVP) tagged with Renilla luciferase (Luc) and 293T-hsACE2
cells were obtained from Integral Molecular, Inc. (Philadelphia, PA). To determine the
SARS-CoV-2 neutralizing activity in human serum and paired cDBS eluates, serum and
DBS eluates were diluted 1:50 in DMEM containing 10mM HEPES [pH 7.4], 10% fetal
bovine serum and 100 U/mL penicillin and 0.1 mg/mL streptomycin (Sigma-Aldrich).
Paired sample types were serially diluted in duplicate in cell medium. RVPs were
thawed at 37°C and immediately diluted per the manufacturer's specification in cell
medium and then 50 µl was added to the assay plate. A total of 50 µl of diluted human
serum or DBS eluate were then transferred from the dilution plate and mixed with RVP
solution. Serum and DBS eluates were incubated with RVPs for 1 hour at 37°C with 5%
$CO_2$. Positive and negative control wells were included on each assay plate and
consisted of RVPs only (100% infectivity control), and no RVPs (0% infectivity control).
During RVP-serum incubation 80-90% confluent 293T-hsACE2 cells were harvested

and cell density was determined using a hemocytometer. Cells were diluted to a density

of $2 \times 10^5$ cells/mL in cell culture media and 100 µl was added to each well of the RVP +

specimen plate and mixed with gentle pipetting. Plates were incubated at 37°C with 5%

$CO_2$ for 72 h. Quantification of SARS-CoV-2 RVP infection was determined using the

*Renilla*-Glo® luciferase assay system (Promega Corp., Madison, WI). Microtiter assay

plates were centrifuged for 5 min at 2000 rpm to prevent cell loss, then supernatant was

removed with a pipet and 30 µl of PBS was added to each well. *Renilla*-Glo® substrate

diluted 1:100 (v/v) into buffer was prepared fresh; 30 µl added per well and plates were

then incubated for 10 min at room temperature. Relative light units (RLU) were detected

on a SpectroMax iD3 (Molecular Devices, LLC, San Jose, CA) and analyzed with

SoftMax Pro 7.0 software. A Z' value was calculated for each plate for assay accuracy

and comparability, plates with a Z'<0.5 were repeated. The RLU mean and standard

deviation (SD) of 100% infected wells was determined, and a minimum limit of detection

was calculated for each plate as = $(Mean– (3xSD))_{100\%Infected.}$ Samples that failed to

reach the LOD cut-off value at the most concentrated dilution were classified as non-

neutralizing (<LOD) and assigned a value of 0 for statistical analysis purposes. For each

plate the RLU mean value for 0% infected wells was subtracted from the mean of 100%

infected wells resulting in a normalized infection value (NIV); % normalized RVP

neutralization was calculated as = $100-(( S_{AVE}/NIV)*100)$, where $S_{AVE}$= the average of

duplicate sample RLU values.

**2.7 Human ACE2 inhibition (iACE2) assay.** Recombinant soluble human ACE2 was

generated by the Wadsworth Center's Protein Expression core facility. The coding

region for human ACE2 (residues 1-615; Genbank #) was cloned into pcDNA3-

sACE2(WT)-8his [(https://www.addgene.org/browse/sequence/298420/] and transfected

into ExpiCHO-S cells using ExpiFectamine, as per manufacturer's instructions

(ThermoFisher (https://www.thermofisher.com/order/catalog/product/A29133). The

recombinant protein was affinity purified on 2 x 1ml HisTrap™ Excel (Cytiva) columns

connected in series. ACE2 was eluted with a linear gradient of 20 mM Tris-HCl, 500 mM

NaCl, 500 mM imidazole pH 8.0 over 20 column volumes. Fractions containing ACE2,

as assessed by SDS-PAGE, were pooled and dialyzed twice against PBS [pH 7.4].

Protein concentration assessed by Bradford assay with bovine albumin as a standard

curve. Biotinylated using EZ-Link NHS Biotin (Thermo Fisher Scientific).

Magplex-C microspheres (Luminex) from separate bead regions were coupled

with RBD and FLT from SARS-CoV-2 isolate USA-WA1/2020 ("wild type") and

resuspended to 2500 beads/region/well. Diluted beads were transferred to wells of a

non-binding 96 well plate; 50 µl of serum or DBS eluate (1:138) was then added to each

well and incubated for 1 h RT with agitation in the dark. Plates were washed twice

manually with wash buffer (190 µl) and then biotinylated-hACE2 (see below; 4 µg/mL)

was added to each well, plates were incubated for an additional 45 min at RT, shaking

in the dark. Plates were washed, probed with streptavidin-PE (1 µg/mL) and incubated

at RT for 30 min. The plates were washed to removed unbound streptavidin-PE then

analyzed using FlexMap3D. The MFI for ACE2 binding (in the absence of antibody) was

derived from a single column of wells containing diluted beads. Inhibition of ACE2

(iACE2) binding calculated as: 190 µl % ACE2 binding-inhibition (iACE2), per sample

well, where % iACE2 = 100-[Raw value/(Average of wells with specific VoC beads and

no serum or DBS eluate) x 100].

**2.8 NAB-Sure™**

The NAB-Sure™ SARS-CoV-2 Neutralizing Antibody Test (Spear Bio, MA, USA) was

performed according to the manufacturer's instructions (www.spear.bio/nabsure). NAB-

Sure™ is a competitive binding assay that measures the ability of antibodies in DBS

samples to block SARS-CoV-2 Spike Protein (S1) from binding to recombinant hACE2

in solution. The assay was validated at the Wadsworth Center using eluates (n=11) from

SARS-CoV-2 infected individuals with range of Spike-specific titers provided by

SpearBio. Contrived Panel D DBS punches (6mm) were incubated in NAB-Sure™

elution buffer at $37^{o}$C for 1 h. Afterwards, eluates (15$\mu$L) were processed according to

the manufacturer's instructions. The test was read using a 7500 Fast Real-Time PCR

System (Thermo Fisher Scientific) in the Wadsworth Center's Bloodborne Virus

Laboratory. CT values were converted to $NT_{50}$ values relative to the negative control

sample that is devoid of SARS-CoV-2 neutralizing antibodies. The cycle difference

between the sample CT value and the negative control CT value was calculated.

Inhibition % was calculated as $= 1-1/2^x$, where $x=$ cycle difference. An interpolated NT50

value was determined using GraphPad Prism software. The final NT50 value$= y*50$,

where y=interpolated NT50 value.

**2.9 Statistical Analysis.**

Statistical analyses were performed using GraphPad Prism 9 software. Correlations

were assessed by the Pearson correlation coefficient. An unpaired, two-tailed Welch's t-

test was performed to determine significance of differences in ACE2 inhibition between

the two groups. Statistical significance of differences in RVP dilutions was assessed

using a 2-way ANOVA. In all cases, the following apply: $*p \leq 0.05$, $**p \leq 0.01$, $***p \leq$

$0.001$, $****p < 0.0001$.

**3. RESULTS**

**3.1 Generation of standard SARS-CoV-2 DBS panel.** The primary objective of this

study was to assess SARS-CoV-2 virus neutralizing activity from DBS eluates

benchmarked against PRNT assays, since performing actual PRNT assays on DBS

eluates is neither practical nor technically feasible. For this purpose, we generated

contrived DBS (cDBS) from a commercial panel ("Panel D") consisting of 30 individuals

with a defined history of SARS-CoV-2 infection during or prior to June 2020. The

participants ranged in age from 25 to 63 (average 45) with 21 females and 9 males. Of

the 30 COVID-19 positive samples, 26 were confirmed positive via SARS-CoV-2 PCR

and four strictly through medical diagnosis$PRNT_{50}$ values for 15 of the 30 samples were

determined using SARS-CoV-2 Washington (WA1), as described (19). The average

$PRNT_{50}$ for the cohort was 351 (+/-271). In addition, the percent SARS-CoV-2

neutralization at a fixed 1:100 serum dilution (average 55 +/- 22%) was extrapolated

based on PRNT curves to enable direct comparison with values derived from DBS

eluates (see below).  A summary of the 15 Panel D participants is provided in **Table 2.**

We also generated cDBS samples from Panel D. As detailed in the Materials and

Methods, serum was combined with an equal volume of human RBCs then spotted (50

$\mu$l) onto Whatman 903 filter paper. The spots were allowed to dry at room temperature

for 4 h then stored with desiccant at -20ºC. Recovery of 90% ($\pm$0.04%) total IgG was

achieved from a single 3 mm punch (1/8'') by incubating the punch in 250 $\mu$l in TBS with 1% casein for 1 h at room temperature (**Figure 1**). The cDBS eluates were prepared fresh for each assay described below. In addition, we also generated cDBS spots from a commercially available panel of serum samples from a healthy cohort (**Panel E)** as well as a fully vaccinated cohort (**Panel H**).

**3.2 Assessment of SARS-CoV-2 S- and N-specific IgG in DBS eluates by MIA.**
Using a recently optimized 8-plex MIA targeting different nucleocapsid (N) and Spike antigens, we examined the relationship between MFIs derived from cDBS eluates and PRNT values. Specifically, the 8-plex MIA includes full length Trimer (FLT), full length spike (FLS), S1, RBD and three N antigens as shown in **Table 1**. As expected, Panel D cDBS eluates had higher median fluorescent intensity (MFI) than Panel E cDBS eluates for all seven SARS-CoV-2 antigens (**Figure S1**). In terms of correlation with PRNT (% neutralization @ 1:100 dilution), the values ranged from 0.76 for FLT (p=0.001) and S1 (p=0.001) to 0.82-0.83 for FLS (p =0.0002) and RBD (p=0.0001), respectively (**Figure 2**). Summing the MFIs for all four spike antigens ("sum Spike") did not greatly increase r values (r = 0.81) (**Figure 2E**). As expected, the correlation between N MFI and PRNT was not significant (r =0.39; p>0.1; **Figure 2F**). We conclude that MFI derived from DBS eluates to a given SARS-CoV-2 spike antigen correlates with virus neutralizing activity as defined by PRNT.

**3.3. Pseudovirus neutralization activity in cDBS eluates.** Non-replicating, lentiviral-based pseudoviruses expressing SARS-CoV-2 trimeric spike proteins are employed widely to assess SARS and SARS-CoV-2 neutralizing antibody titers in plasma and serum (21-24). However, it is unclear whether DBS eluates are amenable with these cell-based assays. To investigate this, we employed a commercially available non-replicating Reporter Virus Particles (RVP) expressing the SARS-CoV-2 spike protein (D648G) and encoding *Renilla* luciferase (**Figure 3**) (25). We confirmed RVP efficacy using the human monoclonal antibody CC12.3, which neutralizes SARS-CoV-2 with a reported $IC_{50}$ of 0.026 $\mu$g/ml (24). In our hands, CC12.3 had an $IC_{50}$ of 0.04 $\mu$g/ml (**Figure 3A**). Moreover, the 15 Panel D serum samples had neutralization activities that

ranged from 15% to 98% with high concordance with PRNT (r=0.67; p=0.007) (**Figure 3B**). Thus, RVP serves as reliable measure of SARS-CoV-2 activity in the context of serum.

To determine if the RVP assay is compatible with cDBS, we collected eluates from three 1/8" (3 mm) punches from each of the 15 cDBS Panel D spots and 7 Panel E spots as negative controls. Eluates were generated by incubating the punches in DMEM-HEPES at room temperature for 2 h with agitation. Eluates were then diluted 1:50 into DMEM-HEPES medium and then mixed with RVPs prior to application to cells 293T cells constitutively expressing human ACE2 (see **Materials and Methods**).

Panel D eluates had 60-90% neutralization activity [average 68%] that were modestly correlated with PRNT (r value of 0.61; p=0.02) (**Figure 3C**). However, the results were confounded because Panel E eluates (negative controls) had "neutralization" (interfering) activity that was essentially equivalent to Panel D (**Figure 3D**). Thus, RVP assay was not able to distinguish between Panel D and Panel E eluates. Efforts to improve the sensitivity of the assay through dilution (>1:3200), alternative elution buffers, and removal of hemoglobin using reagents designed for this purpose (e.g., HemogloBind™) were unsuccessful (**Figure 3D**; **data not shown**). We conclude that cDBS eluates from convalescent (unvaccinated) individuals have insufficient virus neutralizing titers to distinguish them from unvaccinated controls in the pseudovirus (RVP) platform.

**3.4 Relationship between PRNT and iACE2.** We next tested cDBS eluates for the ability to inhibit ACE2-RBD interactions in using a Luminex-based assay (iACE2). In the assay, RBD- and FLT-coupled microspheres were incubated with soluble biotinylated-ACE2 in the absence or presence of competitor antibodies (e.g., monoclonal, antisera or DBS eluates), then probed with streptavidin-PE and analyzed by using FlexMap3D (Luminex). A reduction in MFI was interpreted as antibody-mediated inhibition of ACE2 binding to RBD- and FLT. The utility of the in-house iACE2 assay was confirmed using the well characterized monoclonal antibody CC12.3, which neutralizes SARS-CoV-2 by interfering with ACE2-RBD interactions (24). In the iACE2 assay, CC12.3 had an $IC_{50}$

value for WT FLT and RBD of 0.04 ug/mL and 0.02 ug/mL, respectively (**Figure S2**).

These values are consistent with those reported by Rogers and colleagues (24).

We examined Panel D DBS eluates in the iACE2 assay, with the expectation that

Panel D eluates would have greater inhibition activity than Panel E. Indeed, Panel D

had significantly higher iACE2 values (23% average; range 10%-50%) as compared to

Panel E (9% average; range 0%-15%) (**Figure 4A**). In terms of the correlation with

PRNT, Panel D iACE2 values had an r-value of 0.77 (p<0.001) for RBD-coated beads

and 0.72 (p=0.003) for FLT-coated beads. iACE2 (%) values for FLT-coated beads

ranged from 25% to 40% (avg 31%) (**Figure 4; Figure S2**). Panel D DBS eluates

performed slightly less well than the matched serum samples from which they were

derived, at the same dilution, indicating that a degree of ACE2-RBD inhibitory activity is

lost during the collection of cDBS eluates. Panel D serum samples (1:100 dilution) had

iACE2 activity that correlated with PRNT with r= 0.81 (p<0.001; **Figure S2**). These

results demonstrate that iACE2 assay affords a reliable estimate of SARS-CoV-2

neutralization potential using cDBS eluates.

**3.5. NAB-Sure™.**  NAB-Sure™ is a commercially available cell-free assay based on

successive proximity extension amplification reaction (SPEAR) technology with qPCR

as a readout. Disruption of a proximity extension amplification reaction primed by

recombinant SARS-CoV-2 S1 protein and human ACE2 by neutralizing antibodies in

DBS eluates results in increased CT values. To investigate this technology, we first

evaluated the NAB-Sure™ assay using a set of 11 previously authenticated DBS

eluates provided by the manufacture and which were subjected to five 3-fold serial

dilutions (1 to 243) and tested in duplicate using in-house instrumentation. The NAB-

Sure™ assay revealed $NT_{50}$ values that ranged from 210 to 9500 (avg. 2333) with 17.7

%CV (**Figure 5A**). There was a significant correlation between $NT_{50}$ values from DBS

eluates tested by NAB-Sure™ and paired serum samples tested by PRNT [Pearson's r

= 0.99; p<0.0001] (**Figure 5A**). Thus, NAB-Sure™ was compatible with DBS eluates

and had a high concordance with PRNT.

We then used the NAB-Sure™ assay to evaluate cDBS eluates from Panel D

(#1-14) and E (#1-7). The samples were tested 3 times in independent reactions under

identical assay conditions. Concordance across the different reactions was high, as

evidenced by r values between 0.85-0.92 **(Figure 5B-D)**. The average %CV across

samples was 20.3%. All reactions reached significance when compared to PRNT. From

these results, we conclude that NAB-Sure™ is sufficiently sensitive to detect SARS-

CoV-2 neutralizing titers in DBS eluates, based on high degree of concordance with

PRNT.

**3.6 Assessing SARS-CoV-2 neutralizing activity in vaccinated individuals.** Studies

prior to this point relied on DBS eluates derived from SARS-CoV-2 infected individuals

with relatively low overall Spike-specific antibody titers as well as SARS-CoV-2

neutralizing levels. With an estimated >90% of the United States population seropositive

for SARS-CoV-2 due to vaccination and/or natural infection at this point in time (13), we

wished to revisit the utility of DBS eluates in assessing virus-specific antibody titers and

neutralizing activity in a vaccinated cohort. For this purpose, we generated cDBS from a

commercially available panel (**Panel H**) of serum samples collected from individuals

who had received two doses of the original Moderna (mRNA 1273) vaccine. As shown

in **Figure 6**, Panel H DBS eluates (n=15; 15746 $\pm$ 1790 MFI) had RBD-specific antibody

levels (MFI) what were 1.5 times higher than those in Panel D (10392 $\pm$ 3325 MFI) and

44 times higher than Panel E (351 $\pm$ 313 MFI). DBS eluates were evaluated in the

iACE2 assay (**Figure 6C, D**), RVP (pseudovirus) neutralization assay (**Figure 6E, F**)

and NAB-Sure™ (**Figure 6G, H**). Unfortunately, insufficient volumes of this panel were

available for PRNT assessment. Nonetheless, the results of these analyses

demonstrated that all three platforms (iACE2, RVP, NAB-Sure™) were compatible with

DBS eluates as evidenced by ACE2 inhibition values approaching 100% (**Figure 6C),**

RVP $NT_{50}$ values > 5000 (**Figure 6E**), and NAB-Sure™ $NT_{50}$ values approaching 18,000

(**Figure 6G**). There were statistically significant correlations between RBD binding

antibody levels (MFI; y-axis) and iACE2 values (x-axis; **Figure 6D),** as well as RBD

binding antibody levels (MFI; y-axis) and RVP $NT_{50}$ values (x-axis; **Figure 6F**). This

observation is consistent with the earliest observations serology-based correlates of

immunity elicited by the spike-encoding mRNA vaccines (14).  We did not, however,

observe a significant correlation between RBD antibody levels (MFI) and NAB-Sure™

$NT_{50}$ values (**Figure 6H**) or RVP $NT_{50}$ values and NAB-Sure™ $NT_{50}$ values (r = 0.17;

**data not shown**). These results suggest that NAB-Sure™ may perform optimally with

DBS eluates samples with lower Spike-specific antibody titers, whereas the iACE2 and

RVP assays are better suited for samples with higher Spike-specific antibody titers.

Thus, specific assays may be tailored to specific cohorts and their anticipated SARS-

CoV-2 titers.

**4.0 DISCUSSION**

As the World Health Organization (WHO) and other public health agencies

integrate long term COVID-19 disease management strategies into the broader

landscape of infectious disease, DBS have the potential to serve as a centralized

biospecimen type from which to estimate levels of natural and vaccine-induced

immunity at population scale (11). Indeed, DBS-based seroprevalence and

neutralization assays have been established and deployed to monitor vaccine-induced

immunity for the likes of polio (26), human papillomavirus (HPV) (27) and several

bacterial pathogens and parasites (28). However, with SARS-COV-2 seroprevalence

nearing universality as a result of natural infections/reinfections alongside widespread

vaccination, the future utility of DBS to COVID-19 will depend on the development of

robust methods to assess functional antibody titers alongside established BAU

determinations (14, 29). With this objective in mind, we investigated the compatibility of

DBS eluates with different SARS-COV-2 neutralization, pseudovirus neutralization, and

ACE2-RBD inhibition assays. While neither the SARS-CoV-2 PRNT nor the pseudovirus

(RVP) neutralization assays proved compatible with DBS eluates from convalescent

individuals due to high background, values from ACE2-RBD inhibition assays and NAB-

Sure™ correlated neutralization titers ($NT_{50}$ and %) from paired serum. In a cohort of

SARS-CoV-2 vaccinated individuals, the ACE2-RBD inhibition and RVP neutralization

assays proved more robust than NAB-Sure™, at least under these experimental

conditions and with this limited data set. With this in mind, we would recommend

tailoring SARS-CoV-2 antibody assessments based on cohorts under investigation. In

summary, we conclude that DBS eluates are indeed compatible with multiple SARS-

CoV-2 surrogate neutralizing assays and can be applied in parallel with existing and

forecasted community and population-based serosurveillance efforts.

Itell and colleagues were the first to investigate the use of DBS to access

functional (neutralizing) antibodies from COVID-19 convalescents and SARS-CoV-2

vaccinees (16). In that report, a cell-based SARS-CoV-2 pseudovirus assay proved to

be a reliable measure of neutralizing titers with both mock DBS (venous blood spotted

on Whatman filter paper) and self-collected fingerprick DBS from vaccinated individuals.

In a larger scale study, Roper and colleagues reported success "estimating" SARS-

CoV-2 pseudovirus neutralizing titers from fingerprick DBS eluates but noted that

multiple samples of the 12 tested were eliminated due to values below limit of detection

(30). In their study, fingerprick DBS eluates were not reported at dilutions less than 256,

suggesting to us the possibility of interference at low dilutions. They also used eluates

from a single 19 mm sample disc into PBS, compared to ~9 mm (3 x 3 mm) in our study

eluted into DMEM and 6 mm by Itell into PBS with 0.1 % Tween-20 (16). Collectively,

our results are consistent with DBS eluates being compatible with certain SARS-CoV-2

pseuodovirus assays but only when dilutions exceed ~256 to eliminate blood-based

interference.

McDade and colleagues reported the compatibility of DBS eluates with an ACE2-

RBD inhibition (iACE2) assay on the Mesoscale Diagnostics platform (17). We now

extend those results and demonstrate that both an in-house Luminex-based assay and

a commercial real-time PCR-based assay (NAB-Sure™), give rise to inhibition values

that strongly correlate with neutralization titers derived from PRNT. While we recognize

that the iACE2 assay is strictly a surrogate neutralization assay, both platforms

described herein have the potential to be integrated alongside existing high-throughput

SARS-CoV-2 multiplexed serologic assays. At present, we employ an 8-plex SARS-

CoV-2 microsphere immunoassay (MIA) run in 384 well format on the Luminex platform

with automated liquid handling (12). The in-house iACE2 assay described here is also

run on the Luminex platform using the same magnetic microspheres as the MIA. The

overall costs of the iACE2 assay are at least an order or magnitude less expensive than

cell-based pseudovirus assay with commercial RVPs. Moreover, the in-house iACE2

assay is easily multiplexed to include microspheres coated with RBD from circulating

VoCs and Omicron subvariants side-by-side with ancestral RBD. Indeed, we have
applied a 4-plex RBD assay to several thousand primary DBS samples collected from
an elderly cohort (G. Mirabelle, K. Berman, N. Mantis, unpublished results).

The NAB-Sure™ assay has the advantage of being ultrasensitive and amenable
to scale-up and automation and provided $NT_{50}$ titers similar to PRNT. Moreover, as
NAB-Sure™ is a commercial kit, there were no in-house costs associated with reagent
quality assurance or validation making the kit comparable (cost wise) with most
pseudovirus assays. In our hands, NAB-Sure™ proved most robust with DBS eluates in
which SARS-CoV-2 levels were on the low end (e.g. convalescent samples), while RVP
and iACE2 performed well with high titer samples associated with vaccination. It is
unclear whether this distinction will apply to DBS samples aimed at evaluating antibody
titers against SARS-CoV-2 variants of concern (VoC) like the highly divergent JN.1. In
fact, a tiered approach to interrogating DBS eluates may be in order where MFI values
alone provide a rough idea of neutralizing titers within a given sample especially when
dealing with thousands of samples from diverse cohorts (31). In terms of functional
antibodies, the Luminex-based iACE2 assay is both affordable and adaptable to any
given RBD. In fact, we have performed iACE2 multiplexing with four different RBD
variants simultaneously on hundreds of primary DBS samples from multiple different
cohorts (G. Mirabile, K. Berman, N. Mantis, *manuscript in preparation*). Ultimately, both
the NAB-Sure™ and pseudovirus assays can be employed as a means to establish
quantitative $NT_{50}$ equivalents not achievable with MFI or iACE2 alone.

The use of DBS for infectious disease serology and functional antibody testing is
not without intrinsic challenges, including variability in blood collection by fingerpick,
uniformity of antibody recovery in eluates, and limited sample volumes (11).
Nonetheless, the benefits of DBS as a biospecimen type for surveillance of hard-to-
reach communities are well recognized. For example, in work spearheaded by the CDC,
DBS from ~1,000 children in Bangladesh were sampled for neutralizing antibodies to
polio types 1, 2, and 3 using a modified poliovirus microneutralization assay (26). In the
case of polio, a neutralizing antibody titer of 1:8 correlated with protection and was used
as the threshold definition of seropositivity. Similarly, in a study examining natural
versus vaccine-induced antibody responses to human papillomavirus (HPV),

neutralizing activity was detected in 2 mm DBS punches eluted into 0.2 mL saline (27).

Another consideration is that for SARS-CoV-2 at least there is a strong congruence

between oral fluids and DBS (K. DeRosa, N. Pasaic, C. Heaney, L. Styer, and N.

Mantis, *manuscript in preparation*). Thus, DBS and oral fluids might potentially be used

interchangeably as means to survey large cohorts for immunity and vaccine

responsiveness (32).

**ACKNOWLEDGEMENTS**

We gratefully acknowledge SpearBio, Inc (Woburn, MA) for providing matched serum

and DBS samples as well as assistance in optimizing NAB-Sure™ assay at the

Wadsworth Center. The SARS-CoV-2 reagents provided by MassBiologics (Boston,

MA) were generated under a Project Award Agreement to NJM from the National

Institute for Innovation in Manufacturing Biopharmaceuticals (NIIMBL), United States

and financial assistance award 70NANB20H037 from the US Department of Commerce,

National Institute of Standards and Technology. Research reported in this publication

was supported by the National Cancer Institute of the National Institutes of Health under

Award Number U01CA260508. The content is solely the responsibility of the authors

and does not necessarily represent the official views of the National Institutes of Health.

508
509 **CRediT**
510

| Author | Conceptualization | Curation | Formal analysis | Funding acquisition | Investigation | Methodology | Admin | Resources | Software | Supervision | Validation | Visualization | Writing-original | Writing-review |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| KB | | | x | | x | x | | | | | x | x | x | x |
| GVS | | | | | x | x | | | | | | | | |
| HN | | | | | x | x | | | | | | | | |
| JRM | | | x | | x | x | | | | | x | x | | |
| RB | | | x | | x | x | | | | | x | x | x | x |
| AD | x | x | | | | | | | | | x | | x | x |
| KD | x | | | | x | x | | | | | | | | x |
| GM | | | | | x | x | | | | | | | | |
| RCG | | | x | | x | x | | | | | x | | | |

| | | | | x | x | | | | | x | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| APD | | | | x | x | | | | | x | | | |
| KAM | | | | | | x | x | x | | | | | x |
| MMP | x | | x | | | x | x | x | | | | | |
| LMS | x | | x | | | x | x | x | | | x | x |
| NJM | x | | x | x | | x | x | x | | | x | x |

**Conflicts of Interests (COI).** The authors have no conflicts of interest to declare.

**REFERENCES CITED**

1.  Rosenberg ES, Tesoriero JM, Rosenthal EM, Chung R, Barranco MA, Styer LM, Parker MM, John Leung SY, Morne JE, Greene D, Holtgrave DR, Hoefer D, Kumar J, Udo T, Hutton B, Zucker HA. 2020. Cumulative incidence and diagnosis of SARS-CoV-2 infection in New York. Ann Epidemiol doi:10.1016/j.annepidem.2020.06.004.

2.  Damjanovic A, Styer LM, Nemeth K, Yauney E, Rock JM, Bievenue R, Hoen R, Ehrbar D, Kay DM, Caggana M, Parker MM. 2022. Utility of Newborn Dried Blood Spots to Ascertain Seroprevalence of SARS-CoV-2 Antibodies Among Individuals Giving Birth in New York State, November 2019 to November 2021. JAMA Netw Open 5:e2227995.

3.  Amendola A, Bianchi S, Gori M, Barcellini L, Colzani D, Canuti M, Giacomet V, Fabiano V, Folgori L, Zuccotti GV, Tanzi E. 2021. Dried Blood Spot as an Alternative to Plasma/Serum for SARS-CoV-2 IgG Detection, an Opportunity to Be Sized to Facilitate COVID-19 Surveillance Among Schoolchildren. Pediatr Infect Dis J 40:e46-e47.

4.  Basto-Abreu A, Carnalla M, Torres-Ibarra L, Sanchez-Pajaro A, Romero-Martinez M, Martinez-Barnetche J, Lopez-Martinez I, Aparicio-Antonio R, Shamah-Levy T, Alpuche-Aranda C, Rivera JA, Barrientos-Gutierrez T, Ensanut-Covid C. 2023. SARS-CoV-2 seroprevalence and vaccine coverage from August to November 2021: A nationally representative survey in Mexico. J Med Virol 95:e29038.

5.  Matias WR, Fulcher IR, Sauer SM, Nolan CP, Guillaume Y, Zhu J, Molano FJ, Uceta E, Collins S, Slater DM, Sanchez VM, Moheed S, Harris JB, Charles RC, Paxton RM, Gonsalves SF, Franke MF, Ivers LC. 2023. Disparities in SARS-CoV-2 Infection by Race, Ethnicity, Language, and Social Vulnerability: Evidence from a Citywide Seroprevalence Study in Massachusetts, USA. J Racial Ethn Health Disparities doi:10.1007/s40615-022-01502-4:1-11.

6.  Nemeth KL, Yauney E, Rock JM, Bievenue R, Parker MM, Styer LM. 2023. Use of Self-Collected Dried Blood Spots and a Multiplex Microsphere Immunoassay to Measure IgG Antibody Response to COVID-19 Vaccines. Microbiol Spectr 11:e0133622.

7.  Holt H, Jolliffe DA, Talaei M, Faustini S, Vivaldi G, Greenig M, Richter AG, Lyons RA, Griffiths CJ, Kee F, Sheikh A, Davies GA, Shaheen SO, Martineau AR. 2023. Incidence determinants and serological correlates of reactive symptoms following SARS-CoV-2 vaccination. NPJ Vaccines 8:26.

553    8.      Oelsner EC, Krishnaswamy A, Balte PP, Allen NB, Ali T, Anugu P, Andrews HF,
554           Arora K, Asaro A, Barr RG, Bertoni AG, Bon J, Boyle R, Chang AA, Chen G,
555           Coady S, Cole SA, Coresh J, Cornell E, Correa A, Couper D, Cushman M,
556           Demmer RT, Elkind MSV, Folsom AR, Fretts AM, Gabriel KP, Gallo LC,
557           Gutierrez J, Han MLK, Henderson JM, Howard VJ, Isasi CR, Jacobs DR, Jr.,
558           Judd SE, Mukaz DK, Kanaya AM, Kandula NR, Kaplan RC, Kinney GL,
559           Kucharska-Newton A, Lee JS, Lewis CE, Levine DA, Levitan EB, Levy BD, Make
560           BJ, Malloy K, Manly JJ, Mendoza-Puccini C, et al. 2022. Collaborative Cohort of
561           Cohorts for COVID-19 Research (C4R) Study: Study Design. Am J Epidemiol
562           191:1153-1173.
563    9.      Walmsley SL, Szadkowski L, Wouters B, Clarke R, Colwill K, Rochon P, Brudno
564           M, Ravindran R, Raboud J, McGeer A, Oza A, Graham C, Silva A, Manase D,
565           Maksymowsky P, Parente L, Dayam RM, Simpson J, Pasculescu A, Gingras AC.
566           2023. COVID-19 vaccine antibody responses in community-dwelling adults to 48
567           weeks post primary vaccine series. iScience 26:106506.
568    10.    Wong MP, Meas MA, Adams C, Hernandez S, Green V, Montoya M, Hirsch BM,
569           Horton M, Quach HL, Quach DL, Shao X, Fedrigo I, Zermeno A, Huffaker J,
570           Montes R, Madden A, Cyrus S, McDowell D, Williamson P, Contestable P, Stone
571           M, Coloma J, Busch MP, Barcellos LF, Harris E. 2022. Development and
572           Implementation of Dried Blood Spot-Based COVID-19 Serological Assays for
573           Epidemiologic Studies. Microbiol Spectr 10:e0247121.
574    11.    Lim MD. 2018. Dried Blood Spots for Global Health Diagnostics and
575           Surveillance: Opportunities and Challenges. Am J Trop Med Hyg 99:256-265.
576    12.    Styer LM, Hoen R, Rock J, Yauney E, Nemeth K, Bievenue R, Parker MM. 2021.
577           High-Throughput Multiplex SARS-CoV-2 IgG Microsphere Immunoassay for
578           Dried Blood Spots: A Public Health Strategy for Enhanced Serosurvey Capacity.
579           Microbiol Spectr 9:e0013421.
580    13.    Jones JM, Manrique IM, Stone MS, Grebe E, Saa P, Germanio CD, Spencer BR,
581           Notari E, Bravo M, Lanteri MC, Green V, Briggs-Hagen M, Coughlin MM, Stramer
582           SL, Opsomer J, Busch MP. 2023. Estimates of SARS-CoV-2 Seroprevalence
583           and Incidence of Primary SARS-CoV-2 Infections Among Blood Donors, by
584           COVID-19 Vaccination Status - United States, April 2021-September 2022.
585           MMWR Morb Mortal Wkly Rep 72:601-605.
586    14.    Sobhani K, Cheng S, Binder RA, Mantis NJ, Crawford JM, Okoye N, Braun JG,
587           Joung S, Wang M, Lozanski G, King CL, Roback JD, Granger DA, Boppana SB,
588           Karger AB. 2023. Clinical Utility of SARS-CoV-2 Serological Testing and Defining
589           a Correlate of Protection. Vaccines (Basel) 11.
590    15.    Wang Q, Iketani S, Li Z, Liu L, Guo Y, Huang Y, Bowen AD, Liu M, Wang M, Yu
591           J, Valdez R, Lauring AS, Sheng Z, Wang HH, Gordon A, Liu L, Ho DD. 2023.
592           Alarming antibody evasion properties of rising SARS-CoV-2 BQ and XBB
593           subvariants. Cell 186:279-286 e8.
594    16.    Itell HL, Weight H, Fish CS, Logue JK, Franko N, Wolf CR, McCulloch DJ,
595           Galloway J, Matsen FAt, Chu HY, Overbaugh J. 2021. SARS-CoV-2 Antibody
596           Binding and Neutralization in Dried Blood Spot Eluates and Paired Plasma.
597           Microbiol Spectr 9:e0129821.

17.  Sancilio AE, D'Aquila RT, McNally EM, Velez MP, Ison MG, Demonbreun AR,
McDade TW. 2021. A surrogate virus neutralization test to quantify antibody-
mediated inhibition of SARS-CoV-2 in finger stick dried blood spot samples. Sci
Rep 11:15321.
18.  Sancilio A, Schrock JM, Demonbreun AR, D'Aquila RT, Mustanski B, Vaught LA,
Reiser NL, Velez MP, Hsieh RR, Ryan DT, Saber R, McNally EM, McDade TW.
2022. COVID-19 symptom severity predicts neutralizing antibody activity in a
community-based serological study. Sci Rep 12:12269.
19.  Girardin RC, Dupuis AP, Payne AF, Sullivan TJ, Strauss D, Parker MM,
McDonough KA. 2021. Temporal Analysis of Serial Donations Reveals Decrease
in Neutralizing Capacity and Justifies Revised Qualifying Criteria for Coronavirus
Disease 2019 Convalescent Plasma. J Infect Dis 223:743-751.
20.  Yates JL, Ehrbar DJ, Hunt DT, Girardin RC, Dupuis AP, Payne AF, Sowizral M,
Varney S, Kulas KE, Demarest VL, Howard KM, Carson K, Hales M, Ejemel M, Li
Q, Wang Y, Peredo-Wende R, Ramani A, Singh G, Strle K, Mantis NJ,
McDonough KA, Lee WT. 2021. Serological Analysis Reveals an Imbalanced IgG
Subclass Composition Associated with COVID-19 Disease Severity. Cell Reports
Medicine doi:https://doi.org/10.1016/j.xcrm.2021.100329:100329.
21.  Chandrashekar A, Liu J, Martinot AJ, McMahan K, Mercado NB, Peter L,
Tostanoski LH, Yu J, Maliga Z, Nekorchuk M, Busman-Sahay K, Terry M, Wrijil
LM, Ducat S, Martinez DR, Atyeo C, Fischinger S, Burke JS, Slein MD, Pessaint
L, Van Ry A, Greenhouse J, Taylor T, Blade K, Cook A, Finneyfrock B, Brown R,
Teow E, Velasco J, Zahn R, Wegmann F, Abbink P, Bondzie EA, Dagotto G,
Gebre MS, He X, Jacob-Dolan C, Kordana N, Li Z, Lifton MA, Mahrokhian SH,
Maxfield LF, Nityanandam R, Nkolola JP, Schmidt AG, Miller AD, Baric RS, Alter
G, Sorger PK, Estes JD, et al. 2020. SARS-CoV-2 infection protects against
rechallenge in rhesus macaques. Science 369:812-817.
22.  Corbett KS, Flynn B, Foulds KE, Francica JR, Boyoglu-Barnum S, Werner AP,
Flach B, O'Connell S, Bock KW, Minai M, Nagata BM, Andersen H, Martinez DR,
Noe AT, Douek N, Donaldson MM, Nji NN, Alvarado GS, Edwards DK, Flebbe
DR, Lamb E, Doria-Rose NA, Lin BC, Louder MK, O'Dell S, Schmidt SD, Phung
E, Chang LA, Yap C, Todd JM, Pessaint L, Van Ry A, Browne S, Greenhouse J,
Putman-Taylor T, Strasbaugh A, Campbell TA, Cook A, Dodson A, Steingrebe K,
Shi W, Zhang Y, Abiona OM, Wang L, Pegu A, Yang ES, Leung K, Zhou T, Teng
IT, Widge A, et al. 2020. Evaluation of the mRNA-1273 Vaccine against SARS-
CoV-2 in Nonhuman Primates. N Engl J Med 383:1544-1555.
23.  Corbett KS, Nason MC, Flach B, Gagne M, O'Connell S, Johnston TS, Shah SN,
Edara VV, Floyd K, Lai L, McDanal C, Francica JR, Flynn B, Wu K, Choi A, Koch
M, Abiona OM, Werner AP, Moliva JI, Andrew SF, Donaldson MM, Fintzi J,
Flebbe DR, Lamb E, Noe AT, Nurmukhambetova ST, Provost SJ, Cook A,
Dodson A, Faudree A, Greenhouse J, Kar S, Pessaint L, Porto M, Steingrebe K,
Valentin D, Zouantcha S, Bock KW, Minai M, Nagata BM, van de Wetering R,
Boyoglu-Barnum S, Leung K, Shi W, Yang ES, Zhang Y, Todd JM, Wang L,
Alvarado GS, Andersen H, et al. 2021. Immune correlates of protection by
mRNA-1273 vaccine against SARS-CoV-2 in nonhuman primates. Science
373:eabj0299.

24.    Rogers TF, Zhao F, Huang D, Beutler N, Burns A, He WT, Limbo O, Smith C,
645           Song G, Woehl J, Yang L, Abbott RK, Callaghan S, Garcia E, Hurtado J, Parren
646           M, Peng L, Ramirez S, Ricketts J, Ricciardi MJ, Rawlings SA, Wu NC, Yuan M,
647           Smith DM, Nemazee D, Teijaro JR, Voss JE, Wilson IA, Andrabi R, Briney B,
648           Landais E, Sok D, Jardine JG, Burton DR. 2020. Isolation of potent SARS-CoV-2
649           neutralizing antibodies and protection from disease in a small animal model.
650           Science doi:10.1126/science.abc7520.
651    25.    Xiang Y, Nambulli S, Xiao Z, Liu H, Sang Z, Duprex WP, Schneidman-Duhovny
652           D, Zhang C, Shi Y. 2020. Versatile and multivalent nanobodies efficiently
653           neutralize SARS-CoV-2. Science doi:10.1126/science.abe4747.
654    26.    Estivariz CF, Bennett SD, Lickness JS, Feldstein LR, Weldon WC, Leidman E,
655           Ehlman DC, Khan MFH, Adhikari JM, Hasan M, Billah MM, Oberste MS, Alamgir
656           ASM, Flora MD. 2020. Assessment of immunity to polio among Rohingya
657           children in Cox's Bazar, Bangladesh, 2018: A cross-sectional survey. PLoS Med
658           17:e1003070.
659    27.    Louie KS, Dalel J, Reuter C, Bissett SL, Kleeman M, Ashdown-Barr L, Banwait
660           R, Godi A, Sasieni P, Beddows S. 2018. Evaluation of Dried Blood Spots and
661           Oral Fluids as Alternatives to Serum for Human Papillomavirus Antibody
662           Surveillance. mSphere 3.
663    28.    Gwyn S, Aragie S, Wittberg DM, Melo JS, Dagnew A, Hailu D, Tadesse Z, Haile
664           M, Zeru T, Nash SD, Arnold BF, Martin DL, Keenan JD. 2021. Precision of
665           Serologic Testing from Dried Blood Spots Using a Multiplex Bead Assay. Am J
666           Trop Med Hyg 105:822-827.
667    29.    Clapham H, Hay J, Routledge I, Takahashi S, Choisy M, Cummings D, Grenfell
668           B, Metcalf CJE, Mina M, Barraquer IR, Salje H, Tam CC. 2020.
669           Seroepidemiologic Study Designs for Determining SARS-COV-2 Transmission
670           and Immunity. Emerg Infect Dis 26:1978-1986.
671    30.    Roper KJ, Thomas J, Albalawi W, Maddocks E, Dobson S, Alshehri A, Barone
672           FG, Baltazar M, Semple MG, Ho A, Turtle L, Consortium IC, Paxton WA, Pollakis
673           G. 2023. Quantifying neutralising antibody responses against SARS-CoV-2 in
674           dried blood spots (DBS) and paired sera. Sci Rep 13:15014.
675    31.    Kim JS, Sun Y, Balte P, Cushman M, Boyle R, Tracy RP, Styer LM, Bell TD,
676           Anderson MR, Allen NB, Schreiner PJ, Bowler RP, Schwartz DA, Lee JS,
677           Xanthakis V, Doyle MF, Regan EA, Make BJ, Kanaya AM, Wenzel SE, Coresh J,
678           Isasi CR, Raffield LM, Elkind MSV, Howard VJ, Ortega VE, Woodruff P, Cole SA,
679           Henderson JM, Mantis NJ, Parker MM, Demmer RT, Oelsner EC. 2024.
680           Demographic and Clinical Factors Associated With SARS-CoV-2 Spike 1
681           Antibody Response Among Vaccinated US Adults: the C4R Study. Nat Commun
682           15:1492.
683    32.    Heaney CD, Hempel H, DeRosa KL, Pinto LA, Mantis NJ. 2023. Clinical
684           Assessment of SARS-CoV-2 Antibodies in Oral Fluids Following Infection and
685           Vaccination. Clin Chem doi:10.1093/clinchem/hvad169.
686
687

**FIGURE LEGENDS**

**Figure 1. Generation of contrived DBS eluates.** (**Top**) Schematic showing generation of contrived DBS (cDBS), DBS punches, and eluate collection. (**Bottom**) Summary of Panel D and Panel E samples used in this study. RT = room temperature, TBS = tris buffered saline

**Figure 2. Correlation between PRNT and MIA with cDBS eluates.** (**Top**) Workflow for the 8-plex MIA. (**Bottom**) Subset of Panel D cDBS eluates (n=15) underwent testing via 8-plex MIA and were compared to paired serum tested by PRNT. Pearson correlation of MFI values and SARS-CoV-2 neutralizing activity (%) is shown for the following antigens: (A) FLT; (B) FLS; (C) S1; (D) RBD; (E) sum spike (see text for details); and (F) N. The Pearson r value and p-values are shown as insets within each panel.

**Figure 3. Compatibility of cDBS eluates with RVP neutralizing assay. (Top)** Workflow for the RVP neutralizing assay. (**Bottom**) (A) Neutralizing titers associated with CC 12.3 with $IC_{50}$ shown as text. (B) Correlation between PRNT and RVP with matched Panel D serum samples; (C) Correlation between PRNT (% neutralization) generated with Panel D serum versus RVP generated from matched cDBS eluates; (D) Comparison of RVP neutralization of Panel D and Panel E cDBS eluates during serial dilutions. Asterisks indicate a significant difference between groups by 2-way ANOVA, where *p=0.02, **p=0.009, ***p=0.0002, ****p<0.0001. For (B) and (C), Pearson r and p-value are shown as a panel inset.

**Figure 4. Correlation between Luminex-based ACE2 inhibition and SARS-CoV-2 neutralization activity.** (**Top**) Workflow for the ACE2 inhibition (iACE2) assay. (A) Examination of iACE2 activity in Panel D cDBS eluates (n=15) versus Panel E eluates (n=7). In (A), asterisks indicate a significant difference between groups by Welch's t-test, where ***p=0.0006. (B) Correlation of iACE2 values from Panel D cDBS eluates, as shown in (A), to SARS-CoV-2 neutralizing activity (%) from paired serum analyzed

via PRNT. cDBS eluates were evaluated in triplicate. The Pearson r and p-value are

shown as a panel inset for (B). The dotted line on (B) represents the average iACE2

activity for SARS-CoV-2 negative cohort (Panel E; n=7).

**Figure 5. NAB-Sure exhibits high concordance with PRNT.** (**Top**) Workflow for the

NAB-Sure[TM] assay. (A) Pearson correlation of NAB-Sure [TM] $NT_{50}$ from authentic DBS

sample eluates compared to PRNT paired serum. DBS samples were evaluated in

duplicate. (B-D) Pearson correlation comparing NAB-Sure [TM] $NT_{50}$ for cDBS eluates and

paired serum $NT_{50}$ from PRNT for 14 Panel D samples. (B-D) were exact replicates of

each other. For (A)-(D), Pearson r and p values are shown as a panel insets.

**Figure 6. Assessment of SAR-CoV-2 neutralizing titers in cDBS eluates from**

**vaccinated individuals (Panel H). (A)** Panel H, D and E cDBS eluates were rank

ordered from high to low (left to right) based on $RBD_{WT}$ MFI determined in the 8-plex

MIA. (**B**) Correlation plot between Panel H $RBD_{WT}$ MFI and Diasorin Liaison values

provided from Access Biologics LLC. Panel H cDBS eluate activities in (C) Luminex-

based iACE2 assay, (E) RVP neutralization assay and (G ) NAB-Sure [TM] with

corresponding correlation plots shown in right panels (D, F, H). Pearson r and p values

are shown as a panel insets.

| Table 1. SARS-CoV-3 Reagents and Sources | | | | |
|---|---|---|---|---|
| **Reagent** | **Description** | **Source** | **Catalogue Number** | **Usage** |
| TBS | Tris Buffered Saline with Casein | BioRad | 1610782 | DBS Elution |
| FLT | Full Length Trimeric SARS-CoV-2 Spike | MassBiologics | N/A | 8-Plex MIA/iACE2 |
| FLS | Full length SARS-CoV-2 Spike, His-Tag | Native Antigen | REC31868-100 | 8-Plex MIA |
| Spike Subunit 1 (S1) | SARS-CoV-2 (2019-nCoV) Spike S1, His-Tag | Sino Biological | 40591-V08H | 8-Plex MIA |
| RBD | SARS-CoV-2 Stable RBD, Thrombin-His | MassBiologics | N/A | 8-Plex MIA |
| Nucleocapsid (N) | 2019-nCoV Nucleocapsid Protein, His tag | SinoBiological | 40588-V08B | 8-Plex MIA |
| Nucleocapsid (N-NA) | SARS-CoV-2 Nucleoprotein, His-Tag (E. coli) | Native Antigen | REC31812-100 | 8-Plex MIA |
| Nucleocapsid (NHT) | SARS-CoV-2 (2019-nCoV) Nucleocapsid Protein (His tag) | Sino Biological | 0588-V07E | 8-Plex MIA |
| Internal Control (IC) | Mouse anti-Human IgG3 Secondary Antibody | ThermoFisher | MA1-83242 | 8-Plex MIA |
| WT sRBD | SARS (WA1/2020) RBD, Thrombin-His | MassBiologics | N/A | iACE2 |
| hACE2 | Biotinylated ACE2 | Wadsworth Center Protein Core | N/A | iACE2 |
| CC12.3 | Human Mab RBD-A WT D614G | Scripps | N/A | iACE2/RVP |
| RVP-702L | WT D614G Strain Reporter Virus Particles | Integral Molecular | CL-275A | RVP |
| 293-hsACE2 Cells | 293T cells expressing human ACE2 receptor | Integral Molecular | TA-060520-MC | RVP |

| Table 2. Panel D Summary | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| | | | Serum (PRNT) | | MFI in 8-plex MIA by Antigen | | | | | | iACE2 | RVP | | NAb |
| # | Age | Gender | NT$_{50}$ | %Neut | N[a] | S1[b] | FLS[c] | FLT[d] | RBD[e] | Sum[f] | iACE2[g] | NT50 | %Neut | NT$_{50}$ |
| 1 | 38 | F | >832 | 93 | 6121.0 | 4915 | 10178 | 14238 | 11185 | 40516 | 43 +/- 5.3 | 379 | 89 | 712 |
| 2 | 45 | F | 104 | 51 | 3283.5 | 521 | 1514 | 2557 | 1509 | 6100 | 17 +/- 4.3 | 183 | 65 | 90 |
| 3 | 44 | F | 208 | 42 | 2925.5 | 2377 | 4880 | 6843 | 5306 | 19406 | 22 +/- 2.2 | 151 | 58 | 244 |
| 4 | 48 | F | >832 | 93 | 3396.0 | 4295 | 9836 | 15808 | 11335 | 41274 | 53 +/- 2.7 | 203 | 79 | 557 |
| 5 | 55 | F | 416 | 33 | 544.0 | 872 | 2779 | 4562 | 2868 | 11081 | 14 +/- 3.4 | 435 | 67 | 196 |
| 6 | 65 | F | 208 | 60 | 4560.0 | 3221 | 6926 | 7929 | 5972 | 24047 | 14 +/- 3.4 | 395 | 64 | 270 |
| 7 | 40 | F | 208 | 77 | 7424.0 | 1714 | 4967 | 6497 | 5496 | 18674 | 18 +/- 2.3 | 513 | 67 | 189 |
| 8 | 39 | M | 832 | 70 | 6465.0 | 4561 | 8677 | 16991 | 10578 | 40806 | 42 +/- 7.1 | 262 | 71 | 899 |
| 9 | 34 | F | IND | 30 | 1434.0 | 1147 | 2983 | 3948 | 2873 | 10951 | 17 +/- 1.8 | 239 | 68 | 249 |
| 10 | 46 | F | 208 | 40 | 11565.0 | 433 | 2941 | 9150 | 2051 | 14575 | 15 +/- 4.3 | 229 | 68 | 176 |
| 11 | 30 | M | 208 | 47 | 6989.5 | 3387 | 7843 | 9969 | 6513 | 27711 | 16 +/- 3.4 | 141 | 64 | 356 |
| 12 | 47 | F | 208 | 40 | 2468.0 | 1432 | 3145 | 5014 | 4503 | 14093 | 17 +/- 4.2 | 188 | 65 | 249 |
| 13 | 57 | F | 26 | 21 | 1455.5 | 736 | 1907 | 2948 | 2176 | 7766 | 12 +/- 1.2 | 280 | 68 | 114 |
| 14 | 36 | M | 416 | 60 | 5371.5 | 2682 | 6136 | 9272 | 5869 | 23958 | 24 +/- 5.1 | 209 | 62 | 364 |
| 15 | 30 | F | 208 | 72 | 7691.0 | 2296 | 6687 | 9218 | 5867 | 24068 | 21 +/- 2.9 | 290 | 67 | 321 |

Footnotes: a: N, SARS-CoV-2 Nucleocapsid; b: S1, SARS-CoV-2 Spike subunit 1; c: FLS, SARS-CoV-2 Full Length Spike; d: FLT, SARS-CoV-2 Full length Trimer; e: RBD, SARS-CoV-2 Receptor Binding Domain; f: Sum, represents the total of all Spike antigens ; g: iACE2 indicates inhibition of biotinylated ACE2 binding to RBD or FLT antigen complexed to MagPlex beads.  add MIA = microsphere immunoassay; MFI = median fluorescence intensity; NAb = NabSure neutralizing antibody; RVP = reporter virus assay

**Supplemental Figure Legends**

**Figure S1. Comparison of Panel D and Panel E cDBS.** The full cohort for Panel D (n=30) and Panel E (n=86) samples underwent testing via 8-plex MIA. MFI values for the following antigens are compared: N, N NA, NHT, RBD, S1, FLS and FLT. Asterisks indicate a significant difference between groups by 2-way ANOVA, where ****$p<0.0001$.

**Figure S2. Validation of Luminex-based ACE2 inhibition assay.** (A) MFI values (binding) was assessed in the context of either FLT or RBD with varying concentrations of biotinylated hACE2. (B) ACE2 inhibition titers associated with CC 12.3 in the context of FLT or RBD. $IC_{50}$ is shown as text. (C) Pearson correlation of Panel D serum analyzed by ACE2 inhibition and SARS-CoV-2 neutralizing activity (%), r-value and p-value shown as text. (D) Pearson correlation of Panel D DBS eluates analyzed by ACE2 inhibition and SARS-CoV-2 neutralizing activity (%), in the context of FLT. The r-value and p-value are shown as text. The dotted line on (D) represents the average iACE2 activity for SARS-CoV-2 negative cohort (Panel E; n=7).

**A.**

**B.**

Figure 2

## A. CC 12.3

$IC_{50}= 0.04$ ug/mL

## B. Serum

$r=0.67$
$p=0.007$

## C. DBS Eluates

$r=0.61$
$p=0.02$

## D. Panel D vs Panel E

Panel D
Panel E

**A.**

**A. ACE2**

**B. CC 12.3**

**C. Serum-RBD**

**D. FLT**