Fast volumetric fluorescence lifetime imaging of multicellular systems using single-objective light-sheet microscopy

Valentin Dunsing-Eichenauer[1]#, Johan Hummert[2]#, Claire Chardès[1], Thomas Schönau[2], Léo Guignard[1], Rémi Galland[3], Gianluca Grenci[4], Max Tillmann[2], Felix Koberling[2], Corinna Nock[2], Jean-Baptiste Sibarita[3], Virgile Viasnoff[4], Ivan Michel Antolovic[5], Rainer Erdmann[2], Pierre-François Lenne[1]

1 Aix-Marseille Université & CNRS, IBDM—UMR7288 & Turing Centre for Living Systems, Marseille, France

2 PicoQuant GmbH, Rudower Chaussee 29, 12489 Berlin, Germany

3 Univ. Bordeaux, CNRS, Interdisciplinary Institute for Neuroscience, IINS, UMR 5297, F-33000 Bordeaux, France

4 Mechanobiology Institute, National University of Singapore, Singapore

5 Pi Imaging Technology SA, EPFL Innovation Park, 1015 Lausanne, Switzerland

co-corresponding authors, co-first authors

ABSTRACT

Fluorescence lifetime imaging (FLIM) is widely used for functional and multiplexed bioimaging. The lifetime of autofluorescence or fluorescent sensors encodes physiologically relevant parameters. Thus, FLIM is especially relevant for the investigation of living systems. However, application of FLIM to live specimen is hampered by its slow speed and high phototoxicity. To enable faster and gentler FLIM, we integrated single-objective light-sheet microscopy with pulsed excitation and time-resolved detection on a novel SPAD array detector. We achieved 10-100-fold acceleration compared to confocal FLIM, down to 100 ms acquisition time per image, with excellent quantitative agreement. The massively enhanced speed enables volumetric FLIM acquisitions on live multicellular specimens, which we demonstrate with lifetime-based multiplexing in 3D and time-lapse FLIM of tension probes on living embryonic organoids. We benchmark both scanned and static light-sheet modalities to facilitate adding FLIM capability to a large variety of light-sheet microscopes.

INTRODUCTION

Fluorescence lifetime imaging is a powerful technique that enhances the information content of microscopy images beyond fluorescence intensity. Lifetime information can be utilized to multiplex imaging of different targets[1–3]. Lifetime can also encode biophysically relevant parameters such as pH, temperature, membrane tension, and metabolic activity[4–7]. These sensing capabilities make FLIM especially suited for studies on living cells and multicellular systems. However, in its typical implementation in confocal laser scanning microscopes, the applicability of FLIM to live samples is often limited by the trade-off between speed and phototoxicity. Light-sheet microscopy excels in exactly this trade-off, allowing for fast and gentle imaging.

Until recently, combining light-sheet microscopy and FLIM has been hampered by the availability of nanosecond time-resolved detectors for widefield microscopy. Previously employed detectors for light-sheet FLIM suffer from high dark currents in the case of intensified or modulated cameras[8,9], from limited throughput in the case of multichannel plate-based detectors[10], or limited fill factor and sensor size in the case of a novel SPAD array[11–13]. A promising approach for high-speed FLIM of live samples is time-gating via fast Pockels cells[14,15], but this specialized technology is not readily available. Therefore, light-sheet FLIM has not yet realized its potential to enable faster lifetime imaging than state-of-the-art laser scanning microscopes. In addition, no quantitative comparison of light-sheet and confocal FLIM going beyond analysis of average lifetime has been performed.

Here, we overcome these limitations by integrating single-objective light-sheet microscopy[16,17] (soSPIM) with a newly available SPAD array detector. The soSPIM technology utilizes 45° mirrors embedded in microfabricated imaging devices that create a light-sheet perpendicular to the optical axis and collect the excited fluorescence through the

same high numerical aperture (NA) objective used for illumination. Benefitting from this, super-resolution STORM[16] and imaging fluorescence correlation spectroscopy using soSPIM[17] have been demonstrated. Recently, it has also been shown that soSPIM facilitates high-content 3D imaging of a variety of organoid specimens[18], thanks to dedicated cell culture chips containing arrays of wells, each flanked with four micromirrors (JeWells).

The high collection efficiency of soSPIM enabled by the use of a high NA objective is particularly advantageous for lifetime imaging, significantly improving upon previous light-sheet FLIM implementations with strongly limited collection efficiency[9–12]. With soSPIM-FLIM, we achieve FLIM at unprecedented speed in 3D, enabling multiplexing by lifetime and tension sensing in live organoids.

RESULTS

To extend soSPIM to FLIM, we have equipped a soSPIM microscope with a newly developed pulsed laser and a state-of-the-art SPAD array detector. The laser is a prototype (PicoQuant GmbH) based on fiber amplification of mid-infrared laser diodes and frequency conversion[19]. The detector is a 512x512 pixel gated SPAD array (SPAD512[2], Pi Imaging Technology), enabling time-resolved imaging with a nanosecond global exposure gate that can be shifted with picosecond precision relative to the laser pulse (Fig. 1a).

As a benchmark, we performed soSPIM and confocal FLIM measurements on the same samples, AF488 dye solution and fixed embryonic organoids (gastruloids) immunostained with E-cadherin-AF488. We evaluated soSPIM-FLIM performance in two light-sheet modalities (Fig. 1a): a digital scanned light-sheet generated by fast beam scanning along micromirrors and a static light-sheet created by a cylindrical lens. For confocal measurements, we used a conventional time-correlated single photon counting (TCSPC) system limited to ca. 4 MHz photon count rate and a state-of-the-art rapidFLIM system that supports photon count rates of up to 80 MHz[20]. As a metric for FLIM performance, we quantified the average and standard deviation of lifetime histograms and photon counts in the brightest pixels, with FLIM frame acquisition times ranging from 100 ms to 1 s for light-sheet and 1 s to 2 min for confocal FLIM (Fig. 1b). For organoid measurements, only pixels within the organoid and above a global intensity threshold were considered in the comparison to exclude background pixels. We found excellent agreement between pixel lifetimes measured by confocal and soSPIM-FLIM (Fig. 1c, Fig. S1-3). However, soSPIM measurements provided a far more precise estimate within much shorter acquisition times. Even in comparison to rapidFLIM, ca. 2x narrower lifetime histograms were achieved in 2x shorter acquisition time in scanning light-sheet mode. In static mode, approximately 4x narrower histograms were achieved in 10x shorter acquisition time than in rapidFLIM. Correspondingly, approximately 1000 counts in the brightest pixel and similar histogram widths were achieved in 100 ms with static light-sheet, in 1.1 s with scanning light-sheet, and in 10.9 s with RapidFLIM (Fig. 1c). Strikingly, in comparison to rapidFLIM, high photon counts were generated with 70 and 300 times lower laser power densities for scanning and static soSPIM-FLIM respectively (Table S1).

Figure 1 – Benchmarking of soSPIM-FLIM in 3D embryonic organoids.
(a) Scheme of the FLIM modalities: In confocal FLIM (left), photon arrival times are detected in every pixel of the image via TCSPC detection. This generates the lifetime histogram. In soSPIM-FLIM, the sample is illuminated either using a digital scanned (center) or a static light-sheet (right). Photons are recorded through gated detection, in which multiple images are recorded with different gate delays. **(b)** Exemplary FLIM images of fixed embryonic organoids stained with E-cadherin-AF488, acquired with a conventional or RapidFLIM confocal microscope (top row) or soSPIM-FLIM in scanning or static light-sheet mode. Measurements were performed at 30-50 µm depth. Acquisition times are indicated. Scale bars are 50 µm. **(c)** Average and standard deviation of lifetime histograms obtained for the measurements shown in (b), pooled from all pixels within the organoid. For confocal data, an increasing number of frames were sampled from the same measurement. The reference is the intensity-weighted average lifetime from sum fits of lower power confocal measurements (Fig. S3). The inset shows average maximum photon counts in the brightest pixels (excluding the top 0.1% of pixels).

The shortest acquisition times are achievable only in static light-sheet measurements as the speed of scanning light-sheet FLIM is limited by the available beam scan rates and the potential for pile-up on the SPAD array detector. The scanning light-sheet is much more prone to pile-up due to the shorter illumination of each pixel row, which leads to local saturation of gates with the highest photon count rate in pixels that are illuminated during the passage of the beam. To account for this, we employed an adapted pile-up correction[21] to the scanning light-sheet data (see Methods and SI for details). To evaluate the pile-up effect and the dynamic range of soSPIM-FLIM, we analyzed the lifetime as a function of pixel count rate for different illumination powers (Fig. S1, S6, S7). Although the adapted correction mitigates the impact of pile-up, static light-sheet still achieves approximately 10x faster acquisition than scanning light-sheet at comparable photon counts (Fig. 1, Table S1).

The fast speed of soSPIM-FLIM makes it ideally suited for 3D acquisitions. We performed 3D FLIM measurements in organoids across an 80 µm range, with one image plane every 2 µm, illuminated for 180 ms in static mode and 1.1 s in scanning mode (Fig. 2a). While the average pixel lifetime was constant across depth, we observed a ca. 10% discrepancy between lifetimes determined from static and scanning light-sheet measurements. This is likely explained by residual pile-up in scanning light-sheet in high intensity pixels, also visible as

yellow areas in Fig. 2a. Besides much faster speed, soSPIM-FLIM benefitted from less
photon loss at deeper z planes compared to confocal FLIM, which showed darker regions
inside the organoids at 50 µm depth (Fig. S2). In contrast, soSPIM-FLIM showed lower
optical sectioning and degradation of image quality and photon counts in the more distant part
of the aggregate with respect to the illumination side (Fig. 2a, b). In light-sheet microscopy,
this is commonly circumvented by illuminating the specimen from multiple sides[22,23]. This
can be readily implemented in the JeWell architecture by illumination from two opposing
sides of the JeWell mirrors and fusion of the acquired image stacks to a final, uniform 3D
FLIM image stack (Fig. 2b).

155
Figure 2 − Volumetric (dual-illumination) soSPIM-FLIM in embryonic organoids.
(a) soSPIM-FLIM images obtained in 3D acquisitions with scanning or static light-sheet with 1.1 s (scanning) or
0.18 s (static) exposure time per plane. 3D soSPIM-FLIM stacks were acquired over a range of 80 µm with 2 µm
spacing. The graph on the right shows the average lifetime and standard deviation over the whole image as a
function of imaging depth. **(b)** Dual-illumination soSPIM-FLIM images at 40 µm depth obtained from sequential
3D acquisition from opposing JeWell mirrors at the top or bottom of the image. Image stacks were registered and
fused to obtain a final FLIM image with a more uniform image quality. Right: Average lifetime and standard
deviation as a function of image depth for both illumination sides and the fused image. Scale bars are 50 µm.
164

A powerful application of FLIM is to discriminate spectrally overlapping fluorophores by
their fluorescence lifetime, providing an additional contrast for multi-target imaging[2,24]. In
light-sheet microscopy, lifetime-based multiplexing could allow simultaneous imaging of
multiple targets that would otherwise be imaged sequentially and simplify microscope setups
requiring multiple laser lines and cameras. To test lifetime unmixing, we stained the cell
membranes of live embryonic organoids expressing nuclear H2B-GFP fusion protein with
Flipper-TR (Fig. 3a). We performed 3D soSPIM-FLIM in scanning mode. Lifetime images
obtained with soSPIM showed a clear contrast between nuclei and membrane pixels (Fig. 3b,
c, Video S1). To extract the characteristic lifetimes for both species, we selected two
populations in the phasor plot[25,26] (Fig. 3a). We then mapped selected points back to the
original image, providing masks that showed clear correspondence to either nuclei or
membrane pixels. Fitting a double-exponential model to the summed decays provided a short
and long lifetime component for each species. We performed this analysis in 3 organoids at 3
different depths (10 µm, 30 µm, 50 µm) and obtained robust lifetime estimates of 2.1±0.2 ns,
5.4±0.9 ns for nuclei (i.e. GFP) and 1.0±0.2 ns, 4.8±0.2 ns for membranes (i.e. Flipper-TR)

(Fig. 3b). These values are in good agreement with confocal lifetimes determined from FLIM
images acquired for 4 min on the same organoids by manual ROI selection and two-
component fitting (Fig. S8). The slightly lower lifetimes measured for Flipper-TR with
soSPIM-FLIM are likely a consequence of the lower excitation wavelength (i.e. 488 nm)
compared to confocal FLIM (i.e. 510 nm), as recently shown[27]. Finally, we fitted a
superposition model of both lifetime patterns to each pixel, providing amplitudes that we then
used to calculate unmixed intensity images (Fig. 3c). With this pipeline, we could well
separate nuclei and membranes, i.e. GFP and Flipper-TR signal, which is only separated by
1.2 ns in average lifetime (see SI). Lifetime unmixing worked reliably in different z planes
across 70 µm depth in multiple organoids (Fig. 3d, Video S1). We hereby benefited from the
multi-position imaging capabilities of soSPIM[18], enabling us to perform soSPIM-FLIM on
multiple organoids located in the same JeWells chip (Fig. 3d) without the need for time-
consuming re-alignment or repositioning of the beam for each organoid[18].

194
Figure 3 – Application of soSPIM-FLIM for lifetime multiplexed light-sheet imaging.
(a) Pipeline for generation of lifetime patterns for organoids expressing H2B-GFP and additionally stained with
Flipper-TR (grayscale image). To identify the two species, soSPIM data were transformed into phasor space and
points selected using elliptic cursors drawn around the extremities of the point cloud. Backmapping selected
points to the image generates two masks. **(b)** Two-component fitting was performed to extract characteristic
lifetimes for the two species. Box plots show lifetimes obtained in three different organoids at three different z
planes from measurements with acquisition times of 1 and 2 s in scanning mode. Horizontal lines indicate

reference lifetime values extracted from confocal FLIM performed on the same samples (Fig. S8). **(c)** Lifetime unmixing for H2B-GFP/ Flipper-TR organoids using lifetime patterns described in (b). Pixel-wise fitting with the determined lifetime patterns results in species amplitudes, which are used to generate unmixed intensity images (right), showing successful decomposition into nuclei and membrane signal. **(d)** Composite images obtained by unmixing of slices from 3D soSPIM-FLIM data on the samples described in (a,b), imaged across 70 μm with 2 μm spacing and 1 s acquisition time per plane. Shown are three different organoids from a multi-position acquisition on one JeWell chip (brightfield image of wells shown on the left). Scale bars are 50 μm in all images.

The speed and gentleness of soSPIM-FLIM make it an ideal tool for volumetric time-lapse FLIM on sensitive, dynamic specimens, which is strongly limited with confocal FLIM due to requirements for either high excitation powers or long acquisition times to achieve sufficient photon counts. As a proof-of-concept, we performed 3D time-lapse soSPIM-FLIM on live embryonic organoids stained with Flipper-TR, a mechanosensitive membrane tension sensor recently applied to *in vivo* specimens[28]. Due to the limited photon budget in confocal FLIM, its lifetime is usually determined globally across many cells acquired in one plane for several minutes[29]. With soSPIM-FLIM, we could robustly quantify Flipper-TR lifetimes across z-stacks with only 1-3 s acquisition time per 2D plane (Fig. 4a,b). In agreement with previous studies[29], we obtained two characteristic lifetimes of 1.0 ± 0.1 ns and 5.0 ± 0.1 ns, averaged over 3 organoids and 18 z planes each (Fig. 4b, Table S2). We repeatedly acquired z-stacks over a course of 45 min and analyzed changes in the long lifetime (τ_2) in individual, fastly rearranging cells from, on average, 90 000 photons detected per cell (ca. 15 000 per junction) (Fig. 4c, Fig. S9, Video S2). We observed a decrease in τ_2 in cells tracked over time (Fig. 4d), which is in line with a decrease in membrane tension[29]. While a decrease in lifetime would also be consistent with reduced pile-up due to a decrease in count rate, this alone cannot explain the observed change. Across multiple organoids, the long lifetime only varies by approximately 0.2 ns in the intensity range of the tracked cells over time (Fig. 4b-d). In addition, the initial decrease and later increase of the detected intensity do not map onto changes in the long lifetime (Fig. 4d). In contrast, the observed lifetime change might be a consequence of illumination-induced photosensitization of Flipper-TR, i.e. formation of reactive oxygen species and lipid oxidation which change the properties of the plasma membrane upon continuous illumination, as recently reported[30].

234
Figure 4 – Application of soSPIM-FLIM for live volumetric time-lapse membrane tension imaging in
organoids. (a) FLIM images of Flipper-TR stained live organoids at a depth of 30 μm. The color scale shows the
long lifetime from a two component fit with the short lifetime fixed (see methods). **(b)** Fit results from two
component sum fits at the start of the time-lapse acquisition. The intensity in each z-plane at 10-40 μm depth
was thresholded to select membrane areas and decays fitted with two free lifetimes. Each point is the result from
one z-plane, with averages and standard deviations for each organoid shown as errorbars. **(c)** 3D time-lapse
FLIM imaging of organoid 2 (a) was performed over ca. 45 min in scanning light-sheet mode, with one z-stack
of 60 μm range and 2 μm spacing acquired every 1.5 min. Lifetime images on the left show the long lifetime at
20 μm and 40 μm depth at 0 min (see methods). Cropped images show two individual cells (arrows in the
images on the left) that were segmented and tracked over time. Intensity is displayed in grey and long lifetime in
color code along the cell contour. **(d)** The Flipper-TR lifetime of the tracked cells was determined from the
summed decay of all segmented membrane pixels. Average intensity in these pixels is shown in the plot at the
bottom. The analysis was restricted to a single 2D plane per cell. Scale bars are 50 μm in all images showing
entire organoids and 10 μm in the cropped images in (c).

DISCUSSION

In this work, we have realized fast, optically sectioned wide-field fluorescence lifetime
imaging by integration of pulsed excitation and time-resolved detection into single-objective
light-sheet microscopy. The soSPIM-FLIM technique achieves 10-100x speed advantage
compared to state-of-the-art confocal FLIM, with excellent quantitative agreement of
measured lifetimes. We have demonstrated volumetric lifetime multiplexed light-sheet
imaging and volumetric tension imaging on individual cells, opening up new possibilities for
functional imaging on living samples such as embryos and organoids.

We have benchmarked soSPIM-FLIM in both static and digital scanned light-sheet
modalities, making the results transferable to a broad range of light-sheet systems. Even
though speed and dynamic range of FLIM with scanned illumination is limited by detector
pile-up at currently available scan rates, it provided robust lifetime multiplexed imaging in
3D. However, our benchmark reveals that static light-sheet illumination has clear advantages
over scanned illumination when combined with the time-gated detector architecture employed
here.

While such considerations regarding the interplay between illumination and sensor
architecture are required, the simplicity of adding pulsed excitation and a time-resolved
camera to a light-sheet microscope will enable FLIM in new contexts in the future. In
conjunction with our recently presented high-throughput 3D imaging pipeline[18], soSPIM-
FLIM can be automatized to achieve high-content 3D FLIM imaging, at a throughput of about
50 organoids per hour. While this work has focused on the soSPIM geometry, i.e. a particular
light-sheet architecture employing 45° reflective mirrors and a single objective for
illumination and detection, our implementation is fully compatible with other, e.g. multi-
objective, light-sheet configurations[22,31–33], provided sufficient light collection efficiency can
be achieved. The compatibility with multi-view light-sheet systems allows adapting light-
sheet FLIM to bigger specimens, a potential illustrated by our dual-illumination soSPIM-
FLIM.

Notably, light-sheet FLIM operates at orders of magnitude lower peak illumination powers
than confocal FLIM. We took advantage of this with proof-of-concept 3D time-lapse
membrane tension imaging on single cells in living embryonic organoids. Interestingly, we
observed a lifetime decrease over time, which we attribute to changes of the plasma
membrane environment, induced by repetitive illumination of the Flipper probe[30]. In the
context of this and other live imaging applications of light-sheet FLIM, future work will be
needed to evaluate viability of specimen exposed to prolonged pulsed illumination for
different samples, fluorophores, and experimental conditions.

Overall, light-sheet FLIM not only massively accelerates FLIM imaging, but also provides
new quantitative readouts for light-sheet microscopy beyond fluorescence intensities, to
generate 3D maps of e.g. protein-protein interactions via FLIM-FRET[34,35], metabolic
gradients via lifetime-based label-free imaging[36], or concentrations of, for instance, oxygen or
calcium via FLIM-based sensors[37,38]. Applied to living organisms and the growing number of
organoid models, this approach could thus become a crucial tool for advancing our
understanding of physiology in both health and disease.

MATERIALS AND METHODS

Microscope setups

soSPIM-FLIM microscope

Light-sheet imaging was performed on a custom-made soSPIM microscope built around a Zeiss Axiovert 200 microscope body (Carl Zeiss) following the previously published design[16]. It consists of a custom laser scanning unit, dedicated microfabricated devices[18] containing an array of truncated pyramidal-shaped microwells flanked with 45° mirrors (JeWells), and a soSPIM acquisition module operating in MetaMorph (Molecular Devices). The microscope is equipped with an LED lamp (CooLED, Andover) for brightfield imaging, continuous-wave excitation lasers at 488, 514 and 561 nm and a sCMOS camera (Hamamatsu ORCA Flash 4.0 V3). The sample is illuminated and fluorescence is detected through a 40x 1.2 NA Plan Apochromat water-immersion objective (Carl Zeiss). The light-sheet was generated by either scanning the laser beam laterally along the JeWell mirrors or inserting a cylindrical lens (f=150 mm, Thorlabs) into the optical path and aligning it according to the JeWell mirrors axis[16,17]. 3D imaging is achieved by moving the beam perpendicular to the mirror axis through a pair of galvanometric mirrors (Pangolin SCANMAX 506 actuators) while translating the objective axially using a piezo-driven focus controller (Physik Instrumente). The resulting displacement of the light-sheet waist is compensated by shifting it using an electrical tunable lens (Custom EL-30-10, Optotune). The light-sheet position (axial position and waist) is synchronized with the focal plane of the objective thanks to the soSPIM acquisition module, which ensures optimal 3D sectioning. The dimensions of the light-sheet were controlled by adjusting the laser beam diameter using a tunable diaphragm placed after the fiber output of the coupled excitation lasers. The resulting dimensions for scanning and static light-sheet were ca. 1.5 μm/ 1.9 μm waist and ca. 50 μm/ 80 μm Rayleigh length. The width of the static light-sheet was ca 110 μm.

For FLIM detection, a 512x512 pixel SPAD array detector (SPAD512[2], Pi Imaging Technology) was mounted on the second port of the microscope. The image plane of the detector was aligned with the sCMOS camera and centered on the same field of view. The effective field-of-view of the SPAD array is 210x210 μm^2 with a pixel size of 410 nm. For FLIM excitation, a high power pulsed 488 nm laser prototype[19] (PicoQuant GmbH) was fiber coupled into the optical path. Time gating of the SPAD array detector was triggered by the laser via a NIM to TTL converter (TMF400, PicoQuant GmbH). For 3D FLIM imaging, image acquisition on the SPAD array was synchronized with soSPIM illumination using a TTL trigger signal. To achieve dual-illumination imaging from two sides, two opposing mirrors of a JeWell were illuminated sequentially. The sample was shifted in between to fit it into the field of view of the SPAD array detector. For each side, the waist of the light-sheet was positioned at a distance of about one quarter of the specimen's thickness from the reflecting mirror. To image multiple organoids in different positions, the previously described tools for JeWell chip preview and positioning of individual JeWells were employed using bright field imaging[18]. For live imaging, the JeWell chip was placed into an incubation chamber (Okolab), which was mounted on top of the microscope stage, heated up to 37°C and provided with 5% CO_2 and 85% humidity.

Confocal FLIM microscope

FLIM measurements were performed on a Zeiss LSM880 system (Carl Zeiss) equipped with a time-resolved LSM upgrade (PicoQuant GmbH) using a Plan-Apochromat 40x, 1.2 NA water immersion objective. Images of 512x512 pixels per frame were acquired after excitation with a 510 nm pulsed laser diode operating at 20 MHz repetition rate. Fluorescence was split using a 560 nm dichroic mirror and detected by two PMA Hybrid-40 detectors (PicoQuant) after passing 550/49 nm and 600/50 nm bandpass filters, respectively. The pinhole was set to 1 airy unit. In each measurement, images were accumulated over a time period of 2 min, with a pixel dwell time of 1.23 µs and frame time of 0.6 s.

Rapid FLIM confocal microscope

Rapid FLIM Measurements were performed on a MicroTime 200 confocal microscope (PicoQuant GmbH) equipped with a LDH-D-C-485 picosecond excitation laser, a FLIMBee beam scanner, a PMA Hybrid-40 detector, and a MultiHarp 150 TCSPC unit[39] (all PicoQuant GmbH). RapidFLIM images of fixed organoids stained with E-cadherin-AF488 were acquired with a 60x 1.2 NA water immersion objective (UPLSAPO60XW, Olympus), with 485nm pulsed excitation at a 40 MHz laser repetition rate with 512x512 pixels, 500 nm pixel size, and a dwell time of 1.3 µs.

Data acquisition

Light-sheet FLIM data acquisition was performed with the Pi Imaging software in streaming mode, controlled by self-written Python scripts. Acquisition parameters are listed in Table S2. The light-sheet microscope was controlled via MetaMorph (Molecular Devices). Confocal microscopes were controlled with ZEN (Carl Zeiss) and SymPhoTime64 (PicoQuant) in the case of the LSM880 and with SymPhoTime64 in the case of the MicroTime 200.

Data Analysis

Instrument response function

The combined instrument response function (IRF) of the SPAD array and the laser was characterized by detecting reflected laser light from a mirror placed on the sample holder. The measured IRF is parametrized with an approximated function, given by integration of a Gaussian laser IRF with an infinitely sharp gate edge:

$$IRF(t) = \sqrt{\frac{\pi}{2}} \sigma \left[\text{erf}\left(\frac{t - t_0 + w}{\sqrt{2}\sigma}\right) - \text{erf}\left(\frac{t - t_0}{\sqrt{2}\sigma}\right) \right]$$

with the gate width w, the width of the laser IRF σ, and the gate delay offset t_0 as free parameters. With this IRF model the sigma parameter compensates for the finite edge steepness of the SPAD array gating function, and is a measure for the combined edge steepness given by the laser pulse width and the detector gating function. By parametrizing

the IRF, relative delays between the IRF and light-sheet measurements can be compensated
and the high demands for experimental IRF data quality are relaxed (Fig. S10).

Background and pile-up correction

To limit the impact of dark counts, background measurements (Fig. S10) were taken daily
with the same settings as the fluorescence measurements and subtracted from measured data.
To account for detector saturation, we employed an adapted pile-up correction[40]. The
corrected photon number in 8-bit images is calculated from the measured photon number as:

$$N_{corr} = -ln\left(1 - N * F/255\right) * 255/F$$

with $F=1$ for static light-sheet measurements and $F=3.5$ for scanning light-sheet
measurements. Pile-up correction was applied after background correction. A detailed
discussion of pile-up correction can be found in the SI.

Phasor analysis

Phasor plots were generated using a discretized phasor transformation[26]. From the detected
lifetime decay $N_{m,ij}$ in each pixel ($N_{m,ij}$: photon count in gate delay m in pixel ij), the phasor
coordinates were calculated as follows:

$$G_{ij} = \frac{\sum_{m=0}^{k-1} N_{m,ij} cos\left(\omega\left(m+\frac{1}{2}\right)\Delta T\right)}{\sum_{m=0}^{k-1} N_{m,ij}},$$

$$S_{ij} = \frac{\sum_{m=0}^{k-1} N_{m,ij} sin\left(\omega\left(m+\frac{1}{2}\right)\Delta T\right)}{\sum_{m=0}^{k-1} N_{m,ij}},$$

where $\omega = 2\pi/T$, with the total detection time T, the gate width Δt and k the number of gate
delays. To account for the IRF, the total phasor and the phasor of the IRF were calculated in
complex space and the IRF phasor divided from the total phasor in each pixel. Afterwards, the
real and imaginary parts G_{ij} and S_{ij} were determined.

Fitting

Time-gated FLIM signals are fitted with an approximate model function. This is based on an
analytic approximation function for the convolution of a Gaussian pulse with an exponential
decay, according to[41]:

$$f(t) = \frac{1}{2}\exp\left(\frac{\sigma^2}{2\tau^2}\right)\left(1 - erf\left(\frac{\sigma}{\sqrt{2}\tau} - \frac{t}{\sqrt{2}\sigma}\right)\right)\exp\left(-\frac{t}{\tau}\right)$$
,

403 where σ is the width of the gaussian pulse and τ is the monoexponential decay time. The
404 model function for decays obtained by time-gated detection is then obtained by integrating the
405 decay function with infinitely sharp gate edges, analogous to the model function for the IRF:

406 $$f_{Gated}(t) = A\frac{\tau}{2}\exp\left(\frac{\sigma^2}{2\tau^2}\right)\left[\left(erf\left(\frac{\sigma}{\sqrt{2}\tau} - \frac{t-t_0+w}{\sqrt{2}\sigma}\right) - 1\right)\exp\left(-\frac{t-t_0+w}{\tau}\right) + \exp\left(-\frac{\sigma^2}{2\tau^2}\right)erf\left(\frac{t-t_0+w}{\sqrt{2}\sigma}\right)\right]$$
$$- A\frac{\tau}{2}\exp\left(\frac{\sigma^2}{2\tau^2}\right)\left[\left(erf\left(\frac{\sigma}{\sqrt{2}\tau} - \frac{t-t_0}{\sqrt{2}\sigma}\right) - 1\right)\exp\left(-\frac{t-t_0}{\tau}\right) + \exp\left(-\frac{\sigma^2}{2\tau^2}\right)erf\left(\frac{t-t_0}{\sqrt{2}\sigma}\right)\right]$$
,

where A is the decay amplitude, t_0 can express a time offset between the gate delays and the
laser pulse and w is the width of the approximated gate. The use of this analytic model
function without explicit convolutions not only speeds up pixel-wise fitting, but also
simplifies the use of the parametrized IRF and thus the compensation of IRF variations across
the detector area (see SI). The IRF parameters σ and w were held fixed at the extracted pixel
values from the IRF measurement for all fits. Maximum likelihood estimation of model
parameters from experimental decay signals was performed with the L-BFGS-B
implementation within the scipy optimize package.

Pattern Matching

To decompose FLIM images on organoids containing two fluorophore species, i.e. GFP and
Flipper-TR, phasor plots of FLIM images were generated in three different z planes (i.e. at 10,
30 and 50 μm depth) for three different organoids. In each phasor plot, two elliptic cursors
were drawn around the extremities of the point cloud and selected points mapped back to the
image to define pixel masks for individual fluorophore populations. Lifetime decays were
then summed across all corresponding pixels and a double-exponential decay model fitted to
the overall decay for each species. Averaging over all organoids and z depths, this procedure
yielded two characteristic lifetimes and respective amplitudes for each species, defining their
lifetime patterns. Next, a superposition model, defined as the sum of the two lifetime patterns
(i.e. two double exponential decay functions, each with fixed decay times and amplitudes)
was fitted pixel-wise to the lifetime decay in each pixel of the 3D FLIM stack with species
amplitudes as free fit parameters. Final intensity images for each species were generated by
weighing the total pixel intensity in the original image with the determined species
amplitudes. Before intensity weighting, the amplitudes were smoothed with a gaussian kernel
(sigma=1 pixel) to reduce pixel-to-pixel variations.

Dual-illumination registration and fusion

Image stacks acquired from two opposite illumination sides were first registered in 3D using
the Blockmatching algorithm[42] based on a 3D translation. In the registered image stacks,
aggregates were segmented in a maximum intensity projection after Gaussian blurring with a
sigma of 10 pixels, mean thresholding and dilation of the obtained mask with disk of radius
10. A bounding box surrounding the union of the two segmentation masks was then calculated
and fusion weights $w_{1,2}$ defined using a sigmoid function,

438
$$w_{1,2}(x) = \frac{1}{e^{\pm \frac{x-l/2}{l/8}}+1},$$

439   centered around the mid-line of the bounding box, where l is the width of the bounding box, x
440   the spatial coordinate along the light propagation axis. Image and lifetime stacks were then
441   fused using a weighted sum[43,44],

442
$$I(x,y,z) = \frac{w_1(x)I_1(x,y,z)+w_2(x)I_2(x,y,z)}{w_1(x)+w_2(x)},$$

443
$$\tau(x,y,z) = \frac{w_1(x)\tau_1(x,y,z)+w_2(x)\tau_2(x,y,z)}{w_1(x)+w_2(x)},$$

where $I_{1,2}$ and $\tau_{1,2}$ are the intensity and lifetime stacks acquired from each illumination side
and *(x,y,z)* denote the spatial coordinate of each image voxel.

Flipper-TR time-lapse analysis and cell segmentation

For Flipper-TR measurements, FLIM analysis was performed in two iterative rounds of fitting
with a two-component decay model. First, each 2D plane was intensity thresholded, using
pixels between the 50th and 95th intensity percentile to select in focus areas and exclude very
bright areas to reduce the impact of pile-up. Detected lifetime decays of remaining pixels
were summed and a double-exponential decay model fitted to the overall decay, with both
lifetime components and the amplitude ratio as free fit parameters. Then, the average short
lifetime (1.0 ns) and amplitude ratio (0.45) from all z planes across 3 organoids was
determined (Fig. 4). For all subsequent analysis of changes in the long lifetime (τ_2), the short
lifetime and amplitude ratio were fixed to these values. For individual cell lifetime analysis of
3D time-lapse Flipper-TR data, cells of interest were selected and, for each time point and
cell, the image containing the cell at its central z-position extracted from the z-stack. Cell
membranes were segmented in these images in 2D, using the Tissue Analyzer plugin[45] in
Fiji[46]. Label images containing a separate label for each segmented junction were manually
curated in Napari[47] and junctions belonging to the same cell pooled to obtain segmentation
masks for each time point. Subsequently, cells masks were manually annotated in time.
Finally, masks were dilated with a disk kernel of radius of one pixel to account for the width
of the cell outlines in the images.

Confocal FLIM

All confocal FLIM data were analyzed using the SymPhoTime64 software (PicoQuant
GmbH, Berlin, Germany). For pixelwise single exponent estimation (Ecad-AF488) the fast
lifetime algorithm is used, corresponding to center of mass analysis. For reference lifetimes of
single fluorophore data, bi-exponential re-convolution fitting of sum signals at moderate
intensity is performed (Fig. S3). For reference lifetimes of dual-labelled samples, decay
patterns were generated from manually selected ROIs (Fig. S8) of nuclei and membranes.
Sum decays of the ROIs were fitted with a bi-exponential re-convolution fit to obtain the
reference lifetimes given in Fig. 3.

Sample preparation

Solution measurements

Alexa Fluor 488 (AF-488) NHS Ester (ThermoFisher Scientific) was dissolved in PBS at
200nM and pipetted onto the soSPIM chip for imaging.

Gastruloid culture, staining and mounting

A complete description of the culture conditions and the protocol for making gastruloids is
presented in Baillie-Johnson et al., 2015[48]. In the experiments presented here, gastruloids
were generated from E14Tg2a.4 or H2B-GFP mouse embryonic stem cells (mESC)
(MMRRC, University of California Davis, US) using an initial cell number of 100, and

cultured according to the modified protocol described in Hashmi et al., 2022[49]. After 2 days of
culture, gastruloids were incubated for 1 h in culture medium supplemented with 1 µM
Flipper-TR and manually transferred to JeWell chips with 120 or 170 µm well size (top
opening). E14 gastruloids were fixed using 4% paraformaldehyde and immunostained for E-
cadherin using rat primary antibody and anti-rat-AF-488 secondary antibody, following
published protocols[49]. JeWell culture chips were rinsed with 96% ethanol, passivated with
0.2% lipidure solution and rinsed in PBS after evaporation. To remove air bubbles, chips were
evacuated in a vacuum chamber for 1 h. Afterwards, chips were sterilized under UV
illumination for 30 min and medium was exchanged to cell culture medium using multiple
washes. Live or fixed gastruloids were mounted manually by pipetting under the occular
(source) just before imaging. For RapidFLIM measurements, fixed gastruloids were placed in
a 35mm glass-bottom dish (#1.5, MatTek Corp., Ashland, MA).

AUTHOR INFORMATION

Corresponding Authors

Valentin Dunsing-Eichenauer, valentin.dunsing@univ-amu.fr

Johan Hummert, hummert@picoquant.com

Author Contributions

V.D.-E., J.H. conceived the project idea, coordinated the project, performed all experiments,
wrote the analysis code, analyzed and visualized all data, co-wrote the manuscript. V.D.-E.,
J.H. and C.C. aligned the optical setup. V.D., J.H. and P.-F.L. discussed the results. T.S. and
C.N. designed and built the 488 nm laser prototype. L.G. provided software. R.G. provided
assistance with the optical setup. G.G. conceived and fabricated the JeWell chips. M.T.
assisted with data analysis. R.G., G.G., J.-B.S. and V.V. developed the soSPIM technology.
I.M.A. developed the SPAD array detector. R.E. and P.-F.L. provided funding and equipment.
All authors co-edited the manuscript.

ACKNOWLEDGEMENT

We thank Agathe Rostan (IBDM Marseille) for help with sample preparation, Frank Schnorrer
(IBDM Marseille) for access to the confocal FLIM system. We thank Daniel Sapede and
Brice Detailleur from the IBDM imaging facility for technical assistance. V.D.-E.
acknowledges support by an HFSP long-term postdoctoral fellowship (HFSP LT0058/2022-
L). This work is supported by the French National Research Agency ("Adgastrulo" project
ANR-19-CE13-0022) and the Fondation de la Recherche Médicale (to P.-F.L.

EQU202003010407). We also acknowledge the France-Bioimaging Infrastructure (ANR-10-
INBS-04). J.H. thanks the Federal Ministry of Education and Research (BMBF) for their
financial support of this work in project LIVE2QMIC/13N15953 as part of the program
"Enabling Technologies für die Quantentechnologien". J.-B.S. and R.G. acknowledge the
ANR-22-CE42 DEEPHEPATOSCREEN and the University of Bordeaux's IdEx "Investments
for the Future" program (GPR BRAIN_2030).

CONFLICT OF INTEREST

529 J.H., T.S., M.T., F.K., and C.N. are employed at PicoQuant GmbH, R.E. is the CEO and co-
530 founder of PicoQuant GmbH, I.-M.A. is the CEO and co-founder of Pi Imaging Technology.

REFERENCES

1. Niehörster, T. *et al.* Multi-target spectrally resolved fluorescence lifetime imaging
microscopy. *Nat. Methods* **13**, 257–262 (2016).

2. Frei, M. S. *et al.* Engineered HaloTag variants for fluorescence lifetime multiplexing.
Nat. Methods 2021 191 **19**, 65–70 (2021).

3. Starling, T. *et al.* Multicolor lifetime imaging and its application to HIV-1 uptake. *Nat.*
Commun. 2023 141 **14**, 1–17 (2023).

4. Colom, A. *et al.* A fluorescent membrane tension probe. *Nat. Chem. 2018 1011* **10**,
1118–1125 (2018).

5. Barroso, M., Monaghan, M. G., Niesner, R. & Dmitriev, R. I. Probing organoid
metabolism using fluorescence lifetime imaging microscopy (FLIM): The next frontier
of drug discovery and disease understanding. *Adv. Drug Deliv. Rev.* **201**, 115081
(2023).

6. Okabe, K. *et al.* Intracellular temperature mapping with a fluorescent polymeric
thermometer and fluorescence lifetime imaging microscopy. *Nat. Commun. 2012 31* **3**,
1–9 (2012).

7. Shimolina, L. *et al.* Fluorescence lifetime-based pH mapping of tumors in vivo using
genetically encoded sensor SypHerRed. *Biophys. J.* **121**, 1156–1165 (2022).

8. Liu, X. *et al.* Fast fluorescence lifetime imaging techniques: A review on challenge and
development. *J. Innov. Opt. Health Sci.* **12**, 12 (2019).

9. Mitchell, C. A. *et al.* Functional in vivo imaging using fluorescence lifetime light-sheet
microscopy. *Opt. Lett. Vol. 42, Issue 7, pp. 1269-1272* **42**, 1269–1272 (2017).

10. Hirvonen, L. M. *et al.* Lightsheet fluorescence lifetime imaging microscopy with wide-

field time-correlated single photon counting. *J. Biophotonics* **13**, e201960099 (2020).

11. Nutt, K. J. *et al.* High-efficiency digitally scanned light-sheet fluorescence lifetime
microscopy   (DSLM-FLIM).   *bioRxiv*   2023.06.02.543377   (2023).
doi:10.1101/2023.06.02.543377

12. Samimi, K. *et al.* Light-sheet autofluorescence lifetime imaging with a single-photon
avalanche diode array. *https://doi.org/10.1117/1.JBO.28.6.066502* **28**, 066502 (2023).

13. Nedbal, J. *et al.* A time-correlated single photon counting SPAD array camera with a
bespoke data-processing algorithm for lightsheet fluorescence lifetime imaging (FLIM)
and FLIM videos. *Sci. Reports 2024 141* **14**, 1–17 (2024).

14. Bowman, A. J., Huang, C., Schnitzer, M. J. & Kasevich, M. A. Wide-field fluorescence
lifetime imaging of neuron spiking and subthreshold activity in vivo. *Science (80-.).*
380, 1270–1275 (2023).

15. Bowman, A. J., Klopfer, B. B., Juffmann, T. & Kasevich, M. A. Electro-optic imaging
enables efficient wide-field fluorescence lifetime microscopy. *Nat. Commun. 2019 101*
10, 1–8 (2019).

16. Galland, R. *et al.* 3D high-and super-resolution imaging using single-objective SPIM.
Nat. Methods **12**, 641–644 (2015).

17. Singh, A. P. *et al.* 3D Protein Dynamics in the Cell Nucleus. *Biophys. J.* **112**, 133–142
(2017).

18. Beghin, A. *et al.* Automated high-speed 3D imaging of organoid cultures with multi-
scale phenotypic quantification. *Nat. Methods 2022* 1–12 (2022). doi:10.1038/s41592-
022-01508-0

19. PicoQuant GmbH. High-power picosecond pulses at 488 nm. Available at:
https://www.picoquant.com/scientific/product-studies/high-power-picosecond-pulses-
at-488-nm. (Accessed: 18th March 2024)

20. Loidolt-Krüger, M. *et al.* Visualize dynamic processes with rapidFLIMHiRes, the ultra
fast FLIM imaging method with outstanding 10 ps time resolution. Available at:
https://www.picoquant.com/dl_technotes/AppNote_rapidFLIM_HiRes.pdf. (Accessed:
19th March 2024)

21. Antolovic, I. M., Burri, S., Bruschini, C., Hoebe, R. & Charbon, E. Nonuniformity
analysis of a 65-kpixel CMOS SPAD imager. *IEEE Trans. Electron Devices* **63**, 57–64
(2016).

22. Tomer, R., Khairy, K., Amat, F. & Keller, P. J. Quantitative high-speed imaging of
entire developing embryos with simultaneous multiview light-sheet microscopy. *Nature*
Methods **9**, 755–763 (2012).

23. Krzic, U., Gunther, S., Saunders, T. E., Streichan, S. J. & Hufnagel, L. Multiview light-
sheet microscope for rapid in toto imaging. *Nat. Methods* **9**, 730–733 (2012).

24. Scipioni, L., Rossetta, A., Tedeschi, G. & Gratton, E. Phasor S-FLIM: a new paradigm
for fast and robust spectral fluorescence lifetime imaging. *Nat. Methods 2021 185* **18**,

542–550 (2021).

25. Digman, M. A., Caiolfa, V. R., Zamai, M. & Gratton, E. The phasor approach to fluorescence lifetime imaging analysis. *Biophys. J.* **94**, L14–L16 (2008).

26. Fereidouni, F., Esposito, A., Blab, G. A. & Gerritsen, H. C. A modified phasor approach for analyzing time-gated fluorescence lifetime images. *J. Microsc.* **244**, 248–258 (2011).

27. Ragaller, F. *et al.* Quantifying Fluorescence Lifetime Responsiveness of Environment-Sensitive Probes for Membrane Fluidity Measurements. *J. Phys. Chem. B* **128**, 50 (2023).

28. Royer, C. *et al.* ASPP2 maintains the integrity of mechanically stressed pseudostratified epithelia during morphogenesis. *Nat. Commun. 2022 131* **13**, 1–19 (2022).

29. Roffay, C. *et al.* Tutorial: fluorescence lifetime microscopy of membrane mechanosensitive Flipper probes. *Nat. Protoc. 2024* 1–13 (2024). doi:10.1038/s41596-024-01027-6

30. Torra, J., Campelo, F. & Garcia-Parajo, M. F. Tensing Flipper: Photosensitized manipulation of membrane tension, lipid phase separation and raft protein sorting in biological membranes. *bioRxiv* 2024.06.25.599907 (2024). doi:10.1101/2024.06.25.599907

31. Moos, F. *et al.* Open-top multisample dual-view light-sheet microscope for live imaging of large multicellular systems. *Nat. Methods* **21**, 798–803 (2024).

32. McDole, K. *et al.* In Toto Imaging and Reconstruction of Post-Implantation Mouse Development at the Single-Cell Level. *Cell* **175**, 859-876.e33 (2018).

33. Sapoznik, E. *et al.* A versatile oblique plane microscope for large-scale and high-resolution imaging of subcellular dynamics. *Elife* **9**, 1–39 (2020).

34. Margineanu, A. *et al.* Screening for protein-protein interactions using Förster resonance energy transfer (FRET) and fluorescence lifetime imaging microscopy (FLIM). *Sci. Reports 2016 61* **6**, 1–17 (2016).

35. Long, Y. *et al.* In vivo FRET–FLIM reveals cell-type-specific protein interactions in Arabidopsis roots. *Nat. 2017 5487665* **548**, 97–102 (2017).

36. Stringari, C. *et al.* Phasor approach to fluorescence lifetime microscopy distinguishes different metabolic states of germ cells in a live tissue. *Proc. Natl. Acad. Sci. U. S. A.* **108**, 13582–13587 (2011).

37. Penjweini, R. *et al.* Single cell-based fluorescence lifetime imaging of intracellular oxygenation and metabolism. *Redox Biol.* **34**, (2020).

38. van der Linden, F. H. *et al.* A turquoise fluorescence lifetime-based biosensor for quantitative imaging of intracellular calcium. *Nat. Commun. 2021 121* **12**, 1–13 (2021).

39. Wahl, M. *et al.* Photon arrival time tagging with many channels, sub-nanosecond deadtime, very high throughput, and fiber optic remote synchronization. *Rev. Sci.*

Instrum. **91**, (2020).

40. Pi Imaging. SPAD512^2 system manual. Available at: https://piimaging.com/doc-spad512s. (Accessed: 18th March 2024)

41. Zewail, S. P. A. H. Femtosecond real time probing of reactions XXII Kinetic description of probe absorption fluorescence depletion and mass spectrometry. *Mol. Phys.* **89**, 1455–1502 (1996).

42. Ourselin, S., Roche, A., Prima, S. & Ayache, N. Block matching: A general framework to improve robustness of rigid registration of medical images. *Lect. Notes Comput. Sci. (including Subser. Lect. Notes Artif. Intell. Lect. Notes Bioinformatics)* **1935**, 557–566 (2000).

43. Medeiros, G. De *et al.* Confocal multiview light-sheet microscopy. *Nat. Commun.* **6**, (2015).

44. Guignard, L. *et al.* Contact area-dependent cell communication and the morphological invariance of ascidian embryogenesis. *Science.* **369**, (2020).

45. Aigouy, B., Umetsu, D. & Eaton, S. Segmentation and Quantitative Analysis of Epithelial Tissues. *Methods Mol. Biol.* **1478**, 227–239 (2016).

46. Schindelin, J. *et al.* Fiji: an open-source platform for biological-image analysis. *Nat. Methods* **9**, 676–682 (2012).

47. Ahlers, J. *et al.* napari: a multi-dimensional image viewer for Python. (2023). doi:10.5281/zenodo.8115575

48. Baillie-Johnson, P., van den Brink, S. C., Balayo, T., Turner, D. A. & Arias, A. M. Generation of Aggregates of Mouse Embryonic Stem Cells that Show Symmetry Breaking, Polarization and Emergent Collective Behaviour In Vitro. *JoVE (Journal Vis. Exp.* **2015**, e53252 (2015).

49. Hashmi, A. *et al.* Cell-state transitions and collective cell movement generate an endoderm-like region in gastruloids. *Elife* **11**, (2022).