Quantification of tissue creatine content using capillary electrophoresis

Y. Jane Choi[1,2*], Angelo P Bautista[3], Jessica R Terrill[3], J. Jane Pillow[1,2], Peter G Arthur[3].

[1] School of Human Sciences, University of Western Australia, Perth, WA, Australia

[2] Wal-yan Respiratory Research Centre, Telethon Kids Institute, Perth, WA, Australia

[3] School of Molecular Sciences, University of Western Australia, Perth, WA, Australia

*Corresponding author
Email: jane.choi@uwa.edu.au (YJC)

Metadata

Funding

This work was supported by the Australian National Health & Medical Research Council (GNT1196188; JJP) and Stan Perron Foundation Post-PhD Career Launching Award (YJC). YJC and AB were supported with an Australian Government Research Training Program (RTP) scholarship.

Competing interests

The authors have no conflicts of interest to declare.

Data availability

Data will be made available upon request.

Associated content

DOI: dx.doi.org/10.17504/protocols.io.4r3l22j23l1y/v2

28 **Abstract**

Creatine plays a fundamental role in cellular energy homeostasis. The current protocol
describes an alternative method for creatine quantification in biological tissue samples using
capillary electrophoresis, with high separation efficiency that enables rapid analysis with low
sample volumes. The protocol involves homogenization of snap-frozen tissue in phosphate
buffer, followed by electrophoresis through a bare-fused capillary (75 µm internal diameter)
and measurement at 200 nm on the Agilent 7100 CE system. Under the optimised conditions,
there was excellent linearity in creatine standards between 6.3 – 100 µM. The overall intra-
assay variability for concentrations between 6.3 – 100 µM was 1.5 %, and the inter-assay
variability was 6.4 %, with a limit of detection at 6 nmol/mg protein. The protocol was
further benchmarked against a commercially available enzyme assay kit using lung samples
from lambs that received continuous creatine or saline supplementation. There was good
agreement between the two methods (mean difference = 0.42 [-0.26-1.1] nmol/mg protein).
Importantly, capillary electrophoresis enables reliable detection of creatine in biological
samples from just ~1.5 mg of wet-weight lung tissue. Capillary electrophoresis enables rapid
(<10 minutes) and highly efficient analysis of tissue samples and avoids challenges faced
with traditional enzymatic assays. The current protocol was developed and optimised with
ovine lung tissue, but it can be easily adapted to analyse various tissue types. For tissues with
higher baseline creatine content, such as the skeletal muscles or brain, <1 mg wet weight
tissue would be sufficient to detect creatine using capillary electrophoresis.

1 Introduction

Creatine is a popular nutritional supplement with a fundamental role in energy homeostasis
[1]. The creatine (Cr)/phosphocreatine (PCr) system is intrinsically involved in cellular
energy homeostasis by converting PCr to Cr, which yields a high-energy phosphate group
used to convert adenosine diphosphate (ADP) to adenosine triphosphate (ATP). Increasing
the tissue creatine content, also called creatine 'loading', increases the intracellular pool of Cr
and PCr and provides an alternate pathway to generate ATP rapidly [2]. In skeletal muscles,
creatine exerts ergogenic effects in healthy individuals and therapeutic benefits in various
myopathies [1,3]. In the brain, the benefit of creatine in neurodegenerative diseases (such as
Huntington's and Alzheimer's disease) through improved cerebral bioenergetics is well-
described [4]. More recently, there is growing evidence for the role of creatine as a perinatal
neuroprotectant via energy preservation and antioxidant protection of the fetal brain during
ischemic conditions [5,6].

Several methods are available to measure the creatine content of biological samples. Direct
determination of creatine is often used in a clinical context to diagnose inborn errors of
creatine metabolism [7]. These quantitative methods involve the analysis of patient serum or
urine samples for creatine content via high-performance liquid chromatography [8] or mass-
spectrometry-based approaches [9–13]. In a research context, similar HPLC methods are used
to assess creatine content in tissue or cell culture samples [14–16] or ^{31}P-magnetic resonance
spectroscopy (MRS) to quantify high-energy phosphates in the brain tissue [17].
Alternatively, enzymatic detection of creatine involves fluorometric assay initially described
by Passonneau and Lowry, which involves a multi-step enzymatic reaction to infer creatine
concentration from the decrease in NADH fluorescence [18]. This method was used to
determine creatine content in various tissue types, such as the muscle, brain, heart, liver and
kidneys, from murine and ovine models [19–21]. However, enzymatic assays can be affected
by the presence of interfering substances in the sample, and often require specific acid
extraction during sample preparation that can be time-consuming.

We present an alternative method for creatine quantification in biological tissue samples
using capillary electrophoresis. The advantage of capillary electrophoresis includes high
separation efficiency that enables cost-effective and rapid analysis with low (nanolitre)

sample volumes [22,23]. These features make capillary electrophoresis amenable to use with

human biopsy samples. Capillary electrophoresis provides improved sensitivity at lower

sample concentrations and avoids interference in enzymatic assays. While the detection of

creatine and creatinine using CE was described previously [24], existing methods are limited

to human urinary creatine and creatinine levels. The current protocol describes workflow for

determination of creatine content from snap-frozen tissue samples; from sample

homogenisation and extraction to optimised conditions for capillary electrophoresis. We

developed and optimised the current protocol with ovine lung samples to determine changes

in tissue creatine content after a direct fetal creatine supplementation [25]. However, the

protocol can be easily adapted to analyse various tissue types.

Materials and Methods

The protocol described in this peer-reviewed article is published on protocols.io

(dx.doi.org/10.17504/protocols.io.4r3l22j23l1y/v2) and is included for printing as Supporting

Information file 1 with this article.

The creatine peaks on the electropherogram were identified via enzymatic depletion of the

signal and enhancement of the endogenous signal after spiking with additional creatine.

Creatine concentration was calculated as the area under the peak (AUC) of the

electropherogram and expressed as a ratio compared to the reference DMSO signal.

Standards were injected in technical duplicates and run at the beginning of each day before

sample analysis. The standards were repeated across four days, and a standard curve was

generated from each day to assess the inter-assay variability. Samples were standardised to 1

mg/mL protein concentration and injected in technical duplicates to assess the intra-assay

variability. Creatine concentration was interpolated from the standard curve and expressed as

nmol/mg protein. The results from capillary electrophoresis were subsequently benchmarked

against a commercially available creatine assay kit (MAK079, Sigma-Aldrich), which

measures creatine without phosphocreatine due to a lack of creatine phosphokinase. Samples

were similarly standardised to 1 mg/mL, and the assay was performed according to the

manufacturer's guidelines. Creatine concentrations are similarly expressed as nmol/mg

protein.

Expected results

Using optimised conditions described in the protocol, creatine peak is confirmed after depletion with creatinase enzyme and increased after spiking with additional creatine (**Fig 1**).

Fig 1. Identification of creatine peak by capillary electrophoresis under optimised conditions with UV detection at 200 nm.

Snap-frozen lung tissues were homogenised in PBS with 0.01 % DMSO (red, baseline), and the creatine peak on the electropherogram was identified via enzymatic depletion with creatinase (blue) and spiking (green) with additional creatine (25 µM). A bare-fused capillary (75 µm internal diameter) at an effective length of 36.5 cm (length to detector 8.5 cm) was used. The background electrolyte contained 75 mM NaH_2PO_4 and 150 mM SDS at pH 6. A voltage of 15 kV was applied at 25 ^{O}C with sample injection at 10 mbar for 10 s.

Linearity and assay variability

We assessed the separation performance via the linearity of the standard curve and the intra- and inter-assay variability. The concentration range for the assessment of linearity was determined based on the expected range of concentration for this sample type. A representative image of the electropherograms generated at each standard concentration is illustrated in **Fig 2**, and the linear regression parameters are described in **Fig 3**. Under the optimised conditions tested, there was excellent linearity in standards between 6.3 – 100 µM.

The overall intra-assay variability for concentrations between 6.3 – 100 µM was 1.5 %, and

the inter-assay variability was 6.4 %. The limit of detection was 6 nmol/mg protein, but

variables such as the concentration of the sample injected or the capillary diameter could be

changed to enhance the lower limit of detection.

Fig 2. Representative electropherograms of creatine standards under optimised CE

conditions with UV detection at 200 nm.

The creatine peaks are identified as: (1) 100 µM, (2) 50 µM, (3) 25 µM, (4) 12.5 µM, (5) 6.3

µM, (6) 0 µM. Creatine standards were prepared in homogenisation buffer, which contains 50

mM NaH_2PO_4, 150 mM SDS and 0.01 % DMSO (as the reference signal) at pH 6. The

separation conditions were identical to Fig 1.

Fig 3. Linearity of creatine standards between 6.3 – 100 μM.

Each panel illustrates the standard curve generated on four separate days. The dotted line represents the 95 % confidence interval of the linear regression. The separation conditions were identical to Fig 1.

Determination of creatine content in fetal lamb lungs and benchmarking against commercial enzymatic assay kit

The optimised conditions were used to compare creatine content in ovine fetal lungs in groups that received a direct intravenous supplementation of creatine (6 mg kg^{-1}hr^{-1}) or saline (0.9 % heparinized saline [2 IU/mL]) for 17 days before surgical delivery at 110 days gestation [25]. We tested a random subset of animals from both saline control and creatine-supplemented groups (n = 5/group, based on *a priori* calculation). All samples had creatine levels within the range of detection. Creatine supplementation significantly increased creatine content in the fetal lungs, determined using a commercial enzymatic assay (mean difference = 11.3 [8.9, 13.7] nmol/mg protein; $p < 0.0001$), and capillary electrophoresis (mean difference [95 % CI] = 11.8 [8.7, 14.9] nmol/mg protein; $p < 0.0001$) (**Fig 4A**). Fig 4B illustrates a

Bland-Altman comparison of the two methods. The two methods showed good agreement,

although capillary electrophoresis derived a marginally higher creatine content compared to

the enzymatic assay (mean difference [95% CI] = 0.42 [-0.26 – 1.1] nmol/mg protein). While

this bias was not statistically significant, it could be explained by potential interference in

enzyme assay or lack of completion of the reaction [18]. Regardless, the mean difference was

within the pre-determined range of ± 15 %, and both methods showed a similar magnitude of

difference between groups that did and did not receive creatine supplementation.

Fig 4. Comparison of capillary electrophoresis (CE) and commercial enzymatic assay

kit.

(A) Creatine content in the lung tissue was determined using a commercial enzymatic assay

kit (MAK079, Sigma-Aldrich) and capillary electrophoresis under separation conditions

described in Fig 1. Groups were compared using unpaired t-tests and presented as mean ±

SD. ****p<0.0001. (B) A Bland-Altman comparison between two methods. The dotted lines

on either side of the mean difference denote the 95 % confidence interval of the mean
difference.

Comments on the utility of this protocol

Overall, the current protocol provides a streamlined process to analyse creatine content in
biological tissue samples. Given the excellent sensitivity of the protocol at low tissue creatine
concentrations, we calculated that approximately 1.5 mg tissue (wet weight) would be
sufficient to determine creatine content in a lung sample (assuming ~10 nmol/mg protein in
control conditions and that each mg of tissue yields ~0.2 mg protein). The current protocol
was developed and optimised with ovine lung tissue samples, but it can be easily adapted to
analyse various tissue types. We anticipate that even less amount of tissue would be sufficient
(< 1 mg) to reliably detect creatine in tissues with higher baseline creatine content, such as
the skeletal muscles or brain.

Ethics declarations

142 Animal ethics approval for the collection of fetal lung tissue was obtained at the University of
143 Western Australia (RA/3/100/1607), and animal research was conducted in accordance with
144 the National Health and Medical Research Council Code of Practice for the Care and Use of
145 Animals for Scientific Purposes (Eighth Edition).

Supporting information

147 S1: Step-by-step protocol, also available on protocols.io.

Acknowledgements

149 The schematic illustration of the protocol workflow was created with Biorender.com.

Authors contributions

151 YJC conceptualised, performed the experiments, drafted the protocol and the manuscript. AB
152 conceptualised, performed the experiments and reviewed the protocol. JT interpreted results
153 and reviewed the protocol. JJP obtained funding for the study and reviewed the protocol.
154 PGA conceptualised, interpreted the results, and reviewed the protocol. All authors reviewed
155 and revised the manuscript and approved the final form.

References

[1]   Wyss M, Kaddurah-Daouk R. Creatine and creatinine metabolism. Physiol Rev. 2000;80:1107–1213.

[2]   Wallimann T, Wyss M, Brdiczka D, et al. Intracellular compartmentation, structure and function of creatine kinase isoenzymes in tissues with high and fluctuating energy demands: The "phosphocreatine circuit" for cellular energy homeostasis. Biochem. J. Portland Press Ltd; 1992. p. 21–40.

[3]   Gualano B, Artioli GG, Poortmans JR, et al. Exploring the therapeutic role of creatine supplementation. Amino Acids. 2010;38:31–44.

[4]   Persky AM, Brazeau GA. Clinical pharmacology of the dietary supplement creatine monohydrate. Pharmacol Rev. 2001;53:161–176.

[5]   Dickinson H, Ellery S, Ireland Z, et al. Creatine supplementation during pregnancy: summary of experimental studies suggesting a treatment to improve fetal and neonatal morbidity and reduce mortality in high-risk human pregnancy. BMC Pregnancy Childbirth. 2014;14:150.

[6]   Tran NT, Kelly SB, Snow RJ, et al. Assessing Creatine Supplementation for Neuroprotection against Perinatal Hypoxic-Ischaemic Encephalopathy: A Systematic Review of Perinatal and Adult Pre-Clinical Studies. Cells. 2021;10.

[7]   Schulze A. Creatine deficiency syndromes. Mol. Cell. Biochem. 2003. p. 143–150.

[8]   Carducci C, Birarelli M, Leuzzi V, et al. Guanidinoacetate and creatine plus creatinine assessment in physiologic fluids: An effective diagnostic tool for the biochemical diagnosis of arginine:glycine amidinotransferase and guanidinoacetate methyltransferase deficiencies. Clin Chem. 2002;48:1772–1778.

[9]   Cognat S, Cheillan D, Piraud M, et al. Determination of Guanidinoacetate and Creatine in Urine and Plasma by Liquid Chromatography–Tandem Mass Spectrometry. Clin Chem. 2004;50:1459–1461.

[10]   Carducci C, Santagata S, Leuzzi V, et al. Quantitative determination of guanidinoacetate and creatine in dried blood spot by flow injection analysis-electrospray tandem mass spectrometry. Clin Chim Acta. 2006;364:180–187.

[11]   Carling RS, Hogg SL, Wood TC, et al. Simultaneous determination of guanidinoacetate, creatine and creatinine in urine and plasma by un-derivatized liquid chromatography-tandem mass spectrometry. Ann Clin Biochem. 2008;45:575–584.

[12]   Bodamer OA, Bloesch SM, Gregg AR, et al. Analysis of guanidinoacetate and creatine by isotope dilution electrospray tandem mass spectrometry. Clin Chim Acta. 2001;308:173–178.

[13]   Fernández-Fernández M, Rodríguez-González P, Añón Álvarez ME, et al. Simultaneous determination of creatinine and creatine in human serum by double-spike isotope dilution liquid chromatography-tandem mass spectrometry (LC-MS/MS) and gas chromatography-

192   mass spectrometry (GC-MS). Anal Chem. 2015;87:3755–3763.

193   [14]   Sestili P, Martinelli C, Bravi G, et al. Creatine supplementation affords cytoprotection in
194   oxidatively injured cultured mammalian cells via direct antioxidant activity. Free Radic Biol
195   Med. 2006;40:837–849.

196   [15]   Sestili P, Barbieri E, Martinelli C, et al. Creatine supplementation prevents the inhibition of
197   myogenic differentiation in oxidatively injured C2C12 murine myoblasts. 2009;53:1187–1204.

198   [16]   Ipsiroglu OS, Stromberger C, Ilas J, et al. Changes of tissue creatine concentrations upon oral
199   supplementation of creatine-monohydrate in various animal species. Life Sci. 2001;69:1805–
200   1815.

201   [17]   Adcock KH, Nedelcu J, Loenneker T, et al. Neuroprotection of creatine supplementation in
202   neonatal rats with transient cerebral hypoxia-ischemia. Dev Neurosci. 2002;24:382–388.

203   [18]   Passonneau J, Lowry O. Typical fluorimetric procedures for metabolic assays. In: Passonneau
204   J, Lowry O, editors. A Flex Syst Enzym Anal. New York: Academic Press; 1972. p. 130–131.

205   [19]   Watt KKO, Garnham AP, Snow RJ. Skeletal muscle total creatine content and creatine
206   transporter gene expression in vegetarians prior to and following creatine supplementation. Int
207   J Sport Nutr Exerc Metab. 2004;14:517–531.

208   [20]   Ireland Z, Dickinson H, Snow R, et al. Maternal creatine: does it reach the fetus and improve
209   survival after an acute hypoxic episode in the spiny mouse (Acomys cahirinus)? Am J Obstet
210   Gynecol. 2008;198:431.e1-431.e6.

211   [21]   Tran NT, Muccini AM, Snow RJ, et al. The physiological effects of creatine supplementation
212   in fetal sheep before, during, and after umbilical cord occlusion and global hypoxia. J Appl
213   Physiol. 2021;131:1088–1099.

214   [22]   Gordon MJ, Huang X, Pentoney SL, et al. Capillary electrophoresis. Science (80-.). American
215   Association for the Advancement of Science; 1988. p. 224–228.

216   [23]   Ouimet CM, D'amico CI, Kennedy RT. Advances in capillary electrophoresis and the
217   implications for drug discovery. Expert Opin. Drug Discov. Taylor & Francis; 2017. p. 213–
218   224.

219   [24]   Smith-Palmer T. Separation methods applicable to urinary creatine and creatinine. J
220   Chromatogr B Anal Technol Biomed Life Sci. 2002;781:93–106.

221   [25]   Choi YJ, Williams E, Dahl MJ, et al. Antenatal creatine supplementation reduces persistent
222   fetal lung inflammation and oxidative stress in an ovine model of chorioamnionitis. Am J
223   Physiol – Lung Cell Mol Physiol. 2024;Forthcoming.