Evaluating the Analytical Performance of Direct-to-Consumer Gut Microbiome Testing Services

Stephanie L. Servetas[1], Diane Hoffmann[2], Jacques Ravel[3,4], Scott A. Jackson[1]

[1]Complex Microbial Systems Group, National Institute of Standards and Technology (NIST), Gaithersburg, MD, USA.
[2]University of Maryland School of Law, Baltimore, MD, USA.
[3]Institute for Genome Sciences, University of Maryland School of Medicine, Baltimore, MD, USA.
[4]Department of Microbiology and Immunology, University of Maryland School of Medicine, Baltimore, MD, USA.

Abstract:

Consumer interest in personal microbiome health has given rise to numerous direct-to-consumer (DTC) microbiome testing services despite questions regarding their analytical and clinical validity, and consumer safety. These tests straddle the line between more strictly regulated medical devices and minimally regulated general health and wellness products; a distinction that may not be readily apparent to consumers. To assess the current state of the industry, we evaluated the performance of seven commercial DTC gut microbiome testing services using a standardized NIST – developed human fecal material. Our results reveal major discrepancies, both within and across the different service providers. Significantly, we found variability between providers was on the same scale as biological variability between different donors. We attribute the observed differences to methodological variability and lack of sufficient quality control. Additionally, we highlight that analytical performance is a prerequisite for making sound clinical recommendations. Our results demonstrate the need for standards to ensure analytical validity and regulatory oversight to ensure patient safety.

Introduction:

Over the last decade, the human gut microbiome has been linked to a plurality of health and disease conditions including mental health, cancer, and obesity to name a few [1-3]. Despite extensive research, there are currently no regulatory-approved clinical microbiome diagnostic tests. Concurrently, direct-to-consumer (DTC) personal wellness services are becoming increasingly accessible despite challenges regarding their technical and clinical validity [4-6]. As our current understanding of the microbiome structure-function relationship is still limited, our ability to make confident diagnostic or prognostic predictions based on an individual's microbiome is limited. The French Society of Microbiology, through a publication in Le Monde, has advised against microbiome testing, primarily due to insufficient knowledge and the personalized nature of defining a "healthy" microbiome [7].

Despite these challenges, associations between the gut microbiome and human health are often prominently featured by media outlets, leading many consumers to seek a better understanding of their own gut health. While citizen science projects like The Microsetta Initiative have offered consumers basic information about their gut microbiomes, there remains a demand for more actionable information about the status and implications of personal gut health [8, 9]. In response, numerous commercial DTC microbiome testing services have emerged, providing consumers with profiles of their gut microbiome. Many of these companies

extend their services beyond microbiome profiling by classifying microbes as pathogens or
beneficial, comparing the profile of a consumer's microbiome against a comparative population,
providing an index of gut microbiome health, or offering recommendations for lifestyle changes,
dietary changes and/or dietary supplements.
While these recommendations are often limited to dietary interventions, regulatory
concerns regarding consumer protection have been raised [10]. Notably, there has been a
technological shift from end-to-end diagnostic testing, traditionally confined to clinical settings
and conducted by trained medical professionals, to DTC testing services like 23andMe. These
services have introduced home testing kits that allow consumers (patients) to collect their own
samples, ship them to a laboratory, and access the results without consulting a clinician. A
primary concern is that these at-home tests do not undergo the same rigorous oversight in
validating analytical performance as traditional medical diagnostic tests. Such validation is
crucial to assure clinicians, patients, and regulators of the reliability and actionability of test
results. This lack of validation and the implications may not be apparent to the consumer.
In the age of digital publishing, numerous citizen scientists, reporters, and microbiome
researchers have submitted samples to different companies and published on their results [11-
13]. While these case studies are often intriguing and provide some anecdotal insight, to our
knowledge, a rigorous evaluation of the analytical performance has not yet been conducted.
The National Institute of Standards and Technology (NIST) has developed a suite of candidate
human fecal standards, confirmed to be homogeneous and stable via rigorous multi'omic
analyses [14, 15]. These materials have been designed to enable stakeholders to evaluate the
impact of methodological variability on their microbiome measurements, both qualitatively and
quantitatively. The fecal standards, being homogenous and stable, are ideally suited to assess
the precision (reproducibility) of measurement workflows within and across laboratories.
In this study, we describe our approach to evaluating the performance of seven DTC
microbiome testing services using a NIST-developed human fecal standard. Each service
employed an NGS-based analysis workflow (16S rRNA gene amplicon sequencing or whole
metagenome shotgun sequencing (WMS)) and provided a detailed report of analysis. We
assessed the reproducibility of the results, both within and across these testing services.

Results:

Experimental Design Strategy

As part of its mission to promote U.S. innovation and industrial competitiveness, NIST
has been working with stakeholders in industry, academia, and other government agencies to
develop standards to support the advancement and commercial translation of microbiome
science. To this end, we wished to assess the current state-of-the-art with respect to DTC gut
microbiome testing services.   To achieve this, we deployed a NIST-developed human gut
microbiome (human fecal) standard to assess the precision/reproducibility of seven DTC gut
microbiome testing companies.   For each of the seven companies, three tests/kits were
ordered via the company's website. Upon receipt of the (3 x 7 = 21) sample collection kits, we
used a homogenized pool of human fecal material to serve as the test sample material for each
collection kit.   Collection procedures were followed according to instructions provided by each

company (Table 1). Samples were placed into a company-provided shipping container and
returned to the company for analysis. Results, in the form of company-specific reports, were
received anywhere from 2-8 weeks following shipment of the sample. Taxonomic profiles were
manually extracted from these reports and used for all subsequent analyses. It is also
noteworthy that the companies were not informed of the ongoing assessment until after all the
samples had been processed and final reports were received by NIST.
88

Methodological parameters varied between companies.

The process of translating the composition of stool sample into a microbiome report is
multifaceted. As there are currently no universally accepted best practices for these methods,
each provider likely employs a unique and often proprietary workflow. Microbiome
measurements are impacted by a vast number of methodological variables including 1) sample
collection, storage, and shipping methods, 2) nucleic acid extraction techniques, 3) NGS-library
preparation, 4) sequencing technology, and 5) bioinformatic analyses. Bias can be introduced at
every step, and even minor changes in methodology can lead to significant differences in
results. In this study we had limited knowledge of the workflows that were employed by each
company; however, we were able to infer key differences in both sample collection and
sequencing methodologies.
Table 1 shows a comparison between companies regarding some methodological
variables likely to impact the result (Table 1). Notably, there was considerable variability in
methods for sample collection, including both the sample type (whole bowel movement vs. used
toilet paper) and the sampling implement. Furthermore, some companies instructed users to
add the sampling implement to a buffer, or to resuspend the sample in a buffer and discard the
implement, while others directed users to ship the implement back 'neat', enclosed in a
secondary sterile container. None of the shipping packages included ice, so all samples were
shipped under ambient conditions (the study was conducted in the spring). The exact transit
times between shipping and receipt of samples were not reported by all companies. Typically,
notification of sample receipt or processing was sent 5-7 days after shipping.
Regarding sequencing methodologies, some companies used marker gene amplicon
sequencing while others relied on WMS. Among those using amplicon sequencing, one utilized
16S rRNA gene amplicon sequencing of the V3-4 region and the other two used 16S rRNA
gene amplicon sequencing of the V4 region; additionally, one company included sequencing of
internal transcribed spacer (ITS) amplicon for fungal identification. The read length and depth of
sequencing, which can impact the specificity and sensitivity of taxonomic identification, varied
by company. Although a wide range of read depths was reported, the variability within a single
company was generally narrower (Table 1). Read depth ranged from $< 2 \times 10^4$ to $> 1 \times 10^7$.
Interestingly, higher read depths did not always correlate with a greater number of species
identified. For example, the three replicates from company C each reported over 600 species
with read depths between 2×10^4 and 3×10^4. In contrast, company D reported 95, 98, 99
species for their three replicates, each with read depths exceeding 1×10^7 reads. Although most
details of read processing were not disclosed, we noted different cutoffs (ranging from 0.1% to 8
$\times 10^{-6}$ %) for reporting taxonomic assignment based on the taxonomic tables provided.
124

126 Table 1. Gut Microbiome Testing Methodological Parameters

Company	Sampling Method	Sampling Tool[3]	Storage buffer (Y/N)	Library Prep Method[4]	Read length[4]	Total Reads (M, Million; K, thousand)[4]	Species Identified[5]
A	swipe over TP	a	N	16S (V3-4)	300	186 K – 424 K	66, 72, 121
B	swipe over TP	b	N	16S (V4)	variable	17.5 K – 33.3 K	20, 29, 42
C	swipe over TP	c	Y	16S (V4)	150	23.7 K – 29.4 K	615, 713, 767
D	swipe over TP	a	N	WGS	151	11 M – 15 M	96, 98, 99
E	Proprietary collector; Insert coring brush in 10-12 areas	d	Y	WGS	151	4.2 M – 8.6 M	139, 163, 163
F[1]	Proprietary collector; 1-2mL liquid; 10-12 areas in solid stool	d	Y	WGS	151	4.6 M – 6.4 M	1471, 1626, 1701
G[2]	Proprietary collector; swipe collectors along pad	e,f	Y/N	WGS	Data not provided	Data not provided	259, 259

128
[1] An eye dropper is also supplied for use with liquid stool samples.
[2] Three sampling implements were supplied, one easicollect and 2 swabs. One swab was sent dry, one in buffer.
[3] The numbers in this column correspond to the following specific sampling implements: (a) 4N6 FLOQSwabs, (b) BD BBL CultureSwab EZ, (c)
Puritan Sterile Cotton Tipped Applicator, (d) Cytobrush Plus GT Gentle touch tip cell collector, (e) EasiCollect Plus, (f) Specimen Collection Swab
[4] Data provided upon request
[5] Company A number determined from taxonomic table provided, abundances reported down to 0.01%; Company B number determined from
taxonomic table provided upon request, abundances reported for any taxa containing at least 4 reads (note only 20% of reads were assigned to
the species level); Company C number determined from taxonomic table provided, abundances reported down to <0.0001%; Company D
number determined from taxonomic table provided, abundances reported down to <0.1%; Company E number provided in the report, no cut-
off provided; Company F number determined from taxonomic table provided upon request, abundances reported down to 0.000008%;
Company G number determined from taxonomic table provided, abundances included down to 0.01%.
140

Taxonomic Profiles Varied Across Companies

The process of generating the taxonomic tables from the data provided by each
company required manual curation, as the reports came in varied formats (e.g., PDF, html).
Additionally, a complete taxonomic profile was not part of Company A's report but was obtained
after sending a request to their help center. For comparison, NIST also performed metagenomic
analyses on aliquots of the same material that was shared with the DTC companies. In
addition, taxonomic data from the other 7 unique donors that contributed to the candidate fecal
reference material development were included to highlight the differences associated with
biological variability (donor) *vs.* technical (methodological) variability. Although species-levels
taxa were reported by each DTC company, all subsequent analyses at NIST were conducted
using genus-level taxa assignments. A comparison of alpha diversity revealed a range of
diversity values with no clear patterns based on read depth or sequencing method (Figure 1A-C,
respectively). Bray-Curtis dissimilarity was calculated to compare all replicates from each of the
seven companies and NIST (Figure 1D). Along with the alpha-diversity plots, the ordination plot
can indicate inter-lab reproducibility. Narrower box plots (Figure 1A-C) and tighter distribution on
the ordination plot (Figure 1D) suggest higher reproducibility for a company (*e.g.,* Company F),
while Company A showed poor reproducibility. Moreover, there was no distinctive clustering
based on whether WMS or amplicon sequencing was used (WMS shown with solid lined, and

16S rRNA gene amplicon sequencing with dashed lined ellipses in Figure 1D). One question
raised was how the diversity observed for a single sample analyzed by multiple companies
compared to biological diversity from different donors analyzed by a single workflow. A second
Bray-Curtis dissimilarity index was calculated that included data from seven additional donors
processed with the NIST workflow. Surprisingly, the diversity observed for samples from a
single donor analyzed by different companies (asterisk in Figure 1E, blue ellipses) appeared to
be greater than or equal to that of biologically distinct samples (Figure 1E, other shapes, yellow
ellipse). This demonstrates that methodological variability can be greater than biological
variability.
168

169
170
Figure 1. Diversity Metrics for Genera identified by DTC companies and by NIST
Plots show alpha-diversity calculated using **(A)** Shannon **(B)** Inverse Simpson indices and **(C)** for the total
genera identified. Three replicates are plotted for companies A, B, C, D, E, and F; company G has 2
replicates and company H has 4.
(D) Ordination plot showing beta-diversity based on genera for different companies processing samples
from a single donor. Colors indicated different Companies; ellipses are drawn to further identify samples
processed with the same Company. Ellipses with dashed lines indicates 16S amplicon sequencing was
used; solid lines on the ellipses indicate WMS.
(E) Ordination plot showing beta-diversity based on genera for a single donor processed by different
companies (asterisk) and different donors processed by a single workflow (other symbols). The yellow

ellipses indicate the cluster of other donors processed with a single workflow; blue ellipses indicate
samples from a single donor. Similar to plot D, dashed lined ellipses denote 16S, solid denote WMS
183
In addition to comparing diversity metrics, the taxonomic composition was compared on a
genus-by-genus basis to identify the genera common to all analyses. For this analysis, any taxa
that was not given a specific genus designation was renamed unspecified and these
assignments were then grouped with any unidentified taxa that were noted in the report. While
some companies provided information that allowed us to determine the percentage of
unidentified reads, this was not uniform across companies. In fact, the NIST analysis was based
on Bracken, and in this workflow unidentified reads are dropped prior to calculating relative
abundance. Therefore, while we were able to show this data for some companies, absence of
the unspecified category in Figure 2 does not indicate 100% of the taxa were identified but could
indicate that the unidentified reads were dropped earlier in the process, prior to calculating
relative abundance values for known taxa. Despite vast differences in read depths, it was
hypothesized that the most common genera would be present in all samples, and less abundant
genera would only be seen in samples with higher coverage. The number of identified genera
ranged from 34 to 906, with a relatively narrow range for each single company (Figure 1C).
From all 9 analyses (7 companies and NIST WMS and 16S rRNA gene amplicon sequencing), a
total of 1208 unique taxa were identified. Only 3 genera were found in all samples when
comparing all taxonomic profiles (data not shown). A single replicate from Company A had a
distinctive taxonomic profile (Figure 3B, and observed in Figure 1D as the single red asterisk on
the left), which when removed from the analysis, increased the number of common genera to
17. This accounted for less than 2 % of the total genera identified, or between 2 % and 43 % for
any given sample. In most cases, these common genera constituted a majority of the genera (>
50 % of the abundance for companies A, C, D, G, and both NIST analyses), or the majority of
genus level specified reads (companies B and F). Additionally, the conservation of identified
genera between replicates varied, from as low as 6 % for Company A to as high as 98 %.
Excluding Company A, the genera conserved between replicates from the same company
accounted for at least 95% of the identified sample composition. This indicates that most
variability between replicates from a single company occurs among taxa present at low
abundance and highlights the importance of empirically determined read cut-offs for taxa
inclusion in a report.
213
Replicate 1 from Company A had a significantly different taxonomic profile compared to
the two other replicates from Company A and those from other companies (Figure 3B). This
egregious anomaly prompted further investigation. Compared to replicates 1 and 2 from
Company A, replicate 3 had approximately half the number of reads (186K compared to just
over 400K). This might have contributed to its distinct profile, as only 6 genera were common
between replicate 3 and replicates 1 and 2, corresponding to 10% of the total genera and
abundance for these replicates. In contrast, replicates 1 and 2 shared 69 out of 70 and 71
genera, respectively. Notably, the other 28 genera in replicate 3 were not unique but were
identified in at least one other replicate across all the companies. It is also worth noting that the
percentage of unspecified reads in replicate 3 comprised over 50% of the sample, whereas the
unspecified designation accounted for just over 20% in replicates 1 and 2. These discrepancies

raise concerns not just because the profile was distinct from other replicates originating from the
same sample, but also because it apparently passed all quality control checks from the
company, resulting in a report being generated and sent back to the customer.
228

229
Figure 2. Common genera identified across testing companies (A) The relative abundance
of genera that were identified in all but one test. Each company is indicated by a letter, with
replicate samples indicated by the number after the dash. (B) Relative abundance data for the
one test sample (A-3) was excluded from the overall analysis. All genera that were identified are
included in the plot, (C) Relative abundance of genera found in common across replicates run
by company A. The grey portion in all bars represents taxa that were not given a Genus level
identification; this includes both taxa designated at a higher taxonomic level or that were listed
as unidentified.
238
239

240 *Differences in Health Indicators Across Companies*

241
In addition to providing customers with a taxonomic profile, many reports also include
health metrics, such as comparing a customer's specific taxonomic profile to that of a
comparator group. Notably, the comparator group varies between companies. For instance,
Company E uses an internally developed dataset from a "healthy" cohort, while other
companies rely on external datasets, such as those from the Human Microbiome Project [16] or
American Gut [9]. Beyond providing comparative data, some companies also identify beneficial
or harmful microbes and suggest ways to modulate these microbes' abundance to improve
health. To examine how results vary between companies on a taxa-by-taxa basis, five clinically
relevant genera were selected (Figure 3 **(A)** *Bacteroides* **(B)** *Bifidobacterium* **(C)** *Clostridium* (to
include *Clostridioides*) **(D)** *Faecalibacterium* and **(E)** *Roseburia*). All results were compared to
the criteria provided by two of the companies. One company reported average values of the 5

genera in their reference population (Figure 3, gray dotted line), and the other company
reported an interquartile range (IQR) of their reference population (Figure 3, range between the
25% and 75% bounds shaded in yellow). Interestingly, except for *Bifidobacterium*, the average
used by one company does not fall within the IQR used by the other company. Variability
between companies is observed for all five genera, with no apparent correlation to library
preparation (16S rRNA gene amplicon sequencing vs WMS). In addition, the precision from a
single company varied based on taxa, with some taxa showing indistinguishable points between
replicates (*e.g.*, Figure 3E, *Roseburia* for companies D and F) and other showing wide variability
(Figure 3B, *Bifidobacterium* for company D). The implications of this variability are particularly
evident in the health recommendations commonly included in DTC microbiome testing reports.
The most striking example is of the conflicting recommendations between replicate 3 from
Company A and replicates 1 and 2. For instance, starch breakdown was assessed as a metric
for gut health; replicate 3 scored below average, whereas replicates 1 and 2 scored above. Out
of the 10 functional categories assessed by this company, seven were below average for
replicate 3, compared to only one for replicates 1 and 2, and the below-average category for
replicates 1 and 2 was distinct from the seven observed for replicate 3. The company reported
the microbiome as healthy for replicate 1 and 2, but unhealthy for replicate 3, which could lead
to unnecessary interventions if such a report was received by a consumer. While Company A
provides an extreme example, other discrepancies, including the identification of pathogens,
were observed between different companies. Notably, *Clostridioides difficile*, a primary
causative agent of antibiotic-associated diarrhea for which carriage is a risk factor, was not
consistently identified between companies.
275
276

277
278
Figure 3. Comparison of Relative Abundance for Select Genera Relative abundance for select genera
(A) *Bacteroides* **(B)** *Bifidobacterium* **(C)** *Clostridium* (to include *Clostridioides*) **(D)** *Faecalibacterium* and
(E) *Roseburia* were plotted compared to reported comparators from company G as a grey dashed line
and from company E as a shaded yellow region. The grey dashed line represents the average from the
Human Gut project reported by company G as a comparator. The yellow shaded region indicates the IQR
used to define the healthy population as a comparator used by company E. The symbols indicated

metagenomic analysis: O's designate companies using 16S amplicon sequencing and X's designate
WMS analysis. Three replicates are shown for each Company except for Company F which only had 2
successful samples.

Discussion:

Our rigorous assessment of seven microbiome testing companies has spotlighted the
systemic issue of poor reproducibility that plagues the industry. The primary contributor to this
outcome is methodological variability. The significant impact of methodological variables on the
results of microbiome metagenomic measurements has been demonstrated by us and others,
and DTC microbiome testing companies are not immune to these sources of bias [14, 17-20].
The use of a homogenous stool reference material enabled a systematic evaluation without the
confounding factors of biological or composition heterogeneity that may be found in a traditional
stool sample. Despite these benefits, it is important to emphasize the stool reference material
used for this study does not represent an absolute "ground truth". Consequently, this study
cannot definitively determine which results are closest to the actual microbiome composition,
nor was that the goal of this study. Nonetheless, access to a homogenous and standardized
material, comparable to the complexity to a real-world sample, is sufficient for evaluating the
precision of the sample processing workflows. Overall, this study revealed that the impact of
methodological variability (between company comparisons) was on the same order of
magnitude as biological variability (samples taken from 8 distinct donors). These findings have
major implications for the comparability of tests on a single individual from two different
companies and highlights the need for caution in interpreting and acting on these test results,
especially in the absence of validated diagnostic evidence. Of note, despite the promising
potential of the gut microbiome in human health, as evidenced by recent drug approvals [21-23],
the diagnostic capabilities of these tests remain largely underdeveloped, partly due to the
challenge demonstrated in this study.
Another crucial aspect we considered was the validity of the diagnostic and prognostic
implications derived from these tests. Questions arise as to how a testing service determines
whether a patient's microbiome is healthy or unhealthy, what the relevant diagnostic and
prognostic biomarkers are, and how these biomarkers were identified and validated. These
questions are significant, as the test results may lead consumers to make potentially
unwarranted or unsafe lifestyle changes. Many of the results are reported with respect to an
average or "healthy" microbiome; however, defining a healthy microbiome remains a challenge
due to the heterogeneity of the human population, confounders, and the possibility of multiple
healthy microbiome definitions. Even among the DTC companies evaluated, there was no
standard practice for selecting a comparative 'heathy' population. In some cases, companies
used data generated in-house, either by measuring a cohort specifically recruited to serve as
the comparative population or by aggregating the data from prior customer's test samples.
Alternatively, some companies look to previously constructed datasets such as that from
American Gut or the Human Microbiome Project (HMP); both methods raise concerns. A benefit
of using data generated in-house is that it was likely collected following the same workflow
being used for the test samples, which should reduce the impact of methodological variability as
reproducibility within a workflow is generally very good. However, little information was given on

the demographics of these comparative populations, raising questions about how representative
they are. In contrast, comparisons using data from previous efforts such as the American Gut,
may provide more demographic information and/or generated a larger more representative
dataset than possible for an individual company. As detailed above, methodological variability is
a major hurdle for data comparability; therefore, if the DTC company is not following the same
protocol used to generate the comparative population data, observed differences could be due
either to biology, methodology, or a combination.
In general, the health recommendations of many DTC microbiome testing companies
focus on healthier eating habits, which may seem innocuous and unlikely to cause direct harm
to the consumers. However, it is crucial to recognize that a significant portion of the individuals
who seek these tests are often those with chronic gut conditions, who may have struggled to
find effective treatments. For these individuals, the variability in test results or the lack of
expected outcomes from the recommendations provided, could not only lead to a loss of
consumer faith in the science but also result in the delay of appropriate medical care.
Additionally, some companies recommended that customers start taking costly supplements
(e.g., probiotics) that are sold by the same company and for which there is very little clinical
evidence for efficacy.
Given that adopting a single methodological workflow across the entire microbiome
industry is unrealistic, there are several steps that could be taken to improve result transparency
and interpretation. There are two primary areas to address with respect to DTC testing (1)
clinical validity and (2) analytical performance. With respect to the first, the rapidly expanding
field of microbiome research is constantly improving our understanding of the associations
between health and the microbiome, but much of the data is likely correlative rather than
causative. Without substantial causative evidence, the clinical validity of these tests will remain
a challenge. While not a solution, inclusion of peer reviewed literature citations to support any
claims to health status, or designation of a microbe as beneficial or harmful, would provide
consumers (and clinicians) with a means of understanding the legitimacy of their results.
There are many more readily available resources for addressing analytical performance.
An ideal standard which would enable both accurate and precise testing does not yet exist;
however, as demonstrated by this study, inclusion of standard fecal material could be used to
determine the minimum number of reads required for reproducible results and other minimum
standards to ensure result validity. Additionally, while on a smaller scale than the 100s of
organisms present in a stool sample, mock communities of 10-20 common gut microbes are
available from several commercial outlets that enable end-users to assess the accuracy and
precision of their microbiome measurements. Although beyond the scope of this discussion, in
addition to the reference materials mentioned here, there are numerous publications on
validating analytical performance of a metagenomic workflow [24-26]. The adoption of
standards, and the reporting of the analytical performance gleaned from these standards, would
help customers and other stakeholders (including regulatory bodies) understand and
characterize bias in a given workflow. The industry should consider developing a consensus
document on minimum requirements, with input from testing companies and other stakeholders.
Similar documents, such as MIQE, EMMI, AsqMI, and STORMS have been developed for other
high complexity molecular diagnostic assays [27-30]. Such a document would support test
validity and bolster consumer confidence for DTC gut microbiome testing.

Methods:

Material preparation

As part of the Microbiome Program, researchers at NIST have undertaken the challenge of producing a whole stool gut microbiome reference material. While the final reference material will be comprised of pooled donor material, as part of early investigations NIST obtained homogenized material from individual donors. Having a relatively large quantity of homogenized stool gave NIST the confidence that any deviations seen in the results were due to bias introduced during the sample processing workflow and not biological differences between aliquots. To further ensure a homogenous starting sample, multiple aliquots were pooled in a single 50 mL conical tube, to a final volume of approximately 540 mg/mL stool and a Bristol stool consistency of 4, for sampling. This pooled material was used to inoculate all tests within 1 hour of preparation.

NIST Analysis of the homogenized whole stool sample:

In addition to the analyses provided by the service providers, the standardized fecal sample was also analyzed in the NIST lab using in-house DNA extraction, 16S rRNA and WMS workflows, to provide a better understanding of how reproducibility within a service provider compared to reproducibility within our own lab. Four replicate DNA extractions were carried out using the Zymo Research Quick-DNA Fecal/Soil microbe miniprep extraction kit [cat # D6010] per protocol with the vortex genie adapter for bead beating carried out for 45 mins. WMS and 16S rRNA sequencing library preps were carried out on the same DNA extract. 16S rRNA library prep was carried out as followed: 12 ng of genomic DNA was amplified in a 25 µL reaction comprised of 2x Kapa HiFi Master Mix (Roche cat # KK2601/2), 1 µL of 10 µM V4 515F/806R primers with Illumina adapters (5'-
TCGTCGGCAGCGTCAGATGTGTATAAGAGACAGGTGYCAGCMGCCGCGGTAA
 -3', 5'- GTCTCGTGGGCTCGGAGATGTGTATAAGAGACAGGGACTACNVGGGTWTCTAAT
 -3'), under the following conditions: 95.0 ˚C (3 min), and 18 cycles of amplification 98.0 ˚C (30 sec), 55 ˚C (15 sec), 72 ˚C (20 sec), and a final extraction at 72 ˚C for 5 min. 16S amplification products were purified with SPRIselect beads (Beckman Coulter cat # B23317/ B23318/ B23319) in a 0.8:1 (bead to product) ratio. The cleaned 16S products were indexed using IDT® for Illumina® DNA/RNA UD Indexes with 8 amplification cycles, but the same thermocycling conditions used for the 16S amplification. The indexed 16S amplicons were purified with SPRIselect beads in a 1:1 (bead to product) ratio. Purified indexed amplicons were pooled in equal quantities and adjusted to a 4nM pool for sequencing preparation following the MiSeq Systems Denature and Dilute DNA Libraries Guide, Protocol A (Document # 15039740 v10, February 2019).

WMS library prep was carried out using Covaris focused-ultrasonication for fragmentation
followed by NEB for indexing. 10 μL of the DNA extract (approximately 500-700 ng of DNA)
was added to a 120 μL of TE in a microtube AFA Fiber Pre-slit Snap Cap 6X16mm tube
(Covaris Cat#: 520045). Samples were fragmented using the following settings: Intensity:
10%; Duty cycle: 10%; Cycle/burst (CPB): 100; Time: 1 min. Fragmented DNA was analyzed
using the Tapestation and average peak size was approximately 280 bp. Fragmented DNA
was indexed using the New England Biolab (NEB) NEBNext Ultra II DNA Library Prep Kit for
Illumina with NEBNext Multiplex oligos for Illumina (NEB cat # E7645S/L; E6440) following
the instruction manual (V6.1_5/20) with 50 μL of fragmented DNA as input; this corresponded
to approximately 200 – 300 ng of DNA input. Purified indexed samples were pooled in equal
quantities and adjusted to a 4nM pool for sequencing preparation following the MiSeq Systems
Denature and Dilute DNA Libraries Guide, Protocol A (Document # 15039740 v10, February
2019). Both 16S and WMS sequencing libraries were run using a MiSeq Reagent Kit v3 600-
cycle cartridge (MS-102-3003) with pair-end reads (raw reads will be available at
https://doi.org/10.18434/mds2-3348; contact sls8@nist.gov if the link is not active).

16S reads were analyzed in R (4.1.0) using CutAdapt [31] for primer trimming and DADA2
(1.20.0) [32] with the silva database [33] (v138) for taxonomic classification. The complete R
code is included in Supplemental file 1 as an R Markdown and html (16S_Analysis/Dada2-
16sAnalysis2024).

WMS reads were taxonomically profiled using Kraken2 [34] and bracken [35] with default
parameters against the prebuilt GTDB-tk database(r89) [36, 37]. Finally, read quality was
assessed with FastQC. The integrated snakemake workflow, based on taxonomic profiling from
NBIS-Meta (https://github.com/NBISweden/nbis-meta) can be found at
https://github.com/ikeenum/WW_influent_bias.

Analysis of additional donor samples

WMS was also conducted on the 7 additional donors of the candidate fecal reference material to
serve as comparators for assessing the magnitude of methodological variability compared to
biological (donor) variability. Two replicate DNA extractions from homogenized whole stool from
all 8 unique donors (the donor used for all previous analyses plus 7 additional donors) was
carried out using the Qiagen PowerSoil Pro Kit. WMS library preparation was done using
Illumina NexteraXT followed by sequencing on the Illumina HiSeq X platform (raw reads will be
available at https://doi.org/10.18434/mds2-3348; contact sls8@nist.gov if the link is not active).
WMS reads were profiled using the Kraken and bracken workflows as described above.

Selection and preparation of DTC Gut Microbiome Tests

NIST conducted a web search for "at home direct to consumer microbiome testing kits" and
selected 7 companies for inclusion in the study; 3 kits were ordered from each company to
assess the intra-and inter-laboratory precision (reproducibility). A NIST employee created an
account for each company using a personal email account so that the companies were blinded
to the experiment at the time of testing.

Each test kit came with instructions, collection devices, sample containers, and pre-paid
shippers. To the extent possible, the sampling instructions were followed per the manufacturers'
instructions. There were two notable deviations from the sampling instructions. First, some
companies provided proprietary collection devices to support sampling directly from a bowel
movement, while others instructed users to sample from used toilet paper; for this experiment all
samples were taken directly from the 50mL conical. Second, some companies provided
instructions to sample multiple sections on the bowel movement, which was not possible with
our sample type. As mentioned above, all kits were inoculated within one hour of preparing the
sample and then immediately packaged in the prepaid shippers and dropped off for shipping at
the respective location (e.g. USPS, FedEx) as designated by the shipping label.

Analysis of Taxonomic Profiles

Six of the 7 service providers produced a report of analysis for each kit (n=3) sent. One service
provider produced 2 reports and reported a sample failure. An additional sample was sent, and
this also failed their sample QC (the specific type of failure was not disclosed) and a third
attempt was not made. Provider reports varied in content and format necessitating manual
curation of the taxonomic data from each report to produce a consistent format. In addition to
creating a consistent file and format, some datasets provided taxonomic resolution down to the
species or even strain level whereas other taxa were identified at a higher order, for example
only down to the family level. A genus level taxonomic classification was chosen to facilitate
comparisons between datasets; therefore, the relative abundance of all species and strain level
designations were summed to the genus level. The relative abundances for unclassified taxa
and taxa where a specific genus designation was not given were summed and given an
unspecified designation. Taxonomic tables generated from the NIST 16S and WMS analysis
were put in the same format. All resulting tables were used as the input files for future analysis
conducted in R Studio (Supplemental files 1 and 2, NIST_data and Genus_tables). We are
unable to share the actual reports in a blinded fashion so this version of the data will not be
made publicly available. The complete code for analysis and figure generation can be found in
Supplemental file 3 as an R Markdown and html (2024-analysis).

References
1. Gopalakrishnan, V., et al., *The Influence of the Gut Microbiome on Cancer, Immunity,*
and Cancer Immunotherapy. Cancer Cell, 2018. **33**(4): p. 570-580.
2. Shoubridge, A.P., et al., *The gut microbiome and mental health: advances in research*
and emerging priorities. Mol Psychiatry, 2022. **27**(4): p. 1908-1919.
3. Van Hul, M. and P.D. Cani, *The gut microbiota in obesity and weight management:*
microbes as friends or foe? Nature Reviews Endocrinology, 2023. **19**(5): p. 258-271.
4. Schlaberg, R., *Microbiome Diagnostics.* Clin Chem, 2020. **66**(1): p. 68-76.
5. Nolan, J.J. and E. Ormondroyd, *Direct-to-consumer genetic tests providing health risk*
information: A systematic review of consequences for consumers and health services.
Clin Genet, 2023. **104**(1): p. 3-21.
6. Acharjee, A., et al., *The diagnostic potential and barriers of microbiome based*
therapeutics. Diagnosis (Berl), 2022. **9**(4): p. 411-420.

7.   Microbiologie, S.F.d. *ALERT ON FECAL MICROBIOTA TESTS*. 2023 [cited 2024
5/17/2024]; Available from: https://www.sfm-microbiologie.org/actualites/alerte-sur-les-
tests-de-microbiote-fecal/.
8.   *The Microsetta Initiative* 2023; Available from: https://microsetta.ucsd.edu/.
9.   McDonald, D., et al., *American Gut: an Open Platform for Citizen Science Microbiome
Research.* mSystems, 2018. **3**(3).
10.   Hoffmann, D.E., et al., *The DTC microbiome testing industry needs more regulation.*
Science, 2024. **383**(6688): p. 1176-1179.
11.   Saey, T.H., *Here's the poop on getting your gut microbiome analyzed.* 2014:
ScienceNews.
12.   LaGreca, B., *Viome vs. Biohm: Which Is Best for Microbiome Testing?*, in *Empowering
Patients Through & Beyond Cancer*. 2023.
13.   Wolfe, B., *What Do At-Home Microbiome Tests Tell Us About Microbes Living in Our
Guts?*, in *Science & Technology*. 2022: TuftsNow.
14.   Forry, S.P., et al., *Variability and bias in microbiome metagenomic sequencing: an
interlaboratory study comparing experimental protocols.* Sci Rep, 2024. **14**(1): p. 9785.
15.   Amanda, B., et al., *Multi'omic Characterization of Human Whole Stool RGTMs*. 2023,
NIST Interagency/Internal Report (NISTIR), National Institute of Standards and
Technology, Gaithersburg, MD.
16.   Turnbaugh, P.J., et al., *The human microbiome project.* Nature, 2007. **449**(7164): p.
804-10.
17.   Sinha, R., et al., *Assessment of variation in microbial community amplicon sequencing
by the Microbiome Quality Control (MBQC) project consortium.* Nat Biotechnol, 2017.
35(11): p. 1077-1086.
18.   O'Sullivan, D.M., et al., *An inter-laboratory study to investigate the impact of the
bioinformatics component on microbiome analysis using mock communities.* Sci Rep,
2021. **11**(1): p. 10590.
19.   Amos, G.C.A., et al., *Developing standards for the microbiome field.* Microbiome, 2020.
8(1): p. 98.
20.   Zaiko, A., et al., *Towards reproducible metabarcoding data: Lessons from an
international cross-laboratory experiment.* Mol Ecol Resour, 2022. **22**(2): p. 519-538.
21.   Walter, J. and F. Shanahan, *Fecal microbiota-based treatment for recurrent
Clostridioides difficile infection.* Cell, 2023. **186**(6): p. 1087.
22.   *Live fecal microbiota (Rebyota) for prevention of CDI recurrence.* Med Lett Drugs Ther,
2023. **65**(1671): p. 35-36.
23.   Jain, N., et al., *Advancing therapeutics for recurrent clostridioides difficile infections: an
overview of vowst's FDA approval and implications.* Gut Microbes, 2023. **15**(1): p.
2232137.
24.   McIntyre, A.B.R., et al., *Comprehensive benchmarking and ensemble approaches for
metagenomic classifiers.* Genome Biol, 2017. **18**(1): p. 182.
25.   Sczyrba, A., et al., *Critical Assessment of Metagenome Interpretation-a benchmark of
metagenomics software.* Nat Methods, 2017. **14**(11): p. 1063-1071.
26.   Poussin, C., et al., *Interrogating the microbiome: experimental and computational
considerations in support of study reproducibility.* Drug Discov Today, 2018. **23**(9): p.
1644-1657.
27.   Bustin, S.A., et al., *The MIQE guidelines: minimum information for publication of
quantitative real-time PCR experiments.* Clin Chem, 2009. **55**(4): p. 611-22.
28.   Mirzayi, C., et al., *Reporting guidelines for human microbiome research: the STORMS
checklist.* Nat Med, 2021. **27**(11): p. 1885-1892.

29.   Keenum, I., et al., *Amplicon Sequencing Minimal Information (ASqMI): Quality and Reporting Guidelines for Actionable Calls in Biodefense Applications.* J AOAC Int, 2023. **106**(5): p. 1424-1430.

30.   Borchardt, M.A., et al., *The Environmental Microbiology Minimum Information (EMMI) Guidelines: qPCR and dPCR Quality and Reporting for Environmental Microbiology.* Environ Sci Technol, 2021. **55**(15): p. 10210-10223.

31.   Martin, M., *Cutadapt removes adapter sequences from high-throughput sequencing reads.* 2011, 2011. **17**(1): p. 3.

32.   Callahan, B.J., et al., *DADA2: High-resolution sample inference from Illumina amplicon data.* Nat Methods, 2016. **13**(7): p. 581-3.

33.   Quast, C., et al., *The SILVA ribosomal RNA gene database project: improved data processing and web-based tools.* Nucleic Acids Res, 2013. **41**(Database issue): p. D590-6.

34.   Wood, D.E., J. Lu, and B. Langmead, *Improved metagenomic analysis with Kraken 2.* Genome Biol, 2019. **20**(1): p. 257.

35.   Lu, J., et al., *Bracken: estimating species abundance in metagenomics data.* Peerj Computer Science, 2017.

36.   Chaumeil, P.A., et al., *GTDB-Tk: a toolkit to classify genomes with the Genome Taxonomy Database.* Bioinformatics, 2020. **36**(6): p. 1925-1927.

37.   de la Cuesta-Zuluaga, J., R.E. Ley, and N.D. Youngblut, *Struo: a pipeline for building custom databases for common metagenome profilers.* Bioinformatics, 2020. **36**(7): p. 2314-2315.