Neuroimage Denoiser - a Deep Learning Framework for Removing Noise from Transient Fluorescent Signals in Functional Imaging.

Stephan Weißbach[1,2,#], Michela Borghi[1], Jonas Milkovits[1,2], Carolina Amaral[1], Abderazzaq El Khallouqi[1], Susanne Gerber[2], Martin Heine[1,#]

[1] Institute of Developmental Biology and Neurobiology (iDN), Johannes Gutenberg-University, 55128 Mainz, Germany,

[2] Institute of Human Genetics, University Medical Center, Johannes Gutenberg-University, 55131 Mainz, Germany

[#] To whom the correspondence should be addressed: marthein@uni-mainz.de and s.weissbach@uni-mainz.de

Keywords: glutamate imaging, calcium imaging, functional imaging, iGluSnFR, GCaMP6f, denoising, deep learning

Highlights

- Neuroimage Denoiser is a deep learning framework to remove noise from functional microscopic recordings, particularly trained and tested for glutamate imaging
- Neuroimage Denoiser balances the removal of noise while preserving the amplitude of responses
- Neuroimage Denoiser operates without re-training for different sensors (when the localization is similar) and recording frequencies

Motivation

Accurate measurements of neuronal activity through functional imaging are critical in understanding mechanisms of synaptic plasticity and learning concerning changes in the molecular composition of single synapses. Traditional denoising methods, such as Gaussian or Median filters, indiscriminately smooth entire recordings, reducing temporal and spatial resolutions considerably. Existing frameworks are not suited to remove noise from glutamate recordings due to the fast dynamics of the sensor. Therefore, a specialized tool for the challenges imposed by glutamate recordings, i.e. faster dynamics, and synaptic localization, is needed.

Summary

We developed Neuroimage Denoiser, a novel U-Net-based model that effectively removes noise from microscopic recordings of transient local fluorescent signals. The model makes the denoising process independent of the recording frequency and the kinetics of the sensor used. The framework is easy to use for denoising and training and has minimal hardware requirements, thus, making it accessible for an average laboratory to create a custom version specific to their experimental setup. Neuroimage Denoiser significantly enhances the quality of functional microscopy recordings by effectively removing noise, thereby facilitating a more accurate and reliable analysis of neural activity.

Introduction

Noise is a pervasive phenomenon impacting all experimental systems used in neuroscience. Accurate measurement of neuronal activity is crucial for understanding neuronal function, yet noise in imaging systems can severely distort data[1]. With the development of genetically encoded calcium or neurotransmitter sensors, the possibility reading the activity of neuronal subcellular compartments like synapses became feasible, based on fast time-lapse fluorescent imaging. In camera-based setups, noise affects pixels individually without any spatiotemporal relations, often uneven distributed over the field of view. However, the signal from fluorescent reporters that indicates neuronal activity in synapses or somata exhibits a high spatiotemporal correlation.

Traditional methods, such as Gaussian or Median filters, can be used to remove noise from microscopic recordings. While effective to some extent, these filters indiscriminately smooth the entire image, leading to reduced temporal and spatial resolutions[2]. Additionally, the manual selection of filter parameters can be labor-intensive and subjective, often resulting in suboptimal noise reduction.

There has been a considerable effort to remove noise from calcium imaging data and several frameworks have been proposed. Current denoising approaches for fluorescence imaging largely rely on temporal correlations in the signal and are optimized for large somatic structures. DeepInterpolation[3], developed for somatic GCaMP6f imaging, uses frame interpolation between time points (similar to Noise2Void[4] and Noise2Self[5]). DeepCAD[6] and DeepCAD-RT[7] use a 3D U-Net that is trained on two split sub-stacks consisting of interlaced frames from the raw somatic GCaMP6f recording leveraging the spatio-temporal features of signal in comparison to noise. Training 3D U-Net can be challenging since the high amount of parameters allows overfitting easily[6]. DeepSeMi[8], though more versatile with multiple organelle markers, also incorporates temporal information through its hybrid spatial-temporal network. SUPPORT[9] integrates spatial and temporal information to denoise a range of fast voltage sensors that show a somatic localization. However, the authors show its applicability to other functional imaging sensors such as GCaMP8f and also conduct

limited tests for the synaptic iGluSnFR and iGABASnFR. Recently, Wang and
colleagues introduced CellMincer[10] a two-stage denoising framework consisting of a
frame-wise 2D U-Net combined with a pixel-wise 1D convolutional model to capture
spatial and temporal dimensions respectively. CellMincer is specifically trained to
denoise fast voltage sensor recordings.

Previous approaches mostly share two key limitations. First, they are primarily
optimized for large cellular structures like somata rather than synapses. Second, their
dependence on the temporal dimension necessitates retraining for different recording
frequencies and sensor combinations, which demands substantial data collection and
computational resources for each new experimental condition.

Commonly used sensors like somatic GCaMP6f, which measure calcium influx, exhibit
slower dynamics compared to diffusion-dominated local release of neurotransmitters.
Thus, monitoring synaptic glutamate release requires considerably higher spatial and
temporal resolution[11]—the glutamate sensor iGluSnFR3.v857.SGZ exhibits orders of
magnitude faster kinetics than somatic calcium indicators used in most previous
studies, with signals originating from tiny synaptic release sites orders of magnitude
smaller than somata. This combination of rapid dynamics and synaptic localization
presents unique challenges: the signals are spatially confined, temporally sparse, and
accompanied by higher background noise compared to somatic recordings.

While existing denoising methods excel at enhancing slow somatic signals by
leveraging their temporal continuity, they may introduce artifacts when applied to rapid
synaptic events that appear discontinuous between frames. The need to retrain models
for each combination of sensor and recording frequency further complicates their
application to diverse imaging conditions. Therefore, a more flexible approach that can
adapt to different temporal dynamics without retraining and specifically addresses the
spatial scale of synaptic signals is needed to preserve the fast dynamics and spatial
precision of synaptic glutamate imaging.

To address these specific challenges, we developed Neuroimage Denoiser, a novel
model that effectively removes independent noise from microscopic recordings. Unlike
many recent methods, the training and architecture of Neuroimage Denoiser makes
the denoising process independent of the recording frequency and the kinetics of the
sensor used. This flexibility sets Neuroimage Denoiser apart from previous methods
and makes it a versatile tool for various experimental setups, including fast calcium
imaging.

Neuroimage Denoiser has minimal hardware requirements, making it accessible for
many laboratories to create a custom version tailored to their specific experimental
setup. Neuroimage Denoiser enables researchers to obtain high-quality, noise-
reduced data without the need for extensive computational resources or expertise.

Material and Methods

Primary hippocampal mouse cultures

Co-authors of the paper acquired the data used for different projects. The use of mice in research was approved by District administration Mainz-Bingen, 41a/177-5865-§11 ZVTE, 30.04.2014. Primary mouse hippocampal cultures were generated from newborn pups[12]. Animals were sacrificed according to the European and local government (Rheinland-Pfalz, Germany) regulations for animal welfare. Briefly, newborn pups of either sex were anesthetized on ice and sacrificed using the decapitation method. Following hippocampal isolation, cells were dissociated with trypsin (Thermo Fisher Scientific; Cat#: 15090-046) and plated on poly-L-lysine-coated 18 mm glass coverslips with a density of 70,000 cells per coverslip. After 1 h incubation in Dulbecco's Modified Eagle Medium (Thermo Fisher, Gibco; Cat#: 41966-029/D6429) supplemented with 10% fetal bovine serum and 0.5% glutamate at 37°C, coverslips were transferred to 12-well plates containing 1mL of growth medium composed of Neurobasal-A (Thermo Fisher, Gibco; Cat#: 12349-015) with 2% GlutaMax (Thermo Fisher, Gibco; Cat#: 35050-038), 2% B27 (Thermo Fisher, Gibco; Cat#: 17504-044), 0.1 M sodium pyruvate (MERK, Sigma; Cat#: S8636). Cultures were maintained in a humidified incubator with an atmosphere of 95 % air and 5 % CO_2 at 37°C.

Neurons transfection

To perform functional imaging experiments, neurons were transfected at DIV3-5 using the calcium phosphate method with either the glutamate sensor iGluSnFR3.v857.SGZ (Addgene; Cat#: 178330)[13], iGluSnFR.S72A (Addgene; Cat#: 106176)[14], or the calcium sensor pAAV-hSyn-synaptophysin-GCamp6f (cloned by Dr. Arun Chhikara from Addgene plasmid Cat#: 67634 and phSyn-Synapthophysin-GCamp6f). Briefly, the neuron's growth medium was exchanged with 1 mL of pre-warmed Neurobasal-A medium, and the original medium was stored and kept at 37°C. In the meantime, per 18 mm coverslip, 30 μL of transfection buffer containing in mM: 274 NaCl, 9.5 KCl, 1.4 Na_2HPO_4, 15 Glucose, 15 HEPES, pH 7.14 were added to a solution containing 3 μg of DNA, 3 μl of $CaCl_2$ (2M) and 24 μL of H_2O (q.s. 30 μL). The resulting DNA/$CaPO_4$ mix was incubated for 20 min at RT and then added in a proportion of 60 μl of mix per well. Incubation on neurons was maintained at 37°C until precipitate formation (40 min). Following DNA/$CaPO_4$ incubation, neurons were washed two times for 5 min with 1X HBS buffer containing in mM: 137 NaCl, 4.75 KCl, 0.7 Na2HPO4, 7.5 Glucose, 7.5 HEPES, pH 6.9 and for one time for 5 min with Neurobasal-A medium. Following the washing steps, the original growth medium was added, and neurons were stored in the incubator at 37°C until live imaging experiments were performed at DIV16 - 19.

Functional Imaging

All live experiments were conducted on hippocampal cultures at DIV16 - 19 at 37°C
within a whole microscope incubator. Neurons were perfused with extracellular solution
containing (in mM): 145 NaCl, 2.5 KCl, 10 HEPES, and 10 D-glucose 2 $MgCl_2$, 2 $CaCl_2$
(pH 7.4). Network activity was suppressed by adding 10 μM CNQX (Tocris; Cat#: 1045)
and 10 μM D-AP5 (Tocris; Cat#: 0106) to the extracellular solution as indicated for the
respective experiment. Experiments were performed with an inverted total internal
reflection fluorescence (TIRF) setup composed of the eclipse Ti microscope (Nikon)
equipped with a 60 x Apo TIRF oil objective (1.49 NA; Nikon). Fluorescence was
excited with a combined laser system (Coherent; MPB Communications Inc.). Images
were captured by a scientific CMOS camera (ORCA-Fusion C14440-20UP,
Hamamatsu) at frame rates between 100 to 1000 Hz controlled by NIS-Elements
Advanced Research acquisition software (Nikon). Electrical field stimulation was
carried out using two platinum electrodes placed on the coverslip at a 10 mm distance.
For glutamate imaging experiments, triggered electrical field stimulation was used to
elicit 20 action-potential-like stimuli. The extracellular stimulus sequences (0.9 ms; 50
mA; 1 s interval) were triggered by an isolated pulse generator (A&M Systems, Model
2100) and delivered via a stimulus isolator unit A385 (WPI). Images were acquired by
image streaming for 2000 frames within a defined region of interest (ROI) 800 x 800
pixels (100 Hz acquisition rate) or 250 x 250 pixels (1000 Hz acquisition rate) with a
pixel size of 72nm. For calcium imaging experiments, each region of interest of 512 x
512 pixels was stimulated using 1,3 and 5 action potential-like stimuli (1 ms, 50 mA,
20 ms interval).

Synaptic ROI and Threshold

Circular synaptic ROIs were selected manually using Fiji (version 1.54f)[15] and set to a
diameter of 9 pixels. The average traces along the z-axis of the synapses were
extracted by either Fiji (version 1.54f)[15] or Python (version 3.10.14)[16] with roifile
(version v2024.3.29)[17].
On the average traces, the threshold was computed using numpy (version 1.26.4)[18]
based on the baseline frames up to the first stimulation, using the formula:

$$Threshold(x) = mean(x) + threshold_{mult} * std(x)$$

Whenever the average trace crossed the computed threshold (set by $threshold_{mult}$),
this was considered to be a response.

Signal-to-noise ratio

To assess the denoising performance, we calculated the signal-to-noise ratio (SNR)
for both raw and denoised glutamate and calcium recordings. SNR was computed for
each synaptic response using the mean traces from synaptic ROIs according to the
formula:

$$SNR = \frac{\Delta F}{\sigma_{baseline}}$$

where ΔF represents the signal amplitude (calculated as the difference between the
peak intensity value and mean baseline), and $\sigma_{baseline}$ is the standard deviation of the

baseline. The baseline was defined as the ten frames preceding the initial threshold
crossing. Statistical comparison between raw and denoised recordings was performed
using a two-tailed paired t-test using scipy (version 1.14.1)[19].
We further compared the single-pixel SNR between the raw and denoised recording,
which was defined as the mean fluorescence value divided by the standard deviation
along the temporal axis using numpy (version 1.26.4)[18]. Subsequently, we randomly
sampled 10,000 pixels and plotted the distribution with a seaborn histplot (version
0.13.2)[20].

Preparing Training Data

Training deep learning models is sensitive to imbalances between foreground and
background[21,22]. Different from synaptic GCaMP6f recordings used for the training of
several previously described denoising methods[3,6,7], the microscopic recording of
synaptic glutamate responses is dominated by background due to the small synapses
that sparsely respond when defining a responding synapse as foreground. Therefore,
the available data is filtered for active regions and frames to remove the background
(Figure 1a). We transformed our recording frames into activity maps to filter efficiently.
First, the data is transformed by a pixel-wise rolling window (along 50 frames) z-
normalization (Figure 1b). The rolling window adapts for potential bleaching in the
recording.

$$z\ score_{frame} = \frac{x_{frame} - \mu_{window_{start}:window_{end}}}{\sigma_{window_{start}:window_{end}}}$$

An activity map is calculated framewise and represents the maximum z-score within a
local region for each spatial position in each frame. Each frame is divided into crops of
a predefined size of 32 x 32 pixels to compute the activity map. A crop is assigned with
the maximum value of the kernel-averaged region, obtained by sliding a quadratic
kernel over the crop (step size 1) and computing the average intensity within that kernel
at each position (Figure 1c, right). The kernel size corresponds to the expected minimal
region of interest (ROI) size, reflecting the spatial extent of the structures under
investigation (6 pixels for synapses in our setup). The activity map effectively captures
and highlights regions exhibiting high-intensity fluctuations that are potential indicators
of neuronal activity, by assigning the maximum averaged kernel value to each crop.
Next, these activity maps are filtered for crops that exceed a minimum z-score of 3.
Subsequently, the pixel-wise z-score (without using a rolling window) is computed, and
the transformed crop is stored in an h5 file for training, using h5py (version 3.11.0).

Neuroimage Denoiser can be run on an entire directory storing multiple recordings and
will automatically prepare the training data without manual intervention. Additionally, a
memory-optimized version is implemented that necessitates less RAM at the cost of a
longer run-time.

Model Architecture and Training

The U-Net model[23] is implemented in PyTorch (version 2.1.1.)[24] with a single input and

output channel (Figure 1a). The U-Net architecture consists of an encoder and a

decoder. The encoder comprises blocks containing two 2D convolutional layers with
3x3 kernels, ReLU activation, and batch normalization, followed by a 2D max-pooling
layer for downsampling. These blocks are referred to as down-blocks. Skip
connections connect the corresponding down-blocks to the up-blocks in the decoder
without any processing.
The decoder comprises up-blocks, each consisting of a 2D transpose convolutional
layer for upsampling, followed by a concatenation with the respective feature map from
the corresponding down-block (skip connection). This concatenated output is then
processed by two 2D convolutional layers with 3x3 kernels, ReLU activation, and batch
normalization. The final layer is a 1x1 2D convolutional layer that maps the feature
vector to the denoised image. This architecture allows the network to capture both local
and global features, enabling accurate denoising of the input image. Using a fully
convolutional model without dense layers allows flexible input image sizes during
inference, limited only by GPU memory, overcoming fixed training input constraints.
Training is performed on z-transformed patches containing active synapses for one
epoch using a batch size of 64, the ADAM optimizer with a learning rate of 0.0001, on
a consumer-grade NVIDIA GeForce 2060 super graphics card with 8GB of vRAM. A
grid search for the parameters listed in Table 1 was performed to find a suitable
combination of parameters for the training. The model's objective is to reduce noise in
the recordings while maintaining the integrity of the peak response amplitudes.
To evaluate the model's performance in preserving the amplitude of neural responses,
regions of interest (ROIs) exhibiting clear stimulus-evoked responses were manually
identified in the raw recorded data. The peak response amplitudes within these ROIs
were measured following the presentation of a stimulus. After training the model, the
same ROIs were analyzed in the denoised output, and the corresponding peak
response amplitudes were extracted. The model's accuracy in preserving the response
amplitudes is quantified by calculating the Mean Absolute Percentage Error (MAPE)
using scikit-learn (version 1.5.0)[25] between the raw and denoised peak amplitudes
across all ROIs. The correlation between raw and denoised peak amplitudes was
calculated using scipy (version 1.14.1)[19].

$$MAPE = 100 * \frac{1}{n} \sum_{t=1}^{n} \left| \frac{peak_t^{raw} - peak_t^{denoised}}{peak_t^{raw}} \right|$$

Additionally, it is crucial that the model efficiently removes noise from the microscopic
recordings. This is measured by the standard deviation of non-responding sections of
the recordings.
Both metrics (MAPE and standard deviation) are ranked, and a combined score is
calculated. The best-performing model is selected based on the minimal combined
score achieved (supplementary table 1).

$$score = \frac{(Rank_{std} + 2 * Rank_{MAPE})}{3}$$

283
284 *Table 1 Model Optimization Parameters: This table presents the various parameters tested during the*
285 *optimization of the Neuroimage Denoiser model. For each parameter, different variables were tested to determine*

286 *the most effective configuration. The "Best performing Model" column lists the values that yielded the best*
287 *performance based on the evaluation metrics.*

Parameter	Tested Variables	Best performing Model
Loss Function	L1, Smooth-L1, MSE, Huber	Smooth-L1
Use Gaussian Filter on y_{True}	True, False	True
Sigma Gaussian Filter	0.5, 1.0	1.0
Noise Center	0, 0.5, 1.0	0
Noise Scale	0.5, 1.0, 1.5, 2.0, 3.0	2.0

Inference/Denoising

The denoising (or inference) using Neuroimage Denoiser can be done on single
images or entire directories containing multiple sub-directories while maintaining the
original folder structure. The process is schematically visualized in Figure 1e. First, the
input image is loaded in tiff or nd2 format using tifffile (version 2023.12.9)[26] or nd2
(version 0.8.1) respectively and a z-normalization (along the time axis; no rolling
window) is applied. The mean and standard deviation per pixel are temporarily saved.
Neuroimage Denoiser performs the denoising on the z-normalized frames and
afterward, the resulting frames are transformed back into intensity values using the
saved mean and standard deviation. The denoised image is saved as a tiff file using
tifffile (version 2023.12.9)[26].

Reimplementation of DeepInterpolation

To fairly compare the performance of DeepInterpolation[3] with the approach of
Neuroimage Denoiser, we re-implemented the model in PyTorch (version 2.1.1)[24].
Therefore, the model can be trained on crops of active synapses (32 x 32 pixels) and
comparability was ensured. The number of frames before and after the target frames
was reduced to five frames to account for the faster dynamics of glutamate sensors
(Supplementary Figure 1a). The model was trained using the ADAM optimizer with an
L1 loss function, and a learning rate of 0.0001 on an NVIDIA GeForce GTX2060 super
graphics card.

Data Visualization

310 Data was visualized using Python (version 3.10.14)[16], matplotlib (version 3.9.0)[27], and
311 seaborn (version 0.13.2)[20]. Schematic drawings were made using Adobe Illustrator.

Results

Denoising glutamate imaging data presents unique challenges compared to calcium
imaging, as glutamate sensors are orders of magnitude faster and the synaptic
localization results in considerably less foreground signal relative to somatic calcium
recordings. Here, we present Neuroimage Denoiser, a U-Net-based imaging denoising

tool (Figure 1a) specifically trained to remove noise from microscopic glutamate time-
series recordings.
To curate a balanced training dataset adjusted to the challenges posed by glutamate
imaging data, the recordings are automatically screened to identify frames and regions
exhibiting active synaptic signals (Figure 1a) in the training preparation step without
manual intervention. This screening can be adapted to different ROI sizes representing
the size in pixels of the biological structure where the sensor is localized. We set the
ROI size to 6 pixels to account for the size of synapses harboring the sensor in our
microscopic set-up.
Initially, we explored training a model in the style of DeepInterpolation[3].
DeepInterpolation predicts a target frame based on a set number of frames preceding
and succeeding the target frame (n_{pre}, n_{post})[3]. To account for the fast dynamics of the
iGluSnFR3 sensor compared to Gcamp6f, we set $n_{pre} = 5$ and $n_{post} = 5$ and the model
predicts the target frame in the middle (Supplementary Figure 1a). However, this
temporal interpolation approach, while effective for slower calcium dynamics, did not
adequately preserve the fast glutamate response and led to artifacts in the amplitudes
(Supplementary Figure 1b).
Given the limitations of an approach that leverages the temporal dimension for
denoising for capturing these rapid synaptic signals, we turned to the Noise2Noise
concept proposed by Lehtinen and colleagues[28]. Their work demonstrated image
restoration without ground truth data by training networks to distinguish signals from
noise patterns by adding artificial noise to noisy images. Further, this simple approach
was the best-performing approach in the Fluorescence Microscopy Denoising dataset
benchmarking, outperforming other methods including DnCNN by almost 2dB in
PSNR[29]. Leveraging this approach, Neuroimage Denoiser is trained according to the
Noise2Noise[28] method on individual frames and subsequently circumvents the
dependence on the recording frequency or the dynamics of the sensor. Therefore, a
trained model can be applied to a broad range of experimental set-ups independently
of the recording frequency and sensor dynamics.
346

Figure 2 a) Schematic representation of the architecture used for Neuroimage Denoiser is a U-Net[23] with one input and output channel. The data runs through the encoder and decoder part of the network while skip connections preserve untransformed data. Convolutional layers (orange) use a 3 by 3 filter with ReLu activation. The output layer is a linear layer that does not use an activation function. b) Representative area of a glutamate recording acquired at 100Hz. The stimulus is given in frame 100 and the peak response is in frame 102. After denoising, the amplitude height and time points are well preserved, and the noise level is minimal. c) Neuroimage Denoiser denoised amplitude heights of manually selected ROIs tested against the raw amplitude heights, showing a high correlation (R=0.9966, p<0.0001). d) Single-pixel SNR distributions for raw and denoised recordings. e) Denoising inference speed (frames per second) tested for frame sizes 50 - 800 pixels (averaged values of three repetitions of recordings with 1,000 frames). The image was processed according to Figure 1e)

We conducted a grid search to identify the optimal combination of training parameters (Table 1) and selected the model that best balanced noise removal and amplitude preservation (Supplementary Table 1). The best-performing model was trained using a smooth L1-loss. The training data included Gaussian noise added with a noise center of 0.0 and a noise scale of 2.0. Notably, higher noise scale and center values yielded models that removed noise more effectively but at the cost of less reliable amplitude preservation (Supplementary Table 1). The selected model effectively removes noise from the raw recording, both visually as measured in the z-profile of the synaptic ROI

(Figure 2b), while preserving the amplitudes well (Figure 2c), as evidenced by a
significant correlation between raw and denoised amplitudes (R=0.9966, p < 0.0001).
The model's denoising performance is further characterized by the distribution of
single-pixel SNR values (Figure 2d). While raw recordings show a narrow distribution
concentrated at low SNR values, denoised recordings exhibit a broader distribution
shifted towards higher SNR values, indicating comprehensive noise reduction across
the entire field of view. This global improvement in signal quality is also reflected in the
significant 5.65 ± 0.33 improved SNR for the peaks (mean $\pm$ SEM, p < 0.0001, paired
one-sided t-test).
Neuroimage Denoiser demonstrates fast inference, as tested on three recordings of
1000 frames with varying spatial dimensions (50 to 800 pixels for x and y). The entire
process, including normalization and reverse normalization, as shown in Figure 1e,
requires between 0.0003 and 0.1519 seconds per frame (Figure 2e). Therefore,
denoising 800 x 800 pixels recording of the length of 1,000 frames takes ~152 seconds.
380

Neuroimage Denoiser enables precise Signal Localization across diverse Recording Conditions

Only infrequent spontaneous activity can be observed when recording a neuronal
culture in the absence of electric stimulation. Specifically, it is challenging to
differentiate the components at complex synapses, which comprise an accumulation
of pre-synaptic inputs and post-synaptic receivers. These synapses often involve
multiple release sites in close proximity, and the released glutamate can diffuse from
its initial release site, leading to highly correlated signals at neighboring post-synaptic
sites (Figure 3a, left). However, the maximum localization of a spontaneous release
event indicates the release site. Due to the spatial proximity within these complex
synapses, identifying and differentiating post-synaptic receiving sites is difficult when
the signal is superimposed by noise (Figure 3a, top). After removing the noise using
Neuroimage Denoiser, each of the three distinct sites within the complex synapse
becomes visible and can be accurately localized (Figure 3a, bottom).

Figure 3 a) Complex synapse with three sites that receive spontaneous input from different pre-synaptic sites. The left column shows the raw recording as it was acquired at the microscope, and the right column shows the same frames denoised by Neuroimage Denoiser. The three different sites showing spontaneous activity are visible. The traces at these sites are highly correlated due to their proximity. b) Neuroimage Denoiser works on recordings acquired with very high recording frequencies (tested here with 1000 Hz), although being trained on 100 Hz recordings exclusively. c-e) Single-pixel SNR distributions for raw and denoised recordings at different acquisition frequencies (200 Hz, 500 Hz, and 1000 Hz). f) Correlation between raw and denoised signal amplitudes, color indicates recording frequency (gray: 100 Hz, light blue: 200 Hz, blue: 500 Hz, dark blue: 1000 Hz)

While Neuroimage Denoiser was trained exclusively on recordings done with 100 Hz, the model is capable of denoising recordings acquired at significantly higher acquisition rates, such as 1000 Hz (Figure 3b). The ability to acquire data at such high frequencies is valuable for capturing rapid synaptic events and amplitudes with high temporal

precision. However, high noise levels often complicate these high-frequency
recordings, making it challenging to identify true responses amidst false-positive
threshold crossings (Figure 3b, top). The denoised recording, on the other hand, has
a considerably improved signal-to-noise ratio (Figure 3b, bottom), like the improvement
observed in previous recordings at 100 Hz.
Quantitative analysis of single-pixel SNR distributions across different acquisition
frequencies (200 Hz, 500 Hz, and 1000 Hz) demonstrates consistent noise reduction
capabilities of Neuroimage Denoiser (Figure 3c-e). Further, we also specifically
analyzed the SNR for responses at synapses to evaluate the biologically relevant
signal improvement. At these functional sites, the denoising significantly improved the
SNR with an average 5.21 ± 0.19-fold (mean ± SEM, paired one-sided t-test, p <
0.0001), 7.72 ± 1.18-fold (mean ± SEM, paired one-sided t-test, p < 0.0001), and 8.97
± 1.18-fold (mean ± SEM, paired one-sided t-test, p < 0.0001) enhancement for 200
Hz, 500 Hz, and 1000 Hz recordings respectively.
Therefore, Neuroimage Denoiser can facilitate the analysis of recordings acquired at
very high frequencies without the need to re-train a new model. Acquiring at these high
rates opens the door to various applications, such as analyzing multi-vesicular or
asynchronous releases from a single synapse, which would yield variable amplitudes
and subsequent peaks, or resolving the precise time from stimulus to glutamate
release with higher temporal resolution. Additionally, it enables the study of rapid
synaptic plasticity mechanisms and the precise kinetics of neurotransmitter release
and clearance. The camera-based imaging will allow the comparison of direct multiple
synapses in the field of view, which will outcompete scanning microscopy-based
approaches, where only single synapses can be imaged at necessary high acquisition
rates (>100 Hz).
433

Application of Neuroimage Denoiser on alternative Glutamate and Calcium Sensor to study Synaptic Plasticity

Since Neuroimage Denoiser is not bound to a certain sensor and its associated kinetics
but only to its localization, we applied the denoising to recordings of fast
iGluSnFR.S72A[14]. This version of the sensor has faster kinetics, but a lower signal-to-
noise ratio compared to iGluSnFR3.v857.SGZ[13] particularly due to the high noise level
present in the baseline of the recording (until stimulation at frame 100). This noise in
the baseline elevates the threshold above which a signal is considered a release event.
iGluSnFR.S72A[14] is useful for plasticity studies since the kinetic properties allow the
recording of synaptic release events up to 20 Hz pulse interval. Conventionally such
recordings are done via patch-clamp[14], which will not allow reaching the single synapse
resolution as easily as imaging approaches. Conversely, iGluSnFRs timing and
localization of excitatory synaptic inputs[11] enable a more complete picture of the
synaptic plasticity of a neuron under a given condition, which is of particular interest
for investigating the molecular mechanisms underlying synaptic plasticity. Key
plasticity metrics are the paired-pulse ratio, quantifying facilitation or depression of

consecutive responses, and the failure rate, measuring the proportion of stimuli failing
to evoke a response hence indicating release probability. Nevertheless, due to the low
signal-to-noise level and the fast bleaching of the sensor, the recordings are
superimposed by noise (Figure 4a, left) that considerably complicates the identification
of responding synapses and to differentiate between response and failure within the
average trace of a selected synapse (Figure 4b, top). The given example in Fig. 4
illustrates the impact of noise in fast pulse trains. Even though a very small threshold
multiplier of two is applied to the raw average trace the responses to the stimulus given
at frames 130 and 140 are not detected. Moreover, the noise in the raw data (Figure
4b, top) obscures these responses, while the denoised trace (Figure 4b, bottom)
reveals clear and distinct responses upon each of the five stimuli. Thus, the failure rate
can be determined to be 0% after applying denoising with Neuroimage Denoiser
(threshold multiplier 2), contrary to 40% in the raw trace (threshold multiplier 2).
Additionally, when computed with a multiplier of two (yellow dashed line, Figure 4b),
the low threshold leads to the detection of false positive responses in the raw trace.
Quantitative analysis of the recordings demonstrated significant improvements in
signal quality after denoising. The single-pixel SNR distributions (Figure 4c) show a
marked shift toward higher values in the denoised data compared to the raw
recordings. Peak analysis revealed a 2.00 ± 0.12-fold improvement in SNR (mean $\pm$
SEM, $p < 0.001$, paired t-test). Importantly, the denoising process preserved the
original signal amplitudes, as evidenced by the strong correlation between raw and
denoised measurements ($R = 0.9998$, $p < 0.0001$; Figure 4d).
472

473
Figure 4 a) A crop of the recording with iGluSnFR.S72A[14]. The left panel represents the raw data with high noise
levels, while the right panel shows the denoised data using Neuroimage Denoiser. b) Average intensity trace for
ROI 2 over the recorded 250 frames. Stimulation is applied at frames 100, 110, 120, 130, and 140. The top graph
is extracted from the raw recording and is characterized by significant noise and fluctuations. The bottom graph
shows the denoised average intensity trace from ROI 2, where the noise level is substantially reduced, allowing
clearer identification of responses to stimuli. Particularly, responses to the stimulation in frames 130 and 140,
which are within the noise level in the raw trace, become visible in the denoised trace. The threshold for two low
threshold multipliers is indicated as yellow and orange dashed lines for multipliers two and three, respectively. c)
Single-pixel SNR distributions for raw and denoised recordings d) Correlation between raw and denoised signal
amplitudes.

Next, we tested whether Neuroimage Denoiser can denoise recordings of pAAV-hSyn-
synaptophysin-GCamp6f that emits fluorescence upon the binding of Calcium.
Although this version of GCamp6f is also synaptically located, it shows far slower
dynamics than the two iGluSnFR versions used in this study. Our results demonstrate
that Neuroimage Denoiser effectively preserves GCaMP6f signals across different
stimulation intensities, from 1AP to 5APs (Figure 5a). The denoised fluorescence
traces closely match the temporal profile of the raw signals while substantially reducing
noise, as evidenced by the smoother baseline and improved signal clarity (Figure 5a).
Quantitative analysis revealed a significant improvement in single-pixel SNR in the
denoised recordings (Figure 5b), showing an average 4.47 ± 0.23-fold improvement
(mean ± SEM, $p < 0.0001$, paired one-sided t-test) for the synaptic responses.

Importantly, the denoising process maintained the original signal amplitudes, showing a very high correlation between raw and denoised measurements (R = 0.9995, p < 0.0001; Fig. 5c). This robust preservation of signal integrity was consistent across all stimulations (1, 3 and 5 Aps).

Figure 5 a) GCaMP6f responses to 1, 3, and 5 action potentials (APs) showing raw recordings (left) and corresponding denoised frames by Neuroimage Denoiser (right). Both fluorescence traces and peak response frames demonstrate signal preservation across different response magnitudes. b) Single-pixel SNR distribution for raw and denoised recordings. c) Correlation between raw and denoised signal amplitudes demonstrates high-fidelity preservation of response magnitudes (R = 0.9995, p < 0.0001)

Discussion

In this manuscript, we presented Neuroimage Denoiser – a U-Net-based[23] deep-learning framework to reduce noise levels in microscopic imaging recordings. By

incorporating the concepts of Noise2Noise[28], the Neuroimage Denoiser framework is
highly flexible for a wide array of experimental settings and does not require re-training
for changed parameters. The Neuroimage Denoiser uses a self-supervised training
approach without requiring noise-free ground truth. By automatically selecting training
data on responsive structures like synapses or somata (on both spatial and temporal
dimensions), the framework avoids overrepresenting background regions.
In the past, researchers developed several sensors to study functional neurobiology
for the dynamics of calcium[30], GABA[31], dopamine[32], acetylcholine[33], and glutamate[13,31]
on a single synapse level. These recordings are superimposed with noise regardless
of the sensor and acquisition rate. Consequently, a unified, flexible model capable of
denoising recordings with the same localization would considerably ease the
downstream processing.
Our comprehensive analysis demonstrates the framework's effectiveness through both
SNR improvements for peaks in responding synapses and single-pixel SNR
distributions throughout the entire field of view. We tested our model, trained solely on
iGluSnFR3.v857.SGZ[13] data acquired at 100Hz, by applying the denoising routine to
changed parameters, i.e., recording frequency (200, 500, and 1000 Hz; Figure 3) and
sensor (iGluSnFR.S72A[14] and GCaMP6f; Figure 4 and Figure 5), and obtained
convincing results. Thus, the framework provides out-of-the-box utility for an array of
changing experimental parameters and drastically reduces the need for re-training.
The proposed framework is capable of significantly reducing the noise in experiments
with fast dynamics, as typical for glutamate imaging. One of the advantages of
Neuroimage Denoiser is its ability to preserve the amplitude of synaptic responses, as
evidenced by a high correlation between denoised and raw responses (Figure 2c,
Figure 3f, Figure 4d, and Figure 5b). In all tested conditions, namely, changed
recording frequency and the iGluSnFR3.v857.SGZ[13] or GCamp6f sensor, the SNR for
the responses was significantly improved by 5 to 9-fold depending on the observed
noise level in the raw recordings. Only the iGluSnFR.S72A[14] showed a lower average
2-fold improvement in SNR, likely because of the comparably weak response
amplitudes. This consistent performance across different sensors and experimental
conditions underscores the framework's robustness and flexibility, which is a key
strength of single-frame denoising approaches. Another important aspect is that this
improved signal-to-noise level helps to resolve complex synapses spatially (Figure 3),
can ease the analysis of spontaneous activity, and can be used as a functional readout
for molecular manipulation to define release mechanisms in specific synapses.
We anticipate that Neuroimage Denoiser can be adopted in many laboratories as it
has minimal hardware requirements, and the preparation of the training data is
automatically balanced for optimal results. The framework is easy to use for data
preparation, training, grid search, and inference (denoising) and does not require
additional programming knowledge, making it accessible to a broad range of
researchers in the field of neuroscience. Although glutamate imaging as a neural
recording method can provide rich and diverse datasets, these data have a high
complexity with many disruptive elements, such as noise, bright sensor depositions, or

vesicular transport of sensor, that make the creation of ground truth data for training a model on responding regions of a recording challenging. We tested several potential training parameters (Table 1) to select the most performant parameter combination. However, users can freely choose a combination suited for their experimental set-up and analysis.

While initially developed for fast glutamate imaging responses, Neuroimage Denoiser has proven to be a versatile framework. Its successful application to both fast glutamate dynamics and slower calcium signals, along with consistent improvements in single-pixel SNR and SNR for responses across different experimental conditions, suggests it could serve as an out-of-the-box tool for enhancing the quality of various functional imaging experiments in neuroscience.

Code availability

The code is available at https://github.com/s-weissbach/neuroimage_denoiser.

Data availability

Data used for the training of Neuroimage Denoiser is available on Zenodo https://zenodo.org/records/14213082.

Acknowledgment

The authors acknowledge financial support from the Emergent Algorithmic Intelligence Centre funded by the Carl-Zeiss Foundation. We thank Anita Heine, Nathalie Philipp, and Corinna Werkmann for technical support and excellent hippocampal culture preparation. The authors thank Stanislav Sys for critical feedback and discussion. We thank Arun Chhikara for the cloning of the GCaMP6f construct used in this study.

Authors contribution

SW and MH conceived the idea and conceptualized the work. MB, CA, AK, and MH conducted experiments. SW conceptualized and implemented the Neuroimage Denoiser framework. SW and JM implemented the data preparation routine. SW created visualizations. SW, MB, and MH wrote the manuscript. SG revised the manuscript. MH and SG supervised the project and acquired funding.

Declaration of Interests

The authors declare no competing interests.

Ethics statement

All procedures involving animals were conducted in accordance with the Guidelines for the Care and Use of Laboratory Animals and approved by the District administration Mainz-Bingen (license number 41a/177-5865-§11 ZVTE, 30.04.2014). Animal handling complied with European Union guidelines (Directive 2010/63/EU) and local government (Rheinland-Pfalz, Germany) regulations for animal welfare (Tierschutzgesetz §4). No in vivo experiments were conducted.

References

1.   Wilt, B. A., Fitzgerald, J. E. & Schnitzer, M. J. Photon Shot Noise Limits on Optical Detection of Neuronal Spikes and Estimation of Spike Timing. *Biophys. J.* **104**, 51–62 (2013).

2.   Buades, A., Coll, B. & Morel, J.-M. A non-local algorithm for image denoising. in *2005 IEEE Computer Society Conference on Computer Vision and Pattern Recognition (CVPR'05)* vol. 2 60–65 vol. 2 (2005).

3.   Lecoq, J. *et al.* Removing independent noise in systems neuroscience data using DeepInterpolation. *Nat. Methods* **18**, 1401–1408 (2021).

4.   Krull, A., Buchholz, T.-O. & Jug, F. Noise2Void - Learning Denoising From Single Noisy Images. in *2019 IEEE/CVF Conference on Computer Vision and Pattern Recognition (CVPR)* 2124–2132 (IEEE, Long Beach, CA, USA, 2019). doi:10.1109/CVPR.2019.00223.

5.   Batson, J. & Royer, L. Noise2Self: Blind Denoising by Self-Supervision. in *Proceedings of the 36th International Conference on Machine Learning* 524–533 (PMLR, 2019).

6.   Li, X. *et al.* Reinforcing neuron extraction and spike inference in calcium imaging using deep self-supervised denoising. *Nat. Methods* **18**, 1395–1400 (2021).

7.   Li, X. *et al.* Real-time denoising enables high-sensitivity fluorescence time-lapse imaging beyond the shot-noise limit. *Nat. Biotechnol.* **41**, 282–292 (2023).

8.   Zhang, G. *et al.* Bio-friendly long-term subcellular dynamic recording by self-supervised image enhancement microscopy. *Nat. Methods* **20**, 1957–1970 (2023).

9.   Eom, M. *et al.* Statistically unbiased prediction enables accurate denoising of voltage imaging data. *Nat. Methods* **20**, 1581–1592 (2023).

10.   Wang, B. *et al.* Robust self-supervised denoising of voltage imaging data using CellMincer. 2024.04.12.589298 Preprint at https://doi.org/10.1101/2024.04.12.589298 (2024).

11.   Marvin, J. S. *et al.* An optimized fluorescent probe for visualizing glutamate neurotransmission. *Nat. Methods* **10**, 162–170 (2013).

12.   Kaech, S. & Banker, G. Culturing hippocampal neurons. *Nat. Protoc.* **1**, 2406–2415 (2006).

13.   Aggarwal, A. *et al.* Glutamate indicators with improved activation kinetics and
localization for imaging synaptic transmission. *Nat. Methods* **20**, 925–934 (2023).
14.   Marvin, J. S. *et al.* Stability, affinity, and chromatic variants of the glutamate
sensor iGluSnFR. *Nat. Methods* **15**, 936–939 (2018).
15.   Schindelin, J. *et al.* Fiji: an open-source platform for biological-image analysis.
Nat. Methods **9**, 676–682 (2012).
16.   The Python Language Reference — Python 3.10.12 documentation.
https://docs.python.org/3.10/reference/.
17.   Gohlke, C. cgohlke/roifile: v2024.5.24. Zenodo
https://doi.org/10.5281/zenodo.11266632 (2024).
18.   Harris, C. R. *et al.* Array programming with NumPy. *Nature* **585**, 357–362
(2020).
19.   Virtanen, P. *et al.* SciPy 1.0: fundamental algorithms for scientific computing in
Python. *Nat. Methods* **17**, 261–272 (2020).
20.   Waskom, M. L. seaborn: statistical data visualization. *J. Open Source Softw.* **6**,
3021 (2021).
21.   Saini, M. & Susan, S. Tackling class imbalance in computer vision: a
contemporary review. *Artif. Intell. Rev.* **56**, 1279–1335 (2023).
22.   Chen, J., Wu, Q., Liu, D. & Xu, T. Foreground-Background Imbalance Problem
in Deep Object Detectors: A Review. in *2020 IEEE Conference on Multimedia*
Information Processing and Retrieval (MIPR) 285–290 (2020).
doi:10.1109/MIPR49039.2020.00066.
23.   Ronneberger, O., Fischer, P. & Brox, T. U-Net: Convolutional Networks for
Biomedical Image Segmentation. Preprint at
https://doi.org/10.48550/arXiv.1505.04597 (2015).
24.   Paszke, A. *et al.* PyTorch: An Imperative Style, High-Performance Deep
Learning Library. Preprint at https://doi.org/10.48550/arXiv.1912.01703 (2019).
25.   Pedregosa, F. *et al.* Scikit-learn: Machine Learning in Python. *Mach. Learn.*
PYTHON.
26.   Gohlke, C. cgohlke/tifffile: v2024.5.22. Zenodo
https://doi.org/10.5281/zenodo.11256883 (2024).
27.   Ari, Niyazi & Ustazhanov, Makhamadsulton. Matplotlib: A 2D Graphics
Environment. *IEEE J. Mag.* (2014).
28.   Lehtinen, J. *et al.* Noise2Noise: Learning Image Restoration without Clean
Data. Preprint at http://arxiv.org/abs/1803.04189 (2018).
29.   Zhang, Y. *et al.* A Poisson-Gaussian Denoising Dataset With Real
Fluorescence Microscopy Images. in *2019 IEEE/CVF Conference on Computer*
Vision and Pattern Recognition (CVPR) 11702–11710 (IEEE, Long Beach, CA, USA,
2019). doi:10.1109/CVPR.2019.01198.
30.   Chen, T.-W. *et al.* Ultrasensitive fluorescent proteins for imaging neuronal
activity. *Nature* **499**, 295–300 (2013).
31.   Marvin, J. S. *et al.* A genetically encoded fluorescent sensor for in vivo
imaging of GABA. *Nat. Methods* **16**, 763–770 (2019).
32.   Sun, F. *et al.* A Genetically Encoded Fluorescent Sensor Enables Rapid and
Specific Detection of Dopamine in Flies, Fish, and Mice. *Cell* **174**, 481-496.e19
(2018).
33.   Borden, P. M. *et al.* A fast genetically encoded fluorescent sensor for faithful in
vivo acetylcholine detection in mice, fish, worms and flies. 2020.02.07.939504
Preprint at https://doi.org/10.1101/2020.02.07.939504 (2020).

673