Comparison of Infinium MethylationEPIC v2.0 to v1.0 for human population epigenetics: considerations for addressing EPIC version differences in DNA methylation-based tools

Beryl C. Zhuang† [1,2], Marcia Smiti Jude† [1,2], Chaini Konwar [1,2], Calen P. Ryan [3], Joanne Whitehead [1,2], Hannah-Ruth Engelbrecht [1,2], Julia L. MacIsaac [1,2], Kristy Dever [1,2], Tran Khanh Toan [4], Kim Korinek [5], Zachary Zimmer [6,7], Kim M. Huffman [8], Nanette R. Lee [9], Thomas W. McDade [10,11], Christopher W. Kuzawa [12], Daniel W. Belsky [13,14], Michael S. Kobor* [1,2,11,15]

† These authors contributed equally and should be considered as co-first authors
* Please address all correspondence concerning this article to Dr. Michael S. Kobor

1. BC Children's Hospital Research Institute, 950 West 28th Avenue, Vancouver, BC, V5Z 4H4, Canada
2. Department of Medical Genetics, Faculty of Medicine, University of British Columbia, Vancouver, BC, V6T 1Z3, Canada
3. Robert N. Butler Columbia Aging Center, Mailman School of Public Health, Columbia University, New York, NY 10032, USA
4. Family Medicine Department, Hanoi Medical University, Hanoi, Vietnam
5. Department of Sociology, University of Utah, Salt Lake City, USA
6. Department of Family Studies and Gerontology, Mount Saint Vincent University, Halifax, Canada
7. Canada Research Chair, Global Aging and Community Initiative
8. Duke University School of Medicine, Durham, NC, 27701, USA
9. USC-Office of Population Studies Foundation, Inc., University of San Carlos, Cebu City, Philippines
10. Department of Anthropology, Northwestern University, Evanston, Illinois, USA
11. Program in Child and Brain Development, CIFAR, Toronto, Ontario, Canada
12. Department of Anthropology and Institute for Policy Research, Northwestern University, Evanston, IL 60208, USA
13. Butler Columbia Aging Center, Columbia University Mailman School of Public Health, New York, NY, USA
14. Department of Epidemiology, Columbia University Mailman School of Public Health, New York, NY, USA
15. The Edwin S.H. Leong UBC Healthy Aging Chair-A UBC President's Excellence Chair, University of British Columbia, Canada

Keywords (3-10)

Epigenetics, DNA methylation, Illumina EPIC array, epigenetic clocks, inflammation and lifestyle biomarkers, cell type deconvolution, replicate probes

Abstract

Background:
The recently launched DNA methylation profiling platform, Illumina MethylationEPIC BeadChip
Infinium microarray v2.0 (EPICv2), is highly correlated with measurements obtained from its
predecessor MethylationEPIC BeadChip Infinium microarray v1.0 (EPICv1). However, the
concordance between the two versions in the context of DNA methylation-based tools, including
cell type deconvolution algorithms, epigenetic clocks, and inflammation and lifestyle biomarkers
has not yet been investigated.
50
Findings:
We profiled DNA methylation on both EPIC versions using matched venous blood samples from
individuals spanning early to late adulthood across three cohorts. On combining the DNA
methylomes of the cohorts, we observed that samples primarily clustered by the EPIC version they
were measured on. Within each cohort, when we calculated cell type proportions, epigenetic age
acceleration (EAA), rate of aging estimates, and biomarker scores for the matched samples on each
version, we noted significant differences between EPICv1 and EPICv2 in the majority of these
estimates. These differences were not significant, however, when estimates were adjusted for EPIC
version or when EAAs were calculated separately for each EPIC version.
60
Conclusions:
Our findings indicate that EPIC version differences predominantly explain DNA methylation
variation and influence estimates of DNA methylation-based tools, and therefore we recommend
caution when combining cohorts run on different versions. We demonstrate the importance of
calculating DNA methylation-based estimates separately for each EPIC version or accounting for
EPIC version either as a covariate in statistical models or by using version correction algorithms.

Background

Infinium Methylation BeadChip microarrays have been widely used to cost-effectively measure the human DNA methylome in large scale and population-wide studies[1–3]. The recently developed Illumina MethylationEPIC BeadChip Infinium microarray v2.0 (900K, EPICv2) features a total of 936,866 probes, encompassing ~77% of the probes in the previous version, the MethylationEPIC BeadChip Infinium microarray v1.0 B5 (850K, EPICv1), and over 200,000 new probes designed for increased coverage of enhancers, open chromatin regions, and CTCF-binding domains[4]. EPICv2 also differs from its predecessor in the overall probe content and utility, with annotation to the most recent GRCh38/h38 human genome build, differences in probe design type and strand switches, incorporation of new "nv" probes targeting recurrent somatic cancer mutations, and 65 single nucleotide polymorphism (SNP) sites, compared to the 59 SNPs on EPICv1, for confirmation of sample identity. Unlike EPICv1 where each probe is unique, EPICv2 includes ~5100 probes that each have between 2-10 replicates, differentiated based on their probe names and sequences[5]. Approximately 143,000 poorly performing probes on the EPICv1 have been removed from the EPICv2, ~73% of which are likely to be influenced by underlying sequence polymorphisms[6,7]. Overall, these modifications in EPICv2 intend to provide wider coverage of the DNA methylome, with optimized performance across primary tissues and cancer cell lines, and extended reliability across diverse human populations[6,7].

Previous iterations of Illumina microarrays (27K, 450K and EPICv1) have been widely used to develop DNA methylation-based tools including cell type deconvolution algorithms[8–10], epigenetic clocks[11–18], interleukin 6 (IL-6) and C-reactive protein (CRP) inflammation markers[19,20], and lifestyle biomarker predictors such as smoking and alcohol use [21]. These tools have been trained on one or more of the previous generation of microarrays, and are based on the property that DNA methylation levels at specific cytosine-guanine dinucleotides (CpGs) highly correlate with cell types, chronological age, and biomarker measures, respectively[11–14,16,17], and a majority of predictive CpGs employed by these tools are retained on EPICv2[6]. It has been previously shown that estimates from some of these DNA methylation-based tools, while highly correlated, are significantly different between 450K and EPICv1[22–24]. With the launch of the recent version of the EPICv2 array, we thus sought to examine whether there are any such differences between EPICv1 and EPICv2, we systematically assessed the concordance of inferred estimates between the two EPIC versions. We also tested whether there were any differences in estimates between EPICv1 and EPICv2 when datasets profiled on the two versions are combined, given that this is an important consideration in meta-analyses which harness statistical power that comes from combining multiple cohorts, and longitudinal studies which include samples profiled on different arrays/versions.

To assess whether EPIC version differences have any implications on DNA methylation measures in the context of DNA methylation-based tools, we examined matched venous blood samples from individuals across three diverse populations spanning early to late adulthood, and profiled DNA methylome using the EPICv1 and EPICv2 arrays. To explore if the high agreement between EPICv1 and EPICv2 reported in cell lines is consistent in primary samples from human populations, we tested the concordance of the two EPIC versions at both array and probe levels and evaluated the potential contribution of EPIC version to overall DNA methylation variation. Next, to investigate whether there are differences in estimates between EPICv1 and EPICv2 from

DNA methylation-based tools built on previous arrays/versions, including immune cell type
deconvolution algorithms, epigenetic clocks, inflammation and lifestyle biomarkers, we compared
their estimates between the two EPIC versions. We also examined whether EPIC version
influenced these estimates by combining EPICv1 and EPICv2 samples, and demonstrated the
significance of accounting for EPIC version in these DNA methylation-based estimates.

Results

Unsupervised clustering of samples was primarily determined by EPIC version
To compare the two most recent generations of Illumina MethylationEPIC BeadChip Infinium
microarrays, we measured the DNA methylomes of a subset of venous whole blood samples on
both EPICv1 and EPICv2 across three cohorts: (i) Vietnam Health and Aging Study (VHAS)[25],
(ii) Cebu Longitudinal Health and Nutrition Survey (CLHNS)[26], and (iii) Comprehensive
Assessment of Long-term Effects of Reducing Intake of Energy (CALERIE)[27] (Table 1, Figure
1). We first explored whether samples clustered depending on which EPIC version they were
measured on or the cohort they belong to, given that the three cohorts were unique in their
demographic and biological characteristics such as sex and age range. To do this, we carried out
unsupervised hierarchical clustering on all samples from the three cohorts using the 721,378
probes shared between EPICv1 and EPICv2. We observed a clear demarcation of samples first by
EPIC version despite cohort-specific characteristics (Figure 2A). When clustering was carried out
on subsets of predictive CpGs employed by DNA methylation-based tools, we still noted some
separation of samples by EPIC version, although not to the same extent as the clustering on shared
probes (Supplementary Figure 1).

Spearman correlation between shared EPICv1 and EPICv2 probes was high at array level
and low at the individual probe level
To compare the technical performance of EPICv1 and EPICv2, we first estimated Spearman
correlation of probes between matched samples assessed on the two EPIC versions at the array
level and at the individual probe level for each cohort independently. We identified overall high
array level Spearman correlations between EPICv1 and EPICv2 in the three cohorts (VHAS:
Spearman *rho* = 0.969-0.981, CLHNS: 0.973-0.979, and CALERIE: 0.968-0.981). As expected,
technical replicate samples measured on the same EPIC version showed marginally higher array
level Spearman correlation compared to those measured between versions (within EPICv1: 0.991,
within EPICv2: 0.986, and between versions: 0.976). Similarly, samples from the same individual
measured on EPICv1 and EPICv2 showed higher absolute β value differences for each probe when
compared to differences calculated between technical replicates within each EPIC version.
Specifically, we noted that ~75% of probes have absolute β value differences between EPIC
versions higher than technical replicates within each EPIC version (VHAS: 74.258% (535,682
probes); CLHNS: 76.365% (550,880 probes); and CALERIE: 76.733% (553,821 probes)). On
evaluating probe level Spearman correlations between the two versions, we identified that only a
fourth of the investigated probes exhibited Spearman correlation greater than 0.70, specifically,
VHAS: 29.220% (210,845 probes), CLHNS: 24.965% (180,093 probes), and CALERIE: 19.287%
(139,134 probes)). The remaining ~75% of probes had a low median correlation between EPICv1
and EPICv2, specifically, VHAS: 0.390 (range: -0.648 to 0.991), CLHNS: 0.368 (range: -0.907 to
1.000), and CALERIE: 0.389 (range: -0.744-0.998). One cause for low probe level correlations

may be low inter-sample variability[22]. To assess if observed low individual probe level correlations in the majority of investigated probes can indeed be attributed to low DNA methylation variability across samples, we used a RMSE threshold of < 5% to denote low variability (Figure 2B-D). We observed that of the probes with Spearman correlation < 0.70 between EPICv1 and EPICv2, more than 85% can be attributed to respective RMSE values estimated at less than 5% (VHAS: 89.752% (485,211 probes), CLHNS: 87.994% (476,299 probes), and CALERIE: 91.805% (534,530 probes)). In summary, ~96% of the shared probes between EPICv1 and EPICv2 had either modest-to-high correlation or low correlation explained by low DNA methylation variance, with 87.065% (628,069 probes) commonly identified across three cohorts. When we compared these ~96% probes to platform-bias free and high-confidence mapping probes (n = 542,491) recently identified in cell lines [28], we found that over 95% overlap across our investigated cohorts (VHAS: 96.642% (521,395 probes), CLHNS: 95.011% (512,598 probes), and CALERIE: 97.017% (523,418 probes)).

While the low correlation in the majority of shared EPICv1 and EPICv2 probes has been explained, ~4% probes still remain unexplained by DNA methylation variance (VHAS: 52,322 probes, CLHNS: 64,986 probes; and CALERIE: 47,640 probes; 25,058 common across all cohorts) (Supplementary Table S1). We noted that the vast majority of these probes which have low correlation and high RMSE had % absolute β value differences higher than technical noise measured within each EPIC version (VHAS: 70.217% (36,739 probes), CLHNS: 93.112% (60,510 probes), and CALERIE: 74.733% (35,658 probes)). On comparing these probes to our list of flagged probes with detection p-value > 0.01 or beadcount < 3 in greater than 1% of the samples, we identified an overlap of 4,977, 2,141, and 1,292 probes in the three cohorts, respectively. We next sought to determine whether design type switch or cross hybridization of probes to the genome can explain the observed low correlation and high RMSE. We noted that 14, 17, and 15 probes in VHAS, CLHNS, and CALERIE, respectively, overlapped with the 82 probes that have undergone design type switches in EPICv2[6], and a small fraction of these probes overlapped with previously identified as cross hybridizing to multiple genomic locations or mapped to genetic variant sites (742, 1132, and 633 probes in the Peters reference[5], 624, 980, and 537 probes in the Pidsley reference[29], 683, 1058, and 585 probes in the McCartney reference[30], and 1829, 2370, and 1659 probes in the Price reference[31], in the VHAS, CLHNS, and CALERIE cohorts, respectively).

High coverage of epigenetic clock, biomarker predictor, and cell type deconvolution CpGs was noted on EPICv2

To investigate the applicability of widely used DNA methylation-based tools, which were built on previous arrays/versions, on EPICv2, we noted the percentage of predictive CpGs which are absent in either of the EPIC versions and which are replicated on EPICv2. Across the seven epigenetic clocks investigated, we observed that ~77-96% predictive CpGs were retained on EPICv2, with a reintroduction of several predictive CpGs which were absent in EPICv1 on EPICv2 (Horvath pan-tissue: 14 CpGs; Hannum: 4 CpGs; and epigenetic timer of cancer (epiTOC): 25 CpGs). We noted that in all investigated clocks, except epiTOC, ~1-7% of the predictive clock CpGs have replicate probes on EPICv2. Similarly, across the four investigated biomarker predictors, we noted ~89-94% coverage of predictive CpGs on EPICv2, with 0.86-2.86% also including replicate probes. In the CRP predictor, we additionally observed that 5.84% predictive CpGs were absent in EPICv1,

with 83 of these CpGs reintroduced in EPICv2. Next, on examining the 1200 pre-selected extended
Identifying Optimal DNA methylation Libraries reference (IDOL) probes used in cell type
deconvolution[10], we noted that 99.33% were represented on EPICv2, with 4% also having
replicate probes on EPICv2 (Table 2). On estimating the correlation of β values of predictive CpGs
employed by these DNA methylation-based tools between EPICv1 and EPICv2, weighted by
corresponding coefficients, we identified high Spearman correlation ranging from 0.9615-0.9952
(Supplementary Figure 2). We also noted that average absolute differences in β values of these
predictive CpGs between the EPIC versions range from 0.0157-0.0342, with the Horvath pan-
tissue clock having the lowest average absolute difference (MAD) between the two versions
(Supplementary Table S2-S3).
231

232 Immune cell type proportions inferred by IDOL and auto probe selection methods were
233 significantly different between EPICv1 and EPICv2

Cellular composition is one of the key contributors to whole blood DNA methylation
variation[32,33], and is often included as a covariate in statistical models to account for cell type
heterogeneity. Besides that, cellular composition changes have been associated with phenotypes
of interest and may in fact be integral in the understanding of underlying biological processes[34–
36]. Most studies do not have actual cell counts and rely on predicted values from DNA
methylation-based algorithms. Given that these algorithms use references profiled on previous
arrays/versions, we tested whether there were differences in cell type proportion estimates obtained
from the two EPIC versions using one of the most commonly used cell type deconvolution methods
for whole blood[10]. We compared the proportions of twelve immune cell types, namely, basophils
(Bas), naïve and memory B cells (Bnv, Bmem), naïve and memory CD4+ T cells (CD4nv,
CD4mem), naïve and memory CD8+ T cells (CD8nv, CD8mem), eosinophils (Eos), monocytes
(Mono), neutrophils (Neu), natural killer (NK), and T regulatory cells (Treg), inferred from DNA
methylation measured on EPICv1 and EPICv2 using the estimatedCellCounts2 functions from
FlowSorted.Blood.EPIC R package with two commonly used probe selection methods: IDOL and
"auto". The IDOL method includes 1200 pre-selected probes, eight of which are absent in EPICv2,
and the "auto" method includes 1200 probes independently selected for each EPIC version as
described in Methods. On comparing the probes selected for each EPIC version using the "auto"
method, we identified a large overlap of 1090 probes (90.83%) between the two EPIC versions,
with only 121 EPICv1 (10.983%) and 134 EPICv2 (11.167%) auto-selected probes overlapping
with the 1200 pre-selected IDOL probes (Supplementary Figure 3).
254
Using the IDOL method, we found Pearson correlations between EPICv1 and EPICv2 to be greater
than 0.80 for all predicted cell types, except for Tregs which showed slightly lower correlation
(Pearson r = 0.628-0.715). The proportion of CD8nv was zero for all samples in VHAS and
CLHNS, but averaged at 0.012 (EPICv1) and 0.010 (EPICv2) in CALERIE. On comparing
inferred cell proportions between EPICv1 and EPICv2, we identified significantly higher
estimated proportions of Bas, Mono, and NK cells, and lower proportions of CD8mem and Eos in
EPICv2 compared to EPICv1, with Cohen's d effect size ranging from small to large depending
on the cell type. For all five cell types, MAD between EPIC versions was higher than technical
replicate differences within EPIC versions. CD4nv and CD8nv proportions were not significantly
different between EPIC versions in any of the investigated cohorts. As for the remaining five cell
types, namely Bmem, Bnv, CD4mem, Neu and Tregs, we observed significant proportion

differences between EPIC versions in at least one cohort, however, these findings were not
consistent in all three cohorts (Figure 3, Supplementary Table S4).
268
Using the "auto" method, we found Pearson correlations between EPICv1 and EPICv2 to be
greater than 0.80 for all predicted cell types, with the exception of CD4mem and CD4nv which
showed only a modest Pearson r of 0.788 and 0.778 respectively in VHAS. For all EPICv1
samples, Treg proportions were estimated to be zero in all three cohorts, while CD8nv proportions
were estimated to be zero for all samples in VHAS and CLHNS, and 87.5% in CALERIE. Paired
t-tests showed significantly higher estimated proportions of Bas, Bmem, CD8mem, CD8nv, Mono,
NK and Tregs, and lower proportions of CD4mem and CD4nv in EPICv2 compared to EPICv1.
Meanwhile, we noted no significant differences in Eos between EPICv1 and EPICv2 across all
three cohorts. For all cell types, the MAD between EPIC versions was higher than technical
replicate differences within EPIC versions. Among the remaining cell types, we observed higher
proportions higher proportions of Bnv (d = 0.13-0.22) and Neu (d = -0.06-0.08) in EPICv2 in
CLHNS and CALERIE (Supplementary Figure 4, Supplementary Table S5).
281

Clock EAAs were significantly different between matched samples on EPICv1 and EPICv2
when samples were combined and EAAs were subsequently calculated
Epigenetic clocks are based on the property that DNA methylation levels at specific CpGs highly
correlate with chronological age or age-related outcomes and are often estimated as a linear
combination of weighted sum of the coefficients of predictive CpGs[11–14,16,17]. In order to
evaluate the concordance between EPICv1 and EPICv2 in the context of DNA methylation-based
tools, we compared estimates from widely used first- and second- generation epigenetic clocks,
namely Horvath pan-tissue[11], Hannum[12], Horvath skin and blood clocks[13], PhenoAge[14],
and GrimAge[15], and rate-based clocks, namely pace of aging calculated from the epigenome
(DunedinPACE)[17], DNA methylation estimator of telomere length (DNAmTL)[16], and
epiTOC[18] in the three cohorts VHAS, CLHNS, and CALERIE. We identified high Pearson
correlations between chronological age and epigenetic age obtained from both EPICv1 (VHAS:
Pearson r = 0.701-0.825, CLHNS: 0.819-0.912, and CALERIE: 0.853-0.951) and EPICv2 (VHAS:
Pearson r = 0.583-0.818, CLHNS: 0.807-0.918, and CALERIE: 0.825-0.957) (Supplementary
Table S6). Further, EPICv2 epigenetic ages for the first- and second-generation clocks were highly
correlated with EPICv1 epigenetic ages (VHAS: Pearson r = 0.911-0.980, CLHNS: 0.879-0.990,
and CALERIE: 0.915-0.996), with differences in epigenetic ages between technical replicate
samples being 0.523-4.206 years and 0.546-2.994 years within EPICv1 and EPICv2 respectively
in VHAS, 0.151-3.408 years within EPICv2 in CLHNS (no technical replicates within EPICv1
were available), and 0.13-2.25 years and 0.245-4.142 years within EPICv1 and EPICv2
respectively in CALERIE (Supplementary Table S2). In these three cohorts, we noted significant
differences in epigenetic ages between EPICv2 and EPICv1 in all clocks, except for Horvath pan-
tissue. We identified significantly lower estimated epigenetic ages for Hannum and Horvath skin
and blood clocks and higher estimated epigenetic ages for PhenoAge in EPICv2 when compared
to EPICv1, with mean absolute difference (MAD) greater than technical replicate differences. The
Hannum clock showed the largest MAD and Cohen's d effect size between EPICv1 and EPICv2
in all three cohorts (VHAS: MAD=21.87, d = 3.40, CLHNS: MAD = 19.81, d = 4.14, and
CALERIE: MAD = 21.71, d = 2.86) (Supplementary Table S2, Supplementary Figure 5A).
310

We also estimated epigenetic age acceleration (EAA), a measure of rate of aging which has been associated with health outcomes, by considering samples profiled on each EPIC version separately, as well as by combining matched samples profiled on both versions. When EAA was calculated separately by versions (EPIC versions separate), we noted modest to high correlation of EAA (VHAS: Pearson $r = 0.867-0.947$, CLHNS: 0.577-0.976, and CLAERIE: 0.622-0.961) and no significant differences between EPICv1 and EPICv2 (Supplementary Table S7, Figure 4B, Supplementary Figure 6B, Supplementary Figure 7B). When clock estimates from the EPIC versions were combined prior to EAA calculation (EPIC versions combined), we observed similar modest to high correlation (VHAS: Pearson $r = 0.862-0.947$, CLHNS: 0.574-0.958, and CALERIE: 0.632-0.960), but with significant differences between EPICv1 and EPICv2 for all five clock EAAs, with the exception of Horvath pan-tissue (Supplementary Table S7, Figure 4C, Supplementary Figure 6C, Supplementary Figure 7C). Additionally, we identified significantly lower EAAs for Hannum and Horvath skin and blood, and higher EAAs for PhenoAge in EPICv2 when compared to EPICv1 (Supplementary Table S7). Next, we tested whether the different EPIC versions indeed contribute to observed EAA differences by first combining epigenetic ages in both EPIC versions and then calculating EAAs by including version as a covariate in the linear regression (EPIC versions combined and adjusted). After EPIC version adjustment, we noted modest to high correlation (VHAS: Pearson $r = 0.862-0.947$, CLHNS: 0.574-0.958, and CALERIE: 0.632-0.960) and no significant EAA differences between the two versions in the three cohorts (Supplementary Table S7, Figure 4D, Supplementary Figure 6D, Supplementary Figure 7D). We repeated the same analyses using epigenetic clocks estimated based on principal components (PC clocks) rather than individual predictive CpGs, as they have been shown to overcome unreliability in clock estimates due to technical noise[37]. We similarly noted that there were no significant EAA differences when EAA was calculated for each EPIC version separately. When EAAs were calculated on combined sets of matched samples, we noted significant differences in PCHorvath skin and blood and PCPhenoAge clocks in the three adult cohorts, and these differences were no longer significant with EPIC version adjustment (Supplementary Figures 8-10).

Given the discrepancies known to exist in epigenetic clock estimations using different normalization methods[23,38,39], we sought to examine whether our results, observed in data normalized by functional normalization, remained consistent irrespective of the choice of normalization method. To test this, we combined the common probes on the two EPIC versions and applied Beta-MIxture Quantile (BMIQ) normalization, and subsequently calculated epigenetic ages and EAAs. We still noted significant differences in epigenetic age and EAA between EPICv1 and EPICv2 in Horvath pan-tissue, Horvath skin and blood, PhenoAge, and GrimAge clocks in at least two out of the three cohorts when the two EPIC versions were combined and analyzed (Supplementary Figure 5B and 11C, 12C, and 13C). No significant differences in EAA between the versions were observed when EAAs were calculated separately for EPICv1 and EPICv2 or when the versions were combined and EAAs were adjusted for EPIC version (Supplementary Figure 11-13). Given the contribution of EPIC version in explaining DNA methylation variation, we also examined if applying version correction using batch correction algorithms can circumvent version-specific effects and eliminate the observed differences in DNA methylation-based estimates between EPICv1 and EPICv2. We tested this by combining common probes on EPICv1 and EPICv2 after functional normalization, and sequentially correcting for EPIC version, chip and sample position on chip (row) effects as applicable using the ComBat function implemented in the

sva R package[40]. As expected, when epigenetic ages were calculated using version-corrected input β values, there were no significant differences between the EPIC versions (Supplementary Figure 5C).

On evaluating the DunedinPACE, DNAmTL, and epiTOC rate-based clocks, we observed a high correlation between EPICv1 and EPICv2 when the two were considered separately, as well as when they were combined and adjusted for EPIC versions (VHAS: Pearson *r* = 0.936-0.983, CLHNS: 0.838-0.967, and CALERIE: 0.808-0.941) (Supplementary Table S8). When we examined the EPIC versions separately, we noted that DNAmTL estimates were significantly higher in EPICv2 when compared to EPICv1 in all three cohorts and the differences were higher than technical replicate differences within EPIC versions (VHAS: d = 0.96, CLHNS: d = 1.50, and CALERIE: d = 0.65). DunedinPACE estimates were significantly higher in EPICv2 in VHAS (d = 0.47) and CALERIE (d = 0.65), and epiTOC estimates were significantly lower in EPICv2 in CLHNS (d = 0.43) but higher in CALERIE (d = 0.60) (Supplementary Table S8, Figure 5A, Supplementary Figure 6E, and Supplementary Figure 7E). As expected, we observed no significant differences between EPICv1 and EPICv2 when the rate-based clock estimates were adjusted for EPIC version (Supplementary Table S8, Figure 5B, Supplementary Figure 6F, and Supplementary Figure 7F). We observed similar differences when BMIQ normalized data was used as input for rate-based clocks. When EPICv1 and EPICv2 were considered separately, there were significant differences in DunedinPACE and DNAmTL estimates between versions in all three cohorts, and significant differences in epiTOC estimates in CLHNS and CALERIE, and these differences were no longer significant when EPIC version was accounted for as a covariate (Supplementary Figure 14) or when input β values were adjusted using version correction algorithms upstream (Supplementary Figure 15A). We also repeated the same analyses using matched capillary blood samples available in the VHAS cohort, as both venous and capillary blood sampling are common in epigenetic research[41]. Consistent with our findings in venous blood samples, we observed significant differences between EPIC versions in epigenetic ages as well as EAAs when EPIC versions were combined and EAAs were calculated in four of the investigated first- and second-generation clocks, two PC clocks, and all three rate-based clocks. When EAAs were calculated separately for each version and when EAAs and rate-based clock estimates were adjusted for EPIC versions, there were no significant differences between the EPIC versions (Supplementary Figures 16-19).

Estimated CRP and smoking scores were significantly different between EPICv1 and EPICv2

DNA methylation-based estimates of IL-6, CRP, smoking, and alcohol use are widely used as reliable proxies of actual measurements in epigenetic studies[39], and we tested if there were any significant differences in these estimates between the two EPIC versions. We observed high correlation of IL-6 (VHAS: Pearson *r* = 0.902, CLHNS: 0.912, and CALERIE: 0.85), CRP (VHAS: Pearson *r* = 0.993, CLHNS: 0.992, and CALERIE: 0.944), and smoking scores (VHAS: Pearson *r* = 0.981, CLHNS: 0.99, and CALERIE: 0.72), and modest correlation of alcohol scores (VHAS: Pearson *r* = 0.789, CLHNS: 0.76, and CALERIE: 0.75) between EPICv1 and EPICv2. When estimates were calculated separately by versions (EPIC versions separate), we identified significantly higher CRP scores in EPICv2 compared to EPICv1 in VHAS (d = 0.26) and CLHNS (d = 0.27), with differences higher than technical replicate differences, while there were no

significant differences in IL-6 and alcohol scores in either cohort. We noted significantly higher
IL-6 score and alcohol score in EPICv2 compared to EPICv1 in CALERIE (d = 0.49 and 1.28,
respectively), and significantly lower smoking scores in CLHNS (d = 0.30). When estimates from
the EPIC versions were first combined and then adjusted for version, as expected, we noted no
significant differences between EPICv1 and EPICv2 in any of the DNA methylation-based
predictors tested in the three cohorts (Supplementary Table S3, Supplementary Figure 20). Our
observation of significant EPIC version differences in estimates remained consistent when BMIQ
normalized β was used as input (Supplementary Figure 21), and these differences were no longer
significant when EPIC version was adjusted for as a covariate during estimate calculation or when
version-correction was applied upstream (Supplementary Figure 15B).

Discussion

The newly developed Illumina EPICv2 microarray provides wider coverage of the DNA
methylome compared to its predecessor EPICv1, with some reported differences in overall probe
content. While good correlations between the two EPIC versions have been previously reported
[5–7], a recent study demonstrated that differing probe content between versions influences
epigenetic age estimates, although this was not evaluated in matched samples profiled on EPICv1
and EPICv2[42]. To that end, we measured DNA methylation on EPICv1 and EPICv2 in matched
venous blood samples from adults across three geographically diverse populations, and
comprehensively examined the concordance between versions in a wide range of DNA
methylation-based tools.

We performed unsupervised hierarchical clustering on all samples across the three cohorts using
the shared probes between EPICv1 and EPICv2 and observed that samples clustered based on the
EPIC version they were measured on, despite cohort-specific characteristics. While examining
subsets of probes employed by the different DNA methylation-based tools, we noted a similar
clustering pattern indicating EPIC version as the predominant driver of DNA methylation
variation, followed by biological characteristics such as sex and age, which are known to influence
DNA methylation. Further, when we combined matched EPICv1 and EPICv2 samples in each
cohort and calculated cell type proportions, epigenetic ages, and biomarker scores, we observed
significant systematic differences in the majority of these estimated measures between the EPIC
versions, although some of these differences and their corresponding effect sizes were cohort-
specific. These differences were no longer significant when EPIC version was adjusted for during
the calculation of estimates, and in the case of epigenetic clocks specifically, when EAAs were
calculated in an EPIC version-specific manner.

On comparing the two EPIC versions, we observed overall high correlations at the array level and
low correlation at the individual probe level, in agreement with previous findings[5–7]. Only ~25%
of the shared probes between EPICv1 and EPICv2 showed modest-to-high probe level correlation,
with an additional ~67% showing low correlation explained by low DNA methylation variance.
The majority of these probes have also been reported previously as platform-bias free and high-
confidence mapping in cell lines[28], indicating their reliability between the two EPIC versions
and potentially across different sample types. In the remaining ~4% of probes, the vast majority
had notable differences in β values between EPICv2 and EPICv1, suggesting that these observed
differences may stem from EPIC version beyond technical noise, consistent with previous findings
comparing 450K and EPICv1[22,43,44]. While non-specific probe hybridization to the genome,

underlying genetic variation, or technical differences such as altered EPICv2 probe design may explain the low concordance between the two EPIC versions[5,6,29–31], such probes accounted for only a small proportion. We therefore recommend cautious interpretation of DNA methylation readouts from these probes owing to their differences between the versions, specifically when integrating EPICv1 and EPICv2 data for analyses.

We observed generally high correlation of cell type proportion estimates between EPICv1 and EPICv2, however, significant proportion differences in specific immune cell types were observed. Using pre-selected IDOL and auto methods of probe selection, we found that the proportions of five to nine cell types were significantly different between EPICv1 and EPICv2, with monocytes and NK cells consistently identified in both methods. Within each probe selection method, our observations were generally consistent across all three investigated cohorts indicating systematic differences between the two EPIC versions that are not driven by cohort-specific characteristics. Given that EPICv2 cell type proportions were predicted using the reference library created by profiling purified immune cell types from peripheral blood on EPICv1[10], and similar differences in predicted cell type proportions have been observed previously between 450K and EPICv1[22], the observed differences are likely due to version incompatibility between the reference library used and input samples for EPICv2. These effects may be compounded by the proportional nature of cell type estimates, implying that any changes in the proportion of even one cell type will influence the proportions of other cell types. Together, our findings indicate inherent EPIC version differences influencing cell type proportion estimations, and the need to investigate whether EPIC version-specific cell type reference libraries would enhance the accuracy of cell type deconvolution in EPICv2, and whether the detected EPIC version differences in cell type proportions would affect downstream analyses such as epigenome wide association studies.

Our epigenetic clock analyses showed significant differences in estimates of epigenetic age between EPICv1 and EPICv2 in five investigated clocks despite the high correlation of shared clock CpGs, suggesting systematic differences in epigenetic ages depending on which EPIC version the samples were profiled on. Given that there are varying clock CpGs absent on either EPIC version, it is possible that the subsequent imputation performed by the clock algorithms could influence obtained epigenetic ages and explain some of the observed variation. While using common probes between the two versions as input for clock calculation, however, the variation in epigenetic age estimates was still observed, suggesting that the proportion of missing clock CpGs does not account for the EPIC version differences. This is in line with a previous study that missing values do not compromise clock estimates in first-generation clocks[23]. Although in our analyses the differences in Horvath pan-tissue epigenetic ages between EPICv1 and EPICv2 did not reach statistical significance with functional normalization, these differences were significant when we applied BMIQ normalization, and in both normalization methods, epigenetic ages were higher in EPICv2 compared to EPICv1. Our Horvath pan-tissue clock findings therefore affirm previous reports that choice of normalization method influences clock estimates[23,38,39]. Nevertheless, we still observe significant epigenetic age differences in all other investigated first- and second-generation clocks in both functional and BMIQ normalization, highlighting the contribution of systematic EPIC version differences to epigenetic ages irrespective of the normalization method used. It has been noted that the stochastic changes to single-CpG DNA methylation level as a function of aging explain a non-negligible amount of variation in established epigenetic clocks[45]. However, the individual-level stochastic variation in epigenetic ages is not the focus

492   of our analyses and cannot be measured using the methods employed in this study, and is likely to
493   be minimal relative to cohort and other batch effects.
494
We also tested whether there were EPIC version differences in EAAs, which are more robust to
data preprocessing methods compared to epigenetic ages[23,38,39]. We noted no EAA differences
in the investigated first- and second-generation clocks when EAAs were calculated separately for
the two EPIC versions, regardless of observed systematic shifts in epigenetic ages, lending
additional support to the robustness of EAA as an epigenetic age measure when calculating in a
version-specific manner. On the other hand, with EAA calculation being relative to input samples
and observed epigenetic age differences between EPICv1 and EPICv2, we expectedly observed
EAA differences when samples from both EPIC versions were combined and then EAA was
calculated. These differences were no longer significant when we calculated EAA by regressing
epigenetic age on chronological age and additionally adjusting for EPIC version. In contrast,
calculating epigenetic age differences, by subtracting chronological age from epigenetic age,
would yield significant differences between EPICv1 and EPICv2 due to the observed systematic
EPIC version differences in epigenetic ages, once again highlighting the appropriateness of EAA
in epigenetic age analyses. Our findings of significant differences between EPIC versions hold
true in rate-based clock estimates and biomarker scores as well, and these differences were no
longer significant with EPIC version adjustment, thus emphasizing correcting for EPIC version in
these tools as an effective strategy to obtain comparable estimates.
512
Overall, our results indicate that epigenetic age differences can be ascribed to EPIC version, which
will influence interpretation of findings if not accounted for. While adding EPIC version as a
covariate in statistical models may be an option to handle version-specific discrepancies, this
approach is not as straightforward in cases when EPIC version is confounded with the variable of
interest. Alternatively, as noted in our investigations, version correction by ComBat may also
circumvent observed version differences. However, employing such tools when study designs are
unbalanced can introduce false biological signals, and is not appropriate when biological variables
are confounded with EPIC versions[46–49]. Although the impact of applying batch correction on
DNA methylation-based estimates is not well studied, it is known that data normalization methods
can introduce variation in epigenetic clock estimates[23,39]. Therefore, it is reasonable to
speculate that implementing additional data preprocessing steps such as version correction will
result in biased estimates, specifically when biological variables are not evenly distributed across
the two versions, highlighting an area that warrants further investigation.
526
Our study improves the current understanding of the applicability of DNA methylation-based tools
for EPICv1 and EPICv2, however, there are several limitations. While our analyses may be limited
by the modest sample sizes, our study includes matched samples collected from adults with wide
age range and across diverse geographical regions. We also recognize that the CLHNS cohort
contains female only, however, the consistency of our findings across these demographically
heterogeneous cohorts provides support that the identified differences were not driven by any
cohort-specific characteristics. We also acknowledge that the time between EPICv1 and EPICv2
measurement varied by approximately one month for VHAS and two years in CLHNS. While the
time between array quantification was not consistent, it offers a more realistic representation of
research settings where samples are collected and quantified across multiple batches on different
EPIC versions. Next, given that the training datasets of the investigated epigenetic clocks were

predominantly composed of European populations, the suitability of investigated clocks may not be well established in our investigated Asian and African cohorts[50–52]. However, the differences in epigenetic clock estimates we observed in the VHAS and CLHNS cohorts was similar to our observations in the CALERIE cohort comprising individuals of European and African descent, indicating that our findings were not entirely driven by population-specific effects. Additionally, any population-specific effects which may influence obtained estimates are likely to influence both EPIC versions in a comparable manner. In our analyses, we observed high correlations between clock estimates and chronological age and performed all comparisons between EPICv1 and EPICv2 in a cohort-specific manner, thereby circumventing any potential population-specific effects on calculated clock estimates. Further, while we compared EPICv2 cell type proportions and biomarker scores to EPICv1, we recognize that actual cell counts and biomarker measurements would be more appropriate as the ground truth and would be useful in validating our DNA methylation-based estimates. However, the primary aim of the current work is to investigate the concordance in estimates obtained from EPICv2 and EPICv1. Furthermore, high concordance between actual cell counts or biomarker scores and bioinformatically predicted estimates for EPICv1 has been previously reported[10,19–21], therefore it may be reasonable to rely on these DNA methylation-based estimates. We are also cognizant of the influence of preprocessing methods including normalization and batch correction options on DNA methylation-based estimates[23,38,39]. We have addressed this to a certain degree by employing two different normalization methods, namely functional normalization and BMIQ, as input in our analyses. Our findings in the majority of the tested epigenetic clocks and biomarker predictors remained significant with BMIQ normalization, implying that EPIC version effects still persisted across the two investigated normalization methods. While applying version correction resolved some of these EPIC version differences, it is plausible that currently available batch correction algorithms can result in biased estimates as previously mentioned. Future work to investigate or develop other preprocessing methods could be helpful in harmonizing EPICv1 and EPCIv2 and potentially resolving observed EPIC version differences.

Conclusions

With the rapid generation of DNA methylation data profiled on the newer iterations of the Illumina microarrays, such as EPICv2 and Infinium Methylation Screening Array (MSA), integrating samples across these platforms poses a challenge, owing to discrepancies in probe content among arrays/versions. Our findings highlight significant differences in the new EPICv2 compared to EPICv1, demonstrate the influence of EPIC version on DNA methylation-based tools, and provide recommendations to minimize technical variation which may arise from inconsistencies between versions. We therefore encourage caution when harmonizing and interpreting DNA methylation data when combining multiple arrays/versions to ensure reliability and reproducibility in epigenetics analyses.

Methods

Description of cohorts

To compare the performance of the two most recent generations of MethylationEPIC BeadChip Infinium microarrays in the context of DNA methylation-based clocks, biomarkers and cell type proportion estimates, we measured the DNA methylomes of a subset of venous whole blood

samples on both EPICv1 and EPICv2 selected from three cohorts: (i) Vietnam Health and Aging
Study (VHAS)[25], (ii) Cebu Longitudinal Health and Nutrition Survey (CLHNS)[26], and (iii)
Comprehensive Assessment of Long-term Effects of Reducing Intake of Energy (CALERIE)[27].
The VHAS cohort additionally includes matched capillary blood samples, randomized using the
same array design on both EPIC versions, such that the concordance between EPICv1 and EPICv2
can be assessed in capillary blood as well (n = 24 × 2 blood collection methods × 2 versions). The
time between EPICv1 and EPICv2 array quantification was approximately one month and two
years for VHAS and CLHNS respectively. EPICv1 and EPICv2 array quantification was carried
out at the same time for CALERIE. Given that the demographic characteristics across the cohorts
are different (Table 1), we performed all analyses independently on the cohorts and reported the
findings in a cohort-specific manner, excluding the unsupervised hierarchical clustering analyses
as described below.
592

DNA methylation profiling, sample and probe quality control

Using similar protocols for all three cohorts, DNA was extracted from samples, bisulfite converted
using EZ-96 DNA Methylation kits (Zymo Research, Irvine, CA), hybridized to the EPICv1 and
EPICv2 arrays, and scanned with the Illumina iScan 2000 to obtain IDAT files that capture raw
DNA methylation intensities. IDATs were read using *minfi* R package to obtain β values that
represent DNA methylation intensities for each CpG site ranging from 0 (fully unmethylated) to 1
(fully methylated). Technical replicates derived from the same sample were quantified to monitor
technical variation within each EPIC version, independently for each cohort. Sample quality
control checks were performed as described in previous publications[34,53]. Blood samples
collected in the three cohorts passed all 17 Illumina quality control metrics in the *ewastools* R
package[54,55], and detection *p*-value, beadcount, average methylated and unmethylated intensity
metrics in the *minfi* R package (Supplementary Table S9). We also performed sample identity
checks using unsupervised hierarchical clustering analysis on the 57 SNP probes that are common
to both EPIC versions, and noted that matched samples from the same individual assessed on
EPICv1 and EPICv2 grouped together.
608

In order to identify EPICv1 and EPICv2 probes of poor quality, we performed quality control
checks using the detectionP and beadcount functions in the *minfi* and *wateRmelon* R packages,
respectively. Probes with detection p-value > 0.01 or beadcount < 3 in greater than 1% of the
samples were flagged (VHAS- EPICv1: 59,233, EPICv2: 46,735, common to both versions: 4,826;
CLHNS-EPICv1: 23,793, EPICv2: 17,587, common to both versions: 991; CALERIE-EPICv1:
29,608, EPICv2: 24,872, common to both versions: 1,292), but all probes in EPICv1 and EPICv2
were retained for subsequent analyses.

DNA methylation data preprocessing and replicate probe analyses

To account for color and probe-type bias, we performed functional normalization (funnorm) with
background correction and dye-bias normalization (noob) in the *minfi* R package[56,57]
independently on EPICv1 and EPICv2 samples for each cohort. Cohort-specific technical replicate
sample correlations were used to monitor preprocessing. Between technical replicate samples on
the same EPIC version, improved Spearman correlations of whole array β values and reduced root
mean square error (RMSE) were observed as processing progressed from raw (VHAS: Spearman
rho = 0.9832, RMSE = 0.0306; CLHNS: Spearman *rho* = 0.9890, RMSE = 0.0235, CALERIE:

Spearman *rho* = 0.9859, RMSE = 0.0349) to funnorm normalized data (VHAS: Spearman *rho* =
0.9851, RMSE=0.0288; CLHNS: Spearman *rho* = 0.9917, RMSE = 0.0231, CALERIE: Spearman
rho = 0.9887, RMSE = 0.0223).

To calculate estimates of DNA methylation-based tools, noob corrected data was used as input for
cell type deconvolution, and funnorm normalized data was used as input for epigenetic clocks and
biomarkers. In addition, to test whether normalization methods influence DNA methylation-based
tools, we compared estimates calculated by combining the common probes on EPICv1 and EPICv2
after noob correction in a cohort-wise manner for VHAS, CLHNS, and CALERIE, and
subsequently applying Beta-MIxture Quantile (BMIQ) normalization implemented in the
wateRmelon R package[58]. Due to their type I and type II design switch between EPICv1 and
EPICv2[6], 82 probes were removed prior to BMIQ normalization. To account for any systematic
bias in DNA methylation measurements[40] and subsequently test whether there are differences
in DNA methylation-based estimates between EPIC versions, we applied batch correction for
EPIC version, chip and row on funnorm normalized data using the ComBat function implemented
in the *sva* R package[40]. We applied Spearman correlation to examine concordance of probes
between the two EPIC versions, while in all other analyses we applied Pearson correlation to
evaluate linear relationships between variables.

Given that there are certain probes on EPICv2 having two or more replicates (replicate probes),
we characterized their distribution across the genome, and compared three strategies to collapse
them into a single β value (based on detection *p*-value, mean and median). Our analyses identified
that collapsed β values of EPICv2 replicate probes obtained using all three methods were highly
correlated to corresponding EPICv1 probes, therefore EPICv2 replicate probes with lowest
detection *p*-value were chosen as the representative probe based on previous recommendation[6]
(Supplementary Material), and this approach was used for all the reported subsequent analyses.

Estimation of immune cell type proportions using DNA methylation-based cell type
deconvolution

Cellular composition in heterogeneous tissue such as whole blood is one of the key contributors to
the variation in DNA methylation profiles of bulk tissue[32,33]. In the absence of complete cell
count data for the study samples, we estimated proportions of twelve immune cell types, basophils
(Bas), naïve and memory B cells (Bnv, Bmem), naïve and memory CD4+ T cells (CD4nv,
CD4mem), naïve and memory CD8+ T cells (CD8nv, CD8mem), eosinophils (Eos), monocytes
(Mono), neutrophils (Neu), natural killer (NK), and T regulatory cells (Treg) from matched venous
blood samples measured on the two EPIC versions. We used two methods of probe selection to
estimate these cell type proportions: (i) the extended Identifying Optimal DNA methylation
Libraries reference (IDOL), with probes not represented on EPICv2 removed from the reference
before cell type proportions estimation in EPICv2 samples, and (ii) the "auto" method which
selects the top 100 probes with F-stat *p*-value < 1E-8 for each cell type with the greatest magnitude
of methylation difference, both implemented in the *FlowSorted.Blood.EPIC* R package with noob
corrected values as recommended[10].

Estimation of epigenetic age and epigenetic age acceleration

We compared the performance of eight commonly used epigenetic clocks between EPICv1 and EPICv2 in VHAS, CLHNS, and CALERIE. Epigenetic age of first generation clocks including Horvath pan-tissue[11], Hannum[12], and Horvath skin and blood clocks[13], and second generation clocks including PhenoAge[14], and GrimAge[15] were obtained from the online DNA Methylation Age Calculator (https://dnamage.genetics.ucla.edu/new). For these clocks, we calculated epigenetic age acceleration (EAA), a measure of rate of aging commonly used in epigenetic clock investigations, by employing three approaches independently for each cohort.

1. **EPIC version separate:** we separated epigenetic ages of EPICv1 and EPICv2 samples and then calculated EAA independently on EPICv1 and EPICv2 samples by extracting residuals from the linear regression model: Epigenetic age ~ chronological age.
2. **EPIC versions combined:** we first combined epigenetic ages of EPICv1 and EPICv2 samples and then calculated EAA by extracting residuals from the linear regression model: Epigenetic age ~ chronological age.
3. **EPIC versions combined and version adjusted:** we first combined epigenetic ages of EPICv1 and EPICv2 samples and calculated EAA by extracting residuals from the linear regression model: Epigenetic age ~ chronological age + EPIC version.

To evaluate if there are significant differences in principal components (PC) clock[37] estimates, which are more robust to technical noise, by EPIC version, we first estimated epigenetic ages by the R script provided (https://github.com/MorganLevineLab/PC-Clocks) and then calculated EAAs using the approaches mentioned above.

As for the rate-based clocks, DunedinPACE was calculated using the DunedinPACE R package[17], epiTOC was obtained using the *getEpiTOC* function in the *cgageR* R package[18], and DNA methylation-based estimator of telomere length (DNAmTL)[16] were obtained from the online DNA Methylation Age Calculator (https://dnamage.genetics.ucla.edu/new). Unlike EAA calculation, there were no secondary measures calculated from these rate-based estimates in our analyses, estimates remained the same when calculated by the EPIC versions separate or combined approach, therefore, we employed only two approaches for each cohort:

1. **EPIC version separate:** rate-based clock estimates were calculated for EPICv1 and EPICv2 samples without EPIC version adjustment.
2. **EPIC versions combined and version adjusted:** rate-based clock estimates were first calculated for EPICv1 and EPICv2 samples, and were subsequently adjusted for EPIC versions by regressing out EPIC versions using linear regression model: Rate estimate ~ EPIC version.

Using paired *t*-tests and applying a Bonferroni multiple test correction, we evaluated differences in epigenetic clock estimations and EAAs (as well as biomarker predictor scores and cell type proportions described below) between matched samples assessed on EPICv1 and EPICv2 in a cohort-specific manner. Statistical significance was defined as Bonferroni adjusted *p*-value < 0.05. Effect sizes were measured by Cohen's d, and classified as "small" (d = 0.2-0.49), "medium" (d = 0.5-0.79), and "large" (d ≥ 0.8) based on recommended benchmarks [59].

Estimation of DNA methylation-based inflammation, smoking, and alcohol scores

Among other DNA methylation-based tools are inflammation, smoking and alcohol score predictors, which provide biomarker measures that correlate with levels of inflammatory markers (IL-6, CRP), smoking and alcohol use, respectively. DNA methylation-derived scores of IL-6, CRP, smoking and alcohol use were calculated as a weighted sum of coefficients derived from published lists of predictive CpGs[19–21]. Being that these biomarkers were trained on previous Illumina arrays, we sought out to determine the correlation of derived DNA methylation-based scores between EPICv1 and EPICv2, and compare these scores without version adjustment (EPIC version separate) and with version adjustment (EPIC versions combined and adjusted) using approaches similar to the rate-based epigenetic clocks.

References

1. Battram T, Yousefi P, Crawford G, Prince C, Sheikhali Babaei M, Sharp G, et al. The EWAS Catalog: a database of epigenome-wide association studies. Wellcome Open Res. 2022;7:41.

2. Kuzawa CW, Ryan CP, Adair LS, Lee NR, Carba DB, MacIsaac JL, et al. Birth weight and maternal energy status during pregnancy as predictors of epigenetic age acceleration in young adults from metropolitan Cebu, Philippines. Epigenetics. 2022;17:1535–45.

3. Li M, Zou D, Li Z, Gao R, Sang J, Zhang Y, et al. EWAS Atlas: a curated knowledgebase of epigenome-wide association studies. Nucleic Acids Res. 2019;47:D983–8.

4. Infinium MethylationEPIC v2.0 BeadChip. Infinium MethylationEPIC v2.0 BeadChip. 2023;

5. Peters TJ, Meyer B, Ryan L, Achinger-Kawecka J, Song J, Campbell EM, et al. Characterisation and reproducibility of the HumanMethylationEPIC v2.0 BeadChip for DNA methylation profiling. BMC Genomics. 2024;25:251.

6. Kaur D, Lee SM, Goldberg D, Spix NJ, Hinoue T, Li H-T, et al. Comprehensive evaluation of the Infinium human MethylationEPIC v2 BeadChip. Epigenetics Commun. 2023;3:6.

7. Noguera-Castells A, García-Prieto CA, Álvarez-Errico D, Esteller M. Validation of the new EPIC DNA methylation microarray (900K EPIC v2) for high-throughput profiling of the human DNA methylome. Epigenetics. 2023;18:2185742.

8. Houseman EA, Molitor J, Marsit CJ. Reference-free cell mixture adjustments in analysis of DNA methylation data. Bioinforma Oxf Engl. 2014;

9. Salas LA, Koestler DC, Butler RA, Hansen HM, Wiencke JK, Kelsey KT, et al. An optimized library for reference-based deconvolution of whole-blood biospecimens assayed using the Illumina HumanMethylationEPIC BeadArray. Genome Biol. 2018;19:64.

10. Salas LA, Zhang Z, Koestler DC, Butler RA, Hansen HM, Molinaro AM, et al. Enhanced cell deconvolution of peripheral blood using DNA methylation for high-resolution immune profiling. Nat Commun. 2022;13:761.

11. Horvath S. DNA methylation age of human tissues and cell types. Genome Biol.
2013;14:R115.

12. Hannum G, Guinney J, Zhao L, Zhang L, Hughes G, Sadda S, et al. Genome-wide Methylation
Profiles Reveal Quantitative Views of Human Aging Rates. Mol Cell. 2013;49:359–67.

13. Horvath S, Oshima J, Martin GM, Lu AT, Quach A, Cohen H, et al. Epigenetic clock for skin
and blood cells applied to Hutchinson Gilford Progeria Syndrome and *ex vivo* studies. Aging.
2018;10:1758–75.

14. Levine ME, Lu AT, Quach A, Chen BH, Assimes TL, Bandinelli S, et al. An epigenetic
biomarker of aging for lifespan and healthspan. Aging. 2018;10:573–91.

15. Lu AT, Quach A, Wilson JG, Reiner AP, Aviv A, Raj K, et al. DNA methylation GrimAge
strongly predicts lifespan and healthspan. Aging. 2019;11:303–27.

16. Lu AT, Seeboth A, Tsai P-C, Sun D, Quach A, Reiner AP, et al. DNA methylation-based
estimator of telomere length. Aging. 2019;11:5895–923.

17. Belsky DW, Caspi A, Corcoran DL, Sugden K, Poulton R, Arseneault L, et al. DunedinPACE,
a DNA methylation biomarker of the pace of aging. Deelen J, Tyler JK, Suderman M, Deelen J,
editors. eLife. 2022;11:e73420.

18. Yang Z, Wong A, Kuh D, Paul DS, Rakyan VK, Leslie RD, et al. Correlation of an epigenetic
mitotic clock with cancer risk. Genome Biol. 2016;17:205.

19. Stevenson AJ, Gadd DA, Hillary RF, McCartney DL, Campbell A, Walker RM, et al. Creating
and Validating a DNA Methylation-Based Proxy for Interleukin-6. J Gerontol A Biol Sci Med Sci.
2021;76:2284–92.

20. Wielscher M, Mandaviya PR, Kuehnel B, Joehanes R, Mustafa R, Robinson O, et al. DNA
methylation signature of chronic low-grade inflammation and its role in cardio-respiratory
diseases. Nat Commun. 2022;13:2408.

21. McCartney DL, Hillary RF, Stevenson AJ, Ritchie SJ, Walker RM, Zhang Q, et al. Epigenetic
prediction of complex traits and death. Genome Biol. 2018;19:136.

22. Solomon O, MacIsaac J, Quach H, Tindula G, Kobor MS, Huen K, et al. Comparison of DNA
methylation measured by Illumina 450K and EPIC BeadChips in blood of newborns and 14-year-
old children. Epigenetics. 2018;13:655–64.

23. McEwen LM, Jones MJ, Lin DTS, Edgar RD, Husquin LT, MacIsaac JL, et al. Systematic
evaluation of DNA methylation age estimation with common preprocessing methods and the
Infinium MethylationEPIC BeadChip array. Clin Epigenetics. 2018;10:123.

24. Dhingra R, Kwee LC, Diaz-Sanchez D, Devlin RB, Cascio W, Hauser ER, et al. Evaluating
DNA methylation age on the Illumina MethylationEPIC Bead Chip. PLOS ONE.
2019;14:e0207834.

25. Korinek K, Teerawichitchainan B, Zimmer Z, Brindle E, Nguyen TKC, Nguyen HM, et al.
Design and measurement in a study of war exposure, health, and aging: protocol for the Vietnam
health and aging study. BMC Public Health. 2019;19:1351.

26. Adair LS, Popkin BM, Akin JS, Guilkey DK, Gultiano S, Borja J, et al. Cohort Profile: The
Cebu Longitudinal Health and Nutrition Survey. Int J Epidemiol. 2011;40:619–25.

27. Rochon J, Bales CW, Ravussin E, Redman LM, Holloszy JO, Racette SB, et al. Design and
conduct of the CALERIE study: comprehensive assessment of the long-term effects of reducing
intake of energy. J Gerontol A Biol Sci Med Sci. 2011;66:97–108.

28. Chen BH, Zhou W. mLiftOver: Harmonizing Data Across Infinium DNA Methylation
Platforms [Internet]. bioRxiv; 2024 [cited 2024 Apr 8]. p. 2024.03.18.585415. Available from:
https://www.biorxiv.org/content/10.1101/2024.03.18.585415v1

29. Pidsley R, Zotenko E, Peters TJ, Lawrence MG, Risbridger GP, Molloy P, et al. Critical
evaluation of the Illumina MethylationEPIC BeadChip microarray for whole-genome DNA
methylation profiling. Genome Biol. 2016;17:208.

30. McCartney DL, Walker RM, Morris SW, McIntosh AM, Porteous DJ, Evans KL. Identification
of polymorphic and off-target probe binding sites on the Illumina Infinium MethylationEPIC
BeadChip. Genomics Data. 2016;9:22–4.

31. Price EM, Cotton AM, Lam LL, Farré P, Emberly E, Brown CJ, et al. Additional annotation
enhances potential for biologically-relevant analysis of the Illumina Infinium
HumanMethylation450 BeadChip array. Epigenetics Chromatin. 2013;6:4.

32. Jones MJ, Moore SR, Kobor MS. Principles and Challenges of Applying Epigenetic
Epidemiology to Psychology. 2017;30.

33. Zheng SC, Beck S, Jaffe AE, Koestler DC, Hansen KD, Houseman AE, et al. Correcting for
cell-type heterogeneity in epigenome-wide association studies: revisiting previous analyses. Nat
Methods. 2017;14:216–7.

34. Merrill SM, Gladish N, Fu MP, Moore SR, Konwar C, Giesbrecht GF, et al. Associations of
peripheral blood DNA methylation and estimated monocyte proportion differences during infancy
with toddler attachment style. Attach Hum Dev. 2023;25:132–61.

35. Ryan CP, Jones MJ, Edgar RD, Lee NR, Kobor MS, McDade TW, et al. Immune cell type and
DNA methylation vary with reproductive status in women: possible pathways for costs of
reproduction. Evol Med Public Health. 2022;10:47–58.

36. Kong Y, Rastogi D, Seoighe C, Greally JM, Suzuki M. Insights from deconvolution of cell
subtype proportions enhance the interpretation of functional genomic data. PLOS ONE.
2019;14:e0215987.

37. Higgins-Chen AT, Thrush KL, Wang Y, Minteer CJ, Kuo P-L, Wang M, et al. A computational
solution for bolstering reliability of epigenetic clocks: Implications for clinical trials and
longitudinal tracking. Nat Aging. 2022;2:644–61.

38. Ori APS, Lu AT, Horvath S, Ophoff RA. Significant variation in the performance of DNA
methylation predictors across data preprocessing and normalization strategies. Genome Biol.
2022;23:225.

39. Engelbrecht H-R, Merrill SM, Gladish N, MacIsaac JL, Lin DTS, Ecker S, et al. Sex
differences in epigenetic age in Mediterranean high longevity regions. Front Aging.
2022;3:1007098.

40. Leek JT, Johnson WE, Parker HS, Jaffe AE, Storey JD. The sva package for removing batch
effects and other unwanted variation in high-throughput experiments. Bioinforma Oxf Engl.
2012;28:882–3.

41. Hubens WHG, Maié T, Schnitker M, Bocova L, Puri D, Wessiepe M, et al. Targeted DNA
Methylation Analysis Facilitates Leukocyte Counts in Dried Blood Samples. Clin Chem.
2023;69:1283–94.

42. Garma LD, Quintela-Fandino M. Epigenetic age prediction drifts resulting from next-
generation methylation arrays [Internet]. 2024 [cited 2024 Jul 1]. Available from:
http://biorxiv.org/lookup/doi/10.1101/2024.06.07.597709

43. Sugden K, Hannon EJ, Arseneault L, Belsky DW, Corcoran DL, Fisher HL, et al. Patterns of
Reliability: Assessing the Reproducibility and Integrity of DNA Methylation Measurement.
Patterns. 2020;1:100014.

44. Olstad EW, Nordeng HME, Sandve GK, Lyle R, Gervin K. Low reliability of DNA
methylation across Illumina Infinium platforms in cord blood: implications for replication studies
and meta-analyses of prenatal exposures. Clin Epigenetics. 2022;14:80.

45. Meyer DH, Schumacher B. Aging clocks based on accumulating stochastic variation. Nat
Aging. 2024;1–15.

46. Price EM, Robinson WP. Adjusting for Batch Effects in DNA Methylation Microarray Data,
a Lesson Learned. Front Genet [Internet]. 2018 [cited 2021 Jun 22];9. Available from:
https://www.frontiersin.org/articles/10.3389/fgene.2018.00083/full

47. Buhule OD, Minster RL, Hawley NL, Medvedovic M, Sun G, Viali S, et al. Stratified
randomization controls better for batch effects in 450K methylation analysis: a cautionary tale.
Front Genet. 2014;5:354.

48. Nygaard V, Rødland EA, Hovig E. Methods that remove batch effects while retaining group
differences may lead to exaggerated confidence in downstream analyses. Biostat Oxf Engl.
2016;17:29–39.

49. Zindler T, Frieling H, Neyazi A, Bleich S, Friedel E. Simulating ComBat: how batch correction
can lead to the systematic introduction of false positive results in DNA methylation microarray
studies. BMC Bioinformatics. 2020;21:271.

50. Lin W-Y. Epigenetic clocks derived from western samples differentially reflect Taiwanese
health outcomes. Front Genet. 2023;14:1089819.

51. Horvath S, Gurven M, Levine ME, Trumble BC, Kaplan H, Allayee H, et al. An epigenetic
clock analysis of race/ethnicity, sex, and coronary heart disease. Genome Biol. 2016;17:171.

52. Cronjé HT, Nienaber-Rousseau C, Min JL, Green FR, Elliott HR, Pieters M. Comparison of
DNA methylation clocks in Black South African men. Epigenomics. 2021;13:437–49.

53. Konwar C, Asiimwe R, Inkster AM, Merrill SM, Negri GL, Aristizabal MJ, et al. Risk-focused
differences in molecular processes implicated in SARS-CoV-2 infection: corollaries in DNA
methylation and gene expression. Epigenetics Chromatin. 2021;14:54.

54. Murat K, Grüning B, Poterlowicz PW, Westgate G, Tobin DJ, Poterlowicz K. Ewastools:
Infinium Human Methylation BeadChip pipeline for population epigenetics integrated into
Galaxy. GigaScience. 2020;9:giaa049.

55. Illumina Inc. BeadArray Controls Reporter Software Guide. 2015;

56. Fortin J-P, Labbe A, Lemire M, Zanke BW, Hudson TJ, Fertig EJ, et al. Functional
normalization of 450k methylation array data improves replication in large cancer studies. Genome
Biol. 2014;15:503.

57. Triche TJ, Weisenberger DJ, Van Den Berg D, Laird PW, Siegmund KD. Low-level processing
of Illumina Infinium DNA Methylation BeadArrays. Nucleic Acids Res. 2013;41:e90.

58. Teschendorff AE, Marabita F, Lechner M, Bartlett T, Tegner J, Gomez-Cabrero D, et al. A
beta-mixture quantile normalization method for correcting probe design bias in Illumina Infinium
450 k DNA methylation data. Bioinformatics. 2013;29:189–96.

59. Cohen J. Statistical Power Analysis for the Behavioral Sciences. 2nd ed. New York: Routledge;
1988.

Acknowledgements

We thank Dr. Sarah Merrill, Maggie Fu, and Dr. Kimberly Schmidt for their helpful feedback
during manuscript preparation. We would also like to thank all cohort participants for their
involvement and providing biological samples for research. This research utilized the
FlowSorted.BloodExtended.EPIC software packages developed at Dartmouth College, which are

governed by the licensing terms provided by Dartmouth Technology Transfer. D.W.B. and M.S.K.
are fellows of the CIFAR CBD Network.

Funding

882 This work received supported from the Canadian Institutes of Health Research (CIHR) / Project
883 Grant (Sponsor Reference Number: PJT-175309), National Institute on Aging of the National
884 Institutes of Health under Award Number R01AG052537, National Institutes of Health / Research
885 Project Grant (R01) (Sponsor Reference Number: 5R01AG061006-05), and US National Institute
886 on Aging Grants R01AG061378.
887

Ethics approval and consent to participate

889 This study was approved by the University of British Columbia Research Ethics Boards (H18-
890 03136) and the University of Utah Institutional Review Board (IRB 00098861).

Competing interests

892 D.W.B. is listed as an inventor of the Duke University and University of Otago invention
893 DunedinPACE, which is licensed to TruDiagnostic. The other authors declare that they have no
894 competing interests.

Authors' contributions

Conceptualization: B.C.Z, M.S.J., C.K.; biological data collection/generation: T.K.T., K.K., Z.Z.,
K.M.H., D.W.B., T.W.M., C.W.K., M.S.K.; DNA methylation data generation: J.L.M., K.D.,
M.S.K.; DNA methylation data curation and processing: B.C.Z., M.S.J., H-R.E.; Phenotypic data
curation: T.K.T., K.K., Z.Z., K.M.H., D.W.B., T.W.M, C.W.K.; Data analysis: B.C.Z., M.S.J.;
Interpretation of results and manuscript writing: M.S.J., B.C.Z., C.K.; Figures/visualization:
B.C.Z., M.S.J.; Manuscript review: all authors; Cohort Owners/Custodians: T.K.T., K.K., Z.Z.,
K.M.H., D.W.B., T.W.M., C.W.K.
903
904
905

907 **Figure 1.** Overview of study design and analyses.

909
910

Figure 2. (A) Hierarchical clustering of 721,378 probes common between EPICv1 and EPICv2 of matched VHAS, CLHNS, and CALERIE samples; blue to red color range denotes Spearman correlation from 0.94-1.00. **(B-D)** Spearman correlation and root mean square error (RMSE) of 721,378 probes common between EPICv1 and EPICv2 investigated on **(B)** VHAS, **(C)** CLHNS, and **(D)** CALERIE. The X-axis represents the Spearman correlation and Y-axis represents the RMSE of probes common to both arrays. Dashed horizontal line indicates the RMSE threshold set at 0.05, and the vertical dashed line indicates the correlation threshold set at 0.70. The colors indicate the density of points, such that pink is low density and yellow is high density. Probes unique to either array are not shown.

Figure 3. Differences in DNA methylation-based immune cell type proportions estimated using the IDOL reference on matched samples assessed on EPICv1 and EPICv2 in VHAS, CLHNS, and CALERIE. Statistical significance was defined as Bonferroni adjusted *p*-value <0.05. ** denotes Bonferroni *p* <0.05, *** denotes Bonferroni *p* <0.001, "ns" denotes "not significant", and "d" denotes effect size measured using Cohen's d. A positive Cohen's d indicates higher estimates in EPICv2 compared to EPICv1.

929
930

Figure 4. Epigenetic ages on matched samples assessed on EPICv1 and EPICv2 in VHAS. **(A)** Scatter plot of Horvath pan-tissue, Hannum, Horvath skin and blood, PhenoAge, and GrimAge clock ages (Y axis) and chronological age (X axis) with dotted line indicating x=y, coloured by EPIC version. **(B-D)** Boxplots comparing EPICv1 and EPICv2 EAAs calculated by considering (B) EPIC versions separately, (C) combined, and (D) combined and EPIC version adjusted. Statistical significance was defined as Bonferroni adjusted *p*-value <0.05. ** denotes Bonferroni *p* <0.05, *** denotes Bonferroni *p* <0.001, "ns" denotes "not significant", and "d" denotes effect size measured using Cohen's d. A positive Cohen's d indicates higher estimates in EPICv2 compared to EPICv1.

941
942

Figure 5. Rate-based clock estimates on matched samples assessed on EPICv1 and EPICv2 in VHAS. Boxplots comparing DunedinPACE, DNAmTL and epiTOC estimates calculated by considering EPIC versions separately **(A)** and combined and EPIC version adjusted **(B)**, between EPICv1 and EPICv2. Statistical significance was defined as Bonferroni adjusted p-value <0.05. ** denotes Bonferroni p <0.05, *** denotes Bonferroni p <0.001, "ns" denotes "not significant", and "d" denotes effect size measured using Cohen's d. A positive Cohen's d indicates estimates in EPICv2 compared to EPICv1.

Table 1. Demographics of the three investigated cohorts.

	VHAS	CLHNS	CALERIE
Samples (n)	48 (24 matched venous and capillary samples)	16	24
Country of origin	Vietnam	Philippines	United States of America
Biological sex (% Female)	36.4%	100%	54.2%
Age in years (mean (SD, range))	74.6 (6.7, 66-86)	48.4 (6.1, 40-63)	36.7(8.1, 21.3-51)

Table 2. Summary of predictive CpGs of DNA methylation-based clocks, biomarker predictors, and cell type deconvolution in EPICv1 and EPICv2.

Predictors	CpGs	Probes absent in EPICv1 (%)	Probes absent in EPICv2 (%)	Probes absent in EPICv1 and reintroduced in EPICv2	EPICv2 replicate probes	Training arrays; training age range (year)	Training sample size	Probes absent in Infinium Methylation Screening Array (MSA)	Probes absent in 450K
Horvath pan-tissue*	353	17 (4.82%)	13 (3.68%)	14	4(1.13%)	27K and 450K Age: 0-100	7844	6(1.7%)	0
Hannum*	71	6 (8.45%)	7 (9.86%)	4	5(7.04%)	450K Age: 19-101	656	3(4.23%)	0
Horvath SkinBlood	391	0	17 (4.35%)	0	9(2.3%)	450K and EPICv1 Age: 0-94	896	3(0.77%)	0
PhenoAge	513	0	18 (3.51%)	0	7(1.36%)	27K, 450K, EPICv1 Age: 18-100	9926	3(0.58%)	0
GrimAge**	1030	NA	NA	0	NA	450K and EPICv1 Age: NA (mean 66)	1731	NA	NA
PC Clocks***	78464	0	5801 (7.39%)	0	650 (0.83%)	450K and EPICv1 Age: 0-101	NA	51108 (65.14%)	0

DunedinPACE	173	0	29 (16.76%)	0	2(1.16%)	EPICv1 Age 38 and 45	1037	2(1.16%)	0
DNAmTL	140	0	31 (22.14%)	0	3(2.14%)	450K and EPICv1 Age: 22-93	2256	2(1.43%)	0
epiTOC*	385	31 (8.05%)	26 (6.75%)	25	0	450K Age: 19-101	656	13(3.38%)	0
IL-6 score	35	0	3 (8.57%)	0	1(2.86%)	450K and EPICv1 Age: 67-78	875	12(34.29%)	0
CRP score*	1765	103 (5.84%)	96 (5.44%)	83	41 (2.32%)	27K, 450K and EPICv1 Age: 16-75	22774	357 (20.23%)	0
Smoking score	233	0	23 (9.87%)	0	2(0.86%)	EPICv1 Age: 18-99	5087	27(11.59%)	0
Alcohol score	450	0	49 (10.89%)	0	6(1.33%)	EPICv1 Age: 18-99	5087	47(10.44%)	0
IDOL	1200	0	8 (0.67%)	0	48(4%)	EPICv1 Age: 19-58	NA	20(1.67%)	741 (61.75%)

* CpGs commonly absent in both EPICv1 and EPICv2: Horvath pan-tissue: 3 CpGs; Hannum: 2 CpGs; EpiTOC: 6 CpGs; CRP: 20 CpGs. ** Clock CpGs for GrimAge are not publicly available. *** PC Clocks CpGs include the total CpGs required to calculate PC versions of Horvath pan-tissue, Hannum, Horvath SkinBlood, PhenoAge, and GrimAge.