Uncertainty-Aware Traction Force Microscopy

Adithan Kandasamy[1,2], Yi-Ting Yeh[1,2], Ricardo Serrano[3], Mark Mercola[3], Juan C. del Álamo[1,2,4*].

1 Department of Mechanical engineering, University of Washington, Seattle.

2 Center for Cardiovascular Biology, University of Washington, Seattle.

3 Cardiovascular Institute and Department of Medicine, Stanford University, Stanford.

4 Division of Cardiology, University of Washington, Seattle.

* juancar@uw.edu

Abstract

Traction Force Microscopy (TFM) is a versatile tool to quantify cell-exerted forces by imaging and tracking fiduciary markers embedded in elastic substrates. However, the computations involved in TFM are ill-conditioned, and data smoothing or regularization is required to avoid overfitting the noise in the tracked substrate displacements. Most TFM calculations depend critically on the heuristic selection of regularization (hyper)parameters affecting the balance between overfitting and smoothing. However, TFM methods rarely estimate or account for measurement errors in substrate deformation to adjust the regularization level accordingly. Moreover, there is a lack of tools to quantify how these errors propagate to the recovered traction stresses. These limitations make it difficult to interpret TFM readouts and hinder comparing different experiments. This manuscript presents an uncertainty-aware TFM technique that estimates the variability in the magnitude and direction of the traction stress vector recovered at each point in space and time of each experiment. In this technique, substrate deformation and its uncertainty are quantified using a non-parametric bootstrap PIV method by resampling the microscopy image pixels (PIV-UQ). This information is passed to a hierarchical Bayesian framework that automatically selects its hyperparameters to perform spatially adaptive regularization conditioned on image quality, and propagates the uncertainty to the traction stress readouts (TFM-UQ). We validate the performance of PIV-UQ and TFM-UQ using synthetic datasets with prescribed image quality variations, and demonstrate the application of PIV-UQ and TFM-UQ to experimental datasets. These studies show that TFM-UQ locally balances the level of smoothing, outperforming traditional regularization methods. They also illustrate how uncertainty-aware TFM tools can be used to objectively choose key image analysis parameters like PIV-UQ interrogation window size. We anticipate these tools will allow for decoupling biological heterogeneity from measurement variability, and facilitate automating the analysis of large datasets by parameter-free, input data-based regularization.

1   Introduction

Mechanical forces in the range of pico and nano Newtons have been implicated in fundamental cellular functions such as cell fate determination [1], cell positioning during embryonic development [2], individual and collective cell migration [3], and cardiac contractility [4]. Over the last half-century, biomechanical force measurement methods have enabled the discovery of cytoskeletal force generation and sensing mechanisms [5, 6]. Fluorescence microscopy-based methods such as traction force microscopy (TFM) [7–11], Förster resonance energy transfer (FRET) [12], and DNA-hairpin-based sensors [13, 14] provide spatial maps of traction stress and force fields with cellular and sub-cellular spatial resolution, and high temporal resolution.

Among these techniques, TFM has become a widely used tool to measure cell-exerted traction forces by tracking fiduciary markers (e.g., fluorescent beads) embedded in linearly elastic substrates of known mechanical properties. Typical TFM experiments do not require highly specialized laboratory equipment, and their preparation is relatively simple. In particular, two-dimensional (in-plane) and three-dimensional (out-of-plane) TFM experiments on flat surfaces can be performed with wide-field fluorescence microscopes [15], and image analysis software is widely available and runs in near real-time in off-the-shelf computers [16–19]. There is a vast body of scientific literature using TFM in biomechanics and mechanobiology studies. In particular, TFM has been instrumental in investigating single cell locomotion in planar surfaces [7, 20], focal adhesion dynamics in adherent cells [21, 22], intercellular forces in collectively migrating epithelial monolayers [23–26], human cardiomyocyte contractility [27, 28], platelet spreading [29, 30], leukocyte transmigration and interstitial migration [31–33], and host-pathogen interaction [34].

TFM is a semi-computational method that solves the inverse elastostatic problem of finding the traction stress field giving rise to the observed field of substrate displacements. This inverse problem is ill-posed because of the long-range decay of displacements, making the calculated traction stresses highly sensitive to experimental measurement noise of high spatial frequency [35–37]. The ill-posedness of TFM manifests itself in the numerical discretization of the elastostatic problem, leading to a linear system of equations with a high condition number [7, 21]. Even when this numerical issue is avoided by sparsely block-diagonalizing the elastostatic system and inverting it exactly in Fourier space [8, 9], the slow decay of substrate deformation still causes noise amplification. Therefore, TFM algorithms have traditionally included regularization schemes in the form of implicit priors [7, 19, 35, 36, 38], explicit noise filters [21], etc. Previous works have been very useful to understand the differences among TFM methods [19, 21]. However, there is no unified framework to comprehensively study the sensitivity and variability of TFM methods to the choice of elastostatic solution, numerical discretization, and the form of regularization scheme.

Most regularization methods in TFM involve optimizing a composite norm weighing model accuracy and complexity. This optimization determines the degree of spatial smoothing applied in the inversion process, affecting the dynamical range and the effective resolution of the recovered traction stress maps [35, 36]. Several approaches have been introduced to balance these effects, e.g.,

χ^2-Criterion [7, 36], Morozov discrepancy principle [36, 39], the self-consistency principle [36], the L-curve criterion [36], etc. Recent efforts have adopted Bayesian frameworks [19, 40, 41] following the seminal work of [7], proposing statistical metrics such as generalized cross-validation [42, 43] to tailor the regularization level to each experimental deformation map. However, the optimal value of regularization parameter estimated by the different criteria for the same experimental data has not always been consistent, leaving room for ambiguity in the balance between overfitting and smoothing [36].

Optimization metrics for TFM regularization are usually calculated over an entire microscopy image or region of interest, assuming that noise is spatially uniform (i.e., homoskedastic). Therefore, the ensuing regularization is sensitive to localized anomalies (e.g., large bead aggregates and empty areas without beads). This sensitivity generates the need to introduce quality control steps in TFM analysis pipelines, which are usually binary (i.e., accept vs. reject) and seldom reported, thus limiting the yield and reproducibility of TFM experiments. More generally, fluorescence microscopy images for TFM can contain spatial gradients in bead focus or brightness, leading to spatially dependent noise in the measured substrate deformations. However, spatial gradients in TFM microscopy images are rarely accounted for in the regularization of the inverse elastostatic problem. A recent work reformulating the optical flow method using the elastostatic equation as an optimization constraint is an important step towards effectively linking spatially-dependent image noise and the recovered traction stresses [39].

The vast majority of regularization schemes in TFM do not provide *a posteriori*, point-by-point estimates of experimental uncertainty. A usual approach to make up for this lack is to perform repeat experiments and report the statistics (e.g., mean and standard deviation) of select readouts. However, this approach mixes the measurement uncertainty that is uniquely due to the experimental setup, imaging method, and mathematical model with the biological variability associated with the fact that two different cells do not exert the same traction stresses. The paucity of uncertainty quantification (UQ) tools in TFM is at odds with the considerable UQ efforts in particle image velocimetry (PIV) and other common methods to measure fluorescent bead displacements [44–48]. Consequently, the choice of crucial parameters affecting the resolution and noise in the measured displacements for TFM, e.g., the PIV interrogation window size, must be done without quantitative information about this choice's effect on the traction stresses. This factor adds to the ambiguity in regularization criteria, sensitivity to image artifacts, and impossibility to separate biological and experimental uncertainty, perpetuating the view that TFM is technically involved and challenging to interpret [49, 50], hindering its more widespread adoption.

This manuscript presents a traction force microscopy method with uncertainty quantification (TFM-UQ). The new TFM-UQ method adaptively regularizes the traction stresses in different regions of a measurement image to be self-consistent with the input deformation noise. Furthermore, it propagates the input noise forward to provide spatial maps of uncertainty in the recovered traction stresses. The input noise in TFM-UQ is determined using a novel, non-parametric PIV uncertainty quantification (PIV-UQ) method suitable for fiduciary markers used in cell mechanics

Fig 1. TFM-UQ method improves traction stress quantification and provides error bars. (A) Traction Force Microscopy experiment. Reference image ($\boldsymbol{I}_{ref}$) of the substrate markers is obtained at a stress-free state. Session image(s) ($\boldsymbol{I}_t$) of the substrate marker are obtained at the time point of interest. (B) Existing TFM workflow provides point estimate of forces, but does not consider the variability in the traction stress due to the input image and TFM pipeline. (C) The proposed TFM-UQ method takes into account the local image quality, ambiguity in regularization parameter selection, and the numerical implementation to provide locally adaptive smoothing and error bars.

experiments. To make the PIV-UQ method tractable, a computationally efficient PIV software is developed with GPU acceleration. To validate the TFM-UQ pipleline, we simulated TFM experiments using synthetically generated fluorescent beads directly simulating the sources of image noise. In experimental microscopy images with different cell types (including images with visible imaging artifacts), the TFM-UQ method uncovered spatially heterogeneous PIV measurement noise, uncertainty associated with image processing and finally, provided confidence intervals to the traction stress output. Taken together, TFM-UQ improves the dynamic range of traction stresses and provides a tool to decouple biological heterogeneity from measurement (and procedural) variability. This framework is attractive for enabling reproducible TFM results, and automating quality control and analysis of large datasets.

2 Methods

This section is organized as follows. First, we provide an overview of traditional traction force microscopy (TFM) and introduce the workflow used in TFM with uncertainty quantification (TFM-UQ, §2.1). Then, we describe a new PIV method to measure substrate deformation and estimate its uncertainty (§2.2), a crucial input to TFM-UQ. Next, we describe the elastostatic model used in this work (§2.3). Finally, we describe a hierarchical Bayesian framework (§2.4) to propagate the uncertainty from deformation to traction stress and locally adjust regularization to account for spatially varying uncertainty levels.

2.1 Overview of TFM and TFM-UQ

Traction force Microscopy (TFM) requires a snapshot of fiduciary substrate deformation markers taken under stress, i.e., $\boldsymbol{I}_t$, to be compared with a stress-free reference image, $\boldsymbol{I}_{\mathrm{ref}}$. The stress-free state is achieved by lysing the cells, waiting for cells to move out of the region of interest, or using prior knowledge about fiduciary marker positions. The set of images $\boldsymbol{I} = \{\boldsymbol{I}_{\mathrm{ref}}, \boldsymbol{I}_t\}$ form the experimental data for TFM analysis (Fig 1 A). Traditional TFM is implemented in two independent steps (Fig 1 B). The first step determines substrate deformation using, e.g., Particle Image Velocimetry (PIV), to obtain the point-wise most likely displacement field, $\boldsymbol{u}_{\mathrm{PIV}}$, between $\boldsymbol{I}_t$ and $\boldsymbol{I}_{\mathrm{ref}}$. The second step solves the inverse elastostatic problem to recover the traction stresses $\boldsymbol{t}$. Traditional TFM methods output a point estimate of the $\boldsymbol{u}_{\mathrm{PIV}}$ and $\boldsymbol{t}$ fields as well as integrated quantities (Fig 1 B). These methods suffer from uncertainty propagation from the images to the traction stresses, but their analysis pipeline is not aware of the input data confidence level. Thus, the measurement uncertainty of the desired traction force reading is unavailable in traditional TFM.

Figure 1C summarizes the TFM-UQ analysis pipeline. Given a pair of experimental images, the first step in the pipeline is to estimate the substrate deformation and its uncertainty. To this end, we developed a non-parametric bootstrap PIV method (PIV-UQ, see §2.2) since PIV is the most common method to measure substrate deformation in TFM. In PIV-UQ, the input image pair is perturbed to find a model-free estimation of the empiricial distribution of the deformation field conditional on the $\boldsymbol{I}$ pair. Nevertheless, the TFM-UQ pipeline is not critically dependent on PIV method, and other registration techniques with UQ capabilities [48] could be adapted to work with TFM-UQ.

In the second step of the pipeline, the deformation vector ($\boldsymbol{u}_{\mathrm{PIV}}$) and its covariance matrix quantifying uncertainty ($\boldsymbol{\Sigma}_{\mathrm{PIV}}$) resulting from PIV-UQ are used as inputs to a Bayesian elastostatic inversion framework. We adopt a hierarchical Bayesian approach in which the model hyperparameters such as the level of regularization are treated as random variables. This strategy allows us to regularize the inversion according to the noise in deformation measurements introduced by the spatially varying features of each $\boldsymbol{I}$ pair. In addition, the hierarchial Bayesian approach provides the confidence interval for traction stress output (Fig 1 C).

2.2 PIV Uncertainty Quantification (PIV-UQ) of Substrate Deformations

Particle image Velocimetry (PIV) is a commonly used method to measure substrate deformations from a pair of TFM images $\boldsymbol{I} = \{\boldsymbol{I}_{\mathrm{ref}}, \boldsymbol{I}_t\}$. PIV identifies overlapping rectangular sub-domains $\mathbf{W}_{ij}$ (a.k.a. interrogation windows) in each member of $\boldsymbol{I}$, and registers them using cross-correlation as the similarity metric (Fig 2 A). This operation is performed for each $\mathbf{W}_{ij}$ resulting in a discrete deformation field $\boldsymbol{u}_{\mathrm{PIV}}$. Typically, PIV uses square interrogation windows of size W_L (pixels) separated by a constant distance of W_S pixels across the entire analyzed image. It is customary to set the window spacing to be a fraction of window size, with $W_S = W_L/2$ providing the optimal trade-off between computational cost and effective spatial resolution based on the Nyquist principle.

Data: Interrogation sub-windows, $\mathbf{W}_{ij} = \{\mathbf{W}_{ij,ref}, \mathbf{W}_{ij,t}\} \in \mathbb{R}^{n \times n}$
Result: $\hat{\mathbf{u}} \in \mathbb{R}^2, \sigma_{\hat{u}} \in \mathbb{R}^+$

$\hat{\mathbf{u}} \leftarrow \texttt{PIV}(\mathbf{W}_{ij,ref}, \mathbf{W}_{ij,t})$;
for $b = 1 : n_B$ **do**
    for $\mathbf{W}_{ij}^* = \{\mathbf{W}_{ij,ref}, \mathbf{W}_{ij,t}\}$ **do**
        Pixel indices, $i_p \leftarrow \texttt{U}(1, n^2, n^2/2)$;
        Deletion indices, $i_d \leftarrow [1 : n^2]\backslash i_p$;
        $\mathbf{W}_{ij}^*(i_d) \leftarrow 0$;
    end
    $\hat{\mathbf{u}}_b^* \leftarrow \texttt{PIV}(\mathbf{W}_{ij,ref}^*, \mathbf{W}_{ij,t}^*)$;
end
$\sigma_{\hat{u}} \leftarrow \texttt{ClusterAnalysis}(\{\hat{\mathbf{u}}^*\})$;

Fig 2. Particle Image Velocimetry with Uncertainty Quantification (PIV-UQ) method. Top:
Schematic of PIV and PIV-UQ. **(A)** The image data $\boldsymbol{I}$ is sectioned into overlapping sub-windows $\mathbf{W}_{ij}$ of
length W_L, their centers separated by W_S. PIV computes the most likely displacement for each
corresponding sub-window $\mathbf{W}_{ij}$ that corresponds to the maximum of the cross-correlation metric. Each pixel
of $\mathbf{W}_{ij}$ contributes to the correlation value, hence the maximization procedure. **(B)** PIV-UQ method
bootstraps the contribution of individual pixels to the cross-correlation metric. Pixel indices are randomly
sampled with replacement and the unsampled indices are set to 0 (Black pixels). Backslash denoted set
difference operation and $\texttt{U}(1, n^2, n^2/2)$ represents uniformly distributed, $n^2/2$ samples in the interval $[1, n^2]$.
The bootstrapped sample of size n_B is analyzed for multiple possible clusters and outliers, subsequently
providing a metric of variability for each sub-window (i.e. for each discrete vector of the deformation field).
Bottom: PIV-UQ bootstrap algorithm.

Consequently, the error and resolution of PIV are usually dictated by one parameter, *i.e.*, W_L.   133
Since, for a given image pair, both the PIV resolution and error increase with W_L, this parameter   134
must be chosen carefully, but current PIV methods used for TFM do not provide information about   135
the uncertainty in $\boldsymbol{u}_{\mathrm{PIV}}$. Below, we introduce a PIV-UQ method that estimates this measurement   136
uncertainty that can be used independent of TFM-UQ to inform the choice of parameters for   137
PIV analysis of experimental images, or in conjunction with TFM-UQ to propagate measurement   138

uncertainty to the recovered traction stresses.

PIV-UQ estimates the most likely substrate displacement $\hat{\mathbf{u}}$ and its uncertainty, $\sigma_{\hat{u}}$ by perturbing the input images to obtain a non-parametric, model-free approximation of the empirical distribution of substrate displacements given the images $\boldsymbol{I}$. PIV-UQ applies the bootstrap procedure outlined in Fig 2 (panel B and algorithm). As described in the figure, the bootstrap method uses square interrogation windows. However, this method is generally applicable to any PIV implementation, including more advanced algorithms using iterative window refinement or window deformation, and the various sub-pixelar interpolation methods for finding the correlation peak [51] by replacing the `PIV` function accordingly. For a given pair of interrogation windows, a first nominal PIV run is carried out to obtain the estimator $\hat{\mathbf{u}}$ for the $\boldsymbol{u}_{\mathrm{PIV}}$ field. Next, bootstrapped interrogation windows $\mathbf{W}_{ij}^*$ are built by randomly resampling pixels from $\mathbf{W}_{ij}$ with replacement, and unsampled pixels are left zero valued. The number of unsampled pixels is randomly drawn between 1 and the total number of pixels in the window using a uniform distribution. Then, PIV is performed to obtain the perturbed estimator $\hat{\mathbf{u}}^*$ from $\mathbf{W}_{ij}^*$. This process is repeated for n_B trials so that the set of bootstrap estimators $\{\hat{\mathbf{u}}^*\}$ represent each interrogation window's empirical distribution $p(\hat{\mathbf{u}}|\mathbf{W}_{ij})$.

The boostrapped distributions $p(\hat{\mathbf{u}}|\mathbf{W}_{ij})$ are further processed using density-based `clusterAnalysis` (DBSCAN algorithm [52] with parameters $\epsilon = W_S/10$ and `minPts` $= n_B/10$) to identify outliers from the bootstrap procedure and verify their unimodality. Because the PIV cross-correlation optimization looks for pure translations of $\mathbf{W}_{ij}$ in $\boldsymbol{I}$, motion patterns that significantly depart from translation (e.g., rotation, shear) can lead to multi-modal distributions where more than one value of $\hat{\mathbf{u}}$ is plausible [51]. In addition, image noise, lack of image features, and brightness gradients can bias $p(\hat{\mathbf{u}}|\mathbf{W}_{ij})$ towards a uniform or skewed distribution, producing erroneous estimates of $\hat{\mathbf{u}}$. PIV-UQ utilizes $p(\hat{\mathbf{u}}|\mathbf{W}_{ij})$ to validate $\hat{\mathbf{u}}$ in place of purely heuristic-based metrics such as ratio of correlation peaks [51,53]. The 'bad' windows that fail the validation are deleted and the standard deviation for each of the remaining sub-windows ($\sigma_{\hat{u}}$) is reported . For the purposes of TFM-UQ, the standard deviation of these windows are set to a high number (e.g. a factor of 5 to the maximum $\sigma_{\hat{u}}$). The number of 'bad' windows also serve as a diagnostic tool in selecting the parameter W_L and validation of PIV results for the image $\boldsymbol{I}$. For the rest of the manuscript, $\boldsymbol{u}_{\mathrm{PIV}}(\mathbf{x}) = \hat{\mathbf{u}}(\mathbf{x})$ and $\sigma_{u,\mathrm{PIV}}(\mathbf{x}) = \sigma_{\hat{u}}(\mathbf{x})$ for each interrogation window centered at $\mathbf{x} = (x, y)$. Coefficient of variation (CoV) $= \mathrm{Std}[\sigma_{\hat{u}}]/\mathrm{Mean}[\sigma_{\hat{u}}]$ is used as a metric to assess convergence of $\sigma_{\hat{u}}$ with number of bootstrap iterations n_B (see §3.3 and Fig SI.3).

2.3 Elastostatic model

We consider a flat inertia-less, linearly elastic substrate of Young's modulus E and Poisson's ratio ν, deformed under the action of traction stresses exerted at its free surface $z = h$. The substrate deformation field ($\boldsymbol{u}$) satisfies the elastostatic partial differential equation (PDE)

$$\nabla^2 \boldsymbol{u} + \frac{\nabla(\nabla \cdot \boldsymbol{u})}{1 - 2\nu} = 0, \tag{1}$$

with boundary conditions $\boldsymbol{u}(x, y, z = 0) = \boldsymbol{0}$ and $\boldsymbol{u}(x, y, z = h) = \boldsymbol{u}_\mathrm{h} \approx \boldsymbol{u}_\mathrm{PIV}$. Once this boundary value problem is solved, the constitutive relation for the stress tensor $\mathbf{T} = F(x, y, z; \boldsymbol{u}, E, \nu)$ allows for computing the traction stress vector at $z = h$ as $\boldsymbol{t}(x, y) = \mathbf{T}(x, y, z = h) \cdot \boldsymbol{n}$, where $\boldsymbol{n} \approx (0, 0, 1)$ is the vector normal to the substrate surface. Because the constitutive equation (i.e. F) and the elastostatic PDE are linear, we can denote the relationship between $\boldsymbol{t}$ and $\boldsymbol{u}_\mathrm{h}$ concisely using the linear transformation $\mathbf{M} : \boldsymbol{t} \to \boldsymbol{u}_\mathrm{h}$, where $\mathbf{M}$ is the elastic response operator. $\boldsymbol{u}_\mathrm{h}$ and $\boldsymbol{t}$ can be two- or three-component vector fields, and h can be finite or infinite.

Using this formulation, the forward problem of finding $\boldsymbol{u}_\mathrm{h}$ given a traction stress field $\boldsymbol{t}$ can be written as

$$\boldsymbol{u}_\mathrm{h} = \mathbf{M}\boldsymbol{t} + \boldsymbol{\varepsilon}_\beta + \boldsymbol{\varepsilon}_m, \tag{2}$$

$$\boldsymbol{\varepsilon}_\beta \sim N(0, \beta^{-1}\mathbf{U}), \tag{3}$$

$$\boldsymbol{\varepsilon}_m \sim N(0, \boldsymbol{\Sigma}_\mathrm{PIV}). \tag{4}$$

where $\boldsymbol{\varepsilon}_m$ is the measurement error arising from PIV and $\boldsymbol{\varepsilon}_\beta$ represents the other sources coming from the elastostatic model, numerical discretization, etc. Both error terms are modeled as normally distributed but $\boldsymbol{\varepsilon}_\beta$ is assumed to be global with precision parameter β, whereas $\boldsymbol{\varepsilon}_m$ is characterized by the spatially dependent variance $\boldsymbol{\Sigma}_\mathrm{PIV}$. Therefore, $\mathbf{U}$ is the identity matrix while $\boldsymbol{\Sigma}_\mathrm{PIV}$ is a diagonal matrix. The precision parameter β is unknown and will be modeled as a hyperparameter in our hierarchical Bayesian approach (§2.4). On the other hand, $\boldsymbol{\Sigma}_\mathrm{PIV}$ is measured directly using the PIV-UQ method described in the previous section (§2.2), and therefore, it will be treated explicitly.

The form of $\mathbf{M}$ depends on each model's geometry (e.g., finite vs. infinite h, 2D vs. 3D deformation), solution method (Fourier transform, finite elements, etc), and substrate material properties. The particular form of $\mathbf{M}$ is inconsequential for the TFM-UQ algorithm, as will be seen below. For simplicity, this report considers the Fourier Transform Traction Cytometry (FTTC) response function, $\widetilde{\mathbf{M}}$, to build $\mathbf{M}$. In the Fourier domain, $\widetilde{\mathbf{M}}$ adopts a straightforward block-diagonal form that can be derived analytically for finite and infinite h for various 2D and 3D boundary conditions at the substrate's surface [8, 9, 54, 55]. Without loss of generality, we used Lin *et al.*'s [54] concise derivation for the finite-h case with 2D boundary conditions:

$$\widetilde{M}(\mathbf{k}) = g_1(k) \begin{bmatrix} 1 & 0 \\ 0 & 1 \end{bmatrix} + g_2(k) \begin{bmatrix} k_x^2 & k_x k_y \\ k_x k_y & k_y^2 \end{bmatrix}, \tag{5}$$

where, $g_1 = 1/f_1$, $g_2 = -f_2/f_1(f_1 + k^2 f_2)$ and $k = |\mathbf{k}|$ is the absolute value of the wavenumber vector $\mathbf{k} = (k_x, k_y)$. The scalar functions are defined as

$$f_1 = E\frac{ck}{2(1+\nu)s}, \quad f_2 = E\frac{(3-4\nu)\nu sc^2 - (1-\nu)ckh + (1-2\nu)^2 s + 2(kh)^2}{(2(1-\nu^2)ks)((3-4\nu)sc + kh)},$$

with $c = \cosh kh$ and $s = \sinh kh$. Inverse Fourier transforming $\widetilde{\mathbf{M}}$, we obtain $\mathbf{M} = \mathbf{F}^H\widetilde{\mathbf{M}}\mathbf{F}$, where

$\mathbf{F}$ is the discretized Fourier Transform (DFT) matrix and $\mathbf{F}^H$ its Hermitian transpose.

The finite-dimensional matrix $\mathbf{M}$ has a large condition number because the eigenvalues of $\widetilde{\mathbf{M}}$ $\sim h$ for $kh \to 0$ (pure shear limit) while they vary as k^{-1} for $kh \gg 1$. Using the Nyquist criterion to determine the maximum wavenumber, $k_{max} = 2\pi/W_L$, we obtain $\kappa(\mathbf{M}) \sim h/W_L$, which worsens as the substrate becomes thicker and the spatial resolution becomes finer. Depending on the elastostatic model's geometry, the condition number of $\mathbf{M}$ can become even higher. For instance, under a more realistic assumption that the TFM fluorescent beads are located slightly under the substrate's surface, i.e., at $z = h - \Delta h$, the measured displacements are amplified by an additional factor $\exp(k\Delta h)$ [9] and, consequently, $\kappa(\mathbf{M}) \sim h/W_L \exp\left(2\pi\Delta h/W_L\right)$. Therefore the noise terms ε_β and ε_m are susceptible to amplification when equation 2 is inverted, and the level of amplification may vary significantly among TFM methods.

2.4   Hierarchical Bayesian TFM formulation to propagate PIV uncertainty

This section presents a Bayesian framework to invert the elastostatic problem in the presence of local measurement noise and global model noise. A three-level hierarchical framework, summarized in Fig 3, is formulated. Consistent with the elastostatic model presented in equations (2)–(4), the likelihood of the observed displacement given a traction stress field, $p(\boldsymbol{u}_{\mathrm{h}}|\boldsymbol{t})$, is governed by the elastostatic matrix $\mathbf{M}$. The error due to PIV measurement of $\boldsymbol{u}_{\mathrm{h}}$ is modeled explicitly based on the variance $\boldsymbol{\Sigma}_{\mathrm{PIV}}$ maps obtained from bootstrap PIV-UQ, and the random global error is modeled as additive noise of variance β^{-1}, where β is a hyperparameter. The likelihood of measured displacements, $p(\boldsymbol{u}_{\mathrm{h}}|\boldsymbol{t}, \beta, \boldsymbol{\Sigma}_{\mathrm{PIV}})$, can be expressed as a multivariate Gaussian distribution for the model residual,

$$p(\boldsymbol{u}_{\mathrm{h}}|\boldsymbol{t}, \beta, \boldsymbol{\Sigma}_{\mathrm{PIV}}) = (2\pi)^{-n/2}|\Lambda|^{-1/2} \exp\left[-\frac{1}{2}(\boldsymbol{u}_{\mathrm{h}} - \mathbf{M}\boldsymbol{t})^T \Lambda^{-1}(\boldsymbol{u}_{\mathrm{h}} - \mathbf{M}\boldsymbol{t})\right], \tag{6}$$

Here, $\Lambda = \beta^{-1}\mathbf{U} + \boldsymbol{\Sigma}_{\mathrm{PIV}}$, and this matrix's determinant is expanded as the product of its diagonal elements, $|\Lambda| = \prod_{i=1}^{2n}\left[\beta^{-1} + \boldsymbol{\Sigma}_{\mathrm{PIV},ii}\right]$, where $\boldsymbol{\Sigma}_{\mathrm{PIV},ii}$ occupies the i-th position in the diagonal of $\boldsymbol{\Sigma}_{\mathrm{PIV}}$. The dimensions of the multivariate Gaussian, n, is given by the number of displacement measurements (i.e., the total number of interrogation windows, $2N_w$).

The prior distribution of the traction stress $p(\boldsymbol{t})$ is based on a linear quadratic functional $\boldsymbol{t}^T\mathbf{L}\boldsymbol{t}$, leading to the zero-mean multivariate Gaussian distribution, *i.e.*,

$$p(\boldsymbol{t}|\boldsymbol{u}_{\mathrm{h}}, \alpha, \mathbf{L}) \propto \alpha^{n/2} \exp\left[-\frac{\alpha}{2}\boldsymbol{t}^T\mathbf{L}\boldsymbol{t}\right], \tag{7}$$

where α is a hyper-parameter. We used the discretized Laplacian operator for $\mathbf{L}$, establishing a Tikhonov low-pass filter of the traction stress field.

The hyperparameters α and β control the influence of the prior constraint and model misfit during the inversion process. The case of $\boldsymbol{\Sigma}_{\mathrm{PIV}} = 0$ and fixed hyperparameter values corresponds to the traditional (frequentist or Bayesian) approach, which is governed by the regularization

Fig 3. Hierarchical Bayesian TFM formulation for adaptive and self-consistent regularization. The problem of choosing a regularization parameter is treated in a Bayesian formulation. Substrate deformation ($\boldsymbol{u}_{\mathrm{h}}$) is the observed quantity, and it is modeled as additive contribution of traction stresses ($\boldsymbol{t}$), locally resolved PIV measurement uncertainty $\boldsymbol{\Sigma}_{\mathrm{PIV}}$ and a global model error (β) that is unknown. The hierarchical formulation allows to express unknown hyper-parameters (ϕ_{s}) for the model error and a prior on traction stress field as random variables to be inferred. The priors can encode reasonable assumptions such as smoothness and global force balance. Markov Chain Monte Carlo (MCMC), specifically an hybrid Gibbs sampling is used for inference from the marginal posterior distrbutions.

parameter $\lambda = \alpha/\beta$ [19, 21, 36, 56]. In our formulation, setting $\boldsymbol{\Sigma}_{\mathrm{PIV}} = 0$ would be equivalent to neglecting the PIV measurement error or to assuming it is spatially uniform so it can be lumped into the model error β.

In the hierarchical framework, the hyperparameters α and β are modeled as random variables, using hyperpriors derived from the Gamma distribution, $p(x|\theta, \phi) = \phi^{\theta} x^{\theta-1} \exp\{-\phi x\}/\Gamma(\theta)$, where the term in the denominator is the Gamma function. To keep the hyperpriors non-informative, the shape and rate parameters were fixed to constant values ($\theta = 1, \phi = 10^{-5}$). This hierarchical approach (Fig 3) allows for including an acceptable level of variability of the hyperparameters given the observed $\boldsymbol{u}_{\mathrm{h}}$ distribution.

The conditional posterior distributions $p(\boldsymbol{t}|\boldsymbol{u}_{\mathrm{h}}, \alpha, \beta)$, $p(\alpha|\boldsymbol{t}, \boldsymbol{u}_{\mathrm{h}})$ and $p(\beta|\boldsymbol{t}, \boldsymbol{u}_{\mathrm{h}})$ were obtained from the joint distribution of the traction stress, displacement and the hyperparameters,

$$
p(\boldsymbol{u}_{\mathrm{h}}, \boldsymbol{t}, \alpha, \beta | \boldsymbol{\Sigma}_{\mathrm{PIV}}, \mathbf{L}) = p(\boldsymbol{u}_{\mathrm{h}}|\boldsymbol{t}, \beta, \boldsymbol{\Sigma}_{\mathrm{PIV}}) p(\boldsymbol{t}|\alpha, \mathbf{L}) p(\alpha|\phi_{\alpha}, \theta_{\alpha}) p(\beta|\phi_{\beta}, \theta_{\beta}) \propto
$$

$$
\propto |\Lambda|^{-1/2} \alpha^{n/2+\theta_{\alpha}-1} \beta^{\theta_{\beta}-1} \exp\left\{-\frac{1}{2}\left[(\boldsymbol{u}_{\mathrm{h}} - \mathbf{M}\boldsymbol{t})^{T} \Lambda^{-1}(\boldsymbol{u}_{\mathrm{h}} - \mathbf{M}\boldsymbol{t})\right] - \frac{\alpha}{2}\boldsymbol{t}^{T}\mathbf{L}\boldsymbol{t} - \phi_{\alpha}\alpha - \phi_{\beta}\beta\right\}.
$$

Data: Displacement field, $\boldsymbol{u}_{\mathrm{h}} \in \mathbb{R}^n$; Displacement uncertainty, $\boldsymbol{\Sigma}_{\mathrm{PIV}}$; $\alpha_0, \beta_0 \in \mathbb{R}_{>0}$; Burn-in iterations, $\mathtt{it_b}$, max iterations, $\mathtt{it_m}$

Result: Posterior samples : $\{\boldsymbol{t}\} \in \mathbb{R}^n$; $\{\alpha\}, \{\beta\} \in \mathbb{R}_{>0}$

Initialization;
$\theta_\alpha \leftarrow 1; \theta_\beta \leftarrow 1;$
$\phi_\alpha \leftarrow 10^{-5}; \phi_\beta \leftarrow 10^{-5};$
$\mathtt{it} \leftarrow 1;$
$\alpha \leftarrow \alpha_0; \beta \leftarrow \beta_0$

while $\mathtt{it} < (\mathtt{it_b} + \mathtt{it_m})$ **do**

$\quad\Lambda^{-1} \leftarrow [1/\beta\mathbf{U} + \boldsymbol{\Sigma}_{\mathrm{PIV}}]^{-1};$

$\quad\mathbf{S}_{\mathrm{post}} \leftarrow [\mathbf{M}^T\Lambda^{-1}\mathbf{M} + \alpha\mathbf{L}]^{-1};$

$\quad\mathbf{S}_{\mathrm{post}} \leftarrow \frac{1}{2}[\mathbf{S}_{\mathrm{post}} + \mathbf{S}_{\mathrm{post}}^T];$

$\quad\boldsymbol{t} \sim N(\mathbf{S}_{\mathrm{post}}\mathbf{M}^T\Lambda^{-1}\boldsymbol{u}_{\mathrm{h}}, \mathbf{S}_{\mathrm{post}});$

$\quad\alpha \sim \Gamma\left(\frac{n}{2} + \theta_\alpha, \frac{1}{2}(\boldsymbol{t}^T\mathbf{L}\boldsymbol{t}) + \phi_\alpha\right);$

$\quad$/* Simulate β by inverse transform sampling of discrete p.d.f $\qquad$ */

$\quad\hat{\beta} \leftarrow \arg\min\log(p(\beta|\boldsymbol{u}_{\mathrm{h}}, \boldsymbol{t}, \theta_\beta, \phi_\beta));$

$\quad$Discrete β posterior domain, $\Omega_\beta \in [\hat{\beta} - 3\sigma_{\hat{\beta}}, \hat{\beta} - 3\sigma + \hat{\beta}]$;

$\quad$c.d.f, $F(x) = \int_{\Omega_{\hat{\beta}}} p(\beta|u, t, \theta_\beta, \phi_\beta);$

$\quad\beta \sim F_x^{-1}(\mathtt{U(0,1)});$

$\quad$**if** $\mathtt{it} > \mathtt{it_b}$ **then**

$\quad\quad|$ Write $\{\boldsymbol{t}, \alpha, \beta\}$.

$\quad$**end**

end

Fig 4. Hybrid Gibbs sampler for TFM UQ inference.

The posterior conditional distribution of traction stress is $\qquad\qquad$ ₂₄₂

$$p(\boldsymbol{t}|\boldsymbol{u}_{\mathrm{h}}, \alpha, \beta, \boldsymbol{\Sigma}_{\mathrm{PIV}}, \mathbf{L}) \propto \exp\left[-\frac{1}{2}(\boldsymbol{u}_{\mathrm{h}} - \mathbf{M}\boldsymbol{t})^T\Lambda^{-1}(\boldsymbol{u}_{\mathrm{h}} - \mathbf{M}\boldsymbol{t}) - \frac{\alpha}{2}\boldsymbol{t}^T\mathbf{L}\boldsymbol{t}\right]$$

$$\propto \exp\left[-\frac{1}{2}\left(\boldsymbol{t} - S_{\mathrm{post}}\mathbf{M}^T\Lambda^{-1}\boldsymbol{u}_{\mathrm{h}}^T]\mathbf{S}_{\mathrm{post}}^{-1}[\boldsymbol{t} - \mathbf{S}_{\mathrm{post}}\mathbf{M}^T\Lambda^{-1}\boldsymbol{u}_{\mathrm{h}}^T\right)\right], \qquad (8)$$

with the covariance matrix $\mathbf{S}_{\mathrm{post}} = \left[\mathbf{M}^T\Lambda^{-1}\mathbf{M} + \alpha\mathbf{L}\right]^{-1}$. In turn, the conditional posterior distri- ₂₄₃ bution for the hyperparameter α is ₂₄₄

$$p(\alpha|\boldsymbol{t}, \boldsymbol{u}_{\mathrm{h}}, \mathbf{L}) \propto \alpha^{n/2 + \theta_\alpha - 1}\exp\left[-\frac{\alpha}{2}\left(2\phi_\alpha + \boldsymbol{t}^T\mathbf{L}\boldsymbol{t}\right)\right]. \qquad (9)$$

These two probability density functions correspond to a Gaussian and a Gamma distribution, i.e., ₂₄₅

$$\boldsymbol{t}|\boldsymbol{u}_{\mathrm{h}}, \alpha, \beta, \boldsymbol{\Sigma}_{\mathrm{PIV}}, \mathbf{L} \sim N(\mathbf{S}_{\mathrm{post}}\mathbf{M}^T\Lambda^{-1}\boldsymbol{u}_{\mathrm{h}}^T, \mathbf{S}_{\mathrm{post}}), \qquad (10)$$

$$\alpha|\boldsymbol{t}, \mathbf{L}, \theta_\alpha \sim \Gamma\left(\frac{n}{2} + \theta_\alpha, \frac{1}{2}|\mathbf{L}\boldsymbol{t}|^2 + \phi_\alpha\right), \qquad (11)$$

and can be found in closed form. However, a similar derivation for the conditional posterior ₂₄₆

distribution for β,

$$p(\beta|\boldsymbol{t}, \boldsymbol{u}_{\mathrm{h}}) \propto |\Lambda|^{-1/2}\beta^{\theta_\beta-1} \exp\left[-\frac{1}{2}(\boldsymbol{u}_{\mathrm{h}} - \mathbf{M}\boldsymbol{t})^T\Lambda^{-1}(\boldsymbol{u}_{\mathrm{h}} - \mathbf{M}\boldsymbol{t}) - \phi_\beta\beta\right], \tag{12}$$

did not produce an expression in closed form, since this expression also depended on $\boldsymbol{t}$. Consequently, we used the iterative Markov Chain Monte Carlo (MCMC) method [57,58] to infer $p(\boldsymbol{t}|\boldsymbol{u}_{\mathrm{h}}, \alpha, \beta)$, $p(\alpha|\boldsymbol{t}, \boldsymbol{u}_{\mathrm{h}})$ and $p(\beta|\boldsymbol{t}, \boldsymbol{u}_{\mathrm{h}})$. The MCMC approach also provides a straightforward way to calculate the quantity of interest, i.e., the marginal posterior distribution of the traction stress field, $p(\boldsymbol{t}|\boldsymbol{u}_{\mathrm{h}}) = \int p(\boldsymbol{t}|\boldsymbol{u}_{\mathrm{h}}, \alpha, \beta)\mathrm{d}\alpha\mathrm{d}\beta$ by summing over MCMC draws. The statistics of interest is approximated from the MCMC draws after a burn-in period ($\mathtt{it_b}$). Specifically, the expected (mean) marginal posterior traction stress field is given by $\hat{\mathbf{t}}(\mathbf{x}) = \mathbb{E}[\boldsymbol{t}(\mathbf{x})|\boldsymbol{u}_{\mathrm{h}}] \approx \frac{1}{N_d}\sum_i \boldsymbol{t}(\mathbf{x})_i$ and the marginal posterior standard deviation of the traction stress is $\mathrm{Std}[\boldsymbol{t}(\mathbf{x})|\boldsymbol{u}_{\mathrm{h}}] \approx \{\frac{1}{N_d}\sum_i (\boldsymbol{t}(\mathbf{x})_i - \mathbb{E}[\boldsymbol{t}(\mathbf{x})|\boldsymbol{u}_{\mathrm{h}}])^2\}^{1/2}$, where $\boldsymbol{t}_i$ represents the i^{th} draw of the vector of traction stress (out of N_d draws) from the Hybrid-Gibbs sampler. $\hat{\mathbf{t}}$ is the traction stress field output and the total, pointwise traction stress uncertainty $\sigma_t = \sqrt{\sigma_{tx}^2 + \sigma_{ty}^2}$, where σ_x is $\mathrm{Std}[t_x|\boldsymbol{u}_{\mathrm{h}}]$ and so on.

A hybrid-Gibbs sampler was adopted to sample these distributions using equations (10 - 12), as described in Fig 4. The hyperparameters were initialized to α_0 and β_0, which can be selected from reasonable maximum likelihood estimate values. In each sampling iteration, Gibbs sampling was performed by sequentially sampling the conditional posteriors, $p(\boldsymbol{t}|\boldsymbol{u}_{\mathrm{h}}, \alpha, \beta)$ and $p(\alpha|\boldsymbol{t}, \boldsymbol{u}_{\mathrm{h}})$. The posterior $p(\beta|\boldsymbol{u}_{\mathrm{h}}, \boldsymbol{t})$ was constructed numerically around the maximum *a posteriori* probability of $\beta|\boldsymbol{u}_{\mathrm{h}}, \boldsymbol{t}$ and an inverse transform sampling using the numerical cumulative distribution function was performed for $\beta|\boldsymbol{u}_{\mathrm{h}}, \boldsymbol{t}$. The burn-in was typically set to 20-50 iterations, and sampling was perfomed until convergence as assessed by traceplots and autocorrelation of the sampling chain. The convergence and therefore, the computation time, of Gibbs sampler depends on the size of the number of parameters (N_w) and the strength of spatial correlation in prior (i.e. $\mathbf{L}$) and $\boldsymbol{\varepsilon}_m$. Using a laptop computer with Intel Core i7-10875H CPU, for typical resolution of PIV field ($N_w \sim 30$ and $\mathtt{t_m} = 100$ iterations), it takes less than 3 minutes. For a higher resolution with $Nw \sim 60$, it takes approximately 1.5 hours. Since the matrix inversion is the computational bottleneck, workstations with more CPU cores can be used to speed up the computation.

2.5   Cell culture

Human vascular umbilical endothelial vein cells (HUVECs) (Cell Applications) were cultured in M199 supplemented with 10% (v/v) endothelial growth medium (Cell Applications), 10% (v/v) fetal bovine serum (Gibco), and 1% penicillin-streptomycin (Gibco). Mouse mesenchymal C3H/10T1/2 (ATCC) were cultured in low glucose Dulbecco's Modified Eagle's Medium (DMEM) supplemented with 10% v/v fetal bovine serum (Gibco) and 1% penicillin-streptomycin (Gibco). The cell plasma membranes were stained with CellMask ($5\,\mu\mathrm{g/mL}$ Thermo Fisher Scientific) to visualize the HUVEC cells.

2.6 Micropatterned polyacrylamide gel preparation

3.5 mm glass bottom dishes were pre-treated in an ultraviolet ozone (UVO) box for 5 min, and the glass surface was then activated by using 2 M NaOH for 5 min. The whole dishes were then washed with distilled water, dried, and treated with 3-aminopropyl-trimethoxysilane for 20 min. After removing silane, glass surfaces were rinsed with 100% EtOH, dried, and treated with 0.5% glutaraldehyde for 30 min. The activated surfaces were rinsed with distilled water and kept at room temperature for their use within the same day.

Round 12 mm glass coverslips were pretreated in an UVO box for 5 min. The coverslips were then incubated with a 110 μL drop of 0.2 mg/mL PLL-PEG (poly[L-lysine] grafted with poly[ethylene glycol]) (Nanosoftpolymers) for 30 min at room temperature. After 30 min, PLL-PEG was removed from the coverslip surface by aspiration. The chrome face of the photomask (Advance Reproductions) was activated by UVO for 3 min. The PLL-PEG-coated coverslips were attached to the chrome side of the photomask by sandwiching a 2 μL drop of distilled water between both surfaces. The assembly of the photomask and coverslips was then exposed to UVO light for 5 min. The photomask was detached from the coverslips by adding distilled water and then dried by aspirating excess water. A 110 μL drop of 50 μg/mL fibronectin (FN; Sigma-Aldrich) was placed on the PLL-PEG-coated surface of the coverslip and incubated for 1 hour. After 1 hour, the FN was removed, and the coverslip was washed once with distilled water. At this stage, the coverslip was ready for printing patterned FN on the PA gel surface. A thin film of FN was deposited on the areas of the coverslips that had been exposed to UVO light, whereas PLL-PEG prevented FN adhesion to the glass surface in the shadowed areas [59].

The polyacrylamide gels were fabricated with a mixture of acrylamide and bis-acrylamide following a well-established method [60]. For traction force microscopy, fluorescent 0.2 μm diameter beads were added to the mixture for later use as fiduciary markers of gel deformation. To promote bead distribution toward the surface of the gel, phosphate-buffered saline (PBS) was used instead of distilled water. The Young's modulus was controlled via the amount ratio of acrylamide/bis-acrylamide, as previously described [60]. Once both coverslips were ready, freshly made 10% ammonium sulfate and tetramethyl ethylenediamine were added to the polyacrylamide and bis-acrylamide mixture to initiate gel polymerization. Immediately after, a 2.5 μL drop of the mixture was pipetted on the treated glass bottom dish and sandwiched with the FN-patterned surface of the 12 mm round coverslip. The assembly was polymerized for 45 min before removing the round coverslip. The unpolymerized acrylamide was removed by rinsing twice with PBS. The resulting patterned gels were sterilized under 354 nm light for 5 min before adding the cells. The cells were seeded on top of the gels and allowed to adhere for 30 min. Unattached cells were washed off to avoid overgrowth of the patterns. The medium was reconstituted, and the cells were incubated overnight in the patterned regions.

3 Results

This section presents a computational verification and experimental demonstration of the proposed PIV-UQ and TFM-UQ techniques. PIV-UQ is verified using synthetic fluorescent bead images with prescribed displacement fields and noise sources in the image. PIV-UQ is shown to estimate substrate deformation uncertainty for each interrogation window. This information can be factored into the criteria to choose the critical PIV parameter, i.e., the interrogation window size, W_L. The synthetic simulations are then extended to the entire TFM pipeline to probe the influence of heteroskedasticity in PIV displacement data during the elastostatic inversion step. TFM-UQ applied to experimental data demonstrates the utility of marginal posterior distribution in quantifying the variability due to hyper-parameter selection and numerical implementation, and the automatic identification of high traction stress error areas associated with image artifacts.

3.1 PIV-UQ identifies local uncertainty arising from image features and substrate deformation

The bootstrap PIV-UQ method was verified on synthetically generated fluorescent bead images with spatially varying uncertainty. Fig SI.1 describes in detail the computational pipeline to generate synthetic images for validation. Briefly, 1024×1024 px images containing randomly placed synthetic beads with Gaussian intensity profiles (max. intensity 1000 a.u.) were generated. Each bead's centroid was displaced according to a prescribed substrate deformation field, and different fields including translation and shear were simulated. Gaussian pixel noise was added to the image with varying levels of signal-to-noise ratio (SNR), as shown in the left column of Fig 5A. The process of random bead placement, displacement, and noise addition was repeated to generate ensembles of images. The baseline PIV and PIV-UQ algorithms were tested on all of the realizations and RMS values ($\sigma_{u,\text{PIV}}$ and $\sigma_{u,ens}$ respectively. Total uncertainty, $\sigma = \sqrt{\sigma_{ux}^2 + \sigma_{uy}^2}$) were determined.

Two displacement fields were simulated (units in pixels):

$$\boldsymbol{u_1} = \begin{bmatrix} 2 \\ 0.6 \end{bmatrix}, \quad \boldsymbol{u_2} = \begin{bmatrix} \sin(\alpha) & g\cos^2(\alpha) \\ \cos(\alpha) & g\sin^2(\alpha) \end{bmatrix} \begin{bmatrix} x_0 \\ y_0 \end{bmatrix}. \tag{13}$$

Here, $\alpha = 45°$, $g = 20$ and $x_0, y_0 \in [-1, 1]$ px for the simulation domain. In line with literature [44,48], a uniform displacement field $\boldsymbol{u_1}$ was simulated to understand the statistical convergence of the PIV-UQ bootstrap method for varying levels of image noise, quantified by the signal-to-noise ratio SNR. In addition, a second displacement field, $\boldsymbol{u_2}$, representing shear was used as benchmark to mimic the substrate deformation gradients typically seen in TFM experiments.

The left panels of Figs 5B,C display arrow plots of the ground-truth, ensemble-averaged displacement fields, $\boldsymbol{u}_{ens}$ overlaid on contour plots depicting the ensemble standard deviation, $\sigma_{u,ens}$. The right panels show the displacements and standard deviation obtained by the PIV-UQ bootstrap method, $\boldsymbol{u}_{\text{PIV}}$ and $\sigma_{u,\text{PIV}}$. For the uniform displacement field ($\boldsymbol{u}_1$), $\sigma_{u,ens}$ was spatially uniform

Fig 5. PIV-UQ synthetic validation. (A), (D) Gaussian profile synthetic beads are randomly generated at varying signal-to-noise (SNR) ratio. A uniform (U) displacement field (B,E) or a shear (S) displacement field (C,F) is applied to the synthetic beads and PIV-UQ is performed as described in §2.2. In (B,C,E and F), left columns show ground-truth data consisting of applied displacement field vectors overlaid on contour maps of ensemble standard deviation ($\boldsymbol{u}_{ens}$ and $\sigma_{u,ens}$). The right columns denote represent the same data estimated from PIV-UQ ($\boldsymbol{u}_{PIV}$ and $\sigma_{u,PIV}$). (G) Joint distribution of ($\sigma_{u,ens}, \sigma_{u,PIV}$) for uniform (U) and shear (S) deformation fields for W_L of 32 and 64 pxs. (H) RMS of ensemble standard deviation and PIV UQ bootstrap estimates are compared for varying SNR values. (I) PIV Window size as a function of SNR, derived from data presented in (H) for isolines of $\sigma_{u,PIV}$.

(Fig 5B). On the other hand, the shear displacement field ($\boldsymbol{u}_2$) showed a clear spatial gradient and reaches higher values in upper left corner and lower right corner of the image. Visual comparison of the ground-truth and PIV-UQ data suggests that the bootstrap method was able to estimate PIV uncertainty and its spatial variations, even if $\sigma_{u,\text{PIV}}$ seemed to overestimate $\sigma_{u,ens}$.

    The uncertainty estimation of PIV-UQ became more accurate when the image SNR decreased (Fig 5D–F). To analyze in more detail the ability of PIV-UQ to distinguish regions of varying deformation uncertainty, we plotted scatter plots of *true* standard deviation of each displacement vectors vs. the corresponding PIV-UQ estimates in Fig 5G for uniform (U, circles) and shear (S, triangles) displacement fields, using two different interrogation window lengths, $W_L = 32, 64$. The resulting point clouds approached the identity line $\sigma_{u,\text{PIV}} = \sigma_{u,ens}$ as uncertainty increased but there was a non-zero $\sigma_{u,\text{PIV}}$ for zero true uncertainty, corresponding to the "floor" error level introduced by the bootstrap procedure.

Fig 6. Synthetic pipeline to simulate experimentally relevant, spatially heterogeneous noise levels in TFM. (A) Synthetic traction stress generated from four traction islands of Gaussian profile, t_G, one in each quadrant with a maximum stress of $500\,\mathrm{Pa}$. (B) Synthetic displacement field resulting from traction profile in (A). Substrate of $E = 5$ KPa and Poisson's ratio of 0.45 is used here. (C) Schematic depicting spatially heterogeneous noise addition directly to the synthetic image by varying fluorescent bead density in X-direction. The synthetic image also has image pixel noise equivalent to SNR= 50 (D) Synthetic fluorescent bead image encoding image pixel noise and bead density variations. (E) PIV-UQ deformation measurement u_PIV measured in the synthetically deformed bead images (D). (F) PIV-UQ estimation of uncertainty (standard deviation $\sigma_{u,\mathrm{PIV}}$) showing higher uncertainty increasing with x-direction in agreement with noise addition process (C).

Next, we analyzed the overall PIV uncertainty for the u_1 field vs. SNR for different values of W_L (Fig 5H). As expected, $\sigma_{u,\mathrm{PIV}}$ decreased as SNR increased and this dependence was steeper for smaller interrogation windows. In particular, $\sigma_{u,\mathrm{PIV}}$ followed the ground-truth values closely for $SNR \lesssim 30$ but it saturated for higher SNR values to the floor noise level described above. Of note, Fig 5H illustrates the common knowledge that PIV error decreases as the interrogation window is enlarged, setting a trade-off between spatial resolution and measurement error. Since PIV-UQ can quantify this dependence, it can be practical to plot W_L versus SNR for constant levels of σ, as in Fig 5I. The different curves in that plot indicate how the interrogation window size should be adjusted to keep a desired precision when performing PIV on images of varying quality. This approach offers an objective criterion to choose W_L in TFM experiments.

3.2 Heteroskedasticity of PIV measurement improves TFM regularization

Armed with the ability to estimate the spatially non-uniform (i.e., heteroskedastic) uncertainty in substrate deformation, we investigated how TFM-UQ propagated this uncertainty forward to the

recovered traction stresses and how the regularization process was affected by heteroskedasticity. To establish ground-truth deformation data for this analysis, we used a synthetic traction stress field, $\boldsymbol{t}$, consisting of Gaussian traction islands ($\boldsymbol{t}_G$) i.e.,

$$\boldsymbol{t}_G = (t_{x,G}, t_{y,G}) = \left(t_p \exp \left\{ - \left[(x - x_0)^2 + (y - y_0)^2 \right] / \left(2s^2 \right) \right\}, 0 \right), \qquad (14)$$

centered at (x_0, y_0) with a width equal to s^2 ($s = 15\,\mu\text{m}$) and peak traction stress $t_p = 500\,\text{Pa}$. The stress field had four quadrants, each containing one traction island $\boldsymbol{t}_G$ pointing towards the vertical axis $x = 0$, as depicted in Fig 6A. The corresponding deformation field, determined by solving equation (1) using Lin *et al*'s solution [54], is shown in Fig 6B. Synthetic fluorescent bead images were randomly generated and displaced according to the computed deformation field, as described in §3.1 above. To consider heteroskedasticity, a horizontal image noise gradient was generated by manipulating the bead density (Fig 6C–D) while keeping a spatially constant pixel noise (SNR = 50). The resulting images were used as inputs to PIV-UQ to determine the displacements and the corresponding standard deviation fields, $\boldsymbol{u}_{\text{PIV}}(x, y)$ and $\sigma_{u,\text{PIV}}(x, y)$, are represented in Fig 6E–F. Consistent with the results shown in §3.1, PIV-UQ captured the horizontal gradient of deformation uncertainty, as well localized increases in $\sigma_{\boldsymbol{u},\text{PIV}}(x, y)$ at the four traction islands, where deformation gradients were more intense. The PIV-UQ outputs were used to build $\boldsymbol{u}_{\text{h}} = \{u_{PIV}(x_1, y_1), \ldots, u_{PIV}(x_{Nw}, y_{Nw}), v_{PIV}(x_1, y_1), \ldots, v_{PIV}(x_{Nw}, y_{Nw})\}$ and the covariance matrix $\boldsymbol{\Sigma}_{\text{PIV}} = \text{diag} \left\{ \sigma_{u,\text{PIV}}^2(x_1, y_1), \ldots, \sigma_{u,\text{PIV}}^2(x_{Nw}, y_{Nw}), \sigma_{v,\text{PIV}}^2(x_{Nw}, y_{Nw}), \ldots, \sigma_{v,\text{PIV}}^2(x_{Nw}, y_{Nw}) \right\}$, where N_w is the total number of interrogation windows in each direction of the image. These variables were passed as inputs to the TFM-UQ framework described in §2.4.

The traction stress field recovered by TFM-UQ accurately approximated the prescribed ground-truth $\boldsymbol{t}$ (Fig 7A). It is worth pointing out that TFM-UQ adapted its regularizer to the spatially varying noise of the displacement field, delivering a smooth traction stress map with spatial patterns resembling the ground-truth $\boldsymbol{t}$ field. Consistent with noise being larger on the right-hand side of the image, regularization was stronger on that side. Moreover, TFM-UQ quantified the uncertainty of the recovered traction stresses at each point, defined as $\sigma_t = \sqrt{\sigma_{tx}^2 + \sigma_{ty}^2}$, (Fig 7B). This uncertainty increased from left to right, consistent with the noise gradient introduced in the image and the resulting $\boldsymbol{\Sigma}_{\text{PIV}}$ distribution (Fig 6F). It also showed local maxima at the traction islands, where deformation gradients were highest.

Despite exhibiting a spatial gradient, the left-to-right growth of σ_t saturated quickly and its maximum value, $\sigma_{t,max} = 40$ Pa, was only 43% than its minimum value, $\sigma_{t,min} = 28$ Pa (Fig 7B). This uncertainty saturation was caused by the α-dependent noise reduction prior in eq. 7, as expected from the Bayesian theorem. To illustrate this phenomenon, here we provide an estimate for $\sigma_{t|\alpha,\beta}$ using the diagonal of the covariance matrix,

$$\boldsymbol{S}_{\text{post}} = \left[\boldsymbol{M}^T \left(\beta^{-1} \boldsymbol{U} + \boldsymbol{\Sigma}_{\text{PIV}} \right)^{-1} \boldsymbol{M} + \alpha \boldsymbol{L} \right]^{-1}.$$

Fig 7. TFM-UQ adaptively regularizes based on the local displacement uncertainty. (A) Mean marginal posterior traction stress distribution ($\hat{\mathbf{t}} = \mathbb{E}[\mathbf{t}|\mathbf{u}_\mathrm{h}]$) approximated from Hybrid-Gibbs sampling. (B) Pointwise marginal posterior uncertainty (standard deviation σ_t) of posterior traction stress (C) L-curve to determine Tikhonov regularization parameter λ in traditional TFM methods. λ_D is obtained by finding the corner of the L-curve (maximum curvature) for differential noise simulation (Fig 6). λ_L and λ_H are regularization parameters obtained from L-curve corners of homoskedastic synthetic simulations with spatially uniform low noise (L) or high noise (H). (D,E,F) Tikhonov regularized traction stress field corresponding to L-cruve corner λ_D (D), low noise λ_L (E) and high noise λ_H. **L** is the discretized Laplacian operator. (G,H,I) Tikhonov regularized traction stress field, with $\mathbf{L} = \mathbf{U}$ identity matrix. λ_D, λ_L and λ_H were determined to match 95^{th} percentile of traction stress magnitude with the respective fields in (D-F).

This matrix has a heteroskedastic contribution coming from the PIV measurement variance, $\mathbf{\Sigma}_\mathrm{PIV}$ and a homoskedastic contributions coming from the β-proportional global error. Both contributions are affected by the elastostatic operator $\mathbf{M}$. In addition, $\mathbf{S}_\mathrm{post}$ has a homoskedastic contribution coming from the α-proportional noise reduction prior. Consider the homoskedastic model with identity prior (i.e. $\mathbf{\Sigma}_\mathrm{PIV} = \mathbf{0}$ and $\mathbf{L} = \mathbf{U}$ [19, 41]) equivalent to traditional Tikhonov regularization ($\lambda = \alpha/\beta$). Considering the long wavelength limit $kh \ll 1$ for $\mathbf{M}$, the elastic interaction becomes local and spatially decoupled ($\mathbf{M} \to h/\mu\mathbf{U}$), where $\mu = E/2(1+\nu)$ is the shear modulus of the substrate [54]. In this case, each diagonal element of posterior traction, $\mathbf{S}_{\mathrm{post},ii} \to \beta^{-1}[h^2/\mu^2 + \lambda]^{-1}$. $\mathbf{S}_{\mathrm{post},ii}$ is bound between $\mu^2/h^2\beta^{-1}$ and α^{-1} for $h^2/\mu^2 \gg \lambda$ and $h^2/\mu^2 \ll \lambda$ respectively. The former limit is

the maximum likelihood estimate when prior contribution is minimum (less regularization) and vice versa. When $\varepsilon_m \neq 0$ (neglecting ε_β), the corresponding bounds for $\mathbf{S}_{\text{post},ii}$ are $\mu^2/h^2 \mathbf{\Sigma}_{\text{PIV},ii}$ and α^{-1} when $h^2/\mu^2 \gg \alpha \mathbf{\Sigma}_{\text{PIV},ii}$ and $h^2/mu^2 \ll \alpha \mathbf{\Sigma}_{\text{PIV},ii}$ respectively. This illustrates that regularization in TFM-UQ is local corresponding to a spatially different regularization parameter $\lambda_i \sim \alpha \mathbf{\Sigma}_{\text{PIV},ii}$ compared to traditional regularization.

For finite-thickness substrates, the response of $\mathbf{M}$ depends on the frequency of the displacement field. Despite ε_m inducing heteroskedasticity in the posterior posterior traction stress distribution, the choice of a homoskedastic prior (i.e. global α) sets the global scale bringing the posterior variance in $\mathbf{S}_{\text{post}}$ close to α^{-1} (Fig SI.2 A). For Laplacian $\mathbf{L}$ matrix, a similar results are obtained with $\alpha \nabla^2(\boldsymbol{t}) \mathbf{\Sigma}_{\text{PIV},ii}$ depending on the local smoothness (Fig 6 A,B). This way, the Bayesian framework provides a systematic way to understand the effect of prior construction on both the scale of recovered traction stresses as well as the uncertainty propagation. To palliate this saturation, one could devise a heteroskedastic noise reduction priors such as replacing eq. 7 with

$$p(t|\alpha, \mathbf{L}) \propto \exp\left[-\frac{\alpha}{2}\boldsymbol{t}^T \mathbf{\Sigma}_{\text{PIV}}^{-1} \mathbf{L}\boldsymbol{t}\right],$$

which would make both the minimum and maximum values of the diagonal of $\mathbf{S}_{\text{post}}$ proportional to $\mathbf{\Sigma}_{\text{PIV}}$ under the assumptions outlined above. However, this prior would also provide weaker regularization in high-noise areas. Repeating the analysis in Fig 7A-B for this heteroskedastic noise reduction prior confirmed these predictions (see SI.2 B).

To further understand how TFM-UQ incorporated heteroskedasticity to regularize traction stresses, we compared this method's results with those obtained using a conventional, homoskedastic Tikhonov regularizer based on the same Laplacian kernel $\mathbf{L}$ and the elastostatic operator $\mathbf{M}$. The conventional regularizer depends on one hyperparameter, $\lambda = \alpha/\beta$, which was chosen by analyzing the L-curve, i.e., a log-log plot of $|\boldsymbol{t}^T \mathbf{L}\boldsymbol{t}|$ versus $|(\boldsymbol{u}_{\text{h}} - \mathbf{M}\boldsymbol{t})^T(\boldsymbol{u}_{\text{h}} - \mathbf{M}\boldsymbol{t})|$ (Fig 7C), to balance the solution's smoothness and its fit to the data as described previously [43, 61]. The value of λ at the point where the L-curve has the maximum curvature (the corner of the L-curve) is usually selected to regularize $\boldsymbol{t}$. Higher values of λ underfit and are located to the right of the L-curve corner. Conversely, lower λ values overfit and are located to the right of the corner. This standard approach attempts to find a one-fits-all value of λ for the whole image, which can lead to underfitting and overfitting in different regions of experiments with spatially dependent noise. Here, it is important to remark that the noise not only comes from the image but from the deformation imparted by the cells in the substrate, so local gradients of noise can arise even in high-quality images with uniform bead distributions.

In our validation dataset, we tested the conventional regularizer for three λ values: the optimal value λ_D corresponding to the L-curve corner, as well as a pair of values λ_H and $\lambda_L < \lambda_H$ located on opposite sides of the corner. Compared with the ground-truth traction stresses shown in Fig 6A, the $\mathbf{L}$-based Tikhonov regularizer underestimated and oversmoothed the traction stress field for $\lambda = \lambda_D$ (Fig 7D), showing significantly more spread patterns and lower traction stress values than TFM-UQ,

	True	TFM-UQ	λ_D	λ_L	λ_H
Weighted RMS error in t (Pa)	-	**26.44**	38.48	*64.46*	41.51
RMS σ_t (Pa)	-	**23.36**	N/A	N/A	N/A
Foreground mean t (Pa)	182.22	99.94	77.78	**170.13**	*63.53*
Foreground max t (Pa)	496	252.95	138.57	**477.31**	*110.17*
Background mean t (Pa)	8.65	**14.36**	23.67	*66.47*	22.38
Background max t (Pa)	76.79	**88.55**	119.20	*1428.77*	96.25
SNR	21.05	**6.96**	3.28	*2.56*	2.84

Table 1. Performance comparison of locally regularized TFM-UQ with classical (global) Tikhonov regularization with discrete Laplacian Prior. The **best** and *worst* performing methods are highlighted in bold and italic respectively for each metric.

without completely cancelling the noise (see the defect in the upper right quadrant). For $\lambda = \lambda_L$, **L**-based Tikhonov did not underestimate traction stresses values or spread their pattern (Fig 7E). However, this method produced significant noise and artifacts including spurious concentrations of large traction stress in areas with zero ground-truth values. Finally, **L**-based Tikhonov severely underestimated and oversmoothed the traction stresses for $\lambda = \lambda_L$, producing non-zero values of t almost everywhere in the region of interest (Fig 7F).

For reference, we also recovered traction stress maps for λ_D, λ_L, and λ_H using a Tikhonov regularizer with the identity matrix (i.e., $\mathbf{L} = \mathbf{U}$) instead of the discretized Laplacian operator, since this regularizer was preferred in several previous studies [19, 21]. These maps, displayed in Fig 7G–I, showed a similar trend as those obtained with the Laplacian smoother in their variation with λ. On the other hand, they had less spread stress contours but displayed more noise and spurious traction islands.

The behaviors visualized in Fig 7 manifested quantitatively in metrics of the recovered traction stress fields, summarized in Table 1. TFM-UQ provided the minimum root mean square (RMS) error with respect to the ground truth, and this error was well approximated by the TFM-UQ RMS σ_t estimate. To quantify how different regularizers balanced overfitting and underfitting, we defined the foreground traction map as the region where the true stress magnitude $|t| \geq \mathrm{RMS}(t)$ (i.e., 75 Pa for the synthetic field of 6 A). Likewise, the background traction map was defined as the region where $|t| < \mathrm{RMS}(t)$. We then evaluated the mean and maximum traction stress magnitude for the foreground and background maps. A TFM method with good overfitting-underfitting balance should yield both a large foreground traction magnitude and a low background traction magnitude. This analysis, also summarized in Table 1, showed that TFM-UQ outperformed the standard regularizers in all metrics except in mapping the foreground traction, where it ranked second best after the standard λ_L regularizer. However, λ_L performed worst among all methods in RMS error, keeping a low background traction, and overall SNR of the traction map. Overall, TFM-UQ showed best balance between overfitting and underfitting by taking into account the local deformation uncertainty from PIV-UQ. In addition, TFM-UQ was able to propagate PIV uncertainty to estimate its the error of the recovered traction stresses.

3.3 Experimental Demonstration of PIV-UQ and TFM-UQ

This section describes two sets of live cell experiments demonstrating the practical application of PIV-UQ and TFM-UQ, and illustrate how these methods deal with spatially varying uncertainty in the measured substrate deformation. First, we performed experiments on C3H/10T1/2 fibroblasts seeded on substrates with micropatterned square fibronectin islands of size $75\,\mu$m (Fig 8 A). Micropatterning of extracellular matrix proteins on cellular substrates is a common approach for guiding adherent cells to adopt predefined shapes [59]. In our study, we used patterned substrates as a benchmark for TFM-UQ because theses substrates can exhibit gradients of fluorescent bead brightness and background noise, as shown in Fig 8B. This feature allowed us to objectively assess PIV-UQ and TFM-UQ versus different levels of image noise when observing the same cell at the same instant of time. To this end, we compared the traction stresses recovered from the "raw" images obtained by wide-field fluorescence microscopy as well as from a second dataset (Fig 8 C) obtained by applying a non-linear background subtraction method (Leica Instantaneous Image Clearing [62]) that improved the signal-to-noise of the image by a factor of 6.

Figs 8D,G show the in-plane substrate deformations measured by PIV-UQ for these two datasets using PIV interrogation windows of length $W_L = 128$ pixels, $\boldsymbol{u}_{\mathrm{PIV},128}(x,y)$. These vector maps were overlaid on their associated uncertainty maps, $\sigma_{\boldsymbol{u},\mathrm{PIV},128}(x,y)$. Consistent with the synthetic analyses presented in §3.1, the PIV uncertainty was markedly higher in the "raw" image. Also consistent with the synthetic analyses, regions with deformation gradients such as the edges and corners of the islands, also exhibited higher PIV uncertainty. To determine the number of bootstrap iterations required, coefficient of variation (CoV) can be used to determine convergence. Fig SI.3 shows that the convergence of PIV-UQ is dependent on the resolution as well as the inherent uncertainty. For $W_L = 128$, "raw" images show 10% variability in the bootstrap estimate with n_B 48 iterations whereas "BGS" image only require 7 images.

Figs 8E,H show the traction stress maps recovered from the "raw" and background subtracted ("BGS") images for $W_L = 128$ pixels. The traction stresses recovered from the "raw" images were markedly weaker and their spatial patterns were more spread out as compared to those obtained from the background subtracted images. The differences were due to the more aggressive regularization applied to the PIV vectors from the "raw" images, which had significantly more uncertainty.

These data illustrate that the effective spatial resolution of the traction stresses recovered by TFM-UQ is not uniquely determined by the PIV resolution. While this is a relatively well known consequence of regularization shared with other TFM methods [21, 50], it is worth noting that TFM-UQ uses an objective criterion to automatically select the level of regularization. Moreover, the traction stress uncertainty, σ_t, can be used to quantify the signal-to-noise ratio in the recovered traction stresses, $SNR_{\boldsymbol{t},128} = |\hat{\mathbf{t}}_{128}|/\sigma_{t,128}$. Figs 8F,I display these $SNR_{\boldsymbol{t},128}$ for "raw" and "BGS" images together with vector fans whose width represents the angular uncertainty in the orientation of the $\boldsymbol{t}$ vector. The figures reveal that the largest variability in the direction of $\boldsymbol{t}$ was observed at

Fig 8. TFM-UQ captures variability associated with microscopy image quality. (A) Brightfield image of C3H/10T1/2 cell cultured on fibronectin micropatterned island of length $75\,\mu$m (Red dashed outline). (B) 'Raw' wide-field fluorescent bead image (beads of size $0.2\,\mu$m). (C) 'BGS' Background subtracted image processed from (B). (D,G,J) $\boldsymbol{u}_{\mathrm{PIV}}$ (arrows) overlaid on PIV-UQ uncertainty field ($\sigma_{u,\mathrm{PIV}}$) corresponding to 'Raw' WL-128 (D), 'BGS' WL-128 (G) and 'BGS' WL-64 (J) respectively. Here, WL denotes the PIV interrogation window size W_L. (E,H,K) Mean marginal posterior traction stress field ($\hat{\mathbf{t}}$) corresponding to 'Raw' WL-128 (E), 'BGS' WL-128 (H) and 'BGS' WL-64 (K) respectively. (F,I,L) Traction stress signal-to-noise ratio ($SNR_{\boldsymbol{t}}$) plotted as heatmap corresponding to 'Raw' WL-128 (F), 'BGS' WL-128 (I) and 'BGS' WL-64 (L) respectively. Uncertainty arrows denote the pointwise angular uncertainty corresponding to 1 circular std. dev. of marginal posterior $p(\boldsymbol{t}|\boldsymbol{u}_{\mathrm{h}})$. Scale bar : $25\,\mu$m.

locations of low traction stress magnitude. Moreover, and as expected, these data indicate that $_{519}$ the $SNR_{t,128}$ obtained from the "raw" images was significantly worse than that obtained from the $_{520}$ background subtracted ones with higher contrast. $_{521}$

The results presented above verify that the SNR_t maps obtained by TFM-UQ effectively $_{522}$ capture how image quality affects traction stress recovery. A practical application of SNR_t is $_{523}$ to evaluate how varying W_L affects this recovery. To illustrate this idea, we re-ran PIV-UQ and $_{524}$ TFM-UQ for the "BGS" images at higher resolution, i.e., using $W_L = 64$ pixels, and asked if this $_{525}$ change produced a more precise traction stress map. The answer to this question is not trivial $_{526}$ and depends on whether the gain in PIV resolution (Fig 8J), which is accompanied by an increase $_{527}$ in peak traction stresses at cell-substrate attachments (Fig 8K), compensates the increase in the $_{528}$ PIV measurement error and its propagation to the traction stresses. In this case, comparing Figs $_{529}$ 8I,L indicated that $SNR_{t,64} > SNR_{t,128}$ and, therefore, that refining the PIV window size led to a $_{530}$ more precise representation of the traction stresses. Repeating the analysis for $W_L = 32$ pixels (Fig $_{531}$??) did not yield a higher $SNR_{t,32}$ everywhere, compared to $SNR_{t,64}$, shwoing a notable decrease $_{532}$ in the top-right corner. Likewise, the traction stresses recovered for $W_L = 32$ did not display a $_{533}$ significantly higher spatial resolution than those obtained for $W_L = 32$ as the regularized was more $_{534}$ aggressive for the $W_L = 32$ dataset. This example illustrates how TFM-UQ can be applied to $_{535}$ objectively choose the optimal PIV window size for each experiment, which in this case would be $_{536}$ $W_L = 64$ pixels. $_{537}$

A second set of experiments was performed with human umbilical vascular endothelial cells $_{538}$ (HUVECs) seeded on fibronectin-coated PAA gels to form confluent monolayers (Fig 9A). To $_{539}$ illustrate how TFM-UQ handles images with distinct areas of poor quality, we selected an example $_{540}$ containing a large clump of fluorescent beads and an out-of-focus region, indicated by arrows $_{541}$ in Figure 9B. Figure 9C displays the substrate deformation map obtained by PIV-UQ for this $_{542}$ experiment, showing significant spatial gradients consonant with previously reported data from $_{543}$ HUVEC monolayers [31]. The associated maps of PIV uncertainty in vector magnitude and direction, $_{544}$ shown in Figure 9D, indicate that the PIV errors were mostly localized to regions of poor image $_{545}$ quality and spatial gradients of $\boldsymbol{u}_{\mathrm{PIV}}$, in agreement with the results presented in §3.1. In the bead $_{546}$ clump, PIV-UQ failed to provide deformation vectors leading to the empty white areas in Figs 9C-D. $_{547}$ To feed the TFM-UQ algorithm in this empty region, the displacement vectors were prescribed $_{548}$ using a median filter and $\boldsymbol{\Sigma}_{\mathrm{PIV}}$ was set to be equal to the maximum value from the rest of the $_{549}$ image. TFM-UQ used the $\boldsymbol{u}_{\mathrm{PIV}}$ and $\boldsymbol{\Sigma}_{\mathrm{PIV}}$ fields to recover and regularize the traction stresses (9E), $_{550}$ and mapped the uncertainty in the magnitude and direction of the stress vectors (9F). It is worth $_{551}$ noting that the areas of highest σ_t co-localized with the bead clump and the out-of-focus region $_{552}$ in the fluorescent bead image, as well as with the largest deformation gradients. These results $_{553}$ illustrate how TFM-UQ integrates information from PIV measurements and image quality to map $_{554}$ the confidence of the recovered traction stresses. $_{555}$

Fig 9. TFM-UQ applied to endothelial monolayer experiment demonstrate uncertainty propagation. (A) Membrane labeling of HUVEC monolayers. (B) Fluorescent bead image (beads of size $0.2 \, \mu m$). Arrows indicate bead-related image artifacts (C) PIV-UQ displacement field $\boldsymbol{u}_\mathrm{h}$ (D) $\sigma_{u,PIV}$ map. White regions indicate 'bad' PIV windows that were deleted and replaced as described in the main text. Uncertainty arrows denote the pointwise angular uncertainty corresponding to 1 circular std. dev. of bootstrapped PIV-UQ samples. (E) Inferred mean marginal posterior traction stress, $\hat{\boldsymbol{t}}$ (F) Marginal posterior traction stress uncertainty field (σ_t). Uncertainty arrows denote the pointwise angular uncertainty corresponding to 1 circular std. dev. of marginal posterior $p(\boldsymbol{t}|\boldsymbol{u}_\mathrm{h})$. Scale bar : $25 \, \mu m$

4 Discussion

Since the pioneering silicone membrane wrinkling assay [5] demonstrated the possibility of detecting cell-generated traction stresses, various strategies have been exploited to quantify these stresses. Currently, a common approach involves measuring substrate deformation by tracking fiduciary markers (e.g., fluorescent beads) embedded in elastic substrates of known mechanical properties. Then, the traction stresses corresponding to the measured deformation field are recovered by solving an inverse elastostatic problem. The shortage of tools to quantify how measurement uncertainty propagates to the recovered stresses hampers the objective evaluation of experimental protocols and elastostatic inversion methods. These limitations make it impossible to separate biological variability from measurement error, and promote the view that TFM technique is susceptible to methodological arbitrariness [49, 50, 63].

4.1 Uncertainty Quantification in TFM

One of the most common methods to measure the cell-induced substrate deformations for TFM is particle image velocimetry (PIV), a technique originally developed to measure velocity vectors in fluid flows [51, 64]. PIV measurement errors are associated with the signal-to-noise ratio (SNR) of the image and spatial deformation gradients. In a typical TFM experiment, the image SNR can be optimized by choosing fluorescent beads of the appropriate size, and controlling their density and spatial distribution during substrate fabrication. Image quality can also be tuned by adjusting substrate thickness, the optics and operational settings of the microscope used for acquisition, and post-processing of the acquired images [21, 28, 63, 65]. Batch-to-batch variability in reagents and operator dependence in fabrication and imaging protocols can cause image quality variations within and across experiments. In addition, differences in equipment, experimental, and analysis protocols among laboratories can create experimental uncertainty in TFM. Ultimately, PIV uncertainty is impossible to control *a priori* since it depends on the deformation gradient generated by the cells on the substrate. Therefore, it is necessary to develop tools to estimate PIV uncertainty in individual TFM experiments and propagate this uncertainty to the recovered traction stresses.

The fluid mechanics community has devoted extensive research to quantifying PIV uncertainty [44–46], but significant differences between experimental images generated in elastic substrates and fluid flows impede the transfer of that knowledge to TFM. Typically, TFM images have limited bead brightness due to concerns over phototoxic effects. The substrate deformations are relatively small and experience gradients over micrometric lengthscales. Moreover, ensemble correlation approaches do not improve PIV resolution in TFM since the bead pattern does not vary over time unlike in flowing fluids. In addition, TFM experiments can last hours and are more sensitive to drift and positioning errors along the optical axis. Consequently, image cross-correlation peaks are wider and more irregularly shaped in TFM than in flow experiments. This feature makes the parametric methods widely used to estimate uncertainty in fluid flow less well suited for TFM.

This work introduced a non-parametric bootstrap method to estimate PIV uncertainty in TFM experimental images (PIV-UQ). In PIV-UQ, the images are randomly perturbed multiple times to produce a distribution of measured substrate displacement vectors for each interrogation window, from which the most likely displacement and its standard deviation are calculated. This approach generalizes Kybic's [48] bootstrap method by introducing perturbations directly in the image data instead of in the optimization that calculates the deformation. Therefore, the bootstrap method derived for PIV-UQ in this work could be extended to other deformation measurement techniques like image registration. Apart from TFM experiments, PIV is routinely used in biomechanics and mechanobiology research to estimate *in vitro* and *in vivo* tissue flow [2, 3, 66], and cardiomyocyte contractility assays [28, 67]. The PIV-UQ tools developed here should be directly transferable to those applications.

Bayesian methods model the quantities of interest as random variables, assimilating information regarding physical models, input data, prior constraints, and their variability. Thus, these methods

offer a natural framework to propagate substrate deformation uncertainty to the traction stresses recovered by TFM. Bayesian methods were proposed in Dembo and Wang's seminal TFM work [7], and the interest in adapting them to TFM is recently increasing [19, 39, 68]. Traditionally, Bayesian TFM workflows have focused on elastostatic inversion without explicitly considering the uncertainty in substrate deformation or propagating this uncertainty. Interestingly, Butler *et al*'s pioneering constrained FTTC method [8] considered the possibility of nonphysical displacement vectors and involved an iterative algorithm to correct these measurements, so that the traction stresses satisfied a prior constraint. A more detailed Bayesian TFM model including a cell area as prior constraint has been developed recently [68], however, both of these methods do not propagate measurement errors from the images or the displacement field. To the best of our knowledge, the only effort to propagate uncertainty from the images to traction stresses is Boquet-Pujadas *et al*'s optical flow method for substrate deformation [39], where traction stresses are modeled in the image registration process. This approach considers the traction stresses to be the only cause for deviations in pixel intensities in the image set I. While the spatial variability of error is incorporated in the framework, the work was mostly theoretical and did not consider experimental data. In particular, it would be interesting to assess how the framework handles image error sources that are not explicitly modeled (e.g., fluorescent bead artifacts, camera focus variation, etc).

In this work, we developed TFM-UQ, a hierarchical Bayesian framework that propagates the PIV measurement variabiltiy estimated by PIV-UQ and provides point-by-point estimations of traction stress uncertainty. The resulting uncertainty maps, $\sigma_t(x, y)$, establish error bars for the traction stress recovered at each location, both in magnitude and direction of the stress vector. This information can be helpful in interpreting and analyzing TFM experiments for multiple reasons. Among these, it may help identify outliers automatically and objectively. It allows for separating biological and measurement uncertainties when pooling data from repeat experiments. It can also help pinpoint sources of measurement error and refine experimental protocols and image analysis methods.

Applying PIV-UQ to fluorescent bead images from TFM experiments, we demonstrated this method quantifies how the PIV error depends on the image quality and the size W_L of the interrogation window. The latter is a key PIV parameter affecting measurement error and resolution. TFM protocols often provide loose guidance suggesting W_L should be small enough to resolve the relevant spatial features of cell-generated deformations, and large enough to minimize PIV error [28, 55]. The scarcity of UQ tools in TFM has hampered the development of more quantitative criteria. The PIV-UQ and TFM-UQ tools presented in this manuscript can be used independently and/or together to objectively choose the optimal PIV window size for each experiment. Using PIV-UQ alone without TFM-UQ, one could analyze one or a few fluorescent bead images selected from a batch of frames (e.g., frame of maximum contraction) in a time-lapse experiment to adjust W_L, so that the PIV measurement error remains within expectations and is conserved across experiments. Using PIV-UQ and TFM-UQ together, it is possible to identify the value of W_L that maximizes the resolution and precision of the recovered traction stresses for each experiment.

4.2 Uncertainty-Aware Elastostatic Inversion and Regularization

A variety of mathematical models have been used to represent the elastostatic response of TFM
substrates, and an even wider variety of techniques has been proposed to invert the model equations,
including boundary element methods [7, 19, 21, 38], Fourier expansions providing exact analytical
solutions in frequency space [8, 9, 21, 54, 55], finite element methods [69], Bayesian methods [19,
29, 39, 41, 68], etc. Owing to the physics of elastostatics, this inversion amplifies the small-scale
(high-frequency) features of the measured deformation field. Because measurement noise typically
has a high content of small-scale features, TFM outputs can suffer from significant noise even when
using exact analytical solutions for inversion. In addition, discretized numerical approximations of
the elastostatic response can lead to singular or near-singular linear systems of equations, especially
when the domain size and spatial resolution increase. Therefore, regularization is a crucial aspect
of image analysis in TFM.

Regularization and data smoothing are routinely employed to enforce well-behaved traction
stress fields and suppress amplification of high-frequency noise during the inversion. The specifics
of the regularization algorithm can vary among TFM formulations, but in general, all regularization
schemes introduce a penalty term whose residual must be minimized together with the elastostatic
residual. Regularization methods generally require specifying the values of the hyper-parameters
controlling each residual's weight in the global minimization process. Heuristic approaches to
hyperparameter selection, such as the L-curve criterion [43, 61], have enjoyed popularity, but they
are susceptible to significant user bias and are often difficult to interpret objectively [50, 61, 63].
Self-consistency [36, 39, 63] is an appealing principle to find regularization parameters, but it requires
a priori knowledge of $\mathbf{\Sigma}_{\mathrm{PIV}}$, which was not previously available. The PIV-UQ technique presented
in this manuscript paves the way for a methodical application of this principle.

The interest in Bayesian methods to select hyperparameters has increased in the last decade
[19, 29, 39, 41]. These methods formulate hyperparameter inference as the maximization of the
posterior joint probability $p(\lambda, \boldsymbol{t}|\boldsymbol{u}_{\mathrm{h}})$ or the evidence $p(\lambda|\boldsymbol{u}_{\mathrm{h}})$, where λ is model hyperparameter(s)
and $\boldsymbol{u}_{\mathrm{h}}$ the measured substrate deformation. While this unified framework is attractive and has
built-in self-consistency, the existing works pay little attention to the uncertainty in PIV-measured
substrate deformation or the variability in hyperparameter selection. There is an abundance of
methods (e.g., low-pass filtering, median filtering, etc) to "pre-regularize" the $\boldsymbol{u}_{\mathrm{h}}$ obtained from
PIV or equivalent technique prior to traction stress recovery. However, once the measurements are
pre-regularized, the information about which PIV vectors are less reliable than others is not passed
on to the TFM algorithm. Consequently, existing Bayesian methods make strong assumptions
about the nature of the noise distribution, i.e., that it is spatially uniform and independent of the
effects of cellular forces [19, 21]. These assumptions have not been verified systematically.

To overcome the drawbacks of traditional TFM regularization schemes and the lack of posterior
uncertainty quantification, we developed TFM-UQ – a hierarchical Bayesian framework with
Markov Chain Monte Carlo (MCMC) sampling and non-informative hyperprior distributions

for the hyperparameters. TFM-UQ accounts for the measured PIV uncertainty when enforcing prior constraints, such as the smoothness of the traction stress field, thereby adjusting the level of regularization to the spatial distribution of the measurement errors. We showed that this approach locally balances overfitting and smoothing, and can outperform traditional regularization methods especially in experiments where the quality of the fluorescent bead images or the measured displacements exhibit appreciable gradients.

Uncertainty propagation in TFM-UQ considers the amplification of high-frequency noise in the substrate displacements and the effect of the regularizing prior(s) defined to suppress that noise. As a consequence of the interplay between these factors, the bounds of σ_t are respectively determined by the PIV uncertainty and the strength of the regularizer. This interdependence reflects the fact that regions with higher measurement uncertainty are regularized more heavily, so the prior information of the regularizer overrides the PIV contribution in those regions. The analysis also showed that posterior traction precision is higher for softer gels in the long wavelength (pure shear) limit. Future studies should address the interactions between heteroskedastic measurements and Bayesian priors in the context of TFM as well as the effect of substrate material parameters on the uncertainty propagation. In particular, the possibility of introducing non-global smoothness priors (shrinkage or heteroskedastic priors) deserve further investigation. TFM-UQ can run in a personal computer in few minutes depending on the resolution. Further investigations using approximate inference methods are worthwhile to make the calculations near real-time for high throughtput applications.

5   Conclusion

This manuscript presents a traction force microscopy technique that can estimate the measurement uncertainty associated with image quality and cell-generated forces, incorporating this variability in the regularization of the recovered traction stresses. A noteworthy contribution of this work is the investigation of error bounds for the magnitude and direction of the traction stress vector measured at each point in space and time of an experiment. This feature facilitates the standardization of TFM protocols, the comparison between experiments, and the separation of measurement variability from biological variability, all of which are challenging in traditional TFM methods. We anticipate these new capabilities will pave the way for more robust, high-throughput applications of TFM that bypass user intervention.

Acknowledgments

This work was supported by the National Institutes of Health through grants 1R01HD109149-01, 1R01AI167943-01A1, and 1R01HL170607-01A1. We thank Dr.Alejandro Gonzalo for providing feedback on the manuscript and Dr.Kelly Stevens for providing C3H/10T1/2 fibroblast cell line.

References

1. Kumar A, Placone JK, Engler AJ. Understanding the extracellular forces that determine cell 717
fate and maintenance. 2017;doi:10.1242/dev.158469. 718

2. Smutny M, Ákos Z, Grigolon S, Shamipour S, Ruprecht V, Čapek D, et al. Friction forces 719
position the neural anlage. Nature Cell Biology. 2017;19:306–317. doi:10.1038/ncb3492. 720

3. Ladoux B, Mège RM. Mechanobiology of collective cell behaviours. Nature Reviews Molecular 721
Cell Biology. 2017;18:743–757. doi:10.1038/nrm.2017.98. 722

4. Ehler E, Perriard JC. Cardiomyocyte Cytoskeleton and Myofibrillogenesis in Healthy and Dis- 723
eased Heart. Heart Failure Reviews 2000 5:3. 2000;5:259–269. doi:10.1023/A:1009861504264. 724

5. Harris AK, Wild P, Stopak D. Silicone rubber substrata: a new wrinkle in the study of cell 725
locomotion. Science. 1980;208(4440):177–179. 726

6. Eyckmans J, Boudou T, Yu X, Chen CS. A hitchhiker's guide to mechanobiology. Develop- 727
mental cell. 2011;21(1):35–47. 728

7. Dembo M, Wang YL. Stresses at the cell-to-substrate interface during locomotion of fibroblasts. 729
Biophysical Journal. 1999;76:2307. doi:10.1016/S0006-3495(99)77386-8. 730

8. Butler JP, Toli-Nørrelykke IM, Fabry B, Fredberg JJ. Traction 731
fields, moments, and strain energy that cells exert on their surround- 732
ings. American Journal of Physiology - Cell Physiology. 2002;282:595–605. 733
doi:10.1152/AJPCELL.00270.2001/ASSET/IMAGES/LARGE/H00320871106.JPEG. 734

9. del Alamo JC, Meili R, Alonso-Latorre B, Rodríguez-Rodríguez J, Aliseda A, Firtel RA, 735
et al. Spatio-temporal analysis of eukaryotic cell motility by improved force cytometry. 2007;. 736

10. Koch TM, Münster S, Bonakdar N, Butler JP, Fabry B. 3D traction forces in cancer cell 737
invasion. PLoS ONE. 2012;7:e33476–e33476. doi:10.1371/journal.pone.0033476. 738

11. Steinwachs J, Metzner C, Skodzek K, Lang N, Thievessen I, Mark C, et al. Three-dimensional 739
force microscopy of cells in biopolymer networks. Nature Methods 2015 13:2. 2015;13:171–176. 740
doi:10.1038/nmeth.3685. 741

12. Grashoff C, Hoffman BD, Brenner MD, Zhou R, Parsons M, Yang MT, et al. Measuring 742
mechanical tension across vinculin reveals regulation of focal adhesion dynamics. Nature 743
2010 466:7303. 2010;466:263–266. doi:10.1038/nature09198. 744

13. Blakely BL, Dumelin CE, Trappmann B, McGregor LM, Choi CK, Anthony PC, et al. A 745
DNA-based molecular probe for optically reporting cellular traction forces. Nature Methods 746
2014 11:12. 2014;11:1229–1232. doi:10.1038/nmeth.3145. 747

14. Brockman JM, Blanchard AT, Pui-Yan V, Derricotte WD, Zhang Y, Fay ME, et al. Mapping the 3D orientation of piconewton integrin traction forces. Nature Methods 2017 15:2. 2017;15:115–118. doi:10.1038/nmeth.4536.

15. Hazlett L, Landauer AK, Patel M, Witt HA, Yang J, Reichner JS, et al. Epifluorescence-based three-dimensional traction force microscopy. Scientific Reports 2021 10:1. 2020;10:1–12. doi:10.1038/s41598-020-72931-6.

16. Barrasa-Fano J, Shapeti A, Álvaro Jorge-Peñas, Barzegari M, Sanz-Herrera JA, Ooster-wyck HV. TFMLAB: A MATLAB toolbox for 4D traction force microscopy. SoftwareX. 2021;15:100723. doi:10.1016/J.SOFTX.2021.100723.

17. Martiel JL, Leal A, Kurzawa L, Balland M, Wang I, Vignaud T, et al. Measurement of cell traction forces with ImageJ. Methods in Cell Biology. 2015;125:269–287. doi:10.1016/BS.MCB.2014.10.008.

18. Bauer A, Prechová M, Fischer L, Thievessen I, Gregor M, Fabry B. pyTFM: A tool for traction force and monolayer stress microscopy. PLOS Computational Biology. 2021;17:e1008364. doi:10.1371/JOURNAL.PCBI.1008364.

19. Huang Y, Schell C, Huber TB, Şimşek AN, Hersch N, Merkel R, et al. Traction force microscopy with optimized regularization and automated Bayesian parameter selection for comparing cells. Scientific Reports 2019 9:1. 2019;9:1–16. doi:10.1038/s41598-018-36896-x.

20. Smith LA, Aranda-Espinoza H, Haun JB, Dembo M, Hammer DA. Neutrophil traction stresses are concentrated in the uropod during migration. Biophysical Journal. 2007;92. doi:10.1529/biophysj.106.102822.

21. Sabass B, Gardel ML, Waterman CM, Schwarz US. High Resolution Traction Force Microscopy Based on Experimental and Computational Advances. Biophysical Journal. 2008;94:207–220. doi:10.1529/BIOPHYSJ.107.113670.

22. Elosegui-Artola A, Oria R, Chen Y, Kosmalska A, Pérez-González C, Castro N, et al. Mechanical regulation of a molecular clutch defines force transmission and transduction in response to matrix rigidity. Nature Cell Biology 2016 18:5. 2016;18:540–548. doi:10.1038/ncb3336.

23. Brugués A, Anon E, Conte V, Veldhuis JH, Gupta M, Colombelli J, et al. Forces driving epithelial wound healing. Nature Physics 2014 10:9. 2014;10:683–690. doi:10.1038/nphys3040.

24. Trepat X, Wasserman MR, Angelini TE, Millet E, Weitz DA, Butler JP, et al. Physical forces during collective cell migration. Nature Physics 2009 5:6. 2009;5:426–430. doi:10.1038/nphys1269.

25. Park JA, Kim JH, Bi D, Mitchel JA, Qazvini NT, Tantisira K, et al. Unjamming and cell shape in the asthmatic airway epithelium. Nature Materials 2014 14:10. 2015;14:1040–1048. doi:10.1038/nmat4357.

26. Serrano R, Aung A, Yeh YT, Varghese S, Lasheras JC, Lamo JCDA Three-Dimensional Monolayer Stress Microscopy. 2019;doi:10.1016/j.bpj.2019.03.041.

27. Ribeiro AJS, Ang YS, Fu JD, Rivas RN, Mohamed TMA, Higgs GC, et al. Contractility of Single cardiomyocytes differentiated from pluripotent stem cells depends on physiological shape and substrate stiffness. Proceedings of the National Academy of Sciences of the United States of America. 2015;112:12705–12710. doi:10.1073/PNAS.1508073112/SUPPL$_F ILE/PNAS$.1508073112.SM12.MOV.

28. Sharma A, Burridge PW, McKeithan WL, Serrano R, Shukla P, Sayed N, et al. High-throughput screening of tyrosine kinase inhibitor cardiotoxicity with human induced pluripotent stem cells. Science Translational Medicine. 2017;9. doi:10.1126/SCITRANSLMED.AAF2584.

29. Mollica MY, Beussman KM, Kandasamy A, Rodríguez LM, Morales FR, Chen J, et al. Distinct platelet F-actin patterns and traction forces on von Willebrand factor versus fibrinogen. Biophysical Journal. 2023;122:3738–3748. doi:10.1016/j.bpj.2023.07.006.

30. Hanke J, Probst D, Zemel A, Schwarz US, Köster S. Dynamics of force generation by spreading platelets. Soft Matter. 2018;14:6571–6581. doi:10.1039/C8SM00895G.

31. Yeh YT, Serrano R, François J, Chiu JJ, Li YSJ, del Alamo JC, et al. Three-dimensional forces exerted by leukocytes and vascular endothelial cells dynamically facilitate diapedesis. Proceedings of the National Academy of Sciences of the United States of America. 2018;115:133–138. doi:10.1073/PNAS.1717489115/SUPPL$_F ILE/PNAS$.1717489115.SM01.MOV.

32. François J, Kandasamy A, Yeh YT, Schwartz A, Ayala C, Meili R, et al. The interplay between matrix deformation and the coordination of turning events governs directed neutrophil migration in 3D matrices. Science Advances. 2021;7:3882–3896. doi:10.1126/SCIADV.ABF3882/SUPPL$_F ILE/SCIADV.ABF$3882$_S M.PDF$.

33. Rabodzey A, Alcaide P, Luscinskas FW, Ladoux B. Mechanical Forces Induced by the Transendothelial Migration of Human Neutrophils. Biophysical Journal. 2008;95:1428–1438. doi:10.1529/BIOPHYSJ.107.119156.

34. Bastounis EE, Yeh YT, Theriot JA. Matrix stiffness modulates infection of endothelial cells by Listeria monocytogenes via expression of cell surface vimentin. Molecular Biology of the Cell. 2018;29:1571–1589. doi:10.1091/MBC.E18-04-0228.

35. Dembo M, Oliver T, Ishihara A, Jacobson K. Imaging the traction stresses exerted by locomoting cells with the elastic substratum method. Biophysical Journal. 1996;70:2008–2022. doi:10.1016/S0006-3495(96)79767-9.

36. Schwarz US, Balaban NQ, Riveline D, Bershadsky A, Geiger B, Safran SA. Calculation of Forces at Focal Adhesions from Elastic Substrate Data: The Effect of Localized Force and the Need for Regularization. Biophysical Journal. 2002;83:1380–1394. doi:10.1016/S0006-3495(02)73909-X.

37. Gómez-González M, Latorre E, Arroyo M, Trepat X. Measuring mechanical stress in living tissues. Nature Reviews Physics 2020 2:6. 2020;2:300–317. doi:10.1038/S42254-020-0184-6.

38. Han SJ, Oak Y, Groisman A, Danuser G. Traction microscopy to identify force modulation in subresolution adhesions. Nature Methods 2015 12:7. 2015;12:653–656. doi:10.1038/nmeth.3430.

39. Boquet-Pujadas A, Olivo-Marin JC. Reformulating Optical Flow to Solve Image-Based Inverse Problems and Quantify Uncertainty. IEEE Transactions on Pattern Analysis and Machine Intelligence. 2023;45:6125–6141. doi:10.1109/TPAMI.2022.3202855.

40. Plotnikov SV, Sabass B, Schwarz US, Waterman CM. High-Resolution Traction Force Microscopy. Methods in Cell Biology. 2014;123:367–394. doi:10.1016/B978-0-12-420138-5.00020-3.

41. Nier V, Jain S, Lim CT, Ishihara S, Ladoux B, Marcq P. Inference of Internal Stress in a Cell Monolayer. Biophysical Journal. 2016;110:1625–1635. doi:10.1016/J.BPJ.2016.03.002.

42. Blumberg JW, Schwarz US. Comparison of direct and inverse methods for 2.5D traction force microscopy. PLOS ONE. 2022;17:e0262773. doi:10.1371/JOURNAL.PONE.0262773.

43. Hansen PC. Regularization tools version 4.0 for Matlab 7.3. Numerical algorithms. 2007;46:189–194.

44. Wieneke B. PIV uncertainty quantification from correlation statistics. Measurement Science and Technology. 2015;26:074002. doi:10.1088/0957-0233/26/7/074002.

45. Sciacchitano A. Uncertainty quantification in particle image velocimetry. Measurement Science and Technology. 2019;30:092001. doi:10.1088/1361-6501/AB1DB8.

46. Bhattacharya S, Charonko JJ, Vlachos PP. Particle image velocimetry (PIV) uncertainty quantification using moment of correlation (MC) plane. Measurement Science and Technology. 2018;29:115301. doi:10.1088/1361-6501/AADFB4.

47. Kybic J. Fast no ground truth image registration accuracy evaluation: Comparison of bootstrap and Hessian approaches. 2008 5th IEEE International Symposium on Biomedical Imaging: From Nano to Macro, Proceedings, ISBI. 2008; p. 792–795. doi:10.1109/ISBI.2008.4541115.

48. Kybic J. Bootstrap resampling for image registration uncertainty estimation without ground truth. IEEE Transactions on Image Processing. 2010;19:64–73. doi:10.1109/TIP.2009.2030955.

49. Polacheck WJ, Chen CS. Measuring cell-generated forces: A guide to the available tools; 2016.

50. Zancla A, Mozetic P, Orsini M, Forte G, Rainer A. A primer to traction force microscopy. Journal of Biological Chemistry. 2022;298(5).

51. Adrian RJ, Westerweel J. Particle image velocimetry. 30. Cambridge university press; 2011.

52. Ester M, Kriegel HP, Sander J, Xu X, et al. A density-based algorithm for discovering clusters in large spatial databases with noise. In: kdd. vol. 96; 1996. p. 226–231.

53. Xue Z, Charonko JJ, Vlachos PP. Particle image velocimetry correlation signal-to-noise ratio metrics and measurement uncertainty quantification. Measurement Science and Technology. 2014;25(11):115301.

54. Lin YC, Tambe DT, Park CY, Wasserman MR, Trepat X, Krishnan R, et al. Mechanosensing of substrate thickness. Physical review E, Statistical, nonlinear, and soft matter physics. 2010;82. doi:10.1103/PHYSREVE.82.041918.

55. del Alamo JC, Meili R, Álvarez González B, Alonso-Latorre B, Bastounis E, Firtel R, et al. Three-Dimensional Quantification of Cellular Traction Forces and Mechanosensing of Thin Substrata by Fourier Traction Force Microscopy. PLOS ONE. 2013;8:e69850. doi:10.1371/JOURNAL.PONE.0069850.

56. Molina R, Katsaggelos AK, Mateos J. Bayesian and regularization methods for hyperparameter estimation in image restoration. IEEE Transactions on Image Processing. 1999;8:231–246. doi:10.1109/83.743857.

57. Gelman A, Carlin JB, Stern HS, Rubin DB. Bayesian Data Analysis. Bayesian Data Analysis. 1995;doi:10.1201/9780429258411.

58. Howard M, Luttman A, Fowler M. Sampling-based uncertainty quantification in deconvolution of X-ray radiographs. Journal of Computational and Applied Mathematics. 2014;270:43–51. doi:10.1016/J.CAM.2014.02.024.

59. Vignaud T, Ennomani H, Théry M. Polyacrylamide hydrogel micropatterning. In: Methods in cell biology. vol. 120. Elsevier; 2014. p. 93–116.

60. Tse JR, Engler AJ. Preparation of hydrogel substrates with tunable mechanical properties. Current protocols in cell biology. 2010;47(1):10–16.

61. Hansen PC, O'Leary DP. The use of the L-curve in the regularization of discrete ill-posed problems. SIAM journal on scientific computing. 1993;14(6):1487–1503.

62. Felts L, Kohli V, Marr J, Schumacher J, Schlicker O. An Introduction to Computational Clearing; 2020. `https://www.leica-microsystems.com/science-lab/life-science/an-introduction-to-computational-clearing`.

63. Denisin AK, Kim H, Riedel-Kruse IH, Pruitt BL. Field Guide to Traction Force Microscopy. Cellular and Molecular Bioengineering 2024. 2024; p. 1–20. doi:10.1007/S12195-024-00801-6.

64. Willert CE, Gharib M. Digital particle image velocimetry. Experiments in Fluids. 1991;10:181–193. doi:10.1007/BF00190388/METRICS.

65. Zündel M, Ehret AE, Mazza E. Factors influencing the determination of cell traction forces. PLOS ONE. 2017;12:e0172927. doi:10.1371/JOURNAL.PONE.0172927.

66. Sun Y, Reid B, Zhang Y, Zhu K, Ferreira F, Estrada A, et al. Electric field–guided collective motility initiation of large epidermal cell groups. Molecular biology of the cell. 2023;34(5):ar48.

67. Sharma A, McKeithan WL, Serrano R, Kitani T, Burridge PW, Del Álamo JC, et al. Use of human induced pluripotent stem cell–derived cardiomyocytes to assess drug cardiotoxicity. Nature protocols. 2018;13(12):3018–3041.

68. Fujikawa R, Okimura C, Kozawa S, Ikeda K, Inagaki N, Iwadate Y, et al. Bayesian traction force estimation using cell boundary-dependent force priors. Biophysical Journal. 2023;122:4542–4554. doi:10.1016/J.BPJ.2023.10.032.

69. Colin-York H, Javanmardi Y, Barbieri L, Li D, Korobchevskaya K, Guo Y, et al. Spatiotemporally Super-Resolved Volumetric Traction Force Microscopy. Nano Letters. 2019;19:4427–4434. doi:10.1021/ACS.NANOLETT.9B01196/SUPPL$_F ILE/NL9B01196_S I_0$06.AVI.

Supplementary information

Fig SI.1. PIV-UQ synthetic validation workflow. A fixed, high-resolution ground truth displacement $\mathbf{u}(\mathbf{x})$ is generated. One realization of synthetic images corresponding to the ground truth displacement is obtained as follows : Synthetic beads are simulated by randomly placing Gaussian beads of prescribed density (ρ_b) and maximum brightness of 1000 a.u. $(\boldsymbol{I}_{ref})$. The centroid of the Gaussian beads are shifted according to $\mathbf{u}(\mathbf{x}_{bead})$ to generate $\boldsymbol{I}_t$. PIV is performed to generate displacement vector for each interrogation window, $\hat{\mathbf{u}}$. This process is repeated N times to obtain an ensemble of synthetic image and $\hat{\mathbf{u}}$ realizations that provide the true ensemble standard deviation, $\sigma_{u,ens}$. For each of the i^{th} ensemble image set $\{\boldsymbol{I}_t^{(i)}, \boldsymbol{I}_{ref}^{(i)}\}$, PIV-UQ method is performed by bootstrapping n_B times generating $\sigma_{UQ}^{(i)}$ and subsequently $\sigma_{u,\mathrm{PIV}}^{(i)}$ for each ensemble realization i. The root mean square (RMS) of $\sigma_{u,\mathrm{PIV}}$ is reported over N ensemble realizations. The scatter plot depicts the point-wise uncertainty of true ensemble realizations with that of PIV-UQ estimate for a shear displacement field ($\mathbf{u}(\mathbf{x}) = \mathbf{u_2}$ (13)) showing the ability of PIV-UQ to predict spatial dependence of PIV uncertainty in non-uniform displacement. Overlaid heatmap shows the conditional distribution of $\sigma_{u,\mathrm{PIV}}$ given $\sigma_{u,ens}$.

Fig SI.2. TFM-UQ uncertainty propagation is dependent on the choice of prior. TFM-UQ is applied to synthetic simulation described in Fig 6 with the choice of two different precision matrices : identity $\mathbf{L}' = \mathbf{U}$ (A-B) and heteroskedastic prior matrix $\mathbf{L}' = \mathbf{\Sigma}_{\mathrm{PIV}}^{-1}\mathbf{L}$ (C-D). Here, $\mathbf{L}$ is the discrete Laplacian operator with second-order finite differences. The prior is of the form $p(\boldsymbol{t}|\alpha) \propto \exp\left[\frac{\alpha}{2}\boldsymbol{t}^T\mathbf{L}'\boldsymbol{t}\right]$. (A,C) Mean marginal posterior traction stress magnitude $\hat{\mathbf{t}}$ and (B,D) Marginal posterior traction uncertainty (σ_t). The spatial ange of traction stress uncertainty is increased with the heteroskedastic prior (D) as opposed to $\mathbf{\Sigma}_{\mathrm{PIV}}^{-1}\mathbf{L}$ as compared to $\mathbf{L}$ (Fig 7). However, the elastostatic term dominates resulting in under-regularization (C).

Fig SI.3. Coefficient of variation as a metric to assess PIV-UQ convergence. Coefficient of variation (CoV) $= \mathrm{Std}[\sigma_{u,\mathrm{PIV}}]/\mathrm{Mean}[\sigma_{u,\mathrm{PIV}}]$. "Raw" images tend to need higher iterations for the error in $\sigma_{u,\mathrm{PIV}}$ estimation to decay to 10% compared to "BGS" images in a resolution dependent manner.

Fig SI.4. C3H/10T1/2 fibroblast cultured on fibronectin micropatterned island analyzed with PIV window size $\mathbf{W_L = 32}$. (A) $\boldsymbol{u}_{\mathrm{PIV}}$ (arrows) overlaid on PIV-UQ uncertainty field ($\sigma_{\boldsymbol{u}_{\mathrm{PIV}}}$) corresponding (B) Mean marginal posterior traction stress field ($\hat{\mathbf{t}}$) (C) Traction stress signal-to-noise ratio ($SNR_{\boldsymbol{t}}$). Uncertainty arrows denote the pointwise angular uncertainty corresponding to 1 circular std. dev. of marginal posterior $p(\boldsymbol{t}|\boldsymbol{u}_{\mathrm{h}})$. Red dashed outline indicate cell contour.