Comparing Raman and NanoSIMS for heavy water labeling of single cells

Running title: Comparing Raman and NanoSIMS

George A. Schaible [1,2], John B. Cliff [3#], Jennifer A. Crandall [2,4], Jeremy J. Bougoure [3], Michael N. Mathuri
[5], Alex L. Sessions [5], Joseph Atwood [6], and Roland Hatzenpichler [1,2,4,7*]

[1] Department of Chemistry and Biochemistry, Montana State University, Bozeman, MT 59717
[2] Center for Biofilm Engineering, Montana State University, Bozeman, MT 59717
[3] Environmental Molecular Sciences Laboratory, Pacific Northwest National Laboratory, Richland, WA
99354
[4] Department of Microbiology and Cell Biology, Montana State University, Bozeman, MT 59717
[5] Division of Geological and Planetary Sciences, California Institute of Technology, Pasadena, CA 91125
[6] Department of Agricultural Economics and Economics, Montana State University, Bozeman, MT 59717
[7] Thermal Biology Institute, Montana State University, Bozeman, MT 59717
[#] Current address: National Security Directorate, Oak Ridge National Laboratory, Oak Ridge, TN

* Correspondence to rolandhatzenpichler@gmail.com. phone: +1-406-994-5469. 111 Chemistry and
Biochemistry Building. Montana State University, Bozeman, MT 59717. USA

ORCID: GAS, 0000-0002-1031-4682. JBC, 0000-0002-7395-5604. JAC, 0000-0002-9051-8481. JJB,
0000-0002-4869-035X. MNM, 0000-0003-0652-7789. ALS, 0000-0001-6120-2763. RH, 0000-0002-
5489-3444.

Conflict of interest: The authors declare no conflict of interest.

Funding source: NIH 1R35GM147166-01

Keywords: Cellular heterogeneity, D_2O, heavy water, cell activity, single cell microbiology, technique
comparison, Raman, NanoSIMS

<u>**Abstract**</u> **(250 words)**
Stable isotope probing (SIP) experiments in conjunction with Raman microspectroscopy (Raman) or
nano-scale secondary ion mass spectrometry (NanoSIMS) are frequently used to explore single cell
metabolic activity in pure cultures as well as complex microbiomes. Despite the increasing popularity of
these techniques, the comparability of isotope incorporation measurements using both Raman and
NanoSIMS directly on the same cell remains largely unexplored. This knowledge gap creates uncertainty
about the consistency of single-cell SIP data obtained independently from each method. Here, we conducted
a comparative analysis of 543 *Escherichia coli* cells grown in M9 minimal medium in the absence or
presence of heavy water (^2H_2O) using correlative Raman and NanoSIMS measurements to quantify the
results between the two approaches. We demonstrate that Raman and NanoSIMS yield highly comparable
measurements of ^2H incorporation, with varying degrees of similarity based on the mass ratios analyzed
using NanoSIMS. The $^{12}C^2H/^{12}C^1H$ and $^{12}C_2^2H/^{12}C_2^1H$ mass ratios provide targeted measurements of C-H
bonds but may suffer from biases and background interference, while the $^2H/^1H$ ratio captures all hydrogen

with lower detection limits, making it suitable for applications requiring comprehensive ^{2}H quantification.
Importantly, despite its higher mass resolution requirements, the use of $C_2{}^2H/C_2{}^1H$ may be a viable
alternative to using C^2H/C^1H due to lower background and higher overall count rates. Furthermore, using
an empirical approach to determining Raman wavenumber ranges via the 2nd derivative improved the data
equivalency of ^{2}H quantification between Raman and NanoSIMS, highlighting its potential for enhancing
cross-technique comparability. These findings provide a robust framework for leveraging both techniques,
enabling informed experimental design and data interpretation. By enhancing cross-technique
comparability, this work advances SIP methodologies for investigating microbial metabolism and
interactions in diverse systems.
54
Importance
Accurate and reliable measurements of cellular properties are fundamental to understanding the
function and activity of microbes. This study addresses to what extent Raman microspectroscopy and nano-
scale secondary ion mass spectrometry (NanoSIMS) measurements of single cell anabolic activity can be
compared. Here, we study the relationship of the incorporation of a stable isotope (^{2}H through incorporation
of ^2H_2O) as determined by the two techniques and calculate a correlation coefficient to support the use of
either technique when analyzing cells incubated with ^2H_2O. The ability to discern between the comparative
strengths and limitations of these techniques is invaluable in refining experimental protocols, enhancing
data comparability between studies, data interpretation, and ultimately advancing the quality and reliability
of outcomes in microbiome research.
65
Introduction
Microbial activities and cellular interactions drive human health, global biogeochemical cycles, and
ecosystem function. Stable isotope probing (SIP) is a particularly powerful approach to studying microbial
function and is widely used in microbiology (1-5). By measuring the incorporation of heavy isotopes (*e.g.*,
^{2}H, ^{13}C, ^{15}N, ^{18}O) into biomass, it is possible to assess the activity and physiology of individual cells at a
microbially-meaningful spatial resolution, enabling a broader understanding of community dynamics (4).
While SIP can be applied to all assimilatory pathways, heavy water labeling holds appeal for
microbiologists due to its universal applicability, negating the need for prior knowledge of cell physiology
(2). When combined with the addition of unlabeled substrates, heavy water labeling enables researchers to
pinpoint which cells within complex microbiomes are activate in presence of a substrate (1). Active
microbes grown or incubated in the presence of heavy water will use deuterium (^{2}H) in lieu of hydrogen
(^{1}H) during biosynthesis, resulting in active microbial community members becoming labeled with ^{2}H (6,
7). Because ^{2}H naturally occurs at very low levels (~0.016%), its environmental background is usually
negligible for biological labeling studies (8, 9).
The assimilation of isotopically labeled substrates, including heavy water, into individual cells can be
measured by either Raman microspectroscopy (Raman) or nano-scale secondary ion mass spectrometry
(NanoSIMS). Micro-autoradiography is a cost-effective alternative but is limited to radioactive isotopes
(10, 11). All three techniques can be combined with fluorescence *in situ* hybridization (FISH), which allows
a direct link between cellular identity and function in complex microbiomes (3, 12-14).
Raman is a non-destructive spectroscopic technique based on the inelastic scattering of light (3, 15). It
can be applied directly to environmental samples with minimal preparation (16), allowing for real-time
monitoring of metabolic changes of single cells (17). By exciting molecular bond vibrations, Raman
elucidates the spatial distribution of specific biomolecules, such as lipids, proteins, and nucleic acids within

cells and/or communities on a sub-micrometer scale (3, 16). Incorporation of stable isotopes into
biomolecules changes the vibrational energy of excited molecules through the increased molecular mass,
inducing a red shift towards smaller wavenumbers (16, 18). While Raman requires little or no sample
preparation, the technique suffers from low sensitivity (~10% ^{13}C, ~10% ^{15}N, and ~0.2% ^{2}H cellular
replacement (16, 18, 19)) and relatively low signal-to-noise ratio as compared to mass spectrometry.
However, carbon-deuterium (^{12}C^{2}H) bonds are reliably detected in Raman spectra by a characteristic peak
shift of the abundant ^{12}C^{1}H peak into the mostly Raman-silent region (1,800-2,800 cm^{-1}) of the cellular
spectrum making Raman analysis well suited for deuterium SIP analysis in single cells (2).

NanoSIMS uses a focused ion beam to produce secondary ions from a sample surface that are
subsequently analyzed via a double focusing mass spectrometer to attain spatially resolved elemental and
isotopic information (20, 21). Although NanoSIMS is capable of detecting isotopes at a much greater
sensitivity and spatial resolution (≥50 nm) than Raman, it is destructive to the sample and provides limited
molecular information (21). Furthermore, unlike Raman, the molecular pools of H that the NanoSIMS
interrogates is unclear. NanoSIMS can suffer from recombination of atoms during ion beam sputtering,
which may result in misinterpretation of results. However, in the case of ^{13}C- and ^{15}N-labeling of biological
materials, recombination is not thought to affect the general interpretation of NanoSIMS results (22). At
the whole-cell level, this assumption likely holds for ^{2}H as well. However, as this study highlights,
heterogeneous sampling of cellular components and the choice of ion pair complicate detailed data
interpretation with our current understanding.

Both Raman and NanoSIMS are powerful tools for studying the structure, function, and interactions of
microbes on a single-cell or community level (3, 21). Studies have successfully applied both techniques in
tandem, including analysis of deposition of organic material in diatom fossils (23), SIP experiments of
microbial cells recovered from soils and sediments (24, 25), and *in vitro* SIP of microbial cells for sorting
(2). However, to the best of our knowledge, only two studies ever applied Raman and NanoSIMS on the
same cells (25, 26), and a comprehensive and direct comparison of these techniques was previously
unavailable, making it difficult to compare and interpret results generated by these techniques.

To address this shortcoming, we employed SIP on *Escherichia coli* using deuterium oxide (^{2}H$_2$O) and
compared results of Raman and NanoSIMS based measurements of 543 individual cells. We demonstrate
strong correlation of these measurements, enabling the corroboration of results from both techniques. This
data set also enabled an investigation into the pre-processing of Raman data providing insights on how to
best optimize the methodology when using either technique for ^{2}H incorporation measurements. In addition,
we investigated a statistical approach for calculating wavenumber ranges for ^{12}C^{2}H measurements using
Raman.

Considering that both Raman and NanoSIMS possess unique strengths and limitations, the drawbacks
of one technique can be compensated for by the other, thereby granting a more comprehensive insight into
biological samples at a micrometer scale (25). Ultimately, this comparative investigation contributes to
refining our understanding of Raman and NanoSIMS data, lays the foundation for future integrated
approaches of correlative SIP-Raman-NanoSIMS analyses, and illuminates the strengths and weaknesses
of each technique to allow future investigators more robust information when evaluating the applicability
of both techniques for a given experimental objective.

129
130   **Results and Discussion**
131   **Stable isotope probing**

We employed our recently developed correlative workflow (25) for the study of single *E. coli* K12 cells that had been grown in unamended media (0% added ^2H_2O) or media amended with either 15%, 30%, or 50% ^2H_2O. Lower ^2H_2O amounts (between 0% and 15%) were not tested due to the limited sensitivity of Raman spectroscopy for detecting ^2H_2O uptake at low enrichment levels (25). In such cases, NanoSIMS is the preferred method, because it offers much higher sensitivity and thus enables detection of low levels of anabolic activity (27). Cells were collected during mid-logarithmic growth (Fig. 1 and S1), fixed with paraformaldehyde (PFA) to preserve cells and inhibit change in ^2H content, immobilized on a stainless-steel coupon, stained with 4',6-diamidino-2-phenylindole (DAPI, necessary for visualization during next step), and regions of interest (ROIs) mapped using laser dissection microscopy (Fig. S2). Raman spectra were collected for each individual cell within the defined ROIs before NanoSIMS measurements were performed.

Fig. 1. Representative Raman spectra and NanoSIMS images illustrating the uptake of ^2H-label by *Escherichia coli* K12 cells after incubation in media with varying percentages of ^2H_2O. **A** The characteristic $^{12}C^2H$ and $^{12}C^1H$ regions are shaded in gray. Each Raman spectrum was obtained from a single cell. **B** $^{12}C_2{}^2H/^{12}C_2{}^1H$ mass ratio NanoSIMS images. All isotope fraction images are on the same scale (0-30 atom %). Scales bars are 10 μm.

Validation of natural abundance using IRMS

Discrepancies between techniques for ^{2}H measurements for 0% ^{2}H$_2$O incubations prompted additional analysis to confirm the natural abundance of ^{2}H in these samples. To investigate this, we used isotope ratio mass spectrometry (IRMS) on lipids extracted from the 0% ^{2}H$_2$O incubation samples. The lipids were selected for IRMS analysis because they are reflective of the isotopic signature of cellular material without being influenced by extracellular artifacts. Extracted lipids were esterified to form fatty acid methyl-esters (FAMEs) for analysis, enabling better ionization and separation during IRMS (28). The analysis showed that lipids from the 0% ^{2}H$_2$O incubations were at natural ^{2}H abundance as compared to the VSMOW international standard (Table S1). This finding confirmed that the elevated ^{2}H levels observed in Raman and NanoSIMS 0% ^{2}H$_2$O data were due to conflated values rather than true biological enrichment. Furthermore, previous studies comparing bulk elemental analyzer IRMS measurements to SIMS have shown that IRMS values are generally higher than those derived from SIMS (29), further supporting that the 0% ^{2}H$_2$O incubations did not contain any incorporated ^{2}H.

Pre-processing of Raman spectra

Currently, there is no standard method for the pre-processing of Raman spectra (16), despite compelling evidence demonstrating the influence of processing procedures on data outcomes (30). Optimizing the workflow by testing various pre-processing procedures on Raman spectra before selecting a method is recommended to minimize information loss in the spectra (31, 32). To address the lack of standardization in pre-processing methods, we systematically tested various methods on our Raman spectra to optimize our spectra processing methodology. Our findings showed that pre-processing techniques had little effect on the quantification of ^{2}H from *E. coli* spectra.

Raman spectra spanning 250-3,200 cm^{-1} were acquired from individual cells within the ROIs. To test if peaks in the fingerprint region of spectra (ca. 700-1,800 cm^{-1}) induce a leveraging effect during smoothing and baselining of the spectra, spectra were analyzed as a whole (250-3,200 cm^{-1}) or were truncated to 1,800-3,200 cm^{-1}, as was previously suggested (31, 33). All spectra were smoothed using a Savitzky-Golay smoothing algorithm, a widely used technique for spectroscopy data (16, 32). This algorithm applies a moving window with a local polynomial fitting procedure to reduce noise while preserving peak shapes (34). It has been shown that excessive smoothing of spectra can result in the loss of genuine Raman peaks, thereby producing inaccurate spectra (30). To avoid this, we used a 2nd degree polynomial with a window size of 11, as these settings preserved important peaks across all spectra. Next, several commonly applied baselining methods were tested to identify if background correction influences the final calculation of ^{2}H content of the *E. coli* cells.

The first baselining method applied was asymmetric least squares (ALS), an iterative method that uses asymmetric weighting to adjust the extent in which data influence baselining (35). The second method used was a polynomial fit, where a polynomial of nth degree (we used a 2nd degree polynomial) can be fit to both the spectra and the baseline subtracted to remove background (36). The third method tested was a frequency differentiated non-linear digital filter, commonly referred to as "rolling ball", which is equivalent to rolling a ball with a given ball diameter (*i.e.*, window size) below the spectra and removing the background below the points the ball touches (37). The fourth and final method tested was statistics-sensitive non-linear iterative peak-clipping (SNIP), a process that compresses the data using a log square root operator followed by iterative evaluations using a clipping window, after which the data is transformed back using the inverse log square root (38). In contrast to the other baselining methods, SNIP does not lead to distortions of the baseline at the edges of the spectral interval (31).

Prior to determining the efficacy of the baselining methods, each spectrum was normalized using sum
normalization to maintain relative proportions of peaks for comparisons. Next, the ^{2}H incorporation into
cells was quantified by calculating the ratio between the area under the curve (AUC) of the $^{12}C^2H$ (2,040-
2,300 cm^{-1}) and the $^{12}C^1H$ (2,800-3,100 cm^{-1}) region of each spectrum (Fig. 1A). The AUC was determined
by drawing a linear baseline between the interval of wavenumbers and numerically integrating the area
above the line and below the curve using the trapezoidal rule (Fig. S3). The ^{2}H content was then used to
compare the baselining methods for both truncated spectra (1,800-3,200 cm^{-1}) and whole spectra (250-3,200
cm^{-1}), and the r-squared value used to determine the fit of each model (Fig. S4).

There was no statistically significant difference between the truncated and whole spectra or between
the baselining methods, except for spectra baselined using the rolling ball method (Fig. S5, Table S2). This
indicates that the baselining method had little effect on ^{2}H calculation from Raman spectra of *E. coli*. This
could be due to the low background observed in a pure culture as compared to an environmental sample
and the low natural fluorescence of *E. coli*, resulting in minimal background subtractions from the data. We
elected to use the SNIP baselining method for our analysis because this method has minimal distortion on
data at both ends of the spectrum and has previously been used in single cell SIP-Raman studies (2, 24, 39).

211   Analysis of NanoSIMS ion pairs

The next step in our workflow was to estimate the ^{2}H content of each cell that had previously been
measured by Raman using NanoSIMS (Fig. 1B). Historically, analysis of ^{2}H uptake in cells by NanoSIMS
has been performed using either the mass ratio of $^2H/^1H$ (25, 40, 41), $^{12}C^2H/^{12}C^1H$ (2, 42), or $^{16}O^2H/^{16}O^1H$
(2, 43). Among these, H$^-$ and C$_2$H$^-$ have relatively higher ion yields compared to CH$^-$ (Fig. S6), and the
$^2H/^1H$ ratio requires relatively lower theoretical mass resolving power (MRP) to acquire a clean signal
compared to organic ion pairs, which are more susceptible to overlapping interferences (44, 45). Although
the use of $^{12}C^2H/^{12}C^1H$ and $^{12}C_2{}^2H/^{12}C_2{}^1H$ in polymer films and charcoal has been explored using high MRP,
these studies relied on a high primary beam current, relatively low spatial resolution, and/or nonstandard
instrument configurations to achieve optimum results (44, 45). While MRP is useful for resolving
overlapping peaks, abundance sensitivity provides a more relevant measure of interference between ion
species. However, in multi-species peaks, both metrics are generally impractical to measure directly.
Confounding this difficulty is the use of multiple detectors during NanoSIMS analysis. Indeed, because of
a variety of physical factors, identical ion species usually have different measured MRPs when measured
on different detectors. In this study, we compare the use of $^2H/^1H$, C^{2}H/C^{1}H, and C$_2{}^2$H/C$_2{}^1$H to detect ^2H_2O
incorporation at the single cell level. To achieve the required spatial resolution, we use a relatively small
primary beam current and standard MRP. We show that mass separation based on the concept of MRP
alone is not a strict requirement for obtaining meaningful NanoSIMS data. We discuss our findings in
relation to subcellular chemical morphology and experimental objectives. Further discussion on NanoSIMS
methodology, including theoretical mass resolving power, detector tuning, and sources of analytical
variability, is provided in the Supplemental Text.

Two observations stand out from the $^2H/^1H$ measurements. First, the background associated with cells
grown without ^2H_2O was significantly lower than measurements using the CH$^-$ and C$_2$H$^-$ ion pairs despite
the presence of significant H in the stainless steel wafers. The lower background can be attributed to the
relative ease of resolving isobars, providing evidence that our assigned ROIs did not include excessive
amounts of H signal from the instrument vacuum or the stainless steel substrate. Second, the slope of the
regression line for the % ^2H_2O in the growth medium versus NanoSIMS $^2H/^1H$ data is significantly lower
than the corresponding regressions for the other ion pairs (Fig. S7, Table S3). This difference is likely a

convolution of different IMFs between ion pairs analyzed as well as a difference in H pool being accessed.
Assuming a value for the natural abundance ^{2}H/H of 0.000156 Vienna Standard Mean Ocean Water
(VSMOW), the slope of the regression of ^{2}H/^{1}H versus the concentration of ^{2}H$_2$O in the growth media (in
%) is consistent with an instrumental mass fractionation (IMF) of approximately -380 ‰ compared with
the regressions for ^{12}C^{2}H/^{12}C^{1}H, and ^{12}C$_2$^{2}H/^{12}C$_2$^{1}H (Fig. S7). This fractionation is in the range of values
reported for SIMS analyses in the past (46) and is offset by $\sim$ 170 ‰ from the average value for fatty acids
analyzed by IRMS (Table S1). We do not know what contribution IMF had versus real variability of ^{2}H in
cellular constituents. We would expect δ^2H to change with cellular constituent as a function of depth. Using
a previously published value for sputter rate of *Bacillus* spores using a 16 keV Cs$^+$ primary ion beam of
$\sim$2.5 nm um^2 pA^{-1} s^{-1}, we estimate that the NanoSIMS data presented here were acquired after removing
about 20 nm of cell material indicating cell wall and membrane material likely have been interrogated,
however, these assertions should be verified using careful future experimentation (47). In a complex
environment like a bacterial cell, biochemically different pools account for different δ^2H isotope values are
expected to be encountered as a function of analysis depth. Large H isotope fractionations in fungal, algal,
and plant cellular components have been previously documented as has fractionation of microbial lipids
based on metabolic pathway (48-50). Intriguingly, our analysis showed a statistically significant difference
($p < 0.005$) between ^{2}H/^{1}H estimates for each pair of mass ratios used to measure the ^{2}H content of cells at
each concentration of ^{2}H$_2$O in the media (Fig. S8). On average, the ^{12}C^{2}H$^-$/^{12}C^{1}H$^-$ ratio indicated the highest
apparent relative ^{2}H content in cells, the ^{12}C$_2$^{2}H$^-$/^{12}C$_2$^{1}H$^-$ ratio indicated the second highest, and the ^{2}H$^-$/^{1}H$^-$
ratio showed the lowest. This discrepancy was likely not sample-specific as the ^{12}C^{2}H$^-$/^{12}C^{1}H$^-$ and ^{12}C$_2$^{2}H$^-$
/^{12}C$_2$^{1}H$^-$ mass ratios were concurrently measured on the same cells. Rather, the differences were likely due
to some combination of sampling different ^{2}H populations associated with the individual cell, different IMF
factors, and variable atomic mixing during the sputtering process (22). Removal of isobaric interferences,
such as the subtraction of the ^{13}C^{1}H$^-$ peak from the m/z 14 peak, would lead to lower detection limits for the
^{12}C^{2}H$^-$/^{12}C^{1}H$^-$ analyses (2). It is likely that calibration using a suitable organic ^{2}H/^{1}H standard would
improve accuracy for all ion pairs.
Because DAPI staining was necessary to visualize cells during etching of the sample surface with the
laser dissection microscope used for cell mapping, we investigated its effect on ^{2}H signal dilution. Addition
of hydrogen-rich chemicals, such as the fixatives paraformaldehyde (PFA) and ethanol or DNA-staining
chemicals such as DAPI and SYBR dyes, can result in a dilution of isotopes in cells (43). Comparing ^{2}H-
labeled cells with and without DAPI revealed an average 12.8% decrease in total ^{2}H signal for DAPI treated
cells (Fig. S9). This dilution is expected, as each DAPI molecule contains fourteen hydrogen atoms and
binds to twelve base pairs along the minor groove in DNA (51). Given the approximately 4.6 Mb long
chromosome of *E. coli* K12, DAPI staining could introduce an estimated 5.4 x 10^6 additional ^{1}H per
chromosome copy. Although this number is likely inflated due to the specific binding of DAPI to adenine-
thymine rich regions in the minor groove of DNA, which are not present in every full turn of the minor
groove (51), this example serves to highlight the effect DAPI can have on single cell isotope measurements.
A previous study demonstrated that both DAPI and SybrGreen DNA-stains have a statistically significant
dilution effect on ^{13}C and ^{15}N isotopes (^{2}H was not evaluated). The authors concluded that if environmental
microbes with low metabolic activity are analyzed, pretreatment of cells (PFA-fixation, DNA-staining, etc.)
should be limited as much as experimentally possible (43). In our work, we did not evaluate the ^{2}H-dilution
effect of DAPI on Raman measurements.
281
Single-cell comparison of Raman and NanoSIMS

Due to the variations in apparent ^{2}H content of cells between the different mass ratios, we made our
NanoSIMS ion pair comparisons using all three mass ratios. A comparison of the Raman spectra (analyzed
using the 2040-2300 cm^{-1} wavenumber range for ^{2}H calculations) and NanoSIMS measurements (for ^{2}H/^{1}H,
^{12}C^{2}H/^{12}C^{1}H, and ^{12}C$_2$^{2}H/^{12}C$_2$^{1}H mass ratios) of 543 cells revealed a high degree of comparability between
the two techniques (Fig. 2). The highest correlation of labeling percentages determined by Raman and
NanoSIMS were found using the ^{12}C$_2$^{2}H/^{12}C$_2$^{1}H mass ratio (0.98 r-squared). The ^{2}H/^{1}H and ^{12}C^{2}H/^{12}C^{1}H
mass ratios showed equal correlation (0.97 r-squared) to the Raman results although the NanoSIMS data
for the ^{2}H/^{1}H had a much lower degree of spread for the 0% ^{2}H$_2$O incubations as compared to ^{12}C^{2}H/^{12}C^{1}H.
Analysis of covariance (ANCOVA) was used to evaluate the relationship between the linear models,
showing that each model for each specific mass was significantly different (F-value > 150, p-value < 2x10$^-$
^{16}). Considering the slope of the models, each comparison had a slope >1, indicating that Raman estimated
a higher amount of deuterium in the cells than NanoSIMS. This discrepancy was the greatest with the ^{2}H/^{1}H
mass ratio for which the slope of the model was 1.5. Furthermore, the y-intercept was close to zero for
^{2}H/^{1}H and ^{12}C$_2$^{2}H/^{12}C$_2$^{1}H while the ^{12}C^{2}H/^{12}C^{1}H mass ratio had a y-intercept of -4.86, representative of the
high estimation of ^{2}H in the cells for the 0% ^{2}H$_2$O incubation. Comparison of ^{2}H content in each cell based
on ^{2}H$_2$O incubations as measured by NanoSIMS each or Raman revealed statistically significant differences
between the techniques (Fig. 3). The ^{12}C$_2$^{2}H/^{12}C$_2$^{1}H mass ratio was not significantly different from the
Raman data at the 0% and 15% ^{2}H$_2$O incubations.

Fig. 2. Single cell comparison of the ^{2}H content of cells as measured by Raman and NanoSIMS using specific mass ratios. Each dot represents a single cell analyzed with both techniques for different ^{2}H$_2$O concentrations in the culture medium. The linear model equation and fit (blue line) is shown for each comparison of Raman to NanoSIMS regarding the specific mass ratios.

Fig. 3. Comparison of ^{2}H atom percent as measured by Raman (2040-2300 cm^{-1} wavenumber range for ^{2}H peak) and NanoSIMS using the three different mass ratios. A higher ^{2}H content was measured by Raman as compared to the NanoSIMS ^{2}H/^{1}H mass ratio measurements. The opposite is true for the ^{12}C^{2}H/^{12}C^{1}H mass ratio for which Raman measures a lower ^{2}H content than NanoSIMS. The ^{12}C$_2$^{2}H/^{12}C$_2$^{1}H mass ratio had more comparable results to Raman for the 0% and 15% ^{2}H$_2$O incubations but measured a lower ^{2}H content than Raman for the 30% and 50% incubations. All statistically significant differences are shown: *** = p-value < 1.0 x 10^{-3}, **** = p-value < 1.0 x10^{-10}, ns = not significant.

Analysis of Raman wavenumber range

The wavenumber range of 2,040-2,300 cm^{-1} for Raman analysis of ^{2}H incorporation was originally proposed by Berry *et al.* and has since been ubiquitously applied in similar studies (2, 16). However, we observed that applying this range uniformly to calculate ^{2}H content in cells with lower incorporation levels, such as those incubated in 15% ^{2}H$_2$O, could result in baseline values being incorporated in final deuterium content (Fig 1A and S3). This discrepancy could affect the calculated ^{2}H values. Furthermore, the exact destination of ^{2}H within a cell is dependent on the metabolism of the organism (*e.g.* anaerobic/aerobic and autotrophic/heterotrophic) and specific wavenumber peak can be influenced by additional available substrates in the incubation, highlighting the nuances to individual datasets (2, 52). To address this, we investigated whether an empirically derived wavenumber range would yield results comparable to those obtained by NanoSIMS.

To assess the potential for alternative wavenumber ranges, we used the 128 Raman spectra collected
from cells grown in 50% ^2H_2O. These spectra were processed using Savitzky-Golay smoothing, SNIP
baselining, and sum normalization. The purpose of using the 50% ^2H_2O data was two-fold. First, the higher
^2H_2O concentration produces a more normal distribution of the ^2H peak, thereby reducing the noise that
may influence the processing. Second, it ensured the inclusion of all relevant data since lower ^2H_2O
concentrations would constrain the wavenumber range and likely result in removal of data at higher
concentrations. Raman spectra were truncated to 2,000-2,400 cm^{-1} and further smoothed using a spline
function (53). Spline smoothing is a nonparametric regression technique used to estimate a curve directly
from noisy data without relying on specific parametric models, thus reducing the impact of inherent noise
in the spectra on the wavenumber range calculations. The smoothness of the curve is influenced by the
degrees of freedom (DOF) of the function, which can dramatically affect the calculation of wavenumber
ranges. To evaluate the impact of smoothing on wavenumber ranges, we used a stepwise strategy using
various DOF (Fig. S10). We computed the 2nd derivative from the smoothed curve to identify the inflection
points (*i.e.*, where the 2nd derivative changes signs) and used the wavenumber values at which the inflection
points occurred to determine the wavenumber range used to calculate the AUC (Fig. S11 and S12). This
approach provides a standardized method for determining the wavenumber range used to calculate ^2H,
yielding an average wavenumber value for the DOF of the 128 Raman spectra analyzed (Table S4).
Using the derived wavenumber ranges for each DOF, we calculated the ^2H atom percent for the 50%
^2H_2O Raman data and compared the results to those obtained using the canonical 2,040-2,300 cm^{-1} range
and NanoSIMS data (Fig. S13). Notably, the canonical 2,040-2,300 cm^{-1} wavenumber range produced ^2H
values that closely aligned with NanoSIMS $^{12}C^2H/^{12}C^1H$ mass ratio values, suggesting that historic studies
using this wavenumber range and mass ratio have yielded reliable results. Our analysis indicated that a
narrower wavenumber range calculated from 10 DOF resulted in ^2H concentrations that were statistically
indistinguishable from the $^2H/^1H$ mass ratio (Table S5). Similarly, no statistical difference was found when
comparing the ^2H atom percent in Raman spectra analyzed using 5.57 DOF to the NanoSIMS $^{12}C_2^2H/^{12}C_2^1H$
mass ratio. When applied to the 0%, 15%, 30%, and 50% ^2H_2O datasets, these optimized wavenumbers
yielded ^2H values that were comparable to the NanoSIMS results with lower statistical variance across the
datasets of the respective mass ratio (Fig. 4, S13 and Table S6). Linear regressions of individual cell data
further supported these findings, showing stronger correlations and improved alignment between the two
techniques for all incubation conditions (Fig. S14). As compared to the ^2H atom percent calculations using
the canonical 2,040-2,300 cm^{-1} range shown in Figure 2, wavenumber ranges calculated using the 2nd
derivative approach had closer to 1, suggesting a stronger linear correlation between the data when
employing wavenumber ranges calculated using the second derivative approach (Table S7). This analysis
demonstrates that calculating wavenumber ranges using the second derivative approach improves the
consistency and accuracy of Raman-based ^2H quantification, particularly in studies with varying ^2H_2O
concentrations.
363

Fig. 4. Comparison of ^{2}H atom percent as measured by Raman using wavenumber ranges determined by DOF (see Table S4) compared to the three mass ratios measured using NanoSIMS. Using the wavenumber ranges deduced from the 2nd derivatives resulted in Raman data more closely resembling the NanoSIMS measurements for the ^{2}H/^{1}H and ^{12}C$_2$^{2}H/^{12}C$_2$^{1}H mass ratios. All statistically significant differences are shown: *** = p-value < 1.0 x 10^{-3}, **** = p-value < 1.0 x10^{-10}, ns = not significant.

Conclusion

Single species microbial populations have been shown to exhibit metabolic heterogeneity, requiring analysis on a single cell level to differentiate phenotypes within the population (54-56). The application of SIP in conjunction with Raman or NanoSIMS is a common approach to studying nuanced physiological differences in populations (3, 12, 16) and for determining the metabolic activity of cells (14, 18, 57-59). Our work, together with the recent study by Caro *et al.* (26), enables a direct comparison and statistical analysis of Raman and NanoSIMS performed on single-cells using different concentrations of ^{2}H$_2$O. By quantifying the uptake of ^{2}H into the biomass of individual *E. coli* cells using correlative SIP-Raman-NanoSIMS, we established the extent of data equivalence between the two techniques, improving the comparability of studies employing either method.

Several factors were considered before comparing datasets. First, we assessed statistical pre-processing methods for Raman data showing that for *E. coli* K12, truncation and various baselining methods have little impact on the calculation of ^{2}H content (Fig. S4 and S5). While pre-processing methods had negligible influence on silent region of the spectra, it is likely that there would be a measurable effect on the fingerprint region of the spectra, highlighting the importance of testing the different pre-processing methods prior to data analysis (31). Additionally, we investigated ^{2}H measurement using different masses for NanoSIMS and found that ^{2}H measured in cells varies depending on the mass analyzed (Fig. S8). This variation could be due to a combination of different IMF factors, variable atomic mixing during the sputtering process,

hydrogen background in the instrument, and sampling of different ^{2}H populations pools within the cell (22,
60). For example measurements using ^{2}H/^{1}H mass ratio include all sputtered ^{2}H and ^{1}H, whereas
^{12}C^{2}H/^{12}C^{1}H and ^{12}C$_2$^{2}H/^{12}C$_2$^{1}H mass ratios specifically sample ^{2}H and ^{1}H that is bound to ^{12}C (22, 61). This
difference introduces a selection bias, as these mass ratios exclude hydrogen bound to other atoms, such as
nitrogen or oxygen, that are not proportionally measured. Raman measurements exhibit a similar bias, as
they detect ^{2}H bound to various biomolecules.

Comparison of Raman and NanoSIMS data measuring the anabolic metabolism of ^{2}H in single cells
demonstrated areas of similarity, supported by robust correlation coefficients for data analyzed using both
the canonical Raman wavenumber range for and the 2nd derivative approach (Fig. 2 and S13). Despite these
correlations, the inherent methodological differences between Raman and NanoSIMS highlight the need
for careful consideration of their respective strengths and limitations. These results support the use of either
Raman or NanoSIMS when analyzing cells incubated with ^{2}H$_2$O, offering researchers flexibility in choosing
the most suitable technique for their specific experimental requirements while allowing comparability to
literature data.

The decision to use Raman versus NanoSIMS depends on the experimental question. Modern Raman
instruments enable faster measurements than NanoSIMS and the infrastructure cost is several times lower.
Acquisition times for spontaneous Raman, such as the instrument used in this study, typically range from
1-60 seconds per measurement, but more advanced systems such as Stimulated Raman spectroscopy can
measure a field of view, which can contain hundreds of cells, within a minute (16). Raman analysis directly
targets C-H bonds, simplifying the interpretation of activity measurements. Confocal Raman
microspectroscopy can make point measurements at a resolution of ca. 0.3 μm, making it well suited for
analyzing microbial communities (16). For NanoSIMS analysis, ^{2}H/^{1}H measurements have very low
detection limits enabling natural abundance measurements, despite having a relatively low electron affinity
(0.75 V; (62)). Assuming three standard deviations for detection limits, the uncalibrated detection limit for
the present experiment for NanoSIMS is 0.04 at% compared to >0.2% for Raman. However, NanoSIMS
measurement targets all H in the cell including exchangeable H (*e.g.*, hydroxyl, amino, or thiol groups) as
well as residual water. Thus, isotopic exchange of ^{2}H during incubation with very minimally active cells
could result in ‰ changes that affect the results (26). The ^{12}C^{2}H/^{12}C^{1}H ion pair targets C-H bonds directly,
but high background, presumably due to ^{12}C^{1}H$_2$, prevents low detection limits at the high-transmission
analysis conditions used here. An unknown amount of atomic recombination probably occurs so some
dilution of the ^{12}C$_2$^{2}H signal is possible (22). ^{12}C$_2$^{2}H/^{12}C$_2$^{1}H measurement can be performed at higher
precision, but again recombination and background interferences limit detection and potentially accuracy.
On the positive side, depending on enrichment levels, other peaks could be monitored along with either
polyatomic pair to measure assimilation of other elements beside H (44). When lower detection limits than
those afforded by ^{12}C^{2}H/^{12}C^{1}H or ^{12}C$_2$^{2}H/^{12}C$_2$^{1}H are required or the experiment requires the measurement
of all cellular ^{2}H, we recommend that researchers consider using ^{2}H/^{1}H, in addition to other mass ratios.

Our comparative investigation fills a gap in the current understanding of spectroscopy and spectrometry
techniques in microbiology and offers a foundation for the further exploration and application of SIP
methods in diverse microbial systems. The comparability of Raman and NanoSIMS measurements of single
cells presented in this study paves the way for a more nuanced and comprehensive understanding of
microbial activities, interactions, and responses to isotopic labeling.
430
Materials and Methods
Preparation of isotopically labeled cells

Escherichia coli K12 (DSM498) cultures were grown aerobically at 200 rpm agitation at 37 °C for 13
h in M9 medium (0.38 mM thiamine and 22.2 mM glucose) that had been amended to a final concentration
of either 0%, 15%, 30%, or 50% ^2H_2O (99.9%-^2H; Cambridge Isotope Laboratories). The optical density
(600 nm) of the cultures was measured every hour for the first six hours, after which measurements were
taken every half hour. When the OD_{600} had reached 0.8-0.9, 1 mL of culture was sampled (Fig. S1) and
chemically fixed by adding paraformaldehyde (PFA; Electron Microscopy Science; EM grade) to a final
concentration of 2%. These samples were then incubated for 60 min at room temperature. Cells were then
washed twice with $1\times$ phosphate buffered saline (PBS; pH 7.4) by centrifugation at 16,000 g for 5 min,
after which supernatants were removed, and cell pellets were resuspended in 1 mL of 1x PBS and samples
stored at 4 ºC.

Preparation of stainless steel coupon
Fixed cells were dried to the surface of mirrored stainless steel due to its desired properties as a Raman
substrate (63). The stainless steel coupon was prepared as previously described in Schaible *et al*. (25).
Briefly, the coupon was cleaned by washing with a 1% solution of Tergazyme (Alconox, New York, NY)
rinsed with Milli-Q water, subsequent one-minute washes in acetone and 200 proof ethanol, and then air
dried. To maintain correct orientation of the samples, asymmetric boxes were etched into the mirrored
surface of each coupon using a welder's pen (Fig. S2). 1 µL of each sample was spotted onto the coupon
and air-dried at 46 °C for 1 min, after which the coupon was washed in 1:1 ethanol:MilliQ for 1 minute and
then air dried. The coupon was incubated in a 300 nM (100 ng/mL) solution of DAPI (4',6-diamidino-2-
phenylindole, ThermoFisher, Waltham, MA) for 3 minutes as per the manufacturer's instructions. The
coupon was then rinsed three times in fresh PBS and then briefly dipped them into ice-cold Milli-Q water
to remove salts and air-dried using compressed air. To track single cells across platforms, regions of interest
(ROIs) were drawn around cells on the coupon using a Leica LMD6 Laser Microdissection System
(Danaher Corporation, Washington DC) using a power setting of 7, a speed of 7, and aperture size of 7 (Fig.
S3).

Confocal Raman microspectroscopy and spectral processing
Raman spectra of individual cells were acquired using a LabRAM HR Evolution Confocal Raman
microscope (Horiba Jobin-Yvon, located at Montana State University in Bozeman, Montana) equipped with
a 532 nm laser and 300 grooves/mm diffraction grating. The instrument was calibrated daily before use
with a silica oxide standard. Cell spectra were acquired using a $100\times$ dry objective (NA = 0.90) in the range
of 250–3200 cm^{-1}, with 3 acquisitions of 10 s each, and a laser power of 4.5 mW. Spectra were processed
using LabSpec version 6.5.1.24 (Horiba). The spectra were preprocessed in GNU R (64) using the Alkahest
(65), ptw (66), and Pracma (67) packages. Each spectra was smoothed using the Savitsky-Goley algorithm,
(34) followed by background subtraction and baselining of spectra using either asymmetric least squares
(ALS) (35), polynomial fit (36), rolling ball (37), or statistics-sensitive non-linear iterative peak-clipping
(SNIP) (38). Raman spectra baselined with the SNIP method were used for the comparison to the
NanoSIMS data. Settings used for smoothing and baselining can be found in the R file deposited on GitHub
(https://github.com/georgeschaible/Raman-spectra-processing). Each spectrum was then normalized to the
sum of its absolute spectral intensity, and the incorporation of ^2H into biomass was calculated by
$\left(\frac{^{12}C^2H}{^{12}C^2H + {}^{12}C^1H}\right) \times 100$, using the integration of the area under the curve (AUC) for the $^{12}C^1H$ (2800–3100
cm^{-1}) peak, the canonical $^{12}C^2H$ (2040–2300 cm^{-1}), or the wavenumber ranges calculated using the 2nd

derivative (Table S4). The AUC was calculated using the trapezoidal rule as follows: $AUC \approx \frac{h}{2}[\int (x_0) +$

$2\int (x_1) + 2\int (x_2) + \ldots + 2\int (x_n)]$ where h is the width of each interval, $\int x_i$ is the function value at

each data point, and n is the number of data points.

Nano-scale secondary ion mass spectrometry

To map the elemental composition and the relative isotopic abundances of hydrogen, ion images of the

same individual cells investigated by Raman were acquired using the Cameca 50L NanoSIMS at the

Environmental Molecular Sciences Laboratory at Pacific Northwest National Laboratory in Richland,

Washington. The vacuum gauge pressure in the analytical chamber during all analyses was consistently

less than 3×10^{-10} mbar. Samples were stored in a dry N_2 storage box when not under instrument vacuum.

Although no isotopic standards were available to calibrate the instrument for instrumental mass

fractionation (IMF) similar responses for different multipliers were ensured by setting electron multiplier

(EM) voltages such that the maximum pulse height distribution (PHD) was at a discriminator voltage of -

230 ± 10 mV. $^1H^-$ was used to set the PHDs for $^2H^-/H^-$ analyses and $^{28}Si^-$ was used to set PHDs for $^{12}C_2^1H^-$

$/^{12}C^1H^-$ and $^{12}C_2^2H^-/^{12}C_2^1H^-$ measurements. For $^{12}C_2^1H^-/^{12}C^1H^-$ and $^{12}C_2^2H^-/^{12}C_2^1H^-$ analyses, conditions

included a 200 μm D1 aperture, 20 μm entrance slit, 200 μm aperture slit, and 90 μm exit slits. $^2H^-/^1H^-$

analyses used a 200 μm D1 aperture, 30 μm entrance slit, 350 μm aperture slit, and 90 μm exit slits. All

NanoSIMS images were acquired using a 2 pA (~120 nm) 16 keV Cs^+ primary ion beam at 256 × 256-pixel

resolution with a dwell time of 1 ms px^{-1}. Analysis areas were pre-sputtered with ~5×10^{15} ions cm^{-2} prior

to collecting 15 consecutive frames. Secondary ions ($^{12}C^1H^-$, $^{12}C^2H^-$, $^{12}C_2^-$, $^{12}C_2^1H^-$, $^{12}C_2^2H^-$, $^{35}Cl^-$, and

$^{12}C^{14}N^{16}O^-$ or $^1H^-$ and $^2H^-$) were accelerated to 8 keV and counted simultaneously using the EMs. $^{35}Cl^-$, and

$^{12}C^{14}N^{16}O^-$ were not used for further analyses. Detectors collecting $^2H^-$ and $^1H^-$ ions were situated near the

center of the magnet radius to improve simultaneous secondary centering characteristics, with careful

attention paid to the secondary tuning of the Helmholtz coils to align ^2H and ^1H. Because different tunning

is required for simultaneous $^{12}C^2H^-/^{12}C^1H^-$ and $^{12}C_2^2H^-/^{12}C_2^1H^-$ compared to $^2H^-/^1H^-$, these measurements

were performed on different cells. After pre-sputtering, secondary signal was centered prior to collecting

15 consecutive frames for each image. The OpenMIMS plugin for ImageJ was used to correct images pixel

by pixel, for dead time (44 ns) and Quasi Simultaneous Arrivals (QSA; $\beta = 0.5$), prior to aligning and

summing the 15 frames. There is a small amount of H in the stainless steel wafers that we used for the

analyses. We found that $^2H^-/^1H^-$ data were sensitive to ROI size with respect to the *E. coli* ion image size.

Therefore, to optimize reproducibility, ROIs, of single microbes were acquired using the Bemsen algorithm

in the Auto Local Threshold function of ImageJ on $^{12}C_2^-$ or $^1H^-$ aligned and summed images. Examples of

ROIs are shown in Figure S15. Similar external and average internal uncertainties within treatments for the

$^2H/^1H$ measurements indicate that the degree of H interference from the stainless steel substrate was

minimal (Table S8). Data from single microbe ROIs were exported to a custom spreadsheet for data

reduction. Internal uncertainties were obtained from counting statistics (68).

Preparation of fatty acid methyl esters

The *E. coli* sample was freeze-dried overnight. Preparation of fatty acid methyl esters (FAMEs) was done

following the procedure adapted from Rodriguez-Ruiz et al. (28) , discussed briefly here. 2 mL of a mixture

of 20:1 v/v anhydrous methanol/acetyl chloride and 1 mL of hexane were added to 1.3 mg of the sample in

a glass culture tube. A method blank (culture tube with no sample) was also included to investigate whether

there were any contaminations resulting from the method. The two glass tubes were capped tightly and

heated at 100 °C for 10 min and then allowed to cool to room temperature. 2 mL of deionized water and 2
mL hexane added forming two phases. The hexane (top) phase was removed and collected in a 40 mL VOA
vial. The extraction was repeated with further 2 mL additions of hexane to maximize the extraction yield.
The sample and the blank were concentrated to 0.125 mL under nitrogen gas. The FAMEs peaks present in
the sample and blank were identified via gas chromatography/mass spectrometry (GC/MS) on a
ThermoScientific Trace 1310 TSQ-9000 equipped with a Zebron ZB-XLB column (30 m x 0.25 mm I.D.
x 0.25 μm film thickness) and programmable temperature vaporizing (PTV) injector operated in split mode
(split ratio 41.7), using helium as a carrier gas (flow rate = 1.2 ml/min). The GC oven was held at 100 °C
for 1.5 min, ramped at 20 °C/min to 150 °C with no hold, then ramped at 5 °C/min to 300 °C with a final 2
min temperature hold. Peaks were identified by comparing the relative retention times and mass spectra to
those an eight-compound FAMEs standard mixture, as well as to mass spectra in the NIST MS Library
database.

δ^2H isotope analysis
The δ^2H values of the FAMEs in the sample were measured five times by a gas chromatograph coupled to
an isotope ratio mass spectrometer (Thermo Scientific 253 Plus) using a pyrolysis interface (*i.e.*,
GC/P/IRMS). Chromatographic separation was achieved on a thick-film Agilent DB-XLB column (30 m x
0.25 mm I.D. x 1.00 μm film thickness) and PTV injector operated in splitless mode, using helium as a
carrier gas (flow rate = 1.4 ml/min). The GC oven was held at 120 °C for 1 min, ramped at 120 °C/min to
225 °C with no hold, ramped at 0.5 °C/min to 240 °C, and then ramped at 10 °C/min to 325 °C with a final
5 min temperature hold. Measured isotope ratios were calibrated using hydrogen gas of known isotopic
composition and are reported in δ notation (in units of ‰, or parts per thousand) relative to the Vienna
Standard Mean Ocean Water (VSMOW) international standard ($\delta^2H = R_{AA}/R_{VSMOW} - 1$), where R =
$^2H/^1H$. Additionally, a FAME standard was analyzed to verify instrument accuracy and precision.

Statistical analyses
All datasets were analyzed in GNU R (64) using the tidyverse, rstatix, emmeans, and ggpubr packages
(69-71). Statistical differences between multiple variables were determined using pairwise t-tests with a
Bonferroni p-adjusted method. Calculations to determine the wavenumber range for ^2H used the pspline R
package (72) and in-house written R code for the calculation of 2^{nd} derivative inflection points of smoothed
curves. Code used for analysis is deposited on GitHub (https://github.com/georgeschaible/Raman-spectra-
processingd).

Data availability
Raman spectra have been deposited in the MicrobioRaman database under accession number S-MBRS8.
NanoSIMS raw data is stored long-term on the EMSL data repository (search.emsl).

Author contributions
GS and RH designed the study. GS performed culturing and prepared samples for analysis. JAC
collected the Raman spectra and GS analyzed the spectra. JBC and JB performed NanoSIMS analyses. GS
and JBC compiled the data. MM performed the lipid extraction and esterification, conducted the IR/MS
analysis, and collaborated with AS to analyze the data. JA and GS performed the second derivative
calculations. GS performed statistical analyses and prepared figures. GS, JBC, and RH wrote the manuscript
draft, which was then edited by all authors.

Acknowledgements

This study was funded through NIH award 1R35GM147166-01 to R.H. GS was partially supported by
NASA FINESST fellowship 80NSSC20K1365. JAC was supported by NSF Research Traineeship Program
award 2125748. Montana State University's Confocal Raman microscope was acquired with support by the
NSF MRI program (DBI-1726561) and the M. J. Murdock Charitable Trust (SR-2017331). A portion of
this research was performed under the Facilities Integrating Collaborations for User Science (FICUS)
program (awards DOI: 10.46936/fics.proj.2020.51544/60000211) and used resources at the Environmental
Molecular Sciences Laboratory (https://ror.org/04rc0xn13), which is a DOE Office of Science User
Facilities operated under Contract No. DE-AC05-76RL01830. This manuscript has been authored by UT-
Battelle, LLC, under contract DE-AC05-00OR22725 with the US Department of Energy (DOE). The US
government retains and the publisher, by accepting the article for publication, acknowledges that the US
government retains a nonexclusive, paid-up, irrevocable, worldwide license to publish or reproduce the
published form of this manuscript, or allow others to do so, for US government purposes. DOE will provide
public access to these results of federally sponsored research in accordance with the DOE Public Access
Plan (https://www.energy.gov/doe-public-access-plan). We thank Hope McWilliams (MSU) for assistance
with SIP labeling and Anthony Kohtz (MSU) for helpful input on the analysis of Raman spectra.

References

1.   Hungate BA, Mau RL, Schwartz E, Caporaso JG, Dijkstra P, van Gestel N, Koch BJ, Liu CM, McHugh
TA, Marks JC. 2015. Quantitative microbial ecology through stable isotope probing. Applied and
environmental microbiology 81:7570-7581.
2.   Berry D, Mader E, Lee TK, Woebken D, Wang Y, Zhu D, Palatinszky M, Schintlmeister A, Schmid
MC, Hanson BT, Shterzer N, Mizrahi I, Rauch I, Decker T, Bocklitz T, Popp J, Gibson CM, Fowler
PW, Huang WE, Wagner M. 2015. Tracking heavy water (D2O) incorporation for identifying and
sorting active microbial cells. Proc Natl Acad Sci U S A 112:E194-203.
3.   Hatzenpichler R, Krukenberg, V., Spietz, R.L., Jay, Z.J. 2020. Next-generation physiology approaches
to study microbial community function at the single cell level. Nat Rev Microbiol.
4.   Alcolombri U, Pioli R, Stocker R, Berry D. 2022. Single-cell stable isotope probing in microbial
ecology. ISME Communications 2.
5.   Neufeld JD, Wagner M, Murrell JC. 2007. Who eats what, where and when? Isotope-labelling
experiments are coming of age. The ISME Journal 1:103-110.
6.   Wegener G, Bausch M, Holler T, Thang NM, Prieto Mollar X, Kellermann MY, Hinrichs KU, Boetius
A. 2012. Assessing sub-seafloor microbial activity by combined stable isotope probing with deuterated
water and 13C-bicarbonate. Environmental microbiology 14:1517-1527.
7.   Zhang X, Gillespie AL, Sessions AL. 2009. Large D/H variations in bacterial lipids reflect central
metabolic pathways. Proceedings of the National Academy of Sciences 106:12580-12586.
8.   Friedman I. 1953. Deuterium content of natural waters and other substances. Geochimica et
cosmochimica acta 4:89-103.
9.   Litvak I, Anker Y, Cohen H. 2018. On-line in situ determination of deuterium content in water via
FTIR spectroscopy. RSC advances 8:28472-28479.
10. Lo H-Y, Wink K, Nitz H, Kästner M, Belder D, Müller JA, Kaster A-K. 2023. scMAR-Seq: a novel
workflow for targeted single-cell genomics of microorganisms using radioactive labeling. Msystems
8:e00998-23.
11. Lee N, Nielsen PH, Andreasen KH, Juretschko S, Nielsen JL, Schleifer K-H, Wagner M. 1999.
Combination of fluorescent in situ hybridization and microautoradiography—a new tool for structure-
function analyses in microbial ecology. Applied and environmental microbiology 65:1289-1297.

12. Musat N, Musat F, Weber PK, Pett-Ridge J. 2016. Tracking microbial interactions with NanoSIMS.
611   Curr Opin Biotechnol 41:114-121.

612   13. Harrison JP, Berry D. 2017. Vibrational spectroscopy for imaging single microbial cells in complex
613   biological samples. Frontiers in microbiology 8:261878.

614   14. Musat N, Foster R, Vagner T, Adam B, Kuypers MM. 2012. Detecting metabolic activities in single
615   cells, with emphasis on nanoSIMS. FEMS Microbiol Rev 36:486-511.

616   15. Yuan X, Song Y, Song Y, Xu J, Wu Y, Glidle A, Cusack M, Ijaz UZ, Cooper JM, Huang WE. 2018.
617   Effect of laser irradiation on cell function and its implications in Raman spectroscopy. Applied and
618   Environmental Microbiology 84:e02508-17.

619   16. Lee KS, Landry Z, Pereira FC, Wagner M, Berry D, Huang WE, Taylor GT, Kneipp J, Popp J, Zhang
620   M, Cheng J-X, Stocker R. 2021. Raman microspectroscopy for microbiology. Nature Reviews Methods
621   Primers 1.

622   17. Cui L, Xin Y, Yang K, Li H, Tan F, Zhang Y, Li X, Zhu Z, Yang J, Kao SJ, Ren B, Zhu YG, Musat F,
623   Musat N. 2023. Live tracking metabolic networks and physiological responses within microbial
624   assemblages at single-cell level. PNAS Nexus 2:pgad006.

625   18. Wang Y, Huang WE, Cui L, Wagner M. 2016. Single cell stable isotope probing in microbiology using
626   Raman microspectroscopy. Curr Opin Biotechnol 41:34-42.

627   19. Huang WE, Stoecker K, Griffiths R, Newbold L, Daims H, Whiteley AS, Wagner M. 2007. Raman-
628   FISH: combining stable-isotope Raman spectroscopy and fluorescence in situ hybridization for the
629   single cell analysis of identity and function. Environ Microbiol 9:1878-89.

630   20. Li Q, Chang J, Li L, Lin X, Li Y. 2023. Research progress of nano-scale secondary ion mass
631   spectrometry (NanoSIMS) in soil science: Evolution, applications, and challenges. Sci Total Environ
632   905:167257.

633   21. Nuñez J, Renslow R, Cliff JB, Anderton CR. 2018. NanoSIMS for biological applications: current
634   practices and analyses. Biointerphases 13.

635   22. Mcmahon G, Saint-Cyr HF, Lechene C, Unkefer C. 2006. CN− secondary ions form by recombination
636   as demonstrated using multi-isotope mass spectrometry of 13 C-and 15 N-labeled polyglycine. Journal
637   of the American Society for Mass Spectrometry 17:1181-1187.

638   23. Akse SP, Das G, Agustí S, Pichevin L, Polerecky L, Middelburg JJ. 2021. Imaging of organic signals
639   in individual fossil diatom frustules with nanoSIMS and Raman spectroscopy. Marine Chemistry 228.

640   24. Eichorst SA, Strasser F, Woyke T, Schintlmeister A, Wagner M, Woebken D. 2015. Advancements in
641   the application of NanoSIMS and Raman microspectroscopy to investigate the activity of microbial
642   cells in soils. FEMS Microbiol Ecol 91.

643   25. Schaible GA, Kohtz AJ, Cliff J, Hatzenpichler R. 2022. Correlative SIP-FISH-Raman-SEM-
644   NanoSIMS links identity, morphology, biochemistry, and physiology of environmental microbes.
645   ISME Communications 2.

646   26. Caro TA, Kashyap S, Brown G, Chen C, Kopf SH, Templeton AS. 2024. Single-cell measurement of
647   microbial growth rate with Raman microspectroscopy. FEMS Microbiology Ecology 100:fiae110.

648   27. Inagaki F, Hinrichs K-U, Kubo Y, Bowles MW, Heuer VB, Hong W-L, Hoshino T, Ijiri A, Imachi H,
649   Ito M. 2015. Exploring deep microbial life in coal-bearing sediment down to~ 2.5 km below the ocean
650   floor. Science 349:420-424.

651   28. Rodríguez-Ruiz J, Belarbi E-H, Sanchez JLG, Alonso DL. 1998. Rapid simultaneous lipid extraction
652   and transesterification for fatty acid analyses. Biotechnology techniques 12:689-691.

653   29. Giardina M, Cheong S, Marjo CE, Clode PL, Guagliardo P, Pickford R, Pernice M, Seymour JR, Raina
654   J-B. 2018. Quantifying inorganic nitrogen assimilation by Synechococcus using bulk and single-cell
655   mass spectrometry: a comparative study. Frontiers in Microbiology 9:2847.

656   30. Bocklitz T, Walter A, Hartmann K, Rösch P, Popp J. 2011. How to pre-process Raman spectra for
657   reliable and stable models? Analytica chimica acta 704:47-56.

658   31. Ryabchykov O, Guo S, Bocklitz T. 2018. Analyzing Raman spectroscopic data. Physical Sciences
659   Reviews 4:20170043.

32. Gautam R, Vanga S, Ariese F, Umapathy S. 2015. Review of multidimensional data processing approaches for Raman and infrared spectroscopy. EPJ Techniques and Instrumentation 2:1-38.

33. Guo S, Popp J, Bocklitz T. 2021. Chemometric analysis in Raman spectroscopy from experimental design to machine learning–based modeling. Nature protocols 16:5426-5459.

34. Savitzky A, Golay MJ. 1964. Smoothing and differentiation of data by simplified least squares procedures. Analytical chemistry 36:1627-1639.

35. Eilers PH, Boelens HF. 2005. Baseline correction with asymmetric least squares smoothing. Leiden University Medical Centre Report 1:5.

36. Lieber CA, Mahadevan-Jansen A. 2003. Automated method for subtraction of fluorescence from biological Raman spectra. Applied spectroscopy 57:1363-1367.

37. Kneen M, Annegarn H. 1996. Algorithm for fitting XRF, SEM and PIXE X-ray spectra backgrounds. Nuclear Instruments and Methods in Physics Research Section B: Beam Interactions with Materials and Atoms 109:209-213.

38. Ryan C, Clayton E, Griffin W, Sie S, Cousens D. 1988. SNIP, a statistics-sensitive background treatment for the quantitative analysis of PIXE spectra in geoscience applications. Nuclear Instruments and Methods in Physics Research Section B: Beam Interactions with Materials and Atoms 34:396-402.

39. Matanfack GA, Pistiki A, Rosch P, Popp J. 2021. Raman (18) O-labeling of bacteria in visible and deep UV-ranges. J Biophotonics 14:e202100013.

40. Doughty DM, Dieterle M, Sessions AL, Fischer WW, Newman DK. 2014. Probing the subcellular localization of hopanoid lipids in bacteria using NanoSIMS. PloS one 9:e84455.

41. Kopf SH, McGlynn SE, Green-Saxena A, Guan Y, Newman DK, Orphan VJ. 2015. Heavy water and 15 N labelling with N ano SIMS analysis reveals growth rate-dependent metabolic heterogeneity in chemostats. Environmental microbiology 17:2542-2556.

42. Sáenz JP, Sezgin E, Schwille P, Simons K. 2012. Functional convergence of hopanoids and sterols in membrane ordering. Proceedings of the National Academy of Sciences 109:14236-14240.

43. Meyer NR, Fortney JL, Dekas AE. 2021. NanoSIMS sample preparation decreases isotope enrichment: magnitude, variability and implications for single-cell rates of microbial activity. Environ Microbiol 23:81-98.

44. Slodzian G, Wu T-D, Bardin N, Duprat J, Engrand C, Guerquin-Kern J-L. 2014. Simultaneous hydrogen and heavier element isotopic ratio images with a scanning submicron ion probe and mass resolved polyatomic ions. Microscopy and Microanalysis 20:577-581.

45. Bardin N, Duprat J, Slodzian G, Wu T-D, Baklouti D, Dartois E, Brunetto R, Engrand C, Guerquin-Kern J-L. 2015. Hydrogen isotopic fractionation in secondary ion mass spectrometry using polyatomic ions. International journal of mass spectrometry 393:17-24.

46. Hauri EH, Shaw AM, Wang J, Dixon JE, King PL, Mandeville C. 2006. Matrix effects in hydrogen isotope analysis of silicate glasses by SIMS. Chemical Geology 235:352-365.

47. Ghosal S, Fallon SJ, Leighton TJ, Wheeler KE, Kristo MJ, Hutcheon ID, Weber PK. 2008. Imaging and 3D elemental characterization of intact bacterial spores by high-resolution secondary ion mass spectrometry. Analytical chemistry 80:5986-5992.

48. Abraham W-R, Hesse C. 2003. Isotope fractionations in the biosynthesis of cell components by different fungi: a basis for environmental carbon flux studies. FEMS microbiology ecology 46:121-128.

49. Osburn MR, Dawson KS, Fogel ML, Sessions AL. 2016. Fractionation of hydrogen isotopes by sulfate- and nitrate-reducing bacteria. Frontiers in microbiology 7:1166.

50. Hayes JM. 2001. Fractionation of the Isotopes of the Isotopes of the Isotopes of Carbon and Hy Carbon and Hy Carbon and Hydrogen in Biosynthetic Pr in Biosynthetic Processes. Mineralogical Society of America.

51. Larsen TA, Goodsell DS, Cascio D, Grzeskowiak K, Dickerson RE. 1989. The structure of DAPI bound to DNA. Journal of Biomolecular Structure and Dynamics 7:477-491.

52. Wang Y, Xu J, Kong L, Liu T, Yi L, Wang H, Huang WE, Zheng C. 2020. Raman–deuterium isotope
probing to study metabolic activities of single bacterial cells in human intestinal microbiota. Microbial
biotechnology 13:572-583.
53. Ramsay J, Heckman N, Silverman B. 1997. Spline smoothing with model-based penalties. Behavior
Research Methods, Instruments, & Computers 29:99-106.
54. Schreiber F, Ackermann M. 2020. Environmental drivers of metabolic heterogeneity in clonal
microbial populations. Current opinion in biotechnology 62:202-211.
55. Schreiber F, Littmann S, Lavik G, Escrig S, Meibom A, Kuypers MM, Ackermann M. 2016.
Phenotypic heterogeneity driven by nutrient limitation promotes growth in fluctuating environments.
Nature microbiology 1:1-7.
56. Zimmermann M, Escrig S, Hubschmann T, Kirf MK, Brand A, Inglis RF, Musat N, Muller S, Meibom
A, Ackermann M, Schreiber F. 2015. Phenotypic heterogeneity in metabolic traits among single cells
of a rare bacterial species in its natural environment quantified with a combination of flow cell sorting
and NanoSIMS. Front Microbiol 6:243.
57. Dekas A.E. PRS, Orphan V.J. 2009. Deep-Sea Archaea Fix and Share Nitrogen in Methane-Consuming
Microbial Consortia. Science 326:422-426.
58. Fernando EY, McIlroy SJ, Nierychlo M, Herbst F-A, Petriglieri F, Schmid MC, Wagner M, Nielsen
JL, Nielsen PH. 2019. Resolving the individual contribution of key microbial populations to enhanced
biological phosphorus removal with Raman–FISH. The ISME journal 13:1933-1946.
59. Zhang M, Hong W, Abutaleb NS, Li J, Dong PT, Zong C, Wang P, Seleem MN, Cheng JX. 2020.
Rapid determination of antimicrobial susceptibility by stimulated Raman scattering imaging of D2O
metabolic incorporation in a single bacterium. Advanced Science 7:2001452.
60. Magee C, Botnick E. 1981. Hydrogen depth profiling using SIMS—Problems and their solutions.
Journal of Vacuum Science and Technology 19:47-52.
61. Gnaser H. 2001. Ionization probability of sputtered negative cluster ions: Dependence on surface work
function and emission velocity. Physical Review B 63:045415.
62. Zhou J, Garand E, Neumark DM. 2007. Vibronic structure in C2H and C2D from anion slow electron
velocity-map imaging spectroscopy. The Journal of chemical physics 127.
63. Lewis AT, Gaifulina R, Isabelle M, Dorney J, Woods ML, Lloyd GR, Lau K, Rodriguez-Justo M,
Kendall C, Stone N, Thomas GM. 2017. Mirrored stainless steel substrate provides improved signal for
Raman spectroscopy of tissue and cells. J Raman Spectrosc 48:119-125.
64. Team RC. 2023. A langauge and environment for statistical computing, *on* http://R-project.org/.
Accessed
65. Frerebeau N. 2023. alkahest: Pre-Processing XY Data from Experimental Methods. https://CRANR-
projectorg/package=alkahest R package version 1.1.1.9000.
66. Bloemberg TG, Gerretzen J, Wouters HJ, Gloerich J, van Dael M, Wessels HJ, van den Heuvel LP,
Eilers PH, Buydens LM, Wehrens R. 2010. Improved parametric time warping for proteomics.
Chemometrics and Intelligent Laboratory Systems 104:65-74.
67. Borchers HW. 2023. pracma: Practical Numerical Math Functions https://CRANR-
projectorg/package=pracma R package version 2.4.4.
68. Fitzsimons IC, Harte B, Clark RM. 2000. SIMS stable isotope measurement: counting statistics and
analytical precision. Mineralogical Magazine 64:59-83.
69. McNamara A. 2018. Key attributes of a modern statistical computing tool. The American Statistician.
70. Kassambara A. 2019. Comparing groups: Numerical variables, vol 192. Datanovia.
71. Lenth R. 2022. emmeans: Estimated marginal means, aka least-squares means. R package version 1.7.
2.
72. Ramsey J, Ripley B. 2022. Package 'pspline'. https://CRANR-projectorg/package=pspline R package
version 1.0-19.
757