STANDARDISED FLOW CYTOMETRIC PROCEDURE FOR DEEP CHARACTERIZATION OF NANOPARTICLES

Valentina Tirelli[1][$]*, Felicia Grasso[2][$], Valeria Barreca[3], Deborah Polignano[3], Alessandra Gallinaro[3], Andrea Cara[3], Massimo Sargiacomo[3], Maria Luisa Fiani[3] and Massimo Sanchez[1]*

[1] Core Facilities, Istituto Superiore di Sanità, Rome, Italy
[2] Department of Infectious Diseases, Istituto Superiore di Sanità, Rome, Italy
[3] National Centre of Global Health, Istituto Superiore di Sanità, Rome, Italy

[$]Contributed equally to this work.

*Correspondence authors: valentina.tirelli@iss.it; massimo.sanchez@iss.it

Abstract

In recent years, there has been growing interest in the identification and quantification of submicron-sized particles (nanoparticles, NPs), such as extracellular vesicles (EVs) and viruses. This is particularly relevant in clinical settings, where they can be used for screening biomarkers or therapeutic agents. Additionally, they are used in virology or environmental monitoring to detect contaminants like micro- and nano-plastics.

The analysis of NPs smaller than 200 nm remains a problem linked to the lack of technology capable of giving reliable results. What makes the analysis of NPs difficult is their great heterogeneity in size, composition, biogenesis, and concentration in different biological fluids. Among the different techniques for the accurate study of NPs, Flow cytometry (FCM) stands out as the ideal method. It is characterized by high resolution, low time consumption, non-destructive sampling, and high throughput.

In this article, we propose the optimization of FCM procedures to identify, quantify, and purify sEVs and virus like particles (VLPs).

The protocol aims to reduce artifacts and errors in sEVs counting, overall caused by the swarming effect. Different threshold strategies were compared to ensure result specificity. Additionally, the critical parameters to consider when using conventional FCM outside of the common experimental context of use have also been identified. Finally, a fluorescent-sEVs sorting protocol was also developed with highly reliable results using a conventional cell sorter.

INTRODUCTION

The recent viral outbreak and the growing interest on EVs, recognized as cellular vehicles or biomarkers, have increased the focus on the identification and quantification of nanoparticles (NPs), characterized by a diameter ranging from 50 to 200 nm (1).

EVs are a heterogeneous population exhibiting variability in size and shape and encompassing exosomes, microvesicles or ectosomes, and apoptotic bodies (2, 3). These cell-derived membrane-bound structures can be categorizing into small EVs (<200 nm, sEVs) and medium/large EVs (>200 nm, lEVs) (4).

Functioning as mediators of cellular communication, EVs play roles in both autocrine and paracrine signaling, participating in normal physiological processes as well as pathological conditions. These membrane vesicles carry proteins, nucleic acids, and bioactive lipids resembling the cell of origin. Once released within the extracellular space and entering the circulation, EVs may transfer their cargo to neighboring or distant cells inducing phenotypical and functional changes (5,6). The unique characteristics of EVs identify them as non-invasive diagnostic biomarkers for the diagnosis and prognosis of various diseases, offering potential contributions to the development of next-generation therapeutic nanocarriers (7).

Among nanocarriers, virus-like particles (VLPs) also represent an attractive platform for the delivery of proteins or drugs in both vaccination and therapeutic strategies, combining ease of production, immunogenicity and high biosafety profile. Indeed, VLPs are small particles (20-200nm) able to self-assemble, incorporating viral structural proteins and mimicking the shape, size and antigenic display of the parental virus. Importantly, since VLPs lack the genetic material, they are non-infectious and non-replicating, avoiding the risk of active infection in exposed individuals. Furthermore, VLPs can display on their surface high density of homologous or heterologous membrane-tethered glycoprotein, inducing strong cellular and humoral immune responses (8).

In this context, it is clear that there is a critical need for a tool capable of distinguishing variations in the number of NPs from changes in molecular composition and providing multi-parametric analysis at the level of individual NPs.

Despite the development of various standard methods for isolating EVs from different sources, the accurate characterization and quantification of these vesicles have prompted the exploration of several technologies. Nanoparticle Tracking Analysis (NTA), Tunable resistive pulse sensing (TRPS), Dynamic Light Scattering (DLS) Electron Microscopy (EM), and Flow Cytometry (FCM) are among the technologies employed, each exhibiting its own set of advantages and disadvantages (9-13).

Although FCM is an ideal technique for high throughput, multiparametric characterization of single

events, conventional FCM encounters challenges in detecting particles below the 200 nm, making

precise identification of sub-micron particles difficult (14,15).

FCM is a commonly used technique in both clinical and experimental research for studying single

cells suspension. Its quick analysis of a large number of events makes it a valuable tool for obtaining

precise quantitative and qualitative data on samples. Employing blue laser excitation (488 nm), FCM

evaluates cell morphology through forward scatter (FSC) and side scatter (SSC) signals, measuring

cell size and cellular complexity, respectively. Although classically applied in immunology,

hematology, oncology, and clinical biology or diagnostics, FCM is increasingly finding applications

in non-conventional fields such as microbiology, industry, and the environment (16-19). Furthermore,

due to its high potential, FCM applications are emerging in the identification of submicron sized EVs

(20,21).

Most conventional flow cytometer are designed to detect particles larger than 1µm whereas for

particles lower than 500 nm, like most viruses (size ranging from 30 to 350 nm) and EVs, the physical

rules that determine the FSC or SSC do not apply (22).

Several strategies have been employed to enhance the resolution of FCM. Firstly, fluorescence is

essential for the detection of particles smaller than 200 nm in size (21,23,24). The utilization of

fluorescence enables the reduction of background noise facilitating the confident identification of the

NPs under investigation. Another strategy to improve resolution involves utilizing the 405 nm Violet

Side Scatter (V-SSC) signal instead of the classic 488 nm Blue Side Scatter (B-SSC). As described

by Zucker et al., V-SSC exhibits a higher separation index than B-SSC and shows a smaller

coefficient of variation (CV), thereby enhancing the resolution limit of FCM, allowing for the

identification of 150 nm polystyrene particles (25, 26).

Although technological advances in the design of new flow cytometers have led to a significant

improvement in the size resolution, there are still no instruments capable of clearly identifying and

quantifying NPs, relying solely on physical or morphological parameters. The main difficulty is

represented by the background noise, including optical and electronic noise signals, which can

overlap with NPs below 200 nm in size.

Furthermore, considering the extremely small dimensions of the NPs, the swarm effect should not be

underestimated. This effect occurs when two or more small particles pass through the laser at the

same time and are considered as a single particle, leading to an incorrect evaluation of the number of

nanoparticles present in the sample (27, 28).

The goal of our work is to develop specific and sensitive FCM procedures to identify and to quantify NPs (e.g. viruses and EVs) in biological samples. To achieve this, we will use two different types of well-characterized NPs: small EVs (sEVs) isolated from the supernatant of a melanoma cell line, and Lentivirus based Virus Like Particles (VLPs) (14, 29). Furthermore, the development of the analysis protocol for the correct identification of the NPs allowed us to transfer it to cell sorter, enabling the purification of specific sEVs.

METHODS

Cell culture

Human melanoma cells A375/M and human embryonic kidney 293T cells (HEK293T) were cultured in Dulbecco's modified Eagle's medium. All media were supplemented with 10% heat-inactivated fetal bovine serum (FBS), 100 units/ml penicillin, 100 μg/ml streptomycin and 2 mM L-glutamine (complete media). Additional 1% MEM Non-Essential Aminoacids were added to A375M cell media. All media and supplements were from Euroclone.

HEK 293T Lenti-X human embryonic kidney cells (Clontech, Mountain View, CA, USA), used to VLPs production, were maintained in Dulbecco's modified Eagle's medium, DMEM (Gibco Life Technologies Italia, Monza, Italy), supplemented with 10% fetal bovine serum FBS (Corning, Mediatech Inc., Manassas, VA, USA), and 100 units/mL penicillin/streptomycin (Gibco).

Cell lines were grown at 37°C, under 5% CO_2, in humidified incubators and routinely tested for mycoplasma contamination using the Mycoplasma PCR Detection Kit (ABM, #G238). All cell lines are commercially available (ATCC).

Extracellular Vesicle labelling and collection.

BODIPY-FL-C16 (C16) (4,4-Difluoro-5,7-Dimethyl-4-Bora-3a,4a-Diaza-s-Indacene-3-Hexadecanoic Acid (ThermoFisher, #D3821) or BODIPY 558/568 C12 (C12) (4,4-Difluoro-5-(2-Thienyl)-4-Bora-3a,4a-Diaza-s-Indacene-3-Dodecanoic Acid) (ThermoFisher, #D38355) were complexed with fatty acid-free bovine serum albumin (BSA, Sigma, #A8806) as described previously (Coscia et al., 2016). To isolate sEVs, 60-70% confluent monolayers of A375M cells in exponential growth were incubated with 7 μM C16 or C12 at 37 °C for 4 h in medium containing antibiotics and glutamine and supplemented with 0.3% FBS (cell labelling medium). To obtain double labelled C16/C12 sEVs, A375 cells were incubated with cell labelling medium containing 7 μM C16 and 7 μM C12 for 4 h. To obtain double labelled C16/Ceramide sEVs, A375 cells were incubated with 5

μM BODIPY TR Ceramide (ThermoFisher, # D7540) complexed to BSA for 90 min on ice, washed twice with Hank's balanced salt solution (HBSS) and further incubated with 5 μM C16 in cell labelling medium for 90 min at 37°C. To obtain double labelled Nef^{mut}-GFP/C12 sEVs, HEK293T cells were transfected with a plasmid encoding Nef^{mut}-GFP as previously described (Ferrantelli et al). Briefly, HEK293T cells were transfected with 1 μg/ml plasmid DNA complexed with 3 μg/ml branched Polyethylenimine (PEI) (Sigma-Aldrich # 408727). After 18 h, cells were washed once in HBSS and incubated with 7 μM C12 in cell labelling medium for 4 h at 37°C. At the end of the incubation with fluorescent lipids, cells were washed twice with Hank's balanced salt solution (HBSS) containing 0.1% (w/v) fatty acid-free bovine serum albumin (H-BSA) to remove excess probe and further incubated for 24 h in complete culture medium. The conditioned medium was either immediately processed for sEVs isolation or stored at 4°C for up to one week. Cells were detached from the plates with trypsin/EDTA, and their viability was assessed by Trypan Blue exclusion. Conditioned medium containing fluorescent sEVs was serially centrifuged at 2000g for 20 min at 4°C to discard cells and large debris. The supernatant was then centrifuged at 10,000g for 20 min at 4°C to pellet microvesicles and other debris. To isolate sEVs, the supernatant from previous centrifugation was ultracentrifuged 100,000g for 90 min at 4°C and the pellet (100K pellet) was washed in 12 ml of PBS and centrifuged again at 100,000g. Pellets were resuspended in 100-150 μL of PBS. All ultracentrifugation steps will be performed at 4°C using a SW41 Ti rotor (Beckman Coulter, Brea, CA, USA).

Virus like particles (VLPs) preparation and collection

To produce Lentivirus-based virus like particle (VLPs), 293 T Lenti-X cells were transiently transfected on 10 cm Petri dishes with 8μg of pGagGFP plasmid, encoding the codon-optimized Gag protein fused to the green fluorescent protein (GFP) (30), using JetPrime transfection kit (Polyplus Transfection, Illkirch, France). After 48 h, cell culture supernatants were recovered, cleared from cellular debris by low-speed centrifugation, passed through a 0.45 μM pore size filter (Millipore Corporation, Billerica, MA, USA) and concentrated by ultracentrifugation for 2.5 h at 65,000 × g using a 20% sucrose cushion. VLPs were dissolved in 1× phosphate buffered saline (PBS, Gibco) and stored at −80°C until use.

Flow Cytometric Acquisition of Beads, sEVs and VLPs

All acquisition experiments were performed using a CytoFLEX LX flow cytometer (Beckman Coulter Life Sciences), equipped with five lasers (355, 405, 488, 561 and 640 nm wavelengths). This instrument allows to set the side scatter on Violet laser (V-SSC) in addition to the standard blue laser

side scatter (B-SSC). Although PBS seems an obvious choice for a sheath fluid, PBS may increase the formation of calcium phosphate crystals in the fluidics system, which may cause additional background noise, clogging, or loss of laminar flow. Therefore, we used purified water as a sheath fluid. Anyway, scatter and all fluorescence parameters were set to log scale. We use two different fluorescent reference beads to fine-tune the instrument: Megamix-Plus FSC with the following diameters: 0.1 µm, 0.3 µm, 0.5 µm and 0.9 µm and Megamix-Plus SSC with following diameters: 0.16 µm, 0.20 µm, 0.24 µm and 0.5 µm (BioCytek Marseille, France). The beads were mixed (Megamix-Plus SSC/FSC) and diluted with PBS, following the instructions in the technical datasheet. Beads and all samples (sEVs and VLPs) were diluted in PBS. Before starting the acquisition, the instrument underwent a thorough washing protocol. This procedure is crucial for minimising contaminants and reducing non-specific signals. Acquisition data were analysed in CytExpert v2.3.

Flow Cytometry sorting of sEVs.

Sorting experiments were performed on MoFlo Astrios-EQ flow cytometer (Beckman Coulter), equipped with four lasers (405, 488, 561 and 640 nm wavelengths) and with a high-sensitivity forward scatter detector (488ex-FSC). Furthermore, the Astrios-EQ system is provided with inline sheath canister filters of 40 nm pore size. Forward scatter (FSC) detectors (FSC1 and FSC2) mounted the mask M1 and P1, respectively. Instrument were aligned with Rainbow QC Beads (Spherotech), while the detector settings for nanovesicle detection and fine-tuning adjustments of the alignment were made with Megamix-Plus SSC/FSC beads (BioCytek). The instrument acquisition software was the Summit v6.3.1. Analyses were performed with both the Summit and the CytExpert v2.3.

Statistical analyses

The results are shown as the means ± S.D. The data were evaluated for statistical significance using Student's t-test. Any p value <0.05 was considered statistically significant.

RESULTS

Threshold settings for detection of fluorescent sEVs

We previously described a promising approach for the study of fluorescent sEVs (24, 31), where the evaluation of fluorescent sEVs using a conventional flow cytometer (Gallios, Beckman Coulter) was described. The main strategy was to use their fluorescent emission rather than light-scattering detection, setting the threshold on fluorescence channel. This approach allows the detection of particles regardless of their size.

The evolution of technology led us to transfer this approach to more advanced instruments with enhanced performance, such as the CytoFLEX LX (Beckman Coulter). The initial step involved defining the sensitivity limit of the instrument. Since the CytoFLEX LX is equipped with the violet laser (405nm wavelength), we compared the resolution efficiency in separating small fluorescent beads gathered on violet side scatter (V-SSC) with the more common blue side scatter (B-SSC). Given that FSC is less sensitive to resolving sEVs, a mix of two types of green fluorescent beads, Megamix Plus FSC and SSC, with dimensions between 100 and 900 nm, was employed as a reference to determine the size sensitivity limit of the cytometer. These reference beads enabled the establishment of both the threshold and gain levels, achieving a good compromise between sensitivity limits and the exclusion of background noise. The visualization of all the fluorescent beads was obtained by modulating the threshold and gain values on the FITC fluorescent channel (band pass 525/40), maintaining the background noise of the PBS barely detectable (<10 events/second), as shown in Figure 1A and B, both for V-SSC and the B-SSC. As shown in Figure 1C-F, the threshold set to fluorescence allows all beads to be detected on both V- and B-SSC channels, although the separation of the smallest beads (between 100 nm and 200 nm) is much more efficient on V-SSC compared to B-SSC (Figure 1C, D, red boxes). Furthermore, dot plots 1C and 1D confirm that small beads cannot be detected by their size on the Forward Scatter (FSC) parameter while V-SSC allows a better separation of compared to B-SSC (Figure 1E, F). Then, to identify fluorescent sEVs, both threshold and gain were defined on the appropriate channel using non-fluorescent sEVs (blank sEVs) (Figure 1 G, H). Fluorescent sEVs were purified from the cell culture supernatant of the A375M melanoma cell line pulsed with the green fluorescent fatty acid, BODIPY™ FL C16 (hereinafter as C16, ex/em 505nm/512nm), as previously described. The threshold and gain values on the fluorescence channel (FITC channel, bandpass 525/40) were adjusted until the background noise and blank sEVs were barely visualized (<20 events per second, Figure 1G, H). Note that there is an increase in background noise above 400-500 nm. As shown in Figure 1I-N, the fluorescence distribution and number of C16-sEVs per microliter using both thresholds (V-SSC and B-SSC) are very similar (Figure 1M, N), but a better resolution of C16-sEVs was confirmed, as shown on the V-SSC channel (see red boxes).

An analogous acquisition strategy was applied to sEVs secreted by cells pulsed with the red fluorescent fatty acid, BODIPY™ FL C12 (referred to as C12-sEVs, ex/em 558/598 nm). The threshold was applied to the PE fluorescent channel (band pass 585/42) in this case. To minimize background noise, we used PBS and blank sEVs isolated from un-pulsed cells to set the threshold and gain values on PE channel. The histograms of Figure 2 display events recorded for PBS (Figure 2A and B) and blank sEVs (Figure 2C and D) when applying the threshold on both FITC (on the left) or

PE (on the right) channels. Figure 2E and 2F show the events recorded of C16- and C12-sEVs, acquired with their respective threshold. As further control, we acquired C16- and C12-EVs by crossing the thresholds (Figure 2G and H).

These results demonstrate that thresholds applied to a specific wavelength can efficiently exclude other fluorescence emission associated with sEVs during acquisition, and the presence or absence of signal in a specific acquisition protocol is not an artefact.

In conclusion, this approach for instrument setting allows for an easy and specific evaluation of fluorescent sEVs in each sample, effectively minimizing background noise.

Optimizing analysis conditions for high reliability FCM: Balance between Flow Rate, Abort Rate and Particles Concentrations

Quantification of NPs remains a challenging task, and to address this, we have determined the optimal conditions for sEVs-FCM experiments. High concentrations of sEVs can result in a high number of missed events (high abort rate) and may lead to coincidence of two or more events, detected by the flow cytometer as a single event (swarm effect). To overcome this effect, it is very important to dilute the sample and/or reduce the sample flow rate, ensuring that only individual sEVs exceed the threshold and contribute to the signals of the measured events.

In this context, the abort rate is a critical parameter to monitor in sEVs-FCM experiments to obtain reliable results.

The ratio between flow speed and sample dilution must be carefully controlled. By modulating/lowering both flow rate and sample concentration, it is possible to find the best conditions to ensure linearity and high reliability on the number of acquired sEVs.

Therefore, C16-sEVs were serially diluted in 200 nm filtered PBS. We acquired C16-sEVs at three different flow rates (slow 10 µl/min, medium 30 µl/min, fast 60 µl/min) and we applied the threshold on fluorescence to better detect and estimate the number of particles. The number of events per second, flow rate and the percentage of abort rate were recorded in each dilution and summarized in Figure 3. The results in the graph 3A demonstrate a decrease in linearity when the abort rate is exceeded by 5% (dotted red line) and the events per second are higher of 5,000. High-speed acquisition (black lines) corresponds to a greater deviation from linearity ($r2= 0.9175$) where the maximum acquisition rate is around 18,000 events per second, while in medium- and slow-speed acquisitions the values of r2 are 0.9780 and 0.9994, respectively. However, the graph in Figure 3B shows that it is possible to approach a value of r2 close to one ($r2= 0.9978$) even under conditions of high flow rate (black line), keeping the number of events per second below 5,000 and the abort rate

263 below 5%, regardless of the acquisition speed. The concomitant presence of high events per second,

264 high speed and high abort rate represents the least efficient condition in terms of reliability.

Flow cytometer set up and comparison between V-SSC and Fluorescence applied thresholds.

As demonstrated previously (Figure1), the separation of beads of small size (between 100nm and

200nm) is much more efficient on V-SSC compared with B-SSC. Megamix-Plus SSC/FSC beads, a

mix of beads that includes those below 200 nm, were used to draw a region including events below

200nm (Figure 4A and B). The threshold applied on V-SSC allows the visualization of the

background noise (Figure 4B) and the boundary between background signal and fluorescent sEVs,

necessary to draw the region, is shown in the dot plots of acquired of PBS (Figure 4C and D). C16-

sEVs were then analyzed applying thresholds on the fluorescent channel or V-SSC channel, showing

similar number of events in both conditions (Figure 4E and F, red numbers in brackets).

Virus Like Particles acquisition

As shown previously, we performed a comparison between the two threshold strategies to establish

overlapping results in terms of linearity and high reliability in the number of acquired sEVs. A greater

fluorescent intensity can allow better separation between background noise and fluorescent particles.

Virus Like Particles (VLPs) used for these experiments were designed to carry the fluorescent fusion

protein Gag-GFP. However, this VLPs lack the virulent components of their parent virus, eliminating

the possibility that these virus particles could cause an active infection in an exposed individual. Their

levels of fluorescence intensity are suitable for better separation of VLPs from background signals,

allowing a reliable comparison of event counts by using the two different thresholds (Figures 5A-D).

The fluorescent VLPs were acquired at same flow rate and the results are consistent with the previous

observations (Figures 3 and 4). The number of VLPs are similar with both applied thresholds,

although a full overlap is obtained with a flow rate $\leq 5{,}000$ events per seconds (Figure 5E). A slight

difference in the number of VLPs was observed above those values when setting the threshold on the

fluorescent channel, although not significant (Figure 5E).

sEVs sorting

289 The most challenging goal is to demonstrate the feasibility of sorting sEVs with high efficiency and

290 purity, despite their small size. since sEVs show weak scattering properties and low antigen densities,

291 having a reliable fluorescent tracer that allows clear identification of the vesicles to be separated from

292 the background noise becomes an essential requirement.

Recently, a method referred to as "nanoFACS," has been developed for the detection, analysis, and sorting of sEVs and small viruses by flow cytometry. This method has been optimised for use with a customized MoFlo Astrios-EQ Cell Sorter (Beckman Coulter) (32). We performed a similar experiment by using a MoFlo Astrios-EQ equipped with four lasers (405, 488, 561 and 640 nm), 70 μm nozzle, and with applied pressure of 60 psi. Although Morales-Kastresana et al have suggested a different approach (Morales-Kastresana et al., 2019), we have applied the threshold on 488-SSC (B-SSC) to allow a broader application of the procedure with commonly available instrumental configurations. As a basic strategy, we set a threshold value on B-SSC channel to visualize fluorescent reference beads smaller than 100 nm, as shown in Figure 6. Furthermore, to achieve optimal FCS separation, we tested several masks on dual forward scatter detectors, identifying the M1 and P1 masks as the best combination. Furthermore, we minimize vesicle aggregates and electronic background noise by a gating defined by plotting B-SCC Area channel versus B-SCC Width channel. In detail, as shown in Figure 6A, a region was drawn to exclude events with a higher B-SSC Width signal. The fluorescent beads Megamix-Plus FSC and SSC are shown in the dot plots (Figure 6 B, D) and histograms (Figure 6 C, E), respectively. Having defined the threshold value on the B-SSC channel and the corresponding voltage to reduce the PBS background noise to less than 300 events/sec, we confidently applied a simplified sorting procedure to sort sEVs stained with different fluorescent tracers (Figure 7). In detail, two different productions of sEVs, red fluorescent C12 and green fluorescent C16, were mixed and the sorting protocol was then applied. Of note, as shown in Figure 7B, sEVs released by unstained cells were used as negative control (blank sEVs). It is important to underline that, despite the refinement of the sorting protocol, it is not possible to eliminate the electronic background noise (quadrant Q4, Figure 7) (33). Therefore, the respective purities of sorted sEVs must be evaluated considering exclusively the positive quadrants for the fluorescent signals (Q1, Q2, and Q3). Reanalysis of the C12-sEV and C16-sEVs populations (Figure 7D, E) shown high levels of purity (98.4% Q1 and 96.6% Q3 respectively), while a non-significant number of sEVs was identified outside specific gate (Q2+Q3 or Q2+Q1, respectively) evidencing the consistency of the method.

Moreover, Figure 8 vividly illustrates the reliability of the method we have developed. In details, fluorescent sEVs were sorted from cells incubated with both C16 and Ceramide, a red-fluorescent dye (λex 589nm / λem 616nm). Even though the two tracers, Ceramide and C16, were differently distributed on the sEVs (Q1= 7.8% vs Q3= 83.1%, respectively, Figure 8A), the result of reanalysis shows a remarkable quality of the sorting (96.2% and 95.3%, respectively, Figures 8B and C). Similar experiments were performed using cells transfected with a construct that carries a Nef/mutant-GFP fusion gene and incubated with C12 to obtain different sEVs populations. The results, once again,

327   demonstrate the feasibility of the protocol to sort sEVs, allowing for C12-EVs and GFP-EVs purities

328   of 94.7% and 97.3%, respectively (Figure 8 E and F). In conclusion, the high purity achieved after

329   sorting, on specific EV populations, demonstrates the reliability of the developed protocol and the

330   high quality of the procedure, regardless of the fluorescence channel and the characteristics of the

331   fluorochrome.

DISCUSSION

The EVs are a heterogeneous population of micro- or nano- membrane-bound particles involved in different biological functions. They play a role in communication between cells and organs, by carrying many classes of molecules that act as biochemical or molecular signals. They are characterized by a great heterogeneity in pathway generation, composition of bilayer, their cargo (4, 5). EVs can transport a variety of proteins, including growth factors, receptors, and cytokines, as well as lipids, nucleic acids, and metabolites, depending on their biogenesis. Furthermore, there is also a wide diversity in their size distribution: size range of EVs starts at ~30 nm up to over 300 nm with a peak at a diameter of 100nm (5, 34). According to their dimension they can be classified as small EVs (sEVs) (< 200nm) and medium/large EVs (>200nm) (35). Small EVs have garnered a lot of attention for their potential applications in diagnosis and their therapeutic potential (36). As well as sEVs, VLPs are small particles of 20-200 nm that can be exploited as carriers for small molecules and are an attractive and versatile tool for the design of vaccines and therapeutics. With this premise, the need for a standardized approach arises from the varying identities of these particles, to establish a reliable and accurate results for reproducible measurements and downstream applications in clearly necessary. Minimal requests for a reliable analysis technique are standardization of isolation methods; optimization of EV cargo analysis; accurate size-based quantification (37). Various techniques try to respond to these methodological needs. For the purification of vesicles below 200nm in size, size exclusion chromatography (SEC) and asymmetric flow field-flow fractionation (AF4) are commonly used isolation techniques. Quantification methods such as microscopy (scanning/transmission electron microscopy and atomic force microscopy), dynamic light scattering (DLS), nanoparticle tracking analysis (NTA), FCM, and small-angle X-ray scattering (SAXS) can be cited (10, 11, 38). FCM is a technique that offers several advantages, and as such, it has received significant attention. In detail, FCM is a low-time-consuming and high-throughput technique allowing high reproducibility, and in many cases, it may require low sample manipulation. The primary challenge is the resolution limit, as it has mainly been used in cellular analyses or functional assays. Conventional FCM is not sensitive enough to detect sEVs. Due to their nano-scale dimensions, the rules to determine their size by the forward scatter (FSC) signal do not apply, making the detection of sEVs a challenge (39, 40). However, technological advances in new FCM have led to a significant improvement in the dimensional resolution, although there are still no tools capable of clearly identifying and quantifying NPs based solely on physical or morphological parameters. This study will optimize an FCM procedure for detecting nanoparticles with confidence. A straightforward protocol has been proposed for the reliable detection, quantification, and sorting of sEVs released by cells in culture supernatants. The experience, gained from using the Gallios cytofluorimeter (31), has

been successfully applied to the Cytoflex LX that boasts a wide linear dynamic range, resulting in improved resolution of dim and bright populations. In conventional FCM protocols, a linear scale is used for light scatter measurements whereas a log scale is used for fluorescence signals, excluding some analyzes such as the cell cycle. In contrast, submicron particles are analyzed by non-conventional FCM settings, and scatter signals (FSC or SSC) are visualized with log scales. Furthermore, the threshold is set to remove signals primarily derived from debris and electronic background. Moreover, FCM instruments exhibit high background noise when calibrated for the analysis of small particles (approximately <200 nm). Using SSC downstream of other lasers, such as the violet laser (405nm) or yellow laser (561nm), will significantly improve instrument resolution, as suggested by Morales-Kastresana et al. (32). However, the chances of success increase when employing fluorescence signals, which are essential for identifying particles below 200 nm, separating them from background noise and debris (31, 33). Our study confidently presents results on the detection, quantification, and sorting of sEVs.Identifying extracellular vesicles smaller than 200nm (sEVs) has been the first challenge. However, with the use of a fluorescent tracer, sEVs can be accurately visualized even though conventional flow cytometers lack the capability to identify smaller particles based solely on their physical characteristics (Forward and Side Scatter). The second drawback has been the selection of the parameter on which to apply the threshold. Our experience and the latest publications have highlighted the importance of fluorescence in visualizing nano-sized particles. Nevertheless, it is also feasible to utilize the enhanced sensitivity of SSCs when excited by lasers other than blue, thus eliminating the need for fluorescence. In our experimental setting, we used C16-sEVs, a population of metabolically labeled sEVs, approximately 80 nm in size (14) We decided to exclude direct-labeled sEVs with fluorescent tracers or with specific fluorochrome-conjugated antibodies because standardized procedures to eliminate or reduce the presence of non-specific events are still under investigation (33, 41). As shown in Figure1, we demonstrated that we successfully detected fluorescent particles around 100 nm by using V-SSC. Additionally, we compared the number of events (Bodipy-C16-sEVs) obtained by applying the threshold on the side scatter parameter with that applied on the fluorescence channel (Figure 4). In summary, our data confirm that applying the threshold on the violet laser (V-SSC) is more efficient in separating beads of different sizes than applying the threshold on the blue laser (B-SSC). Furthermore, we can clearly identify the sEVs with a high level of correspondence by applying the threshold on the physical parameters, even without fluorescence. The main pitfall for a correct sEVs quantification is the swarm effect and efforts should be made to avoid or greatly reduce multiple detections of sEVs. Our studies have revealed that a harmonious balance of speed, concentration, and abort rate is essential in obtaining precise measurements of the particles being analyzed. In particular, the acquisition speed of individual events

per time unit is significant in reducing the percent of abort rate (Figure 3). This parameter is crucial for establishing a window in which it is possible to define linearity conditions between flow rate, abort rate and sEVs concentrations. In our working condition, the limit of abort rate to maintaining linearity has been set at 5%: by utilizing this value, we can enhance the acquisition speed while maintaining a linear regime (Figure 3B). Furthermore, to demonstrate the method's robustness, we confidently enumerated C16-sEVs using SSC or fluorescence thresholds, resulting in an excellent match of recorded events (Figure 4). The same protocol was successfully applied to analyze GFP-VLPs. Once again, the robustness of the method is highlighted, which exactly enumerates the particles analyzed with both the V-SSC and fluorescent thresholds (Figure 5). Therefore, the variability of the number of events counted is strongly influenced by the lack of control over the quantitative impact of background noise, which can be effectively reduced. Applying the threshold on the corresponding fluorescent channel for the quantitative evaluation of fluorescent sEVs significantly reduces the background noise during acquisition compared to the background noise detected with standard protocols (SSC threshold). The use of a fluorescent threshold has already been demonstrated in the detection of sEVs using conventional non-customized flow cytometer. New instruments with increased resolution and sensitivity enable better visualization of events below 200nm. Although standardization of light scattering and fluorescence data between different flow cytometers remains a challenge, new instruments with increased performance in terms of resolution and sensitivity allow better visualization of events below 200nm. Applying these suggestions has enabled the development of sorting strategies that utilize the threshold on the blue laser (B-SSC) with confidence. Our results demonstrate that sorting of sEVs can provide reliable and accurate results even when using the B-SSC threshold. By separating vesicles labelled with different fluorescent tracers, a high level of purity can be achieved using cell sorters with well-defined parameters. As shown in Figure 7 and Figure 8 we reached a post sorting purity over 95%. Of note, it is fundamental to maintain the integrity of sorted particles and preserve their structure for downstream studies. In conclusion, our work emphasizes the importance of a pragmatic approach to instrument settings for the accurate identification and quantification of sEVs.

ACKNOWLEDGEMENTS

The authors acknowledge Dr. Francesco Manfredi (Istituto Superiore di Sanità, Rome) for providing the plasmid encoding Nef^{mut}-GFP.

FUNDING

The author(s) declare financial support was received for the research, authorship, and/or publication of this article. The study received support from internal funds provided by the Istituto Superiore di Sanità.

CONFLICT OF INTEREST

The authors declare that the research was conducted in the absence of any commercial or financial relationships that could be construed as a potential conflict of interest.

References

1. Jeppesen, D. K., Zhang, Q., Franklin, J. L., & Coffey, R. J. (2023). Extracellular vesicles and nanoparticles: emerging complexities. In *Trends in Cell Biology* (Vol. 33, Issue 8, pp. 667–681). Elsevier Ltd. https://doi.org/10.1016/j.tcb.2023.01.002

2. Van Niel, G., D'Angelo, G., & Raposo, G. (2018). Shedding light on the cell biology of extracellular vesicles. In *Nature Reviews Molecular Cell Biology* (Vol. 19, Issue 4, pp. 213–228). Nature Publishing Group. https://doi.org/10.1038/nrm.2017.125

3. van Niel, G., Carter, D. R. F., Clayton, A., Lambert, D. W., Raposo, G., & Vader, P. (2022). Challenges and directions in studying cell–cell communication by extracellular vesicles. *Nature Reviews Molecular Cell Biology*, 23(5), 369–382. https://doi.org/10.1038/s41580-022-00460-3

4. Zhou, X., Jia, Y., Mao, C., & Liu, S. (2024). Small extracellular vesicles: Non-negligible vesicles in tumor progression, diagnosis, and therapy. In *Cancer Letters* (Vol. 580). Elsevier Ireland Ltd. https://doi.org/10.1016/j.canlet.2023.216481

5. Bordanaba-Florit, G., Royo, F., Kruglik, S. G., & Falcón-Pérez, J. M. (2021). Using single-vesicle technologies to unravel the heterogeneity of extracellular vesicles. In *Nature Protocols* (Vol. 16, Issue 7, pp. 3163–3185). Nature Research. https://doi.org/10.1038/s41596-021-00551-z

6. Kilinc, S., Paisner, R., Camarda, R., Gupta, S., Momcilovic, O., Kohnz, R. A., Avsaroglu, B., L'Etoile, N. D., Perera, R. M., Nomura, D. K., & Goga, A. (2021). Oncogene-regulated release of extracellular vesicles. *Developmental Cell*, 56(13), 1989-2006.e6. https://doi.org/10.1016/j.devcel.2021.05.014

7. Duan, L., Li, X., Ji, R., Hao, Z., Kong, M., Wen, X., Guan, F., & Ma, S. (2023). Nanoparticle-Based Drug Delivery Systems: An Inspiring Therapeutic Strategy for Neurodegenerative Diseases. In *Polymers* (Vol. 15, Issue 9). MDPI. https://doi.org/10.3390/polym15092196

8. Nooraei, S., Bahrulolum, H., Hoseini, Z.S. et al. Virus-like particles: preparation, immunogenicity and their roles as nanovaccines and drug nanocarriers. J Nanobiotechnol 19, 59 (2021). https://doi.org/10.1186/s12951-021-00806-7

9. Akbar, A., Malekian, F., Baghban, N., Kodam, S. P., & Ullah, M. (2022). Methodologies to Isolate and Purify Clinical Grade Extracellular Vesicles for Medical Applications. In *Cells* (Vol. 11, Issue 2). MDPI. https://doi.org/10.3390/cells11020186

10. Diehl, J. N., Ray, A., Collins, L. B., Peterson, A., Alexander, K. C., Boutros, J. G., Ikonomidis, J. S., & Akerman, A. W. (2023). A standardized method for plasma extracellular vesicle isolation and size distribution analysis. *PLoS ONE*, 18(4 April). https://doi.org/10.1371/journal.pone.0284875

11. Dumčius, P., Mikhaylov, R., Zhang, X., Bareford, M., Stringer, M., Errington, R., Sun, C., Gonzalez, E., Krukovski, T., Falcon-Perez, J. M., Liang, D., Fu, Y. Q., Clayton, A., & Yang, X. (2023). Dual-Wave Acoustofluidic Centrifuge for Ultrafast Concentration of Nanoparticles and Extracellular Vesicles. *Small*, 19(35). https://doi.org/10.1002/smll.202300390

12. Gul, B., Syed, F., Khan, S., Iqbal, A., & Ahmad, I. (2022). Characterization of extracellular vesicles by flow cytometry: Challenges and promises. In *Micron* (Vol. 161). Elsevier Ltd. https://doi.org/10.1016/j.micron.2022.103341

13. Hendrix, A., Lippens, L., Pinheiro, C., Théry, C., Martin-Jaular, L., Lötvall, J., Lässer, C., Hill, A. F., & Witwer, K. W. (2023). Extracellular vesicle analysis. *Nature Reviews Methods Primers*, 3(1). https://doi.org/10.1038/s43586-023-00240-z

14. Barreca, V., Boussadia, Z., Polignano, D., Galli, L., Tirelli, V., Sanchez, M., Falchi, M., Bertuccini, L., Iosi, F., Tatti, M., Sargiacomo, M., & Fiani, M. L. (2023). Metabolic labelling of a subpopulation of small extracellular vesicles using a fluorescent palmitic acid analogue. *Journal of Extracellular Vesicles*, 12(12). https://doi.org/10.1002/jev2.12392

15. Lucchetti, D., Battaglia, A., Ricciardi-Tenore, C., Colella, F., Perelli, L., Maria, R. De, Scambia, G., Sgambato,
A., & Fattorossi, A. (2020). Measuring extracellular vesicles by conventional flow cytometry: Dream or
reality? *International Journal of Molecular Sciences*, *21*(17), 1–15. https://doi.org/10.3390/ijms21176257
16. De Alteriis, E., Incerti, G., Cartenì, F., Chiusano, M. L., Colantuono, C., Palomba, E., Termolino, P.,
Monticolo, F., Esposito, A., Bonanomi, G., Capparelli, R., Iannaccone, M., Foscari, A., Landi, C.,
Parascandola, P., Sanchez, M., Tirelli, V., De Falco, B., Lanzotti, V., & Mazzoleni, S. (2023). Extracellular DNA
secreted in yeast cultures is metabolism-specific and inhibits cell proliferation. *Microbial Cell*, *10*(12
December). https://doi.org/10.15698/mic2023.12.810
17. Palomba, E., Tirelli, V., de Alteriis, E., Parascandola, P., Landi, C., Mazzoleni, S., & Sanchez, M. (2021). A
cytofluorimetric analysis of a Saccharomyces cerevisiae population cultured in a fed-batch bioreactor.
PLoS ONE, *16*(6 June). https://doi.org/10.1371/journal.pone.0248382
18. Wei, Y., Zhao, Z., Wan, S., Li, Y., Feng, J., Liu, Q., Zhai, J., Wang, J., Zhao, Y., Li, X., Liu, Y., Wang, T., Li, P., Lu,
Q., Zhang, B., Hu, S., & Peng, R. (2023). Flow cytometry-based cell cycle synchronization and
transcriptome analysis in cotton (Gossypium arboretum L.). *Industrial Crops and Products*, *201*.
https://doi.org/10.1016/j.indcrop.2023.116889
19. Xu, Q., Liu, S., Ji, S., Wang, Z., Wang, M., Liu, Y., Gong, X., Fu, B., Ye, C., Chang, H., & Sui, Z. (2022).
Development and application of a flow cytometry-based method for rapid and multiplexed quantification
of three foodborne pathogens in chicken breast. *LWT*, *163*. https://doi.org/10.1016/j.lwt.2022.113487
20. Arkesteijn, G. J. A., Lozano-Andrés, E., Libregts, S. F. W. M., & Wauben, M. H. M. (2020). Improved Flow
Cytometric Light Scatter Detection of Submicron-Sized Particles by Reduction of Optical Background
Signals. *Cytometry Part A*, *97*(6), 610–619. https://doi.org/10.1002/cyto.a.24036
21. van der Pol, E., de Rond, L., Coumans, F. A. W., Gool, E. L., Böing, A. N., Sturk, A., Nieuwland, R., & van
Leeuwen, T. G. (2018). Absolute sizing and label-free identification of extracellular vesicles by flow
cytometry. *Nanomedicine: Nanotechnology, Biology, and Medicine*, *14*(3), 801–810.
https://doi.org/10.1016/j.nano.2017.12.012
22. Bettin, B. A., Varga, Z., Nieuwland, R., & van der Pol, E. (2023). Standardization of extracellular vesicle
concentration measurements by flow cytometry: the past, present, and future. In *Journal of Thrombosis
and Haemostasis* (Vol. 21, Issue 8, pp. 2032–2044). Elsevier B.V.
https://doi.org/10.1016/j.jtha.2023.04.042
23. Arraud, N., Gounou, C., Turpin, D., & Brisson, A. R. (2016). Fluorescence triggering: A general strategy for
enumerating and phenotyping extracellular vesicles by flow cytometry. *Cytometry Part A*, *89*(2), 184–195.
https://doi.org/10.1002/cyto.a.22669
24. Ferrantelli, F., Tirelli, V., Barreca, V., & Manfredi, F. (2022). Generation, Characterization, and Count of
Fluorescent Extracellular Vesicles. In *Methods in Molecular Biology* (Vol. 2504).
https://doi.org/10.1007/978-1-0716-2341-1_15
25. McVey, M. J., Spring, C. M., & Kuebler, W. M. (2018). Improved resolution in extracellular vesicle
populations using 405 instead of 488 nm side scatter. *Journal of Extracellular Vesicles*, *7*(1).
https://doi.org/10.1080/20013078.2018.1454776
26. Zucker, R. M., Ortenzio, J. N. R., & Boyes, W. K. (2016). Characterization, detection, and counting of metal
nanoparticles using flow cytometry. *Cytometry Part A*, *89*(2), 169–183.
https://doi.org/10.1002/cyto.a.22793
27. Libregts, S. F. W. M., Arkesteijn, G. J. A., Németh, A., Nolte-'t Hoen, E. N. M., & Wauben, M. H. M. (2018).
Flow cytometric analysis of extracellular vesicle subsets in plasma: impact of swarm by particles of non-
interest. *Journal of Thrombosis and Haemostasis*, *16*(7), 1423–1436. https://doi.org/10.1111/jth.14154
28. Van Der Pol, E., Van Gemert, M. J. C., Sturk, A., Nieuwland, R., & Van Leeuwen, T. G. (2012). Single vs.
swarm detection of microparticles and exosomes by flow cytometry. *Journal of Thrombosis and
Haemostasis*, *10*(5), 919–930. https://doi.org/10.1111/j.1538-7836.2012.04683.x

29. M. Severa et al., "Functional diversification of innate and inflammatory immune responses mediated by
antibody fragment crystallizable activities against SARS-CoV-2," iScience, vol. 27, no. 5, p. 109703, May
2024, doi: 10.1016/j.isci.2024.109703

30. Gallinaro, A., Borghi, M., Pirillo, M. F., Cecchetti, S., Bona, R., Canitano, A., Michelini, Z., Di Virgilio, A.,
Olvera, A., Brander, C., Negri, D., & Cara, A. (2020). Development and Preclinical Evaluation of an Integrase
Defective Lentiviral Vector Vaccine Expressing the HIVACAT T Cell Immunogen in Mice. *Molecular Therapy*
Methods and Clinical Development, *17*, 418–428. https://doi.org/10.1016/j.omtm.2020.01.013

31. Coscia, C., Parolini, I., Sanchez, M., Biffoni, M., Boussadiaa, Z., Zanetti, C., Fiani, M. L., & Sargiacomo, M.
(2016). Generation, Quantification, and Tracing of Metabolically Labeled Fluorescent Exosomes. *Methods*
Mol Biol, *1448*, 2017–2235. https:// doi: 10.1007/978-1-4939-3753-0_16

32. Morales-Kastresana, A., Musich, T. A., Welsh, J. A., Telford, W., Demberg, T., Wood, J. C. S., Bigos, M., Ross,
C. D., Kachynski, A., Dean, A., Felton, E. J., Van Dyke, J., Tigges, J., Toxavidis, V., Parks, D. R., Overton, W. R.,
Kesarwala, A. H., Freeman, G. J., Rosner, A., … Jones, J. C. (2019). High-fidelity detection and sorting of
nanoscale vesicles in viral disease and cancer. *Journal of Extracellular Vesicles*, *8*(1).

33. Morales-Kastresana, A., Telford, B., Musich, T. A., McKinnon, K., Clayborne, C., Braig, Z., Rosner, A.,
Demberg, T., Watson, D. C., Karpova, T. S., Freeman, G. J., Dekruyff, R. H., Pavlakis, G. N., Terabe, M.,
Robert-Guroff, M., Berzofsky, J. A., & Jones, J. C. (2017). Labeling extracellular vesicles for nanoscale flow
cytometry. *Scientific Reports*, *7*(1). https://doi.org/10.1038/s41598-017-01731-2

34. van der Pol E, Coumans FA, Grootemaat AE, Gardiner C, Sargent IL, Harrison P, Sturk A, van Leeuwen TG,
Nieuwland R. Particle size distribution of exosomes and microvesicles determined by transmission
electron microscopy, flow cytometry, nanoparticle tracking analysis, and resistive pulse sensing. J Thromb
Haemost. 2014 Jul;12(7):1182-92. doi: 10.1111/jth.12602.

35. Welsh JA, Goberdhan DCI, O'Driscoll L, Buzas EI, Blenkiron C, Bussolati B, Cai H, Di Vizio D, Driedonks TAP,
Erdbrügger U, Falcon-Perez JM, Fu QL, Hill AF, Lenassi M, Lim SK, Mahoney MG, Mohanty S, Möller A,
Nieuwland R, …; Théry C, Witwer KW. Minimal information for studies of extracellular vesicles.
(MISEV2023): From basic to advanced approaches. J Extracell Vesicles. 2024 Feb;13(2): e12404. doi:
10.1002/jev2.12404. PMID: 38326288; PMCID: PMC10850029

36. Lee, I., Choi, Y., Shin, D. U., Kwon, M., Kim, S., Jung, H., Nam, G. H., & Kwon, M. (2023). Small Extracellular
Vesicles as a New Class of Medicines. In *Pharmaceutics* (Vol. 15, Issue 2). MDPI.
https://doi.org/10.3390/pharmaceutics15020325

37. van der Pol E, Welsh JA, Nieuwland R. Minimum information to report about a flow cytometry experiment
on extracellular vesicles: Communication from the ISTH SSC subcommittee on vascular biology. J Thromb
Haemost. 2022 Jan;20(1):245-251. doi: 10.1111/jth.15540. Epub 2021 Nov 26. PMID: 34637195; PMCID:
PMC8729195.

38. Gholizadeh, S., Shehata Draz, M., Zarghooni, M., Sanati-Nezhad, A., Ghavami, S., Shafiee, H., & Akbari, M.
(2017). Microfluidic approaches for isolation, detection, and characterization of extracellular vesicles:
Current status and future directions. In *Biosensors and Bioelectronics* (Vol. 91, pp. 588–605). Elsevier Ltd.
https://doi.org/10.1016/j.bios.2016.12.062

39. de Rond, L., van der Pol, E., Bloemen, P. R., Van Den Broeck, T., Monheim, L., Nieuwland, R., van Leeuwen,
T. G., & Coumans, F. A. W. (2020). A Systematic Approach to Improve Scatter Sensitivity of a Flow
Cytometer for Detection of Extracellular Vesicles. *Cytometry Part A*, *97*(6), 582–591.
https://doi.org/10.1002/cyto.a.23974

40. Imanbekova, M., Suarasan, S., Lu, Y., Jurchuk, S., & Wachsmann-Hogiu, S. (2022). Recent advances in
optical label-free characterization of extracellular vesicles. In *Nanophotonics* (Vol. 11, Issue 12, pp. 2827–
2863). De Gruyter Open Ltd. https://doi.org/10.1515/nanoph-2022-0057

41. Fortunato, D., Mladenović, D., Criscuoli, M., Loria, F., Veiman, K. L., Zocco, D., Koort, K., & Zarovni, N.
(2021). Opportunities and pitfalls of fluorescent labeling methodologies for extracellular vesicle profiling
on high-resolution single-particle platforms. *International Journal of Molecular Sciences*, *22*(19).
https://doi.org/10.3390/ijms221910510

Figure legend.

Figure 1. Comparison between Violet side scatter (V-SSC) and Blue side scatter (B-SSC) resolutions for fluorescent sEVs detection.

Panels A-B: FCM analysis of 0.22 μm filtered PBS acquired with threshold applied on FITC fluorescent channel (band pass 525/40).
Panels C-E: Megamix Plus FSC/SSC polystyrene beads are used to define the resolution limit of the instrument and to compare the efficiency in separating smallest beads on violet side scatter (V-SSC) compared to the conventional blue side scatter (B-SSC). Threshold was applied on FITC fluorescent channel.
Panels G-H: FCM analysis of unstained EVs released by A375M melanoma cell line (Blank sEVs).
Panels I-N: FCM analysis of fluorescent EVs released by A375M melanoma cell line pulsed with the green fluorescent fatty acid, BODIPY™ FL C16 (C16 EVs). Red boxes include beads between 100 nm and 200 nm in size.

Figure 2. FCM profile of sEVs purified from surnatant of A375 melanoma cell line.

The histograms represent FCM analysis of 0.22 μm filtered PBS (A, B), unstained EVs, (C, D), C16-labeled sEVs (E, H) and C12-labeled EVs (F, G). Fluorescent EVs were obtained pulsing cells with BODIPY™ FL C16 or BODIPY™ FL C12. Panel A, C, E, G: the threshold was applied on green fluorescent channel (B525 FITC channel). Panel B, D, F, H: the threshold was applied on red fluorescent channel (Y585 PE channel). The threshold is set to reduce the background signal as much as possible (A, B, C, D) without losing the resolution of the appropriate fluorescent sEVs (E, F). The histograms demonstrate the robustness of the analysis protocol: no events were found when C12-labeled EVs were analysed through green fluorescence threshold (G) as well as no C16- labeled sEVs were found through PE fluorescence threshold (H).

Figure 3. **Effect of flow rate, abort rate and sample concentration on analysis accuracy**.

Several halving dilutions of C16-sEVs was analysed at three different flow rates (red line-slow: 10μl/min, blue line-medium: 30 µl/min, black line-fast: 60 µl/min) and the number of events per second and the percentage of abort rate were evaluated.

The dots on the continuous lines represent the number of events per second (y-axes on the left) at the indicated dilution (x-axes); dashed lines represent the fitted data with exponential model; shaded area represents the abort rate expressed as a percentage (y-axes on the right).

A) High speed of acquisition (black line) combined with higher sEVs concentration corresponds to a greater deviation from linearity (r^2= 0.9175). Decreasing the flow rate increases the fitting of the model (r^2= 0.9780) until slow flow rate (r^2= 0.9994). Decrease in linearity is observed in all the flow rate conditions when the abort rate is exceeded by 5% (dotted line).

B) The graph shows that it is possible to approach a value of r^2 close to 1 even in conditions of high flow rate (black line), keeping the number of events/second below 5,000 and the abort rate below 5% (dotted black line).

Results are representative of three independent experiments performed in duplicate.

Figure 4. Comparison between V-SSC and Fluorescence applied thresholds.

Representative density plots of reference beads (Megamix Plus FSC/SSC), analyzed applying the threshold on the fluorescence or on the V-SSC channels, are shown in A and B, respectively. PBS filtered at 0.22 µm was used to exclude the background noise (C, D). The black boxes were drawn to include the smallest sized beads (< 200 nm) and to exclude the background noise. The C16 sEVs that fall in the black box were enumerated (E, F). The numbers in red, inside the dot plots C, D, E, and F, indicate the events per µl.

Figure 5. High reliability on the number of acquired nanoparticles.

VLPs that carry fluorescent fusion protein Gag-GFP were quantified through the two different threshold strategies: A, B) threshold on green fluorescent channel; C, D) Threshold on V-SSC channel. PBS buffer 0.22µm filtered (A, C) was used to determine instrument background noise. The black boxes were drawn to include fluorescent VLPs (< 200 nm) and to exclude the background noise.
E) The graph summarizes serial dilution of VLPs count using the two different threshold strategies. The two series of data have a good fit ($r^2 > 0.995$) as demonstration of the robustness of the method through counting with the two different threshold strategies. Count is expressed as percentage of starting VLPs preparation. Each experimental session was run with same acquisition settings (flow rate of 10 µl/min, threshold and gain value). The grey (see also B) and black (D) error bars indicate the standard deviation of samples acquired with the threshold applied on fluorescent or V-SSC channels, respectively. No significant difference was observed.

Figure 6. Gate strategy for sEVs sorting.

The B-SCC Area signal versus B-SCC Width signal allowed the exclusion of vesicle aggregates and background noise from the population of singlet particles (A). Representative dot plots and respective histograms of singlet gated Megamix-Plus FSC (B, C) and SSC (D, E) polystyrene beads with size between 100 nm and 900 mm (0.1µm, 0,16µm, 0,2µm, 0,24µm, 0.3 µm, 0.5 µm and 0.9 µm). The analysis was performed on the events falling in the singlet region (see black arrow).

Figure 7. Sorting of fluorescent mixed C16 and C12 sEVs.

Representative dot plot of background noise assessed by PBS acquisition (A) sEVs released by unpulsed cells were used as negative control for the fluorescent populations (B). Dot Plot of mixed C12- and C16- labeled sEVs are shown in C. Reanalysis of sorted C12- and C16- labeled sEVs are shown in D and C, respectively. The small tables on the left and on the right of D and F indicate the values of the positive quadrants for fluorescent signals (Q1, Q2, and Q3). The value of the positives is expressed as a percentage of the sum of the positive quadrants. The limitations of current flow cytometry technology cannot eliminate systemic background noise present in the quadrant Q4 that is excluded from the calculation of the positive sEVs.

Figure 8. High resolution Sorting of Ceramide-C16 EVs and C12-Nef GFP EVs.

The reliability of sorting tested on sEVs sample obtained from cells incubated with C16 and Ceramide (A-C), and sEVs sample released by cells transfected with a construct that carries a Nef/mutant-GFP fusion gene and incubated with C12 (D-E). The sEVs populations before sorting are shown on the dot plots A and D, respectively. Reanalysis of sorted sEVs populations positive for a single fluorescence are shown on dot plots B, C, E, and F. The small tables on the left and on the right of the dot plots indicate the values of the positive quadrants for fluorescent signals (Q1, Q2, and Q3) and (Q5, Q6, and Q7). The limitations of current flow cytometry technology cannot eliminate systemic background noise in the quadrant Q4 and Q8 that are therefore excluded from the calculation of the positive sEVs.

Figure 1

Figure 2

Figure 3

Megamix-Plus SSC

A

B

0-200nm

0-200nm

PBS

C

(13)

D

(3)

0-200nm

0-200nm

C16 sEVs

E

(4407)

F

(4387)

0-200nm

0-200nm

V-SSC

FITC Channel

Figure 4

Figure 5

Figure 6

Figure 7

Figure 8