A novel method (RIM-Deep) enhances imaging depth and resolution stability of deep-cleared brain tissue in inverted confocal microscopy

Yisi Liu[1], Pu Wang[2], Junjie Zou[3], Hongwei Zhou[1,4] *

[1]Microbiome Medicine Center, Department of Laboratory Medicine, ZhuJiang Hospital, Southern Medical University, Guangzhou, Guangdong, China.

[2]Nikon Precision, Shanghai, China.

[3] Department of Neurosurgery, ZhuJiang Hospital, Southern Medical University, The National Key Clinical Specialty, Guangzhou, Guangdong, China.

[4] State Key Laboratory of Organ Failure Research, Southern Medical University, Guangzhou, Guangdong, China.

Correspondence: Hongwei Zhou, E-mail: hzhou@smu.edu.cn

Postal address: Microbiome Medicine Center, Department of Laboratory Medicine, Zhujiang Hospital, Southern Medical University, Guangzhou, Guangdong, 510282, China.

Funding: This work was supported by grants from the National Key R&D Programme of China (2022YFA0806400), Guangzhou Key Research Programme on Brain Science (202206060001) and the National Natural Science Foundation of China (82130068) to H- WZ.

Contributors: Yisi Liu and Pu Wang designed and supervised the study. Yisi Liu collected and analysed the data. Yisi Liu and Junjie Zou collected samples. Yisi Liu wrote the manuscript. Hongwei Zhou contributed to text revision and discussion. All authors discussed the results and approved the manuscript.

Competing interests: None declared.

Abstract

The increasing use of tissue clearing techniques underscores the urgent need for cost-

effective and simplified deep imaging methods. While traditional inverted confocal

microscopes excel in high-resolution imaging of tissue sections and cultured cells,

they face limitations in deep imaging of cleared tissues due to refractive index

mismatches between the immersion media of objectives and sample container. To

overcome these challenges, the RIM-Deep was developed to significantly improve

deep imaging capabilities without compromising the normal function of the confocal

microscope. This system facilitates deep immunofluorescence imaging of the

prefrontal cortex in cleared macaque tissue, extending imaging depth from 2 mm to 5

mm. Applied to an intact and cleared Thy1-EGFP mouse brain, the system allowed for

clear axonal visualization at high imaging depth. Moreover, this advancement enables

large-scale, deep 3D imaging of intact tissues. In principle, this concept can be

extended to any imaging modality, including existing inverted wide-field, confocal,

and two-photon microscopy. This would significantly upgrade traditional laboratory

configurations and facilitate the study of connectomics in the brain and other tissues.

Introduction

Tissue clearing techniques enable 3D imaging of whole tissues and organs [1]. Unlike traditional histology, these techniques reduce light scattering and absorption, enhancing depth imaging capabilities and achieving single-cell resolution in deep samples [2]. These approaches are now widely used in neurobiology, developmental biology, immunology, and oncology. To achieve deep imaging of fully cleared tissues, various optical sectioning techniques for three-dimensional imaging have rapidly developed, including light sheet microscopy, two-photon microscopy, and confocal microscopy [1].

For effective 3D imaging of cleared tissues, it is crucial to utilize microscopes equipped with objectives that offer both low to medium magnification and long working distances [3]. Objectives with low numerical aperture (NA) are often used to achieve these long working distances, but they face significant challenges in maintaining high axial resolution for deep tissue imaging [4]. The main issue arises from light scattering and reflection at the interfaces within the sample and between the sample and the lens, leading to image distortion and blurring. This phenomenon, known as spherical aberration, occurs when light rays entering the periphery of the lens focus at a different point on the optical axis compared to those entering near the center, leading to a significant degradation in fluorescence signal intensity and image resolution [5]. Using dry objectives with high-refractive-index mounting media or cleared tissues can exacerbate this issue, as the sample itself contributes to the aberration [6]. Therefore, matching the refractive index (RI) of the objective's immersion medium to that of the sample's immersion medium is essential to mitigate axial distortion caused by spherical aberration.

Many immersion objectives feature correction collars that can be adjusted to compensate for RI mismatches. Achieving an exact match between the refractive indices of the immersion medium and the sample medium is often challenging. Although dipping caps with glass windows can transform air objectives into immersion lenses by enabling the front lens to be directly inserted into the sample

medium, this approach helps to avoid axial deviation during imaging. However,
immersion objectives on inverted microscopes encounter limitations [7]. These
objectives cannot be inserted into sample chambers and must move up and down
outside the chamber, which restricts the imaging depth.

To overcome these limitations, several alternative methods have been developed. One
approach involves moving the sample within a chamber filled with the mounting
medium while keeping the imaging objective fixed outside the chamber [8]. Another
technique utilizes adaptive optics, such as deformable mirrors or spatial light
modulators, to correct aberrations across the field of view [9]. Additionally, radially
symmetric phase masks are used to balance spherical aberration at various depths
within the sample, thereby simplifying the image restoration process through post-
capture deconvolution [10].

However, inverted confocal microscopes encounter issues due to the gravitational
dispersion of the RI buffer between the objective and the confocal dish, resulting in
RI mismatches with air and complicating high-resolution imaging. Additionally, the
use of long-working-distance air objectives causes severe spherical aberrations due to
RI mismatches, further limiting imaging depth and resolution.

Here, we developed the Refractive Index Matching-Deep (RIM-Deep) system to
address these challenges, which includes an immersion chamber designed around the
optical components of the objective lens and a specimen holder integrated with a
motorized x–y–z stage. This design effectively stabilizes the refractive index between
the objective and the sample medium. When paired with an immersion objective, the
RIM-Deep system enables high-resolution, deep imaging, significantly enhancing the
capabilities of inverted confocal microscopy. This innovation provides a practical
solution for achieving high-quality 3D imaging of cleared tissues, paving the way for
more detailed and accurate biological studies.

Materials and Methods

Animals

We used 6 years male Macaca fascicularis, male C57BL/6J mice and male Thy1-EGFP mice gifted from Prof. Xiong Cao (South Medical University). After anesthetizing the mice with sodium pentobarbital (50 mg/kg, intraperitoneal injection), we opened the thoracic cavity to expose the heart for perfusion and fixation with saline and 4% formaldehyde. The brain tissue was then removed and fixed overnight in 4% formaldehyde. All animal procedures were approved by the Southern Medical University Animal Ethics Committee and conducted in compliance with the guidelines of the Chinese Council on Animal Care to minimize animal suffering and reduce the number of animals used.

Sample preparation for clearing

We employed the CUBIC method for clearing Thy1-EGFP mouse brain tissue [11], with the process carried out at the Shared Instrument Platform of the School of Life Sciences at Tsinghua University.

Sample pretreatment and immunolabeling

As previously described [12], The Macaca fascicularis fixed brain underwent a series of washes and treatments as follows: Initially, they were washed twice for one hour each in PBS, followed by sequential hour-long immersions in 50%, 80%, and 100% methanol solutions, with the latter repeated. The samples were then bleached overnight at 4°C using a 5% hydrogen peroxide solution in a 20% DMSO/methanol mix (prepared with 1 volume of 30% H_2O_2, 1 volume of DMSO, and 4 volumes of methanol, all kept ice cold). Post-bleaching, the samples were washed twice in methanol for an hour each, followed by two one-hour washes in a 20% DMSO/methanol solution, then in 80% and 50% methanol solutions for an hour each, and finally, twice in PBS for an hour. The last step involved two one-hour washes in a PBS solution containing 0.2% Triton X-100, preparing the samples for subsequent

staining procedures.

The samples underwent a series of incubation and washing processes as described:

Initially, they were incubated overnight at 37°C in a solution of PBS with 0.2% Triton

X-100, 20% DMSO, and 0.3 M glycine. Following this, they were blocked using a

PBS solution containing 0.2% Triton X-100, 10% DMSO, and 6% Donkey Serum at

37°C for 3 days. Subsequently, the samples were washed twice for an hour each in a

PBS solution with 0.2% Tween-20 and 10 µg/ml heparin (PTwH), then incubated in

primary antibody (Anti-CD31, ab281583) dilutions in a PTwH solution with 5%

DMSO and 3% Donkey Serum at 37°C for 4 days. After this incubation, the samples

were washed in PTwH for one day, followed by an incubation in secondary antibody

(Alexa Fluor® 488, ab150073) dilutions in a PTwH solution with 3% Donkey Serum

at 37°C for 4 days. The final step involved washing the samples in PTwH for two

days before proceeding with clearing and imaging.

The immunolabeled tissues underwent a clearing process using the iDISCO technique

[13]. This method and the solutions employed were adapted from protocols detailed in

a previously published study. The entire procedure was conducted at the Shared

Instrument Platform located in the School of Life Sciences at Tsinghua University.

Vascular labeling and tissue clearing

Adult male C57BL/6J mice (8 weeks) were used for the creation of the

photothrombotic stroke model. Mice were anesthetized with sodium pentobarbital (50

mg/kg, IP). The head was secured in a stereotaxic device, and the skull was exposed.

A photoactive dye, Rose Bengal was injected into the tail vein. Focused 561 nm laser

irradiation was then applied to the cortical area to activate the dye, causing thrombosis

and inducing ischemia by obstructing blood vessels. After surgery, mice were placed

in warm cages for recovery and subsequently returned to their cages. Three days later,

mice underwent deep sedation via intraperitoneal administration of sodium

pentobarbital at a dosage of 50 mg/kg. Following this, vascular labeling was executed

utilizing the VALID protocol [14]. Initially, mice were infused with 0.01 M

phosphate-buffered saline (PBS) for the removal of blood from the vascular system.

Subsequently, a volume of 10–15 mL of VALID working solution, warmed to

maintain its fluid state, was perfused through the circulatory system. Post perfusion,

the mice were refrigerated at 4°C overnight to solidify the gel within the vasculature.

The next phase involved organ extraction and further post-fixation in 4%

paraformaldehyde (PFA) for an extended period. Extreme care was taken to detach the

skull from the perfused mouse body to avoid distortion of the sample. During this

process, the dura mater was carefully excised. Finally, the removed skull was post-

fixed in 4% PFA for an overnight duration.

MACS was executed for demonstrated previously [15]. Fixed samples were serially

incubated in MACS-R0, MACS-R1, and MACS-R2 solutions, with gentle shaking at

room temperature. MACS-R0 was prepared by mixing 20% (vol/vol) MACS with 15%

(wt/vol) sorbitol (Sigma, 85529) in dH2O; MACS-R1 was prepared by mixing 40%

(vol/vol) MXDA with 30% (wt/vol) sorbitol dissolved in 1× PBS; MACS-R2 was

prepared by mixing 40% (vol/vol) MXDA with 50% (wt/vol) sorbitol in dH2O.

Microscope and objectives

For image acquisition, we used the Nikon AX inverted laser scanning confocal

microscope, paired with a Nikon 10x/0.5 NA immersion objective and Nikon

10x/0.45 NA air objective.

Design principle of the RIM-Deep

The RI adapter consists of a solution reservoir (Supplementary figure 2A), a specimen

holder and support bracket (Supplementary figure 2B). The solution reservoir and

specimen holder are made of glass. The nested arrangement of the solution and

specimen holder forms a semi-closed space at the outer reservoir, which is filled with

imaging buffer (Supplementary figure 2C-F). The specimen holder is centrally

positioned on the top surface of the support bracket, ensuring that its base is parallel

to the base of the support bracket. The bottom of the specimen holder is 0.17 mm

thick cover glasses. The solution reservoir is nested in the cap of the objective.

Image acquisition

Cleared tissue was placed in a 35 mm confocal dish or the dual-reservoir nested adapter. To ensure that the sample does not slide during X-Y scanning, the sample was fixed in place with transparent glass glue before imaging. Unless otherwise specified, the Z intensity correction module was used to offset the fluorescence signal degradation with depth. The NIS-Elements Advance research software was used for post-processing of images. The processed data were then imported into Imaris (Version 9.0.1, Bitplane AG) for the next step.

Result

Official adapters for 10X immersion objectives in inverted confocal microscopy significantly enhance cleared tissue imaging, but RI buffer leakage still occurs, causing RI mismatch and limiting imaging depth (Supplementary Figure 1A). In response to this challenge, the RIM-Deep was designed based on a light sheet microscopy imaging chamber. It features a media reservoir affixed to the objective and filled with RI buffer, enabling 3D imaging of the sample within an imaging buffer. To further optimize imaging performance, a support bracket integrated into the motorized stages elevates the imaging platform and centrally positions a specimen holder (Supplementary Figure 1B and 2). The effectiveness and utility of this adapter configuration have been substantiated through a comprehensive series of experimental validations (Supplementary Figure 3).

To demonstrate the effectiveness of the RIM-Deep, imaging resolution was characterized by comparing 3-µm diameter fluorescence beads using 10X long-working-distance air objectives versus 10X immersion objectives with the RIM-Deep (Figure 1A and B). At depths beyond 100 micrometers, the 10X long-working-distance air objectives exhibited significant losses in axial (XZ) resolution. In contrast, the 10X immersion objectives with the RIM-Deep maintained optimal axial resolution throughout (Figure 1C-E). However, there was no significant difference in lateral (XY) resolution between the two types of objectives (Figure 1C-E).

Subsequently, the performance of the RIM-Deep for high-depth imaging was

evaluated and compared to that of a conventional confocal dish. The imaging was

conducted on a section of a Macaca fascicularis brain that had been cleared using the

iDISCO method (Figure 1F). Gravity affected the refractive index matching fluid at

the tip of the objective, limiting the imaging depth to less than 2 mm (Figure 1G). As

observed, with the RIM-Deep, the specimen holder was fully immersed in the RI

liquid, the depth of images is obviously enhanced to 5 mm, and the vessel structures

could thereby be observed clearly from 0.5 mm to 5 mm (Figures 1H-I). Furthermore,

a 5-mm 3D network of monkey brain vasculature was successfully constructed within

a single field of view (Figures 1J).

To further demonstrate the resolution and depth robustness of the RIM-Deep,

experiments were conducted on Thy1-eGFP mouse brains cleared with the CUBIC

method (Figure 2A). The use of the RIM-Deep also significantly extended the

imaging depth to approximately 5 mm (Figure 2B, Supplementary Movie 1). This

extension enabled the clear visualization and reconstruction of neuronal soma and

axon structures in the hippocampus and thalamus (Figure 2C, E, F,G, I), maintaining

resolution at the micron level (Figure 2D and H).

A 3x3 stitching mode with multi-layer Z-axis scanning was used to examine a cleared

brain segment (Figure 3A-B, Supplementary Movie 2). This method enabled

visualization of neuronal cell bodies and axons up to nearly 5 mm in X-Y optical

slices (Figure 3C). For large tissue imaging, the RIM-Deep was used on a cleared

half-brain, achieving comprehensive imaging (Figure 3D). Its stable and precise

positioning ensured accurate alignment across all optical slices, facilitating detailed

and thorough imaging (Figure 3E-I).

Stroke, caused by disrupted brain blood supply, induces cell death and significant

vascular changes[16]. Generally, confocal microscopy is limited to thin sections and

cannot fully study vascular networks. However, with the assistance of the RIM-Deep,

complete observation of vascular damage is possible. Therefore, mouse brains with

induced cortical ischemia were examined to thoroughly investigate this process.

Using VALID labeling and MACS tissue clearing techniques, the cerebral vasculature

was reconstructed, revealing a significantly reduced vessel density in the ischemic

cortex (Figure 3J-L, Supplementary Movies 3-5). These findings demonstrate the

effectiveness of the RIM-Deep in visualizing disease-related changes in vascular

structures, making it a valuable tool in neuroscience research.

Discussion

Inverted confocal microscopy often faces challenges such as limited axial resolution

and spherical aberration when imaging cleared tissues. The RIM-Deep, featuring a

media reservoir filled with RI buffer, was designed to address these issues.

Experimental validations showed that the RIM deep maintained optimal axial

resolution when imaging fluorescent beads. Using the iDISCO method on Macaca

fascicularis brains, imaging depths reached up to 5 mm with clear visualization of

vascular structures. Similarly, in Thy1-eGFP mouse brains, the RIM deep extended

imaging depth to approximately 5 mm, allowing clear visualization of neuronal and

axonal structures. The RIM deep also effectively visualized vascular damage in

mouse brains with cortical ischemia, proving valuable in neuroscience research.The

development of tissue clearing has revolutionized neuroscientific research. However,

two-photon and light sheet microscopes are prohibitively expensive and their

operation and maintenance costs are high, limiting their application scenarios. Many

hospitals and laboratories, although capable of performing tissue clearing

independently, are unable to conduct imaging.

The inverted point-scanning confocal microscope, preferred for live cell and tissue

slice imaging, offers advantages in deep imaging of cleared tissues over upright

microscopes, including reduced sample contact and buffer use, lowering the risk of

damage and cost. However, the use of long-working-distance air objectives introduces

severe spherical aberrations due to RI mismatch, limiting imaging depth and

resolution. Our RIM-Deep addresses these issues, extending imaging depth without

compromising the microscope's normal functionality.

Wide-field microscopes, although efficient for rapid imaging, often face limitations in
depth and resolution when applied to cleared tissues [17]. The RIM-Deep can also be
employed in these cases to enhance imaging performance. When using wide-field
microscopy with the RIM-Deep, it is necessary to apply deconvolution algorithms or
other techniques to eliminate out-of-focus signals [17]. However, it is important to
note that this system may reduce fluorescence intensity in deeper regions of the
sample.

Limitations of the study and prospect

Despite the advantages offered by the RIM-Deep, several limitations remain. The
imaging speed of inverted point-scanning confocal microscopy is slower compared to
light sheet microscopy, necessitating advanced post-processing algorithms for high-
speed 3D imaging. To address this, resonance scanning methods combined with
artificial intelligence-based image analysis can significantly improve the efficiency
and accuracy of imaging, providing a more robust solution for detailed analysis [18].
Iterative reconstruction algorithms can enhance resolution and mitigate imaging
artifacts. However, achieving uniform resolution across different depths remains
challenging, particularly for very dense samples.

Furthermore, our research primarily focuses on neural tissues, and its effectiveness in
imaging other tissue types remains to be explored. What's more, the objective lens
used is limited to a correction collar for refractive indices between 1.33 and 1.51.
Techniques like iDISCO and aqueous CUBIC/MACS, with refractive indices of 1.56
and 1.51/1.52 respectively, facilitate depth imaging to nearly 5 mm. However, if the
medium's refractive index exceeds the lens's correction range, it could compromise
resolution and image quality. Recently, laboratories have developed objectives with
long working distances, high numerical apertures, and broad refractive index
correction capabilities [19]. The Schmidt immersion objective, composed of spherical
mirrors and aspheric correction plates, achieves an NA of 1.08 at a refractive index of
1.56, FOV of 1.1 mm, and WD of 11 mm [19].

Future work should explore machine learning-based techniques to further optimize
image quality and resolution [20]. Moreover, as imaging depth increases, maintaining
a high signal-to-noise ratio (SNR) becomes more challenging. Advanced denoising
algorithms will be critical to address these issues, ensuring high-quality imaging of
thick tissue samples [21, 22].

In summary, the RIM-Deep represents a significant advancement in deep tissue
imaging for inverted confocal microscopy, addressing key challenges related to axial
resolution and spherical aberration. However, further refinements and integration with
advanced computational techniques are necessary to optimize its performance and
expand its applicability in various research fields.

Figure 1. RIM-Deep on an inverted confocal microscope for increased imaging

depth of cleared tissues

(A and C) The 10X long working distance objective demonstrates significant losses in axial resolution due to the deleterious effects of mismatched refractive index.

(B and D) 10X immersion objective paired with RI adapter provides superior axial resolution throughout due to refractive index matching.

(E) Axial resolutions for a 10X long working distance objective and a 10X immersion objective paired with RIM-Deep at different axial positions. The resolution is estimated by FWHMs of intensity profiles with a Gaussian fit for 3-μm-diameter beads embedded in 1% agarose dissolved in CLARITY mounting solution (with more than 10 beads per plane). Data are presented as mean ± s.e.m.

(F) Experimental scheme illustrates the process of brain structure visualization in Macaca fascicularis, involving fixation, immunolabeling, and optical clearing with iDISCO (RI=1.56).

(G) The official adapter displays a three-dimensional reconstruction of a cleared brain up to a depth of 2 mm.

(H) RIM-Deep showcases a deeper 3D reconstruction of the cleared brain, extending to za depth of 5 mm.

(I) Progressive 2D slices of the reconstruction in (G) at varying depths, ranging from 0.5 mm to 5 mm.

(J) A top-down view of the vascular network (left) and the maximum intensity projection (MIP) of the XZ side view (right) from (H).

Figure 2. RIM-Deep improves imaging depth at micron-level resolution in Thy1-
EGFP transgenic mouse brain

(A) A Thy1-EGFP mouse brain undergoes CUBIC clearing for subsequent 3D
imaging;

(B) 3D reconstruction of the ~5 mm deep in the mouse brain (left), MIP of XZ side
view (middle) and MIP of YZ side view (right) ;

(C and G) 3D reconstructions of neuronal structures within the hippocampus and
thalamus, respectively, as indicated in B.

(D and H) Lateral slices through the indicated lateral planes in (C and G). Zoom-in
views of the selected areas in top right.

(E-F) MIP of XY side view and XZ side view of C.

(I-J) MIP of XY side view and YZ side view of G.

Figure 3. The 3D imaging and reconstruction of neural and vessel structures in

intact tissues

(A) 3x3 stitching pattern of deep imaging of a cleared brain;

(B) MIP of YZ side view of A;

(C) 2D planar image of (B). The magnified view in the top left white box distinctly
illustrates the discernibility of neuronal somas and fibers.

(D) 3D reconstruction of a half cleared brain in a Thy1-eGFP mouse brain. The white
box represents MIP;

(E-I) Selected stitched single layer images in the Z-direction;

(J) 3D imaging of the entire brain vasculature in ischemic stroke mice. The images
along the z stack are colored by spectrum;

(K) MIP of vascular imaging in the ischemic region;

(L) MIP of vascular imaging in the contralateral region.

Supplementary figure 1. Comparative illustration highlighting the distinctions between the RIM-Deep and the official adapter

(A) A conceptual illustration of cleared tissue imaging with standard adapter. Challenges remain with gravitational RI buffer dispersion between the objective and specimen, leading to RI mismatch.

(B) A conceptual illustration of cleared tissue imaging with the RI adapter, which ensures the stability and consistency of the refractive index during deep imaging.

Supplementary figure 2. Design and configuration of RIM-Deep for inverted confocal microscope

(A) A schematic diagram of a solution reservoir. ① A side view of the solution reservoir. ② A top view of the solution reservoir;

(B) A schematic diagram of a support bracket and specimen holder. ① A side view of

the support bracket. ② Top view of the support bracket.③ The front view of the

support bracket. ④Three-dimensional diagram of the specimen holder;

(C) The solution reservoir integrates with the cap;

(D) The specimen holder can be individually customized based on the size and

morphological characteristics of the samples and is strategically positioned at the

precise center of support bracket;

(E) The solution reservoir with the cap, affixed to the immersion objective, is

designed to accommodate various imaging buffers;

(F) The specimen holder is placed on the stage through the adapter.

Supplementary figure 3. Flowchart for evaluation the performance of RIM-Deep in cleared tissue imaging.

References:

[1]. Ueda, H.R., et al., Tissue clearing and its applications in neuroscience. Nat Rev Neurosci, 2020. 21(2): p. 61-79.

[2]. Susaki, E.A., et al., Whole-brain imaging with single-cell resolution using chemical cocktails and computational analysis. Cell, 2014. 157(3): p. 726-39.

[3]. Jonkman, J., et al., Tutorial: guidance for quantitative confocal microscopy. Nat Protoc, 2020. 15(5): p. 1585-1611.

[4]. Birk, U., J.V. Hase and C. Cremer, Super-resolution microscopy with very large working distance by means of distributed aperture illumination. Scientific Reports, 2017. 7(1): p. 3685.

[5]. Diel, E.E., J.W. Lichtman and D.S. Richardson, Tutorial: avoiding and correcting sample-induced spherical aberration artifacts in 3D fluorescence microscopy. Nature Protocols, 2020. 15(9): p. 2773-2784.

[6]. Gibson, S.F. and F. Lanni, Experimental test of an analytical model of aberration in an oil-immersion objective lens used in three-dimensional light microscopy. J Opt Soc Am A, 1992. 9(1): p. 154-66.

[7]. Dodt, H.U., et al., Ultramicroscopy: development and outlook. Neurophotonics, 2015. 2(4): p. 041407.

[8]. Engelbrecht, C.J. and E.H. Stelzer, Resolution enhancement in a light-sheet-based microscope (SPIM). Opt Lett, 2006. 31(10): p. 1477-9.

[9]. Ji, N., Adaptive optical fluorescence microscopy. Nat Methods, 2017. 14(4): p. 374-380.

[10]. Patwary, N., et al., Reducing effects of aberration in 3D fluorescence imaging using wavefront coding with a radially symmetric phase mask. Opt Express, 2016. 24(12): p. 12905-21.

[11]. Susaki, E.A., et al., Advanced CUBIC protocols for whole-brain and whole-body clearing and imaging. Nat Protoc, 2015. 10(11): p. 1709-27.

[12]. Renier, N., et al., iDISCO: a simple, rapid method to immunolabel large tissue samples for volume imaging. Cell, 2014. 159(4): p. 896-910.

[13]. Tomer, R., et al., Advanced CLARITY for rapid and high-resolution imaging of intact tissues. Nat Protoc, 2014. 9(7): p. 1682-97.

[14]. Zhu, J., et al., A versatile vessel casting method for fine mapping of vascular networks using a hydrogel-based lipophilic dye solution. Cell Rep Methods, 2023. 3(2): p. 100407.

[15]. Zhu, J., et al., MACS: Rapid Aqueous Clearing System for 3D Mapping of Intact Organs. Adv Sci (Weinh), 2020. 7(8): p. 1903185.

[16]. Sekerdag, E., I. Solaroglu and Y. Gursoy-Ozdemir, Cell Death Mechanisms in Stroke and Novel Molecular and Cellular Treatment Options. Curr Neuropharmacol, 2018. 16(9): p. 1396-1415.

[17]. Shaw, P.J., Comparison of Widefield/Deconvolution and Confocal Microscopy for Three-Dimensional Imaging, in Handbook Of Biological Confocal Microscopy, J.B. Pawley, J.B. Pawley^Editors. 2006, Springer US: Boston, MA. p. 453-467.

[18]. Elliott, A.D., Confocal Microscopy: Principles and Modern Practices. Curr Protoc Cytom, 2020. 92(1): p. e68.

[19]. Voigt, F.F., et al., Reflective multi-immersion microscope objectives inspired by the Schmidt telescope. Nat Biotechnol, 2024. 42(1): p. 65-71.

[20]. Melanthota, S.K., et al., Deep learning-based image processing in optical microscopy. Biophysical Reviews, 2022. 14(2): p. 463-481.

437   [21]. Mandracchia, B., et al., Optimal sparsity allows reliable system-aware restoration of fluorescence

438   microscopy images. Sci Adv, 2023. 9(35): p. eadg9245.

439   [22]. Li, X., et al., Spatial redundancy transformer for self-supervised fluorescence image denoising.

440   Nat Comput Sci, 2023. 3(12): p. 1067-1080.

441