Neural Mechanism Underlying Successful Classification of Amnestic Mild Cognitive Impairment Using Multi-Sensory-Evoked Potentials

Lei Zhang[1], Malcom Binns[1,2], Ricky Chow[1], Rahel Rabi[1], Nicole D. Anderson[1,3,6], Jing Lu[4,5], Morris Freedman[1,7,8,9], and Claude Alain[1,3,10,11]

[1]Rotman Research Institute, Baycrest Academy for Research and Education, Ontario, Canada, M6A 2E1

[2]Dalla Lana School of Public Health, University of Toronto, Ontario, Canada, M5T 3M7

[3]Department of Psychology, University of Toronto, Ontario, Canada, M5G 3G3

[4]MOE Key Lab for Neuroinformation, School of Life Science and Technology, University of Electronic Science and Technology of China, Chengdu, China, 611731

[5]Center for Information in Medicine, School of Life Science and Technology, University of Electronic Science and Technology of China, Chengdu, China, 611731

[6]Department of Psychiatry, University of Toronto, Ontario, Canada, M5T 1R8

[7]Department of Medicine (Neurology), Baycrest Centre, Ontario, Canada, M6A 2E1

[8]Department of Medicine (Neurology), Mount Sinai Hospital, Ontario, Canada, M5G 1X5

[9]Department of Medicine (Neurology), University of Toronto, Ontario, Canada, M5S 3H2

[10]Institute of Medical Science, University of Toronto, Ontario, Canada, M5S 1A8

[11]Music and Health Science Research Collaboratory, University of Toronto, Ontario, Canada, M5S 2C3

Correspondance: Claude Alain, Ph.D.
                       Rotman Research Institute
                       Baycrest Academy for Research and Education
                       3560 Bathurst Street
                       Toronto, Ontario, M6A 2E1
                       Canada
                       (416) 785-2500 ext. 3523 (telephone)
                       (416) 785-2862 (fax)
                       Email: calain@research.baycrest.org

                       Lei Zhang, Ph.D.
                       Rotman Research Institute
                       3560 Bathurst Street
                       Toronto, Ontario, M6A 2E1
                       Canada
                       (416) 785-2500 ext. 2103 (telephone)
                       Email: lzhang@research.baycrest.org

Abstract

Introduction

The diagnosis, prognosis, and management of amnestic mild cognitive impairment (aMCI)

remains challenging. Early detection of aMCI is crucial for timely interventions.

Method

This study combines scalp recordings of auditory, visual, and somatosensory stimuli with a

flexible and interpretable support vector machine classification pipeline to differentiate

individuals diagnosed with aMCI from healthy controls.

Results

Event-related potentials (ERPs) and functional connectivity (FC) matrices from each modality

successfully predicted aMCI. We got optimal classification accuracy (96.1%), sensitivity (97.7%)

and specificity (94.3%) when combining information from all sensory conditions than when

using information from a single modality. Reduced ERP amplitude, higher FC in frontal region

which predicted worse cognitive performance, and lower FC in posterior regions from delta to

alpha frequency in aMCI contributed to classification.

Conclusions

The results highlight the clinical potential of sensory-evoked potentials in detecting aMCI, with

optimal classification using both amplitude and oscillatory-based FC measures from multiple

modalities.

Background

Amnestic mild cognitive impairment (aMCI) represents an intermediate stage between healthy aging and dementia, typically due to Alzheimer's disease (AD), the most common cause of dementia in the aging population. Individuals with aMCI show impairment in episodic memory and sometimes in additional cognitive domains such as executive functioning and language,[1-3] with preserved functional independence in daily activities.[4] The diagnosis, prognosis, and management of aMCI remain challenging. Timely identification and intervention in aMCI are pivotal for its effective clinical management. The present study contributes to this critical need by introducing a cutting-edge framework that integrates the power of support vector machine (SVM) algorithm with a novel multisensory evoked potential (MSEP) paradigm.

Scalp-recording of event-related potentials (ERP) may help identify individuals with aMCI. Relative to healthy controls, individuals with MCI (including amnestic and non-amnestic subtypes that are prodromal to all-cause dementia) show significant reductions in N2 and P3 amplitude, as well as delayed P3 latency, signifying deficits in inhibitory control.[5-12] Multivariate classification of cognitive ERPs has been shown to correctly discriminate participants with MCI from healthy controls with an accuracy ranging from 70% to 80% depending on the task.[13-15] However, Cognitive ERP recordings can be time-consuming due to variability in amplitude and latency, requiring numerous trials for reliable signal-to-noise ratios. Additionally, cultural factors like language, education, and literacy can influence performance on cognitive ERP paradigms.[16]

To overcome these limitations, we developed the MSEP paradigm,[17] which provides a means to delineate normal from impaired sensory processing. In this paradigm, lateralized

auditory, visual, and somatosensory stimuli are presented sequentially at a high rate (e.g., 3-5

stimuli per second) and in random order. This approach minimizes neural adaptation, thereby

ensuring reliable and fairly quick (~15 minutes) measurements of sensory-evoked responses

from three sensory modalities. These three measurements may then have potential for

differentiating individuals with aMCI from healthy older adults. Because these sensory-evoked

responses are recorded while participants sit quietly in the absence of a task (i.e., no response

required), there are no concerns related to task complexity, or difficulty understanding complex

instructions, making it easy to administer to older adults. Therefore, the design of this paradigm

provides practicality in clinical settings as this paradigm is short in length, accessible and

inclusive for older adults, and can provide reliable measures of brain functioning.

The MSEP paradigm has been developed in light of prior ERP studies using either

auditory,[5,18,19] visual[20-22] or somatosensory[23,24] stimuli, which have revealed aberrant

(hypersensitive and/or delayed responses) sensory-evoked responses in older adults. Sensory-

evoked response amplitude may be linked to widespread structural changes in prefrontal

regions of individuals with MCI.[25-35] Notably, lesions to prefrontal cortices have been associated

with an increase in amplitude of auditory, visual and somatosensory ERPs, which was attributed

to deficits in inhibitory control.[36,37] Moreover, sensory evoked response amplitude have been

associated with deficits in memory tasks.[38-40] Given these findings from studies measuring

sensory evoked response amplitude from a single sensory modality, we foresee that using

measurements from auditory, visual and somatosensory systems could significantly enhance

the sensitivity of EEG measures in detecting aMCI.

Evidence suggests that SVM algorithm can successfully identify older adults with MCI

(including amnestic and non-amnestic subtypes) from healthy controls with accuracy ranging

from 75% to 98% using speech-evoked responses[41] or resting-state EEG[42-44] depending on the

number and type of EEG features analyzed. A recent study applying parameter optimized SVM

classifiers to oscillatory-based functional connectivity data showed hearing status could be

decoded during a speech categorization task solely using network-level descriptions.[45]

Extending prior work,[45] we used adjusted connectome-based predictive modeling with SVM

classifiers to distinguish older adults with a diagnosis of aMCI from healthy controls using EEG

functional connectivity measures during processing of auditory, visual, and somatosensory

stimuli. Given prior findings of aberrant sensory evoked responses in single sensory modalities,

we hypothesize that correct classification of aMCI will increase when combining information for

multiple sensory modalities compared to using information from a single sensory modality.

Methods

Participants

Participants were recruited if they were native English speakers or learned English

before the age of five, had normal or corrected-to-normal vision (e.g., no history of

degenerative conditions, glaucoma, cataracts significant enough to impede vision, or color

blindness), and reported no significant hearing loss, no history of learning disabilities, stroke,

transient ischemic attack, traumatic brain injury with loss of consciousness greater than five

minutes, substance abuse disorder, history of intracranial surgery, and any other diagnosis of

major neurological or psychiatric disorder. Participants were excluded if they had a history of

myocardial infarction, coronary artery disease, or bypass surgery. Participants were also

excluded if they were prescribed hearing aids, taking medication known to possibly affect

cognitive functioning including antidepressants, anticonvulsants, antipsychotics, or consuming

recreational drugs either concurrently or within the year prior to testing. An inclusion criterion

for older adults was scoring above the cut-off on the Telephone Interview for Cognitive Status-

Modified (TICS-M).[46] Our final sample included 44 individuals with aMCI (62-88 years, 23

males), and 35 healthy older adult controls (60-88 years, 20 males). Participants were recruited

from the Rotman Research Institute participant database. The study protocol was approved by

the Research Ethics Board of the Rotman Research Institute at Baycrest. All participants

provided informed written consent, and were compensated a modest stipend for their

participation and out-of-pocket expenses incurred (e.g., parking).

Neuropsychological assessment

Preceding the recording of neuroelectric activity, participants were administered a

battery of standardized neuropsychological tests in the domains of intellectual functioning,

memory, language, processing speed, and executive functioning. The neuropsychological

assessment took place on a separate day from measurements of sensory-evoked responses to

prevent fatigue effects. Diagnosis of single-domain or multiple-domain aMCI was verified by a

registered neuropsychologist. A full description of the administered neuropsychological

assessments have been previously reported in Chow et al[47] and Rabi et al.[48] Demographic,

neuropsychological, and clinical data for individuals with aMCI and age-matched controls are

displayed in Table 1. The aMCI and control groups did not statistically differ in age, sex, or years

of education.

Stimuli and multi-sensory evoked paradigm

The visual stimulus was a 5 by 10 square black and white checkerboard pattern

presented at 50ms in duration that spanned the entire height and approximately 40% of the

width of a 19-inch computer monitor that extended from either the left or right edge of the

monitor. The entire checkerboard stimulus subtended a visual angle of 14.3° by 28.1°, with

each checkerboard square subtending a visual angle of 2.9° by 2.9°. The auditory stimulus was a

monoaural harmonic complex tone (50 ms in duration, 5 ms rise/fall time). It comprised five

tonal elements (200, 400, 600, 800, and 1000 Hz) and had a fundamental frequency of 200 Hz.

The tone was presented to the left or right ear through ER-3A insert earphones (Etymotic

Research Inc., Elk Grove, IL, USA). The sound intensity was set at 85 decibels (dB) sound

pressure level (SPL) as measured by a Larson Davis 824 SPL meter using a 2cc coupler. The

somatosensory stimulus consisted of pneumatic tactile stimulation delivered simultaneously to

the tips of all four fingers (thumb excluded) of the left or right hand. The somatosensory

stimulation was delivered from a puff-type pressure stimulator from a compressed air tank with

electromagnetic air valves kept constant at 30 psi through plastic tubes to 8 pneumatically

driven inflatable circular plastic membranes with a diameter of 1 cm each. Each membrane was

worn on the participant's fingers at the distal interphalangeal joint. During somatosensory

stimulation, the valves were activated to periodically inflate the membrane and deliver brief

pressure pulses of 50 ms duration to the fingertips.

Procedure

To measure sensory-evoked potentials, we used a continuous, rapid serial event-related

paradigm in which lateralized visual, auditory, and somatosensory stimuli were presented

sequentially in a randomized order. The paradigm was administered with participants seated in

a sound-attenuated booth. Before recording sensory evoked potentials, participants were

presented with a 10-second sample of the paradigm to ensure that they could clearly perceive

the auditory, visual, and somatosensory stimulation from both left and right hemispace. During

the stimulus presentation, participants were instructed to gaze on a fixation dot 0.60 cm in

diameter at the center of a 19-inch computer screen set at 100% contrast level and placed 60

cm from participants. The stimulus onset asynchrony varied between 200 and 340 ms (20 ms

steps, rectangular distribution). The stimulus presentation lasted approximately 12 minutes,

with a total of 1440 stimuli (480 trials per sensory modality) evenly distributed between left

and right laterality. Stimuli were presented using Presentation software (version 13,

Neurobehavioral Systems, Albany, CA).

EEG Acquisition and Preprocessing

Neuroelectric brain activity was recorded continuously using a 76-channel BioSemi

ActiveTwo acquisition system (BioSemi, Amsterdam, The Netherlands) with a sampling rate of

512 Hz. Sixty-six electrodes were positioned on the scalp using a BioSemi headcap according to

the standard 10-20 system with a pair of Common Mode Sense active electrode and Driven

Right Leg passive electrode serving as ground. Ten additional electrodes were placed below the

hairline (i.e., mastoids, pre-auricular points, two lateral ocular sites, two inferior ocular sites,

and two additional frontal-lateral electrodes) to monitor eye movements and cover the whole

scalp evenly. EEG recordings were preprocessed offline using FieldTrip toolbox[49] and custom

scripts.

The EEG data were first re-referenced to the average of all electrodes. Then,

independent component analysis was performed to remove the artifacts caused by eye blinks,

eye movements, and heartbeat. Next, the EEG data were bandpass filtered from 1 to 40Hz.

Filtered data were parsed into 500 ms epochs time-locked to stimulus onset, including 100 ms

of pre-stimulus activity and baseline corrected using pre-stimulus interval (-100 to -20 ms).

Lastly, the preprocessed data were visually inspected to identify trials contaminated with

artifacts and noisy electrodes; these trials were then rejected from subsequent analysis. Noisy

electrodes were interpolated, with no more than 10% of the channels interpolated per

participant.

We did not have a priori hypotheses regarding possible stimulus lateralization. Hence, as

a dimension reduction method and to improve the signal-to-noise ratio, epochs of lateralized

sensory-evoked potentials were averaged together in a transposed montage such that right-

lateralized responses (elicited by left-lateralized stimuli) were kept untransposed, and left-

lateralized evoked potentials (elicited by right-lateralized stimuli) were transposed across

laterality. This transposition method allowed us to combine the evoked responses from

lateralized stimuli, thereby significantly increasing the number of trials per modality.

Electrode-level functional connectivity matrix construction

A surface Laplacian method was applied to reduce the effect of volume conduction on

functional connectivity estimates.[50] The surface Laplacian is a spatial filter that computes the

second spatial derivative of the scalp potential, which provides a measure of the local current

source density. Then, for each sensory modality, phase locking value (PLV), which estimated the

phase synchronization of two signals, was calculated for estimating the functional connectivity

between pairs of electrodes across 12 frequency bands ranging from 2.5 Hz to 30 Hz, with each

band being 2.5 Hz in width (i.e., 2.5 Hz, 5 Hz, 7.5 Hz … 27.5 Hz, 30 Hz). Consequently, for each

participant, twelve 72 by 72 functional connectivity matrices were created for each stimulus

modality per participant.

Adjusted connectome-based predictive modeling (CPM)

We adjusted the procedure from CPM[51] and used it with SVM learning to classify each

participant as belonging to either the aMCI group or control group (Fig. 1). Functional

connectivity matrices of each modality and frequency band were separately entered into the

classification analysis. Leave-one-subject-out (LOSO) cross-validation was employed to evaluate

classification performance (i.e., classification accuracy) for each electrode under each sensory

condition. Specifically, for each cross-validation fold, matrices from 78 participants were used

to build the model and the matrix from the left-out participant was used to test model

performance. For each matrix, we first performed a two-sample independent t-test (two-tailed)

on each edge between aMCI and healthy control groups using the training set. Edges that

showed sufficient group difference ($p_{uncorrected}$ < 0.05) were selected for training the SVM

classification model with a linear kernel. A parameter search was performed using training data

to optimize the cost value, which balanced the margin and training error, for each model and

each stimulus modality (thereby maximizing the average accuracy of the top three frequency

bands over the range $2^{-6}, 2^{-5}, ..., 2^{5}$). After training the model, the edges from the left out

participant were used to test the model. With the LOSO cross validation procedure, every

participant would be predicted to be in the aMCI or control group using the model trained by

the selected functional connectivity edges of all other participants. Classification accuracy was

calculated by averaging the accuracy across all folds. SVM classification analysis was conducted

using the LIBSVM package.[52] The selected edge masks in each fold were kept for visualizing the

important edges that contributed to the classification.

We performed permutation testing to determine the statistical significance of the

classification accuracy of each frequency band for each sensory modality. The labels of the

participants were shuffled and used to perform the classification analysis described above. The

number of permutations was set at 2000, which allowed us to separately construct the null

distributions of classification accuracy for each frequency band under three sensory modalities.

Permutation-based statistics were calculated based on the following formula:

$$\frac{accuracy_{real} - mean(accuracy_{permutation})}{standard\ deviation\ (accuracy_{permutation})}$$

The frequency-level permutation-based statistics was followed by cluster-based

permutation tests to correct for multiple comparisons across frequency bands. The sums of the

permutation-based statistics of contiguous frequencies that were significant ($p_{permutation}$ <

0.05) were computed. Then, 2000 permutations were conducted and the largest permutation-

based statistics-mass for each permutation was extracted to get the null distribution of the

cluster level mass. Significance was set at a family-wise corrected p-value of p_{fwe} < 0.05 cluster-

mass corrected.

SVM classification with sensory-evoked response amplitude

Besides functional connectivity matrices, we also performed SVM classification analysis

using the amplitude of sensory-evoked potentials (Fig. 1). Averaged amplitudes for each

modality were separately entered in the classification analysis. For each electrode, classifiers

were trained using SVM algorithm with a linear kernel separately for the three modalities.

Similarly, LOSO cross-validation was employed to evaluate classification performance (i.e.,

classification accuracy) for each electrode in each sensory modality. A parameter search was

also performed using training data to optimize the cost value for each model and each stimulus

modality (thereby maximizing the average accuracy of top ten electrodes over the range

$2^{-3}, 2^{-2}, \ldots, 2^5$). We used cluster-based permutation testing to determine the statistical

significance of the classification accuracy of each electrode for each modality. Channel

neighbors were defined by the FieldTrip function "ft_prepare_neighbours" with the

"triangulation" method.[49] The sums of the permutation-based statistics of contiguous channels

that were significant ($p_{permutation}$ < 0.05) were computed. Then, 2000 permutations were

conducted and the largest permutation-based statistics-mass for each permutation was

extracted to get the null distribution of the cluster-level mass. Significance was set at p_{fwe} <

0.05 cluster-mass corrected.

Voting method for classification combining multiple modalities

We used voting strategies to combine the classification results from different modalities

and indices. The frequency bands or electrodes that showed highest classification accuracy in

each sensory modality were selected and entered in the voting procedure. We employed two

strategies in the voting procedure: majority rule or one-vote veto rule. The majority rule

denotes that whether a given participant is classified as belonging to the aMCI or control group

is determined by most supporters (i.e., a participant was classified as belonging to one group if

they were also identified as belonging to the same group in two or all three modalities). As for

the one-vote veto rule, a given participant was classified as belonging to the aMCI or control

group only when all involved classifiers agree on aMCI status. The left-out participants were

labelled as undetermined participants and were not included in the accuracy calculation.

Group comparison of evoked response from three modalities

Averaged sensory-evoked responses in the significant clusters under each modality were

entered into the group comparison analysis between aMCI and control groups. Note that the

participants with more than one channel that misclassified the participants were not included

in the group comparison since they could not be classified with high consistency across

channels. Independent-sample *t*-tests were performed to compare the group-averaged

sensory-evoked amplitudes among significant clusters between aMCI and controls. Cluster-

based permutation testing was performed to find clusters that showed significant group

difference of sensory-evoked response amplitude between the two groups.

Brain-behavior correlation

Linear regression models were fit to examine the relationship between neural indices

(functional connectivity estimates of the important edges, sensory-evoked response amplitude

of significant clusters) and composite scores from the neuropsychological assessment (Table 1).

Group variable (aMCI/healthy older adults) was set as the control variable in the linear

regression models.

Results

Classification with unisensory prediction models

Firstly, with the adjusted CPM approach, we trained the SVM learning classifiers to

predict whether a given participant belonged to the aMCI or control group separately with

unisensory functional connectivity matrices (visual, auditory, or somatosensory). We found

frequency clusters that showed significantly greater classification accuracy than chance for all

three sensory modalities (Fig. 2A). For visual stimuli, functional connectivity matrices from 2.5

Hz to 10 Hz and from 15 Hz to 17.5 Hz significantly predicted correct classification (i.e., whether

a participant belonged to the aMCI or healthy control group, 2.5hz-10hz: $p_{fwe} = 0.007$, 15hz-

17.5hz: $p_{fwe} = 0.040$). For auditory stimuli, the significant frequency bands were from 2.5 Hz

to 7.5 Hz and 12.5 Hz (2.5hz-7.5hz: $p_{fwe} < 0.001$, 12.5 hz: $p_{fwe} < 0.001$). For somatosensory

stimuli, the significant frequency bands were from 2.5 Hz to 12.5 Hz ($p_{fwe} < 0.001$).

Figure 2A shows classification accuracy per frequency band. Classification accuracies

tended to be higher for lower frequencies (i.e., delta, theta, and alpha) in all three modalities,

with the highest classification accuracy at 5 Hz (Fig. 2D, 69.6% for visual stimuli, $p_{permutation} =$

0.004, 89.9% for auditory stimuli, $p_{permutation} < 0.001$, 92.4% for somatosensory stimuli,

$p_{permutation} < 0.001$). The sensitivities of all three modalities were high, while the

specificities were relatively low (Fig. 2D, for visual stimuli, 79.5% vs. 57.1%; for auditory stimuli,

95.5% vs. 82.9%; for somatosensory stimuli, 95.5% vs. 88.6%). In other words, the fitted model

seemed to show good detection of participants with aMCI (i.e., better able to detect true

positives) but relatively poor detection of healthy older adults (i.e., resulting in a few false

positives).

We employed a voting method to combine the predictive power from three modalities.

With majority rule, the accuracy, sensitivity, and specificity of the model was similar to the

somatosensory modality, which showed the best performance (i.e., Fig. 2E, Accuracy: 91.1%;

Sensitivity: 95.5%; Specificity: 85.3%). The classification accuracy was calculated based on the

subset of participants for whom prediction from all involved classifiers were consistent.

Therefore, there exists accuracy-proportion of predictable participant trade-off with one-vote

veto rule. We found a slight increase of the performance compared to somatosensory classifier,

but only 68.4% of participants could be accurately identified (Fig. 2E, F, Accuracy: 94.4%;

Sensitivity: 100%; Specificity: 85.7%). With a voting procedure that combined three modalities,

we found that specificity was still relatively low.

Functional connectivity strength predicted cognitive performance

We extracted all of the masks used for each fold. Since LOSO cross-validation was used

in SVM classification, 79 masks were generated in the analysis. Masks were slightly different for

each fold because training sets varied for each fold. We kept the edges that were preserved in

most folds (> 90%), which were more important for the classification analysis than the rest

edges.

We divided the electrode montage into five regions of interest (ROIs): frontal

(frontopolar, anterior-frontal and frontal channels), temporal (frontotemporal and temporal-

parietal channels), central (frontocentral and central channels), parietal (centroparietal and

parietal channels), and occipital (parietal-occipital and occipital channels). Then, we

summarized the positive and negative edge numbers that were kept according to the five ROIs

separately for each sensory modality and significant frequency band. Note that, positive edges

refer to the edges where aMCI showed higher functional connectivity estimates than healthy

older adults, while negative edges refer to the edges where aMCI showed lower functional

connectivity estimates.

Within lower frequency bands (i.e., delta, theta, and alpha bands), we found similar

pattern of connectivity for the three modalities (Fig. 2B and 2C). That is, the aMCI group

showed stronger functional connectivity than healthy older adults within frontal electrodes, as

well as connectivity from frontal and central electrodes to posterior electrodes. Meanwhile,

within beta band, we also found that aMCI showed greater functional connectivity than

controls across the whole brain with visual stimuli. Besides the increased connectivity, we also

observed weaker edges in the aMCI group that connected posterior regions like occipital,

temporal, and parietal areas within lower frequency bands. Within beta band, we found

increased frontal-central and temporal-parietal connectivity in the aMCI group.

Linear regression analysis was performed to investigate the relationship between

functional connectivity of the important edges and cognitive performance from

neuropsychological assessments. We separately summed the positive or negative important

edges for three modalities as the independent variable in the linear regression analysis. Group

variable was inputted to the model as the control variable. We found that functional

connectivity that was higher in the aMCI group than healthy older adults predicted worse

cognitive performance within both frequency bands (Table2, EF: Delta-Theta-alpha band: Visual:

β = -4.37, p = 0.005; Somatosensory: β = -3.60, p = 0.017; Beta band: Visual: β = -5.39, p =

0.010). The higher functional connectivity in aMCI may reflect a loss of complexity in EEG

dynamics, which has been associated with difficulties to adapt to new situations and could be a

potential biomarker for aMCI diagnosis.[53-55] However, for functional connectivity that was lower

in the aMCI group, we observed different results. Within beta band, weaker functional

connectivity between frontal and posterior areas predicted better performance on tests of

language (Table2, Language: β = -5.92, p = 0.006).

Classification with unisensory evoked response amplitude

We also trained the SVM classifiers to predict whether the participant belonged to the

aMCI or control group using the amplitude of sensory evoked responses (visual, auditory, or

somatosensory stimuli). Compared to a null distribution of classification accuracy generated by

permutation testing, we found significant clusters that showed significantly greater

classification accuracy than null distribution under all three sensory modalities conditions (Fig.

3A-C). With visual stimuli, we found clusters that showed significant classification accuracy in

left central-frontal areas (FP1, AF7, AF3, F3, FC3, $p_{fwe} = 0.003$) and right parietal areas (CP6,

CP4, P6, CP2, P4, $p_{fwe} = 0.001$). With auditory stimuli, we found significant clusters in

bilateral central areas (FC5, FC3, C1, Cz, C2, $p_{fwe} < 0.001$), left temporal-parietal areas (P7,

PO7, PO9, TP9, $p_{fwe} = 0.020$, and right parietal areas (P6, P8, PO8, PO4, O2, $p_{fwe} < 0.001$).

With somatosensory stimuli, we found significant cluster in right occipital areas (Iz, O2,

$p_{fwe} = 0.040$).

As shown in Fig. 2D, with three modality stimuli, electrodes that showed highest

classification accuracy were for visual stimuli (P8, classification accuracy = 79.7%; sensitivity =

81.8%; specificity = 77.1%), with auditory stimuli (P4, classification accuracy = 76.0%; sensitivity

= 88.6%; specificity = 60.0%), and with somatosensory stimuli (Iz, classification accuracy = 79.7%;

sensitivity = 84.1%; specificity = 74.3%). Again, similar to the model trained by functional

connectivity matrices, specificity of the model is lower than the sensitivity.

Models trained with unisensory evoked response amplitude were not as good as models

trained with unisensory functional connectivity. However, as for EEG connectivity, model

performance increased significantly after combining three modalities using the voting

procedure. With majority rule among classifiers of all three modalities, classification accuracy

increased to 87.3% (Fig. 3E, accuracy = 87.3%, sensitivity = 93.2%, specificity = 80.0%). With

one-vote veto rule, the model performance slightly increased compared to the model trained

by functional connectivity. Classification accuracy increased to 97.5% (Fig. 3E-F, sensitivity =

100%, specificity = 92.3%) at the expense that only 50.6% of participants could be determined.

Reduced sensory evoked response amplitude in aMCI

To unravel what contributes to the predictability of the SVM classifier trained by

averaged evoked responses from three sensory modalities, we compared the averaged sensory-

evoked response amplitude among the significant clusters between the aMCI group and

healthy older adults. Only clusters that contained no more than one electrode that generated

false prediction were included.

With visual stimuli, we found a larger early positive component in healthy older adults

than those with aMCI (Fig. 3G, from -28ms to 50ms, $p_{fwe} = 0.003$) over the frontocentral

region ipsilateral to stimulus presentation. However, we did not find a significant difference in

contralateral parietal areas. With auditory stimuli, we found that older adults showed greater

amplitude in positive and negative waveforms peaking at about 200 and 300 ms after sound

onset, referred to as P2 and N3, respectively (Fig. 3H, P2: from 171ms to 238ms, $p_{fwe} = 0.004$;

N3: from 294ms to 338ms, $p_{fwe} = 0.004$) in central cluster, greater P2 amplitude (Fig. 3I, from

172ms to 249ms, $p_{fwe} = 0.002$) in a left temporal-parietal cluster, and greater P2 amplitude

(Fig. 3J, from 238ms to 283ms, $p_{fwe} = 0.016$) in right parietal cluster. With somatosensory

stimuli, we failed to find any difference between two groups in lateral occipital regions.

Next, we employed linear regression to investigate the relationship between significant

clusters and cognitive performance in older adults. After controlling for group as the control

variable, we found that a greater N3 amplitude of the auditory evoked response in central

regions significantly predicted higher speed composite scores ($\beta_{AEP} = -0.36, p = 0.029$).

Combining classification models trained by functional connectivity and ERPs

Since both models that were trained by connectivity measures and ERP amplitude after

combining three modalities and the specificity were both relatively low with majority rule, we

combined the six models (3 modalities × 2 neural indices) with voting procedure. With one-vote

veto rule, only 35.4% of participants could be determined and therefore the model showed

little utility. With majority rule, classification accuracy was 96.1%, and we got balanced

sensitivity and specificity (Fig. 4A-B, sensitivity = 97.7%, specificity = 94.3%).

Discussion

In this study, we employed the SVM algorithm to classify individuals with aMCI and healthy

older adults with electrode-wise functional connectivity matrices and sensory-evoked

responses from auditory, visual, and somatosensory systems. We found that healthy older

adults and those with aMCI could be reliably classified using neural indices for each sensory

modality alone. However, we observed a significant increase in classification accuracy after

combining both neural indices under all three sensory conditions with the majority rule voting

procedure, which yielded optimal classification accuracy (96.1%), sensitivity (97.7%), and specificity (94.3%). With adjusted CPM analysis, we identified increased functional connectivity within frontal and central regions in a lower frequency band and whole-brain functional connectivity in the beta band, which were the major contributing features in classification. Meanwhile, a group comparison of sensory-evoked response amplitude between correctly classified participants with aMCI and healthy older adults showed smaller response amplitude in the aMCI group than healthy older adults at central, frontal, parietal, and temporal electrodes. Notably, these increases and reductions in connectivity were correlated with better cognitive performance, namely executive functioning and speed, which suggest these features could be sensitive to cognitive impairment that is often associated with aMCI.

Since individuals with aMCI often experience difficulty with complex decision-making, experimental paradigms that are easy, fast, and require low cognitive demands are often used in clinical assessment of aMCI. Hence, investigating resting-state brain activity, which is simple and task-free, is popular in studies aiming to develop biomarkers for distinguishing aMCI from healthy controls.[56,57] However, recent studies found that task-induced functional connectivity may capture more behaviour-related information than resting-state functional connectivity, even if the tasks during scanning are not directly relevant to the behavioral tasks.[58-60] For example, Finn and Bandettini[59] found that functional connectivity during movie watching could better predict watchers' trait-like phenotypes in both cognition and emotion domains than resting-state connectivity. In the current study, we designed a paradigm for the passive perception of simple stimuli from three different modalities. Our MSEP paradigm is as simple, fast, and low in cognitive demand as the resting-state paradigm. Besides the well-performed

classification, we also found robust correlations of sensory-induced functional connectivity to

cognitive performance. These results support that sensory-induced neural activities could

capture some changes in cognitive performance associated with aMCI, even though these

cognitive functions are irrelevant to the perceptual stimuli per se.

Neural indices from all three modalities provide information for distinguishing between

older adults with and without a diagnosis of aMCI. However, the amount of information varies

between modalities, leading to different classification accuracies ranging from 69.6% to 92.4%

(sensitivity: from 79.5% to 95.5%; specificity: from 57.1% to 88.6%). This is consistent with

previous finding that neural responses to single modal stimuli could be used to classify aMCI

from healthy controls.[41,61-63] With our MSEP paradigm, we could obtain neural responses to

different modalities in a single recording session, which enables us to combine informative

component from different modalities for a better classification performance. Previous studies

merging features from different modalities or sources usually combine features in the feature

selection and model training stage and predict samples using the combined features. [41,42,61] The

present study combines different features (i.e., amplitude of the evoked response, functional

connectivity) under three sensory conditions in the prediction output session. With our strategy,

model prediction performance markedly improved in accuracy, sensitivity, and specificity

(accuracy = 96.1%, sensitivity = 97.7%, specificity = 94.3%).

The first advantage of our pipeline is the high flexibility. In the model training and test

stages, the SVM algorithm could be replaced with any other machine learning or deep learning

algorithms such as logistic regression or artificial neural networks. Then, permutation testing

and the voting procedure could be performed in the same way as this study. Furthermore, the

voting procedure is a highly flexible method to combine information from different features

because it does not require merging different features before or during the model training and

testing phases. Therefore, any model or pipeline that produces a prediction label for each

sample will apply to the voting procedure. This feature provides the advantage that different

neural indices could be trained and tested using different algorithms since the optimal

algorithm is not necessarily the same across different neural indices due to their divergent data

characteristics and task types.[64]

One of the main challenges in applying machine learning methods to psychological or

medical research is interpretability.[65] Another advantage of our analysis pipeline is providing

insight into the neural mechanisms underlying successful classification. The adjusted CPM

provides important edges that contribute to successful classification. A group comparison of

neural indices among the correctly classified samples identifies significant components that

distinguish the two groups. Therefore, these extracted components enable brain-behavior

analyses to explore how changes in functional connectivity seen in aMCI are associated with

cognitive impairment in this population.

Using PLV, we found that aMCI was associated with greater functional connectivity within

central-frontal electrodes and greater functional connectivity from central-frontal electrodes to

temporal electrodes including parietal, temporal, and occipital electrodes in the theta to alpha

frequency bands. Meanwhile, older adults diagnosed with aMCI exhibited greater whole-brain

functional connectivity in the beta band. Our findings complement prior research on resting-

state functional connectivity in MCI, showing hypo- and hyper-connectivity in MCI compared to

controls.[56] A recent meta-analysis of functional near infrared spectroscopy studies showed that

aMCI was associated with weaker functional connectivity between prefrontal and posterior

parietal regions during resting state than healthy older adults.[66] However, for task-induced

functional connectivity, older adults with aMCI showed stronger connectivity between medial

left frontal cortex and temporal regions compared to healthy controls during tasks of working

memory and naming.[66] These opposing results may reflect different constructs of brain

functioning captured by resting-state functional connectivity and task-induced functional

connectivity. Weaker functional connectivity at rest may suggest that aMCI-related deficits in

the stability of intrinsic organization, while higher functional connectivity may suggest a loss of

neural complexity in aMCI in service of sensory perception and goal-oriented tasks. Our

correlation analyses of the most important edges to cognitive tasks also supports that the

increased functional connectivity was correlated to impaired cognitive function in the aMCI

group.

Besides functional connectivity, we also found that sensory-evoked response amplitude

under all three sensory conditions are significant predictors of aMCI status. By comparing group

differences in sensory-evoked responses, we found reduced amplitude at various latencies in all

three sensory modalities. These results are consistent with previous studies investigating

auditory- and visual-evoked response alterations in older adults diagnosed with MCI and early

Alzheimer's disease compared to healthy controls.[67-70]

Previous EEG studies found that older adults displayed enhanced early sensory evoked

response amplitude compared to young adults, which has been attributed to reflecting changes

in sensory regulation mediated by the PFC.[17,36] The decreased amplitude of sensory-evoked

responses in individuals with aMCI may also reflect further deficits in frontal regulation of

sensory responses attributed to early Alzheimer's disease-related degeneration of frontal and

temporal areas. This result also is consistent with deficits in frontal functioning in aMCI, such as

inhibition,[2] working memory,[71] and executive function.[72]

In this study, we introduced the novel MSEP paradigm that is suitable for the aMCI

population and a classification analysis pipeline that enables researchers to flexibly combine

features from different modalities and further explore the neural mechanism underlying the

classification performance. Here, we employed the MSEP paradigm and classified older adults

with aMCI from healthy older adults. Classification performance increased substantially after

combining features from evoked potentials of multiple sensory modalities. Additionally, we

found that the reduced sensory ERP amplitudes and increased functional connectivity in

individuals with aMCI were the main contributors to the successful classification. These findings

provide new insights into how EEG may be used to support the diagnosis and identification of

neural biomarkers for aMCI and potentially other neurological disorders. Until now, past

research has focused on examining sensory-perceptual processing in a single modality. Our

study is the first to combine modalities providing a more comprehensive picture of sensory-

processing deficits which improves classification accuracy of aMCI.

References

1. Kramer JH, Nelson A, Johnson JK, et al. Multiple cognitive deficits in amnestic mild
cognitive impairment. *Dement Geriatr Cogn Disord*. 2006;22(4):306-11. doi:10.1159/000095303
2. Rabi R, Vasquez BP, Alain C, Hasher L, Belleville S, Anderson ND. Inhibitory Control
Deficits in Individuals with Amnestic Mild Cognitive Impairment: a Meta-Analysis. *Neuropsychol*
Rev. Mar 2020;30(1):97-125. doi:10.1007/s11065-020-09428-6
3. Ribeiro F, de Mendonca A, Guerreiro M. Mild cognitive impairment: deficits in cognitive
domains other than memory. *Dement Geriatr Cogn Disord*. 2006;21(5-6):284-90.
doi:10.1159/000091435
4. Petersen RC. Mild Cognitive Impairment. *Continuum (Minneap Minn)*. Apr 2016;22(2
Dementia):404-18. doi:10.1212/CON.0000000000000313
5. Golob EJ, Irimajiri R, Starr A. Auditory cortical activity in amnestic mild cognitive
impairment: relationship to subtype and conversion to dementia. *Brain*. Mar 2007;130(Pt
3):740-52. doi:awl375 [pii] 10.1093/brain/awl375

6. Golob EJ, Johnson JK, Starr A. Auditory event-related potentials during target detection
are abnormal in mild cognitive impairment. Clinical Trial Research Support, U.S. Gov't, P.H.S.
Clinical neurophysiology : official journal of the International Federation of Clinical
Neurophysiology. Jan 2002;113(1):151-61.

7. Bennys K, Portet F, Touchon J, Rondouin G. Diagnostic value of event-related evoked
potentials N200 and P300 subcomponents in early diagnosis of Alzheimer's disease and mild
cognitive impairment. *Journal of clinical neurophysiology : official publication of the American*
Electroencephalographic Society. Oct 2007;24(5):405-12. doi:10.1097/WNP.0b013e31815068d5
8. Cespon J, Galdo-Alvarez S, Diaz F. Electrophysiological correlates of amnestic mild
cognitive impairment in a simon task. Research Support, Non-U.S. Gov't. *PLoS One*.
2013;8(12):e81506. doi:10.1371/journal.pone.0081506
9. Cespon J, Galdo-Alvarez S, Diaz F. Inhibition deficit in the spatial tendency of the
response in multiple-domain amnestic mild cognitive impairment. An event-related potential
study. *Front Aging Neurosci*. 2015;7:68. doi:10.3389/fnagi.2015.00068
10. Papadaniil CD, Kosmidou VE, Tsolaki A, Tsolaki M, Kompatsiaris IY, Hadjileontiadis LJ.
Cognitive MMN and P300 in mild cognitive impairment and Alzheimer's disease: A high density
EEG-3D vector field tomography approach. *Brain Research*. Oct 01 2016;1648(Pt A):425-33.
doi:10.1016/j.brainres.2016.07.043
11. Missonnier P, Herrmann FR, Richiardi J, et al. Attention-related potentials allow for a
highly accurate discrimination of mild cognitive impairment subtypes. Research Support, Non-
U.S. Gov't. *Neuro-degenerative diseases*. 2013;12(2):59-70. doi:10.1159/000338815
12. Missonnier P, Deiber MP, Gold G, et al. Working memory load-related
electroencephalographic parameters can differentiate progressive from stable mild cognitive
impairment. Research Support, Non-U.S. Gov't. *Neuroscience*. Dec 05 2007;150(2):346-56.
doi:10.1016/j.neuroscience.2007.09.009

13.   Bidelman GM, Lowther JE, Tak SH, Alain C. Mild cognitive impairment is characterized by
deficient brainstem and cortical representations of speech. *J Neurosci*. Mar 07 2017;doi:3700-
16 [pii] 10.1523/JNEUROSCI.3700-16.2017 JNEUROSCI.3700-16.2017 [pii]

14.   Cespon J, Galdo-Alvarez S, Pereiro AX, Diaz F. Differences between mild cognitive
impairment subtypes as indicated by event-related potential correlates of cognitive and motor
processes in a Simon task. Research Support, Non-U.S. Gov't. *Journal of Alzheimer's disease :*
JAD. 2015;43(2):631-47. doi:10.3233/JAD-132774
15.   Chapman RM, McCrary JW, Gardner MN, et al. Brain ERP components predict which
individuals progress to Alzheimer's disease and which do not. Research Support, N.I.H.,
Extramural. *Neurobiology of aging*. Oct 2011;32(10):1742-55.
doi:10.1016/j.neurobiolaging.2009.11.010
16.   Pegado F, Comerlato E, Ventura F, et al. Timing the impact of literacy on visual
processing. *Proc Natl Acad Sci U S A*. Dec 9 2014;111(49):E5233-42.
doi:10.1073/pnas.1417347111
17.   Alain C, Chow R, Lu J, et al. Aging Enhances Neural Activity in Auditory, Visual, and
Somatosensory Cortices: The Common Cause Revisited. *J Neurosci*. Jan 12 2022;42(2):264-275.
doi:10.1523/JNEUROSCI.0864-21.2021
18.   Irimajiri R, Michalewski HJ, Golob EJ, Starr A. Cholinesterase inhibitors affect brain
potentials in amnestic mild cognitive impairment. Clinical Trial Research Support, N.I.H.,
Extramural. *Brain Research*. May 11 2007;1145:108-16. doi:10.1016/j.brainres.2007.01.120

19.   Green DL, Payne L, Polikar R, Moberg PJ, Wolk DA, Kounios J. P50: A candidate ERP
biomarker of prodromal Alzheimer's disease. Research Support, Non-U.S. Gov't. *Brain Research*.
Oct 22 2015;1624:390-7. doi:10.1016/j.brainres.2015.07.054
20.   Fix ST, Arruda JE, Andrasik F, Beach J, Groom K. Using visual evoked potentials for the
early detection of amnestic mild cognitive impairment: a pilot investigation. *International*
journal of geriatric psychiatry. Jan 2015;30(1):72-9. doi:10.1002/gps.4117
21.   Yamasaki T, Horie S, Ohyagi Y, et al. A Potential VEP Biomarker for Mild Cognitive
Impairment: Evidence from Selective Visual Deficit of Higher-Level Dorsal Pathway. *Journal of*
Alzheimer's disease : JAD. May 23 2016;53(2):661-76. doi:10.3233/JAD-150939
22.   Stothart G, Kazanina N, Naatanen R, Haworth J, Tales A. Early visual evoked potentials
and mismatch negativity in Alzheimer's disease and mild cognitive impairment. Research
Support, Non-U.S. Gov't. *Journal of Alzheimer's disease : JAD*. 2015;44(2):397-408.
doi:10.3233/JAD-140930
23.   Stephen JM, Montano R, Donahue CH, et al. Somatosensory responses in normal aging,
mild cognitive impairment, and Alzheimer's disease. Research Support, N.I.H., Extramural
Research Support, Non-U.S. Gov't. *Journal of neural transmission*. Feb 2010;117(2):217-25.
doi:10.1007/s00702-009-0343-5

24.   Metzger FG, Polak T, Aghazadeh Y, Ehlis AC, Hagen K, Fallgatter AJ. Vagus
somatosensory evoked potentials--a possibility for diagnostic improvement in patients with
mild cognitive impairment? *Dement Geriatr Cogn Disord*. 2012;33(5):289-96.
doi:10.1159/000339359000339359 [pii]

25.   Van Der Flier WM, Van Den Heuvel DM, Weverling-Rijnsburger AW, et al. Cognitive
decline in AD and mild cognitive impairment is associated with global brain damage.
Comparative Study Research Support, Non-U.S. Gov't. *Neurology*. Sep 24 2002;59(6):874-9.

26.   de Mendonca A, Ribeiro F, Guerreiro M, Garcia C. Frontotemporal mild cognitive
impairment. Case Reports. *Journal of Alzheimer's disease : JAD*. Feb 2004;6(1):1-9.
27.   Bozzali M, Filippi M, Magnani G, et al. The contribution of voxel-based morphometry in
staging patients with mild cognitive impairment. *Neurology*. Aug 08 2006;67(3):453-60.
doi:10.1212/01.wnl.0000228243.56665.c2
28.   Dai W, Lopez OL, Carmichael OT, Becker JT, Kuller LH, Gach HM. Mild cognitive
impairment and alzheimer disease: patterns of altered cerebral blood flow at MR imaging.
Radiology. Mar 2009;250(3):856-66. doi:10.1148/radiol.2503080751
29.   Fennema-Notestine C, Hagler DJ, Jr., McEvoy LK, et al. Structural MRI biomarkers for
preclinical and mild Alzheimer's disease. Research Support, N.I.H., Extramural. *Human brain*
mapping. Oct 2009;30(10):3238-53. doi:10.1002/hbm.20744
30.   Machulda MM, Senjem ML, Weigand SD, et al. Functional magnetic resonance imaging
changes in amnestic and nonamnestic mild cognitive impairment during encoding and
recognition tasks. Research Support, N.I.H., Extramural Research Support, Non-U.S. Gov't.
Journal of the International Neuropsychological Society : JINS. May 2009;15(3):372-82.
doi:10.1017/S1355617709090523

31.   Lee ES, Yoo K, Lee YB, et al. Default Mode Network Functional Connectivity in Early and
Late Mild Cognitive Impairment: Results From the Alzheimer's Disease Neuroimaging Initiative.
Alzheimer disease and associated disorders. Oct-Dec 2016;30(4):289-296.
doi:10.1097/WAD.0000000000000143
32.   Yu J, Li R, Jiang Y, Broster LS, Li J. Altered Brain Activities Associated with Neural
Repetition Effects in Mild Cognitive Impairment Patients. *Journal of Alzheimer's disease : JAD*.
May 11 2016;53(2):693-704. doi:10.3233/JAD-160086
33.   Long Z, Jing B, Yan H, et al. A support vector machine-based method to identify mild
cognitive impairment with multi-level characteristics of magnetic resonance imaging.
Neuroscience. Sep 07 2016;331:169-76. doi:10.1016/j.neuroscience.2016.06.025
34.   Ribeiro LG, Busatto GF. Voxel-based morphometry in Alzheimers disease and mild
cognitive impairment: Systematic review of studies addressing the frontal lobe. *Dement*
Neuropsychol. Apr-Jun 2016;10(2):104-112. doi:10.1590/S1980-5764-2016DN1002006
35.   Wei R, Li C, Fogelson N, Li L. Prediction of Conversion from Mild Cognitive Impairment to
Alzheimer's Disease Using MRI and Structural Network Features. *Front Aging Neurosci*.
2016;8:76. doi:10.3389/fnagi.2016.00076
36.   Knight RT, Staines WR, Swick D, Chao LL. Prefrontal cortex regulates inhibition and
excitation in distributed neural networks. Research Support, Non-U.S. Gov't Research Support,
U.S. Gov't, Non-P.H.S. Research Support, U.S. Gov't, P.H.S. Review. *Acta psychologica*. Apr
1999;101(2-3):159-78.

37.   Alain C, Woods DL, Knight RT. A distributed cortical network for auditory sensory
memory in humans. *Brain Res*. Nov 23 1998;812(1-2):23-37.

38.   Pekkonen E, Hirvonen J, Jaaskelainen IP, Kaakkola S, Huttunen J. Auditory sensory
memory and the cholinergic system: implications for Alzheimer's disease. Clinical Trial
Comparative Study Randomized Controlled Trial Research Support, Non-U.S. Gov't.
Neuroimage. Aug 2001;14(2):376-82. doi:10.1006/nimg.2001.0805

39.   Gazzaley A, Clapp W, Kelley J, McEvoy K, Knight RT, D'Esposito M. Age-related top-down
suppression deficit in the early stages of cortical visual memory processing. *Proc Natl Acad Sci U*
S A. Sep 2 2008;105(35):13122-6. doi:0806074105 [pii] 10.1073/pnas.0806074105

40.   Thomas C, vom Berg I, Rupp A, et al. P50 gating deficit in Alzheimer dementia correlates
to frontal neuropsychological function. Research Support, Non-U.S. Gov't. *Neurobiology of*
Aging. Mar 2010;31(3):416-24. doi:10.1016/j.neurobiolaging.2008.05.002
41.   Khatun S, Morshed BI, Bidelman GM. A Single-Channel EEG-Based Approach to Detect
Mild Cognitive Impairment via Speech-Evoked Brain Responses. *IEEE transactions on neural*
systems and rehabilitation engineering : a publication of the IEEE Engineering in Medicine and
Biology Society. May 2019;27(5):1063-1070. doi:10.1109/TNSRE.2019.2911970
42.   Ieracitano C, Mammone N, Hussain A, Morabito FC. A novel multi-modal machine
learning based approach for automatic classification of EEG recordings in dementia. *Neural*
Netw. Mar 2020;123:176-190. doi:10.1016/j.neunet.2019.12.006
43.   Trinh TT, Tsai CF, Hsiao YT, Lee CY, Wu CT, Liu YH. Identifying Individuals With Mild
Cognitive Impairment Using Working Memory-Induced Intra-Subject Variability of Resting-State
EEGs. *Front Comput Neurosci*. 2021;15:700467. doi:10.3389/fncom.2021.700467
44.   Siuly S, Alcin OF, Kabir E, et al. A New Framework for Automatic Detection of Patients
With Mild Cognitive Impairment Using Resting-State EEG Signals. *IEEE Trans Neural Syst Rehabil*
Eng. Sep 2020;28(9):1966-1976. doi:10.1109/TNSRE.2020.3013429
45.   Bidelman GM, Mahmud MS, Yeasin M, Shen D, Arnott SR, Alain C. Age-related hearing
loss increases full-brain connectivity while reversing directed signaling within the dorsal-ventral
pathway for speech. *Brain structure & function*. Jul 25 2019;doi:10.1007/s00429-019-01922-9
46.   Welsh KA, Breitner JC, Magruder-Habib KM. Detection of dementia in the elderly using
telephone screening of cognitive status. *Cognitive and Behavioral Neurology*. 1993;6(2):103-
110.
47.   Chow R, Rabi R, Paracha S, et al. Reaction Time Intraindividual Variability Reveals
Inhibitory Deficits in Single- and Multiple-Domain Amnestic Mild Cognitive Impairment. *J*
Gerontol B Psychol Sci Soc Sci. Jan 12 2022;77(1):71-83. doi:10.1093/geronb/gbab051
48.   Rabi R, Chow R, Paracha S, et al. Time of Day Effects on Inhibitory Functioning: Cognitive
and Neural Evidence of Sundowning in Amnestic Mild Cognitive Impairment. *J Alzheimers Dis*.
2022;90(2):869-890. doi:10.3233/JAD-220580
49.   Oostenveld R, Fries P, Maris E, Schoffelen JM. FieldTrip: Open source software for
advanced analysis of MEG, EEG, and invasive electrophysiological data. *Comput Intell Neurosci*.
2011;2011:156869. doi:10.1155/2011/156869
50.   Srinivasan R, Winter WR, Ding J, Nunez PL. EEG and MEG coherence: measures of
functional connectivity at distinct spatial scales of neocortical dynamics. *J Neurosci Methods*.
Oct 15 2007;166(1):41-52. doi:10.1016/j.jneumeth.2007.06.026

51.   Shen X, Finn ES, Scheinost D, et al. Using connectome-based predictive modeling to
predict individual behavior from brain connectivity. *Nat Protoc*. Mar 2017;12(3):506-518.
doi:10.1038/nprot.2016.178
52.   Chang CC, Lin CJ. LIBSVM: a library for support vector machines. *ACM transactions on
intelligent systems and technology (TIST)*. 2011;2(3):1-27.
53.   Bruna R, Poza J, Gomez C, Garcia M, Fernandez A, Hornero R. Analysis of spontaneous
MEG activity in mild cognitive impairment and Alzheimer's disease using spectral entropies and
statistical complexity measures. Research Support, Non-U.S. Gov't. *Journal of neural
engineering*. Jun 2012;9(3):036007. doi:10.1088/1741-2560/9/3/036007
54.   Cantero JL, Atienza M, Cruz-Vadell A, Suarez-Gonzalez A, Gil-Neciga E. Increased
synchronization and decreased neural complexity underlie thalamocortical oscillatory dynamics
in mild cognitive impairment. *Neuroimage*. Jul 15 2009;46(4):938-48. doi:S1053-
8119(09)00237-7 [pii] 10.1016/j.neuroimage.2009.03.018

55.   Nakagawa TT, Jirsa VK, Spiegler A, McIntosh AR, Deco G. Bottom up modeling of the
connectome: linking structure and function in the resting brain and their changes in aging.
Research Support, Non-U.S. Gov't Review. *Neuroimage*. Oct 15 2013;80:318-29.
doi:10.1016/j.neuroimage.2013.04.055

56.   Eyler LT, Elman JA, Hatton SN, et al. Resting State Abnormalities of the Default Mode
Network in Mild Cognitive Impairment: A Systematic Review and Meta-Analysis. *J Alzheimers
Dis*. 2019;70(1):107-120. doi:10.3233/JAD-180847
57.   Yang X, Wu H, Song Y, et al. Functional MRI-specific alterations in frontoparietal network
in mild cognitive impairment: an ALE meta-analysis. *Front Aging Neurosci*. 2023;15:1165908.
doi:10.3389/fnagi.2023.1165908
58.   Zhao W, Makowski C, Hagler DJ, et al. Task fMRI paradigms may capture more
behaviorally relevant information than resting-state functional connectivity. *Neuroimage*. Apr
15 2023;270:119946. doi:10.1016/j.neuroimage.2023.119946
59.   Finn ES, Bandettini PA. Movie-watching outperforms rest for functional connectivity-
based prediction of behavior. *Neuroimage*. Jul 15 2021;235:117963.
doi:10.1016/j.neuroimage.2021.117963
60.   Finn ES, Scheinost D, Finn DM, Shen X, Papademetris X, Constable RT. Can brain state be
manipulated to emphasize individual differences in functional connectivity? *Neuroimage*. Oct
15 2017;160:140-151. doi:10.1016/j.neuroimage.2017.03.064
61.   Jiang J, Yan Z, Sheng C, et al. A Novel Detection Tool for Mild Cognitive Impairment
Patients Based on Eye Movement and Electroencephalogram. *J Alzheimers Dis*. 2019;72(2):389-
399. doi:10.3233/JAD-190628
62.   Toth L, Hoffmann I, Gosztolya G, et al. A Speech Recognition-based Solution for the
Automatic Detection of Mild Cognitive Impairment from Spontaneous Speech. *Curr Alzheimer
Res*. 2018;15(2):130-138. doi:10.2174/1567205014666171121114930
63.   Ruzzoli M, Pirulli C, Mazza V, Miniussi C, Brignani D. The mismatch negativity as an index
of cognitive decline for the early detection of Alzheimer's disease. *Sci Rep*. Sep 12
2016;6:33167. doi:10.1038/srep33167

64.   Arena S, Florian E, Sgarbossa F, Sølvsberg E, Zennaro I. A conceptual framework for
machine learning algorithm selection for predictive maintenance. *Engineering Applications of*
Artificial Intelligence. 2024;133:108340.
65.   Vellido A. The importance of interpretability and visualization in machine learning for
applications in medicine and health care. *Neural computing and applications*.
2020;32(24):18069-18083.
66.   Wang S, Wang W, Chen J, Yu X. Alterations in brain functional connectivity in patients
with mild cognitive impairment: A systematic review and meta-analysis of functional near-
infrared spectroscopy studies. *Brain Behav*. Apr 2024;14(4):e3414. doi:10.1002/brb3.3414
67.   Tarawneh HY, Mulders W, Sohrabi HR, Martins RN, Jayakody DMP. Investigating
Auditory Electrophysiological Measures of Participants with Mild Cognitive Impairment and
Alzheimer's Disease: A Systematic Review and Meta-Analysis of Event-Related Potential Studies.
J Alzheimers Dis. 2021;84(1):419-448. doi:10.3233/JAD-210556
68.   Gu L, Zhang Z. Exploring Potential Electrophysiological Biomarkers in Mild Cognitive
Impairment: A Systematic Review and Meta-Analysis of Event-Related Potential Studies. *J*
Alzheimers Dis. 2017;58(4):1283-1292. doi:10.3233/JAD-161286
69.   Yamasaki T, Horie S, Muranaka H, Kaseda Y, Mimori Y, Tobimatsu S. Relevance of in vivo
neurophysiological biomarkers for mild cognitive impairment and Alzheimer's disease. Research
Support, Non-U.S. Gov't Review. *Journal of Alzheimer's disease : JAD*. 2012;31 Suppl 3:S137-54.
doi:10.3233/JAD-2012-112093

70.   Paitel ER, Samii MR, Nielson KA. A systematic review of cognitive event-related
potentials in mild cognitive impairment and Alzheimer's disease. *Behav Brain Res*. Jan 1
2021;396:112904. doi:10.1016/j.bbr.2020.112904
71.   Deiber MP, Ibanez V, Herrmann F, et al. Face short-term memory-related
electroencephalographic patterns can differentiate multi- versus single-domain amnestic mild
cognitive impairment. Research Support, Non-U.S. Gov't. *Journal of Alzheimer's disease : JAD*.
2011;26(1):157-69. doi:10.3233/JAD-2011-110170
72.   Cid-Fernandez S, Lindin M, Diaz F. Effects of amnestic mild cognitive impairment on N2
and P3 Go/NoGo ERP components. Research Support, Non-U.S. Gov't. *Journal of Alzheimer's*
disease : JAD. 2014;38(2):295-306. doi:10.3233/JAD-130677

Acknowledgments

Funding

This research was supported by grants from the Natural Sciences and Engineering Research

Council of Canada (NSERC, 194536, 05241), Canadian Institute for Health Research (PJT

183614), Alzheimer Society of Canada, and The Lorraine Johnson Foundation to CA; RR was

supported by a postdoctoral fellowship from the Alzheimer Society of Canada. Morris

Freedman receives support from the Saul A. Silverman Family Foundation as a Canada

International Scientific Exchange Program and Morris Kerzner Memorial Fund. Reprint requests

should be sent to Claude Alain, Rotman Research Institute, Baycrest Centre, 3560 Bathurst

Street, Toronto, Ontario, Canada, M6A 2E1 or via email: calain@research.baycrest.org

Conflicts of Interest

The authors declare no conflicts of interest.

Consent Statement

All participants provided informed consent for participation.

Keywords: Aging, Evoked Potentials, EEG, sensory perception, mild cognitive impairment

Table 1

Participant Demographic and Neuropsychological Data

Variable	aMCI Means (SD) n = 44		HC Means (SD) n = 35	
	Raw	Scaled	Raw	Scaled
Demographics				
Age (years)	76.00 (5.82)	--	75.86 (6.99)	--
Education (years)	15.36 (2.76)	--	16.34 (2.29)	--
Sex (F:M)	20:20	--	15:20	--
TICS-m *	32.53 (3.49)	--	37.03 (3.04)	--
MoCA *	22.43 (2.97)	--	27.09 (2.34)	--
Estimates of IQ				
WAIS-III Matrix Reasoning *	21.30 (5.15)	13.09 (2.27)	24.14 (5.26)	14.40 (2.67)
Shipley Vocabulary	34.41 (3.77)	11.39 (2.93)	35.94 (3.28)	12.54 (2.70)
Memory				
CVLT-II Learning *	26.02 (8.13)	5.83 (2.14)	49.29 (9.77)	13.20 (2.96)
CVLT-II Short Delay FR *	2.43 (2.35)	3.83 (2.97)	10.31 (3.26)	12.40 (3.07)
CVLT-II Long Delay FR *	2.85 (2.47)	4.05 (2.89)	10.57 (3.32)	11.54 (2.82)
WMS-R Visual PA I *	7.61 (4.23)	8.89 (2.62)	11.97 (3.29)	11.71 (2.47)
WMS-R Visual PA II *	3.14 (1.85)	9.80 (2.21)	4.97 (1.40)	11.97 (1.72)
WMS-R Verbal PA I *	10.72 (6.40)	5.42 (2.73)	16.03 (2.96)	9.94 (2.25)
WMS-R Verbal PA II *	4.42 (1.92)	8.19 (3.07)	6.77 (1.17)	11.54 (2.49)
WAIS-III Digit Symbol IL FR *	4.70 (2.22)	7.07 (3.06)	7.31 (0.93)	10.43 (1.07)
WAIS-III Digit Symbol IL PR *	4.36 (4.50)	6.52 (3.47)	11.89 (4.32)	10.40 (1.42)
Language				
BNT-15	50.18 (8.71)	9.82 (3.38)	53.71 (4.93)	10.94 (2.82)
Phonemic Fluency (FAS)	40.41 (12.28)	9.93 (3.12)	47.40 (13.16)	11.46 (3.46)
Semantic Fluency (Animals) *	13.16 (4.56)	6.66 (3.18)	19.00 (5.27)	10.47 (3.59)
Executive Functioning and Processing Speed				
WAIS-III Digit Symbol *	49.70 (15.44)	10.57 (2.95)	60.97 (14.42)	12.80 (2.77)
D-KEFS Trails Numbers *	51.74 (21.40)	11.02 (3.51)	38.97 (12.18)	12.94 (2.18)
D-KEFS Trails Letters *	54.90 (23.32)	10.71 (3.11)	37.74 (11.41)	12.89 (1.64)
D-KEFS Trails N-L Switch *	142.60 (62.42)	9.18 (4.34)	98.97 (37.95)	12.03 (2.51)
D-KEFS CWIT Colour *	34.40 (7.88)	10.00 (3.03)	30.55 (5.59)	11.54 (2.22)
D-KEFS CWIT Word	25.82 (8.27)	10.20 (3.59)	22.85 (4.56)	11.17 (2.93)
D-KEFS CWIT Inhibition *	75.29 (26.53)	10.59 (3.24)	57.59 (13.69)	12.91 (1.90)
Alpha Span Stop Point	21.41 (6.19)	8.46 (2.58)	28.51 (9.28)	10.60 (3.10)
WCST Perseverative Errors % *	23.12 (12.65)	9.93 (3.43)	14.12 (10.19)	13.20 (4.41)

WCST Categories	3.93 (2.29)	--	4.66 (1.94)	--
Questionnaires				
HADS Anxiety	4.74 (3.67)	--	4.69 (2.89)	--
HADS Depression	2.74 (2.25)	--	2.51 (2.16)	--
EPW	5.81 (4.15)	--	6.69 (2.96)	--
PSQI	5.05 (2.74)	--	5.76 (2.99)	--
MAC Abilities *	59.85 (9.29)	8.22 (2.41)	68.86 (9.75)	10.69 (2.44)
FAQ	1.81 (2.29)	--	--	--

Note. TICS-m = modified Telephone Interview of Cognitive Status; MoCA = Montreal Cognitive

Assessment; WAIS-III = Wechsler Adult Intelligence Scale, 3rd edition; CVLT-II = California Verbal

Learning Test, 2nd edition; FR = Free Recall; WMS-R = Wechsler Memory Scale – Revised; PA =

Paired Associates; PR = Paired Recall; IL= Incidental Learning; BNT-15 = Boston Naming Test,

short form; FAS = phonemic fluency to the letters F, A, and S; D-KEFS = Delis Kaplan Executive

Functioning System; CWIT = Colour-Word Interference Test; N-L = Number-Letter; WCST =

Wisconsin Card Sorting Test; HADS = Hospital Anxiety and Depression Scale; EPW = Epworth

Sleepiness Scale; PSQI = Pittsburgh Sleep Quality Index; MAC = Memory Assessment Clinics Self-

Rating Scale; FAQ = Functional Assessment Questionnaire. Scaled scores: mean =10, SD = 3. For

normed assessments, tests of significance were run on scaled scores. *aMCI ≠ HC, $p < .05$

795
796

Table 2

798 Linear regression results. Regression coefficients and corresponding p values (in parenthesis) between

799 summed functional connectivity and cognitive performance after controlling group.

	Edge Strength	Memory	Speed	EF	Language
Delta-Theta-Alpha band	V: aMCI > HC	0.03 (0.982)	-2.09 (0.099)	-4.37 (0.005)	-2.07 (0.264)
	A: aMCI > HC	1.00 (0.554)	-1.57 (0.306)	-3.57 (0.063)	-2.38 (0.284)
	S: aMCI > HC	-0.28 (0.833)	-1.18 (0.328)	-3.60 (0.017)	-1.74 (0.321)
	A: aMCI < HC	2.13 (0.243)	0.50 (0.764)	1.19 (0.569)	0.34 (0.889)
	S: aMCI < HC	1.88 (0.272)	2.44 (0.115)	2.68 (0.172)	1.29 (0.571)
Beta band	V: aMCI > HC	-2.71 (0.143)	-3.17 (0.058)	-5.39 (0.010)	-3.65 (0.135)
	V: aMCI < HC	-0.01 (0.993)	-2.15 (0.154)	-2.55 (0.180)	-5.92 (0.006)

800 *Note:* EF = executive functioning; HC = healthy control;

Fig. 1. Classification procedure illustration. We separately performed SVM classification analyses using functional connectivity matrices or ERP for each sensory modality (for details, please see methods). With leave-one-subject-out cross validation, we obtain classification accuracy of each frequency or electrode for each sensory modality. Next, we performed permutation-based test for the significance and statistics of each frequency or electrode for each sensory modality (for details, please see methods). Lastly, we employed the voting procedures to combine the output from difference neural indices evoked by stimuli from difference sensory modalities to perform the prediction.

Fig. 2. Classification results using functional connectivity matrices. A. Classification accuracy at

different frequency bands for three modalities. *denotes the significant frequencies for each

modality after cluster-based correction. B-C, Summarized the positive (B) and negative (C)

important edge numbers that were kept according to the five ROIs (frontal ROI: frontopolar,

anterior-frontal and frontal channels, temporal ROI: frontotemporal and temporal-parietal

channels, central ROI: frontocentral and central channels, parietal ROI: centroparietal and

parietal channels, and occipital ROI: parietal-occipital and occipital channels), Important

connectivity edges were those edges that preserved in more than 90% folds. D-F, Classification

accuracy, sensitivity, and specificity of the top frequencies (D), voting results combining three

modalities (E), predictable subject percent (F).

Fig. 3. Classification results using ERP. A-C, Topoplot (left: unthresholded accuracy map; right: significant clusters) of prediction accuracy for visual (A), auditory (B), and somatosensory (C) modalities. D-F, Classification accuracy, sensitivity, and specificity of the top electrodes (D), voting results combining three modalities (E), predictable subject percent (F). G-J, Averaged ERP of the significant clusters.

Fig. 4. Voting results combining six models. We employed majority rules and one vote veto

rules to make predictions based on the predictions from the models trained by functional

matrices or ERP evoked by visual, auditory, and somatosensory modalities. A-B, Classification

accuracy, sensitivity, and specificity of voting results combining three modalities (A),

predictable subject percent (B).