Proteonano™: A Robust Platform for Large Cohort Deep Plasma Proteomic Studies

Hao Wu*,[1], Yonghao Zhang[1], Yi Wang[1], Xiehua Ouyang[1]

[1]Nanomics Biotechnology Co., Ltd.

*Corresponding author: haowu@nanomics.bio

Keywords: Proteonano Ultraplex Proteomics Platform, Deep Plasma Proteomics, Robust and Fast MS-

based Proteomics Workflow

10
11

Abstract

Complete profiling of human plasma proteome is an immerse source for disease biomarker discovery.

Cutting-edge mass spectrometers, like ThermoFisher's Orbitrap Astral, have promised unprecedented

insights into the exploration of multiple protein biomarkers from human plasma samples. However, large-

scale, deep profiling of the human plasma proteome, especially low-abundant proteins (LAPs, <10 ng.mL⁻

¹), in a robust and fast way remains challenging. This is largely due to the lack of standardized and

automated workflows including LAPs enrichment, reduction, and enzymatic digestion procedures. Until

now, these complex procedures have not been incorporated into a streamlined workflow to achieve

reproducibility, high-throughput, and deep proteome coverage.

Here we report the Proteonano Ultraplex Proteomics (PUP) Platform for large cohort plasma proteomics

studies with robustness and fast throughput by standardizing workflow by incorporating the Proteonano

platform and high-resolution mass spectrometers, including Orbitrap Exploris™ 480, Orbitrap Astral™, and

timsTOF Pro. This pipeline demonstrates excellent stability and reproducibility, with tunable balance

between proteome coverage and throughput. We further demonstrate the utility of this platform for

potential biomarker discovery in a neurodegenerative disease cohort. This harmonized method enables

robust, fast and large-cohort plasma proteomics studies to meet the need to discover new biomarkers.

Introduction

The development of untargeted bottom-up mass spectrometry-based proteomics has evolved over the

past two decades. Proteomic studies have played a pivotal role in identifying disease biomarkers, which

are now integral to precision medicine, contributing to disease detection, patient stratification,

therapeutic treatment monitoring, and predictive analysis[1–9]. Blood plasma, a rich repository of novel

protein biomarkers, presents unique challenges due to the wide dynamic range (>10 orders of magnitude)

in protein concentrations. High-abundance plasma proteins often overshadow the detection of medium-

and low-abundance proteins during mass spectrometry-based proteomic analyses[10–12].

Therefore, prior to mass spectrometry, enrichment and fractionation of low-abundance proteins from

plasma samples are commonly employed to mitigate the interference caused by high-abundance

proteins[10,11,13]. One widely used approach involves the immunodepletion of the most abundant plasma

proteins using immunoassay columns, such as the High-Select Top 14 mini columns. Immunodepletion

techniques typically enhance protein detection depth, achieving a median 4-fold increase in identified

protein groups. However, studies have shown that depleted plasma primarily contains medium- to high-

abundance proteins, with only 5–6% identified as low-abundance[12]. Additionally, immunodepletion poses

challenges such as non-specific binding leading to unintended loss of low-abundance proteins, complex

sample handling, high costs, and batch effects, limiting its scalability for large-cohort plasma proteomics

studies[12].

Another substantial barrier to the widespread adoption of mass spectrometry (MS)-based proteomics in

large cohort studies is the reproducibility of distinct mass spectrometry configurations across laboratories.

Ruedi et al.[14] assessed the quantitative reproducibility, dynamic range, and sensitivity of 11 mass

spectrometers globally. Their study quantified over 4,000 proteins from HEK293 cells, achieving a median

coefficient of variation (CV) of 57.6% before normalization and 22.0% after normalization. For plasma

proteomics, sample-related biases also greatly impact data reproducibility and subsequent biomarker

discovery. To address this issue, Geyer et al.[15] developed three quality marker panels to evaluate whether

plasma samples were contaminated by coagulation, platelets, or erythrocytes.

The aforementioned challenges have resulted in most MS-based proteomic studies being small and

specialized. In contrast, affinity binder technologies, despite their higher cost, have been widely adopted

for larger proteomics cohorts[16]. Consequently, there is a pressing need for robust, efficient, and high-

throughput MS-based proteomic workflows that are cost-effective and suitable for large-cohort studies.

Recently, the advancement of several cutting-edge technologies has driven substantial progress in MS-

based plasma proteomics. These technologies include nanoparticle enrichment, automated sample

preparation, and high-throughput mass spectrometry. Firstly, surface-engineered superparamagnetic

nanoparticles have emerged as a promising technology for enriching low-abundance plasma

proteins[10,11,17–19]. These nanoparticles interact with proteins in biological fluids to form a thin layer known

as the protein corona, and extensive studies have shown the critical role of protein corona in nanoparticle

pathophysiology[20–27]. Secondly, unlike traditional manual methods such as immunodepletion or

chromatographic fractionation, magnetic nanoparticle reagent-based protein capture, combined with

peptide identification via mass spectrometry, is well-suited for automation using common liquid handling

workstations equipped with magnet modules and additional components such as heater-shaker units.

Notably, affinity binder-based protein capture integrated with signal readout through next-generation

sequencing (NGS)-based proteomics technologies like Olink PEA has already been widely employed in

large cohort plasma proteomic studies, offering highly automated workflows. Lastly, newly commercialized

high-resolution mass spectrometers, including ThermoFisher's Orbitrap Astral, have significantly

enhanced detection depth and proteome coverage, achieving a maximum throughput of 180 samples per

day per instrument[28].

Thanks to these advancements, mass spectrometry-based nanoparticle-enriched plasma proteomics has

seen remarkable advancements in recent years. Blume et al.[10] introduced a 5-nanoparticle panel,

functionalized with chemical ligands, that successfully identified over 2,000 proteins from 141 plasma

samples using an automated sample preparation workflow. However, the reliance on multiple

nanoparticles, sample fractionation, and high starting plasma volume imposes limitations on throughput

and substantially increases overall costs. To address such challenges, MacCoss and colleagues[11] developed

the Mag-Net enrichment strategy, which leverages strong-anion exchange magnetic microparticles to

enrich membrane-bound particles, enabling the profiling of over 4,000 plasma proteins. Nevertheless, the

heterogeneity of extracellular vesicles across plasma samples and impact of centrifugation conditions may

pose challenges for achieving reproducibility between batches, a topic that receives minimal attention in

the manuscript. For instance, Beimers et al.[12] compared six plasma proteomics methods and reported that

the Mag-Net strategy achieved a detected depth of 1,200 protein groups.

While significant progress has been achieved, notable challenges continue to persist. A recent meta-

analysis of 2,134 published manuscripts in the proteomics analyses of nanoparticle corona revealed

substantial variability in proteome coverage[18]. Gharibi et al.[17] highlighted this issue by distributing three

technical replicates of a protein corona sample to 17 mass spectrometry core facilities across the United

States, uncovering notable discrepancies in plasma proteome coverage. They further demonstrated that

employing a standardized database search could effectively minimize these discrepancies across distinct

sites[19]. Therefore, there is a critical need to establish standardized nanoparticle protein corona workflows

encompassing nanoparticle characterization, sample preparation, and automation, along with mass

spectrometry instrumentation, data analysis pipelines, and achieving an optimal balance between

proteome coverage and throughput.

In this study, we introduced the Proteonano Ultraplex Proteomics (PUP) Platform, a comprehensive system

integrating proprietary monodispersed magnetic nanoparticle-based reagents for low-abundance protein

enrichment, an automated sample preparation workstation for parallel sample preparation (Nanomation

G1), and high-resolution mass spectrometry instruments for measurement, as well as built-in quality

control procedures to ensure consistent and reliable data monitoring throughout the workflow. The PUP

platform was rigorously evaluated for key features, including proteome coverage, the balance between

protein depth and throughput, compatibility with various mass spectrometry configurations, and batch-

to-batch quantitative reproducibility. Notably, it demonstrated the capacity to profile up to 6,000 proteins

from pooled human plasma samples, with an intra-batch quantitative median CV of 15%. Its versatility was

further tested across ten LC-MS/MS systems, showcasing its ability to maintain a balance between

throughput and proteome coverage. Finally, the platform's feasibility was validated in a cohort of

participants with cognitive impairment, demonstrating its potential for novel biomarker discovery in

plasma proteomics studies. To our knowledge, this study presents a fully integrated, robust, and

standardized mass spectrometry-based deep plasma proteomics platform. We believe this innovation

holds great promise for fostering the widespread adoption of nanoparticle-enriched MS-based platforms,

thereby accelerating advancements in plasma biomarker discovery on a broader scale.

Results

Overview of the Proteonano workflow

The Proteonano platform is intended to be a standardized, high-throughput solution for large-cohort

plasma studies, combining robustness with efficiency. It integrates proprietary plasma proteome

enrichment reagents, the Nanomation G1 automated workstation, and high-resolution LC-MS/MS systems

(Fig. 1). The reagents feature hierarchically structured peptide-conjugated nanoparticles (PCNPs) to

selectively enrich low-abundance proteins while depleting medium- and high-abundance proteins during

sample preparation. Details of peptide properties are included in the Supplementary Table 1. The

Nanomation G1 workstation includes a heater-shaker module for nanoparticle-plasma incubation, a

magnetic module for protein corona separation, and an 8-channel liquid handler for parallel pipetting. The

top view of the workstation layout can be seen in the Supplementary Information Fig. S1). For peptide

identification, the platform incorporates three high-resolution mass spectrometers—Orbitrap 480, Astral

instruments (ThermoFisher Scientific), and timsTOF Pro 2 (Bruker Corporation)—ensuring compatibility

with widely established proteomics infrastructure (Fig. 1)

Figure 1. An overall view of the Proteonano Ultraplex Proteomics platform.

The generation of mass spectrometry-based proteomic data is often influenced by various technical and

nontechnical factors. To address these challenges, we introduce the Quality Control System (QCS) for the

Proteonano workflow—a framework designed to evaluate and ensure data quality and reproducibility. As

illustrated in Fig. 2, the QCS encompasses three key control steps: the **Incubation Control** characterizes

the proteome enrichment process via nanoparticle-plasma incubation; the **Detection Control** assesses and

monitors systematic variations from LC-MS/MS instruments using premade peptide samples; and the **Data**

Control includes algorithms for missing data imputation and normalization are employed to detect and

correct batch-to-batch variations effectively.

Fig. 2: The Quality Control System of the Proteonano Platform for batch-to-batch data monitoring.

In a 96-well plate experiment, six QC samples are used to monitor workflow performance. QC 1 consists

of a lyophilized peptide mix derived from pooled plasma, processed with nanoparticle enrichment

followed by conventional proteomic preparation steps. It is included in triplicate per plate to assess LC-

MS/MS reliability. QC 2 uses neat plasma, untreated by nanoparticles, that undergoes standard sample

preparation to evaluate the quality of conventional procedures. QC 3 employs pooled plasma processed

nanoparticle enrichment and standard preparation steps to monitor the performance of the complete

pipeline. By comparing results from replicates of QC1, QC2 and QC3, the protein enrichment process's

performance can be evaluated. The recommended positions of the six QC samples on the 96-well plate

can be seen in the Supplementary Fig. S2.

Preparation and Characterization of Nanoparticles

The Proteonano nanoparticle is designed to enrich low-abundance proteins from biological samples

through nano-bio interactions. As illustrated in Fig. 3a, photographs depict the dispersion of peptide-

conjugated nanoparticles in water both before and after peptide conjugation. To fully characterize the

multi-layered structures of the PCNPs, we conducted comprehensive measurements on key features,

including size, size distribution, surface charge, elemental distribution, and surface ligand density. These

analyses were achieved using a combination of advanced techniques including dynamic light scattering

(DLS, Fig. 3b, d-g), transmission electron microscopy (TEM, Fig. 3h-n), , zeta potential measurements (Fig.

3c), and high-angle annular dark-field imaging (HAADF-STEM, Fig. 3o-t), and fluorescein isothiocyanate

analysis (FTIC, Fig. 3u). The experimental details can be found in the Methods section.

Figure 3. Characterization of peptide-conjugated nanoparticles (PCNPs). a Photos of the PCNPs reagent (right) and

its precursor (left). **b-k** The hydrodynamic radius and zeta potential (**b**), size distribution (**d-g**), and TEM images (**h-k**)

of the Fe3O4@PDA@Au, Fe_3O_4@PDA@Au-PP1, Fe_3O_4@PDA@Au-PP2, and Fe_3O_4@PDA@Au-PP3. **i-n** Representative

TEM images of nanoparticles Fe_3O_4, Fe_3O_4@PDA, and Fe_3O_4@PDA@Au. **o Representative** HAADF-STEM images

Fe_3O_4@PDA@Au. **p-t** EDS elemental mapping of the dispersion of all elements combined (**p**) and individual element

Au (**q**), Fe (**r**), C (**s**) and N (**t**). **u** FITC characterization showing the peptide conjugation density on the surface of

nanoparticles.

First, TEM, DLS and zeta potential measurements were utilized to characterize three peptide-conjugated

nanoparticles, namely Fe_3O_4@PDA@Au-PP1, Fe_3O_4@PDA@Au-PP2, and Fe_3O_4@PDA@Au-PP3, as well as

their precursor (Fe_3O_4@PDA@Au). The Fe_3O_4 nanoparticles exhibit a relatively coarse surface and a regular

spherical morphology with an average size of 200 nm (Fig. 3b), which is in line with the hydrodynamic

radius and size distribution shown in Fig. 3b, d. A thin layer of polydopamine (PDA) was coated on the

surface of Fe_3O_4 nanoparticles to stabilize it. Upon peptide modification, the particle size increased to

approximately 300 nm (Fig. 3b, e-g, i-k). The surface properties of the PCNPs were further assessed using

zeta-potential measurement. Initially, the nanoparticles displayed a negative zeta-potential of -41.09 ±

1.38 mV, which shifted to around -30 mV following peptide modification, indicating changes in surface

charge due to peptide conjugation (Fig. 3c).

Second, HAADF-STEM imaging (Fig. 3l–o) and EDS elemental mapping were performed to investigate the

nanoscale structures and elemental distributions of the PCNPs. HAADF-STEM images (Fig. 3l, m, n) show

the progression from Fe_3O_4 nanoparticles (l) to Fe_3O_4@PDA nanoparticles (m) to Fe_3O_4@PDA@Au

nanoparticles (n). A noticeable surface smoothing was observed after coating Fe_3O_4 with a thin PDA layer.

The subsequent attachment of Au onto Fe_3O_4@PDA was evident as discrete nanospheres, rather than a

continuous Au layer. EDS elemental maps (Fig. 3q–t) reveal that Au (Fig. 3q) and N (Fig. 3t) are

predominantly located on the nanoparticle's outer surface, while Fe (Fig. 3r) and C (Fig. 3s) are distributed

throughout. The combined elemental map (Fig. 3p) and the analysis effectively validate the structure and

composition of the synthesized nanoparticles.

Third, the density of peptide ligands on the outer layer was quantified using FITC. As shown in Fig. 2u, at

an initial FITC-polypeptide concentration of 24.70 μg.mg^{-1}, the input level of fluorescent polypeptide was

relatively low, yet the binding efficiency to magnetic beads was high, achieving a conjugation rate of 93%.

As the FITC-polypeptide concentration increased, the effective binding efficiency gradually declined. At 53

μg.mg^{-1}—consistent with the ligand density on the NP surface (PP1, PP2, PP3)—the conjugation rate

decreased to 86%, yielding an estimated ligand density of approximately 46 μg.mg^{-1}.

In conclusion, PCNPs can be visualized as a composition of three functional layers. The core layer consists

of superparamagnetic Fe_3O_4 nanoparticles with a uniform diameter of 200 nm and remarkable

homogeneity. The middle layer features a thin coating of Au particles, deposited as discrete nanospheres.

The outer layer comprises chemically modified peptide ligands with a density of approximately 46.0 μg.mg^{-}

^{1}. These features collectively endow each nanoparticle with excellent size uniformity and a high density of

binding sites, enabling the selective enrichment of low-abundance proteins within an extremely small

volume. Simultaneously, high-abundance proteins are depleted during sample preparation, facilitating the

detection of peptides from rare proteins in subsequent MS measurements.

Assessment of the plasma proteome coverage

To assess the proteomic detection depth of plasma samples processed with the Proteonano platform,

pooled healthy donor plasma samples were analyzed using both the conventional neat plasma pipeline

and the Proteonano workflow, followed by LC-MS/MS on the Orbitrap Astral instrument. As shown in Fig.

4a, the Proteonano workflow significantly improved plasma proteome coverage, identifying nearly 4000

protein groups compared to 900 with the conventional pipeline—a ~330% increase. Among these, ~2800

protein groups overlapped with proteins cataloged in the Human Plasma Proteome Project (HPPP),

whereas only ~800 overlapped with the neat plasma pipeline results (Fig. 4b). Additionally, the Proteonano

workflow detected 123 FDA-cleared circulating plasma biomarkers spanning eight orders of magnitude

from a single pooled plasma sample, markedly outperforming the neat plasma pipeline (Fig. 4c).

Given the plasma proteome's composition of secreted proteins from various tissues—both healthy and

diseased—it serves as an ideal matrix for disease biomarker discovery and pathogenesis research[29]. We

further investigated the secreted proteins linked to the hallmarks of cancer, ten defining traits of cancerous

cells: evading growth suppression, immune destruction, genome instability, replicative immortality,

reprogrammed energy metabolism, resisting cell death, sustained angiogenesis, proliferative signaling,

tissue invasion and metastasis, and tumor-promoting inflammation[30,31]. As shown in Fig. 4d, for each

hallmark, the Proteonano workflow identified 3–5 times more secreted proteins compared to the

unenriched process. The broader coverage of the plasma secretome associated with cancer progression

may position the Proteonano platform as a promising proteomic strategy for early diagnosis, subtyping,

and prognosis of various cancers.

Figure 4. Proteomic coverage of the Proteonano workflow in comparison to neat digestion workflow. a Protein

groups identified from a pooled healthy donor plasma sample processed using the Proteonano workflow and the

neat plasma pipeline. DIA LC-MS/MS measurement was conducted on an Orbitrap Astral instrument (1% protein and

peptide FDR; details in the Methods section). **b** Venn diagram comparing plasma proteomic coverage, highlighting

overlaps between protein groups identified by the two workflows and those cataloged in the HPPP. **c** Intensity

distribution of proteins groups detected matching the HPPP database. Intensities for proteins from neat plasma are

shown in the upper panel. Intensities for the Proteonano are shown in the bottom panel. Black circles indicated FDA-

approved protein biomarkers. **d** A comparison of the Proteonano workflow and the neat workflow regarding the

coverage of secreted proteins associated with 10 hallmarks of cancer progression.

Balance between proteome coverage and throughput on ten LC-MS/MS systems.

Besides proteome coverage, throughput (typically defined as samples per day, SPD) is a crucial

consideration when adopting proteomics platforms, as it directly affects the speed and scalability of data

generation. High-throughput platforms are essential for large-scale or time-sensitive studies, such as

biomarker discovery or population analyses. However, there is often a trade-off between throughput and

proteome coverage; some platforms prioritize speed, while others focus on achieving comprehensive

proteomic analysis at a slower pace. Selecting the right LC-MS/MS system depends on the specific goals,

budget, and requirements of the study.

To demonstrate the versatility of the Proteonano workflow, we assessed its coverage-throughput balance

across 10 distinct LC-MS/MS systems. These systems featured various combinations of liquid

chromatography columns and high-resolution mass spectrometry instruments. Detailed specifications are

provided in Supplementary Information Table S2.

We first evaluated the plasma proteome coverage of the Proteonano workflow using the Orbitrap Astral

instrument paired with three widely used LC columns (ES906, ES75550, μPac 110 cm), varying throughput

from 180SPD to 7SPD. As illustrated in Fig. 5a, even at the shortest gradient (180SPD), the workflow

identified 3168±9 protein IDs (AVG±SE, n=3) from a single pooled healthy plasma sample. Gradual

increases in gradient time enhanced proteome coverage, yielding 4945±5 protein IDs (AVG±SE, n=3) at

24SPD and 6410±4 protein IDs (AVG±SE, n=3) at 7SPD. Furthermore, with MS/MS parameters kept

constant, the proteomic detection depth was evaluated using the ES75500 and μPac 110 cm long LC

columns. At comparable throughputs (14SPD), both columns exhibited similar numbers of detected

protein groups, indicating the stability of the Proteonano workflow.

We further evaluated the impact of the Proteonano workflow on proteomic detection depth and

throughput using three mass spectrometers—Orbitrap Exploris 480, Orbitrap Astral, and timsTOF Pro 2—

all coupled with ES906 liquid chromatography. Technical assessment was conducted using three replicates

of a pooled healthy plasma sample. As shown in Fig. 5b, at 24SPD, Orbitrap Exploris 480 identified 2,320

± 11 protein IDs (AVG ± SE, n=3), while timsTOF Pro 2 detected 2,914 ± 22 protein IDs (AVG ± SE, n=3).

Orbitrap Astral exhibited the best performance, identifying 2,987 ± 12 protein groups (AVG ± SE, n=3) at

180SPD and 3,817 ± 4 protein groups (AVG ± SE, n=3) at 100SPD. All mass spectrometers showed low and

similar inter-sample quantitative CVs, with Orbitrap Astral achieving slightly lower CVs (8.10% at 100SPD

and 8.80% at 180SPD) compared to timsTOF Pro 2 (12.71% at 24SPD) and Orbitrap Exploris 480 (11.97%

at 24SPD) (See Fig. 5c).

These results demonstrate the excellent stability of the Proteonano workflow across various LC-MS/MS

systems. Additionally, its tunable and broad depth-throughput bandwidth, ranging from over 3000 protein

IDs at 180SPD and up to 6400 at 7SPD, makes it a desirable MS-based platform for deep plasma profiling

across different instrument configurations and study requirements.

Figure 5. Tunability of proteome coverage and throughput across 10 distinct LC-MS/MS systems. a Number of

proteins groups identified using different LC columns coupled with the Orbitrap Astral Instrument (DIA mode, 1%

protein and peptide FDR). Bar height indicates the mean protein groups detected of three replicates of a pooled

plasma sample. Individual dots represent protein groups detected in each replicate. **b** Venn diagram of protein groups

of identified in A. **b** Proteome coverage of different mass spectrometry coupled with the same ES906 liquid

chromatography column. **c** Distribution of CV values of protein intensity across different LC-MS/MS systems

described in **b**.

Technical Reproducibility of Proteonano Workflow

Affinity-based, non-MS multiplexed proteomics platforms have established comprehensive quality control measures as a fundamental approach to ensure data reproducibility. For instance, Olink implements a rigorous quality control system built on its Proximity Extension Assay (PEA) technology, incorporating validated biomarker panels and automated processes to reduce manual handling while maintaining data integrity. These measures include protein expression normalization and stringent panel performance validation[32]. Similarly, SomaLogic adopts a Total Quality Management approach focused on safety, efficacy, and reproducibility. Their quality control framework integrates normalization techniques, calibration, and adaptive signal adjustments to ensure consistent and accurate proteomic data[33].

We first evaluated whether variations between PCNP batches influenced proteomic analysis, as reproducibility of the enrichment reagent is vital for quality control. Supplementary Fig. 4 shows high pairwise Pearson correlation coefficients between two batches, confirming the strong reproducibility and consistency of PCNPs in enrichment-based plasma proteomics workflows.

We next assessed the baseline reliability of the LC-MS/MS system (Vanquish Neo liquid chromatography loaded with ES906 column, coupled with Orbitrap Astral) using 12 equal aliquots of the QC1 plasma sample from six batches, which were processed in parallel with the Proteonano workflow. The resulting freeze-dried peptides were dissolved, pooled, and consecutively injected for LC-MS/MS measurements (detailed in the Methods section). As shown in Fig. 6a and 6b, 3641±15.9 protein groups (AVG±SE, n=12) were identified, with an inter-plate coefficient of variation (CV) of 7.69% before and 3.38% after quantile normalization, based on the number of identified protein groups. We further compared the technical CVs of the low abundance proteins versus the medium- to high- abundance proteins. As shown in the boxplot

in Fig. 6c, low-abundance proteins displayed higher median CVs (6.65%, <10 ng.mL^{-1}) compared to

medium- (4.32%, 10–100 ng.mL^{-1}) and high-abundance proteins (3.86%, ≥100 ng.mL^{-1}). It is anticipated

that the higher abundance proteins would be more quantitatively reproducible, as there is a lower noise-

to-signal ratio for more reliable quantification.

Second, we evaluated the inter-batch reproducibility of the complete Proteonano workflow across six

independent batches with each batch including three replicates of QC3 sample. As the results shown in

Fib. 6d and 6e, the average number of protein groups identified in each batch kept relatively consistent,

ranging from the lowest of 4858 to the highest of 5170. We next compared the median CV values of the 6

batches. The median CV of the raw unadjusted intensities were in the range of 18.2% to 26.9%. After

quantile normalization, the median CVs dropped to 7.4% to 10.2%. Compared to the CVs in Fig. 6c, we

noticed a clear increase, which is probably due to the additional variations introduced by the sample

preparation steps prior to mass spectrometry measurements. It is reasonable to infer that the coefficient

of variation for the complete Proteonano workflow (Fig. 6e) represents the combined variability stemming

from the sample preparation process and the LC-MS/MS measurement (Fig. 6c). Based on this rationale,

the average CV attributed to the sample preparation procedure was determined to be 16.5%.

Third, the reproducibility of differentially expressed proteins (DEPs) across two independent projects was

assessed. Project 1 comprised six batches, each containing three replicates of the QC3 sample (P1QC3),

while Project 2 included three batches, each with three replicates of a different QC3 sample (P2QC3). Nine

P1QC3 samples were randomly divided into three groups, with the same grouping applied to the nine

P2QC3 samples. Pairwise comparisons were conducted between the three groups from Project 1 and the

three groups from Project 2. DEPs were identified based on $|\log_2(\text{fold change})| > 1.2$, and categorized as

Log2FC_1, Log2FC_2, and Log2FC_3, as illustrated in Fig. 6f. The analysis showed that over 90% of DEPs

were reproducibly measured across both projects. Furthermore, the Jaccard Similarity Index (JSI), a

measure of quantitative similarity between datasets, was calculated for the three Log2FC groups, resulting

in a value of 0.718. This high JSI score reflects excellent quantitative reproducibility of DEPs between the

independent projects.

In addition, all the replicates of QC samples analyzed here were examined for potential contamination

risks from platelets, erythrocytes, or coagulation, as highlighted by Geyer et al[15]. Quality marker panels

across all replicates of the two projects were systematically checked, and none exhibited signs of obvious

contamination (Supplementary Fig. S5, S6). Moreover, we compared the quantitative accuracy and fold-

change accuracy between the reference neat digestion workflow and the Proteonano workflow, the

results are Detailed results are presented in Supplementary Information, Fig. S7 and S8.

Figure 6. Investigation on the batch-to-batch reproducibility of the Proteonano workflow. a Baseline LC-MS/MS system reliability—number of identified protein groups (AVG±SE, n=12). **b** Inter-plate CVs before and after quantile normalization. **c** Technical variability comparison—boxplot illustrating median CVs of low-, medium-, and high-abundance proteins. **d** Inter-batch protein identification consistency—average number of protein groups across six independent batches. **e** Median CVs across batches and assessment of sample preparation variability. **g** Differentially

expressed proteins (DEPs)-pairwise comparisons of Log2FC groups across two independent projects. **f** Evaluation of

reproducibility using Jaccard Similarity Index.

Use of Proteonano workflow for neurodegenerative biomarker discovery

Emerging immunoassay-based and immuno-MS-based technologies have demonstrated excellent

accuracy in detecting well-established blood-based biomarkers linked to cognitive decline, such as plasma

Aβ, p-Tau, GFAP, NfL, and TDP-43[34,35]. However, these approaches are suitable for detecting a few validated

biomarkers for the purpose of clinical diagnosis for a specific disease. Typically, they fall short in providing

an unbiased discovery of novel plasma biomarkers, which could enable more nuanced patient stratification

and facilitate early screening—not solely for Alzheimer's pathology, but for neurodegenerative diseases in

general.

Affinity-based multiplexed proteomics platforms like Olink and SomaLogic have been employed in the

discovery of blood-based plasma biomarkers associated with cognitive decline[36]. However, these two

approaches generally rely on panels of pre-selected proteins and are unable to differentiate post-

translational modifications (PTMs), such as phosphorylation. In contrast, mass spectrometry offers an

unbiased approach to profiling the plasma proteome and PTMs and has the potential to discover novel

biomarkers. Nevertheless, it often faces challenges in detecting low-abundance blood-based biomarkers.

Therefore, MS is traditionally applied to brain tissues and cerebrospinal fluids rather than plasma for

neurological biomarker detection[37]. Thus, there remains a critical need for discovery tools that provide

deep proteomic coverage, robust performance, and cost-effectiveness, addressing current gaps in

biomarker identification.

We conducted a proof-of-concept study to demonstrate the feasibility of utilizing the Proteonano

workflow for the discovery of potential plasma biomarker candidates. 183 plasma samples were collected

from the Hubei Memory and Aging cohort Study (HAMCS) (ChiCTR; Registration number:

ChiCTR1800019164), which included two tests for global cognitive function (the Mini-Mental State

Examination [MMSE] and the basic Montreal Cognitive Assessment). The details of the population

characteristics can be seen in the previous publications[38,39] and Supplementary Table 3. In this experiment,

a total of 195 plasma samples (Control: n = 125; Cognitive decline: n = 58; QC: n=12) were processed in six

batches using the Proteonano workflow. The complete workflow was visualized in Fig. 7a.

We first assessed the quality of data obtained from 12 QC samples. As shown in Fig. 7b, an average of 3788

± 16 protein groups (AVG±SE, n=12) were identified, with a median CV of 16.9%, indicating high

quantitative consistency. For the cohort samples, an average of 2298 ± 38 protein groups (AVG±SE, n=183)

were identified. However, the median CV for protein raw intensities prior to quantile normalization was

48.7% (Fig. 7c), reflecting significant biological variability compared to technical variability. In their study

on biomarker discovery for Parkinson's disease via cerebrospinal fluid proteome profiling, Mann and

colleagues[40] highlighted that achieving technical variability lower than biological variability is crucial for

identifying disease-specific proteomic signatures. This observation emphasizes the importance of

producing high-quality and consistent data in biomarker research.

In addition, through differential protein expression analysis (Fig. 7d), we identified 8 upregulated proteins

groups and 49 downregulated proteins groups (FDR corrected p value < 0.05 and |log2FC| > 1) between

samples with and without cognitive function decline. Additionally, we matched these 57 DEPs with the

abundance data from the HPPP database and revealed that 13 of these proteins fall within the low

abundance range, defined as less than 10 ng mL^{-1}. The DEPs are listed in Supplementary Table 4. For

comparison, only 5 of the 13 low-abundance DEPs were detected in neat plasma without the use of

Proteonano workflow. Further, we conducted a cross-examination to determine whether these 13 low

abundance proteins have been detected by other plasma proteome profiling platforms. Notably, Soni et

al.[41] reported that 12 of the 13 DEPs were detectable using Seer's Proteograph XT platform.

Figure 7. The feasibility of using the Proteonano workflow for the discovery of potential biomarker candidates. a

Study design. **b** Number of proteins identified in the Cognitive Decline sample group and Control group. **c**

Quantification precision assessed by calculating the CVs of the proteins in samples of the cognitive decline and

control group. **d** Volcano plot of differentially expressed proteins in the control group and the cognitive decline group.

e ROC curves of best multivariate model based on features selected by random forest method. **f** Illustration of

selected proteins that are either increased (EIF1AX) or decreased (CFI, GSTO1, and HBB) in plasma with cognitive

decline when compared with the ones in the Control Group.

Furthermore, we found that several proteins among the DEPs have been previously linked to cognitive

decline in earlier research, indicating their potential significance in neurodegenerative processes. We

highlight seven proteins—GPX1, RAC1, RAC2, AZU1, S100A4, SUCLG1, and CDC42, all of which have been

implicated in various neurodegenerative processes. GPX1, a selenium-dependent antioxidant enzyme, has

shown protective effects against oxidative stress in neurodegenerative disorders[42]. RAC2, known for its

role in cytoskeletal dynamics, regulates neurite outgrowth, and its function is impaired when γ-adducin

cleavage leads to its downregulation[43]. CAP37 (AZU1), a protein released by neutrophils, promotes

microglial activation and neuroinflammation in Alzheimer's disease, with elevated levels in AD brains

correlating with disease progression[44]. Similarly, S100A4 is associated with mitochondrial dysfunction and

the accumulation of reactive oxygen species (ROS) in astrocytes affected by Alzheimer's disease,

contributing to neurodegenerative pathology[45]. SUCLG1, an essential enzyme in the TCA cycle, shows

decreased modifications in Alzheimer's disease; however, treatment with β-hydroxybutyrate has been

shown to enhance its activity and ATP production, suggesting potential therapeutic avenues[46]. Lastly,

CDC42, a key regulator of cell signaling, is involved in Alzheimer's disease-related cognitive deficits and

dendritic spine loss, with its decline directly correlating with disease progression and cognitive

impairment[47,48]. These findings emphasize the significant roles of these proteins in neurodegenerative

processes, highlighting them as potential targets and biomarkers for further research and therapeutic

development.

Although many proteins require further investigation, we explored whether a small panel of proteins could

reliably distinguish individuals with cognitive decline from those without. Differential protein expression

data were analyzed using a random forest approach, which identified key features separating the two

cohort groups. Based on these features, optimal multivariate models were constructed using the Akaike

Information Criterion. The best model incorporated four proteins—EIF1AX, CFI, GSTO1, and HBB—and

achieved an excellent AUC value of 0.92. Fig. 7e presents a receiver operating characteristic (ROC) curve,

demonstrating the model's ability to distinguish participants in the Cognitive Decline Group from the

Control. Notably, GSTO1's role in modulating oxidative stress stands out as a significant factor that may

contribute to the underlying pathophysiological mechanisms of neurodegeneration[49]. This suggests its

potential as a proteomic signature for patient stratification. However, transition from biomarker discovery

to validation is a lengthy, time-consuming, and oftentimes expensive process. It involves critical steps of

study design, including cohort selection, evaluation of statistical power, sample blinding, randomization,

and rigorous data quality control[50–53]. While a detailed discussion of these best practices exceeds the scope

of this study. Further efforts on the validation of their candidacy as potential biomarkers are necessary

and will be the focus of our future work.

For cross-referencing, another nanoparticle-based MS proteomics platform, known as Proteograph XT, was

utilized by Lucar et al.[54] to analyze plasma samples from 379 AD patients and 240 healthy controls using a

linear mixed-effects model to identify AD-associated proteins. Their findings revealed that proteins

upregulated in AD included DKK2, MAPK3, HLA-DRA, LPL, and ACSL4, among others, while downregulated

proteins included ETFA, TFRC, ADH1B, MGST3, and TNFRSF6B. Conversely, the Mag-Net enrichment

strategy examined plasma samples from 10 AD patients and 10 healthy controls, applying a two-sided

Mann-Whitney U test to identify AD-associated proteins[11]. This approach identified upregulated proteins

in AD such as INPPL1, VPS13A, DNAJC10, S100B, and HTT, while downregulated proteins included VCP,

MAP2K2, USP15, JMJD6, and CD36. Notably, our study observed a consistent trend in the differential

expression of VCP with results reported by the Mag-Net strategy. However, none of the well-established

blood biomarkers—such as Aβ, p-Tau, GFD, and TDP-42—were identified by any of those three

nanoparticle-enriched mass spectrometry workflows. We hypothesize that factors like cohort

characteristics, sample size, population diversity, and nanoparticle enrichment methodology may have

contributed to this outcome. Further investigations with refined experimental designs are warranted to

explore these observations more comprehensively.

Discussion

This study developed and systematically evaluated the Proteonano Ultraplex Platform for robust, deep

plasma proteomic analysis. When paired with advanced mass spectrometers, the platform demonstrated

exceptional capability for deep proteomic analysis of low-input plasma samples, achieving extensive

proteome coverage, high throughput, a dynamic range spanning nine orders of magnitude, and minimized

batch-to-batch variability. Furthermore, we demonstrated that the Proteonano workflow exhibited

remarkable stability and reproducibility across ten distinct and widely adopted DIA LC-MS/MS systems,

showcasing flexibility in balancing protein depth and sample processing speed. Finally, the platform's

performance was validated using plasma samples from an elderly community cohort, including individuals

with neurodegenerative diseases.

The PUP Platform described here offers several advantages for MS-based plasma proteomics studies,

especially in the discovery phase of biomarker screening. Unlike traditional immunodepletion methods,

our platform is highly standardized, robust, high-throughput, user-friendly, cost-effective, and compatible

with various mass spectrometry configurations across different laboratories. Notably, the workflow

requires only 40 µL of plasma per sample to achieve optimal performance, whereas previously reported

technologies typically require at least 100 µL[12]. This critical feature, which has not been emphasized

enough, plays a key role in the practical adoption of new technologies, especially given the common

limitations on sample volumes. In comparison, Olink PEA technology, widely adopted by dozens of

biopharmaceutical companies and applied to millions of samples globally, requires only 10 µL of plasma.

Secondly, the PUP Platform achieved cutting-edge deep plasma proteome coverage and data quality by

effectively combining multiple nanoparticles into a single, uniform reagent, a strategy akin to those utilized

in multiplexed immunoassays but without the challenges of antibody cross-reactions. This innovative

approach removes the need for sample fractionation and multiple mass spectrometry runs for individual

samples. As a result, it significantly minimizes time, cost, and batch effect in plasma proteomics studies,

making it particularly advantageous for large cohorts involving hundreds or even thousands of samples. In

the future, we plan to enhance nanoparticle technology by integrating a wider range of compositions into

a single reagent to enrich a broader range of low abundance plasma proteins. This innovation aims to

achieve deeper plasma proteome coverage through MS-based proteomics, bringing its capabilities closer

to those of non-MS multiplex technologies regarding proteome coverage, such as SomaScan 7K and 11K[55].

Thirdly, the PUP Platform has implemented a fully standardized workflow, starting with nanoparticle-

plasma incubation at the beginning of the experiment and extending to data analysis across various

batches. This carefully structured system incorporates a robust quality control framework, guarantees

compatibility with multiple mainstream mass spectrometers, supports deep plasma coverage, and

minimizes batch-to-batch effect over independent studies. This level of standardization significantly

improves reliability and efficiency, solidifying its applicability to large-scale plasma proteomics studies.

That said, we have not included discussions of other considerations that may be associated with the

proteome coverage and data reproducibility of nanoparticle-enriched plasma proteomics. For instance,

the effect of nanoparticle size and size distribution on the compositions of protein corona is less studied.

Notably, Tenzer et al.[23] demonstrated the significance of nanoparticle size in shaping corona layers by using

monodispersed silica nanoparticles of varying sizes (20, 30, and 100 nm). Their findings revealed that

nanoparticle size critically influenced the composition of the corona, which accounted for 37% of the

identified proteins in their study[23]. Yet, the effect of size uniformity on the coverage and the reproducibility

of proteomic analysis remains unexamined. In this context, the surface-modified nanoparticles presented

here offer an average diameter of 300 nm and exhibit exceptional size uniformity (PDI < 0.05, Fig. 3b). In

comparison, particles documented by Blume[10] and Wu[11] are either characterized by relatively high

polydispersity or, in the case of microparticles, lack detailed data on size distribution. Alternative

approaches, such as graphene oxide-modified materials, have also been explored for proteome

enrichment[56]. However, the highly heterogeneous nature of these flake-like nanomaterials may pose

substantial challenges on quality control, particularly when employed in large cohort plasma studies.

Although the precise mechanisms remain unclear, we hypothesize that employing nanoparticles with

defined size and low polydispersity is crucial for ensuring quality control and enhancing reproducibility in

plasma proteomics applications[57].

In addition, we recognize that the proof-of-concept study using the PUP Platform for biomarker discovery

is subject to several limitations. First, mass spectrometers inherently face significant challenges in

detecting low-abundance proteins in plasma. This limitation is particularly evident for biomarkers of

neuropathology, which are present at very low concentrations in plasma (typically a few pg.mL^{-1}). That

explains partly why none of the three reported nanoparticle-enriched and MS-based workflows detected

low abundant biomarkers like p-Tau181. We cannot rule out the possibility that these AD-related

biomarkers were also present in this cohort but at concentrations below the detection sensitivity of mass

spectrometry, even after enrichment. While the PrecivityAD2 assay also relies on mass spectrometry, it

incorporates the use of antibody-conjugated microbeads to enrich specific AD biomarkers from plasma

prior to LC-MS/MS measurements. Second, it is evident that the cohort is relatively small and lacks

diversity. Mann and colleagues proposed a "rectangular" plasma proteome profiling strategy that

suggested [29] typically 100 to 1000 samples should be used in the phase of plasma biomarker discovery.

Therefore, we anticipate that PUP Platform could serve as a valuable discovery-phase tool, enabling the

profiling of thousands of plasma proteins to identify potential biomarkers. These identified biomarkers can

then undergo further validation in larger cohorts using targeted proteomics techniques such as MS-based

Parallel Reaction Monitoring (PRM) and single-molecule immunoassays like Simoa, or even low-cost ELISA.

In summary, our study highlights the Proteonano Ultraplex Proteomics Platform as an innovative solution

for MS-based deep plasma proteomics. This platform offers extensive proteomic coverage, high-

throughput capabilities, cost efficiency, and a standardized data production workflow. It supports minimal

plasma input (40 μL), demonstrates compatibility with various LC-MS/MS systems and throughput levels,

and ensures batch-to-batch consistency, reproducibility, and precision in protein identification and analysis

in large-cohort studies. These advancements offer exceptional potential to push the frontiers of plasma

proteomics and pave the way for the discovery of novel protein biomarkers and therapeutic targets across

a range of diseases. In future endeavors, we aim to expand our research by evaluating the performance of

the PUP platform across multiple laboratories and conducting larger, multi-centered plasma proteomics

studies to further validate its robustness and scalability.

Methods

Blood Samples

For most of the experiments, pooled healthy donor blood samples were obtained from Oricells (Ori

Biotech, Shanghai, China), Whole blood was collected in K2-EDTA containing tubes. Samples were

centrifuged at 1500 g for 15 min at 4 °C. Supernatant was transferred from centrifuge tubes. Plasma

samples from multiple donors were pooled, aliquoted into microfuge tubes, and stored at −80 °C until

analysis.

Plasma samples collected from a community cohort aimed at understanding cognitive deficiencies in

elderly patients in Hubei Province (Hubei Memory and Aging Cohort Study, HMACS) were utilized. Ethical

approval for blood sample collection was received at respective institutions for the study.

Preparation and Characterization of Nanoparticles

The magnetic Fe_3O_4 particles were prepared according to a solvothermal approach. Typically, $FeCl_3 \cdot 6H2O$

(3.91 g) and trisodium citrate (0.46 g) were first dissolved in ethylene glycol (70 mL). Then, NaAc (9.42 g)

was added, and the mixture was vigorously stirred to form a transparent solution. Afterward, the solution

was transferred to a 150 mL Teflon-lined stainless-steel autoclave. The autoclave was sealed and heated

at 200 °C and maintained for 12 h, and then it was allowed to cool to room temperature. The products

were washed with ethanol and deionized water several times and dried at 60 °C for 12 h.

In order to prepare monodispersed Fe_3O_4@PDA (polydopamine) core-shell nanospheres, 180 mg of the

hydrochloride dopamine was dispersed in 90 ml Boric acid-borax buffer (0.5 $mol.L^{-1}$, pH=8.0). Then, 90

mg of the as-prepared Fe_3O_4 was added, stirring overnight at room temperature. The resultant product

was separated and collected with a magnet and then washed with ultrapure water three times.

In a typical preparation of Fe_3O_4@PDA@Au, firstly, 20 mg of the Fe_3O_4@PDA was added into 12.5 mL

deionized water, sonicate for 10 min. Then, 0.5 mL dopamine solution (0.76 mg of the hydrochloride

dopamine was dispersed in 1 mL deionized water) and 83.5 µL sodium hydroxide standard solution (0.6M)

were added in sequence. Typically, 0.8 mL 1 wt% aqueous solution of $HAuCl_4$ (1 g $HAuCl_4$ was dispersed in

100 mL deionized water) was quickly added and reaction was allowed for overnight at room temperature

under shaking at 1000 rpm. Finally, the product was separated by a magnet and washed with water.

The nanomaterial sample was ultrasonically dispersed in ethanol, and a drop of the solution was then

placed onto an ultra-thin carbon film. After drying, the sample was analyzed using a transmission electron

microscope (TEM). The instrument used was an FEI Tecnai G2 F20 with a maximum acceleration voltage

of 200 kV. It has a point resolution of 0.24 nm, a line resolution of 0.102 nm, and an information resolution

of ≤0.14 nm. Energy dispersive spectroscopy (EDS) was performed using an Oxford XPLORE system.

Well dispersed Fe_3O_4 microspheres were prepared by a solvothermal method[58]. Fe_3O_4 microspheres were

dispersed in chloroauric acid containing sodium citrate aqueous solution and stirred at 90 °C for 0.5 h.

Resulting Au NPs-deposited Fe_3O_4 (Fe_3O_4@PDA@Au) were magnetically separated from the suspension

and subsequently washed dried under vacuum.

For peptide conjugation, peptides were dissolved in water. Dissolved peptide was incubated with Tris(2-

carboxyethyl) phosphine hydrochloride (TCEP) solution, followed by adding of Fe_3O_4@PDA@Au

synthesized above. The reaction mixture was incubated at 25°C overnight. Reaction product was washed

by ethanol then deionized water and stored at refrigerator at 4°C. Three peptides, PP1

(HKAATKIQASFRGHITRKKLC), PP2 (DIEEVEVRSKYFKKNERTVEC), and PP3 (DIEEVEVRSKYFKKNERTVEC) were

custom synthesized by GenScript (Suzhou, China). The details can be seen in the Supporting Information

S1 section.

Following peptide conjugated $Fe_3O_4@PDA@Au$ synthesis, transmission electron microscopy (TEM) was

performed by using a Tecnai 12 electron microscope (ThermoFisher Scientific) at an accelerating voltage

of 200 kV. DLS and Zeta potential of the particles were examined by using a NanoBrook 90Plus PALS

instrument (Brookhaven Instruments, Holtsville, NY).

Peptide Conjugation Density

First, 125 μL of TCEP solution (200 mM) and 5 mL of FITC-peptide solution (0.2 mg.mL^{-1}) were added to a

centrifuge tube, mixed thoroughly, and incubated in the dark for 2 hours. Then, varying volumes (0, 200,

400, 600, and 1200 μL) of the TCEP-treated FITC-peptide solution (fluorescence intensity denoted as I_0)

were transferred into separate 5 mL centrifuge tubes. Each tube was supplemented with 1000 μL of

$Fe_3O_4@PDA@Au$ NPs (1.5 mg.mL^{-1}). Ultra-pure water was then added to bring the final volume to 2300

μL, followed by ultrasonic mixing for 5 minutes. The reaction mixtures were incubated at 25°C under

shaking conditions for 17 hours. After the reaction, the samples were washed three times with ultra-pure

water, and the supernatants from the three washes were collected (fluorescence intensity denoted as I_1).

All fluorescence intensity values were measured in triplicate using a Tecan SPARK multifunctional

microplate reader.

Sample Preparation by the Proteonano Workflow

Automatic sample preparation was performed in batches utilizing Nanomation G1 workstation, equipped

with both a magnetic module and a heater shaker module. Plasma samples were dispensed on 96 well flat

bottom assay plates for subsequent processing. For most experiments, 40 µL of human plasma was diluted

to a final volume of 100 µL by using 1 x PBS at pH7.4 and was subsequently combined with PCNPs (as

synthesized above) from the Proteonano reagent in a 1:1 volumetric ratio. The mixture was incubated at

25 °C and agitated for 60 min. Following magnetic immobilization, PCNPs were washed thrice with 1 x PBS.

Proteins captured on the PCNPs were reduced by 20 mM DTT at 37 °C for 60 min. Alkylation was performed

using 50 mM iodoacetamide (IAA) at room temperature in darkness for 30 min. Trypsin (Promega

Corporation, Madison, WI, USA) digestion was carried out at 37°C for a duration of 16 hours with shaking

at 1000 rpm. Post-digestion, peptides were purified using desalting C18 columns in micropipette tip format

(ThermoFisher Scientific) and lyophilized with a LyoQuest freeze dryer (LyoQuest, Telstar, Terrassa, Spain).

Lyophilized peptides were then reconstituted in 0.1 % formic acid prior to mass spectrometry. Peptide

concentrations were measured with a Nano300 microvolume spectrophotometer (Allsheng Instruments,

China).

LC-MS/MS Experiments

Multiple LC-MS/MS instruments were used in the study. For most studies, Orbitrap Astral (ThermoFisher

Scientific) mass spectrometer was used. In some cases, Orbitrap Exploris 480 (ThermoFisher Scientific)

mass spectrometer equipped with FAIMS, or timsTOF Pro 2 mass spectrometer (Bruker Instruments) were

used.

For studies using Orbitrap Astral instrument, 300 ng or 500 ng peptides dissolved in 0.1% formic acid were

separated by a Vanquish™ Neo UHPLC system (ThermoFisher Scientific) followed by mass spectrometry.

Data were acquired in data independent acquisition mode (Detailed experimental setup are listed in

Supplementary Table S5-14).

For studies using Orbitrap Exploris 480, 500 ng of peptides were separated by an Easy-nLC1200 reverse-

phase HPLC system (ThermoFisher Scientific) using a precolumn (homemade, 0.075 mm × 2 cm, 1.9 μm,

C18) and a self-packed analytical column (0.075 mm × 20 cm, 1.9 μm, C18) over a 48 min gradient before

nano-electrospray on Orbitrap Exploris 480 mass spectrometer equipped with FAIMS (ThermoFisher

Scientific). Solvent A was 0.1 % formic acid and solvent B was 80 % acetonitrile (ACN)/0.1 % formic acid.

Gradient conditions were 3–7 % B (0–1 min), 7–30 % B (1–36 min), 30–95 % B (36–38 min), and 95 % B

(38–48 min). Mass spectrometer was operated in DIA mode. Spray voltage was set to 2.4 kV, RF Lens level

at 40 %, and heated capillary temperature at 320 °C. Full MS resolutions were set to 60,000 at m/z 200

and full MS automatic gain control (AGC) target was 100 % with an injection time (IT) of 50 ms. Mass range

was set to m/z 350–1200. The AGC target value for fragment spectra was set at 1,000 %. Resolution was

set to 30,000 and IT to 54 ms. Normalized collision energy was set at 30 %. Default settings were used for

FAIMS with voltages applied as -45 V, except gas flow, which was applied with 3.5 L/min.

For studies using timsTOF Pro 2 instrument, 200 ng of peptides was subjected to a nanoElute® 2 nanoLC

system coupled with a Bruker timsTOF Pro 2 mass spectrometer using a trap-and-elute configuration. First,

the peptides were loaded on an Acclaim PepMap 100 C18 trap column (0.3 mm ID × 5 mm), then separated

on a self-packed analytical column (0.075 mm × 25 cm, 1.8 μm, C18) at a flow rate of 300 nL min^{-1}. Solvent

A was 0.1 % formic acid and solvent B was 80 % ACN/0.1 % formic acid. Gradient conditions were 2–22 %

B (0–45 min), 22–35 % B (45–50 min), 35–80 % B (50–55 min), 80 % B (55–60 min). The spray voltage was

set to 1.5 kV. The mass spectrometer was operated in diaPASEF mode using ion mobility range of 0.76–

1.29 Vs/cm^2 with 100 ms accumulation time. MS1 mass spectrometry scans 452-1152 m/z with peak

height above 2,500 before being detected. The 452-1152 m/z range was divided into four steps. Each step

was divided into seven windows, the Number of Mobility Windows was set to 2, and a total of 56 windows

for continuous window shattering and information gathering. The splitting mode was CID and the splitting

energy was set 20-59 eV. The mass width of each window was 25 Da and the cycle time for a DIA scan was

1.59 s.

Proteomic Data Analysis

The *.raw* or *.d* files were obtained directly from mass spectrometers, or *mzML* files converted from .raw

files by msConvert[59] (Version 3.0) software were searched using DIA-NN (Version 1.8.1) in library free mode.

For each sample, spectra were searched against a UniProt *Homo sapiens* reviewed proteome dataset, or

other datasets. DIA-NN search parameters were: 10 ppm mass tolerance for mass accuracy, one missed

cleavages of trypsin, carbamidomethylation of cysteine as fixed modification, and methionine oxidation as

the only variable modification. The rest of the parameters were set to default. The FDR cutoffs at both

precursor and protein level were set to 0.01.

The .raw or .d files generated directly from mass spectrometers, or mzML files converted from .raw files

using msConvert (Version 3.0) software, were analyzed with DIA-NN (Version 1.8.1) in library-free mode.

For each sample, spectra were matched against a UniProt Homo sapiens reviewed proteome dataset or

other relevant datasets. The DIA-NN search parameters included a 10 ppm mass tolerance for mass

accuracy, one missed cleavage for trypsin, carbamidomethylation of cysteine as a fixed modification, and

methionine oxidation as the sole variable modification. All other parameters were set to default, with false

discovery rate (FDR) cutoffs of 0.01 applied at both the precursor and protein levels.

Statistical Analysis and Data Visualization

Coefficient of variation (CV) for each protein was determined by dividing its empirical standard deviation

by its empirical mean, and CV analyses were performed on raw intensities and quantile-normalized

intensities, respectively. Median values were reported as overall coefficient of variation. Pearson

correlation analyses and linear regression were conducted using Pingouin[59]. No missing value imputation

was performed during the above analysis, and calculations were performed only for proteins that were

identified in all samples. For replicate experiments, protein identifications are expressed as AVG±SE.

Seaborn was used to generate bar charts and violin charts[60]. The matplotlib-venn[61] package

(https://pypi.org/project/matplotlib-venn/) and the Venn[62] (https://pypi.org/project/venn/) package

were used to plot VENN[39] plots. MS-DAP was used to screen for differentially expressed proteins, and

ggVolcano[63] (https://github.com/BioSenior/ggVolcano) was used to generate volcano plot. KEGG pathway

enrichment analysis was performed using the GSEApy[64] software package in Python. Feature selection was

performed using MetaboAnalystR[65]. Multivariate analysis was performed using MASS[66] in R. Variables

were first processed by glm (MASS package in R), followed by Akaike information criterion (AIC)

determination using stepAIC (MASS package in R) with forward selection and Wald test. Multivariate

Receiver operating characteristic(ROC) analysis was subsequently performed using roc (glmtoolbox[67]

package in R) to assess the predictive performance of the logistic regression model selected by step AIC.

ROC curves were graphed using ROC-utils[68].

Data Availability

All raw and processed proteomics data generated in this study have been deposited to the iProX repository

and are available for review. The data can be accessed at

https://www.iprox.cn/page/PSV023.html?url=1744530654097Tmnc using the password dO25. The data

are currently under controlled access for peer review and will become publicly available after the embargo

period. Alternatively, the data can be made available upon reasonable request by contacting the

corresponding author.

References

1.   Zhong, W. *et al.* Next generation plasma proteome profiling to monitor health and disease. *Nat.*
Commun. **12**, 2493 (2021).

2.   Deng, Y.-T. *et al.* Atlas of the plasma proteome in health and disease in 53,026 adults. *Cell* **188**,
253-271.e7 (2025).

3.   He, B., Huang, Z., Huang, C. & Nice, E. C. Clinical applications of plasma proteomics and
peptidomics: Towards precision medicine. *Proteomics Clin. Appl.* **16**, e2100097 (2022).

4.   Giudice, G. & Petsalaki, E. Proteomics and phosphoproteomics in precision medicine:
applications and challenges. *Brief. Bioinform.* **20**, 767–777 (2017).

5.   Yu, X., Schwenk, J. M., Xu, P. & LaBaer, J. Advances in plasma proteomics: Moving from
technology to precision medicine. *PROTEOMICS – Clin. Appl.* **16**, 2200083 (2022).

6.   Soni, R. K. Frontiers in plasma proteome profiling platforms: innovations and applications. *Clin.*
Proteomics **21**, 43 (2024).

7.   Schmidt, I. M. *et al.* Plasma proteomics of acute tubular injury. *Nat. Commun.* **15**, 7368 (2024).

8.   Liu, W.-S. *et al.* Plasma proteomics identify biomarkers and undulating changes of brain aging.
Nat. Aging **5**, 99–112 (2025).

9.   Guo, T., Steen, J. A. & Mann, M. Mass-spectrometry-based proteomics: from single cells to
clinical applications. *Nature* **638**, 901–911 (2025).

10.   Blume, J. E. *et al.* Rapid, deep and precise profiling of the plasma proteome with multi-
nanoparticle protein corona. *Nat. Commun.* **11**, 3662 (2020).

11.   Wu, C. C. *et al.* Mag-Net: Rapid enrichment of membrane-bound particles enables high coverage
quantitative analysis of the plasma proteome. *BioRxiv Prepr. Serv. Biol.* 2023.06.10.544439 (2024)
doi:10.1101/2023.06.10.544439.

12.   Beimers, W. F. *et al.* A Technical Evaluation of Plasma Proteomics Technologies. *BioRxiv Prepr.*
Serv. Biol. 2025.01.08.632035 (2025) doi:10.1101/2025.01.08.632035.

13.   Liu, X. *et al.* Plasma-to-tumour tissue integrated proteomics using nano-omics for biomarker
discovery in glioblastoma. *Nat. Commun.* **16**, 3412 (2025).

14.   Collins, B. C. *et al.* Multi-laboratory assessment of reproducibility, qualitative and quantitative
performance of SWATH-mass spectrometry. *Nat. Commun.* **8**, 291 (2017).

15.   Geyer, P. E. *et al.* Plasma Proteome Profiling to detect and avoid sample-related biases in
biomarker studies. *EMBO Mol. Med.* **11**, e10427 (2019).

16.   Gan, Y.-H. *et al.* Large-scale proteomic analyses of incident Parkinson's disease reveal new
pathophysiological insights and potential biomarkers. *Nat. Aging* (2025) doi:10.1038/s43587-025-00818-
0.

17.   Ashkarran, A. A. *et al.* Measurements of heterogeneity in proteomics analysis of the nanoparticle
protein corona across core facilities. *Nat. Commun.* **13**, 6610 (2022).

18.   Hajipour, M. J. *et al.* An Overview of Nanoparticle Protein Corona Literature. *Small Weinh.*
Bergstr. Ger. **19**, e2301838 (2023).

19.   Gharibi, H. *et al.* A uniform data processing pipeline enables harmonized nanoparticle protein
corona analysis across proteomics core facilities. *Nat. Commun.* **15**, 342 (2024).

20.   Walkey, C. D. & Chan, W. C. W. Understanding and controlling the interaction of nanomaterials
with proteins in a physiological environment. *Chem. Soc. Rev.* **41**, 2780–2799 (2012).

21.   Walkey, C. D. *et al.* Protein Corona Fingerprinting Predicts the Cellular Interaction of Gold and
Silver Nanoparticles. *ACS Nano* **8**, 2439–2455 (2014).

22.   Tenzer, S. *et al.* Rapid formation of plasma protein corona critically affects nanoparticle
pathophysiology. *Nat. Nanotechnol.* **8**, 772–781 (2013).

23.   Tenzer, S. *et al.* Nanoparticle size is a critical physicochemical determinant of the human blood
plasma corona: a comprehensive quantitative proteomic analysis. *ACS Nano* **5**, 7155–7167 (2011).

24.   Ritz, S. *et al.* Protein Corona of Nanoparticles: Distinct Proteins Regulate the Cellular Uptake.
Biomacromolecules **16**, 1311–1321 (2015).

25.   Hofmann, D. *et al.* Mass Spectrometry and Imaging Analysis of Nanoparticle-Containing Vesicles
Provide a Mechanistic Insight into Cellular Trafficking. *ACS Nano* **8**, 10077–10088 (2014).

26.   Landgraf, L. *et al.* A plasma protein corona enhances the biocompatibility of Au@Fe3O4 Janus
particles. *Biomaterials* **68**, 77–88 (2015).

27.   Chithrani, B. D. & Chan, W. C. W. Elucidating the Mechanism of Cellular Uptake and Removal of
Protein-Coated Gold Nanoparticles of Different Sizes and Shapes. *Nano Lett.* **7**, 1542–1550 (2007).

28.   Heil, L. R. *et al.* Evaluating the Performance of the Astral Mass Analyzer for Quantitative
Proteomics Using Data-Independent Acquisition. *J. Proteome Res.* **22**, 3290–3300 (2023).

29.   Geyer, P. E., Holdt, L. M., Teupser, D. & Mann, M. Revisiting biomarker discovery by
plasma proteomics. *Mol. Syst. Biol.* **13**, 942 (2017).

30.   Hanahan, D. & Weinberg, R. A. The hallmarks of cancer. *Cell* **100**, 57–70 (2000).

31.   Hanahan, D. & Weinberg, R. A. Hallmarks of cancer: the next generation. *Cell* **144**, 646–674
(2011).

32.   Wik, L. *et al.* Proximity Extension Assay in Combination with Next-Generation Sequencing for
High-throughput Proteome-wide Analysis. *Mol. Cell. Proteomics MCP* **20**, 100168 (2021).

33.   Candia, J. *et al.* Assessment of Variability in the SOMAscan Assay. *Sci. Rep.* **7**, 14248 (2017).

34.   Chatterjee, P. *et al.* Plasma Aβ42/40 ratio, p-tau181, GFAP, and NfL across the Alzheimer's
disease continuum: A cross-sectional and longitudinal study in the AIBL cohort. *Alzheimers Dement. J.*
Alzheimers Assoc. **19**, 1117–1134 (2023).

35.   Schindler, S. E. *et al.* Head-to-head comparison of leading blood tests for Alzheimer's disease
pathology. *Alzheimers Dement.* **20**, 8074–8096 (2024).

36.   Teunissen, C. E. *et al.* Methods to Discover and Validate Biofluid-Based Biomarkers in
Neurodegenerative Dementias. *Mol. Cell. Proteomics MCP* **22**, 100629 (2023).

37.   Johnson, E. C. B. *et al.* Large-scale proteomic analysis of Alzheimer's disease brain and
cerebrospinal fluid reveals early changes in energy metabolism associated with microglia and astrocyte
activation. *Nat. Med.* **26**, 769–780 (2020).

38.   Li, L. *et al.* The Hubei Memory and Aging Cohort Study: Study Design, Baseline Characteristics,
and Prevalence of Cognitive Impairments. *J. Alzheimers Dis. JAD* **85**, 561–571 (2022).

39.   Cheng, G.-R. *et al.* Prevalence and risk factors for subjective cognitive decline and the correlation
with objective cognition among community-dwelling older adults in China: Results from the Hubei
memory and aging cohort study. *Alzheimers Dement. J. Alzheimers Assoc.* **19**, 5074–5085 (2023).

40.   Karayel, O. *et al.* Proteome profiling of cerebrospinal fluid reveals biomarker candidates for
Parkinson's disease. *Cell Rep. Med.* **3**, 100661 (2022).

41.   Soni, R. K. Frontiers in plasma proteome profiling platforms: innovations and applications. *Clin.*
Proteomics **21**, 43 (2024).

42.   Sharma, G. *et al.* Glutathione peroxidase-1 and neuromodulation: Novel potentials of an old
enzyme. *Food Chem. Toxicol.* **148**, 111945 (2021).

43.   Xiong, M. *et al.* A γ-adducin cleavage fragment induces neurite deficits and synaptic dysfunction
in Alzheimer's disease. *Prog. Neurobiol.* **203**, 102074 (2021).

44. Stock, A. J., Kasus-Jacobi, A. & Pereira, H. A. The role of neutrophil granule proteins in
neuroinflammation and Alzheimer's disease. *J. Neuroinflammation* **15**, 240 (2018).

45. Lampinen, R. *et al.* Neuron-astrocyte transmitophagy is altered in Alzheimer's disease.
Neurobiol. Dis. **170**, 105753 (2022).

46. Han, W. *et al.* Ketogenic β-hydroxybutyrate regulates β-hydroxybutyrylation of TCA cycle-
associated enzymes and attenuates disease-associated pathologies in Alzheimer's mice. *Aging Cell* **24**,
e14368 (2025).

47. Zhu, M. *et al.* Cdc42GAP deficiency contributes to the Alzheimer's disease phenotype. *Brain* **146**,
4350–4365 (2023).

48. Zhang, Y. & Niu, C. Relation of CDC42, Th1, Th2, and Th17 cells with cognitive function decline in
Alzheimer's disease. *Ann. Clin. Transl. Neurol.* **9**, 1428–1436 (2022).

49. Allen, M. *et al.* Glutathione S-transferase omega genes in Alzheimer and Parkinson disease risk,
age-at-diagnosis and brain gene expression: an association study with mechanistic implications. *Mol.*
Neurodegener. **7**, 13 (2012).

50. Nakayasu, E. S. *et al.* Tutorial: best practices and considerations for mass-spectrometry-based
protein biomarker discovery and validation. *Nat. Protoc.* **16**, 3737–3760 (2021).

51. Frantzi, M., Bhat, A. & Latosinska, A. Clinical proteomic biomarkers: relevant issues on study
design & technical considerations in biomarker development. *Clin. Transl. Med.* **3**, 7 (2014).

52. Mischak, H. *et al.* Recommendations for Biomarker Identification and Qualification in Clinical
Proteomics. *Sci. Transl. Med.* **2**, 46ps42-46ps42 (2010).

53. Ignjatovic, V. *et al.* Mass Spectrometry-Based Plasma Proteomics: Considerations from Sample
Collection to Achieving Translational Data. *J. Proteome Res.* **18**, 4085–4097 (2019).

54. Lacar, B. *et al.* Identification of Novel Biomarkers for Alzheimer's Disease and Related Dementias
Using Unbiased Plasma Proteomics. 2024.01.05.574446 Preprint at
https://doi.org/10.1101/2024.01.05.574446 (2024).

55. Kirsher, D. Y. *et al.* The Current Landscape of Plasma Proteomics: Technical Advances, Biological
Insights, and Biomarker Discovery. 2025.02.14.638375 Preprint at
https://doi.org/10.1101/2025.02.14.638375 (2025).

56. Zhang, B. *et al.* Nitrogen-modified reduced graphene oxide for serum enrichment of N-glycans
and MALDI-TOF MS-based identification of HCC biomarkers. *The Analyst* **150**, 650–660 (2025).

57. Mahmoudi, M. The need for improved methodology in protein corona analysis. *Nat. Commun.*
13, 1–4 (2022).

58. Deng, H. *et al.* Monodisperse Magnetic Single-Crystal Ferrite Microspheres. *Angew. Chem. Int.*
Ed. **44**, 2782–2785 (2005).

59. Vallat, R. Pingouin: statistics in Python. *J. Open Source Softw.* **3**, 1026 (2018).

60.   Waskom, M. L. seaborn: statistical data visualization. *J. Open Source Softw.* **6**, 3021 (2021).

61.   Tretyakov, K. konstantint/matplotlib-venn. (2024).

62.   Grigorev, K. LankyCyril/pyvenn. (2024).

63.   BioSenior/ggVolcano: The goal of `ggVolcano` is to help users make a beautiful volcano map
more easily, including general volcano plot(`ggvolcano`), gradient color volcano plot(`gradual_volcano`)
and GO term volcano plot(`term_volcano`). https://github.com/BioSenior/ggVolcano.

64.   Fang, Z., Liu, X. & Peltz, G. GSEApy: a comprehensive package for performing gene set
enrichment analysis in Python. *Bioinforma. Oxf. Engl.* **39**, btac757 (2023).

65.   Pang, Z., Chong, J., Li, S. & Xia, J. MetaboAnalystR 3.0: Toward an Optimized Workflow for Global
Metabolomics. *Metabolites* **10**, 186 (2020).

66.   Ripley, B. *et al.* MASS: Support Functions and Datasets for Venables and Ripley's MASS. (2024).

67.   Vanegas, L. H., Rondón, L. M. & Paula, G. A. glmtoolbox: Set of Tools to Data Analysis using
Generalized Linear Models. (2024).

68.   hirsch-lab/roc-utils. Digital Health Group (2024).

Acknowledgements

This work was financially supported by Nanomics Biotechnology., Ltd.

Author contributions

H.W. designed the study and drafted the manuscript. Y.W. led product development. Y.H.Z. performed the

proteomics data analysis. X.H.O. conducted the sample preparation and mass spectrometry experiments.

Competing interests statement

Y.W., Y.H.Z., X.H.O. and H.W. are employees of Nanomics Biotechnology.