An automated microfluidic platform for toxicity testing based on *Caenorhabditis*

elegans

Jiaying Wu[a,b,c,†], Jing Xi[a,b,†,*], Zhouhai Zhu[a,†], Wei Zhao[d], Qiyuan Peng[a], Chunbo Liu[a],

Ziqian Wan[e], Yiyi Cao[b], Xiang Chen[f*], Yang Luan[a,b,*]

[a] Joint Institute of Tobacco and Health, 41 Keyi Road, Kunming, Yunnan, China

[b] School of Public Health, Shanghai Jiao tong University School of Medicine, 227

South Chongqing Road, Shanghai, China

[c] Shanghai Municipal Center for Disease Control and Prevention, 1380 West

Zhongshan Road, Shanghai, China

[d] Technology Center, China Tobacco Yunnan Industrial Co., Ltd. 181 Hongjin

Road,Kunming 650231

[e] Biological science department, Xi'an jiaotong-Liverpool University, 111 Renai Road,

Suzhou, Jiangsu, China

[f] National Key Laboratory of Science and Technology on Micro/Nano Fabrication,

Department of Micro/Nano Electronics, Shanghai Jiao Tong University, 800

Dongchuan Road, Shanghai, China

[†] Jiaying Wu, Jing Xi and Zhouhai Zhu contributed equally to this work.

*Correspondence to:

Jing Xi, School of Public Health, Shanghai Jiao tong University School of Medicine,

227 South Chongqing Road, 200025 Shanghai, China. E-mail: xijing@shsmu.edu.cn

Xiang Chen, National Key Laboratory of Science and Technology on Micro/Nano

Fabrication, Department of Micro/Nano Electronics, Shanghai Jiao Tong University,

800 Dongchuan Road, Shanghai, China. xiangchen@sjtu.edu.cn

Yang Luan, School of Public Health, Shanghai Jiao tong University School of Medicine,

227 South Chongqing Road, 200025 Shanghai, China. E-mail: yluan@sjtu.edu.cn

Abbreviations[1]

Abstract

Humans are frequently exposed to a multitude of chemicals daily, necessitating

efficient methods for rapidly assessing toxicity and potential health risks.

Microfluidics has shown promise as an intelligent tool for rapid compound testing,

owing to its flexibility in integrating with automated devices. The article introduces

an automated microfluidic platform, based on *Caenorhabditis elegans (C. elegans)*,

designed for chemical toxicity testing. This platform consists of three modules –

worm culture, monitoring, and image analysis – which enable automated worm

culturing, drug delivery, periodic monitoring, and automated phenotypic analysis.

Researchers have designed a bridged microfluidic chip that permits worms to move

freely during experiments and established an economical monitoring module for

long-term tracking and periodic imaging. Furthermore, they have developed an

automated image analysis algorithm to automatically determine worm bending

frequency. The platform was subsequently utilized for long-term toxicological

assessments of the organophosphate pesticide and environmental pollutants. Results

indicated that the platform can effectively evaluate the general and developmental

---

[1]   **Abbreviations:** C. elegans, Caenorhabditis elegans; MCP, monocrotophos; BDE-47, 2,2',4,4'-Tetrabromodiphenyl ether; MEMS, Micro-Electro-Mechanical Systems; PDMS, Polydimethylsiloxane; NGM, nematode growth medium; 5-FUdR, 5-Fluoro-20-deoxyuridine; DMSO, dimethyl sulfoxide;; SE, standard error; PBDEs, polybrominated diphenyl ethers;

impacts of chemicals. The automated microfluidic worm analysis platform holds

significant potential for applications in drug safety assessmentand drug screening

research, contributing to human health and industry advancement.

Keywords: automated microfluidic platform, Caenorhabditis elegans, chemical

toxicity testing

Introduction

With the development of technology, large numbers of newly registered chemicals

discharge into the environment every day (https://www.cas.org/). Humans are

consistently exposed to various chemicals, which demands efficient methods to

rapidly evaluate the toxicity and potential health risks to humans. Microfluidics has

shown promise as an intelligent tool for rapid testing of compounds due to its

flexibility to integrate with automation devices. Current researches based on *in vitro*

organ chips or microphysiological systems cannot simulate *in vivo* dynamic process in

human bodies[1, 2]. *Caenorhabditis elegans (C. elegans)*, a classic model organism, is a

powerful alternative *in vivo* assay system that has proven to be a useful tool in

toxicology studies[3, 4]. Recently, platforms combining microfluidic chip with *C.*

elegans are promising for toxicity testing of numerous chemicals[5, 6].

Numerous researchers have developed microfluidic platforms for short-term

observation of worms in acute toxicity testing assays[7, 8]. These platforms exhibit

significant benefits, including minimal reagent usage and the capacity for seamless

integration with automated systems. The microchambers can be configured to

prevent the overlap of worms along the Z-axis, thereby simplifying the process of

automated imaging and monitoring throughout the experiments. In chronic toxicity

studies, periodic observation and phenotype analysis are needed to provide

information on the cumulative effect of toxins, thus demanding long-term cultivation

with periodic drug delivery and medium replenishment. Sahand *et al*. developed a

platform for age synchronization and analyzing subtle aging phenotypes of worms,

on which worms were fed manually outside and transferred daily into the chip for

progeny filtration[9]. Researchers have reported the development of platforms

capable of sustaining the culture of worms and conducting time-lapse imaging over

extended periods within microfluidic chips[10-12]. More recently, Le *et al*. introduced a

microfluidics-based automated system known as HeALTH, designed to facilitate the

lifelong cultivation and behavioral monitoring of worms[13]. Traditionally, the

observation and analysis of phenotypic characteristics have involved immobilizing

worms within microfluidic chips[9, 10, 14]. However, there has been a growing trend

towards the creation of chips with specialized designs that permit the free

movement of worms, thereby enabling the observation of behavioral phenotypes in

their entirety. To this end, researchers have also devised algorithms capable of

automatically identifying, tracking, and analyzing the behavioral traits of worms[15-18].

Despite these advancements, the approaches necessitate substantial computing

resources and are computationally intensive for standard computers. In recent times,

artificial intelligence (AI) has been integrated into the image processing of worms[19],

yet there are limited studies on the use of deep learning for the analysis of worm

phenotypes, particularly in images with complex backgrounds such as those from

microfluidic chips. In summary, while the aforementioned platforms have made

strides in the automation of toxicity testing assays using worms, there remains

considerable room for further development.

In this study, we presented a microfluidic platform designed for the automated

assessment of chemical toxicity using *C. elegans*, which consists of worm cultivation,

drug delivery, regular monitoring and automated phenotype analysis. Initially, we

designed a bridge-like microfluidic chip for worm culturing, which allows worms to

move freely form larval stage to adult. Additionally, we constructed an economical

monitoring module capable of longitudinal tracking and periodic image capture of

the worms. . The bend frequency of worms can be automatically determined using

an image analysis algorithm previously developed in our study[20]. Subsequently, we

carried out a whole-life-covered worm culture (17 days) and toxicity evaluation of

monocrotophos (MCP) and 2,2',4,4'-Tetrabromodiphenyl ether (BDE-47) using this

microfluidic platform, to validate the capacity of the platform for long-term toxicity

testing assays on worms.

Materials and methods

Microfluidic device fabrication

The microfluidic chip mold was developed by MEMS (Micro-Electro-Mechanical

Systems) process, including metal sputtering and electroplating[21]. Liquid PDMS

mixture (Sylgard 184 Silicone Elastomer Base : Curing Agent = 10:1) was cast on a

silicon mold and cured overnight at 70 °C to fabricate a PDMS chip. The inlet and

outlet were punched with needles. Finally, the PDMS chip was bonded to a clean

glass by air plasma treatment and stored overnight at 135 °C. For more details, see

supplementary materials.

Worm strains and maintenance

The *C. elegans* strains used in this study were wild-type Bristol N2，which was

obtained from the Caenorhabditis Genetics Center (St. Paul, Minnesota, USA).

Nematodes were cultured in the nematode growth medium (NGM, 17 g/L Agar B, 3

g/L NaCl, 2.5 g/L peptone, 25 mM KPO_4 buffer pH 6.0, 1 mM $MgSO_4$, 1 mM $CaCl_2$, 5

mg/L cholesterol in ethanol) or in M9 buffer (17 mM K_2HPO_4, 42 mM Na_2HPO_4, 85

mM NaCl, 1 mM $MgSO_4$) with the *E. coli* strain OP50 as nutrient at 20 °C. To obtain a

synchronized nematode population, a bleach-sodium hydroxide mixture (0.5 M

NaOH and 1% (v/v) NaClO) was used with an established method[22].

Chemicals

5-Fluoro-20-deoxyuridine (5-FUdR) and dimethyl sulfoxide (DMSO) were

purchased from Sigma-Aldrich (St. Louis, Missouri, USA). Monocrotophos (MCP) was

purchased from LGC Science (ShangHai Ltd) and dissolved in ultrapure water.

2,2',4,4'-Tetrabromodiphenyl ether (BDE-47) was obtained from Matrix Scientific

(Columbia, South Carolina, USA) and dissolved in DMSO.The stock solutions were

subsequently diluted with M9 buffer to obtain the working solutions. Ultrapure

water was produced using a Milli-Q system (Millipore, Bedford, MA, USA).

Lifespan assays

The lifespan assay was performed on age-synchronized adult worms (48 h after

hatching) in 96-well plates and PDMS microfluidic chips. Worms were cultured in

culturing medium (*E. coli* OP50:M9 buffer 1:200, v/v) and 5-FUdR (final concentration

of 20 mg/mL) was added to inhibit reproduction. In the PDMS microfluidic chips,

culturing medium was refreshed every day to provide the worms with sufficient

nutrients. The worms were stimulated with an acupuncture needle (96-well plates)

or by increasing the flow rate (microfluidic chips) every 24 h. If a worm did not

respond to any stimulation within 1 min, it was considered dead.

Developmental toxicity assays

For comparison of the developmental toxicity assays in PDMS microfluidic chips

and multi-well plates, N2 worms were treated with MCP (0, 250, 500 μM and 1 mM)

and BDE-47 (0, 3, 10 and 30 μg/ml) from the L1 stage (no feeding after hatching).

After 24 h, videos of worms in each chamber were recorded and analyzed with

automated image analysis module. The same assay was repeated with multi-well

plates, and the body length of worms were calculated manually.

Neurotoxicity assays

Synchronized N2 worms in L4 stage (feeding for 48 h after hatching) were treated

with chemicals for 24 h in this assay, and 20 mg/mL 5-FUdR was added to inhibit

reproduction. The concentrations of MCP were 0, 250, 500 μM and 1 mM, BDE-47

were 0, 3, 10 and 30 μg/ml. After treatment, worms were rinsed with M9 and

recorded by monitoring module. The bend frequencies of worms were calculated

with the image analysis module.

Statistical analysis

GraphPad Prism version 8.0 (GraphPad Software, La Jolla, California, USA) and IBM

SPSS Statistics version 25 (SPSS, Chicago, IL, USA) were used for statistical analysis. All

data are expressed as mean ± standard error (SE). The log-rank test was used to

compare the survival curves of worms cultured on PDMS microfluidic chips and

micro-well plates. One-way analysis of variance (one-way ANOVA) followed by the

Dunnett t-test were used to determine the differences between the treatment

groups and control group. A value of *P < 0.05* was considered statistically significant.

Results

Overview of microfluidic platform for worm toxicity testing

Since current microfluidic platforms for worm toxicity testing usually operate at a

low level of automation, we developed an integrated microfluidic platform, which

consists of three modules (Fig. 1A). The worm cultivation module includes a

microfluidic chip for worm culturing and observation, equipped with pressure-based

delivery system. Based on this module, diets and test substances can be

automatically added at regular intervals, and worms can be cultured in microfluidic

chip until death. During chronic toxicity testing assays, the phenotypes of worms in

bright filed can be monitored and collected regularly with the monitoring module,

which includes a motion stage and a Pi camera. After capturing images or videos of

worms, the body length and bend frequency can be calculated automatically with the

image analysis module. Fig.1B shows the physical photograph of the entire platform,

which is controlled through the control center.

Fig.1. Schematic of the microfluidic platform. (A) Diagram of the designed platform consisting of the "worm

cultivation module" "monitoring module" and "image analysis module". **(B)** Physical photograph of the whole

platform. Inside the yellow dotted box (graph on the right) is magnified photograph of "worm culturing module".

The bridge-like microfluidic chip for worm cultivation

We designed a bridge-like microfluidic chip that allows worms to move freely

during the experiment (Fig. 2A). Five parallel chambers were mounted on one chip,

and each chamber had a bridge-like structure to allow medium replacement and

prevent the worms from escaping. To meet the requirements of worm experiments

from the larval stages to adult, the gap should be designed to arrest the worms in L1

stage. Since L1 worms are approximately 10-15 μm wide and sufficiently flexible to

squeeze through narrow gaps, an ideal gap should not only arrest the worms but also

allow the medium to flow smoothly under relatively low pressure. We tested the

working conditions with gap sizes of 2, 5 and 10 μm and confirmed that a 5 μm gap

can meet the above requirements (see supplementary materials). To prevent the

"bridge deck" from collapsing, pillars were set as support structures at intervals of

100 μm. The designed height of the chambers is 40 μm (approximately equal to the

diameter of adult worms) to avoid overlapping more than two worms along the

Z-axis, which is beneficial for observation and automated analysis. Five inlets and

outlets are merged on each chip (Fig. 2C).

The operation procedure of the microfluidic chip is demonstrated as Fig. 2B. Firstly,

synchronized worm suspension in the reservoir was loaded into microfluidic chip.

During the experiment, culturing medium can be replenished and consistently loaded

into the chip at a low flow rate, to provide sufficient nutrients for worms.

To validate the biocompatibility of this microfluidic platform, we cultured worms in

multi-well plates and PDMS microfluidic chips until death. Briefly, 24 and 17 worms

were included in the multi-well plate group and PDMS microfluidic chips. As depicted

in Fig. 2D, the growth status of worms cultured on microfluidic platform are

consistent with those in traditional multi-well plates. The average lifespan (± SD) of

the worms in multi-well plates and microfluidic chips were 12 (± 4) days and 11 (± 3)

days, respectively, which revealed no significant difference between the survival

curves of two groups ($P > 0.05$) (Fig. 2D). The above results indicated that the

platform is suitable for worm growth and automated long-term cultivation, which

can relive labor and time compared with traditional assays on multi-well plates.

Fig.2. Design and operation of the microfluidic chip. (A) Five structures are paralleled in one chip. The inlet and

outlet are 150 μm wide and the chamber is 40 μm in height. The "bridge deck" has a thickness of 35 μm and a 5

μm gap underneath to prevent the worms from escaping while allowing the culturing medium to flow. Pillars are

set as support structures at intervals of 100 μm. Five inlets and outlets are merged and soaked in liquid to avoid

water evaporation. **(B) Procedure of worm loading and nutrient replenishment.** Step 1: valves 1 is closed (red

bar) and valve 2 is open (green bar), then worm suspension is loaded into the chip. Step 2: when reagents are

exhausted, valves 1 is open and valve 2 is closed. Step 3: nutrient is loaded into the chip. **(C) The physical**

photograph of the microfluidic chip. (D) Comparison of worm lifespan on multi-well plate and PDMS chip. The

survival curves of the two groups have no significant difference ($p > 0.05$).

Low-cost monitoring module for longitudinal observation and image acquisition

The monitoring module is a low-cost optical imaging system that can ensure

stability for the long-term culturing of worms. We built a low-cost three-axis

displacement platform using aluminum profiles. Cost of the monitoring module was

less than US$ 200. The displacement platform was controlled by an Arduino circuit

connected to a Raspberry Pi CPU. The GRBL code library was pre-written in Arduino.

The stepper motor and its driver were connected. The Raspberry Pi could send the G

code to Arduino to control the motion of the stage. The motion accuracy of this

module can easily reach 10 μm. We used a 4× lens and a low-cost Raspberry Pi CCD

for imaging.

Automated image analysis module for phenotype analysis of worms

With the large amount of image data generated in toxicity testing assays,

automated data processing can significantly improve the efficiency of the experiment.

Therefore, we used the automated image analysis algorithm developed in our

previous study[20] to analyze the body length and bend frequency of worms. We used

deep learning to extract worm contours from videos that contain chamber structure

and impurities in the culture environment. The following process is described as

before.

Toxicity evaluation of MCP and BDE-47 on the platform

Based on this microfluidic platform, we chose MCP and BDE-47 to demonstrate

the feasibility of the platform for automated toxicity testing of exogenous chemicals.

MCP is an organophosphorus pesticide widely used in agricultural production, the

residues of which may be detected frequently in food. As a cholinesterase inhibitor

acting on an acetylcholine receptor, MCP is highly neurotoxic[23]. As a kind of

commercial flame retardant, BDE-47 is the most abundant and prevalent

polybrominated diphenyl ethers (PBDEs) congeners in the environment, which has

been known to introduce neurodevelopmental toxicity[24].

We treated N2 worms with MCP and BDE-47 for 72 h and evaluated the body

length and bend frequency of worms at 24 h. Fig. 3A showed that treatment of MCP

and BDE-47 induced significant general toxicity in worms on this platform. The body

length was used as the toxic endpoints to evaluate the development toxicity. As

shown in Fig. 3B&E, MCP and BDE-47 treatment caused dose-dependent decreases

($p < 0.001$), implying evident developmental toxicity to nematodes. Fig. 3C&F showed

that the bend frequency of worms decreased significantly ($p < 0.001$) with the

increase in dose after 24 h treatment.

The above results demonstrated that this platform can be applied to automated

toxicity testing of numerous chemicals that may cause neurotoxicity and

development toxicity. In the future, more phenotypes such as brood size and cell

apoptosis, can be measured to expand the scope of toxicity evaluation.

Fig. 3. Toxicity evaluation of MCP and BDE-47 on the platform. (A-C) The survival curve, body length and bend

frequency of worms treated with MCP. (D-F) The survival curve, body length and bend frequency of worms treated with BDE-47.

Discussion

With the development of science and technology, microfluidic technology is more and more widely used in the biomedical field. Microfluidic technology has the characteristics of high throughput, low consumption and precise control and is widely used in cell culture, drug screening, disease diagnosis and other fields. As a model organism, *C. elegans* has the characteristics of short life cycle, simple genome and relatively simple nervous system and are ideal models for studying neurotoxicology and neural effects of drugs. In this study, we introduced an automated microfluidic platform based on *C. elegans*, which can realize automation of worm toxicity testing assays including worm culturing, monitoring and phenotype analysis.

Toxicity assays require periodic rinsing and recording of the worms' phenotypes to evaluate the toxicity of compounds with extended exposure times. In experiments lasting more than 24 hours, the diet-containing culturing medium should also be replenished regularly. Therefore, an automated liquid delivery and switching system is needed. NemaLife has introduced a product that can realize programmable regimens of feeding and reagent addition for toxicity testing (https://www.nemalifeinc.com/infinity-system), but in the latest report, they still use hand-held syringes to manually perform fluid manipulation[25]. Some researchers that developed an open-source liquid handling device to realize reagent delivery from large reservoirs to microfluidic chips were faced with the problems of complicated

motion platforms and extra on-chip backpressure designs[26]. Other researchers

applied rotary valves[27] or robotic devices[28] to liquid manipulation, but these fluid

switch and delivery systems are limited to complex structures with high-cost, thus

limiting the automation of chemical screening. In our study, we designed an

automated medium replenishment device to relive the tedious operation of manual

liquid change, the cost of which is low enough for laboratories.

Combined with microfluidic technology and nematode model, the neural effects of

chemicals or drugs can be studied efficiently, quickly and accurately. Through the

platform, the exposure environment of different concentrations of chemicals or

drugs to nematodes can be simulated, and the behavioral changes of nematodes,

such as movement trajectory, movement speed, reaction time can be observed and

recorded. Through the analysis of these data, the neural effects of chemicals or drugs

can be evaluated and the scientific basis for the safety assessment can be provided.

In general, the automated microfluidics nematode analysis platform is expected to

play a crucial role in the safety assessment of tobacco products, drug screening and

neurotoxicology research, contributing greatly to people's health and the sustainable

development of the industry.

Supplementary Information

Microfluidic device fabrication

Determination of the gap height of the microfluidic chips

Declarations

Availability of data and material

The data used and analyzed in the present study are available from the corresponding author on reasonable request.

Competing interests

The authors declare that they have no competing interests.

Funding

This work was sponsored by the Open fund of the Joint Institute of Tobacco and Health [grant number 2021539200340048].

Reference

[1] NIESKENS T T G, SJöGREN A K. Emerging In Vitro Systems to Screen and Predict Drug-Induced Kidney Toxicity [J]. Semin Nephrol, 2019, 39(2): 215-26.

[2] SHINOZAWA T, KIMURA M, CAI Y, et al. High-Fidelity Drug-Induced Liver Injury Screen Using Human Pluripotent Stem Cell–Derived Organoids [J]. Gastroenterology, 2021, 160(3): 831-46.e10.

[3] HUNT P R. The C. elegans model in toxicity testing [J]. J Appl Toxicol, 2017, 37(1): 50-9.

[4] QUEIRóS L, PEREIRA J L, GONçALVES F J M, et al. Caenorhabditis elegans as a tool for environmental risk assessment: emerging and promising applications for a "nobelized worm" [J]. Critical Reviews in Toxicology, 2019, 49(5): 411-29.

[5] BEN-YAKAR A. High-Content and High-Throughput In Vivo Drug Screening Platforms Using Microfluidics [J]. Assay Drug Dev Technol, 2019, 17(1): 8-13.

[6] KAMILI F, LU H. Recent Advances and Trends in Microfluidic Platforms for C. elegans Biological Assays [J]. Annu Rev Anal Chem (Palo Alto Calif), 2018, 11(1): 245-64.

[7] DING X, NJUS Z, KONG T, et al. Effective drug combination for Caenorhabditis elegans

nematodes discovered by output-driven feedback system control technique [J]. Sci Adv, 2017,

3(10): eaao1254.

[8] LOCKERY S R, HULME S E, ROBERTS W M, et al. A microfluidic device for whole-animal

drug screening using electrophysiological measures in the nematode C. elegans [J]. Lab Chip,

2012, 12(12): 2211-20.

[9] SABERI-BOSARI S, HUAYTA J, SAN-MIGUEL A. A microfluidic platform for lifelong

high-resolution and high throughput imaging of subtle aging phenotypes in C. elegans [J]. Lab

Chip, 2018, 18(20): 3090-100.

[10] KEIL W, KUTSCHER L M, SHAHAM S, et al. Long-Term High-Resolution Imaging of

Developing C. elegans Larvae with Microfluidics [J]. Dev Cell, 2017, 40(2): 202-14.

[11] ATAKAN H B, CORNAGLIA M, MOUCHIROUD L, et al. Automated high-content

phenotyping from the first larval stage till the onset of adulthood of the nematode

Caenorhabditis elegans [J]. Lab Chip, 2018, 19(1): 120-35.

[12] ATAKAN H B, XIANG R, CORNAGLIA M, et al. Automated Platform for Long-Term

Culture and High-Content Phenotyping of Single C. elegans Worms [J]. Sci Rep, 2019, 9(1):

14340.

[13] LE K N, ZHAN M, CHO Y, et al. An automated platform to monitor long-term behavior

and healthspan in Caenorhabditis elegans under precise environmental control [J]. Commun

Biol, 2020, 3(1): 297.

[14] BERGER S, LATTMANN E, AEGERTER-WILMSEN T, et al. Long-term C. elegans

immobilization enables high resolution developmental studies in vivo [J]. Lab Chip, 2018,

18(9): 1359-68.

[15] YEMINI E, JUCIKAS T, GRUNDY L J, et al. A database of Caenorhabditis elegans

behavioral phenotypes [J]. Nat Methods, 2013, 10(9): 877-9.

[16] LIN J L, CHEN Y S, HUANG Y H, et al. Approach to the Caenorhabditis Elegans

Segmentation from Its Microscopic Image; proceedings of the 2018 IEEE International

Conference on Systems, Man, and Cybernetics (SMC), F 7-10 Oct. 2018, 2018 [C].

[17] WERKHOVEN Z, ROHRSEN C, QIN C, et al. MARGO (Massively Automated Real-time

GUI for Object-tracking), a platform for high-throughput ethology [J]. PLoS One, 2019, 14(11):

e0224243.

[18] JAVER A, CURRIE M, LEE C W, et al. An open-source platform for analyzing and sharing

worm-behavior data [J]. Nat Methods, 2018, 15(9): 645-6.

[19] ITSKOVITS E, LEVINE A, COHEN E, et al. A multi-animal tracker for studying complex

behaviors [J]. BMC Biol, 2017, 15(1): 29.

[20] WU J, GAO Y, XI J, et al. A high-throughput microplate toxicity screening platform

based on Caenorhabditis elegans [J]. Ecotoxicol Environ Saf, 2022, 245: 114089.

[21] JIN X, CHEN R, ZHAO S, et al. An efficient method for CTCs screening with excellent

operability by integrating Parsortix™-like cell separation chip and selective size amplification

[J]. Biomed Microdevices, 2018, 20(2): 51.

[22] STIERNAGLE T. Maintenance of C. elegans [J]. WormBook, 2006: 1-11.

[23] SALIM C, RAJINI P S. Glucose feeding during development aggravates the toxicity of

the organophosphorus insecticide Monocrotophos in the nematode, Caenorhabditis elegans

[J]. Physiol Behav, 2014, 131: 142-8.

[24] ZHENG S, LIU C, HUANG Y, et al. Effects of 2,2',4,4'-tetrabromodiphenyl ether on

neurobehavior and memory change and bcl-2, c-fos, grin1b and lingo1b gene expression in

male zebrafish (Danio rerio) [J]. Toxicol Appl Pharmacol, 2017, 333: 10-6.

[25] RAHMAN M, EDWARDS H, BIRZE N, et al. NemaLife chip: a micropillar-based

microfluidic culture device optimized for aging studies in crawling C. elegans [J]. Sci Rep, 2020,

10(1): 16190.

[26] LAGOY R C, ALBRECHT D R. Automated fluid delivery from multiwell plates to

microfluidic devices for high-throughput experiments and microscopy [J]. Sci Rep, 2018, 8(1):

6217.

[27] MORIOKA K, SATO H, MORITA K, et al. Development of an on-chip sample injection

system with a 6-port valve incorporated in a microchip [J]. RSC Adv, 2020, 10(59): 35848-55.

[28] LONGWELL S A, FORDYCE P M. micrIO: an open-source autosampler and fraction

collector for automated microfluidic input-output [J]. Lab Chip, 2020, 20(1): 93-106.