***S*timulus-modulated *a*pproach to *s*teady-*s*tate: A new paradigm for event-related fMRI.**

Renil Mathew, Amr Eed, L. Martyn Klassen, Stefan Everling, Ravi S Menon

Functional MRI (fMRI) studies discard the initial volumes acquired during the approach
of the magnetization to its steady-state value. Here, we leverage the substantially higher
temporal signal-to-noise ratio (tSNR) of these initial volumes to increase the sensitivity of
event-related fMRI experiments. To do this, we introduce Acquisition Free Periods
(AFPs) that allow for the full recovery of the magnetization, followed by equal-length task
or baseline acquisition blocks (AB) of fMRI volumes. An appropriately placed stimulus
in the AFP produces a Blood Oxygenation-Level-Dependent (BOLD) response that peaks
during the initial high tSNR phase of the AB, yielding an ~50% reduction in the number
of trials (relative to conventional fMRI) needed to achieve a given statistical threshold.
The AFP allows flexibility for the presentation of auditory stimuli, uncontaminated
recording of electrophysiological activity, physiological gating, as well as the ability to
average aperiodic fMRI data using stimulus or response-locked signal acquisition.

BOLD fMRI has been a mainstay of non-invasive functional brain imaging for over 30

years[1–3]. fMRI experiments rely almost exclusively on versions of the echo-planar imaging

(EPI) sequence to map resting-state or stimulus-evoked BOLD signal changes over time[4,5].

Typically, a dozen or so initial volumes, acquired during the steady-state equilibrium of the

magnetization are either discarded or treated as dummy scans. Unfortunately, these are the

images with the highest SNR. The approach to steady-state of the magnetization is described

by the Bloch equations[6] and is influenced by the longitudinal relaxation time constant (T1) of

the tissue. The MRI signal intensity decays from a fully-relaxed maximum to reach steady-

state within 5-10 seconds depending on the T1 of the tissue and flip angle used. Since T1

increases with field strength, this approach has particular benefit at ultra-high magnetic fields.

While generally considered as an annoyance for paradigm design, stimulus triggering or

analysis pipelines, EPI volumes acquired during the approach to equilibrium can offer valuable

insights into tissue-specific relaxation parameters and have been proposed for improving

grey/white matter tissue discrimination, resulting in a reduction of false positives in event-

related activation maps[7].

In this study, we capitalize on the substantially higher magnetization available during

the initial decay period to achieve enhanced stimulus-evoked BOLD signal detection.

Recurring AFPs with no gradients or radio frequency pulses present allow for the full recovery

of the magnetization and are followed by ABs consisting of multiple rapidly acquired volumes

of a typical fMRI sequence such as EPI. The EPI volumes acquired have an initially higher

SNR that decays to the usual steady-state value. Implementing this into a task paradigm, we

structure ABs into identical length independent baseline and task ABs. To achieve optimum

signal detection sensitivity for the stimulus, we present the stimulus such that the peak of the

hemodynamic response that underlies the BOLD signal is captured in task AB volumes during

the higher SNR phase of the magnetization decay. Because the *s*timulus is modulating the

*a*pproach to *s*teady-*s*tate, we abbreviate this paradigm as SASS. Here we demonstrate the

advantages of using the SASS paradigm and compare the detection sensitivity with the

traditionally used steady-state (SS) EPI sequence, keeping either the experiment time or

number of trials constant. The experiments were conducted on awake common marmosets

(*Callithrix jacchus*) in a 9.4 T MRI, with visual and auditory stimuli.

RESULTS

Acquisition block and tSNR calculations. The basic building block for SASS paradigm

consists of two elements, an AFP followed by an AB. This building block is repeated multiple

times with (task AB) or without (baseline AB) a presented stimulus (Fig.1a) (Methods). Within

an AB, apparent transverse relaxation time (T2*)-weighted EPI volumes with a given repetition

time (TR) and flip angle (FA) result in a magnetization that decays to a steady-state value

governed by the T1 relaxation in each voxel. To characterize the noise in each voxel over the

course of the AB, we computed individual voxel-wise tSNR maps for each volume within an

AB (Fig.1b) because tSNR (rather than SNR) is an important determinant of the ability to detect

BOLD signal over the noise. For our specific acquisition and tissue parameters at 9.4 T, the

grey matter tSNR maps across the AB (block length = 16 s, TR = 1 s) demonstrated a higher

initial tSNR before reaching the steady-state tSNR value after 8-10 seconds (~4-5 times the T1

constant of grey matter[8]) (Fig.1c). As a simple tissue discrimination approach, the tSNR maps

were segmented using Independent Component Analysis (ICA) into grey and white matter

voxels from the raw timeseries of acquired functional volumes, where the inherent difference

in their T1 relaxation decays can be used for tissue discrimination (Extended Data Fig.3a, b).

Notably, we observe that the grey matter tSNR of the first volume was 61% higher than the last

volume in the AB (where the steady-state has been achieved) (Fig.1d). Averaged over the first

five volumes (5 s) where we expect the BOLD response to peak, the tSNR gain was ~25%, and

will be dependent on TR. The higher tSNR in the initial AB volumes gave us insight into how

a stimulus response captured during these initial volumes can be detected with higher signal

sensitivity compared to a steady-state condition.

The SASS paradigm was used to detect a stimulus-evoked BOLD response by separating

the ABs into alternating identical length task and baseline blocks. The stimulus was presented

such that the subsequent BOLD response was captured during a task AB (Fig.1a). The figure

shows two different stimulus timings used in the study, with a stimulus presented either during

the AFP immediately before the task AB (pre-AB stimulus) or within the first volume of the

task AB (in-AB stimulus). To distinguish the stimulus-evoked BOLD response from the

decaying magnetization signal while maintaining the fidelity of the BOLD response, we used

the baseline ABs from each preprocessed fMRI run to calculate the mean and standard

deviation of each voxel timepoint across all baseline ABs (Fig.1b). The baseline values were

used to z-score normalize all volumes in all ABs, thereby representing each voxel timepoint

within a task or baseline AB in terms of the standard deviation from their respective baseline

AB mean values (Equation 2 - Methods). This volume-by-volume z-score normalization

process is necessary because the imaging signal varies in time, unlike the conventional steady-

state case where it is constant. The resultant z-score transformed runs are free from the

influence of T1 decay, maintain the shape fidelity of the BOLD response and can be used for

further statistical analysis with conventional fMRI methods. Fig.2a shows the mean raw

timeseries of active voxels from a single fMRI run, reflecting the characteristic T1 relaxation

of each AB within the run and Fig.2b shows the baseline z-score normalized timeseries of the

same. Task ABs are coloured in orange.

Comparison of experimental data to theoretical modeling. We compared the acquired raw

data using the SASS paradigm with theoretical predictions using a series solution of the Bloch

equations for the magnetization behaviour in an AB (Equation 1 – Methods). The mean task

and baseline magnetization decay of active voxels for a single run containing 15 baseline ABs

and 15 task ABs was computed (Fig.2c). The task was a 1 s presentation of a visual stimulus

in the AFP immediately before the start of each task AB acquisition (pre-AB stimulus). A clear

difference between the baseline and stimulus-evoked magnetization decay curves is observed.

In a conventional steady-state fMRI experiment, the time-varying percent signal change curve

accurately mirrors the shape of the BOLD response because the baseline is constant. However,

in the case where the magnetization is decaying in time, the shape of the time-varying percent

signal change is distorted, and the SNR also varies in time. To correct for these factors, we z-

score normalized all AB voxels using the baseline ABs to estimate the temporal variance as

described above. This baseline z-score normalized signal correctly reflects the shape of the

BOLD response for a pre-AB stimulus, allowing for standard fMRI processing tools to be used
in further analysis (Fig. 2d).

We computed the tissue magnetization decay within a baseline AB using the solution
to the Bloch equations (Equation 1 - Methods) (Fig.2e). We modeled the decay with T1 = 2.2
s, FA = 50 degrees (to match the nominal FA of the experiment) and TR = 1 s. We then further
modeled the T2*-weighted BOLD response to a 1 s pre-AB visual stimulus by convolving a
standard double-gamma hemodynamic response function (HRF) peaking at 3 s with a 1 s
stimulus (Fig.2f). By multiplying the calculated baseline magnetization with the resulting
BOLD response[9] (truncated stimulus-HRF to match pre-AB stimulus onset) for every volume
in the AB, we generated an estimate of the stimulus-driven magnetization decay (Fig.2g). Our
theoretical modeling of baseline and task magnetization decays (Fig. 2g) closely match the
observed signal characteristics of the decay and stimulus-evoked HRFs observed in the
acquired data (Fig. 2c).

Comparing in-AB BOLD detection sensitivity between SS and SASS. We investigated the
difference in stimulus-evoked BOLD signal sensitivity between a traditional SS paradigm and
our SASS paradigm using a simple visual stimulation experiment with awake marmoset
monkeys (n=3, Methods) using identical numbers of volumes or stimulus trials. The voxel-
based GLM analysis, with a corrected (for multiple comparisons) P < 0.025, revealed activation
maps for each individual subject's runs for both methods, representing stimulus driven
activations. At the group level, we observed stronger brain wide activations using the SASS
paradigm, compared to the SS paradigm (Fig.3a). SASS showed greater numbers of active
voxels at higher z-value compared to those from the SS paradigm. For voxels confined to the
primary visual cortex (V1) and medial superior temporal area of cortex (MST), the SASS
paradigm showed a similar trend with a greater number of active voxels at higher z-values
compared to the SS paradigm (Fig.3b, c). Mean z-score transformed voxel timeseries of each
subject run from all active voxels at z-value> 5.1 were used to plot the average timeseries
across all subjects, exhibiting characteristic BOLD responses from both paradigms.

The mean timeseries show robust responses in the SASS paradigm (Fig.3d) and the SS
paradigm (Fig.3e). For both paradigms, the mean z-score timeseries of task regions from active
voxels exhibits a characteristic BOLD response across subject runs (Extended Data Fig.2a, b).
Based on these observations, we conclude that there exists an enhanced detection sensitivity to
BOLD signal changes when utilizing SASS compared to a traditional SS paradigm, when
comparing matched numbers of volumes (however, with SASS taking 2 minutes longer than
SS due to the AFPs). To match these two approaches by scan time, we also compared 480

volumes with 15 stimulus trials per run for SS paradigm to 352 volumes with 11 stimulus trials

per run for the SASS paradigm. Despite having 27% fewer trials, the SASS paradigm still

showed better activation maps with higher z-value voxels (Extended Data Fig.1a, b). We

extended the decimation of the number of stimulus trials (z-score normalization using

predetermined baseline values from 15 baseline ABs) until we achieved a SASS map that was

comparable to the full SS map. Only 8 stimulus trials in SASS paradigm were needed to match

the activation in V1 ROI from the two paradigms (Extended Data Fig. 1a, b).

Comparing pre-AB BOLD detection sensitivity between SS and SASS. In our auditory

experiment paradigm, high frequency pure tones were presented for 1 second during the MRI

noise-free AFPs immediately before the start of each task AB (pre-AB stimulus - Methods).

The voxel-based GLM analysis using the truncated stimulus-HRF (Fig.2f), with a corrected

P<0.025, revealed z-value maps for each individual run. At the group level (n=2), corrected

voxel for a P<0.025 shows robust localized activations in auditory cortex corresponding to the

stimulus[10] (Fig.4a). The mean z-score voxel timeseries from active voxels shows robust BOLD

responses to the auditory stimulus presented in the AFP (Fig.4b). Comparison against an

auditory stimulus presented in the SS paradigm was not possible due to the superposition of

the scanner noise, which leads to very poor maps. To further validate the reliability for detecting

a BOLD response to a stimulus presented during the AFP, we replicated our visual stimulus

paradigm placing the 1 s stimulus in the AFP (pre-AB stimulus) rather than coincident with the

onset of the AB. The resulting group activation map (n=3) (Fig.4c) and time course (Fig.4d) is

consistent with our in-AB visual results, noting however that the activation maps are in theory

not comparable considering the difference in pre-AB stimulus presented during a quiet period

and in-AB stimulus influenced by scanner noise. For both auditory and visual pre-AB stimulus

experiments, the mean z-score timeseries of task regions from active voxels exhibits a

characteristic BOLD response across subject runs with an earlier peak (Extended Data Fig.2c,

d). From these observations, we conclude that the BOLD response to a stimulus presented

during the AFP can be extracted accurately from ABs when an appropriate truncated stimulus-

HRF (Fig.2f) is used in the GLM analysis.

DISCUSSION

MRI noise-free regions in periodic sequences operating in the steady-state regime have

been previously utilized to facilitate auditory stimulus presentation[11–14]. Such approaches use

either sparse sampling or continue applying the radio frequency (RF) pulses to maintain the

steady-state of the magnetization during the quiet periods. Sparse sampling limits the temporal

resolution of the HRF, while the RF in the silent periods is in fact difficult to implement for the

ubiquitous slice-multiplexed fMRI pulse sequences[15,16] (such as multi-band which require

gradients) and interferes with multi-modal fMRI, such as fMRI-EEG. Our approach is distinct

in that we use the AFPs to both fully recover the magnetization and present stimuli, after which

we read out the BOLD response to the stimulus with an AB. This approach enhances the

detection of the BOLD response due to the enhanced tSNR and requires no periodicity for the

stimulus presentation. The AFP just needs to be long enough to allow full recovery of the

magnetization, namely at least 4 times the T1 of the grey matter. The stimulus should be placed

such that the peak of the BOLD response occurs during the period of enhanced tSNR. The

earlier the BOLD signal peak during the decay, the higher the peak detection sensitivity. This

places some constraints on the duration and location of the stimulus in the SASS paradigm

design. A short duration (few seconds) stimulus placed in the AFP immediately before the start

of the AB is appropriate. If characterization of the full HRF shape is important, an in-AB

stimulus coincident with the start of the AB may be more appropriate. These types of short

stimulus presentation paradigms are characteristic of event-related designs in fMRI studies.

Through judicious presentation of the stimulus in the AFP, the peak of the BOLD HRF

can be timed to be in the AB region where tSNR is up to 61% higher than the steady-state

tSNR. Integrating the volume-by-volume tSNR values multiplied by the volume-by-volume

HRF yields an approximate doubling of detection efficiency for our chosen parameters. In other

words, approximately half the number of trials is needed relative to a steady-state sequence to

achieve the same z-value map. At faster TR values, this enhancement increases further as many

samples of the HRF are acquired at high tSNR, aiding better GLM fit (data not shown).

In the demonstrations here we have used equal numbers of alternating task and baseline

ABs. This is not a requirement of the SASS paradigm. One simply needs enough baseline ABs

to accurately determine the temporal standard deviation of each volume in the AB to allow for

accurate z-score normalization, either acquiring them in between task ABs (as done here) or

having an independent run to characterize the baseline AB and then using only task ABs in

subsequent runs. However, to simplify (in a statistical sense) the z-scoring process, the number

of volumes in the task and baseline ABs should ideally be the same. Acquisition of baseline

ABs can be completely forgone if other approaches to estimating the tSNR in a voxel are used.

It is tempting to compare the almost doubling of the peak detection sensitivity of the

pre-AB visual stimulus with the in-AB visual stimulus. These are not quite the same experiment

from a psychophysical perspective, the major difference being that in the pre-AB condition the

visual stimulus is presented in a period of silence before the AB starts. The presentation of

temporally congruent auditory and visual stimuli in the in-AB condition, is known to enhance

activation in the sensory cortices and elsewhere due to multimodal integration, inhibition and
disinhibition[17–19]. These arguments should make for a more robust in-AB map, yet the pre-AB
map is visually comparable or superior in part because of the presentation of the visual stimulus
in isolation and in large part because the HRF peak of such a pre-AB stimulus plays out during
the highest tSNR section of the magnetization decay.

With the RF coil and acquisition parameters used in this study, we are in the thermal
noise regime, wherein tSNR is proportional to the available magnetization. This is the regime
in which many ultra-high magnetic field laminar level fMRI studies in humans and animal
models are conducted. At lower field strengths or spatial resolutions, physiological noise
dominates, in which case the tSNR could be somewhat independent of the magnetization and
therefore relatively uniform from volume to volume in the AB. However, the SASS paradigm
still has advantages in this application because the stimulus presentation can be triggered by a
physiological signal. The ability to synchronize the stimulus presentation and subsequent AB
to a parameter such as cardiac and/or respiration phase would dramatically increase the tSNR
in fMRI experiments[20]. Likewise, triggering acquisitions based on a behavioral event (e.g. eye
fixation, saccade, key press or EEG waveform) could significantly decrease inter-trial
variability. This approach would enable image acquisition to occur solely in response to
triggers from the subject, potentially giving the subject control over the stimulus presentation.

The imaging gradients and the radio frequency pulses are not present during the AFP.
This has many advantages if the stimulus is presented in the AFP. Auditory signals or
instructions can be presented to the subject with no competing MRI sound interference. Speech
or vocalization responses can be recorded with high fidelity during this time as well. In multi-
modal experiments, such as simultaneous electrophysiology and fMRI[21–23], electrophysiology
data acquired are heavily affected by MRI gradient artifacts and it is challenging to clean the
data and maintain waveform fidelity. By presenting a stimulus in the AFP, we enable the unique
ability to record undistorted single-unit, local field potential or evoked potential responses to
the stimulus without the interference of MR gradient or RF artifacts. The AB following the
AFP can still capture the BOLD response, because the response peaks with a hemodynamic
delay of several seconds. In our simple demonstrations, a pre-AB and in-AB stimulus both
produce HRFs that occur during the higher tSNR phase of the decay. This allows HRF
measurement in a variety of designs where a stimulus and a response occur close together.
Alternatively, where the stimulus and response are far apart (e.g. a working memory task), the
fMRI acquisition can be triggered by the stimulus or the response. Incorporating stimulus
presentation during the AFP offers other additional benefits, such as minimizing stimulus-

driven motion artifacts in awake animal experiments where immediate motion due to the
stimulus does not affect the ABs that follow. The same is true for motor tasks performed during
the AFP, such as pointing or operating a joystick, where head motion and magnetic field
distortions are not salient.

Multiple repetitions of the imaging volume with a short enough TR to well characterize
the BOLD response within an AB can be challenging, depending on the required spatial
resolution. Using modern multi-band fMRI approaches, it is certainly possible to obtain whole
brain fMRI volumes in ~1 s at cortex-respecting spatial resolutions on human MRI systems[16,24].
For acquisitions with short TR, the average tSNR gain in an AB will be substantially higher as
more volumes are acquired at higher SNR phase of the decay before reaching a steady-state. If
higher spatial resolution is required, a smaller stack of slices can be acquired with the cost of
additional stimulus presentations, also utilizing the additional tSNR gain due to shorter TR.
Alternately, the stimulus can be temporally jittered between presentations to effectively
subsample the BOLD response.

The use of the SASS paradigm has the additional benefit of enhanced tissue
discrimination for grey matter segmentation directly from the functional images, without the
reliance of a T1 weighted image. This is because tissue components such as grey matter and
white matter decay with different T1 relaxation time constants and can therefore be segmented
from their decay curve properties. EPI images generally have some geometric distortion
relative to conventional high resolution anatomical images and need sophisticated registration
to be placed in the typically T1-weighted anatomical image space. By using the tissue
segmented SASS images, one can improve the registration between the functional and
structural images. The segmentation could be further improved using timeseries clustering or
curve fitting methods, possibly allowing for differentiating not just grey or white matter voxels
but also larger blood vessels[7].

Finally, the method offers promise for layer-fMRI investigations, where the tSNR gain
in acquired volumes becomes particularly valuable in delineating layer differences in BOLD
activation. These potential applications underscore the versatility of SASS paradigm alongside
the improved signal sensitivity.

METHODS
Animal preparations. Experimental procedures followed the guidelines outlined by the
Canadian Council of Animal Care policy and were conducted under a protocol approved by the
Animal Care Committee of the University of Western Ontario Animal Care Committee #2021-
111. We ensured all ethical regulations governing animal use are followed. Ultra-high field

fMRI data used for all experiments were obtained from 6 awake common marmoset monkeys

(*Callithrix jacchus*): two females and four males. (6.7±3.7 years and 403±33g) (Visual

experiment n = 3, auditory experiment n = 2, tSNR calculation n = 1).

To prevent head motion during MRI acquisition, animals were surgically implanted

with an MR-compatible machined PEEK (polyetheretherketone) head post[25], conducted under

anesthesia and aseptic conditions. During the surgical procedure, the animals were first sedated

and intubated to ensure they remained under gas anesthesia, maintained by a mixture of O2

and isoflurane (0.5–3%). With their heads immobilized in a stereotactic apparatus, the head

post was positioned on the skull following a midline skin incision along the skull. The head

post was secured in place using a resin composite (Core-Flo DC Lite; Bisco). Heart rate,

oxygen saturation, and body temperature were continuously monitored throughout the surgery.

Two weeks post-surgery, the monkeys were acclimated to the head-fixation system and the

MRI environment through a three-week training period in a mock scanner[26].

MRI and fMRI procedures. MRI experiments were performed on a 9.4-T/31-cm small

animal magnet (Agilent, Palo Alto, CA, USA) interfaced to a Bruker Avance NEO console

(Bruker BioSpin Corp, Billerica, MA) and equipped with an in-house designed and built 15-

cm gradient coil of 450mT/m strength[27], with Paravision-360 v3.3 software at the Centre for

Functional and Metabolic Mapping located within the Robarts Research Institute at the

University of Western Ontario. An 8-channel phased-array receive coil was used in

combination with a birdcage transmit coil for data acquisition. We employed a gradient-echo-

based single-shot echo-planar imaging (EPI) sequence with following parameters: repetition

time (TR) = 1s, echo time (TE)= 15 ms, flip angle = 40°, field of view (FOV) = 48x64 mm^2,

matrix size = 96x128, isotropic resolution of 0.5 mm^3, 25 coronal slices, bandwidth = 400 kHz,

and a GRAPPA acceleration factor of 2 (left-right). Additionally, an extra set of EPIs with

opposite phase-encoding direction (right-left) were acquired for EPI-distortion correction

purposes. For each session, the T2-weighted anatomical images were acquired for each session

using the TurboRARE2D pulse sequence (6 averages, 42 slices, slice thickness = 500 μm, FOV

48×64 mm^2, matrix size 192×192, TE = 39 ms, TR = 7 s, echo spacing = 13 ms, and RARE

factor 8).

For the SASS paradigm built from the above-mentioned EPI sequence, we incorporated

recurring 8-second-long AFPs after every 16 (visual stimulus paradigm) or 12 (auditory

stimulus paradigm) second long ABs. The inclusion of AFPs increased the scan time compared

to a SS paradigm as the number of volumes acquired was kept the same in both sequences (480

volumes in each run). For scan time comparison with SS paradigm, we reduced the number of

volumes per run in the SASS paradigm to best match the duration of the SS paradigm. The number of slices in both visual and auditory paradigms were set to 25 slices, covering respective regions of interest according to the stimulus paradigm. For the visual stimulus paradigm, data from each animal was collected during a single session in the given order: SASS paradigm (in-AB stimulus), SS paradigm, SASS paradigm (pre-AB stimulus).

Acquisition blocks and acquisition free period. We introduced Acquisition-Free Periods (AFPs) to induce magnetization relaxation within an fMRI run. AFPs represent silent periods in the sequence where no volumes are acquired, and no RF is applied. The duration of the AFP is made long enough for full magnetization recovery, typically 4-5 times the T1 relaxation time constant of grey matter (2.0-2.2 seconds in marmosets)[28]. The volume acquired immediately after an AFP captures the magnetization at its maximum for a given flip angle (FA), with the subsequent volumes capturing the magnetization decay before reaching a steady-state intensity. The acquisition period following the AFP is called an acquisition block (AB), within which a series of volumes are acquired with a specified repetition time (TR) and FA. ABs, each of identical length, form independent periods of the time series which serve as the basic repeating unit for designing task paradigms. The AFP and AB durations can be modified according to the task design making sure there is enough time for full magnetization recovery and characterization of the BOLD hemodynamic response (only if necessary).

For our fMRI experiments, ABs are categorized into alternating task and baseline blocks. The length of ABs is tailored to capture the hemodynamic response associated with a stimulus presented either before (pre-AB stimulus) or within a task block (in-AB stimulus). We use baseline blocks to characterize the magnetization decay curve, permitting the z-score normalization of task blocks for subsequent analysis. For our respective stimulus paradigms, we employed two different block lengths (16 seconds for visual experiment and 12 seconds for auditory experiments) with the same minimum AFP duration of 8 seconds (~4*T1).

Theoretical modeling. The relaxation process of the magnetization vector in the presence of an external magnetic field is described by the Bloch equations[6]. In the context of T2*-weighted EPI acquisitions, our solutions derived from the Bloch equations for the magnetization $M(n)$ after the n^{th} excitation in a series of excitations is governed by imaging and tissue parameters given in the equation (Equation 1). Using the equation, we simulated data for the magnetization relaxation values at each TR (1 second) for an AB length of 16 seconds (Fig.2c). The maximum magnetization M_0 (after a single 90-degree pulse from the fully relaxed condition) value was set to 100. We used a θ (flip angle) of 50 degrees (Ernst angle) in the simulation to match the

nominal FA of the experiment over the slice. The grey matter T_1 time was set to 2.2 seconds

matching the grey matter T_1 relaxation time constant in marmosets ranging between 2.0 and

2.2 seconds at 9.4T MRI.

To simulate the influence of a stimulus on the decay, we generated an HRF using the

double gamma function defined in the FSL toolbox (time to peak - 3 s and time to peak of

undershoot - 8 s). The maximum BOLD percent signal change due to the stimulus was fixed

to 4% (typical signal change observed in 9.4T fMRI experiments). The baseline magnetization

curve was multiplied with this HRF function convolved with a stimulus duration of 1 second.

The resultant magnetization values represent a stimulus-modulated approach to steady-state.

Because of this multiplication, the initial magnetization values during a task are slightly higher

than those during the baseline as the BOLD signal is a few percent higher than baseline

(Fig.2g). When the stimulus occurs during the AFP, the generated HRF should be truncated

before multiplying with the baseline magnetization, as the onset of stimulus doesn't coincide

with the start of AB (Fig.2f).

353
$$M(n) = \left(M_0(E_1 cos\theta)^n + M_0(1 - E_1) \sum_{k=0}^{n-1} (E_1 cos\theta)^k \right) sin\theta$$

354
$$E_1 = e^{-TR/T_1}$$

355
$$M_0 = \text{Thermal equilibrium magnetization}$$

356
$$\theta = \text{Flip angle}$$

357
$$n = \text{excitation number (k=0 is the first excitation)} \quad (\text{Eq.}1)$$

Temporal Signal-to-Noise (tSNR) calculation. tSNR values in AB volumes were calculated

using preprocessed data from a single subject across four runs (30 ABs per run). For each

timepoint in an AB, the mean and standard deviation were computed for each voxel across all

blocks. Independent Component Analysis (ICA) using FSL's melodic ICA was applied to the

time series, revealing inherent differences in magnetization decay between grey and white

matter voxels (Extended Data Fig.3a). This resulted in independent components corresponding

to grey and white matter voxels (Extended Data Fig.3b). The white matter components were

removed, and the mean grey matter temporal signal-to-noise ratio (tSNR) was calculated using

the mean and standard deviation obtained for each timepoint in an AB (Fig.1c).

Baseline z-score normalization of acquisition blocks. To extract the BOLD signal changes

evoked by the stimulus, we needed to eliminate the influence of the T1 governed magnetization

decay from the ABs. Prior methods have utilized a global decay regressor for regressing the

decay[29], which is inappropriate due to the inherent variability in T1 decay curves across

different voxels, influenced by their tissue characteristics and RF coil inhomogeneity that
causes variable FAs. For SASS paradigm runs, we utilized alternating baseline blocks within
each run to standardize all blocks uniformly. Within each run, standard preprocessed fMRI
volumes are initially segmented into task and baseline blocks. For each voxel timepoint within
the AB, the mean and standard deviation are calculated from all baseline blocks. These
statistical measures are then utilized to calculate standardized z-scores for both individual task
and baseline ABs (Equation 2). This transformation enables the expression of voxel signal
fluctuations between each time points in terms of their respective standard deviation units from
the baseline mean values, thereby eliminating the decay influence from each block timepoint.
For each voxel the resultant z-scores are mean shifted to their respective temporal mean from
all baseline blocks. Adding the mean as a constant was only necessary to be able to utilize the
conventional FSL tools (for registration of subject's functional to anatomical to template
space). All z-score transformed blocks are concatenated accordingly and are fit for further
analysis using the FSL toolbox. The same method was used to z-score normalize SS paradigm
data by utilizing the well-defined baseline and task regions in the raw timeseries. For SS runs,
the first baseline region (16 volumes) during approach to steady-state was replaced with the
subsequent baseline region before normalization.

388
$$Z(n, AB) \;=\; \frac{M(n, AB) - \bar{M}(n)}{\sigma(n)}$$

389
$$\bar{M}(n) = \frac{1}{N} \sum_{b=1}^{N} M(n, b)$$

390
$$\text{Standard Deviation, } \sigma(n) = \sqrt{\frac{1}{N} \sum_{b=1}^{N} (M(n, b) - \bar{M}(n))^2}$$

391   n = Excitation number within acquisition block (*Same as the n in Eqn. 1*)

392   AB = Acquisition block number to be z-score transformed

393   N = Number of baseline blocks in a run

394   b = Individual baseline acquisition block index

395            $\bar{M}(n)$ = Mean across all baseline acquisition blocks.

396            $M(n)$ = Magnetization in the n^{th} volume in the acquisition blocks

397            $\sigma(n)$ = Standard deviation across all baseline acquisition blocks (Eqn. 2)

398   **Stimulus design.** We utilized two different stimulus designs for the study. For all visual
399   stimulus experiments, a 10 Hz flickering LED light was presented for a duration of 1 second.
400   Fifteen task ABs were collected in each run, interleaved with 15 baseline ABs. The ABs were

16 s long. For the SASS paradigm, two different stimulus designs were used such that the stimulus was presented coincident with the start of the AB (in-AB stimulus) or during the AFP immediately before task AB (i.e. end of stimulus coincident with the start of task AB) for the other (pre-AB stimulus). For the visual experiment with the SS paradigm, the stimulus was presented every 32 seconds allowing well-defined baseline and task regions (16 s). The FA and the TR was the same in both SASS and SS runs so that the steady-state magnetization was identical.

For the auditory experiments using SASS paradigm, we utilized an AB length of 12 seconds with auditory stimulus presented during AFP (pre-AB stimulus). The design had alternating baseline and task ABs, with a total of 20 stimulus trials in each run. We presented alternating (50ms ON/OFF) high frequency pure tones (16K Hz - 20K Hz) for 1 second every 40 seconds before the start of each task AB, which includes the 8 s long AFPs between ABs.

The monkeys were positioned in a horizontal MR system operating at 9.4T, adopting a sphinx position within an MRI-compatible restraint system with their heads secured using a head post[26]. An MR-compatible camera (model 12M-i, MRC Systems GmbH, Heidelberg, Germany) was integrated in the MRI chair to monitor the animals during data acquisition. Both the onset of visual and auditory stimuli was synchronized with the MRI using TTL (transistor-transistor logic) pulses from the scanner and a custom-written Python program running on a Raspberry Pi 4 model B (Raspberry Pi Foundation, Cambridge, UK). The TTL pulses were utilized to ensure precise synchronization of stimulus onsets and volume acquisition. For the visual stimuli, we custom-built a setup with two bilaterally attached white LEDs, positioned approximately 10 cm from the animals' head location. For the auditory stimuli, following the head fixation steps, MRI-compatible auditory tubes (S14, Sensimetrics, Gloucester, MA) were inserted into the animals' ear canals bilaterally and secured using sound-attenuating silicone earplugs (Amazon) and self-adhesive veterinary bandage[26,30].

Data analysis and statistics. The DICOM functional images were converted to NifTi format using AFNI's[31] dcm2niix function. EPIs acquired with opposite phase-encoding direction were used for EPI distortion correction (top-up) using custom-written MATLAB code. Using the FSL toolbox[32], the raw functional images were preprocessed by applying SUSAN noise reduction, MCFLIRT motion correction and spatial smoothing at full width at half-maximum Gaussian kernel (FWHM) of 0.65mm. For each run, brain extracted functional images were linearly registered to the respective manually skull stripped T2-weighted anatomical images of the animal using FSL's FLIRT function. The T2-weighted anatomical images were registered to the NIH marmoset brain atlas[33] via affine linear registration (12 degrees of freedom) using

FSL's FLIRT function. All preprocessed fMRI data for both SS and SASS paradigms were

then z-score normalized using the baseline ABs or regions within each run and mean shifted to

respective baseline temporal mean values.

The z-score transformed fMRI runs were used for statistical analysis using the general

linear model (GLM) approach in the FSL toolbox[32]. The data underwent slice-time correction

as the slices were acquired in interleaved order (0, 2, 4, …1, 3, 5, …). For comparison study

between SASS (in-AB) and SS paradigms, we used a gamma HRF peaking at 3 seconds (with

temporal derivates to find the best fit) for generating a stimulus onset regressor for GLM

analysis. For our pre-AB stimulus condition, where the stimulus is presented immediately

before the start of task AB, we use a custom truncated stimulus-HRF (without time-derivatives)

(Fig.2f) to match the stimulus onset to capture the characteristic BOLD peaking a second

earlier than in-AB stimulus. The pre-AB analysis used a fixed HRF (shown in Fig.2f) due to

limitations in the analysis software package for incorporating time-derivate for custom HRFs.

For each individual run, corrected voxels (Gaussian Random Field Theory (GRF) correction

for multiple comparisons using one-tailed t-test) with a 5.1 (visual) and 3.1 (auditory) z-value

thresholds at P<0.025 were considered active voxels. The z-value maps were generated per run

for each subject for all different experimental conditions. These maps were used for FSL's

fixed-effects group analysis for each condition resulting in group z-value maps for P<0.025.

All the group z-value maps were registered to the NIH marmoset brain atlas[33] template space

to visualize the differences in voxel responses across difference conditions.

REFERENCES

1. Kwong, K. K. *et al.* Dynamic magnetic resonance imaging of human brain activity during

primary sensory stimulation. *Proc. Natl. Acad. Sci.* **89**, 5675–5679 (1992).

2. Bandettini, P. A., Wong, E. C., Hinks, R. S., Tikofsky, R. S. & Hyde, J. S. Time course

EPI of human brain function during task activation. *Magn. Reson. Med.* **25**, 390–397 (1992).

3. Ogawa, S. *et al.* Intrinsic signal changes accompanying sensory stimulation: functional

brain mapping with magnetic resonance imaging. *Proc. Natl. Acad. Sci.* **89**, 5951–5955

(1992).

4. Mansfield, P. Multi-planar image formation using NMR spin echoes. *J. Phys. C: Solid*
State Phys. **10**, L55 (1977).

5. Buckner, R. L. *et al.* Detection of cortical activation during averaged single trials of a
cognitive task using functional magnetic resonance imaging. *Proc. Natl. Acad. Sci.* **93**,
14878–14883 (1996).

6. Bloch, F. Nuclear Induction. *Phys. Rev.* **70**, 460–474 (1946).

7. Mazaheri, Y., Biswal, B. B., Ward, B. D. & Hyde, J. S. Measurements of tissue T1 spin-
lattice relaxation time and discrimination of large draining veins using transient EPI data sets
in BOLD-weighted fMRI acquisitions. *NeuroImage* **32**, 603–615 (2006).

8. Miller_IM2011.pdf. https://users.fmrib.ox.ac.uk/~karla/pubs/Miller_IM2011.pdf.

9. Hung, C.-C. *et al.* Functional MRI of visual responses in the awake, behaving marmoset.
Neuroimage **120**, 1–11 (2015).

10. Toarmino, C. R. *et al.* Functional magnetic resonance imaging of auditory cortical fields
in awake marmosets. *NeuroImage* **162**, 86–92 (2017).

11. Peelle, J. E. Methodological challenges and solutions in auditory functional magnetic
resonance imaging. *Front. Neurosci.* **8**, 253 (2014).

12. Schmidt, C. F. *et al.* Silent and continuous fMRI scanning differentially modulate
activation in an auditory language comprehension task. *Hum. Brain Mapp.* **29**, 46–56 (2008).

13. Schwarzbauer, C., Davis, M. H., Rodd, J. M. & Johnsrude, I. Interleaved silent steady
state (ISSS) imaging: A new sparse imaging method applied to auditory fMRI. *NeuroImage*
29, 774–782 (2006).

14. Nebel, K. *et al.* Sparse imaging and continuous event-related fMRI in the visual domain:
A systematic comparison. *Hum. Brain Mapp.* **24**, 130–143 (2005).

15. Martino, F. D., Moerel, M., Ugurbil, K., Formisano, E. & Yacoub, E. Less noise, more
activation: Multiband acquisition schemes for auditory functional MRI. *Magn. Reson. Med.*
74, 462–467 (2015).

16. Feinberg, D. A. & Setsompop, K. Ultra-fast MRI of the human brain with simultaneous

multi-slice imaging. *J. Magn. Reson.* **229**, 90–100 (2013).

17. Mazo, C., Baeta, M. & Petreanu, L. Auditory cortex conveys non-topographic sound

localization signals to visual cortex. *Nat. Commun.* **15**, 3116 (2024).

18. Ibrahim, L. A. *et al.* Cross-Modality Sharpening of Visual Cortical Processing through

Layer-1-Mediated Inhibition and Disinhibition. *Neuron* **89**, 1031–1045 (2016).

19. Williams, A. M., Angeloni, C. F. & Geffen, M. N. Sound Improves Neuronal Encoding

of Visual Stimuli in Mouse Primary Visual Cortex. *J. Neurosci.* **43**, 2885–2906 (2023).

20. Hutton, C. *et al.* The impact of physiological noise correction on fMRI at 7T.

NeuroImage **57**, 101–112 (2011).

21. Logothetis, N. K., Pauls, J., Augath, M., Trinath, T. & Oeltermann, A.

Neurophysiological investigation of the basis of the fMRI signal. *Nature* **412**, 150–157

(2001).

22. Pan, W.-J. *et al.* Broadband Local Field Potentials Correlate with Spontaneous

Fluctuations in Functional Magnetic Resonance Imaging Signals in the Rat Somatosensory

Cortex Under Isoflurane Anesthesia. *Brain Connectivity* **1**, 119–131 (2011).

23. Yu, Y. *et al.* Sleep fMRI with simultaneous electrophysiology at 9.4 T in male mice. *Nat*

Commun **14**, 1651 (2023).

24. Moeller, S. *et al.* Multiband multislice GE-EPI at 7 tesla, with 16-fold acceleration using

partial parallel imaging with application to high spatial and temporal whole-brain fMRI.

Magn. Reson. Med. **63**, 1144–1153 (2010).

25. Gilbert, K. M. *et al.* A radiofrequency coil to facilitate task-based fMRI of awake

marmosets. *J. Neurosci. Methods* **383**, 109737 (2023).

26. Zanini, A. *et al.* In vivo functional brain mapping using ultra-high-field fMRI in awake

common marmosets. *STAR Protoc.* **4**, 102586 (2023).

27. Handler, W. B. *et al.* Design and construction of a gradient coil for high resolution

marmoset imaging. *Biomed. Phys. Eng. Express* **6**, 045022 (2020).

28. Haga, Y. *et al.* MR Imaging Properties of ex vivo Common Marmoset Brain after

Formaldehyde Fixation. *Magn. Reson. Méd. Sci.* **18**, 253–259 (2019).

29. Bodin, C. *et al.* Functionally homologous representation of vocalizations in the auditory

cortex of humans and macaques. *Curr. Biol.* **31**, 4839-4844.e4 (2021).

30. Jafari, A. *et al.* A vocalization-processing network in marmosets. *Cell Rep.* **42**, 112526

(2023).

31. Cox, R. W. AFNI: software for analysis and visualization of functional magnetic

resonance neuroimages. *Comput. Biomed. Res., Int. J.* **29**, 162–73 (1996).

32. Smith, S. M. *et al.* Advances in functional and structural MR image analysis and

implementation as FSL. *NeuroImage* **23**, S208–S219 (2004).

33. Liu, C. *et al.* A digital 3D atlas of the marmoset brain based on multi-modal MRI.

NeuroImage **169**, 106–116 (2018).

Data availability

The raw MR data acquired, and sequence protocol used for the study is available from

the corresponding author upon reasonable request.

Code availability

Custom python and MATLAB codes for stimulus paradigm, simulation and z-score

normalization are available from the corresponding author upon reasonable request.

Acknowledgements

We would like to thank all the members of the Centre for Functional and Metabolic

Mapping (CFMM) at Robarts Research Institute, Western University for their valuable

contributions to this project. We thank Dr Alex Li for valuable input on sequence design and

acquisition protocol. We thank Dr Kyle Gilbert and Peter Zeeman for the design and building

RF coil, MR animal chair and stimulus paradigm setup. We thank our veterinary staff Miranda

Bellyou, Whitney, Cheryl and Hannah for their help with conducting the animal experiments.

This work was supported by funding from the Canadian Institutes for Health Research (FDN-

148453), a Brain Canada Platform Support Grant and a Canada First Research Excellence Fund

award to BrainsCAN.

Author information

Authors and Affiliations

Graduate Neuroscience program, Schulich School of Medicine and Dentistry, Western

University, London, Canada.

Renil Mathew

Center for functional and metabolic mapping (CFMM), Robarts Research Institute,

Western University, London, Canada.

Renil Mathew, Amr Eed, L. Martyn Klassen, Ravi S Menon

Department of Medical Biophysics, Schulich School of Medicine and Dentistry, Western

University, London, Canada.

Ravi S Menon

Department of Physiology and Pharmacology, Western University, London, Canada.

Stefan Everling

Contributions

R.M and R.S.M involved in conceptualization of paradigm. R.M and AE designed and

performed the NHP experiments. R.S.M. wrote the simulation code. L.M.K developed the

sequence for Bruker scanner. S.E performed the surgeries. R.M analyzed the data. R.M and

R.S.M wrote the paper with additional contributions from of L.M.K.

Corresponding author

Correspondence to Ravi S Menon (rmenon@uwo.ca)

Ethics declarations

Competing interests

The authors declare no competing interests.

Fig.1 | Acquisition blocks and tSNR calculations. a, Schematics of basic recurring unit in an SASS paradigm with ABs characterized into baseline and task blocks separated by AFPs. Schematics also represent pre-AB and in-AB stimulus timings used in the study. **b,** Schematics for calculating baseline mean and standard deviation from baseline ABs, used for baseline z-score normalization and tSNR calculation (tSNR calculated from all ABs). **c,** Mean grey matter tSNR variance across AB volumes with maximum value in the first volume before approaching steady-state value in the last few volumes. **d,** Bar graph showing the tSNR gain in percentage relative to last volume in steady-state.

Fig.2 | Z-score normalization and comparison of raw data to theoretical modeling. a, Raw mean timeseries of task active voxels showing characteristic magnetization decays in all ABs, task AB's in orange color (number of volumes per run =480, TR 1 s). **b,** Mean zscore timeseries of task active voxels after baseline z-score normalization. **c,** Mean magnetization decays from baseline and task ABs of task active voxels (for pre-AB stimuli) from a single fMRI run (30 ABs), both decays normalized to the maximum value of mean task AB. **d,** Mean of z-score normalized task ABs of task active voxels from a fMRI run, normalized to maximum. **e,** Simulated baseline magnetization decay from Bloch's equations for TR:1 s, T1:2.2 s and FA:50 degrees, normalized to maximum of task magnetization. **f,** Truncated stimulus-HRF peaking at 3 second for a 1 second pre-AB stimulus (maximum BOLD percent signal change : 4%). **g,** Simulated baseline magnetization plotted with task magnetization (baseline magnetization x stimulus-HRF), normalized to maximum of task magnetization.

Fig.3 | Comparison study with visual stimulus paradigm. a, SASS paradigm group activation map (n=3) for an in-AB visual stimulus train and SS paradigm group activation map (n=3, P<0.025, z-value>5.1) from identical visual stimulus paradigm. **b,** Voxel count at respective z-values from group maps of primary visual cortex (V1) for both paradigms, z-value threshold at 5.1 (dashed line). **c,** Voxel count at respective z-values from group maps of medial superior temporal area of cortex (MST) for both paradigms, z-value threshold at 5.1 (dashed line).**d,** Normalized mean z-score timecourse of active voxels above z-value > 5.1 (P<0.025) across all subject runs of SASS paradigm (number of volumes per run = 480, TR 1s). **e,** Normalized mean z-score timecourse of active voxels above z-value > 5.1 (P<0.025) across all subject runs of SS paradigm (number of volumes per run = 480, TR 1s). (MG-medial geniculate, PG-parietal area)

a

b

c

d

Fig.4 l Pre-AB stimulus paradigms. a, Group activation maps (n= 2, z-value> 3.1, P< 0.025) for high-frequency pure tone auditory stimuli presented during the AFP using SASS paradigm. **b,** Auditory experiment: normalized mean zscore timeseries from active voxels z-value> 3.1 across all subject runs (number of volumes = 480, TR=1 s). **c,** Group activation maps (n= 3, z-value> 5.1,P< 0.025) for pre-AB visual stimuli during silent AFP. **d,** Visual experiment: normalized mean zscore timeseries from active voxels above z-value> 5.1 across all subject runs (number of volumes = 480, TR=1 s) .

Extended Data Fig.1 | Reduced trails comparision a. Group activation map (n=3) for in-AB visual stimulus experiment using SASS paradigm with different number of trials per run. First row shows group activation from 5 stimulus trials per run (160 volumes), second row with 8 stimulus trials per run (256 volumes) and third row showing 11 trials per run (352 volumes - matching the acquisition time with SS paradigm). **b.** Average voxel count at each z-values for V1 ROI with different reduced number of trials from SASS IN-AB paradigm with SS paradigm voxel count for 15 trials (480 volumes).

Extended Data Fig.2 | Mean of task active voxels across task AB/region. a, SASS-paradigm: In-AB visual stimuli BOLD response as mean z-score (normalized to maximum value) of task ABs across all subject runs (n= 3, P< 0.025, z-value> 5.1,). **b,** SS paradigm: Visual stimuli BOLD response as mean z-score (normalized to maximum value) of task regions across all subject runs (n= 3, P< 0.025, z-value> 5.1). **c,** SASS paradigm: pre-AB visual stimuli BOLD response as mean z-score (normalized to maximum value) of task ABs across all subject runs (n= 3, P< 0.025, z-value> 5.1) **d,** SASS paradigm: pre-AB auditory stimuli BOLD response as mean z-score (normalized) of task ABs across all subjects runs (n=2, P< 0.025, z-value> 3.1).

Extended Data Fig.3 | Segmentation using T1 decays. a, Mean magnetization decays of gray matter and white matter voxels across all ABs in a run, normalized to the maximum value.**b,** ICA component mask for white matter voxels from raw timeseries of one run.