Rapid identification of seven bacterial species using microfluidics, time-lapse phase-contrast microscopy, and deep learning

Erik Hallström[1*], Vinodh Kandavalli[2], Carolina Wählby[1], Anders Hast[1]

1 Department of Information Technology, Uppsala University, Uppsala, Sweden

2 Department of Cell and Molecular Biology, Uppsala University, Uppsala, Sweden

* erik.hallstrom@it.uu.se

Abstract

For effective treatment of bacterial infections, it is essential to identify the species causing the infection as early as possible. Currently, the main techniques for determining the species require hours of overnight culturing of a bacterial sample and a larger quantity of cells to function effectively. This study uses one-hour phase-contrast time-lapses of bacterial growth collected from a microfluidic chip, also known as a "mother machine". Then, this data is used to train deep artificial neural networks to identify the species. Both video and image classification models of the Vision Transformer (ViT) and Convolutional Neural Network (CNN) families were evaluated in this study. We have previously demonstrated this approach to classify four different species, which is now extended to seven species: *Pseudomonas aeruginosa*, *Escherichia coli*, *Klebsiella pneumoniae*, *Acinetobacter baumannii*, *Enterococcus faecalis*, *Proteus mirabilis*, and *Staphylococcus aureus*. The models are then trained and evaluated using subsampled images, simulating using lower-resolution microscopy in a potential clinical setting. The method can be used simultaneously with phenotypic antibiotic susceptibility testing (AST) in the microfluidic chip. The best models attained an average precision of 93.5% and a recall of 94.7% (0.997 AUC) on a trap basis in a separate, unseen experiment with mixed species within 70 minutes. However, in a

real-world scenario, one can assume many traps will contain the actual species causing the infection, improving the reliability. The experiments suggest spatiotemporal features can be learned from video data of bacterial cell divisions, and both textures and morphology contribute to the final performance of the models. Rapidly identifying responsible species causing acute infection and simultaneously performing drug sensitivity in a much shorter time than today — reduced from days to hours or minutes — could lead to a paradigm shift in how initial treatments for severe bacterial infections are chosen. These developments could contribute to the fight against antibiotic resistance, improve patient outcomes, and ultimately save lives.

Author Summary

Acute bacterial infections are initially treated by administering many different antibiotics (broad-spectrum) to a patient. Specimens taken from the patient will then be sent to a microbiology laboratory to be cultivated on agar plates, after which the species and the resistance profile to different antibiotics can be obtained. Depending on this information, the patient's treatment will be adjusted, but the process is work-intensive and typically takes more than 24 hours. However, prompt treatment with suitable antibiotics is critical for the patient's survival, especially for multi-resistant strains. This study demonstrates a method to speed up this species identification. It utilizes a microfluidic chip, also known as a "mother machine", to film seven different bacterial species for one hour using phase-contrast microscopy. Artificial neural network models are then trained to identify the bacterial species in an unseen experiment, only using microscopy video data of bacteria reproducing inside traps of the microfluidic chip. The best model achieved a precision of 93.% and a recall of 94.7%. Furthermore, a large part of the performance was retained when the network was trained to recognize bacteria at low resolution, demonstrating the potential to use the method in a clinical setting using lower-magnification microscopy. The technique opens the door for more effective and targeted treatment of acute bacterial infections and could ultimately save lives.

1 Introduction

Since its introduction in the mid-20th century, antibiotics have been pivotal for modern healthcare, saving countless lives [1]. It plays an integral role, not only in treating bacterial infections, but also in preventing them in patients with weakened immune systems, such as those undergoing chemotherapy for cancer treatment and transplant recipients [2]. Currently, antimicrobial resistance is on the rise [3] [4] [5] [6] [7], partly due to the overuse of broad-spectrum antibiotics [8] [9]. Rapidly identifying the species causing the infection is necessary to mitigate this, as it helps select the best-optimized antibiotic therapy and reduces the likelihood of generating new resistance with antibiotic misuse. Also, for sepsis patients, prompt treatment with appropriate antibiotics is critical for survival, with every hour counting [10].

Today, the primary bacterial identification method is Matrix-Assisted Laser Desorption/Ionization Time-of-Flight (MALDI-TOF) mass spectrometry [11]. Although very accurate, it requires overnight pre-cultivation (18-24 hours) on agar plates for large enough colonies to form and is rather work-intensive and costly. Several techniques have been explored to reduce the detection time. One possibility is using microscopy to analyze time-lapses of micro-colony growth, reducing the detection time to 6-12 hours [12] [13]. Raman spectroscopy, illuminating bacteria samples with laser and analyzing the scattered wavelength spectrum, typically also requires a few hours for the microcolonies to form [14] [15]. Another method is genotyping by fluorescence microscopy [16]. While not requiring pre-cultivation, the method is destructive as the cell walls must be dissolved so the fluorescent stain can enter the bacteria, also, the number of species that can be detected is limited.

Microfluidic devices for monitoring single-cell bacteria growth (example visualized in S9 Video) are fabricated with oblong traps with a width of around 1.5µm, laid out orthogonal to a larger center channel providing growth medium [17]. The traps accommodate only a single column of cells, with the mother cell located at the top of the trap, pushing the descendants down and eventually out into the center channel, hence the name "mother machine". A modification to this chip design introduced by Baltekin et. al. [18] had a physical stop, protruding from the side of the trap wall, stopping the bacterial cells but allowing fluids to pass by. This allowed the microfluidic

chip to be loaded faster using pressure techniques instead of diffusion methods used ³²
before. Furthermore, growth medium or other substances, such as antibiotics, could be ³³
directly administered to the cells. ³⁴

Data from this particular chip design introduced by Baltekin et. al. was used in this ³⁵
study. The overall experiment setup is shown in Fig 1. Note that fluorescence ³⁶
microscopy is only used to obtain labels for the dataset; the classification is performed ³⁷
using phase-contrast time-lapses of living cells reproducing in the traps. A schematic of ³⁸
the bacteria reproducing in a mother machine is shown in Fig 1A, with the physical ³⁹
stop protruding out from the top-left side of the trap wall. Phase-contrast ⁴⁰
microscopy [19] is then used to collect a time-lapse containing single-cell growth inside ⁴¹
the traps of the mother machine Fig 1B. After capturing the growth, cells are ⁴²
permeabilized, and fluorescently labeled probes binding to species-specific nucleic acid ⁴³
are administered to the cells for genotyping. Following the application of laser light, the ⁴⁴
cells emit light, revealing the species of bacteria in each trap. This procedure is referred ⁴⁵
to as fluorescence in situ hybridization (FISH) Fig 1C. Each trap is then cropped out, ⁴⁶
containing a phase-contrast time-lapse of bacterial growth. The fluorescence signal in ⁴⁷
the final frame provides the ground truth label. Then, deep learning video or image ⁴⁸
classification models are trained to classify the species growing inside each trap using ⁴⁹
only the phase-contrast data, illustrated in Fig 1D. We have previously demonstrated ⁵⁰
this technique for four species [20], using a separate fluorescent stain for each species. ⁵¹
Recently Kandavalli et al. [16] have developed a method to identify seven species using ⁵²
fluorescence microscopy where the fluorescent probes attach to one or two species, ⁵³
combining the signal, where *E. coli* is visible in Cy3, *K. pneumoniae* in Cy3 and Cy5, *E.* ⁵⁴
faecalis in Alexa488, *P. aeruginosa* in Alexa488 and Txr, *A. baumannii* in Alexa488 ⁵⁵
and Cy3, *P. mirabilis* in Alexa488 and *S. aureus* in Cy5 and Txr. Thus the dataset ⁵⁶
contains two cocci species with a circular shape and five rod-shaped species. This study ⁵⁷
uses the dataset available from this publication for testing and evaluation. A ⁵⁸
phase-contrast microscopy time-lapse followed by the four fluorescence frames of one ⁵⁹
cropped-out trap with *A. baumannii* one whole position is shown Fig 2. One whole ⁶⁰
position with the procedure is visualized in S9 Video. ⁶¹

A. Microfluidic mother machine   B. Phase-contrast timelapse

C. Fluorescence   D. Neural network

Fig 1. Experimental setup A: Schematic of a bacteria reproducing in a mother machine, the arrow on the right shows the growth direction. B: Phase-contrast time-lapse images of bacteria growing and reproducing in a mother machine. C: The fluorescence images captured after the final phase-contrast frame reveal the species in each trap (used as ground truth labels), with *E. coli* in green, *K. pneumoniae* in cyan, *E. faecalis* in magenta, *P. aeruginosa* in orange, *A. baumannii* in yellow, *P. mirabilis* in red and *S. aureus* in blue. Examples of cropping targets (traps containing a single species) for each of the seven species are outlined with dashed lines and pointed to by arrows in corresponding colors. D: Input to the neural network is phase-contrast only, a single frame or time-lapse video. Borders are color-coded according to the respective species.

Fig 2. One trap phase-contrast time-lapse and final fluorescence images
Image of a cropped-out trap in the mother machine growing for 70 minutes and imaged every two minutes, resulting in 36 phase-contrast frames in the time-lapse (numbered 0-35), and the final fluorescence images for each channel are shown side-by-side. This sample is *A. baumannii* from Experiment 14 of the training set (see Table 1 in Materials and Methods).

Since the early 2010s, deep artificial neural networks (referred to as "deep learning") [21] have emerged as the most prominent technique for analyzing data across many modalities such as images [22] [23], text [24] [25], video [26] and audio [27]. In this study, both image and video classification methods were trained and evaluated from the two main families, Vision Transformer (ViT) [23] and Convolutional Neural Network (CNN) [28] [29]. The models were trained on a phase-contrast dataset of 18101 traps obtained from 16 experiments and tested on a hold-out experiment containing 1399 traps. Each trap contains single-species growth captured with phase-contrast microscopy, around 1-hour time-lapses, 2 minutes per frame. The whole dataset consists of 621,931 images. The best model attained a precision of 93.5% and a recall of 94.7% (0.997 AUC) on a trap basis within 60 minutes. Furthermore, the performance was only slightly degraded when subsampling, allowing it to be implemented in a low-resolution clinical setting. Thus, the overall aim of this study - to identify seven bacterial species

solely using phase-contrast time-lapses was achieved.   75

2 Results   76

In this study, two CNN-based video classification networks were used: R(2+1)D with   77
31.5M parameters and R3D Video ResNet with 33.4M parameters from [26]. In   78
addition, three CNN-based image classification networks: ResNet 18 [22] with 11.7M   79
parameters, RegNetY 016 [30] with 11.2M parameters, EfficientNet B3 [31] with 12.23M   80
parameters, and one ViT-based FastViT SA12 with 11.6 parameters [32]. The entire   81
trap had a width of 52 pixels and a height of 1500 pixels, but the spatial input to the 82
models was set to 52x114 pixels. The growth extent in each trap could be obtained   83
using a semantic segmentation model. Using this vertical growth, a number of crops,   84
either still images (S1 Fig) or video clips (S2 Fig), could be extracted in a tiling window   85
fashion. The final output for a trap was obtained by summing the output logits from   86
each crop (image or video clip). The networks were retrained for each setting to ensure 87
reproducibility: five times for the scaling experiments (except for the data scaling   88
experiment) and 30 times for the categorical experiments.   89

The evaluation metrics used were precision, recall hold-one-out, and micro-averaged   90
Area Under the Curve (AUC) (see Metrics in Materials and Methods section). One   91
unseen experiment was used for testing data, from which 447 *A. baumannii*, 287 *K   92
pneumoniae*, 175 *S. aureus*, 154 *P. mirabilis*, 116 *E. coli*, 139 *E. faecalis*, and 81 *P.   93
aeruginosa* traps were extracted. The test set had 35 phase-contrast frames captured   94
during 70 minutes.   95

All the following visualizations use Strp plot or Beeswarm plot from the Seaborn   96
visualization library [33], which are scatter plots with jitter added to the points for   97
readability. Each point represents a trained neural network. The lines are added to   98
track the mean values in the scaling experiments and are not meant for interpolation or 99
statistical inference. Vertical lines in the categorical plot show standard deviation. The   100
trap visualizations use the Viridis colormap [34] to show the grayscale one-channel   101
phase-contrast images.   102

The main results are shown in Fig 3, testing the AUC gradually increasing the   103
frames in the time-lapse, each frame corresponding to 2 minutes. Using video   104

classification is better than aggregating still images after a few minutes. In this case, 105
ResNet 18 outperforms the modern FastViT. No substantial improvements were 106
observed using the other more modern CNN designs. The video classification network 107
reaches 0.99 AUC within 30 minutes. Notably, spatiotemporal features are important to 108
differentiating the rod species *E. coli*, *K. pneumoniae*, and *P. mirabilis*. 109

Fig 3. Time evaluation Comparing model and Video ResNet R(2+1)D species
classification performance over time, gradually using more frames in the time-lapse.

A confusion matrix of the Video ResNet R(2+1)D model trained using 35-frame 110
time-lapses is shown in Fig 4. A few misclassifications of this model are shown in 111
S3-S8 Figs. 112

Fig 4. Video classification performance Classification performance of Video ResNet R(2+1)D utilizing the full crop size 52x114 pixels, using all 35 frames (full time-lapse). Left: Confusion matrix for the classifier with the median AUC among the five retrainings. Right: Receiver operating characteristic curves for the same model. The legend displays the mean and standard deviation of the AUC across the five retrainings.

2.1 Subsampling experiments

The networks were retrained with the images spatially subsampled to test the possibility of using the models in a lower resource setting. Lancoz [35] subsampling was used to simulate using lower-resolution microscopy. In total, 28 subsampling steps were used from 52x114 pixels to 3x7 pixels, visualized in S1-S7 Videos. The evaluation was performed on the whole time-lapse (35 frames representing 70 minutes). Video ResNet R(2+1)D again attained the best performance, AUC only dropping to 0.96 when using 35-frame time-lapses at 3x7 pixels (very low resolution) shown in Figs 5.

Fig 5. Subsampling model comparison Retraining the networks with the image input spatially subsampled during test and training, testing on the whole time-lapse. Left: Model comparison. Using spatiotemporal features (video classifier) is especially important at very low resolution. Right: Species comparison using Video ResNet R(2+1)D at 5x11 pixels. Four species attain classification performance of 0.99 AUC even at very low resolution.

A confusion matrix of the Video ResNet R(2+1)D model trained using 35-frame time-lapses operating at 5x11 pixels is shown in Fig 6, indicating an increasing number of rod-shaped species confused. A few misclassifications of this model are shown in S9-S14 Figs.

Fig 6. Video classification performance at 5x11 pixels Classification performance of Video ResNet R(2+1)D utilizing spatially subsampled images 5x11 pixels, using all 35 frames (full time-lapse). Left: Confusion matrix for the classifier with the median AUC among the five retrainings. Right: Receiver operating characteristic curves for the same model. The legend displays the mean and standard deviation of the AUC across the five retrainings.

Furthermore, time evaluations were performed to assess the performance degradation ₁₂₅
of Video ResNet R(2+1)D, training and testing using gradually more spatially ₁₂₆
subsampled time-lapses, shown in Fig 7. ₁₂₇

Fig 7. Video classification performance at various subsampling steps Left:
Degradation of performance during time evaluations using Video ResNet R(2+1)D
trained with spatially subsampled data. Distinct performance jumps occur at specific
time points for the highly subsampled data, highlighting the significance of
spatiotemporal features at low resolutions. Right: Time evaluation of Video ResNet
R(2+1)D operating at spatially subsampled crop size of 5x11, visualizing species-specific
AUC.

A comparison of all video and image classification models is shown in Fig 8, notably ₁₂₈
Video Resnet R(2+1)D using decomposed convolutions in spatial (2D) and temporal ₁₂₉
(1D) domains performs better and is more stable than Video ResNet 3D utilizing 3D ₁₃₀
convolutions (see Deep learning model training in Materials and Methods for details). ₁₃₁
There is no apparent improvement when using more modern CNN designs such as ₁₃₂
EfficientNet, RegNetY, or the modern ViT model, FastViT. ₁₃₃

Fig 8. Video classification performance at various subsampling steps Model comparison between video and image classification at gradually lower resolutions. Video classification performs better, especially at lower resolutions.

2.2 Augmentation ablation study

Several augmentation methods from the Albumetation library [36] were used during the training process. Additionally, two custom developed methods were employed: Random video frame erasings and Segment (see Augmentations in Materials and Methods for details). A range of experiments were performed using Video ResNet R(2+1)D, removing one of these augmentations during training. In one experiment, the models were trained from scratch with randomly initialized weights instead of using pretrained weights used by default. Finally, one experiment was conducted removing all train augmentations.

Visualized in Fig 9, RandomBrightnessContrast had the largest impact on performance, followed by Label smoothing [37], suggesting that the models easily overfit to experiment settings (such as illumination and contrast). Removing all augmentations was detrimental to the performance.

Fig 9. Removing augmentation The performance was tested by removing one augmentation and retraining the Video ResNet R(2+1)D. The baseline with all augmentation present is shown using a dashed red line. The plots are split up for readability due to the removal of RandomBrightnessContrast having a larger impact. Also, using label smoothing was beneficial, preventing the networks from overfitting. The removal of all augmentations experiment used randomly initialized weights and no label smoothing.

2.3   Texture, morphology, and division pattern

The full-resolution Video ResNet R(2+1)D and ResNet 18 models were retrained and evaluated using only pre-generated segmentation masks (see Segmentation model in Materials and Methods). Hence, texture information was removed. Then, this was repeated, but instead replacing each pixel in the images with the row-wise average, meaning that the models could only see division patterns and cell length (refer to Image Modifications in Materials and Methods for more details).

The results are visualized in Fig 10. The video model, which can observe growth and division speed through row-wise mean modifications, has a much smaller performance drop than the still-frame image classification model, which can only infer cell length from the static images.

Fig 10. Distortions Assessing the significance of texture, morphology, cell length, and growth and division speed by comparing static images classified using ResNet 18 with video clips classified using Video ResNet R(2+1)D.

2.4 Data scaling experiments

Data scaling experiments were performed using ResNet 18 training on subsampled images at 12x26 pixels. The network was retrained multiple times with a progressively larger training set by increasing the number of included traps per species. It should be noted that the traps were sorted based on the data quality score from Algorithm 1 (see Label Adjustments in Materials and Methods), thus high-quality samples were included first when scaling. The networks were also trained without data augmentation. As shown in Fig 11, applying data augmentation was essential for this dataset. The performance plateaus for the non-augmented models, implying that the variability in the training set may be low (refer to Table 1 in Materials and Methods). A training set collected from multiple experiments with mixed species could yield better non-augmented results.

Fig 11. Data scaling Performance when including more traps per species in the dataset. The lines are added moving averages with a rolling window size of 5 to track the general trend. Augmentation was essential across all dataset sizes for the model's ability to generalize, indicating low variability in the training set. The models also appear to overfit when trained on too much non-augmented data.

3 Discussion

This study performs rapid seven-species bacteria classification in an unseen experiment in a microfluidic device, the "mother machine", only using phase-contrast microscopy time-lapses analyzed by deep learning models. The setup in this paper was more challenging for the models than our previous study [20], which used seven experiments, all with four mixed species loaded into each chip section, testing on a random 15% train-test split (splitting on a trap basis). On the contrary, the dataset in this study had only 1-2 species per chip and chip section in the training data, and testing was performed using a separate experiment with mixed species in both sections. Also, *K. pneumoniae* and *E. coli* were more distinguishable in our previous study (*K. pneumoniae* being slightly thicker and bulkier). This study introduced two new rod species, *P. mirabilis*, and *A. baumannii*, which were visually almost identical to *E. coli* and *K. pneumoniae*. Thus, the networks must learn subtle features that are not always visible to the human eye. Furthermore, this study expands upon our previous work by examining how texture, morphology, cell length, division speed, dataset size, augmentations, and increase in the number of testing frames (corresponding to minutes

in a potential clinical setting) affect performance. 186

The methodology can be applied prior to or concurrently with phenotypic AST testing and give an early warning before results of other approaches to species identification, such as MALDI-TOF, arrive from the medical laboratory. Additionally, the growth can be monitored for an extended time longer than one hour, updating the estimate as time progresses. Similarly to our previous study, a significant proportion of the performance is retained at low resolution, especially when using video classification networks, enabling the usage of simpler microscopes in a clinical setting.

The experiments suggest that both morphology, textures, and spatiotemporal features contribute to the final performance of the models and that learning spatiotemporal is especially important at lower resolutions. Augmentation and label smoothing greatly improved the performance, indicating the models easily overfitted to the trainset. The overfitting occurred possibly due to each chip only being loaded with 1-2 species, shown in Table 1. Collecting a larger, more diverse dataset could enhance model performance and robustness. Interestingly, modern neural network architectures did not outperform ResNets in this study.

No validation set was used, which could be considered a limitation of this study. However, no hyperparameter optimization was performed, all models were trained using identical settings, including augmentation, batch size, learning rate, and number of epochs.

Another limitation of this study is that the experiments use bacterial lab isolates, genetically identical and growing in the same type of media. Future work would continue investigating the methodology using other species and, importantly, different clinical isolates, as these can exhibit varied phenotypes.

Although this study utilizes a specific microfluidic chip, the "mother machine," one can envision classifying bacterial species based on spatiotemporal features from single-cell growth in other settings captured through different image modalities. This enables future work to extend beyond the particular setup used in this paper.

The study further demonstrates the potential of training deep learning models to identify bacterial species in a much shorter time than routine methods today. Instead of days to obtain positive cultures, typing, and phenotypic drug sensitivity testing results, these new methods decrease the waiting time to hours or even minutes. Such

developments could lead to a paradigm shift in how to treat acute bacterial infections, moving from the current "try-and-test" approach to more knowledge-based decisions. This could improve patient outcomes, minimize antibiotic misuse, and ultimately save lives.

4   Materials and methods

The following sections outline the main components of the software and data. However, for full implementation details, refer to the released replication package [38], which contains documentation, functions to reproduce the experiments, as well as pre-trained models. Python version 3.10.12 was used for all software, apart from the libraries stated below the software utilized NumPy [39], Pandas [40], and Matplotlib [41].

4.1   Dataset

The dataset was obtained from Kandavalli. et al. [16], for detailed information on the biological experiment, please refer to their publication. The data consisted of 17 different experiments, each loading a microfluidic chip with a bacterial sample of 1-7 species and monitoring bacterial growth for 58-74 minutes, where each frame was taken every 2 minutes, thus 30-36 phase-contrast frames in each time-lapse. The phase-contrast microscopy, shown in Fig 1B, and final fluorescence images, shown in Fig 1C, were captured at each position on the chip where a position consisted of 42 traps and had an image size of around 2000x4000 pixels (one position also shown in S9 Video). Each trap was cropped out using a cropping program outlined below, resulting in crops with an image size of 52x1500 pixels, 30-36 phase-contrast frames, and four fluorescence frames per crop. Empty traps and traps containing multiple species were discarded using a labeling tool described below. Each microfluidic chip contained two sections, "Up" and "Down", where a section could be loaded with a specific species. The crosstable of the dataset used for training the neural networks is shown in Table 1, and the corresponding number of species extracted from each experiment. A number of sample images from the dataset are shown in Fig 12.

Table 1. Data summary

	Frames	A. baumannii Up	A. baumannii Down	E. coli Up	E. coli Down	E. faecalis Up	E. faecalis Down	K. pneumoniae Up	K. pneumoniae Down	P. mirabilis Up	P. mirabilis Down	P. aeruginosa Up	P. aeruginosa Down	S. aureus Up	S. aureus Down	All species Total
Experiment 1	33	0	0	0	0	0	0	415	466	0	0	0	0	0	0	881
Experiment 2	32	0	0	0	0	0	0	0	0	0	0	955	1169	0	0	2124
Experiment 3	32	0	0	0	0	136	1211	0	0	0	0	0	0	0	0	1347
Experiment 4	31	0	0	733	830	0	0	0	0	0	0	0	0	0	0	1563
Experiment 5	30	0	0	0	21	0	0	40	0	0	0	0	0	0	0	61
Experiment 6	30	0	0	0	0	390	0	0	0	0	0	0	47	0	0	437
Experiment 7	32	0	183	0	0	0	0	0	0	528	0	0	0	0	0	711
Experiment 8	30	0	0	0	0	0	0	0	0	0	0	0	0	656	664	1320
Experiment 9	30	0	0	134	128	66	156	0	0	0	0	0	0	0	0	484
Experiment 10	30	0	0	0	0	0	0	813	866	0	0	0	0	520	670	2869
Experiment 11	35	442	604	0	0	0	0	0	0	0	0	1	66	0	0	1113
Experiment 12	30	0	0	0	64	0	0	0	0	572	651	0	0	0	0	1287
Experiment 13	29	0	0	49	45	79	211	0	0	0	0	0	0	0	0	384
Experiment 14	36	674	697	0	0	0	0	0	0	0	0	120	242	0	0	1733
Experiment 15	31	0	0	179	174	94	257	0	0	0	0	0	0	0	0	704
Experiment 16	31	0	0	0	243	0	0	0	0	399	441	0	0	0	0	1083
Experiment 17	35	210	237	53	63	73	66	118	169	60	94	37	44	69	106	1399
All experiments		1326	1721	1148	1568	838	1901	1386	1501	1559	1186	1113	1568	1245	1440	19500

Time-lapses extracted from all seventeen experiments; each entry corresponds to a microfluidic trap containing only one species of bacteria. The last experiment highlighted and marked in bold was used for testing.

Fig 12. Samples from the dataset Ten concatenated sample images (52x512 pixels) of each bacterial species taken from randomly selected traps. Visually distinguishing between the rod-shaped bacteria in this dataset is challenging, even for trained humans. Specifically, differentiating between *E. coli*, *P. mirabilis*, and *K. pneumoniae*. *P. aeruginosa* often has a characteristic thin-curved shape, while *A. baumannii* is comparatively shorter and thicker, however, these are not always reliable features. *S. aureus* had a larger diameter than *E. faecalis* for these isolates.

4.2   Image modifications

The networks were trained and evaluated using three different modifications on the phase-contrast data, visualized in Fig 13. The modifications are displayed side-by-side in S8 Video.

- **Default:** Original images without any modifications.

- **Mask:** Semantic segmentation masks pre-generated from the segmentation model (refer to the Segmentation Model section below).

- **Mean:** The pixel value of every pixel was replaced by the row-wise mean.

Furthermore, a final mode, "Segment" where the background was removed with a probability, was used as an augmentation step during training.

Fig 13. Samples from the dataset Four modifications visualized for a part of a single trap (52x256 pixels, 15 frames). In the row-wise mean images, the cell height and division speed is visible, but not morphology and texture information. In the mask, no texture information is visible, and for the segment, all background is blocked out, aiming to reduce the risk of models overfitting to experimental settings.

4.3   Stabilizing and cropping out traps

Our previous 4-species study [20] included software to crop and stabilize the phase contrast timelapse, but it encountered frequent failures when applied to this dataset. Therefore, a new pipeline was developed using classical image processing methods to align, register, and crop out traps from the raw images. The main technique used in the pipeline was projecting the raw phase-contrast images on the vertical and horizontal axis. The resulting signal contained peaks where the traps (seen in Fig 1B and S9 Video) were located on the horizontal axis.

The pipeline started with rotating the images since the chip had a slight tilt in the raw microscopy images, usually around 1-3 degrees. When the chip structure was perfectly horizontal, the peaks in the projections were at maximum values. The correct rotation was determined through an adaptive grid search, with progressively refined step sizes of 1, 0.1, and finally 0.05 degrees.

The rotation step was followed by a custom-developed rough registration method for aligning the phase-contrast time-lapse frames and fluorescence images. This was necessary because standard registration methods occasionally mixed up the vertical traps due to their regular striped pattern, especially when significant horizontal drift occurred during the timelapse. The custom registration started by performing an

adaptive threshold of the phase-contrast images revealing the chip structure. Each image in the phase-contrast stack was then shifted a number of pixels horizontally and vertically until aligned to the first image in the time-lapse (evaluating on the minimum of bitwise XOR between the binary images). This procedure was repeated, but instead registered each phase-contrast frame to the previous frame in the time-lapse. Then, the fluorescence images and phase-contrast images were projected on the horizontal axis, observing peaks that revealed the horizontal location of the traps. The fluorescence images were aligned to the last phase-contrast frame, shifting horizontally until the projections were aligned, evaluating using the absolute difference.

After this rotation and registration step, the phase-contrast images were projected onto the horizontal axis, and the horizontal location of each trap was extracted. The traps were cropped vertically, using a distance of 26 pixels from the center position on each side of a peak, extracting crops of 52 pixels width. A procedure was developed to crop the images horizontally right at the physical stop protruding from the side of a wall at the top of each microfluidic trap (visible as a circular blob in the top-right of each phase-contrast frame in Fig 2). The procedure utilized the horizontal asymmetry introduced in the images due to this physical stop. The resulting crops had a size of 52x1500 pixels, showing a frame of a single trap. Finally, a fine registration was performed on each cropped-out phase-contrast stack using StackReg [42] using only translation (no scaling, rotating, or shear).

Barcodes were evenly laid out to locate traps in the chip, as seen on the sides and middle in Fig 1B and S9 Video. Typically, there were three barcodes visible in each position, but one on the side could go out of view if there was significant horizontal drift during capture. This drift introduced a problem since the barcodes generated peaks similar to a trap during horizontal projection. The cropped-out phase-contrast stacks extracted from the sides of the chip were projected onto the vertical axis to determine if they were barcodes. If the resulting 1D signal contained strong peaks with the width of a barcode dot, it was classified as a barcode. Also, the median pixel intensity was lower if the crop contained barcodes, which was used as an additional predictor.

The pipeline also generated debug images on each raw phase-contrast and fluorescence frame, highlighting the crop locations to ensure no errors occurred during the preprocessing.

4.4 Labeling tool

A program was developed to perform fluorescent signal aggregation and assign species labels to each trap. Examples of the four fluorescence channels are visualized on the very left in Fig 2. In the dataset, *E. coli* was visible in Cy3, *K. pneumoniae* in Cy3 and Cy5, *E. faecalis* in Alexa488, *P. aeruginosa* in Alexa488 and TxR (Texas Red), *A. baumannii* in Alexa488 and Cy3, *P. mirabilis* in Alexa488 and *S. aureus* in Cy5 and TxR. Assigning a label proved challenging, as the images were often noisy and exhibited significant variations in fluorescence intensity across the fluorescent channels. The tool used standard background removal followed by applying a fluorescence segmentation model, selecting regions with bacterial illumination. The fluorescence images of the trap were then projected on the vertical axis, which converted the images into a 1D signal. At each vertical coordinate a species label was extracted using fine-tuned thresholds comparing the signal of each channel. The experiments were often performed with a few species per position, as shown in Table 1, and the labeling tool could use this prior information when assigning a label.

In total, 73706 traps were labeled using the labeling tool, and 38570 of these were retained containing single-species growth, discarding traps with multiple species or empty traps.

4.5 Data cleaning and label adjustments

A number of issues with the staining process were observed in the data. The cells in the final phase-contrast frame were not always vertically aligned with the fluorescence images. This misalignment may be due to cell shrinkage when dissolving the cell walls and the time lag between probe addition and fluorescence image capture.

Additionally, disparities in probe absorption were observed among the species; probes of one stain had a reduced absorption in certain species. This resulted in some fluorescent signals being notably dimmer or emitting no signal at all. Examples of these are shown in S15-S18 Figs.

After reviewing the output, 95 traps were discarded from the test set where the labeling tool clearly made an error, and the trap contained multiple species. Furthermore, 14 test traps were discarded due to either disconnecting from the top

during growth or being empty for more than 15 consecutive frames (equivalent to 30 minutes). Additionally, 20 traps were relabeled as *K. pneumoniae* containing a dim but present signal in Cy5, and one trap was relabeled as *P. mirabilis*. Due to the low absorption of the fluorescent probe Alexa488, 25 samples were relabeled as *K. pneumoniae* that were previously unlabeled. For transparency, all test-trap label adjustments are displayed in the S1 Appendix.

Similarly, a total of 355 traps were adjusted or discarded in the training set. These mislabelings, in both the training and test sets, were identified by analyzing network misclassifications. One intrinsic property of neural networks is their robustness to label noise [43]. Network misclassifications in both the training and test sets were often due to erroneous labeling.

After these adjustments, the training set was significantly unbalanced with 9719 *E. coli*, 7687 *P. aeruginosa*, 5957 *K. pneumoniae*, 4570 *A. baumannii*, 3408 *E. faecalis*, 2591 *P. mirabilis* and 2510 *S. aureus*. The ResNet 18 and Video ResNet R(1+2)D were then trained according to the following settings:

- Training on the full imbalanced dataset, with traps from each species randomly sampled for each batch in proportion to their fraction in the dataset, using a Random Weighted Sampler in Pytorch [44].

- Training on a balanced, capped dataset, containing a maximum of 2600 traps per species, selected randomly.

- Training on a balanced, capped dataset, containing a maximum 2600 of traps per species as previously, but first ranking the traps based on data quality outlined in Algorithm 1. For species with more than 2,600 samples (*E. coli*, *P. aeruginosa*, *K. pneumoniae*, *A. baumannii* and *E. faecalis*), this meant retaining the 2600 traps with the highest data quality scores.

The dataset was evaluated on the same hold-out test set as before, Experiment 17 in Table 1. The AUC results over thirty runs were for ResNet 18: 0.97 ± 0.004, 0.976 ± 0.002, and 0.987 ± 0.001, respectively, and for Video ResNet R(1+2)D: 0.983 ± 0.003, 0.984 ± 0.004 and 0.997 ± 0.001, visualized in Fig 14. Thus, using data quality scoring significantly improved performance. Consequently, all experiments conducted in this

paper used a training set obtained by limiting at 2600 traps per species following the

data scoring outlined in Algorithm 1.

Algorithm 1 Pseudocode for grading the data quality of each trap Traps with steady growth through many small steps, without abrupt jumps, and a clear fluorescent signal extending the full length of the trap in the final frame were assigned higher scores. Furthermore, the score was set to zero if a trap contained too many empty frames, the sum of the growth measurement array was negative, or the last frame had a growth measurement of less than 250 pixels.

Require: A An array of vertical growth measurements (in pixels) from a trap at each time-point (frame)
Require: F The vertical extent (in pixels) of fluorescent content after the final frame
Output: $Score$ The data quality score of the trap

1: $Score \leftarrow 0$
2: $Growth \leftarrow [\ ]$
3: **for** $i \leftarrow 2$ to length(A) **do**
4:     $current\_growth \leftarrow A[i] - A[i-1]$
5:     $Growth[i-1] \leftarrow current\_growth$ ▷ Store the growth measurement
6:     $Score \leftarrow Score + F \times \log(\max(current\_growth, 1))$
7: **end for**
8:
9: $n\_empty \leftarrow 0$
10: **for** $i \leftarrow 1$ to length(A) **do**   ▷ Count the number of empty frames
11:     **if** $A[i] < 25$ **then**
12:         $n\_empty \leftarrow n\_empty + 1$
13:     **end if**
14: **end for**
15: **if** $\sum(Growth) < 0$ **or** last$(A) < 250$ **or** $n\_empty > 25$ **then**
16:     $Score \leftarrow 0$ ▷ Set the score to zero if abnormal events occur
17: **end if**

Fig 14. Results using data quality scoring The ResNet 18 and Video ResNet R(2+1)D networks were retrained 30 times using three different approaches: (1) using all data in the training set but utilizing a random weighted sampler, (2) limiting the number of traps per species to 2600 using random selection, (2) limiting the number of traps per species to 2600 but using data quality score to select the traps.

4.6 Segmentation model

The segmentation model used in this paper was an ensemble model consisting of a U-Net [45] and a Segformer [46] model. A pixel was activated only when there was an agreement between both models (logical AND). The segmentation model was used to measure the vertical growth in each trap. Empirically, it was observed that U-Net performed better segmenting fine-grained local edges and textures, and the Segformer had better global awareness, presumably due to the global receptive field of the transformer architecture due to the self-attention mechanism [23].

Prior to using the segmentation model, classical image processing methods, such as histogram equalization [47] and Otsu's thresholding [48], were used to segment bacterial content in a trap. However, these methods proved unreliable, especially in the presence of brightness variations and out-of-focus areas across the image. Then Omnipose [49] was tested with the standard phase-contrast bacteria configuration but did not work, presumably due to the unusual elongated shape of the cropped trap images.

The data training for the ensemble model was obtained by first training a standard U-Net [45] model using segmentation masks published in [16], segmenting bacteria cells on 512x512 pixel crops. This model was then used to generate 50,000 binary segmentation masks from our phase-contrast dataset consisting of 52x1500 pixel crops. These masks were enhanced by using custom logic excluding regions with either too small area or minimal width that do not resemble bacterial cells. This dataset was then used to train the final ensemble model used in this paper.

The fluorescence images were often noisy, and standard background removal methods were frequently insufficient. Therefore, 1,000 segmentation masks of fluorescence images were manually labeled and used to train a segmentation model (the same Segformer U-Net ensemble outlined above) to identify regions with bacterial fluorescence. This fluorescence segmentation model was then integrated into the labeling tool.

4.7 Augmentations

The Albumentations library [36] was used with the augmentations ShiftScaleRotate, RandomBrightnessContrast, the erasings CoarseDropout, GridDropout and PixelDropout, Blur, HorizontalFlip, VerticalFlip, and RandomCrop. Furthermore, a custom-developed "Random video frame erasing" was developed that erased (set all pixels to zero) 1-2 consecutive frames at 1-3 positions in the timelapse. Another custom augmentation method, "Segment", removed the background using pre-generated masks from the segmentation model. It was observed that RandomBrightnessContrast was of high importance, so it was always applied (100% probability). ShiftScaleRotate and horizontal and vertical flips were applied with a 50% probability, while the remaining augmentations were applied with a 10% probability in the augmentation pipeline. Standard parameters were used for each augmentation with only minor adjustments.

4.8 Deep learning model training

The deep learning models were trained using the Pytorch [44] library, and the image models were obtained from the Pytorch Image Models (timm) library [50] and the video classification models from the Torchvision library [51]. The models were trained on images with a spatial size of 52x114 pixels, larger images caused the networks to overfit

for this dataset. The image classification variants were chosen to have 11-12 million ₄₁₀ parameters: ResNet 18, EfficientNet B3, RegNetY 016, and FastVit SA12. The video ₄₁₁ classification models ResNet R(2+1)D and Video ResNet R3D had around 30 million ₄₁₂ parameters. ₄₁₃

The main idea behind ResNet variants is the use of residual connections, which ₄₁₄ mitigates performance degradation as networks become deeper [22]. Video ResNet ₄₁₅ R(2+1)D extends this concept to video processing. The model also decomposes ₄₁₆ standard 3D convolutions into separate 2D spatial convolutions followed by 1D ₄₁₇ temporal convolutions. In contrast, the Video ResNet R3D employs regular 3D ₄₁₈ convolutions to simultaneously detect spatial and temporal features. ₄₁₉

RegNets [30] are a family of models proposed through an automated search, ₄₂₀ optimizing network width and depth to achieve optimal performance with a given ₄₂₁ number of floating point operations (FLOPs) budget for the inference of one sample. ₄₂₂ Similarly, EfficientNets [31] results from a comparable search to find optimal ₄₂₃ architectures. The EfficientNet models also use depth-wise separable convolutions to ₄₂₄ reduce the number of parameters and, thus, computational cost. ₄₂₅

FastViT [32] is a recently developed vision transformer [23] incorporating several ₄₂₆ optimizations to enhance computational efficiency. ₄₂₇

The models were trained in a standard supervised fashion using Adam optimizer [52], ₄₂₈ batch size 32, learning rate 0.0001, and cross-entropy loss function. A label smoothing ₄₂₉ of 0.15 was used to mitigate overfitting, distributing 15% of the probability mass to the ₄₃₀ other classes. This adjusted the one-hot encoded vectors to have a value of 0.85 for the ₄₃₁ true class and 0.025 for the false classes [37]. Weights were initialized by pretrained ₄₃₂ models by default. The image models were pretrained on the ImageNet dataset [53], ₄₃₃ and the video models were pretrained on the Kinetics-400 dataset [54]. ₄₃₄

During the image model training, a random non-empty frame was sampled from a ₄₃₅ trap. The trap was initially vertically cropped according to the vertical growth obtained ₄₃₆ from the segmentation model. This procedure ensured that the subsequent 52x114 pixel ₄₃₇ RandomCrop always contained bacterial content, avoiding empty empty sections of the ₄₃₈ trap. The video models were trained such that each sample was a 10-frame video clip ₄₃₉ randomly sampled from the timelapse. The trap was similarly cropped vertically to ₄₄₀ ensure that at least 75% of the video frames were non-empty when extracting a video ₄₄₁

clip from the trap using RandomCrop.

The trap had an original width of 52 pixels and a height of 1500 pixels. During testing, a number of 52x114 pixel crops (image or video clips) were extracted from the traps in a tiling window arrangement, as seen in S1 and S2 Figs. The final output was obtained by summing the output logits from each crop. Initially, the output aggregation was performed by voting, but this did not improve performance. Furthermore, summing the logits simplified the Receiver Operating Characteristics (ROC) computations.

Time testing evaluations (Figs 3 and 7) were conducted by including all frames up to the target frame and then running testing. Video classification models can handle variable-length video inputs by applying global average pooling across the time dimension before the final fully connected layer [55]. The image models were tested by including crops from all frames leading up to the target frame.

The video classification networks were trained for 150 epochs, and the image classification networks for 300 epochs. Nvidia A100 GPU:s were used for the training using MIG reservation for the less demanding subsampled models. Training times ranged from 1-2 hours for low-resolution ResNet 18 models to around 10 hours for the Video ResNet R(2+1)D at full resolution 52x114 pixels. Inference times were 0.00744 seconds per sample for Video ResNet R(2+1)D at full resolution and 6.351e-5 seconds per sample for ResNet 18 at full resolution, using a batch size of 32 on the A100 GPU.

The models were retrained five times for the scaling experiments, and 30 times for the augmentation ablation, and texture, morphology and division pattern experiments. Each retraining was conducted using a predetermined random seed for reproducibility.

4.9   Metrics

The precision and recall were calculated for each species i, using the confusion matrix (evaluated at argmax).

$$\text{Precision}_i = \frac{\text{TP}_i}{\text{TP}_i + \text{FP}_i} \quad i \in C \tag{1}$$

$$\text{Recall}_i = \frac{\text{TP}_i}{\text{TP}_i + \text{FN}_i} \quad i \in C \tag{2}$$

Where TPi and FPi, and FNi are the number of true positives, false positives and false negatives for the respective species i, C are the set of classes.

Then, Receiver Operating Characteristics (ROC) curves and the corresponding Area Under the Curve (AUC) were computed using varying thresholds. AUC was selected as the evaluation metric for all experiments in this paper, as it is threshold-independent and resilient to class imbalance. Also, comparison becomes easier when using only one metric for all tests. The AUC was obtained using the Scikit-Learn library [56] and calculated for each species in a one-vs-rest fashion, treating it as a binary classification problem. The average AUC reported used the micro-averaged ROC curve obtained using the following formula, evaluated at each threshold:

$$\text{TPR} = \frac{\sum_i \text{TP}_i}{\sum_i (\text{TP}_i + \text{FN}_i)}, \quad i \in C \tag{3}$$

$$\text{FPR} = \frac{\sum_i \text{FP}_i}{\sum_i (\text{FP}_i + \text{TN}_i)} \quad i \in C \tag{4}$$

4.10   Spatial subsampling

The models were trained and tested using subsampled images, simulating using lower-resolution microscopy in a possible clinical setting. The Lanczos method was used for subsampling, considered the most physically accurate subsampling method [35]. In total, 28 subsampling steps were used: 52x114, 48x105, 44x96, 40x88, 36x79, 32x70, 28x61, 26x57, 24x53, 22x48, 20x44, 19x42, 18x39, 17x37, 16x35, 15x33, 14x31, 13x28, 12x26, 11x24, 10x22, 9x20, 8x18, 7x15, 6x13, 5x11, 4x9 and 3x7 pixels.

4.11   Use of artificial intelligence tools and technologies

The tools Grammarly and ChatGPT were used for grammar checking, as a synonym book, and as rephrasing tools when writing this article. No original content was generated by the models. Furthermore, the coding assistant GitHub Copilot was used to generate type annotations and docstrings for the software in the released replication package.

References

1. Monosson E, Monosson E. Discovery: Antibiotics and the Rise of the Superbug. Unnatural Selection: How We Are Changing Life, Gene by Gene. 2015; p. 9–31.

2. Leekha S, Terrell CL, Edson RS. General principles of antimicrobial therapy. In: Mayo clinic proceedings. vol. 86. Elsevier; 2011. p. 156–167.

3. Murray CJL, Ikuta KS, Sharara F, Swetschinski L, Robles Aguilar G, Gray A, et al. Global burden of bacterial antimicrobial resistance in 2019: a systematic analysis. The Lancet. 2022;399(10325):629–655. doi:https://doi.org/10.1016/S0140-6736(21)02724-0.

4. Makary MA, Kaczmarski K, Nachman K. A call for doctors to recommend antibiotic-free foods: agricultural antibiotics and the public health crisis of antimicrobial resistance. The Journal of Antibiotics. 2018;71(8):685–687.

5. Nadimpalli ML, Chan CW, Doron S. Antibiotic resistance: a call to action to prevent the next epidemic of inequality. Nature medicine. 2021;27(2):187–188.

6. Larsson D, Flach CF. Antibiotic resistance in the environment. Nature Reviews Microbiology. 2022;20(5):257–269.

7. Hofer U. The cost of antimicrobial resistance. Nature Reviews Microbiology. 2019;17(1):3–3.

8. Malik B, Bhattacharyya S. Antibiotic drug-resistance as a complex system driven by socio-economic growth and antibiotic misuse. Scientific reports. 2019;9(1):9788.

9. Holmes AH, Moore LS, Sundsfjord A, Steinbakk M, Regmi S, Karkey A, et al. Understanding the mechanisms and drivers of antimicrobial resistance. The Lancet. 2016;387(10014):176–187.

10. Liu VX, Fielding-Singh V, Greene JD, Baker JM, Iwashyna TJ, Bhattacharya J, et al. The timing of early antibiotics and hospital mortality in sepsis. American journal of respiratory and critical care medicine. 2017;196(7):856–863.

11. Claydon MA, Davey SN, Edwards-Jones V, Gordon DB. The rapid identification of intact microorganisms using mass spectrometry. Nature biotechnology. 1996;14(11):1584–1586.

12. Wang H, Ceylan Koydemir H, Qiu Y, Bai B, Zhang Y, Jin Y, et al. Early detection and classification of live bacteria using time-lapse coherent imaging and deep learning. Light: Science & Applications. 2020;9(1):118.

13. Paquin P, Durmort C, Paulus C, Vernet T, Marcoux PR, Morales S. Spatio-temporal based deep learning for rapid detection and identification of bacterial colonies through lens-free microscopy time-lapses. PLOS Digital Health. 2022;1(10):e0000122.

14. Espagnon I, Ostrovskii D, Mathey R, Dupoy M, Joly PL, Novelli-Rousseau A, et al. Direct identification of clinically relevant bacterial and yeast microcolonies and macrocolonies on solid culture media by Raman spectroscopy. Journal of biomedical optics. 2014;19(2):027004–027004.

15. Ho CS, Jean N, Hogan CA, Blackmon L, Jeffrey SS, Holodniy M, et al. Rapid identification of pathogenic bacteria using Raman spectroscopy and deep learning. Nature communications. 2019;10(1):1–8.

16. Kandavalli V, Karempudi P, Larsson J, Elf J. Rapid antibiotic susceptibility testing and species identification for mixed infections. bioRxiv. 2021; p. 2021–11.

17. Wang P, Robert L, Pelletier J, Dang WL, Taddei F, Wright A, et al. Robust growth of Escherichia coli. Current biology. 2010;20(12):1099–1103.

18. Baltekin Ö, Boucharin A, Tano E, Andersson DI, Elf J. Antibiotic susceptibility testing in less than 30 min using direct single-cell imaging. Proceedings of the National Academy of Sciences. 2017;114(34):9170–9175.

19. Zernike F. Phase contrast, a new method for the microscopic observation of transparent objects part II. Physica. 1942;9(10):974–986.

20. Hallström E, Kandavalli V, Ranefall P, Elf J, Wählby C. Label-free deep learning-based species classification of bacteria imaged by phase-contrast microscopy. PLoS Computational Biology. 2023;19(11):e1011181.

21. LeCun Y, Bengio Y, Hinton G. Deep learning. nature. 2015;521(7553):436–444. ₅₄₆

22. He K, Zhang X, Ren S, Sun J. Deep residual learning for image recognition. In: ₅₄₇ Proceedings of the IEEE conference on computer vision and pattern recognition; ₅₄₈ 2016. p. 770–778. ₅₄₉

23. Dosovitskiy A, Beyer L, Kolesnikov A, Weissenborn D, Zhai X, Unterthiner T, ₅₅₀ et al. An Image is Worth 16x16 Words: Transformers for Image Recognition at ₅₅₁ Scale. In: International Conference on Learning Representations; 2021.Available ₅₅₂ from: `https://openreview.net/forum?id=YicbFdNTTy`. ₅₅₃

24. Vaswani A, Shazeer N, Parmar N, Uszkoreit J, Jones L, Gomez AN, et al. ₅₅₄ Attention is All you Need. In: Guyon I, Luxburg UV, Bengio S, Wallach H, ₅₅₅ Fergus R, Vishwanathan S, et al., editors. Advances in Neural Information ₅₅₆ Processing Systems. vol. 30. Curran Associates, Inc.; 2017.Available from: ₅₅₇ `https://proceedings.neurips.cc/paper_files/paper/2017/file/` ₅₅₈ `3f5ee243547dee91fbd053c1c4a845aa-Paper.pdf`. ₅₅₉

25. Brown T, Mann B, Ryder N, Subbiah M, Kaplan JD, Dhariwal P, et al. ₅₆₀ Language Models are Few-Shot Learners. In: Larochelle H, Ranzato M, Hadsell ₅₆₁ R, Balcan MF, Lin H, editors. Advances in Neural Information Processing ₅₆₂ Systems. vol. 33. Curran Associates, Inc.; 2020. p. 1877–1901. Available from: ₅₆₃ `https://proceedings.neurips.cc/paper_files/paper/2020/file/` ₅₆₄ `1457c0d6bfcb4967418bfb8ac142f64a-Paper.pdf`. ₅₆₅

26. Tran D, Wang H, Torresani L, Ray J, LeCun Y, Paluri M. A closer look at ₅₆₆ spatiotemporal convolutions for action recognition. In: Proceedings of the IEEE ₅₆₇ conference on Computer Vision and Pattern Recognition; 2018. p. 6450–6459. ₅₆₈

27. Radford A, Kim JW, Xu T, Brockman G, McLeavey C, Sutskever I. Robust ₅₆₉ speech recognition via large-scale weak supervision. In: International conference ₅₇₀ on machine learning. PMLR; 2023. p. 28492–28518. ₅₇₁

28. LeCun Y, Bottou L, Bengio Y, Haffner P. Gradient-based learning applied to ₅₇₂ document recognition. Proceedings of the IEEE. 1998;86(11):2278–2324. ₅₇₃

29. Fukushima K. Neocognitron: A self-organizing neural network model for a mechanism of pattern recognition unaffected by shift in position. Biological cybernetics. 1980;36(4):193–202.

30. Radosavovic I, Kosaraju RP, Girshick R, He K, Dollár P. Designing network design spaces. In: Proceedings of the IEEE/CVF conference on computer vision and pattern recognition; 2020. p. 10428–10436.

31. Tan M. Efficientnet: Rethinking model scaling for convolutional neural networks. arXiv preprint arXiv:190511946. 2019;.

32. Vasu PKA, Gabriel J, Zhu J, Tuzel O, Ranjan A. FastViT: A fast hybrid vision transformer using structural reparameterization. In: Proceedings of the IEEE/CVF International Conference on Computer Vision; 2023. p. 5785–5795.

33. Waskom ML. Seaborn: statistical data visualization. Journal of Open Source Software. 2021;6(60):3021.

34. Garnier, Simon, Ross, Noam, Rudis, Robert, et al.. viridis(Lite) - Colorblind-Friendly Color Maps for R; 2024. Available from: https://sjmgarnier.github.io/viridis/.

35. Mazzoli F. Lánczos interpolation explained; 2023. Available from: https://doi.org/10.5281/zenodo.7761401.

36. Buslaev A, Iglovikov VI, Khvedchenya E, Parinov A, Druzhinin M, Kalinin AA. Albumentations: Fast and Flexible Image Augmentations. Information. 2020;11(2). doi:10.3390/info11020125.

37. Szegedy C, Vanhoucke V, Ioffe S, Shlens J, Wojna Z. Rethinking the inception architecture for computer vision. In: Proceedings of the IEEE conference on computer vision and pattern recognition; 2016. p. 2818–2826.

38. Hallström E. Phase-contrast time-lapses of seven bacterial species growing in microfluidic mother machine traps.; 2024. Available from: https://doi.org/10.5281/zenodo.13321089.

39. Harris CR, Millman KJ, van der Walt SJ, Gommers R, Virtanen P, Cournapeau D, et al. Array programming with NumPy. Nature. 2020;585(7825):357–362. doi:10.1038/s41586-020-2649-2.

40. pandas development team T. pandas-dev/pandas: Pandas; 2020. Available from: https://doi.org/10.5281/zenodo.3509134.

41. Hunter JD. Matplotlib: A 2D graphics environment. Computing in Science & Engineering. 2007;9(3):90–95. doi:10.1109/MCSE.2007.55.

42. Thevenaz P, Ruttimann UE, Unser M. A pyramid approach to subpixel registration based on intensity. IEEE transactions on image processing. 1998;7(1):27–41.

43. Rolnick D, Veit A, Belongie S, Shavit N. Deep Learning is Robust to Massive Label Noise; 2018. Available from: https://arxiv.org/abs/1705.10694.

44. Paszke A, Gross S, Massa F, Lerer A, Bradbury J, Chanan G, et al. PyTorch: An Imperative Style, High-Performance Deep Learning Library. In: Wallach H, Larochelle H, Beygelzimer A, d'Alché-Buc F, Fox E, Garnett R, editors. Advances in Neural Information Processing Systems. vol. 32. Curran Associates, Inc.; 2019.Available from: https://proceedings.neurips.cc/paper_files/paper/2019/file/bdbca288fee7f92f2bfa9f7012727740-Paper.pdf.

45. Ronneberger O, Fischer P, Brox T. U-net: Convolutional networks for biomedical image segmentation. In: Medical image computing and computer-assisted intervention–MICCAI 2015: 18th international conference, Munich, Germany, October 5-9, 2015, proceedings, part III 18. Springer; 2015. p. 234–241.

46. Xie E, Wang W, Yu Z, Anandkumar A, Alvarez JM, Luo P. SegFormer: Simple and efficient design for semantic segmentation with transformers. Advances in neural information processing systems. 2021;34:12077–12090.

47. Gonzalez RC, Woods RE. Digital image processing. Upper Saddle River, N.J.: Prentice Hall; 2008. Available from: http://www.amazon.com/Digital-Image-Processing-3rd-Edition/dp/013168728X.

48. Otsu N. A Threshold Selection Method from Gray-Level Histograms. IEEE Transactions on Systems, Man, and Cybernetics. 1979;9(1):62–66. doi:10.1109/TSMC.1979.4310076.

49. Cutler KJ, Stringer C, Lo TW, Rappez L, Stroustrup N, Brook Peterson S, et al. Omnipose: a high-precision morphology-independent solution for bacterial cell segmentation. Nature methods. 2022;19(11):1438–1448.

50. Wightman R. PyTorch Image Models; 2019. https://github.com/rwightman/pytorch-image-models.

51. maintainers T, contributors. TorchVision: PyTorch's Computer Vision library; 2016. https://github.com/pytorch/vision.

52. Kingma DP, Ba J. Adam: A method for stochastic optimization. arXiv preprint arXiv:14126980. 2014;.

53. Deng J, Dong W, Socher R, Li LJ, Li K, Fei-Fei L. Imagenet: A large-scale hierarchical image database. In: 2009 IEEE conference on computer vision and pattern recognition. Ieee; 2009. p. 248–255.

54. Kay W, Carreira J, Simonyan K, Zhang B, Hillier C, Vijayanarasimhan S, et al. The kinetics human action video dataset. arXiv preprint arXiv:170506950. 2017;.

55. Lin M. Network in network. arXiv preprint arXiv:13124400. 2013;.

56. Pedregosa F, Varoquaux G, Gramfort A, Michel V, Thirion B, Grisel O, et al. Scikit-learn: Machine Learning in Python. Journal of Machine Learning Research. 2011;12:2825–2830.

5 Supporting information

S1 Appendix. Testset adjustments and data cleaning The document shows all traps that either had the label manually changed or were removed from the test set after processing of the labeling tool.

5.1 Crop extraction

The following figures show crop extraction from a time-lapse.

S1 Fig.   Crops extracted from a trap for image classification. The vertical growth extent of the bacteria in each trap was measured using the segmentation model, indicated by the horizontal markings on the tip of the lowest cell in each trap. Crops were extracted in a tiling-window fashion, shown by outlined yellow squares. In total, 202 52x114 pixel crops were extracted from this trap. The criteria for including a crop is that at least 25% of its vertical extent contains bacteria.

S2 Fig.   Crops extracted from a trap for video classification The vertical

growth extent of the bacteria in each trap was measured using the segmentation model,

indicated by the horizontal markings on the tip of the lowest cell in each trap. Each

video crop is outlined with a yellow square. The vertical dotted lines between adjacent

frames indicate they are part of the same video crop. In total, five 52x114x35 pixel

video crops were extracted from this trap. The criteria for including a video crop is that

at least 75% of the frames contain bacterial content.

5.2 Misclassifications

S3 Fig. Trap misclassified by Video ResNet R(2+1)D at full resolution. *E. coli* and *P. mirabilis* are visually very similar in shape, both being rods. There is some overcrowding in the trap with overlapping cells.

S4 Fig.   Trap misclassified by Video ResNet R(2+1)D at full resolution *E. coli* and *K. pneumoniae* are visually very similar in shape, both being rods. There is some overcrowding in the trap with overlapping cells at the end of the time-lapse.

S5 Fig. Trap misclassified by Video ResNet R(2+1)D at full resolution K. penumaniae and *E. coli* are visually very similar in shape, both being rods. There is some dislocation in the trap at the end of the time-lapse, causing one of the video-crops to have empty frames.

S6 Fig. **Trap misclassified by Video ResNet R(2+1)D at full resolution** *S. aureus* and *E. faecalis* are visually very similar in shape, both being cocci. There is a dislocation in the top, one of the video crops has only empty frames. This trap should have been discarded according to the discarding criteria (no dislocation from the top of the trap), but it was missed during the test set inspection.

S7 Fig. **Trap misclassified by Video ResNet R(2+1)D at full resolution** The
model mistakes the circular physical stop at the top of the trap for *E. faecalis*. This
trap should have been discarded according to the discarding criteria since the trap is
empty up to frame 15 (cells need to be loaded within 30 minutes), but it was missed
during the test set inspection.

S8 Fig. Trap misclassified by Video ResNet R(2+1)D at full resolution A rod cell is present at the beginning of the time-lapse, which then dissolves and is not visible in the fluorescence images.

S9 Fig.   Trap misclassified by Video ResNet R(2+1)D at subsampled 5x11 resolution *P. mirabilis* and *E. coli* are visually very similar in shape, both being rods. Classifying at very low resolution is significantly more challenging.

S10 Fig. Trap misclassified by Video ResNet R(2+1)D at subsampled 5x11 resolution *K. pneumoniae* and *E. coli* are visually very similar in shape, both being rods. Classifying at very low resolution is significantly more challenging.

S11 Fig. Trap misclassified by Video ResNet R(2+1)D at subsampled 5x11 resolution *A. baumannii* and *E. coli* are visually very similar in shape, both being rods. Classifying at very low resolution is significantly more challenging.

S12 Fig. Trap misclassified by Video ResNet R(2+1)D at subsampled 5x11 resolution This trap was also misclassified by Video ResNet in the full resolution. *E. coli* and *P. aeruginosa* are visually very similar in shape, both being rods. Classifying at very low resolution is significantly more challenging.

S13 Fig. Trap misclassified by Video ResNet R(2+1)D at subsampled 5x11 resolution This trap was also misclassified by Video ResNet in the full resolution. *S. aureus* and *E. faecalis* are visually very similar in shape, both being cocci. There is a dislocation in the top, one of the video crops has only empty frames. This trap should have been discarded according to the discarding criteria (no dislocation from the top of the trap), but it was missed during the test set inspection. Classifying at very low resolution is significantly more challenging.

S14 Fig. Trap misclassified by Video ResNet R(2+1)D at subsampled 5x11 resolution *P. aeruginosa* and *A. baumannii* are visually very similar in shape, both being rods. This trap should have been discarded according to the discarding criteria as it is empty until frame 25 (cells must be loaded within 30 minutes), but it was missed during the test set inspection. Classifying at very low resolution is significantly more challenging.

5.3 Combination staining issues

The following figures show examples of challenges in assigning labels to traps due to irregularities in the staining process.

S15 Fig. **K. pneumonie** For K. pneumonie the cells are visible in Cy3 and Cy5 in the fluorescent staining

S16 Fig. K. pneumonie The cells should be visible in Cy3 and Cy5, but there was almost no absorption of Cy5 fluorophores. The trap originated from the same experiment and chip section as S15 Fig, which contained only *K. pneumoniae*, allowing the labeling tool to infer the correct label.

S17 Fig. ***P. aeruginosa*** The species should be visible in Alexa488 and TxR, but there was no absorption of Alexa488 fluorophores. It can not be *S. aureus* (Cy5, TxR) being a cocci; hence, the label was manually assigned.

S18 Fig. *S. aureus* The species should be visible in Cy5 and TxR, but there was no absorption of Cy5 fluorophores. The chip section contained only *S. aureus*, allowing the labeling tool to infer the correct label.

5.4 Videos

The following video clips show the same trap with all the subsampling steps used in the resolution scaling experiments. The number in the top row indicates the pixel width.

S1 Video. *P. aeruginosa*

S2 Video. *E. coli*

S3 Video. *K. pneumoniae*

S4 Video. *A. baumannii*

S5 Video. *E. faecalis*

S6 Video. *P. mirabilis*

S7 Video. *S. aureus*

The following are other supporting videos for the article.

S8 Video. **Image distortion modes** The two image distortion modes "Mean" and "Mask" are visualized for a single trap on the left, and the default phase-contrast images on the right.

S9 Video. **Microfluidic Mother Machine** Raw unstabilized phase-contrast microscopy timelapse of a chip position in the microfluidic mother machine followed by fluorescence microscopy capturing four stains, Alexa488, Cy3, Cy5, and Texas Red (TxR). The single-channel raw grayscale images are shown here (no color map applied).