QUANTIFICATION OF THE EFFECT OF HEMODYNAMIC OCCLUSION IN TWO-PHOTON IMAGING

Baba Yogesh[1,2], Matthias Heindorf[1], Rebecca Jordan[3], & Georg B. Keller[1,2,4]

[1]*Friedrich Miescher Institute for Biomedical Research, Basel, Switzerland.*
[2]*Faculty of Natural Sciences, University of Basel, Basel, Switzerland.*
[3]*Simons Initiative for the Developing Brain, University of Edinburgh, 1 George Square, EH8 9JZ, Edinburgh, UK.*
[4]*Correspondence to: georg.keller@fmi.ch*

ABSTRACT

The last few years have seen an explosion in the number of tools available to measure neuronal activity using fluorescence imaging (Chen et al., 2013; Feng et al., 2019; Jing et al., 2019; Sun et al., 2018; Wan et al., 2021). When performed in vivo, these measurements are invariably contaminated by hemodynamic occlusion artifacts. In widefield calcium imaging, this problem is well recognized. For two-photon imaging, however, the effects of hemodynamic occlusion have only been sparsely characterized. Here we perform a quantification of hemodynamic occlusion effects using measurements of fluorescence changes observed with GFP expression using both widefield and two-photon imaging. We find that in many instances the magnitude of signal changes attributable to hemodynamic occlusion is comparable to that observed with activity sensors. Moreover, we find that hemodynamic occlusion effects were spatially heterogeneous, both over cortical regions and across cortical depth, and exhibited a complex relationship with behavior. Thus, hemodynamic occlusion is an important caveat to consider when analyzing and interpreting not just widefield but also two-photon imaging data.

INTRODUCTION

Optical imaging of neuronal activity and neuromodulator concentration often involves recording the changes in fluorescence intensity of calcium or neuromodulator sensors expressed in neurons. In vivo, the hemodynamic changes in tissue around at the imaging location can modulate light transmission. Changes in blood volume through vasodilation and vasoconstriction as well as hemoglobin oxygenation change the transmission properties of the tissue. This is referred to as hemodynamic occlusion. Often, these

hemodynamic changes are driven by local neuronal activity, a phenomenon known as neurovascular

coupling (Attwell et al., 2010; Iadecola and Nedergaard, 2007). Hemodynamic changes are known to

contribute significantly to activity measurements in techniques like fiber photometry (Zhang et al., 2022)

and widefield microscopy (Ma et al., 2016; Waters, 2020). For these techniques, several methods have

been proposed to control for the hemodynamic contribution to these activity measurements (Lohani et

al., 2022; Valley et al., 2020; Zhang et al., 2022). However, the potential contribution of hemodynamic

signals to activity measurements acquired using two-photon imaging remains largely unexplored.

Here, we characterized hemodynamic signals in two-photon imaging of mouse cortex. We expressed an

activity independent marker (GFP) in mouse cortical neurons and measured fluorescence changes using

two-photon imaging while mice were interacting with a virtual environment. We found changes in GFP

fluorescence in response to locomotion and visual stimuli of a magnitude comparable to that measured

with common activity sensors like GCaMP and GRAB variants. GFP signals were apparent both in

population average responses and at individual neuron level. Moreover, these signals were

heterogeneous over dorsal cortex and cortical layers. We compared the size of GFP signals to the calcium

responses reported by GCaMP6 sensors, and to the neuromodulator signals reported by the GRAB sensors

for dopamine (GRAB-DA1m), serotonin (GRAB-5HT1.0), acetylcholine (GRAB-ACh3.0), and norepinephrine

(GRAB-NE1m). We found that a large fraction of the GRAB response to the different neuromodulators was

explained by the GFP signal, consistent with an earlier study characterizing this relationship for GRAB-

5HT1.0 and GRAB-5HT3.0 (Ocana-Santero et al., 2024). Based on our findings, we speculate that the

primary driver of the hemodynamic signals in two-photon imaging is hemodynamic occlusion. If so, then

this problem is likely primarily mitigated by using high dynamic range sensors with low baseline

fluorescence, and best controlled for using activity independent fluorescence measurements.

RESULTS

To quantify hemodynamic occlusion in two-photon microscopy we imaged GFP in mouse cortex during

behavior. We used an AAV vector (AAV2/1-Ef1α-eGFP-WPRE) injected in primary visual cortex (V1) and

anterior cingulate cortex (ACC) to express GFP in cortical neurons. Given that the Ef1α promoter biases

expression to excitatory neurons, we estimate that most (95%) of the GFP-expressing neurons were

excitatory (Attinger et al., 2017). We then used a two-photon microscope to record GFP signals while mice

explored a virtual tunnel. Mice were head-fixed on a spherical treadmill surrounded by a toroidal screen

and exposed to a set of visuomotor conditions known to activate neurons in V1 and ACC (Attinger et al.,

2017; Keller et al., 2012; Leinweber et al., 2017; Zmarz and Keller, 2016). First, mice were exposed to a

closed loop condition during which running speed was coupled to visual flow speed in a virtual tunnel.

Mice were also exposed to an open loop condition, during which locomotion and visual flow in the virtual

tunnel were not coupled, to darkness, and to the presentation of full-field drifting gratings. Throughout

all experimental conditions, mice were free to locomote on the spherical treadmill and did so voluntarily.

Unless stated otherwise, we will use 'response' to refer to the observed change in GFP fluorescence.

Locomotion onset and visual stimuli resulted in strong changes in GFP population responses

At the onset of locomotion, we found a transient increase in apparent GFP fluorescence in the population

response of layer 2/3 (L2/3) neurons in V1 (**Figure 1A**). This increase in fluorescence of approximately 1%

$\Delta F/F_0$ was only marginally smaller than the fluorescence changes typically observed at locomotion onset

with genetically encoded calcium indicators (approximately 2% $\Delta F/F_0$ for GCaMP6f (Widmer et al., 2022;

Yogesh and Keller, 2023), approximately 1.5% $\Delta F/F_0$ for GCaMP3 (Keller et al., 2012), etc.). By contrast,

presentation of full-field drifting gratings resulted in a decrease of the GFP population signal (**Figure 1B**).

We observed a similar decrease in GFP population signal on the presentation of an optogenetic

stimulation light (637 nm, 10mW after the objective) directed at V1 with either the virtual reality turned

on or in the dark (**Figure 1C**). This response was likely visually driven as the decrease was twice as strong

in dark than when the virtual reality was on, in line with the stimulation light being more visually salient

in dark. Visuomotor mismatches (pauses in the visual flow feedback during locomotion), resulted in an

average increase of apparent GFP fluorescence (**Figure 1D**). By design, visuomotor mismatch only occurs

during times of locomotion. Because apparent GFP fluorescence increases during locomotion, we

estimated the increase expected by chance using random triggers during locomotion to compute the 95%

confidence interval in **Figures 1D, 1H, and 1L**. GFP responses to mismatch were only briefly outside these

95% confidence intervals.

To test whether stimulus triggered GFP responses depend on imaging depth and cortical area, we

repeated these experiments in layer 5 (L5) of V1 and in L2/3 of ACC. In L5, we found that apparent

fluorescence changes at locomotion onset were smaller than those observed in L2/3 (**Figure 1E**). As in

L2/3, full-field drifting grating stimuli (**Figure 1F**) and optogenetic stimulation light (**Figure 1G**) both

resulted in a significant decrease in apparent GFP fluorescence, while visuomotor mismatch responses

were only barely above chance (**Figure 1H**). In L2/3 of ACC, locomotion onset resulted in a strong increase

of apparent GFP fluorescence (**Figure 1I**), while we found only small grating responses (**Figure 1J**) and a

positive response to optogenetic stimulation light (**Figure 1K**). We again found no evidence of visuomotor

mismatch responses (**Figure 1L**). Thus, apparent GFP fluorescence changes can vary both as a function of

recording depth and cortical area.

The observation that average population responses of GFP and calcium indicators are similar does not

mean that the dynamic range of individual neurons is similar. When comparing the distribution of peak

$\Delta F/F_0$ responses, it is evident that peak responses are much larger for calcium indicators (**Figures 2A and**

2B). This is still the case when comparing trial averaged responses (**Figures 2C-2E**), but the difference

between GFP and calcium indicator response distributions is smaller, in particular for locomotion onsets

(**Figure 2C**). The main driver for this effect is likely the highly variable stimulus triggered calcium responses

of many neurons.

Individual neurons showed significant GFP responses to different stimuli

Next, we tested whether GFP signals are strong enough to result in significant responses when looking at

individual neurons. Raw GFP traces exhibited fluctuations reminiscent of calcium responses but tended to

be of lower peak amplitude (**Figure 2A**). It is possible that small, correlated changes in GFP signal, when

averaged over the population, result in a significant response indistinguishable from calcium signals

reported with GCaMP but are negligible at the level of individual neurons. Thus, we quantified the fraction

of neurons that are significantly responsive to different stimuli. Commensurate with the population

averages (**Figure 1**), we found a surprisingly large fraction of neurons in V1, in both L2/3 (**Figure 3A**) and

L5 (**Figure 3B**) responsive to locomotion onset and grating stimuli. In both locations, the fraction of

neurons responsive to visuomotor mismatch was not different from chance. In ACC L2/3 (**Figure 3C**), a

significant fraction of GFP labelled neurons was responsive to locomotion onset, while the fraction of

neurons responsive to gratings and visuomotor mismatch was not different from chance. Thus, even at

the single neuron level, hemodynamic influence on GFP remains measurable.

Changes in GFP fluorescence correlated with blood vessel dilation

We speculated that the dominant change in GFP fluorescence is driven by hemodynamic occlusion. As

blood vessels dilate and constrict, they influence the transmission of both excitation and emission light.

One testable prediction of this hypothesis is that the cross section of blood vessels visible in a given field

of view should correlate inversely with the apparent GFP fluorescence of surrounding cells. This makes

the simplifying assumption that the cross section of surface blood vessels, which likely have the strongest

contribution to occlusion, correlates with that of blood vessels in the imaging plane. To quantify the

correlation between blood vessel cross section and apparent GFP fluorescence, we first estimated the

cross section of selected blood vessels in the imaging plane (**Figures 4A and 4B**). Blood vessels in V1 dilated

systematically, for example upon light stimulation (**Figures 4A1 and 4B1**). The estimation of blood vessel

cross section worked well for high contrast boundaries but failed in many instances of low contrast

boundaries due to noise in the boundary estimates in single imaging frames. A proxy for blood vessel cross

section that does not rely on thresholding is the average fluorescence of a region of interest of fixed size

and fully contained within the blood vessel (**Figures 4A2 and 4B2**). Comparing both metrics to the average

apparent GFP fluorescence changes of the surrounding cells (**Figures 4A3 and 4B3**), we found that both

blood vessel area and fluorescence explain approximately 80% of the variance of apparent GFP

fluorescence changes (**Figures 4A4 and 4B4**) in these example data. Quantifying this for all data, we found

that the fraction of variance explained was 50% or higher for most event triggered averages (**Figure 4C**)

as well as across the entire recording duration (**Figures 4D and S2**). The variance explained tended to be

lower in ACC than in V1; we suspect that this is caused by the superior sagittal sinus contributing more

strongly to hemodynamic occlusion in ACC, and that the diameters of local blood vessels are less well

correlated with that of this large blood vessel.

GFP responses were visuomotor context sensitive and could not be explained by a linear combination of

responses to constituent stimuli

Next, we investigated how stereotypical GFP responses are across different visuomotor conditions and

cortical regions, and whether GFP responses can be explained by a linear combination of a locomotion

driven component and a visually driven component. We found that in V1 L2/3, locomotion onset

responses were similar in closed loop and open loop, but different in the dark (**Figure 5A**). During a closed

loop locomotion onset, there is concurrent onset of visual flow with the locomotion onset. During open

loop and dark locomotion onsets there is no concurrent onset of visual flow, and they only differ in the

average visual input. The similarity between open and closed loop running onsets could be explained by

a dominant locomotion related signal and a negligible visual onset response. If so, the smaller GFP

responses in dark locomotion would need to be the consequence of the lower average visual input. This

demonstrates that GFP responses, similar to neuronal calcium responses, can be non-linear combinations

of component responses. This is perhaps not surprising, assuming local blood flow is primarily related to

local neuronal activity. In L5 of V1, locomotion onset responses in open loop and dark conditions were

comparable, while those in closed loop conditions had a delayed transient dip (**Figure 5B**).   Thus,

hemodynamic responses can exhibit a strong depth dependence. In ACC L2/3 neurons all three types of

locomotion onsets resulted in increases of apparent fluorescence (**Figure 5C**). Thus, hemodynamic

responses are difficult to predict from component signals (**Figure S1**), and do not generalize across depth

or cortical areas. Finally, while there appears to be some correspondence to local neuronal activity in that

locomotion onsets are more strongly modulated by visual context in V1 than in ACC, it is not immediately

obvious how the GFP signals relate to responses measured with calcium indicators.

Locomotion increased correlation in GFP signals

Another metric frequently used in the analysis of two-photon data are correlation-based analyses.

Locomotion, for example, has been shown to decorrelate the activity of excitatory neurons in cortex

(Aydın et al., 2018; Dadarlat and Stryker, 2017; Erisken et al., 2014; Yogesh and Keller, 2023). Given the

sizable fraction of neurons significantly responsive to locomotion, we speculated that correlations

between neurons likely are influenced by hemodynamic responses. We tested this by quantifying the

effect of locomotion on the pairwise correlations of GFP signals between individual neurons. We found

that pairwise correlations were systematically increased during locomotion (**Figure 6**). This was true for

neurons in V1 L2/3 (**Figure 6A**), V1 L5 (**Figure 6B**), and ACC L2/3 (**Figure 6C**). Thus, hemodynamic responses

can have a strong influence on pairwise correlations, and in the case of locomotion, this effect is opposite

to that observed with calcium indicators (Aydın et al., 2018; Dadarlat and Stryker, 2017; Erisken et al.,

2014; Yogesh and Keller, 2023). This is consistent with the interpretation that hemodynamic changes on

locomotion occur at a larger spatial scale than changes in neuronal activity patterns and thus tend to

increase correlations between neurons.

GFP responses in widefield imaging were similar in magnitude to those in two-photon imaging

In widefield calcium imaging, hemodynamic responses are better characterized and widely recognized as

a confounding problem (Allen et al., 2017; Ma et al., 2016; Valley et al., 2020). To directly contrast

hemodynamic widefield signals with those observed in two-photon imaging, we repeated our

experiments by performing widefield imaging of GFP responses. We expressed GFP pan-neuronally across

dorsal cortex using either a retro-orbital injection of AAV-PHP.eB-EF1α-eGFP, or transgenic mice that

express GFP under a *c-fos* promoter (**Table S2**). As with two-photon imaging (**Figure 1**), and consistent

with previous work (Allen et al., 2017; Valley et al., 2020), we found strong GFP responses across dorsal

cortex (**Figure 7**). Locomotion and grating onsets resulted in GFP responses in both V1 (**Figures 7A and**

7B) and ACC (**Figures 7D and 7E**). During visuomotor mismatch, we found no significant response in either

of these regions (**Figures 7C and 7F**). All these responses were greatly attenuated or absent when we

imaged the cortex without expressing a fluorescent indicator (**Figure S3**), arguing in favor of hemodynamic

occlusion of fluorescence from the genetically expressed indicators as the underlying cause for most of

these responses in the GFP signal.

GRAB sensors in cortex displayed similar responses to those recorded in GFP imaging

Finally, we compared the magnitude of hemodynamic responses to those recorded with the relatively low

dynamic range GRAB sensors using both two-photon and widefield imaging. For this we used the GRAB-

DA1m dopamine sensor (Sun et al., 2018), the GRAB-5HT1.0 serotonin sensor (Wan et al., 2021), the

GRAB-ACh3.0 acetylcholine sensor (Jing et al., 2019), and the GRAB-NE1m norepinephrine sensor (Feng

et al., 2019). We first imaged GRAB-DA1m, GRAB-5HT1.0, and GRAB-ACh3.0 using two-photon imaging in

L2/3 of V1. To estimate hemodynamic responses from GRAB sensors, we compared responses of regions

of interest (ROI) selected in neuropil and those selected in blood vessels (**Figures 8A, 8E and 8I**). We

reasoned that the signal in neuropil regions is a combination of a true GRAB signal and hemodynamic

occlusion, while the signal in a ROI selected inside of a blood vessel should be dominated by the blood

vessel diameter. We found that the GRAB-DA1m response to locomotion and full-field grating onsets

(**Figure 8B and 8C**) was similar to GFP responses in V1 L2/3 (**Figure 1**). There were responses to locomotion

and grating onset in the GRAB-DA1m signal (**Figure 8B and 8C**). In both cases neuropil and blood vessel

ROI responses were similar. Also, similar to the GFP responses, we found no evidence of a response to

visuomotor mismatch in the GRAB-DA1m signal (**Figure 8D**). We observed similar responses when imaging

GRAB-5HT1.0 in V1 (**Figures 8F-8H**). However, these responses were smaller in amplitude, likely driven by

the overall dimmer fluorescence of the GRAB-5HT1.0 sensor. Thus, it is likely that both GRAB-DA1m and

GRAB-5HT1.0 responses recorded with two-photon imaging are dominated by hemodynamic occlusion.

A response that was distinct from GFP, GRAB-DA1m, and GRAB-5HT1.0 signals was the response in GRAB-

ACh3.0 on locomotion onset (**Figure 8J**). Here we found a prominent increase in GRAB-ACh3.0 signal that

was three-fold higher than either the GFP signal or the GRAB-ACh3.0 signals measured in blood vessel

ROIs. This makes it more likely that the observed fluorescence changes are driven by a mixture of

hemodynamic occlusion and GRAB-ACh3.0 responses. The GRAB-ACh3.0 response to grating onset (**Figure**

8K), however, was similar to GFP responses. And again, we found no response to visuomotor mismatch in

GRAB-ACh3.0 (**Figure 8L**). To test whether the absence of response differences between GRAB-DA1m,

GRAB-5HT1.0 and GFP signals was simply a result of imaging in V1, where both dopamine and serotonin

release is weaker than in frontal areas of cortex (Berger et al., 1991; Hamada et al., 2023), we repeated

these experiments in ACC. Also here, we found no evidence of GRAB responses that could not be explained

by hemodynamic occlusion (**Figure S4**). A strong influence of hemodynamic signals was also apparent in

widefield imaging of GRAB-NE1m. We found that the widefield GRAB-NE1m responses (**Figure 9**) were

similar to GFP responses (**Figure 7**). Thus, in certain cases of low dynamic range sensors, one can readily

measure responses, but these are not always easily distinguished from those driven by hemodynamic

occlusion.

DISCUSSION

Changes in neuronal activity in the cortex are associated with hemodynamic changes through

neurovascular coupling (Ruff et al., 2024) that are layer specific (Mächler et al., 2021). Since blood absorbs

light, hemodynamic occlusion can affect fluorescence intensity measurements. This complicates the

interpretation of signals measured with fluorescent activity indicators as blood flow varies on similar time

scales as those of neuronal activity indicators. We quantified the magnitude of hemodynamic occlusion

by imaging GFP and found the responses to be comparable to those recorded with commonly used

calcium and neuromodulator sensors in population averages. We further found that these signals are

heterogenous across the cortex, present at single neuron level, and non-linearly related to behavioral

events. All this calls for greater caution in interpreting imaging data, particularly when the signal-to-noise

ratio of the sensor is not substantially higher than the hemodynamic occlusion artifact.

Hemodynamic influence on fiber photometry (Zhang et al., 2022) and widefield imaging (Ma et al., 2016;

Scott et al., 2018; Valley et al., 2020) are well documented, and there are correction methods published

(Allen et al., 2017; Ma et al., 2016). But hemodynamic influence on two-photon imaging is relatively

unexplored. Here we show that not only is this influence substantial, but in certain cases it is nearly

indistinguishable from signals acquired with calcium or neuromodulator fluorescent sensors. Both the

excitation and the emission photons need to pass through intervening vascularized tissue between the

objective lens and the imaging plane deep in brain. Thus, changes in blood vessel diameter will influence

both stimulation and emission light. Given that the absorption of light is wavelength dependent, it is not

entirely trivial to estimate hemodynamic effects concurrently with two-photon imaging. There are likely two approaches: estimate the hemodynamic contribution in a separate set of experiments using GFP, or use a second imaging channel for concurrent isosbestic illumination of the sensor. For the calcium indicator GCaMP, isosbestic imaging has been shown to work in widefield imaging (Allen et al., 2017; Couto et al., 2021). However, imaging concurrently with two-photon imaging would require a second two-photon laser, and a way to de-multiplex or spectrally filter the two emission signals that are spectrally very close.

Based on both the correlation between blood vessel diameter and hemodynamic signals as well as the overall strength of the signals, we have here assumed that the primary contribution to the effects we measured is from hemodynamic occlusion. Previous work has shown that imaging neurons singly electroporated with a structural marker (Alexa Fluor 594) in rat visual cortex show a decrease in fluorescence on the presentation of visual stimuli when the neuron is present under an arteriole, but not a venule (Shen et al., 2012). Thus, we suspect the primary contributor to hemodynamic occlusion are alterations in arteriolar diameter.

The similarity in response of many of the GRAB sensors to the GFP signal is a cause for concern and calls for careful distinction between responses driven by the sensor and those resulting from hemodynamic occlusion (Ocana-Santero et al., 2024). In the case of neuromodulator sensors, this problem is further complicated by the fact that norepinephrine and acetylcholine also directly influence the diameter of arterioles. With improvement in the signal-to-noise ratio of these sensors these problems will be reduced. However, given that even for very high dynamic range sensors like GCaMP, the population average responses of hemodynamic occlusion are comparable in magnitude to the sensor signals, it is unlikely that the problem will disappear anytime soon.

 METHODS

 Key Resources Table

Reagent type (species) or resource	Designation	Source or reference	Identifier	Additional information
Strain, strain background (adeno-associated virus)	AAV2/1-Ef1α-eGFP-WPRE (10^{12} GC/ml)	FMI vector core	vector.fmi.ch	
Strain, strain background (adeno-associated virus)	AAV2/1-Ef1α-GCaMP6f-WPRE (10^{11}-10^{14} GC/ml)	FMI vector core	vector.fmi.ch	
Strain, strain background (adeno-associated virus)	AAV2/9-hSyn-GRAB-DA1m (10^{12} GC/ml)	FMI vector core	vector.fmi.ch	
Strain, strain background (adeno-associated virus)	AAV2/9-hSyn-GRAB-5HT1.0 (10^{13} GC/ml)	FMI vector core	vector.fmi.ch	
Strain, strain background (adeno-associated virus)	AAV2/9-hSyn-GRAB-ACh3.0 (10^{13} GC/ml)	FMI vector core	vector.fmi.ch	
Strain, strain background (adeno-associated virus)	AAV-PHP.eB-EF1α-eGFP (10^{13} GC/ml)	FMI vector core	vector.fmi.ch	
Strain, strain background (adeno-associated virus)	AAV-PHP.eB-hSyn-GRAB-NE1m (10^{13} GC/ml)	FMI vector core	vector.fmi.ch	
Chemical compound, drug	Fentanyl citrate	Actavis	CAS 990-73-8	Anesthetic compound
Chemical compound, drug	Midazolam (Dormicum)	Roche	CAS 59467-96-8	Anesthetic compound
Chemical compound, drug	Medetomidine (Domitor)	Orion Pharma	CAS 86347-14-0	Anesthetic compound
Chemical compound, drug	Ropivacaine	Presenius Kabi	CAS 132112-35-7	Analgesic compound
Chemical compound, drug	Lidocaine	Bichsel	CAS 137-58-6	Analgesic compound
Chemical compound, drug	Buprenorphine	Reckitt Benckiser Healthcare	CAS 52485-79-7	Analgesic compound
Chemical compound, drug	Humigel	Virbac	-	Ophthalmic gel
Chemical compound, drug	Flumazenil (Anexate)	Roche	CAS 78755-81-4	Anesthetic antagonist
Chemical compound, drug	Atipamezole (Antisedan)	Orion Pharma	CAS 104054-27-5	Anesthetic antagonist
Chemical compound, drug	Metacam	Boehringer Ingelheim	CAS 71125-39-8	Analgesic compound
Chemical compound, drug	N-Butyl-2-cyanoacrylate	Braun	CAS 6606-65-1	Histoacryl
Chemical compound, drug	Dental cement (Paladur)	Heraeus Kulzer	CAS 9066-86-8	

Genetic reagent *Mus musculus*	*C57BL/6*	Charles River	-	
Genetic reagent *Mus musculus*	*B6J.129S6-Chat^{tm2(Cre)Lowl}/MwarJ* Alias used here: ChAT-IRES-Cre	Jackson Laboratories	RRID:IMSR_JAX:028861	Cre expression in cholinergic neurons
Genetic reagent *Mus musculus*	*B6J.FVB(Cg)-Tg(Tlx3-cre)PL56Gsat/Mmucd* Alias used here: Tlx3-Cre	MMRRC	RRID:MMRRC_041158-UCD	Cre expression in a subset of L5 neurons
Genetic reagent *Mus musculus*	*B6.Cg-Tg(Fos/EGFP)1-3Brth/J* Alias used here: fosGFP	Jackson Laboratories	RRID:IMSR_JAX:014135	Cre expression in a subset of cortical neurons
Software, algorithm	MATLAB (2023b)	The MathWorks	RRID:SCR_001622	Data analysis
Software, algorithm	LabVIEW	National Instruments	RRID:SCR_014325	Hardware control
Software, algorithm	Two-photon acquisition software	Keller laboratory	sourceforge.net/p/iris-scanning/	Data acquisition
Software, algorithm	Image data processing software	Keller laboratory	sourceforge.net/p/iris-scanning/calliope	Data processing
Software, algorithm	Python	python.org	RRID:SCR_008394	Virtual reality
Software, algorithm	Panda3D	panda3d.org	N/A	Virtual reality
Other	Virtual reality and two-photon setup	(Leinweber et al., 2014, 2017)	DOI: 10.3791/50885, 10.1016/j.neuron.2017.08.036	Hardware setup
Other	OBIS 673 nm LX	Coherent	Cat#1187194	Optogenetic stimulation laser
Other	LED	Prizmatix	UHP-T-595	Sham stimulation
Other	Titanium headplate	FMI/ETHZ workshop	N/A	Mice head-fixation
Other	Dental drill	Meisinger	N/A	For craniotomy

Mice. We used a total of 35 C57BL/6 mice, 4 fosGFP mice, 6 Tlx3-Cre mice, and 23 ChAT-IRES-Cre mice. Both male and female mice, 6 - 16 weeks old at the start of the experiment, were used. See **Table S2** for details of mouse inclusion for different figures. Between experiments, mice were group-housed in a vivarium (light/dark cycle: 12/12 hours). All animal procedures were approved by and carried out in accordance with the guidelines laid by the Veterinary Department of the Canton of Basel-Stadt, Switzerland.

Surgery. For all surgical procedures, mice were anesthetized with a mixture of fentanyl (0.05 mg/kg;

Actavis), midazolam (5.0 mg/kg; Dormicum, Roche), and medetomidine (0.5 mg/kg; Domitor, Orion)

injected intraperitoneally. Analgesics were applied perioperatively (2% lidocaine gel, meloxicam 5 mg/kg)

and postoperatively (buprenorphine 0.1 mg/kg, meloxicam 5 mg/kg). Eyes were covered with ophthalmic

gel (Virbac Schweiz AG). Depending on the experiment, we either implanted a cranial window to perform

two-photon imaging or used crystal skull (or clear skull preparation in case of GRAB-NE1m imaging) for

widefield imaging. Cranial windows were implanted over V1 and ACC as previously described (Keller et al.,

2012; Leinweber et al., 2014). Briefly, using a dental drill, a 4 mm craniotomy was made over the right V1,

centered 2.5 mm lateral and 0.5 mm anterior to lambda. A second craniotomy was made over right ACC,

centered at midline, 0.5 mm anterior to bregma. After injection of an AAV vector carrying the reporter

(see **Table S2** for details of virus), the exposed cortex was sealed with a 3 mm or 4 mm circular glass

coverslip and glued in place using gel superglue (Ultra Gel, Pattex). The remaining exposed surface of the

skull was covered with Histoacryl (B. Braun), and a titanium head bar was fixed to the skull using dental

cement (Paladur, Heraeus Kulzer). For widefield experiments, we injected an AAV vector with PHP.eB

capsid retro-orbitally (6 µl per eye of at least 10^{13} GC/ml) to drive expression throughout cortex, or imaged

GFP expressed in a fosGFP mice, or imaged without any fluorophore. For crystal skull surgery, we surgically

removed the skull plate overlying the dorsal cortex and superglued a crystal skull coverslip over the

craniotomy. For clear skull preparation, the skull was cleared with a three-component polymer (C&B

Metabond, Parkell), and the crystal skull coverslip was directly attached to the skull. An epifluorescence

image was taken to mark reference points on cortical surface. As with 4mm cranial window implantation,

a titanium head bar was fixed to the skull using dental cement. After surgery, anesthesia was antagonized

by a mixture of flumazenil (0.5 mg/kg; Anexate, Roche) and atipamezole (2.5 mg/kg; Antisedan, Orion

Pharma) injected intraperitoneally. Imaging commenced earliest 2 weeks after head bar implantation or

3 weeks after retro-orbital AAV injection.

Virtual reality environment. The virtual reality setup is based on the design of Dombeck and colleagues

(Dombeck et al., 2007). Briefly, mice were head-fixed and free to run on an air-supported spherical

treadmill. The rotation of the ball was restricted around the vertical axis with a pin. The virtual reality

environment was projected onto a toroidal screen covering approximately 240 degrees horizontally and

100 degrees vertically of the mouse's visual field, using a projector (Samsung SP-F10M) synchronized to

the resonant scanner of the two-photon microscope. The virtual environment consisted of an infinite

corridor with walls patterned with vertical sinusoidal gratings with a spatial frequency of approximately

0.04 cycles per degree (Leinweber et al., 2014). In closed loop sessions, the locomotion of the mouse was

coupled to movement along a virtual tunnel. In open loop sessions, we uncoupled the two and replayed

the visual flow from a preceding closed loop session. In grating sessions, we presented full-field drifting

gratings (0°, 45°, 90°, 270°, moving in either direction) in a pseudo-random sequence. Grating stimuli were

presented for between 1 s and 8 s depending on the experiment. In the inter-stimulus interval, mice were

shown a gray screen with average luminance matched to that of the grating stimuli. For optogenetic light

stimulation, we used 1 s pulses (10 mW after the objective) of a red laser (637 nm) directed at visual cortex

through a cranial window.

Two-photon imaging. Two-photon calcium imaging was performed using custom-built microscopes

(Leinweber et al. 2014). The illumination source was a tunable femtosecond laser (Insight, Spectra Physics

or Chameleon, Coherent) tuned to 930 nm. Emission light was band-pass filtered using a 525/50 filter for

GCaMP and a 607/70 filter for tdTomato/mCherry (Semrock) and detected using a GaAsP photomultiplier

(H7422, Hamamatsu). Photomultiplier signals were amplified (DHPCA-100, Femto), digitized (NI5772,

National Instruments) at 800 MHz, and band-pass filtered at 80 MHz using a digital Fourier-transform filter

implemented in custom-written software on an FPGA (NI5772, National Instruments). The scanning

system of the microscopes was based on a 12 kHz resonant scanner (Cambridge Technology). Images were

acquired at a resolution of 750 x 400 pixels (60 Hz frame rate), and a piezo-electric linear actuator (P-726,

Physik Instrumente) was used to move the objective (Nikon 16x, 0.8 NA) in steps of 15 µm between frames

to acquire images at 4 different depths. This resulted in an effective frame rate of 15 Hz. The field of view

was 375 µm x 300 µm.

Widefield imaging. Widefield imaging experiments were conducted on a custom-build macroscope with

commercially available objectives mounted face-to-face (Nikon 85 mm/f1.8 sample side, Nikon 50

mm/f1.4 sensor side). A 470 nm LED (Thorlabs) powered by a custom-build LED driver was used to excite

GFP through an excitation filter (SP490, Thorlabs) reflected off a dichroic mirror (LP490, Thorlabs) placed

parfocal to the objectives. The fluorescence was collected through a 525/50 emission filter on a sCMOS

camera (PCO edge 4.2). LED illumination was adjusted with a collimator (Thorlabs SM2F32-A) to

homogenously illuminate cortical surface through the cranial window. An Arduino board (Arduino Mega

2560) was used to synchronize LED onset with frame trigger signal of the camera. The duty cycle of the

470 nm LED was 90%. Images were acquired at 50 Hz or 100 Hz effective frame rate. Raw images were

cropped on the sensor and data was stored to disk with custom-written LabVIEW (National Instruments)

software, resulting in an effective pixel size of 60 μm^2 at a resolution of 1108 pixels x 1220 pixels (1.35

MP).

Extraction of neuronal activity. Calcium imaging data were processed as previously described (Keller et

al., 2012) and all data analysis was done in MATLAB (MathWorks). Briefly, raw images were full-frame

registered to correct for lateral brain motion. All data with visible z-motion were excluded. Raw

fluorescence traces were corrected for slow drift in fluorescence using an 8^{th}-percentile filtering with a 66

s (or 1000 frames) window (Dombeck et al., 2007). $\Delta F/F_0$ traces were calculated as mean fluorescence in

a selected region of every imaging frame, subtracted and normalized by the median fluorescence of the

trace. ROIs were placed relative to readily identifiable anatomical landmarks, resulting in two 20 pixels x

20 pixels ROIs per hemisphere. We calculated $\Delta F/F_0$ traces as for two-photon imaging.

Data analysis. All data analysis was done using custom scripts written in MATLAB (MathWorks). To

quantify the average population response traces, we first calculated the average event-triggered

fluorescence trace for each ROI. The responses of all ROIs were then averaged and baseline-subtracted.

Locomotion onset was defined as running speed crossing a threshold of 0.25 cm/s for at least 1 s, while

having been below the threshold during the preceding 1 s. The same criteria were used to define visual

flow onsets in the open loop condition using visual flow speed. Visuomotor mismatch responses were

probed by presenting brief 1 s full-field visual flow halts in the closed loop condition. For a visuomotor

mismatch event to be included in analysis, mice had to be locomoting uninterrupted above threshold

(0.25 cm/s) from -0.5 s to +1 s after the event onset. Additionally, for a ROI to be included for analysis of

the response to a particular event, it had to have at least 3 onsets to the event. The GFP response was

baseline subtracted using a -0.5 s to 0 s window relative to onset. The same was -1 s to -0.5 s in case of

locomotion onset to account for preparatory activity.

In **Figures 2C-2E**, to compare GFP and GCaMP responses, we trial averaged the response of individual

neurons and used the mean response over a time-window of 0 s to +2.0 s relative to locomotion or

visuomotor mismatch onset, with a baseline subtraction window of -1 s to -0.5 s. For grating responses,

we used the mean of the trial averaged response over +0.5 s to +3.0 s relative to stimulus onset, with a

baseline subtraction window of -0.5 s to 0 s.

In **Figure 3**, for each neuron, we averaged the response from +0.5 s to +1.5 s relative to onset for individual

locomotion onset and visuomotor mismatch trials and tested if the mean of this distribution is significantly

different from 0 using a t-test with a chance threshold of 5% . In a similar analysis, the response window

for grating responsive neurons was set at +0.5 s to +2.5 s to account for the 2 s long grating stimuli.

In **Figure 4**, the blood vessel diameter and the average GFP signal from both blood vessel ROIs and neuropil

ROIs were smoothed with a 0.5 s moving average and the variance explained was computed on the -2.0 s

to +6.0 s window relative to stimulus onset. In **Figure 4C**, data points were drawn as outliers if they were

larger than 75^{th} percentile + 1.5*(75^{th} percentile - 25^{th} percentile) or smaller than 25^{th} percentile - 1.5*(75^{th}

percentile - 25^{th} percentile) of the distribution and were omitted from the figure.

In **Figures 7** and **S3**, due to insufficient triggers, we relaxed the locomotion threshold for mismatch to 0.12

cm/s, and in **Figure S3**, we relaxed the sitting window before locomotion onset to -0.5 s to 0 s.

Statistical Analysis. All statistical information for the tests performed in the manuscript is provided in

Table S1. Unless stated otherwise, the shading indicates the error in mean estimation to within 1 standard

deviation. For analysis where the experimental unit was neurons (or ROIs), we used hierarchical bootstrap

(Saravanan et al., 2020) for statistical testing due to the nested nature (neurons (or ROIs) and mice) of the

data. Briefly, we trial averaged the data for the respective event per neuron (or ROI), and first resampled

the data (with replacement) at the level of imaging sites, and then, from the selected sites, resampled for

neurons. We then computed the mean of this bootstrap sample and repeated this N times to generate a

bootstrap distribution of the mean estimate. For all statistical testing the number of bootstrap samples

(N) was 10 000, for plotting bootstrap mean and standard error response curves it was 1000. The

bootstrap standard error is the 68% confidence interval (1 SD, 15.8^{th} percentile to 84.2^{nd} percentile) in the

bootstrap distribution of the mean.

 SUPPLEMENTARY INFORMATION

 Table S1. Statistical information on all analysis.

Figure panel	Value compared	Test	P-value	N (ROIs)	N (Sites)	N (Mice)
1A	Locomotion onset response in V1 L2/3	-	-	2647	6	6
1B	Grating onset response in V1 L2/3	-	-	2647	6	6
1C	Visual response to opto stim light in V1 L2/3	-	-	2647	6	6
1D	Visuomotor mismatch response in V1 L2/3	-	-	2218	5	5
1E	Locomotion onset response in V1 L5	-	-	1463	6	6
1F	Grating onset response in V1 L5	-	-	1463	6	6
1G	Visual response to opto stim light in V1 L5	-	-	1463	6	6
1H	Visuomotor mismatch response in V1 L5	-	-	947	4	4
1I	Locomotion onset response in ACC L2/3	-	-	2417	6	6
1J	Grating onset response in ACC L2/3	-	-	2417	6	6
1K	Visual response to opto stim light in ACC L2/3	-	-	2417	6	6
1L	Visuomotor mismatch response in ACC L2/3	-	-	2417	6	6
2B-2E	GCaMP6f *vs* GFP	-	-	15 103, 6527	51, 18	15, 6
3A	Fraction of L2/3 neurons in V1 classified responsive to locomotion onsets based on GFP signals	Hierarchical bootstrap	$<10^{-4}$	2647	6	6
	Fraction of L2/3 neurons in V1 classified responsive to grating onsets based on GFP signals	Hierarchical bootstrap	$<10^{-4}$	2647	6	6
	Fraction of L2/3 neurons in V1 classified responsive to mismatch onsets based on GFP signals	Hierarchical bootstrap	0.34	2218	5	5
	Fraction of L2/3 neurons in V1 classified responsive to locomotion onsets *vs* grating onset based on GFP signals	Hierarchical bootstrap	$<10^{-4}$	2647, 2647	6,6	6,6
	Fraction of L2/3 neurons in V1 classified responsive to grating onsets *vs* mismatch onset based on GFP signals	Hierarchical bootstrap	$<10^{-4}$	2647, 2218	6,5	6,5
	Fraction of L2/3 neurons in V1 classified responsive to locomotion onsets *vs* mismatch onset based on GFP signals	Hierarchical bootstrap	$<10^{-4}$	2647, 2218	6,5	6,5
3B	Fraction of L5 neurons in V1 classified responsive to locomotion onsets based on GFP signals	Hierarchical bootstrap	$<10^{-4}$	1463	6	6
	Fraction of L5 neurons in V1 classified responsive to grating onsets based on GFP signals	Hierarchical bootstrap	$<10^{-4}$	1463	6	6
	Fraction of L5 neurons in V1 classified responsive to mismatch onsets based on GFP signals	Hierarchical bootstrap	0.09	947	4	4

	Description	Test	p-value	N		
	Fraction of L5 neurons in V1 classified responsive to locomotion onsets *vs* grating onset based on GFP signals	Hierarchical bootstrap	0.08	1463, 1463	6,6	6,6
	Fraction of L5 neurons in V1 classified responsive to grating onsets *vs* mismatch onset based on GFP signals	Hierarchical bootstrap	$<10^{-4}$	1463, 947	6,4	6,4
	Fraction of L5 neurons in V1 classified responsive to locomotion onsets *vs* mismatch onset based on GFP signals	Hierarchical bootstrap	$<10^{-4}$	1463, 947	6,4	6,4
3C	Fraction of L2/3 neurons in ACC classified responsive to locomotion onsets based on GFP signals	Hierarchical bootstrap	$<10^{-4}$	2417	6	6
	Fraction of L2/3 neurons in ACC classified responsive to grating onsets based on GFP signals	Hierarchical bootstrap	0.18	2417	6	6
	Fraction of L2/3 neurons in ACC classified responsive to mismatch onsets based on GFP signals	Hierarchical bootstrap	0.34	2417	6	6
	Fraction of L2/3 neurons in ACC classified responsive to locomotion onsets *vs* grating onset based on GFP signals	Hierarchical bootstrap	$<10^{-4}$	2417, 2417	6,6	6,6
	Fraction of L2/3 neurons in ACC classified responsive to grating onsets *vs* mismatch onset based on GFP signals	Hierarchical bootstrap	0.57	2417, 2417	6,6	6,6
	Fraction of L2/3 neurons in ACC classified responsive to locomotion onsets *vs* mismatch onset based on GFP signals	Hierarchical bootstrap	$<10^{-4}$	2417, 2417	6,6	6,6
4C-4D	Variance explained in V1 L2/3 neuron GFP signal	-	-	2647	6	6
4C-4D	Variance explained in V1 L5 neuron GFP signal	-	-	1463	6	6
4C-4D	Variance explained in ACC L2/3 neuron GFP signal	-	-	2417	6	6
5A	Locomotion onset response under different conditions in V1 L2/3	-	-	2647	6	6
5B	Locomotion onset response under different conditions in V1 L5	-	-	1463	6	6
5C	Locomotion onset response under different conditions in ACC L2/3	-	-	2417	6	6
6A	Pairwise correlation in GFP signal in V1 L2/3 neurons, stationary *vs* locomoting	Hierarchical bootstrap	$<10^{-3}$	2647	6	6
6B	Pairwise correlation in GFP signal in V1 L5 neurons, stationary *vs* locomoting	Hierarchical bootstrap	$<10^{-4}$	1276	6	6
6C	Pairwise correlation in GFP signal in ACC L2/3 neurons, stationary *vs* locomoting	Hierarchical bootstrap	$<10^{-4}$	2417	6	6
7A	Locomotion onset response in V1	-	-	48	24	11
7B	Grating onset response in V1	-	-	52	26	12
7C	Visuomotor mismatch response in V1	-	-	48	24	11
7D	Locomotion onset response in ACC	-	-	48	24	11
7E	Grating onset response in ACC	-	-	52	26	12

7F	Visuomotor mismatch response in ACC	-	-	48	24	11
8B	Locomotion onset response in GRAB-DA1m in V1 – neuropil and blood vessel ROI response	-	-	168, 34	7, 5	7, 5
8C	Grating onset response in GRAB-DA1m in V1 – neuropil and blood vessel ROI response	-	-	168, 34	7, 5	7, 5
8D	Visuomotor mismatch response in GRAB-DA1m in V1 – neuropil and blood vessel ROI response	-	-	124, 20	5, 3	5, 3
8F	Locomotion onset response in GRAB-5HT1.0 in V1 – neuropil and blood vessel ROI response	-	-	220, 113	10, 10	9, 9
8G	Grating onset response in GRAB-5HT1.0 in V1 – neuropil and blood vessel ROI response	-	-	205, 101	9, 9	9, 9
8H	Visuomotor mismatch response in GRAB-5HT1.0 in V1 – neuropil and blood vessel ROI response	-	-	193, 101	9, 9	8, 8
8J	Locomotion onset response in GRAB-ACh3.0 in V1 – neuropil and blood vessel ROI response	-	-	178, 118	8,8	8,8
8K	Grating onset response in GRAB-ACh3.0 in V1 – neuropil and blood vessel ROI response	-	-	178, 118	8,8	8,8
8L	Visuomotor mismatch response in GRAB-ACh3.0 in V1 – neuropil and blood vessel ROI response	-	-	178, 118	8,8	8,8
9A	Locomotion onset response in GRAB-NE1m in V1	-	-	64	32	5
9B	Grating onset response in GRAB-NE1m in V1	-	-	12	6	5
9C	Visuomotor mismatch response in GRAB-NE1m in V1	-	-	18	9	5
9D	Locomotion onset response in GRAB-NE1m in ACC	-	-	64	32	5
9E	Grating onset response in GRAB-NE1m in ACC	-	-	12	6	5
9F	Visuomotor mismatch response in GRAB-NE1m in ACC	-	-	18	9	5
S1A	Visual flow onset response with GFP in V1 L2/3	-	-	2647	6	6
S1B	Locomotion onset response in dark with GFP in V1 L2/3	-	-	2647	6	6
S1C	Visual flow onset response in closed loop with GFP in V1 L2/3	-	-	2647	6	6
S3A	Locomotion onset response with widefield and no fluorophore in V1	-	-	24	12	6
S3B	Grating onset response with widefield and no fluorophore in V1	-	-	24	12	6
S3C	Visuomotor mismatch response with widefield and no fluorophore in V1	-	-	16	8	4
S3D	Locomotion onset response with widefield and no fluorophore in ACC	-	-	24	12	6
S3E	Grating onset response with widefield and no fluorophore in ACC	-	-	24	12	6
S3F	Visuomotor mismatch response with widefield and no fluorophore in ACC	-	-	16	8	4

S4A	Locomotion onset response in GRAB-DA1m in V1 – neuropil and blood vessel ROI response	-	-	122, 61	7, 7	7, 7
S4B	Grating onset response in GRAB- DA1m in V1 – neuropil and blood vessel ROI response	-	-	122, 61	7, 7	7, 7
S4C	Visuomotor mismatch response in GRAB- DA1m in V1 – neuropil and blood vessel ROI response	-	-	80, 50	5, 5	5, 5
S4D	Locomotion onset response in GRAB-5HT1.0 in ACC – neuropil and blood vessel ROI response	-	-	123, 79	7, 7	6, 6
S4E	Grating onset response in GRAB-5HT1.0 in ACC – neuropil and blood vessel ROI response	-	-	109, 69	6, 6	6, 6
S4F	Visuomotor mismatch response in GRAB-5HT1.0 in ACC – neuropil and blood vessel ROI response	-	-	91, 56	5, 5	5, 5

380 **Table S2. Number of mice in different experimental groups.**

Genotype	Virus	Site of injection	Nbr. Mice	Nbr. ROIs	Figures
C57BL/6	AAV2/1-Ef1α-eGFP-WPRE	Primary visual cortex – L2/3	6	3072	Figures 1, 2, 3, 4, 5, 6, S1, S2
C57BL/6	AAV2/1-Ef1α-eGFP-WPRE	Primary visual cortex – L5	6	1710	Figures 1, 2, 3, 4, 5, 6, S1, S2
C57BL/6	AAV2/1-Ef1α-eGFP-WPRE	Anterior cingulate cortex – L2/3	6	2739	Figures 1, 2, 3, 4, 5, 6, S1, S2
ChAT-IRES-Cre	AAV2/1-Ef1α-GCaMP6f-WPRE	Primary visual cortex – L2/3 and L5	15	15 103	Figure 2
C57BL/6	AAV-PHP.eB-EF1α-eGFP	Retro-orbital	8	32	Figure 7
fosGFP	-	-	4	20	Figure 7
Tlx3-Cre	-	-	6	24	Figure S3
ChAT-IRES-Cre	AAV2/9-hSyn-GRAB-ACh3.0	Primary visual cortex	8	609	Figure 8
C57BL/6	AAV2/9-hSyn-GRAB-DA1m	Anterior cingulate cortex	7	183	Figures 8, S4
		Primary visual cortex	7	202	
C57BL/6	AAV2/9-hSyn-GRAB-5HT1.0	Anterior cingulate cortex	9	202	Figures 8, S4
		Primary visual cortex	6	333	
C57BL/6	AAV-PHP.eB-hSyn-GRAB-NE1m	Retro-orbital	5	156	Figure 9

   FIGURES

Figure 1. GFP responses in two-photon imaging of neurons in V1 and ACC.

(**A**) Average change in apparent GFP fluorescence in L2/3 neurons in V1 on locomotion onset. Mean (solid lines) and the bootstrap standard error (shading) are calculated as hierarchical bootstrap estimate for each time bin.

(**B**) As in **A**, but for grating onsets in L2/3 neurons in V1. Green shading marks the duration of the gratings.

(**C**) As in **A**, but for responses to optogenetic stimulation light in L2/3 neurons in V1, with VR on (light red) or in dark (dark red). Pink shading marks the duration of the light stimulus.

(**D**) As in **A**, but for visuomotor mismatch in L2/3 neurons in V1. Orange shading marks the duration of visuomotor mismatch. As mismatch events are defined to occur during times of locomotion, and locomotion itself drives GFP responses, we would expect to find an increase in these signals by chance at mismatch. To correct for this, we quantified the distribution of signal amplitude on random triggers during locomotion (95% confidence interval, gray shading).

(**E**) As in **A**, but for L5 neurons in V1.

(**F**) As in **B**, but for L5 neurons in V1.

(**G**) As in **C**, but for L5 neurons in V1.

399   (**H**) As in **D**, but for L5 neurons in V1.

400   (**I**) As in **A**, but for L2/3 neurons in ACC.

401   (**J**) As in **B**, but for L2/3 neurons in ACC.

402   (**K**) As in **C**, but for L2/3 neurons in ACC.

403   (**L**) As in **D**, but for L2/3 neurons in ACC.

Figure 2. Peak responses of GCaMP6f were larger than those observed with GFP.

(**A**) Example GFP response traces (black) of three neurons with corresponding running speed traces (red).

(**B**) Peak responses of GCaMP6f (blue) were typically an order of magnitude larger than those observed with GFP (orange). Inset is a zoomed-in version of the same data.

(**C**) Histogram of trial averaged responses on locomotion onset measured with two-photon imaging of calcium indicator (blue) and GFP (orange). Dotted lines indicate the mean response of the distribution of locomotion onset response in calcium indicator (blue) and GFP (orange).

(**D**) As in **C**, but for grating onset.

(**E**) As in **C**, but for visuomotor mismatch.

Figure 3. Fraction of neurons with significant GFP responses to different stimuli.

(**A**) The fraction of GFP expressing V1 L2/3 neurons responsive to locomotion, full-field grating, and visuomotor mismatch onset, quantified for all imaging sites. Each dot is an imaging site, jittered along the x-axis to make them visible. Whiskers mark the SEM. Dashed line indicates chance levels (p = 0.05). n.s.: not significant; *: p < 0.05; **: p < 0.01; ***: p < 0.001; see **Table S1** for all statistical information.

(**B**) As in **A**, but for V1 L5 neurons.

(**C**) As in **A**, but for ACC L2/3 neurons.

Figure 4. Correlation between blood vessel diameter and GFP responses.

(**A**) Example average fluorescence image from an imaging site in L2/3 of V1 under a two-photon microscope. The fluorescence signal is averaged over 22 trials of optogenetic stimulation light directed at V1 while the mouse was in the dark. Shown on the right is an example blood vessel before (upper panel) and after (lower panel) stimulation.

(**A1**) Average area of the example blood vessel shown in **A** (in blue box) on light stimulation. The area of the indicated blood vessel is extracted from the trial averaged two-photon image. Shading in pink indicates the optogenetic light presentation.

(**A2**) Average change in GFP signal from a ROI positioned within the blood vessel (shown in **A** in red). Mean (red) and SEM (shading) were calculated over light stimulation trials.

(**A3**) Average change in GFP signal from all neurons (268 neurons) simultaneously imaged in the same site. Mean (red) and SEM (shading) were calculated over light stimulation trials.

(**A4**) Variance in average GFP signal in this example L2/3 site in V1 explained by area of the example blood
vessel and by the change in GFP signal from a ROI within the blood vessel.

(**B**) As in **A**, from an imaging site in L5 of V1.

(**C**) Percentage of variance in neuronal GFP signal to individual events explained by the change in GFP
signal from blood vessel ROIs. Boxes show 25^{th} and 75^{th} percentile, central mark is the median, and the
whiskers extend to the most extreme datapoints not considered outliers.

(**D**) Percentage of variance in neuronal GFP signal over the entire recording explained by blood vessel GFP
signal per imaging location. Each datapoint is one imaging site. Boxes show 25^{th} and 75^{th} percentile, central
mark is the median, and the whiskers extend to the most extreme datapoints not considered outliers.

Figure 5. GFP responses at locomotion onset depended on visuomotor condition and cortical region.

(**A**) Average GFP response in L2/3 neurons in V1 on locomotion onset in closed loop (in red), in open loop
(in blue), and in darkness (in grey). Here and in subsequent panels, bins with a significant difference (p <
0.01) are indicated by a black line above the plot (lines compared are marked on the left by pairs of colored
horizontal bars); those with p > 0.01 are marked gray. Mean (solid lines) and the bootstrap standard error
(shading) are calculated as hierarchical bootstrap estimate for each time bin.

(**B**) As in **A**, but for L5 neurons in V1.

(**C**) As in **A**, but for L2/3 neurons in ACC.

Figure 6. Locomotion increased pairwise correlation of GFP signals between neurons.

(**A**) Average pairwise correlation in GFP signal of V1 L2/3 neurons while mice were stationary and while they were locomoting. Each dot is the mean pairwise correlation of one neuron to all other neurons in the same imaging site. Colors indicate data from different mice.

(**B**) As in **A**, but for V1 L5 neurons.

(**C**) As in **A**, but for ACC L2/3 neurons.

Figure 7. Signals in widefield fluorescence imaging of V1 and ACC in mice expressing GFP.

(**A**) Average change in signal from V1 on locomotion onset. Mean (solid lines) and the bootstrap standard error (shading) are calculated as hierarchical bootstrap estimate for each time bin.

(**B**) As in **A**, but for grating onset in V1. Grating duration were randomly selected between 4 s and 8 s (green shading).

(**C**) As in **A**, but on visuomotor mismatch in V1. Orange shading marks the duration of visuomotor mismatch. Gray shading marks the 95% confidence interval.

(**D**) As in **A**, but for change in signal from ACC on locomotion onset.

(**E**) As in **B**, but for change in signal from ACC on grating onset.

(**F**) As in **C**, but for change in signal from ACC on visuomotor mismatch.

Figure 8. GRAB signals in V1 with two-photon imaging.

(**A**) Example two-photon image taken in V1 of mouse expressing GRAB-DA1m, with neuropil ROIs (in violet) and blood vessel ROIs (in red).

(**B**) Average change in GRAB-DA1m signal in V1 on locomotion onset in ROIs over neuropil (in black) and in ROIs in blood vessels (in red). Mean (solid lines) and the bootstrap standard error (shading) are calculated as hierarchical bootstrap estimate for each time bin.

(**C**) As in **B**, but for response to grating onset. Green shading marks the duration of the gratings.

(**D**) As in **B**, but for response to visuomotor mismatch. Orange shading marks the duration of visuomotor mismatch.

(**E**) Example two-photon image taken in V1 of mouse expressing GRAB-5HT1.0, with neuropil ROIs (in green) and blood vessel ROIs (in red).

(**F**) Average change in GRAB-5HT1.0 signal in V1 on locomotion onset in ROIs over neuropils (in black) and in ROIs in blood vessels (in red).

(**G**) As in **F**, but for response to grating onset.

(**H**) As in **F**, but for response to visuomotor mismatch.

(**I**) Example two-photon image of mouse expressing GRAB-ACh3.0 in V1, with neuropil ROIs (in green) and
blood vessel ROIs (in red).

(**J**) Average change in GRAB-ACh3.0 fluorescence in V1 in response to locomotion onset in ROIs over
neuropil (in black) and in ROIs over blood vessels (in red).

(**K**) As in **J**, but for response to grating onset.

(**L**) As in **J**, but for response to visuomotor mismatch.

Figure 9. Signals of V1 and ACC in widefield imaging in mice expressing GRAB-NE1m.

(**A**) Average change in signal from V1 on locomotion onset. Mean (solid lines) and the bootstrap standard error (shading) are calculated as hierarchical bootstrap estimate for each time bin. Here and below, GFP response (in gray) from **Figure 7** overlaid.

(**B**) As in **A**, but for grating onset in V1. Green shading marks the duration of the gratings.

(**C**) As in **A**, but for visuomotor mismatch in V1. Orange shading marks the duration of visuomotor mismatch. Gray shading marks the 95% confidence interval.

(**D**) As in **A**, but for change in signal from ACC on locomotion onset.

(**E**) As in **B**, but for change in signal from ACC on grating onset.

(**F**) As in **C**, but for change in signal from ACC on visuomotor mismatch.

504 **SUPPLEMENTARY FIGURES**

505
Figure S1. GFP response in V1 L2/3 on locomotion onset in closed loop was not a linear sum of
component responses. Related to Figure 1.

(**A**) Average change in GFP signal from L2/3 in V1 in response to visual flow onset in open loop during
stationary periods. Mean (solid lines) and the bootstrap standard error (shading) are calculated as
hierarchical bootstrap estimate for each time bin.

(**B**) As in **A**, but locomotion onset in dark.

(**C**) As in **A**, but for locomotion onset in closed loop.

Figure S2. The variance of neuronal GFP signal explained by blood vessel GFP signal as a function of smoothing window. Related to Figure 4D.

Black arrow indicates the smoothing window used in **Figure 4D**.

Figure S3. Signals in widefield fluorescence imaging of V1 and ACC in mice not expressing a fluorophore. Related to Figure 7.

(**A**) Average change in signal from V1 on locomotion onset. Mean (solid lines) and the bootstrap standard error (shading) are calculated as hierarchical bootstrap estimate for each time bin, in mice not expressing a fluorophore.

(**B**) As in **A**, but on grating onset in V1. Grating duration were randomly selected between 4 s and 8 s (green shading).

(**C**) As in **A**, but on visuomotor mismatch in V1. Orange shading indicates the duration of visuomotor mismatch.

(**D**) As in **A**, but for change in signal from ACC on locomotion onset.

(**E**) As in **B**, but for change in signal from ACC on grating onset.

(**F**) As in **C**, but for change in signal from ACC on visuomotor mismatch.

530

Figure S4. GRAB signals in ACC recorded with two-photon imaging. Related to Figure 8.

(**A**) Average change in GRAB-DA1m signal in ACC on locomotion onset in ROIs over neuropil (in black) and in ROIs over blood vessels (in red). Mean (solid lines) and the bootstrap standard error (shading) are calculated as hierarchical bootstrap estimate for each time bin.

(**B**) As in **A**, but for response to full-field grating onset. Green shading marks the duration of the gratings.

(**C**) As in **A**, but for response to visuomotor mismatch.

(**E**) Average change in GRAB-5HT1.0 signal in ACC on locomotion onset in ROIs over neuropil (in black) and in ROIs over blood vessels (in red).

(**F**) As in **G**, but for response to grating onset.

(**G**) As in **G**, but for response to visuomotor mismatch.

ACKNOWLEDGEMENTS

We thank all the members of the Keller lab for discussion and support. This project has received funding from the Swiss National Science Foundation (GBK), the Novartis Research Foundation (GBK), and the European Research Council (ERC) under the European Union's Horizon 2020 research and innovation programme (grant agreement No 865617) (GBK).

AUTHOR CONTRIBUTIONS

BY designed the experiments. BY performed the two-photon imaging experiments, RJ performed the widefield imaging of GRAB-NE1m experiments, and RJ and MH performed the widefield imaging of GFP and with no fluorophore experiments. BY analyzed the data. BY and GK wrote the manuscript.

DECLARATION OF INTERESTS

The authors declare no competing financial interests.

REFERENCES

Allen, W.E., Kauvar, I.V., Chen, M.Z., Richman, E.B., Yang, S.J., Chan, K., Gradinaru, V., Deverman, B.E., Luo, L., Deisseroth, K., 2017. Global Representations of Goal-Directed Behavior in Distinct Cell Types of Mouse Neocortex. Neuron 94, 891-907.e6. https://doi.org/10.1016/j.neuron.2017.04.017

Attinger, A., Wang, B., Keller, G.B., 2017. Visuomotor Coupling Shapes the Functional Development of Mouse Visual Cortex. Cell 169, 1291-1302.e14. https://doi.org/10.1016/j.cell.2017.05.023

Attwell, D., Buchan, A.M., Charpak, S., Lauritzen, M., MacVicar, B.A., Newman, E.A., 2010. Glial and neuronal control of brain blood flow. Nature 468, 232–243. https://doi.org/10.1038/nature09613

Aydın, Ç., Couto, J., Giugliano, M., Farrow, K., Bonin, V., 2018. Locomotion modulates specific functional cell types in the mouse visual thalamus. Nat. Commun. 9, 4882. https://doi.org/10.1038/s41467-018-06780-3

Berger, B., Gaspar, P., Verney, C., 1991. Dopaminergic innervation of the cerebral cortex: unexpected differences between rodents and primates. Trends Neurosci. 14, 21–27. https://doi.org/10.1016/0166-2236(91)90179-x

Chen, T.-W., Wardill, T.J., Sun, Y., Pulver, S.R., Renninger, S.L., Baohan, A., Schreiter, E.R., Kerr, R.A., Orger, M.B., Jayaraman, V., Looger, L.L., Svoboda, K., Kim, D.S., 2013. Ultrasensitive fluorescent proteins for imaging neuronal activity. Nature 499, 295–300. https://doi.org/10.1038/nature12354

Couto, J., Musall, S., Sun, X.R., Khanal, A., Gluf, S., Saxena, S., Kinsella, I., Abe, T., Cunningham, J.P., Paninski, L., Churchland, A.K., 2021. Chronic, cortex-wide imaging of specific cell populations during behavior. Nat. Protoc. 16, 3241–3263. https://doi.org/10.1038/s41596-021-00527-z

Dadarlat, M.C., Stryker, M.P., 2017. Locomotion enhances neural encoding of visual stimuli in mouse V1. J. Neurosci. 37, 3764–3775. https://doi.org/10.1523/JNEUROSCI.2728-16.2017

Dombeck, D.A., Khabbaz, A.N., Collman, F., Adelman, T.L., Tank, D.W., 2007. Imaging Large-Scale Neural Activity with Cellular Resolution in Awake, Mobile Mice. Neuron 56, 43–57. https://doi.org/10.1016/j.neuron.2007.08.003

Erisken, S., Vaiceliunaite, A., Jurjut, O., Fiorini, M., Katzner, S., Busse, L., 2014. Effects of Locomotion Extend throughout the Mouse Early Visual System. Curr. Biol. 24, 2899–2907. https://doi.org/10.1016/j.cub.2014.10.045

Feng, J., Zhang, C., Lischinsky, J.E., Jing, M., Zhou, J., Wang, H., Zhang, Y., Dong, A., Wu, Z., Wu, H., Chen, W., Zhang, P., Zou, J., Hires, S.A., Zhu, J.J., Cui, G., Lin, D., Du, J., Li, Y., 2019. A Genetically Encoded Fluorescent Sensor for Rapid and Specific In Vivo Detection of Norepinephrine. Neuron 102, 745-761.e8. https://doi.org/10.1016/j.neuron.2019.02.037

Hamada, H.T., Abe, Y., Takata, N., Taira, M., Tanaka, K.F., Doya, K., 2023. Optogenetic activation of dorsal raphe serotonin neurons induces brain-wide activation. https://doi.org/10.1101/2022.08.07.503074

Iadecola, C., Nedergaard, M., 2007. Glial regulation of the cerebral microvasculature. Nat. Neurosci. 10, 1369–1376. https://doi.org/10.1038/nn2003

Jing, M., Li, Yuexuan, Zeng, J., Huang, P., Skirzewski, M., Peng, W., Qian, T., Tan, K., Wu, R., Zhang, S., Pan, S., Xu, M., Li, H., Saksida, L.M., Prado, V.F., Bussey, T., Prado, M.A.M., Cheng, H., Li, Yulong, 2019. An optimized acetylcholine sensor for monitoring in vivo cholinergic activity 1–37.

Keller, G.B., Bonhoeffer, T., Hübener, M., 2012. Sensorimotor Mismatch Signals in Primary Visual Cortex of the Behaving Mouse. Neuron 74, 809–815. https://doi.org/10.1016/j.neuron.2012.03.040

598   Leinweber, M., Ward, D.R., Sobczak, J.M., Attinger, A., Keller, G.B., 2017. A Sensorimotor Circuit in
599   Mouse Cortex for Visual Flow Predictions. Neuron 95, 1420-1432.e5.
600   https://doi.org/10.1016/j.neuron.2017.08.036
601   Leinweber, M., Zmarz, P., Buchmann, P., Argast, P., Hübener, M., Bonhoeffer, T., Keller, G.B., 2014. Two-
602   photon Calcium Imaging in Mice Navigating a Virtual Reality Environment. J. Vis. Exp. 1–6.
603   https://doi.org/10.3791/50885
604   Lohani, S., Moberly, A.H., Benisty, H., Landa, B., Jing, M., Li, Y., Higley, M.J., Cardin, J.A., 2022.
605   Spatiotemporally heterogeneous coordination of cholinergic and neocortical activity. Nat.
606   Neurosci. 25, 1706–1713. https://doi.org/10.1038/s41593-022-01202-6
607   Ma, Y., Shaik, M.A., Kim, S.H., Kozberg, M.G., Thibodeaux, D.N., Zhao, H.T., Yu, H., Hillman, E.M.C., 2016.
608   Wide-field optical mapping of neural activity and brain haemodynamics: considerations and
609   novel approaches. Philos. Trans. R. Soc. B Biol. Sci. 371, 20150360.
610   https://doi.org/10.1098/rstb.2015.0360
611   Mächler, P., Fomin-Thunemann, N., Thunemann, M., Sætra, M.J., Desjardins, M., Kılıç, K., Şencan, I., Li,
612   B., Saisan, P., Cheng, Q., Weldy, K.L., Boas, D.A., Buxton, R.B., Einevoll, G.T., Dale, A.M., Sakadžić,
613   S., Devor, A., 2021. Microscopic Quantification of Oxygen Consumption across Cortical Layers.
614   https://doi.org/10.1101/2021.10.13.464176
615   Ocana-Santero, G., Packer, A.M., Sharp, T., Butt, S.J.B., 2024. In Vivo Two-Photon Microscopy Reveals
616   Sensory-Evoked Serotonin (5-HT) Release in Adult Mammalian Neocortex. ACS Chem. Neurosci.
617   15, 456–461. https://doi.org/10.1021/acschemneuro.3c00725
618   Ruff, C.F., Anaya, F.J., Dienel, S.J., Rakymzhan, A., Altamirano-Espinoza, A., Couey, J.J., Fukuda, M.,
619   Watson, A.M., Su, A., Fish, K.N., Rubio, M.E., Hooks, B.M., Ross, S.E., Vazquez, A.L., 2024. Long-
620   range inhibitory neurons mediate cortical neurovascular coupling. Cell Rep. 43.
621   https://doi.org/10.1016/j.celrep.2024.113970
622   Saravanan, V., Berman, G.J., Sober, S.J., 2020. Application of the hierarchical bootstrap to multi-level
623   data in neuroscience. Neurons Behav. Data Anal. Theory 3,
624   https://nbdt.scholasticahq.com/article/13927-application-of-the-hierarchical-bootstrap-to-
625   multi-level-data-in-neuroscience.
626   Scott, B.B., Thiberge, S.Y., Guo, C., Tervo, D.G.R., Brody, C.D., Karpova, A.Y., Tank, D.W., 2018. Imaging
627   cortical dynamics in GCaMP transgenic rats with a head-mounted widefield macroscope. Neuron
628   100, 1045-1058.e5. https://doi.org/10.1016/j.neuron.2018.09.050
629   Shen, Z., Lu, Z., Chhatbar, P.Y., O'Herron, P., Kara, P., 2012. An artery-specific fluorescent dye for
630   studying neurovascular coupling. Nat. Methods 9, 273–276.
631   https://doi.org/10.1038/nmeth.1857
632   Sun, F., Zeng, J., Jing, M., Zhou, J., Feng, J., Owen, S.F., Luo, Y., Li, F., Wang, H., Yamaguchi, T., Yong, Z.,
633   Gao, Y., Peng, W., Wang, L., Zhang, S., Du, J., Lin, D., Xu, M., Kreitzer, A.C., Cui, G., Li, Y., 2018. A
634   Genetically Encoded Fluorescent Sensor Enables Rapid and Specific Detection of Dopamine in
635   Flies, Fish, and Mice. Cell 174, 481-496.e19. https://doi.org/10.1016/j.cell.2018.06.042
636   Valley, M.T., Moore, M.G., Zhuang, J., Mesa, N., Castelli, D., Sullivan, D., Reimers, M., Waters, J., 2020.
637   Separation of hemodynamic signals from GCaMP fluorescence measured with wide-field
638   imaging. J. Neurophysiol. 123, 356–366. https://doi.org/10.1152/jn.00304.2019
639   Wan, J., Peng, W., Li, X., Qian, T., Song, K., Zeng, J., Deng, F., Hao, S., Feng, J., Zhang, P., Zhang, Y., Zou, J.,
640   Pan, S., Shin, M., Venton, B.J., Zhu, J.J., Jing, M., Xu, M., Li, Y., 2021. A genetically encoded
641   sensor for measuring serotonin dynamics. Nat. Neurosci. 24, 746–752.
642   https://doi.org/10.1038/s41593-021-00823-7

Waters, J., 2020. Sources of widefield fluorescence from the brain. eLife 9, 1–13.
https://doi.org/10.7554/eLife.59841
Widmer, F.C., O'Toole, S.M., Keller, G.B., 2022. NMDA receptors in visual cortex are necessary for
normal visuomotor integration and skill learning. eLife 11, e71476.
https://doi.org/10.7554/eLife.71476
Yogesh, B., Keller, G.B., 2023. Cholinergic input to mouse visual cortex signals a movement state and
acutely enhances layer 5 responsiveness. eLife 12. https://doi.org/10.7554/eLife.89986
Zhang, W.-T., Chao, T.-H.H., Yang, Y., Wang, T.-W., Lee, S.-H., Oyarzabal, E.A., Zhou, J., Nonneman, R.,
Pegard, N.C., Zhu, H., Cui, G., Shih, Y.-Y.I., 2022. Spectral fiber photometry derives hemoglobin
concentration changes for accurate measurement of fluorescent sensor activity. Cell Rep.
Methods 2, 100243. https://doi.org/10.1016/j.crmeth.2022.100243
Zmarz, P., Keller, G.B., 2016. Mismatch Receptive Fields in Mouse Visual Cortex. Neuron 92, 766–772.
https://doi.org/10.1016/j.neuron.2016.09.057
656