In situ structural analysis of mammalian cells using a 200 kV electron cryomicroscope – implications for research infrastructure.

Piotr Szwedziak[†*]

Center for Microscopy and Image Analysis
University of Zurich

[†]corresponding author:

Piotr Szwedziak

[†]Center for Microscopy and Image Analysis
University of Zurich
Winterthurerstrasse 190
8057 Zurich
Switzerland

[*]ReMedy-International Research Agenda Unit
University of Warsaw
Stefana Banacha 2c
02-097 Warsaw
Poland

email: piotr.szwedziak@zmb.uzh.ch

keywords: electron cryomicroscopy, electron cryotomography, *in situ*, subtomogram averaging, core facility, research infrastructure

Abstract

Electron cryotomography is a powerful imaging technique allowing for studying

functional cellular modules in their native environment with macromolecular

resolution. However, it requires access to complex and expensive instrumentation,

typically a 300 kV electron cryomicroscope equipped with an energy filter.

Simpler and cheaper 200 and 100 kV instruments have been successfully used for

single particle cryoEM analyses, which has helped to democratize the technique

and broaden access. Here, we set out to investigate if a 200 kV electron

cryomicroscope without an energy filter can be utilized for *in situ* structural

studies of mammalian cells by electron cryotomography. We demonstrate that the

resulting tomograms of thin edges of U2OS cells are of sufficient quality to

annotate the contents of the cell and observe spatial inter-relationships among

macromolecules. In particular, we undertook a molecular sociology analysis of

lipid droplets and addressed their subcellular distribution and interactions with

other organelles. Finally, we examined geographical distribution and scientific

output of the two most common electron cryomicroscopy platforms and deduced

that 200 kV instruments are heavily underutilized with respect to electron

cryotomography applications. Given the favorable ratio of their versatility versus

costs we foresee that 200 kV electron cryomicroscopes will become workhorses

of local electron cryomicroscopy facilities.

Introduction

It has been a decade since the resolution revolution in the electron
cryomicroscopy (cryoEM) field occurred (Kühlbrandt, 2014). Single particle
analysis (SPA) cryoEM has been a tremendously successful structural biology
technique (as evidenced by the exponentially growing number of relevant entries
in the Electron Microscopy Data Bank) enabling routine determination of high-
resolution 3D structures of purified macromolecules such as heterogenous or
flexible protein complexes (Punjani & Fleet, 2023; Schwab et al., 2024).

However, biological functions are not executed by individual molecules in
isolation but rather complex networks of interactions between all the cellular
components, including proteins, nucleic acids and lipids, are responsible for
maintaining cellular homeostasis. Therefore, the need for structural studies by
cryoEM *in situ* (i.e. in unperturbed cellular environments) is steadily growing
(Nogales & Mahamid, 2024). Electron cryotomography (cryoET) is a cryoEM
modality that provides 3D molecular resolution images of cellular environments.
CryoET preserves the full spectrum of conformations and interactions of each
molecule, enabling detailed ultrastructural analyses of cells and tissues in health
and disease that has been described as 'visual proteomics' (Beck & Baumeister,
2016; Förster et al., 2010). If structures of interest are present in multiple copies
within a tomogram, they can be extracted, aligned and averaged in a process called
subtomogram averaging, which is analogous to SPA but makes use of 3D volumes
rather than 2D projections. In favorable conditions this can yield a near-atomic
resolution reconstructions (Tegunov et al., 2021). Additionally, the position and
context of molecules and how they relate to the cellular context is available.

A critical requirement for cellular cryoET is that sample thickness must be less
than ~300 nm, given that the mean free path for inelastic scattering for 300 kV
electrons is ~300-400 nm (Martynowycz et al., 2021; Rice et al., 2018). Therefore,
initially the analyses were limited to *in vitro* specimens, bacteria or thin edges of
eukaryotic cells (Medalia et al., 2002; Szwedziak et al., 2014; Szwedziak & Pilhofer,
2019). In the last ten years techniques based on dual-beam FIB-SEM instruments

have become available to prepare thin lamellae of cells and tissues for subsequent

cryoET analyses (Rigort & Plitzko, 2015).

Complex sample preparation techniques, fragmented data processing pipelines

and lower throughput, as compared to SPA cryoEM, are primary reasons why

cryoET still lags SPA in terms of maturity and output. Cellular cryoET is a costly

technique as it frequently requires access to high-end 300 kV TEMs equipped with

energy filters that remove inelastically scattered electrons in order to increase the

signal-to-noise ratio. In the future the costs will become even more prohibitive

given the implementation of laser phase plates and Cc correctors that will yield an

ultimate, cryoET purpose-built instrument (Russo et al., 2022).

Given that the mean free path at 200 kV is ~260 nm and inspired by the study

(Martynowycz et al., 2021) suggesting that useful cryoEM data can be collected up

to twice the inelastic mean free path without the need for energy filters, we set out

to investigate if a 200 kV TEM equipped with a direct electron detector is sufficient

to provide meaningful cryoET data.

Here, we imaged by cryoET thin edges of mammalian U2OS cells using a 200 kV

Glacios G1 cryoTEM fitted with a Falcon 3EC camera without an energy filter. The

tomograms enabled us to observe a plethora of cellular structures ranging from

mitochondria through cytoskeletal filaments to macromolecular complexes, e.g.

ribosomes. By combining segmentation with subtomogram averaging we

generated detailed ultrastructural cellular snapshots that might be useful in

addressing various cell biology problems, e.g. lipid droplets (LDs) metabolism.

LDs are central to lipid and energy homeostasis and are important storage

organelles (Olzmann & Carvalho, 2019). LDs are known to originate in the

endoplasmic reticulum (ER) and associate with other cellular structures and a

myriad of proteins (Olzmann & Carvalho, 2019). To better understand the role of

LDs in cellular metabolism we undertook a molecular sociology analysis of LDs

and addressed their subcellular distribution and interactions with other

organelles. We also provide visual evidence for direct interactions between LDs

and membrane trafficking proteins.

The current trend towards democratizing cryoEM in the SPA field is gaining

momentum and it has been demonstrated that 200 kV (Herzik et al., 2017) and

100 kV (McMullan et al., 2023) cryoTEMs are sufficient to obtain high-quality SPA

cryoEM data. Now, we have proven that 200 kV TEMs without an energy filter can

be utilized for cellular cryoET investigations. We suggest that a 200 kV cryoTEM

can be a cost-effective workhorse cryoEM instrument capable of providing

complimentary information from both major cryoEM modalities.

Results

CryoET imaging of mammalian cells.

CryoET has been frequently employed to address cell biology problems that
otherwise were not accessible to other imaging techniques, e.g. coronavirus
replication (Wolff et al., 2020) or detailed membrane architecture (Wozny et al.,
2023). To ensure optimal performance and data quality these studies made use of
complex 300 kV TEMs equipped with energy filters. Due to prohibitive costs of
purchase and maintenance, such instruments are predominantly available in
major research centers with limited access. Given that the mean free path for
inelastic scattering at 200 kV is ~260 nm, which is comparable to the average
cryoET specimen thickness, we set out to investigate if a 200 kV TEM without an
energy filter can deliver interpretable cryoET results. Specifically, we were
interested in understanding the interplay and spatial arrangement of LDs in the
context of intact cells.

U2OS cells were seeded on Quantifoil grids and vitrified by plunge-freezing. These
cells are known to adopt a flat, pancake-like shape that offers a significant cryoET
imaging surface over thin cells areas without the need for micromachining by
cryoFIB. The grids were imaged by a Glacios G1 cryoTEM equipped with a Falcon
3EC camera without an energy filter. The cells could be easily recognized at low
magnifications (**Fig. 1a, left**) with the cell nucleus surrounded by the electron
transparent cytoplasmic regions. This allowed for a relatively straightforward
setup of automated tilt series acquisition. It was possible to acquire multiple tilt
series from a single cell, and the final thickness of the 3D volume was dictated by
the local cell thickness (**Fig. 1a, small panels**). The imaged areas were
structurally well preserved and showed no sign of crystalline ice contamination.
In total, 64 cells were imaged over which we recorded almost 230 tomograms with
an average thickness of 240 ± 75 nm (**Fig. 1b**). Most tomograms were between
200 and 250 nm thick, with the thinnest being 50 nm and thickest 400 nm (**Fig.
1c**). Attempts to record tilt series on thicker areas failed due to the low contrast
of resulting images and poor tracking. To assess the quality of the data
quantitively we fitted the contrast transfer function (CTF) on the 2D zero tilt

images and using CTFFIND4 we estimated the resolution up to which the CTF is

fitted reliably. We found that the precision of the fit is linearly dependent on the

local cell thickness (**Fig. 1d**). Encouraged by these results we moved on to a more

detailed inspection of the tomograms.

Molecular atlases of native cellular environments.

Since the aim of this study was to generate low nanometer resolution 3D volumes

of intact mammalian cells in order to qualitatively and quantitatively describe LDs

molecular neighborhoods, we adopted an appropriate data collection strategy

(**Table 1**): we utilized relatively high defocus (range from -5 to -7 μm) and high

total dose (140 eÅ^{-2}) to boost image contrast and large pixel size (5.1 Å) to benefit

from the broad cellular context provided by the large field of view (~2 μm). The

quality of the tomograms can be judged from the resolving power of known

features of mammalian cells, e.g.: microtubules, intermediate filaments, actin

filaments, lipid bilayers, ribosomes (**Fig. 2a, left panels**). A representative

tomogram of thickness of 180 nm is shown in **Fig. 2a (left large panel)** and **Movie**

S1. In the total volume of ~0.72 μm^3 we could find the three classes of intracellular

filaments (microtubules, intermediate filaments, actin) intertwined with each

other as well as with membranous structures of the endoplasmic reticulum (ER)

and ribosome particles. Subsequently we trained a neural network within the

EMAN2 software package (Chen et al., 2017) to recognize lipid bilayers,

intermediate filaments, double membranes of mitochondria, microtubules and

ribosomes and applied it to the tomograms (**Fig. 2a, middle panels**). This allowed

for reproducible identification of features of interest and their separation from the

background. By applying a neural network, we identified and automatically

extracted ∼1500 ribosomal particles and subjected them to subtomogram

averaging. This resulted in a moderate resolution 3D reconstruction at 38 Å with

the conservative threshold of 0.5 for FSC (**Fig. 2a, small panel**). Including more

particles in the average didn't improve the resolution suggesting that the dataset

is limited by the data collection strategy used (high defocus and large total

electron dose) and sample thickness. Importantly, resolution is not the best

criterion for data quality in cryoET as the contextual information provided by

cellular tomograms can be of higher significance. Accordingly, by mapping the

averaged structure to the calculated positions and orientations of each particle in

the tomograms and combining it with segmentations of other cellular components

we generated 3D molecular maps of thin edges of intact mammalian cells (**Fig. 2a,**

right panel, Movie S2). Such maps are visually appealing to human perception

and easier in interpretation than raw tomograms. We find the prospect of

producing detailed annotations of 3D cryoET volumes obtained by imaging with a

200 kV cryoTEM without an energy filter particularly exciting since such

segmentations aid interpretation and are essential for many downstream cryoET

analysis steps, e.g. observing spatial inter-relationships among macromolecules.

We went on to evaluate if the 200 kV cryoET data allow for reliable identification

of other known subcellular structures (**Fig. 2b**). In 13 tomograms we recognized

mitochondria that adopted either spherical or rod-shaped morphologies (**Fig. 2b**

(I), Movie S3 and Fig. S1). The measured cristae width was 12.8 nm, which stays

in accordance with the previously published 14.5±5.8 nm (Fry et al., 2024). Within

the mitochondrial matrix we could observe abundant granules characterized by

high contrast – we attributed those densities to calcium phosphate storage that is

used to regulate calcium homeostasis and to maintain mitochondrial function.

Similar densities were observed by cryoET in neuronal mitochondria (Wu et al.,

2023). CryoET at 200 kV could be an excellent diagnostic platform for *in situ*

organelle phenotyping as it has been shown that mitochondrial ultrastructure,

including its membrane architecture, has profound implications for human

disease (Navaratnarajah et al., 2021).

CryoET at 200 kV was sufficient to resolve coated vesicles/buds that play a role in

transport between various cellular compartments (**Fig. 2b (II), Movie S4**). The

overall morphology and the localization in the cell peripheries indicate that the

presented membrane-bound compartment is an early endosome. A closer look at

the dense protein coat and comparison to the previously published cryoET data

(Robinson et al., 2017) identified it as COPI. COPI coat is known to play an

important role in mediating transport between the Golgi and ER within the

secretory pathway but it has also been demonstrated to act in early endosome

function (Daro et al., 1997). The observed buds/vesicles had an average diameter

of 55±7 nm (n=7), which stays in accordance with other reports (Faini et al., 2012;

Robinson et al., 2017). To fully appreciate the complex membrane architecture,

we recommend seeing **Movie S4**.

In the imaged cells we found extensive areas of ER. The observed ER is

characterized by the lumen deprived of other intracellular

organelles/macromolecular complexes and is tightly coated on the cytoplasmic

face by ribosomes where individual particles could be discerned (**Fig. 2b (III) and**

Movie S5). The ER network consisted of cisterna interconnected by an array of

branching tubules. In 10 tomograms we also identified a recently reported form

of ER – ribosome associated vesicles (RAVs). RAVs were originally described as

dynamic ER subcompartments found primarily in the periphery of secretory cells

(Carter et al., 2020). Here, we found numerous RAVs interacting tightly with

microtubules that in the most severe cases triggered membrane flattening (**Fig.**

2b (IV) and Movie S6 and Fig. S2). RAV-bound ribosomes were excluded from

the RAV-microtubule interface. The observed RAVs were interconnected via

membrane tubules that were 8 nm in diameter (**Fig. 2b (V) and Movie S7**).

Membrane-bound polyribosome, which is a cluster of ribosomes bound to an

mRNA, was found on the RAV surface (**Fig. S3 and Movie S8**). Given that actively

translating mRNAs are frequently associated with multiple ribosomes we suspect

that RAVs might be sites of membrane/extracellular protein synthesis, similarly

to the canonical ER.

We also observed multilamellar compartments containing vesicles as well as

membrane sheets and other cellular debris/dense deposits. These are most likely

degradative compartments, for example autophagosomes (**Fig. 2b (VI) and**

Movie S9) or endolysosomes (**Fig. 2b (VII) and Movie S10**), as described in

(Foster et al., 2022). The luminal density within these organelles was highly

variable, suggesting that their content is different and not well defined.

Detailed analysis of LDs.

In 84 tomograms we noticed the presence of LDs. They are characterized by an

amorphous, electron-dense core made of neutral lipids with a single phospholipid

layer on the outside, which is manifested as the electron-dense line delineating

the LDs circumference in the tomograms (**Fig. 3a, top panel and Movie S11**). In

total we recognized 231 LDs that were 205±70 nm in diameter. LDs were often

found in clusters (**Fig. 3a, bottom panels and Movie S12**) and 29% of LDs stayed

in direct contact with each other, which frequently resulted in shape changes and

flattening (**Fig. 3a, bottom right panel and Fig. S4**), as reported previously

(Ganeva et al., 2023). ER double leaflet is the site of LDs synthesis (Fujimoto &

Parton, 2011) and they remain tightly associated with the ER throughout their life

cycle (**Fig. 3b, top right panel and Movie S13**): indeed, 69% of the cytoplasmic

LDs were in contact with ER and the interaction had impact on the ER membrane

curvature (**Fig. 3b, top left panel, Movie S14**). Interestingly, 5.6% of the

observed LDs were positioned in the ER lumen (**Fig. 3b, top right panel and Fig.**

S5). The luminal LDs are the least characterized members of the mammalian LD

family, and more investigations are needed to establish what triggers LDs

movement towards either cytoplasmic or luminal side. LDs association with ER

was described as very promiscuous (Valm et al., 2017) and we could observe

several distinct instances of this interaction. For example, luminal and cytoplasmic

LDs in direct contact presumably during material exchange (**Fig. 3b, bottom**

middle panel) or cytoplasmic LDs impacting the ER membrane and remaining at

a distance of 10-12 nm that is sufficient to accommodate proteins mediating the

interaction (**Fig. 3b, bottom right panel**). The lipid monolayer is a busy hub for

interaction with proteins that frequently are LD-specific (Olarte et al., 2022) and

mediate LDs interactions with various organelles (Olzmann & Carvalho, 2019): for

example, 3% of all LDs interacted with mitochondria (**Fig. 3c, left panel and**

Movie S15) and this association has been reported before as being crucial for

metabolism (Fan & Tan, 2024). 40% of the identified LDs interacted with

microtubules (**Fig. 3c, middle bottom panel, Movie S16 and Fig. S6**). Direct

physical contact between LDs and microtubules can aid with transport of LDs and

facilitate finding their cognate organelles. Rapid LD movement on microtubules

has been observed by time-lapse fluorescent light microscopy imaging (Gross et

al., 2000; Kilwein & Welte, 2019). A small fraction (1.7% of total observed LDs) of

LDs interacted with lysosomes and MVBs (**Fig. 3c, middle top and right panels**).

These LDs are most likely destined for degradation by lysosomal acid lipase and

the resulting fatty acids can enter metabolic pathways in mitochondria (Zhang et

al., 2022). We could discern direct association of membrane trafficking proteins

on the LDs surface. COPI has been demonstrated to act directly on lipid droplets

in vitro (Thiam et al., 2013) and *in vivo* to enable their connection to the ER for

protein targeting (Wilfling et al., 2014). Here, we observed the dense COPI coating

on the surface of an LD that promoted budding of lipid monolayer resulting in an

empty void between the LD core and the monolayer (**Fig. 3d left panels and**

Movie S17). It is likely that COPI-induced membrane deformation could help

establish continuous membrane bridges between the LD and outer ER membrane

leaflet (Wilfling et al., 2013), which has been observed in our tomograms (**Fig. 3b,**

bottom left panel). Here, the LD is tightly embedded within the lipid bilayer.

Another kind of vesicle protein coat, clathrin, has been detected in LDs by mass

spectrometry (Bartz et al., 2007; Nonoyama et al., 2019) and our study provides

visual confirmation of this interaction (**Fig. 3d, middle and right panels**). The

triskelion building block assembling into pentagonal and hexagonal lattice was

readily visible and the overall arrangement was similar to clathrin cage

architecture reported by SPA cryoEM (Fotin et al., 2006). The role of clathrin in

association with LDs is unclear but it has been suggested that it might be

important for lipid hydrolysis as it facilitates lysosome formation that is essential

for lipid droplet digestion (Nonoyama et al., 2019).

LDs are ubiquitous in the eukaryotic world and essential for lipid and energy

metabolism. With the emergence of cryoET imaging we are now beginning to

appreciate how LDs interact with other cellular compartments (summarized in

Table 2), and how these interactions play a role in cellular homeostasis. We

believe that structural insights into LDs *in situ* biology obtained by cryoET imaging

at 200 kV might further advance the field.

Discussion
2
CryoET is a versatile and unique structural biology technique as it can bridge
information from several orders of magnitude (Å to μm) and is an ideal tool for
integrative studies (Ziemianowicz & Kosinski, 2022). In this investigation we
sought to evaluate whether a 200 kV cryoTEM equipped with a direct electron
detector without an energy filter can deliver interpretable cryoET data that could
be used to address structural cell biology problems. Our primary objective was to
utilize the mammalian U2OS cell line as the model specimen, compare obtained
findings into cellular architecture with the previously published data acquired at
300 kV cryoTEMs fitted with direct electron detectors and energy filters and
possibly provide novel insights. This work resulted in several significant
outcomes: firstly, the utilized setup allowed for cryoET tilt series acquisition over
thin edges of mammalian cells with thicknesses ranging from 50 to 300 nm that
could subsequently be reliably reconstructed into 3D volumes; secondly, the
resulting tomograms could be reproducibly segmented and all the major cellular
components readily recognized; thirdly, the images of RAVs, LDs and other
cellular compartments were qualitatively comparable to previously published
tomograms acquired by 300 kV cryoTEMs with energy filters (Dudka et al., 2024;
Ganeva et al., 2023; Mahamid et al., 2019); fourthly, quantitative analysis of LDs
in their native cellular environment revealed the molecular sociology of these
organelles. Additionally, in this work we provide molecular resolution direct
visual evidence of LDs interaction with membrane trafficking components COPI
and clathrin. We expect that a 200 kV electron cryomicroscope without an energy
filter will produce similar results when imaging lamella of cellular/tissue
specimens obtained by cryo-focused ion beam milling using either gallium or
plasma sources which typically are micromachined to the final thickness of 150-
200 nm. It remains to be seen how a 200 kV cryoTEM without an energy filter can
perform regarding high-resolution subtomogram averaging, including both *in*
vitro and cellular specimens.

The field of *in situ* structural studies by cryoET is lagging behind SPA cryoEM – as
of August 2024 83.4% EMDB entries were derived from SPA analyses vs. 11.4%
obtained by cryoET approaches. However, the dynamics of the growth for the

cryoET modality is greater – in 2023 year-over-year growth was 50% as

compared to 24% for SPA. This rapid growth has been partially possible due to

higher throughput data collection strategies (Eisenstein et al., 2023, 2024) and

attempts to streamline data processing (Burt et al., 2024; Chen et al., 2019a).

Current hardware developments including C_c correction, laser phase plates and

reduced radiation damage at liquid helium temperatures will further boost the

field of cryoET but with a hefty price tag. Such cryoET-purpose built instruments

will be prohibitively expensive to purchase and maintain and will be available in

a handful of institutions only.

To better understand how current cryoEM resources are utilized we reviewed the

output (based on the EMDB database) and distribution (based on the map of high-

end cryoEM worldwide (Noble, 2024)) of the two most popular 200 and 300 kV

cryoEM paltforms: Thermo Fisher Glacios and Thermo Fisher Krios, respectively

(**Fig. 4a**). In 2023 there were 5379 EMDB entries attributed to SPA and 837

entries attributed to cryoET obtained from the Krios platform. Whereas Glacios

contribution to SPA cryoEM was only 0.95% of Krios in 2020, it steadily rose to

5.9% in 2023. In contrast, the Glacios platform delivered only 4 EMDB entries in

the whole of 2023.

As of August 2024, there were ~364 Krioses and ~230 Glacioses around the world

suggesting that 200 kV cryoTEMs are predominantly used as screening devices

and heavily underutilized for primary data collection, as evidenced by the number

of relevant EMDB entries. In further analysis we focused on Europe that is home

to 102 Krioses and 74 Glacioses. As shown on the map, the distribution of these

instruments across the continent is not equal. 17 European countries are in

possession of either Krios or Glacios or both, 2 countries have access to other

cryoEM platforms whereas the 25 remaining countries don't have access to

cryoEM instrumentation at all (**Fig. 4b, c**). The maturity and complexity of cryoEM

infrastructure reflects to a certain degree economic development - the top 3

Glacios/Krios owner countries are also 3 top countries by nominal GDP (according

to IMF, August 2024). Next, we surveyed the distribution of the Glacios and Krios

cryoEM platforms across academic hubs defined here as cities or agglomerations

characterized by the presence of research institutes, either academic or

commercial. The Krioses were spread across 39 academic hubs whereas Glacioses

across 49, which accounts for 2.6 Krios per hub and 1.5 Glacios per hub. Such

clustering of state of the art 300 kV cryoTEMs and better spatial distribution of

200 kV microscopes is related to the fact that the former ones are usually

maintained by national/international facilities (eBIC UK, EMBL Germany, NeCEN

Netherlands, SciLifeLab Sweden). This trend is visible worldwide, e.g. in the

United States there are 3 major SPA cryoEM centers and 4 cryoET centers under

the umbrella of NIH Common Fund Transformative High Resolution Cryo-Electron

Microscopy Program Centers, each containing multiple high-end instruments.

Such centers provide not only access to state-of-the-art microscopes but also to

know-how and sample preparation services that, particularly for cellular cryoET,

are complex and expensive. The clustering of cryoEM research infrastructure will

become more prominent with the arrival of another generation of cryoET

purpose-built cryoTEMs. Such microscopes can cost >5 M CHF with added costs

for space, service contracts and qualified staff.

200 kV cryoTEMs have been demonstrated to deliver high-quality SPA data, e.g.

(Brouwer et al., 2024; Träger et al., 2024) as well as cryoET data from *in vitro*

samples, e.g. (Tian et al., 2020). Here, we show that these instruments are capable

of delivering excellent cryoET images of cellular specimens. We believe that the

200 kV platform can be a viable option for smaller universities and research

institutes that will help democratize cryoEM access and can serve as workhorse

for both major cryoEM modalities: SPA and cryoET. We expect that these 1st

degree cryoEM access points will further boost the field by training independent

users/operators,   facilitating   methods   development   and   fostering

interdisciplinary research in local academic centers and will greatly complement

the 2nd degree cryoEM access points offering the latest and ultimate technological

solutions (**Fig. 4d**).

Acknowledgements

We thank dr Ishier Raote (CRG Barcelona) for providing the U2OS cell line, dr

Tomasz Góral (University of Warsaw) for the microscope maintenance and Sonia

Ngati (University of Warsaw) for the cell culture technical assistance. Members of

the Center for Microscopy and Image Analysis at the University of Zurich are

acknowledged for insightful discussions. This work was supported by the

"Regenerative Mechanisms for Health-ReMedy" grant MAB/20172 from the

Foundation for Polish Science.

Author contributions

PS: conception and design, acquisition of data, analysis and interpretation of data,

drafting and revising the article

Figure Legends

Figure 1. CryoET imaging of intact mammalian cells using a 200 kV cryoTEM

without an energy filter.

a)   A low magnification overview of a U2OS cell seeded on a Quantifoil grid

and vitrified in liquid ethane-propane. The electron opaque area encircled

in blue corresponds to the nucleus/perinuclear area. The cell boundaries

are encircled in red. Six tomograms were acquired in the electron-

transparent area corresponding to the thin cell peripheries. Central slices

of each tomogram are shown in the panels on the right and color-coded to

mark the acquisition areas. Tomograms thicknesses are displayed. Scale

bars: 10 μm and 100 nm.

b)   Table summarizing the dataset acquired in this study.

c)   Histogram of the tomogram thicknesses. The thickness varied from 50 to

400 nm with the most populated bin in the range of 200-250 nm.

d)   Plot of resolution up to which CTF was reliably fitted for a central tilt in

every tilt series against the tomogram/local cell thickness. The red line

shows a linear fit through all data points.

Figure 2. CryoET at 200 kV provides molecular resolution atlases of intact

mammalian cells.

a)   Central slice (14 nm thick) through a tomogram of an intact U2OS cell with

a local thickness of 180 nm (left large panel). Small panels on the left

display common mammalian cell features that were identified in the

tomogram, from top to bottom: microtubules, intermediate filaments,

actin, membrane tubules and ribosomes (red arrows). By employing

machine learning-based approaches we could identify, extract and average

subvolumes containing ribosome particles and map the resulting low

nanometer resolution average back in the tomogram (middle panels). The

quality of the tomogram was sufficient to segment other components of the

cell: membranes, mitochondria, intermediate filaments and microtubules

(middle panels). By combining all the segmentations, we generated a

simplified model of the cell that can aid interpretation (right panel). Scale

bars: small panels – 25 nm, main panel – 100 nm.

b) Examples of some of the functional cellular modules that could be readily

identified in the tomograms: I) Mitochondrion with calcium phosphate

inclusions (red arrow). Measured cristae diameter is highlighted in white.

II) COPI-coated vesicles budding off an early endosome. Color-coded are

enlarged views of highlighted areas at a different tomographic slice. COPI

coat is characterized by a very dense and compact appearance. III)

endoplasmic reticulum (ER) with its empty lumen and the cytoplasmic

surface covered with ribosomes (red arrow). IV) ribosome-associated

vesicle (RAV) covered densely with ribosome particles except where it

interacts with a microtubule filament that ultimately leads to a local

ribosome deprivation and membrane flattening of otherwise spherical

structure (red arrow). V) RAVs and ER subcompartments are

interconnected by a dense network of thin membrane tubules that are 8

nm in diameter, as highlighted in the enlarged view that is a snapshot at a

different tomographic slice of the area encircled in blue. VI)

Autophagosome with internal membrane debris. VII) Characteristic, bell-

like shaped endolysosome that is a result of the fusion of a lysosome with

a multivesicular body. Scale bars: major panels – 100 nm, enlarged views –

25 nm.

Figure 3. Molecular sociology of lipid droplets (LDs) in U2OS cell periphery as

revealed by cryoET at 200 kV.

a) LDs were found as spherical, amorphous entities characterized by distinct

contrast (top left panel). One of the unique structural features of LDs is

their lipid monolayer surface that manifests itself as an electron dense line

around perimeter (enlarged view, red arrow). LDs were frequently found

near each other (bottom left panel), including direct contact that resulted

in flattened and deformed morphology (bottom right panel, red arrow).

Scale bars: major panels – 100 nm, enlarged view – 25 nm.

b) LDs lifecycle is inherently connected to ER, which is the site of their
synthesis. 69% of cytoplasmic LDs were in close association with the ER
membrane that impacted membrane curvature (top left panel, blue
arrowheads). Three distinct modes of LDs interactions with ER are shown
on the top right panel and in enlarged views in the bottom panels, color-
coded accordingly. The LD highlighted in orange is embedded within the
ER lipid bilayer, which splits to accommodate the bulky volume of the LD
(red arrow); in blue are highlighted luminal and cytoplasmic LDs
interacting across the ER lipid bilayer (red arrow) and in red a cytoplasmic
LD impacting the ER membrane and remaining at a distance of 10-12 nm
that is sufficient to accommodate proteins mediating the interaction
(Hugenroth & Bohnert, 2020). Scale bars: major panels – 100 nm, enlarged
view – 25 nm.

c) LDs remain in close association (marked with red arrows) with a myriad
of cellular modules: mitochondria (left panel), microtubules (middle
bottom panel), MVBs (middle top panel), lysosomes (right panel). Scale
bars: 100 nm.

d) LDs interact with membrane trafficking machinery. Top left panel depicts
an LD with apparent lipid monolayer deformation (blue square). More
detailed analysis revealed (bottom left panel, which is a view of the region
marked with the blue square at a different tomographic slice) that the lipid
monolayer is coated with the dense COPI protein layer (red arrow, see also
Fig. 2b (II) for comparison).

Top middle and right panels show tomographic slices through central
sections of two LDs and below their corresponding views at different
tomographic slices. The surface of the two LDs is covered with clathrin that
adopts its characteristic architecture with the triskelion forming both
hexagonal and pentagonal units (red arrows). Scale bars: top left panel –
100 nm, other panels – 25 nm.

Figure 4. Current utilization of high-end cryoTEMs.

a) Table summarizing scientific output generated by two most common 200
kV and 300 kV cryoTEMs: Glacios and Krios, respectively. Whereas the
Glacios platform contributed less than 1% of the Krios SPA output in 2020,
this fraction rose to 5.9% in 2023. With this respect the Glacios platform is
heavily underutilized in cryoET with only a marginal contribution. Based
on the EMDB database.

b) Geographical distribution of the Glacios and Krios platforms in Europe. As
of August 2024, there were ~102 Krioses and ~74 Glacioses scattered
across 17 European countries. Eastern and South-Eastern Europe regions
are underequipped. Based on the map of high end cryoEM worldwide as of
August 2024 (Noble, 2024).

c) Number of Glacios and Krios cryoTEMs in European countries. The
countries are listed from left to right according to the nominal gross
domestic product (according to IMF, August 2024).

d) Implications of this study for core facilities: it is expected that 200 kV
instruments will become versatile workhorses in the majority of local
cryoEM facilities. These 1st degree access facilities are essential for
democratizing access to cryoEM. The 2nd degree access facilities that
operate at the national/international level will offer access to cutting-edge
imaging technology and advanced sample preparation techniques for
cryoET.

Table 1. CryoET data collection.

Microscope	Thermo Fisher Glacios G1
Camera	Falcon 3EC
Energy filter	N/A
Magnification	28000
Voltage (kV)	200
Electron exposure (eÅ^{-2})	140
Defocus range (μm)	5.0-7.0
Pixel size (Å)	5.1
Tilt range	±54^0
Increment	3^0

3   **Table 2. Lipid droplets interactions summary.**

#total LDs	231
#cytoplasmic LDs	218
#luminal LDs	13
#LDs in contact with each other	68
#LDs in contact with microtubules	94
#LDs in contact with ER	151
#LDs in contact with mitochondria	7
#LDs in contact with lysosomes/MVBs	4

Supplementary Figures and Movies Legends

Figures:

Figure S1. Representative cryoET slices showing mitochondria with calcium chloride deposits. Corresponding to **Fig. 2b (I)**. Scale bars: 100 nm.

Figure S2. Representative cryoET slice showing a RAV closely associated with a microtubule (red arrow). Corresponding to **Fig. 2b (IV)**. Scale bar: 100 nm.

Figure S3. Representative cryoET images showing a RAV at two corresponding views at different tomographic slices. Red arrow highlights a polysome associated with the surface of the RAV. Corresponding to **Fig. 2b (V)** and **Movie S8**. Scale bars: 100 nm.

Figure S4. Representative cryoET slice showing two LDs closely associated with each other (red arrow). Corresponding to **Fig. 3a**. Scale bar: 100 nm.

Figure S5. Representative cryoET slices showing LDs residing in the ER lumen. Corresponding to **Fig. 3b**. Scale bars: 100 nm.

Figure S6. Representative cryoET slice showing LD closely associated with a microtubule (red arrow). Corresponding to **Fig. 3c**. Scale bar: 100 nm.

Movies (available here):

Movie S1. Tomogram of a thin edge of an U2OS cell. Corresponding to **Fig. 2a**. Scale bar: 100 nm.

Movie S2. Segmentation of a thin edge of an U2OS cell. Corresponding to **Fig. 2a** and **Movie S1**.

Movie S3. Tomogram of a mitochondrion found in the thin edge of an U2OS cell. Corresponding to **Fig. 2b (I)**. Scale bar: 100 nm.

Movie S4. Tomogram of COPI-coated membranes found in the thin edge of an U2OS cell. Corresponding to **Fig. 2b (II)**. Scale bar: 100 nm.

Movie S5. Tomogram of ribosome-coated endoplasmic reticulum found in the thin edge of an U2OS cell. Corresponding to **Fig. 2b (III)**. Scale bar: 100 nm.

Movie S6. Tomogram of a ribosome-associated vesicle attached to a microtubule found in the thin edge of an U2OS cell. Corresponding to **Fig. 2b (IV)**. Scale bar: 100 nm.

Movie S7. Tomogram of two ribosome-associated vesicles connected by a thin membrane tubule found in the thin edge of an U2OS cell. Corresponding to **Fig. 2b (V)**. Scale bar: 100 nm.

Movie S8. Tomogram of a large ribosome-associated vesicle with an internal LD and with polysomes positioned on its top surface found in the thin edge of an U2OS cell. Corresponding to **Fig. S3**. Scale bar: 100 nm.

Movie S9. Tomogram of autophagosome found in the thin edge of an U2OS cell. Corresponding to **Fig. 2b (VI)**. Scale bar: 100 nm.

Movie S10. Tomogram of an endolysosome found in the thin edge of an U2OS cell. Corresponding to **Fig. 2b (VII)**. Scale bar: 100 nm.

Movie S11. Tomogram of a single lipid droplet found in the thin edge of an U2OS cell. Corresponding to **Fig. 3a**. Scale bar: 100 nm.

Movie S12. Tomogram of a bunch of lipid droplets found in the thin edge of an U2OS cell. Corresponding to **Fig. 3a**. Scale bar: 100 nm.

Movie S13. Tomogram of lipid droplets closely interacting with an endoplasmic reticulum found in the thin edge of an U2OS cell. Corresponding to **Fig. 3b**. Scale bar: 100 nm.

Movie S14. Tomogram of a lipid droplet impacting the lipid bilayer of an endoplasmic reticulum found in the thin edge of an U2OS cell. Corresponding to **Fig. 3b**. Scale bar: 100 nm.

Movie S15. Tomogram of a lipid droplet closely interacting with a mitochondrion found in the thin edge of an U2OS cell. Corresponding to **Fig. 3c**. Scale bar: 100 nm.

Movie S16. Tomogram of a lipid droplet closely interacting with microtubules found in the thin edge of an U2OS cell. Corresponding to **Fig. 3c**. Scale bar: 100 nm.

1   **Movie S17.** Tomogram of a lipid droplet coated with COPI found in the thin edge

2   of an U2OS cell. Corresponding to **Fig. 3d**. Scale bar: 25 nm.

3   **Movie S18 and Movie S19.** Tomograms of lipid droplets coated with clathrin

4   found in the thin edge of an U2OS cell. Corresponding to **Fig. 3d**. Scale bars: 25 nm.

Methods

Sample preparation

Quantifoil Au200 R2/2 grids were glow discharged for 90 s at 20 mA, coated with
fibronectin (25 μg/ml) and transferred to a 12 well plate in 3D printed cell culture
grid holders to facilitate handling (Fäßler et al., 2020). After sterilization for 30'
under UV light, the grids were seeded with U2OS cells (~6000 per well) and
incubated at 37^0C overnight. Before plunge freezing the grids were briefly washed
in PBS and 2.5 μl of 10 nm colloidal gold fiducial markers was applied.
Subsequently, the grids were back-blotted and vitrified in liquid ethane using a
Thermo Fisher Vitrobot.

Data collection

Electron cryotomography was performed using a Thermo Fisher Glacios G1 TEM
operating at 200 kV, equipped with a 4k x 4k Falcon 3EC direct electron detection
camera at a magnification of 28k, corresponding to a pixel size of 5.1 Å at the
specimen level. Samples were tilted between approximately -54^0 and $+54^0$ with a
3^0 increment using the dose symmetric scheme. The defocus was set between -5
and -7 μm and the total dose for each tilt series was 140 eÅ^{-2}.

Data analysis

Tomographic reconstruction from tilt series were calculated using RAPTOR (Amat
et al., 2008) and the IMOD tomography reconstruction package (Kremer et al.,
1996) followed by Weighted Back Projection or Simultaneous Iterative
Reconstruction Technique (SIRT) with the tomo3d package (Agulleiro &
Fernandez, 2011). Measurements of distances were carried out within IMOD.
Single threshold surface representation and movies were prepared using Chimera,
with a similar threshold level to exclude most background noise for each data set.
EMAN2 (Chen et al., 2017, 2019b) was used for tomograms
segmentations/annotations and subtomogram averaging. CTFFIND4 (Rohou &
Grigorieff, 2015) was used to obtain information about CTF fit.

Data deposition

Representative 3D volumes have been deposited in the Electron Microscopy Data
Bank (EMDB) under accession numbers: EMD-51957, EMD-51959, EMD-51960,
EMD-51961 and EMD-51962.

References

Agulleiro, J. I., & Fernandez, J. J. (2011). Fast tomographic reconstruction on
multicore computers. *Bioinformatics*, *27*(4), 582–583.
https://doi.org/10.1093/bioinformatics/btq692

Amat, F., Moussavi, F., Comolli, L. R., Elidan, G., Downing, K. H., & Horowitz, M.
(2008). Markov random field based automatic image alignment for electron
tomography. *Journal of Structural Biology*, *161*(3), 260–275.
https://doi.org/10.1016/j.jsb.2007.07.007

Bartz, R., Zehmer, J. K., Zhu, M., Chen, Y., Serrero, G., Zhao, Y., & Liu, P. (2007).
Dynamic activity of lipid droplets: Protein phosphorylation and GTP-
mediated protein translocation. *Journal of Proteome Research*, *6*(8), 3256–
3265. https://doi.org/10.1021/pr070158j

Beck, M., & Baumeister, W. (2016). Cryo-Electron Tomography: Can it Reveal the
Molecular Sociology of Cells in Atomic Detail? In *Trends in Cell Biology* (Vol.
26, Issue 11, pp. 825–837). Elsevier Ltd.
https://doi.org/10.1016/j.tcb.2016.08.006

Brouwer, P. J. M., Perrett, H. R., Beaumont, T., Nijhuis, H., Kruijer, S., Burger, J. A.,
Bontjer, I., Lee, W.-H., Ferguson, J. A., Schauflinger, M., Müller-Kräuter, H.,
Sanders, R. W., Strecker, T., van Gils, M. J., & Ward, A. B. (2024). Defining
bottlenecks and opportunities for Lassa virus neutralization by structural
profiling of vaccine-induced polyclonal antibody responses. *Cell Reports*,
43(9), 114708. https://doi.org/10.1016/j.celrep.2024.114708

Burt, A., Toader, B., Warshamanage, R., von Kügelgen, A., Pyle, E., Zivanov, J.,
Kimanius, D., Bharat, T. A. M., & Scheres, S. H. W. (2024). An image
processing pipeline for electron cryo-tomography in RELION-5. *FEBS Open*
Bio. https://doi.org/10.1002/2211-5463.13873

Carter, S. D., Hampton, C. M., Langlois, R., Melero, R., Farino, Z. J., Calderon, M. J.,
Li, W., Wallace, C. T., Han Tran, N., Grassucci, R. A., Siegmund, S. E.,
Pemberton, J., Morgenstern, T. J., Bartolini, F., Volchuk, A., Murray, S. A.,
Aridor, M., Fish, K. N., Walter, P., … Freyberg, Z. (2020). Ribosome-associated
vesicles: A dynamic subcompartment of the endoplasmic reticulum in
secretory cells. In *Despoina Aslanoglou* (Vol. 10). https://www.science.org

Chen, M., Bell, J. M., Shi, X., Sun, S. Y., Wang, Z., & Ludtke, S. J. (2019a). A complete
data processing workflow for cryo-ET and subtomogram averaging. *Nature*
Methods, *16*(11), 1161–1168. https://doi.org/10.1038/s41592-019-0591-8

Chen, M., Bell, J. M., Shi, X., Sun, S. Y., Wang, Z., & Ludtke, S. J. (2019b). A complete
data processing workflow for cryo-ET and subtomogram averaging. *Nature*
Methods, *16*(11), 1161–1168. https://doi.org/10.1038/s41592-019-0591-8

Chen, M., Dai, W., Sun, S. Y., Jonasch, D., He, C. Y., Schmid, M. F., Chiu, W., & Ludtke,
S. J. (2017). Convolutional neural networks for automated annotation of

cellular cryo-electron tomograms. *Nature Methods*, *14*(10), 983–985. https://doi.org/10.1038/nmeth.4405

Daro, E., Sheff, D., Gomez, M., Kreis, T., & Mellman, I. (1997). Inhibition of Endosome Function in CHO Cells Bearing a Temperature-sensitive Defect in the Coatomer (COPI) Component-COP. In *The Journal of Cell Biology* (Vol. 139, Issue 7). http://www.jcb.org

Dudka, W., Salo, V. T., & Mahamid, J. (2024). Zooming into lipid droplet biology through the lens of electron microscopy. In *FEBS Letters* (Vol. 598, Issue 10, pp. 1127–1142). John Wiley and Sons Inc. https://doi.org/10.1002/1873-3468.14899

Eisenstein, F., Fukuda, Y., & Danev, R. (2024). Smart parallel automated cryo-electron tomography. *Nature Methods*, *21*(9), 1612–1615. https://doi.org/10.1038/s41592-024-02373-9

Eisenstein, F., Yanagisawa, H., Kashihara, H., Kikkawa, M., Tsukita, S., & Danev, R. (2023). Parallel cryo electron tomography on in situ lamellae. *Nature Methods*, *20*(1), 131–138. https://doi.org/10.1038/s41592-022-01690-1

Faini, M., Prinz, S., Beck, R., Schorb, M., Riches, J. D., Bacia, K., Brügger, B., Wieland, F. T., & Briggs, J. A. G. (2012). The structures of COPI-coated vesicles reveal alternate coatomer conformations and interactions. *Science*, *336*(6087), 1451–1454. https://doi.org/10.1126/science.1221443

Fan, H., & Tan, Y. (2024). Lipid Droplet–Mitochondria Contacts in Health and Disease. In *International Journal of Molecular Sciences* (Vol. 25, Issue 13). Multidisciplinary Digital Publishing Institute (MDPI). https://doi.org/10.3390/ijms25136878

Fäßler, F., Zens, B., Hauschild, R., & Schur, F. K. M. (2020). 3D printed cell culture grid holders for improved cellular specimen preparation in cryo-electron microscopy. *Journal of Structural Biology*, *212*(3). https://doi.org/10.1016/j.jsb.2020.107633

Förster, F., Han, B. G., & Beck, M. (2010). Visual Proteomics. In *Methods in Enzymology* (Vol. 483, Issue C, pp. 215–243). Academic Press Inc. https://doi.org/10.1016/S0076-6879(10)83011-3

Foster, H. E., Santos, C. V., & Carter, A. P. (2022). A cryo-ET survey of microtubules and intracellular compartments in mammalian axons. *Journal of Cell Biology*, *221*(2). https://doi.org/10.1083/jcb.202103154

Fotin, A., Kirchhausen, T., Grigorieff, N., Harrison, S. C., Walz, T., & Cheng, Y. (2006). Structure determination of clathrin coats to subnanometer resolution by single particle cryo-electron microscopy. *Journal of Structural Biology*, *156*(3), 453–460. https://doi.org/10.1016/j.jsb.2006.07.001

Fry, M. Y., Navarro, P. P., Hakim, P., Ananda, V. Y., Qin, X., Landoni, J. C., Rath, S., Inde, Z., Lugo, C. M., Luce, B. E., Ge, Y., Mcdonald, J. L., Ali, I., Ha, L. L.,

Kleinstiver, B. P., Chan, D. C., Sarosiek, K. A., & Chao, L. H. (2024). *In situ*
architecture of Opa1-dependent mitochondrial cristae remodeling.
https://doi.org/10.1101/2023.01.16.524176

Fujimoto, T., & Parton, R. G. (2011). Not just fat: The structure and function of the
lipid droplet. *Cold Spring Harbor Perspectives in Biology*, *3*(3), 1–17.
https://doi.org/10.1101/cshperspect.a004838

Ganeva, I., Lim, K., Boulanger, J., Hoffmann, P. C., Muriel, O., Borgeaud, A. C.,
Hagen, W. J. H., Savage, D. B., & Kukulski, W. (2023). The architecture of
Cidec-mediated interfaces between lipid droplets. *Cell Reports*, *42*(2).
https://doi.org/10.1016/j.celrep.2023.112107

Gross, S. P., Welte, M. A., Block, S. M., & Wieschaus, E. F. (2000). Dynein-mediated
Cargo Transport In Vivo: A Switch Controls Travel Distance. In *The Journal*
of Cell Biology (Vol. 148, Issue 5). http://www.jcb.org

Herzik, M. A., Wu, M., & Lander, G. C. (2017). Achieving better-than-3-Å
resolution by single-particle cryo-EM at 200 keV. *Nature Methods*, *14*(11),
1075–1078. https://doi.org/10.1038/nmeth.4461

Hugenroth, M., & Bohnert, M. (2020). Come a little bit closer! Lipid droplet-ER
contact sites are getting crowded. In *Biochimica et Biophysica Acta -*
Molecular Cell Research (Vol. 1867, Issue 2). Elsevier B.V.
https://doi.org/10.1016/j.bbamcr.2019.118603

Kilwein, M. D., & Welte, M. A. (2019). Lipid droplet motility and organelle
contacts. *Contact (Thousand Oaks)*, *585*, 276–296.

Kremer, J. R., Mastronarde, D. N., & Mcintosh, J. R. (1996). *Computer Visualization*
of Three-Dimensional Image Data Using IMOD.

Kühlbrandt, W. (2014). The resolution revolution. In *Science* (Vol. 343, Issue
6178, pp. 1443–1444). American Association for the Advancement of
Science. https://doi.org/10.1126/science.1251652

Mahamid, J., Tegunov, D., Maiser, A., Arnold, J., Leonhardt, H., Plitzko, J. M., &
Baumeister, W. (2019). Liquid-crystalline phase transitions in lipid droplets
are related to cellular states and specific organelle association. *Proceedings*
of the National Academy of Sciences of the United States of America, *116*(34),
16866–16871. https://doi.org/10.1073/pnas.1903642116

Martynowycz, M. W., Clabbers, M. T. B., Unge, J., Hattne, J., & Gonen, T. (2021).
Benchmarking the ideal sample thickness in cryo-EM. *118*, 2108884118.
https://doi.org/10.1073/pnas.2108884118/-/DCSupplemental

McMullan, G. I., Naydenova, K. I., Mihaylov, D., Yamashita, K. I., Peet ID, M. J.,
Wilson, H. I., Dickerson ID, J. L., Chen, S. I., Cannone, G. I., Lee, Y. I., Hutchings
ID, K. A., Gittins, O. I., Sobhy ID, M. A., Wells, T. I., El-Gomati ID, M. M., Dalby,
J., Meffert, M. I., Schulze-Briese, C. I., Henderson, R. I., & Russo, C. J. (2023).

Structure determination by cryoEM at 100 keV.
https://doi.org/10.1073/pnas

Medalia, O., Weber, I., Frangakis, A. S., Nicastro, D., Gerisch, G., & Baumeister, W.
(2002). Macromolecular Architecture in Eukaryotic Cells Visualized by
Cryoelectron Tomography. *Science*, *5596*, 1209–1213.
https://doi.org/DOI:10.1126/science.1076184

Navaratnarajah, T., Anand, R., Reichert, A. S., & Distelmaier, F. (2021). The
relevance of mitochondrial morphology for human disease. In *International
Journal of Biochemistry and Cell Biology* (Vol. 134). Elsevier Ltd.
https://doi.org/10.1016/j.biocel.2021.105951

Noble, A. J. (2024). *HighEnd CryoEM Worldwide.*

Nogales, E., & Mahamid, J. (2024). Bridging structural and cell biology with cryo-
electron microscopy. In *Nature* (Vol. 628, Issue 8006, pp. 47–56). Nature
Research. https://doi.org/10.1038/s41586-024-07198-2

Nonoyama, T., Nojima, D., Maeda, Y., Noda, M., Yoshino, T., Matsumoto, M.,
Bowler, C., & Tanaka, T. (2019). Proteomics analysis of lipid droplets
indicates involvement of membrane trafficking proteins in lipid droplet
breakdown in the oleaginous diatom Fistulifera solaris. *Algal Research*, *44*.
https://doi.org/10.1016/j.algal.2019.101660

Olarte, M. J., Swanson, J. M. J., Walther, T. C., & Farese, R. V. (2022). The CYTOLD
and ERTOLD pathways for lipid droplet–protein targeting. In *Trends in
Biochemical Sciences* (Vol. 47, Issue 1, pp. 39–51). Elsevier Ltd.
https://doi.org/10.1016/j.tibs.2021.08.007

Olzmann, J. A., & Carvalho, P. (2019). Dynamics and functions of lipid droplets. In
Nature Reviews Molecular Cell Biology (Vol. 20, Issue 3, pp. 137–155). Nature
Publishing Group. https://doi.org/10.1038/s41580-018-0085-z

Punjani, A., & Fleet, D. J. (2023). 3DFlex: determining structure and motion of
flexible proteins from cryo-EM. *Nature Methods*, *20*(6), 860–870.
https://doi.org/10.1038/s41592-023-01853-8

Rice, W. J., Cheng, A., Noble, A. J., Eng, E. T., Kim, L. Y., Carragher, B., & Potter, C. S.
(2018). Routine determination of ice thickness for cryo-EM grids. *Journal of
Structural Biology*, *204*(1), 38–44.
https://doi.org/10.1016/j.jsb.2018.06.007

Rigort, A., & Plitzko, J. M. (2015). Cryo-focused-ion-beam applications in
structural biology. In *Archives of Biochemistry and Biophysics* (Vol. 581, pp.
122–130). Academic Press Inc. https://doi.org/10.1016/j.abb.2015.02.009

Robinson, M. S., Bykov, Y. S., Schaffer, M., Dodonova, S. O., Albert, S., rgen Plitzko,
J. M., Baumeister, W., Engel, B. D., & Briggs, J. A. (2017). *The structure of the
COPI coat determined within the cell.*
https://doi.org/10.7554/eLife.32493.001

Rohou, A., & Grigorieff, N. (2015). CTFFIND4: Fast and accurate defocus estimation from electron micrographs. *Journal of Structural Biology*, *192*(2), 216–221. https://doi.org/10.1016/j.jsb.2015.08.008

Russo, C. J., Dickerson, J. L., & Naydenova, K. (2022). Cryomicroscopy in situ: what is the smallest molecule that can be directly identified without labels in a cell? *Faraday Discussions*, *240*, 277–302. https://doi.org/10.1039/d2fd00076h

Schwab, J., Kimanius, D., Burt, A., Dendooven, T., & Scheres, S. H. W. (2024). DynaMight: estimating molecular motions with improved reconstruction from cryo-EM images. *Nature Methods*. https://doi.org/10.1038/s41592-024-02377-5

Szwedziak, P., & Pilhofer, M. (2019). Bidirectional contraction of a type six secretion system. *Nature Communications*, *10*(1). https://doi.org/10.1038/s41467-019-09603-1

Szwedziak, P., Wang, Q., Bharat, T. A., Tsim, M., & Löwe, J. (2014). Architecture of the ring formed by the tubulin homologue FtsZ in bacterial cell division. *ELife*, *3*. https://doi.org/10.7554/eLife.04601

Tegunov, D., Xue, L., Dienemann, C., Cramer, P., & Mahamid, J. (2021). Multi-particle cryo-EM refinement with M visualizes ribosome-antibiotic complex at 3.5 Å in cells. *Nature Methods*, *18*(2), 186–193. https://doi.org/10.1038/s41592-020-01054-7

Thiam, A. R., Antonny, B., Wang, J., Delacotte, J., Wilfling, F., Walther, T. C., Beck, R., Rothman, J. E., & Pincet, F. (2013). COPI buds 60-nm lipid droplets from reconstituted water-phospholipid- triacylglyceride interfaces, suggesting a tension clamp function. *Proceedings of the National Academy of Sciences of the United States of America*, *110*(33), 13244–13249. https://doi.org/10.1073/pnas.1307685110

Tian, Y., Liang, R., Kumar, A., Szwedziak, P., & Viles, J. H. (2020). 3D-visualization of amyloid-β oligomer and fibril interactions with lipid membranes by cryo-electron tomography. In *bioRxiv*. https://doi.org/10.1101/2020.07.21.214072

Träger, T. K., Kyrilis, F. L., Hamdi, F., Tüting, C., Alfes, M., Hofmann, T., Schmidt, C., & Kastritis, P. L. (2024). Disorder-to-order active site capping regulates the rate-limiting step of the inositol pathway. *Proceedings of the National Academy of Sciences of the United States of America*, *121*(34). https://doi.org/10.1073/pnas.2400912121

Valm, A. M., Cohen, S., Legant, W. R., Melunis, J., Hershberg, U., Wait, E., Cohen, A. R., Davidson, M. W., Betzig, E., & Lippincott-Schwartz, J. (2017). Applying systems-level spectral imaging and analysis to reveal the organelle interactome. In *Nature* (Vol. 546, Issue 7656, pp. 162–167). Nature Publishing Group. https://doi.org/10.1038/nature22369

Wilfling, F., Thiam, A. R., Olarte, M. J., Wang, J., Beck, R., Gould, T. J., Allgeyer, E. S.,
Pincet, F., Bewersdorf, J., Farese, R. V., & Walther, T. C. (2014). Arf1/COPI
machinery acts directly on lipid droplets and enables their connection to the
ER for protein targeting. *ELife*, *2014*(3).
https://doi.org/10.7554/eLife.01607

Wilfling, F., Wang, H., Haas, J. T., Krahmer, N., Gould, T. J., Uchida, A., Cheng, J. X.,
Graham, M., Christiano, R., Fröhlich, F., Liu, X., Buhman, K. K., Coleman, R. A.,
Bewersdorf, J., Farese, R. V., & Walther, T. C. (2013). Triacylglycerol
synthesis enzymes mediate lipid droplet growth by relocalizing from the ER
to lipid droplets. *Developmental Cell*, *24*(4), 384–399.
https://doi.org/10.1016/j.devcel.2013.01.013

Wolff, G., Limpens, R. W. A. L., Zevenhoven-Dobbe, J. C., Laugks, U., Zheng, S., De
Jong, A. W. M., Koning, R. I., Agard, D. A., Grünewald, K., Koster, A. J., Snijder,
E. J., & Bárcena, M. (2020). A molecular pore spans the double membrane of
the coronavirus replication organelle. *Science*, 1395–1398.
https://doi.org/10.1126/science.abd3629

Wozny, M. R., Di Luca, A., Morado, D. R., Picco, A., Khaddaj, R., Campomanes, P.,
Ivanović, L., Hoffmann, P. C., Miller, E. A., Vanni, S., & Kukulski, W. (2023). In
situ architecture of the ER–mitochondria encounter structure. *Nature*,
618(7963), 188–192. https://doi.org/10.1038/s41586-023-06050-3

Wu, G. H., Smith-Geater, C., Galaz-Montoya, J. G., Gu, Y., Gupte, S. R., Aviner, R.,
Mitchell, P. G., Hsu, J., Miramontes, R., Wang, K. Q., Geller, N. R., Hou, C.,
Danita, C., Joubert, L. M., Schmid, M. F., Yeung, S., Frydman, J., Mobley, W.,
Wu, C., … Chiu, W. (2023). CryoET reveals organelle phenotypes in
huntington disease patient iPSC-derived and mouse primary neurons.
Nature Communications, *14*(1). https://doi.org/10.1038/s41467-023-
36096-w

Zhang, S., Peng, X., Yang, S., Li, X., Huang, M., Wei, S., Liu, J., He, G., Zheng, H., Yang,
L., Li, H., & Fan, Q. (2022). The regulation, function, and role of lipophagy, a
form of selective autophagy, in metabolic disorders. In *Cell Death and
Disease* (Vol. 13, Issue 2). Springer Nature.
https://doi.org/10.1038/s41419-022-04593-3

Ziemianowicz, D. S., & Kosinski, J. (2022). New opportunities in integrative
structural modeling. In *Current Opinion in Structural Biology* (Vol. 77).
Elsevier Ltd. https://doi.org/10.1016/j.sbi.2022.102488

Figure 1

b)

Data collection summary	
#imaged grids	9
#imaged cells	64
#tomograms	228
average thickness	240±75 nm

Figure 2

Figure 3

Figure 4

a)

	SPA cryoEM		cryoET		Others	
Year	Glacios	Krios	Glacios	Krios	Glacios	Krios
2020	24	2519	0	243	0	112
2021	50	3239	0	260	1	165
2022	149	4233	0	449	3	248
2023	318	5379	4	837	18	333

EMDB entries by cryoTEM instrument and modality

b)

c)

d)

1º **Local cryoEM facilities**
200 kV instruments suitable for SPA cryoEM and cryoET data collection and screening

2º **National/international cryoEM hubs**
300 kV cryoET purpose-built instruments fitted with laser phase plate/Cc corrector/liquid helium cryostage.
Complex cryoET sample prep (HPF, cryo-pFIB-SEM with lift-out)

Figure S1

Figure S2

Figure S3

Figure S4

Figure S5

Figure S6