Fast in vivo deep-tissue 3D imaging with selective-illumination NIR-II light-field microscopy and aberration-corrected implicit neural representation

Fenghe Zhong[1,4], Xue Li[2,4], Mian He[1,4], Yihang Huang[1], Chengqiang Yi[1], Shiqi Mao[1], Xin Huang[1], Kui Ren[2], Miaomiao Kang[2], Dong Wang[2], Zhijun Zhang[2*], Dongyu Li[1,3*], Peng Fei[1,3*]

[1] *School of Optical and Electronic Information-Wuhan National Laboratory for Optoelectronics, Huazhong University of Science and Technology, 430074, Wuhan, China*

[2] *Center for AIE Research, Guangdong Provincial Key Laboratory of New Energy Materials Service Safety, College of Materials Science and Engineering, Shenzhen University, Shenzhen 518060, China*

[3] *Advanced Biomedical Imaging Facility, Huazhong University of Science and Technology, 430074, Wuhan, China*

[4] *These authors contributed equally: Fenghe Zhong, Xue Li, Mian He*

Correspondence: feipeng@hust.edu.cn; li_dongyu@hust.edu.cn; zhijunzhang@szu.edu.cn

Abstract

Near-infrared II (NIR-II) microscopy, which enables *in vivo* deep-tissue visualization of vasculature and cell activities, has been a promising tool for understanding physiological mechanisms. However, the volumetric image speed of the current NIR-II microscopy is hindered by scanning strategy, causing limitations for observing instantaneous biological dynamics in 3D space. Here, we developed a NIR-II light-field microscopy (LFM) based on selective illumination and self-supervised implicit neural representation (INR)-reconstruction, which allows ultra-fast 3D imaging (20 volumes/s) in deep tissue. Through integrating INR with view-wise aberration correction, our strategy could conquer the artifacts induced by the angular subsampling and refractive index variation problems, achieving single-cell resolution at a reconstruction volume of 550 μm diameter and 200 μm thickness. The volumetric selective illumination overcomes the influence of out-of-focus background on reconstruction, together with the low scattering advantage of NIR-II wavelength, extending the imaging depth to 600 μm. The developed aberration-corrected implicit neural representation reconstruction (AIR) NIR-II LFM showcases its capability by monitoring hemodynamics of mouse brain under norepinephrine and flow redistribution of ischemic stroke in 3D vasoganglion, as well as noninvasively tracking immune cell activities inside subcutaneous solid tumor through intact skin. This approach represents a significant advancement in 3D *in vivo* imaging, holding great potential in both biomedical research and preclinical studies.

Introduction

In vivo optical microscopy has revolutionized biomedical research by enabling real-time visualization of cellular and even subcellular events in living animals, providing valuable insights on physiological mechanisms, disease pathogenesis, and therapeutic interventions. The continuous development of cutting-edge biomedical research demands that *in vivo* imaging technologies possess the capabilities of observing clearly, quickly, and deeply in 3D, posing a significant challenge due to the complex environment in the living body. A series of imaging modalities that could conquer the server tissue scattering have been intensively developed for understanding *in vivo* activities. Multiphoton microscopy achieved high-resolution at large depth using near-infrared (NIR) excitation light with less scattering and the corresponding point scan mechanism[1–3]. Photoacoustic microscopy, taking advantage of the much longer wavelength of ultrasound waves and its low transmission speed, achieves deep tissue volumetric mapping of vasculature in live animals[4–6]. As an emerging technique, NIR-II (900-1880 nm) fluorescence imaging explosives in biomedical research due to its long scattering length and remarkable penetration depth[7–10]. At the microscopic level, researchers realized optical sectioning from the cortical surface to 800 μm with long fluorescent wavelength at the NIR-II region and the confocal mechanism[11–13]. However, the above point-scanning-based acquisition strategies significantly degrade the temporal resolution, limiting their applications in highly dynamic life activities such as hemodynamics, especially for 3D monitoring.

Multiple approaches have been devised to reduce the scanning times to enhance the volumetric acquisition speed. Multi-focus two-photon microscopy splits a single excitation pulse into 30 spatially and temporally isolated pulses, enhancing the scanning speed by 30 times and achieving a volumetric imaging speed of 2 Hz[14]. By employing a spatial light modulator, temporal focusing technique excites multiple regions of interest simultaneously rather than sequentially, and captures images of different areas using one high-speed sCMOS camera, markedly boosting the image speed in 3D space[15]. Nevertheless, the severe scattering of visible light significantly compromises imaging depth. NIR-II light-sheet microscopy[16,17], which allows for selective-plane illumination and acquisition, achieves a high frame rate with large imaging depth owning to the long wavelength. However, its volumetric imaging speed remains constrained by the inherent scanning-based strategy.

Light-field microscopy (LFM), as a rising star in biological research[18–20], offers unique advantages for monitoring subcellular interactions and recording neuronal activities with high temporal resolution. By capturing projection images from multiple angles in a single shot, LFM facilitates volumetric reconstruction, achieving ultra-high temporal resolution 3D imaging[18,21]. The integration of NIR-II microscopy with LFM holds great potential for large-volume high-speed imaging of hemodynamics, tumor immunization, and neuronal activities. However, several major challenges inhibit the full integration of these techniques for *in vivo* deep-tissue imaging. Firstly, the crosstalk fluorescence from out-of-focus areas poses a significant issue[22,23], particularly in thick organs such as the brain and tumors, where the fluorescence collected from beyond the focal zone can compromise accuracy and introduce

artifacts. Additionally, besides the angular subsampling problem in LFM, which causes artefacts in 3D

reconstruction[24,25], the limited angle range recorded with low NA objectives, often required for *in vivo*

observations to achieve a sufficient field of view, exacerbates the missing cone problem[26,27] and further

impacts reconstruction accuracy. Furthermore, non-uniform distributions of refractive index in deep-

tissue generate spatially varying dynamic aberrations that impair imaging performance[28]. Aberration-

induced degradation is particularly detrimental to LFM, as the cumulative effects of aberrations from

multiple views aggregate. Although hardware-based adaptive optics have demonstrated efficacy in

correcting aberrations in microscopic imaging[29,30], they are constrained by computational time and

device refresh rates, resulting in correction times of several dozen seconds prior to each snapshot. This

limits their applicability in high-speed LFM. Recently, a high-speed digital adaptive optics (DAO)

algorithm tailored for LFM has been introduced[31]. This algorithm compensates for multi-angle

aberrations through post-processing, thereby enhancing the reconstructed spatial resolution on *in vitro*

cells and diverse tissue and organ surfaces. However, DAO simplifies the aberration model as linear

translations in the light-field phase domain, with a solution space restricted to just two parameters. This

approximation hinders its capability to address more intricate aberrations, thereby limiting its utility in

deeper and more complex *in vivo* imaging contexts.

It should be also noted that, the ultra-high imaging speed relies on an ultra-limited exposure time,

presenting a significant challenge to the brightness of the fluorescent probe. Unfortunately, the rarity of

NIR-II fluorophores with high quantum yield (QY) complicates the development of efficient NIR-II

LFM.

Here, we propose a NIR-II LFM for the first time, based on selective illumination and aberration

corrected implicit neural representation (INR)-reconstruction, capable of deep-tissue volumetric

imaging in living mice at video rate and cellular resolution. To meet the stringent requirements for ultra-

short exposure times in ultra-high-speed imaging, we developed a novel NIR-II aggregation-induced

emission luminogen (AIEgen), named TTT8,4-B, which exhibits a high fluorescence QY of 0.94%

(using the standard NIR-II dye IR-26 as a reference) when formulated into nanoparticles (NPs). With

one of the highest QY values among all reported NIR-II AIEgens, TTT8,4-B NPs help to enable

volumetric imaging at an unprecedented speed of 20 Hz. To eliminate out-of-focus fluorescence and

enhance the 3D reconstruction quality, we employ selective volumetric illumination achieved through

rapid scanning of a light-sheet within the duration of a single snapshot. Instead of traditional

deconvolution-based reconstruction methods, we developed an AIR network. Taking advantage of the

continuity of the INR function and geometry consistency of synthesized views, our approach

significantly reduces the impact of the angular subsampling problem and missing cone problem, and

realizes higher axial resolution and fewer artifacts. Further integrated with a self-supervised view-wise

aberration correction module in the network, our approach achieves minimized artifacts after

reconstruction of a large volume (550 μm diameter and 200 μm thickness) with only 9 views. Our AIR

NIR-II LFM facilitated the observation of 3D hemodynamics in the mouse brain due to its video-rate

imaging speed and promising imaging quality, enabling us to analyze the response of the vasoganglion

to clinical pressor norepinephrine and ischemic stroke. Furthermore, we functionalized TTT8,4-B NPs

with specific antibodies to target immune cells *in vivo*, and utilized our imaging approach to achieve,

for the first time, *in vivo* 3D tracking of immune cells in the tumor.

Results

The principle of AIR selective-illumination NIR-II LFM

Interrogating the fast 3D dynamics in live animals is crucial for revealing biological mechanisms[32,33].

However, such complex imaging condition brought by light scattering and attenuation rises the

challenges on the volumetric imaging speed, imaging depth and anti-aberration ability of microscopes.

These metrics entangled together manifested as the tradeoff of imaging quality and imaging speed in

scanning-based microscope[34,35].

Here, to capture the fast 3D time-varied signals, we proposed a NIR-II light field microscopy with

selective volume illumination and aberration-corrected INR reconstruction algorithm to divide-and-

conquered this problem by optimizing the signal generation, optical encoding, and algorithm's decoding.

To achieve ultra-fast *in vivo* deep-tissue 3D imaging using the aforementioned selective-

illumination NIR-II light-field microscopy, NIR-II fluorescent probes with high fluorescence brightness

are indispensable, as ultra-high-speed imaging demands ultra-short exposure times. Motivated by the

bright fluorescence that AIEgens typically exhibit in an aggregated state, we focus on designing NIR-

II AIEgens with high emission intensity. The excellent emission behavior of AIEgens is attributed to

their unique twisted structures, which effectively extend molecular packing distances in the aggregated

state, thereby significantly reducing undesirable intermolecular interactions and preserving more

excited energy for fluorescent emission[36–40]. As shown in Fig. 1a, the typical triphenylamine (TPA) unit

as well as the attached triphenylethylene moiety were introduced. Their inherently distorted

conformation is conducive to the AIE feature and luminescence efficiency of the final compound

TTT8,4-B. Beyond that, the secondary branched carbon alkyl chain was also introduced to further twist

the molecular skeleton[41] and provide spatial separation of TTT8,4-B in the aggregated state, which

finally helps to elevate the fluorescence QY. In addition to serving as a twisted building block, the

electron-rich TPA group also functions as an effective electron donor (D), which together with the

thiophene unit (working as the second D and π-bridge) constitutes a robust electron-donating part[42,43].

On the other hand, the electron-deficient benzobisthiadiazole (BBTD) was designed as the powerful

electron acceptor (A)[44] to endow the final TTT8,4-B with intensive electron D–A interaction. Strong

D–A interaction as well as the large π-conjugation are expected to remarkably reduce the electronic

bandgap of TTT8,4-B, thus potentially extending the emission wavelength to the NIR-II region.

In order to facilitate the *in vivo* application, the hydrophobic TTT8,4-B was formulated into

TTT8,4-B NPs *via* the nanoprecipitation method[45] using amphiphilic copolymers (PS5000-mPEG5000)
as an encapsulation matrix. The particle size of TTT8,4-B NPs was determined to be approximately 52
nm as measured by dynamic light scattering (DLS) (Fig. 1b). Transmission electron microscopy (TEM,
Fig. 1b, inset) revealed that the particles were consistently spherical, with an average diameter ranging
from 50 to 60 nm, corroborating the findings obtained from DLS. Compared to the absorption observed
in the THF solution, TTT8,4-B NPs exhibited a notable red-shift in the maximum absorption, reaching
732 nm, which is advantageous for deep tissue excitation and minimizing photodamage to biological
systems[46,47]. Additionally, TTT8,4-B NPs demonstrated intense fluorescence emission within the 800-
1300 nm wavelength range, with a distinct emission peak at 1000 nm, highlighting their potential as a
superior NIR-II imaging probe (Fig. 1c). The relative QY of TTT8,4-B NPs reached 0.94% under 808
nm excitation, which is 18.8 times higher than that of IR-26, a standard NIR-II reference dye with a QY
of 0.05% in 1,2-dichloroethane (Fig. 1d and Supplementary Fig. 1). Last but not least, the biosecurity
of TTT8,4-B NPs was comprehensively evaluated. The cytotoxicity assessment on normal cell lines
and hemolysis assay (Supplementary Fig. 2) indicated their excellent biocompatibility *in vitro*. The *in*
vivo biosafety of TTT8,4-B NPs were also substantiated by the non-differential physiological
parameters, including body weight (Supplementary Fig. 3), hepatic function markers, renal function
markers (Supplementary Fig. 4), blood routine parameters (Supplementary Table 1), and the
hematoxylin and eosin (H&E) staining results of major organs (Supplementary Fig. 5), between
TTT8,4-B NPs and PBS treated mice. Collectively, these results substantiate the capability of TTT8,4-
B NPs as a promising candidate for ultra-fast NIR-II fluorescence imaging *in vivo*.

To efficiently collect the 3D information in living animals in a snapshot, one plausible approach
lies in the development of light-field microscopy[48]. However, this method alone suffered from the cross-
talk due to the out-focus signals in thick tissue. Sequential scanning-based approach[49] could solve this
problem but sacrificed the volumetric imaging speed severely. Taking both into consideration, we have
developed the selective illumination-armed Fourier light field microscopy (FLFM, Light field encoding
in Fig. 1e, Fig. 1f and Supplementary Fig. 6). This innovative technique rapidly scans the imaging
volume with light-sheet excitation in one light-filed imaging duration, thereby achieving both the high-
efficiency volumetric encoding and out-of-focus crosstalk minimization. This improved excitation
strategy benefited the signal to background ratio (SBR) and weak signal detection (Fig. 1h
Supplementary Fig. 7). To generate the selective volume illumination, a cylindrical lens is employed to
generate a light sheet, which is subsequently scanned by a galvanometer. With the aim of observing
living organisms at single-cell resolution, we customized a microlens array with square arrangement to
fully split the optical aperture into 9 FLFM views, which has an effective NA of 0.14 for each microlens
(Methods), producing 4.27 μm × 9.5 μm spatial resolution and ~200 μm depth of field (DOF), and
enables capturing the vascular dynamics in the mouse ear, iris, brain and tumor (Supplementary Fig. 8).
The Rayleigh range of the light sheet was designed as 69 μm to maximal the remove of the crosstalk
out of the 200 μm DOF while maintaining an acceptable signal amplitude in the region of interest. We
characterized NIR-II LFM by imaging fluorescent beads embedded in agarose gel. The system can

achieve a spatial resolution of 4.7 μm × 11 μm (Supplementary Fig. 9).

After acquiring the enhanced light-field captures, accurately decoding the 3D signals from the

encoded 2D multi-view projections remains a significant challenge for existing light-field

reconstruction algorithms[18]. Besides of the information compression in light-field encoding, which

cause a severer resolution degradation, the presence of optical aberrations, derived from the dynamic

refractive index inhomogeneities in *in vivo* deep-tissue imaging, further complicates the 3D

reconstruction inverse problem, manifested as the view-wise wavefront distortion[31]. While supervised

learning approaches[18] would achieve resolution improvement, they suffered the generalization errors

and struggle to build an accurate degradation model especially under this unknown complex aberration

(varied with time and view-angle). In contrast, classic-model based approach[31] mitigated the aberration

without prior knowledge, but remained limited by their simplified aberration models.

Here, we employ an AIR network, which contains an INR reconstruction module and a view-wise

Aberration prediction (VAP) module, to reduce the artifacts raised by sparse angular sampling rate and

aberrations in *in vivo* imaging (AIR-network in Fig. 1e and Fig. 1g). First, a 3D volume was generated

with coordinates input during Multi-Layer Perceptron (MLP) inference. The wave-optics model was

then introduced to project the 3D volume onto a 2D estimated light-field map which act as a constrain.

To obtain an accurate forward model, the VAP module incorporates learnable Zernike coefficients into

light-field PSF to estimate the view-wise aberrations (Supplementary Note 1). The estimated phase of

different views in Fig. 1g described the wavefront distortion from objective space (biological sample)

to imaging system (FLFM system) in an end-to-end manner, achieving higher fidelity of PSF estimation

(Supplementary Fig. 10) and 3D reconstruction than conventional aberration-correction approach

(Supplementary Fig. 11). By calculating the loss between the LF projections with estimated aberration

and the raw LF captures, the INR network could gradually remove the artifacts induced by aberrations

in 3D reconstruction (Supplementary Fig. 12). Through offline processing, our method could estimate

the phase of each FLFM view and produce accurate reconstructions during *in vivo* FLFM imaging (Fig.

1g). Notably, our approach does not rely on hardware correction devices or additional prior information,

offering a highly efficient and accessible solution for aberration correction and 3D reconstruction

especially in acute biological dynamics.

We then demonstrated the performance advances of AIR through a comparison with DAO and

conventional deconvolution-based reconstruction method on in vivo mouse brain vasculature. As shown

in Fig. 1i and Supplementary Movie 1, in deconvolution-based reconstruction image, the reconstructed

vessels have severe artifacts, as well as elongations along the axis. When DAO is used in FLFM, it can

alleviate the problem of artifacts to a certain extent. However, there are still vascular discontinuities due

to signal loss. Also, the DAO does not solve the axial elongation caused by the missing cone problem.

In sharp contrast, AIR reconstructs more details both horizontally and axially. Furthermore, due to the

intrinsic advantage of continuity of the INR function and geometry consistency of synthesized views,

our method reduces the axial elongation while reconstructing 3D image.

Fig. 1. The NIR-II LFM based on selective illumination and Aberration-corrected INR Reconstruction (AIR). a. Chemical structure of TTT8,4-B and the fabrication of TTT8,4-B NPs. **b.** Size distribution and TEM image of TTT8,4-B NPs. **c.** Absorption and emission spectra of TTT8,4-B NPs. **d.** The QY of TTT8,4-B NPs relative to IR-26. **e.** The schematic of our AIR NIR-II LFM. **f.** Schematic of the light field encoding of selective illumination NIR-II LFM. A cylindrical lens is employed to generate a light sheet, which is subsequently scanned by a galvanometer to enable selective-volume illumination. The excited fluorescence signal with limited crosstalk is modulated by microlens array and subsequently captured by a high-sensitivity camera. **g.** The algorithm decoding process of AIR which contains INR-reconstruction, VAP module and loss computation. **h.** The xy and xz projected mouse brain vasculature and the corresponding zoom in reconstructed with selective illumination and without selective illumination. Right bottom: The cross-sectional profiles of the dashed lines. Scale bar: 100 μm (left), 50 μm (right). **i.** The comparison of reconstructions by deconvolution, DAO and AIR. The white boxed zoom in (right) indicates the noticeable errors in deconvolution and DAO based reconstruction

232 result which is accurately resolved by the AIR method. The blue boxed region indicates the region of the xz

233 projection (bottom). Scale bar: 100 μm (left, right).

NIR-II and visible LFM comparison in phantom

To showcase the capability of our AIR NIR-II LFM in obtaining high-resolution volumetric images in

deep tissue, we prepared leaf phantom immersed in lipids, mimicking the tissue environment (Fig. 2a).

The phantom was stained with a mixture of Rhodamine B and TTT8,4-B NPs solution. Imaging of the

phantom was conducted using both visible LFM (532 nm) and NIR-II LFM, through lipid agarose gel

layers with varying thicknesses. The results reveal that NIR-II LFM can clearly resolve the intricate leaf

vasculature from the surface down to a depth of 500 μm under lipids (Fig. 2b). In contrast, the quality

of vasculature reconstruction in visible LFM significantly deteriorates with the addition of just a 100

μm lipid layer, and the vessels became indistinguishable with artifacts when the lipid layer reaches 200

μm (Fig. 2c). We further quantified the structure similarity index measurement (SSIM) between the

surface MIP images and the deeper layer images (Fig. 2d). Compared to visible group, NIR-II group

exhibits consistent SSIM values, highlighting the deep penetration and reconstruction capabilities of

NIR-II LFM in scattering tissue. Conversely, the SSIM of images captured by visible LFM decreases

markedly.

NIR-II and visible LFM *in vivo* mouse brain comparison

The unique advantage of our NIR-II LFM over traditional visible LFM was further validated with *in*

vivo mouse brain imaging. We performed a side-by-side comparison between the two imaging

modalities in the same region. As shown in Fig. 2e, the NIR-II LFM achieved high imaging quality with

little resolution loss at different depths (spectrograms in the right bottom), while the resolution of visible

microscopy degraded significantly from 5.5 μm to 8.8 μm. Cross-sectional profiles in Fig. 2f show that

there are microvessel missing using the visible LFM due to the short scattering length of visible light

and the low signal-to-background ratio (SBR) of the acquired images.

Fig. 2. Visible and NIR-II LFM comparison on leaf phantom and *in vivo* mouse brain. a. The procedure of leaf phantom preparation. **b.** TTT8,4-B NPs and Rhodamine B co-dyed leaf vasculature imaging with NIR-II LFM. Scale bar: 50 μm. **c.** The leaf vasculature imaging with the visible LFM. Scale bar: 50 μm. **d.** The degradation of SSIM while imaging with different depths of lipid phantom covered. **e.** Visible and NIR-II co-registered *in vivo* mouse brain imaging. Scale bar: 50 μm. **f.** The cross-sectional profiles of the dashed white lines in **e.**

Deep tissue hemodynamic imaging

Fig. 3a presents a mosaic of partially overlapping reconstruction volumes, seamlessly integrating to form a continuous vasculature map that rivals the imaging depth capabilities of other scanning-based imaging techniques[50]. The vasculature is distinctly resolved, spanning from the cortical surface down to a depth of 600 μm. Fig. 3b illustrates the vessels at three distinct depths (surface, 300 μm, and 600 μm), with the depth sectioning further demonstrated in Supplementary Movie 2.

Non-invasive capture of hemodynamics and blood flow *in vivo* offers profound insights into neuroscience and cancer studies that highly relate to metabolism. Our AIR NIR-II LFM, as a high temporal-resolution volumetric imaging tool, presents an accurate method for blood flow quantification in 3D spaces. Fig. 3c showcases the targeted region within the volume image, with the dashed line highlighting the vessel of interest, and its magnified view presented in Fig. 3d. Operating a high volumetric imaging rate at 20 Hz (550 μm diameter and 200 μm thickness), a flowing particle cluster

(probably TTT8,4-B NPs conjugated with hemoglobin or other biomacromolecules in blood) can be resolved by tracking successive volumes. The MIP images clearly illustrate volumetric movement in both lateral and axial directions. Notably, the high reconstruction quality allows for the analysis of each vessel branch's depth and penetrating direction (Fig. 3e). The time-lapse 3D path of the flowing target could be extracted and presented in Fig. 3f. The incorporation of depth information enhances the accuracy of flow velocity analysis compared to conventional 2D kymograph analysis methods[51]. The differences in the flow velocities quantified by these two methods are illustrated in Fig. 3h. With the depth-incorporated approach, the flow velocities of selected vessels were quantified and shown in the color-encoded image in Fig. 3i.

Fig. 3. The performance of AIR NIR-II LFM. a. The volume rendering of the *in vivo* mouse brain vasculature from the cortical surface to 600 μm. Scale bar: 100 μm. **b.** The vasculature at the motor cortex surface, 300 μm and 600 μm depth. Scale bar: 100 μm. **c.** The selected region for high-speed monitoring. Scale bar: 100 μm. **d.** Time-lapse xy (top) and xz (bottom) MIP images of the vessel outlined by the white dashed line in **c**. The blue arrows indicate the location of a flowing particle cluster. Scale bar: 50 μm. **e.** Depth map of the vasculature. Scale bar: 100 μm. **f.** The time-lapse 3D path of the flowing particle cluster. **h.** Flow velocity comparison between 2D (v_{2D}) and 3D (v_{3D}) quantification methods. **i.** Color-encoded flow velocity in the mouse brain. Scale bar: 100 μm.

Mouse brain *in vivo* imaging under norepinephrine

To demonstrate the capability in capturing acute brain activities, norepinephrine, which is used to increase the artery pressure in the clinic, was injected while imaging (Fig. 4a). A contraction in the artery and vein were observed (Fig. 4b). Time-lapsed cross-sectional profiles of the selected artery and

vein are shown in Fig. 4c. After norepinephrine injection, the artery underwent contraction, reducing its

volume to 60% of its resting state. The vein exhibited a more limited reduction in size (~10%).

Additionally, vessel contraction was evident along the axial direction, as illustrated in Fig. 4d. The

dynamic response of the vessel, captured with high temporal resolution, is presented in Supplementary

Movie 3. We further quantified the corresponding flow velocity, as depicted in Fig. 4e, revealing a

significant increase. The concomitant occurrence of vessel contraction and elevated flow velocity

suggests an increase in blood pressure within the brain vessels, aligning with clinical prediction.

Monitoring the formation of ischemic stroke

Ischemic stroke is the most common type of stroke, which can cause permanent brain damage and be

fatal. Monitoring the formation of clots with high spatial-temporal resolution presents a unique

opportunity for the development of potential therapies for ischemic stroke.

As shown in Fig. 4f, artery-targeted photothrombosis was applied while monitoring with the AIR

NIR-II LFM. Prior to the experiments, Rose Bengal and TTT8,4-B NPs were administered. Fig. 4g

presents the vasculature before and after stroke induction. The targeted artery was occluded using a 532

nm laser for excitation. The subsequent clot formation, accompanying changes in vessel morphology,

and 3D flow redistribution were observed and illustrated in Fig. 4h and Supplementary Movie 4.

Noninvasive *in vivo* imaging of T cells in subcutaneous solid tumor through intact skin

The high penetration depth and high temporal-spatial resolution of the AIR NIR-II LFM enable

noninvasive 3D *in vivo* imaging and real-time tracking of T cells in the subcutaneous solid tumor

through intact skin. For ease of observation, cytosine-phosphate-guanine (CpG) was intratumorally

injected to activate T cells previously because it is the ligand of Toll-like receptor 9 (TLR9) and is

capable of inducing systemic immune responses upon binding with TLR9[52,53]. To facilitate the targeted

monitoring of T cells within the tumor microenvironment, TTT8,4-B NPs were functionalized with

αOX40 to yield TTT8,4-B NPs-αOX40 (Fig. 4i). αOX40 is the monoclonal antibody of OX40, which

is a co-stimulatory molecule expressed on activated effector T cells and regulatory T cells[54,55].

With the assistance of T cells-targeted TTT8,4-B NPs-αOX40, the movements of the T cells were

imaged by the NIR-II LFM with 500 ms integration time each 15 s (Supplementary Movie 5). In this

case, we observed the cell movements and mapped its trajectories. As shown in Fig. 4k, a location shift

during 30 min can be observed volumetrically. Fig. 4l shows the time-lapsed trajectories of the three

target T cells. Such results suggest our approach holds great potential in the investigation of

immunotherapy of solid tumors.

Fig. 4. Mouse brain hemodynamic monitoring and T-cell tracking in a solid tumor. a. Schematic of norepinephrine experiment. **b.** MIP of the vasculature before and after norepinephrine injection. Left: before injection, right: after injection. Top: xy projection, bottom: xz projection of the selected region. The dashed lines indicate the vessel diameter changes before and after injection. Scale bar: 100 μm. **c.** The volumetric change of the target artery and vein during the experiment. The blue arrow indicates the time of injection. The red line indicates the normalized diameter before the injection. **d.** The cross-sectional profiles of the white arrow positioned region in **b**. **e.** The flow velocity comparison before and after the norepinephrine injection. **f.** Schematic of ischemic stroke experiment. **g.** The volume renderings and the xy projected images before and after the ischemic stroke. Scale bar: 100 μm. **h.** Time-lapse monitoring of the generation of ischemic stroke and flow redistribution. Scale bar: 50 μm. **i.** A schematic illustration for preparation of the TTT8,4-B NPs-αOX40. **j.** Schematic of the tumor experiment. **k.** T cells monitoring in subcutaneous a tumor. The white arrows indicate the three moving cells. Scale bar: 100 μm. **l.** T cell volumetric tracking in a subcutaneous solid tumor. Scale bar: 100 μm.

Discussion

Observation of biological dynamics *in vivo* in deep tissue is essential for investigating physiological

mechanisms, disease pathogenesis, and therapeutic interventions. However, it is challenging due to the inherent trade-off among visualizing clearly, deeply, and quickly. Scanning-based microscopy offers high resolution and large penetration depth, but severely sacrifices the imaging speed, thereby limiting the capability of capturing fast biological dynamics in 3D spaces. Our newly developed NIR-II LFM strategy circumvents this compromise between speed, resolution, and depth in *in vivo* imaging based on a fully self-supervised view-wise aberration corrected INR reconstruction and volumetric selected illumination. With the integration of ultra-fast LFM and deep NIR-II detection, our strategy provides a volumetric imaging speed of 20 Hz for a volume of 550 μm diameter and 200 μm thickness at cellular resolution and finally achieves a large imaging depth of 600 μm in the mouse brain, while visible LFM could not obtain distinguishable vessels over only 100 μm depth.

It is worth noting that such an efficient combination is based on a series of technical advances. First, to address the challenge of poor 3D reconstruction quality stemming from sparse and limited angular captures, particularly pronounced in large-scale *in vivo* observations due to the low NA of the objective, we introduce physics-based INR for 3D reconstruction. Leveraging the continuity of the INR function and geometry consistency of synthesized views, our approach significantly reduces artefacts and improve axial resolution. More importantly, our VAP module, the self-supervised digital AO, view-wisely corrects the abbreviations introduced by both the optical system and complex samples, without compromising capturing speed as hardware AO does, allowing LFM to fully exploit its advantage of high temporal resolution. Furthermore, it is crucial to prevent out-of-focus fluorescence for LFM reconstruction in thick tissues. Our scanning lightsheet-based volumetric selective illumination approach significantly enhances imaging quality and facilitates the differentiation of small structures that would otherwise be obscured without selective illumination.

Benefiting from the unique advantage of AIR selective-illumination NIR-II LFM, we were able to track vessel dynamics and blood flow in the 3D vasoganglion with video rate, providing an accurate flow velocity quantification. Our approach also allowed to capture the artery contraction and vein dilation under the challenge of norepinephrine, as well as to monitor hemodynamic redistribution during the formation of ischemic stroke through photothrombosis. Moreover, the 3D movement of immune cells inside a subcutaneous tumor was clearly observed through intact skin and trajectory tracking could be easily performed. Therefore, AIR selective-illumination NIR-II LFM can be a powerful tool in brain science and tumor science. Potential investigations of neuronal activity and tumor environment would further elevate the impact of this technique.

While NIR-II LFM exhibits enhanced performance over existing *in vivo* optical microscopy techniques, particular in terms of 3D imaging speed, it still faces some limitations. First, TTT8,4-B NPs emits fluorescence with a peak located at 1000 nm, although belongs to NIR-II region, it still suffers from relatively higher tissue scattering compared to longer wavelengths. Meanwhile, although scanning lightsheet illumination can significantly increase the optical power density within the reconstructed depth of field compared to out-of-focus areas, it does not completely eliminate out-of-focus

fluorescence. Addressing these challenges would facilitate a deeper imaging capability for NIR-II LFM. From a material standpoint, the development of bright, biocompatible, multi-color, and targeted fluorophores with longer emission wavelengths would enhance imaging quality at greater depths. Optically, integrating a confocal strategy into NIR-II LFM could further mitigate out-of-focus background, thereby improving overall image performance.

Methods

NIR-II LFM

A schematic diagram of the experimental arrangement is shown in Supplementary Fig. 6. The laser beam from a NIR CW laser (808 nm, 200 mw, Changchun New Industries Optoelectronics Tech, Co., Ltd.) is collimated with an objective lens. A beam expander, which is constructed by two lenses (L1, f = 45 mm, OLD2435-T3; L2, f = 60 mm, OLD2438-T3 JCOPTIX), is used to reshape the beam diameter. A cylindrical lens (CL1, f = 200 mm, OLYC2601661-T3, JCOPTIX) and multiple doublets (L3, f = 30 mm, OLD2430-T3; L4, f = 60 mm, OLD2438-T3; L5, f = 100 mm, OLD2444-T3, JCOPTIX) are selected to adjust the beam waist in two directions. The NIR beam is steered by a high-speed galvanometer (6240H, Cambridge Technology, INC.) and passes through a dichroic mirror (DM1, OFD1LP-950, JCOPTIX). A doublet (L6, f = 250 mm) and a tube lens (TL, f = 200 mm, ACT508-200-C, Thorlabs) relay the galvanometer to the back aperture of the objective lens (25x, 1.0NA, XLPLN25XSVMP2, Olympus). The generated fluorescent light is collected by the objective and then reshaped by the TL and lens (L7, f = 200 mm, ACT508-200-C, Thorlabs). A micro lens array (pitch size 2.4 mm, f = 30 mm) is carefully adjusted to generate view images from 9 directions. In this study, we enhance the imaging depth-of-field without compromising the field-of-view by designing a mask for each microlens to reduce their light-gathering aperture (d = 1.7 mm). The images are collected by relay lens L8 (f = 75 mm, OLD3240-T3, JCOPTIX) and L9 (f = 75 mm, OLD3240-T3, JCOPTIX) on camera (pixel = 15 um, Ninox Ultra 640, Raptor Photonics). Another CW laser at 532 nm (MGL-U-532-20mW, Changchun New Industries Optoelectronics Tech, Co., Ltd.) is coupled into the system by the DM1 to generate a tightly focused spot on the mouse brain for photothrombosis.

Selective illumination

Fluorescence microscopy suffers from the out-of-focus background signals due to the linearity of single photon absorption. This image quality degradation, especially on thick tissue, casts doubt on the accuracy of light field reconstruction. Two-photon scanning LFM[49] eliminates crosstalk by two-photon absorption, but it is time consuming and the ultrafast system is relatively expensive. Confocal alignment with a specially designed mask[33] shows excellent performance but also places high requirements on design, fabrication and alignment.

To improve the reconstruction quality, out-of-focus crosstalk was mitigated through carefully designed selective illumination. As shown in Fig. 1a and Supplementary Fig. 6, the thin illumination plane is generated by a cylindrically focused lens. By forming a focused light-sheet plane and scanning

with a galvanometer, only the on-focus reconstruction region was illuminated with high fluence. The
beam waist of the light sheet is designed to 3 μm which corresponds to a Rayleigh range of 69 μm.

Network implementation of AIR

The iterative process of our AIR network can be delineated into four distinct steps: 1) position encoding;
2) INR reconstruction; 3) View-wise aberration prediction; and 4) loss computation and network
optimization.

Position encoding. Deep networks exhibit a predisposition toward learning lower frequency
functions. By employing position encoding functions to map 3D coordinates into a higher-dimensional
space before they are input into the network, this approach facilitates more precise data fitting processes
involved in high-frequency variations[24]. The expression of the Fourier position encoding function
employed in this paper is provided below:

427
$$\gamma(p_{xy}) = \{(sin2^0\pi p_{xy}, cos2^0\pi p_{xy}), (sin2^1\pi p_{xy}, cos2^1\pi p_{xy}), ..., (sin2^{L_{xy}-1}\pi p_{xy}, cos2^{L_{xy}-1}\pi p_{xy})\}$$

428
$$\gamma(p_z) = \{(\sin2^0\pi p_z, \cos2^0\pi p_z), (\sin2^1\pi p_z, \cos2^1\pi p_z), ..., (\sin2^{L_z-1}\pi p_z, \cos2^{L_z-1}\pi p_z)\}$$

Where p_{xy} represents the lateral coordinates (x, y), while p_z denotes the axial coordinate. The
parameters L_{xy} and L_z correspond to the lateral and axial encoding orders, respectively. Following
the position encoding process, a high-dimensional feature vector, denoted as $V \subseteq R^{N \times L}$ is generated,
where N signifies the number of input coordinates and L represents the number of channels of the
encoded features determined by L_{xy} and L_z. Each element V_i within the vector V can be expressed
as follows:

435
$$V_i = \left[\gamma(p_{xy}), \gamma(p_z), x, y, z \right]$$

INR reconstruction. Our INR (Implicit Neural Representation) reconstruction network utilizes a
multi-layer perceptron (MLP) structure with ReLU activation. 3D coordinates are firstly flattened into
a one-dimensional vector, transforming the spatial structure into a format suitable for the MLP
architecture. Next, the 3D coordinates undergo Fourier position encoding, after which they are fed into
the INR reconstruction network. The INR reconstruction network predicts the intensity values at each
coordinate. This process can be represented as follows:

442
$$I_i = Relu\left(MLP(V_i)\right)$$
$$Vol_{est} = Reshape(\{I_i\}_{i=1}^N)$$

Here, i denotes the index of a specific voxel, and I_i represents the corresponding grayscale value.
The ReLU activation function is employed to truncate any predicted values that are less than zero.
Following the inference process of the network, the predicted voxel values are subsequently reshaped

into a 3D estimated volume (Vol_{est}).

View-wise aberration prediction. In order to constrain the 3D estimated volume of the INR

reconstruction network using the 2D input raw light field images, we must generate a PSF that aligns

with the parameters of the optical system to project 3D estimated volume, thereby producing the 2D

estimated light field image. Subsequently, the estimated light field image is compared with the input

raw light field image to compute a loss function, which serves to constrain the output of the INR

reconstruction network. The use of a PSF based on wave optics does not account for aberrations and

distortions caused by the sample and optical system, leading to artifacts and ghosting in the

reconstructed image (Fig. 1). Addressing this issue, optimizing the Zernike coefficients and using linear

combinations of Zernike terms to represent aberrations has been an effective approach for removing

aberrations in wide-field imaging[25]. Here, we introduce a view-wise aberration prediction network with

learnable parameters $\alpha_{n,v}^{l}$ representing Zernike coefficients. Here, v denotes the viewing angle index,

ranging from 1 to 9. while n and l represent radial and azimuthal indices, respectively, identifying the

specific Zernike term (spherical or coma, etc.). A total of 12 terms are employed (ANSI standards 4th

to 15th). Thus, the estimated aberration for each view (ϕ_v) can be computed as follows:

461
$$\phi_v = \sum_{n,l} \alpha_{n,v}^{l} \cdot Z_n^{l}$$

The integration of estimated aberration into the light field PSF is an important problem. Yi Zhang

et al. addressed this by introducing optical aberrations on the native image plane and simulating the PSF

after passing through a microlens array modulation using Fresnel diffraction[49]. However, as the PSF for

Fourier LFM is essentially 5D, it requires substantial computational resources, particularly for deep

learning networks where the process needs to be repeated every iteration. Therefore, in this study, we

chose to directly input the ideal PSF from wave optics and add aberrations to the PSF in the spatial

frequency domain for each view. The resulting aberrated PSF was then inverse Fourier-transformed to

yield the final PSF with aberrations (psf_v'). This approach not only saves unnecessary computational

resources but also enables the independent addition of aberrations to the PSF for each view (psf_v). The

process can be represented as follows:

472
$$psf_v' = \text{ifft}\left(\text{fft}(psf_v) \cdot e^{\frac{j2\pi\phi_v}{\lambda}} \right)$$

loss computation and network optimization. Having obtained a PSF with aberrations, forward

projection of the 3D estimated volume produced by the INR reconstruction network is performed. This

involves a slice-wise convolution of the 3D estimated volume with the estimated PSF along the depth

direction, followed by sum projection to generate the 2D estimated light field image (LF_{est}). The error

between this estimated light field image and the input raw light field image (LF_{raw}) is quantified using

a designated loss function:

$$Loss = L_2(LF_{est}, LF_{raw}) + \lambda_1(1 - SSIM(LF_{est}, LF_{raw})) + \lambda_2 \nabla_{zz}(Vol_{est}) + \lambda_3 L_1(Vol_{est})$$

Where L_2 is L_2 loss function. SSIM represents the calculation of the structural similarity
between two images. ∇_{zz} and L_1 are the regularization term to avoid model overfitting. λ_1, λ_2,
λ_3 are weighting coefficients.

An Adam optimizer then utilizes the gradients derived from this loss function to jointly update the
parameters of both the INR reconstruction network and the view-wise aberration prediction network.
This constitutes a single iteration of the AIR network. With increasing iterations, the INR reconstruction
network produces a 3D structure of higher fidelity, while the aberration prediction network refines its
prediction of the aberrations based on the input raw light field image.

Synthesis and characterization of TTT8,4-B

The synthetic route of TTT8,4-B is shown in Supplementary Fig. 13.

Synthesis of intermediate **1**:

A solution of 1-(4-bromophenyl)-1,2,2-triphenylethylene (2.05 g, 5 mmol), aniline (600 mg, 6.5
mmol), tri-tert-butylphosphine (16 mg, 0.08 mmol), $Pd_2(dba)_3$ (64 mg, 0.07 mmol), and sodium tert-
butoxide (625 mg, 6.5 mmol) were refluxed under nitrogen (N_2) in dry toluene (30 mL) at 110 $^\circ$C for
24 h. After cooling to room temperature, the solvent was removed by evaporation under reduced
pressure. H_2O (30 mL) and chloroform (200 mL) were then added. The organic layer was separated and
washed with brine, dried over anhydrous $MgSO_4$, and evaporated to dryness under reduced pressure.
The crude product was purified by column chromatography on silica gel using hexane/chloroform (v/v
= 5/1) as the eluent to afford intermediate **1**. ^{1}H NMR (600 MHz, chloroform-d) δ 7.26 (d, J = 7.6 Hz,
2H), 7.23-6.99 (m, 18H), 6.93 (d, J = 8.3 Hz, 3H), 6.84 (d, J = 8.2 Hz, 2H).

Synthesis of intermediate **2**:

Intermediate **1** (3.77 g, 8.9 mmol), 4-bromoiodobenzene (2.97 g, 10.6 mmol), and 1,10-
phenanthroline (270 mg, 1.5 mmol) were dissolved in toluene (30 mL). After the solution was heated
to 100 $^\circ$C, CuCl (150 mg, 1.5 mmol) and KOH (1.23 g, 22.0 mmol) were added under N_2 purge. The
mixture was refluxed for 48 h at 120 $^\circ$C. After cooling to room temperature, the solvent was removed
by evaporation under reduced pressure. H_2O (30 mL) and chloroform (40 × 3 mL) were then added.
The organic phase was separated and washed with brine, and dried over anhydrous Na_2SO_4. After
removal of the solvent, the residual was purified on a silica gel column with ethyl acetate/petroleum
(v/v = 1/40) as the eluent, giving intermediate **2**. ^{1}H NMR (400 MHz, chloroform-d) δ 7.32 (d, J = 8.9
Hz, 2H), 7.29-7.23 (m, 2H), 7.17-7.02 (m, 18H), 6.91 (dd, J = 8.8, 2.9 Hz, 4H), 6.80 (d, J = 8.6 Hz,
2H).

Synthesis of intermediate **3**:

Intermediate **2** (2 g, 3.46 mmol), bis(pinacolato)diboron (1.93 g, 7.61 mmol), Pd(dppf)$_2$Cl$_2$ (119.3

mg, 0.16 mmol), and KOAc (1.02 g, 10.2 mmol) were added into a two-necked flask. Then 1,4-dioxane

was added under the N$_2$ atmosphere. The mixture was reacted at 90 $^{\circ}$C for 12 h. After being cooled to

room temperature, the solvent was evaporated under reduced pressure. Subsequently, 30 mL of H$_2$O

and 200 mL of chloroform were added. The organic layer was separated and subjected to brine washing,

followed by drying with anhydrous MgSO$_4$. Finally, the mixture was evaporated to dryness under

reduced pressure. The crude product was purified by silica gel column (petroleum

ether/dichloromethane = 3/1, v/v) to give a yellow powder (1.25 g, yield: 61.4%). ^{1}H NMR (600 MHz,

chloroform-*d*) δ 7.69 (d, *J* = 8.0 Hz, 2H), 7.27 (d, *J* = 7.8 Hz, 2H), 7.21-7.04 (m, 18H), 7.01 (d, *J* = 8.0

Hz, 2H), 6.93 (d, *J* = 8.1 Hz, 2H), 6.86-6.83 (m, 2H), 1.38 (d, *J* = 2.8 Hz, 12H).

Synthesis of **TTT8,4-B**:

The synthesis of **TTT8,4-B** followed the procedure outlined in the previous report[56]. Briefly,

intermediate **3** (630 mg, 1 mmol), BBTD-T (8,4) (340 mg, 0.4 mmol), Pd$_2$(dba)$_3$ (22.9 mg, 0.025 mmol),

P(o-tol)$_3$ (60.9 mg, 0.2 mmol), toluene (1.5 mL), and H$_2$O (0.3 mL) were added and sealed in a

microwave tube. The reaction mixture was heated to 130 $^{\circ}$C for 1 h in a Biotage microwave reactor.

After being cooled to room temperature, the solvent was evaporated under reduced pressure.

Subsequently, 30 mL of H$_2$O and 60 mL of dichloromethane were added. The organic layer was

separated and subjected to brine washing, followed by drying with anhydrous MgSO$_4$. Finally, the

mixture was evaporated to dryness under reduced pressure. The crude product was quenched and

extracted with dichloromethane. The crude product was purified by column chromatography to obtain

a dark green solid (388 mg, yield: 40%). ^{1}H NMR (500 MHz, chloroform-*d*) δ 7.47 (d, *J* = 8.4 Hz, 4H),

7.20 (s, 4H), 7.17 (s, 2H),7.12-6.94 (m, 40H), 6.84 (d, *J* = 7.9 Hz, 4H), 6.77 (s, 4H), 2.50 (d, *J* = 6.8 Hz,

4H), 1.34 (dd, *J* = 12.7, 6.6 Hz, 2H), 1.08 (h, *J* = 7.1 Hz, 4H), 0.95-0.82 (m, 26H), 0.73 (t, *J* = 7.3 Hz,

6H), 0.60-0.54 (m, 6H). ^{13}C NMR (126 MHz, CD2Cl2) δ 153.23, 147.42, 147.21, 146.07, 145.64,

144.80, 143.99, 143.82, 143.60, 140.79, 140.69, 138.67, 132.16, 131.27, 131.25, 131.22, 129.28, 128.96,

128.10, 127.63, 127.60, 126.49, 126.42, 126.38, 126.31, 125.32, 124.67, 123.39, 123.29, 116.25, 39.07,

34.76, 33.25, 32.97, 31.80, 29.53, 28.60, 28.57, 26.32, 26.29, 22.82, 22.65, 13.86, 13.80. MS: m/z: [M]$^+$

calculated for C$_{114}$H$_{108}$N$_6$S$_4$: 1689.75911; Found: 1689.75607.

More details of the synthesis and characterization of TTT8,4-B are shown in Supplementary Note

2 and 3.

Preparation of TTT8,4-B NPs and TTT8,4-B NPs-αOX40

For the preparation of TTT8,4-B NPs, 1 mL of THF solution containing 1 mg of TTT8,4-B and 10 mg

of PS5000-mPEG5000 was first prepared. This solution was then mixed with 9 mL of ultrapure water

under continuous ultrasonication for 2 minutes using a microtip probe sonicator (45% output power).

After ultrasound, the reaction mixture was transferred into a dialysis bag (MWCO 3500 Da) and

dialyzed against ultrapure water for 24 h, with the water refreshed every 4 h to ensure the complete

removal of THF. Following dialysis, the resulting NPs dispersed in ultrapure water were concentrated
using an ultrafiltration tube and either stored at 4 °C or used immediately.

To prepare TTT8,4-B NPs-αOX40, maleimide-modified TTT8,4-B NPs (TTT8,4-B NPs-Mal)
were first synthesized. The procedure for preparing TTT8,4-B NPs-Mal was identical to that of TTT8,4-
B NPs, except that the 1 mL THF solution contained 8 mg of PS5000-mPEG5000, 2 mg of PS5000-
PEG5000-Mal, and 1 mg of TTT8,4-B. Subsequently, TTT8,4-B NPs-Mal were functionalized with
thiol-modified αOX40 (prepared according to the literature[57,58]) via sulfhydryl-maleimide coupling to
yield the final TTT8,4-B NPs-αOX40. The resulting TTT8,4-B NPs-αOX40 in PBS were concentrated
and stored at 4 °C or used immediately.

The concentrations of TTT8,4-B NPs and TTT8,4-B NPs-αOX40, based on TTT8,4-B content,
were quantified using a pre-established standard absorption curve of TTT8,4-B in THF.

More details of the synthesis and characterization of TTT8,4-B NPs are shown in Supplementary
Note 2 and 3.

Phantom preparation

The dried leaf was firstly dyed in TTT8,4-B NPs (1.7 mmol/mL) and Rhodamine B (1 mg/mL) mixture
for 30 min, which provides fluorescent signals simultaneously in visible and near-infrared. Then, it was
fixed with a thin layer of transparent agarose (0.8%). After the acquisition of the surface image without
scattering tissue, the lipid (0.167%) and agarose (0.8%) mixture was applied on the top of the phantom.
The depth of the covered lipid tissue was estimated by calculating the volume and the area of the disk.
Once a new layer of lipid agarose was applied, the phantom was placed on a flat surface and waited for
solidification. The photos of the leaf vessels under different depths of lipid can be found in Fig. 2a.

Hemodynamic imaging

Supplementary Fig. 14 is a schematic of the process of measuring flow speed with the NIR-II LFM
strategy. First, the depth map was calculated by searching through the points with the maximum value
in the vessel. Then, a skeleton was drawn on an interested vessel in the MIP images. By combining the
depth information and the top view 2D location matrix, we can derive the volumetric skeleton location
coordinates. A kymograph was relocated and constructed with the volumetric coordinates. The 3D blood
flow velocity of the interested vessel was quantified based on the relationship between the slope with
the real speed.

Animal preparation for *in vivo* brain imaging

We used BALB/C mice (mice, male, 8 weeks old) for the *in vivo* studies. Throughout the experiments,
the mice were maintained under anesthesia with 1.5% isoflurane and the mice body were kept at 37 °C
using a heating pad. Craniotomy surgery was performed to remove the skull, and a thin glass plate was
used to seal the brain. Before imaging, erythromycin cream was applied to mice eyes to keep them

moist. The laser power was measured with a power meter and controlled carefully. All experimental
procedures were carried out in conformity with the animal protocol approved by the laboratory animal
center at Huazhong University of Science and Technology.

Mouse imaging under norepinephrine

586 To validate the system performance in monitoring acute hemodynamic change. The mouse was imaged
587 under the challenge of norepinephrine. The mouse was anesthetized and placed on a heating pad. The
588 AIE NP solution was injected before imaging and the region was selected under LFM. A syringe was
589 injected intravenously without drug delivery. After acquiring baseline vasculature and flow speed for 1
590 min. We pumped norepinephrine (0.02 mg/mL, 200 μL) and monitored for another 1.5 mins.

Ischemic stroke

We used 532 nm CW laser and a photosensitizer (Rose Bengal, Sigma; concentration 10 mg/mL, 0.2
mL per mouse) to generate blood clots in target regions. Specifically, the mixture of Rose Bengal and
TTT8,4-B NPs were injected intravenously. Then, the mouse was placed under the objective. Water was
applied to the top of the head and the objective was soaked. The NIR laser was turned on with a safe
level output power. The region of interest was located with the aid of NIR-II LFM. Then, the brain
region was continuously monitored for 1 min at a 20 Hz frame rate and the 532 nm laser was launched
to the target vessel. The 532 nm laser dwelled for 30 seconds for the stable generation of stroke. Finally,
the *in vivo* mouse brain was recorded for another minute.

Tumor inoculation and *in vivo* NIR-II LFM imaging of tumors

5-week-old BALB/c mice were shaved using hair removal lotion and inoculated with ~ 2 million CT26
cancer cells near the inguinal region for tumor growth. When the tumor size reached ~ 4 mm, 100 μL
aOX40-TTT8,4-B NPs were intravenously injected into the BALB/c mouse through tail vein 24 hours
after 50-μg CpG (ODN 2395, InvivoGen) intratumoral treatment. NIR-II LFM was performed to track
the 3D movement of aOX40-TTT8,4-B NPs-stained T cells in the CT26 tumor 24 hours post the probe
injection.

Supplementary information

608 Supplementary Figs. 1-29, Notes 1-3 and Table 1.

Data availability

610 The raw data will be made available upon request to the corresponding author. The remaining data are
611 available within the Article, Supplementary Information.

Acknowledgements

613 We thank Yao Zhou from Huazhong University of Science and Technology for her kind advise. This
614 work was partly supported by the National Natural Science Foundation of China (62375095,

32361133552, 62405099, 22275124, 22475134); the Natural Science Foundation of Wuhan (NO. 2024040801020204); the China Postdoctoral Science Foundation (2024M750994); the Interdisciplinary Research Program of HUST (2024JCYJ064); Postdoctor Project of Hubei Province under Grant Number (2024HBHCXA015); Start-up Grant from Shenzhen University (868-000001032113, 868-000001032219); Shenzhen University 2035 Program for Excellent Research (868-000003011036).

Author contributions

D.L. conceived the idea. F.Z. and M.H. led the experiment with supervision from D.L. and P.F.. M.H., F.Z. and Y.H. built the optical setup. Z.Z., X.L. and K.R. designed, synthesized and characterized the TTT8,4-B. Z.Z. and X.L. designed, fabricated, and characterized the TTT8,4-B nanoparticles (NPs) and TTT8,4-B NPs-αOX40. M.H. and F.Z. discussed and designed the experimental system and control system. F.Z., M.H. and S.M. handled the animal preparation. F.Z., M.H., Y.H., C.Y. and X.H. processed and analyzed microscopic data. Z.Z, D.L. and P.F. manage the project. F.Z., X.L., M.H., C.Y., M.K., D.W., Z.Z., D.L. and P.F. wrote and revised the paper.

Competing interest: The authors declare no competing interests.

References

1. Hoover, E. E. & Squier, J. A. Advances in multiphoton microscopy technology. *Nature Photonics* **7**, 93–101 (2013).

2. Samanta, S. *et al.* AIE-active two-photon fluorescent nanoprobe with NIR-II light excitability for highly efficient deep brain vasculature imaging. *Theranostics* **11**, 2137–2148 (2021).

3. Horton, N. G. *et al.* In vivo three-photon microscopy of subcortical structures within an intact mouse brain. *Nature Photonics* **7**, 205–209 (2013).

4. Wang, L. V. & Yao, J. A practical guide to photoacoustic tomography in the life sciences. *Nature Methods* **13**, 627–638 (2016).

5. Zhong, F., Huang, X., Sun, M., Li, D. & Fei, P. High-Speed Hemodynamic Imaging with Low-Fluence Photoacoustic Microscopy and Self-Supervised Single Volume Denoising. *Laser and Photonics Reviews* **19**, 2401291 (2025).

6. Lan, B. *et al.* High-speed widefield photoacoustic microscopy of small-animal hemodynamics. *Biomedical Optics Express* **9**, 4689 (2018).

7. Zi-Han, C. *et al.* NIR-II Anti-Stokes Luminescence Nanocrystals with 1710 nm Excitation for in vivo Bioimaging. *Angewandte Chemie International Edition* **e202416893**, 1–7.

8. Xiangdan, M. *et al.* Recent Advances in Near-Infrared-II Fluorescence Imaging for Deep-Tissue Molecular Analysis and Cancer Diagnosis. *Small* **18**, 1–26 (2022).

9. Feng, Z. *et al.* Engineered NIR-II fluorophores with ultralong-distance molecular packing for
high-contrast deep lesion identification. *Nature Communications* **14**, 5017 (2023).

10. Feng, Z. *et al.* Perfecting and extending the near-infrared imaging window. *Light Science &*
Applications **10**, 197 (2021).

11. Hong, G. *et al.* Through-skull fluorescence imaging of the brain in a new near-infrared
window. *Nature Photonics* **8**, 723–730 (2014).

12. Qi, J. *et al.* Aggregation-Induced Emission Luminogen with Near-Infrared-II Excitation and
Near-Infrared-I Emission for Ultradeep Intravital Two-Photon Microscopy. *ACS Nano* **12**, 7936–
7945 (2018).

13. Welsher, K. *et al.* A route to brightly fluorescent carbon nanotubes for near- infrared imaging
in mice. *Nature Nanotechnology* **4**, 773–780 (2010).

14. Demas, J. *et al.* High-speed, cortex-wide volumetric recording of neuroactivity at cellular
resolution using light beads microscopy. *Nature Methods* **18**, 1103–1111 (2021).

15. Papagiakoumou, E., Ronzitti, E. & Emiliani, V. Scanless two-photon excitation with temporal
focusing. *Nature Methods* **17**, 571–581 (2020).

16. Wang, F. *et al.* Light-sheet microscopy in the near-infrared II window. *Nature Methods* **16**,
545–552 (2019).

17. Wang, F. *et al.* In vivo NIR-II structured-illumination light-sheet microscopy. *Proceedings of*
the National Academy of Sciences of the United States of America **118**, e2023888118 (2021).

18. Wang, Z. *et al.* Real-time volumetric reconstruction of biological dynamics with light-field
microscopy and deep learning. *Nature Methods* **18**, 551–556 (2021).

19. Zhang, Y. *et al.* Long-term mesoscale imaging of 3D intercellular dynamics across a
mammalian organ. *Cell* **187**, 6104-6122.e25 (2024).

20. Yi, C. *et al.* Video-rate 3D imaging of living cells using Fourier view-channel-depth light field
microscopy. *Communications Biology* **6**, 1259 (2023).

21. Yoon, Y.-G. *et al.* Sparse decomposition light-field microscopy for high speed imaging of
neuronal activity. *Optica* **7**, 1457 (2020).

22. Truong, T. V. *et al.* High-contrast, synchronous volumetric imaging with selective volume
illumination microscopy. *Communications Biology* **3**, 74 (2020).

23. Hsu, F.-C. *et al.* Light-field microscopy with temporal focusing multiphoton illumination for
scanless volumetric bioimaging. *Biomedical Optics Express* **13**, 6610 (2022).

24. Prevedel, R. *et al.* Simultaneous whole-animal 3D imaging of neuronal activity using light-

field microscopy. *Nature Methods* **11**, 727–730 (2014).

25. Wagner, N. *et al.* Instantaneous isotropic volumetric imaging of fast biological processes.

Nature Methods **16**, 497–500 (2019).

26. Campagnola, P. J. High-speed 3D mapping of nonlinear structures. *Nature Photonics* **14**,

527–528 (2020).

27. Gustafsson, M. G. L. *et al.* Three-Dimensional Resolution Doubling in Wide-Field

Fluorescence Microscopy by Structured Illumination. *Biophysical Journal* **94**, 4957–4970 (2008).

28. Booth, M. J. Adaptive optical microscopy: the ongoing quest for a perfect image. *Light*

Science & Applications **3**, e165–e165 (2014).

29. Tang, J., Germain, R. N. & Cui, M. Superpenetration optical microscopy by iterative

multiphoton adaptive compensation technique. *Proceedings of the National Academy of*

Sciences of the United States of America **109**, 8434–8439 (2012).

30. Wang, C. *et al.* Multiplexed aberration measurement for deep tissue imaging in vivo. *Nature*

Methods **11**, 1037–1040 (2014).

31. Wu, J. *et al.* Iterative tomography with digital adaptive optics permits hour-long intravital

observation of 3D subcellular dynamics at millisecond scale. *Cell* **184**, 3318-3332.e17 (2021).

32. Gottschalk, S. *et al.* Rapid volumetric optoacoustic imaging of neural dynamics across the

mouse brain. *Nature Biomedical Engineering* **3**, 392–401 (2019).

33. Zhang, Z. *et al.* Imaging volumetric dynamics at high speed in mouse and zebrafish brain

with confocal light field microscopy. *Nature Biotechnology* **39**, 74–83 (2021).

34. Chen, W. *et al.* High-throughput volumetric mapping of synaptic transmission. *Nature*

Methods **21**, 1298–1305 (2024).

35. Zhong, F. & Hu, S. Thin-film optical-acoustic combiner enables high-speed wide-field multi-

parametric photoacoustic microscopy in reflection mode. *Optics Letters* **48**, 195 (2023).

36. Sun, Z. *et al.* CXCL9+ Macrophage-targeted NIR-II aggregation-induced emission

nanoprobes for the early diagnosis of myocarditis. *Nano Today* **54**, 102107 (2024).

37. Yang, S. *et al.* More Is Better: Dual-Acceptor Engineering for Constructing Second Near-

Infrared Aggregation-Induced Emission Luminogens to Boost Multimodal Phototheranostics.

Journal of the American Chemical Society **145**, 22776–22787 (2023).

38. Zhang, Z. *et al.* The fast-growing field of photo-driven theranostics based on aggregation-

induced emission. *Chemical Society Reviews* **51**, 1983–2030 (2022).

39. Li, X. *et al.* Aggregation-induced emission materials: a platform for diverse energy

transformation and applications. *Journal of Materials Chemistry A* **11**, 4850–4875 (2023).

40. Cai, Y. *et al.* Near-infrared fluorophores with absolute aggregation-caused quenching and negligible fluorescence re-illumination for in vivo bioimaging of nanocarriers. *Aggregate* **4**, e277 (2023).

41. Li, Y. *et al.* Design of AIEgens for near-infrared IIb imaging through structural modulation at molecular and morphological levels. *Nature Communications* **11**, 1255 (2020).

42. Kang, M. *et al.* Aggregation-enhanced theranostics: AIE sparkles in biomedical field. *Aggregate* **1**, 80–106 (2020).

43. Xu, Z. *et al.* An NIR-II Two-Photon Excitable AIE Photosensitizer for Precise and Efficient Treatment of Orthotopic Small-Size Glioblastoma. *Advanced Materials* **37**, 2413164 (2025).

44. Yan, D. *et al.* An All-Rounder for NIR-II Phototheranostics: Well-Tailored 1064 nm-Excitable Molecule for Photothermal Combating of Orthotopic Breast Cancer. *Angewandte Chemie International Edition* **63**, e202401877 (2024).

45. Lin, R. *et al.* Type I Photosensitization with Strong Hydroxyl Radical Generation in NIR Dye Boosted by Vigorous Intramolecular Motions for Synergistic Therapy. *Advanced Materials* **35**, 2303212 (2023).

46. Qin, Y. *et al.* Modular Construction of AIE-Active Supramolecular Cages with Tunable Fluorescence for NIR-II Blood Vessel Imaging. *ACS Materials Lett.* **5**, 1982–1991 (2023).

47. Qin, Y. *et al.* Long-Term In Vivo Fluorescence Analyses and Imaging-Guided Tumor Surgery in the Second Near-Infrared Window Using a Supramolecular Metallacage. *Aggregate* **0**, e708 (2024).

48. Broxton, M. *et al.* Wave optics theory and 3-D deconvolution for the light field microscope. *Optics Express* **21**, 25418 (2013).

49. Zhao, Z. *et al.* Two-photon synthetic aperture microscopy for minimally invasive fast 3D imaging of native subcellular behaviors in deep tissue. *Cell* **186**, 2475-2491.e22 (2023).

50. Yao, J. *et al.* High-speed label-free functional photoacoustic microscopy of mouse brain in action. *Nature Methods* **12**, 407–410 (2015).

51. Shih, A. Y. *et al.* Two-Photon Microscopy as a Tool to Study Blood Flow and Neurovascular Coupling in the Rodent Brain. *Journal of Cerebral Blood Flow & Metabolism* **32**, 33 (2012).

52. Kawarada, Y. *et al.* Peritumoral CpG DNA Elicits a Coordinated Response of CD8 T Cells and Innate E ectors to Cure Established Tumors in a Murine Colon Carcinoma Model. *The Journal of Immunology* **167**, 5247–5253 (2001).

53. Wagner, H. The immunogenicity of CpG-antigen conjugates. *Advanced Drug Delivery*
Reviews **61**, 243–247 (2009).

54. Ruby, C. E., Montler, R., Zheng, R., Shu, S. & Weinberg, A. D. IL-12 Is Required for Anti-OX40-
Mediated CD4 T Cell Survival. *The Journal of Immunology* **180**, 2140–2148 (2008).

55. Curti, B. D. *et al.* OX40 Is a Potent Immune-Stimulating Target in Late-Stage Cancer Patients.
Cancer Research **73**, 7189–7198 (2013).

56. Song, S. *et al.* Molecular engineering of AIE luminogens for NIR-II/IIb bioimaging and
surgical navigation of lymph nodes. *Matter* **5**, 2847–2863 (2022).

57. Howard, M. D. *et al.* Targeting to Endothelial Cells Augments the Protective Effect of Novel
Dual Bioactive Antioxidant/Anti-Inflammatory Nanoparticles. *Molecular Pharmaceutics* **11**, 2262–
2270 (2014).

58. Parhiz, H. *et al.* PECAM-1 directed re-targeting of exogenous mRNA providing two orders
of magnitude enhancement of vascular delivery and expression in lungs independent of
apolipoprotein E-mediated uptake. *Journal of Controlled Release* **291**, 106–115 (2018).

759