From Movement to METs: A Validation of ActTrust® for Energy Expenditure Estimation and Physical Activity Classification in Young Adults

Elias dos Santos Batista[1,3,4], Stephania Ruth Basilio Gomes[2], Ayrton Bruno de Morais Ferreira[1], Lucas G. S. França[5,6], John Fontenele Araújo[2,4], Arnaldo Luis Mortatti[1], Mario André Leocadio-Miguel[7*]

[1]Programa de Pós-Graduação em Educação Física, Universidade Federal do Rio Grande do Norte, Natal, 59078-970, Rio Grande do Norte, Brazil.
[2]Programa de Pós-Graduação em Psicobiologia, Universidade Federal do Rio Grande do Norte, Natal, 59078-970, Rio Grande do Norte, Brazil.
[3]Instituto Federal de Educação Ciência e Tecnologia do Rio Grande do Norte, Caicó, 59300-000, Rio Grande do Norte, Brazil.
[4]Programa de Pós Graduação Multicêntrico em Ciências Fisiológicas, Universidade Federal do Rio Grande do Norte, Natal, 59078-970, Rio Grande do Norte, Brazil.
[5]Department of Computer and Information Sciences, Faculty of Engineering and Environment, Northumbria University, Newcastle upon Tyne, NE1 8ST, United Kingdom.
[6]Department of Forensic and Neurodevelopmental Science, Institute of Psychiatry, Psychology & Neuroscience, King's College London, London, SE5 8AF, United Kingdom.
[7*]Department of Psychology, Faculty of Health and Life Sciences, Northumbria University, Newcastle upon Tyne, NE1 8ST, United Kingdom.

*Corresponding author(s). E-mail(s): mario.miguel@northumbria.ac.uk;
Contributing authors: elias.batista@ifrn.edu.br;
stephaniaruth.basilio@gmail.com; ayrtonbruno12@hotmail.com;

lucas.franca@northumbria.ac.uk; john.araujo@ufrn.br;
amortatti@gmail.com;

Abstract

Physical activity (PA) is recognised for providing several health benefits in humans, mainly preventing and controlling chronic non-communicable diseases. However estimating PA is a challenging and expensive task. An alternative would be to devise a model to estimate PA using actigraphy devices calibrated from an initially validated model. This has been previously done to a number of devices, including ActiGraph® GT3X+. In this study we aimed at validating ActTrust® against the widely used GT3X+ and comparing activity counts to metabolic equivalents (METs) derived from indirect calorimetry during treadmill walking and running. Fifty-six young adults (34 men, 22 women) participated in controlled effort exercises including light, moderate, vigorous, and very vigorous activity intensities. We developed a general linear model to estimate energy expenditure (EE) from movement count of combinations of GT3X+ and ActTrust devices placed at hip or wrist. We then estimated cut-off points for each intensity range. Our results showed correlations between treadmill speed, METs, and movement counts across all devices and placements combinations. Our proposed model performed well with balanced accuracies above 0.77 for all intensity ranges and over 0.9 for light and moderate activity. This is the first study to model estimate and validate PA intensity thresholds on ActTrust® devices. Our findings support the use of ActTrust® devices in PA estimation as a low complexity and cost approach to allow 24-hour assessments of EE.

Keywords: indirect calorimetry, accelerometer, actigraphy cut-points, activity intensity

1 Introduction

Physical activity (PA) is recognised for providing several health benefits in humans, mainly preventing and controlling chronic non-communicable diseases [1]. Furthermore, the practice of PA is associated with improved subjective and mental well-being and physical self-perception[2]. The World Health Organization (WHO) recommends 150-300 minutes of moderate-intensity aerobic PA, or at least 75-150 minutes of vigorous-intensity aerobic PA, or an equivalent combination of PA of both intensities during the week for adults. Also important is the WHO's global action plan on physical activity, which aims to decrease the prevalence of sedentarism in adults and adolescents by 15 per cent by 2030 [3].

One way to define the intensity of PA is through the metabolic equivalent (MET) of task. This mechanism expresses energy expenditure (EE) and indicates the oxygen needs for different activities, which at rest is equivalent to 3.5 ml O_2/ kg/min [4, 5]. The gold standard methodological approach to assessing energy expenditure is the doubly labelled water (DLW) technique [6]. However, there are limitations mainly related

to this method's high cost, complexity and scalability. Alternatively, accelerometry emerges as a practical, scalable and economically viable method to estimate energy expenditure (EE) and quantify PA based on movement and body acceleration, generating numerical variables of frequency, duration and intensity of the actions performed [7, 8]. The challenges inherent in using accelerometry include the diversity of designs in accelerometer models, the disparities in signal processing between the different sensors and the multiplicity of options regarding possible placements[9–11]. Furthermore, it is possible to verify a variation in the algorithms that calculate the counts, both of the activity and other extracted parameters on different devices [12].

The ActiGraph® GT3X+ accelerometer has the most significant number of validation studies and published thresholds and has been widely used to validate devices with the same purpose [13]. The ActTrust® (Condor Instruments, São Paulo, Brazil) is a device that has a triaxial accelerometer and is specific for inferring sleep from locomotor activity [14]. However, research on the accuracy of these devices for measuring PA is limited and not validated.

The validation of accelerometers through indirect calorimetry is crucial for developing mathematical equations capable of predicting energy expenditure using these devices and establishing cutoff points for classifying the intensity of PA derived from accelerometer data [15]. Based on this context, the two main objectives of this study are: (i) to define equations for estimating EE using ActTrust in young adults, and (ii) to establish cutoff points for the ActTrust movement variable (light, moderate, vigorous and very vigorous PA).

2 Methods

2.1 Participants

Healthy young adults were recruited for this study by online advertising. We established musculoskeletal or cardiovascular diseases that could hinder PA as excluding criteria. Participants were also excluded if they had any other contraindications to exercise or were taking medication that affected their metabolic rate. All participants completed the Physical Activity Readiness Questionnaire (PAR-Q) [16–18], and no one was excluded from the study for answering yes to one or more questions. The statistics of the characteristics of the participants are presented in Table 1.

The study was approved by the Human Ethics Committee of the Federal University of Rio Grande do Norte, under protocol number 3.672.214 (CAAE: 15857419.5.0000.5537). It was performed according to the declaration of Helsinki and complied with the Ethical Standards in Sport and Exercise Science Research. Participants provided written consent to participate in the study.

2.2 Measuring activity

The *ActiGraph GT3X+* (ActiGraph Corporation, headquartered in Pensacola, FL, Firmware version 4.1.0) is a lightweight device that employs triaxial accelerometers and offers a dynamic range ranging from -6 g to +6 g. For the current study, a sampling rate of 30 Hz was chosen, as this range is expected to capture the majority

of accelerations arising from human movement sufficiently [19]. After digitisation, the signal passed through a digital filter that restricts the accelerometer's frequency range to 0.25–2.5 Hz. We used the vector magnitude (VM) data (vector representing the summed value of the three axes). Data from the GT3X+ device were retrieved and processed using *ActiLife version 6.13.4*, software provided by ActiGraph Corporation based in Pensacola, FL.

The *ActTrust* device (Condor Instruments, São Paulo, SP, Brazil) is an actigraphy tool that features a rechargeable battery, a digital tri-axial accelerometer, a skin temperature sensor and a light sensor. It samples acceleration data at 25 Hz, ranging from 0.03 to 4 g. First, this data undergoes processing through a band-pass filter (0.5 - 2.7 Hz), after which the norm is calculated from the three axes. The resultant value is then integrated over an epoch (1 to 60 seconds) and stored in memory. This equipment is designed for activity and sleep monitoring and processes activity using parameters such as PIM (proportional integration mode), ZCM (zero crossing mode) and TAT (time above threshold) [20]. The *Act Studio software* (Condor Instruments, São Paulo, Brazil) downloads, visualises, and exports the collected data.

2.3 Measuring oxygen uptake

Oxygen uptake was measured 'breath-by-breath' continuously during each condition using indirect calorimetry (*Quark CPET*, Cosmed®, Italy). Before testing each subject, we calibrated the metabolic cart with a known gas mixture (16% O_2 and 5% CO_2) and volume. One test had to be repeated seven days later due to an error in the treadmill's security system (*Super ALT*, Inbramed®, Porto Alegre, Brazil). Occasional errant breath episodes (e.g., due to coughing, swallowing or talking) were excluded from the data set when exceeding three standard deviations of the mean, the latter being defined as the average of two following and two preceding sampling intervals [21].

2.4 Experimental procedure

All *GT3X+* and *ActTrust* devices were initialiased via computer interfaces to collect data in 1-second epochs [22]. The GT3X+ was initialised to collect the VM data, and the ActTrust to collect activity in the PIM mode. Each participant wore two pairs of GT3X+ and ActTrust devices, securely placed on the hip on the side of the participant's dominant lower limb through an elastic belt and the other on the non-dominant wrist.

Participants refrained from exercise and fasted for at least 4 hours before the testing session. The protocol consisted of 5 conditions (of 10-min duration each), interspersed with 5-minute rest periods: (i) resting; (ii) treadmill (Quasar Med 4.0, h p cosmos, Nussdorf-Traunstein, Germany) walking at 3 $km \cdot h^{-1}$; (iii) treadmill walking at 5 $km \cdot h^{-1}$; (iv) treadmill walking or running at 7 $km \cdot h^{-1}$; and (v) treadmill running at 9 $km \cdot h^{-1}$.

2.5 Data analysis

Based on the protocol used by [8], we calculated the mean activity counts from the four central minutes of the Vector Magnitude (VM) for GT3X+ and the counts (proportional integral mode - PIM) for *ActTrust* to estimate activity levels. METs (Metabolic Equivalents) were individually calculated (VO^2) similarly.

First, we assessed the effect of the different devices and placement with an ANOVA according to the model Treadmill speed$\sim$ Activity$^{1/2}$: Device$_{placement}$. Secondly, we then modelled the association between METs and activity for each device using a general linear model defined in equation 1.

$$\text{MET}^{1/2} \sim \text{Activity}^{1/2} * \text{Device}_{\text{placement}} \tag{1}$$

We evaluated two devices according to their placement, hence our model can be written as in the equation 2.

$$\text{MET}^{1/2} = \beta_0^{\dagger} + \beta_1^{\dagger}\text{Activity}^{1/2} \tag{2}$$

with $\beta_0^{\dagger} = \beta_0 + \beta_2\text{ACTT}_{hip} + \beta_3\text{ACTT}_{wrist} + \beta_4\text{GT3X+}_{wrist}$ and $\beta_1^{\dagger} = \beta_1 + \beta_5\text{ACTT}_{hip} + \beta_6\text{ACTT}_{wrist} + \beta_7\text{GT3X+}_{wrist}$. ACTT_{hip}, ACTT_{wrist}, and GT3X+_{wrist} are binary variables set to adjust the model values to each configuration. GT3X+_{hip} was defined as the reference level.

Analyses were conducted using R v. 4.4.1 (R Foundation for Statistical Computing, Vienna, Austria) [23] and auxiliary packages *tidyverse* [24], *caret* [25], *MuMIn* [26], *MetBrewer* [27], *cowplot* [28], *pROC* [29], and *broom* [30]. The significance level was set at $\alpha = 0.05$. The data and scripts used in the analyses featured in this article are available at https://github.com/circadia-bio/ACTT_validation_study [31].

We estimated metabolic rate in METs to derive the cut-points for moderate, vigorous and very vigorous exercise intensity levels according to the intervals (3.0-5.99, 6.0-8.99 and above 9.0 METs).

We employed performance metrics for model diagnostics and computed R^2 and confusion matrix. Finally, we evaluated the ability of the new cut points from the linear regression model to accurately classify the physical activity intensity levels derived from EE objective measurements, according to the different MET bands with sensitivity, specificity, balanced accuracy, and area-under-the-curve (AUC) bench marks in a similar fashion to previous works in literature [8].

3 Results

The sample comprised young adults (34 men, 22 women) aged 18–35. The participants' characteristics are displayed in Table 1. Energy expenditure values obtained with indirect calorimetry reflected each condition, increasing proportionally according to the treadmill speed, see Figure 1.

Table 1 Demographic and physical characteristics of the participants. Mean (st. dev.).

Sex	Sample size	Age (y)	Height (cm)	Weight (kg)	Body mass index (kg/m²)
Male	34	28.3 (4.6)	172 (6.2)	78.2 (11.9)	23.1 (1.6)
Female	22	26.8 (5.0)	160 (7.3)	59.9 (9.4)	26.7 (2.5)

3.1 Activity counts and treadmill velocity

We performed a two-way ANOVA to understand the outputs of each device (GT3X+ and ActTrust) according to the placement and condition (treadmill velocity). Movement counts from the GT3X+ devices were mostly distinct ($F(19, 1087) = 1347$, $p < 0.001$, $\eta^2 = 0.96$) with significantly lower activity magnitudes than those from the ActTrust devices at 3, 5, 7 and 9 $km \cdot h^{-1}$ ($p < 0.05$). In contrast, placement did not affect the measures for GT3X+ regardless of hip or wrist acquisitions for 3 $km \cdot h^{-1}$ and 5 $km \cdot h^{-1}$. Movement counts of both devices (GT3X+ and ActTrust) placed on the hip or wrist increased with the increment of the treadmill speed (see Figure 1). Statistical details of the ANOVA analysis and post-hoc Tukey's Honestly Significant Difference (HSD) [32, 33] tests were included in the supplementary material.

3.2 Definition of equations to estimate energy expenditure

Our main objective was to analyse the relationship between motion measurements obtained with the ActTrust® in comparison to the GT3X+ and metabolic equivalent based on gas exchange in young adults. First, we explored how treadmill speed is associated with METs and movement counts. Specifically, speed was highly correlated to both METs ($r = 0.95, p < 0.05$) and movement counts from both GT3X+ and ActTrusts devices placed either on the hip ($r = 0.94, p < 0.05; r = 0.93, p < 0.05$) or on the wrist ($r = 0.88, p < 0.05; r = 0.88, p < 0.05$), respectively. Our second step was to develop a linear regression model to estimate METs based on the movement counts recorded by each device at each placement (see Figure 2). The resulting equations are crucial for estimating physical activity in METs from the movement counts of the GT3X+ and ActTrust devices placed on the hip and wrist (Table 2).

3.3 GT3X+ and ActTrust movement count cut points to classify physical activity intensity

Finally, we derived cut-points based on MET ranges [0,3) or light, [3,6) or moderate, [6,9) or vigorous, [9,∞) or very vigorous physical activity intensity ranges based on the movement counts from both devices and placement. We then assessed the ability of the new cut points from our linear regression model to accurately classify the physical activity intensity levels objectively derived from EE measurement (METs). We estimated the sensitivity (recall), specificity, balanced accuracy, area-under-the-curve (AUC), and confusion matrix for each intensity zone (for metrics see Table 3). The obtained confusion matrices show a consistent high accuracy for the [0,3); [3,6); and [9, ∞) classes, featuring true positive rates above 75%, both when aggregating all the devices and placements (see Figure 2c) and for specific combinations of device and

Fig. 1 Scores for each treadmill speed band. a) Metabolic equivalents (VO^2) obtained from all participants for each speed band. b) Activity counts obtained for each speed band for every combination of placement and device, i.e., *GT3X+ (hip)*, *ACTT (hip)*, *ACTT (wrist)*, *GT3X+ (wrist)*. The scores across the different devices show a similar profiles increasing towards higher speeds, which is compatible to the variation exhibited by the ergo-spirometry metabolic equivalents measures.

placement (see Figure 3). Sensitivity and balanced accuracy are reduced for vigorous and very vigorous activity but remained well above 0.6, see Table 3.

Fig. 2 Cross device score analyses. a) Pearson correlation between the activity counts magnitudes measured at each device and placement combination. b) Model of Activity count and METs following the relationship *Treadmill speed*$^{1/2} \sim$ *Activity*$^{1/2} \times$ *Device*$_{placement}$. c) Confusion matrix relating predicted intensity ranges from model in b) and metabolic equivalents (VO^2) obtained from simultaneous ergo-spirometry measurements, with a high true positive rate for the classes [0,3), [3,6) and [9,∞)

Table 2 Model equations and cut-off points for each device and placement combination.

		Movement counts		
Device (placement)	Equation	3	6	9
GT3X+ (hip)	$\mathrm{MET}^{1/2} = 1.06 + 0.02\mathrm{Activity}^{1/2}$	1132	4853	9468
ACTT (hip)	$\mathrm{MET}^{1/2} = 1.11 + 0.01\mathrm{Activity}^{1/2}$	5057	23339	46410
ACTT (wrist)	$\mathrm{MET}^{1/2} = 1.23 + 0.01\mathrm{Activity}^{1/2}$	3761	22368	47203
GT3X+ (wrist)	$\mathrm{MET}^{1/2} = 1.20 + 0.01\mathrm{Activity}^{1/2}$	1698	9503	19787

Table 3 Sensitivity, specificity, balanced accurancy, and area under the curve for the [0,3), [3,6), [6,9), and [9,∞) intensity ranges

Metric	Class: [0,3)	Class: [3,6)	Class: [6,9)	Class: [9,∞)
Sensitivity	0.8810	0.9190	0.6630	0.6441
Specificity	0.9762	0.9182	0.8932	0.9556
Balanced Accuracy	0.9286	0.9186	0.7781	0.7998
AUC	0.9820	0.6307	0.7813	0.9550

Fig. 3 Confusion matrices for each combination of device and placement, i.e., a) *GT3X+ (hip)*, b) *GT3X+ (wrist)*, c) *ACTT (hip)*, *ACTT (wrist)*. The individual devices confusion matrices reproduce a similar profile to the one observed for the entire model, with a higher true positive rate for the classes [0,3), [3,6) and [9,∞)

4 Discussion

This study compared the *ActTrust®* and *ActiGraph GT3X+* activity counts during resting, treadmill walking and running. The main findings are as follows. Firstly, we established the suitability of our indirect calorimetry setup by measuring the metabolic equivalent as a function of the conditions (treadmill speed, resting). The speed of the treadmill was highly and significantly correlated with the estimated METs. Secondly, we demonstrated that movement counts increased as speed increased, and both *ActTrust®* and *GT3X+* could differentiate among the different conditions, independently of the placement, either on the wrist or the hip. As a next step, we then modelled the association between METs and activity for each device X placement using a general linear model. In addition, we established cut points for movement counts of GT3X+ and ActTrust devices to classify the intensity of physical activity. These points were derived for the categories of moderate, intense and very intense physical activity (see Table 2), considering both the movement counts of the devices and the placement. Finally, we evaluated the ability of the new cut points from the linear regression models to accurately classify the intensity levels of physical activity objectively derived from EE measurement (METs). For this analysis, we calculated the balanced accuracy, sensitivity (recall), and specificity for each cut point and each linear model (device X placement).

Acceleromery-based activity monitors are light, cost-effective, non-invasive devices that provide objective physical activity indicators. They can estimate the intensity and frequency of the activity practiced by the user and are easy to use for extended periods of time [15]. Our results demonstrated that the movement count of both devices (GT3X+ and ActTrust), placed on both the hip and the wrist, increased with the increase in the velocity of the treadmill, as did the energy expenditure (METs) obtained from indirect calorimetry. Our results are consistent with similar validation studies [34, 35]. Nevertheless, this is the first study to develop movement-based cut points for wrist and hip-worn *ActTrust®* devices that reflect moderate, intense, and very intense physical activity categories. This study established cut points for classifying physical activity intensity across the spectrum, demonstrating excellent accuracy. However, as this is the first article to report cut points for the ActTrust, it's not possible to compare these values with other studies.

If we compare our results with studies which used different devices [8, 22], while our model's sensitivity and accuracy are reduced for vigorous and very vigorous activity, it still achieves accuracies above 0.90 and higher AUC scores than previous studies [8] for identifying moderate physical exercise. Therefore, we found that the model we proposed is more accurate with a simpler method for the estimation of energy expenditure (EE) than the equations previously developed [8]. Additionally, our approach is fully explainable - as opposed to neural network models.

Accelerometers are not only used to estimate exercise or energy expenditure; they are extremely useful to assess the sleep-wake cycle based on rest-activity data. Actigraphy devices, such as the *ActTrust®* accelerometers have been used in sleep and circadian research for decades [36, 37]. Actually, the sleep and wake states have been classically inferred by algorithms based on activity counts, provided frequency, duration and intensity of movement [38]. Although sleep studies are designed to collect

24-hour data, including rest and physical activity throughout the day, the simultaneous analysis of sleep/circadian data, physical activity, and sedentary behaviour is limited [13]. It contrasts with the combined strong evidence that supports recommendations for regular exercise, reduced sedentary behaviour, healthy sleep patterns and circadian rhythms to enhance longevity, improve health, and prevent disease [39, 40].

Despite this clear dependence between physical activity and sleep, it is essential to move one step further and take advantage of the complete 24-hour rest-activity cycle data obtained through actigraphy. Precise health guidelines can only be determined by combining measurements of sleep, sedentary behaviour, and physical activity [13]. That said, expanding the availability of cut-off points for physical activity based on multiple actigraphy devices increases the potential to improve recommendations, guidelines and, ultimately, health and longevity. Therefore, studies like ours have the potential to stimulate future research by providing tools to extract physical activity and sleep information, even retrospectively. In fact, one of the main advantages of using accelerometry over self-reported (retrospective) physical activity is that it is not prone to recognition or memory biases, as recall measures are [41]. The validation of devices that were exclusively dedicated to inferring sleep-wake cycles as tools to identify and classify physical activity according to the metabolic equivalent opens up the possibility of using such data not only prospectively, but also retrospectively. Studies such as ours are crucial to developing personalised programmes that maximise long-term adherence to physical activity and the long-term benefits of exercise. We acknowledge that adults should aim for at least 150 to 300 minutes of moderate-intensity aerobic physical activity, or 75 to 150 minutes of vigorous-intensity aerobic physical activity, or an equivalent combination throughout the week, to achieve substantial health benefits [42]. Likewise, adults should usually aim for 7 or more hours of regular sleep each night to promote optimal health [43]. However, only validated accelerometer devices can provide information on these two essential proxies of health simultaneously (sleep and physical activity).

It is important to acknowledge the limitations in the findings of this study. Although a significantly large number of participants (N=56) was considered, the participants were young healthy adults, with no functional limitations. As a result, the findings may not be applicable to individuals with health conditions and other age groups. In addition, the study was carried out in a laboratory setting and focused on specific physical activities, walking and running, which may limit the generalizability of our findings. Although these activities are common among populations around the world, it is crucial to assess physical activities in real life and home environments. Further work is needed to define appropriate cut-off points for other groups, such as children and older people, as well as different device placements, as our findings are limited to devices worn on the wrist and hip.

5 Conclusion

In this study we have for the first time validated measures of the ActTrust®. When paired with a simple linear model, we can estimate energy expenditure and classify physical activity intensity. Our model was able to distinguish between light, moderate,

vigorous, and very vigorous with comparable or better benchmark metric to those available in literature - whilst adopting a much simpler and explainable model.

These results demonstrated the feasibility of using ActTrust® as a cost effective and scalable tool to estimate PA in naturalistic settings for research projects. It also has the potential for use in clinical practice, following adequate clinical tests. Our results also provide hints on how to harmonise outcomes from different devices whilst estimating PA and analysing activity measures. Future work should extend our analyses to free-living conditions and more naturalistic activity patterns.

Supplementary information.

Declarations

Author contribution. Conceptualization: ESB, ML-M. Formal analysis: LGSF, ML-M. Investigation: ESB, ABMF. Methodology: ESB, ABMF, LGSF, ML-M. Project administration: ML-M. Supervision: ML-M, JFA, ALM. Writing – original draft: ESB, ABMF, SRBG, LGSG, ML-M. Writing – review & editing: JFA, ALM, LGSF, ML-M.

Funding. We acknowledge the support from the Brazilian National Council for Scientific and Technological Development (CNPq).

Ethics approval and consent to participate. This study was approved by the Human Ethics Committee of the Federal University of Rio Grande do Norte, under protocol number 3.672.214 (CAAE: 15857419.5.0000.5537). Participants provided written consent to participate in the study.

Conflict of interest/Competing interests. The authors declare no competing interests.

Data availability. The data used in the analyses featured in this article are available at https://github.com/circadia-bio/ACTT_validation_study

Materials availability. Not applicable.

Code availability. The code used in the analyses featured in this article are available at https://github.com/circadia-bio/ACTT_validation_study

References

[1] Anderson, E., Durstine, J.L.: Physical activity, exercise, and chronic diseases: A brief review. Sports Medicine and Health Science **1**(1), 3–10 (2019) https://doi.org/10.1016/j.smhs.2019.08.006

[2] Buecker, S., Simacek, T., Ingwersen, B., Terwiel, S., and, B.A.S.: Physical activity and subjective well-being in healthy individuals: a meta-analytic review. Health Psychology Review **15**(4), 574–592 (2021) https://doi.org/10.1080/17437199.2020.1760728 https://doi.org/10.1080/17437199.2020.1760728. PMID: 32452716

[3] Bull, F.C., Al-Ansari, S.S., Biddle, S., Borodulin, K., Buman, M.P., Cardon, G., Carty, C., Chaput, J.-P., Chastin, S., Chou, R., Dempsey, P.C., DiPietro, L., Ekelund, U., Firth, J., Friedenreich, C.M., Garcia, L., Gichu, M., Jago, R., Katzmarzyk, P.T., Lambert, E., Leitzmann, M., Milton, K., Ortega, F.B., Ranasinghe, C., Stamatakis, E., Tiedemann, A., Troiano, R.P., Ploeg, H.P., Wari, V., Willumsen, J.F.: World health organization 2020 guidelines on physical activity and sedentary behaviour. British Journal of Sports Medicine **54**(24), 1451–1462 (2020) https://doi.org/10.1136/bjsports-2020-102955 https://bjsm.bmj.com/content/54/24/1451.full.pdf

[4] deJong, A.: The metabolic equivalent: reevaluating what we know about the met. ACSM's Health & Fitness Journal **14**(4), 43–46 (2010)

[5] Franklin, B.A., Brinks, J., Berra, K., Lavie, C.J., Gordon, N.F., Sperling, L.S.: Using metabolic equivalents in clinical practice. The American journal of cardiology **121**(3), 382–387 (2018)

[6] Speakman, J.: Doubly Labelled Water: Theory And Practice. Springer

[7] Migueles, J.H., Cadenas-Sanchez, C., Ekelund, U., Delisle Nyström, C., Mora-Gonzalez, J., Löf, M., Labayen, I., Ruiz, J.R., Ortega, F.B.: Accelerometer data collection and processing criteria to assess physical activity and other outcomes: a systematic review and practical considerations. Sports medicine **47**, 1821–1845 (2017)

[8] Santos-Lozano, A., Santin-Medeiros, F., Cardon, G., Torres-Luque, G., Bailon, R., Bergmeir, C., Ruiz, J.R., Lucia, A., Garatachea, N.: Actigraph gt3x: validation and determination of physical activity intensity cut points. International journal of sports medicine **34**(11), 975–982 (2013)

[9] Hees, V.T., Thaler-Kall, K., Wolf, K.-H., Brønd, J.C., Bonomi, A., Schulze, M., Vigl, M., Morseth, B., Hopstock, L.A., Gorzelniak, L., *et al.*: Challenges and opportunities for harmonizing research methodology: Raw accelerometry. Methods of information in medicine **55**(06), 525–532 (2016)

[10] Rosenberger, M.E., Haskell, W.L., Albinali, F., Mota, S., Nawyn, J., Intille, S.: Estimating activity and sedentary behavior from an accelerometer on the hip or wrist. Medicine and science in sports and exercise **45**(5), 964 (2013)

[11] Redenius, N., Kim, Y., Byun, W.: Concurrent validity of the fitbit for assessing sedentary behavior and moderate-to-vigorous physical activity. BMC medical research methodology **19**, 1–9 (2019)

[12] Neishabouri, A., Nguyen, J., Samuelsson, J., Guthrie, T., Biggs, M., Wyatt, J., Cross, D., Karas, M., Migueles, J.H., Khan, S., *et al.*: Quantification of acceleration as activity counts in actigraph wearable. Scientific Reports **12**(1), 11958 (2022)

[13] Rosenberger, M.E., Buman, M.P., Haskell, W.L., McConnell, M.V., Carstensen, L.L.: 24 hours of sleep, sedentary behavior, and physical activity with nine wearable devices. Medicine and science in sports and exercise **48**(3), 457 (2016)

[14] Bellone, G.J., Plano, S.A., Cardinali, D.P., Chada, D.P., Vigo, D.E., Golombek, D.A.: Comparative analysis of actigraphy performance in healthy young subjects. Sleep Science **9**(4), 272–279 (2016)

[15] Chen, K.Y., David R Bassett, J.: The technology of accelerometry-based activity monitors: current and future. Medicine & Science in Sports & Exercise **37**(11), 490–500 (2005)

[16] Chisholm, D., Collis, M., Kulak, L., Davenport, W., Gruber, N.: Physical activity readiness. BC Med J **17**(2), 375–8 (1975)

[17] Warburton, D.E., Gledhill, N., Jamnik, V.K., Bredin, S.S., McKenzie, D.C., Stone, J., Charlesworth, S., Shephard, R.J.: Evidence-based risk assessment and recommendations for physical activity clearance: Consensus document 2011. Applied Physiology, Nutrition, and Metabolism **36**(S1), 266–298 (2011)

[18] Schwartz, J., Oh, P., Takito, M.Y., Saunders, B., Dolan, E., Franchini, E., Rhodes, R.E., Bredin, S.S., Coelho, J.P., Dos Santos, P., *et al.*: Translation, cultural adaptation, and reproducibility of the physical activity readiness questionnaire for everyone (par-q+): the brazilian portuguese version. Frontiers in cardiovascular medicine **8**, 712696 (2021)

[19] Matlary, R.E.D., Holme, P.A., Glosli, H., Rueegg, C.S., Grydeland, M.: Comparison of free-living physical activity measurements between actigraph gt3x-bt and fitbit charge 3 in young people with haemophilia. Haemophilia **28**(6), 172–180 (2022)

[20] Ancoli-Israel, S., Cole, R., Alessi, C., Chambers, M., Moorcroft, W., Pollak, C.P.: The role of actigraphy in the study of sleep and circadian rhythms. Sleep **26**(3), 342–392 (2003)

[21] Lamarra, N., Whipp, B.J., Ward, S.A., Wasserman, K.: Effect of interbreath fluctuations on characterizing exercise gas exchange kinetics. Journal of Applied Physiology **62**(5), 2003–2012 (1987)

[22] Sasaki, J.E., John, D., Freedson, P.S.: Validation and comparison of actigraph activity monitors. Journal of science and medicine in sport **14**(5), 411–416 (2011)

[23] R Core Team: R: A Language and Environment for Statistical Computing. R Foundation for Statistical Computing, Vienna, Austria (2024). R Foundation for Statistical Computing. https://www.R-project.org/

[24] Wickham, H., Averick, M., Bryan, J., Chang, W., McGowan, L.D., François,

R., Grolemund, G., Hayes, A., Henry, L., Hester, J., Kuhn, M., Pedersen, T.L., Miller, E., Bache, S.M., Müller, K., Ooms, J., Robinson, D., Seidel, D.P., Spinu, V., Takahashi, K., Vaughan, D., Wilke, C., Woo, K., Yutani, H.: Welcome to the tidyverse. Journal of Open Source Software **4**(43), 1686 (2019) https://doi.org/10.21105/joss.01686

[25] Kuhn, Max: Building predictive models in r using the caret package. Journal of Statistical Software **28**(5), 1–26 (2008) https://doi.org/10.18637/jss.v028.i05

[26] Bartoń, K.: MuMIn: Multi-Model Inference. (2024). R package version 1.48.4. https://CRAN.R-project.org/package=MuMIn

[27] Mills, B.R.: MetBrewer: Color Palettes Inspired by Works at the Metropolitan Museum of Art. (2022). R package version 0.2.0. https://CRAN.R-project.org/package=MetBrewer

[28] Wilke, C.O.: Cowplot: Streamlined Plot Theme and Plot Annotations for 'ggplot2'. (2024). R package version 1.1.3. https://CRAN.R-project.org/package=cowplot

[29] Robin, X., Turck, N., Hainard, A., Tiberti, N., Lisacek, F., Sanchez, J.-C., Müller, M.: proc: an open-source package for r and s+ to analyze and compare roc curves. BMC Bioinformatics **12**, 77 (2011)

[30] Robinson, D., Hayes, A., Couch, S.: Broom: Convert Statistical Objects Into Tidy Tibbles. (2024). R package version 1.0.6. https://CRAN.R-project.org/package=broom

[31] Santos Batista, E., Basilio Gomes, S.R., Morais Ferreira, A., França, L.G.S., Fontenele Araújo, J., Mortatti, A.L., Leocadio-Miguel, M.A.: From Movement to METs: A Validation of ActTrust® for Energy Expenditure Estimation and Physical Activity Classification in Young Adults. Zenodo (2025). https://doi.org/10.5281/zenodo.15315735

[32] Miller, R.G.: Simultaneous Statistical Inference. Springer Series in Statistics. Springer, New York, NY (1981). https://doi.org/10.1007/978-1-4613-8122-8

[33] Yandell, B.S.: Practical Data Analysis for Designed Experiments, 1st edn. Routledge, ??? (2017). https://doi.org/10.1201/9780203742563

[34] Davoudi, A., Wanigatunga, A.A., Kheirkhahan, M., Corbett, D.B., Mendoza, T., Battula, M., Ranka, S., Fillingim, R.B., Manini, T.M., Rashidi, P., *et al.*: Accuracy of samsung gear s smartwatch for activity recognition: Validation study. JMIR mHealth and uHealth **7**(2), 11270 (2019)

[35] Skotte, J., Korshøj, M., Kristiansen, J., Hanisch, C., Holtermann, A.: Detection of physical activity types using triaxial accelerometers. Journal of physical activity

and health **11**(1), 76–84 (2014)

[36] Yuan, H., Hill, E.A., Kyle, S.D., Doherty, A.: A systematic review of the performance of actigraphy in measuring sleep stages. Journal of Sleep Research **33**(5), 14143 (2024)

[37] Sadeh, A.: The role and validity of actigraphy in sleep medicine: an update. Sleep medicine reviews **15**(4), 259–267 (2011)

[38] Fekedulegn, D., Andrew, M.E., Shi, M., Violanti, J.M., Knox, S., Innes, K.E.: Actigraphy-based assessment of sleep parameters. Annals of Work Exposures and Health **64**(4), 350–367 (2020)

[39] Morris, C.J., Yang, J.N., Scheer, F.A.: The impact of the circadian timing system on cardiovascular and metabolic function. Progress in brain research **199**, 337–358 (2012)

[40] Duncan, M.J., Murphy, L., Oftedal, S., Fenwick, M.J., Vincent, G.E., Fenton, S.: The associations between physical activity, sedentary behaviour, and sleep with mortality and incident cardiovascular disease, cancer, diabetes and mental health in adults: a systematic review and meta-analysis of prospective cohort studies. Journal of Activity, Sedentary and Sleep Behaviors **2**(1), 19 (2023)

[41] Burchartz, A., Anedda, B., Auerswald, T., Giurgiu, M., Hill, H., Ketelhut, S.I., Kolb, S., Mall, C., Manz, K., Nigg, C.R., *et al.*: Assessing physical behavior through accelerometry–state of the science, best practices and future directions. Psychology of Sport and Exercise **49**, 101703 (2020)

[42] Burtscher, J., Millet, G.P., Burtscher, M.: Pushing the limits of strength training. American journal of preventive medicine **64**(1), 145–146 (2023)

[43] Panel, C.C., Watson, N.F., Badr, M.S., Belenky, G., Bliwise, D.L., Buxton, O.M., Buysse, D., Dinges, D.F., Gangwisch, J., Grandner, M.A., *et al.*: Recommended amount of sleep for a healthy adult: a joint consensus statement of the american academy of sleep medicine and sleep research society. Journal of Clinical Sleep Medicine **11**(6), 591–592 (2015)