Title: Cortex-Wide, Cellular-Resolution Volumetric Imaging with a Modular Two-Photon Imaging Platform

Jiahao Hu[1#], Yanfeng Zhu[1#], Shoupei Liu[1#], Chengyu Li[1], Min Zhang[1], Xinyang Gu[2], Jingchuan Wu[2], Fang Xu[3], Liang Chen[2], Ying Mao[2], Bo Li[1,2*]

[1]Huashan Hospital, MOE Frontiers Center for Brain Science, State Key Laboratory of Brain Function and Disorders, Institutes for Translational Brain Research, Jinshan Hospital, Fudan University, Shanghai 200032, China

[2]Department of Neurosurgery, Huashan Hospital, Shanghai Medical College, Fudan University, Shanghai 200040, China.

[3]Interdisciplinary Center for Brain Information, Brain Cognition and Brain Disease Institute, Shenzhen Institutes of Advanced Technology, Chinese Academy of Sciences; Shenzhen, 518055, China.

*Correspondence: liboresearch@gmail.com (B. L.)

Abstract

Mapping cortex-wide neuronal activity at single-cell resolution has been limited by the physical trade-off between numerical aperture and field-of-view (FOV) in two-photon microscopes. We present Meso2P, a modular two-photon platform that decouples excitation and detection by introducing a lateral paraboloid fluorescence collector. The design sustains an effective NA 0.87 over a contiguous 6×6 mm² FOV at high speed ($2{,}048 \times 2{,}048$ pixels at 7.67 Hz). The modular platform can be upgraded with optional modules for simultaneous multi-plane imaging (1-4 planes at full resolution and speed), volumetric imaging ($6 \times 6 \times 0.5$ mm³, $2{,}048 \times 2{,}048 \times 28$ voxels at 1 Hz capturing $> 210{,}000$ neurons), and holographic two-photon optogenetic stimulation for targeted perturbations. To handle the resulting large-scale data, we provide an open-source deep-learning pipeline that automates motion correction, segmentation, and spike inference. We demonstrate cortex-wide sensory responses, layer-specific network synchrony during anaesthesia, and in-vivo tracking of micro- and nanoplastic distribution. Meso2P therefore provides a reproducible route to high-throughput volumetric imaging across almost the entire cortex with high detection efficiency.

Introduction

Understanding how spatially distributed cortical circuits cooperate to generate perception and behaviour[1-3] requires imaging tools that record activity from multiple brain regions simultaneously and with single-cell resolution[4]. Two-photon (2P) microscopy offers intrinsic optical sectioning, micron-scale resolution and compatibility with genetically encoded indicators, and has therefore become the work-horse of cellular neurophysiology[5-12]. Yet conventional designs image less than 1 mm² per frame, obliging researchers to trade breadth for detail and to rely on sequential recordings when probing cortex-wide dynamics[13].

Large-field-of-view (FOV) mesoscopes have begun to relax this constraint by scaling scan angles and using custom objectives with entrance pupils exceeding 25 mm. These instruments can span several millimetres, but the required long focal lengths inevitably lower their numerical aperture (NA ≈ 0.4-0.5)[13-21], reducing fluorescence collection efficiency, compromising deep-tissue sensitivity[22] and introducing field curvature[23]. Add-on volumetric modules—Bessel beams[24], spatial multiplexing[25], or light-field optics[17]—restore 3D coverage but inherit the parent system's limited NA and are often incompatible with 2P optogenetic stimulation. In parallel, the massive data streams (> 100 GB h⁻¹) produced by centimetre-scale imaging strain existing analysis pipelines, which were designed for thousands rather than hundreds of thousands of neurons[17, 26-28].

At the heart of these challenges lies a simple physical coupling: for a single objective with shared excitation and detection paths, NA × FOV is approximately proportional to the objective diameter ($D \approx 2 \times f \times NA$). As FOV grows, maintaining high NA demands impractically large optics and scan apertures. Because fluorescence signal in scattering tissue scales with the square of NA, any drop in collection efficiency disproportionately penalises volumetric imaging, where excitation power must already be divided across multiple planes. Breaking the NA-FOV coupling is therefore essential for a practical, cortex-wide cellular imager.

Here, we introduce Meso2P, a modular 2P platform that eliminates this trade-off

by physically separating excitation and detection. Its key advances are: first, a paraboloid lateral collection module decouples detection from excitation, preserving an effective NA 0.87 and uniform diffraction-limited resolution across a 6 × 6 mm² cortical window at high speed (2048 × 2048 pixels at 7.67 Hz); second, a modular imaging architecture supports simultaneous multi-plane imaging (1-4 planes at full resolution and speed), volumetric imaging (6 × 6 × 0.5 mm³, 2048 × 2048 × 28 voxels at 1 Hz), and two-photon optogenetic stimulation; third, an open-source deep-learning pipeline (NeuroPixelAI) motion-corrects, segments and deconvolves calcium signals from > 200,000 neurons in 100 GB datasets on a standard workstation.

Results

Implementation of Meso2P for High-Efficiency, Mesoscale Two-Photon Imaging

Meso2P integrates four plug-and-play modules for a centimetre-scale, cellular-resolution imaging and perturbation platform (Fig. 1): (1) a custom wide-angle scan relay for micron-resolution imaging across a 6 × 6 mm² FOV; (2) a lateral paraboloid fluorescence collection module that decouples detection from excitation and raises the effective NA to 0.87; (3) a multiplexing module that enables simultaneous multi-plane imaging, which can be further upgraded to volumetric imaging by adding a piezo-driven axial scanning stage; (4) a two-photon optogenetics module for cellular-resolution perturbation.

The custom scan relay delivers diffraction-limited, uniform imaging over centimetre-scale fields. Galvo-resonant mirrors scanning through ± 13° are relayed by a custom-designed optics chain comprising several scan lenses, a tube lens, an aberration compensation module, and a 4× objective (Fig. 1A, Extended Data Fig. 1B). Wavefront errors remain < 0.1 λ RMS throughout the scan range, with uniform image brightness preserving subcellular fidelity across the entire FOV (Extended Data Fig. 1C,D). Unlike widely used mesoscopes that exhibit pronounced field curvature (~ 66.7 μm/mm)[23], Meso2P maintains negligible curvature (~ 5.57 μm/mm) as validated by uniform fluorescence from thin ruler imaging (Extended Data Fig. 1E). The resulting configuration achieves diffraction-limited lateral and axial resolutions of 0.6

Fig. 1 Schematic overview of the Meso2P system. (A) Optical schematic of Meso2P. **(B)** Working principle of the lateral paraboloid collection module. **(C)** Multi-plane free imaging enabled by a time-division multiplexing module combined with piezoelectric z-scanning. **(D)** Optical layout of the SLM-based two-photon holographic stimulation module. S, shutter; HWP, half-wave plate; SL1-2, scan lenses; SLM, spatial light modulator; ZB, zero order block. **(E)** NeuroPixelAI, an integrated deep learning-based analysis pipeline for high-throughput segmentation and calcium signal extraction from mesoscale two-photon datasets. ***See also*** **Extended Data Fig. 1,2 and Video S1**.

µm and 11.7 µm, respectively, across a 6 × 6 mm² FOV (Extended Data Fig. 1F).

Alternative objectives, such as a commercial 16×, can be integrated to flexibly trade

spatial resolution for imaging area (Extended Data Fig. 1G).

Decoupling fluorescence detection from excitation breaks the NA-FOV coupling,

significantly enhancing photon collection. Conventional large-FOV two-photon

systems enlarge objective diameters to increase NA, yet remain fundamentally

constrained by the NA × FOV relationship ($D \approx 2 \times f \times NA$). Meso2P circumvents

this limit by introducing an aluminium-coated paraboloid fluorescence collection

module positioned laterally between the objective and sample (Fig. 1B, Extended

Data Fig. 1H). Excitation via the 4× objective at NA 0.30 provides adequate

micron-level resolution for large-scale imaging, while fluorescence collection benefits

from an effective NA of 0.87 across a 6×6 mm^2 area, outperforming conventional

and custom large-diameter objectives (Extended Data Fig. 1A). Uniformity in

point-spread-function peak intensity and collection efficiency (< 10% variation)

ensures consistent high-quality imaging without oversized optics (Extended Data Fig.

1I). A quantitative side-by-side benchmarking against representative large-FOV

two-photon platforms is provided in Supplementary Note 1 and Table 1-3.

The multiplexing module enables safe, high-throughput and simultaneous

multi-plane imaging. Each laser pulse is divided into four temporally separated

sub-pulses (~ 3.125 ns apart), simultaneously exciting multiple focal planes separated

by ~ 130 μm, adjustable to cortical layers of interest (Fig. 1C, Extended Data Fig.

2A,B). Crucially, the inter-plane spacing is adjustable, allowing precise targeting of

anatomically defined cortical layers—such as layers 2/3 and 5 (Fig. 5)—and enabling

concurrent imaging of functionally distinct networks at different depths. This flexible

multi-plane capability is particularly valuable for studies requiring high-resolution,

cross-laminar interrogation of distributed brain regions, a configuration that has

remained unattainable with previous large-FOV systems (Table 1). Furthermore, the

system can be upgraded to volumetric imaging by combining four-plane multiplexing

with piezo-driven axial scanning. Piezo-driven axial scanning further divides each

plane into seven axial steps (18 μm spacing), creating a $6 \times 6 \times 0.5$ mm^3 volume ($3 \times$

3×18 μm^3 voxel size) imaged at 1 Hz. Efficient single-point excitation combined

with high-NA detection enables robust imaging at a total laser power of ~ 250

mW—substantially lower than previous systems[25] and below the threshold for

heat-induced tissue damage, as confirmed by immunolabelling assays (Extended Data

Fig. 2C-F). Two-plane multiplexed imaging exhibits negligible crosstalk from

fluorescence lifetime overlap, while minor crosstalk in the four-plane mode is

effectively mitigated through a demixing process followed by analysis with the

NeuroPixelAI pipeline (Extended Data Fig. 2G).

Integrated two-photon optogenetic module ensures precise perturbation

capability. A spatial-light modulator generates user-defined holographic stimulation

patterns, dynamically scanned by galvanometric mirrors in spiral trajectories (Fig. 1D, Extended Data Fig. 1B). Imaging and stimulation beams are seamlessly co-registered via a dichroic mirror, achieving sub-micron spatial accuracy without additional alignment steps.

NeuroPixelAI provides automated, scalable processing of large-scale calcium imaging datasets. To efficiently handle extensive neuronal recordings—typically containing tens of thousands of neurons across hundreds of thousands of frames—we developed NeuroPixelAI, a dedicated computational pipeline leveraging deep learning (Fig. 1E). The pipeline automates motion correction, image denoising, neuron segmentation, calcium trace extraction, and spike inference, delivering rapid, robust, and fully automated analysis with minimal user intervention.

Characterization of Meso2P for high-speed, high-resolution mesoscale imaging

Using Camk2α-GCaMP6s mice, we first assessed single-plane performance. At 7.67 Hz and 2,048 × 2,048 pixels/frame, the system resolved individual layer 2/3 neurons 250 μm below the dura across the entire 6 × 6 mm² FOV (Fig. 2A). high-resolution mode (4,096 × 4,096 pixels, 3.83 Hz) is available when finer morphology is required. Fluorescence intensity and point-spread function were spatially uniform, enabling automated extraction of 13,586 neurons from 2000 frames in a representative animal (Fig. 2B,C).

Side-collection proved essential for this throughput. Under identical excitation power, the paraboloid mirror increased photon yield from ~ 4% (standard epi-detection) to 53%, equivalent to NA 0.87 (Fig. 2D-F). This translated directly into a ~ 10 dB increase in signal-to-background ratio and a 2.5-fold improvement in neuron segmentation yield (Fig. 2G-I, Extended Data Fig. 3A), demonstrating that decoupled detection directly enhances population imaging fidelity.

We next assessed volumetric performance. By capturing 28 axial planes across a 0.5 mm depth at 1 Hz, the system achieved high-speed, large-scale volumetric recording (Fig. 2J, Extended Data Fig. 3B). In awake mice, this configuration captured approximately 211,000 neurons spanning over 13 cortical areas (Fig. 2K, Extended Data Fig. 3C)—an order-of-magnitude improvement compared to

Fig. 2 Experimental characterizations and benchmarking of Meso2P. (A) *In vivo* calcium imaging of a transgenic mouse expressing GCaMP6s in excitatory neurons using Meso2P. Top, high-resolution mode (4,096 × 4,096 pixels, 3.83 Hz); Bottom, high-speed mode (2,048 × 2,048 pixels, 7.67 Hz). A total of 13,586 neurons were detected. **(B)** Zoom-in views of the regions marked in (A). **(C)** Heatmaps of calcium activity traces from neurons shown in (A); the zoom-in panel shows traces from 50 example neurons. **(D)** Comparison of image quality between fluorescence signals collected via the parabolic module (left) and direct detection through the objective (right). Three zoom-in views are included. **(E)** Representative neurons and calcium

traces obtained using the parabolic module (top) versus the objective (bottom). The red arrow indicates the recording time of the frames. **(F)** Comparisons of simulated collection efficiency obtained by the parabolic module (purple) and objective (pink). **(G)** Violin plot of the signal-to-noise ratios (SNR) from 10,466 neurons imaged with and without the parabolic module. ***$p < 0.001$; Tukey test. **(H)** Quantification of segmented neurons from eight randomly selected 700 μm × 700 μm regions in (D). Data presented as mean ± SD. Significance assessed by Tukey test: *$p < 0.05$,**$p < 0.01$,***$p < 0.001$; n.s., not significant. **(I)** Normalized number of segmented neurons across various segmentation parameters for data with and without parabolic module. **(J)** 3D rendering of 10-minute recording from Meso2P. Neurons are color-coded by depth. **(K)** Histogram of neuron counts extracted across 12 cortical regions. **(L)** Heatmaps of spike probabilities from 212,300 neurons in (J). Zoom-in panel shows spike probabilities of 10,000 neurons. **(M)** Proof-of-concept two-photon photostimulation in layer 2/3 of the retrosplenial cortex. Four neurons (white circles) were simultaneously targeted for stimulation. **(N)** Left, calcium transients from the four photostimulated neurons showing the responses on all individual trials. Right, simultaneously recorded activity of four non-stimulated neurons. ***See also* Extended Data Fig. 3, Supplementary Video S2**.

one-photon light-field methods imaging similar volumes[17]. Calcium traces maintainedhigh signal-to-noise ratios throughout the volume (Fig. 2L).

Finally, we verified optogenetic compatibility. Fluorescent test slides confirmed the galvo/SLM path accurately generated arbitrary holographic patterns (Extended Data Fig. 3D). *In vivo*, simultaneous two-photon stimulation of four C1V1-expressing neurons produced synchronous calcium transients restricted to targeted cells, with off-target neurons remaining inactive (Fig. 2M,N). This proof-of-concept demonstrates sub-cellular co-registration of imaging and perturbation channels within the same optical core.

NeuroPixelAI enables scalable and automated analysis of mesoscale calcium imaging

NeuroPixelAI is an open-source, GPU-accelerated pipeline that converts raw mesoscale imaging data into spike trains via automated motion registration, deep-learning denoising, region-adaptive segmentation, baseline-corrected trace extraction, and spike inference (Fig. 3A,B, Extended Data Fig. 4A). A streaming phase-correlation algorithm with dynamic memory allocation processes a 30-GB file in a single pass on a standard 128-GB workstation, while plug-in support for neural networks such as DeepCAD-RT[29] and SRDTrans[30] effectively suppresses shot noise

Fig. 3 Development and Benchmarking of NeuroPixelAI pipeline. (A) Schematic of the
NeuroPixelAI processing pipeline, comprising five sequential modules: registration, denoising,
regionalized segmentation, signal extraction, and spike inference. **(B)** Representative user
interface of NeuroPixelAI. **(C)** Comparative segmentation results on a large FOV dataset. Top left,
mean projection of raw fluorescence image. Top right, zoomed-in standard deviation projection
with NeuroPixelAI segmentation. Bottom left, Suite2p segmentation (max / mean projection).
Bottom right, CaImAn batch segmentation (mean projection). **(D)** F1 score comparison of neuron
segmentation performance across the three pipelines on a simulated small FOV (300 μm × 300 μm)
dataset generated by NAOMI with Gaussian-Poisson mixed noise. **(E)** F1 score comparison on a

large FOV dataset (6,000 μm × 6,000 μm). Ten subregions (1,500 μm × 1,500 μm) were randomly
sampled for evaluation. Mean ± SD; ***$p < 0.001$; Tukey test. **(F)** Comparisons of spike inference
outputs of NeuroPixelAI, Suite2p and CaImAn batch on the same dataset. **(G)** Correlations
between inferred spike trains and electrophysiological ground truth across NeuroPixelAI, Suite2p
and CaImAn batch. Mean ± SD; ***$p < 0.001$; Tukey test. **(H)** NeuroPixelAI achieves higher
overall performance (defined as average F1 score × correlation with ground truth) while
maintaining a compact set of seven parameters, only 2–4 of which typically require user
adjustment, outperforming other pipelines. ***See also*** **Extended Data Fig. 4, Supplementary**
Video S3.

and detector artefacts (Extended Data Fig. 4B,C). Segmentation leverages Cellpose

2.0 within a region-adaptive mask, ensuring reliable soma identification across

heterogeneous cortical areas (Extended Data Fig. 4D-F). Cubic-spline correction

(Extended Data Fig. 4G) removes baseline drift before Cascade-based spike inference

(Extended Data Fig. 4H-J). An iterative, label-free ranking module allows real-time

visualization of population activity (Extended Data Fig. 4K).

Compared to Suite2p and CaImAn[26, 27], NeuroPixelAI processes large files

directly without manual file splitting. It achieves substantial improvements in

segmentation accuracy (F1-score) on large-FOV data—by ~ 50% over Suite2p and ~

300% over CalmAn (Fig. 3C-E), and nearly doubles spike-inference fidelity on

GENIE datasets (Fig. 3F,G)[31, 32]. These advantages are achieved with only seven

adjustable parameters (typically 2-4 require minor adjustments per dataset), ensuring

a lightweight, robust workflow (Fig. 3H). Practically, a one-hour Meso2P dataset (>

200,000 neurons, > 100 GB) can be automatically processed into analysis-ready spike

trains within a few hours, transforming large-scale calcium imaging analysis from a

bottleneck into a routine laboratory task.

Cortex-wide neuronal dynamics during whisker stimulation

The mammalian neocortex integrates sensory information through distributed

cortical networks, establishing robust yet flexible regional co-fluctuation patterns

critical for perception[22, 33-37]. However, it remains poorly understood whether the

adaptive reorganization induced by repeated sensory stimuli propagates exclusively

within subnetworks or elicits global network-wide reorganization[36-39]. Conventional

wide-field approaches capture the spatial extent of such dynamics but are limited by

Fig. 4 Reorganization of whole-cortex population coding patterns during repetitive whisker stimulation. **(A)** Experimental schematic of the whisker stimulation paradigm. Each session began with a 300 s baseline, followed by ten trials of 15 s air-puff stimulus and 45 s inter-trial intervals. **(B)** Imaging field and anatomical landmarks. Scale bar, 1 mm. Inset: magnified cortical map. PL, prelimbic cortex; ACA, anterior cingulate area. M, motor cortex; S, somatosensory cortex; V1, primary visual cortex; RSP, retrosplenial cortex. The same abbreviations are used in all subsequent figures. **(C)** Standard deviation projection of the calcium imaging video showing > 10,000 neurons, overlaid with cortical area boundaries. **(D)** Heatmaps of calcium signals extracted from neurons in (C). Pink dashed lines indicate stimulus onset in each trial. **(E)** Mean numbers of neurons detected per brain region across mice (11,941 ± 3,193 neurons, mean ± SD; n = 3 mice). Gray dots represent individual animals. Inset: histogram of neurons active across all 10 trials. **(F)** Normalized activity of 4 cortical regions aligned to locomotion onset (vertical dashed line; n = 3 mice). Shaded area represents mean ± SD. **(G)** Mean classification accuracy of stimulus identity

across trials using SVM based on regional activity. Dashed lines denote stimulus onset and delay
time. Data are presented as mean ± SD (n = 3 mice). **(H)** Trial-by-trial calcium traces from
representative neurons across 4 cortical regions. Dark shading, delay period; Light shading,
stimulation period. **(I)** Bar plot of average calcium event rate of the 4 cortical regions over
baseline and 10 trials. **(J)** Top, bar plots of normalized neuron spike probability at each time point
and anterior-posterior bin in trial 1. The blue line indicates the direction of propagation of the
peak-signal arrival time. Bottom, same analysis in trial 10. The transparent pink walls mark the
peak-signal arrival time in trial 1 and trial 10, respectively. **(K)** 3D CEBRA-Time visualization
(on spike probabilities) of neural recording across ten trials in a representative mouse. The gray
arrows indicate the direction of change in the trial status. **(L)** The projected activity of each area in
the CCA dimensions identified. Top, CCA-projected activity in trial 1. Bottom, CCA-projected
activity in trial 10. Signals were projected onto pairwise CCA-derived dimensions across brain
regions and averaged (n = 3 mice). A typical change of CCA-projected activity for each area is
shown in Fig. S5K. **(M)** Each matrix shows normalized correlation coefficients, r, for CCA mode
between two of ten trials. A large matrix element value indicates that the source cofluctuated with
the target using a similar activity mode; small values imply distinct cofluctuation modes. Results
are shown for the largest CCA modes for each source-target pair, averaged over three mice.
Solid-outlined boxes highlight areas with lower relevance. Dashed-outlined box indicates stronger
correlation between adjacent trials. ***See also* Extended Data Fig. 5.**

sparse labeling and lack of optical sectioning, restricting depth resolution and the
number of neurons sampled[39, 40]. Leveraging Meso2P's large FOV and optical
sectioning, we recorded calcium dynamics from densely labeled layer 2/3 neurons
(11,941 ± 3,193 neurons; n = 3 mice) during repeated whisker stimulation (Fig. 4A-C),
enabling us to directly probe functional connectivity changes within large-scale neural
networks with high precision.

Single air-puff stimuli triggered widespread, cortex-wide activation across
multiple areas (Fig. 4D), where neurons displayed diverse stimulus tuning profiles
(Fig. 4E, Extended Data Fig. 5A,B) but with significant co-activation across different
areas (Fig. 4F). Population decoding via support-vector machines revealed stimulus
presence with > 90% accuracy for 1.5-2 s post-stimulus (Fig. 4G), and accuracy
scaled positively with neuron count (Extended Data Fig. 5C,D). Consistently, S1 and
V1 contributed more significantly and stably to the overall response pattern (Fig.
4H-I), and the stimulus-evoked information was predominantly driven by
inter-regional shared fluctuations between S1 and V1 (Extended Data Fig. 5K).

Repeated stimuli induced adaptive reorganization of cortical population activity,

decreasing calcium event rates and peak spike probabilities across all brain regions,

stabilizing after several trials (Fig. 4I, Extended Data Fig. 5E-G). Posterior cortical

areas (AP ≈ –3 to –2.5 mm) exhibited advanced response onsets (~ 0.5 s) in later trials,

indicating a temporal reorganization (Fig. 4J, Extended Data Fig. 5H,I). CEBRA

embeddings of neuronal trajectories showed directional shifts converging onto stable

attractors with repeated stimulation (Fig. 4K).

Canonical correlation analysis (CCA) revealed region-specific reorganization of

neuronal population activity (Extended Data Fig. 5J). Although stimulus information

was primarily encoded by locally shared fluctuation patterns between S1 and V1 (Fig.

4L, Extended Data Fig. 5K), global coordination across cortical regions emerged

during adaptation to repeated stimulation. This global coordination exhibited

regionally heterogeneous temporal engagement: Motor cortex (M1), S1, and V1

maintained sustained involvement, while the retrosplenial cortex (RSP) underwent

rapid reconfiguration during mid-adaptation (Fig. 4M). Critically, analysis of neuronal

loading weights across regions demonstrated that adaptive reorganization to sensory

stimuli involved not merely changes in network response magnitude, but also

restructuring of information transfer architecture within the network. This

restructuring was evidenced by directional dynamic changes in unit-specific weights,

which were no longer driven by specific local subnetworks but instead reflected

global coordinated shifts (Extended Data Fig. 5L,M).

Thus, leveraging Meso2P technology, we achieved cortex-wide

cellular-resolution recordings at high temporal resolution in densely labeled mice,

revealing network-level changes potentially underlying robust perceptual adaptation.

Its large FOV design enables high-fidelity, high-throughput spatially extended

characterization of functionally distinct cortical regions at mesoscopic scales,

facilitating neural network architecture analyses.

Simultaneous dual-layer imaging reveals laminar dynamics during anesthesia

Conscious processing is thought to rely on activity within the cerebral cortex[41-45],

whereas general anesthesia induces large-scale cortical network dynamics

characterized by widespread neuronal synchrony, particularly involving layers 2/3 and

Fig. 5 Cross-laminar recording reveals dynamic signal propagation within and across cortical layers during anesthesia induction. (A) Experimental schematic showing simultaneous two-plane calcium imaging during general anesthesia, with acquisition depths at 300 μm (cortical layer 2/3) and 560 μm (layer 5). **(B)** Representative two-plane imaging examples acquired simultaneously at dual depths, with zoom-in views showing neurons in the same cortical area at both depths. **(C)** Mean calcium event rates of all tracked neurons across different periods and depths, calculated from the initial 3-minute window per period (n = 3 mice). Error bars represent 95% confidence intervals. **(D)** Bar plot of the average number of segmented active neurons across different periods and depths. **(E)** Spike probabilities in a representative mouse. Black trace, averaged signals. Red bands, state transition periods. **(F-G)** Neuronal synchrony of all tracked neurons. Top, anesthesia protocol. Middle, neuronal synchrony in 15-s rolling window. Bottom, probability of mouse being in various states. Purple shading indicates P(awake) > P(iso); yellow shading indicates P(iso) > P(awake). Gray vertical bar marks crossing of these states. Data shown

as mean ± SD across n = 3 mice. **(H)** Violin plot of neuronal synchrony across different periods and depths (n = 21387 active cells from three mice). ***$p < 0.001$; Tukey test. **(I)** Top, Granger causality matrix across 16 cortical bins during anesthesia. First column shows causality from the most posterior bin to other bins. Causality with $p < 0.05$ (Granger F test) was shown after correction for multiple comparisons to control false discovery rate. Bottom, summary map of causality between cortical bins. Red arrows indicate inferred causality direction. **(J)** Schematic of synchronized signal propagation direction in a representative mouse. Top, mean spike probabilities during anesthesia induction. Red arrows indicate synchronized spikes expanded in the middle and bottom panels. Middle, bar plots of normalized neuronal spike probability across anterior-posterior positions, showing peak timing sequences in four synchronized events. Bottom, single-neuron-resolution response delay map. Neurons are color-coded by their peak time relative to population activity. (Yellow: early-peaking source regions; purple: late-peaking target regions). Full replication in mice shown in Fig. S5C (n = 3 mice). **(K)** Boxplot of signal propagation direction. For each detected synchronized event, peak timing differences were computed between anterior (8 bins) and posterior (8 bins) regions, then averaged to determine information flow (n = 3 mice), *$p < 0.05$, Tukey test. **(L)** Heatmap of normalized CCA mode correlation matrices between consecutive synchronized event clusters, revealing gradual reorganization of neuronal ensemble activity patterns during anesthesia. **_See also_ Extended Data Fig. 6, Supplementary Video S4.**

5. Notably, synchrony changes in layer 5 neurons are temporally aligned with the loss and recovery of consciousness[41-45]. Capturing this phenomenon across multiple cortical layers simultaneously, at cellular resolution and mesoscale coverage, remains challenging with existing approaches.

Using Meso2P, we performed simultaneous mesoscale imaging of layers 2/3 and 5 in head-fixed, isoflurane-anesthetized mice, with imaging planes at ~ 300 μm and ~ 560 μm below the cortical surface covering prefrontal and posterior cortices (Fig. 5A-B). Throughout both induction and recovery phases, the system reliably captured large populations of active neurons from both layers with high signal fidelity (Fig. 5C-D). Notably, both layers exhibited prominent large-scale synchrony across millimeter-scale regions during anesthetic induction (Fig. 5E)—consistent with previous studies[45-48]—with this effect being particularly pronounced in layer 5 (Fig. 5F-H). Unexpectedly, leveraging Meso2P's cross-layer, high-speed and large-scale imaging capability, we observed asynchronous information flow within global synchronization events: individual "synchronous" events contained distinct spatiotemporal origins and propagation pathways, with near-complete global synchrony emerging only during late anesthesia. Granger causality analysis revealed a

predominant posterior-to-anterior directional bias during induction (Fig. 5I). To

validate this finding, we analyzed the spatiotemporal sequence of individual

synchronized events by aligning normalized calcium activity across 16 contiguous

cortical windows within a $\pm$ 2 s window centered on peak activation (Fig. 5J,

Extended Data Fig. 6A,B). Although activity appeared globally synchronized,

consistent temporal offsets revealed heterogeneity in initiation sites and propagation

trajectories. As anesthesia deepened, this variability progressively diminished, with

Layer 5 neurons achieving near-complete synchronization exclusively during

late-stage events (Fig. 5J, Extended Data Fig. 6B,C). Quantitative analysis of spike

onset distributions confirmed posterior-to-anterior propagation as the dominant

direction (Fig. 5K).

CCA analysis further revealed progressive divergence in global synchrony

patterns and substantial reconfiguration of neuronal loading weights across trials (Fig.

5L; Extended Data Fig. 6D).

Dual-layer imaging also allowed direct interrogation of interlaminar interactions.

Activity predominantly propagated in a bottom-up manner from layer 5 to layer 2/3 (p

< 0.005), with occasional bidirectional interactions (Extended Data Fig. 6E,F).

In summary, these findings show that anesthesia-induced cortical synchrony

emerges primarily in deep layers and propagates in a posterior-to-anterior and

bottom-up fashion, accompanied by dynamic reconfiguration of interlaminar

interactions—significantly departing from the minimal delay synchronization

proposed in prior studies[41-45]. With its capability for high-speed, high-precision,

large-scale and multi-layer free imaging, Meso2P exhibits unique potential for

elucidating information propagation mechanisms within macroscopic cortical

networks. This provides novel perspectives for investigating laminar organization

during anesthesia and other altered states of consciousness.

Dynamic monitoring of MNP distribution and vascular occlusion *in vivo*

Micro- and nanoplastics (MNPs) have rapidly become widespread environmental

contaminants, increasingly detected in human tissues including the brain[49].

Conventional analytical methods (e.g., Py-GC-MS, Raman imaging) quantify MNPs

Fig. 6 Dynamic monitoring of MNP distribution and vascular occlusion in the living brain.
(A) 28-layer MNP imaging in head-fixed mice using Meso2P. **(B)** Vascular resolution in 28-layer
Meso2P imaging: FITC-Dextran (green) and RITC-Dextran (red) labeled vessels. **(C)** Distribution
pattern of 5-µm MPs in the brain after injection: vessels (green), MPs (red), co-localization
(yellow). **(D)** MP distribution at different time points in volumes of 6,000 × 6,000 × 450 µm³ (top)
and 1,600 × 1,600 × 450 µm³ (bottom): vessels (green), MPs (red), co-localization (yellow),
co-localized MPs observed at both time points (pink arrows). **(E)** Comparison of blood flow
velocity in MP-treated vessels; data sources a-d shown in Fig. S7B. **(F)** Depth distribution of MPs

in vessels of different diameters across cortical layers 1–4 (top to bottom) at two time points. **(G)** Spatiotemporal distribution of MPs in vessels of different diameters at two time points. **(H)** Turnover rates of MPs across layers, and proportions of occlusion, clearance, and re-occlusion between layers. **(I)** Distribution pattern of 200-nm NPs in the brain after injection: NPs adhering to vascular walls (pink arrows), within NP-laden cells (yellow arrows), extravascular distribution (white dashed circles), and occluding vessels (white dashed rectangles); vessels (red), NPs (green). **(J)** NP uptake by NP-laden cells: pre-uptake (white solid line) and post-uptake (white dashed line) position comparison (yellow arrows). **(K)** NP distribution at different time points in volumes of 6,000 × 6,000 × 450 μm³ (top) and 1,600 × 1,600 × 450 μm³ (bottom): vessels (red), NPs (green), co-localization (yellow), co-localized NPs at both time points (green arrows). **(L)** Comparison of vascular occlusion between MPs and NPs. *Scale bars*: 500 μm (B, D, K); 100 μm (C); 20 μm (I, J). ***See also* Extended Data Fig. 7 and Supplementary Videos S5-9**

but lack temporal resolution, limiting their use in dynamic biological studies[50]. Standard two-photon microscopy, constrained by its small imaging field, also struggles to capture real-time, three-dimensional MNP distribution and cerebrovascular interactions.

We leveraged Meso2P's large-FOV volumetric imaging to first delineate cortical vasculature of C56BL/6J mice recovered from surgery with vascular dyes (Extended Data Fig. 7A), then dynamically tracked fluorescent 5-μm microplastics (MPs) and 200-nm nanoplastics (NPs), revealing their distinct spatiotemporal accumulation patterns and cerebrovascular consequences (Fig. 6A,B).

MPs entered cortical vessels rapidly (~ 13 s post-injection), causing immediate vascular occlusions (Fig. 6C, Supplementary Video 5). Occluded vessels exhibited reduced blood flow, compensated by increased flow in adjacent vessels (Fig. 6D; Extended Data Fig. 7B). Compared to the 12-hour time point, partial vascular occlusions were resolved by 24 hours, while some persisted for at least 12 hours (Fig. 6E). MPs preferentially accumulated in smaller vessels (< 30 μm diameter), particularly within deeper cortical layers, and either persisted or re-occluded after partial clearance (Fig. 6F-H, Extended Data Fig. 7C, Supplementary Video 6).

In contrast, smaller-sized NPs diffused broadly upon vascular entry, forming diverse aggregates (granular, aggregated, and cellular) within vessels and surrounding tissues (Fig. 6I, Extended Data Fig. 7D, Supplementary Video 7). NPs adhered to vascular walls, occasionally extravasated, and progressively accumulated within

putative phagocytes with time-dependent increase ("NP-laden cells", Fig. 6J, Extended Data Fig. 7E, Supplementary Video 8). NP-induced vascular occlusions, though fewer, were notably persistent (~ 60% unresolved; Fig. 6K,L). Vessel wall adhesion promoted recurrent clustering, stabilizing occlusions despite NPs' small size (Extended Data Fig. 7F, Supplementary Video 9).

Thus, Meso2P offers a powerful platform for dynamic, cellular-resolution analysis of complex *in vivo* processes. Its volumetric imaging capacity captures three-dimensional particle distributions and biological responses in real time, enabling new investigations into a wide range of questions that demand large-scale, three-dimensional coverage.

Discussion

Meso2P resolves a long-standing optical bottleneck in two-photon microscopy by physically separating excitation and detection, effectively breaking the inverse relationship between numerical aperture and FOV. A compact paraboloid collector boosts the effective detection NA to 0.87, while excitation remains delivered through a standard 4× objective. As a result, diffraction-limited performance is maintained uniformly across a contiguous 6 × 6 mm² imaging window, avoiding the use of oversized custom optics. This innovative approach converts high-NA collection from a mechanical limitation into a flexible optical design choice.

The volumetric implementation pushes single-beam point scanning toward its physical limits. Time-division multiplexing increases the effective laser repetition rate from 80 MHz to 320 MHz, remaining just below the 3.125-ns fluorescence lifetime of common calcium indicators, thus minimizing inter-plane crosstalk. Piezo-driven axial stepping distributes excitation pulses across a 520-μm depth range at 1 Hz. The resulting 6 × 6 × 0.5 mm³ imaging volume optimally balances sampling density, tissue heating, and indicator kinetics. Achieving greater imaging speed would likely necessitate parallel excitation approaches, potentially compromising resolution and signal fidelity in scattering tissues. Crucially, Meso2P's modular design—comprising lateral collection, volumetric imaging, and holographic stimulation modules—allows

users to flexibly adapt the instrument according to experimental requirements and to

incorporate additional functionalities as imaging technologies evolve.

Handling the data rate of cortex-wide imaging motivated the development of

NeuroPixelAI. By streaming registration, frame-wise deep denoising and

region-adaptive segmentation through memory-aware kernels, the pipeline processes

full-session files on a single workstation and returns spike trains with higher

segmentation and inference fidelity than widely used alternatives. Its open-source

code base and modular design invite adoption of new models as they appear and

lower the barrier for laboratories that lack extensive computational expertise.

Although we primarily demonstrated Meso2P's capabilities through

neuroscience applications, its optical core is versatile and equally suited to imaging

dynamic biological processes such as immune-cell trafficking, tumor-stromal

interactions, or embryonic morphogenesis, wherever large-scale coverage with

single-cell resolution is required.

Several limitations remain. The 11.7-μm axial point-spread function approaches

the diameter of a cortical soma, so applications that demand sub-laminar separation

will still benefit from high-NA objectives. The collector sits above the cranial window

and currently precludes integration with head-mounted devices, which will require

miniaturisation or fold-mirror designs. The 3.125-ns inter-plane delay is a

compromise between crosstalk and multiplexing depth; extending it to six

nanoseconds would sharpen temporal separation but halve the number of

simultaneous planes. NeuroPixelAI analyses each imaging plane independently; a

fully three-dimensional model could further improve source separation in densely

labeled tissue. Finally, while we confirmed accurate co-registration of imaging and

holographic stimulation, we did not explore closed-loop perturbations, leaving ample

scope for future circuit manipulation experiments.

Collectively, Meso2P demonstrates that decoupling fluorescence detection from

excitation is a powerful strategy to achieve NA-limited performance across mesoscale

FOV. Its modular hardware–software architecture provides a reproducible and flexible

framework for next-generation instruments aimed at comprehensive imaging,

perturbation, and analysis of biological activity across extensive cortical volumes in real time.

Methodological Details

Animals

All mouse experiments were approved by the Institutional Animal Care and Use Committee of the Department of Laboratory Animal Science, Fudan University, Shanghai, China (ethical approval no. 2022JSITBR-013). Mice were housed in a controlled environment (~ 20 °C, 45% relative humidity) under a 12 h light/dark cycle (lights on 08:00, off 20:00). Unless specified, both male and female mice transgenically expressing GCaMP6s in excitatory neurons (*Camk2α-Cre*, Cyagen Biosciences #C001015 × *fl-GCaMP6s*, Jackson Labs #031562) underwent cranial window implantation at 8-10 weeks. Briefly, under 1.5% isoflurane (RWD Life Science #R510-22-10) anesthesia (mixed with oxygen), craniotomy was drilled (7 mm). The skull flap was removed while preserving meningeal integrity. After debris clearance, a cranial window (6.5-mm coverslip bonded to a glass ring: 7.5-mm outer diameter, 6-mm inner diameter) and a metal headpiece fixed with dental cement/cyanoacrylate glue were implanted. Mice with major intraoperative hemorrhage were excluded. After 2-week recovery (daily dexamethasone, Beyotime Biotechnology #C001015 and cefazolin, Shanghai Macklin Biochemical Technology #HY-B0712A, 5 mg/kg, i.p.), mice were adapted to head fixation for 2 weeks before imaging. Unless specified, head-fixed imaging used a parabolic lateral collection module and custom 4× objective. At the end of the experimental procedures, unless specified, experimental animals were euthanized by CO_2 asphyxiation followed by cervical dislocation. Neuronal imaging signal extraction was performed using NeuroPixelAI.

System Configuration

Laser Source

The laser system consisted of two components: an imaging laser source and a photostimulation laser source. For imaging, we employed the Chameleon Discovery

NX laser system (Coherent, Inc.), which features a tunable wavelength range from 660 nm to 1,320 nm, a pulse duration of 100 fs, and a repetition rate of 80 MHz. At a wavelength of 920 nm, the system delivers an average output power of 3,200 mW. First, we used the laser's integrated dispersion compensation module to generate an anomalous group delay dispersion (GDD) of -17,358 fs^2, compensating for material dispersion from system components. We then adjusted the laser output power using a rotatable half-wave plate (AHWP05M-980, Thorlabs) combined with a polarizing beam splitter (PBS255, Thorlabs), equipped with a precision rotation mount (PRM1) and a voice coil motor controller (KVC101, Thorlabs). For photostimulation, we used a Monaco 1035 laser system (Coherent)..

Spatiotemporal Multiplexing Module

To increase the system's sampling rate to 320 MHz, thus reaching the sampling limit for fluorescence imaging, we constructed a polarization-based spatiotemporal multiplexing system (Extended Data Fig. 2A,B). The system initially divides the incident light into four beams using a combination of three polarizing beam splitters (PBS623, LBTEK) and half-wave plates (HWP25-905A-M, LBTEK). Each half-wave plate is mounted on a rotatable adjustment frame (CRM-1AS) to control power allocation among the optical paths. Each optical path was then assigned a distinct time delay, resulting in the sequential emission of the four light beams at 3.125 ns intervals. Furthermore, the system incorporated a pair of cemented lenses (AC250-150-B-ML and AC250-300-B-ML, Thorlabs) arranged as a magnification telescope. This arrangement enabled beam enlargement while allowing focal position adjustment using a precision translation stage (A-LSQ-E, Zaber), thereby creating four adjustable focal points on the imaging plane. Finally, the optical paths were combined using two beam splitters (BS1255-B, LBTEK) and a polarizing beam splitter, producing a composite light beam characterized by a 320 MHz sampling frequency, 3.125 ns temporal intervals, and a 130 μm spatial separation between focal points. The power of each beam within the composite light and the spacing between focal points could be adjusted via the corresponding half-wave plates and by moving the lenses on the translation stage, respectively, to meet varying imaging requirements.

Excitation Optical Pathway

The excitation optical pathway of Meso2P is shown in Fig. 1A and Extended Data Fig. 1B,D. The optical path begins with a Y-axis scanning galvanometer mirror (6215H, Cambridge Technology) at the system forefront. The beam then passes through a telescope comprising of two telecentric flat field scanning lenses (LSM54-850, Thorlabs) before being incident on a resonant mirror (CRS8K, Cambridge Technology) for X-axis scanning. This separated-axis design prevents sampling depth changes from affecting the XY field of view (FOV). After the scanning process, the beam undergoes aperture magnification through a 3.4x magnification telescope composed of a telecentric flat field scanning lens (LSM54-850, Thorlabs) and two doublet achromatic lenses (AC508-300-B), followed by a custom aberration compensation module before entering the rear aperture of the objective lens.

Aberration Compensation Module

In Meso2P, most optical modules employ commercial low-aberration components, but cumulative aberrations from optical mismatches still significantly degrade image quality. To correct multiple types of aberrations present in the system and enhance the fluorescence excitation efficiency, we designed a symmetric cemented lens pair with large air gaps, inspired by the Petzval lens structure. This pair of lenses, after design optimization, features an extended focal length as given by:

$$\phi = \phi_1 + \frac{h_2}{h_1}\phi_2 \tag{1}$$

here, ϕ_1 and ϕ_2 represent the focal powers of the front and rear lens groups, respectively, and ϕ denotes the total focal power of the module. This design allows for significant optimization of spherical aberration, field curvature, and astigmatism without affecting the original focal position of the system. The design primarily focuses on optimizing the flatness of the imaging field quality, ensuring that the overall composite aberration RMS across the imaging field remains within 0.1 ± 0.01 λ at the operational wavelength of 920 nm $\pm$ 18 nm.

Parabolic lateral collection module

The parabolic lateral collection module consists of six components: a parabolic

reflector module, a first focusing lens (LA1145-A, f = 75 mm, Thorlabs), a filter cube

module (including custom-sized low-pass filter ET720sp-2p8, high-pass filter

T556lpxs-UF3, and band-pass filters for red and green wavelengths ET605/70m and

CT512/60bp, Chroma Technology), a second focusing lens, a third focusing lens

(48-244-INK, f = 50 mm, Edumund), and a photomultiplier module (H13543-300,

HAMAMATSU) (Fig. 1D and Extended Data Fig. 1G).

Due to the constraints imposed by the aperture of optical components, the

imaging field of view (FOV), and the focal length of the objective, the parabolic

reflector module was meticulously designed (Extended Data Fig. 1H). The focal

length p and the distance behind the focus L are subject to the following limitations:

611
$$\begin{cases} f_{\mathrm{obj}} < 33\mathrm{mm}: p < 33, \\ W_{\mathrm{laser}} = 8\mathrm{mm}: p > 8, L > 4, \\ D_{\mathrm{Rlens}} = 52\mathrm{mm}: y(p + L/2) < 26. \end{cases} \tag{2}$$

Furthermore, to ensure that the fluorescence collection efficiency is nearly uniform

across different points on the imaging plane, the distance before reflection along path

1 should be relatively long, necessitating a maximally large p. Additionally, to enable

the collection module to gather fluorescence at larger angles, the angle θ along path 2

should be relatively small, which requires L to be maximally large while keeping p

as small as possible. A grid search method was employed to balance these constraints,

ultimately selecting the module parameters p = 20 and L = 8.

Given the absence of commercially available 4× long-working-distance

objectives suitable for large-field multiphoton fluorescence imaging with a working

distance greater than 30mm, a custom objective lens was designed and fabricated to

accommodate the parabolic lateral collection module (Fig. 1B and Extended Data Fig.

1G). This objective uses a five-group, seven-element design, equipped with an M32 ×

0.75 thread, and is mounted on a piezo objective scanner (PFM450E, Thorlabs),

allowing for independent axial scanning. The objective has a focal length f = 45 mm,

a numerical aperture NA = 0.296, and a working distance of 31 mm. Targeted

optimizations were performed to minimize spherical aberration, astigmatism, and

meridional field curvature, ensuring that the composite wavefront aberration RMS

across the entire field at the operational wavelength of 920 nm remains below 0.02 λ.

Using Zemax-based optical system simulations, this lens approaches the performance

of an ideal objective. Notably, due to the customizable structure of the self-designed

objective, adjustments were made during simulation to further enhance image quality,

aligning the lens optimally with the system.

Multiplexed Acquisition of Quadruple Electrical Signals

Due to the frequency limitations of the data acquisition card (vDAQ, MBF

Bioscience), the maximum signal acquisition rate of the system is constrained to 80

MHz. To facilitate the sampling of a 320 MHz electrical signal using this acquisition

card, a strategy involving the multiplexed acquisition of quadruple electrical signals

was implemented. During the imaging process, the fluorescence signal is first

converted into an electrical signal by a photomultiplier tube (PMT, H7422-40,

Hamamastu), which is then amplified by an amplifier. The output from the amplifier

(C9663, Hamamastu) is fed into a four-way power splitter (PD-0/6-4S,

TD-MICROWAVE), which divides the signal into four pathways. To achieve

temporal displacement among the multiple acquisition points, different cable lengths

are used based on the rate of electrical transmission through the conductor. This

arrangement allows the four separate channels of the acquisition card to capture the

four components of the 320 MHz signal sequentially. The difference in cable lengths

can be estimated using the following equations:

$$v = \frac{1}{\sqrt{\varepsilon_0 \varepsilon_r \mu_0 \mu_r}} \tag{3}$$

$$\Delta L = v \cdot \Delta t \tag{4}$$

651 where ε_0 is the permittivity of free space, 8.89×10^{-12} F/m8, ε_r is the relative

652 permittivity of the material, μ_0 is the permeability of free space, $4\pi \times 10^{-7}$ H/m, μ_r

653 represents the material's relative permeability, and Δt is 3.125 ns.

Denoising Algorithm Testing

655 To evaluate the performance of various denoising algorithms, we compared their

656 effects on data with a 0 dB signal-to-noise ratio (SNR) across different frame rates

657 (Extended Data Fig. 4B), where SNR is defined as the ratio of signal power to noise

658   power:

$$\text{SNR} = 10lg\frac{P_\text{S}}{P_\text{N}} \tag{5}$$

where P_S represents the signal power, and P_N denotes the noise power.

We first acquired clean calcium imaging videos at 100 Hz frame rate, 200 μm

depth, and 300 × 300 μm² FOV using NAOMI[51]. These datasets were then

downsampled to generate frame rates spanning 1-30 Hz, followed by noise injection

to simulate experimental conditions. The noise inherent in calcium imaging data

predominantly consists of photon shot noise, dark current noise, and readout noise[52].

Among these, readout noise is characterized by a Gaussian distributio and the power

of dark current noise is relatively low and stable[53]. The principal component of noise,

however, is the unavoidable shot noise, which occurs due to quantum fluctuations

during the photon acquisition process and is also known as Poisson noise. The

theoretical noise model can thus be described as :

$$I_\text{acq} = P(I_\text{clean}) + N(0, \sigma^2) \tag{6}$$

672   Considering digital resolution(image bit depth), the generation of MPGnoise can be

673   reformulated as:

$$I_\text{acq} = \text{MPG}(I_\text{clean}; Q, n, \sigma^2) = \frac{2^n-1}{Q} * P(\frac{I_\text{clean}*Q}{2^n-1}) + N(0, \sigma^2) \tag{7}$$

here, I_clean represents the photon distribution in the pristine data, and I_acq represents

the corresponding noise-contaminated photon distribution. P is a function operator

of the signal-related Poisson noise, N represents signal-independent zero-mean

Gaussian noise with a variance characteristic σ^2, n denotes the bit depth of the

captured image (set here as n = 16), $\text{MPG}(\cdot)$ is the operator generating MPG noise,

used to simplify the mathematical expression and Q represents the relative number

of photons per pixel captured by the detector. In this work, different Q values were

used to simulate calcium imaging data with various SNRs. Given that Gaussian noise

constitutes a smaller component of the total noise, its mean is set to zero, and its

variance is set to $\sigma^2 = 1000$, hence $\sigma = 31.623$.

To calculate SNR in calcium imaging videos, for a noise-free calcium imaging

video, we calculate the signal power by squaring the intensity of each pixel in every

frame throughout the video and then dividing by the total number of pixels. This can

be expressed specifically as:

689
$$P_{\mathrm{s}} = \sum\nolimits_{\mathrm{pixs}} \frac{I^2}{\mathrm{pixs}} \tag{8}$$

690   where I represents the pixel value of a single pixel in a particular frame, and pixs is

691   the sum of the number of pixels across all frames. After calculating the signal power,

692   similarly, the noise power is determined by computing the squared difference between

693   the pixel values of the noisy video and the noise-free video at each point, summed and

694   then divided by the number of pixels:

695
$$P_{\mathrm{N}} = \sum\nolimits_{\mathrm{pixs}} \frac{(I - I_n)^2}{\mathrm{pixs}} \tag{9}$$

where I_n denotes the pixel value in the noisy video for a single pixel in a particular

frame. Finally, the SNR is calculated according to Equation (5).

Assessment of Cell Segmentation Algorithms

Neuronal spatial segmentation is typically framed as a multi-class task, assigning

each pixel to its corresponding neuron. To evaluate algorithm performance, we used

standard metrics: Precision, Recall, and the F1 Score (Extended Data Fig. 4D-F).

Furthermore, the computational speed of the segmentation algorithms was included as

a performance benchmark.

Precision, also known as positive predictive value, represents the proportion of

true positives among the samples predicted as positive. It can be described by the

following equation:

707
$$\mathrm{Precision} = \frac{\mathrm{TP}}{\mathrm{TP+FP}} \tag{10}$$

708   where TP is true positive, FP is false positive.

709   Recall, also referred to as sensitivity, indicates the proportion of true positives

710   out of all actual positive samples. It can be represented as:

711
$$\mathrm{Recall} = \frac{\mathrm{TP}}{\mathrm{TP+FN}} \tag{11}$$

712   where FN is false negative.

713   The F1 Score, which emerges from the concepts of Precision and Recall,

714   typically presents a trade-off between these metrics: high Precision often coincides

715   with lower Recall, and vice versa. To provide a balanced consideration of these

716   metrics, the concept of the F1 Score was introduced:

717
$$F1 = \frac{2*\text{Precision}*\text{Recall}}{\text{Precision}+\text{Recall}} \tag{12}$$

Regarding the processing speed of the algorithms, we have simply evaluated it in

frames per second.

The specific method of testing employed cross-validation. In the experimental

setup, we used six sets of calcium imaging videos at 3.51 Hz with 6,500 frames (512

× 512 pixels), manually annotating nearly 600 ROIs as ground truth using Cellpose

2.0. For algorithms requiring network training, five sets of data were used for training

and the remaining set served as the test set, completing the evaluation across all six

video sets in a repetitive manner. For algorithms that did not require training,

parameter adjustments were made prior to testing. During each algorithm's testing

phase per video set, we concurrently recorded Precision, Recall, F1 Score, and image

processing speed as metrics of evaluation.

Baseline Correction Algorithm

In certain experiments, we observed a concurrent reduction in neuronal action

potential firing and total $[Ca^{2+}]_i$ fluctuations, manifesting as baseline variations in the

recorded data. To accurately extract spikes from individual neurons despite ongoing

baseline fluctuations, we incorporated a baseline correction algorithm based on Cubic

Spline Approximation (Smoothing) (csaps) into our software. Upon inputting data, the

smoothing parameter of csaps is adjusted to smooth = 1e-8. The data fitted using this

method serve as the baseline for calculating the change in neuronal fluorescence,

ΔF/F0:

738
$$\text{dff}_{\text{sig}} = \frac{\text{raw}_{\text{sig}}-\text{baseline}}{\text{baseline}} \tag{13}$$

739   where baseline is the baseline obtained by fitting the original data using the csaps

740   algorithm, raw_{sig} represents the original data, and dff_{sig} denotes the final calculated

741   change in neuronal fluorescence intensity *ΔF/F0*.

Spike inferrence Algorithm Testing

To evaluate algorithm performance, we computed the Pearson correlation coefficient between electrophysiologically recorded membrane potentials and inferred spike trains from original calcium imaging data processed by each algorithm. Analyses were performed across multiple standardized noise levels (Extended Data Fig. 4H-J).

Specifically, each algorithm processed original calcium signals to generate inferred spike probability time series. These inferred traces were then compared to simultaneously recorded membrane potential signals, which were converted into standardized spike probability estimates using established methods. Correlation coefficients served as the primary metric to: (1) compare algorithm performance across different standardized noise levels, and (2) assess the stability of each algorithm's performance at a fixed noise level.

The definition of standardized noise level is given by[54]:

$$\upsilon = \frac{\sigma_{\Delta F/F}}{\sqrt{f_\mathrm{r}}} = \frac{Median_t|\Delta F/F_{t+1} - \Delta F/F_t|}{\sqrt{f_\mathrm{r}}} \tag{14}$$

here, based on optical imaging principles with finite granularity noise where the detector collects N photons per second, the noiae magnitude is $\sqrt{N}$. Thus, the baseline noise level is expressed as: $\sigma = \frac{1}{\sqrt{N}}$ (in percentage units). Considering the frame rate, f_r, for each frame, the formula's N can be replaced entirely by N/f_r, resulting in a baseline noise level per frame of: $\sigma_f = \frac{\sqrt{f_\mathrm{r}}}{\sqrt{N}}$. To eliminate the factor of frame rate, the standardized noise level υ is defined as the Equation (14).

To convert the potential signals into spike probabilities, we performed binarization followed by Gaussian kernel convolution. The original spike timing data in Fig. 3F and Extended Data Fig. 4H show discrete signals. After binarization and convolution processing, the resulting spike probabilities are used to evaluate the performance of the algorithms.

Iterative Correlation Sorting Algorithm

Even in the absence of recorded specific stimulus times during experiments,

neuronal signals can be sorted by their inherent temporal characteristics of local firing

potentials. Initially, the data from all neurons are aggregated into a single time series

(representing the overall temporal signal of all neurons). The100 largest-amplitude

time points from this series as template for cross-correlation-based alignment of all

neuronal signals. Subsequently, the average of the top 10 most highly correlated

neuronal signals post-alignment is taken as the new template to reorder all neuronal

signals. After 30 iterations, the resulting correlation sorting plot appears in the

lower-right image frame of software Tab 2.

Development and Implementation of NeuroPixelAI for Large-Field Image

Analysis on a High-Efficiency Software Platform

In order to address the challenges associated with processing extremely large

image datasets from wide-field observations and to furnish a platform capable of

one-click, high-speed processing, we developed NeuroPixelAI using PyQt5. The

workflow is compartmentalized into six distinct stages: motion registration, image

denoising, cell segmentation, signal extraction (baseline correction), spike inference,

and correlation-ranking clustering. Given the dynamic changes inherent to biological

activity, tissue expansion, and mechanical fixation during imaging processes, there

often occurs displacement within the imaging field. To counteract this, our initial step

involves motion registration. We adopted a phase-correlation-based registration

method, similar to that used in Suite2p[26], which iteratively computes the average of

the most correlated frames in a video to serve as a reference image. Subsequent

frames are then aligned with this reference image to achieve motion registration. We

enhanced this algorithm by incorporating dynamic memory allocation, allowing it to

handle larger data structures effectively. Moreover, we reduced the adjustable

parameters from twenty to four input parameters, and introduced options for non-rigid

registration and bidiphase adjustment, significantly reducing parameter redundancy

and simplifying usability. The second step is image denoising. Calcium imaging data,

due to the transient, low-amplitude accumulations of intracellular calcium ions and

the rapid dynamics involved, coupled with the limitations of electronic detection

equipment, are susceptible to various types of noise contamination, such as shot noise,

dark current noise, and readout noise[53]. This poses significant challenges for

high-performance imaging, such as dendritic imaging and deep tissue imaging. We

referred to algorithms such as DeepCAD-RT[29] and SRDTrans[30] (Extended Data Fig.

4B,C), which are known for their speed, precision, and parameter-free operation.

Upon completion of their training phase, these algorithms can effectively eliminate

various types of white noise from multiphoton image data. Consequently, we

integrated the denoising function with the prior motion registration into a unified

module (Image Enhancement) within NeuroPixelAI and included visualization

capabilities. The third step involves cell segmentation. To effectively handle the

diverse cellular morphologies and distributions across different brain regions and to

facilitate efficient one-click processing, we developed an optimized processing

pipeline. Building upon Cellpose 2.0[55], we implemented parameter optimization and

established a region-adaptive framework to achieve high-precision, high-throughput

cellular segmentation across large fields of view. The streamlined workflow enables

efficient one-click processing within NeuroPixelAI, while retaining manual curation

capabilities for user-guided refinement. Algorithm performance was validated against

multiple alternatives (CNMF-E[27, 56], Cellpose[26, 55], SUNS[57]) using both raw and

denoised data (Extended Data Fig. 4D-F). The fourth stage in our methodology is

signal extraction. Based on the cellular masks generated during the third stage of cell

segmentation, we sum the pixel values within each ROI in every frame to obtain the

raw temporal data of individual neurons. Additionally, we offer an optional baseline

correction method utilizing the cubic spline interpolation algorithm. This approach is

designed to correct baseline fluctuations in the signals, which may arise due to

changes in the concentration of optical indicators within cells, photobleaching, or

specific biological activities (Extended Data Fig. 4G). This method is particularly

advantageous for subsequent spike inference during extended recordings of multiple

signals. The final stage involves spike inference. Commonly used optical imaging

signals, particularly calcium signals, serve as indirect, typically non-linear, and

low-pass filtered representations of more fundamental variables of interest—namely, somatic action potentials (spikes)[54, 58-60]. Historically, many studies have analyzed raw calcium signals directly[39, 61] or employed various deconvolution algorithms, such as OASIS (Suite2p, CaImAn)[61-64] or custom-written toolbox[65], to approximate signals closer to the original action potentials for analysis. However, in practical imaging scenarios, the single-pulse responses of different neurons, and even different temporal segments within the same neuron, vary significantly[65]. Moreover, parameters like pulse width, often preset in deconvolution algorithms, are challenging to estimate accurately prior to analysis. This variability leads to substantial standard deviation fluctuations in the precision of traditional deconvolution algorithms, especially when processing data with low SNRs. We tested various deconvolution algorithms[26, 27, 66] and the deep learning algorithm Cascade using datasets provided by CRCNS[31, 32] and to restore calcium signals to electrophysiological signals (Extended Data Fig. 4H-J). Our findings demonstrate that the deep learning-based Cascade algorithm significantly outperforms traditional deconvolution methods when dealing with calcium imaging data of varying SNR levels. We have integrated Cascade's testing methodology into NeuroPixelAI, along with enhancements such as visualization features for neuronal signals during processing and a signal-correlation iterative sorting algorithm (Extended Data Fig. 4K). These enhancements allow for an intuitive observation of the temporal signals and overall correlation characteristics of neurons post-processing.

Instrument (Meso2P) Validation

For in-house bred *Camk2α-Cre::fl-GCaMP6s* mice, imaging comparisons were conducted on the same cortical plane (250-μm depth) without and with the parabolic lateral collection module. Images were acquired in high-speed (2,048 × 2,048 pixels/frame, 7.67 Hz) or high-resolution mode (4,096 × 4,096 pixels/frame, 3.83 Hz) using 130-mW excitation power. Fluorescence collection efficiency was assessed using the parabolic lateral collection module at 330-μm depth below the cortex with 90-mW imaging power. Twenty-eight-layer volumetric imaging was achieved using parabolic lateral collection, time-division multiplexing, and a piezo-driven z-scanning

module. This enabled simultaneous imaging at 2,048 × 2,048 resolution across a 6,000 × 6,000 × 500 μm³ volume (from 60 to 560 μm below the cortical surface) at 1.1 Hz with 250-mW excitation power. For 8-10-week-old *Camk2α-Cre::fl-GCaMP6s* mice, the red-shifted opsin pAAV-EF1α-DIO-C1V1(t/t)-Ts-mCherry-WPRE (Obio Technology) , a widely used optogenetic actuator, was injected before cranial window implantation. After a 4-week recovery period for viral expression, two-photon single-cell optogenetic stimulation imaging used 50-ms light pulses every 16 s (10 cycles; cycles 4-9 as control) at 220-μm depth (Nikon, 16×/0.8 NA objective). Imaging power: 20 mW; stimulation: 50 mW.

Photodamage Assessment

Uniform 28-layer imaging of 500-μm-thick cortex was performed for 1 h at 0 mW (control), 250 mW and 350 mW excitation (parabolic lateral collection module, 4× objective). After 20 h, mice were anesthetized with 1% pentobarbital sodium and transcardially perfused with heparinized saline (20 U/mL) followed by 4% paraformaldehyde (PFA). Brains were post-fixed overnight in 4% PFA at 4 °C, cryoprotected 20% and 30% sucrose/PBS, and sectioned at 25-μm thickness. Photodamage was assessed by immunofluorescence: sections blocked in 5% normal donkey serum/PBST (PBS + 0.3% Triton X-100) for 1 h, incubated with primary antibody (HSP70/HSP72 monoclonal antibody, Enzo Life Sciences #ADI-SPA-810, 1:400) overnight at 4 °C, and labeled with secondary antibody (Alexa Fluor® 594 AffiniPure® Donkey Anti-Mouse IgG (H+L), Jackson #715-585-151) for 2 h at room temperature, and sealed with Antifade Mounting Medium with DAPI (Beyotime Biotechnology #P0131) Six cortical regions per brain were imaged using a confocal microscope. Thermal injury areas were quantified using ImageJ software (v1.53t, NIH)[67] with the "Threshold" function (normalized to 0-mW control). Data analyzed by one-way ANOVA with Tukey's post hoc test (n = 3 mice per group).

Optical resolution calibration.

To calibrate the optical resolution of our multiphoton imaging system, we computed the 1/e intensity radius of the focal spot based on Gaussian functions fit to integral representation of the electric field near the focus of a diffraction-limited

888   focus[68]:

889   Lateral radius:

$$\omega_{xy} = \frac{0.31\lambda}{\sqrt{m}NA} \qquad (15)$$

891   Axial radius:

$$\omega_z = \frac{0.266\lambda}{n_i\sqrt{m}\sin^2(\alpha/2)} = \frac{0.532\lambda}{\sqrt{m}}\left[\frac{1}{n_i - \sqrt{n_i^2 - NA^2}}\right] \qquad (16)$$

893    $\alpha = \arcsin(NA/n)$ in radians. ω is the 1/e radius. For the $1/e^2$ radius multiply by

894    $\sqrt{2}$. For the FWHM multiply by $2\sqrt{\ln 2}$ = 1.665. m = order of excitation: m = 1 for

895   conventional excitation, m = 2 for 2 photon excitation, etc. Experimentally, we

896   imaged 0.1 μm fluorescent beads and derived the 1/e radius from measured FWHM

897   values.

Quantification and statistical analysis

Image Processing and Statistical Analysis

All image data were processed using NeuroPixelAI for motion correction,

denoising, cell segmentation, signal extraction, and spike inference, with additional

baseline correction and iterative sorting applied based on data characteristics. Outputs

include but are not limited to: enhanced images, neuron masks and properties, raw

pixel values, $\Delta F/F0$ signals, spike probabilities and inferred spikes. Statistical

analyses used MATLAB® R2023a (v9.14.0, Mathworks). Normally distributed

continuous data are presented as mean ± s.d. All statistical details (the size and type of

individual samples, n) are specified in figure legends.

Cascade Model Training

Given variable imaging rates across experiment, we trained multiple

spike-inference models using the "Global EXC" dataset[54] at diverse frame rates and

Gaussian smoothing kernel parameters. This dataset covers multiple indicators (e.g.,

GCaMP6f, GCaMP6s, OGB-1, GCaMP5k, Cal-520, R-CaMP1.07, and jRCaMP) and

brain regions (e.g. the visual cortex, somatosensory cortex, hippocampus, several

regions of the zebrafish forebrain, and the olfactory bulb), including some inhibitory

neurons. The training noise levels ranged from standard noise level v 2 to 9. The

smoothing parameter was empirically set t othe inverse of the frame rate multiplied by

1.5; for instance, at 7.67 Hz, the smoothing parameter was 195 ms, and at 30 Hz, it

was 50 ms. The causal ground truth smoothing kernel was set to 0 ms.

Cortex-wide neuronal dynamics during whisker stimulation

Experiment Paradigms

An air-puff nozzle was positioned in front of head-fixed

Camk2α-Cre::fl-GCaMP6s mice. After a 5-min baseline recording, mice received 10

cycles of air-puff stimulation (15-s duration each) with 45-s interstimulus intervals.

Air-puff delivery was synchronized with imaging using a programmable solenoid

valve controller. Neuronal calcium signals at 300-μm depth within a $6,000 \times 6,000$

μm^2 cortical area were continuously recorded with the parabolic lateral collection

module ($2,048 \times 2,048$ pixels, 7.67 Hz).

Brain Region Division and Regional Neuron Statistics

929 Brain regions were delineated and masks were created referencing the Allen

930 Brain Atlas[69]. Imaging areas were proportionally registered based on coronal and

931 horizontal planes, and all neurons were assigned to brain regions based on centroid

932 coordinates processed by NeuroPixelAI (Fig. 4C and Extended Data Fig. 5A).

Definition of Active (Coding) Cells

Active neurons were identified based on the inferred spikes output from

NeuroPixelAI. Neurons that exhibited an average action potential firing rate more

than twice the resting state firing rate of mice during ten stimulus trials were

considered "active". Under Gaussian distribution, non-coding neurons should have a

95% probability of firing at less than twice the resting rate in a single experiment (Fig.

4E and Extended Data Fig. 5C,D).

940
$$P\left(\mu - 1.96 \times \frac{\sigma}{\sqrt{n}} \leq \bar{x} \leq \mu + 1.96 \times \frac{\sigma}{\sqrt{n}}\right) = 0.95 \qquad (17)$$

Decoding Neural Population Activity Using SVM Decoders

To investigate the encoding responses of neural population activity in various

brain regions to air puff stimuli, we employed SVM decoders with a linear kernel for

population decoding[70]. For each mouse, we utilized the neural activity data collected

from 10 s before to 10 s after each trial, comparing it to resting state activity.

Decoders were trained on 80% of trials using a sliding 2-s time window and tested for

accuracy on the withheld 20% of trials. To avoid bias and overfitting, we repeated

decoding 512 times with random trial selections and averaged the measured values.

In tests examining the relationship between decoding accuracy and the number of

neurons, we randomly sampled increments of 200 (5%) neurons from a range of 200

to 4,000 neurons for decoding tests. The average decoding accuracy over the 15 s

duration of the air puff stimulus was recorded as the decoding accuracy for each

group of neurons. Each mouse underwent this random sampling 10 times to mitigate

the random errors caused by uneven distribution of neurons across brain regions.

Global Neural Peak Activity Conduction Bar Graph

In each trial, we observed specific directional inter-regional transmission of

neural group signals. To visualize fluctuations, we divided the AP = [-3 mm, 3 mm]

area into 16 bins and categorized the neurons within each. Since calcium signaling

serves as a low-pass, non-linear proxy for the sequence of cellular action potentials

(spikes), and the spike probabilities inferred by the cascade algorithm can be

considered a Gaussian kernel convolution of the spike sequence, generally featuring a

relatively higher temporal resolution, we averaged the spike probabilities S ($N_b \times N_t$,

where N_b is the number of regional bins and N_t is the number of time points) across

all neurons in each bin to obtain S_{bin}. This was then normalized to the peak values

within a 2 s window before and after each stimulation to generate the peak

transmission matrix T:

$$T = S_{bin}/S_{peak} \tag{18}$$

Subsequently, we generated a 3D visualization of matrix T, showing normalized

signal fluctuations across regions and the sequence and timing of peak arrivals in

response to a single stimulus (Fig. 4J and Extended Data Fig. 5H,I).

Dimensional Reduction and Visualization of Neural Representations Using

CEBRA

To investigate whether global network-wide reorganization of activity patterns

occurs during adaptation to repeated sensory stimuli, we employed the CEBRA[71]

algorithm for dimensional reduction and visualization of neural representations. For

our study, the CEBRA algorithm was applied to neuron signals that vary over time,

specifically in the CEBRA-Time mode. The input consisted of spike probabilities

from neurons across 10 trials. During the application of the CEBRA algorithm, we

employed a nonlinear convolutional neural network model, the "offset10-model," for

self-supervised training. The batch size for training data was set at 512, with a

learning rate of 0.0003 for the optimizer, and a maximum iteration count of 10,000.

The multi-dimensional neural signals were reduced to a three-dimensional space, with

both the training and testing phases accelerated by high-performance graphics

processing units. To enhance the visualization of the dimensional reduction results, we

employed pseudocolor coding that changes over time to represent each time point in

the reduced data, as shown in Fig. 4K.

Canonical Correlation Analysis and Principal Component Analysis of Pattern

Shifts and Neural Activity Weight Changes

To investigate the correlation between activity patterns across different cortical

areas and their gradual representational changes over successive experiments, we

employed Canonical Correlation Analysis (CCA) to analyze the shifts in shared

activity patterns within pairs of experiments and brain regions. For each experiment,

the spike probability was captured within a 1.5 s window before and after stimulus

presentation, segregated by brain region. For a given pair of trials within a specific

brain region, we represented the dynamical features of this area during the two trials

with matrices X and Y, sized $N_{t1} \times N_1$ and $N_{t2} \times N_2$, respectively. Here, N_{t1} and

N_{t2} denote the total number of time points in the group dynamic matrices for the two

trials, while N_1 and N_2 represent the number of neurons in each trial. Following the

standard methodology of CCA, we identified two sets of loading vectors, $\{w_i\}$ and

$\{v_i\}$. Each vector represents the weight distribution of each neural ensemble in a

shared activity pattern observed across the two trials. The index i, ranging from 1 to

the minimum of the dimensions of X and Y, denotes each mode. We determined these

modes such that the projections of the neural activity fluctuations, X and Y onto $\boldsymbol{w}_i$

and $\boldsymbol{v}_i$, achieved maximum correlation between the two ensembles.

The maximization of the correlation between these projections is expressed as:

$$\text{Maximize}_{\boldsymbol{w}_i, \boldsymbol{v}_i} (X\boldsymbol{w}_i)^{\mathrm{T}} (Y\boldsymbol{v}_i) \tag{19}$$

subject to the normalization constraint: $\boldsymbol{w}_i^{\mathrm{T}} X^{\mathrm{T}} X \boldsymbol{w}_i = \boldsymbol{v}_i^{\mathrm{T}} Y^{\mathrm{T}} Y \boldsymbol{v}_i = 1$. Given this

normalization condition, the quantity $(X\boldsymbol{w}_i)^{\mathrm{T}} \cdot (Y\boldsymbol{v}_i)$ equals the correlation

coefficient r_i of the activity modes, $X\boldsymbol{w}_i$ and $Y\boldsymbol{v}_i$, in the two different trials.

In order to explore the origins of gradual shifts in global activity patterns in

response to stimuli and the involvement of different brain regions in these changes,

we initially normalized the correlation coefficient r_1 for mode 1 extracted between

different trials for each brain region, and visualized this data using a correlation

heatmap (Fig. 4M). Subsequently, for each trial, we conducted cross-CCA between

brain regions. Specifically, within a given independent trial, we performed cross-CCA

on the dynamic matrices of the M, S, V, and RSP regions sequentially, obtaining

pairwise correlated modes w_{areaij} and v_{areaij}. These modes were then averaged

according to brain region:

$$w_{\text{area}i_{\text{global}}} = \frac{\sum_j w_{\text{area}ij_{\text{global}}}}{N_{\text{area}}} \tag{20}$$

this calculation yielded a single-trial local loading vector $w_{\text{area}i_{\text{global}}}$ for each brain

region, where N_{area} denotes the number of brain regions. Here, $w_{\text{area}i_{\text{global}}}$ reflects

the absolute contribution of all neurons in that region to the global activity pattern

during that trial. We posited that as trials progressed, the variation in these patterns

would manifest in slow changes in the weight distribution of neuronal responses

across different regions (Extended Data Fig. 5M). Accordingly, we mapped the

changes in neuronal contribution weights $w_{\text{weight}_{\text{change}}}$ across trials for different brain

regions using the weight change matrix:

$$w_{\text{weight}_{\text{change}}} = \left[w_{\text{area1}_{\text{global}}}, w_{\text{area2}_{\text{global}}}, \cdots w_{\text{area10}_{\text{global}}} \right] \tag{21}$$

these changes were visualized (Extended Data Fig. 4M) and further analyzed using

PCA to reduce the dimensions of the weight change matrix. The dominant latent

component, PC1 (Extended Data Fig. 5L), revealed gradual shifts in the contribution

weights of neurons across different brain regions.

Additionally, we focused on the collective contributions of each brain region to

the global activity pattern across ten trials. We projected the group activity of different

regions onto the global activity pattern and averaged it across neurons, i.e.,

$$1036 \qquad \text{Projection signal} = \text{mean} \left(X w_{\text{area}i_{\text{global}}} \cdot w_{\text{area}i_{\text{global}}} T \right) \qquad (22)$$

1037 It was observed that, despite all regions participating in the encoding response to

1038 air-puff stimuli and undergoing similar changes in neuronal contribution weights

1039 through successive adaptations, the activity patterns in regions S and V contributed

1040 more significantly to global encoding compared to regions M and RSP (Fig. 4L and

1041 Extended Data Fig. 5K).

Dissecting layer 2/3 and layer 5 activity simultaneously across cortex during anesthesia

Experiment Paradigm

A facemask was positioned in front of head-fixed *Camk2α-Cre::fl-GCaMP6s*

mice and adjusted to fit snugly (gas flow 1.0 L/min). After a 30-min adaptation,

10-min awake baseline imaging was performed. Isoflurane was then increased to 2%

for 20-min anesthesia induction, followed by a 10-min stabilization. Isoflurane was

reduced to 0% for the recovery imaging. Simultaneous dual-layer imaging at 300-μm

and 560-μm depth (6,000 × 6,000 μm^2 FOV) used the parabolic lateral collection

module at 2,048 × 2,048-pixels resolution (7.67 Hz) .

Analysis of Synchrony in Neuronal Population Activity

1053 To investigate changes in synchrony between deep and shallow neuronal activity

1054 during various stages of isoflurane anesthesia, we computed the Pearson correlation

1055 coefficient, R, for individual neurons against the remaining population activity within

1056 each window. This was performed using a 15 s time window with a 3 s step size:

$$1057 \qquad R = \frac{E[(x - \mu_x)(y - \mu_y)]}{\sigma_x \sigma_y} \qquad (23)$$

where x represents the activity of a single neuron within the window, y represents

the activity of the excluded neuronal population and μ_x, μ_y denote the means of x

and y, respectively, while σ_x and σ_y represent their standard deviations. E

signifies the arithmetic mean. The mean correlation across all neurons was then used

to quantify the synchrony of population activity within that window (Fig. 5F,G).

Subsequently, we defined the probability of anesthesia-related events based on

synchrony, using a 650 ms time window. Specifically, the peak synchrony during

wakefulness and the lowest synchrony during anesthesia were used as thresholds to

calculate the probability of population activity synchrony crossing these thresholds

across multiple experiments.

To compare the synchrony differences between deep and shallow layers under

different anesthetic states, we also analyzed the synchrony within the first 180 s

following the onset of synchronized activity, typically between 40 s and 220 s after

administration and cessation of isoflurane. Both during the induction and the

awakening phases, neurons in both layers exhibited synchronized activity, with deeper

layers showing higher synchrony than shallow layers, and induction phases showing

higher synchrony than anesthesia phases (Fig. 5H, $p < 0.001$; Tukey test; neurons =

$19,808 \sim 23,684$; n = 3 mice).

Granger Causality Analysis

Based on experimental observations, we hypothesized that multiple

"synchronized activities" during the induction of anesthesia actually represent specific

directional flows rather than complete synchrony. To infer causality in activity

between anterior and posterior (AP) cortical regions during individual synchronized

events (synchronization peaks), we uniformly divided the cortical populations within

an AP range of [-3 mm, 3 mm] into 16 segments. We then analyzed Granger causality

among these 16 groups of neurons using the MVGC Multivariate Granger Causality

Toolbox (https://github.com/lcbarnett/MVGC1)[72]. The input matrix included three

dimensions: AP bins, time points, and peaks number, recording activity of different

AP bins' neuronal populations within a 4 s interval (2 s before to 2 s after each peak).

This multivariate Granger causality toolkit, based on a vector autoregressive model,

calculates Granger causality where the model order (time lags) was estimated by

Akaike Information Criteria with the maximum set to 20. Pairwise conditional

causality was then calculated, where causality includes directional information: the

direction "from" an X module "to" a Y module indicates that activity in X preceded

that in Y. Granger causality values that reached $p < 0.05$ (Granger F test) after

applying the false discovery rate for multiple-comparison correction (Fig. 5I).

Analysis of Global Neural Activity Transmission During Synchronous Peaks

During periods of anesthesia, multiple "synchronous" peaks are not entirely

synchronized and exhibit various directions of transmission. It is only during deep

anesthesia that these "synchronous" peaks approach complete synchronization. To

verify this, similar to the repetitive air-puff whisker stimulation experiment, we

initially utilized the spike probability outputs from NeuroPixelAI to perform regional

peak detection, resulting in the generation of a spike transmission matrix T_{peaks}. Here,

peak detection was conducted using a MATLAB-based local thresholding algorithm,

with parameters set at a threshold of 0.001, minimum peak height of 0.04, and

minimum peak distance of 20. The structure of T_{peaks} is represented as $N_b \cdot N_t \cdot$

$peaks_{num}$, where N_b denotes the number of regional bins, N_t indicates the number

of time points (spanning [-2 s, 2 s] across 31 frames at a frame rate of 7.67), and

$peaks_{num}$ represents the detected number of peaks. A three-dimensional visualization

of the signals within T_{peaks} facilitates the generation of a global map of neural peak

activity transmission (Fig. 5J and Extended Data Fig. 6C).

To further visualize the dynamic transmission process of "synchronous" signals

throughout the anesthesia induction, we rendered a global dynamic transmission map

of peak activities at the resolution of individual neurons (Fig. 5J, Extended Data Fig.

6A,C, and Supplementary Video). In this representation, yellow signifies the origin

regions of the signals, while purple denotes the receiving regions. Specifically, for

each "synchronous" event, we normalized the spike probability for each neuron

around the event ([-2 s, 2 s]), based on peak values. We then calculated the response

delay, which is the difference between the peak positions of individual neurons and the peak position of the group activity. If the response delay is less than zero, the neuron is identified as the source of the event; if greater than zero, the neuron is considered a receiver. Based on the response delays of various neurons, we were able to ascertain the global dynamics of peak transmission within the imaging field area $(6{,}000 \times 6{,}000 \ \mu m^2)$:for each layer in T_{peaks}, or each peak, we determined the times t_{AP} at which different AP bins reached their peak and calculated the standard deviation across these bins:

$$\text{Global response delay} = \sqrt{\frac{\sum_i (t_{APi} - \bar{t}_{APi})^2}{N_b - 1}} \tag{24}$$

the global response delay reflects the fluidity of each spike's global response, illustrating the progression from incomplete to almost complete synchronization of multiple "synchronous" peaks during the entire period of anesthesia induction (Extended Data Fig. 6B). To mitigate the impact of spurious spikes detected during peak detection, we applied a sliding window averaging and normalization over 20 spikes.

Through the assessment of global response delays, we hypothesize that during the process of anesthesia induction, there exists a gradual change in neural group activities and a redistribution of neuronal activity weights, associated with an increase in 'synchronized' peak events. To investigate this, we categorized the detected peaks into deciles, representing 10% of the peaks per group, and initially computed the normalized CCA mode1 correlation matrices between these groups (Fig. 5L). Subsequently, we demonstrated the matrices of changes in neuronal weight across these ten groups of peaks (Extended Data Fig. 6D). The correlation coefficient matrices were derived through pairwise CCA of each group of peaks. The CCA of the neuronal weight change matrices encompassed two elements: the neural dynamics matrix for the period [-2s, 2s] around a specific group of peaks, and the average neural dynamics matrix for all other peaks except for the group in question. Following each CCA, the weight matrices w_i were normalized (ensuring the sum of weights for all neurons equaled one) and subjected to iterative correlation ranking.

In addition, we analyzed the proportions of forward versus backward transmission peaks (Fig. 5L; $p < 0.05$) and the transmission from deep to shallow layers versus from shallow to deep layers during the entire anesthesia induction process (Extended Data Fig. 6F; $p < 0.01$). For the anterior-posterior peak transmission, we calculated the difference in response delays between the first 50% of bins and the last 50% of each event to determine whether the anterior or posterior regions reached peak values first. For the deep-to-shallow layer transmission, we calculated the differences in response delays across 16 bins corresponding to different depths(Extended Data Fig. 6E). A negative response delay indicated that, for a given event, the signal was transmitted from deeper to more shallow layers; conversely, a positive response delay suggested transmission from shallow to deeper layers. These findings align with previous studies[73-76], indicating that during the anesthesia induction phase, global signal propagation predominantly occurs from posterior to anterior regions, with notable instances of reverse transmission, primarily occurring from deeper to more shallow layers.

Dynamic monitoring of MNP distribution and vascular occlusion in the living brain

Experiment Paradigm

Prior to imaging, male *C57BL/J* mice (Charles River Laboratories #219) under isoflurane anesthetized received 200-µL retro-orbital injections of 5% FITC-Dextran or RITC-Dextran. Twenty-eight-layer structural vascular imaging at 450-µm depth cover $6,000 \times 6,000 \times 450$ µm³ (4× objective) or $1600 \times 1600 \times 450$ µm³ volumes (16× objective), using the parabolic lateral module at $2,048 \times 2,048$ pixels resolution (7.67 Hz/layer). For real-time MNPs tracking, dye-injected mice received 100-µL tail vein injections of 1 mg/mL 5-µm polystyrene (PS) microplastics (MPs, BEISILE #7-1-0500) or 200-nm PS nanoplastics (NPs, BEISILE #7-3-0020) during single-layer imaging. Repeat 28-layer structural imaging was performed at 12-h and 24-h post-injection.

Blood Flow Velocity Detection

Blood flow velocity was measured using an established method[77]. Fluorescent dye injection rendered red blood cells (RBCs) as dark particles. RBC displacement was tracked by 1k Hz high-speed line-scanning within ROIs. Velocity was calculated from pixel-wise displacement over 10-s line-scanning recordings. Given MNP occlusion-prone capillaries, we analyzed single-RBC-perfused vessels. Measurements occurred at: 1-h post-anesthesia (avoiding stress hyperemia), and 12/24-h post-injection. Blood flow was not assessed during tail vein injections to eliminate artifacts.

Blood-Brain Barrier Integrity Assay

Male *C57BL/J* mice were divided into: sham-operated, 5-day post-surgery, and 14-day post-surgery groups. Each received 125-μL tail vein injections of 2% Evans Blue (Beyotime Biotechnology #ST3273) 1.5 h before brain collection. Mice were anesthetized with 1% pentobarbital sodium (i.p.), then transcardially perfused with heparinized saline (20 U/mL) and 4% PFA. Evans Blue extravasation was visually assessed in whole brains to qualitatively evaluate BBB integrity.

References

1.   Li, K., Nakajima, M., Ibañez-Tallon, I. & Heintz, N. A Cortical Circuit for Sexually Dimorphic Oxytocin-Dependent Anxiety Behaviors. *Cell* **167**, 60-72.e11 (2016).

2.   Buzsáki, G. & Draguhn, A. Neuronal Oscillations in Cortical Networks. *Science* **304**, 1926-1929 (2004).

3.   Vyas, S., Golub, M.D., Sussillo, D. & Shenoy, K.V. Computation Through Neural Population Dynamics. *Annual Review of Neuroscience* **43**, 249-275 (2020).

4.   Kim, T.H. & Schnitzer, M.J. Fluorescence imaging of large-scale neural ensemble dynamics. *Cell* **185**, 9-41 (2022).

5.   Denk, W., Strickler, J.H. & Webb, W.W. Two-Photon Laser Scanning Fluorescence Microscopy. *Science* **248**, 73-76 (1990).

6.   Helmchen, F. & Denk, W. Deep tissue two-photon microscopy. *Nature Methods* **2**, 932-940 (2005).

7.   Cahalan, M.D., Parker, I., Wei, S.H. & Miller, M.J. Two-photon tissue imaging: seeing the immune system in a fresh light. *Nature Reviews Immunology* **2**, 872-880 (2002).

8.   Wu, J., Ji, N. & Tsia, K.K. Speed scaling in multiphoton fluorescence microscopy. *Nature Photonics* **15**, 800-812 (2021).

9.   Hampson, K.M. et al. Adaptive optics for high-resolution imaging. *Nature Reviews Methods Primers* **1**, 68 (2021).

10.   Pittet, Mikael J. & Weissleder, R. Intravital Imaging. *Cell* **147**, 983-991 (2011).

11. Ji, N., Magee, J.C. & Betzig, E. High-speed, low-photodamage nonlinear imaging using
passive pulse splitters. *Nature Methods* **5**, 197-202 (2008).

12. Zipfel, W.R., Williams, R.M. & Webb, W.W. Nonlinear magic: multiphoton microscopy in the
biosciences. *Nature Biotechnology* **21**, 1369-1377 (2003).

13. Stirman, J.N., Smith, I.T., Kudenov, M.W. & Smith, S.L. Wide field-of-view, multi-region,
two-photon imaging of neuronal activity in the mammalian brain. *Nature Biotechnology* **34**,
857-862 (2016).

14. Tsai, P.S. et al. Ultra-large field-of-view two-photon microscopy. *Opt. Express* **23**,
13833-13847 (2015).

15. Ota, K. et al. Fast, cell-resolution, contiguous-wide two-photon imaging to reveal functional
network architectures across multi-modal cortical areas. *Neuron* **109**, 1810-1824.e1819
(2021).

16. Clough, M. et al. Flexible simultaneous mesoscale two-photon imaging of neural activity at
high speeds. *Nature Communications* **12**, 6638 (2021).

17. Zhang, Y. et al. Long-term mesoscale imaging of 3D intercellular dynamics across a
mammalian organ. *Cell* **187**, 6104-6122.e6125 (2024).

18. Mengke, Y. et al. Ultra-wide-field, deep, adaptive two-photon microscopy. *bioRxiv -*
Neuroscience (2025).

19. Yu, C.-H. et al. The Cousa objective: a long-working distance air objective for multiphoton
imaging in vivo. *Nature Methods* **21**, 132-141 (2024).

20. Yu, C.-H., Stirman, J.N., Yu, Y., Hira, R. & Smith, S.L. Diesel2p mesoscope with dual
independent scan engines for flexible capture of dynamics in distributed neural circuitry.
Nature Communications **12**, 6639 (2021).

21. Mok, A.T. et al. A large field-of-view, single-cell-resolution two- and three-photon microscope
for deep and wide imaging. *eLight* **4**, 20 (2024).

22. Rumyantsev, O.I. et al. Fundamental bounds on the fidelity of sensory cortical coding. *Nature*
580, 100-105 (2020).

23. Sofroniew, N.J., Flickinger, D., King, J. & Svoboda, K. A large field of view two-photon
mesoscope with subcellular resolution for in vivo imaging. *eLife* **5**, e14472 (2016).

24. Lu, R. et al. Rapid mesoscale volumetric imaging of neural activity with synaptic resolution.
Nature Methods **17**, 291-294 (2020).

25. Demas, J. et al. High-speed, cortex-wide volumetric recording of neuroactivity at cellular
resolution using light beads microscopy. *Nature Methods* **18**, 1103-1111 (2021).

26. Pachitariu, M. et al. Suite2p: beyond 10,000 neurons with standard two-photon microscopy.
bioRxiv, 061507 (2017).

27. Giovannucci, A. et al. CaImAn an open source tool for scalable calcium imaging data analysis.
eLife **8**, e38173 (2019).

28. Lu, J. et al. MIN1PIPE: A Miniscope 1-Photon-Based Calcium Imaging Signal Extraction
Pipeline. *Cell Reports* **23**, 3673-3684 (2018).

29. Li, X. et al. Real-time denoising enables high-sensitivity fluorescence time-lapse imaging
beyond the shot-noise limit. *Nature Biotechnology* **41**, 282-292 (2023).

30. Li, X. et al. Spatial redundancy transformer for self-supervised fluorescence image denoising.
Nature Computational Science **3**, 1067-1080 (2023).

31. Chen, T.-W. et al. Ultrasensitive fluorescent proteins for imaging neuronal activity. *Nature* **499**,
295-300 (2013).

32. Akerboom, J. et al. Optimization of a GCaMP Calcium Indicator for Neural Activity Imaging.
The Journal of Neuroscience **32**, 13819 (2012).

33. Faisal, A.A., Selen, L.P.J. & Wolpert, D.M. Noise in the nervous system. *Nature Reviews*
Neuroscience **9**, 292-303 (2008).

34. Lütcke, H., Margolis, D.J. & Helmchen, F. Steady or changing? Long-term monitoring of
neuronal population activity. *Trends in Neurosciences* **36**, 375-384 (2013).

35. Stein, R.B., Gossen, E.R. & Jones, K.E. Neuronal variability: noise or part of the signal?
Nature Reviews Neuroscience **6**, 389-397 (2005).

36. Zohary, E., Shadlen, M.N. & Newsome, W.T. Correlated neuronal discharge rate and its
implications for psychophysical performance. *Nature* **370**, 140-143 (1994).

37. Driscoll, L.N., Pettit, N.L., Minderer, M., Chettih, S.N. & Harvey, C.D. Dynamic
Reorganization of Neuronal Activity Patterns in Parietal Cortex. *Cell* **170**, 986-999.e916
(2017).

38. Montijn, Jorrit S., Meijer, Guido T., Lansink, Carien S. & Pennartz, Cyriel M.A.
Population-Level Neural Codes Are Robust to Single-Neuron Variability from a
Multidimensional Coding Perspective. *Cell Reports* **16**, 2486-2498 (2016).

39. Ebrahimi, S. et al. Emergent reliability in sensory cortical coding and inter-area
communication. *Nature* **605**, 713-721 (2022).

40. Dong, Y., Li, J., Zhou, M., Du, Y. & Liu, D. Cortical regulation of two-stage rapid eye
movement sleep. *Nature Neuroscience* **25**, 1675-1682 (2022).

41. Brown, R., Lau, H. & LeDoux, J.E. Understanding the Higher-Order Approach to
Consciousness. *Trends in Cognitive Sciences* **23**, 754-768 (2019).

42. Franco, S.J. et al. Fate-Restricted Neural Progenitors in the Mammalian Cerebral Cortex.
Science **337**, 746-749 (2012).

43. Mashour, G.A., Roelfsema, P., Changeux, J.-P. & Dehaene, S. Conscious Processing and the
Global Neuronal Workspace Hypothesis. *Neuron* **105**, 776-798 (2020).

44. Muindi, F. et al. Electrical stimulation of the parabrachial nucleus induces reanimation from
isoflurane general anesthesia. *Behavioural Brain Research* **306**, 20-25 (2016).

45. Bharioke, A. et al. General anesthesia globally synchronizes activity selectively in layer 5
cortical pyramidal neurons. *Neuron* **110**, 2024-2040.e2010 (2022).

46. Steriade, M., Nuñez, A. & Amzica, F. A novel slow (<1 Hz) oscillation of neocortical neurons
in vivo: Depolarizing and hyperpolarizing components. *The Journal of Neuroscience* **13**,
3252-3265 (1993).

47. Pal, D. et al. Differential Role of Prefrontal and Parietal Cortices in Controlling Level of
Consciousness. *Current Biology* **28**, 2145-2152.e2145 (2018).

48. Gao, S., Proekt, A., Renier, N., Calderon, D.P. & Pfaff, D.W. Activating an anterior nucleus
gigantocellularis subpopulation triggers emergence from pharmacologically-induced coma in
rodents. *Nature Communications* **10**, 2897 (2019).

49. Nihart, A.J. et al. Bioaccumulation of microplastics in decedent human brains. *Nature*
Medicine **31**, 1114-1119 (2025).

50. Thompson, R.C. et al. Twenty years of microplastic pollution research—what have we learned?
Science **386**, eadl2746 (2024).

51. Song, A., Gauthier, J.L., Pillow, J.W., Tank, D.W. & Charles, A.S. Neural anatomy and optical
microscopy (NAOMi) simulation for evaluating calcium imaging methods. *Journal of*
Neuroscience Methods **358**, 109173 (2021).

52. Meiniel, W., Olivo-Marin, J.C. & Angelini, E.D. Denoising of Microscopy Images: A Review
of the State-of-the-Art, and a New Sparsity-Based Method. *IEEE Transactions on Image*
Processing **27**, 3842-3856 (2018).

53. Li, X. et al. Reinforcing neuron extraction and spike inference in calcium imaging using deep
self-supervised denoising. *Nature Methods* **18**, 1395-1400 (2021).

54. Rupprecht, P. et al. A database and deep learning toolbox for noise-optimized, generalized
spike inference from calcium imaging. *Nature Neuroscience* **24**, 1324-1337 (2021).

55. Pachitariu, M. & Stringer, C. Cellpose 2.0: how to train your own model. *Nature Methods* **19**,
1634-1641 (2022).

56. Pnevmatikakis, Eftychios A. et al. Simultaneous Denoising, Deconvolution, and Demixing of
Calcium Imaging Data. *Neuron* **89**, 285-299 (2016).

57. Bao, Y., Soltanian-Zadeh, S., Farsiu, S. & Gong, Y. Segmentation of neurons from
fluorescence calcium recordings beyond real time. *Nature Machine Intelligence* **3**, 590-600
(2021).

58. Rose, T., Goltstein, P.M., Portugues, R. & Griesbeck, O. Putting a finishing touch on GECIs.
Frontiers in Molecular Neuroscience **Volume 7 - 2014** (2014).

59. Wei, Z. et al. A comparison of neuronal population dynamics measured with calcium imaging
and electrophysiology. *PLOS Computational Biology* **16**, e1008198 (2020).

60. Farhan, A. & Alex, C.K. Interpreting *in vivo* calcium signals from neuronal cell bodies,
axons, and dendrites: a review. *Neurophotonics* **7**, 011402 (2019).

61. Yeh, P.-T. et al. Discrete photoentrainment of mammalian central clock is regulated by
bi-stable dynamic network in the suprachiasmatic nucleus. *Nature Communications* **16**, 3331
(2025).

62. Guyoton, M. et al. Cortical circuits for cross-modal generalization. *Nature Communications*
16, 4230 (2025).

63. Shao, L.-X. et al. Psilocybin's lasting action requires pyramidal cell types and 5-HT2A
receptors. *Nature* **642**, 411-420 (2025).

64. Russell, L.E. et al. All-optical interrogation of neural circuits in behaving mice. *Nature*
Protocols **17**, 1579-1620 (2022).

65. Koukouli, F. et al. Nicotine reverses hypofrontality in animal models of addiction and
schizophrenia. *Nature Medicine* **23**, 347-354 (2017).

66. Deneux, T. et al. Accurate spike estimation from noisy calcium signals for ultrafast
three-dimensional imaging of large neuronal populations in vivo. *Nature Communications* **7**,
12190 (2016).

67. Schneider, C.A., Rasband, W.S. & Eliceiri, K.W. NIH Image to ImageJ: 25 years of image
analysis. *Nature Methods* **9**, 671-675 (2012).

68. Wolf, E. Electromagnetic Diffraction in Optical Systems. I. An Integral Representation of the
Image Field. *Proceedings of the Royal Society of London. Series A, Mathematical and*
Physical Sciences **253**, 349-357 (1959).

69. Lein, E.S. et al. Genome-wide atlas of gene expression in the adult mouse brain. *Nature* **445**,
168-176 (2007).

70. Bellafard, A., Namvar, G., Kao, J.C., Vaziri, A. & Golshani, P. Volatile working memory representations crystallize with practice. *Nature* **629**, 1109-1117 (2024).

71. Schneider, S., Lee, J.H. & Mathis, M.W. Learnable latent embeddings for joint behavioural and neural analysis. *Nature* **617**, 360-368 (2023).

72. Barnett, L. & Seth, A.K. The MVGC multivariate Granger causality toolbox: A new approach to Granger-causal inference. *Journal of Neuroscience Methods* **223**, 50-68 (2014).

73. Pullon, R.M., Yan, L., Sleigh, J.W. & Warnaby, C.E. Granger Causality of the Electroencephalogram Reveals Abrupt Global Loss of Cortical Information Flow during Propofol-induced Loss of Responsiveness. *Anesthesiology* **133** (2020).

74. Huo, J. et al. Orexin signalling in the nucleus accumbens promotes arousal from isoflurane anaesthesia and restores communication between the nucleus accumbens and frontal cortex. *British Journal of Anaesthesia* (2025).

75. Yin, M. et al. Synchronicity of pyramidal neurones in the neocortex dominates isoflurane-induced burst suppression in mice. *British Journal of Anaesthesia* **134**, 1122-1133 (2025).

76. Song, X.-J. & Hu, J.-J. Neurobiological basis of emergence from anesthesia. *Trends in Neurosciences* **47**, 355-366 (2024).

77. Kim, T.N. et al. Line-Scanning Particle Image Velocimetry: An Optical Approach for Quantifying a Wide Range of Blood Flow Speeds in Live Animals. *PLOS ONE* **7**, e38590 (2012).

Resource availability

Lead contact

Further information and requests for resources and reagents should be directed to and will be fulfilled by the lead contact, Bo Li (liboresearch@gmail.com).

Materials availability

This study did not generate new unique reagents.

Data and code availability

The codes generated during this study with example data are available at GitHub with the following link: https://github.com/Yourswear/Meso2P. NeuroPixelAI are available with the following link: https://github.com/Yourswear/NeuroPixelAI. The published article includes all data generated during this study and all the raw data is also available from the corresponding author upon request. Any additional information required to reanalyze the data reported in this paper is available from the lead contact upon request.

Acknowledgements

This work was supported by the Shanghai Municipal Education Commission (2024RGZNB03), Fudan University (FudanX24AI048, yg2023-04), the National Natural Science Foundation of China (T2222006), and the Science and Technology Commission of Shanghai Municipality (20JC1419500).

Author Contributions

Conceptualization: B.L.; methodology: B.L., J.H.H., S.P.L., C.Y.L., M.Z.; biological experiments: Y.F.Z., J.H.H, X.Y.G.; Zemax optical simulation: J.H.H., F.X.; investigation: J.H.H., Y.F.Z.; visualization: J.H.H., Y.F.Z., M.Z.; funding acquisition: B.L., Y.M., L.C.; project administration: B.L.; writing: B.L., J.H.H, Y.F.Z. Y.M., L.C., J.C.W.

Declaration of interests

The authors declare no competing interests.

Extended data

Extended Data Fig. 1 Modular design schematic and performance evaluation of

Meso2P imaging, related to Figures 1 and 2 and Methods.

 (A) Comparative analysis of Meso2P with other two-photon mesoscopes. Left, Meso2P
 overcomes traditional trade-offs between NA and FOV. Right, Meso2P demonstrates superior
 fluorescence collection efficiency and a smaller objective diameter compared to other mesoscopes.
 The NA and FOV were all derived from explicit or displayed parameters in the cited reference
 paper. **(B)** SolidWorks-rendered optical path geometry (excitation in blue, stimulation in orange).
 (C) Field-dependent aberration quantified by point spread function (PSF) profiles across the entire

FOV. **(D)** Optical simulation performance of Meso2P transverse scanning module. (i) Zemax optical path diagram. (ii) Field-dependent RMS wavefront aberration (0°: FOV center; 13°: 3 mm off-axis). (iii) Wavelength-dependent RMS wavefront aberration (blue: center; green: 1.5 mm off-axis; red: 3 mm off-axis). (iv) Spot diagram at focal plane demonstrating simulated 6.13 mm FOV. (v) Field curvature and distortion across entire FOV. **(E)** FOV calibration using fluorescent-etched scale bars. **(F)** Fundamental imaging characteristics of Meso2P with 4× objective. Left, example images of 0.1 μm fluorescent beads. Right, imaging demonstration with zoomed-in views of two regions (white boxes). **(G)** Same as (F), but for 16× objective. **(H)** Parabolic collection module: 3D model and imaging performance. Top, SolidWorks-rendered model (inset: custom 4× objective in dashed box). Bottom, Imaging comparison: objective-only vs. parabolic module collection (fluorescent-etched scale bar sample) with pixel intensity distribution. **(I)** Top, Optical principle of parabolic collection module. LPF, low-pass filter. BPF, band-pass filter. Bottom, field uniformity analysis (± 3 mm FOV): PSF peak intensities and collection efficiencies, with comparisons of PSF at the center and 3 mm edge.

 Extended Data Fig. 2 Schematic implementation of diverse and flexible

 volumetric imaging using Meso2P, related to Figure 2 and Methods.

1413

(A) 3D SolidWorks schematic of the dual-extension module for volumetric imaging. **(B)**
Dual-extension module for volumetric imaging. BS: 50/50 beam splitter; PBS: polarizing beam
splitter; HWP: rotatable half-wave plate. All lenses are mounted on programmable stages for focal
adjustment. **(C-E)** Representative images of brain sections showing immunolabeling for thermal
damage marker (anti-HSP70/72, red) and DNA stain (DAPI, blue) after exposure to the laser
power listed below. (C) control, no laser exposure; (D) 250 mw; (E) 350 mw. Scale bars: 500 μm.
(F) Intensity of immunolabeling corresponding to imaging area as a fraction compared to mean of
control samples. (n = 3 mice per group, 6 cortical regions per brain). n.s, no significance; ***$p <$
0.001; Tukey test. **(G)** Data processing pipeline for large-volume 3D imaging on the basis of
NeuroPixelAI. Key steps: Four-layer image crosstalk elimination, neuron information extraction
via NeuroPixelAI, mask generation and volumetric reconstruction in Imaris.

 Extended Data Fig. 3 Experimental validation of the Meso2P extension module,

 related to Figure 2

1427

(A) Paraboloid collection module enhances neural imaging versus conventional objective. (i)
Image quality comparison with a zoom-in region. (ii) Comparison of image quality and signal
traces in single neurons (white arrows: example neurons; black/red traces,
raw/NeuroPixelAI-enhanced signals). (iii) Segmentation improvement of neurons in (i). (iv) SNR
enhancement of temporal signals from three marked neurons in (ii). **(B)** Example 3D segmentation
masks across 28 imaging planes (layers 21-28; 430-560 μm below the surface). Right, example
image at deepest layer (28) with zoomed-in neuronal regions (yellow boxes) and corresponding
temporal signals. **(C)** 3D rendering of extracted neurons across nine representative brain regions
(28 layers total). Neurons are colored by depth. **(D)** The representative holographic pattern
presented by the galvo/SLM path.

Extended Data Fig. 4 Algorithm benchmarking and implementation schematics

for NeuroPixelAI, related to Figure 3 and Methods.

1440

(A) Schematic workflow of the standard NeuroPixelAI processing pipeline. Key steps include
motion registration, denoising, segmentation, signal extraction, baseline correction, spike
inference, spike discretization, and iterative sorting. Batch processing of selected steps is enabled
using predefined processing masks. **(B)** Performance comparison of denoising algorithms on noisy
data (0 dB SNR). Bottom panels show zoomed-in views of example neurons. **(C)** Denoising
algorithm performance across varying imaging frame rates (1–30 Hz). SRDTrans was trained on
datasets at each specific frame rate; DeepCAD-RT utilized its published large pretrained model.
(D) Comparison of segmentation algorithm performance. Boxed regions highlight areas exhibiting
significant differences. **(E)** Bar plot showing F1 scores of segmentation algorithms. Data derived
from two-photon and three-photon imaging at depths of 100–1000 μm (n = 5 mice). **(F)**
Comparison of processing speed versus F1 score for segmentation algorithms. **(G)** Schematic and
efficacy of baseline correction using a cubic spline interpolation algorithm. Correction (right)
reduces erroneous spike inference caused by baseline drift. **(H)** Performance comparison of spike
inference algorithms. Predicted spikes (black) versus ground truth (yellow). **(I)** Correlation

between inferred and ground-truth action potentials for spike inference algorithms across standardized noise levels. Datasets: Simultaneous electrophysiology and calcium imaging recordings from single neurons (GENIE datasets; n = 8 neurons). **(J)** Box plot statistics of inference correlation coefficients for all algorithms at a standardized noise level of 3.3. **(K)** Workflow and output of the label-free iterative sorting algorithm. Gray dashed lines indicate clear sorting results.

Extended Data Fig. 5 Large-scale population neural response dynamics during

repetitive whisker stimulation, related to Figure 4 and Methods.

1463

(A) Decoding performance as a function of neuronal population size. Left, performance versus
absolute number of neurons. Right, performance versus the percentage of extracted neurons. Data
are shown for three mice. The dashed line indicates the mean proportion of active neurons across
mice, mean ± SD. (B) Comparisons of decoding accuracy using active neurons versus all neurons
(n = 3 mice), 5 trials per mouse. (C) Calcium signal heatmaps across ten trials for three mice.
Gray dashed lines demarcate the 4 s peri-stimulus period. Pink solid lines separate signals from

the first and last trials. The heatmaps reveal concurrent changes in response intensity and neuronal

loading weights across trials. **(D)** Cortical distribution of segmented stimulus-responsive neurons.

Colors denote corresponding brain regions. **(E)** Normalized neuronal spike probabilities over time

and position for trials 1 and 10. Bar plots show the normalized spike probability at each time point

and anterior-posterior bin for the first (Trial 1) and last (Trial 10) trials. Small neuronal

populations in the posterior cortical areas of all three mice exhibited a significant temporal shift in

peak activity in Trial 10 compared to Trial 1. **(F)** Left, trial-by-trial changes of peak spiking

probability across 10 trials in three mice. Right, changes in calcium event rate (mean ± SD) within

a 2-s window surrounding stimulus onset for individual recorded brain regions. **(G)** Average

neural activity Z-scores across 10 trials for each brain region. Stimulus onset is indicated by the

red vertical line. **(H)** Bar plots of normalized spike probability across 10 trials, averaged across

three mice. **(I)** Trial-dependent dynamics of regional brain activity. Individual panels display

neural activity Z-scores per brain region in single trials. Stimulus onset is marked by red vertical

lines. **(J)** Schematic of structural reorganization in neural networks. Illustrates trial-dependent

changes in single-neuron loading weights despite identical sensory inputs. **(K)** The projected

activity of each area in the CCA dimensions identified across 10 trials. Population activity in S1

and V1 cortices exhibits dominant contributions to the stimulus-evoked cofluctuation mode, with

trial-dependent modulation in response magnitude and phase shifts. **(L)** PC1 derived from PCA of

neuronal loading weight matrices. **(M)** Heatmap of neuronal loading weights across trials. Depicts

trial-dependent reorganization of loading weights within the network. **(K-M)** Representative data

from one mouse.

 Extended Data Fig. 6 Schematic representation and assessment of intra- and

 interlaminar signaling dynamics during anesthesia induction using Meso2P

 microscopy, related to Figure 5 and Methods.

1494

(A) Data processing pipeline for anesthetized imaging experiments. Raw image data were

processed by NeuroPixelAI to extract single-neuron spike probabilities. Global spike probability

traces underwent peak detection to identify multiple "synchronous" event timings. For each

synchronized event, fluctuation delays were quantified at both individual anterior-posterior (AP)

bins and single-neuron levels, followed by spatiotemporal visualization. **(B)** Normalized global

response delays per mouse. Standard deviation (SD) of peak times in spike probabilities across 16

anterior-posterior (AP) bins was computed during anesthesia progression. This reveals a gradual

transition from spatially propagating delays (indicating asynchronous activity) to uniform
synchrony across cortical regions within individual synchronous events. **(C)** Schematics of
synchronized signal propagation direction in two additional mice. **(D)** Heatmap of neuronal CCA
loading weights across synchronous events in an example mouse. Synchronous events were
divided into 10 consecutive time bins. For each bin, cross-bin CCA identified stable activity
modes during anesthesia induction. Late-phase bins exhibited significant reorganization of loading
weights, concomitant with gradual uniform synchronization of cortical activity. **(E)** Heatmap of
synchronized signal propagation direction between deep and shallow cortical layers during
anesthesia induction. This indicates the predominant direction of signal flow during each
synchronous event within individual AP bins. Red, Deep-to-shallow propagation; Blue,
Shallow-to-deep propagation. **(F)** Boxplot of signal propagation direction proportions between
deep and shallow cortical layers during anesthesia induction (n = 3 mice), **$p < 0.01$, Tukey test.

Extended Data Fig. 7 Supplementary data for in vivo MNP imaging.

1515

(A) Blood-brain barrier (BBB) integrity assay to determine optimal MNP imaging window; compromised BBB (red arrows). (B) Comparison of blood flow velocity before and 12–24 h after MP injection: vessels (green), MPs (yellow), unoccluded vessels (a, b), occluded vessel (c), adjacent vessel (d). (C) Cumulative distribution of MPs across vascular layers 12 h post-injection: vessels (green), MPs (red), co-localization (yellow). MPs occluding vessels < 30 μm (red circles), > 30 μm (blue circles), extravascular MPs (white dashed circles), unresolved MPs in

vessels < 30 μm (pink arrows) and > 30 μm (blue arrows). **(D)** Structural morphology of NPs in

blood vessels: vessels (red), NPs (green), co-localization (yellow). **(E)** Time-dependent

accumulation of NP-laden cells in the uppermost cortical layer following NP injection: vessels

(red). **(F)** NP aggregation and clearance dynamics in blood vessels: NPs (green), aggregation foci

(yellow arrows). **(G)** Comparison of blood flow velocity in vessels treated with NPs. **(H)**

Schematic illustration of differential brain distribution patterns between 5-μm MPs and 200-nm

NPs. *Scale bars*: 500 μm (C, E); 100 μm (D, F). ***See also*** **Supplementary Videos S5–9.**

 Supplementary Tables

Platform	FOV×NA	FOV×NA/D	Frame resolution (pixels)	Pixel rate (Mpixel/s)	Volume Imaging	Two-photon optical stimulation
Meso2P	31.32	0.85	4096×4096	64.26	YES	YES
ULTRA[18]	32.00	0.38	4400×4800	17.53	NO	NO
Light beads[25]	19.44	0.22	1080×1200	2.85	YES	NO
2p-RAM[23]	23.76	0.26	3667×4154	10.66	NO	NO
Ultra-large FOV 2p[14]	22.40	0.62	2000×500	0.32	NO	NO
DEEP scope[21]	6.05	0.09	1024×1024	6.04	YES	NO
FASHIO-2p[15]	7.20	0.09	2048×2048	31.46	NO	NO
Quadroscope[16]	1.17	0.03	1126×600	20.27	NO	NO
Diesel2p[20]	8.10	0.09	2048×4096	32.21	NO	NO
Trepan 2p[13]	5.27	0.10	2048×2048	0.34	NO	NO

1530 **Supplementary Table 1: Comparative benchmarking of Meso2P against existing**

1531 **large-FOV two-photon platforms for single plane imaging.**

Platform	Volume size (mm)	Volume resolution (voxels)	Volume rate (Hz)	Simultaneous imaging plane	Voxel rate (Mpixel/s)
Meso2P	6×6×0.5	2048×2048×28	1	28	117.44
Light beads[25]	5.4×6×0.5	1080×1200×30	2.2	30	85.54
DEEP scope[21]	3.23×3.23×0.08	1030×515×5	2.2	5	5.83

1533 **Supplementary Table 2: Comparative benchmarking of Meso2P against existing**

1534 **large-FOV two-photon platforms for volume imaging.**

Platform	FOV (mm^2)	NA	Objective diameter (mm)	Frame rate (Hz)
Meso2P	36	0.87	37	3.83
ULTRA[18]	64	0.5	85	0.83
Light beads[25]	32.4	0.6	90	2.2
2p-RAM[23]	23.76	1	90	0.7
Ultra-large FOV 2p[14]	80	0.28	36	0.32
DEEP scope[21]	10.4329	0.58	67	5.76
FASHIO-2p[15]	9	0.8	84	7.5
Quadroscope[16]	2.25	0.52	38.21	30
Diesel2p[20]	15	0.54	90	3.84
Trepan 2p[13]	12.25	0.43	50.8	0.08

1536 **Supplementary Table 3: Specific parameter values of comparison benchmark**
1537 **between Meso2P and existing large FOV two-photon platforms.**

Supplementary Notes

Supplementary Notes 1: Quantitative benchmarking of Meso2P against large-FOV two-photon microscopes

We first compared single-plane imaging performance of Meso2P with state-of-the-art mesoscale 2P systems (Supplementary Tables 1 and 2). Using the conventional figure-of-merit FOV × NA, Meso2P exceeds all existing platforms and matches the ULTRA microscope, yet it does so with a standard 4× objective. When normalised by objective diameter (FOV × NA / D), Meso2P leads the field by a substantial margin.

In throughput, Meso2P delivers a pixel rate twice that of the next fastest system. It can operate at 4,096 × 4,096 pixels per frame (3.83 Hz), providing the combination of resolution and speed required for large-scale calcium imaging.

Among currently available mesoscopes, Meso2P is one of the few that can be upgraded to volumetric operation. Its voxel rate outperforms all reported volumetric scanners (Supplementary Table 3), and it is the only platform that maintains single-plane spatial resolution and frame rate while simultaneously acquiring 2–4 planes.

Finally, Meso2P is one of the rare large-FOV instruments that integrates two-photon optogenetic stimulation without compromising imaging performance, enabling tightly co-registered perturbation and read-out within the same optical core.