Indocyanine green excitation-emission matrix characterization: spectral shifts and application-specific spectra

Alberto J. Ruiz,[a*] **Sophie A. Lyon,**[a] **Ethan P.M. LaRochelle,**[a] **Kimberley S. Samkoe,**[b]

[a]QUEL Imaging, White River Junction, VT, USA
[b]Dartmouth Engineering, Hanover, NH, USA

Abstract

Significance: Indocyanine Green (ICG) is the most widely used fluorophore in fluorescence-guided surgery (FGS), yet its spectral characteristics can significantly vary in different microenvironments. This variability impacts the design and effectiveness of fluorescence sensing systems used in medical imaging and diagnostics.

Aim: Provide the first excitation-emission matrix (EEM) characterization of ICG in different microenvironments to comprehensively understand spectral shifts, including dimethyl sulfoxide (DMSO), bovine serum albumin (BSA) solutions, and 3D-printed (3DP) resin.

Approach: EEMs and absorbance spectra of ICG in DMSO, BSA solutions, and 3DP resin were acquired using a CCD-based fluorescence and absorbance spectrometer. The study investigated the impact of these microenvironments and varying concentrations on ICG's fluorescence behavior.

Results: ICG in DMSO exhibited symmetric spectra across varying excitation wavelengths, confirming conventional fluorophore behavior. In contrast, ICG in BSA solution and 3DP resin showed a notable 'rotation' of central spectral features, indicative of red-edge excitation shifts (REES), such that fluorescence emission varied with excitation wavelength. Furthermore, varying ICG concentration measurements showed fluorescence quenching, concentration-dependent red shifts (CDRS), and disparities between absorbance and emission spectra due to inner filter effects (IFE).

Conclusions: This study provides the first EEM characterization of ICG alongside the first report of REES in FGS fluorophores. These results highlight the importance of considering excitation wavelengths in spectral comparisons, illustrating that EEM data offers more comprehensive analysis than conventional fluorescence spectra collected at a single excitation wavelength. This study lays the groundwork for improved fluorophore characterizations for advancing fluorescence sensing in clinical applications, including FGS imaging.

Keywords: indocyanine green, ICG, excitation-emission matrix, fluorescence-guided surgery, FGS, red-edge excitation shift.

*Alberto J. Ruiz, E-mail: alberto@quelimaging.com

1 Introduction

Understanding the spectral characteristics of a fluorophore is crucial for fluorescence sensing,[1] including fluorescence-guided surgery (FGS).[2] The microenvironment of a fluorophore can significantly influence its physical and chemical behaviors, with factors such as local viscosity, polarity, temperature, redox conditions, and acidic-base status playing important roles.[3] In fluorescence imaging, the most relevant changes associated with the fluorophore's microenvironment are the excitation-emission spectral shifts and variations in the quantum yield,[4] since they affect the sensitivity and specificity of fluorescence sensing devices. Solvatochromism in fluorophores, which is the change in its spectral properties based on their microenvironment, generally involves shifts in absorbance or excitation-emission spectra due to solvent effects.[5,6] As FGS continues to develop, understanding how fluorophore signals change in various microenvironments can help advance optical system design, inter-system comparisons, and facilitate the development of imaging phantoms and targets for the clinical translation of fluorescence imaging systems.[7–9]

Indocyanine green (ICG) is the most widely used fluorophore in FGS, utilized for tissue perfusion, cardiac flow indication, and lymph node mapping.[10] Its prominence in fluorescence guidance can be attributed to its absorption and emission in the near-infrared (NIR) range, low toxicity, and extensive medical use for more than half a century.[10–13] Understanding the variability of ICG's spectral characteristics within *in vivo* environments can help optimize imaging system design and cross-system comparisons. This includes standard solvent effects as well as non-linear shifts, such as spectral broadening and changes in fluorescence emission at different excitation wavelengths, a phenomenon known as a red-edge excitation shift (REES).[14,15] These non-linear effects could be particularly significant in the body's complex environment, where proteins such as blood albumin are present.[16]

Given that system performance can affect clinical decision-making,[17] there is a compelling need for imaging standards that can help provide 'ground truth' for system characterization.[7,18] Understanding how the excitation and emission of fluorescence imaging targets can mimic *in vivo* fluorescence is essential to provide relevant one-to-one equivalence to clinical fluorescence imaging applications. Differences in fluorophore excitation and emission characteristics within different microenvironments can be quantified by acquiring its excitation-emission matrix (EEM).[19]

Fig 1 The excitation-emission matrix (EEM) of a fluorophore is generated by **(a)** acquisition of fluorescence emission spectra at various excitation wavelengths to produce the **(b)** three-dimensional EEM data set. The **(c)** EEM can be used to obtain conventional **(d)** excitation and emission spectra by isolating data along a single wavelength on the excitation or emission axis, respectively.

EEMs are three-dimensional data sets that describe the fluorescence intensity across a range of excitation and emission wavelengths. These matrices are typically acquired by collecting fluorescence emission spectra over a range of discrete excitation wavelengths (**Fig 1a,b**). Conceptually, they can be visualized as a series of emission spectra at different excitation wavelengths or, alternatively, as a series of excitation spectra at different emission wavelengths (**Fig 1c,d**). They offer insights into how fluorescence emission varies across excitation wavelengths, delivering more information than conventional excitation-emission graphs collected at a single probing wavelength. Additionally, EEMs provide the necessary data to understand phenomena such as solvachromic effects, red-edge shifts, and intensity variations across wavelengths.[15] In FGS applications, EEMs enable the generation of system-specific fluorescence spectra based on the system's excitation illumination. A significant limitation of EEMs is the long acquisition times associated with obtaining fluorescence spectra across various excitation wavelengths. However, recent advances in CCD-based fluorescence spectrometers have enabled the rapid acquisition of these

data sets. These developments provide a promising pathway for utilizing EEMs as a superior alternative

to individual spectra capture.

Here, we investigate the excitation and emission spectra changes of ICG associated with dimethyl

sulfoxide (DMSO), bovine serum albumin (BSA) solution, and 3D-printed (3DP) resin microenvironment

s, providing the first EEM characterization of ICG as well as the first reported instance of REES in FGS

fluorophores. This study provides a comprehensive analysis of changes in spectral and fluorescence

intensity within the context of the EEM, overcoming the limitations of previously reported studies,[8,9,20–24]

which focused on singular fluorescence spectral snapshots.

2   Methods

2.1 Fluorescence excitation-emission matrix acquisition

Fluorescence EEMs and absorbance spectra were collected using a CCD-based fluorescence and

absorbance spectrometer (Duetta, HORIBA Scientific). This spectrometer used a xenon lamp

monochromator to excite the samples, sweeping wavelengths from 650-900 nm at 1 nm steps.

Fluorescence emission was detected by the CCD sensor from 600-1000 nm. All spectra were collected

using a 3 nm bandpass slit size, a 0.5 nm (1 pixel) emission increment, and three measurement

accumulations. The reported max fluorescence emission values are assumed to have $\pm$ 0.5 nm uncertainty,

and the excitation values are assumed to have a $\pm$ 1 nm uncertainty, given the corresponding acquisition

step size parameters. The uncertainty of the reported Stokes shifts is assumed to be $\pm$ 1.1 nm from adding

the excitation and emission uncertainties in quadrature. A 0.75 s – 2 s integration time was used for all

samples. Integration times were kept identical for comparative measurements, enabling consistent cross-

sample fluorescence emission comparison. For the varying concentration measurements, the 3 µM and 10

µM samples were collected with a half integration time to avoid detector saturation such that their

corresponding measured intensity counts were doubled during data processing to account for integration

time differences.

The reported EEMs are calculated by subtracting the acquired data of the corresponding control (0 nM

ICG) sample from the fluorescent sample. The emission spectra at a given excitation were obtained from

plotting the intensity values along the emission axis of the EEM (**Fig 1c,d**). The excitation spectra at a

"monitored" fluorescence wavelength can be derived from plotting the intensity values along the

excitation axis of the EEM (**Fig 1c,d**).

EEMs were acquired for 1 µM and 0 µM (control) concentrations of IR-125 laser dye (Exciton Inc.,

09030) in DMSO, BSA solution, and 3DP photocurable resin. IR-125 is the laser dye marketing name of

ICG (sharing the same Chemical Abstracts Service number, 3599-32-4), with identical chemical, quantum

yield, and fluorescence spectral characteristics.[25] IR-125 dye was utilized for this study, rather than

clinical ICG dye, due to the availability of purity specifications from the respective manufacturers.

Further concentrations of IR-125 in DMSO were used to study concentration-dependent EEM effects.

Each of these samples was prepared as per the respective methods sections below. Throughout the

remainder of this study, the terms IR-125 and ICG are used synonymously.

2.2 Absorbance spectra acquisition

Absorbance spectra were collected over the 600-1000 nm range at 1 nm step increments utilizing the

spectrometer silicon photodiode detector. For all samples, an integration time of 0.1 s and a band pass slit

size of 3 nm were used alongside corresponding controls of DMSO, BSA, and 3DP resin as blanks. The

reported max absorbance values are assumed to have a $\pm$ 1 nm uncertainty, given the corresponding

acquisition step size parameters Absorbance data is reported with units of optical density (OD), where the

measured transmission $= 10^{-OD}$.

2.3 Data processing

The EEM and absorbance spectra were processed using robust local regression smoothing with a span of

40 nm. The particular method utilized, named RLOESS, uses robust locally weighted linear least squares

regression with a 2^{nd} degree polynomial model.[26,27] This robust smoothing allows for the elimination of

outlier measurements and generally provided better peak estimations than traditional smoothing

techniques (i.e. Savitzky-Golay filtering) when tested against Gaussian and skewed Gaussian fits to the

spectral data. Furthermore, the robust smoothing provided the ability to eliminate scattering peaks

introduced by the BSA and 3D printed resin. An example of this scattering artifact elimination by the

smoothing is shown in **Fig 2**. RLOESS can be implemented through the use of the MATLAB function

smooth() or in Python with the Pypi.org *loess* project[28,29] that provided equivalent robust weights to the

native MATLAB implementation.

Fig 2 EEMs for samples with inherent scattering (3DP resin and albumin solutions) exhibit a scattering peak along the excitation-emission line **(a)** which can be corrected through the use of the RLOESS smoothing algorithm to result in the post-processed EEMs **(b)**.

2.4 Preparation of DMSO solution samples

EEM and absorbance measurements for the DMSO and BSA solutions were performed on standard 4.5 mL cuvettes (12x12x45 mm, 10 mm pathlength). A stock solution of 1000 μM ICG was prepared in DMSO (Sigma-Aldrich, 472301) to pre-suspend the fluorophore and minimize error during spectral comparisons.

The 1 μM ICG in DMSO cuvette sample was prepared by further diluting the 1000 μM DMSO stock to prepare 100 mL of 1 μM ICG in DMSO. 3.5 mL of this 1 μM solution was deposited into a cuvette for data collection with a corresponding control cuvette of 3.5 mL of DMSO (0 μM).

To study concentration-dependent EEM effects, an additional dilution of 100 μM was prepared from the 1000 μM DMSO stock. This 100 μM ICG stock was used to create, through serial dilution, 10 mL samples at concentrations of 10, 3, 1, 0.3, 0.1, and 0.03 μM of ICG in DMSO. Volumes of 3.5 mL of each of these solutions were deposited into cuvettes for data collection with an additional control cuvette of 3.5 mL of DMSO (0 μM ICG).

2.5 Preparation of BSA solution samples

The 1 μM ICG in BSA solution (44 mg/mL) cuvette sample was prepared using the 1000 μM ICG in DMSO stock, 150 g/mL BSA stock, and phosphate-buffered saline (PBS). 10x concentrated PBS (Sigma-Aldrich, P7059) was diluted tenfold with distilled water to create a 1x PBS stock appropriate for use. BSA lyophilized powder (Sigma-Aldrich, A2153) was dissolved with the 1x PBS to create a BSA stock of 150 mg/mL. 100 mL of the 1 μM ICG in BSA solution (44 mg/mL) was prepared from the 1000 μM

ICG in DMSO stock, BSA stock, and 1x PBS, 3.5 mL of which was transferred to a cuvette for data

collection. A corresponding BSA (44 mg/mL) control cuvette sample (0 µM ICG) was also prepared. The

ICG in BSA solution was incubated at room temperature and protected from light for an hour prior to

measurement. It should be noted that the 1000 µM ICG in DMSO stock solution was used in this

preparation to mitigate the aggregation of dye in PBS solution, which was experimentally observed at

these high concentrations. The 44 mg/mL BSA concentration was used as a representative concentration

of albumin in human plasma.[30,31]

To study the effect of BSA concentration on spectral properties, 1 µM ICG cuvette samples with varying

BSA concentrations were prepared using the 150 mg/mL BSA stock, a 9 µM ICG in water stock, and

distilled water. The 1000 µM ICG in DMSO stock was diluted with distilled water to make a 9 µM ICG

in water stock solution. BSA was diluted with distilled water to create two sets of 3.5 mL of BSA at

concentrations of 100, 50, 25, 10, and 5 mg/mL. One set was kept as a control (0 µM) with the other set

using 389 µL of the 9 µM ICG stock to create 3.5 mL cuvette samples of 1 µM ICG in BSA solutions of

100, 50, 25, 10, and 5 mg/mL. The samples containing ICG were incubated at room temperature and

protected from light for an hour prior to measurement.

2.6 Preparation of 3D-printed samples

The 1 µM ICG in 3DP resin and the control (0 µM) samples were prepared using a proprietary clear

photocurable resin (MML-REPC-001, QUEL Imaging) and the 1000 µM ICG in DMSO stock solution.

The 3DP samples were prepared following a previously published method.[8] In brief, 100 µL of the ICG

stock was mixed into 100 mL of the clear resin and printed using layer-by-layer stereolithography

printing (405 nm curing) to produce a fluorescent 3DP cuvette (12 x 12 x 40 mm^3) to match the

dimensions of the standard cuvette samples; a control cuvette (0 µM) was also printed using the same

method. After printing, the cuvettes were cleaned with isopropyl alcohol and post-cured utilizing a high-

radiance 385 nm light (Solarez, 88903). To minimize scattering of light, the progressively wet-sanded

with distilled water using 1500, 2000, and 3000 grit silicon carbide sandpaper, followed by a two-step

liquid polisher (Novus 7100). Sanding and polishing resulted in an optically clear surface, enabling data

collection using the Duetta spectrometer comparable to liquid samples prepared in a standard cuvette.

3   Results

3.1 ICG in DMSO

The acquired EEM for 1 µM ICG in DMSO and associated spectra are shown in **Fig 3**. The EEM (**Fig 3a**)

maxima for excitation and emission were measured as 791 nm and 826.5 nm, respectively. The associated

conventional excitation and emission spectra at these maxima are plotted in **Fig 3b**, showing a measured
Stokes shift of 35.5 nm**.**

To quantify spectral shifts in the fluorescence emission and intensity, spectra from varying excitation
wavelength in the 760-820 nm range (10 nm steps) were extracted from the EEM data (**Fig 3c,d**). The
identical spectral shapes of the fluorescence emission (**Fig 3d**) indicate there are no significant shifts in
the fluorescence spectral characteristics for varying excitation wavelengths for ICG in DMSO solution.
The measured maximum emissions and normalized intensities for the various excitations, reported as
(excitation, emission maxima, relative emission intensity) were: (760 nm, 827.0 nm, 0.57), (770 nm,
826.5 nm, 0.72), (780 nm, 826.5 nm, 0.90), (790 nm, 826.5 nm, 1.0), (800 nm, 826.5 nm, 0.95), (810 nm,
826.5 nm, 0.75), and (820 nm, 827.0 nm, 0.43).

Fig. 3 EEM and associated spectra for 1 µM ICG in DMSO: **(a)** Top-down plot of the acquired EEM. **(b)**
Conventional excitation and emission spectra extracted from the EEM maxima. **(c)** Emission spectra plots at varying
excitations in the 760-820 nm range at 10 nm steps to visualize intensity shifts and **(d)** normalized counterparts to
visualize spectra maxima shifts.

The acquired EEMs of ICG in DMSO at varying concentrations of 0.03, 0.1, 0.3, 1.0, 3.0, and 10 µM are

provided in **Figure S1**. These measurements showed fluorescence quenching at the 10 µM concentration

(**Fig 4a**), concentration-dependent red shifts (CDRS) (**Fig 4b**), and disparities between absorbance and

emission spectra due to inner filter effects (IFE) (**Fig 4c,d**). The log-log plot of maximum fluorescence

intensity vs. concentration (**Fig 4a**) shows a constant increase in fluorescence emission vs. concentration

for the 0.03 - 3 µM range with a drop in intensity for the 10 µM concentration. This drop in fluorescence

intensity at the highest concentration is most likely attributed to both primary and secondary inner filter

effects (IFE), which correspond to quenching caused by attenuation of the excitation beam and re-

absorption of emitted fluorescence, respectively.[32]

To simplify discussion of concentration dependent fluorescence shifts, normalized fluorescence emission

spectra at 785 nm excitation were extracted from the EEMs (**Fig 4b**). The corresponding measured

maximum emission wavelengths and normalized intensities for these plots were: (0.03 µM, 822.0 nm,

0.020), (0.1 µM, 823.0 nm, 0.066), (0.3 µM, 824.0 nm, 0.19), (1 µM, 826.0 nm, 0.52), (3 µM, 830.5 nm,

1.0), and (10 µM, 839.5 nm, 0.89). This increase in fluorescence emission maxima (red-shift) for varying

concentrations is termed CDRS, an effect that that primarily results from the re-absorption of emitted

fluorescence.[32]

The measured absorbance data is plotted in **Fig 4c,** showing equivalent spectra from the full 0.03-10 µM

concentration range. Data from the 0.03 µM sample was excluded from **Fig 4c** due to the low signal-to-

noise ratio of its absorbance acquisition. **Table S1** contains summarized OD absorbance measurements

for the varying concentrations, showing no significant variation in calculated molar extinction coefficients

(**Table S2**) for the 0.1-10 µM range. In contrast to absorbance, the excitation spectra (**Fig 4d**) at varying

concentrations, generated from EEMs at the **Fig 4b** emission peaks, shows broadening and quenching for

the 3 µM and 10 µM concentrations. This discrepancy between the absorbance and excitation spectra is

caused by secondary IFE effects from the re-absorption of fluorescence emission at these high

concentrations.[32] with the calculated molar extinction coefficients provided in.

Fig. 4: Concentration-dependent spectral effects for ICG in DMSO. **(a)** Plot of maximum fluorescence intensity vs.

concentration showing a consistent increase for the 0.03 - 3 µM range with a drop-off in intensity for the 10 µM

concentration. **(b)** Normalized fluorescence emission spectra at 785 nm excitation showing concentration-dependent

red shifts. **(c)** Absorbance spectra show consistent spectral features over the entire concentration while the **(d)**

excitation spectra show broadening and changes due to fluorescence re-absorption for the 3 µM and 10 µM

concentrations.

3.2 ICG in BSA Solution

The acquired EEM for 1 µM ICG in BSA solution (44 mg/mL) and associated spectra are shown in **Fig 5**.

Compared to the ICG in DMSO EEM (**Fig 3a**), the ICG in BSA solution EEM (**Fig 5a**) showed a

significant 'rotation' of the central spectra feature, indicating an excitation-dependent REES which causes

significant shifts in the emission spectral characteristics for varying excitation wavelengths.[15] The EEM

maxima for excitation and emission were measured as 794.0 nm and 813.5 nm, respectively. The

associated conventional excitation and emission spectra at these maxima are plotted in **Fig 5b**, showing a

measured Stokes shift of 19.5 nm.

To quantify spectral shifts in the fluorescence emission and intensity, spectra from varying excitation

wavelengths in the 760-820 nm range (10 nm steps) were extracted from the EEM data (**Fig 5c,d**). The

fluorescence emission spectra for varying BSA concentrations (**Fig 5d**) indicate that there are excitation-dependent changes in the central peak wavelength but no significant changes in photon distribution (spectral broadening) for the ICG in BSA solution. The measured maximum emissions and normalized intensities for the various excitations, reported as (excitation, emission maxima, relative emission intensity) were: (760 nm, 808.0 nm, 0.44), (770 nm, 807.0 nm, 0.62), (780 nm, 807.5 nm, 0.84), (790 nm, 811.5 nm, 0.99), (800 nm, 816.0 nm, 1.0), (810 nm, 819.0 nm, 0.84), and (820 nm, 823.0 nm, 0.48). The average red-emission shift observed over the 780-820 nm excitation range was 0.39 nm of emission per nm of excitation.

Fig. 5 EEM and associated spectra for 1 μM ICG in BSA solution (44 mg/mL). **(a)** Top-down plot of the acquired EEM showing a 'rotation' of the central spectral feature indicating excitation-dependent spectral shifts. **(b)** Conventional excitation and emission spectra extracted from the EEM maxima. **(c)** Emission spectra plots at varying excitations in the 760-820 nm range at 10 nm steps to visualize intensity shifts and **(d)** normalized counterparts to visualize spectra maxima shifts.

The acquired ICG EEMs at different concentrations of BSA in the 5 mg/mL – 100 mg/mL range are provided in **Figure S2**. The EEM measurements showed no significant changes in spectral characteristics for varying BSA mg/mL concentration within the tested range, which covers the full range of biologically

relevant concentrations.[30,31] To simplify discussion of BSA concentration dependent fluorescence shifts,

normalized fluorescence emission spectra at 785 nm excitation were extracted from the EEMs (**Fig 6**).

The relative fluorescence emission plots (**Fig 6a**) show no significant changes in the emission intensities,

and the normalized plots (**Fig 6b**) show no significant changes in the photon distributions of the emission.

The measured maximum emissions and normalized intensities for the various concentrations of BSA

were: (5 mg/mL, 816.0 nm, 0.96), (10 mg/mL, 815.5 nm, 0.96), (25 mg/mL, 815.5 nm, 0.97), (50 mg/mL,

815.0 nm, 0.94), and (100 mg/mL, 815.5 nm, 1.0).

Fig. 6 Concentration-dependent fluorescence emission measurements for the BSA solutions at 785 nm excitation for

(a) the relative emissions to visualize changes in intensity and **(b)** normalized emissions to visualize spectral

distribution changes.

3.3 ICG in 3DP Resin

The acquired EEM for 1 µM ICG in 3DP resin and associated spectra are shown in **Fig 7**. Compared to

the ICG in DMSO EEM (**Fig 3a**), the ICG in 3DP resin EEM (**Fig 7a**) showed a significant 'rotation' of

the central spectra feature, indicating an excitation-dependent REES which causes significant shifts in the

emission spectral characteristics for varying excitation wavelengths.[15] The EEM maxima for excitation

and emission were measured as 809 nm and 822.0 nm, respectively. The associated conventional

excitation and emission spectra at these maxima are plotted in **Fig 7b**, showing a measured Stokes shift of

13 nm**.**

To quantify spectral shifts in the fluorescence emission and intensity, spectra from varying excitation

wavelength in the 760-820 nm range (10 nm steps) were extracted from the EEM data (**Fig 7c,d**). The

varying spectral shapes of the fluorescence emission (**Fig 7d**) indicate there are changes in the peak

wavelength and photon distribution (spectral broadening) with excitation wavelength for ICG in 3DP

resin. The measured maximum emissions and normalized intensities for the various excitations, reported

as (excitation, emission maxima, relative emission intensity) were: (760 nm, 818.5 nm, 0.34), (770 nm, 809.0 nm, 0.44), (780 nm, 811.5 nm, 0.61), (790 nm, 815.0 nm, 0.80), (800 nm, 818.0 nm, 0.96), (810 nm, 823.0 nm, 1.0), and (820 nm, 830.0 nm, 0.82). The average red-emission shift observed over this excitation range was 0.45 nm per nm of excitation.

Fig. 7 EEM and associated spectra for 1 µM ICG in 3DP resin. **(a)** Top-down plot of the acquired EEM showing a 'rotation' of the central spectral feature indicating excitation-dependent spectral shifts. **(b)** Conventional excitation and emission spectra extracted from the EEM maxima. **(c)** Emission spectra plots at varying excitations in the 760-820 nm range at 10 nm steps to visualize intensity shifts and **(d)** normalized counterparts to visualize spectra maxima shifts.

3.4 ICG EEM Comparison

The absorbance spectra, excitation at peak emission spectra, and emission spectra at 785 nm and 805 nm excitation for ICG in DMSO, BSA solution (44 mg/mL), and 3DP resin are shown in **Fig 8** alongside summarized data in **Table 1**.

The absorbance spectra (**Fig 8a**) measured absorbance maxima of 792, 795, and 808 nm for DMSO, BSA

solution, and 3DP resin, respectively, with similar OD for their maxima showing <5% difference in

transmission (**Table 1**). The excitation spectra at peak emission (**Fig 8b**) show similar differences in peak

emissions but with changing photon distributions for the BSA solution and 3DP resin. These

measurements demonstrate the discrepancies between absorbance and excitation spectra caused by REES

in fluorophore microenvironments.

Fig. 8 Spectral comparisons of ICG in DMSO, BSA solution (44 mg/mL), and 3DP solution: (a) Absorbance

spectra, (b) excitation spectra at peak emission, (c) fluorescence emission spectra at 785 nm excitation, and (d)

fluorescence emission spectra at 805 nm excitation.

Table 1: Summarized data for the comparison of the spectral behavior of ICG in DMSO, BSA solution and 3DP

resin microenvironments.

	Absorbance		Excitation		Emission (785 nm excitation)		Emission (805 nm excitation)	
	Max (nm)	OD	Max (nm)	Intensity (au)	Max (nm)	Intensity (au)	Max (nm)	Intensity (au)
DMSO	792	0.167	791	n/a	826	1.00	827	1.00
Albumin	795	0.142	794	n/a	810	0.478	817	0.538
3DP	808	0.158	802	n/a	813	0.487	821	0.772

Given that fluorescence sensing relies on fluorescence emission detection, excitation-specific spectra provide the most relevant comparisons for different microenvironments. The fluorescence emission spectra for 785 nm and 805 nm excitation are provided in **Fig. 8c,d**, which are two of the most common wavelengths used for ICG excitation in fluorescence imaging applications. These spectra demonstrate notable overlap between the BSA solution and 3DP resin, with almost identical peak emission intensities at the 785 nm excitation.[7] It should be noted that despite the < 5% difference in transmittance between the three samples, the emission spectra showed ~50% decrease in fluorescence for the BSA solution and 3DP resin when compared to DMSO at 785 nm excitation (**Table 1**). Furthermore, the excitation-dependent fluorescence intensity and photon distributions highlight the need to consider the excitation wavelength and emission collection of the fluorescence sensing device to provide relevant comparisons within varying fluorophore microenvironments and system design parameters.

4   Discussion

Here we presented the first EEM measurements of ICG in DMSO, BSA solutions, and 3DP resin, allowing for a comprehensive understanding of fluorescence excitation-emission shifts in these microenvironments alongside the first reported case of REES for FGS fluorophores.

4.1 ICG in DMSO

The EEM for 1 µM ICG in DMSO (**Fig 3**) provided 'symmetric' spectra throughout the varying excitation wavelengths with measured emission peaks of 826.5 nm ± 0.5 nm over the 760 – 820 nm excitation range. This EEM confirms the conventional fluorophore behavior that assumes invariant spectral photon distribution of fluorescence emission with excitation wavelength.

The EEMs and absorbance at varying ICG concentrations in DMSO (**Fig 4**) show the effects of

concentration-dependent fluorescence shifts including fluorescence quenching, CDRS, and disparities

between the absorbance and excitation spectra due to IFEs. Increases in fluorophore concentration lead to

a proportional increase in fluorescence emission intensity within a linear range, followed by a decrease at

higher concentrations due to IFEs or fluorophore aggregation.[32,33] The measured linear range for ICG in

DMSO was 0.3 – 3 µM with the 10 µM sample showing significant quenching, emission shifts (CDRS)

and distortion of the excitation spectra. The 3 µM sample exhibited minor quenching, CDRS, and

excitation spectra broadening. Within the linear range, the measured EEMs (**Fig S1**) demonstrated no

significant differences, displaying an anticipated increase in emission intensity proportional to

concentration across the entire wavelength matrix. The spectral distribution of the absorbance spectra was

consistent along the full concentration range, which indicates that the observed fluorescence shifts are due

to IFEs and not fluorophore aggregation at these high concentrations. The photon distribution of the

absorbance and excitation spectra were equivalent within the identified linear range. It is worth noting

that this measured linearity range is specific to the geometry of the spectral measurement, which utilized

an orthogonal excitation-emission collection with a 10 mm pathlength. These results highlight the need to

understand the linearity range of a fluorophore to adequately choose concentrations for EEM and

fluorescence spectra characterization.

4.2 ICG in BSA solution

In contrast to the ICG in DMSO EEM, the 1 µM ICG in BSA solution (**Fig 5**) showed a 'rotation' of the

central spectral feature indicating the presence of REES and the dependence of emission spectra on

excitation wavelength. This nonlinear effect is most likely attributed to albumin binding[16,34] causing

changes in solvent relaxation times, where the excitation red shift is observed because relaxation is not

complete before fluorescence emission occurs.[15] These results emphasize the importance of EEM

fluorophore measurements to characterize spectral effects associated with different microenvironments,

including REES.

EEM measurements for ICG in varying BSA concentration solutions (**Fig 6, Figure S2**) showed the

same REES, with no significant spectral differences within the biologically-relevant albumin

concentrations in the 5 - 100 mg/mL range.[30,31] Consideration of REES fluorescence emission shifts for

ICG in BSA solutions could help optimize fluorescence sensing designs including FGS imagers given the

relevancy to perfusion and *in vivo* applications. The measured ICG EEM in the BSA solution (**Fig 5**)

provides spectral information for albumin concentrations found in whole blood,[31,35] thus serving as a good

estimator of *in vivo* ICG spectra for perfusion applications. Furthermore, the invariability of the EEM

within the wide albumin concentration range indicates that it might be a suitable estimator of interstitial
ICG spectra.[35]

373
4.3 ICG in 3DP resin

Similarly to the ICG in BSA solution EEM, the ICG in 3DP resin EEM (**Fig 7**) showed a 'rotation' of the
central spectral feature indicating the presence of REES and the dependence of emission spectra on
excitation wavelength. This nonlinear effect is most likely attributed to the embedding of the fluorophore
within a solid polymer matrix, causing changes in relaxation times, where the excitation red shift is
observed because relaxation is not complete before fluorescence emission occurs.[15] Given the intended
use of these 3DP fluorescent materials as 'ground-truth' measurements, in the form of reference targets
and phantoms,[7–9,18] the presence of REES underscores the need to consider the excitation wavelength for
accurate, application-specific, and system-specific characterization and comparisons.

383
4.4 ICG EEM comparison

The EEM of ICG in DMSO, BSA solution, and 3DP resin**,** along their corresponding fluorescence and
absorption spectra (**Fig 8**), revealed shifts in spectral emission wavelengths and distributions, including
excitation-dependent fluorescence effects in the BSA and 3DP samples. As fluorescence sensing
applications predominantly focus on emission, the overlap in emission spectra at specific excitation
wavelengths offers a more accurate representation of fluorescent behavior in different micro-
environments. Notably, the ICG in BSA solution and 3DP resin samples displayed significant emission
overlap (**Fig 8c,d**), particularly at excitation wavelengths commonly used in FGS devices (785-805 nm
range). This indicates that the ICG in 3DP material can serve as an adequate reference for ICG in BSA
solution when accounting for both system excitation and emission collection parameters.

Considering the potential shifts in both spectral distribution and peak emission wavelengths across
different microenvironments and excitation wavelengths, the most accurate method for comparing
fluorescence emission intensity involves integration under the curve, tailored to the specific collection
parameters of the system (such as long-pass emission collection, band-pass emission collection, etc.).
This dataset facilitates application and system-specific comparisons, incorporating considerations of both
excitation and emission collection parameters. It is important to note that the provided EEMs are directly
comparable in their reported intensities, as they were acquired using the same acquisition parameters.
Although not explored in our analysis, it is worth noting that 'weighted emission spectra' can be
generated from the provided EEMs for broadband excitation sources.

403
4.5 Future work

The presented EEM and absorbance measurements provided comprehensive insights into ICG spectral
shifts in varying microenvironments but were limited to room-temperature conditions for the solvent and
3DP samples. Further insights could be gained by including whole blood in the suite of tested solvents.
Although the high absorption and scattering in whole blood may pose measurement challenges,
understanding REES in this context would improve the *in vivo* spectral estimations. Moreover,
conducting *in vivo* fluorescence spectral measurements, which would require specialized equipment,
could further understanding the predominance of albumin binding on *in-vivo* ICG EEM shifts.
Understanding of temperature effects on EEMs would also provide valuable insight. Additionally, IFE
corrections[36] and spectral unmixing of EEMs could further help advance fluorophore characterization,
including understanding of non-linear and aggregation effects.

As FGS targeted fluorophores become available, measuring EEM spectra for bound fluorophores that
mimic the *in vivo* environment will be crucial to understanding their behavior. This is complicated further
by the chemical environments of the targeted binding site. If solvent effects for a fluorophore are found to
be limited, this could indicate 'robust' spectral properties, with no significant emission and absorption
shifts observed despite changes in microenvironment. Conversely, significant spectral changes due to
solvent effects would be important to identify prior to clinical application and could inform the imaging
system design or further modification of the fluorophore structure.

5 Conclusion

This study presents the first detailed EEM characterization of ICG, including DMSO, BSA solutions, and
3DP resin microenvironments. Our research provides new insights into the spectral shifts of ICG, notably
the presence of REES in FGS fluorophores. The results highlight the significant impact of different
solvents and environments on the fluorescence behavior of ICG, emphasizing the necessity of accounting
for these variations in fluorescence sensing applications. The observed spectral changes, dependent on
excitation wavelengths and specific microenvironments, stress the importance of precise fluorophore
concentration selection and a thorough understanding of spectral properties for effective biomedical
imaging. Furthermore, this underscores the importance of considering excitation wavelength in spectral
comparisons, showing that EEM data, which captures spectral variations across a range of excitation
wavelengths, provides a more comprehensive analysis than conventional fluorescence spectra collected at
a single wavelength. This study lays the groundwork for improved fluorophore characterizations to
further explore these spectral shifts under varied conditions, including temperature effects and *in vivo*

environments, which can help optimize fluorophore chemistry and fluorescence-based imaging system
design for clinical settings.

Disclosures
AJR, SAL, and EPML work for QUEL Imaging, which is a medical imaging phantom company.
KSS declares no conflicts of interest.

Acknowledgments
This work was partially funded by the NIBIB R44 grant EB029804-02.

Data, Materials, and Code Availability
Data and code are available upon reasonable request.

References

1. "Chemical Sensing via Fluorescence," in Molecular Fluorescence, pp. 409–478, John Wiley & Sons,
Ltd (2012) [doi:10.1002/9783527650002.ch14].

2. "Autofluorescence and Fluorescence Labeling in Biology and Medicine," in Molecular Fluorescence,
pp. 479–505, John Wiley & Sons, Ltd (2012) [doi:10.1002/9783527650002.ch15].

3. Z. Yang et al., "Macro-/micro-environment-sensitive chemosensing and biological imaging," Chem
Soc Rev **43**(13), 4563–4601 (2014) [doi:10.1039/C4CS00051J].

4. N. O. Mchedlov-Petrossyan, N. A. Vodolazkaya, and A. O. Doroshenko, "Ionic Equilibria of
Fluorophores in Organized Solutions: The Influence of Micellar Microenvironment on Protolytic and
Photophysical Properties of Rhodamine B," 14.

5. J. R. Lakowicz, Ed., "Solvent and Environmental Effects," in Principles of Fluorescence
Spectroscopy, pp. 205–235, Springer US, Boston, MA (2006) [doi:10.1007/978-0-387-46312-4_6].

6. "Environmental Effects on Fluorescence Emission," in Molecular Fluorescence, pp. 109–140, John
Wiley & Sons, Ltd (2012) [doi:10.1002/9783527650002.ch5].

7. M. I. Ochoa et al., "Assessment of open-field fluorescence guided surgery systems: implementing a standardized method for characterization and comparison," J. Biomed. Opt. **28**(9), 096007, SPIE (2023) [doi:10.1117/1.JBO.28.9.096007].

8. A. J. Ruiz et al., "3D printing fluorescent material with tunable optical properties," Sci. Rep. **11**(1), 17135 (2021) [doi:10.1038/s41598-021-96496-0].

9. A. J. Ruiz et al., "Indocyanine green matching phantom for fluorescence-guided surgery imaging system characterization and performance assessment," J. Biomed. Opt. **25**(05), 1 (2020) [doi:10.1117/1.JBO.25.5.056003].

10. J. A. Zelken and A. P. Tufaro, "Current Trends and Emerging Future of Indocyanine Green Usage in Surgery and Oncology: An Update," Ann. Surg. Oncol. **22**(S3), 1271–1283 (2015) [doi:10.1245/s10434-015-4743-5].

11. M. Koch and V. Ntziachristos, "Advancing Surgical Vision with Fluorescence Imaging," Annu. Rev. Med. **67**(1), 153–164 (2016) [doi:10.1146/annurev-med-051914-022043].

12. J. T. Alander et al., "A Review of Indocyanine Green Fluorescent Imaging in Surgery," International Journal of Biomedical Imaging, 2012, [doi:10.1155/2012/940585] (accessed 25 November 2019).

13. M. V. Marshall et al., "Near-Infrared Fluorescence Imaging in Humans with Indocyanine Green: A Review and Update," Open Surg. Oncol. J. Online **2**(2), 12–25 (2010) [doi:10.2174/1876504101002010012].

14. A. Chattopadhyay and S. Mukherjee, "Fluorophore environments in membrane-bound probes: A red edge excitation shift study," Biochemistry **32**(14), 3804–3811 (1993) [doi:10.1021/bi00065a037].

15. J. R. Lakowicz, Ed., "Dynamics of Solvent and Spectral Relaxation," in Principles of Fluorescence Spectroscopy, pp. 237–276, Springer US, Boston, MA (2006) [doi:10.1007/978-0-387-46312-4_7].

16. S. Yoneya et al., "Binding properties of indocyanine green in human blood.," Invest. Ophthalmol. Vis. Sci. **39**(7), 1286–1290 (1998).

17. M. Koch, P. Symvoulidis, and V. Ntziachristos, "Tackling standardization in fluorescence molecular imaging," Nat. Photonics **12**(9), 505 (2018) [doi:10.1038/s41566-018-0221-5].

18. E. P. M. LaRochelle et al., "3D-Printed Tumor Phantoms for Assessment of In Vivo Fluorescence Imaging Analysis Methods," Mol. Imaging Biol. **25**(1), 212–220 (2023) [doi:10.1007/s11307-022-01783-5].

19. R. S. DaCosta, H. Andersson, and B. C. Wilson, "Molecular Fluorescence Excitation-Emission
Matrices Relevant to Tissue Spectroscopy¶," Photochem. Photobiol. **78**(4), 384–392 (2007)
[doi:10.1562/0031-8655(2003)0780384MFEMRT2.0.CO2].

20. M. L. Landsman et al., "Light-absorbing properties, stability, and spectral stabilization of
indocyanine green," J. Appl. Physiol. **40**(4), 575–583 (1976) [doi:10.1152/jappl.1976.40.4.575].

21. R. H. Patel et al., "Multifunctionality of indocyanine green-loaded biodegradable nanoparticles for
enhanced optical imaging and hyperthermia intervention of cancer," J. Biomed. Opt. **17**(4), 046003
(2012) [doi:10.1117/1.JBO.17.4.046003].

22. E. I. Altınoğlu et al., "Near-Infrared Emitting Fluorophore-Doped Calcium Phosphate Nanoparticles
for *In Vivo* Imaging of Human Breast Cancer," ACS Nano **2**(10), 2075–2084 (2008)
[doi:10.1021/nn800448r].

23. U. Kanniyappan et al., "Near-infrared fluorescence image quality test methods for standardized
performance evaluation," in Design and Quality for Biomedical Technologies X **10056**, pp. 91–98,
SPIE (2017) [doi:10.1117/12.2255867].

24. T. Desmettre, J. M. Devoisselle, and S. Mordon, "Fluorescence Properties and Metabolic Features of
Indocyanine Green (ICG) as Related to Angiography," Surv. Ophthalmol. **45**(1), 15–27 (2000)
[doi:10.1016/S0039-6257(00)00123-5].

25. S. Reindl et al., "Quantum yield of triplet formation for indocyanine green," J. Photochem. Photobiol.
Chem. **105**(1), 65–68 (1997) [doi:10.1016/S1010-6030(96)04584-4].

26. I. Gijbels and I. Prosdocimi, "Loess," WIREs Comput. Stat. **2**(5), 590–599 (2010)
[doi:10.1002/wics.104].

27. W. S. Cleveland and S. J. Devlin, "Locally Weighted Regression: An Approach to Regression
Analysis by Local Fitting," J. Am. Stat. Assoc. **83**(403), 596–610, Taylor & Francis (1988)
[doi:10.1080/01621459.1988.10478639].

28. M. Cappellari, "loess: LOESS: smoothing via robust locally-weighted regression in one or two
dimensions," Python.

29. M. Cappellari et al., "The ATLAS3D project - XX. Mass-size and mass-σ distributions of early-type
galaxies: bulge fraction drives kinematics, mass-to-light ratio, molecular gas fraction and stellar
initial mass function," Mon. Not. R. Astron. Soc. **432**, 1862–1893 (2013) [doi:10.1093/mnras/stt644].

30. "5 Human Albumin," Transfus. Med. Hemotherapy **36**(6), 399–407 (2009).

31. S. Choi et al., "A rapid, simple measurement of human albumin in whole blood using a fluorescence
immunoassay (I)," Clin. Chim. Acta **339**(1–2), 147–156 (2004) [doi:10.1016/j.cccn.2003.10.002].

32. S. Kumar Panigrahi and A. Kumar Mishra, "Inner filter effect in fluorescence spectroscopy: As a
problem and as a solution," J. Photochem. Photobiol. C Photochem. Rev. **41**, 100318 (2019)
[doi:10.1016/j.jphotochemrev.2019.100318].

33. J. R. Lakowicz, "Quenching of Fluorescence," in Principles of Fluorescence Spectroscopy, J. R.
Lakowicz, Ed., pp. 257–301, Springer US, Boston, MA (1983) [doi:10.1007/978-1-4615-7658-7_9].

34. L. Bai et al., "Super-stable cyanine@albumin fluorophore for enhanced NIR-II bioimaging,"
Theranostics **12**(10), 4536–4547 (2022) [doi:10.7150/thno.71443].

35. M. Ellmerer et al., "Measurement of interstitial albumin in human skeletal muscle and adipose tissue
by open-flow microperfusion," Am. J. Physiol.-Endocrinol. Metab. **278**(2), E352–E356 (2000)
[doi:10.1152/ajpendo.2000.278.2.E352].

36. T. Wang, L.-H. Zeng, and D.-L. Li, "A review on the methods for correcting the fluorescence inner-
filter effect of fluorescence spectrum," Appl. Spectrosc. Rev. **52**(10), 883–908, Taylor & Francis
(2017) [doi:10.1080/05704928.2017.1345758].

<u>Caption List</u>

Fig 1 The excitation-emission matrix (EEM) of a fluorophore is generated by **(a)** acquisition of fluorescence
emission spectra at various excitation wavelengths to produce the **(b)** three-dimensional EEM data set. The **(c)** EEM
can be used to obtain conventional **(d)** excitation and emission spectra by isolating data along a single wavelength
on the excitation or emission axis, respectively.

Fig 2 EEMs for samples with inherent scattering (3DP resin and albumin solutions) exhibit a scattering peak along
the excitation-emission line **(a)** which can be corrected through the use of the RLOESS smoothing algorithm to
result in the post-processed EEMs **(b)**.

Fig. 3 EEM and associated spectra for 1 µM ICG in DMSO: **(a)** Top-down plot of the acquired EEM. **(b)**
Conventional excitation and emission spectra extracted from the EEM maxima. **(c)** Emission spectra plots at varying
excitations in the 760-820 nm range at 10 nm steps to visualize intensity shifts and **(d)** normalized counterparts to
visualize spectra maxima shifts.

Fig. 4: Concentration-dependent spectral effects for ICG in DMSO. **(a)** Plot of maximum fluorescence intensity vs. concentration showing a consistent increase for the 0.03 - 3 µM range with a drop-off in intensity for the 10 µM concentration. **(b)** Normalized fluorescence emission spectra at 785 nm excitation showing concentration-dependent red shifts. **(c)** Absorbance spectra show consistent spectral features over the entire concentration while the **(d)** excitation spectra show broadening and changes due to fluorescence re-absorption for the 3 µM and 10 µM concentrations.

Fig. 5 EEM and associated spectra for 1 µM ICG in BSA solution (44 mg/mL). **(a)** Top-down plot of the acquired EEM showing a 'rotation' of the central spectral feature indicating excitation-dependent spectral shifts. **(b)** Conventional excitation and emission spectra extracted from the EEM maxima. **(c)** Emission spectra plots at varying excitations in the 760-820 nm range at 10 nm steps to visualize intensity shifts and **(d)** normalized counterparts to visualize spectra maxima shifts.

Fig. 6 Concentration-dependent fluorescence emission measurements for the BSA solutions at 785 nm excitation for **(a)** the relative emissions to visualize changes in intensity and **(b)** normalized emissions to visualize spectral distribution changes.

Fig. 7 EEM and associated spectra for 1 µM ICG in 3DP resin. **(a)** Top-down plot of the acquired EEM showing a 'rotation' of the central spectral feature indicating excitation-dependent spectral shifts. **(b)** Conventional excitation and emission spectra extracted from the EEM maxima. **(c)** Emission spectra plots at varying excitations in the 760-820 nm range at 10 nm steps to visualize intensity shifts and **(d)** normalized counterparts to visualize spectra maxima shifts.

Fig. 8 Spectral comparisons of ICG in DMSO, BSA solution (44 mg/mL), and 3DP solution: (a) Absorbance spectra, (b) excitation spectra at peak emission, (c) fluorescence emission spectra at 785 nm excitation, and (d) fluorescence emission spectra at 805 nm excitation.

Table 1: Summarized data for the comparison of the spectral behavior of ICG in DMSO, BSA solution and 3DP resin microenvironments.