In vivo microvascular flow quantification in the mouse brain using Row-Column Ultrasound Localization Microscopy and directed graph analysis

Adrien Bertolo[1,2], Jeremy Ferrier[2], Oscar Demeulenaere[1], Alexandre Dizeux[1], Tanguy Delaporte[2], Bruno Osmanski[2], Mickael Tanter[1], Mathieu Pernot[1*], Thomas Deffieux[1*]

[1] Physics for Medicine, ESPCI-PSL, Inserm U1273, CNRS, 10 rue Oradour sur Glane, Paris 75015, France

[2] Iconeus, Paris 75014, France

*These authors contributed equally to this work

Abstract

Brain perfusion relies on a complex vascular network of arteries, veins, and capillaries to meet its constant demand for oxygen and nutrients. Disruption of this microvascular system is a hallmark of many neurological disorders, including small vessel disease, stroke, and brain tumors. As such, high-resolution in vivo imaging of cerebral microvascular flow and structure remains critical to understanding these pathologies. Among them, Ultrasound Localization Microscopy (ULM) allows noninvasive imaging of microvascular network at subwavelength resolution using injected microbubbles, but the approach remains mainly limited to 2D imaging with few volumetric implementations. In this study, we explore in vivo transcranial 3D ULM of the mouse brain using Row-Column Arrays (RCA) and introduce an analysis framework to build a flow-directed vascular graph from the ULM microbubble tracking data, allowing to differentiate between subgraphs of artery-like and vein-like vascular segments. Combined with Allen-based and Radius-based

24   segmentations, we extract metrics including cerebral blood flow (CBF), microbubble (MB) velocity,

25   flowrate, CBV fraction, vascular segment length and tortuosity in sixty-six different vascular classes.

26   This high-sensitivity framework enables in vivo microvascular imaging and quantification in mice

27   and provides a scalable platform for preclinical neurovascular studies in health and disease.

Keywords

29   Microvascular imaging, Ultrasound Localization Microscopy, Row column array, Vascular graph,

30   Mouse

Introduction

Preclinical neuroimaging has become an indispensable tool in neuroscience research for imaging the

brain's vascular network, a highly intricate system responsible for regulating cerebral blood flow,

oxygen and nutrient delivery, and metabolic waste clearance. This network is fundamental to

maintaining brain health, yet many aspects of its structure and function remain poorly characterized.

Emerging evidence suggests that alterations in microvascular structure and function may serve as

early biomarkers for a range of neurological disorders[1]. For instance, small-vessel diseases and

neurodegenerative disorders - such as Alzheimer's disease - are often preceded by microvascular

alterations[2], and such alterations may contribute to disease progression before overt symptoms appear.

In vascular stroke, phenomena such as vasospasms, aneurysms, and delayed ischemia disrupt blood

flow across multiple spatial and temporal scales and are critical areas of study for developing and

testing effective recovery and intervention strategies[3]. In brain tumors, the formation of aberrant and

complex vascular networks is essential for supporting rapid tumor growth and survival. Imaging these

networks is not only vital for understanding tumor biology but also for developing targeted

therapeutic approaches to disrupt their blood supply and limit progression.

Several imaging modalities have been developed for neurovascular mapping, including magnetic

resonance angiography (MRA)[4,5] and computed tomography angiography (CT) using contrast agents

but also optical imaging techniques[6]. However, MRA and CT imaging techniques are limited in their

ability to provide both high-resolution imaging of microvascular architecture and blood flow

dynamics, while optical imaging is restricted to surface and cannot image deeper brain regions. While

advanced ex vivo techniques, such as optical clearing methods (e.g., iDISCO, CLARITY)[7] and

synchrotron-based phase-contrast tomography[8-10], excel in providing unparalleled spatial resolution

for microvascular imaging, they come with significant limitations. These techniques are highly

effective for detailed structural visualization of cerebral vasculature, often achieving resolutions at the

sub-micron to micron scale but are inherently restricted to fixed tissue samples and thus lack the

ability to capture dynamic flow information, such as blood velocity or cerebral blood flow (CBF).

These ex-vivo approaches are not suitable for longitudinal studies, as they require tissue extraction

and processing, preventing repeated measurements in alive subjects.

These limitations highlight the need for novel imaging modalities that can provide both high-

resolution structural imaging and dynamic functional assessment of the brain's microvasculature in

vivo. Ultrasound Localization Microscopy (ULM) is an emerging imaging technique that can map

microvascular flow at high spatial and temporal resolution. ULM relies on the imaging, localization,

and tracking of FDA-approved gas microbubbles circulating in the blood flow[11] using ultrafast

ultrasound imaging. The technique has been applied to map cerebral microvascular flow in rodents

and in humans using 2D imaging approaches allowing subwavelength resolution in depth and

providing quantitative velocity estimates[12]. Recently, functional ULM (fULM) has been introduced to

map functional hyperemia in the rodent brain at high spatial resolution and in time[13], access to high

spatial resolution maps allows to probe different vascular compartments and further investigate the

neurovascular coupling along different scales in depth.

While Ultrasound Localization Microscopy (ULM) has significant potential to increase our

understanding of cerebral microvasculature, its generic two-dimensional implementation presents

important limitations. Single-plane imaging not only limits the spatial regions that can be examined

simultaneously but also introduces systematic biases in velocity measurements, as only two

components of the three-dimensional velocity vector can be reliably captured. These constraints have

motivated the development of three-dimensional (3D) ULM approaches using dense ultrasonic matrix

arrays, composed of thousands of piezoelectric elements arranged in two-dimensional configurations

instead of linear configuration for rodent brain imaging including stroke and tumor models [14–18].

Despite these early proofs of concepts, matrix arrays still face substantial practical and technical

challenges. Manufacturing these arrays remains costly and technically demanding, particularly at the

high frequencies above 10 MHz typically required for preclinical neuroimaging applications such as

rodents, and a fundamental trade-off exists between acoustic element pitch, which governs sidelobe

levels and overall sensitivity, and the practical challenges of array fabrication and precision[19]. This

compromise makes scaling to larger fields of view particularly challenging without exponentially

increasing the number of elements, thereby escalating costs and system complexity. While electronic

multiplexers can mitigate some of these requirements by dynamically switching element[17,20], this

approach inevitably reduces the effective volume rate due to circuit switching times and the sequential

addressing of array elements, compromising microbubble tracking efficiency[21] in the case of ULM

and overall image quality[22].

To address these limitations, Row-Column Array (RCA) probes have emerged as a promising

alternative for three-dimensional high volume rate ultrasound imaging. Conceptually, RCA probes

consist of two overlapped orthogonal one-dimensional arrays which offers a simplified, yet efficient

approach compared to matrix arrays for 3D ultrasound imaging. RCA enables larger field-of-view

imaging with reduced complexity while maintaining improved sensitivity and can be manufactured

with finer pitches even at high frequencies. RCA enables a significant reduction of channels compared

to matrix arrays (N+N rather than N², N being the number of elements along one side of the array),

which remains compatible with standard 256-channel preclinical ultrasound scanners. The versatility

of RCA probes in preclinical neuroimaging has been demonstrated across various modalities,

including Doppler imaging[23], functional ultrasound imaging (fusi)[24] and non-linear contrast imaging

of gas vesicles[25]. Their potential has been explored in proof-of-concept studies of 3D ultrasound

localization microscopy (ULM)[26,27] including recently 3D imaging of the mouse brain by Wu and

colleagues[28]as and the rat brain by Sun et al[29]. While RC-ULM shows clear promise for three-

dimensional ULM imaging in rodent brains, further advancements in acquisition protocols,

reconstruction methods, and analytical frameworks are still needed toward achieving a high-

throughput, benchside microvascular flow quantification technique suitable for in vivo preclinical

studies.

Our study explores transcranial 3D Ultrasound Localization Microscopy (ULM) in mouse models

using a dedicated Row-Column Array (RCA) probe through the combination of optimized acquisition

sequences, advanced reconstruction algorithms, and automated flow-directed graph-based analysis of

vessels. We first demonstrate the capability of this approach to reconstruct the 3D cerebral

microvasculature through the skull with high sensitivity and spatial resolution, including the local

estimation of blood velocities, the 3D vessel reconstruction and its registration to the Allen Brain

Atlas. We further show that these data can enable the construction of a flow-directed vascular graph

which locally encodes flow direction and can then be used for labeling artery-like and vein-like

vascular segments based on local branching and flow gradient parameters. By combining directed

graph with local vessel radii estimation, we further extend the analysis to the automated quantification

of dynamic features (cerebral blood flow, microbubble velocity) and vessel structural characteristics

(segment length, tortuosity) across both distinct vascular compartments and anatomical regions across

the brain.

Material and Methods

Ethics

The institutional and regional committees approved the study for animal care (Comité d'éthique pour

l'expérimentation animale \#59 - "Paris Centre et Sud," Protocole \# 2017-23) and adhered to the

ARRIVE recommendations. All animals received humane care in accordance with the European

Union Directive of 2010 (2010/63/EU).

Mice were naive, seven weeks old at the time of the experiment, and randomly placed in the cage;

none were excluded. The individual animal is the experimental unit in this investigation.

Animal preparation

Five male C57Bl/6 mice (Janvier Labs; Le Genest St Isle, France) were utilized in this study. The mice were housed under controlled conditions (22 1 ° C, 60% relative humidity, 12/12h light/dark cycle, ad libitum food and drink) for a week prior to the start of the experiment. The mice were anesthetized with an initial administration of ketamine (Imalgene, 100 mg/kg) and Xylazine (Rompun) intramuscularly to render them unconscious, and anesthesia was maintained with a 1.5% isoflurane supply and the animal physiology was monitored. The animals were then placed in a stereotaxic frame. The skin and the skull were kept intact. The eyes of the mouse were protected using an ointment (Ocry-gel, TVM, UK). Body temperature was controlled with a rectal probe connected to a heating pad set at 37°C. Respiration and heart rate were monitored using a PowerLab data acquisition system with the LabChart software (ADInstruments, USA). Additional IP ketamine/xylazine doses (25 mg/kg and 2.5 mg/kg, respectively) were infused intermittently (every 90 to 120 min), as deemed necessary based on changes in physiological parameters. ULM was performed using Sonovue bolus injections in the tail vein.

RC-ULM acquisition

Row-Column Array, sequences and scanner

A piezoelectric 15 MHz row-column array (Iconeus, Paris, France) consisting of 80 rows and 80 columns with a 0.110-millimeter pitch was used to perform the 3D power Doppler imaging and 3D Ultrasound Localization Microscopy acquisitions. The array was connected to a 256-channel functional ultrasound scanner (Iconeus One, Iconeus, Paris, France) driving the probe at 12.5 MHz instead of 15 MHz to limit the ultrasonic wave aberrations induced during the propagation through the skull bone.

For ultrafast power Doppler, the imaging sequence was defined by transmitting 20 plane waves along the rows followed by 20 plane waves with the columns (angular range of 15°). A four-half-cycle pulse with a duty cycle of 67% and a tension of 25 V was used with a volume rate of 500 Hz. Each block of 200 images (0.4 s) was acquired at 1 Hz to limit the probe heating.

The probe was positioned perpendicular to the animal head's surface, and a four-axis motor module

(Iconeus, Paris, France) was used to ensure proper positioning of the probe using a real-time 3-view

Doppler imaging mode.

The ultrafast RC-ULM sequence was defined by transmitting 14 plane waves along the rows followed

by 14 plane waves along the columns at a PRF of 28 kHz, yielding a high volume rate of 1000Hz

(angular range, 10°). A two-half-cycle pulse was sent at a frequency of 12.5 MHz and a tension of 20

V. The ultrafast sequence comprised 500 volumetric images acquired in 500 milliseconds, repeated

400 times for a total accumulation time of 200 seconds. The acquisition time was 6 minutes and 40

seconds, with 0.5-second intervals between each block of 500 ultrafast volumes.

Beamforming and XDoppler reconstruction

RCA beamforming was implemented on two GPUs (A6000 boards, Nvidia, USA) on the scanner and

optimized to reach real-time 3D beamforming of the ultrafast volumes including GPU-based clutter

filtering. This allowed to perform both the live power Doppler volumetric imaging with a 3-view

rendering for the positioning but also perform the online reconstruction of the microbubble volumes

using the XDoppler scheme[30] down to the prelocalization of microbubbles.

More specifically, the volume RC (transmission along Rows and reception along Columns) was

beamformed using a conventional Delay-And-Sum beamformer optimized for the dual GPUs. The

beamformer grid was set to ($\lambda \times \lambda \times \lambda/2$) and the different angles transmitted along the row were

coherently compounded. A SVD clutter filter was then applied on this stack with an ensemble length

of 500 frames. The same processing was applied to the CR volume (transmission along Columns and

reception along Rows). The filtered RC and CR stacks were then cross correlated with a window size

of 5 frames for ensemble length using the XDoppler scheme. This correlation approach allows to

reduce the signal from the lobes, compared to a synthetic summation approach like orthogonal plane

wave compounding (OPW)[30].

Within each ultrafast block, MB were prelocalized in real time using a local maximum algorithm and

a neighborhood of (5×5×5) voxels around each maximum was extracted and stored on disk for further

processing and refined localization. All those steps were implemented on the GPU in Cuda and

running in real time.

3D ULM processing

The localizations were then refined offline using a 3D interpolation with a gaussian kernel of each

neighborhood volumes followed by a 3D gaussian filter. The sub voxel localization was performed

using a 3D paraboloid fit, the resulting position table was stored and a tracking algorithm was used to

pair microbubbles from one frame to another within given physical constraints on pairing distance and

velocity ranges using a Hungarian linker[31]. The tracks were subsequently rasterized to generate

microbubble (MB) count volumes and MB velocity component volumes (10 µm isotropic).

For visual distinction between cortical arterioles and venules, we also reconstructed a signed MB

velocity map in which we flipped the sign of velocity norm for every voxel whose axial velocity was

negative.

Comparison with Multiplexed Fully Populated Matrix arrays (MUX-FPM)

We compared the RC-ULM scheme with a MUX-FPM, centered at 15 MHz, on the same animal.

After the RC-ULM acquisition, a pause of 30 min allowed the mouse to recover and to clear the

remaining MB from the circulation. During this time, the RCA probe was replaced with the MUX-

FPM array. A second imaging session of the same acquisition duration (240 blocks of 0.5 each) was

then performed using the same ultrasound frequency and the same voltage. The imaging sequence was

adapted from the work of Chavignon and colleagues [20] where 10 transmit/receive combinations were

repeated per tilted plane wave. The PRF was fixed at 12 kHz and 6 tilted plane waves were

transmitted using the following angular sequence: (-5°, 0°); (0°; 0°); (0°, 5°); (0°, -5°); (0°, 0°); (0°,

5°). The resulting framerate of 200 Hz allowed the accumulation of 100 compounded frames per 0.5s

block. The SVD cutoff for the clutter filter was set to 10/100 after optimization. After localization and

tracking, we generated volumetric density and velocity maps (resolution 10 µm). The resulting ULM

volume was realigned to the RCA volume automatically using the transformation found by intensity

based registration between the two corresponding power Doppler volumes. The MEX-FPM ULM

density map was resampled in the RCA space for side-by-side visual comparisons (see supplementary

figure 6).

Spatial resolution assessment

To evaluate the spatial resolution of the proposed vascular imaging modality, we first measured the

Fourier Shell Correlation (FSC) between two MB count volumes, obtained by rasterizing two

independent sub-groups of MB tracks split from the same acquisition. Another relevant assessment of

the spatial resolution was then performed by evaluating the capacity to separate the MB velocity

profiles from small and close cortical vessels in the mouse brain. The velocity profile selection was

performed in a region of interest containing two penetrating arterioles using Amira software. Velocity

profiles were extracted on three different axial slices (thickness 20 µm), taken at three different depths

to intersect different sub-branches of the same penetrating arteriole. Across each profile, the MB

velocity was extracted using the Line Probe function of Amira Avizo, using a sampling step of 5

microns.

Microvascular graph reconstruction and analysis

Graph construction from ULM maps

As a first step, the MB density volumes were filtered using a hessian-like 3D spatial filter to enhance

vessels (3D Jerman enhancement filter[32]). The resulting volume was then thresholded and

skeletonized to extract the vessel centerline coordinates. Radius estimation was performed on these

centerline coordinates using the VesselVio[33] toolbox. VesselVio allowed to build a graph structure

representing the vascular tree, described as a list of edges (describing a vascular segment) connecting

vertices with a connectivity of 2 between branching nodes (vertices with a connectivity of at least 3).

Each edge was hence characterized by a list of coordinates associated to these vertices. Vascular

segment morphological features such as radius, length, surface, volume and tortuosity features such as

MB density and 3D MB velocity vector were extracted and stored in the graph data. The radius and

centerline MB velocity estimations were used to further estimate the flowrate in each segment of the

vascular graph, as proposed in our previous work[34] which was also stored for each vascular segment

in the graph data.

Graph orientation from local blood flow direction

To orient the graph of vascular segments, we took advantage of the locally measured flow direction

and the graph geometry. The orientation of each edge followed the direction of the flow, meaning that

the sign of the scalar product between the average MB velocity vector field and the orientation vector

of each segment was always positive. By running the Breadth First Search (BFS) algorithm as

implemented in the *igraph* library in Python on this directed graph object, perfused or drained

territories could be easily revealed from any seed segments. A graph traversal algorithm was also

implemented in order to reveal the tree structure of any cluster of connected components and store the

branching level of each segment in this tree.

Artery-like and vein-like labeling from graph analysis

Then, vascular segments were classified in either artery-like or vein-like groups using a multistep

multiscale approach. The first step, more global, consisted in identifying seed segments with a

flowrate higher than 3 µL/min, supposed to be potential upstream initiators of perfused territories or

downstream attractors of drained territories. We could then trace the path of blood flow circulation

and identify the descendants (perfused segments) or ancestors (feeding segments) of each seed

segment and estimate the flowrate gradient along the vascular path. Segments associated with a

negative flowrate gradient along the vascular path were labelled as artery-like segments, whereas

segments associated with a positive flowrate gradient were labelled as vein-like segments. The second

step, more local, allowed us to confirm and to complement the results of the first step, and consisted

in estimating for each node the difference between its outward connectivity (number of descendants)

and its inward connectivity (number of ancestors). The nodes with a diverging behavior were detected

with a positive branching difference, and their connected vascular segments were classified as artery-

like. Alternatively, the nodes with a merging behavior detected with a negative branching difference,

and their connected vascular segments classified as vein-like.

Vascular compartments labeling from vascular type and local diameters

To further classify the different parts of the directed vascular tree, we separated the different branches based on their diameters in three classes: large, medium and small segments. Diameter thresholds ($\emptyset\_T1$, $\emptyset\_T2$) were chosen to obtain a clear distinction of the three classes of segments (S class: ($\emptyset<\emptyset\_T1$) - M class: ($\emptyset\_T1<\emptyset<\emptyset\_T2$) – L class: ($\emptyset>\emptyset\_T2$), with $\emptyset\_T1 = 50$ μm and $\emptyset\_T2 = 70$ μm. These thresholds were chosen to group together penetrating arterioles and venules in the S class, arteries and veins including pial vessels in the M class, and major arteries and veins such as arteries from the polygon of Willis or veins from the different sinuses in the L class.

Template registration and Allen-based region labeling

The geometric transformation aligning the brain vascular volume to the Allen mouse brain atlas[35] was automatically estimated using the Iconeus Brain Positioning System, as described by Nouhoum et al.[36], and implemented in the IcoStudio software (Iconeus, Paris, France). This process involved aligning a power Doppler volume - extracted from the first 20 seconds of the continuous ULM acquisition - to a mouse brain power Doppler angiography template which was already pre-aligned with the Allen Brain atlas[35]. Once the optimal affine transformation was identified, it could be used for ULM quantification in any anatomical region from the Allen atlas.

Vascular segments were then assigned to different anatomical regions of interest based on the template registration results. We chose 11 regions in the study (see figure 6.i.). For each chosen region, we tested if the coordinates of the graph centerline points were inside or outside the closed mesh of said region using *intriangulation* [37]. This further allowed us to classify the different vascular segments by region in addition to their diameter and vascular type.

Vascular metrics extraction by class

Vascular segments were classified according to three previously defined criteria: vascular type (artery-like or vein-like), anatomical regions (based on the 11 main Allen atlas areas), and vessel diameter (small, medium, or large), resulting in sixty-six distinct classes. For each class, we extracted in vivo metrics encompassing both morphology- and flow-related parameters. The morphology-related metrics[38] included segment length, segment tortuosity, and the cerebral blood volume (CBV) fraction, while the flow-related metrics comprised microbubble (MB) velocity, flow rate, and cerebral blood flow (CBF).

The CBV fraction in each anatomical region was defined as the ratio of vascular volume to the total volume of the corresponding region. CBF was defined as the total flowrate delivered to the tissue per unit volume within an anatomical region of interest (μL/min/mm³).

Data visualization

The 3D renderings of the MB density and signed velocity volumes (Figure 3 and Figure 4) were performed using the Amira software (Amira v.6.0.1 software, Visualization Sciences Group, USA). The 2D slices represented in the Figure 3 were reconstructed with a prototype of IcoStudio software featuring 3D ULM tracks visualization (Iconeus, Paris, France). The 3D vascular graph renderings were rendered using PyVista visualization toolkit (Figure 5) or using Matlab (Figure 6). Flowrate for artery-like subgraph was rendered in red whereas it was rendered in blue for vein-like subgraphs (Figure 5 and Figure 6).

Results

3D RC-ULM imaging

The proposed RC-ULM approach allows for a precise localization and tracking of microbubbles, enabling the generation of detailed microbubble density and velocity volumes that are displayed in 3D and in 2D slices in figure 2. The MB density revealed a complex architecture of the vascular network in the mouse brain. Important arteries and veins were identified visually and are indicated in the

Figure 2. Thanks to the high temporal resolution of the sequence that reaches 1000 volumes per

second, we were able to reconstruct the arteries of the circle of Willis (CoW) and high velocity

arteries such as the PCA with MB velocities reaching 10 cm/s, as displayed in the axial slices of the

Figure 2.ii.a.. We also measured lower flowrates in small penetrating arterioles originating from the

MCA and ACA branching territories, with MB velocities ranging between 1 and 10 mm/s (Figure

2.ii.b.). The binary vesselness mask and the corresponding skeleton are represented in supplementary

Figures 1 and 2.

We also compared those results to the matrix-array MUX-based acquisition. Results shown in

supplementary Figure 4 illustrate the gain in sensitivity between the two approaches on the same

mouse with the same injection protocol.

Spatial resolution and velocity profiles assessment

The spatial resolution of the RC-ULM volumes was first estimated at 20,26 µm, using the Fourier

Shell Correlation and a 0.5 resolution cutoff (Figure 4.i). To illustrate the ability to measure the

flowrate in small cortical vessels, we also selected one region of interest (ROI) around one arteriole

and extracted the velocity profile at three different depths, chosen at three different levels of

branching. The region of interest and the corresponding velocity profiles are shown in Figure 4.iii-iv.

These velocity profiles demonstrated the ability to estimate the velocity in arterial subbranches,

separated from less than half a wavelength. The FWHM and velocity amplitude peaks for each vessel

crossed by each profile are reported in the legend. The minimum FWHM among the three examined

profiles was 22.9 µm.

Flow-directed graph analysis and Artery-like and Vein-like labeling

To illustrate and investigate our flow-directed graph analysis and labeling method, we first focused on

two vascular territories of interest. By visually comparing our data with other microvascular volumes

from the ex-vivo mouse brain vascular imaging literature[39–41], we selected a seed segment just before

the bifurcation node of the Medial Cerebral Artery (MCA) into three sub-branches, and another seed

segment near to the confluence node of the two Galeno Veins (GV). These two seed segments were

used to initiate the BFS algorithm, that revealed 353 descendants downstream of the MCA seed

segment, delineating its perfused vascular territory, and 53 ancestor segments upstream to the GV

seed segment, defining its drained vascular territory. Each resulting subgraph is presented in two

different representations in the Figure 5.i-ii, the first one displaying the branching level of each

vascular segment with respect to the seed position, and the second one showing the corresponding

segment flowrate. While the flowrate decreases in the case of the MCA sub-graph, it increases in the

case of the GV sub-graph, and the same tendency is also verified in the plots representing the

evolution of the average flowrate by branching level, with an average flowrate gradient of -1.35

µL/min/mm in the MCA sub-graph case, and an average flowrate of +0.22 µL/min/mm in the MCA

sub-graph case. Moreover, the average branching difference was found to be constantly positive in the

arterial subgraph (MCA), and negative in the venous subgraph (GV), reflecting their respective

diverging and merging behaviors along the vascular path. Using these combined metrics, 90% of the

graph segments were labeled, with 5% of these labels automatically corrected to prevent biologically

implausible vein-to-artery transitions. The remaining 10 % were discarded. The resulting artery-like

and vein-like segments are represented in red and blue colors in the Figure 5.iii.

To validate our labelling method, we extracted a region of interest in the cortex where arterioles and

venules can be recognized based on the axial velocity sign and verified that our branching

classification results matched the axial velocity classification. A matching score of 85% between

those two resulting graphs was obtained (Figure 5.iv). The comparisons in the MCA and GV regions

are presented in supplementary Figure 3.

Automatic extraction of morphometric and flowrate features by vascular compartments

Graph and label rendering

Labelization according to the three different criteria (branching, atlas and diameter) can be visualized directly on the 3D graph renderings (Figure 6.i) whereas morphometric and flowrate features for each vascular class are presented using pie plots, bar plots and split violin plots (Figure 6.ii).

Regional flow rate differences

Total CBF was measured at 0.53 mL/min/mm^3. Although MB velocity and flowrate distributions (figure 6.ii.d-e) were most importantly correlated to vessel diameter, we still found important regional differences. Higher flow values were detected on the thalamus, hypothalamus, midbrain and olfactory areas. In these regions, flowrate in the artery-like vessels of the L class exceeded that of the corresponding vein-like vessels, which remains consistent with the presence of major feeding arteries passing through the areas from the CoW. CBF was more homogenous across brain regions for the S class (standard deviation (SD) = 2.8 µL/min/mm³), compared to the M (SD = 4.4 µL/min/mm³) and L (SD = 7.5 µL/min/mm³) classes (figure 6.ii.a) indicating fewer regional variations for the smaller vessels. Moreover, in those smaller vessels, flowrate and MB velocity were more balanced between arterial- and vein-like segments, leading to a local perfusion role.

Volumes of different vascular compartments

The mean CBV fraction was 2.04% (figure 6.ii.b). The S and M classes contributed most to this volume fraction (mean CBV fraction of 0.70 % and 0.85%, respectively) compared to the L class (0.47 %). Finally, artery-like vessels were detected more often than vein-like vessels (mean CBV fraction of 1.41% and 0.61%, respectively). Small and medium vessels also exhibited slightly higher tortuosity than larger ones (figure 6.ii.c).

Discussion

The cerebral microvasculature is central to both neurophysiology and pathology, yet in vivo characterization remains a formidable technical challenge even in preclinical research. Ultrasound Localization Microscopy (ULM) has recently emerged as a promising microvascular imaging technique, offering the potential to non-invasively map microvascular networks with super-resolution precision including local blood velocity measurements. Recent advancements have demonstrated the feasibility of 3D ULM in preclinical models, leveraging matrix array technologies[14,16,19] and more recently multiplexed matrix arrays[20,42] and Row-Column (RC) arrays[28,29] to reduce channel count requirements. While those works proved 3D ULM in vivo feasibility, including in rodent brains, 3D ULM remains challenging in practice due to the massive raw data-rate throughput from the arrays, intensive reconstruction processing and the complex analysis needed to make sense of the 3D ULM data. Building on these recent advances, we thus developed an optimized RCA-based ULM framework, compatible with 256-channel scanners, that enable high quality imaging, full microvascular graph reconstruction, flow-direction labeling, and region-specific quantification. We demonstrated high-resolution, high-sensitivity imaging of cerebral vessels down to 20 micrometers and the reconstruction of a flow-directed microvascular graph from RCA-based 3D ULM to extract quantitative morphological and hemodynamic metrics across anatomical regions in the mouse brain.

Our approach is well-adapted to image diverse range of vessel sizes (20 to 200 µm) and flow velocities (about 2 to 100 mm/s) which remain limited with other 3D ULM approaches such as multiplexed matrix arrays due to the lower volume rate.

Although classified as 'small', the smallest vessels in this dataset have diameters above the typical capillary range (~5–8 µm) and thus likely correspond to precapillary arterioles and postcapillary venules rather than true capillaries.

We registered the reconstructed vessel volumes to the Allen Brain atlas using a preregistered mouse brain Doppler template. The high-contrast of the 3D RC-ULM data further allows for automatic graph

reconstruction using standard tools developed for optical imaging[33]. By incorporating blood velocity

direction in the vascular graph, we could orientate the reconstructed graph for each segment. We

could then separate artery-like and vein-like subgraphs by assessing the branching or merging

behavior along the vascular path, thanks to the estimations of the flowrate gradient and the difference

between inward and outward connectivity for each node. This classification was compared to known

vessels in the brain and particularly in the cortex, where the direction of vertical venules and arterioles

is well known from the axial velocity direction.

Together with the classification in three diameters and in template-derived brain regions, this

approach allows for automatic detailed anatomical and quantitative analysis of the microvascular

compartments' features at different scales. Namely, we could effectively quantify essential

microvascular parameters such as flow velocity, vessel diameter, vessel tortuosity, CBV fraction and

flow rate for those vascular compartments and regions.

We further hypothesize that adding such dynamic flow information into directed microvascular

graphs will further enable researchers to introduce robust and scalable tools for brain microvascular

studies beyond raw ULM data volumes.

The preclinical applications of RC-ULM extend from neurovascular research to drug discovery. It

could enable high-throughput microvascular phenotyping in genetic mouse models, quantification of

drug-induced vascular changes, and integration of local perfusion data with molecular profiles. RC-

ULM could also provide localized microvascular assessment in acute neurovascular disorders (stroke,

aneurysms), neurodegenerative diseases (Alzheimer's, Parkinson's), tumor angiogenesis, and

cerebrovascular small vessel disease.

The translational potential of this technology for clinical investigations of microcirculation is also

significant, especially in perioperative settings where a skull piece is removed during the surgery.

Those applications include assessing tumoral vascular network in the brain to help surgeons better

define the tumor margins and guide the resection or tools in deep and small brain regions without

hitting larger vessels.

A key benefit of the RCA technology is its capacity to deliver high-quality 3D ULM using fewer

channels (less than 256) as opposed to traditional matrix arrays that require specialized research

scanner with more than a thousand electronic channels. This simplification not only reduces the

complexity but also increases the technology's accessibility and adaptability for broader applications.

Compared to a multiplexed array, also the RCA does not require any additional electronics and can be

plugged directly to an ultrafast scanner with a smaller cable. Furthermore, using a multiplexor

requires to transmit and to receive the channels in different steps which leads to a strong diminution of

the volume rate by a factor up to 16 when using a 1:4 multiplexor. In our own tests, RC-ULM

consistently outperformed the multiplexed matrix array of similar frequency in vessel imaging quality,

especially in high flowrate regions like the CoW (Supplementary figure 4).

There remain limitations to the RC-ULM approach. For one, while many steps from beamforming to

prelocalization could be implemented in real time on the scanner GPUs in this study, performing 3D

ULM including tracking and motion correction remains computationally intensive. To address these

challenges, we are exploring the development of deep learning algorithms and online GPU-based

tracking algorithms. Another limitation is the restricted "en face" field of view of the RCA due to

intrinsic inability to steer the row and column beams on the same side. To overcome this limitation

especially for clinical applications such as cardiac imaging, researchers have proposed the use of

diverging lens[43-45] or curved RCA arrays[46]. Finally, another limitation of the RC-ULM approach is its

reduced flexibility in correcting skull-induced aberrations. While individual row and column

corrections is possible along each dimension, full 2D adaptive correction as with dense matrix arrays

remain impossible. In our case, we observed reduced imaging quality below the main skull sutures, it

remains to be understood if this is due to wavefront distortions from said structures and if they could

be further corrected to improve imaging in those areas.

Overall, RC-ULM with directed graph reconstruction demonstrates robust capability for high-

resolution imaging and whole-brain microvascular quantification in mice compatible with preclinical

functional ultrasound systems. This framework could enable large-scale acquisition and analysis of

3D ULM, facilitating comprehensive microvascular flow quantification for preclinical neurovascular

research and complementing existing imaging techniques.

Acknowledgements

455   This study was supported by the MICROVASC project funded by the European Innovation Council

456   and SMEs Executive Agency under grant agreement No 101070917. We acknowledge the ART

457   (Technological Research Accelerator) biomedical ultrasound program of INSERM.

Author contributions

M.P, T.D, B.O, and M.T conceived the study. A.B, O.D, T.D developed sequence acquisitions, O.D,

A.B, J.F acquired data. A.B, T.D, A.D., M.T and M.P. performed data processing. A.B, T.D, M.P, J.F,

interpreted the results. T.D, A.B, M.P and J.F. wrote the first draft of the manuscript with substantial

contribution from M.T and B.O

All authors edited and approved the final version of the manuscript. Underlying data have been

verified by M.P and T.D

Disclosure

466   M.T, M.P, B.O and T.D are co-founders and stockholders of Iconeus and have received fundings from

467   Iconeus for research on functional ultrasound imaging. B.O, J.F and A.B are employees of Iconeus.

Data sharing statement

469   MB volumes and microvascular graphs are available at a data repository on zenodo. Other data that

470   support the findings of this study are available from the corresponding author upon reasonable

471   request. Researchers wishing to obtain the raw data must contact the Office of Research Contracts at

472   INSERM to initiate a discussion on the proposed data transfer or use.

References

1. Schneider JA, Bennett DA. Where Vascular Meets Neurodegenerative Disease. *Stroke*; 41. Epub ahead of print October 2010. DOI: 10.1161/STROKEAHA.110.598326.

2. Torre JC. Alzheimer Disease as a Vascular Disorder: Nosological Evidence. *Stroke* 2002; 33: 1152–1162.

3. Vivien D. New Opportunities for Diagnosis and Prognosis of Stroke: The Benefits of Across Border Approaches. *Hämostaseologie* 2021; 41: 022–024.

4. Hope MD, Purcell DD, Hope TA, et al. Complete Intracranial Arterial and Venous Blood Flow Evaluation with 4D Flow MR Imaging. *Am J Neuroradiol* 2009; 30: 362–366.

5. Howles GP, Ghaghada KB, Qi Y, et al. High-resolution magnetic resonance angiography in the mouse using a nanoparticle blood-pool contrast agent: MRA of the Mouse Brain With Blood-Pool Contrast. *Magn Reson Med* 2009; 62: 1447–1456.

6. Choi WJ. Optical coherence tomography angiography in preclinical neuroimaging. *Biomed Eng Lett* 2019; 9: 311–325.

7. Kirst C, Skriabine S, Vieites-Prado A, et al. Mapping the Fine-Scale Organization and Plasticity of the Brain Vasculature. *Cell* 2020; 180: 780-795.e25.

8. Miettinen A, Zippo AG, Patera A, et al. *Micrometer-resolution reconstruction and analysis of whole mouse brain vasculature by synchrotron-based phase-contrast tomographic microscopy*. Preprint, Neuroscience. Epub ahead of print 16 March 2021. DOI: 10.1101/2021.03.16.435616.

9. Quintana DD, Lewis SE, Anantula Y, et al. The cerebral angiome: High resolution MicroCT imaging of the whole brain cerebrovasculature in female and male mice. *NeuroImage* 2019; 202: 116109.

10. Starosolski Z, Villamizar CA, Rendon D, et al. Ultra High-Resolution In vivo Computed Tomography Imaging of Mouse Cerebrovasculature Using a Long Circulating Blood Pool Contrast Agent. *Sci Rep* 2015; 5: 10178.

11. Errico C, Pierre J, Pezet S, et al. Ultrafast ultrasound localization microscopy for deep super-resolution vascular imaging. *Nature* 2015; 527: 499–502.

12. Demené C, Robin J, Dizeux A, et al. Transcranial ultrafast ultrasound localization microscopy of brain vasculature in patients. *Nat Biomed Eng* 2021; 5: 219–228.

13. Renaudin N, Demené C, Dizeux A, et al. Functional ultrasound localization microscopy reveals brain-wide neurovascular activity on a microscopic scale. *Nat Methods* 2022; 19: 1004–1012.

14. Demeulenaere O, Bertolo A, Pezet S, et al. In vivo whole brain microvascular imaging in mice using transcranial 3D Ultrasound Localization Microscopy. *EBioMedicine* 2022; 79: 103995.

15. Heiles B, Correia M, Hingot V, et al. Ultrafast 3D Ultrasound Localization Microscopy Using a 32x32 Matrix Array. *IEEE Trans Med Imaging* 2019; 38: 2005–2015.

16. Heiles B, Chavignon A, Bergel A, et al. Volumetric ultrasound localization microscopy of the whole rat brain microvasculature. *IEEE Open J Ultrason Ferroelectr Freq Control* 2022; 2: 261–282.

17. Chavignon A, Hingot V, Orset C, et al. 3D transcranial ultrasound localization microscopy for
discrimination between ischemic and hemorrhagic stroke in early phase. *Sci Rep* 2022; 12:
14607.

18. McCall JR, Santibanez F, Belgharbi H, et al. Non-invasive transcranial volumetric ultrasound
localization microscopy of the rat brain with continuous, high volume-rate acquisition.
Theranostics 2023; 13: 1235–1246.

19. McCall JR, Chavignon A, Couture O, et al. Element Position Calibration for Matrix Array
Transducers with Multiple Disjoint Piezoelectric Panels. *Ultrason Imaging* 2024;
01617346241227900.

20. Chavignon A, Heiles B, Hingot V, et al. 3D Transcranial Ultrasound Localization Microscopy
in the Rat Brain with a Multiplexed Matrix Probe. *IEEE Trans Biomed Eng* 2021; 69: 1–1.

21. Guo X, Ta D, Xu K. Frame rate effects and their compensation on super-resolution microvessel
imaging using ultrasound localization microscopy. *Ultrasonics* 2023; 132: 107009.

22. Bertolo A, Ferrier J, Cazzanelli S, et al. High sensitivity mapping of brain-wide functional
networks in awake mice using simultaneous multi-slice fUS imaging. *Imaging Neurosci* 2023;
1: 1–18.

23. Sauvage J, Flesch M, Ferin G, et al. A large aperture row column addressed probe for in vivo
4D ultrafast doppler ultrasound imaging. *Phys Med Biol* 2018; 63: 215012.

24. Sauvage J, Porée J, Rabut C, et al. 4D Functional imaging of the rat brain using a large aperture
row-column array. *IEEE Trans Med Imaging* 2019; 39: 1884–1893.

25. Heiles B, Nelissen F, Terwiel D, et al. Nonlinear sound-sheet microscopy: imaging opaque
organs at the capillary and cellular scale. 2024; 2024.07.31.605825.

26. Jensen JA, Ommen ML, Oygard SH, et al. Three-Dimensional Super-Resolution Imaging Using
a Row–Column Array. *IEEE Trans Ultrason Ferroelectr Freq Control* 2020; 67: 538–546.

27. Jensen JA, Schou M, Ommen ML, et al. 3-D Super Resolution Imaging using a 62+62 Elements
Row-Column Array. In: *2019 IEEE International Ultrasonics Symposium (IUS)*, pp. 1838–
1841.

28. Wu A, Porée J, Ramos-Palacios G, et al. 3D transcranial Dynamic Ultrasound Localization
Microscopy in the mouse brain using a Row-Column Array. *IEEE Trans Biomed Eng* 2025; 1–
12.

29. Sun Q, Hou S, He R, et al. Single-Crystal Row–Column Array-Based Rat Brain 3-D Ultrasound
Localization Microscopy. *IEEE Trans Ultrason Ferroelectr Freq Control* 2025; 72: 698–708.

30. Bertolo A, Sauvage J, Tanter M, et al. XDoppler: Cross-Correlation of Orthogonal Apertures for
3D Blood Flow Imaging. *IEEE Trans Med Imaging* 2021; 40: 3358–3368.

31. Tinevez J-Y. tinevez/simpletracker, https://github.com/tinevez/simpletracker (2024, accessed 23
July 2024).

32. timjerman.   timjerman/JermanEnhancementFilter,
https://github.com/timjerman/JermanEnhancementFilter (2024, accessed 23 July 2024).

33. Bumgarner JR, Nelson RJ. Open-source analysis and visualization of segmented vasculature
datasets with VesselVio. *Cell Rep Methods*; 2. Epub ahead of print 25 April 2022. DOI:
10.1016/j.crmeth.2022.100189.

34. Demeulenaere O, Bertolo A, Pezet S, et al. In vivo whole brain microvascular imaging in mice
using transcranial 3D Ultrasound Localization Microscopy. *eBioMedicine* 2022; 79: 103995.

35. Wang Q, Ding S-L, Li Y, et al. The Allen Mouse Brain Common Coordinate Framework: A 3D
Reference Atlas. *Cell* 2020; S0092867420304025.

36. Nouhoum M, Ferrier J, Osmanski B-F, et al. A functional ultrasound brain GPS for automatic
vascular-based neuronavigation. *Sci Rep* 2021; 11: 15197.

37. Korsawe J. intriangulation, https://www.mathworks.com/matlabcentral/fileexchange/43381-
intriangulation-vertices-faces-testp-heavytest.

38. Nicolas N, Dinet V, Roux E. 3D imaging and morphometric descriptors of vascular networks on
optically cleared organs. *iScience* 2023; 26: 108007.

39. Dorr A, Sled JG, Kabani N. Three-dimensional cerebral vasculature of the CBA mouse brain: A
magnetic resonance imaging and micro computed tomography study. *NeuroImage* 2007; 35:
1409–1423.

40. Ghanavati S, Lerch JP, Sled JG. Automatic anatomical labeling of the complete cerebral
vasculature in mouse models. *NeuroImage* 2014; 95: 117–128.

41. Xiong B, Li A, Lou Y, et al. Precise Cerebral Vascular Atlas in Stereotaxic Coordinates of
Whole Mouse Brain. *Front Neuroanat*; 11. Epub ahead of print 19 December 2017. DOI:
10.3389/fnana.2017.00128.

42. Lowerison MR, Chandra Sekaran NV, Dong Z, et al. *Super-resolution ultrasound imaging of
cerebrovascular impairment in a mouse model of Alzheimer's disease*. Preprint, Neuroscience.
Epub ahead of print 7 October 2022. DOI: 10.1101/2022.10.05.511008.

43. Audoin M, Salari A, Tomov BG, et al. Diverging Polymer Acoustic Lens Design for High-
Resolution Row–Column Array Ultrasound Transducers. *IEEE Trans Ultrason Ferroelectr Freq
Control* 2024; 71: 202–213.

44. Engholm M, Beers C, Bouzari H, et al. Increasing the field-of-view of row–column-addressed
ultrasound transducers: implementation of a diverging compound lens. *Ultrasonics* 2018; 88:
97–105.

45. Favre H, Pernot M, Tanter M, et al. Transcranial 3D ultrasound localization microscopy using a
large element matrix array with a multi-lens diffracting layer: an in vitro study. *Phys Med Biol*
2023; 68: 075003.

46. Caudoux M, Demeulenaere O, Porée J, et al. Curved toroidal row column addressed transducer
for 3D ultrafast ultrasound imaging. *IEEE Trans Med Imaging* 2024; 1–1.

Titles and legends to figures

Figure 1: 3D ULM acquisition setup with the RCA probe. i. Iconeus One driving the 160 elements RCA probe centered at 13 MHz. ii. Experimental setup for the 3D ULM experiment in the anesthetized mouse brain, with intact skull and skin.

Figure 2: i. Synthetic focusing with the RCA probe geometry. The columns (C) are represented in blue, and the rows (R) are represented in red. To obtain an isotropic PSF, the final volume is reconstructed by combining the backscattered echoes acquired with the rows (after plane wave transmission through the columns) and the columns (after plane wave transmission through the rows) using the XDoppler strategy. ii. The ultrafast BMode movie of the MB is processed by the ULM processing pipeline. After tracking, the MB positions are used to correct the potential spatial drifts before rasterization. A vesselness filter is then applied on the resulting rasterized volume, before skeletonization. iii. Finally, a graph-based processing pipeline is applied to extract morphometric and flowrate parameters at the segment level in different classes of vascular segments. The first level of classification is defined by the anatomical segmentation, the second level of classification is based on diameter estimations: S class: ($\emptyset < \emptyset\_T1$) - M class: ($\emptyset\_T1 < \emptyset < \emptyset\_T2$) − L class: ($\emptyset > \emptyset\_T2$), and the last level of classification is based on the branching (artery-like segments) or merging (vein-like segments) behaviors observed on the directed graph.

Figure 3: 3D ULM visualization i. Sagittal Maximum Intensity Projection (MIP) of the 3D MB density volume (i.) and of the 3D signed MB velocity volume (ii.) are rendered with Amira software. The ULM tracks are rendered in the following 2D slices with IcoStudio software. The axial slices displayed in ii.a. and in ii.b. have a thickness of 2 mm. The scalebars represent 1 mm. The coronal and sagittal slices presented in iii. and in iv. have a slice thickness of 1.5 mm. Abbreviations: CoW : Circle of Willis ; ICA : Internal Cerbral Artery ; PCA : Posterior Cerebral Artery ; MCA : Medial

Cerebral Artery ; Azc : Asygos ; SS : Sagittal Sinus ; TS: Transversal Sinus ;   PSV : Posterior

Superficial Vein ; GV : Galeno Vein.

Figure 4: Spatial resolution assessment: Characterization of the MB velocity on different branching

levels of the same penetrating arteriole. (i) The spatial resolution measured by the Fourier Shell

Correlation is 20,26 µm (ii) Zoom of the MB density volume in the cortex of the mouse brain

(corresponding to the green bounding box in Figure 3.i.). The penetrating arterioles ROI (white box)

of size (4λ x 4λ x 10λ) was selected within the ACA territory. (iii) The selected region contains two

arterioles (ArtA and ArtB). Three planes were placed at z = 325 µm (green), z = 635 µm (red) and z =

895 µm (blue). On each plane, the velocity intensity was extracted along the 1D profiles represented

with the orange lines (ii). For each depth, a cropped view showing the selected plane is represented in

(iii), and the velocity profile is shown in a transparent overlay. The corresponding velocity profile is

plotted next to each cropped volume (iv). At z = 325 µm, the profiles crosses ArtA and ArtB and

reveals a distance of 137 µm between them. At z = 635 µm and z = 895 µm, the velocity profile

crosses new branches of ArtA, which are separated from 53 µm. The FWHM values are: A = 43.5

µm; B = 99 µm; C = 31.6 µm; D = 65 µm ; E = 25 µm; F = 29.1 µm

Figure 5: Mapping perfused and drained vascular territories, and labeling artery-like or vein like

vascular segments using directed graph analysis. (i) The vascular territory perfused by the MCA is

identified by selecting a seed segment (red sphere), upstream of the MCA bifurcation, and extracting

all downstream (descendants) segments. Segments are represented over the MB density volume. Tube

colors encode branching level (relative to the seed) on the left, and flowrate on the middle. The right

plot shows a negative slope of average flowrate across branching levels, alongside a positive

branching difference at the node level (red color code), reflecting the diverging pattern of artery-like

vascular territories. (ii) In contrast, the drained vascular territory of the GV is identified by selecting a

seed near the GV confluence node of the GV and tracing upstream (ancestor) segments. Here,

segments merge at successive level, with increasing flowrate along the vascular path. The positive

gradient is accompanied by a negative branching difference at the node level (blue color code),

highlighting the converging behavior of vein-like vascular territories. (iii) The sign of flowrate

renderings in the whole mouse brain is based on the branching behavior. The artery-like subgraph

(red) is overlayed to the vein-like sub-graph (blue) on the left rendering, and each subgraph is

represented separately on the middle and right renderings. (iv) The left rendering represents the MIP

of the signed MB velocity volume in a small region of interest with penetrating arterioles and venules,

in which a good match is observed between the sign assignment based on the axial MB velocity

(middle) and the sign assignment based on the branching classification (right).

Figure 6: Automatic extraction of morphometric and flowrate parameters in different vascular

compartments. (i) Sixty-six classes were defined using three different levels of classifications:

anatomical (left), radius (middle) and branching (right). (ii) Results of the automatic flowrate and

morphometric extractions. (ii.a) The global CBF repartition over the different anatomical regions is

represented with pie plots, for each radius class (S-M-L). Each anatomical region is represented using

the color-code already used in (i). (ii.b) The CBV fraction in the sixty-six different classes is

represented in a bar plot figure. For each anatomical region, the left bar corresponds to the artery-like

segments while the right bar corresponds to the vein-like segments. In each bar, the three sub-bars

correspond to the radius-based classification (S-M-L), with distinct gray-scale colors. For each

category, the distributions of the vascular segment tortuosity, MB velocity and flowrate are displayed

(ii.c-e). For each anatomical region, three split-viol violin plots are represented, for each radius class.

The left part (red) corresponds to the artery-like segments whereas the right part (blue) corresponds to

the vein-like segments. The mean value of each distribution is indicated by the white dots.

i. Vascular segments perfused by MCA

Seed: Upstream attractor

Branching level (relative to seed)

Flowrate (µL/min)

Branching level (relative to seed)

Flowrate (µL/min)

Average branching difference

ii. Vascular segments drained by GV

Seed: Downstream attractor

Branching level (relative to seed)

Flowrate (µL/min)

Branching level (relative to seed)

Flowrate (µL/min)

Average branching difference

iii. Differentiating artery-like and venous-like sub-graphs

Vein-like   Artery-like

Artery-like

Vein-like

Signed Flowrate (µL/min)

Vein-like   Artery-like

iv. Comparison with axial velocity sign for penetrating arterioles and venules

Pial vessels

arterioles

venules

MB signed velocity

Sign : Axial MB velocity

Signed Flowrate (µL/min)

Sign : Branching behavior

Arterylike

Vein-like

Signed Flowrate (µL/min)

i. Vascular segment classification

Atlas based

- Isocortex
- Olfactory areas
- Hippocampus
- Cortical subplate
- Striatum
- Pallidum
- Thalamus
- Hypothalamus
- Midbrain
- Fibers
- Other

Radius based

S · M · L

Branching based

Signed Flowrate (µL/min)
5.00 Artery-like
-5.00 Vein-like

ii. Automatic extraction of flowrate and morphometric metrics

a. CBF Repartition (Total = 0.53 mL/min·mm³)

S Class
Total CBF: 0.14 mL/min·mm³

M Class
Total CBF: 0.19 mL/min·mm³

L Class
Total CBF: 0.21 mL/min·mm³

b. Average CBV Fraction: 2.04%

S class
M class
L class

c. Vascular segment tortuosity

Artery-like
Vein-like

d. Vascular segment MB velocity

Artery-like
Vein-like

e. Vascular segment MB flowrate

Artery-like
Vein-like

i.

256 échographic channels scanner

8.8 mm

8.8 mm

RCA probe
160 elements (80 + 80)
F0 = 13 MHz

ii.

RCA probe

8.8 mm

Stereotaxic frame

Heating pad

Syringe pump