1
2
3 **Non-invasive longitudinal imaging reveals aging-associated changes in neuroimmune**
4 **cells in adult zebrafish**
5
6 **Running Title: Longitudinal Neuroimaging in Fish**
7
8 Elizabeth M. Haynes[1,2], Corey Steinhauser[2,4], Robert Swader[2], George Petry[2], Tyler K. Ulland[3],
9 Kevin W. Eliceiri[2,4]
10
11   1.   Department of Cell and Regenerative Biology, University of Wisconsin-Madison
12
13 2. Morgridge Institute for Research, Madison WI
14
15 3. Department of Pathology, University of Wisconsin-Madison
16
17 4. Center for Quantitative Cell Imaging, University of Wisconsin-Madison
18
19
20
Corresponding author:

Elizabeth M. Haynes

Department of Cell and Regenerative Biology

1111 Highland Ave

Madison, WI 53705

liz.haynes@wisc.edu

608-262-1199

Keywords: zebrafish, intravital imaging, neurobiology, microglia, multiphoton, aging

Abstract
Zebrafish are a powerful model for imaging studies of development and neurobiology. However, most studies have focused on developing zebrafish due to technical and biological challenges of imaging adult stages. These include increased tissue opacity and illumination depth limitations, and difficulty maintaining life support and anesthesia in a 2-4cm long fish. There are currently limited tools for intravital imaging of the adult zebrafish brain. The ability to image the brain in the same individual repeatedly without physical damage would allow zebrafish to be better utilized to study aging and neurodegenerative disease. We designed and applied a 3D-printable device for non-invasive, repeatable multiphoton neural imaging of genetically non-pigmented adult zebrafish from 2 months to 2 years old. Animals successfully recovered after multi-hour imaging sessions and could be imaged repeatedly over periods of weeks to years. We show the utility of this approach through imaging the neuroimmune system, revealing that microglia in aged zebrafish have enhanced cellular dynamics. This technique could be widely used and beneficial for other cell-scale neuroimaging studies in the adult fish.

Summary Statement: An open-source, 3D printable device that enables non-invasive neuroimaging in adult zebrafish reveals aging-associated changes in neuroimmune cell behavior.

Introduction

As we discover the expansive roles of microglia in the brain, there is increased interest in studying microglial biology during aging and disease progression in relevant animal models. Intravital imaging of dynamic behaviors of microglia is difficult in mammalian models and generally requires surgery or skull thinning. These procedures require a large skill and time investment and can alter the local environment of the brain(Cramer et al., 2021; Kalmbach and Waters, 2012). Small thinned skulled vertebrate teleost fish such as zebrafish and danionella are an excellent alternative model for the study of longitudinal, dynamic neuroimmune biology. In this work we describe a method to perform repeated imaging sessions over periods of weeks to months with minimal physical and behavioral stress on the animal. Several animal health concerns limit the ability to serially image individual animals, including physical trauma from mounting and unmounting, hypoxia from inadequate oxygenation, and the possibility of neurological effects from repeated sedation. To avoid some of these complications we endeavored to design an open-source 3D printable device that allows for quick and reproducible positioning of a fish in an upright position with minimal handling and use of force. Our device's design allows the fish's head to remain accessible to a water dipping objective in an upright orientation and keeps the gills free from any obstruction. Intubation of fish has been previously demonstrated (Barbosa et al., 2016; Castranova et al., 2022; Chow et al., 2020; Dray et al., 2015) to allow sustained oxygenation, sedation, and long-term imaging. Thus, our design is compatible with intubation as well as circulation of additional water to provide adequate water temperature control (in the range of 26-29 °C) within the chamber. Additionally, the design needed to accommodate adult fish across age ranges and length scales to facilitate longitudinal imaging in a continuously growing animal. No currently published teleost imaging device was appropriate for our specific applications (see discussion below), so this novel design contributes a valuable, easily adopted, and adaptable tool for the zebrafish research community.

There have been a handful of published approaches for imaging juvenile and adult Danio species. While zebrafish are still the most prevalent model, *Danionella* are becoming popular due to their small size, optical accessibility, and diverse behaviors including vocalization(Lam,

2022; Schulze et al., 2018; Tatarsky et al., 2022). Limited work has been done showing the
ability to image other species, such as cave fish (Castranova *et al.*, 2022).

Previous approaches have varied greatly in their mounting strategy, achieved depth, and goals
for imaging. All approaches have used some form of intubation to provide for sedation and
oxygenation. Some designs modified existing lab supplies to create a temporary mounting
device (Barbosa et al., 2016, 2015; Chow et al., 2020; Dray et al., 2015). The drawback to these
approaches is a lack of reproducibility both between individual experiments and across different
labs, though this work demonstrated that multiphoton microscopy could be used to image brain
regions of both adult zebrafish and *Danionella.*

In 2022, Castranova et al. created the first reproducible imaging chamber for studying wound
healing in live adult zebrafish using an inverted microscope (Castranova *et al.*, 2022). The study
demonstrated that zebrafish could be imaged while under anesthesia for 20 hours while still
being recoverable.

Lam et.al subsequently demonstrated a reproducible imaging chamber for *D. cerebrum* that is
compatible with both upright and inverted imaging. Unlike previous examples, Lam did not rely
on anesthesia to prevent movement and instead secured the fish in place by encasing it in low
melting point agarose, though intubation is still used to provide oxygenation. Fish could be
imaged up to 3 hours using this method. Conveniently, the smaller size of Danionella and their
lack of a skull roof plate allows for high resolution neural imaging with just a standard confocal.

Development of re-usable, reproducible chambers that maintain animal viability over long period
of time and repeated imaging sessions is the first step in increasing adoption of adult intravital
imaging in fish species. Unfortunately, none of these strategies was highly suited to the needs
of our research. Namely, we wanted a design that was non-invasive, allowed for highly
reproducible upright positioning, and could maintain the health of the fish during both long-term
imaging and serial imaging. We built upon these existing strategies to make an easy-to-use
open-source 3D printable imaging chamber for adult zebrafish neural imaging at the cell-scale
using multiphoton microscopy. Here we demonstrate its utility for non-invasive neuroimaging
over hours to months.

Results

Non-invasive imaging of microglia distribution and dynamics in living zebrafish

To confirm the utility of our designed device (Fig. 1A, B, Sup. Fig. 2) for long-term
neuroimaging, we tested the ability to image cells of the neuroimmune system in the zebrafish
telencephalon and optic tectum (Fig. 1C). These areas are superficially positioned and
accessible to imaging. We imaged transgenic zebrafish expressing a marker of the vasculature
(*kdrl*:mCherry) along with a marker of phagocytic immune cells (*mpeg1*:eGFP) in a pigmentless
Casper background (White *et al.*, 2008). The inclusion of a secondary structural marker such as
the vasculature is helpful to correctly identify the same regions over longitudinal imaging
periods, especially since the zebrafish brain is capable of continuous neuroproliferation and
growth throughout the fish's lifespan. However, autofluorescent or second-harmonic signal in
the bone or other tissues may also be suitable for re-identifying regions across time (Dray *et al.*,
2015). Using two Bruker multiphoton microscopes setups, a Bruker Ultima and Ultima 2P+, we
were able to image at depths of up to 500µM using a Spectra Physics Insight ultrafast laser.
Laser power measured at the imaging plane at generally used settings was 10-13mW. The
depth we measured from the top of the skin to the start of brain tissue varied by age and brain

region, ranging from as little as 25 µM in the telencephalon of a 3-month-old fish to 225µM in
the telencephalon of a 19-month old fish. We found the optic tectum tends to remain closer to
the skull and thus can be easier to image in older animals. We were able to visualize both the
vasculature of the telencephalon and neuroimmune cells through the intact skin and skull of
adult zebrafish (Fig. 1 D, D', Supplemental Movie 1). *Mpeg1*:eGFP positive neuroimmune cells
ranged in morphology from small, rounded vessel-associated cells to larger and more ramified
cells that could be close to vessels or in the parenchyma. The population of neuroimmune cells
that could be imaged at laser powers generally accepted to be non-destructive tended to be
superficial and just below what we presume to be the meninges, generally within the first 25-
50µms of the brain tissue, but bright cells could be identified as deep as 200µm within the brain
tissue. We focused on the larger, deeper, and more ramified cells as being more consistent with
morphology associated with microglia. We achieved sufficient resolution to be able to identify
individual cells of interest to target for imaging at greater magnification (Fig. 1E, E',
arrowheads). Cellular behavior was live imaged for 10-30 minutes for analysis (Fig. 1F, F',
Supplemental Movie 2).

In agreement with previous intravital imaging studies of microglia, we found that ramified
mpeg1:eGFP positive cells in young adult animals are highly dynamic in their protrusive
behaviors (Fig. 2). We recorded no significant translocation of the cell bodies of these cells
during the imaging window. We also found that total cell volume did not change dramatically but
instead was balanced between cycles of protrusion and retraction. This was consistent between
cells in the telencephalon (Fig. 2A) and the optic tectum (Fig. 2B). These results confirm that
adequate resolution for the study of the cell biology and dynamic behavior of neuroimmune cells
can be achieved with non-invasive 2-photon imaging. It is likely that different cell types, such as
other glia and neurons, could also be successfully imaged non-invasively using this method.

Repeated imaging of the same brain region in the same individual over months

Since we are interested in following changes in neuroimmune cell morphology over time and
during disease progression, we wanted to confirm that we could reproducibly image the same
brain region serially in the same individual. We tested if longitudinal imaging could be achieved
non-invasively, without skin removal or skull thinning. We imaged a series of individual animals,
including the same individual from Figures 1 and 2, every two weeks for 6 weeks in both the
telencephalon and optic tectum (Fig. 3). Animals were imaged for approximately 2 hours each
session. Rough estimates of imaging area were obtained through transmitted light visualization,
then adjusted based on imaging of the vasculature and autofluorescence of the skull and
meninges. The vasculature could be used successfully to identify specific regions and reliably
re-identify them over weeks (Fig. 3C). Local neuroimmune cells were more difficult to reliably re-
identify, as morphology and location changed significantly over the scale of the intervals
between imaging sessions.

Observing changes in neuroimmune cell morphology and behavior across the lifespan

Finally, we tested whether much older and larger animals were suitable for non-invasive
intravital imaging. We were able to image animals ranging from 1-2 years old and obtain a
similar degree of detail and ability to capture dynamic cellular behavior (Fig. 4, Supplemental
Movies 3,4). No skull thinning or tissue removal was performed to obtain these results. Imaging
older fish does pose some additional challenges, as we observed the distance between the skull
and the first layers of brain tissue becomes larger. There is also an increase in structures we
believe to be subcutaneous fat pockets that can have a distorting effect. The best imaging
outcomes can be obtained by using brightfield imaging to identify areas with minimal fat pockets

before imaging. We found that both the telencephalon and optic tectum remained accessible for
imaging in these older animals. We were able to image a similar field of view for the
telencephalon in aged adult animals as in younger animals (Fig. 4 A). Specific sub-regions of
the telencephalon were identified (Fig. 4 A', A'') and selected for live imaging for periods of up to
30 minutes (Supplemental Movie 4, 5). Phagocytic neuroimmune cells within the brains of aged
zebrafish displayed increased dynamic behavior and motility compared to that of younger
animals. We were able to observe examples of cell-cell interaction (Fig. 4 B) and retraction of
cellular process (Fig. 4 C) during our imaging window. Unlike in young animals, where little
translocation of cell bodies was observed, tracked cells in aged animals showed increased cell
motility as measured by average velocity (Fig. 4D).

184
Discussion
We have developed and tested a chamber intended for live intravital neuroimaging of adult
zebrafish. Zebrafish are a cost-effective and efficient model organism to conduct longitudinal
non-invasive imaging for the study of neuroimmune cell biology. We have demonstrated that it is
possible to non-invasively image the superficial layers of two regions of the zebrafish brain, the
telencephalon and optic tectum, longitudinally across the lifespan of the zebrafish. We are
particularly interested in non-invasive imaging since we are studying the cell biology of the
neuroimmune system, where creating wounds or environmental disturbances can complicate
interpretation of imaging data. However, in cases where this is not a concern, improved imaging
depth could likely be achieved through previously demonstrated techniques like scale removal
or skull thinning (Barbosa *et al.*, 2016). We have shown that non-invasive imaging is suitable for
imaging eGFP expressing cells up to 200µm deep, however the use of brighter and more red-
shifted fluorophores or 3-photon illumination could further improve depth of imaging.
Additionally, photoconvertible fluorophores may be necessary to accurately identify individual
cells in local populations longitudinally. Further study is needed to compare the behavior and
morphology of *mpeg1*:eGFP positive neuroimmune cells in superficial layers versus deeper
layers. This technique could be expanded for imaging of other neuronal cell types, and the
modular design of the restraint device design means it could be easily modified and redesigned
to accommodate different mounting orientations for imaging different structures.

Finally, we have also shown preliminary evidence that like mammals, neuroimmune cells in
zebrafish undergo detectable changes in morphology and behavior during aging. These
changes in morphology and behavior are similar to observed morphological changes associated
with microglial dysfunction in mammals (Hefendehl *et al.*, 2014). While intriguing, further
research is necessary to identify whether all neuroimmune populations across zebrafish brain
regions show similar signs of aging, or if there are regional differences. Better transgenic lines
with expression more restricted to mark microglia are also necessary to confirm the identity of
specific neuroimmune populations. Since zebrafish are well known for their lifelong
neuroproliferative ability, these preliminary findings pose many compelling questions about how
neuroproliferative regions may influence neuroimmune aging. Though there is abundant work
remaining to be done, we believe this device and these data strengthen and broaden the utility
of the adult zebrafish model as a tool for neuroimaging and neurobiological study.

Materials and Methods
217
Chamber Design and Manufacture
We designed a device to facilitate fast and reproducible mounting of adult zebrafish fish with
minimal restraint. This device can be 3D printed in biocompatible resin, Nylon-12, or produced
in glass-filled nylon (preferable for long-term use). The design features a removable **restraint**

device (Fig. 1 A, B, Sup. Fig. 2, 3) which secures the fish in place via a V-shaped channel and
a small well for low-melting point agarose. The use of the well confines the agarose to only the
tail of the fish and allows the fish to be freed from the chamber with minimal effort. Importantly,
this prevents low melting point agarose from contacting the fish's gills. Dental wax (widely
available from drugstores) must be affixed over top of the fish's back for further stabilization
(Fig. 1B, Sup. Fig. 3 A'')
228
The restraint device docks securely via magnets into the **inner reservoir** (Fig. 1A, Sup. Fig.
2G). We have designed the reservoir to include a second **overflow reservoir** (Fig. 1A), and two
water outlet channels to prevent overflows. For unattended imaging, we recommend adding
water overflow sensors (Castranova *et al.*, 2022). The inner reservoir also features a dock for a
temperature monitor probe (Fig. 1A). The fish is intubated during the imaging period to allow for
adequate oxygenation and delivery of anesthetic. To increase the ease of the intubation
process, we have added a thumbscrew-adjustment on the intubation tubing support (Fig. 1A,
Sup. Fig. 2C) to change the position of the tubing quickly and accurately. This allows for fine
adjustments of tube position to accommodate different lengths/ages of fish.
238
This imaging device was designed in SolidWorks to fit an ASI stage insert and can easily be
modified for other stages. CAD files were sent for 3D printing (Midwest Prototypes) in glass-
filled Nylon 12 using laser sintering. Glass-filled Nylon 12 is both biocompatible and strong and
rigid making it resistant to warping. It can also be easily sterilized via soaking in 10% bleach
without degradation of the material. A complete diagram of assembly steps can be found in
Supplemental Figure 2, and additional assembly instructions are included in supplemental
methods and online (https://morgridge.org/research/labs/fablab/designs/).
246
Animals
Adult zebrafish (*Danio rerio*) were kept in a 14/10 hr. light/dark cycle and fed Gemma Micro 500.
Serially imaged fish were separated into distinguishable pairs in 1L tanks on a quarantine rack
to allow for accurate re-identification of individual animals. Transgenic lines Tg(*kdrl*:mCherry,
mpeg1:eGFP)(Jin *et al.*, 2005) and Tg(*mpeg1*:eGFP) (Ellett *et al.*, 2011) were maintained in a
pigmentless Casper roy[a9], nacre[w1] double mutant (Lister *et al.*, 1999; Ren *et al.*, 2002; White *et*
al., 2008). We have found that nacre[w1] mutants were also suitable for intravital imaging, and the
presence of iridophores does not prevent imaging. Equal numbers of male and female animals
were used. All animals in these studies were handled in accordance with the National Institutes
of Health Guide for the care and use of laboratory animals, and the University of Wisconsin
Institutional Animal Care and Use Committee (IACUC). These studies were approved by the
University of Wisconsin IACUC (protocol L006536).
259
Anesthesia
General considerations for anesthesia
As previously noted, multiple strategies have been used for anesthetizing adult zebrafish for
microscopy. The goals for sustained anesthesia for long term microscopy differ from those of
general anesthesia for brief procedures. MS-222 (tricaine) has been widely adopted for use in
zebrafish for anesthesia. However, there are no standardized guidelines for adult dosage based
on age and/or weight in zebrafish. We found that for animals over 60 days of age and above
180mg, maintenance concentrations of tricaine in the range of 160mg/L to 170mg/L is effective
and well-tolerated (Supplemental Figure 1). Tricaine is tolerated very well in embryonic, larval,
and adult zebrafish. However, we and others have found that the metamorphic stages in
juveniles (roughly ages 21-70 days) tolerate tricaine poorly and have an increased risk of death

during periods of anesthesia regardless of length of exposure or concentration of MS-222
(Owen and Kelsh, 2021). Therefore, we do not recommend imaging of fish in juvenile stages
using MS-222 if it is vital that the fish be recovered. Alternative anesthetic strategies (such as
combined use of isoflurane with tricaine) could also be considered (Huang *et al.*, 2010;
Lockwood *et al.*, 2017; Owen and Kelsh, 2021). Use of a chamber that allows full encasement in
agarose in place of tricaine to restrain the adolescent during imaging is also a viable strategy
(Lam, 2022).

A major drawback to the use of tricaine as an anesthetic for neural imaging is that it can
increase respiration (and thus gill movement) (Huang *et al.*, 2010). This can result in increased
jitter, drift, and imaging artifacts. At our current resolution, most jitter can be resolved
computationally. However, in certain instances where it is crucial to have continuity of frame
over time, paralytic agents such as pancuronium bromide or full encasement in low-melting
point agarose could be considered. Both methods will likely affect the length of time the fish can
be viably imaged, and the ability to recover the fish after imaging.

285
Anesthesia Protocol
Tricaine anesthetic (Syncaine by Syndel) was prepared at a stock concentration of 10mg/mL
and gradually adjusted to pH 7.0 with Tris buffer (pH 9.0). Tricaine is highly acidic when
dissolved in water so the pH should be carefully controlled to ensure comfortable anesthesia. If
tricaine is not dissolved correctly, it can have an appearance like oil droplets on the surface.
This can happen if pH is brought up too quickly, or not properly dissolved. We recommend
discarding the solution if this occurs and consider a lower stock concentration if continued
problems occur. Water for initial induction is prepared in a 1L tank at a concentration of 180-
200mg/L in fresh system water. The maintenance dose of tricaine is prepared in the main
reservoir at a concentration range of 150-170mg/L. Fish were fasted for at least 15 hours prior
to imaging to avoid regurgitation. Adult fish were anesthetized by placement into the induction
tank until loss of righting response was observed and the fish becomes unresponsive to a gentle
tail pinch. After induction was achieved the fish remained deeply anesthetized for several
minutes, during which mounting and intubation can be performed. If intubation is not achieved
within ~3-5 minutes, the fish may begin to show signs of movement and will have to be re-
induced. The goal is to maintain a very light anesthesia for maximum chance of routine survival.
Body standard length has been previously reported as an indicator of anesthesia tolerance in
embryonic and larval zebrafish (Owen and Kelsh, 2021), but length becomes more variable in
adults older than 90 days and is not reliable. We assessed anesthesia dosage per body weight
and found animals weighing 180mg and above to be the most reliable for anesthesia at tricaine
concentrations of 160-170mg/L (Sup. Fig. 1). Animals in the metamorphic (juvenile) state, where
reliance on passive oxygen diffusion through the skin begins to wane in favor of active oxygen
uptake via the gills, are more susceptible to tricaine overdose and exhibit unpredictable
responses to tricaine(Owen and Kelsh, 2021). Therefore, we included only fully adult animals in
this study. The age of the tricaine stock is also very important. Tricaine made from powder
nearing the end of its shelf-life and frozen tricaine stocks nearing their expiration date have
resulted in less predictable outcomes in our hands.
313
Mounting
All materials were prepared and at hand prior to initial induction of anesthesia in the fish, since
the time between induction and intubation should only be brief (2-3 minutes) . Prior to starting
the imaging device was securely placed on the microscope stage, the peristaltic pump and
tubing were attached to the device, and water was warmed to 28° C and circulating in the
chamber.

320
Low melting point (LMP) agarose (2% in PBS, Sigma-Aldritch) is prepared in aliquots, melted at
75° C and maintained at 42° C in a heat block in preparation for mounting. The restraint device
is prepared by wrapping parafilm around the agarose reservoir that will secure the tail (Sup. Fig.
3A). After induction of anesthesia in the fish was confirmed (detailed in Anesthesia section) the
fish was removed from the induction tank by using a pair of soft tipped tweezers to grip the tail
fin and immediately placed into the groove of the restraint device (Sup. Fig. 3A). The device was
tilted from horizontal to vertical and softly tapped on the counter to allow the fish's tail to slide
back into the agarose reservoir (Sup. Fig. 3B). If the tail was caught on top of the parafilm, soft-
tipped forceps or a pipette tip were used to position the tail into the agarose reservoir. The fish
should easily fall into an upright position due to the device shape, but upright positioning can be
ensured by gently using soft-tipped forceps to adjust. Once the position was correct, LMP
agarose is immediately added to the reservoir area of the restraint device while it is still held
vertically (Sup. Fig. 3B). Maintaining a mostly vertical position ensures that agarose was kept
away from the gills. No agarose should be beyond the anal fin of the fish (Sup. Fig. 3A', dotted
line). The LMP agarose was allowed to solidify. An ice pack can be helpful to hasten this
process as long as care is taken to ensure that the fish does not come in direct contact with the
ice pack. Once the agarose is solidified, the chamber was immediately placed into the inner
chamber containing water with Tricaine at the maintenance dosage. Dental wax was draped
over the upper body of the fish and pinched onto the sides of the restraint device to further
stabilize the fish for imaging (Sup. Fig. 3A''). The fish was carefully intubated by placing the
0.5mm ID, 1mm OD silicone tubing into the fish's mouth. Water flow from the intubation tube
was kept at ~6ml/min, corresponding to a pump rate of 26.5 rotations per minute with our
specified pump and tubing. A more detailed protocol is available online at
(https://morgridge.org/research/labs/fablab/designs/).
345
Live Imaging and Recovery
All imaging was performed with a Bruker Ultima or Ultima 2P+ equipped with a Spectra Physics
Insight ultrafast laser and a Nikon CFI75 LWD 16X water-dipping objective. The period between
frames for live imaging varied based on stack size but typically ranged from 30-45 seconds for a
total of 10-20 minutes per movie. Fish were imaged for a maximum of 2 hours each imaging
session. Temperature at the chamber was maintained between 27° C and 28° C, and was
consistently monitored via a temperature probe. The imaging plane was identified by focusing
past the skin and skull (identified by autofluorescent signal and appearance of skin
macrophages) and to the point where characteristic brain blood vessels and sparse immune
cells could be identified. After imaging was completed, the intubation tubing was removed, and
animals were recovered from the chamber by unwrapping the parafilm and using the end of a
forceps or wooden applicator stick to dislodge the agarose from the restraint device. The fish
can then be easily slid out of the device into an individual tank containing fresh water for
resuscitation. The agarose will generally fall off as soon as the fish resumes movement, though
use of a transfer pipette to agitate the water around the fish's tail can help this process. Fish
were housed in as individuals or in distinguishable pairs and monitored for an hour after
recovery. Fish were checked again 12 hours after imaging to ensure no unusual behavior or
injuries were present. We have found that fish are quick to recover from light anesthesia, but if a
fish does not begin to show signs of movement within a minute, a transfer pipette can be used
to force water past the fish's gills to increase oxygenation and speed recovery. All fish
successfully recovered from imaging survived for over a year until they were euthanized
according to normal stock maintenance practices in our animal protocol.
368
369 **Acknowledgements**

We would like to acknowledge the useful discussions that have made this work possible: D. Castranova for the initial inspiration and valuable feedback to refine this imaging protocol, the Zebrafish Rock community for helpful feedback surrounding zebrafish anesthesia practices, and K. Stephenson and L. Krugner-Higby for assistance developing our animal protocol. This work would also not be possible without technical assistance from J. V. Chacko.

Competing Interests
None to report.

Funding
This work was funded by a Morgridge Post-Doctoral Fellowship to E. M. Haynes, a Simons Fellows-to-Faculty award to E. M. Haynes, and and a Wisconsin Alzheimer's Disease Research Center Developmental Project Grant (subawarded from P30AG062715).

References

2. Barbosa, J.S., Di Giaimo, R., Götz, M., Ninkovic, J., 2016. Single-cell in vivo imaging of adult neural stem cells in the zebrafish telencephalon. Nat. Protoc. 11, 1360–1370.
3. Barbosa, J.S., Sanchez-Gonzalez, R., Di Giaimo, R., Baumgart, E.V., Theis, F.J., Götz, M., Ninkovic, J., 2015. Neurodevelopment. Live imaging of adult neural stem cell behavior in the intact and injured zebrafish brain. Science 348, 789–793.
4. Castranova, D., Samasa, B., Galanternik, M.V., Gore, A.V., Goldstein, A.E., Park, J.S., Weinstein, B.M., 2022. Long-Term Imaging of Living Adult Zebrafish. Development. https://doi.org/10.1242/dev.199667
5. Chow, D.M., Sinefeld, D., Kolkman, K.E., Ouzounov, D.G., Akbari, N., Tatarsky, R., Bass, A., Xu, C., Fetcho, J.R., 2020. Deep three-photon imaging of the brain in intact adult zebrafish. Nat. Methods 17, 605–608.
6. Cramer, S.W., Carter, R.E., Aronson, J.D., Kodandaramaiah, S.B., Ebner, T.J., Chen, C.C., 2021. Through the looking glass: A review of cranial window technology for optical access to the brain. J. Neurosci. Methods 354, 109100.
7. Dray, N., Bedu, S., Vuillemin, N., Alunni, A., Coolen, M., Krecsmarik, M., Supatto, W., Beaurepaire, E., Bally-Cuif, L., 2015. Large-scale live imaging of adult neural stem cells in their endogenous niche. Development 142, 3592–3600.
8. Ellett, F., Pase, L., Hayman, J.W., Andrianopoulos, A., Lieschke, G.J., 2011. Mpeg1 promoter transgenes direct macrophage-lineage expression in zebrafish. Blood 117, e49-56.
9. Hefendehl, J.K., Neher, J.J., Sühs, R.B., Kohsaka, S., Skodras, A., Jucker, M., 2014. Homeostatic and injury-induced microglia behavior in the aging brain. Aging Cell 13, 60–69.
10. Huang, W.-C., Hsieh, Y.-S., Chen, I.-H., Wang, C.-H., Chang, H.-W., Yang, C.-C., Ku, T.-H., Yeh, S.-R., Chuang, Y.-J., 2010. Combined use of MS-222 (tricaine) and isoflurane extends anesthesia time and minimizes cardiac rhythm side effects in adult zebrafish. Zebrafish 7, 297–304.
11. Jin, S.-W., Beis, D., Mitchell, T., Chen, J.-N., Stainier, D.Y.R., 2005. Cellular and molecular analyses of vascular tube and lumen formation in zebrafish. Development 132, 5199–5209.
12. Kalmbach, A.S., Waters, J., 2012. Brain surface temperature under a craniotomy. J. Neurophysiol. 108, 3138–3146.
13. Lam, P.-Y., 2022. Longitudinal in vivo imaging of adult Danionella cerebrum using standard confocal microscopy. Dis. Model. Mech. 15. https://doi.org/10.1242/dmm.049753

14. Lister, J.A., Robertson, C.P., Lepage, T., Johnson, S.L., Raible, D.W., 1999. Nacre encodes a zebrafish microphthalmia-related protein that regulates neural-crest-derived pigment cell fate. Development 126, 3757–3767.

15. Lockwood, N., Parker, J., Wilson, C., Frankel, P., 2017. Optimal Anesthetic Regime for Motionless Three-Dimensional Image Acquisition During Longitudinal Studies of Adult Nonpigmented Zebrafish. Zebrafish 14, 133–139.

16. Owen, J.P., Kelsh, R.N., 2021. A suitable anaesthetic protocol for metamorphic zebrafish. PLoS One 16, e0246504.

17. Ren, J.Q., McCarthy, W.R., Zhang, H., Adolph, A.R., Li, L., 2002. Behavioral visual responses of wild-type and hypopigmented zebrafish. Vision Res. 42, 293–299.

18. Schulze, L., Henninger, J., Kadobianskyi, M., Chaigne, T., Faustino, A.I., Hakiy, N., Albadri, S., Schuelke, M., Maler, L., Del Bene, F., Judkewitz, B., 2018. Transparent Danionella translucida as a genetically tractable vertebrate brain model. Nat. Methods 15, 977–983.

19. Tatarsky, R.L., Guo, Z., Campbell, S.C., Kim, H., Fang, W., Perelmuter, J.T., Schuppe, E.R., Conway, K.W., Reeve, H.K., Bass, A.H., 2022. Acoustic and postural displays in a miniature and transparent teleost fish, Danionella dracula. J. Exp. Biol. 225. https://doi.org/10.1242/jeb.244585

20. White, R.M., Sessa, A., Burke, C., Bowman, T., LeBlanc, J., Ceol, C., Bourque, C., Dovey, M., Goessling, W., Burns, C.E., Zon, L.I., 2008. Transparent adult zebrafish as a tool for in vivo transplantation analysis. Cell Stem Cell 2, 183–189.

 Figure Legends

Figure 1. Non-invasive neuroimaging in the adult zebrafish. A) The life-support chamber is
a modular design consisting of a base that docks into the microscope stage. The base includes
an inner water reservoir and an overflow reservoir that allows water to drain and recirculate.
Temperature of inner reservoir water can be continuously monitored via a thermoprobe. The fish
is mounted in a restraint device by fixation of the tail with low melting point agarose. The
dimensions of the restraint device can be modified to accommodate multiple ages and sizes of
fish. **B)** An adult fish mounted in life-support chamber ready to be imaged. **C)** A diagram of
regions of the adult zebrafish brain. Imaging data is presented anterior facing left. TE =
telencephalon, OT = optic tectum, CE = cerebellum, HB = hind brain. **D)** A maximum intensity
projection of both lobes of the telencephalon of the adult zebrafish imaged by multiphoton
microscopy. Magenta = vasculature, as marked by mCherry driven by the *kdrl* promotor. Green
= neuroimmune cells marked by eGFP driven by *mpeg1* promotor. **D')** The eGFP channel from
D isolated to better show neuroimmune cells. **E)** An enlargement of the inset marked by the
white dotted line in D. Cells of ramified and amoeboid morphology are present. **E')** The eGFP
channel from E isolated to better show neuroimmune cells. White arrows indicate cells imaged
in F. **F)** A still from a movie taken at 3x zoom showing a subset of cells present in E' (white
arrows). **F')** A still of the individual cell demarcated by the white dashed box in F showing the
amount of cellular detail that can be captured in live intravital imaging.

Figure 2. Intravital imaging allows monitoring of highly dynamic processes in
***mpeg1:eGFP*+ neuroimmune cells across brain regions. A)** Representative images from
approximately 13 minutes of timelapse intravital imaging from the same animal depicted in
figure 1D. Cells were traced and the difference between depicted frames was determined by
overlaying cell traces (middle panel). The area of the cell volume gained (green) or lost
(magenta) between the represented frame and the previous frame was calculated and is
represented in the graphs to the right. The same procedure was performed in **(B)** except the
imaged cell was in the optic tectum. No significant translocation of cell bodies was detected in
either brain region.

Figure 3. Longitudinal multiphoton imaging can reliably follow changes in the same brain
region over weeks and months. The same animal as depicted in figures 1 and 2 was imaged
across several weeks. Imaris was used to make projections of brain regions with skull and skin
layers removed. **A)** The telencephalon of an individual animal with transgenic labeling of
vasculature (*kdrl:mCherry*, magenta) and phagocytic neuroimmune cells (*mpeg1:eGFP*, green)
was imaged at 2.5 months and again at 3 months. The vasculature can be compared to confirm
similar areas. The morphology of the neuroimmune cells changes significantly over weeks, in
agreement with previous data showing the dynamicity of their processes. **B)** A region of the
optic tectum is imaged in the same individual animal at 3 months and again at 4 months old. **C)**
Regions of the vasculature from the optic tectum of the same animal imaged at 3 months, 3.5
months, and 4 months old demonstrating that the same region can be reliably identified across
imaging periods.

Figure 4. Longitudinal intravital imaging can detect changes in neuroimmune cell
behavior between young and aged animals. A) The telencephalon of a 19-month old

zebrafish with vasculature in magenta and neuroimmune phagocytes in green. **A', A''** are
enlarged insets of the regions marked by white dashed boxes in A. **B-C)** Representative stills
from a 14 minute long movie of the region depicted in A'. **B** shows two individual cells (C1, C2)
touching and then elongating in response to contact. **C** shows an individual cell that retracts its
protrusions over the course of the imaging period. **D)** To demonstrate that these methods allow
us to detect differences in neuroimmune cell dynamics across ages, cells in two sample movies
from a 3-month old animal were tracked and compared to cells from two sample movies from a
19-month old animal. Tracks were corrected for drift. Cells in the 19-month old animal
demonstrated greater velocity and distance translocated than cells in the 3-month old animal.
Velocity was calculated by summing the distance traveled per frame per cell, subtracting the
drift, and then dividing the distance traveled by the imaging period. P=0.0099, unpaired T-test
with Welch's correction. Points represent individual cells.

Supplemental Figure 1. **A) Fish subjected to longitudinal imaging did not demonstrate**
weight loss and continued to grow during the imaging period, suggesting that imaging
did not interfere with normal feeding behavior or cause excessive stress. **B)** The time
required to resuscitate an individual animal after imaging was not significantly correlated with
any measured parameters. Time to Resuscitate vs Age: p = 0.115, slope = -1.22, Time to
Resuscitate vs Time Imaged: p = 0.859, slope = -0.419, Time to Resuscitate vs Weight: p =
0.071, slope = 28.82. **C)** Table of tricaine dosage based on stage of anesthesia for fish older
than 60 days old (past metamorphic stage).

Supplemental Figure 2. Chamber assembly and full feature diagram. A) The inflow hole
and outflow holes (**H.1**) are shown. All holes are tapped using a 1/16-27 NPT standard tap and
threading-to-barb fittings (1/16" ID tube) are screwed in. **B)** Holes for screws in the top of the
support (H.2) are tapped using a #4-40 UNC tap. Socket screws are screwed in to tighten and
loosen the clamp as needed. **C)** Steel dowel pins are placed in H.3, going through the outer
chamber and the support piece. A thumb screw is placed in H.4, with only the part of the hole
through the support piece being tapped. A M3x0.5 tap is used. The thumb screw must have a
0.5 mm groove starting 3.5 mm from the collar and heading away from the collar. The resulting
diameter in the groove must be 3/32". **D)** Set screws to secure the dowel pins are placed in the
side holes (**H.5**). These side holes are tapped with a M2x0.4 tap. **E)** An external retaining ring is
placed in the groove on the thumb screw to secure it in place. **F)** A small amount of epoxy is
placed in H.6, then magnets are inserted. More epoxy is placed over the magnets to protect
them from corrosion. **G)** Magnets are placed in H.7 in the inner chamber using the same
method as the previous step. The magnets snap the inner chamber into place but still allow for
easy removal. **H)** The digital thermometer probe is placed in the holder (**H.8**) in the inner walls
and the blunt end needle and Luer lock adapter is secured in the support (**H.9**).

Supplemental Figure 3. Mounting of fish and imaging setup. A) The restraint device is
prepared by wrapping parafilm around the low melting point agarose reservoir. **A')** The
anesthetized fish is placed into the prepared restraint device and tilted upwards (tapping on the
benchtop helps fish's tail slide into the agarose reservoir). Low melting point agarose can be
added to reservoir by pipetting and should not fill the reservoir past the dotted line to prevent
interference with gills. Caution should be taken to ensure that agarose is an appropriate
temperature to avoid burns to fish. **A'')** A blanket of dental wax can be draped across the fish's
back to further immobilize, with care taken to leave head and gills exposed. Dental wax can be
pinched to side of restraint device to secure. Once agarose is set, the device is ready to be

placed into the inner reservoir of imaging chamber. **B)** Example image of fish in restraint device
prior to adding dental wax. **C)** Fish placed in restraint device in imaging chamber without
addition of dental wax. Intubation tube is placed in the fish's mouth once it is inside chamber.
Caution: take care not to position the intubation tube too deeply in the fish's mouth to prevent
damage or death to animal. **D)** Schematic of total intubation setup. A perfusion pump is used to
circulate warmed system water containing tricaine anesthetic through a pulse dampener and
into the imaging chamber. Water is continuously circulated and overflows the top of the inner
chamber into the overflow chamber. Water outflows from the overflow chamber and returns to
the main carboy where it can be rewarmed and oxygenated.

Figure 1

A.
Restraint Device
Overflow Reservoir
Inner Reservoir
Thermoprobe Slot
Intubation Tubing Guide & Fine Adjustment

B.

C.
TE
OT
CE
HB
Anterior
Posterior

D. 200 µm

E. 50 µm

D'. 200 µm

E'. 50 µm

F. 50 µm

F'. 15 µm

Figure 2

Figure 3

Figure 4

A. Telencephalon - 19 months

A'

A''

100 μm

50 μm

50 μm

D.

Velocity

μm/min

**

3 mo.

19 mo.

B C1 C2 T_0 C1 C2 T_7 C1 C2 T_{14} 15 μm

C T_0 T_7 T_{14} 15 μm

Supplemental Figure 1

A.

B.

C.

Stage of Anesthesia	Final Concentration (mg/L)	Amount of 10mg/ml Tricaine to add to System Water
Induction	180-200	9-10ml to 500ml water
Maintenance (low)	150	22.5ml to 1.5L water
Maintenance (medium)	160	24ml to 1.5L water
Maintenance (high)	170	25.5ml to 1.5L water

Supplemental Figure 2

Supplemental Figure 3

A.

A'.

A".

B.

C.

D.

Digital Probe Thermometer

Upright 2-Photon with 16x Water Dipping Objective

16xW LWD 0.7 NA

to intubation coupling

chamber inflow

outflow tubing

Pulse Dampener

inflow tubing

4-line Peristaltic Pump at a 6ml/min flow rate

Carboy of system water containing:
- pharma. tricaine
- aquarium heater
- bubbler
- thermometer