Quality-control Normalization of Fluorescence Microscopy Morphometry and Colocalization Measurements for Improved Accuracy and Cross-instrument Reproducibility

John Oreopoulos[1], Glyn Nelson[2], Matthew Gastinger[3], Christopher L. Morrison[1], Sabine Thomas[4], Michael Kiebler[4], Aidan Boyce[1], Bernhard Goetze[1]

[1] Andor Technology Ltd., an Oxford Instruments company, Belfast, United Kingdom

[2] Faculty of Medical Sciences, Newcastle University, Newcastle upon Tyne, United Kingdom

[3] Oxford Instruments Imaris, Zurich, Switzerland

[4] Biomedical Center, Ludwig Maximilian University of Munich

Abstract

Quantitative fluorescence microscopy is more reproducible when instrument performance is measured and incorporated into the analysis. We show that routinely monitored quality-control (QC) metrics like the system resolution and inter-channel co-registration are determinant variables that can be used to normalize common image readouts and thereby separate instrument-induced variation from genuine biological changes. For intra-channel morphometry, a Gaussian approximation of the fluorescence imaging process yields analytical factors that predict how geometric measurements (length, separation distance, area, volume, etc.) inflate and scale with resolution blur due to optical misalignments or natural optical quality variations. We validate this behavior by deliberately perturbing the system resolution and by exploiting the natural resolution differences in three nominally equivalent objective lenses configured to image the exact same synapses in cultured hippocampal neurons, where structural differences subtle by eye nonetheless produced statistically significant shifts in measured synaptic puncta volumes. For dual-channel colocalization (overlap) measurements, we normalize the inter-channel co-registration QC metric by the measured point-spread function (PSF) resolutions (rather than theoretical limits associated with the objective lens) and demonstrate how fluorescent pre/postsynaptic cleft protein overlap signals decay in a predictable, exponential fashion as the PSF-normalized registration error increases, with the decay rate depending on the imaged object relative to the PSF size ratio. Mapped field-of-view gradients in channel registration also explain feature orientation flips/rotations and overlap loss without any underlying biological change. Finally, we outline a simple QC-aware microscope normalization

workflow where each image measurement dataset is paired with its session PSFs and local co-registration error to remove instrument bias and optionally re-project the results to a declared reference PSF without altering the raw images. This approach improves image measurement accuracy and cross-instrument comparability of experiments and reframes light microscope QC from a passive certification of instrument health into a practical normalization that links the acquisition state to quantitative outcomes, thus ensuring the reliability and interpretability of morphometric and colocalization data in fluorescence microscopy.

Introduction

Light microscopy is a cornerstone technique of the biological and material sciences, enabling researchers to visualize structures and understand molecular interactions at microscopic scales[1–4]. However, the reliability of the observations and measurements obtained through this technique hinges on quality control (QC) practices[5–12]. Among the many different instrument QC factors that can potentially influence data quality and reproducibility, two critical metrics that come into play in many experimental outcomes are the microscope resolution and chromatic channel co-registration[13–17]. Unchecked variation in these parameters can propagate into image readouts, making comparisons across fields of view (FOV), instruments, and laboratories almost impossible.

Resolution is the ability of the microscope to distinguish between two closely spaced objects and ultimately determines the perceived sharpness of the images[18,19]. Even with the best optical design, this is physically limited by the wavelength of light and the numerical aperture (NA) of the respective objective lens[20]. Typical diffraction-limited performance with a high-NA oil-immersion objective (NA ≈ 1.3–1.4, λ ≈ 520–600 nm) yields lateral resolution on the order of ~200–250 nm and axial resolution ~500–800 nm. At these length scales, one can cleanly identify and separate cell bodies, nuclei, mitochondria and lysosomes at the organelle level, trace cytoskeletal tracks and cellular processes such as neurites, and quantify diffraction-limited puncta (e.g., synaptic markers, endocytic pits) as objects; however, sub-organelle features such as synaptic vesicles (40–60 nm), the 20–40 nm synaptic cleft, nuclear-pore substructure, or receptor nanoclusters and similarly sized nanoscopic molecular complexes remain below the resolvable limit and appear blended or enlarged. Paths toward improved resolution include computational deconvolution to restore contrast and moderately sharpen features[21], structured-illumination microscopy (SIM) achieving ~100 nm lateral and improved axial resolution[22,23], stimulated emission depletion (STED) routinely reaching ~50 nm laterally[24,25], and single-molecule localization methods (PALM/STORM/DNA-PAINT) achieving ~10–30 nm precision (with effective resolution set by sampling density and reconstruction)[26–29]. These approaches have revealed previously inaccessible biology such as the layered architecture of focal adhesions[30], the periodic actin–spectrin lattice in axons[31], nanoscale receptor nanodomains and their alignment with presynaptic release sites ("nanocolumns")[32], and the ring-like organization of the nuclear pore complex[33]. Yet each route to superior resolution is accompanied by trade-offs and additional costs, including higher light doses and phototoxicity, demanding labeling density and photophysics, longer acquisitions and drift sensitivity, complex

reconstruction/calibration routines, and for some methods, limited compatibility with thick or
dynamic live samples.

Chromatic co-registration refers to the 3D overlay accuracy between a microscope's distinct
spectral channels and determines how faithfully spatial relationships and interactions between
fluorescently labeled biomolecules or targeted molecular species are represented[34–36].
Departures from ideal overlay arise from both longitudinal (axial focal shifts) and lateral
(color-dependent magnification/translational) chromatic aberrations, as well as dispersion
mismatch, cover-glass thickness errors, filter tilt/wedge angles, and misalignments in the
excitation or detection paths. These image registration shift mechanisms motivate two classes
of remedies: hardware designs and alignments that minimize color error at the source, and
software corrections that estimate and remove the remaining offsets. Hardware strategies to
improve registration include apochromatic or super-apochromatic objectives with extended color
correction, matching tube lenses and relay optics to objective design, minimizing wedge/tilt in
filters, selecting multi-band dichroics engineered for low lateral color, and using immersion
media whose dispersion matches the specimen[37]. Recently, reinforcement learning combined
with tunable lenses have also been employed for real-time correction of chromatic aberrations[38].
Careful alignment of excitation and detection paths plus routine microsphere-based checks help
maintain both on-axis and off-axis overlay[12]. Software approaches register channels
post-acquisition or in real-time using mapped microsphere fields, embedded fiducial markers,
and biological landmarks to estimate global rigid transforms (translation/rotation/scale) or
spatially varying warps (polynomials, thin-plate splines, piecewise-affine), extended to 3D to
handle axial chromatic focal shifts[39–44]. Image cross-correlation correction methods that do not
require fiducial markers or mapped microsphere offsets have also been used[45]. When calibrated
per instrument and objective, these software approaches can bring channels into sub-pixel
agreement across much of the FOV and markedly stabilize two-color measurements. Even so,
chromatic corrections have limitations as well: Chromatic shifts often vary with imaging depth
and refractive-index mismatch, so a single 2D mapping may fail in thick samples; non-rigid
warps can subtly distort morphology, and pixel intensity resampling introduces blur and pixel
correlations. Furthermore, transformations of image datasets may not perfectly transfer to
heterogeneous specimens, and for live-cell imaging, the added processing can reduce temporal
resolution or fail under drifting illumination, lateral stage, and axial focus conditions.

Despite the availability of super-resolution and sophisticated chromatic-correction solutions,
conventional, diffraction-limited fluorescence imaging remains the workhorse of cell biology. It is
fast, comparatively gentle on live samples, broadly compatible with common labels and thick
specimens, inexpensive to acquire and maintain, and straightforward to train across
laboratories. It also scales to high-throughput screens and longitudinal studies while preserving
continuity with the vast legacy of archival datasets. As light sources, advanced optics, and
cameras become cheaper and more uniform across core facilities and labs, the limiting factor
for experimental comparability is increasingly practice, not hardware, thus making standardized
QC both timely and impactful. Indeed, the continued expansion of the consortium for Quality
Assessment and Reproducibility for Instruments and Images in Light Microscopy (QUAREP-
LiMi), comprising over 500 members across academia and industry globally, is a sign of the

growing recognition of the critical importance of and desire for QC in microscopy related
research[46,47]. Without shared and standardized practices, though, variability across instruments
and associated protocols will obscure biological effects and limit cross-study comparability.
Accordingly, our goal in this paper is to provide an unambiguous, experimentally grounded and
theoretically supported demonstration of why routine light-microscope QC is essential. We
image the same biological structures while deliberately varying two core instrument metrics—
resolution and chromatic co-registration—and show that these QC values predict the direction
and magnitude of changes in standard morphometry and colocalization readouts. To make this
link explicit and usable, we develop a compact, analytic normalization treatment that connects
QC to measurement bias and supply a practical rationale for normalizing conventional
fluorescence imaging measurements when those measurements are close in scale to the
related QC metrics. To operationalize this motivation, we adopt QUAREP-LiMi–aligned
procedures to characterize instrument performance and investigate corresponding
consequences for standard readouts[48,49]. Specifically, we quantify resolution along with
inter-channel offsets and then re-image matched cellular structures under distinct microscope
performance conditions to test the predicted effects on morphometry and colocalization.

As a biologically-grounded model system for linking these two QC metrics to image readouts,
we focus on presynaptic and postsynaptic protein puncta in cultured hippocampal neurons.
These synaptic compartments have a well-established nanometer-scale organization from
electron microscopy[50,51] and super-resolution light microscopy studies[52–54], yet, when imaged
with conventional, diffraction-limited light microscopy[55], they exist at the edge of optical
resolution. In excitatory synapses, the active zone (AZ) and postsynaptic density (PSD) face
one another across a ~20–40 nm cleft. Commonly used antibody epitopes (e.g., vGLUT,
Syanpasin, and Bassoon at the AZ and Homer/PSD-95 at the PSD) are offset in 3D by roughly
120–160 nm. Consequently, two-color fluorescence microscopy image readouts that appear
coincident at the pixel level often reflect tightly apposed but non-identical structures—exactly
within the geometric regime where traditional light microscopy resolution and chromatic co-
registration most strongly modulate measured volumes, overlaps, and pixel-wise coefficients.
Image assays of these neuronal structures are also methodologically mature and widely
adopted. For example, fixed-cell synaptogenesis workflows typically immunolabel a presynaptic
marker (e.g., Bassoon) and a postsynaptic marker (e.g., Homer or PSD-95) and then quantify
either (i) punctum volume/intensity per channel or (ii) pre/post colocalization (overlap fractions,
object-pair counts, or pixel-wise coefficients). Popular pipelines such as the ImageJ Puncta
Analyzer define a synapse as an apposed pre/post pair (often with an "overlap-pixel" or sub-
diffraction centroid-distance rule), making the readout both standardized and conveniently re-
analyzable under different QC regimes[56]. Biologically, these structural metrics are meaningful.
Puncta volume and intensity correlate well with synaptic organization and function (e.g., AZ
scaffold size and release probability presynaptically; PSD scaffold/receptor content
postsynaptically)[57], while colocalization-based synapse counts track developmental, activity-
dependent, and disease-related changes[58]. Many groups pair the structural readouts with
functional measurements (mEPSC frequency/amplitude, pHluorin/iGluSnFR/GCaMP reporters)
or ultrastructural validation, reinforcing their interpretability[59,60]. Critically for this study, the same
features that make synapses biologically interesting also make them QC-sensitive. Because

true pre/post separations are sub-diffraction, yet non-zero in size, even modest changes in the
PSF (resolution) can inflate apparent punctum volumes and artificially increase overlap,
whereas tens-of-nanometers of chromatic mis-registration can erase true appositions or create
spurious false-positive ones. Leveraging instrument-derived QC maps, we show that these
biases predictably reshape common observations of synaptic morphology across the FOV and
across objective lenses, enabling principled normalization and improved experimental design.
Although we illustrate the analytic formalism and companion workflow for such observations, the
same QC-to-measurement relationship applies to any image assay where feature sizes and
separations are on the order of the system's resolution.

Methods

Notation

Throughout this work, uppercase axis labels (X, Y, Z) are used to denote the physical axes of
the microscope hardware (e.g., motorized XY stage translations or axial Z-scans), whereas
lowercase symbols (x, y, z) are used for spatial coordinates in the image and analysis domain.
This convention distinguishes instrument motions from mathematical variables. In the present
study, the instrument and coordinate systems are aligned such that X, Y, and Z correspond
directly to x, y, and z, respectively. In particular, the Z-axis of the focus drive coincides with the
optical axis of the imaging system.

Microscope equipment

All experiments were carried out on an Andor Technology / Oxford Instruments Benchtop
Confocal Microscope (BC43) equipped with three different plan-apochromat 60× oil immersion
objective lenses having nearly the same NA, but were of different optical design and model
generation (Objective 1: Nikon 1.42 NA CFI Plan Apochromat Lambda D series, Model#
MRD71670; Objective 2: Nikon Plan Apochromat XA 1.40 NA, model# MRD01600; Objective 3:
Nikon 1.40 NA CFI Plan Apochromat Lambda series, Model# MRD01605). This spinning-disk
confocal microscope[61,62] uses a 4.2 megapixel (2048×2048 pixels) scientific complementary
metal-oxide-semiconductor (sCMOS) camera detector having 6.5 μm×6.5 μm size pixel
elements. At 60× magnification, the back-projected sample plane pixel size was almost exactly
100×100 nm and was confirmed by imaging the "field of rings pattern" of the Argolight Argo-HM
calibration slide[63] (Argolight SA, France). The BC43 has four distinct spectral channels utilizing
a common multi-band dichroic mirror in combination with directly modulated lasers for
fluorescence excitation and bandpass interference filters loaded in a rotating filter wheel for
viewing fluorescence emission. Only the "Green" (488 nm excitation; 521/24 nm emission) and
"Yellow" (561 nm excitation; 593/31 nm emission) channels were used for this study. All images
were acquired using the confocal imaging mode of the microscope (50 μm pinhole physical size
on the disk). The same immersion oil was used for all three objective lenses at room
temperature (Thorlabs OILCL30; Cargille Type LDF with refractive index 1.518).

Neuronal cell culture

Cell cultures from primary rat hippocampal neurons were generated as previously described[64]. In short, embryos of timed pregnant Sprague-Dawley rats (Charles River Laboratories) were used to isolate hippocampi of embryonic day 17 (E17). Cells were dissociated and plated on Poly-L-Lysine coated cover slips and cultured in NMEM+B27 supplemented medium (Invitrogen). Immunostaining was performed with cultured neurons on day 17 in vitro (DIV).

Immunostaining and sample preparation

Neurons were fixed for 15 minutes with 4% paraformaldehyde (PFA) and immunostained as previously described[65]. The following primary antibodies were used for incubation at 4°C overnight: polyclonal guinea pig anti-vGLUT 1:500 dilution (Synaptic Systems) and polyclonal mouse anti-Homer 1:500 dilution (Synaptic Systems). On the next day, the following secondary antibodies were incubated for 1 hour at room temperature: anti-guinea pig AlexaFluor488 and anti-mouse Cy3 conjugated antibodies 1:1,000 dilution (Life Technologies). DAPI staining was carried out for 5 minutes at room temperature. Coverslips were mounted on microscope slides with ProLong Diamond Mounting Medium (Life Technologies).

Quality control and instrument characterization

The most common and practical method for experimentally measuring the microscope PSF involves imaging sub-resolution fluorescent microspheres[13,12,48]. The microspheres act as idealized point sources of fluorescence emission and their image, when captured by the microscope, directly represents the system's PSF. In a well-aligned system using high-quality objective lenses and optics, the PSF response near the middle of the imaged FOV is constant (isoplanatic), thereby producing the most reliable imaging results. However, even in systems such as these, it is not uncommon to observe deviations from the central, on-axis PSF response in the image corners as well as at greater depths within inhomogeneous samples[66]. The microscope's channel pair-dependent chromatic co-registration can also be measured in a similar manner using multi-color emitting microspheres and localizing the 3D center of the microsphere in each spectral channel to measure the positional offset of different wavelengths[49]. Similarly to how the microscope's PSF can exhibit a spatial dependence across the FOV, so too will there often be differences in channel registration across the FOV, and these spatial variations in channel overlap can considerably impact upon measures of co-localization or overlap between two fluorescently labelled biomolecules.

Since this report attempts to directly and unambiguously demonstrate how changes in a light microscope's system QC metrics can have a direct impact on experimental outcomes, stringent efforts were made to image the exact same FOV within all examined samples with each objective lens using identical laser powers and exposure times while simultaneously limiting and minimizing light exposure to avoid specimen photobleaching. Before imaging, the microscope was allowed to reach thermal equilibrium, ensuring negligible axial (Z) drift during data acquisition. Sequential dual-channel 3D confocal Z-stacks (10 μm Z-range, 100 nm Z-step) of

200 nm multi-color TetraSpeck fluorescent microspheres mounted in Saint Gobain BC-600 Optical Cement clear epoxy resin (refractive index 1.56) between a 170 µm thick glass coverslip - 1 mm glass slide sandwich (Thermo-Fisher T14792) were acquired. Acquisition FOVs were carefully selected for those exhibiting well-dispersed single microspheres on the coverslip surface with good coverage right to the image FOV corners and avoiding FOVs possessing obviously bright microspheres (which were likely to be aggregates). Once selected, the same FOV was scanned with each 60× objective lens to ensure that the characterization of each objective in use with the BC43 microscope would be completely fair and equal. Multiple FOVs were imaged in the same manner with each objective lens in sequence as replicates to confirm the repeatability of the image measurements taken with the first FOV. These image datasets were subsequently used to create quantified spatial heatmaps of the (Green channel) resolution and (Green-to-Yellow channel) chromatic co-registration of each objective lens (see below for more details). Z-stacks of the Argolight Argo-HM slide "field of rings" pattern were also acquired with each objective lens using the same acquisition settings as an alternative sample type to assess and confirm measurements of chromatic co-registration made using the TetraSpeck microsphere sample.

Full FOV lateral (XY) and axial (Z) heatmaps of the system PSF in conjunction with each 60× objective lens were created using the freely available and open-source software package PSFj[14]. Very briefly, this program processes 3D Z-stacks of fluorescent microspheres by first automatically identifying individual microspheres through size and intensity thresholds and then subsequently fitting a 3D Gaussian ellipsoid to each microsphere using non-linear regression techniques. After accounting for the sample and immersion medium refractive index, as well as the voxel size and microsphere size (100 nm x 100 nm x 100 nm and 200 nm diameter respectively in our case), PSFj extracts the minimum and maximum lateral full-width at half-maximum ($FWHM_{min}$ and $FWHM_{max}$ respectively) as well as the axial full-width at half-maximum ($FWHM_z$) Gaussian model fit parameters associated with each identified microsphere. Since the Airy disk diffraction pattern of a fluorescent point-source object is well approximated near its center by a Gaussian function, the fitted Gaussian widths (FWHM) are commonly used as estimates of the lateral and axial resolution[12,19,48] (see also Supplementary Text). For this study, we ignored the $FWHM_{min}$ values and instead focused on the more conservative $FWHM_{max}$ values to denote the measured lateral resolution. Having also measured and tabulated the 3D sub-voxel localization coordinates of each identified microsphere, the program then outputs color-coded, continuous heatmaps of the lateral and axial resolution using a sliding-window median filter that assigns each heatmap pixel the local median FWHM value of nearby microspheres. PSFj also possesses a dual-channel analysis mode which can conduct the same image analysis pipeline on image datasets of multi-color microspheres and can create dual-channel colored heatmaps of the lateral and axial chromatic mis-registration by computing the lateral and axial separation distances between identical microspheres seen in both channel datasets. For single-channel and dual-channel analysis modes, PSFj also outputs a full spreadsheet file listing the found coordinates of each processed microsphere in both channels along with the measured lateral and axial FWHM values. A custom-written Python programming language-based executable application (see download link in Supplementary Text) was used to convert the listed dual-channel microsphere sub-voxel localized coordinates into measured

lateral and axial chromatic separation distances and then subsequently transformed into linearly
interpolated co-registration heatmaps according to the definition of co-registration set out by
Faklaris *et al.*[12]. Just like with the resolution heatmaps, interpolated spatial representations of a
microscope system's co-registration like this allow for easy look-up of where within the FOV the
inter-channel 3D mis-registration is minimal and to what quantifiable degree. We also
normalized the spreadsheet $FWHM_{max/z}$ values against the theoretical lateral/axial resolutions
for each objective lens and replotted them against each other on 2D scatter graphs, as was
demonstrated by Faklaris *et al.*[12]. Lateral and axial channel separation distances were
normalized by the *measured* lateral and axial resolutions and similarly plotted against each
other. Transformations of the PSFj data output like this permit a quick and easy comparison of
system resolution and co-registration across different microscope instruments (or equivalently in
our case, the same system using different objective lenses).

Cellular imaging

The cultured neurons were imaged in the same manner as the QC calibration samples.
Sequential dual-channel (Green; Alexa488-vGLUT and Yellow; Cy3-Homer) 3D confocal Z-
scans of the fluorescent pre- and postsynaptic proteins in the neuronal cells were captured with
all three objective lenses consecutively, again taking great care to maintain the laser power and
exposure time acquisition settings within the exact same sample FOV each time. For some
cells, after acquiring the initial Z-scan, the sample was immediately translated with the
motorized XY-stage such that a dendrite of interest in the middle of the FOV was moved to the
upper left-hand corner of the FOV and another Z-scan was acquired. For these cells imaged in
the middle and corner of the FOV, the same procedure was also immediately followed with the
other two objective lenses. These paired middle and corner FOV-dependent datasets were used
for the co-registration studies described in the results section.

Neuronal cell image analysis

All 3D surface objects were created using Imaris 10.2 (Oxford Instruments). The creation
parameters were constructed using the built-in Imaris machine learning pixel segmentation
method. To provide a consistent and reproducible surface creation, the detection algorithm was
trained on the same puncta from all 3 objective lenses simultaneously. Training pixels within
the XY and XZ planes yielded accurate detection of the fluorescent edges of the puncta. This
training process was followed independently on both the Green and Yellow channels. Each
successfully identified Yellow channel puncta was paired with a corresponding Green channel
puncta. Following object creation, the resulting surfaces were filtered by intensity, shape, and
size. For each objective lens and region of interest (ROI), the same puncta were rendered and
analyzed. Inter-channel, (object-based) colocalized puncta volume overlap was defined by
measuring the percentage overlap enclosed between the pre- and postsynaptic puncta surfaces

while colocalization was calculated using the traditional Pearson's correlation coefficient (PCC)
definition[15].

Operational definition of puncta (vGLUT and Homer)

In this study, a *punctum* denotes a contiguous 3D fluorescence object segmented from a single
channel's voxels after background correction, thresholding, and size filtering. When the Green
channel reports a successful segmentation (vGLUT), the resulting punctum represents a single
presynaptic site at light microscope resolution—i.e., the AZ scaffold and/or its associated
vesicle-rich cluster, not an individual vesicle or molecule. In contrast, when the Yelllow channel
reports a segmentation (Homer), the resulting punctum is the local postsynaptic specialization—
i.e., the excitatory (PSD) or inhibitory (scaffold/receptor) density, not an individual receptor
complex. In all image analyses, "puncta" refer to these segmented 3D masks. Volume and
overlap/colocalization image measurements were computed directly from the associated binary
masks and intensity counts within their voxel supports.

Results

How resolution impacts sub-cellular volume image measurements

We designed two experiments based around one microscope stand to directly demonstrate the
effect of relatively minor optical misalignments and component alterations on intra-channel
morphometric image measurements. In the first experiment, acting as a positive control, we
purposefully reduced the effective NA of Objective 1 by constricting the back-aperture of the
objective lens with an internal iris positioned near an equivalent/conjugate back-focal plane in
the fluorescence emission pathway relay optics (Fig. 1). This motorized iris normally opens and
closes automatically via firmware settings when the image mode is toggled between
fluorescence and differential phase contrast[67] (DPC) modes to suppress background noise and
enhance contrast respectively. The aim of these artificially induced resolution changes was to
be deliberately subtle with the system perturbations to mimic real world scenarios where a
microscope may well have some minor aberration or misalignment error in the light path, but the
user would not necessarily notice through visual inspection of the images unless they had
performed a quantitative QC check. This restriction of the objective back aperture in the
emission pathway was expected to increase the lateral and axial FWHM compared to the Full
NA unrestricted condition, and indeed we measured a 12% lateral and 18% axial widening of
the PSF associated with a single microsphere located in the middle of the FOV (ie: reduced
resolution) in the most restricted condition (Restricted NA 2, see Fig. 2).

Objective lens 1

Equivalent Back-focal plane in emission pathway optics

Full NA   Restricted NA 1   Restricted NA 2

Fig. 1 | Schematic representation of the resolution change positive control experiment. Fluorescence images of QC microsphere calibration samples and fluorescent neurons were captured with Objective 1 using an increasingly restricted emission pathway beam through means of a motorized iris positioned in an equivalent back-focal plane of the microscope optical train. Radial constriction of the emission beam diameter in this infinity space is equivalent to reducing the effective emission numerical aperture associated with the objective lens and therefore should lead to a reduction in image resolution as judged by the PSF extent of the microspheres or the volumes of small dendritic spine puncta in the neurons. **a,** The Full NA of the objective lens is utilized to image the sample since the completely open iris exceeds the beam size and back aperture diameter of the objective. **b,** A slightly closed iris position (Restricted NA 1) vignettes the periphery of the emission beam and induces a small, but measurable resolution degradation. **c,** The iris was closed down further (Restricted NA 2) to make the resolution degradation more obvious and greater in magnitude. Schematic not to scale.

By eye, the resolution change was subtle and barely measurable under the Restricted NA 1
setting but became more pronounced under Restricted NA 2, especially in the axial (XZ)
projection of the microsphere. This behavior was expected since axial resolution scales
approximately as $1/NA^2$, whereas lateral resolution scales as $1/NA$[19]. Even with the larger
degradation at Restricted NA 2, the magnitude of the resolution change remained difficult to
estimate by visual inspection alone, underscoring that qualitative impressions can diverge from
quantitative PSF-based measurements. This is the first example we show where visual
impressions of an image do not correlate well with quantified image measurements.
352

Fig. 2 | PSF analysis of a single fluorescent microsphere observed with varying emission pathway NA conditions as described in Fig. 1. a, Lateral XY (top row) and axial XZ (bottom row) maximum intensity projections of the microsphere for each emission NA setting. Scale bar = 0.5 µm. **b**, Normalized lateral and, **c**, axial fitted Gaussian intensity line profiles of the microsphere for each NA setting. **d**, The associated absolute and normalized FWHM profile values are tabulated with the relative difference to the Full NA condition.

353   We then measured the microscope PSF over the entire FOV to obtain a better statistical
354   sampling and look for signs of any spatial dependencies of the system's resolution (Fig. 3). The
355   average lateral and axial FWHMs for the two restricted NA conditions across the FOV showed

	Full NA	**Restricted NA 1**	**Resticted NA 2**
Avg. FWHM max (nm)	248 +/-11	254 +/-8	294 +/-6
Avg. FWHM max/xy$_{theory}$	1.31 +/-0.06	1.34 +/-0.04	1.55 +/-0.03
Avg. FWHM max %Change		2	16
Avg. FWHM z (nm)	625 +/-21	661 +/-16	772+/-32
Avg. FWHM z/z$_{theory}$	1.16 +/-0.04	1.22 +/-0.03	1.43 +/-0.06
Avg. FWHM z %Change		5	19

Fig. 3 | Full FOV microscope lateral and axial resolution characterization of Objective 1 under the varying emission pathway NA conditions. a, Green channel maximum intensity Z-projection, full FOV image of 200 nm fluorescent microspheres acquired with Objective 1 at Full NA. **b**, PSFj lateral (FWHM$_{max}$) and axial (FWHM$_z$) resolution heatmaps of Objective 1 for each emission pathway iris condition of the resolution control experiment. Scale bar in **a** and **b** is 20 μm. **c**, For each NA setting, the lateral and axial FWHM values for each isolated microsphere in **a** were normalized by their theoretical values for Objective 1 (FWHM$_{xy,theory}$ = 190 nm; FWHM$_{z,theory}$ = 541 nm) and plotted against each other. The green box delimiting the 1.5x theoretical resolution ratios match the recommended microscope resolution performance tolerance limits of Faklaris et al.[12]. **d**, The centroids or each data "cloud" in **c** plotted the same way representing the average resolution performance across the FOV for each NA setting of the control experiment. **e**, Average FOV resolution FWHM measurements in tabular form, with the relative change in relation to the Full NA scenario.

significant increase in lateral and axial spread of the PSF (16% and 19% respectively for the
most restricted NA (Fig. 3e), again in line with the expected resolution reduction due to a purely
reduced detection NA. Generally, the microsphere's PSF lateral extent uniformly inflated across
the FOV because of the induced aberration, whereas the axial extent showed a slight spatial
dependency with the PSF broadening, biased mostly in the middle region of the FOV (Fig. 3b
and e).

Normalizing the measured $FWHM_{max}$ and $FWHM_z$ for each microsphere in the FOV by Objective 1's Full NA theoretical lateral and axial resolution ($FWHM_{xy,theory} = 0.51\lambda_{EM}/NA$, λ_{EM} = emission wavelength; $FWHM_{z,theory} = \lambda_{EM}/n\text{-}\sqrt{(n^2 - NA^2)}$, n = immersion media refractive index) and plotting these data points against each other on a Cartesian graph produced a "cloud" of lateral/axial resolution pair data points which allowed us to assess the overall resolution behaviors for each of the three NA conditions of the positive control experiment (Fig. 3c). Furthermore, we determined the average resolution behaviors for each of these NA conditions by plotting the centroid of each data cloud (Fig. 3d). With these data transformations, we observed that the resolution changes due to the purposefully "misaligned" emission pathway iris closure was sufficient to push the judged resolution performance of Objective 1 from a level of "acceptable" at the Full NA condition (both FWHM ratios <1.5) to an "unacceptable" level according to the tolerance recommendations of Faklaris et al.[12] and the QUAREP-LiMi standards community at the Restricted NA 2 condition (lateral $FWHM_{max/theory}$ ratio >1.5).

An alternative investigation of the PSF's lateral broadening was performed using Fourier domain (spatial frequency) analysis techniques yielding a similar degraded resolution trend under progressively closed down detection NAs by examining the radially averaged modulation transfer function (MTF) resulting from the Fourier transform of maximum intensity Z-projection images of the microspheres[68–71] (see Supplementary Fig. 1). Although this complementary method offers a global assessment of resolution changes and reinforces the findings described for lateral resolution, it lacks the 3D spatial context that traditional spatial domain PSF fitting techniques provide.

Having successfully established the changing resolution behaviors observed using the QC microsphere sample, we then carefully imaged dendritic spine presynaptic puncta via fluorescently tagged AZ vGLUT proteins in cultured hippocampal neurons within the exact same FOV for each corresponding NA condition (Fig. 4). As with the QC sample, for this positive control experiment we expected to observe changes to the cell images that would correlate with the degraded resolution of the restricted NA conditions. In particular, since the immunostained neurons contain punctate spatial features which were small and comparable to the resolution limit associated with Objective Lens 1, the volumes of these features should correspondingly enlarge due to the resolution blurring induced by the emission pathway iris closing down. Alterations of the spine puncta shape and size were visually subtle and difficult to spot (again emphasizing the need for quantitative measurement and how our visual impressions of changing image features often *do not* correlate with quantitative variations in distance-based image measurements), but were discernible in the extreme case of Restricted NA 2, especially when examining zoomed-in, axial side views of individual puncta (Fig. 4f) or when spine puncta were surface rendered in 3D and presented side-by-side (Fig. 4g). Fourier-domain MTF analysis of these cellular images showed a decrease in the lateral cutoff frequency (Supplementary Fig. 2), mirroring the microsphere results (Supplementary Fig. 1) and confirming that stopping down the emission pathway iris also increased blur in the fixed sample images.

Fig. 4 | Changing dendritic spine puncta volumes of cultured neurons imaged under systematically reduced microscope resolution conditions. a, Full FOV Green channel maximum intensity Z-projection image of neurons with fluorescently tagged AZ vGLUT proteins acquired with Objective 1. Scale bar = 20 μm. **b-d**, Zoomed in views of a single dendritic process from the red boxed ROI in **a** for the Full NA, Restricted NA 1 and Restricted NA 2 emission pathway iris conditions respectively. **e**, Further digitally zoomed views of individual puncta from the yellow boxed ROIs in **b**. **f**, Side-view XZ intensity projections and **g**, 3D Imaris surface renderings of the same dendritic spine puncta from **e** under the various NA settings.

Comparing the same presynaptic puncta volume populations between the two extreme NA
conditions (Full NA and Restricted NA 2) showed that both could be described by a log-normal
distribution (fitted solid curves, Fig. 5a) with the majority of puncta having volumes in 0.150 to
0.300 fL range, consistent with previous reports imaging the same protein by cryo-electon
tomography and super-resolution stimulated emission depletion (STED) microscopy[72,73]. When
the iris was closed down to the Restricted NA 2 condition, we observed a measurable increase
in puncta volumes, but mostly biased towards the smaller puncta volumes, as indicated by the
shifted curve peak position between the two distributions. In conjunction with the model
presented in the Supplementary Text, the more pronounced volume growth of smaller puncta
makes sense: Resolution degradations mostly impact measurements of image features which
are comparable in size to the resolution limit. Beyond this size, the blurring effect on the
volumes of spherical objects having a diameter that exceed the resolution limit by ~2-3x is
negligible (see Equation S10 in Supplementary Text) and explains the distribution range of the
tail regions of both histograms appearing similar (Fig. 5a). Both the modal puncta volume of the
Full NA condition (Fig. 5b - 0.190 fL, vertical dashed line) and the peak distribution shift (19%,
horizontal dashed line) almost intersect perfectly with the theoretical curve, and the theoretical
curve itself passes roughly through the middle of the scattergram data points, thus
demonstrating good agreement between theory and measurement for this control experiment.

To visualize the size-dependent puncta inflation depicted in Fig. 5 and predicted by
Supplementary Equation S10, we simulated 3D images of fluorescent spheres convolved with
3D Gaussian ellipsoid-shaped PSFs of increasing width (reduced effective NA). The apparent
simulated volumes follow the analytic trend where features whose diameters are comparable to
the PSF extent expand markedly as blur increases, whereas objects larger than ~0.5 µm are
minimally affected (see Supplementary Fig. 3). This explains why the shift in Fig. 5a is
concentrated among smaller puncta and why the volume-change ratio in Fig. 5b decays with
initial size.

Our second resolution experiment replicated the imaging procedures of the control experiment.
but utilizing 3 similar objective lenses mounted on the same microscope instead of directly
manipulating the resolution of one objective lens. The three objectives were all Nikon 60× Plan
Apochromats with an NA of 1.40 (Objective 2 and 3) or 1.42 (Objective 1, Fig. 6). As before, we
used 200 nm sized fluorescent microspheres to characterize the PSF FWHM lateral and axial
extents associated with each objective lens and found they all exhibited relatively spatial
homogenous resolutions across the FOV, but with distinct and measurable differences across
objectives, reflecting the natural variations in objective lens quality (Fig. 7). The average lateral
resolutions of each objective lens were comparable and about 23-32% larger than their
theoretical values based on their operation wavelength (529 nm) and NA (1.40 for Objective 2
and 3, 1.42 for Objective 1). Objective 2 showed slightly larger axial fits and visual inspection of
the images showed an asymmetric "banana" shaped aberration in its PSF, possibly indicative of
damaged or defective internal lens elements (Fig. 6a and Fig. 7a). Surprisingly, Objective 3
possessed the best (smallest) overall lateral and axial FWHM, with measured resolution data
pairs clustering in a cloud closest to the 1-1 region of the FWHM scattergram (Fig. 7b and c),
even though Objective 1 had a slightly higher NA and was the latest generation model of this

lens (LambdaD vs. Lambda series). At the same time Objective 1's lateral and axial resolution
was more consistent across the full FOV (Fig. 7a). Interestingly, all three of these objective
lenses would be considered "acceptable" in terms of resolution performance according to the
QUAREP-LiMi recommendation standards.

Fig. 5 | Dendritic spine puncta volume inflation for the resolution control experiment. a, Histograms of measured puncta volumes for the Full NA and Restricted NA 2 conditions extracted from the full FOV neuron images presented in Fig. 4. **b**, Scattergram plot of puncta volume change ratio vs. initial Full NA puncta volume for every identified dendritic spine punctum. The solid black curve is the theoretical curve associated with these resolution-induced volume changes (Equation S11, Supplementary Text) for different sized spherical objects. The vertical dashed line shows the measured peak (mode) puncta volume of the Full NA condition (0.207 fL) while the horizontal dashed line is the distribution peak shift from Full NA to Restricted NA 2 (19%).

Fig. 6 | PSF comparison across three 60× oil immersion objective lenses imaging the exact same fluorescent microsphere. a, XY (top row) and XZ (bottom row) maximum intensity projections of the exact same 200 nm fluorescent microsphere imaged on the same microscope with three different 60× oil immersion objective lenses. Scale bar = 0.5 μm. **b**, Photographs of the objective lenses used for this study.

449 Again, lateral Fourier domain MTF analysis of the microsphere maximum intensity Z-projection
450 images validated these spatial domain PSF findings for the three objective lenses and showed
451 smaller resolution differences than we obtained in the control experiment by restricting the NA of
452 Objective 1 (Supplementary Fig. 4).

Fig. 7 Full FOV lateral and axial resolution characterization of Objective lenses 1-3. a, PSFj lateral (FWHM$_{max}$) and axial (FWHM$_z$) resolution heatmaps of Objective 1, 2, and 3. Scale bars = 20 µm. **b**, For each objective, the lateral and axial FWHM values of each isolated microsphere from the FOV were normalized by their theoretical values and plotted against each other. The green box delimiting the 1.5x theoretical resolution ratios match the recommended microscope resolution performance tolerance limits of Faklaris et al.[12] **c**, The centroids of each data "cloud" in **b** plotted the same way representing the average resolution performance across the entire FOV for each objective lens. **d**, Average FOV resolution FWHM measurements in tabular form.

The table in panel d:

	Objective 1 (NA 1.42)	Objective 2 (NA 1.4)	Objective 3 (NA 1.4)
Avg. FWHM max (nm)	250 +/-12	245 +/-4	237 +/-15
Avg. FWHM max/xy$_{theory}$	1.32 +/-0.06	1.27 +/-0.02	1.23 +/-0.08
Avg. FWHM z (nm)	612 +/-25	772 +/-29	578+/-12
Avg. FWHM z/z$_{theory}$	1.13 +/-0.05	1.35 +/-0.05	1.01 +/-0.02

Given the subtle differences when measuring well-defined microspheres, we expected that the

images of our biological specimen acquired with each objective lens would appear nearly

identical, and indeed the example neuron images shown in Fig. 8 are nearly impossible to

distinguish visually (although the banana-shaped PSF of Objective 2 does manifest itself more

clearly in the orthogonal and 3D surface rendered views of the smallest presynaptic puncta (Fig.
8f and g).

Fig. 8 | Comparative neuron imaging across Objectives 1-3. a, Whole FOV image of presynaptic immunolabelled vGLUT proteins in cultured hippocampal neurons from Objective 1. Scale bar = 20 µm. **b-d**, Zoom in view of red boxed ROI from panel A for all three objectives. **e-g**, Yellow boxed sub-ROIs 1-3 from panels **b-d** for each objective lens depicting individual vGLUT puncta shown laterally **e**, orthogonally **f**, and as surface rendered volumes **g**.

However, an equivalent statistical analysis of the measured puncta volumes observed in the
images acquired with each objective lens reveals that the volume distribution differences are
significant and correlate with the ranking of measured objective lens resolutions (Objective 3 <
Objective 1 < Objective 2), as illustrated in Fig. 9. Plotting all measured puncta volumes for
each objective, we were still able to observe measurable volume differences when comparing
the same central FOV with significant differences in Mann-Whitney U tests between Objective 1
and 3 and Objective 2 and 3, but not between Objective 1 and 2 (Fig. 9c). The results of this
second resolution experiment show again that apparent puncta volumes scale with the
instrument's PSF and motivate resolution-based instrument normalization of intra-channel
image readouts (using measured PSF FWHM) to ensure that measured feature differences
obtained across systems reflect biology rather than optical blur.

b

Objective	Peak Puncta Volume (fL)	Mean Puncta Volume (fL)	Variance (fL)
1	0.296	0.525	0.122
2	0.314	0.561	0.143
3	0.281	0.479	0.091

c

Comparison	Mann-Whitney U Statistic	P-Value
Objective 1 vs. Objective 2	239682.0	0.025
Objective 1 vs. Objective 3	286612.0	0.005
Objective 2 vs. Objective 3	391037.5	9.120 e-08

Fig. 9 | Objective-dependent differences in measured presynaptic vGLUT puncta volumes. a, Distributions of presynaptic puncta volumes extracted from the whole FOV shown in Fig. 8 across the 3 objectives. All three objectives showed similar distribution shapes. **b**, Statistical analysis of the modal (peak) and mean volumes and their variance for each objective. **c**, Mann-Whitney U pairwise tests of the three objective volume results, demonstrating that the measured differences are statistically significant in two of the three pairwise cases.

How channel co-registration impacts sub-cellular inter-channel image measurements of colocalization and volume overlap

To isolate the optical contribution to two-channel image measurement readouts, we first established where each objective lens was best registered and how much the channel mis-registration varied across the detector FOV. Full FOV chromatic aberration maps derived from two-color microsphere QC calibration tests showed that lateral chromatic shifts dominate and grow toward the periphery, while axial shifts for these three objective lenses were smaller and more uniform (Fig. 10a). These lateral/axial chromatic aberrations are unified into a single *scalar* Chromatic Displacement Index (CDI) ratio, χ – a channel pair co-registration QC metric that combines lateral and axial offsets after normalizing by the measured PSF widths (FWHM), so that χ reflects mis-registration in units of the system's actual resolving power (see Supplementary Text and Equation S13). This is in contrast to the original metric definition recommended by Faklaris *et al.* with the co-registration ratio applying the *theoretical* lateral and axial resolutions for normalization instead in an instrument-agnostic manner[12]. Switching the CDI normalization to the *measured* PSF FWHM (per-objective and per-FOV location) advantageously makes χ an instrument- and site-specific detection accurate QC metric. It displays whether the mis-registration is large relative to what the microscope actually resolved in that image dataset. This adjustment is also consistent with the quantitative framework presented in the Supplementary Text, which shows that two-channel colocalization/overlap readouts depend exponentially on inter-channel mis-registration using PSF-normalized units. A broader PSF lessens apparent registration penalties, whereas a sharper PSF amplifies them. Thus, measured PSF normalization aligns the metric with the true sensitivity of the acquisition rather than an ideal theoretical limit that the system may not reach (e.g., due to aberrations or depth-dependent blur), and it avoids over- or under-penalizing the same geometric shift across different instruments or ROIs. For continuity with prior practice, we also report the theory-normalized CDI co-registration ratio (χ_{theo}) spatial maps (using the shorter-wavelength channel theoretical FWHMs) in Supplementary Fig. 5. Theoretical resolution normalization values facilitate instrument/optical component benchmarking, whereas the measured-PSF normalization we adopt for modeling is a more appropriate scaling for cross-study comparisons, as well as estimating experimental detectability and colocalization bias in the image data actually acquired. Objective 3 maintained low χ over the largest area making it the best overall objective lens for inter-channel overlap image measurements, Objective 1 was intermediate, and Objective 2 exhibited a far off-center "sweet spot" of near-perfect registration with poorer regions elsewhere. These co-registration heatmap trends and observations were corroborated by the similarly measured PSF FWHM-normalized axial (r_z) vs. lateral (r_{xy}) channel displacement scatterplot (Fig. 10b). This plot again suggested the superiority of Objective 3 for colocalization studies since the centroid of the data cloud associated with this objective lens lay closest to the origin (which represents the state of perfect channel pair registration). In contrast, the centroids of the data clouds associated with Objective 1 and 2 were about equidistant from the scatterplot origin and indicated their differences in lateral versus axial co-registration performance. While such scatterplots are useful for rapid objective lens comparisons, they lack the spatial context that the CDI heatmaps provide, which is critical for choosing inter-channel analysis regions.

Guided by these co-registration QC heatmaps, we then acquired matched biological data from
the same fixed pre- and postsynaptic immunolabeled neuron at two positions: the FOV center
(ROI A1) and, after a motorized XY translation, the upper-left corner (ROI A2), repeating this
center/corner confocal Z-stack pair acquisition for all three objectives under identical settings
(Fig. 11). This experiment design held the biology constant while intentionally sampling different
local channel registration states known from the QC CDI χ heatmaps. Visually, the merged
vGLUT-Homer overlays suggested more overlap/colocalization at the image center than at the
corner (Fig. 12), consistent with the expectation that registration errors can modulate the
apparent colocalization or volume overlap.

Fig. 10 | Full field of view objective lens chromatic co-registration characterization with measured PSF normalization. a, Green-Yellow channel pair heatmaps derived from two-color microsphere QC calibrations for objective lenses 1-3. Top row: Absolute lateral chromatic displacement magnitude r_{xy} (nm) map. Middle row: Absolute axial chromatic displacement magnitude r_z (nm) map. Bottom row: Scalar CDI ratio χ (unitless) map, which combines lateral and axial shifts after normalization by the corresponding measured PSF FWHM values for the respective objective lens. The CDI ratio heatmaps use a different 16-step discretized color look-up-table (LUT) spanning a range of 0.2-1.0; Lower χ values (black < blue < green … < white) indicate better inter-channel registration. Scale bars = 20 μm. Across objectives, lateral offsets dominate and increase radially toward the periphery, whereas axial offsets are small and comparatively uniform. Objective 3 exhibits the lowest χ over the largest fraction of the FOV (extensive dark blue contoured zone, indicating superior co-registration performance; Objective 2 shows a pronounced region of near-perfect co-registration (black zone) that is significantly offset from the FOV center; Objective 1 is radially symmetric like Objective 3 with χ values still within the recommended tolerance over the entire FOV. These heatmaps enable users to select both the objective and the specific FOV sub-region that meets the precision required for a given experiment. **b,** Scatterplot of the axial (r_z) vs. lateral (r_{xy}) channel displacements of every measured microsphere within the FOV normalized by the respective objective lens' measured lateral and axial PSF FWHM (FWHM$_z$ and FWHM$_{xy}$). The shaded green box denotes the acceptance window (as recommended by Faklaris et al.[12]) where both normalized shifts are < 1. Microsphere channel displacements from Objective 3 cluster closest to the origin, consistent with panel **a,** identifying this objective as the best choice for colocalization studies among the three optics, while Objectives 1 and 2 remain usable since their data clouds are well within the tolerance zone as well.

Fig. 11 | **Central vs. edge two-color neuron images and ROI selection for colocalization analysis under spatially varying chromatic co-registration.** Single-channel maximum-intensity Z-projections from 3D confocal stacks of fixed neurons immunolabeled for vGLUT (**a**, Green channel, green LUT) and Homer (**b**, Yellow channel, magenta LUT), acquired with Objective 1. Scale bars = 20 µm. **c**, Color-merged overlay of panels **a**–**b** with cyan box labeled A1 marking the ROI used for quantitative colocalization analysis. **d**, The same neuron re-imaged after a motorized XY-stage translation that moved the neuron into the upper-left corner of the FOV. Both channels were re-acquired with identical settings. Cyan box labeled A2 denotes the ROI that was spatially identical to A1 in panel **c**. These ROIs were chosen to sample zones with different local chromatic co-registration quality (low-χ center vs higher-χ edge) as determined by QC-derived CDI co-registration heatmaps (see Fig. 10).

Fig. 12 | Visual effects of objective- and position-dependent co-registration on synaptic puncta overlap and apparent cleft orientation, with matched bead-QC orthogonal views (vGLUT, green LUT; Homer, magenta LUT; overlap appears white). a, Two-color confocal maximum intensity Z-projection overlay acquired with Objective 1; cyan boxes (a–c) mark sub-ROIs chosen for orthogonal inspection from ROI A shown in Fig. 11. Scale bar = 5 µm **b**, Close-ups of sub-ROI b from the center (A1b, left) and upper-left corner (A2b, right) of the FOV acquired with Objective 2. Cyan/orange arrowheads indicate the apparent pre- versus postsynaptic sides to show that the cleft orientation can change between positions, which is a consequence of differences in the local co-registration vector (direction/sign). Subtle differences in puncta overlap (white) are also evident with each objective lens and position. Scale bar = 1 µm **c**, For each sub-ROI pair (rows A1a/A2a, A1b/A2b, A1c/A2c), XY and XZ views of the synaptic puncta are shown for Objectives 1–3, alongside matched microsphere-QC XY/XZ profiles taken from the same FOV locations in the calibration data. Center ROIs generally show tighter apposition (more white) and more symmetric XZ profiles, whereas corner ROIs show increased channel separation and orientation bias that mirror the local QC measurements (see Fig. 10a).

Identical image pairs of the exact same neuron at the same two FOV positions as shown for
Objective 1 in Fig. 12 were also acquired with Objectives 2 and 3. These paired ROIs provide a
controlled test of how spatially varying co-registration affects vGLUT–Homer colocalization
readouts. Orthogonal views (as depicted in Fig. 12c) made these microscope channel
registration dependencies explicit at the single dendrite punctum scale. Sub-ROIs drawn from
A1 (center) and A2 (corner) revealed (i) visible changes in punctum overlap and, in some cases,
(ii) changes in apparent cleft orientation for the same pre/post punctum pair, emphasizing that
these effects were purely instrument-driven. For Objective 2 in particular, ROI center-to-corner
comparisons showed an orientation change of the vGLUT↔Homer vector and a reduced
colocalization/overlap. The matched microsphere QC views taken from the same FOV locations
with this objective exhibited the same axial "banana-shaped" XZ signature, directly linking the
biological appearance of the punctum to the objective's aberration behavior, as was presented
in Fig. 6a and Fig. 8.

These emergent biases of vGLUT-Homer protein complexes under the varying co-registration
imaging conditions align with the analytical framework developed in the Supplementary Text.
For covariance-like (e.g., Pearson's correlation coefficient, PCC) and geometric inter-channel
overlap type image measurements (ie: a typical 2-channel synaptogenesis assay), channel mis-
registration enters as a multiplicative exponential attenuation of the idealized "no-shift" or zero
channel shift value. The model predicts two robust behaviors: (i) For fixed object sizes, worse
co-registration (larger channel displacements offsets) decreases intensity product-based
(Hadamard) image measurements (PCC, volume overlap, etc.[74]), and (ii) measurement
sensitivity depends on object-to-PSF size via a size comparison factor, α. Resolution blur (larger
effective width, larger α) reduces the impact of a given channel shift, which is why an imaging
system with more resolution blur can *appear* to have higher colocalization for the same
underlying biology and registration error.

In Fig. 13, we quantify this experiment to model link by plotting the measured *change* in
colocalization between ROIs A1 and A2 against the measured change in local co-registration for
each objective lens, using Equation S23 (derived in the Supplementary Text). For each punctum
we computed a Dice symmetric volume-overlap colocalization score[75,76] and compared the
edge-to-center colocalization ratio to the change in co-registration ($\Delta\chi^2$) between ROIs A2 and
A1. We then overlaid theoretical predictions from our Gaussian overlap model using object-to-
PSF size ratios (α) derived from the measured PSF and puncta volumes. The analysis indicated
that across objectives the A1→A2 colocalization differences follow the trend extrapolated from
Equation S23 within error bars for all three lenses, consistent with the monotonic, size-
modulated attenuation predicted by the model. Reporting results with measured PSF
normalization (Fig. 13a) yielded the best alignment between theory and data and most faithfully
reflected instrument performance.

Together, Fig. 10 - Fig. 12 establish the instrumental origin of the observed changes in inter-
channel puncta overlap and orientation for the same neuron, and Fig. 13 provides the
quantitative, model-consistent calibration that ties local co-registration to the magnitude of
colocalization loss - another demonstration of how a critical microscope QC metric couples to
and influences a real sample image measurement. As was noted with intra-channel distance-
based measurements (e.g. volume), the QC metric matters most when the cross-channel
feature measurement is comparable in size to the QC metric value (lower α).

Fig. 13 | Inter-channel registration error predicts loss of puncta colocalization overlap. a, Dice (C) volume overlap coefficient ratio (C_{A2}/C_{A1}) versus change in co-registration error $\Delta(\chi^2)=\chi_{A2}^2-\chi_{A1}^2$ with measured PSF FWHM normalization. Here, C denotes the Sørensen–Dice overlap coefficient between the 3D binary masks of the green and yellow channel punctum for each matched object:

$$C \equiv \text{Dice} = \frac{2\,|G \cap Y|}{|G| + |Y|}$$

For each objective, 15 paired vGLUT-Homer puncta (A1↔A2; total N=45) visible in both channels were analyzed. Data points show the median of the per-punctum ratios and vertical uncertainty bars are bootstrap 95% confidence intervals for the median, while horizontal uncertainty bars are standard errors (SE) for $\Delta(\chi^2)$ propagated from the ROI-level uncertainties. Scalar co-registration χ values were computed as the mean +/- standard deviation (SD) of pixel values in the co-registration heatmaps of Fig. 10a, restricted to the same A1/A2 ROIs used in Fig. 12c–d. These ROI means (and SDs) feed $\Delta(\chi^2)$ and its SE. Dotted black curves show the Equation S23 reference family curves for different object to PSF size ratios for α={0.5,1,1.5,2,3,4}. Dashed colored curves are per-objective fits of α. **b**, The same analysis as in panel **a**, but with theoretical PSF FWHM normalization. Replacing measured PSFs with theoretical values shifts the points rightward (larger $\Delta(\chi^2)$), and the fitted α values correspondingly increase (two lenses move to α~1.1–1.2; one remains <1).

Discussion

A central aim of this study was to demonstrate directly and unambiguously that routine light-microscope QC metrics are not merely ancillary indicators of instrument "health" or guardrails for data consistency, but also determinants of quantitative image readouts. Since QC metrics set the scales at which features are sampled, changes in a QC value propagate predictably into the reported measurement with the direction and magnitude of the bias governed by how the feature compares to the metric (e.g., object size relative to the PSF; inter-channel displacement relative to object extent/overlap). This transformation of the data is bidirectional: biological inferences drawn from an image readout are valid and interpretable *only to the extent that the relevant QC metric remains stable over the measurement interval*. We showed this link concretely for two archetypal metric–measurement pairs: Image resolution and intra-channel, distance-based morphometrics (e.g., volume) and inter-channel co-registration and covariance-like and overlap-based indicators of colocalization.

In our constricted objective NA, resolution-reducing control experiment, the apparent presynaptic puncta volumes intuitively increased as the resolution of the system decreased. A more important question is if these correlations could have been predicted based on the observed changes in the QC samples alone. A 3D Gaussian convolution model of the microscope PSF which depends on the variance-addition properties of this function is presented in the Supplementary Text and explains how intra-channel distance-based image measurements (feature volume in this case) of punctate dendritic spines *intrinsically depend* on the system resolution, *but in a size-dependent manner*. Practically, even a minor broadening of the PSF, or a small optical misalignment that widens the effective blur made these objects appear >30% larger than the Full NA, well-aligned conditions, even for structures larger than the diffraction limit. Such inflation is easy to miss visually but can flip or confound experimental conclusions if the instrument state drifts between conditions. The three 60× objective lenses we tested differed only subtly and all of them would have passed common resolution tolerance cut-offs (e.g., FWHM-based thresholds recommended in recent community guidelines), underscoring that "within tolerance" does not mean "measurement-equivalent". Routine reporting of resolution via a simple FWHM from sub-diffraction sized fluorescent microspheres provides the minimal calibration needed to interpret and compare volume measurements across sessions and instruments.

We note that Objective 2, despite its clear inherent aberration, could be perfectly acceptable for certain imaging scenarios and depending on the nature of the experiment. As we showed in the control experiment, the measurement of volume is affected by characteristics of the PSF only when the specimen features under study are similar in size/separation distance to the PSF itself. In contrast, when the relevant structures under study are several PSF widths across (or separated by many PSF widths), the intensity signal is dominated by broad, slowly varying features that are essentially unchanged by modest blur, and accordingly, differences in resolution or mild aberrations have negligible impact on the readout (assuming adequate digital sampling and signal to noise ratio).

The same normalization approach can be applied to any intra-channel, distance-based image measurement and is not limited to neuronal cell structures. Many cell-biological readouts from fluorescence microscopy like organelle morphometry (length, area, volume, cluster size, and other spatial statistics such as nearest-neighbor/centroid separations) of mitochondria, lysosomes, endosomes, focal adhesions, clathrin-coated pit sizes, nuclear foci/condensates, etc. depend on fluorescence segmentation and are therefore susceptible to PSF-induced size inflation and sampling biases. A recommended and lightweight best practice then is to pair PSF calibration with the same variance-addition modelling used here to compute lens- and session-specific correction factors. This "instrument bias normalization" improves accuracy without reprocessing images and offers an alternative to full deconvolution (which can be time-consuming, expensive, and assumption-laden) while simultaneously avoiding over-sharpening artefacts and deconvolution pipeline dependencies. Because the model is explicit in PSF units, one can also re-project a measured volume/area/length/etc. into the PSF of another system or time point provided steps are taken to determine the object/feature to PSF size ratio (α) by a resolution/co-registration perturbation through some auxiliary experiment (like we have demonstrated here), or if α is known/estimated through some other empirical means (electron microscopy or super-resolution imaging for example). This allows a researcher to ask the question, "What would I have measured on a microscope with PSF X instead of Y?" This re-projection of data into the characteristics of another microscope enables cross-system comparability and corrects the very biases that would otherwise confound quantitative fluorescence measurements, without altering the underlying images. Indeed, in our second set of resolution experiments where the exact same fluorescent neuronal synapses were imaged with three different objective lenses with all other things being equal, we measured three distinct puncta distributions, again stressing that the measurement only makes sense and converges to agreement when cited along with or corrected by the measured system resolution characteristics according to the instrument bias normalization approach outlined in the Supplementary Text.

As a theoretical illustration of the proposed normalization practice, suppose Lab 1 images vGLUT puncta with a 60×, 1.42 NA objective lens and Lab 2 uses a 100×, 1.30 NA objective. Both labs segment volumes from the raw fluorescence. Reported medians differ by ~15%, leaving reviewers to juggle two volume lists and two PSF tables to guess whether the data gap is biological or optical in origin. Re-projecting both datasets to a common reference PSF (e.g., the measured PSF from Lab 1, or a community "standard PSF") removes the instrument term analytically using the closed-form inflation factors in the Supplementary Text (variance-addition formulas for length/area/volume under isotropic or anisotropic PSFs). If, after re-projection, the two volume distributions still exhibit a residual difference, that difference can confidently be identified as biological in origin (within the sampling error) rather than instrumental. Crucially, this is achieved without altering images: The correction applies to the numbers using each dataset's measured $FWHM_{xy,z}$, and the results become directly comparable across experimental setups, time, and independent labs. By coupling every quantitative intra-channel image measurement to its intrinsically related microscope QC metrics, the reporting of imaging results then becomes explicit and FAIR (findable, accessible, interoperable, and reusable)[77,78], enabling reviewers and downstream users to compare effects across microscopes.

Across our dual-channel datasets we observed that the exact same pre- and postsynaptic
punctum, imaged with different objective lenses or at different FOV locations, could show
measurable changes in apparent pre→post volume overlap. These shifts co-varied with the QC
scalar maps of local chromatic co-registration χ. In this purposefully arranged experiment, we
also confirmed that modest, spatially varying offsets could reorient the AZ↔PSD vector without
any biological change, rendering both the true cleft orientation and protein colocalization
ambiguous in the absence of related QC information, as well as potentially causing biased
downstream interpretations of this synaptogenesis assay. Again, the approach to disentangle
inter-channel instrumental dependencies from real sample features is not limited to neuronal
cells. Applying Δχ-derived corrections (or constraining analysis to low-|Δχ| regions) stabilizes
orientation and overlap image readouts across objectives and positions. This behavior matches
the 2-channel Gaussian PSF overlap model presented in the Supplementary Text (Equations
S20 and S23). In PSF-normalized units, mis-registration enters as a predictable, multiplicative
exponential on the idealized no-shift inter-channel measurement,

664
$$\mathcal{I}_{meas}(x, y) = \mathcal{I}_0 \exp \left[-f(\alpha) \, \chi(x, y)^2 \right]$$

Thus, once χ is mapped and the object-to-PSF scale α is known or reasonably estimated via
complementary experimental techniques, one can *a priori* forecast precision loss for any
overlap/colocalization readout. As an illustrative back-of-the-envelope estimate, a ~150 nm
chromatic shift dampens the overlap of diffraction-limited synaptic puncta by ≈40%, but by only
≈6% for structures ~3× larger than the PSF core. In this sense, QC co-registration heatmaps act
like transfer functions that guide ROI selection and objective choice, making cross-system
experimental comparisons interpretable when Δχ and α are reported.

A striking implication of the model is that a microscope with broader PSF (due to aberrations or
imperfect alignment) can show higher apparent two-channel overlap/colocalization for the *same*
underlying registration error and biology. That is not a virtue of the instrument, but rather a
manifestation of the α term in the equation above. The same $\Delta\chi^2$ produces a smaller mis-
registration penalty when the effective width of the feature is larger relative to the PSF (larger α).
Intuitively, a system experiencing blur due to degraded resolution acts as a low-pass filter,
acting to smear intensity variations such that a fixed physical mis-registration is less
visible/quantifiable to overlap/covariance-type measurements. Therefore, without accounting for
PSF and object size, cross-system or cross-day comparisons can be biased and confounding.
The remedy is not to accept blur, but to measure it and either standardize it across datasets or
explicitly correct for its effect using the model. Although most microscopists conducting research
at the cellular level likely understand this effect, its quantitative connection to the microscope's
co-registration QC metric which we have demonstrated here is under-appreciated.

Our analytic links between QC and measurement rest on two simplifications. First, we treat both
the PSF and the effective object as Gaussian distributions that are separable across axes. In
most day-to-day conditions this is a useful and reasonable approximation since optical
aberrations primarily alter the pre-exponential factor without changing the exponential
dependence on $\Delta(\chi^2)$. However, strong PSF asymmetries, field tilt, or pronounced axial–lateral

coupling would warrant a more complex axis-resolved treatment (for example by replacing the
scalar form with lateral/axial terms, e.g., $f(\alpha_{xy})$, $f(\alpha_z)$, and use a vectorized registration field rather
than a single scalar χ value). Second, we analyzed ROIs where the biological offset between
channels, δ, averages to ~0 (e.g., pre/post markers centered across the cleft).. When δ≠0
because of orientated structures, the observed overlap/colocalization reflects the combined
separation |**δ**±**Δr**|. In those cases and depending on the geometry, microscope channel mis-
registration can mask true offsets, or amplify them. The model framework generalizes by adding
δ to the mis-registration term and, for mixed populations, modeling the distribution of **δ**. In
practice, reporting lateral and axial PSF extents, the vectorized co-registration map, and any
known biological offsets alongside α and $\Delta(\chi^2)$ keeps interpretations of these more complex
imaging scenarios rational.

Conclusion

Our study reframes routine microscope QC as determinants of quantitative readouts rather than
just auxiliary health checks of an instrument. We derived analytical links from measured PSF
blur and chromatic mis-registration to the bias they induce in common image measurements
(volume, overlap/colocalization). We also showed that colocalization signals depend
exponentially on inter-channel mis-registration in PSF-normalized units, and that normalizing by
the measured PSF instead of the theoretical limit is both necessary and sufficient to bring data
and theory into agreement across objectives and FOV positions. We described an instrument-
normalization / re-projection workflow that removes the instrument contributions to these
measurement errors, allowing one to ask, *"How would this dataset change and appear on a*
microscope with PSF and co-registration X instead of Y?" without altering or transforming any of
the acquired images. Together these elements convert QC from a passive checklist of
instrument health into a predictive design tool for setting tolerances, choosing
acquisition/analysis ROIs, and making cross-system comparisons that are genuinely biological.

Practically, the recommended application of QC metrics into an experimental workflow is simple.
First measure lateral/axial PSFs and the co-registration across the FOV using fluorescent
microspheres and report these metrics alongside any related image analysis, and—when object
scale relative to the PSF (α) is known or estimated—apply the closed-form corrections or re-
project to a declared reference PSF size/co-registration level. Where α is not directly available,
brief auxiliary calibrations or transparent computational surrogates could suffice. In this manner,
QC-aware normalization turns instrument variability from a confound into a controllable variable,
improving interpretability, reproducibility, and comparability of fluorescence microscopy data at
scale.

Finally, we note that the principles of this QC normalization strategy also align closely with the
"reverse-thinking" framework proposed by Wait *et al*. for hypothesis-driven quantitative
microscopy[79]. Their recommended approach for using light microscopy in research emphasizes
beginning with a clearly articulated hypothesis, defining the specific data needed to test it, and
then determining the experimental parameters and microscope configuration required to
generate those data. By providing objective, quantitative metrics of instrument performance

along with recorded image readouts, our suggested instrument normalization practice fits
naturally within the experimental design phase of this framework and enables investigators to
select imaging modalities and acquisition/analysis settings that are properly matched to the
analytical image measurements required to test their hypotheses. Moreover, the central insight
of this work is that image-derived biological measurements are not independent of instrument
performance; rather, they are inherently constrained by measurable microscope QC
parameters. Here, we have explicitly demonstrated these relationships for distance-based
image morphometry and inter-channel overlap and colocalization measurements, but the
concept is general. Many common quantitative image measurements can, in principle, be linked
mathematically back to a corresponding or related opto-mechanical microscope performance
QC parameter. Making these dependencies explicit provides a experimental mindset that builds
performance validation into experimental design, so biological conclusions rely on measured,
instead of assumed, imaging capability.

References

1. Yuste, R. Fluorescence microscopy today. *Nat. Methods* **2**, 902–904 (2005).

2. Lichtman, J. W. & Conchello, J.-A. Fluorescence microscopy. *Nat. Methods* **2**, 910–919

(2005).

3. Piston, D. W. The impact of technology on light microscopy. *Nat. Cell Biol.* **11**, S23–S24

(2009).

4. Swedlow, J. R. What's next for 'just having a look'? *Nat. Cell Biol.* **11**, S25–S26 (2009).

5. Waters, J. C. Accuracy and precision in quantitative fluorescence microscopy. *J. Cell Biol.*

185, 1135–1148 (2009).

6. Yamada, K. M. & Hall, A. Reproducibility and cell biology. *J. Cell Biol.* **209**, 191–193 (2015).

7. Baker, M. 1,500 scientists lift the lid on reproducibility. *Nature* **533**, 452–454 (2016).

8. Lee, J.-Y. & Kitaoka, M. A beginner's guide to rigor and reproducibility in fluorescence

imaging experiments. *Mol. Biol. Cell* **29**, 1519–1525 (2018).

9. Jost, A. P.-T. & Waters, J. C. Designing a rigorous microscopy experiment: Validating

methods and avoiding bias. *J. Cell Biol.* **218**, 1452–1466 (2019).

10. Montero Llopis, P. *et al.* Best practices and tools for reporting reproducible fluorescence

microscopy methods. *Nat. Methods* **18**, 1463–1476 (2021).

11. Nelson, G. *et al.* QUAREP-LiMi: A community-driven initiative to establish guidelines for

quality assessment and reproducibility for instruments and images in light microscopy. *J.*

Microsc. **284**, 56–73 (2021).

12. Faklaris, O. *et al.* Quality assessment in light microscopy for routine use through simple

tools and robust metrics. *J. Cell Biol.* **221**, e202107093 (2022).

13. Cole, R. W., Jinadasa, T. & Brown, C. M. Measuring and interpreting point spread functions

to determine confocal microscope resolution and ensure quality control. *Nat. Protoc.* **6**,

1929–1941 (2011).

14. Theer, P., Mongis, C. & Knop, M. PSFj: know your fluorescence microscope. *Nat. Methods*

11, 981–982 (2014).

15. Bolte, S. & Cordelières, F. P. A guided tour into subcellular colocalization analysis in light

microscopy. *J. Microsc.* **224**, 213–232 (2006).

16. Cordelières, F. P. & Bolte, S. Experimenters' guide to colocalization studies: Finding a way

through indicators and quantifiers, in practice. in *Methods in Cell Biology* vol. 123 395–408

(Elsevier, 2014).

17. Mascalchi, P. & Cordelières, F. P. Which Elements to Build Co-localization Workflows?

From Metrology to Analysis. in *Computer Optimized Microscopy* (eds Rebollo, E. & Bosch,

M.) vol. 2040 177–213 (Springer New York, New York, NY, 2019).

18. Inoué, S. Foundations of Confocal Scanned Imaging in Light Microscopy. in *Handbook Of*

Biological Confocal Microscopy (ed. Pawley, J. B.) 1–19 (Springer US, Boston, MA, 2006).

doi:10.1007/978-0-387-45524-2_1.

19. Amos, B., McConnell, G. & Wilson, T. 2.2 Confocal Microscopy. in *Comprehensive*

Biophysics 3–23 (Elsevier, 2012). doi:10.1016/B978-0-12-374920-8.00203-4.

20. Wilson, T. Resolution and optical sectioning in the confocal microscope. *J. Microsc.* **244**,

113–121 (2011).

21. Sarder, P. & Nehorai, A. Deconvolution methods for 3-D fluorescence microscopy images.

IEEE Signal Process. Mag. **23**, 32–45 (2006).

22. Gustafsson, M. G. L. Surpassing the lateral resolution limit by a factor of two using

structured illumination microscopy. *J. Microsc.* **198**, 82–87 (2000).

23. Gustafsson, M. G. L. *et al.* Three-Dimensional Resolution Doubling in Wide-Field

Fluorescence Microscopy by Structured Illumination. *Biophys. J.* **94**, 4957–4970 (2008).

24. Hell, S. W. & Wichmann, J. Breaking the diffraction resolution limit by stimulated emission:

stimulated-emission-depletion fluorescence microscopy. *Opt. Lett.* **19**, 780 (1994).

25. Willig, K. I., Rizzoli, S. O., Westphal, V., Jahn, R. & Hell, S. W. STED microscopy reveals

that synaptotagmin remains clustered after synaptic vesicle exocytosis. *Nature* **440**, 935–

939 (2006).

26. Betzig, E. *et al.* Imaging Intracellular Fluorescent Proteins at Nanometer Resolution.

Science **313**, 1642–1645 (2006).

27. Rust, M. J., Bates, M. & Zhuang, X. Sub-diffraction-limit imaging by stochastic optical

reconstruction microscopy (STORM). *Nat. Methods* **3**, 793–796 (2006).

28. Jungmann, R. *et al.* Single-Molecule Kinetics and Super-Resolution Microscopy by

Fluorescence Imaging of Transient Binding on DNA Origami. *Nano Lett.* **10**, 4756–4761

(2010).

29. Jungmann, R. *et al.* Multiplexed 3D cellular super-resolution imaging with DNA-PAINT and

Exchange-PAINT. *Nat. Methods* **11**, 313–318 (2014).

30. Kanchanawong, P. *et al.* Nanoscale architecture of integrin-based cell adhesions. *Nature*

468, 580–584 (2010).

31. Xu, K., Zhong, G. & Zhuang, X. Actin, Spectrin, and Associated Proteins Form a Periodic

Cytoskeletal Structure in Axons. *Science* **339**, 452–456 (2013).

32. Tang, A.-H. *et al.* A trans-synaptic nanocolumn aligns neurotransmitter release to receptors.

Nature **536**, 210–214 (2016).

33. Szymborska, A. *et al.* Nuclear Pore Scaffold Structure Analyzed by Super-Resolution

Microscopy and Particle Averaging. *Science* **341**, 655–658 (2013).

34. Manders, E. M. M., Verbeek, F. J. & Aten, J. A. Measurement of co-localization of objects in

dual-colour confocal images. *J. Microsc.* **169**, 375–382 (1993).

35. Zinchuk, V., Zinchuk, O. & Okada, T. Quantitative Colocalization Analysis of Multicolor

Confocal Immunofluorescence Microscopy Images: Pushing Pixels to Explore Biological

Phenomena. *ACTA Histochem. Cytochem.* **40**, 101–111 (2007).

36. Oheim, M. & Li, D. Quantitative Colocalisation Imaging: Concepts, Measurements, and

Pitfalls. in *Imaging Cellular and Molecular Biological Functions* (eds Shorte, S. L. &

Frischknecht, F.) 117–155 (Springer Berlin Heidelberg, Berlin, Heidelberg, 2007).

doi:10.1007/978-3-540-71331-9_5.

37. Seward, G. H. *Optical Design of Microscopes*. (SPIE, 1000 20th Street, Bellingham, WA

98227-0010 USA, 2010). doi:10.1117/3.855480.

38. Schmidt, K., Guo, N., Wang, W., Czarske, J. & Koukourakis, N. Chromatic aberration

correction employing reinforcement learning. *Opt. Express* **31**, 16133 (2023).

39. Thevenaz, P., Ruttimann, U. E. & Unser, M. A pyramid approach to subpixel registration

based on intensity. *IEEE Trans. Image Process.* **7**, 27–41 (1998).

40. Sorzano, C. O. S., Thevenaz, P. & Unser, M. Elastic Registration of Biological Images Using

Vector-Spline Regularization. *IEEE Trans. Biomed. Eng.* **52**, 652–663 (2005).

41. Arganda-Carreras, I. *et al.* Consistent and Elastic Registration of Histological Sections Using

Vector-Spline Regularization. in *Computer Vision Approaches to Medical Image Analysis*

(eds Beichel, R. R. & Sonka, M.) vol. 4241 85–95 (Springer Berlin Heidelberg, Berlin,

Heidelberg, 2006).

42. Matsuda, A., Schermelleh, L., Hirano, Y., Haraguchi, T. & Hiraoka, Y. Accurate and fiducial-

marker-free correction for three-dimensional chromatic shift in biological fluorescence

microscopy. *Sci. Rep.* **8**, 7583 (2018).

43. Leiwe, M. N., Fujimoto, S. & Imai, T. Post hoc Correction of Chromatic Aberrations in Large-

Scale Volumetric Images in Confocal Microscopy. *Front. Neuroanat.* **15**, 760063 (2021).

44. López, A. C., Conesa, P., Oña Blanco, A. M., Gómez-Pedrero, J. A. & Sorzano, C. O. S.

Real-time correction of chromatic aberration in optical fluorescence microscopy. *Methods*

Appl. Fluoresc. **11**, 045001 (2023).

45. Hodgson, L., Shen, F. & Hahn, K. Biosensors for Characterizing the Dynamics of Rho

Family GTPases in Living Cells. *Curr. Protoc. Cell Biol.* **46**, (2010).

46. Nelson, G. *et al.* QUAREP-LiMi: A community-driven initiative to establish guidelines for

quality assessment and reproducibility for instruments and images in light microscopy. *J.*

Microsc. **284**, 56–73 (2021).

47. Boehm, U. *et al.* QUAREP-LiMi: a community endeavor to advance quality assessment and

reproducibility in light microscopy. *Nat. Methods* **18**, 1423–1426 (2021).

48. Monitoring the point spread function for quality control of Fluorescent microscopes v2.

protocols.io https://www.protocols.io/view/monitoring-the-point-spread-function-for-quality-c-

d4pn8vme.

49. Ensuring accurate co-registration measurement for quality control of Single Point Confocal

Laser Scanning Microscopes. *protocols.io* https://www.protocols.io/view/ensuring-accurate-

co-registration-measurement-for-cgqttvwn.

50. Harris, K. M. & Weinberg, R. J. Ultrastructure of Synapses in the Mammalian Brain. *Cold*

Spring Harb. Perspect. Biol. **4**, a005587 (2012).

51. Südhof, T. C. The Presynaptic Active Zone. *Neuron* **75**, 11–25 (2012).

52. Nair, D. *et al.* Super-Resolution Imaging Reveals That AMPA Receptors Inside Synapses

Are Dynamically Organized in Nanodomains Regulated by PSD95. *J. Neurosci.* **33**, 13204–

13224 (2013).

53. MacGillavry, H. D., Song, Y., Raghavachari, S. & Blanpied, T. A. Nanoscale Scaffolding

Domains within the Postsynaptic Density Concentrate Synaptic AMPA Receptors. *Neuron*

78, 615–622 (2013).

54. Tang, A.-H. *et al.* A trans-synaptic nanocolumn aligns neurotransmitter release to receptors.

Nature **536**, 210–214 (2016).

55. Verstraelen, P. *et al.* Systematic Quantification of Synapses in Primary Neuronal Culture.

iScience **23**, 101542 (2020).

56. Ippolito, D. M. & Eroglu, C. Quantifying Synapses: an Immunocytochemistry-based Assay to

Quantify Synapse Number. *J. Vis. Exp.* 2270 (2010) doi:10.3791/2270.

57. Perez de Arce, K. *et al.* Topographic Mapping of the Synaptic Cleft into Adhesive

Nanodomains. *Neuron* **88**, 1165–1172 (2015).

58. Prange, O. & Murphy, T. H. Correlation of Miniature Synaptic Activity and Evoked Release

Probability in Cultures of Cortical Neurons. *J. Neurosci.* **19**, 6427–6438 (1999).

59. Goddard, C. A., Butts, D. A. & Shatz, C. J. Regulation of CNS synapses by neuronal MHC

class I. *Proc. Natl. Acad. Sci.* **104**, 6828–6833 (2007).

60. Turrigiano, G. G. The Self-Tuning Neuron: Synaptic Scaling of Excitatory Synapses. *Cell*

135, 422–435 (2008).

61. Toomre, D. & Pawley, J. B. Disk-Scanning Confocal Microscopy. in *Handbook Of Biological*

Confocal Microscopy (ed. Pawley, J. B.) 221–238 (Springer US, Boston, MA, 2006).

doi:10.1007/978-0-387-45524-2_10.

62. Oreopoulos, J., Berman, R. & Browne, M. Spinning-disk confocal microscopy: present

technology and future trends. in *Methods in Cell Biology* vol. 123 153–175 (Elsevier, 2014).

63. Royon, A. & Converset, N. Quality Control of Fluorescence Imaging Systems: A new tool for

performance assessment and monitoring. *Opt. Photonik* **12**, 22–25 (2017).

64. Macchi, P. *et al.* A GFP-based System to Uncouple mRNA Transport from Translation in a

Single Living Neuron.

65. Goetze, B. *et al.* The brain-specific double-stranded RNA-binding protein Staufen2 is

required for dendritic spine morphogenesis. *J. Cell Biol.* **172**, 221–231 (2006).

66. Kohli, A. *et al.* Ring deconvolution microscopy: exploiting symmetry for efficient spatially

varying aberration correction. *Nat. Methods* **22**, 1311–1320 (2025).

67. Tian, L. & Waller, L. Quantitative differential phase contrast imaging in an LED array

microscope. *Opt. Express* **23**, 11394 (2015).

68. Williams, C. S. & Becklund, O. A. *Introduction to the Optical Transfer Function.* (SPIE,

2002). doi:10.1117/3.2265056.

69. Rowlands, C. J., Bruns, O. T., Bawendi, M. G. & So, P. T. C. Objective, comparative

assessment of the penetration depth of temporal-focusing microscopy for imaging various

organs. *J. Biomed. Opt.* **20**, 061107 (2015).

70. Wang, H. *et al.* Deep learning enables cross-modality super-resolution in fluorescence

microscopy. *Nat. Methods* **16**, 103–110 (2019).

71. Zhang, Y. *et al.* Virtual staining of label-free tissue in imaging mass spectrometry. *Sci. Adv.*

11, eadv0741 (2025).

72. Siksou, L. *et al.* Three-Dimensional Architecture of Presynaptic Terminal Cytomatrix. *J.*

Neurosci. **27**, 6868–6877 (2007).

73. Nishimune, H., Badawi, Y., Mori, S. & Shigemoto, K. Dual-color STED microscopy reveals a

sandwich structure of Bassoon and Piccolo in active zones of adult and aged mice. *Sci.*

Rep. **6**, 27935 (2016).

74. Gonzalez, R. & Woods, R. *Digital Image Processing.* (Pearson, New York, NY, 2018).

75. Zou, K. H. *et al.* Statistical validation of image segmentation quality based on a spatial

overlap index1. *Acad. Radiol.* **11**, 178–189 (2004).

76. Taha, A. A. & Hanbury, A. Metrics for evaluating 3D medical image segmentation: analysis,

selection, and tool. *BMC Med. Imaging* **15**, 29 (2015).

77. Wilkinson, M. D. *et al.* The FAIR Guiding Principles for scientific data management and

stewardship. *Sci. Data* **3**, 160018 (2016).

78. Sarkans, U. *et al.* REMBI: Recommended Metadata for Biological Images—enabling reuse

of microscopy data in biology. *Nat. Methods* **18**, 1418–1422 (2021).

79. Wait, E. C., Reiche, M. A. & Chew, T.-L. Hypothesis-driven quantitative fluorescence

microscopy – the importance of reverse-thinking in experimental design. *J. Cell Sci.* **133**,

(2020).

Competing Interests

J.O., C.L.M., A.B. and B.G. are employees of Andor Technology (Oxford Instruments), the
manufacturer of the BC43 microscope used in this work. M.G. is an employee of Oxford
Instruments Imaris (Bitplane/Imaris), developer of the image-analysis software used in this
study. Andor/Oxford Instruments offers an Installation Qualification/Operational Qualification
(IQ/OQ) quality-control service program for BC43 microscopes. Since this manuscript examines
QC metrics and workflows relevant to BC43 performance, we disclose this potential conflict.
G.N., S.T., and M.K. declare no competing interests.

949   *Supplementary Fig. 1*

950   *Supplementary Fig. 2*

951   *Supplementary Fig. 3*

952   *Supplementary Fig. 4*

953   *Supplementary Fig. 5*