1

Deep Learning Enables Automated Segmentation and Quantification of Ultrastructure from

Transmission Electron Microscopy Images

Anqi Zou[1,#], Winston Tan[2,#], Jiayi Ji[2,#], Florencia Rojas-Miguez[3], Laura Dodd[2], Emily Oei[2], Sarah

R. Vargas[4], Stephen P. Berasi[4], Hongying Yang[4], Hui Chen[3], Joel M. Henderson[3,5,6], Xueping

Fan[2,*], Weining Lu[2,3,*], Chao Zhang[1,*]

[1]Computational Biomedicine Section, Department of Medicine, Boston University Chobanian &

Avedisian School of Medicine, Boston, MA, USA

[2]Nephrology Section, Department of Medicine, Boston University Chobanian & Avedisian

School of Medicine, Boston Medical Center, Boston, MA, USA

[3]Department of Pathology and Laboratory Medicine, Boston University Chobanian & Avedisian

School of Medicine, Boston Medical Center, Boston, MA, USA

[4]Global Pathology and Centers for Therapeutic Innovation, Pfizer, Inc., Groton, CT, and

Cambridge, MA, USA.

[5]Department of Pathology, Beth Israel Deaconess Medical Center, Boston, MA, USA

[6]Department of Pathology, Harvard Medical School, Boston, MA, USA.

[#] These authors contributed equally to this work

[*] Corresponding authors

E-mail addresses: chz2009@bu.edu (C. Zhang), wlu@bu.edu (W. Lu), xpfan@bu.edu (X. Fan)

Abstract

Transmission electron microscopy (TEM) has become an essential technique for observing

subcellular ultrastructure, and is widely used in both clinical diagnosis and biomedical research.

However, analysis of TEM data remains extremely labor-intensive and often inconsistent across

operators due to the lack of dedicated computational methods. Here, we present TEAMKidney, a

deep learning framework for accurate and scalable measurement of ultrastructures in TEM images

across species, magnifications, and instrument platforms. We collected 12,991 TEM images from

patients with multiple kidney diseases and from different animal models. By combining a self-

training-based semantic segmentation stage with a TEM-tailored panoptic segmentation model,

we address two major challenges in TEM data analysis: the lack of accurately labeled training data

and the difficulty of achieving high segmentation accuracy for complex ultrastructure. Application

of TEAMKidney to both human and animal images successfully reveals disease-associated

changes in two critical glomerular ultrastructures: the glomerular basement membrane and

podocyte foot processes. In addition to significantly outperforming existing tools, TEAMKidney

shows close agreement with pathological expert measurements used in clinical assessment

protocols. By reducing dependence on manual tracing while preserving expert-level accuracy,

TEAMKidney demonstrates that deep learning can substantially reduce the burden of image

analysis in both clinical pathology and biomedical research settings.

Introduction

Transmission electron microscopy (TEM) provides high-resolution visualization of cellular ultrastructure beyond the resolving capabilities of light microscopy, immunofluorescence (IF) staining, immunohistochemistry, or molecular assays[1, 2] (Fig 1A). Beyond its extensive use in biomedical research and drug discovery, TEM remains an important diagnostic tool in many clinical areas, including infectious diseases, oncology, ophthalmology, neurology, and especially renal pathology[3-5].

Chronic kidney disease (CKD) affects over 37 million people in the US[6] and 850 million worldwide[7, 8]. It has been recognized as a leading public health burden in the US. By 2027, the direct healthcare costs for kidney diseases in the US were projected to reach \$173.4 billion[9]. Diabetic kidney disease and glomerular disease account for more than 50% of CKD cases[10, 11] and are the leading cause of kidney failure[12, 13]. These disorders encompass a broad spectrum of conditions characterized by distinct patterns of glomerular ultrastructural damage, which are associated with different therapeutic strategies and clinical outcomes. In the past decade, TEM has become an essential and often decisive component of the diagnostic modalities for kidney diseases[14]. Some major nephrology and pathology centers now utilize TEM as a routine diagnostic tool, applying it to more than 80% renal biopsies[14, 15].

Despite its value, significant barriers limit the broader adoption of TEM. The technique remains highly labor-intensive[14] and producing a full pathology report can take up to 26 working days on average[5], primarily due to the need for expert renal pathologists to analyze the ultrastructure, such as glomerular basement membrane (GBM) and podocyte foot processes (PFP). Measurement of GBM and PFP widths is widely used for both research and clinical diagnosis, yet researchers and renal pathologists still rely primarily on manual measurement of TEM images (Fig

1B). This process is time-consuming and restricts the advancement of CKD research and clinical

diagnosis.[16-19]. Manual measurement also introduces significant inter-operator variation and bias[20-

22]. Due to the inherent characteristics of TEM data, such as grayscale-only imaging, low contrast,

and fuzzy boundaries between ultrastructural features, which make conventional artificial

intelligence (AI) models struggle to delineate fine details accurately. Currently, no available tools

achieve results as accurate as manual labeling (Fig 1C).

During the past decade, semi-automated methods have been piloted for GBM and PFP

measurements[20, 23, 24] (Table 1). However, these methods have suffered from low accuracy and still

require expert user intervention for each image. In parallel, several fully automated deep learning-

based approaches have been proposed for TEM image analysis, targeting the segmentation of

GBM[25, 26], PFP[21], or both[27-29], as well as other related ultrastructural components[25, 30, 31]. Most of

these methods adopt U-Net-based semantic segmentation frameworks. While such models

demonstrate improved efficiency over traditional pipelines, they are inherently limited in their

ability to distinguish individual PFP instances, a critical requirement for accurate morphometric

analysis. Moreover, many of these studies are trained and evaluated on single-center or disease-

specific datasets, limiting generalizability across species, disease types, and imaging protocols[25,

26, 28]. In addition, suboptimal segmentation accuracy further constrains their applicability for

precise downstream ultrastructural measurements. More recently, the automated model FPW-DL

was developed to measure PFP width[31], with publicly released code and a valuable human disease

TEM dataset. The method demonstrates stable performance when applied to human samples

during testing. However, its reliance on identifying slit diaphragms as proxies for PFP width

measurement imposes intrinsic methodological limitations, leaving substantial room for

improvement in measurement accuracy (Fig. 1D). The Segment Anything Model (SAM)[32], the

recently published model, has become one of the most successful deep learning models for general

image segmentation. A variant fine-tuned for medical images, called MedSAM[33], has also been

released. While both models outperform most others in many biomedical image segmentation

tasks, their performance in TEM image analysis is unreliable, even when assisted with manual

prompting (Yellow boxes in Fig 1E, 1F). The need for precise prompts underscores the challenge

of applying SAM to TEM images. Building upon these developments, a recent study proposed

TEM-AID[34], a multi-stage TEM analysis framework integrating detection, prompt-based SAM,

measurement, and classification. However, its reliance on sequential processing and human-in-

the-loop refinement increases system complexity and limits scalability, and its quantitative

analysis primarily focuses on GBM, with limited support for simultaneous PFP quantification.

Significant variance across different diseases, the lack of high-quality labeled training data, high

resolution images with multi-scale and fine-grained features, and boundary ambiguity make TEM

image analysis extremely challenging for general AI models. Together, these challenges highlight

the urgent need for a specialized digital pathology platform capable of efficient and precise TEM

image analysis.

In response to these challenges, we introduce TEAMKidney, a novel deep-learning

framework designed for precise segmentation and quantification of TEM ultrastructure. Through

extensive generation, collection, and curation of approximately 13,000 TEM images representing

various kidney diseases across human, mouse, and rat samples, TEAMKidney is the first deep

learning model capable of robustly analyzing TEM images across different magnifications, disease

states, and species. To address the challenge of limited high-quality labeled data, in addition to

having experienced pathologists meticulously annotating the ultrastructure of glomeruli in 315

TEM images, we incorporated a self-training strategy to leverage large amounts of unlabeled data.

Furthermore, our specifically designed panoptic segmentation model can accurately distinguish individual ultrastructural components even in the absence of clear boundaries, which represents a major challenge for existing methods.

To validate the framework, we conducted extensive quantitative evaluations. In addition to comparisons with expert manual annotations, TEAMKidney significantly outperforms existing methods. Robust performance across data from different sources and species demonstrates the strong generalizability of TEAMKidney. Successful application to both animal models and patient datasets with diverse disease states further indicates its ability to capture subtle ultrastructural changes in real-world TEM data. Overall, this approach has the potential to enhance diagnostic accuracy, facilitate large-scale research on kidney disease, and support clinical decision-making in nephrology. Moreover, it can readily be extended to other clinical domains, making it a promising step toward practical deep learning-based analysis for general TEM imaging.

Results

The Overview of the TEAMKidney Framework

TEAMKidney is a deep learning-based automated framework for TEM image analysis, designed

to precisely quantify ultrastructure, such as GBM and PFP widths (Fig. 1G). The framework

comprises three key components: (1) integrating HRNet with self-training for initial semantic

segmentation (Fig. 1H), (2) developing a novel model for panoptic segmentation of glomerular

ultrastructure (Fig. 1H), and (3) performing post-segmentation analysis to robustly quantify

detailed ultrastructure, such as GBM width and PFP width (Fig. 1I-K).

To train and validate our framework across diverse pathological conditions and species,

our study conducted an analysis of both animal and human models under normal and diseased

states. For the mouse, the *Integrin linked kinase* podocyte-specific conditional knockout (ILK

cKO)[35, 36] was used for model training, whereas the type IV collagen subunit alpha 3 knockout

(Col4a3 KO) model was reserved as an independent evaluation dataset to assess model

generalizability. For rats, the passive Heymann nephritis model (PHN)[37] was used, which was

originally developed to characterize Membranous nephropathy (MN). For human samples, we

focused on the comparison among healthy donors, diabetic kidney disease (DKD)[38, 39] and Fabry

disease[40, 41]. For each species and disease model, a corresponding normal control dataset was

included to enable both training and validation.

This framework was implemented to automatically segment, label, and measure

ultrastructure in TEM images, thereby eliminating the need for manual tracing and reducing

observer variability. This design allows the framework to address diverse measurement tasks,

including interspecies comparisons, disease-associated ultrastructural changes, and quantitative

evaluation of experimental and clinical datasets.

Two-Stage Segmentation Model for Improved Fine Ultrastructure Recognition

Due to the high magnification and resolution of grayscale TEM images, ultrastructural features

often appear irregular and have unclear boundaries. In addition, TEM images are acquired across

multiple magnification levels and spatial resolutions, resulting in substantial scale variation of

ultrastructural components within and across datasets. These challenges prevent the direct

adaptation of existing segmentation software for TEM image analysis (Fig. 1D-F)[31-33]. To address

this, we design a novel deep learning model, Glom2Mask, that fuses both semantic and panoptic

segmentation approaches, enabling precise segmentation of ultrastructural features across species

and imaging scales.

In the semantic segmentation stage (Fig. 2A), we evaluated three backbone networks for

comparison: DeepLabV3[42-44], PSPNet[45], and HRNet[46]. All three models are effective in capturing

global context and fusing multi-scale features. Overall, the models achieved comparable

performance (Fig. 2B). Among them, HRNet attained the highest Dice scores across most

categories and demonstrated markedly greater training efficiency. Although the average Dice

scores of the three methods were similar, HRNet exhibited the most robust and consistent

performance. Therefore, we selected HRNet as the backbone for our panoptic segmentation model

in the second stage (Fig. 2C). In contrast, DeepLabV3 often failed on low-quality images (Fig. 2C,

Supplementary Fig. 1), resulting in unreliable segmentation, while PSPNet tended to misclassify

unrelated ultrastructure (Fig. 2C, Supplementary Fig. 1), particularly mistaking epithelial cells on

glomerular capillaries for foot processes.

Although our semantic segmentation model accurately recognized major ultrastructure,

direct post-segmentation analysis to identify individual instances, such as PFP and slit diaphragms,

remained highly challenging and often inaccurate, similar to most existing models[28, 29]. To

overcome this limitation, we introduced a panoptic segmentation stage (Fig. 1H) described in our

proposed Glom2Mask model. This architecture utilizes the extracted features from HRNet trained

in the first step. In addition, using annotations generated during the semantic segmentation stage,

along with manually refined corrections, we constructed an expanded and higher-quality labeled

training dataset to improve panoptic segmentation accuracy.

The panoptic segmentation performance of the proposed Glom2Mask and the baseline

Mask2Former models were compared using standard panoptic segmentation metrics: Panoptic

Quality (PQ), Segmentation Quality (SQ), and Recognition Quality (RQ), evaluated for both things

(PFPs) and stuff (GBM) classes. The results show that Glom2Mask outperformed the baseline

Mask2Former (Fig. 2D), achieving higher PQ, SQ, and RQ scores for both PFP and GBM while

requiring fewer training iterations (66,000 vs. 360,000). The quantitative results corresponding to

Fig. 2D are summarized in Table S1. These findings demonstrate superior segmentation accuracy

and training efficiency of the proposed model. Fig. 2E shows representative segmentation results

from Glom2Mask and Mask2Former. The visualizations demonstrate that Glom2Mask accurately

distinguishes continuous GBM structures (blue) from discrete PFP instances (various shades of

red), further supporting its robustness and generalizability across species.

To evaluate the robustness of our method to variations in image resolution, which may

mimic the differences in images generated by different institutes or instruments, we conducted a

controlled downsampling experiment using TEM images at multiple scales. As shown in

Supplementary Fig. 2A-B, representative segmentation results and corresponding input images

demonstrate that meaningful GBM and PFP boundaries can already be identified at moderate

downsampling levels, with progressively improved delineation as resolution increases.

Quantitative analysis further confirms this observation (Supplementary Fig. 2C), showing that

Dice scores for both GBM and PFP increase rapidly with higher resolution and reach a stable

plateau beyond a downsampling ratio of approximately 0.1. Notably, segmentation performance

remains consistently high across a broad range of resolution scales, indicating that the proposed

method is robust to substantial resolution degradation and generalizes well across varying imaging

conditions.

Concordance of TEAMKidney Predictions with Ground Truth Across Species

To evaluate how segmentation accuracy affects ultrastructure measurement, we generated ground

truth labels for a validation dataset comprising 48 images from 8 mouse glomeruli, 107 images

from 12 rat glomeruli, and 91 images from 5 human glomeruli. All validation images were

excluded from model training. GBM and PFP annotations were manually labeled by three human

experts to generate precise ultrastructure masks. Then, the post-segmentation quantification

algorithm was applied to both the TEAMKidney predictions and the expert-generated ground truth

masks in this validation dataset. Quantified GBM and PFP width values obtained from both ground

truth labels and TEAMKidney predictions were compared across mouse, rat, and human samples

at the individual image level. No statistically significant differences were observed between the

ground truth and TEAMKidney measurements for either GBM or PFP width across all three

species (Fig. 2F), demonstrating strong concordance and reliability of the proposed framework.

These results indicate that our method provides highly reliable GBM and PFP width measurements

across diverse datasets.

Distinguishing Ultrastructural Changes in Multiple Animal Models of Kidney Disease

Animals, especially rodents, are widely used to model a variety of kidney diseases. Ultrastructural

changes reflect disease progression and serve as indicators of treatment efficacy during drug

development[47]. To test the robustness of our framework for detecting pathophysiological changes

in the glomerular filtration barrier, we applied our automated method to measure two well-

established mouse models and a rat model of glomerular diseases (Fig. 3A, B). In the mouse

models, we used Col4a3 KO, an animal model of Alport syndrome[48] and ILK cKO mice, an animal

model of focal segmental glomerulosclerosis (FSGS), to represent two distinct mechanisms of

kidney glomerular injury[35, 36, 48]. Col4a3 KO mice exhibit severe GBM disorganization due to the

lack of type IV α3 collagen, which is a crucial component of the mature GBM, whereas ILK cKO

mice develop podocyte detachment and cytoskeletal abnormalities with the absence of ILK

protein, which is required for podocyte adhesion to the GBM (Fig. 3A, C). Consistent with these

pathological features, our automated analysis revealed significantly increased GBM widths in

Col4a3 KO mice compared with wild-type (WT) controls. Similarly, PFP widths were markedly

widened for the ILK cKO mouse model. These quantitative differences align with the expected

ultrastructural alterations associated with abnormal GBM of the Alport mouse model and podocyte

injury in the ILK cKO FSGS mouse model (Fig. 3C, D).

We next examined the rat PHN model of immune-mediated podocyte injury[37]. In this

model, rats were injected with anti-Fx1a antibody to induce the formation of immune complexes

that deposit beneath podocytes and trigger complement-mediated podocyte injury, leading to

progressive PFP effacement and GBM thickening (Fig. 3B, E). Automated measurements

faithfully captured these ultrastructural changes: compared with WT animals, PHN rats showed a

striking increase in GBM width (Fig. 3F). These differences are consistent with classical TEM-

based characterizations of PHN, an animal model of human membranous nephropathy, confirming

that our approach is sensitive to detecting pathological remodeling of the glomerular filtration

barrier.

Identifying Ultrastructural Remodeling in Human Kidney Diseases

TEM-based ultrastructural quantification is commonly used as diagnostic evidence in clinical renal

pathology practice. To investigate the translational utility of our framework in human kidney

disease states, we analyzed TEM images from patients with DKD and Fabry disease (Fig. 4A).

DKD is characterized by hyperglycemia-induced hyperfiltration, GBM thickening, diffuse

mesangial matrix expansion, leading to podocyte stress and progressive glomerular injury, PFP

effacement, glomerular hypertrophy, and eventually podocyte loss and severe nodular

glomerulosclerosis[49-51]. Fabry disease results from *GLA* mutations that cause accumulation of

globotriaosylceramide (Gb3) in podocytes and subsequent structural damage, PFP effacement,

podocyte detachment, and global or segmental glomerulosclerosis[52, 53] (Fig. 4A). Representative

TEM images predicted by our TEAMKidney model demonstrate marked distinct pathological

features in both disease cohorts compared with normal controls (Fig. 4B).

Quantitative measurements revealed disease-specific ultrastructural alterations (Fig. 4C).

In DKD patients, GBM thickness was significantly increased. Notably, the mean PFP width in

DKD showed only a slight increase compared with normal samples. The relatively preserved PFP

width in our DKD cohort warrants further investigation, as it may reflect the disease stage of the

included patients or capture an early phase where GBM pathology precedes overt podocyte

architectural disruption. Longitudinal studies correlating PFP measurements with clinical disease

staging would help clarify the temporal evolution of these ultrastructural changes[39, 54]. In contrast,

patients with Fabry disease exhibited significant increases in both GBM thickness and PFP width,

reflecting the more severe podocyte injury characteristic of lysosomal storage disorders[55, 56].

Furthermore, sex-stratified analysis of Fabry patients revealed that both male and female

patients exhibited substantial GBM and PFP widening (Fig. 4D, E), with more pronounced changes

in Fabry males. These findings are consistent with the known pathology of Fabry disease, in which

hemizygous males typically exhibit more severe phenotypes due to the X-linked inheritance

pattern[57, 58].

These results highlight the ability of our automated framework to detect clinically relevant

and disease-specific ultrastructural alterations in human glomerulopathies. By accurately capturing

the differential patterns of GBM thickening and PFP effacement in DKD and Fabry disease, the

method demonstrates strong potential for application in both diagnostic pathology and clinical

research, where reproducible and scalable measurements are critical for disease characterization,

staging, patient stratification and management.

To assess robustness across imaging scales, we evaluated our method on TEM images

acquired at multiple magnifications in the Fabry dataset. As shown in Fig. 4F, the predicted GBM

and PFP mean widths remain consistent across different magnifications, despite substantial

differences in image resolution. These results indicate that our method is stable and effective across

imaging scales and disease conditions. Similar results for the DKD and normal datasets are shown

in Supplementary Fig. 3.

Comparison of PFP Quantification with a Benchmark Model

To better evaluate the performance, we benchmarked TEAMKidney against existing automated

TEM quantification approaches using human, mouse, and rat kidney samples. As FPW-DL[31] is the

only method with publicly available code or model, and is designed exclusively to measure PFP

width, the comparison was limited to PFP width quantification.

Considering FPW-DL was trained only on human samples, our comparison focused on

human diseases. Even though all images were obtained from the FPW-DL study, TEAMKidney

consistently outperformed FPW-DL across both normal and disease cohorts (Fig. 5A, B,

Supplementary Fig. 4). Scatter plots show strong agreements between TEAMKidney and the

ground truth for both PFP width and counts (Fig. 5A, B). In contrast, FPW-DL tended to

overestimate PFP width and underestimate PFP counts, with this bias being more pronounced in

DKD than in the normal cohort. Mean error of PFP width was reduced from 24.80 pixels (FPW-

DL) to 4.46 pixels using our framework (Supplementary Fig. 4).

Further investigation revealed that PFP miscounting in FPW-DL arose from the use of slit

diaphragm detection (yellow dots in Fig. 5C) rather than direct PFP detection. Under disease

conditions or variable imaging quality, PFP effacement and overlap can obscure slit diaphragms,

leading to overestimation of PFP width when slit diaphragms are missed, or preferential detection

of relatively normal-appearing PFPs while pathologically altered regions are overlooked. Accurate

ultrastructural quantification, therefore, critically depends on high-quality segmentation, which is

the primary design goal of TEAMKidney. Representative TEM images show that our approach

yields more continuous and anatomically faithful segmentation of podocyte FPs and the GBM

compared with alternative methods (Fig. 5C), resulting in robust performance across both human

disease states and animal models.

The performance gap between the two methods became more pronounced in animal sample

(Fig. 5A, B), where FPW-DL, which was trained exclusively on human kidney data, showed

substantially larger deviations from ground truth (Fig. 5A). This limitation stems from two primary

factors: first, morphological differences between species, particularly the thinner GBM in rodents

compared to humans, poses additional challenges for accurate structure identification; second,

differences in imaging resolution and scale between our rodent datasets and FPW-DL's training

data further compromised its generalizability. In contrast, TEAMKidney maintained robust

performance across species, demonstrating superior adaptability to diverse morphological features

and imaging conditions.

Agreement of TEAMKidney and Manual Measurements Across Species

Manually tracing and labeling each ultrastructural component is not feasible in practice. Instead,

pathologists and biomedical scientists typically quantify ultrastructure by measuring multiple

regions within an image to obtain representative values (see Methods). To assess discrepancies,

we compared TEAMKidney measurements with manual measurements across normal, DKD, and

Fabry disease (male and female) from human cohorts (Fig. 6) and animal models (Supplementary

Fig. 5). For GBM widths, TEAMKidney quantification showed no significant differences from

manual measurements in all groups (Fig. 6A, Supplementary Fig. 5), confirming that the

framework preserves accuracy for this ultrastructural feature. Similarly, no significant differences

were detected in PFP counts between TEAMKidney and manual counting across all groups (Fig.

6B, Supplementary Fig. 5), indicating that the framework maintains biological fidelity in

enumerating PFP.

We observed a significant discrepancy in PFP width measurements between the two

methods, with mean values from TEAMKidney consistently lower than those obtained by manual

measurement. This difference likely arises from two factors. First, due to the irregular shape of

PFPs, in the manual approach, PFP width is not measured directly but instead estimated by dividing

the GBM length by the PFP count. GBM length is often overestimated, particularly in disease

models with pronounced morphological alterations. Second, manual measurements include the slit

diaphragm as part of the PFP, which inflates the estimated PFP width. In contrast, the

TEAMKidney measurement strategy segments individual PFPs and slit diaphragm separately;

therefore, the slit diaphragm is excluded from the PFP measurement, resulting in higher accuracy

compared with the manual method (Fig. 6C, Supplementary Fig. 5).

Together, these results demonstrate strong concordance between TEAMKidney and

manual measurements, establishing the framework as a reliable substitute for labor-intensive

tracing. By accurately replicating expert-level quantification across both healthy and diseased

human samples, the method enhances reproducibility and offers a scalable solution for

ultrastructural kidney assessment.

Comparison of Clinical Protocol and TEAMKidney for GBM and PFP Quantification

Although manual measurement is widely used in basic and clinical research, area selection is

subject to individual judgment, leading to substantial inter-operator variability. In clinical settings,

more stringent area-selection protocols are applied to improve reproducibility (see Methods). To

contextualize the advantages of our automated approach, we compared it with the conventional

diagnostic manual workflow commonly used for ultrastructural measurements of GBM and PFP

widths (Fig. 7). Pathologists rely on ImageJ (or FIJI) based protocols, which require calibration

with scale bars, grid overlays, and labor-intensive tracing of GBM thickness at multiple grid

intersections, followed by foot process identification and marking of lateral edges (Fig. 7A). This

procedure demands at least 200 measurements from three independent glomeruli per sample to

achieve statistical robustness, making it time consuming and highly dependent on operator

expertise.

Representative TEM images further illustrate the differences in measurement strategies

(Fig. 7B-E). Manual quantification requires repetitive grid-based annotation and point-by-point

width extraction (Fig. 7C and D), whereas our automated framework directly performs panoptic

segmentation and downstream computational measurement of GBM and PFP widths (Fig. 7E).

This not only eliminates observer variability but also enables high-throughput analysis of large

datasets with consistent reproducibility.

Collectively, these comparisons underscore the limitations of traditional manual workflows,

including their subjectivity, inefficiency, and limited scalability, and demonstrate that the proposed

automated framework effectively overcomes these challenges, providing a practical and objective

solution for ultrastructural kidney analysis.

Discussion

In recent years, a global shortage of pathologists has been widely reported[59, 60], affecting even large medical centers and resource-rich regions. The preparation and interpretation of tissue samples for TEM imaging require additional specialized training, particularly for renal pathologists, further exacerbating this workforce constraint. To address the urgent needs in nephrology research and clinical pathology, our team of experienced renal pathologists, nephrologists, biologists, and computational scientists designed and implemented TEAMKidney. Precise measurement of the GBM and PFP provides a much-needed solution to address labor-intensive and often subjective assessments of glomerular diseases. GBM and PFP are not only essential ultrastructure for diagnosing and monitoring chronic kidney diseases, but they also represent two distinct classes of TEM features. GBM is characterized by its continuous structure with a relatively regular shape and ambiguous boundaries throughout the image. In contrast, PFP consists of small, isolated structures with clear boundaries and irregular shapes. Accurate segmentation and quantification of these structures demonstrate the capability of the framework to extend beyond nephrology and renal pathology to general TEM image analysis across other fields.

High-quality labeled data is critical for training deep learning models, but manual annotation of TEM images is extremely time-consuming and requires specialized pathological expertise. To overcome this challenge, we adopted semi-automated pre-labeling strategies to accelerate the annotation workflow and implemented a multi-stage quality control process, ensuring that each label was independently reviewed and refined. These steps significantly reduced inter-operator variability and improved the overall reliability of the training data. Combined with a self-training strategy, our framework achieves high accuracy with only a few hundred finely annotated images, making it feasible to adapt to other tasks efficiently.

Previous TEM analysis methods have primarily focused on segmenting a single

ultrastructural feature, often failing to generalize to other ultrastructures. Most studies have relied

on U-Net based or semi-automated approaches (Table 2), which reduce the need for large amounts

of training data but are well known for their limited generalizability to diverse morphologies.

Because these methods typically use data from a single source and a single species, their practical

applicability in animal disease modeling and clinical settings is further constrained. In addition to

collecting and generating diverse datasets across multiple sources, species, and disease states, our

specifically designed panoptic segmentation model offers two key advantages over directly

applying generic computer vision models to this task. First, semantic segmentation provides

spatially consistent, region-level guidance, reducing noisy or fragmented instance predictions

during pseudo-labeling. Second, the trained HRNet model serves as the backbone for the panoptic

segmentation network, enabling direct transfer and reuse of feature representations learned during

semantic segmentation. This strategy not only accelerates model convergence but also improves

feature consistency between semantic and panoptic tasks, thereby enhancing boundary precision

and instance recognition. Overall, this hierarchical training strategy achieves greater stability and

accuracy, particularly for densely packed ultrastructural components such as podocyte foot

processes.

TEM images inherently exhibit substantial variation in magnification and spatial resolution

across different species, disease models, and experimental setups. Such differences often lead to

inconsistencies in the visual scale and morphological representation of ultrastructural components,

posing a major challenge for deep learning models trained on limited magnification ranges. To

address this issue, we employed a series of strategies at multiple stages to enhance the model's

robustness to magnification variability. First, we collected TEM samples acquired at different

magnifications across species and disease models to ensure comprehensive coverage of structural

and contextual diversity. Second, we organized and standardized these multi-magnification

samples by applying systematic cropping and rescaling operations to generate datasets

representing diverse spatial scales. This step provided the model with balanced exposure to both

high- and low-resolution features of the glomerular ultrastructure. Third, we incorporated random

scaling, cropping, and other image augmentation techniques during model training to simulate

continuous magnification changes and prevent overfitting to specific resolutions. Collectively,

these strategies substantially improved the model's ability to generalize across heterogeneous

imaging conditions, yielding consistent segmentation quality and reliable quantitative

measurements across a wide range of magnification differences.

Our proposed framework demonstrates high accuracy and reproducibility in segmenting

and quantifying GBM and PFP, and generalizes effectively across species and disease states while

accommodating substantial morphological and magnification variability. Nonetheless, there

remains room for further improvement. Because TEM sections are two-dimensional

representations of inherently three-dimensional structures, the spatial orientation of PFP can

influence quantitative measurements, although this effect is unlikely to substantially alter overall

conclusions when analyzed across whole images. To further enhance measurement accuracy, we

are developing new methods that explicitly incorporate PFP orientation analysis. In parallel, we

are extending the framework to include additional ultrastructural features relevant to other kidney

diseases, such as immune deposits, slit diaphragms, and endothelial cells. As the first framework

to achieve both segmentation and quantitative measurement with expert-level precision,

TEAMKidney enables accurate characterization of glomerular morphological changes across

diverse clinical conditions, ranging from normal kidneys and common diseases such as DKD to

rare disorders such as Fabry disease. This work establishes a solid foundation for expanding the

framework to quantify the full spectrum of glomerular diseases and to support future applications,

including disease detection, classification, and automated report generation.

While open-source software is standard practice in computer science, it remains

uncommon in the field of TEM image analysis. With the exception of FPW-DL, all other existing

studies have not released their code or trained models, making independent evaluation impossible

and significantly limiting the broader adoption and impact of these methods. We greatly respect

the FPW-DL authors for openly sharing their code, models, and data, and for their prompt and

helpful responses to our inquiries. We strongly believe that making our own code and models

publicly available will foster community-wide experimentation, accelerate methodological

advances, and ultimately benefit patient care. This openness also reflects our commitment to

transparency and confidence in the robustness of our approach.

References

1. Koster, A.J. & Klumperman, J. Electron microscopy in cell biology: integrating structure and function. *Nat Rev Mol Cell Biol* **Suppl**, SS6-10 (2003).

2. Curry, A., Appleton, H. & Dowsett, B. Application of transmission electron microscopy to the clinical study of viral and bacterial infections: present and future. *Micron* **37**, 91-106 (2006).

3. Roingeard, P., Raynal, P.I., Eymieux, S. & Blanchard, E. Virus detection by transmission electron microscopy: Still useful for diagnosis and a plus for biosafety. *Rev Med Virol* **29**, e2019 (2019).

4. Graham, L. & Orenstein, J.M. Processing tissue and cells for transmission electron microscopy in diagnostic pathology and research. *Nat Protoc* **2**, 2439-2450 (2007).

5. L'Imperio, V. et al. Digital pathology for the routine diagnosis of renal diseases: a standard model. *J Nephrol* **34**, 681-688 (2021).

6. Murray, R. et al. Kidney-Related Research in the United States: A Position Statement From the National Kidney Foundation and the American Society of Nephrology. *Am J Kidney Dis* **78**, 161-167 (2021).

7. Assady, S., Benzing, T., Kretzler, M. & Skorecki, K.L. Glomerular podocytes in kidney health and disease. *Lancet* **393**, 856-858 (2019).

8. Faucon, A.L. CKD Beyond the Clinic: A Holistic Projection of its Global Burden. *Kidney Int Rep* **10**, 2524-2526 (2025).

9.   Chadban, S. et al. Projecting the economic burden of chronic kidney disease at the patient

level (Inside CKD): a microsimulation modelling study. *EClinicalMedicine* **72**, 102615

(2024).

10. Ramachandran, R. et al. Challenges in Diagnosis and Management of Glomerular Disease

in Resource-Limited Settings. *Kidney Int Rep* **7**, 2141-2149 (2022).

11. George, C., Mogueo, A., Okpechi, I., Echouffo-Tcheugui, J.B. & Kengne, A.P. Chronic

kidney disease in low-income to middle-income countries: the case for increased screening.

BMJ Glob Health **2**, e000256 (2017).

12. Rajapurkar, M.M. et al. What do we know about chronic kidney disease in India: first report

of the Indian CKD registry. *BMC Nephrol* **13**, 10 (2012).

13. Sharma, M., Doley, P. & Das, H.J. Etiological profile of chronic kidney disease: A single-

center retrospective hospital-based study. *Saudi J Kidney Dis Transpl* **29**, 409-413 (2018).

14. Yamashita, M., Lin, M.Y., Hou, J., Ren, K.Y.M. & Haas, M. The Continuing Need for

Electron Microscopy in Examination of Medical Renal Biopsies: Examples in Practice.

Glomerular Dis **1**, 145-159 (2021).

15. May, R.M. et al. A multi-center retrospective cohort study defines the spectrum of kidney

pathology in Coronavirus 2019 Disease (COVID-19). *Kidney Int* **100**, 1303-1315 (2021).

16. Pisarek-Horowitz, A. et al. Loss of Roundabout Guidance Receptor 2 (Robo2) in Podocytes

Protects Adult Mice from Glomerular Injury by Maintaining Podocyte Foot Process

Structure. *Am J Pathol* **190**, 799-816 (2020).

17. Veron, D. et al. Overexpression of VEGF-A in podocytes of adult mice causes glomerular

disease. *Kidney Int* **77**, 989-999 (2010).

18.   Sethi, S. et al. Mayo Clinic/Renal Pathology Society Consensus Report on Pathologic
493   Classification, Diagnosis, and Reporting of GN. *J Am Soc Nephrol* **27**, 1278-1287 (2016).

494   19.   Levin, A. et al. Global kidney health 2017 and beyond: a roadmap for closing gaps in care,
495   research, and policy. *Lancet* **390**, 1888-1917 (2017).

496   20.   Rangayyan, R.M., Kamenetsky, I. & Benediktsson, H. Segmentation and analysis of the
497   glomerular basement membrane in renal biopsy samples using active contours: a pilot
498   study. *J Digit Imaging* **23**, 323-331 (2010).

499   21.   Unnersjo-Jess, D. et al. Deep learning-based segmentation and quantification of podocyte
500   foot process morphology suggests differential patterns of foot process effacement across
501   kidney pathologies. *Kidney Int* **103**, 1120-1130 (2023).

502   22.   Geraghty, M. et al. Sources of Variability in Podocyte Foot Process Width Measurements
503   and Approaches to Mitigation: PO2004. *Journal of the American Society of Nephrology*
504   **31**, 617 (2020).

505   23.   Kamenetsky, I., Rangayyan, R.M. & Benediktsson, H. Analysis of the glomerular basement
506   membrane in images of renal biopsies using the split-and-merge method: a pilot study. *J*
507   *Digit Imaging* **23**, 463-474 (2010).

508   24.   Wu, H.S., Dikman, S. & Gil, J. A semi-automatic algorithm for measurement of basement
509   membrane thickness in kidneys in electron microscopy images. *Comput Methods*
510   *Programs Biomed* **97**, 223-231 (2010).

511   25.   Wang, Y. et al. Segmentation and thickness calculation of glomerular basement membrane
512   using RADS-Net in glomerular microscopic images. *Biomedical Signal Processing and*
513   *Control* **88**, 105557 (2024).

514   26.   Laudon, A. et al. Digital pathology assessment of kidney glomerular filtration barrier
515   ultrastructure in an animal model of podocytopathy. *Biol Methods Protoc* **10**, bpaf024
516   (2025).

517   27.   Lin, G. et al. GCLR: A self-supervised representation learning pretext task for glomerular
518   filtration barrier segmentation in TEM images. *Artif Intell Med* **146**, 102720 (2023).

519   28.   Yamashita, M. et al. AI-driven glomerular morphology quantification: a novel pipeline for
520   assessing basement membrane thickness and podocyte foot process effacement in kidney
521   diseases. *Comput Methods Programs Biomed* **268**, 108842 (2025).

522   29.   Zhang, Z. et al. Glo-UMF: A Unified Multi-model Framework for Automated
523   Morphometry of Glomerular Ultrastructural Characterization. *arXiv preprint*
524   *arXiv:2508.10351* (2025).

525   30.   Yang, J., Hu, X., Pan, H., Chen, P. & Xia, S. Multi-scale attention network for segmentation
526   of electron dense deposits in glomerular microscopic images. *Microsc Res Tech* **85**, 3256-
527   3264 (2022).

528   31.   Smerkous, D. et al. Development of an automated estimation of foot process width using
529   deep learning in kidney biopsies from patients with Fabry, minimal change, and diabetic
530   kidney diseases. *Kidney Int* **105**, 165-176 (2024).

531   32.   Kirillov, A. et al. in Proceedings of the IEEE/CVF international conference on computer
532   vision 4015-4026 (2023).

533   33.   Ma, J. et al. Segment anything in medical images. *Nat Commun* **15**, 654 (2024).

534   34.   Ma, P. et al. AI-Based System for Analysis of Electron Microscope Images in Glomerular
535   Disease. *JAMA Netw Open* **8**, e2534985 (2025).

536   35.   Huang, M. et al. Integrin-Linked Kinase Deficiency in Collecting Duct Principal Cell

537   Promotes Necroptosis of Principal Cell and Contributes to Kidney Inflammation and

538   Fibrosis. *J Am Soc Nephrol* **30**, 2073-2090 (2019).

539   36.   El-Aouni, C. et al. Podocyte-specific deletion of integrin-linked kinase results in severe

540   glomerular basement membrane alterations and progressive glomerulosclerosis. *J Am Soc*

541   *Nephrol* **17**, 1334-1344 (2006).

542   37.   Jefferson, J.A., Pippin, J.W. & Shankland, S.J. Experimental Models of Membranous

543   Nephropathy. *Drug Discov Today Dis Models* **7**, 27-33 (2010).

544   38.   Yang, H. et al. Podocyte programmed cell death in diabetic kidney disease: Molecular

545   mechanisms and therapeutic prospects. *Biomed Pharmacother* **177**, 117140 (2024).

546   39.   Marshall, C.B. Rethinking glomerular basement membrane thickening in diabetic

547   nephropathy: adaptive or pathogenic? *Am J Physiol Renal Physiol* **311**, F831-F843 (2016).

548   40.   Bokhari, S.R.A. & Safdar, A. in StatPearls (Treasure Island (FL); 2026).

549   41.   Levstek, T. et al. Identification of genetic variants associated with Fabry nephropathy

550   progression using whole-exome sequencing. *Mol Genet Metab* **145**, 109191 (2025).

551   42.   Chen, L.-C., Papandreou, G., Schroff, F. & Adam, H. Rethinking atrous convolution for

552   semantic image segmentation. *arXiv preprint arXiv:1706.05587* (2017).

553   43.   Chen, L.-C., Papandreou, G., Kokkinos, I., Murphy, K. & Yuille, A.L. Deeplab: Semantic

554   image segmentation with deep convolutional nets, atrous convolution, and fully connected

555   crfs. *IEEE transactions on pattern analysis and machine intelligence* **40**, 834-848 (2017).

556   44.   Chen, L.-C., Zhu, Y., Papandreou, G., Schroff, F. & Adam, H. in Proceedings of the

557   European conference on computer vision (ECCV) 801-818 (2018).

558   45.   Zhao, H., Shi, J., Qi, X., Wang, X. & Jia, J. in Proceedings of the IEEE conference on

559   computer vision and pattern recognition 2881-2890 (2017).

560   46.   Sun, K., Xiao, B., Liu, D. & Wang, J. in Proceedings of the IEEE/CVF conference on

561   computer vision and pattern recognition 5693-5703 (2019).

562   47.   Beck, L.H., Jr. et al. PODO: Trial Design: Phase 2 Study of PF-06730512 in Focal

563   Segmental Glomerulosclerosis. *Kidney Int Rep* **6**, 1629-1633 (2021).

564   48.   Cosgrove, D. et al. Collagen COL4A3 knockout: a mouse model for autosomal Alport

565   syndrome. *Genes Dev* **10**, 2981-2992 (1996).

566   49.   Lin, Y.C., Chang, Y.H., Yang, S.Y., Wu, K.D. & Chu, T.S. Update of pathophysiology and

567   management of diabetic kidney disease. *J Formos Med Assoc* **117**, 662-675 (2018).

568   50.   Sugahara, M., Pak, W.L.W., Tanaka, T., Tang, S.C.W. & Nangaku, M. Update on diagnosis,

569   pathophysiology, and management of diabetic kidney disease. *Nephrology (Carlton)* **26**,

570   491-500 (2021).

571   51.   Mora-Fernandez, C. et al. Diabetic kidney disease: from physiology to therapeutics. *J

572   Physiol* **592**, 3997-4012 (2014).

573   52.   Alroy, J., Sabnis, S. & Kopp, J.B. Renal pathology in Fabry disease. *J Am Soc Nephrol* **13

574   Suppl 2**, S134-138 (2002).

575   53.   Eng, C.M. et al. Fabry disease: guidelines for the evaluation and management of multi-

576   organ system involvement. *Genet Med* **8**, 539-548 (2006).

577   54.   Najafian, B. et al. Progressive podocyte injury and globotriaosylceramide (GL-3)

578   accumulation in young patients with Fabry disease. *Kidney Int* **79**, 663-670 (2011).

579   55.   Wise, A.F. et al. Fabry Disease Podocytes Reveal Ferroptosis as a Potential Regulator of

580   Cell Pathology. *Kidney Int Rep* **10**, 535-548 (2025).

56.   Monte Neto, J.T.D. & Kirsztajn, G.M. The role of podocyte injury in the pathogenesis of

Fabry disease nephropathy. *J Bras Nefrol* **46**, e20240035 (2024).

57.   Mehta, A. & Hughes, D.A. in GeneReviews((R)). (eds. M.P. Adam et al.) (Seattle (WA);

1993).

58.   Germain, D.P. Fabry disease. *Orphanet J Rare Dis* **5**, 30 (2010).

59.   Osman, M.A. et al. Global nephrology workforce: gaps and opportunities toward a

sustainable kidney care system. *Kidney Int Suppl (2011)* **8**, 52-63 (2018).

60.   Collaboration, G.B.D.C.K.D. Global, regional, and national burden of chronic kidney

disease, 1990-2017: a systematic analysis for the Global Burden of Disease Study 2017.

Lancet **395**, 709-733 (2020).

61.   Salant, D.J. & Cybulsky, A.V. Experimental glomerulonephritis. *Methods Enzymol* **162**,

421-461 (1988).

62.   Russell, B.C., Torralba, A., Murphy, K.P. & Freeman, W.T. LabelMe: A Database and Web-

Based Tool for Image Annotation. *International Journal of Computer Vision* **77**, 157-173

(2008).

63.   Lahdenkari, A.T. et al. Podocytes are firmly attached to glomerular basement membrane in

kidneys with heavy proteinuria. *J Am Soc Nephrol* **15**, 2611-2618 (2004).

64.   Shea, S.M. & Morrison, A.B. A stereological study of the glomerular filter in the rat.

Morphometry of the slit diaphragm and basement membrane. *J Cell Biol* **67**, 436-443

(1975).

Methods

Sample Preparation and Image Acquisition

ILK podocyte-specific knockout (ILK cKO) and Col4a3 knockout (Col4a3 KO) models. The ILK

cKO mice were generated by crossing ILK[flox/flox] mice with Podocin-Cre[+] mice[36]. Col4a3 KO mice

were purchased from the Jackson Laboratory (RRID:IMSR_JAX:002908). The littermate WT

mice were used as healthy controls. Mice were sacrificed at 16 weeks of age, and the kidneys were

dissected. The fresh kidney samples were fixed in Karnovsky's fixative and processed according

to the published protocol[4]. In brief, small pieces of kidney samples (<1 mm thick) were fixed in

buffered 2.5% glutaraldehyde. Kidney specimens were then post-fixed in 1% osmium tetroxide

(Electron Microscopy Sciences), dehydrated, embedded in propylene oxide/Epon resin, and

sectioned (60 to 90 nm thick) using the ultramicrotome. Collected sections were post-stained with

uranyl acetate and lead citrate (Electron Microscopy Sciences), according to the published

protocol[4]. Images were obtained using the Jeol JEM-1011 transmission electron microscope (JEOL

USA Inc.) equipped with an Erlangshen ES100W digital camera (Gatan, Pleasanton, CA). The

operator was blinded as to the genotype and the injury status of the animals at the time of image

collection and analyses[16].

Rat PHN injury model. A modified PHN injury model was induced in Lewis rats (strain code 004;

Charles River Laboratories, Wilmington, MA) by one dose of i.v. injection [1.5 to 2 mL/kg sheep

anti-rat Fx1A serum (PTX-002S; Probetex, San Antonio, TX)], according to Probetex and

published protocol[61]. Rats were euthanized at day 16 after anti-rat Fx1A serum injection, and

kidneys were collected and fixed in Karnovsky's fixative for TEM analysis. Sterile 0.9% sodium

chloride (normal saline) was used in control rats. A total of six animals in each experimental group

625 were studied. Kidney samples were prepared, and TEM imaging was performed as described

626 above.

627 *Human diseases.* TEM images used in this study were obtained from an open-source dataset

628 described in a recent publication[31]. The dataset includes kidney biopsy samples from patients with

629 Fabry disease, type 2 diabetes, and minimal change disease, as well as healthy controls from living

630 kidney transplant donors. All biopsies were processed at eight EM laboratories following

631 standardized protocols. Ultrathin glutaraldehyde fixed sections were stained with toluidine blue

632 for glomerular identification, and high magnification transmission EM images (original

633 magnification ×30,000) were captured using a systematic uniform random sampling strategy.

635 **Label Generation, Manual Measurement, and Image Prepressing**

636 *Ground Truth Label.* All TEM images were divided into training, validation, and testing datasets

637 as summarized in Table 1. The training/validation and testing sets were intentionally sourced from

638 different cohorts, encompassing distinct species and disease types, to ensure that the testing dataset

639 could robustly evaluate the generalization capability of our algorithm. For the training and

640 validation sets, regions corresponding to the GBM and PFP were manually annotated using

641 LabelMe software[62]. The annotations were performed by trained researchers using high-resolution

642 TEM images and subsequently reviewed by renal pathologists to ensure consistency and structural

643 accuracy. This manual labeling process served as the foundation for model training and validation,

644 while the independently collected testing data were reserved for assessing model performance

645 across unseen imaging and disease conditions.

646 *Manual measurement for GBM and FP widths.* Manual podocyte foot process width was traced

647 and measured on ×12,000 high magnification TEM images using ImageJ software version 1.47v,

according to published methods[16]. Briefly, the total length of the GBM spanning the area of the

slit diaphragms was measured in each image using ImageJ, according to published methods[17, 63,

64]. The results were then pasted on an Excel sheet and adjusted for the scale bar. After that, all slit

diaphragms covering the GBM were counted, and the PFP width was calculated using the total

length of GBM divided by the total number of slit diaphragms in each image. The GBM width was

measured as the distance from the podocyte membrane to the endothelial membrane. Each image

was randomly measured in 20 areas around the GBM, and the harmonic mean was calculated for

each image.

Preparing training and testing datasets for the deep learning framework. Images were normalized

by mean-std normalization and tiled into [size, e.g., 1024×1024, 2624*2624] with 50% overlap.

Scale bars and metadata were parsed to record the pixel-to-nm factor (e.g., 4.122 nm/pixel for

mice, 3.88 nm/pixel for rats) used for unit conversion. To avoid leakage, splits were patient/animal-

level stratified into train/validate with an 80/20 ratio while preserving class balance per species

and disease model. All preprocessing scripts are provided (see Code availability).

Hybrid Data Preparation. To improve the model's adaptability across images of different

resolutions, we adopted a stage-wise data preparation and training strategy. High-resolution TEM

images were first used to pretrain the backbone network, enabling the model to learn fine-grained

structural features of GBM and PFP. Subsequently, lower resolution images were used for fine-

tuning to encourage scale-invariant feature learning and improve cross-magnification

generalization. During both training stages, various data augmentation techniques were applied,

including random scaling, random cropping, horizontal and vertical flipping, and random

adjustments of brightness and contrast. These augmentations helped prevent overfitting and further

enhanced the model's robustness to variations in image scale and quality.

Ultrastructure Quantification Based on Clinical Protocols. GBM thickness and PFP width were

measured using TEM images analyzed in ImageJ (FIJI). Images were calibrated using reference

scale bars to ensure accurate morphometric analysis. For GBM thickness measurement, a

standardized 1.5 µm × 1.5 µm grid was superimposed on micrographs for systematic sampling. At

each grid intersection, perpendicular lines were drawn from the endothelial cell membrane to the

outer margin of the lamina rara externa beneath epithelial foot processes. Measurements were

taken only in well-preserved regions, excluding artifacts due to sectioning angle, folding, or

detachment. The Orthogonal Intercept Method (OIM) was applied to calculate mean GBM

thickness. For PFP width measurement, the lateral edges of individual PFPs were manually

identified, and width was measured as the linear distance between adjacent borders. Only clearly

defined and intact PFPs were included in the analysis.

Glom2Mask: TEM Image Panoptic Segmentation Model Architecture

Our proposed Glom2Mask model is designed based on the Mask2Former panoptic segmentation

framework, enabling comprehensive segmentation of GBM and PFP in TEM images. Given the

distinct structural characteristics of these components, we categorize GBM as a "stuff" class for

semantic segmentation and PFP as a "things" class for instance segmentation. This formulation

ensures precise differentiation between the continuous structure of the GBM and the discrete PFP

instances within the glomerular filtration barrier.

To enhance segmentation accuracy, we integrate HRNet as the backbone for high-

resolution feature extraction. HRNet is well-suited for this task due to its ability to preserve spatial

details throughout the network, which is essential for capturing fine structural differences in TEM

images. Unlike traditional backbones that downsample feature maps at early stages, HRNet

maintains high-resolution representations, allowing for better delineation of GBM boundaries and

accurate identification of individual PFP structures.

Our model follows a transformer-based approach, where multi-scale features extracted by

HRNet are processed through multi-head attention mechanisms to generate high-quality

segmentation masks. The network outputs a set of learned queries, each corresponding to a GBM

region or a PFP instance, enabling robust panoptic segmentation across diverse TEM datasets,

including mouse, rat, and human kidney samples.

By leveraging HRNet for high-resolution feature extraction and Mask2Former for unified

panoptic segmentation, our model effectively segments GBM and PFP structures, providing the

necessary groundwork for precise automated width measurement in kidney disease analysis.

Model Training for the Deep Learning Framework

Our training process consists of two stages: a self-training-based semantic segmentation stage and

a panoptic segmentation stage using Glom2Mask. The first stage aims to generate additional

pseudo-labeled data while simultaneously learning a high-resolution backbone network. In the

second stage, we utilize the HRNet trained in the first stage as the backbone of a modified

Mask2Former model for panoptic segmentation, thereby improving both training efficiency and

segmentation accuracy.

Self-Training for Semantic Segmentation. In the first stage, we employ a self-supervised learning

approach to HRNet-based semantic segmentation.

For cases with a limited number of annotated images (e.g., 19, 48, or 64 labeled samples),

we employed a self-training strategy to improve segmentation performance. Self-training is a

semi-supervised learning technique that utilizes both labeled and unlabeled data. Initially, the

model is trained using only the labeled dataset to capture fundamental feature representations.

After this initial training, the trained model generates pseudo labels for the unlabeled data. These

pseudo-labeled samples, combined with the original labeled dataset, are then used to retrain the

model iteratively, progressively refining its segmentation accuracy and generalization capabilities.

The model performance was evaluated using the mean Dice score. For each manual

segmentation mask, per category, was compared to the prediction mask of the model. Dice score

was calculated by,

724
$$Dice = \frac{2|A_i \cap P_i|}{|A_i| + |P_i|}$$

Where A_i as the ground truth segmentation result for class i, P_i as the predicted segmentation

result for class i. Dice score values range from 0 to 1, where 0 represents no pixel overlap, and 1

represents perfect pixel overlap between the two masks.

During training, we applied a series of data augmentation techniques to improve model

robustness, including random scaling within a range of 0.5-2.0, random rotation within ±10°,

horizontal and vertical flipping with a probability of 0.5 each, random cropping to a size of

1024×1024 pixels, and random intensity jitter with brightness and contrast factors between 0.8 and

1.2. The model was optimized using the AdamW optimizer with a learning rate of 6×10^{-5} and a

weight decay of 0.05. A polynomial learning rate decay schedule with a linear warm-up phase of

1,500 iterations was employed. Training was conducted with a batch size of 16 for a total of 80,000

iterations. To address class imbalance, a combination of cross-entropy loss and focal loss ($\gamma = 2.0$,

$\alpha = 0.25$) was used.

Panoptic Segmentation with Glom2Mask. In the second stage, the training setup follows the same

training validation split as in the first step, ensuring consistency in dataset partitioning. Data

augmentation techniques, including random scaling, cropping, rotation, flipping, and contrast

adjustments, are applied during training to enhance generalization similar to step 1. The model

was optimized using the AdamW optimizer with a learning rate of 1×10^{-5} and a weight decay of

0.05 for a total of 160,000 iterations. The training objective combined three loss functions,

including classification cross entropy, mask cross entropy, and Dice loss, with relative weights of

2.0: 5.0: 1.0. Early stopping and checkpoint selection were determined based on the PQ metric

evaluated on the validation set.

The loss functions for training are as follows: Segmentation classification head: Cross-

entropy loss, ensuring accurate categorical predictions for GBM and PFP. Mask prediction head:

A combination of cross-entropy loss and Dice loss, improving both pixel-wise accuracy and mask

shape consistency. To quantitatively assess model performance, we employ the following panoptic

segmentation metrics:

• Panoptic Quality (PQ): Measures the overall segmentation accuracy, balancing

segmentation precision (true positive detection) and quality (IoU based mask accuracy).

Defined as:

754
$$PQ = \frac{\sum_{(p,g)\in TP} IoU(p,g)}{|TP| + 0.5|FP| + 0.5|FN|}$$

755   where TP (true positives) are correctly matched masks, FP (false positives) are extra masks,

756   and FN (false negatives) are missed masks.

757   • Segmentation Quality (SQ): Measures the average mask accuracy of correctly predicted

758   objects:

759
$$SQ = \frac{\sum_{(p,g)\in TP} IoU(p,g)}{|TP|}$$

760   • Recognition Quality (RQ): Measures the correctness of mask instance association and is

761   computed as:

$$RQ = \frac{|TP|}{|TP| + 0.5|FP| + 0.5|FN|}$$

Training Environment and Reproducibility

All experiments were run on Ubuntu 22.04 with Python 3.9, PyTorch 2.1.0, CUDA 12.1, cuDNN

8.9, MMDetection 2.0.1, and MMCV. Hardware:1*NVIDIA A100-80G GPU, Intel Xeon Gold

6338 CPU (40 cores). Deterministic behavior was enforced using a fixed seed (seed=42) and

disabling nondeterministic cuDNN ops. Full environmental files, exact config YAMLs, and trained

checkpoints are provided (see Code availability).

Post Segmentation Processing and Quantification

GBM Width Measurement. The second stage of our proposed method consists of a Python image

processing algorithm that takes input of the GBM and PFP segmentation tiles predicted by the

deep learning model to measure the mean GBM width and PFP width. We will describe the GBM

width measurement first here. The mean GBM width for each tile is then calculated as its total

GBM area divided by its total GBM central length according to the following equation

$$GBM\ width = \frac{GBM\ area}{GBM\ central\ length}$$

Where the area of the GBM is straightforward to compute, it involves directly counting the

number of pixels in a section of the GBM. The process for calculating the length of the GBM is as

follows: First, we obtain the skeleton of the GBM using the skeletonize method with the Python

package scikit-image. The general skeleton obtained through the skeletonize method, in addition

to the central length of the GBM, also includes many small branching points (Fig. 1J). To eliminate

these branching points, we construct an undirected cyclic graph of the GBM skeleton using the

Python package NetworkX. Here, there are two scenarios: if the GBM forms a loop structure, we

iteratively prune nodes with an in-degree of one from the undirected cyclic graph of the GBM

skeleton until no nodes with an in-degree of one remain. We then calculate the perimeter of the

resulting graph to determine the central length of the GBM; if the GBM forms a strip-like structure,

we calculate the shortest distance between any two endpoints and identify the two endpoints with

the greatest distance as the central length of the GBM. Lastly, the specimen mean GBM width was

converted from pixels to distance with a conversion factor derived from the TEM image scale bar,

for example, 3.88 nm/pixel for rat model data.

PFP Width Measurement. Following this, the PFP width calculation is more direct. Each PFP

instance is represented by a unique integer starting from 1, with the background assigned a value

of 0.

Next, we use morphological operations in OpenCV to dilate the GBM, enabling us to obtain

its contour. In this processed image, the background remains coded as 0, and the contour positions

are coded as 1.

Finally, we calculate the width of each PFP by multiplying the PFP instance matrix by the

matrix of the dilated GBM contours. This operation yields the width by computing the pixel length

of PFPs. Similarly, the specimen PFP widths were converted from pixels to distance with a

conversion factor derived from the TEM image scale bar, for example, 3.88 nm/pixel for rat model

data.

Author contributions

C.Z., W.L., A.Z., and X.F. conceived the study. A.Z. developed the deep learning model. W.L., X.F., S.R.V., S.P.B., H.Y., and H.C. acquired the images from animal models. X.F., A.Z., W.T., J.J., L.D., and E.O. labeled, curated, and measured the images. A.Z., W.T., J.J., and C.Z. performed the data analysis. F.R.M. and J.M.H. provided clinical consultation and measurements. A.Z., C.Z., X.F., W.T., and J.J. generated the figures and wrote the original draft. C.Z., W.L., X.F., and A.Z. supervised the work and revised the final manuscript. All authors reviewed and approved the final manuscript.

Data availability

The datasets supporting this study's findings are publicly accessible and include three infrared and visible image datasets: human disease dataset (https://github.com/najafian-lab/fpw-dl-v1), mice and rat disease model dataset (https://github.com/CZCBLab/TEAMKidney)

Code availability

The code is available at https://github.com/CZCBLab/TEAMKidney.

Disclosures

SRV, SPB, and HY are current and former Pfizer employees who helped generate the PHN rat kidney TEM images. WL and JMP received a research grant from Pfizer Centers for Therapeutic Innovation. All other authors have nothing to disclose.

Acknowledgments

We thank Casey Ritenour, Cheryl Tyszkiewicz, Seo-Kyoung Hwang, Chang-Ning Liu, and Lindsay Tomlinson of Pfizer Global Pathology and Comparative Medicine and Drug Safety R&D for help with rat PHN models. We thank Richa Sharma, Sudhir Kumar, and Boston University's Animal Science Center for their help in generating Col4a3 and ILK knockout mice. This work is supported by research grants from NIH/NIDDK grants R21DK143406 (CZ, WL), R01DK133940 (WL), R01DK078226 (WL), and a DOD grant E01 HT9425-23-1-1058 (WL). This work is also supported by BIT-ARC and B4D-ARC grants from Boston University Evans Center (WL, CZ, JMH), a grant from Pfizer Centers for Therapeutic Innovation (WL), a BU CTSI Integrated Pilot Grant funded by the Department of Medicine (CZ, WL), and the Boston University Undergraduate Research Opportunities Program (JJ). This publication is also supported in part by the National Center for Advancing Translational Sciences, National Institutes of Health, through BU-CTSI Grant Number 1UL1TR001430. Its contents are solely the authors' responsibility and do not necessarily represent the official views of the NIH. The funders had no role in study design, data collection and analysis, decision to publish, or preparation of the manuscript

843 **Table 1. Data Collection and Selection Criteria.**

Group	Total	Manually labeled	Manually measured	Training & Validation	Evaluation
Mouse models					
WT	107	48	20	48	61
ILK cKO	241	67	20	67	174
Col4A3	93	0	20	0	93
Rat models					
WT	69	0	20	0	69
PHN	584	540	20	540	44
Human models					
Normal	1361	71	20	0	1361
Fabry males	4301	76	20	468	3833
Fabry females	2487	40	20	0	2487
DKD	2463	33	20	0	2463

846 **Table 2. Comparison Between Proposed Model and Other Published Methods.**

Method	Measurement of FP	Measurement of GBM	Label Dependency	Segmentation Target	DL model	Open Source	Year
Wu et al.[22]	No	Yes	N/A	N/A	N/A	No	2010
Yang et al.[28]	No	No	Supervised	GBMDD	U-Net	No	2022
Smerkous et al.[29]	Yes	No	Supervised	PGBMI, SD	U-Net	Yes	2023
Unnersjo-Jess et al.[19]	Yes	No	Supervised	PFP, SD	U-Net	No	2023
Lin et al.[25]	No	No	Self-supervised	GBM, PFP	GCLR	No	2023
Wang et al.[23]	No	Yes	Supervised	GBM	RADS-Net	No	2024
Laudon et al.[24]	Yes	Yes	Supervised	GBM	U-Net	Yes	2025
Yamashita et al.[26]	Only Effacement	Yes	Supervised	GBM, PFP	DeepLab V3	No	2025
Zhang et al.[27]	No	Yes	Supervised	GBM, PFP	U-Net	No	2025
Ma et al.[32]	No	Yes	Supervised	GBM	YOLO+SAM	No	2025
Our Method	**Yes**	**Yes**	**Supervised**	**GBM, PFP**	**Glom2Mask**	**Yes**	**2025**

Figure legends

Fig. 1. Lack of existing models for accurate TEM image analysis and overview of the proposed framework. (A) Example TEM image of a podocyte-specific ILK conditional knockout mouse. (B) The boxed region of the glomerular filtration barrier in (A) with manually measured PFP in red and GBM in blue. (C) Ground truth segmentation of PFP and GBM. Segmentation results obtained using (D) FPW-DL, (E) SAM, and (F) MedSAM models. (G) Workflow of the proposed TEAMKidney framework for augmentation, segmentation, and quantitative analysis of mouse, rat, and human TEM images. (H) The proposed Glom2Mask network architecture for panoptic segmentation. (I) Example image showing post-segmentation of GBM and PFP. (J) Boxed region in (I) illustrating the initial skeleton extraction. (K) Demonstration of skeleton extraction error correction by the post-segmentation analysis algorithm.

Fig. 2. Detailed structure and evaluation of the TEAMKidney framework. (A) Workflow of the segmentation stage incorporating a self-training strategy. (B) Performance and training time comparison of different backbone networks for semantic segmentation across datasets. (C) Representative TEM images from Rat samples segmented by the three backbone networks for qualitative comparison. (D) Performance and training time comparison between the proposed Glom2Mask model and the benchmark Mask2Former for panoptic segmentation. (E) Representative image showing segmentation results from Glom2Mask and Mask2Former. The GBM is segmented as a stuff class (blue), and PFP as thing instances (various shades of red). (F) Comparison of GBM and PFP measurements between the TEAMKidney algorithm and ground truth across different species.

Fig. 3. Application of TEAMKidney to three animal models of kidney disease. (A) Schematic illustrating two mouse models of glomerular disease: knocking out Col4a3 or specific podocyte

ILK. (B) Schematic of a rat model of glomerular disease (PHN) induced by anti-Fx1a antibody

injection. (C) Representative TEM images analyzed by the TEAMKidney model for WT, Col4a3

KO, and ILK cKO mice. (D) Violin plots showing quantitative measurements of PFP and GBM

widths across WT, Col4a3 KO, and ILK cKO mice. (E) Representative TEAMKidney predicted

images for WT and PHN rat samples. (F) Violin plots comparing PFP and GBM widths between

WT and PHN groups.

Fig. 4. Application of TEAMKidney to human kidney diseases. (A) Schematic illustration of

the pathogenesis of DKD and Fabry disease. (B) Representative clinical TEM images from normal,

Fabry disease, and DKD patients were analyzed using the TEAMKidney model. (C) Violin plots

showing the average GBM and PFP widths per glomerulus across normal, Fabry disease, and DKD

samples. Each point represents the mean width of a single glomerulus. (D) Representative

TEAMKidney predicted images from male and female Fabry disease patients. (E) Violin plots

comparing GBM and PFP widths between male and female Fabry disease groups. (F) Violin plots

illustrating the distributions of GBM and PFP widths across different magnifications in human

kidney TEM images.

Fig. 5. Comparison of PFP quantification between TEAMKidney and FPW-DL methods.

Scatter plots with diagonal reference lines show the agreement between model predictions and

ground truth measurements for (A) average PFP width and (B) PFP counts. (C) Representative

TEM images with segmentation masks generated by TEAMKidney (left) and FPW-DL (right).

From top to bottom: normal human kidney, human DKD, ILK cKO mouse model, and PHN rat

model.

Fig. 6. Evaluation of concordance between expert manual measurements and TEAMKidney

predictions on human clinical TEM images. (A) GBM mean width. (B) PFP mean width. (C)

PFP count. From left to right: Normal kidney samples, DKD samples, Fabry disease samples from

male patients. (D) Fabry disease samples from female patients.

Fig. 7. TEAMKidney aligned with clinical routine measurement protocols. (A) Schematic

illustration of the clinical manual annotation process for GBM and PFP. Representative example

showing (B) the original TEM image, (C) GBM measurement following the clinical manual

procedure, (D) PFP counting following the clinical manual procedure, and (E) segmentation masks

predicted by TEAMKidney.

Table 1. Data Collection and Selection Criteria.

Table 2. Comparison Between Proposed Model and Other Published Methods.

Fig. 1

Fig. 2

Fig. 3

Fig. 4

Fig. 5

Fig. 6

Fig. 7

A

Acquire TEM Image

Open ImageJ (or FIJI)

Calibrate Image Using Scale Bar

GBM Measurement | **PFP Measurement**

GBM Measurement:
- Overlay 1.5 μm x 1.5 μm Grid
- Measure GBM Thickness at Each Grid Intersection
- Gather ≥ 200 Measurements from ≥ 3 Glomeruli
- Analyze Data Using OIM

PFP Measurement:
- Identify Foot Processes
- Mark Lateral Edges
- Measure Foot Process Width
- Collect and Analyze Data

B

C

D

E