OCR-Stats: Robust estimation and statistical testing of mitochondrial

respiration activities using Seahorse XF Analyzer

Vicente A. Yépez M.[a,b], Laura S. Kremer[c,d], Arcangela Iuso[c,d], Mirjana Gušić[c,d],

Robert Kopajtich[c,d], Eliška Koňaříková[c,d], Agnieszka Nadel[c,d], Leonhard Wachutka[a],

Holger Prokisch[c,d] and Julien Gagneur[a,b,*]

a. Department of Informatics, Technical University Munich, Boltzmannstr. 3, 85748 Garching,
Germany.
b. Quantitative Biosciences Munich, Gene Center, Department of Biochemistry, Ludwig-
Maximilians Universität München
c. Institute of Human Genetics, Helmholtz Zentrum München, Ingolstädter Landstr. 1, 85764
Neuherberg, Germany.
d. Institute of Human Genetics, Klinikum rechts der Isar, Technical University Munich,
Ismaninger Str. 22, 81675 München, Germany.

*Correspondence to: J.G. (gagneur@in.tum.de)

Abstract

Accurate quantification of cellular and mitochondrial bioenergetic activity is of great

interest in many medical and biological areas. Mitochondrial stress experiments

performed with Seahorse Bioscience XF Analyzers allow estimating 6 bioenergetics

measures by monitoring oxygen consumption rates (OCR) of living cells in multi-well

plates. However, detailed statistical analyses of OCR measurements from XF

Analyzers have been lacking so far. Here, we performed 126 mitochondrial stress

experiments involving 203 fibroblast cell lines to understand how OCR behaves

across different biosamples, wells, and plates; which allowed us to statistically model

OCR behavior over time. We show that the noise of OCR is multiplicative and that

outlier data points can concern individual measurements or all measurements of a

well. Based on these insights, we developed a novel statistical method, OCR-Stats,

that: i) models multiplicative noise, ii) automatically identifies outlier data points and

outlier wells, and iii) takes into account replicates both within and between plates.

This led to a significant reduction of the coefficient of variation across experiments of

basal respiration by 36% ($P = 0.004$), and of maximal respiration by 32% ($P = 0.023$).

Also, we propose an optimal experimental design with a minimum number of well

replicates needed to obtain confident results. Finally, we use statistical testing taking

into account the inter-plate variation to compare the bioenergetics measures of two

samples.

Keywords: Oxygen Consumption Rate (OCR); mitochondrial respiration;
bioenergetics; statistical testing; outlier detection.

1. Introduction

Mitochondria are double membrane enclosed, ubiquitous, maternally inherited,
cytoplasmic organelles present in most eukaryotic organisms (Gorman et al., 2016).
They are the powerhouses of the cell (Bhola et al., 2016; Sun et al., 2016), and are
also involved in regulating reactive oxygen species (Wallace, 2007), apoptosis (Bhola
et al., 2016), amino acid synthesis (Birsoy et al., 2015; Sullivan et al., 2015), cell
proliferation (Sullivan et al., 2015), cell signaling (Zong et al., 2016), and in the
regulation of innate and adaptive immunity (Weinberg et al., 2015). It follows that a
decline in mitochondrial function, reflected by a diminished electron transport chain
activity, is implicated in many human diseases ranging from rare genetic disorders
(Titov, Cracan et al., 2016) to common disorders such as cancer (Wallace, 2012;
Zong et al., 2016), diabetes (Dunham-Snary et al., 2014), neurodegeneration (Yao et
al., 2009), and aging (Sun et al., 2016). One of the most informative assessments of
mitochondrial function is the quantification of cellular respiration, as it directly reflects
electron transport chain impairment (Titov, Cracan et al., 2016) and depends on
many sequential reactions from glycolysis to oxidative phosphorylation (Koopman et
al., 2016). Estimations of oxygen consumption rates (OCR) expressed in pmol/min,
which are mainly driven by mitochondrial respiration through oxidative
phosphorylation, and extracellular acidification rates (ECAR) expressed in mpH/min,
which reflect glycolysis (Divakaruni et al., 2014; Ferrick et al., 2008; Koopman et al.,
2016), are more conclusive for the ability to synthesize ATP and mitochondrial
function than measurements of intermediates (such as ATP or NADH) and potentials
(Brand et al., 2011; Dmitriev et al., 2012).
OCR was classically measured using a Clark-type electrode, which required a
substantial amount of purified mitochondria, was time consuming, and did not allow
automated injection of compounds (Wu et al., 2007). Also, experimentation with
isolated mitochondria is ineffective because cellular regulation of mitochondrial
function is removed during isolation (Hill et al., 2012). In the last few years, a new
technology using fluorescent oxygen sensors (Gerencser et al., 2009) in a microplate
assay format has been developed by the company Seahorse Bioscience (now part of
Agilent Technologies) (Ribeiro et al., 2015). It allows simultaneous, real-time
measurements of both OCR and ECAR in multiple cell lines and conditions, reducing
the amount of required sample material and increasing the throughput (Divakaruni et
al., 2014; Ribeiro et al., 2015).
Typically, OCR and ECAR are measured using the Seahorse XF Analyzer in 96 (or
24) well-plates at multiple time steps under three consecutive treatments (Fig. 1B), in
a procedure known as mitochondrial stress test (Agilent Technologies, 2017). Under
basal conditions, complexes I-IV exploit energy derived from electron transport to

pump protons across the inner mitochondrial membrane. The thereby generated proton gradient is subsequently harnessed by complex V to generate ATP. Blockage of the proton translocation through complex V by oligomycin represses ATP production and prevents the electron transport throughout complexes I-IV due to the unexploited gradient. Administration of FCCP, an ionophor, subsequently dissipates the gradient uncoupling electron transport from complex V activity and increasing oxygen consumption to a maximum level. Finally, mitochondrial respiration is completely halted using the complex I inhibitor Rotenone. This approach is label-free and non-destructive, so the cells can be retained and used for further assays (Ferrick et al., 2008). OCR differences between different stages of these procedures provide estimation of six different bioenergetics measures: basal respiration, proton leak, non-mitochondrial respiration, ATP production, spare respiratory capacity, and maximal respiration (Brand et al., 2011; Divakaruni et al., 2014) (Figure 1). Increase in proton leak and decrease in maximum respiratory capacity are indicators of mitochondrial dysfunction (Brand et al., 2011). ATP production, basal respiration, and spare capacity alter in response to ATP demand, which is not necessarily mitochondrion-related as it may be the consequence of deregulation of any cellular process altering general cellular energy demand.

Current literature describing the Seahorse technology addressed experimental aspects regarding sample preparation (Dranka et al., 2011; Zhang et al., 2012), the amount of cells to seed (Zhang et al., 2012; Zhou et al., 2012), and compound concentration in different organisms (Dranka et al., 2011; Koopman et al., 2016; Shah-Simpson et al., 2016). However, studies regarding statistical best practices for determining OCR levels and testing them against another are lacking. The sole definition of bioenergetic measure varies between authors, as well as the number of time points in each interval (one time point in (Dranka, Hill, & Darley-Usmar, 2010), two time points in (Chacko et al., 2014) and four or more time points in (Dunham-Snary et al., 2014)); and whether differences (Invernizzi et al., 2012; Koopman et al., 2016; Sullivan et al., 2015), ratios (Yao et al., 2009; Zhang et al., 2011), or both (Shah-Simpson et al., 2016; Zhou et al., 2012) should be computed. Consequently, comparison of results across studies is difficult. Moreover, statistical power analyses for experimental design are often not provided. Differences in OCR between distinct biosamples (e.g. patient vs. control, or gene knockout vs. WT) can be as low as 12 – 30% (Almontashiri et al., 2014; Mitsopoulos et al., 2015; Stroud et al., 2016). Therefore, to design experiments with appropriate power to significantly detect such differences, it is important to know the source and amplitude of the variation within each sample, and reduce it as much as possible.

Here, we developed statistical good practices to support experimentalists in
designing, analyzing, and reporting results of Seahorse mitochondrial stress
experiments. To this end, we analyzed a large dataset of 126 mitochondrial stress
experiments in 96-well plate format involving 203 different fibroblast cell lines (Table
S1). The large amount of between-plate and within-plate replicates allowed us to
statistically characterize the nature and amount of biases and random variations in
these data. Based on these insights, we developed a statistical procedure, called
OCR-Stats, to extract robust and accurate oxygen consumption rates for each well,
which translates into robust summarized values of the multiple replicates inside one
plate and across plates. OCR-Stats includes normalization of raw data and outlier
identification and controls for well and plates biases, which led to significant
increased in accuracy over state-of-the-art methods. Between-well and between-
plate biases, as well as random variations, were found to be essentially multiplicative.
This motivated for a definition of bioenergetics measures based on ratios. We
formally defined 5 such measures: ETC-dependent OC proportion, ATPase-
dependent OC proportion, ETC-dependent proportion of ATPase-independent OC,
and Maximal OC fold change (Fig. 1A). We provide estimators for each one that were
empirically normally distributed, which permitted using linear regression models for
assessing statistical significance of bioenergetics measures comparisons.
Furthermore, our study provides experimental design guidance by i) showing that
between-plate variation largely dominates within-plate variation, implying that it is
important to seed the same biosamples in multiple plates, and ii) providing estimates
of variances within and between plates for each bioenergetic measure allowing for
statistical power computations. A free and pose source implementation of OCR-stats
in the statistical language R is provided at github.com/gagneurlab/OCR-Stats.

2. Results

2.1 Experimental design and raw data

We derived OCR, ECAR, and cell number for 203 dermal fibroblast cultures derived
from patients suffering from rare mitochondrial diseases, and control cells from
healthy donors (normal human dermal fibroblasts - NHDF, Methods, Table S1).
These were assayed in 126 plates, all using the same protocol (Methods). We grew
27 cell lines multiple times and placed them in more than one plate. We will refer to
these growth replicates as different biosamples. The NHDF cell line was seeded in all
plates for assessment of potential systematic plate biases. All four corners of each
plate were left as blank, i.e. filled with media but no cells to control for changes in
temperature (Dranka et al., 2011). The typical layout of a plate is depicted in Fig. 1C,

showing how each biosample is present in many well replicates. We seeded between
3 and 7 biosamples per plate (median = 4). This variation reflects typical set-ups of
experiments in a lab performed over multiple years.
We used the standard mitochondrial stress test assay (Fig. 1A, (Agilent
Technologies, 2017)) leading to four time intervals with three time points each and
denoted by Int_1 (resting cells), Int_2 (after oligomycin), Int_3 (after FCCP) and Int_4 (after
Rotenone). Wells for which the median OCR level did not follow the expected order,
namely, $median(OCR(Int_3))$ > $median(OCR(Int_1))$ > $median(OCR(Int_2))$ >
$median(OCR(Int_4))$, were discarded (977 wells, 10.47%). We also excluded from the
analysis contaminated wells and wells in which the cells got detached (461 wells,
4.94%, Methods).
161
162 **2.2 Random and systematic variations between replicates within plates**
163 Typical replicate time series are shown in Fig. 2A, with data from 12 wells for a single
164 biosample in a single plate. It shows the kinds of variations that we observed.
165
First, outlier data points occurred frequently. We distinguished two different types of
outliers: entire series for a well (e.g., well G5 in Fig. 2A) and individual data points
(e.g., well B6 at time point 6 in Fig. 2A). In the latter case, eliminating the entire
series for well B6 would be too restrictive, and would result in losing valuable data
from the other 11 valid time points. Therefore, methods to find outliers considering
these two possibilities must be devised.
172
Second, we noticed that the higher the OCR value, the higher the variance between
replicates, suggesting that the error is multiplicative. Unequal variance, or
heteroscedasticity, can strongly affect the validity of statistical tests and the
robustness of estimations. We therefore suggest modeling OCR on a logarithmic
scale, where the dependency between variance and mean disappears (Figs. 2B, 2C).
Respiratory chain enzyme activities such as NADH-ubiquinone reductase have
already been shown to obey log-normal distributions (Hautakangas et al., 2016).
180
Third, we observed systematic biases in OCR between wells (e.g., OCR values of
well C6 are among the highest while OCR values of well B5 are among the lowest at
all time points, Fig. 2A). Variations in cell number, initial conditions, treatment
concentrations, and fluorophore sleeve calibration can lead to systematic differences
between wells, which we refer to as well biases. To investigate whether well biases
could be mostly corrected using cell number as suggested in (Dranka et al., 2010),

we counted the number of cells after the experiments using Cyquant (Methods). As

expected, median OCR for each interval grows linearly with cell number measured at

the end of the experiment (Spearman rho between 0.32 and 0.47, P < 2.2e-16, Fig.

S1A). However, the relation is not perfect reflecting important additional sources of

variations, and also possible noise in measuring cell number. Strikingly, dividing OCR

by cell count led to a higher coefficient of variation (standard deviation divided by the

mean) between replicate wells than without that correction (Fig. S1B). This analysis

showed that normalization for cell number should not be done simply by a blunt

division by raw cell counts and motivated us to derive another method to capture well

biases.

2.3 A statistical model of OCR

Building on these insights, we next introduced a statistical model of OCR within plate.

For a given biosample in one plate, we modeled the logarithm of OCR $y_{w,t}$ of well w

at time point t as a sum of well bias, interval effects and noise, i.e.,:

$$y_{w,t} = \alpha_{i(t)} + \beta_w + \varepsilon_{w,t}. \quad (1)$$

201   The term $\alpha_{i(t)}$ is the effect of the interval $i(t)$ of time point t. The term β_w is the relative

202   bias of well w compared to a reference well, which is set arbitrarily and corresponds

203   to the first well in alphabetical order. The term $\varepsilon_{w,t}$ is the error.

204   We defined the OCR levels (θ_i) as the expected log OCR per interval, averaged over

205   all wells:

$$\hat{\theta}_i = \alpha_i(t) + \frac{\sum_w \beta_w}{n}, \quad (2)$$

where n is the number of wells.

Note that the well bias is modeled independently for each plate, i.e., the bias of a

certain well in one plate is different from the bias of the well at the same location in

another plate.

We present now our OCR-Stats algorithm, for a given plate:

1. Fit the log linear model (1) using the least-squares method, which consists in

212   minimizing $\sum_w \sum_t (y_{w,t} - \alpha_{i(t)} - \beta_w)^2$, thus obtaining the coefficients α_i, β_w; and

213            $\hat{\theta}_i$ using (2).

214   2. For each time point t in interval i and well w, define the OCR residual:

215            $e_{w,t} = y_{w,t} - \hat{\theta}_{i(t)}$, which is used to identify outliers (Methods).

216   3. Identify and remove well level outliers, fit again, iteratively, until no more are

217   found.

218   4. Identify and remove single point outliers, fit again, iteratively, until no more

219   are found.

220   5. Scale back to natural scale in order to compute the bioenergetics measures
221   (e.g.: Basal respiration $= e^{\theta_1} - e^{\theta_4}$, Maximal respiration $= e^{\theta_3} - e^{\theta_4}$, etc.), or
222   take the difference in the logarithmic scale to obtain the metrics from Table 1.

2.4 Variations within plates

We were then interested in determining the amplitude of the variance inside each plate in order to compute the number of wells needed to obtain robust estimates $\hat{\boldsymbol{\theta}}$. Using only the controls NHDF, we computed the standard deviation $\sigma_{i,j}^w$ of the logarithm of OCR across all wells for each plate j and interval i. Then, we computed the median across plates, thus obtaining one value σ_i^w per interval ($\sigma_1^w = 0.10, \sigma_2^w = 0.13, \sigma_3^w = 0.12, \sigma_4^w = 0.16$). As we worked in the logarithmic scale, the error in the natural scale becomes multiplicative and relative. The standard error of the estimates $\hat{\boldsymbol{\theta}}$ can be expressed as $\sigma_{\hat{\theta}_i} = \sigma_i^w / \sqrt{n_w}$, where n_w is the number of wells. The highest value of σ_i^w was 0.16, so in order to get a relative error of 5%, cells should be seeded in 10 wells. This result comes from a variation after removing outliers, so considering that around 16.5% of wells were found to be outliers, then ideally we should use $10/(1 - 0.165) \approx 12$ wells per biosample.

2.5 Variations between plates

After analyzing the variation among wells inside plates, we set up to study the variation across multiple plates. Using data from the controls NHDF, we found that the variability between plates for all four intervals is much larger than between wells (Table S2, Fig. S4). We next asked whether there exists a systematic plate bias that could be corrected for. We indeed observed a similar increase in OCR on the interval 1 for both biosamples on plate #20140430 with respect to #20140428 (Fig. 3A). To test whether this tendency held across every repeated biosample, we compared all replicate pairings with their respective NHDF controls and found a positive correlation (Fig. 3B). These differences can come from changes in temperature or the use of different sensor cartridges (Koopman et al., 2016). Because the plate biases are systematic, we can correct for them using a log linear model (Methods). Nonetheless, the biases do not explain all the between plate variation as the remaining variance is large (relative variance of the residuals: I_1: 49.8%, I_2: 51.6%, I_3: 65.6% and I_4: 55.9%). It is therefore important to perform multiple plate analyses to be able to conclude for a reproducible systematic difference between biosamples.

2.6 Statistical comparison between biosamples

254 In order to compare the bioenergetics measures of two biosamples, we first need to
255 decide if it is better evaluating differences or ratios of the OCR levels in the natural
256 scale. Even after correcting for well biases, there is a remaining cell number effect
257 (Fig. 3C); therefore, we recommend working with ratios of OCR levels (or differences
258 in the logarithmic scale). We propose the following definitions:

OCR ratios	Abbr.	Metrics	Tested differences d	Equivalent
ETC-dependent OC proportion	E/I – proportion	$$\frac{OCR_1 - OCR_4}{OCR_1} =$$ $$1 - \exp(\theta_{Ei} - \theta_I)$$	$(\theta_{I,f} - \theta_{Ei,f})$ $- (\theta_{I,\text{ctrl}}$ $- \theta_{Ei,\text{ctrl}})$	Basal Respiration
ATPase-dependent OC proportion	A/I – proportion	$$\frac{OCR_1 - OCR_2}{OCR_1} =$$ $$1 - \exp(\theta_{Ai} - \theta_I)$$	$(\theta_{I,f} - \theta_{Ai,f})$ $- (\theta_{I,ctrl}$ $- \theta_{Ai,ctrl})$	ATP Production
ETC-dependent proportion of ATPase-independent OC	E/Ai - proportion	$$\frac{OCR_2 - OCR_4}{OCR_2} =$$ $$1 - \exp(\theta_{Ei} - \theta_{Ai})$$	$\left(\theta_{Ai,f} - \theta_{Ei,f}\right)$ $- (\theta_{Ai,ctrl}$ $- \theta_{Ei,ctrl})$	Proton Leak
Maximal OC fold change	M/I – fold change	$$\frac{OCR_3}{OCR_1} =$$ $$\exp(\theta_M - \theta_I)$$	$(\theta_{M,f} - \theta_{I,f})$ $- (\theta_{M,ctrl}$ $- \theta_{I,ctrl})$	Spare Capacity
Maximal over ETC-independent OC fold change	M/Ei – fold change	$$\frac{OCR_3}{OCR_4} =$$ $$\exp(\theta_M - \theta_{Ei})$$	$(\theta_{M,f}$ $- \theta_{Ei,f})$ $- (\theta_{M,ctrl}$ $- \theta_{Ei,ctrl})$	Maximal Respiration

Table 1: *OCR ratios, metrics and equivalents*

259 Then, for any given OCR ratio *b* (eg. M/Ei - fold change), we test differences of log

260 OCR ratios of patient versus a control cell line (Table 1) using the following linear

261 model:

$$d_{b,f,p} = \mu_{b,f} + \epsilon_{b,f,p}, (3)$$

where $d_{b,f,p}$ corresponds to the difference of ratio *b* of a cell line *f* and the respective

control on plate *p*. We solve it using linear regression, thus obtaining one value $\mu_{b,f}$

per each ratio *b* and cell line *f*. We then compare these $\mu_{b,f}$ values (which follow a t-

Student distribution) against the null hypothesis $\mu_{b,f} = 0$ to obtain p-values and

confidence intervals (Figs. 4A, 4B, Methods).

2.7 Benchmark of OCR-Stats algorithm

In order to benchmark the OCR-Stats algorithm, we computed the coefficient of

variation (standard deviation divided by mean) of the six bioenergetics measures in

the natural scale of all repeated biosamples across plates. The lower the coefficient

of variation among replicates, the better the method. We cannot test using the final

estimates $\hat{\theta}^f$ after correcting for plate effect, because we would fall into circularity as

correcting using $\beta_{i,p}$ forces replicates to have a closer value. Therefore, just for

benchmarking purposes, we corrected for plate effect using only the data from the

controls NHDF *c* of each plate, namely:

$$y_{i,p}^c = \beta_0^c + \beta_i^c + \beta_p^c + \varepsilon_{i,p}. (4)$$

We solved (4) using linear regression and used the effects β_p^c as offsets in (1), and

recomputed $\hat{\theta}_i$ values accordingly. We scaled back to natural scale to calculate the

bioenergetics measures and the coefficient of variation of all repeated biosamples

(except the control to avoid circularity) using: i) the default Extreme Differences (ED)

method (Methods) provided by the vendor, ii) the log linear (LL) corresponding to

steps 1 and 2 of the OCR-Stats algorithm, iii) complete OCR-Stats (LL + outlier

removal), and iv) OCR-Stats after correcting for plate effect (OCR-PE) using (4).

Each step contributed to lowering the coefficient of variation, obtaining a final

significant reduction of 36% and 32% in basal and maximal respiration, respectively,

from OCR-PE with respect to ED ($P < 0.03$, one-sided Wilcoxon test) (Fig. 5).

2.8 Benchmark of OCR-Stats statistical testing method

We applied OCR-Stats, Extreme Differences with Wilcoxon test within each plate

(within-plate ED), and Extreme Differences with Wilcoxon test across plates (across-

plate ED) to obtain the M/Ei ratio and maximal respiration (MR) of all the 26 cell lines

that were seeded in more than one plate (Methods). For every approach we

computed p-values for significant fold-changes against the controls. Six of these cell
lines come from patients with rare variants in genes associated with an established
cellular respiratory defect, allowing for assessing the sensitivity of each approach
(Table S3, (Haack et al., 2013; Hildick-Smith et al., 2013; Kremer et al., 2017;
Pronicka et al., 2016; Van Haute et al., 2016)). Also, two cell lines (#73901 and
#91410) that showed no significant respiratory defects in earlier studies (Powell et
al., 2015) (Kremer et al., 2016) served as negative controls.
298
The within-plate ED method reported significantly higher or lower MR for
56/69=81.2% biosamples (Figs. 4A, 4B, Table S3). Moreover, every cell line was
found to be significant on at least one plate, despite large variation in M/Ei fold
change between plates (Fig. 4A). Also, for 11 cell lines, one plate at least also gave
non-significant differences. These results show the importance of assessing
differences using multiple plates and advocate for a more robust approach than
within-plate ED.
306
307 One approach to take multiple plates into accounts is to perform a Wilcoxon test
308 based on per plate average ED values (across-plate ED, Methods). However, this
309 approach requires samples to be seeded in at least five plates in order to obtain
310 significant results. Here, only one cell line, #78661, was found significant this way.
311
In contrast, significance with the OCR-Stats statistical algorithm can be reached by
seeding a biosample in one plate only; provided there were other between-plate
replicates to compute the inter-plate variance. On this data, OCR-Stats was much
more conservative than within-plate ED and found only 7/26=27% cell lines to have
aggregated significantly lower M/Ei than the control. There was no evidence against
the normality and homoscedasticity assumption of OCR-stats as the quantile-quantile
plots of the residuals aligned well along the diagonal (Figs. 4C, S4). All the 6 positive
control cell lines were reported to have significantly lower M/Ei than control by OCR-
Stats (Figs. 4A, 4B, Table S3). Moreover, OCR-Stats did not report significant M/Ei
differences for the two negative controls. Altogether, these results show that OCR-
Stats successfully identifies and removes variation within and between plates,
providing more stable results which translates into less false positives.

Discussion and conclusion

Mitochondrial studies using extracellular fluxes (specifically the XF Analyzer from
Seahorse) are gaining popularity; therefore, it is of paramount importance to have a

proper statistical method to estimate the OCR levels from the raw data. In this paper,
we have developed such a model, which includes approaches to control for well and
plate biases, and automatic outlier identification. By doing so, we were able to
significantly reduce the coefficient of variation of replicates across plates. After
analyzing the intra-plate variation, we found that the minimum number of wells per
biosample should be 12.
We found that dividing cellular OCR by cell number was introducing more noise than
was seen for uncorrected data. Here, we seeded always the same number of cells.
Hence, the variations we observed in cell number at the end of the experiments are
largely overestimated by noise in measurements. In other experimental settings, in
which different numbers of cells are seeded, we suggest to include an offset term to
the model (1) equal to the logarithm of the seeded cell number to control for this
variation by design. Also, the Seahorse XF Analyzer can be used on isolated
mitochondria and on isolated enzymes, where a normalization approach is to divide
OCR by mitochondrial proteins or enzyme concentration (Seahorse Bioscience,
2014). However, as described here for cellular assays, robust normalization
procedures require careful analysis.

To use XF Seahorse Analyzers for large-scale experiments, one needs to be able to
compare biosamples measured on different plates. Our investigation showed that
there is roughly multiplicative bias between plates that can be controlled to some
extent by including control biosamples across plates, as we did here with NHDF. We
proposed an extension of our intra-plate robust linear regression approach to multiple
plates that can handle model this plate bias. However, we also noticed that the
assumption of a multiplicative plate bias is not sufficient as there are other sources of
variation. Therefore, for comparing two biosamples statistically, they need to be
placed on the same plate, and repeated multiple times. We demonstrated that it is
better to compare OCR ratios rather than differences as this eliminates sources of
variation like cell number. We proposed another linear model that takes into account
the inter-plate variation, which we showed to agree with previous results of patients
diagnosed with mitochondrial disorders.

We also encourage users to understand the biological meaning of each OCR ratio
(Table 1). For example, cell line #73387 was found to have a lower, but non
significantly ($P < 0.10$), M/Ei ratio (the most common metric used throughout the
literature, Table S3), but when analyzing its E/I proportion, we found that it was
drastically lower than the control ($P < 1.2x10^{-7}$). This result is consistent with its

genetic diagnosis (Table S1, (Oláhová et al., 2015)). For visualizing OCR ratios, raw
OCR vs. time plots are useful in both logarithmic and natural scales.

In principle, OCR-Stats should be able to estimate ECAR levels. Nevertheless,
similar analyses as performed here should be done beforehand in order to guarantee
that the method is indeed applicable. Preliminary investigations suggest that the
nature of noise (outliers, multiplicative) is similar than for OCR.

Finally, it is important to understand further sources of variations between plates, cell
cultures, treatments and other factors in order to correct for them. Here, we found
that gender does not significantly influence OCR levels (Fig. S5), but age (for which
we have no register), may play a role.

Methods

Biological material

374 All biosamples come from primary fibroblast cell lines of humans suffering from rare
375 mitochondrial diseases, established in the framework of mitoNet and GENOMIT. The
376 controls used are primary patient fibroblast cell lines, normal human dermal
377 fibroblasts (NHDF) from neonatal tissue, commercially available from Lonza, Basel,
378 Switzerland.

Measure of extracellular fluxes using Seahorse XF96

We seeded 20,000 fibroblasts cells in each well of a XF 96-well cell culture
microplate in 80 ml of culture media, and incubated overnight at 37°C in 5% CO_2.
The four corners were left only with medium for background correction. Culture
medium is replaced with 180 ml of bicarbonate-free DMEM and cells are incubated at
37°C for 30 min before measurement. Oxygen consumption rates (OCR) were
measured using a XF96 Extracellular Flux Analyzer (Agilent Technologies, 2017).
OCR was determined at four levels: with no additions, and after adding: oligomycin (1
μM); carbonyl cyanide 4-(trifluoromethoxy) phenylhydrazone (FCCP, 0.4 μM); and
rotenone (2 μM) (additives purchased from Sigma at highest quality). After each
assay, manual inspection was performed on all wells using a conventionally light
microscope.

Cell number quantification

Cell number was quantified using the CyQuant Cell Proliferation Kit (Thermo Fisher
Scientific, Waltham, MA, USA) according to the manufacturer's protocol. In brief,
cells were washed with 200 μL PBS per well and frozen in the microplate at -80°C to
ensure subsequent cell lysis. Cells were thawed and resuspended vigorously in 200
μL 1x cell-lysis buffer supplemented with 1x CyQUANT GR dye per well.

Resuspended cells were incubated in the dark for 5 min at RT whereupon

fluorescence was measured (excitation: 480 nm, emission: 520 nm).

Extreme Differences (default) Method to compute bioenergetics measures

On every plate independently, for each well, on interval 1 take the OCR

corresponding to the last measurement, on intervals 2 and 4 take the minimum and

on interval 3 the maximum OCR value (Divakaruni et al., 2014). Then, do the

corresponding differences to estimate the bioenergetics measures. Report the results

per patient as the mean across wells plus standard deviation or standard error,

separately for each plate.

Outlier Removal

For each sample s and well w, compute the mean across time points of its squared

residuals: $r_w := \mathrm{mean}_t(e_{w,t}{}^2)$, thus obtaining a distribution r. Identify as outliers the

wells whose $r_w > \mathrm{median}(r) + 5 \cdot \mathrm{mad}(r)$, where mad, median absolute deviation, is

a robust estimation of the standard deviation (Fig. S2A). We found that deviations by

5 mad from the median were selective enough in practice. Compute the vector of

estimates $\hat{\theta}$ using the remaining wells and iterate this procedure until no more wells

are identified as outliers. It required 8 iterations until convergence and around 16.5%

of all the wells were found to be outliers (Fig. S2B).

Single point outliers are identified in a similar way. After discarding the wells that

were found to be outliers in the previous step, categorize as outliers single data

points whose $e_{w,t}{}^2 > \mathrm{median}_t(e_{w,t}{}^2) + 5 \cdot \mathrm{mad}_t(e_{w,t}{}^2)$ (Fig. S2C). Iterate until no

more outliers are found. It required 19 iterations until convergence and approximately

6.1% of single points were found to be outliers (Fig. S2D).

Plate effect model

In an attempt to correct for plate effect, we propose a log linear model where the

levels θ' depend on interval i, samples s and plate p:

$$\theta'_{i,s,p} = \alpha_{i,s} + \beta_{i,p} + \varepsilon_{i,s,p},$$

thus obtaining one coefficient $\beta_{i,p}$ for each plate-interval combination. These effects

are added to the previous estimates: $\hat{\theta}^f_{i,s,p} = \hat{\theta}_{i,s} - \beta_{i,p}$, obtaining the final estimates

$\hat{\theta}^f$. As for (1), the model is solved using linear regression.

Multi-plate averaging method

In case of inter-plate comparisons, the multi-plate averaging methods takes the

average and standard error of the bioenergetics measures obtained using the ED

method of all repeated biosamples across plates (Agilent Technologies, 2016).

Statistical Testing

To evaluate the OCR ratios between a fibroblast f and a control, we use the

corresponding tested difference d (Table 1). For a fibroblast f located on a plate p, we

define $\mu_{i,j,f,p} := \left[\hat{\theta}_i - \hat{\theta}_j\right]_{f,p} - \left[\hat{\theta}_i - \hat{\theta}_j\right]_{NHDF,p}$, where i and j are any two different

intervals. From there, we can obtain a t-statistic: $t_{\hat{d}} = \frac{\mu - d_0}{se(\mu)}$, where $d_0 = 0$ as that is the

value against we want to compare μ against, and se is the standard error. The t-

statistic follows a t-distribution with $n - 2$ degrees of freedom, from which we can

obtain p-values. Moreover, we can obtain confidence intervals: $[\mu - se(\mu)t_{n-2}^{\alpha},\ \mu +$

$se(\mu)t_{n-2}^{\alpha}]$, where $(1 - \alpha)$ is the confidence level and t_{n-2}^{α} the $(1 - \alpha/2)$ quantile of

the t_{n-2} distribution. Note that the normality assumption holds for the residuals $\epsilon_{b,f,p}$

(Figs. 4C, S4).

Acknowledgements

We would like to thank Daniel Bader, Žiga Avsec, Jun Cheng and Paula Fernández-

Guerra for valuable discussions and manuscript revision. This study was supported

by the German Bundesministerium für Bildung und Forschung (BMBF) through the E-

Rare project GENOMIT (01GM1603 and 01GM1207, H.P. and T.M.), through the

Juniorverbund in der Systemmedizin 'mitOmics' (FKZ 01ZX1405A J.G., L.W. and

V.A.Y.M.) and the DZHK (German Centre for Cardiovascular Research, L.S.K.). A

Fellowship through the Graduate School of Quantitative Biosciences Munich (QBM)

supports V.A.Y.M.. H.P. is supported by EU FP7 Mitochondrial European Educational

Training Project (317433). J.G., V.A.Y.M. and H.P. are supported by EU Horizon2020

Collaborative Research Project SOUND (633974). We thank the Cell lines and DNA

Bank of Pediatric Movement Disorders and Mitochondrial Diseases of the Telethon

Genetic Biobank Network (GTB09003).

Author contributions

J.G. and H.P. planned the project and overviewed the research. H.P. designed the

experiments. V.A.Y.M. curated and analyzed the data. J.G. devised the statistical

analysis. L.S.K., A.I., E.K., M.G., and A.N. performed the mitochondrial stress test

experiments and cell number quantification. V.A.Y.M., L.W. and J.G. made the

figures. V.A.Y.M. and J.G. wrote the manuscript. All authors performed critical

revision of the manuscript.

Figure legends

462   **Figure 1. Principle of the mitochondrial stress test assay** (**A**) Cartoon illustration

463   of OCR levels (y-axis) versus time (x-axis). Injection of the three compounds

464   oligomycin, FCCP and rotenone delimit four time intervals within which OCR is

465   roughly constant. (**B**) Targets of each compound in the electron transport chain. (**C**)

466   Typical layout of a mitochondrial stress test 96-well plate.

Figure 2. OCR behavior over time. (**A**) Typical time series replicates inside a plate. Behavior of OCR expressed in pmol/min (y-axis) of Fibro_VY_017 over time (x-axis). Colors indicate the row, and shape the column of 12 well replicates. Variation increases for larger OCR values, OCR has a systematic well effect and there exist two types of outliers: well-level and single-point. (**B**) Scatterplot of standard deviation (y-axis) vs. mean (x-axis) of all 3 time replicates of each interval, well and plate of OCR of NHDF only, shows a positive correlation ($n = 409$). (**C**) Same as (B) but for the logarithm of OCR, where the correlation disappears.

Figure 3. Plate bias. (**A**) Log of OCR in interval 3 (y-axis) for the cell lines #65126 and NHDF (x-axis) which were seeded in 2 different plates (color-coded). The similar increase in OCR from plate #20140128 to #20140430 in both biosamples suggests that there is a systematic plate bias. (**B**) Scatterplots of the differences of the logarithm of OCR levels θ of all possible 2 by 2 combinations of repeated biosamples across experiments (y-axis) against their respective controls (NHDF) (x-axis) show that there exists a positive correlation (I_1: $\rho= 0.64$, $P < 2.3 \times 10^{-8}$, I_2: $\rho= 0.65$, $P < 1.2 \times 10^{-8}$, I_3: $\rho= 0.52$, $P < 1.2 \times 10^{-5}$, I_4: $\rho= 0.64$, $P < 1.4 \times 10^{-8}$), suggesting a systematic plate bias ($n = 63$). **(C)** Scatterplot of the difference of log OCR levels of patients vs. control NHDF (both axes) of every interval with respect to another. All intervals correlate with each other even after removing plate bias (by subtracting control values).

Figure 4. Statistical testing of M/Ei fold change patient vs. control on multiple plates. (A) Ratio of M/Ei fold change (y-axis) of all cell lines repeated across plates (x-axis) and their respective control, sorted by p-value obtained using the OCR-Stats method. Left of the red dashed line are cell lines with significantly lower M/Ei fold change using OCR-Stats. Dots in orange represent cell lines with significantly lower or higher M/Ei fold change using the ED method. Highlighted positive (+) and negative (-) controls. **(B)** Similar as (A), but depicting the p-value in logarithmic scale (y-axis) using OCR-Stats. Red dashed line at $P = 0.05$. Dots in red represent biosamples with significantly lower M/Ei fold change using the OCR-Stats method. **(C)** Quantile-quantile theoretical (x-axis) vs. observed (y-axis) plot of the residuals of the linear model (3) applied to M/Ei fold change.

Figure 5. Benchmark using coefficient of variation. Coefficient of variation (CV = standard deviation / mean, y-axis) of replicates across experiments ($n=26$) using different methods (x-axis) to estimate the 6 bioenergetics measures. In all, except for Spare Capacity, OCR-Stats with plate effect showed significantly lower variation with

respect to the Extreme Differences method. P-values obtained from one-sided paired

Wilcoxon test.

References

Agilent Technologies. (2016). Multi-File XF Report Generator, User Guide. Retrieved
from http://www.agilent.com/cs/library/usermanuals/public/Report Generator
User Guide_Seahorse XF Cell Mito Stress Test_MultiFile_RevA.pdf

Agilent Technologies. (2017). Mito Stress Test Kit, User Guide, (103016–400).
Retrieved from
https://www.agilent.com/cs/library/usermanuals/public/XF_Cell_Mito_Stress_Te
st_Kit_User_Guide.pdf

Almontashiri, N. A. M., Chen, H. H., Mailloux, R. J., Tatsuta, T., Teng, A. C. T.,
Mahmoud, A. B., … Stewart, A. F. R. (2014). SPG7 Variant Escapes
Phosphorylation-Regulated Processing by AFG3L2, Elevates Mitochondrial
ROS, and Is Associated with Multiple Clinical Phenotypes. *Cell Reports*, *7*(3),
834–847. http://doi.org/10.1016/j.celrep.2014.03.051

Bhola, P. D., & Letai, A. (2016). Mitochondria-Judges and Executioners of Cell Death
Sentences. *Molecular Cell*, *61*(5), 695–704.
http://doi.org/10.1016/j.molcel.2016.02.019

Birsoy, K., Wang, T., Chen, W. W., Freinkman, E., Abu-Remaileh, M., & Sabatini, D.
M. (2015). An Essential Role of the Mitochondrial Electron Transport Chain in
Cell Proliferation Is to Enable Aspartate Synthesis. *Cell*, *162*, 540–551.
http://doi.org/10.1016/j.cell.2015.07.016

Brand, M. D., & Nicholls, D. G. (2011). Assessing mitochondrial dysfunction in cells.
The Biochemical Journal, *435*(2), 297–312. http://doi.org/10.1042/BJ20110162

Chacko, B. K., Kramer, P. a, Ravi, S., Benavides, G. a, Mitchell, T., Dranka, B. P., …
Darley-Usmar, V. M. (2014). The Bioenergetic Health Index: a new concept in
mitochondrial translational research. *Clinical Science*, *127*(6), 367–373.
http://doi.org/10.1042/CS20140101

Divakaruni, A. S., Paradyse, A., Ferrick, D. A., Murphy, A. N., & Jastroch, M. (2014).
Analysis and interpretation of microplate-based oxygen consumption and pH
data. Methods in Enzymology (Vol. 547). http://doi.org/10.1016/B978-0-12-
801415-8.00016-3

Dmitriev, R. I., & Papkovsky, D. B. (2012). Optical probes and techniques for O 2
measurement in live cells and tissue. *Cellular and Molecular Life Sciences*, *69*,
2025–2039. http://doi.org/10.1007/s00018-011-0914-0

Dranka, B. P., Benavides, G. A., Diers, A. R., Giordano, S., Blake, R., Reily, C., …
Darley-Usmar, V. M. (2011). Assessing bioenergetic function in response to
oxidative stress by metabolic profiling. *Free Radical Biology and Medicine*,
51(9), 1621–1635. http://doi.org/10.1016/j.freeradbiomed.2011.08.005

Dranka, B. P., Hill, B. G., & Darley-Usmar, V. M. (2010). Mitochondrial reserve
capacity in endothelial cells: The impact of nitric oxide and reactive oxygen
species. *Free Radical Biology and Medicine*, *48*(7), 905–914.
http://doi.org/10.1016/j.freeradbiomed.2010.01.015

Dunham-Snary, K. J., Sandel, M. W., Westbrook, D. G., & Ballinger, S. W. (2014).
Redox Biology A method for assessing mitochondrial bioenergetics in whole
white adipose tissues. *Redox Biology*, *2*, 656–660.
http://doi.org/10.1016/j.redox.2014.04.005

Ferrick, D. A., Neilson, A., & Beeson, C. (2008). Advances in measuring cellular
bioenergetics using extracellular flux. *Drug Discovery Today*, *13*(March).
http://doi.org/10.1016/j.drudis.2007.12.008

Gerencser, A. A., Neilson, A., Choi, S. W., Edman, U., Yadava, N., Oh, R. J., …
Brand, M. D. (2009). Quantitative Microplate-Based Respirometry with

Correction for Oxygen Diffusion. *Analytical Chemistry*, *81*(16), 6868–6878.

Gorman, G. S., Chinnery, P. F., DiMauro, S., Hirano, M., Koga, Y., McFarland, R., …
Turnbull, D. M. (2016). Mitochondrial diseases. *Nature Reviews Disease*
Primers, *2*. http://doi.org/10.1038/nrdp.2016.80

Haack, T. B., Rolinski, B., Haberberger, B., Zimmermann, F., Schum, J., Strecker, V.,
… Prokisch, H. (2013). Homozygous missense mutation in BOLA3 causes
multiple mitochondrial dysfunctions syndrome in two siblings. *Journal of*
Inherited Metabolic Disease, *36*(1), 55–62. http://doi.org/10.1007/s10545-012-
9489-7

Hautakangas, M. R., Hinttala, R., Rantala, H., Nieminen, P., Uusimaa, J., &
Hassinen, I. E. (2016). Evaluating clinical mitochondrial respiratory chain
enzymes from biopsy specimens presenting skewed probability distribution of
activity data. *Mitochondrion*, *29*, 53–58.
http://doi.org/10.1016/j.mito.2016.05.004

Hildick-Smith, G. J., Cooney, J. D., Garone, C., Kremer, L. S., Haack, T. B., Thon, J.
N., … Paw, B. H. (2013). Macrocytic anemia and mitochondriopathy resulting
from a defect in sideroflexin 4. *American Journal of Human Genetics*, *93*(5),
906–914. http://doi.org/10.1016/j.ajhg.2013.09.011

Hill, B. G., Benavides, G. A., Jr, J. R. L., Ballinger, S., & Italia, L. D. (2012).
Integration of cellular bioenergetics with mitochondrial quality control and
autophagy. *Biological Chemistry*, *393*(12), 1485–1512.
http://doi.org/10.1515/hsz-2012-0198

Invernizzi, F., D'Amato, I., Jensen, P. B., Ravaglia, S., Zeviani, M., & Tiranti, V.
(2012). Microscale oxygraphy reveals OXPHOS impairment in MRC mutant
cells. *Mitochondrion*, *12*(2), 328–335. http://doi.org/10.1016/j.mito.2012.01.001

Koopman, M., Michels, H., Dancy, B. M., Kamble, R., Mouchiroud, L., Auwerx, J., …
Houtkooper, R. H. (2016). A screening-based platform for the assessment of
cellular respiration in Caenorhabditis elegans. *Nature Protocols*, *11*(10), 1798–
1816. http://doi.org/10.1038/nprot.2016.106

Kremer, L. S., Bader, D. M., Mertes, C., Kopajtich, R., Pichler, G., Iuso, A., …
Prokisch, H. (2017). Genetic diagnosis of Mendelian disorders via RNA
sequencing. *Nature Communications*, *8*, 15824.
http://doi.org/10.1038/ncomms15824

Kremer, L. S., Distelmaier, F., Alhaddad, B., Hempel, M., Iuso, A., Küpper, C., …
Haack, T. B. (2016). Bi-allelic Truncating Mutations in TANGO2 Cause Infancy-
Onset Recurrent Metabolic Crises with Encephalocardiomyopathy. *American*
Journal of Human Genetics, *98*(2), 358–362.
http://doi.org/10.1016/j.ajhg.2015.12.009

Mitsopoulos, P., Chang, Y.-H., Wai, T., König, T., Dunn, S. D., Langer, T., &
Madrenas, J. (2015). Stomatin-like protein 2 is required for in vivo mitochondrial
respiratory chain supercomplex formation and optimal cell function. *Molecular*
and Cellular Biology, *35*(10), 1838–47. http://doi.org/10.1128/MCB.00047-15

Oláhová, M., Hardy, S. A., Hall, J., Yarham, J. W., Haack, T. B., Wilson, W. C., …
Taylor, R. W. (2015). LRPPRC mutations cause early-onset multisystem
mitochondrial disease outside of the French-Canadian population. *Brain*,
138(12), 3503–3519. http://doi.org/10.1093/brain/awv291

Powell, C. A., Kopajtich, R., D'Souza, A. R., Rorbach, J., Kremer, L. S., Husain, R.
A., … Minczuk, M. (2015). TRMT5 Mutations Cause a Defect in Post-
transcriptional Modification of Mitochondrial tRNA Associated with Multiple
Respiratory-Chain Deficiencies. *American Journal of Human Genetics*, *97*(2),
319–328. http://doi.org/10.1016/j.ajhg.2015.06.011

Pronicka, E., Piekutowska-Abramczuk, D., Ciara, E., Trubicka, J., Rokicki, D.,
Karkucińska-Więckowska, A., … Płoski, R. (2016). New perspective in
diagnostics of mitochondrial disorders: two years' experience with whole-exome
sequencing at a national paediatric centre. *Journal of Translational Medicine*,

14(1), 174. http://doi.org/10.1186/s12967-016-0930-9

Ribeiro, S. M., Giménez-cassina, A., & Danial, N. N. (2015). Measurement of
Mitochondrial Oxygen Consumption Rates in Mouse Primary Neurons and
Astrocytes. *Methods in Molecular Biology*, *1241*, 59–69.
http://doi.org/10.1007/978-1-4939-1875-1

Seahorse Bioscience. (2014). Normalizing XF metabolic data to cellular or
mitochondrial parameters, User Guide. Retrieved from
http://hpst.cz/sites/default/files/attachments/appnote-normalizing-metabolic-
data.pdf

Shah-Simpson, S., Pereira, C. F. A., Dumoulin, P. C., Caradonna, K. L., & Burleigh,
B. A. (2016). Bioenergetic profiling of Trypanosoma cruzi life stages using
Seahorse extracellular flux technology. *Molecular and Biochemical Parasitology*,
208(2), 91–95. http://doi.org/10.1016/j.molbiopara.2016.07.001

Stroud, D. A., Surgenor, E. E., Formosa, L. E., Reljic, B., Frazier, A. E., Dibley, M. G.,
… Ryan, M. T. (2016). Accessory subunits are integral for assembly and
function of human mitochondrial complex I. *Nature*, *538*(7623), 1–17.
http://doi.org/10.1038/nature19754

Sullivan, L. B., Gui, D. Y., Hosios, A. M., Bush, L. N., Freinkman, E., &
Vander Heiden, M. G. (2015). Supporting Aspartate Biosynthesis Is an Essential
Function of Respiration in Proliferating Cells. *Cell*, *162*(3), 552–563.
http://doi.org/10.1016/j.cell.2015.07.017

Sun, N., Youle, R. J., & Finkel, T. (2016). The Mitochondrial Basis of Aging.
Molecular Cell, *61*(5), 654–666. http://doi.org/10.1016/j.molcel.2016.01.028

Titov, D. V, Cracan, V., Goodman, R. P., Peng, J., Grabarek, Z., & Mootha, V. K.
(2016). Complementation of mitochondrial electron transport chain by
manipulation of the NAD+/NADH ratio. *Science*, *352*(6282), 231–235.
http://doi.org/10.1126/science.aad4017

Van Haute, L., Dietmann, S., Kremer, L., Hussain, S., Pearce, S. F., Powell, C. A., …
Minczuk, M. (2016). Deficient methylation and formylation of mt-tRNAMet
wobble cytosine in a patient carrying mutations in NSUN3. *Nature*
Communications, *7*(May), 12039. http://doi.org/10.1038/ncomms12039

Wallace, D. C. (2007). Why do we still have a maternally inherited mitochondrial
DNA? Insights from evolutionary medicine. *Annual Review of Biochemistry*,
76(1), 781–821. http://doi.org/doi:10.1146/annurev.biochem.76.081205.150955

Wallace, D. C. (2012). Mitochondria and cancer. *Nature Reviews Cancer*, *12*(10),
685–698. http://doi.org/10.1038/nrc3365

Weinberg, S. E., Sena, L. A., & Chandel, N. S. (2015). Mitochondria in the regulation
of innate and adaptive immunity. *Immunity*, *42*(3), 406–417.
http://doi.org/10.1016/j.immuni.2015.02.002

Wu, M., Neilson, A., Swift, A. L., Moran, R., Tamagnine, J., Parslow, D., … Chomicz,
S. (2007). Multiparameter metabolic analysis reveals a close link between
attenuated mitochondrial bioenergetic function and enhanced glycolysis
dependency in human tumor cells. *The American Journal of Physiology - Cell*
Physiology, *292*, 125–136. http://doi.org/10.1152/ajpcell.00247.2006.

Yao, J., Irwin, R. W., Zhao, L., Nilsen, J., Hamilton, R. T., & Brinton, R. D. (2009).
Mitochondrial bioenergetic deficit precedes Alzheimer's pathology in female
mouse model of Alzheimer's disease. *PNAS*, *106*(34), 14670–14675.

Zhang, J., Khvorostov, I., Hong, J. S., Oktay, Y., Vergnes, L., Nuebel, E., … Teitell,
M. A. (2011). UCP2 regulates energy metabolism and differentiation potential of
human pluripotent stem cells. *The EMBO Journal*, *30*(24), 4860–4873.
http://doi.org/10.1038/emboj.2011.401

Zhang, J., Nuebel, E., Wisidagama, D. R. R., Setoguchi, K., & Hong, J. S. (2012).
Measuring energy metabolism in cultured cells, including human pluripotent
stem cells and differentiated cells. *Nature Protocols*, *7*(6).
http://doi.org/10.1038/nprot.2012.048.Measuring

Zhou, W., Choi, M., Margineantu, D., Margaretha, L., Hesson, J., Cavanaugh, C., …
Ruohola-Baker, H. (2012). HIF1α induced switch from bivalent to exclusively
glycolytic metabolism during ESC-to-EpiSC/hESC transition. *The EMBO*
Journal, *31*(9), 2103–2116. http://doi.org/10.1038/emboj.2012.71
Zong, W.-X., Rabinowitz, J. D., & White, E. (2016). Mitochondria and Cancer.
Molecular Cell, *166*(3), 555–566. http://doi.org/10.1016/j.cell.2016.07.002
670

Fig. 1

Fig. 2

Fig. 3

Fig. 4

Fig. 5