Transcranial photoacoustic imaging of NMDA-evoked focal circuit dynamics in rat hippocampus

Jeeun Kang[1, 2], Shilpa D. Kadam[3], Joshua S. Elmore[1], Brennan J. Sullivan[4,*], Heather Valentine[1,*],
Adarsha P. Malla[5], Maged M. Harraz[5], Arman Rahmim[1, 6], Jin U. Kang[2, 7,] Leslie M. Loew[8], Michael
Baumann[9], Anthony A. Grace[10], Albert Gjedde[1,11,12,13], Emad M. Boctor[1,2,7, 14,†], Dean F.
Wong[1,5,15,16,†]

[1] Russell H. Morgan Department of Radiology and Radiological Science, Johns Hopkins University,
Baltimore, MD, USA.

[2] Laboratory of Computational Sensing and Robotics, Whiting School of Engineering, Johns Hopkins
University, Baltimore, MD, USA

[3] Neuroscience Laboratory, Hugo W. Moser Research Institute at Kennedy Krieger & Department of
Neurology, Johns Hopkins Medical Institutions, Baltimore, MD, USA

[4] Neuroscience Laboratory, Hugo W. Moser Research Institute at Kennedy Krieger, Baltimore, MD,
USA

[5] The Solomon H. Snyder Department of Neuroscience, Johns Hopkins University School of
Medicine, Baltimore, MD, USA

[6] Departments of Radiology and Physics, University of British Columbia, Vancouver, BC, Canada

[7] Department of Electrical and Computer Engineering, Whiting School of Engineering, Johns
Hopkins University, Baltimore, MD, USA

[8] Department of Cell Biology, University of Connecticut Health, Farmington, CT, USA

[9] National Institute on Drug Abuse, Intramural Research Program, NIH, Baltimore, MD

[10] Departments of Neuroscience, Psychiatry and Psychology, University of Pittsburgh, Pittsburgh,
PA, USA

[11] Department of Neurology and Neurosurgery, McGill University, Montreal, Quebec, Canada

[12] Departments of Clinical Research and Nuclear Medicine, Odense University Hospital and
University of Southern Denmark, Odense, Denmark

[13] Department of Neuroscience, University of Copenhagen, Copenhagen, Denmark

[14] Department of Computer Science, Whiting School of Engineering, Johns Hopkins University,
Baltimore, MD, USA

[15] Department of Psychiatry and Behavioral Sciences, Johns Hopkins Medical Institutions, Baltimore,
MD, USA

[16] Department of Neurology, Johns Hopkins Medical Institutions, Baltimore, MD, USA

* Equal contribution

**† These authors share senior authorship and address all correspondence to: Dean F. Wong, E-
mail: dfwong@jhmi.edu and to Emad M. Boctor, E-mail: eboctor1@jhmi.edu.**

**Keywords: Photoacoustic, neuroimaging, voltage-sensitive dye imaging, microdialysis,
quantitative electroencephalogram, N-methyl-D-aspartate, glutamate, hippocampus**

Abstract

Transcranial functional photoacoustic (fPA) voltage-sensitive dye (VSD) imaging promises to overcome current temporal and spatial limitations of current neuroimaging modalities. The technique previously distinguished global seizure activity from control neural activity in groups of rats. To validate the focal specificity of transcranial fPA neuroimaging *in vivo*, we now present proofs-of-concept that the results differentiate between low- and high-dose N-methyl-D-aspartate (NMDA) evoked neural activity in rat hippocampus. Concurrent quantitative EEG (qEEG) and microdialysis recorded real-time circuit dynamics and glutamate concentration change, respectively. We hypothesized that location-specific fPA VSD contrast would identify the neural dynamics in hippocampus with the correlation to NMDA evoked focal glutamate release and time-specific EEG signals. To test the hypothesis, we infused 0.3 to 3.0 mM NMDA at 2 μl/min over 60 min via an implanted microdialysis probe. The dialysate samples collected every 20 min during the infusion were analyzed for focal changes in extracellular glutamate release. Transcranial fPA VSD imaging provided NMDA-evoked VSD responses with positive correlation to extracellular glutamate concentration change at the contralateral side of the microdialysis probe. The graded response represents the all-or-none gating system of the dentate gyrus (DG) in hippocampus. Quantitative EEG (qEEG) successfully confirmed induction of focal seizure activity during NMDA infusion. We conclude that transcranial fPA VSD imaging distinguished graded DG gatekeeping functions, based on the VSD redistribution mechanism sensitive to electrophysiologic membrane potential. The results suggest the potential future use of this emerging technology in clinics and science as an innovative and significant functional neuroimaging modality.

Introduction

Whereas electrophysiological and invasive neurochemical techniques have been valuable in assessing activity within specific neural circuits in brain, there is a distinct need for a non-invasive means to image real-time regional functional activity in deep brain *in vivo*. The hippocampal circuit has rows of excitatory pyramidal neurons all aligned in close proximity with similar orientation. The dentate gyrus (DG) is classically considered as a gate keeper for the propagation of excitatory activity into the hippocampus as the excitatory neurons of the DG are less excitable than other types of hippocampal neurons. Therefore, low amplitude excitatory inputs do not generate action potentials in DG neurons and fail to open the DG gate [1]. In contrast to sub-threshold excitatory stimulus, strong repetitive excitatory inputs can break the DG gate and induce seizure activity [1]. This all-or-none gating function of the DG is consistent with the hippocampus being the lowest seizure threshold region of the mammalian brain.

Previous optical imaging approaches to local cortical circuits [2-4] applied invasive procedures to overcome superficial imaging depth, and further extension to deeper brain regions such as the hippocampus was prohibitive. The imaging of electrophysiological or neurochemical dynamics within the hippocampus primarily has been accomplished by optical imaging of freshly-sliced brain tissue [5] or magnetic resonance imaging spectroscopy (MRS) for non-invasive quantification of glutamate at high spatial resolution [6,7]. However, optical neuroimaging and MRS suffer from shallow imaging depth and from slow imaging speed, respectively. The application of two-photon microscopy to measure calcium ion dynamics in specific deep brain structures, including hippocampus, has poor temporal resolution [8]. Therefore, non-invasive, real-time neuroimaging modality having a high spatiotemporal resolution would be a significant advance, but such an approach would require neurochemical and electrophysiological characterization to validate the neural activity.

Photoacoustic (PA) imaging is a hybrid approach combining optics and acoustics where the signal corresponding to the neural activity is detected in the form of acoustic transcranial imaging with optical absorbance as an image contrast [9,10]. PA imaging is based on thermo-elastic perturbation of a target evoked by light absorbance from a pulsed laser illumination, which generates radio-frequency (RF) acoustic pressure waves detected by piezoelectric ultrasound transduction or optical interferometry. With this unique imaging mechanism, several attractive applications have been proposed for preclinical and clinical research with tomographic and microscopic imaging modes, including detection of endogenous contrast of cancer indicators, e.g., melanoma [11], breast microcalcifications [12-14], monitoring of cancer angiogenesis [15], oxygen metabolism [16], and quantification of lipid content [17,18], among others. Recently, we presented the functional PA (fPA) neuroimaging of near-infrared VSD redistribution mechanism differentiating the graded membrane potential in lipid vesicle model and chemo-convulsant seizure in rodent brain *in vivo* [19].

Here, we validated the transcranial fPA neuroimaging of hippocampal electrophysiology activated by stimulation of N-methyl-D-aspartate (NMDA) receptors in rat brain *in vivo* through intact skull and scalp. The tri-modal neural sensing approach utilized microdialysis, quantitative electroencephalography (qEEG), and transcranial fPA neuroimaging to simultaneously monitor electrophysiological changes associated with glutamatergic neurotransmission (Figure 1). To stimulate the hippocampus, we infused NMDA by reverse microdialysis that triggered localized changes of extracellular glutamate concentration. We quantified the change of extracellular glutamate concentration evoked by reverse microdialysis by direct microdialysis and high-performance liquid chromatography (HPLC). We combined fPA neuroimaging and qEEG with microdialysis to evaluate

103 the effects associated with graded DG gatekeeping at hippocampus, in the presence of glutamatergic
104 excitation. Concurrently, the electrophysiological signatures were captured with transcranial fPA
105 neuroimaging. Each of the three modalities thus provided information with different degrees of
106 specificity, including spatiotemporal specificity from fPA imaging, temporal specificity from qEEG,
107 and neurochemical specificity from microdialysis in the target neural circuit.

Results

Transcranial fPA neuroimaging of rat hippocampus. Transcranial fPA neuroimaging of
hippocampal circuit dynamics was performed during NMDA infusion at rat hippocampus (Figure 2).
The representative fPA sagittal planes obtained at 790 nm presented a sufficient sensitivity at the
depths of interest, i.e., 3.6 mm, at the contralateral position to the microdialysis probe (Figure 3a).
The sagittal PA imaging plane of hippocampus revealed the cross-sections of transverse sinus and
inferior cerebral vein in the superficial depth range (Figure 3b). From the temporal dimension, time-
averaged VSD responses were reconstructed pixel-by-pixel for pre-injection phase (-10 – 0 min)
followed by the results for time bins in 10 min interval: 0 – 10 min, 10 – 20 min, and 20 – 30 min.
Note that the total recording duration (40 min) was limited by internal memory of the fPA
neuroimaging system. Figure 3c shows the representative VSD responses in hippocampal cross-
sections collected during 0.3 mM and 3.0 mM NMDA infusion. As a result, the hippocampus with
0.3 mM NMDA infusion did not present such a significant change ($n = 3$), which implies a failure to
overcome the high activation threshold of DG gatekeeping with 0.3 mM NMDA infusion. Otherwise,
the substantial amount of circuit dynamics was detected at hippocampus with 3.0 mM NMDA
infusion ($n = 6$). In the 3.0 mM NMDA infusion group, the peak VSD responses were presented
during the first or second 10 min durations, and their peak intensities corresponds to the maximal
glutamate concentration change measured in the microdialysis, e.g., 734.48 % and 493.91 % (center
and right columns in Figure 3c). Further quantitative multi-modal correlation will be given in the
following subsections.

NMDA-evoked glutamatergic neurotransmission. Intrahippocampal infusion of NMDA diluted
into artificial cerebrospinal fluid at the required concentrations was performed through the
microdialysis probe. Concentrations of glutamate release in the dialysate were calculated from
values of three baseline samples collected at 20 min intervals before NMDA infusion. During
infusion into hippocampus (bregma -3.8 mm, lateral 3 mm, depth 3.6 mm; Figure 4a), samples
continued to be collected every 20 min for the remainder of the study. Glutamate release was
calculated as a % of basal values. Figure 4b shows the effect of NMDA infusion on
glutamate release in the hippocampus of anesthetized rats. The results demonstrate the dose-related
response of NMDA infusion at 0.3 ($n = 3$), 1.0 ($n = 1$), and 3.0 mM ($n = 6$) into the hippocampus,
where 0.3 mM NMDA caused a 17.93±19.05 % increase in glutamate levels that remained elevated
for the duration of the infusion; the 1.0-mM NMDA infusion appears to cause doubling of the
glutamate levels in the hippocampus compared to baseline (i.e., 112.42 %). This increase observed at
the 1.0-mM remained elevated through to the end of the NMDA infusion. However, at 3.0-mM, the
NMDA infusion raised the glutamate level up to 392.99±250.54 %, above baseline level, which
peaked during the second 20-min sampling period.

EEG of electrophysiological activity in hippocampus. The DG is a hippocampal region
specifically subjected to a barrage of excitatory inputs. However, the majority of excitatory activity
does not propagate through the DG and into the hippocampus as the DG performs a gating function [1].

Excessive activation of the DG disrupts its gating function and induces acute seizures in naïve
animals [1]. Hippocampal infusion of 3.0 mM NMDA induced a maximal VSD and significant rise in
extracellular glutamate concentration, suggesting a potential break in the DG gate that may lead to
acute seizure activity. We therefore utilized qEEG during the same hippocampal 3.0 mM NMDA
infusion protocol to identify if the maximal VSD response and extracellular glutamate concentrations
were associated with significant electrophysiological activity, a tri-modal approach that validates the
temporal and spatial resolution of fPA imaging during focal NMDA infusion.

In patients with focal epilepsy, gamma and theta activity from scalp EEG are an indicator of
the seizure onset zone and ictal onset [20,21]. Here, the temporal specificity of qEEG recorded circuit
responses in real-time, before and after focal NMDA infusion. The 3.0 mM NMDA infusion into the
hippocampal circuit induced focal seizure activity recorded on qEEG (Figure 4c black asterisk and
S1). When NMDA was delivered to the hippocampal circuit by the microdialysis probe (Figure 4a)
[22], a significant change in qEEG spectral power was presented, especially in the gamma range (Figure
4c). In the theta range, spectral power increased during the seizure (Figure 4d) further supporting the
identification of the focal seizure activity. The seizure trace on EEG demonstrated both a gamma and
theta component (Figure 4f). Importantly, the ictal event occurred during the same temporal window
as the maximum VSD response and the greatest increase in extracellular glutamate concentration.
The dose-related VSD and extracellular glutamate concentration response with qEEG suggest that the
3.0-mM NMDA infusion resulted in a break in the DG gate associated with focal seizure activity.

However, qEEG lacks spatial resolution, and recordings from deep brain structures cannot be
isolated without implantation of invasive depth electrodes. Therefore, in order to maintain the tri-
modal experimental paradigm we opted not to place depth electrodes. We made the stimulation
focal, according to the previous proof-of-concept results by fPA detection onset of generalized
seizures through an intact skull and scalp [19].

Quantitative multi-modal correlation. The fPA VSD responses at the contralateral side to the
microdialysis probe (blue dots in Figure 3c) were measured and plotted as a function of the
corresponding changes in extracellular glutamate concentration (Figure 5). Note that the regions-of-
interest ROIs to quantify the VSD responses were in 1 x 1 mm^2 size in the sagittal cross-section of rat
hippocampus. The reference phase for VSD response quantification was obtained from the baseline
phase: 5 – 10 min. In fPA imaging, 3.0 mM NMDA infusion at hippocampus yielded significant
elevation of VSD response: 0.08±0.35, 2.24±1.06, and 1.17±1.56 for baseline (-10 – 0 min), NMDA1
(0 – 20 min), and NMDA2 (20 – 30 min) phases, respectively (Figure S2a). The increase of
glutamate concentration change was correspondingly presented from -4.67±4.48 %, 369.75±254.91
%, to 392.99±250.54 %. In addition, when re-analyzing the data for -50 – 100 %, 100 – 500 %, and
500 – 1,000 % bins of fractional glutamate concentration change, high positive correlation was
obtained with VSD responses: 0.38±0.55 (n_{sample} = 9), 0.97±1.40 (n_{sample} = 5), and 3.17±0.32 (n_{sample}
= 4) for 11.04±29.60 %, 279.85±141.64 %, to 666.49±73.63 % of glutamate concentration changes
(red lines in Figure 5). The goodness of fit (R^2) among the mean values was 0.95 with slope and y-
intercept at 0.12 and 0.00, respectively. The qEEG found the increased bursts of gamma power
during 3.0 mM NMDA infusion, reaching up to ~261 % increase with a focal seizure (Figure 4c and
S1).   Otherwise, focal 0.3 mM NMDA infusion at hippocampus did not presented any circuit
activity increase statistically significant. The VSD responses were -0.11±0.02, -0.15±1.71, and -
0.07±0.64 in baseline, NMDA1, and NMDA2 phases, respectively, which are correlated to the low
glutamate concentration changes in hippocampus: -1.65±6.90 %, 3.90±10.41 %, and 17.93±19.05 %
(Figure S2b). Also, the VSD responses presented strong positive correlation when re-analyzed in -10
– 0 %, 0 – 10 %, and 10 – 50 % bins of fractional glutamate concentration changes: -1.00±1.00,

193 0.10 ± 0.35, and 0.58 ± 0.48 (n_{sample} = 3 for each) with -6.84 ± 3.38 %, 4.43 ± 4.05 %, and 22.59 ± 11.64 %
194 of fractional glutamate concentration changes, respectively (blue lines in Figure 5). The R^2 was 0.88
195 with slope and y-intercept at 0.05 and -0.46, respectively.

196 **Brain histology.** Brain tissue was extracted after all *in vivo* experimental protocols. The brain was
197 frozen-sectioned into 300-μm thickness slices and evaluated to confirm probe placement using the
198 hemorrhage caused by the microdialysis probe insertion for the NMDA infusion and collections as a
199 marker. Figure 6a shows the bright-field images of the coronal plane of the hippocampus at -3.8 mm
200 from bregma, respectively. White arrows indicate the hemorrhage caused by the microdialysis probe
201 positioned at the following coordinates in hippocampus: bregma -3.8 mm, lateral 3 mm, and depth
202 3.6 mm. The VSD staining of brain tissue was confirmed with near-infrared fluorescence
203 microscopy. Uniform VSD fluorescence was found in the VSD perfusion animal, while negative
204 control (VSD-) presented negligible fluorescence emission (Figure 6b), which again confirms the
205 results in our previous publication [23].

Discussion

208 The application of fPA neuroimaging was expanded to neuroscience [24-26] as label-free transcranial
209 fPA imaging of neurovascular coupling proposed as a means to quantify hemodynamic changes [27,28].
210 However, this approach does not yield quantitative neural activities that directly correspond to
211 electrical activity. Hemoglobin provides an effective contrast signal in fPA neuroimaging, but the
212 neurovascular coupling in brain is comparatively slow compared to electrophysiological neural
213 activities. Instead, there has been extensive investigations into more effective exogenous contrast
214 agents [29-33]. This approach has enabled several neuroimaging approaches with functional voltage
215 sensors. Deán-Ben et al. showed real-time PA tomography of a genetically encoded calcium
216 indicator, GCaMP5G, using zebrafish *in vivo* [34]. Sheryl Roberts, et al. also proposed a new
217 metallochromic calcium sensor for PA imaging (CaSPA) which has a high extinction coefficient, a
218 low quantum yield, and high photo-bleaching resistance for brain and heart imaging [35]. Ruo et al.
219 reported PA imaging of neural activity evoked by electrical stimulation and 4-aminopyridine-induced
220 epileptic seizures using hydrophobic anions such as dipicrylamine (DPA) in mouse brain [36].
221 However, these voltage sensors requires PA imaging at the visible spectral range (488 nm and 530
222 nm for GCaMP5G; 550 nm for CaSPA; 500 nm and 570 nm for DPA), which are suboptimal when
223 imaging deep brain such as hippocampus positioned at 5 mm – 8 mm depth including intact scalp and
224 cortex in rat [22,37].

225 Recently, we proposed transcranial fPA recordings of brain activity *in vivo* with a near-
226 infrared VSD, delivered through the blood-brain barrier (BBB) via pharmacological modulation, as a
227 promising tool to transcend optical neuroimaging limitations, particularly as it relates to sensing
228 depth [19,23,38]. The studies demonstrated that transcranial fPA neuroimaging distinguishes *in vivo*
229 seizure activity in stimulated rat brains from that of control groups in real time. However, the results
230 were limited by the use of a global chemo-convulsant, causing perturbation across the entire brain
231 caused by the intraperitoneal administration of penetylenetetrazole (PTZ). In this paper, we presented
232 follow-on advances in fPA VSD neuroimaging by focal neural stimulation of heterogeneous neural
233 circuits, with concomitant validations from qEEG and glutamate quantification using microdialysis,
234 respectively. The set of experiments described here yield key findings as follows: (1) The
235 microdialysis-dependent low- and high-dose NMDA infusion into the central nervous system (CNS)
236 lead to a wide range of focal extracellular glutamate concentration increase in hippocampus up to

~800 %. (2) The neurochemical response (microdialysis) was well-correlated to the phenotypes in the
electrophysiological sensing (qEEG). The NMDA activation in the hippocampus triggers an all-or-
none type of circuit dynamics that lead to the initiation of a focal seizure in the hippocampal circuit.
(3) Transcranial fPA neuroimaging data successfully identified hot spots of focal NMDA receptor
activation, as presented in the qEEG recordings. The hippocampal circuitry provided the proportional
excitation of glutamatergic neurotransmission with concomitant NMDA infusion. The DG of the
hippocampus is positioned as a gatekeeper to regulate the vast excitatory cortical inputs from
propagating into the hippocampus [1]. Characteristically, the DG displays a high activation threshold; a
trait that is mediated by its profuse innervation by inhibitory GABAergic neurons and relatively
hyperpolarized resting membrane potential of its pyramidal neurons [39]. In epilepsy the DG fails to
gate the propagation of excitatory inputs into the hippocampus, resulting in overexcitation and
seizures. In naïve animals, *in vivo* optogenetic activation of the DG disrupts its gating function and
induces seizures that increase in severity depending on the duration of the stimulus [1]. In this study,
the disruption of DG gating by strong stimuli has been clearly demonstrated by utilizing the fPA
VSD neuroimaging techniques during focal NMDA infusion.

Further investigations are required to advance our current perspectives available with tri-
modal sensing, including fPA, qEEG, and microdialysis. Glutamate produces fast-rising brief
depolarizations in pyramidal neurons. Therefore, the use of fPA neuroimaging will enable us to more
precisely assess glutamate and GABA dynamics in order to formulate a more complete profile of
circuit activation. Once homeostasis is disrupted, neuronal activity is sensitive to changes both of
excitatory and inhibitory mechanisms. Faster neurochemical recording is another approach that may
prove useful in assessing the impact of these measures. Although microdialysis successfully
provided quantitative, focal neurochemical concentrations, the sampling rate of 1 sample per 20 min
was slow. Techniques offering faster temporal resolution may allow more meaningful comparison of
the neurochemical changes yielded by microdialysis and the electrophysiological events monitored
by qEEG and fPA neuroimaging [40]. One such technique, using custom built hardware and the
Amplex Red method, achieved fluorescence-based quantification of glutamate from samples taken
every 5 seconds, though reliability appeared to be limited when higher glutamate concentrations were
measured [41]. Implantable glutamate biosensors allow sub-second readouts of neurochemical
concentrations. However, current limitations include sensitivity, selectivity, and high cost; recent
developments in materials, effective modeling, and sensor design may soon alleviate some of these
limitations [42,43].

From the results, transcranial fPA neuroimaging was able to differentiate the circuit activity
defined with qEEG and microdialysis. However, future developments should serve to further advance
the efficacy of the fPA neuroimaging in neuroscience. (1) We expect that improved signal processing
for extracting neural activity from the ubiquitous blood context will enable better characterization of
brain function. The present *in vivo* experiments confirmed the possibility of background suppression,
as also presented in our previous study [19]. Enhanced signal processing and/or use of multi-spectral
wavelengths may allow significantly improved spectral differentiation of electrophysiological
activities in the brain at higher temporal resolution, leading to development of novel quantitative
metrics for real-time brain activity measures. Having isotropic resolution with 2-D PA probes would
be also an interesting direction to pursue as a follow up to the present work. The use of 2-D PA probe
would not only allow real-time volumetric information, but also enable the suppression of off-axis
interference. Even though we presented that neural activity can be successfully discerned with our
current 1-D PA probe, its sensitivity might be affected by off-axis interferences especially from the
elevation direction because of the limited acoustic lens focusing at a fixed depth. The neuroimaging
using 2-D PA probe would reject those interferences by the advanced electrical beamforming

capability in axial, lateral, and elevation directions. Having an improved PA imaging system would
provide significant breakthrough in terms of spatiotemporal resolution in fPA neuroimaging. Even
though our current laser system yields both 4 fps (frame-per-second) of temporal resolution and PA
signal sensitivity at rat hippocampus, further optimization of temporal resolution would provide finer
spatiotemporal specificity. On the other hand, we consider employing larger animal model for this
fPA VSD neuroimaging research. Pig models have been an ideal subject to pave the way to human
translation of neuroengineering technologies, thanks to their analogous brain structure and
physiology and scalp and skull thicknesses to those in humans, with alleviated ethical issues [44]. We
already validated the transcranial fPA neuroimaging in the pig model, and will continue to pursue
this research direction [45].

In all, the transcranial fPA neuroimaging at hippocampus in *in vivo* rat brain was successfully
correlated with electrophysiologic and neurochemical measurements using qEEG and microdialysis:
focal NMDA infusion triggers glutamate release that excites the neural circuit, and at threshold doses
it causes runaway excitation in the hippocampus by overcoming DG gating. This is reflected in the
lower seizure threshold of the hippocampus. Therefore, the transcranial fPA neuroimaging is a
promising technology for the visualization of focal neural events in real time.

Material and Methods

Animal preparation. For the proposed *in vivo* experiments, 8-9-week-old male Sprague Dawley rats
(Charles Rivers Laboratory, Inc., MA, United States) weighing 275-390g were used. The use of
animals for the proposed experimental protocol was approved by the Institutional Animal Care and
Use Committee of Johns Hopkins Medical Institute (RA16M225). Rats were housed in groups of 3
per cage with free access to food and water and maintained on a 12hr light / 12hr dark cycle.

On the day of the study the rats were weighed and anesthetized with urethane (1.2mg/kg,).
Urethane was given incrementally with alternating intra-peritoneal (ip) and subcutaneous (sc) dosing.
Three (3) ml of isotonic saline was given sc on each side of the body to keep the animal hydrated
during the experimental procedure. Body temperature was maintained until animal was fully
anesthetized and ready for surgery. For fPA and qEEG studies, an iv catheter was inserted into a tail
vein prior to surgery for dye administration during the studies. Once a stable plane of anesthesia was
established, hair was shaved from the scalp of each rat to have acoustic coupling for transcranial fPA
recording. The rat was placed into a stereotaxic device (Stoeling Co. Wood Dale, IL). This fixation
procedure was required to prevent any unpredictable movement during fPA or EEG recording of
neural activities. A CMA12 microdialysis probe (Harvard Apparatus, Holliston, MA, USA) was
implanted into the CA_3 region of the right hippocampus (stereotaxic coordinates: 3 mm lateral and
3.8 mm posterior to bregma, and 3.6 mm below the surface of the dura, Figure 3a) [22]. The probe
active exchange surface was 2×0.5 mm. The probe was secured to the skull using dental acrylic
cement. The fPA and qEEG probes were placed on the contralateral side of the microdialysis probe.

Fluorescence quenching-based near-infrared voltage-sensitive dye. In the present *in vivo* study,
we used the fluorescence quenching-based near-infrared cyanine VSD, IR780 perchlorate (576409,
Sigma-Aldrich Co. LLC, MO, United States) as used in our previous *in vivo* study differentiating a
chemo-convulsant seizure activity [19], and it has the analogous chemical structure of PAVSD800-2,
our new VSD validated in our previous *in vitro* study [38]. This VSD yields fluorescence quenching
and de-quenching depending on membrane polarization and subsequent change in the local VSD

molecule density, leading to a reciprocal change of PA contrast with non-radiative relaxation of
absorbed energy.

Functional fPA neuroimaging. We used real-time PA data acquisition to record
electrophysiological neural activities *in vivo* as in our previous study [19]: an ultrasound research
system consisted of an ultrasound linear array transducer connected to a real-time data acquisition
system (SonixDAQ, Ultrasonix Medical Corp., Canada). To induce the PA signals, pulsed laser light
generated by a second-harmonic (532 nm) Nd:YAG laser pumping an optical parametric oscillator
(OPO) system (Phocus Inline, Opotek Inc., USA) provided 690-900 nm of tunable wavelength range
and 20 Hz of the maximum pulse repetition frequency. A bifurcated fiber optic bundle, each 40 mm
long and 0.88 mm wide, was used for laser pulse delivery. The PA probe was situated between the
outlets of the bifurcated fiber optic bundles using a customized, 3-D printed shell for evenly
distributing laser energy density in the imaging field-of-view. The alignment of outlets was focused
specifically at 20 mm depth. The PA probe was positioned in the contralateral sagittal plane of
microdialysis probe (3 mm) to cover the hippocampal cross-section. The distance between the PA
probe and the rat skin surface was 20 mm filled with acoustic gel, and the resultant energy density
was at ~3.5 mJ/cm^2 , which is far below the maximum permissible exposure (MPE) of skin to laser
radiation by the ANSI safety standards [46]. A wavelength of 790 nm was used, at which sufficient
absorbance can be obtained by the near-infrared VSD, i.e., IR780 perchlorate. Also, excitation at that
wavelength prevented the undesired time-variant change of blood oxygen saturation, since the
wavelength corresponds to the isosbestic point of Hb and HbO_2 absorption spectra. Detailed
information of neural activity reconstruction using normalized time-frequency analysis can be found
in our previous publication [19].

***In vivo* microdialysis**. *In vivo* microdialysis sampling was carried out as previously described [47,48].
For infusion experiments, NMDA (Sigma-Aldrich Chemicals, St. Louis, Mo) was weighed,
solubilized, and diluted to the desired concentration in artificial cerebrospinal fluid (NaCl, 147
mmol/L; KCl, 2.7 mmol/L; $CaCl_2$, 1.2 mmol/L; $MgCl_2$, 0.85 mmol/L) (Harvard Apparatus, Holliston,
MA, USA) on the study day. Once the probe was inserted and secured, it was perfused with artificial
cerebrospinal fluid pumped at a flow rate of 2 μl/min. Samples were collected at 20 min intervals,
and immediately transferred to a -80°C freezer until assayed. To allow sufficient time for the
glutamate levels to equilibrate, three baseline samples were collected an hour following initiation of
infusion. Following these samples, NMDA was infused into the brain directly through the dialysis
probe with the same pump parameters as used for the baseline samples. Dialysate samples were
assayed for glutamate by a two-step process using HPLC-ECD on an Eicom HTEC-500 system
(EICOM, San Diego, CA, USA). After passing the samples through a separation column, they were
processed via a column containing immobilized L-glutamate oxidase enzyme, resulting in the release
of hydrogen peroxide. The hydrogen peroxide concentration was then determined using a platinum
working electrode. Chromatographic data were acquired online and exported to an Envision software
system (EICOM, San Diego, CA, USA) for peak amplification, integration, and analysis.

Quantitative EEG. All EEG recordings utilized a three-electrode paradigm: 1 recording, 1 reference
(aligned to the site of activation) and 1 ground over the rostrum. The electrodes (IVES EEG; Model #
SWE-L25 – IVES EEG solutions, MA, USA) were fixed with minimal cyanoacrylate adhesive
(KrazyGlue), similar to previous protocols [49]. Data acquisition was performed using Sirenia software
(Pinnacle Technologies Inc., Kansas, USA) with synchronous video capture. Data acquisition had a
14-bit resolution, 400 Hz sampling rate, and a band pass filter between 0.5 Hz and 50 Hz. The
acquisition files were stored in an .EDF format and scored manually, using real-time annotations

from the experiments. EEG power for 2-second epochs was done using an automated fast Fourier
transformation module in Sirenia software [50].

***In vivo* experimental protocol.** The *in vivo* protocols were designed for simultaneous multi-modal
sensing of the neural activity at hippocampus: microdialysis-qEEG and microdialysis-fPA
neuroimaging. Figure 2 shows a detailed schematic protocol for each group representing the response
to the administration of NMDA, Lexiscan and VSD (i.e., IR780 perchlorate). fPA and qEEG data
acquisition were performed for 40 min to correlate with three microdialysis samples collected at 20-
min intervals. Graded NMDA infusion concentrations were applied to identify the dose-dependent
glutamatergic excitation of hippocampal circuit: 0.3 mM ($n = 3$) and 3.0 mM ($n = 6$). VSD and
Lexiscan followed the data acquisition sequence with 3-min delay, thereby 5 min of baseline phase
was guaranteed for the VSD response reconstruction in fPA neuroimaging before starting NMDA
infusion. The dosing protocol for Lexiscan and VSD administration was as follows: through an iv tail
vein catheter, 150 µl of Lexiscan (0.4mg/5ml) was injected, followed by 200 µl of VSD at 2 mg/ml
concentration, flushed immediately with 150 µl of 0.9% isotonic saline. The EEG signal was
recorded an identical preparation procedure as the fPA neuroimaging, including animal preparation
and administration of IR780, Lexiscan and experimental duration time for all recordings.

Brain histology. Rats used for the above protocol were sacrificed, and whole brains immediately
harvested and placed in 10 % formalin. All brains were allowed to fix in fresh 10 % formalin for at
least 48 hours with gentle agitation on a conical rotator. Subsequently, the brains were processed
through a series of sucrose gradients (15 %, 20 %, 30 % for 12-24 hours each) for cryoprotection.
Brains were sectioned frozen at 300 µm thickness. Tissue sections were mounted on slides in
ProLong Diamond Anti-face mountant. Slides with sections were imaged using an Olympus OM-D
E-M5 Mark II for bright field image and using LI-COR Odyssey for fluorescence visualization.

Author Contributions

DFW, AAG and MB originally conceived of the NMDA administration and of microdialysis with PA idea and helped to design the overall research plan with initial funding for the *in vivo* experiments. DFW critically revised the focus of the results and final version and interpretation. EMB, LML, and SDK helped to design the overall research plan, helped plan specific experiments and all contributed to the review and writing of the manuscript. Jeeun K, SDK, JSE, BJS, HV planned and carried out *in vivo* experiments, analyzed the research outcomes, and wrote key elements of the first draft of the manuscript. Jeeun K analyzed and interpreted the PA measurements and completed the first manuscript, HV and JE the NMDA dosing and microdialysis, and SDK and BS the EEG experiments. MB provided resources and personnel and vital collaboration for the microdialysis part of the experiment. HV and LML devised the VSD vehicle preparation. APM performed histopathological analysis. MMH planned and supervised confirmation of VSD penetration into brain tissue also contributed to the final version of the manuscript. AAG critically revised experimental design, draft, and final versions of manuscript, and interpretation of results. Jin K developed, funded, and participated in the current PA system design. AR participated in early planning and critically read and edited the manuscript. He has contributed both technically and materially to support this research. EMB led the development, system specification, design specification, and funding of the current PA imaging system. Secured the funding of the needed imaging experiments throughout the lifetime of the project, including taking responsibility of 2 full-time research members specifically for this work, Jeeun K and APM. Further, he has contributed intellectually by mentoring these members and providing input on manuscripts, and participating on PI meetings.

Funding

This work was supported by the NIH BRAIN Initiative under Grant No. R24 MH106083-03 (DFW, AR, AAG, EMB, HV, JE) and the NIH National Institute of Biomedical Imaging and Bioengineering under Grant No. R01EB01963 (LML); NIH National Institute of Child Health and Human Development (NICHD) for R01HD090884 (SDK); NIH National Institute of Heart, Lung and Blood (NHLBI) under grant number R01HL139543 (Jeeun K, APM, EMB); National Cancer Institute (NCI) under grant number R21CA202199 and its equipment supplement (EMB). Funding of the PA equipment was provided via resources of Jin K and EMB NSF Career award #1653322. Jeeun K was partially supported by the Basic Science Research Program through the National Research Foundation of Korea (NRF) funded by the Ministry of Education #2018R1A6A3A03011551.

Conflict of Interest Statement

The subject matter described in this article is included in patent applications filed by the University of Connecticut and Johns Hopkins University. LML is a founder and owner of Potentiometric Probes LLC, which sells voltage sensitive dyes.

The remaining authors declare that the research was conducted in the absence of any commercial or financial relationships that could be construed as a potential conflict of interest.

435   **Figures**

437   **Figure 1 Tri-modal sensing of rat hippocampus.** Red dotted rectangular describes dentate gyrus (DG)
438   gating breakdown at hippocampus to a focal NMDA infusion. NMDA, N-methyl-d-aspartate; VSD, voltage-
439   sensitive dye.

442   **Figure 2 *In vivo* experimental protocol.** Tri-modal monitoring of rat hippocampus using reverse/forward
443   microdialysis, transcranial PA imaging, and qEEG. ACF, artificial cerebrospinal fluid.

445

Figure 3. Transcranial PA neuroimaging of the hippocampal circuit dynamics following focal NMDA infusion. (**a**) Illustration of stereotaxic coordinates for the microdialysis probe and sagittal cross-sections for real-time PA recording. (**b**) The sagittal PA imaging plane was selected in the contralateral side of the microdialysis probe infusing NMDA into the brain (3 mm lateral). (**c**) Time-averaged VSD response maps during -10 – 0 min (baseline phase); 0 – 10 min, 10 – 20 min, and 20-30 min (NMDA infusion phases). Note that the blue points indicate microdialysis probe in contra-lateral positions. Maximal glutamate concentration increases for representative fPA images were 34.23 % (left), 734.48 % (center) and 493.91 % (right).

453

Figure 4. Extracellular glutamate concentration and concomitant electrophysiology during NMDA-evoked activity in the hippocampus. (**a**) Illustrates stereotaxic coordinates for the microdialysis probe and recording leads. Ground electrode was implanted over the rostrum. (**b**) 0.3 ($n = 3$), 1.0 ($n = 1$), and 3.0 mM ($n = 6$) NMDA infusion into the hippocampus caused maximal glutamate increases of 17.93±19.05%, 112.42%, and 392.99±250.54% respectively as compared to % baseline. Gray dotted line indicates the baseline. (**c**) Gamma (35-50Hz) power shows a 5-min baseline recording (black trace), followed by two consecutive 5 min traces (blue trace) following focal NMDA infusion in the hippocampus. EEG demonstrated a 261% increase in gamma power, as compared to % baseline, and an onset of epileptiform discharges after focal NMDA infusion (see inset bar graph; asterisk denotes epileptiform activity). (**d**) Respective theta power after NMDA infusion represents the theta component of the focal seizure. (**e**) 15-min spectral power heat map demonstrates the spectral power changes associated with NMDA infusion and subsequent epileptiform activity (denoted by black asterisk in e). (**f**) The representative raw EEG trace during the occurrence of the epileptiform event (solid black line in e) for full (05.-50Hz), gamma (0.5-50Hz), and theta (5.5-8.5Hz) power; respectively. For the expanded time scale of the focal hippocampal seizure see Supplemental 1.

Figure 5. fPA VSD response at hippocampus as a function of extracellular glutamate concentration change. Grey dotted lines indicate the basal level in fPA VSD response and extracellular glutamate concentration.

Figure 6. Histopathological confirmation (**a**) Microdialysis probe at hippocampus. White arrow indicates the wound caused by microdialysis probe installation. (**b**) Frozen-sectioning histopathological confirmation of systematic VSD delivery throughout brain tissue region.

Figure S1. Expanded time scale of EEG recording during NMDA infusion into the hippocampus. (a) 10 sec expanded time scale of raw EEG traces during hippocampal NMDA infusion. (I) Baseline (II) immediately after onset of NMDA infusion (III) immediately before ictal event (IV) during focal seizure event (V) and sustained short duration high frequency hippocampal discharges after the focal seizure event.

488 **Figure S2. fPA VSD responses in baseline, NMDA1, and NMDA2 phases as a function of extracellular**
489 **glutamate concentration change at hippocampus.** (**a**) 3.0 mM NMDA infusion. (**b**) 0.3 mM NMDA
490 infusion. 0.3 mM NMDA data in red rectangular is magnified and presented together (see inset graph). Grey
491 dotted lines indicate the basal level in fPA VSD response and extracellular glutamate concentration.

References

1   Magnuson EK, Armstrong C, Bui A, Lew S, Oijala M, Soltesz I. In vivo evaluation of the
dentate gate theory in epilepsy. *The Journal of Physiology* 2015; **593**: 2379–2388.

2   Derdikman D, Hildesheim R, Ahissar E, Arieli A, Grinvald A. Imaging spatiotemporal
dynamics of surround inhibition in the barrels somatosensory cortex. *J Neurosci* 2003; **23**:
3100–3105.

3   Petersen CCH, Grinvald A, Sakmann B. Spatiotemporal Dynamics of Sensory Responses in
Layer 2/3 of Rat Barrel Cortex Measured In Vivo by Voltage-Sensitive Dye Imaging Combined
with Whole-Cell Voltage Recordings and Neuron Reconstructions. *Journal of Neuroscience*
2003; **23**: 1298–1309.

4   Petersen CC, Sakmann B. Functionally independent columns of rat somatosensory barrel cortex
revealed with voltage-sensitive dye imaging. *J Neurosci* 2001; **21**: 8435–8446.

5   Okubo Y, Sekiya H, Namiki S, Sakamoto H, Iinuma S, Yamasaki M *et al.* Imaging
extrasynaptic glutamate dynamics in the brain. *Proceedings of the National Academy of*
Sciences 2010; **107**: 6526–6531.

6   Crescenzi R, DeBrosse C, Nanga RPR, Byrne MD, Krishnamoorthy G, D'Aquilla K *et al.*
Longitudinal imaging reveals subhippocampal dynamics in glutamate levels associated with
histopathologic events in a mouse model of tauopathy and healthy mice. *Hippocampus* 2017;
27: 285–302.

7   Voets NL, Hodgetts CJ, Sen A, Adcock JE, Emir U. Hippocampal MRS and subfield volumetry
at 7T detects dysfunction not specific to seizure focus. *Sci Rep* 2017; **7**: 1–14.

8   Kondo M, Kobayashi K, Ohkura M, Nakai J, Matsuzaki M. Two-photon calcium imaging of the
medial prefrontal cortex and hippocampus without cortical invasion. *Elife* 2017; **6**: 592.

9   Wang LV, Hu S. Photoacoustic tomography: in vivo imaging from organelles to organs. *Science*
2012; **335**: 1458–1462.

10   Wang X, Pang Y, Ku G, Xie X, Stoica G, Wang LV. Noninvasive laser-induced photoacoustic
tomography for structural and functional in vivo imaging of the brain. *Nat Biotechnol* 2003; **21**:
803–806.

11   Zhang HF, Maslov K, Stoica G, Wang LV. Functional photoacoustic microscopy for high-
resolution and noninvasive in vivo imaging. *Nat Biotechnol* 2006; **24**: 848–851.

12   Kang J, Kim E-K, Kim GR, Yoon C, Song T-K, Chang JH. Photoacoustic imaging of breast
microcalcifications: a validation study with 3-dimensional ex vivo data and spectrophotometric
measurement. *J Biophoton* 2015; **8**: 71–80.

13   Kang J, Chung WY, Kang S-W, Kwon HJ, Yoo J, Kim E-K *et al.* Ex Vivo Estimation of
Photoacoustic Imaging for Detecting Thyroid Microcalcifications. *PLoS ONE* 2014; **9**.
doi:10.1371/journal.pone.0113358.

14   Kim GR, Kang J, Kwak JY, Chang JH, Kim SI, Youk JH *et al.* Photoacoustic imaging of breast
microcalcifications: a preliminary study with 8-gauge core-biopsied breast specimens. *PLoS*
ONE 2014; **9**: e105878.

15   Yao J, Maslov KI, Wang LV. In vivo Photoacoustic Tomography of Total Blood Flow and
Potential Imaging of Cancer Angiogenesis and Hypermetabolism. *Technol Cancer Res Treat*
2012; **11**: 301–307.

16   Hu S. Listening to the Brain With Photoacoustics. *IEEE J Select Topics Quantum Electron*
2016; **22**: 117–126.

17   Hui J, Cao Y, Zhang Y, Kole A, Wang P, Yu G *et al.* Real-time intravascular photoacoustic-
ultrasound imaging of lipid-laden plaque in human coronary artery at 16 frames per second. *Sci*
Rep 2017; **7**: 1417.

18   Wu M, Springeling G, Lovrak M, Mastik F, Iskander-Rizk S, Wang T *et al.* Real-time
volumetric lipid imaging in vivo by intravascular photoacoustics at 20 frames per second.
Biomed Opt Express 2017; **8**: 943–953.

19   Kang J, Zhang HK, Kadam SD, Fedorko J, Valentine H, Malla AP *et al.* Transcranial Recording
of Electrophysiological Neural Activity in the Rodent Brain in vivo Using Functional
Photoacoustic Imaging of Near-Infrared Voltage-Sensitive Dye. *Frontiers in Neuroscience*
2019; **13**: 1–14.

20   Andrade-Valenca LP, Dubeau F, Mari F, Neurology RZ, 2011. Interictal scalp fast oscillations
as a marker of the seizure onset zone. *Neurology* 2011; **77**: 524–531.

21   Foldvary N, Klem G, Hammel J, Bingaman W, Najm I, Lüders H. The localizing value of ictal
EEG in focal epilepsy. *Neurology* 2001; **57**: 2022–2028.

22   Paxinos G, Watson C. *The Rat Brain in Stereotaxic Coordinates*. Seventh edition. Elsevier
Academic Press, 2014.

23   Pak RW, Kang J, Valentine H, Loew LM, Thorek DL, Boctor EM *et al.* Voltage-sensitive dye
delivery through the blood brain barrier using adenosine receptor agonist Regadenoson. *Biomed*
Opt Express 2018; **9**: 3915–3922.

24   Nie L, Cai X, Maslov K, Garcia-Uribe A, Anastasio MA, Wang LV. Photoacoustic tomography
through a whole adult human skull with a photon recycler. *J Biomed Opt* 2012; **17**: 110506.

25   Li W, Chen R, Lv J, Wang H, Liu Y, Peng Y *et al.* In Vivo Photoacoustic Imaging of Brain
Injury and Rehabilitation by High-Efficient Near-Infrared Dye Labeled Mesenchymal Stem
Cells with Enhanced Brain Barrier Permeability. *Adv Sci (Weinh)* 2018; **5**: 1700277–11.

26   Liu Y, Yang Y, Sun M, Cui M, Fu Y, Lin Y *et al.* Highly specific noninvasive photoacoustic
and positron emission tomography of brain plaque with functionalized croconium dye labeled
by a radiotracer. *Chem Sci* 2017; **8**: 2710–2716.

27   Hu S, Wang LV. Neurovascular photoacoustic tomography. *Front Neuroenergetics* 2010; **2**: 10.

28 Zhang P, Li L, Lin L, Hu P, Shi J, He Y *et al.* High-resolution deep functional imaging of the
whole mouse brain by photoacoustic computed tomography in vivo. *J Biophoton* 2018; **11**:
e201700024–6.

29 Luke GP, Yeager D, Emelianov SY. Biomedical Applications of Photoacoustic Imaging with
Exogenous Contrast Agents. *Ann Biomed Eng* 2011; **40**: 422–437.

30 Weber J, Beard PC, Bohndiek SE. Contrast agents for molecular photoacoustic imaging. *Nat
Meth* 2016; **13**: 639–650.

31 Huynh E, Jin CS, Wilson BC, Zheng G. Aggregate enhanced trimodal porphyrin shell
microbubbles for ultrasound, photoacoustic, and fluorescence imaging. *Bioconjugate Chem*
2014; **25**: 796–801.

32 Ju K-Y, Kang J, Pyo J, Lim J, Chang JH, Lee J-K. pH-Induced aggregated melanin
nanoparticles for photoacoustic signal amplification. *Nanoscale* 2016; **8**: 14448–14456.

33 Moon H, Kang J, Sim C, Kim J, Lee H, Chang JH *et al.* Multifunctional theranostic contrast
agent for photoacoustics- and ultrasound-based tumor diagnosis and ultrasound-stimulated local
tumor therapy. *Journal of Controlled Release* 2015; **218**: 63–71.

34 Deán-Ben XL, Sela G, Lauri A, Kneipp M, Ntziachristos V, Westmeyer GG *et al.* Functional
optoacoustic neuro-tomography for scalable whole-brain monitoring of calcium indicators. *Light
Sci Appl* 2016; **5**: e16201–7.

35 Roberts S, Seeger M, Jiang Y, Mishra A, Sigmund F, Stelzl A *et al.* Calcium Sensor for
Photoacoustic Imaging. *J Am Chem Soc* 2018; **140**: 2718–2721.

36 Ruo B, Zhang R, Li L, Shao J-Y, Wang LV. Photoacoustic imaging of voltage responses beyond
the optical diffusion limit. *Sci Rep* 2017; **7**: 2560.

37 Nowak K, Mix E, Gimsa J, Strauss U, Sriperumbudur KK, Benecke R *et al.* Optimizing a
Rodent Model of Parkinson's Disease for Exploring the Effects and Mechanisms of Deep Brain
Stimulation. *Parkinson's Disease* 2011; **2011**: 1–19.

38 Zhang HK, Yan P, Kang J, Abou DS, Le HND, Jha AK *et al.* Listening to membrane potential:
photoacoustic voltage-sensitive dye recording. *J Biomed Opt* 2017; **22**: 045006.

39 Szabo GG, Du X, Oijala M, Varga C, Parent JM, Soltesz I. Extended Interneuronal Network of
the Dentate Gyrus. *Cell Rep* 2017; **20**: 1262–1268.

40 Johnston MV, Ammanuel S, O'Driscoll C, Wozniak A, Naidu S, Kadam SD. Twenty-four hour
quantitative-EEG and in-vivo glutamate biosensor detects activity and circadian rhythm
dependent biomarkers of pathogenesis in Mecp2 null mice. *Front Syst Neurosci* 2014; **8**: 118.

41 Medina-Ceja L, Pardo-Peña K, Morales-Villagrán A, Ortega-Ibarra J, López-Pérez S. Increase
in the extracellular glutamate level during seizures and electrical stimulation determined using a
high temporal resolution technique. *BMC Neurosci* 2015; **16**: 11.

42   Clay M, Monbouquette HG. A Detailed Model of Electroenzymatic Glutamate Biosensors To
Aid in Sensor Optimization and in Applications in Vivo. *ACS Chem Neurosci* 2018; **9**: 241–251.

43   Moon J-M, Thapliyal N, Hussain KK, Goyal RN, Shim Y-B. Conducting polymer-based
electrochemical biosensors for neurotransmitters: A review. *Biosens Bioelectron* 2018; **102**:
540–552.

44   Sauleau P, Lapouble E, Val-Laillet D, Malbert CH. The pig model in brain imaging and
neurosurgery. *Animal* 2009; **3**: 1138–1151.

45   Kang J, Boctor EM, Adams S, Kulikowicz E, Zhang HK, Koehler RC *et al.* Validation of
noninvasive photoacoustic measurements of sagittal sinus oxyhemoglobin saturation in hypoxic
neonatal piglets. *J Appl Physiol* 2018; **125**: 983–989.

46   American National Standard for the Safe Use of Lasers. *Annals of Internal Medicine* 1975; **82**:
132.

47   Baumann MH, Ayestas MA, Partilla JS, Sink JR, Shulgin AT, Daley PF *et al.* The designer
methcathinone analogs, mephedrone and methylone, are substrates for monoamine transporters
in brain tissue. *Neuropsychopharmacology* 2012; **37**: 1192–1203.

48   Elmore JS, Dillon-Carter O, Partilla JS, Ellefsen KN, Concheiro M, Suzuki M *et al.*
Pharmacokinetic Profiles and Pharmacodynamic Effects for Methylone and Its Metabolites in
Rats. *Neuropsychopharmacology* 2017; **42**: 649–660.

49   Airan RD, Meyer RA, Ellens NPK, Rhodes KR, Farahani K, Pomper MG *et al.* Noninvasive
Targeted Transcranial Neuromodulation via Focused Ultrasound Gated Drug Release from
Nanoemulsions. *Nano Lett* 2017; **17**: 652–659.

50   Kang SK, Ammanuel S, Thodupunuri S, Adler DA, Johnston MV, Kadam SD. Sleep
dysfunction following neonatal ischemic seizures are differential by neonatal age of insult as
determined by qEEG in a mouse model. *Neurobiol Dis* 2018; **116**: 1–12.

a Microdialysis
Bregma -3.8 mm
Lateral 3 mm
Depth 3.6 mm

Lateral 3 mm

b Sagittal section

Motor cortex

Sensory cortex

Visual cortex

Hippocampus

DG

1 mm

PA intensity at 790nm

c

Baseline

Microdialysis

-10 – 0 min

1 mm

NMDA infusion

0 – 10 min

10 – 20 min

20 – 30 min

0.3 mM NMDA

3 mM NMDA

0   fPA VSD response [%]   3

a Microdialysis

Bregma -3.8 mm
Lateral 3 mm
Depth 3.6 mm

Signal

Reference

b

Change in glutamate (% basal)

- 0.3 mM NMDA
- 1 mM NMDA
- 3 mM NMDA
- Baseline

NMDA

Time (min)

c

γ power (μv²/Hz)

3.0 mM NMDA

*

+261%

Time (min)

d

θ power (μv²/Hz)

*

+13%

Time (min)

e

50Hz

*

0Hz

0 min   15 min

200 μv²/Hz

0 μv²/Hz

f

(0.5-50Hz)

100 μV
2 sec

γ (35-50Hz)

50 μV
2 sec

θ (5.5-8.5Hz)

50 μV
2 sec

I

II

III

IV

V

Time (10 sec)

80 µV

1sec