Automated High-Throughput Light-Sheet Fluorescence Microscopy of Larval Zebrafish

Savannah L. Logan[†], Christopher Dudley[†], Ryan P. Baker[†], Michael J. Taormina, Edouard A. Hay, Raghuveer Parthasarathy*

Materials Science Institute , Institute of Molecular Biology, and Department of Physics, The University of Oregon, Eugene, OR 97403

[†] Equal contributors

* Corresponding author: raghu@uoregon.edu

ABSTRACT

Light sheet fluorescence microscopy enables fast, minimally phototoxic, three-dimensional imaging of live specimens, but is currently limited by low throughput and tedious sample preparation. Here, we describe an automated high-throughput light sheet fluorescence microscope in which specimens are positioned by and imaged within a fluidic system integrated with the sheet excitation and detection optics. We demonstrate the ability of the instrument to rapidly examine live specimens with minimal manual intervention by imaging fluorescent neutrophils over a nearly 0.3 mm^3 volume in dozens of larval zebrafish. In addition to revealing considerable inter-individual variability in neutrophil number, known previously from labor-intensive methods, three-dimensional imaging allows assessment of the correlation between the bulk measure of total cellular fluorescence and the spatially resolved measure of actual neutrophil number per animal. We suggest that our simple experimental design should considerably expand the scope and impact of light sheet imaging in the life sciences.

INTRODUCTION

Light sheet fluorescence microscopy (LSFM) is a powerful tool for examining the three-dimensional structural and temporal dynamics of living systems. In LSFM, a thin sheet of laser light excites fluorophores in a sample. Scanning the sample through the sheet enables fast, three-dimensional imaging with low phototoxicity, high resolution, and a wide field of view (1–5). Imaging with LSFM has enabled numerous studies of embryonic development (6,7), neural activity (8), microbial dynamics (9–11), and other phenomena. A large body of work has focused on improving the optical capabilities of light sheet imaging, for example using structured illumination (12,13), multiple lens pairs (6), two-photon excitation (14), and

other techniques. However, current light sheet fluorescence microscopes have significant constraints related to throughput and sample handling that, we argue, have placed much greater limitations on their scientific utility than issues of spatial or temporal resolution. The majority of existing light sheet fluorescence microscopes, both commercial and non-commercial, are designed to hold a single specimen. A few instruments (including one from the authors of this paper) can hold up to six specimens for sequential imaging. Moreover, light sheet fluorescence microscopy typically requires extensive sample preparation and manual sample mounting, most commonly by embedding specimens in agarose gels. Examination of large numbers of specimens is therefore slow and difficult, which is especially important given the high level of inter-individual variability found in many complex biological processes. Increasing the pace of insights into developmental biology, multicellular biophysics, or microbial community structure will require faster and simpler acquisition of three-dimensional imaging datasets. To date, there exists only one report of an automated light sheet microscope that makes use of fluidic positioning of live animals (15); its throughput (specimens per hour) is not stated, and though its ability to image larval zebrafish is clear, the total number of animals examined was only twelve. In contrast, automated, high-throughput methods have been integrated with other types of microscopes, including confocal microscopes (16–18), about which we comment further in the Discussion. Our system adopts and builds upon these, especially the confocal-based setup of Ref. (16).

To address the issues described above, we developed a light sheet fluorescence microscope capable of automated, high-throughput imaging of live specimens. Our instrument uses fluidic control and image-based registration to rapidly but precisely position specimens for light sheet scans and subsequently remove them from the imaging area. We

characterize the optical quality of our instrument, and demonstrate its capabilities by rapidly

imaging immune cells in dozens of larval zebrafish. While the spatial resolution of our

microscope does not equal that of current "low-throughput" light sheet microscopes, it is

more than sufficient for determining cellular distributions. Moreover, we argue that the

tradeoff of lower resolution for higher throughput is worthwhile given the large variance in

most biological datasets.

We illustrate the utility of the instrument by imaging neutrophils in dozens of larval

zebrafish. Neutrophils are an important and highly dynamic cell type of the innate immune

system. These cells migrate to sites of damage or infection and recognize and kill pathogenic

microbes (19). Zebrafish are a well recognized model organism for studying neutrophil

responses (20,21), and prior work has uncovered changes in neutrophil distributions in

response to wound-induced chemotaxis (22), drugs that lead to symptoms similar to human

enterocolitis (23), and stimulation by gut microbes or bacterial lipopolysaccharide (LPS)

(24,25), to list a few of many examples. Quantifing even basic properties such as neutrophil

abundances over large extents for the dozens of specimens required given the high variance

between individuals often necessitates slow confocal imaging or painstaking histological

sectioning or gut dissection (24–26). These sorts of studies will be greatly facilitated by

instruments that can provide large-scale 3D imaging of automatically detected and imaged

specimens.

We show, as expected, a high degree of variation between fish in total neutrophil

number and an increase in its mean value following LPS exposure, as well as spatial

clustering of neutrophils in two distinct regions near the swim bladder. While our instrument

is optimized for imaging of larval zebrafish, the design could easily be modified for rapid

imaging of a wide range of biological and non-biological samples, which should broaden the

impact of light sheet microscopy in a variety of fields.

RESULTS

Instrument design

The light sheet portion of the microscope closely follows the design of Keller et al (1); a

rapidly-scanned galvanometer creates a sheet of light for fluorescence excitation, and emitted

light is captured by a camera perpendicular to the plane of the sheet (Fig. 1A). In

conventional light sheet fluorescence microscopes, gel-mounted specimens are introduced

vertically in between horizontal lenses. To achieve high throughput, we use a continuous

fluidic path through plastic tubing and glass capillaries for transport as well as imaging,

detailed below. If the fluidic path were oriented vertically, specimens would gravitationally

drift during imaging. Therefore, we adopted a geometry in which specimens are transported

horizontally and the sheet plane is vertical (Fig. 1A,B). To allow this arrangement, we

designed an elongated sample chamber with windows oriented below and perpendicular to

the sample (Fig. 1B,C). Before entering the imaging chamber, specimens flow through a

system of 0.7 mm diameter plastic tubing at a typical flow rate of 1 ml/min, or 4 cm/sec

(Fig. 1). Flow speed and direction are controlled by a syringe pump (Fig. 1); see

Supplemental Methods for a parts list and descriptions.

Figure 1. Instrument design. (A) Schematic of the instrument design, with

labels corresponding to the parts list in Table 1. See also Supplementary

Movie 1. The excitation laser line is selected by an acousto-optic tunable

filter ($AOTF_2$) , then directed to a galvanometer mirror (G_3) and objective

lens (L_4) to create a time-averaged sheet of light in the sample chamber (C)

via a prism (Pr_5). Specimens flow through a system of tubing controlled by a

syringe pump (Pu_9) and valves (V_{10}) and are automatically positioned in a

square-walled capillary (Cap_{11}) for imaging. Bright field images are used for

positioning the sample and are illuminated with an LED (LED_6). After

imaging, specimens are directed into a reservoir (R). **(B)** Schematic of the

imaging area. The 3D-printed sample chamber (C), prism (Pr_5), and imaging

capillary (Cap_{11}) are apparent. **(C)** Photograph of the imaging area

corresponding to the schematic in (B).

Inside the imaging chamber, specimens flow into a square-walled glass capillary in front of

the imaging objective where they are automatically detected by bright field microscopy.

Specimens are rapidly stopped using computer-controlled valves on either side of the

imaging chamber, with a precision of approximately 1 mm in position, comparable to the

length of a larval zebrafish. Fine positioning is performed by iterated movement of the

capillary by a computer-controlled stage, brightfield imaging, and comparison of images with

a previously assembled image library (see Methods) (Fig. 2 A,B). The travel range of the

capillary on the stage allows movement of up to 30 mm in the x-direction. Like many

studies, ours make use of larval zebrafish as a model organism; strong features such as eyes

and the swim bladder enable straightforward correlation-based registration, with a precision

of about 20 μm as described below. This approach should be applicable to any specimen

with a roughly sterotypical anatomy. The specimen is not rotated about the tube axis; we

comment further on this in the Discussion section.

Once positioned, specimens are automatically imaged using LSFM. The imaging chamber

has sufficient depth, 35 mm, that the capillary can be scanned through the sheet by a

motorized stage. In our setup, repeated scans with up to three excitation wavelengths are

possible; this is limited simply by the number of available laser lines. The precision of the

automated positioning enables scans to be taken of particular regions, for example the larval

gut, as shown below. After imaging, specimens flow into a collection reservoir and

subsequent specimens are automatically positioned for imaging. We provide a movie of the

instrument in operation as Supplementary Video S1. A complete parts list is provided as

Table 1, in Supplemental Methods.

Optical quality

Our instrument uses glass capillaries for specimen mounting, rather than more conventional

gel embedding. The square cross-section of these capillaries should lead to less distortion

than more common cylindrical capillaries. To assess the optical quality of our setup, we

measured the point spread function (PSF) by imaging 28 nm diameter fluorescent

microspheres dispersed in oil in these capillaries (see Methods for details). The diffraction-

limited width of the particles in the sheet plane (xy), assessed as the standard deviation of a

Gaussian function fit to the particle's intensity profile, is 0.6 μm, and the width along the

detection axis (z) is 3.4 ± 0.6 μm, consistent with the expected sheet thickness of our setup

(Fig. 2 C,D).

Figure 2. Specimen positioning and image quality. (A) Composite

brightfield image of a larval zebrafish positioned in a glass capillary. Scale bar:

50 μm. (B) Normalized intensity averaged along the short axis of the

brightfield image, and the intensity of the template image that best matches

the fish in (A). Cross-correlation with the template is used to automatically

position the fish for light sheet fluorescence imaging. (C) Light sheet

fluorescence images of a 28 nm diameter fluorescent microsphere, showing

x-y and *z-y* planes centered on the particle. (D) Line-scan of intensity along

the detection axis (*z*) through a fluorescent microsphere, with a Gaussian fit

showing a width of approximately 3 μm.

Data collection capabilities

Using this system, we can image approximately 30 larval zebrafish per hour, obtaining from each a 666 x 431 x 1060 μm (x, y, z) three-dimensional scan, a marked improvement over manual mounting and imaging that, even by a skilled researcher, is limited to about 5 fish per hour. The triggering accuracy is about 90%, with roughly 10% of detected objects being bubbles or debris that are easily identified after imaging. On average, 81% of larval fish are automatically positioned correctly in front of the imaging objective. The remaining 19% correspond to multiple fish being in the field of view, or other positioning errors. Importantly, the majority of the run time of the instrument is spent obtaining light sheet fluorescence images, and is not dominated by specimen positioning. In the batch of N=41 fish whose neutrophil distributions were imaged in experiments described below, for example, the flow, detection, and positioning of the specimens occupied only approximately 30 seconds per fish. In the limit of zero imaging time (e.g. for very bright signals or small regions of interest), the system could therefore record data from up to about 120 specimens per hour in the absence of triggering or positioning errors, or about 90 specimens per hour with the present system performance.

Neutrophils in larval zebrafish

To demonstrate the capabilities of the instrument, we imaged fluorescent neutrophils in larval zebrafish at 5 days post-fertilization (dpf), focusing especially on the number of these immune cells and their distribution near the anterior of the intestine. Specifically, these were fish engineered to express green fluorescent protein (GFP) under the promoter myeloperoxidase, an enzyme primarily produced in neutrophils (27).

Positioning accuracy

To assess the positioning accuracy of our instrument, we performed automated scans of

larval zebrafish with GFP-expressing neutrophils, and then reloaded the same fish and re-

scanned them. Changes in the neutrophil positions are due to both imaging error and to

motion of the neutrophils during the intermediate time. (Neutrophils are highly motile cells,

crawling through tissue and also entering or leaving tissue via the bloodstream.) For five

twice-scanned fish, we manually identified three neutrophils that were unambiguously the

same in each scan (i.e. not newly entered or departed). For these neutrophils, the within-fish

standard deviations of the changes in position provide a measure of the biological variation,

e.g. from neutrophil motion. These were 5.3 ± 3.5 μm, 11.6 ± 9.0 μm, and 40.0 ± 36.4 μm,

for x, y, and z, respectively, where x is the flow direction and z is perpendicular to the light

sheet. The standard deviation between fish of the mean neutrophil positions provides a

measure of the instrumental variation, e.g. from imperfect positioning. These were $18.9 \pm$

6.7 μm, 56.1 ± 19.9 μm, and 52.6 ± 18.6 μm for x, y, and z, respectively. Along the capillary

axis, therefore, the fish positioning is reproducible to within about 20 μm. The larger

variance in y and z is expected, as we are not explicitly detecting location along these axes,

and because there can be specimen rotation about the x-axis. Overall, therefore, global

positioning uncertainty is on the order of a cell diameter for cells such as neutrophils.

Neutrophil number and variance

In total, we imaged 41 fish, obtaining from each a single 666 x 431 x 1060 μm three-

dimensional image in which neutrophils were readily evident (Fig. 3A and Supplementary

Movie S2). Brightfield images are captured and saved prior to fluorescence imaging; a set of

four such images, demonstrating their appearance and variance, are provided as

Supplemental Figure 1. The strong GFP signal enabled automated identification of

neutrophils by standard segmentation methods. Corroborating previous work done by

manual dissection of zebrafish (24), we found a high degree of variation in neutrophil

number between specimens, with the standard deviation being 30% of the mean (Fig. 3B).

Furthermore, we found that neutrophils tend to cluster in two distinct regions: adjacent to

the swim bladder on both the anterior and posterior (Fig. 3C). Notably, the total GFP

intensity in a fish is weakly correlated with the number of neutrophils, with a coefficient of

determination $R^2 = 0.4$, indicating that a simple measure of overall brightness, as could be

assessed without three-dimensional microscopy, would provide a poor diagnostic of the

actual abundance of immune cells (Fig. 3D). We also compared neutrophil counts

determined from two-dimensional maximum intensity projections of the full three-

dimensional image stacks, mimicking images that would be obtained from simple widefield

fluorescence microscopy. This yielded a number of neutrophils that was on average 0.76 ±

0.02 of that from the three-dimensional counts, indicating as expected that some neutrophils

are behind others in the three-dimensional space of the specimen, and hence require three-

dimensional imaging for accurate assessment (Supplemental Figure 2).

Figure 3. Imaging neutrophils in larval zebrafish. (A) A maximum intensity projection of a three-dimensional light sheet fluorescence image of GFP-expressing neutrophils near the intestine of a 5 dpf larval zebrafish. The 3D scan is provided as Supplemental Movie 2. The intestine and swim bladder are roughly outlined by the yellow dotted lines. Scale bar: 100 μm. (B) The total number of neutrophils in each fish; the x-axis is ordered by neutrophil count. (C) Neutrophil count along the anterior-posterior dimension, summed over all fish examined (*N*=41). The x-axis corresponds approximately to the horizontal range of (A). (D) The total intensity of the detected neutrohils per fish vs the total number of neutrophils in that fish. The two measures are weakly correlated with a coefficient of determination $R^2 = 0.4$.

In addition, we examined neutrophil number in response to exposure of larval

zebrafish to soluble LPS, the major component of the outer membrane of gram negative

bacteria, known from earlier studies to stimulate an inflammatory immune reaction (25,28).

As above, we detected and scanned transgenic zebrafish with GFP-expressing neutrophils,

exposed to LPS in their flask water at a concentration of 150 µg/ml for 0 (control), 2, and 24

hours prior to imaging. As above, neutrophil number shows a large degree of variability. At 5

dpf, we detected 85.0 ± 17.0 ($N = 13$) and 87.6 ± 27.6 ($N = 14$) (mean $\pm$ standard

deviation) neutrophils in fish subjected to no LPS and LPS for 2 hours, respectively;

indicating no discernible change. At 6 dpf, we detected 95.1 ± 23.4 ($N = 21$) and $114.2 \pm$

38.1 ($N = 19$) in the control and 24-hour LPS-treated fish, respectively, indicating an

increase in the mean neutrophil count by a factor of 1.2 ± 0.1.

Figure 4. Neutrophil counts after exposure to LPS. The number of

neutrophils counted from light sheet fluorescence images of larval zebrafish

after exposure to 150 µg/ml LPS for 0 (control), 2, or 24 hours. At 2 hours

post-exposure, there was no discernible difference between the LPS-treated

group and the control group. At 24 hours post-exposure, the LPS group

displayed an increase in mean neutrophil count of 1.2 ± 0.1.

DISCUSSION

In recent years, many instruments for high content and high throughput imaging of small

organisms, such as embryonic and larval zebrafish, have been developed (15–18,29–32), and

have demonstrated their utility for fast, accurate imaging and screening. Our setup has

similarities and differences to several of these, and advantages and disadvantages for

particular applications. Several instruments are designed to image specimens held in multi-

well plates (29–32), including commercial instruments such as the ImageXpress High-

Content Screening System (33) and the Acquifer Imaging Machine (34). These plate-based

systems enable integration with standard specimen containment, and also facilitate

automated delivery of chemicals or other perturbations. Imaging is typically provided by

widefield microscopy, which is rapid and simple but which prohibits three-dimensional

imaging. While confocal microscopy is in principle possible, the thickness of standard plate

bottoms would make this difficult except for specialized setups. Three-dimensional confocal

fluorescence microscopy is, however, attained in automated instruments such as those of

Refs. (16) and the commercial VAST device (35), which use capillary- and tubing-based

automated specimen handling and positioning. While successful, and enabling a wide range

of studies, there are applications for which the advantages of light sheet microscopy

compared to confocal microscopy, namely its rapid speed and low phototoxicity (1–5,36),

are important, such as the imaging of active cells in dynamic environments like the intestine.

The requirement of orthogonal accessible light paths for light sheet imaging make its

integration with existing high-throughput instruments non-trivial, motivating the work

presented here. Our design of a relatively freestanding capillary amid a vertical excitation

sheet and horizontal detection axis is effective, but it need not be the only solution. For

example, one could imagine integration with single-lens-based light sheet techniques (37,38),

and it will be interesting to see if such schemes are developed. We note that our device will

have lower throughput than plate-based instruments, and will not integrate with commercial

confocal microscopes as existing tubing-based methods do, but rather will enhance studies

for which light sheet fluorescence microscopy allows insights into dynamic biological

phenomena that are otherwise unattainable.

While our microscope is optimized for rapid imaging of larval zebrafish, the design is

general and opens numerous possibilities for imaging a wide range of specimens, such as

organoids, drosophilia embryos, small marine invertebrates, and more, with the appropriate

expansion of the positioning image library.

Our design does not rotate the specimen about the travel axis, a design choice that

has advantages and disadvantages for applications. It is certainly possible to vary the

specimen orientation by rotating the glass capillary, as, for example, in the VAST instrument

(35). Rotation would enhance image quality, both by allowing the selection of particular

orientations with minimal sheet distortion and by enabling the fusion of multiple views to

gain isotropic resolution. We note, moreover, that variation in image quality due to the

uncontrolled orientations of zebrafish likely contributes, along with true biological variability,

to the total scatter in neutrophil counts (Figs. 3, 4). The ability to examine a large number of

specimens allows averaging over both sources of variation. Though it could be implemented,

rotational positioning of each animal takes time and, at least in our design, significantly

reduces the throughput of the instrument. Furthermore, we believe that many useful

applications, e.g. the neutrophil counting demonstrated here, can be realized without

specimen rotation, and that the simplicity of the setup as presented can hopefully foster

widespread adoption. We note that rotation can be decoupled from the imaging area, as in

Ref. (35), so that one specimen can be oriented while another is imaged, preventing a

reduction in throughput. Such additions to the instrument described, though carrying a cost

of greater complexity, may be worthwhile.

312   While the present design provides only a single three-dimensional image of each

specimen, we envision future integration of a closed-loop fluidic circuit, with which

specimens can be automatically loaded, imaged, and circulated repeatedly, allowing for high-

throughput acquisition of multiple snapshots of the same specimen over time. In general,

tackling the challenge of automated, high-throughput specimen handling will allow the

technique of light sheet fluorescence microscopy to maximize its impact on the life sciences.

METHODS

Hardware

The majority of the instrument was constructed with off the shelf parts. Custom parts were

either laser cut from acrylic sheets or were 3D printed (see Parts List).

Fluorescence excitation is provided by various solid state lasers, selected by an acousto-optic

tunable filter (AOTF, AA Opto-electronic) for coupling into a fiber launch to a

galvanometer mirror (Cambridge Technology), which oscillates with a triangular waveform

at 1 kHz to sweep the beam into a sheet. An objective lens (Mitutoyo 5X) and a prism route

the sheet to the water-filled sample chamber where it intersects the specimen (Fig. 1).

Detection is provided by a 20x water immersion objective (Olympus 20XW), mounted in the

side of the chamber and sealed with an o-ring, its corresponding tube lens, and an sCMOS

camera (Hamamatsu Orca Flash 4.0). Exposure times were 25 ms for all experiments

presented here. Instrument control software was written in MATLAB.

Ethics statement

All zebrafish experiments were carried out in accordance with protocols approved by the

University of Oregon Institutional Animal Care and Use Committee (39).

Zebrafish husbandry

The zebrafish line *Tg[BACmpo:gfp]* (27) was used for neutrophil imaging. Larval zebrafish

were raised at a density of one embryo per milliliter and kept at a constant temperature of 28

ºC. Embryos were not fed during experiments and were sustained by their yolks.

Sample preparation

Larval zebrafish were placed in dishes containing sterile embryo media with 0.05%

methylcellulose and anaesthetized using 240 µg/ml tricaine methanesulfonate. This

anaesthetic concentration is higher than the standard dosage, but was necessary likely

because of permeation through the plastic tubing. Specimens are initially loaded into the

tubing system by manual aspiration using a syringe connected to the opposite end of the

tubing, maintaining a spacing between specimens of approximately 6 inches. Fish are pulled

into the tubing head-first to ensure high and consistent flow speeds. (Fish travel better

forwards than backwards). During experiments, larvae flowed through the tubing and were

automatically stopped and positioned by a syringe pump and a series of valves. After

imaging, larvae flowed into a dish containing sterile embryo medium.

LPS Treatment

A filter-sterilized LPS (*E. coli* serotype 0111:B4, Sigma) solution was injected into flasks

containing 15 zebrafish larvae at 5 dpf for a final concetration of 150 µg/ml of LPS in each

flask. Fish were incubated in the LPS solution for 2-24 hours as indicated in the text, after

which they were removed from the solution and imaged.

Imaging Procedure

For live imaging, each larval zebrafish flows through 0.7 mm inner diameter silicone tubing

to a 50 mm long section of round 0.7 mm inner diameter, square exterior cross section, glass

tubing in the specimen chamber. The contrast between the specimen and background in

brightfield images is used to detect the specimen, stop the pump, and close off tubing using

pinch valves to prevent specimen drift. Detecting the specimen and halting the flow places

the specimen within three to five millimeters of the desired location. Fine positioning of the

specimen is performed by translating the specimen along the capillary axis ("x") in roughly

30 steps of 0.3 mm size, capturing a brightfield image at each position, tiling the brightfield

images into one large image, and comparing the result with a previously assembled "library"

image. The comparison can be done in one of two ways, which yield indistinguishable

positioning accuracies, both of which begin by averaging the image intensity along the "y"

direction, giving a one-dimensional intensity profile. The profile is then compared with the

profile derived from average library images either (i) by cross-correlation, in which one

profile is offset in position and then multiplied by the other profile, and integrated

(summed), giving a correlation value as a function of offset value, or (ii) the location of the

profile's intensity minimum is determined, and compared to the location of the average

library profile's minimum. In (i) the peak in the correlation function and in (ii) the difference

between the minima gives the displacement between the profile and the library profile, and

therefore indicates where the specimen should be precisely positioned. Following

positioning, bright field illumination is switched off and the desired laser wavelength is

selected by the AOTF. The *xyz* arm is then scanned perpendicular to the sheet plane,

generating a 3D image stack. The scan can be repeated for another region or another

wavelength before the pump is directed to send the specimen out of the chamber, and bring

in the next specimen.

Point Spread Function

For measurements of the point spread function, 28 nm diameter fluorescent carboxylate-

modified polystyrene spheres (Thermo Fisher cat. #F8787), with peak excitation and

emission wavelengths 505 and 515 nm, respectively, were dispersed in oil with a similar

index of refraction as water (Zeiss Immersol W 2010) inside the imaging capillaries. Three-

dimensional scans were taken, with a z-spacing of 0.5 μm.

Image-based neutrophil quantification

Neutrophils were detected from image stacks using custom code written in Python,

involving a coarse and fine thresholding. First, the 3D stack was thresholded using a low

intensity value followed by the morphological operations of closing and erosion with a range

of 1 pixel (40). Connected above-threshold pixels were identified as objects using scikit-

image's label function. Then, each detected object was further thresholded using an Otsu

filter and labeled again. This re-thresholding process was repeated twice, and objects below 3

μm^3 in volume were discarded as unphysical. The total intensity within each identified object

represents the total fluorescence intensity of that neutrophil. For a full 3D dataset, the

processing takes 1-2 minutes.

Raw data

Raw image data, roughly 150 GB in total size, is available from the authors on request.

ACKNOWLEDGEMENTS

We thank Rose Sockol, Kyleah Murphy, and the University of Oregon Zebrafish Facility

staff for fish husbandry, and Karen Guillemin and Brandon Schlomann for useful comments

and conversations. Research reported in this publication was supported in part by the

National Science Foundation (NSF) under award 1427957 (to RP), by the the M. J. Murdock

Charitable Trust, and by the National Institutes of Health (NIH) as follows: by the National

Institute of General Medical Sciences under award number P50GM098911 and by the

National Institute of Child Health and Human Development under award P01HD22486,

which provided support for the University of Oregon Zebrafish Facility. The content is

solely the responsibility of the authors and does not represent the official views of the NIH,

NSF, or other funding agencies.

REFERENCES:

1. Keller PJ, Schmidt AD, Wittbrodt J, Stelzer EHK. Reconstruction of Zebrafish Early
Embryonic Development by Scanned Light Sheet Microscopy. Science. 2008 Nov
14;322(5904):1065–9.

2. Power RM, Huisken J. A guide to light-sheet fluorescence microscopy for multiscale
imaging. Nat Methods. 2017 Apr;14(4):360–73.

3.   Keller PJ, Ahrens MB, Freeman J. Light-sheet imaging for systems neuroscience
[Internet]. Nature Methods. 2014 [cited 2018 Mar 22]. Available from:
https://www.nature.com/articles/nmeth.3214

4.   Huisken J. Slicing embryos gently with laser light sheets. BioEssays [Internet]. 2012
[cited 2018 Mar 22];34. Available from:
https://www.readcube.com/articles/10.1002/bies.201100120

5.   Santi PA. Light Sheet Fluorescence Microscopy. J Histochem Cytochem. 2011 Feb
1;59(2):129–38.

6.   Tomer R, Khairy K, Amat F, Keller PJ. Quantitative high-speed imaging of entire
developing embryos with simultaneous multiview light-sheet microscopy. Nat
Methods. 2012 Jul;9(7):755–63.

7.   Weber M, Scherf N, Meyer AM, Panáková D, Kohl P, Huisken J. Cell-accurate optical
mapping across the entire developing heart. eLife. 2017 Dec 29;6:e28307.

8.   Keller PJ, Ahrens MB. Visualizing Whole-Brain Activity and Development at the Single-
Cell Level Using Light-Sheet Microscopy. Neuron. 2015 Feb;85(3):462–83.

9.   Wiles TJ, Jemielita M, Baker RP, Schlomann BH, Logan SL, Ganz J, et al. Host Gut
Motility Promotes Competitive Exclusion within a Model Intestinal Microbiota. PLOS
Biol. 2016 Jul 26;14(7):e1002517.

10. Logan SL, Thomas J, Yan J, Baker RP, Shields DS, Xavier JB, et al. The Vibrio cholerae
Type VI Secretion System Can Modulate Host Intestinal Mechanics to Displace
Commensal Gut Bacteria. bioRxiv. 2017 Nov 29;226472.

11. Parthasarathy R. Monitoring microbial communities using light sheet fluorescence
microscopy. Curr Opin Microbiol. 2018 Jun;43:31–7.

12. Chen B-C, Legant WR, Wang K, Shao L, Milkie DE, Davidson MW, et al. Lattice light-
sheet microscopy: Imaging molecules to embryos at high spatiotemporal resolution.
Science. 2014 Oct 24;346(6208):1257998.

13. Chang B-J, Meza VDP, Stelzer EHK. csiLSFM combines light-sheet fluorescence
microscopy and coherent structured illumination for a lateral resolution below 100 nm.
Proc Natl Acad Sci. 2017 May 9;114(19):4869–74.

14. Truong TV, Supatto W, Koos DS, Choi JM, Fraser SE. Deep and fast live imaging with
two-photon scanned light-sheet microscopy. Nat Methods. 2011 Sep;8(9):757–60.

15. Gualda EJ, Pereira H, Vale T, Estrada MF, Brito C, Moreno N. SPIM-fluid: open source
light-sheet based platform for high-throughput imaging. Biomed Opt Express. 2015
Nov 1;6(11):4447–56.

16. Pardo-Martin C, Chang T-Y, Koo BK, Gilleland CL, Wasserman SC, Yanik MF. High-
throughput *in vivo* vertebrate screening. Nat Methods. 2010 Aug;7(8):634–6.

17. Yanik MF, Rohde CB, Pardo-Martin C. Technologies for Micromanipulating, Imaging,
and Phenotyping Small Invertebrates and Vertebrates. Annu Rev Biomed Eng.
2011;13(1):185–217.

18. Hwang H, Lu H. Microfluidic tools for developmental studies of small model organisms
— nematodes, fruit flies, and zebrafish. Biotechnol J. 2013 Feb;8(2):192–205.

19. Mayadas TN, Cullere X, Lowell CA. The multifaceted functions of neutrophils. Annu
Rev Pathol. 2014;9:181–218.

20. Meijer AH, Spaink HP. Host-pathogen interactions made transparent with the zebrafish
model. Curr Drug Targets. 2011 Jun;12(7):1000–17.

21. Galindo-Villegas J. Recent findings on vertebrate developmental immunity using the
zebrafish model. Mol Immunol. 2016 Jan 1;69:106–12.

22. Kadirkamanathan V, Anderson SR, Billings SA, Zhang X, Holmes GR, Reyes-Aldasoro
CC, et al. The Neutrophil's Eye-View: Inference and Visualisation of the
Chemoattractant Field Driving Cell Chemotaxis In Vivo. PLoS ONE. 2012 Apr
26;7(4):e35182.

23. Oehlers SH, Flores MV, Okuda KS, Hall CJ, Crosier KE, Crosier PS. A chemical
enterocolitis model in zebrafish larvae that is dependent on microbiota and responsive
to pharmacological agents. Dev Dyn. 2011 Jan 1;240(1):288–98.

24. Rolig AS, Parthasarathy R, Burns AR, Bohannan BJM, Guillemin K. Individual
Members of the Microbiota Disproportionately Modulate Host Innate Immune
Responses. Cell Host Microbe. 2015 Nov 11;18(5):613–20.

25. Bates JM, Akerlund J, Mittge E, Guillemin K. Intestinal alkaline phosphatase detoxifies
lipopolysaccharide and prevents inflammation in zebrafish in response to the gut
microbiota. Cell Host Microbe. 2007 Dec 13;2(6):371–82.

26. Kanther M, Tomkovich S, Xiaolun S, Grosser MR, Koo J, Flynn EJ, et al. Commensal
microbiota stimulate systemic neutrophil migration through induction of serum
amyloid A. Cell Microbiol. 2014 Jul;16(7):1053–67.

27. Renshaw SA, Loynes CA, Trushell DMI, Elworthy S, Ingham PW, Whyte MKB. A
transgenic zebrafish model of neutrophilic inflammation. Blood. 2006 Dec
15;108(13):3976–8.

28. Bates JM, Mittge E, Kuhlman J, Baden KN, Cheesman SE, Guillemin K. Distinct signals
from the microbiota promote different aspects of zebrafish gut differentiation. Dev
Biol. 2006;297(2):374–86.

29. Spomer W, Pfriem A, Alshut R, Just S, Pylatiuk C. High-throughput screening of
zebrafish embryos using automated heart detection and imaging. J Lab Autom. 2012
Dec;17(6):435–42.

30. Letamendia A, Quevedo C, Ibarbia I, Virto JM, Holgado O, Diez M, et al. Development
and Validation of an Automated High-Throughput System for Zebrafish In Vivo
Screenings. PLOS ONE. 2012 May 15;7(5):e36690.

31. Liu X, Painter RE, Enesa K, Holmes D, Whyte G, Garlisi CG, et al. High-throughput
screening of antibiotic-resistant bacteria in picodroplets. Lab Chip. 2016 Apr
26;16(9):1636–43.

32. Westhoff JH, Giselbrecht S, Schmidts M, Schindler S, Beales PL, Tönshoff B, et al.
Development of an Automated Imaging Pipeline for the Analysis of the Zebrafish
Larval Kidney. PLOS ONE. 2013 Dec 4;8(12):e82137.

33. ImageXpress Micro XLS Widefield High-Content Analysis System | Molecular Devices.
Available from: https://www.moleculardevices.com/products/additional-
products/imagexpress-micro-xls-widefield-high-content-analysis-system#gref

34. ACQUIFER. Screening. ACQUIFER. Available from:
https://www.acquifer.de/screening/

35. Biometrica S 180 Blackfly Interactive, Union. Union Biometrica | VAST BioImager™
Platform Overview. 2010. Available from: http://www.unionbio.com/vast/

36. Jemielita M, Taormina MJ, Delaurier A, Kimmel CB, Parthasarathy R. Comparing
phototoxicity during the development of a zebrafish craniofacial bone using confocal
and light sheet fluorescence microscopy techniques. J Biophotonics. 2013;6:920–8.

37. Dunsby C. Optically sectioned imaging by oblique plane microscopy. Opt Express. 2008
Dec 8;16(25):20306–16.

38. Bouchard MB, Voleti V, Mendes CS, Lacefield C, Grueber WB, Mann RS, et al. Swept
confocally-aligned planar excitation (SCAPE) microscopy for high-speed volumetric
imaging of behaving organisms. Nat Photonics. 2015 Feb;9(2):113–9.

39. Westerfield M. The zebrafish book. A guide for the laboratory use of zebrafish (Danio
rerio). Vol. 5th. Eugene, OR: University of Oregon Press; 2007. Available from:
http://zfin.org/zf_info/zfbook/zfbk.html

40. Gonzalez RC, Woods RE. Digital Image Processing. Upper Saddle River, N.J.: Prentice
Hall; 2008.

SUPPORTING INFORMATION CAPTIONS:

Supplemental Movie 1. Video of the automated, high-throughput light sheet microscope in operation. After a larval zebrafish, flowing through tubing, is detected and positioned, 3D scans in two colors in two different regions, are obtained. The final image (0:57-1:03) is of tdTomato-labeled bacteria and GFP-labeled neutrophils.

Supplemental Movie 2. A light sheet fluorescence scan through the anterior intestine of a larval zebrafish with GFP-expressing neutrophils evident as very bright objects above autofluorescent background. The total image depth is 475 µm (2.5 µm/slice x 190 slices). The maximum intensity projection of this scan is shown in Figure 2E.

Supplemental Figure 1. Sample brightfield images of four larval zebrafish, captured and saved prior to light sheet fluorescence imaging. Notably, the orientation is random in the capillary, with a tendancy toward either "gut down" or "gut up." Scale bar: 100 µm.

Supplemental Figure 2. The number of neutrophils in 67 larval zebrafish, assessed from two-dimensional maximum intensity projections and from the full three-dimensional light sheet fluorescence scans. The former is 0.76 ± 0.02 of the latter, indicating that three-dimensional imaging is necessary to capture all the cells in a three-dimensional volume.