Super-resolution fight club: A broad assessment of 2D & 3D single-molecule localization microscopy software

Daniel Sage[*+1], *Thanh-An Pham*[+1], *Hazen Babcock*[2], *Tomas Lukes*[3], *Thomas Pengo*[4], *Jerry Chao*[5], *Ramraj Velmurugan*[5], *Alex Herbert*[6], *Anurag Agrawal*[7], *Silvia Colabrese*[1,8], *Ann Wheeler*[9], *Anna Archetti*[10], *Bernd Rieger*[11], *Raimund Ober*[5], *Guy M. Hagen*[12], *Jean-Baptiste Sibarita*[13], *Jonas Ries*[14], *Ricardo Henriques*[15], *Michael Unser*[1], *Seamus Holden*[*+16]*

*Corresponding authors: daniel.sage@epfl.ch, seamus.holden@ncl.ac.uk.

+Equal contribution

1: Biomedical Imaging Group, School of Engineering, Ecole Polytechnique Fédérale de Lausanne (EPFL), Switzerland
2: Harvard Center for Advanced Imaging, Harvard University, Cambridge, Massachusetts, USA
3: Laboratory of Nanoscale Biology & Laboratoire d'Optique Biomédicale, STI - IBI, EPFL, Lausanne, Switzerland
4: University of Minnesota Informatics Institute, University of Minnesota Twin Cities, USA
5: Electrical Engineering, University of Texas Dallas, Richardson, Texas, USA
6: MRC Genome Damage and Stability Centre, School of Life Sciences, University of Sussex, Brighton, UK
7 : Double Helix LLC, Boulder, Colorado, USA
8 : Istituto Italiano di Tecnologia, Genova, Italy
9: Edinburgh Super-Resolution Imaging Consortium, University of Edinburgh, UK
10 : Laboratory of Experimental Biophysics, École Polytechnique Fédérale de Lausanne (EPFL), Lausanne, Switzerland
11: Department of Imaging Physics, Faculty of Applied Sciences, Delft University of Technology, The Netherlands
12: UCCS center for the Biofrontiers Institute, University of Colorado at Colorado Springs, Colorado, USA
13: Institut Interdisciplinaire de Neurosciences, University of Bordeaux, France
14: European Molecular Biology Laboratory (EMBL), Cell Biology and Biophysics Unit, Heidelberg, Germany
15: Quantitative Imaging and Nanobiophysics Group, MRC Laboratory for Molecular Cell Biology, University College London, UK
16: Centre for Bacterial Cell Biology, Institute for Cell and Molecular Biosciences, Newcastle University, UK

ABSTRACT

With the widespread uptake of 2D and 3D single molecule localization microscopy, a large set of different data analysis packages have been developed to generate super-resolution images. To guide researchers on the optimal analytical software for their experiments, we have designed, in a large community effort, a competition to extensively characterise and rank these options. We generated realistic simulated datasets for popular imaging modalities – 2D, astigmatic 3D, biplane 3D, and double helix 3D – and evaluated 36 participant packages against these data. This provides the first broad assessment of 3D single molecule localization microscopy software, provides a holistic view of how the latest 2D and 3D single molecule localization software perform in realistic conditions, and ultimately provides insight into the current limits of the field.

INTRODUCTION

Image processing software is central to single molecule localization microscopy (SMLM), which delivers an order of magnitude resolution improvement on diffraction limited conventional fluorescence microscopy, from 250 nm to approximately 20 nm resolution, by temporal separation of fluorophores within a sample[1–3]. Efficient and automated image processing is essential to extract the super-resolved positions of individual molecules from thousands of raw microscope images, containing millions of blinking fluorescent spots.

Improvements in SMLM image processing algorithms have been crucial in maximizing spatial resolution and in reducing the imaging time of SMLM for compatibly with live cell imaging[4–6]. If SMLM is to achieve a resolving power approaching that of electron microscopy, the analysis software employed needs to be robust, accurate, and performing at current algorithmic limits. This can only be achieved through rigorous quantification of SMLM software performance.

The first localization microscopy software challenge was carried out in 2013, to enable robust benchmarking of 2D localization microscopy software packages[7]. But biology is not just a 2D problem, and a key focus of localization microscopy is the imaging of 3D imaging of nanoscale cellular processes[8,9]. 3D localization microscopy is a more difficult image processing problem than 2D SMLM. In addition to finding the center of diffraction limited spots to super-resolve lateral position, 3D SMLM algorithms must also extract axial information from the image, usually by measuring small changes in the shape of a fluorophore's PSF[10].

There are roughly three common approaches for 3D SMLM. First, point spread function engineering, where the axial asymmetry of the microscope point spread function (PSF) is increased by introducing intentional aberrations in the system, ranging from simple astigmatism[10] to more complex PSF manipulation such as the double helix PSF method[11]. Second, biplane or multiplane imaging, where axial position is measured based on simultaneous measurement of PSF shape at two or more focal planes[12]. Third, dual objective based interferometry, where Z-position is calculated from single photon interference between opposing objectives[13]. Multiplane and PSF engineering methods typically obtain axial resolutions on the order of 50 nm[10,11]. Interferometry achieves the best axial resolution, 10-20 nm[13], but is not yet widely adopted.

Despite the widespread use of 3D localization microscopy, and challenging nature of 3D SMLM image processing, the performance of software for 3D single molecule localization microscopy has previously only been assessed for 2 or 3 software packages at a time, and without standard test data or metrics[14–17]. In the absence of common reference datasets and reliable assessment procedure of 3D software performance, it is not possible to objectively assess how different software affects final image quality, or which algorithmic approaches are most successful. Crucially, end-users cannot determine which 3D SMLM software package and imaging modality is optimal for their application.

We therefore ran the first 3D localization microscopy software challenge, to assess the performance of 3D SMLM software. We generated synthetic datasets for three popular 3D SMLM modalities: astigmatic imaging, biplane imaging and double helix point spread function microscopy. We also ran a second 2D localization microscopy software challenge, to reassess the 2D SMLM software state-of-the art on new, tougher, more realistic datasets.

Our simulations incorporate experimentally acquired point spread functions for maximal authenticity, used signal and noise levels based closely on common experimental conditions, and incorporated a realistic 4-state model of fluorophore photophysics[18]. Our synthetic data was designed to mimic two common classes of cellular structure: narrow line-like microtubules (MT) and larger tubes similar to the endoplasmic reticulum (ER) or mitochondria. Our simulations also included conditions with low density (LD) of active fluorophores, used experimentally to obtain maximal resolution, and with high density (HD) of active fluorophores, used experimentally for fast or live cell imaging.

RESULTS

Competition design

We established a large committee from within the SMLM research community, including experimentalists and software developers, to define the scope of the challenge, ensure realism of the datasets and define analysis metrics. We further opened this discussion to the whole community, through an open forum, discussing best practices for the implementation of this contest[19].

Thirty-six software packages have been entered in the competition thus far (Table S1). Excitingly, participation in the competition actually led at least 8 teams to their software to support additional 3D SMLM modalities, showing how competition fosters microscopy software development.

In 2016, we ran a first round of the 3D SMLM competition with explicit submission deadlines, with 30 competitor teams, culmination in a special session at the 6th annual Single Molecule Localization Microscopy Symposium (SMLMS 2016). Since then, the challenge has been opened to continuously accept new entries. We have had 12 new registrations of which 5 have submitted localizations, including a multiple best-in-class performer (SMAP-2018[20], an updated version of previously entered software) demonstrating the utility of the competition as an evolving measure of the state of the field.

Realistic 3D simulations

Testing super-resolution software on experimental data lacks the ground truth information required for rigorous quantification of software performance. Therefore, realistic simulated 3D SMLM datasets are required. After comparison of simulated microscope PSFs with multiple experimental PSFs from SMLM microscopes around the world, we observed that a critical challenge to realistic 3D SMLM simulations was to accurately model the experimental microscope PSF for each 3D modality. Even experimental 2D PSFs showed significant aberrations away from the focal plane (Fig S9).

3D SMLM inherently involves addition of aberrations to the microscope PSF to encode the Z-position of the molecule. For the PSF models included in the competition: 2D, astigmatic (AS), double helix (DH), and biplane (BP), we observed that the PSFs showed complex aberrations not well described by simple analytical models (Fig S9). We thus combined experimental 3D PSFs with simulated ground truth by performing simulations using PSFs directly derived from experimental calibration data (Fig 1, *Methods*). The experimental PSFs used to generate the simulated data are available online (*Methods*). As the goal of this study was to compare software obtained on typical SMLM microscopes, we deliberately chose PSFs representative of common implementations of each 3D modality. However, additional PSF engineering should improve results of any specific modality, for example adaptive-optics corrected astigmatism[21], or reduced Z-range, higher SNR DH-PSF designs[22].

For the 3D competition, we simulated synthetic 25 nm diameter microtubules (Fig 1). For the 2D competition, in addition to synthetic microtubules (MT), we simulated larger diameter 150 nm cylinders, designed to approximate larger cellular structures such as mitochondria and the endoplasmic reticulum (ER) (Fig 1). We incorporated a 4-state model of fluorophore photophysics, including a transient dark state (dye "blinking") and a bleaching pathway (Fig S1C).

As performance at different density of active emitters is a key challenge for SMLM software, we generated 3D competition datasets at both sparse emitter density (0.2 mol. [molecule] μm^{-2}) and high emitter density (2 mol. μm^{-2}). We additionally generated a very high density dataset (5 mol. μm^{-2}) for the 2D competition.

We generated data at three different signal-to-noise ratio (SNR) levels, based on real signal to noise levels encountered under common SMLM experimental scenarios: fixed cells antibody labelled with organic dye[10], fluorescent protein labelling[1], and live cell affinity dye labelling[23,24].

Together, these simulations closely resemble experimental 3D and 2D data under a range of challenging conditions of SNR, spot density, axial thickness and test structure summarized in Table S2.

139 In addition, we provide also a z-stack of extremely bright beads for software calibration. The
140 competition datasets are available online (*Methods*).

Quantitative performance metrics for comparison of 3D software

We assessed software performance by 24 quality metrics (*Supplementary Note 2*), in four categories:
1) single molecule localization error, 2) ability to successfully detect molecules, 3) image-based
resolution metrics and 4) image-based signal to noise ratio *(Methods).* We also recorded software run
time. The complete set of summary statistics, axially resolved performance and super-resolved images
is available for each competition software on the competition website. We generated an online
leaderboard[25], allowing easy ranking of each software by each metric. Software results can be
accessed interactively through a visualization interface that allows side-by-side comparison of results
for multiple software packages (Fig S10).

In order to rank overall software performance, we performed a principal component analysis of core
metrics to identify key variables (Fig S14). A correlation matrix identified four major blocks of metrics
showing strong codependency (Fig S14B) corresponding closely to the manually identified categories
above. We chose to focus further analysis primarily on the metrics directly derived from single
molecule localizations, rather than image derived metrics, which we reasoned would be sensitive to
additional factors such as image rendering method. We thus chose representative metrics from the
first two blocks:

1. Single molecule localization error. The foremost consideration for localization software is how
accurately it finds the position of labelled molecules. This was quantified as the root mean squared
localization error (RMSE) between the measured molecule position and the ground truth, in both the
lateral (XY) and axial (Z) dimensions.

2*. Ability to successfully detect fluorescent molecules.* In addition to localization precision, SMLM
image resolution also depends critically on number of localized molecules[26], so it is crucial for SMLM
software to accurately detect a large fraction of molecules in a dataset, and minimize false
localizations. For every frame, we identified the localizations that are close enough to a ground-truth
position as true-positives (TP), the spurious localizations as false-positives (FP) and the undetected
molecules as false-negatives (FN). We then computed the *Jaccard index* (JAC, %), which measures the
fraction of correctly detected molecules in a dataset:

$$JAC = 100 \frac{TP}{TP + FP + FN}$$

The average JAC, lateral RMSE and axial RMSE measured the performance of a software. A very good
RMSE should always read in context of the Jaccard index to check if good RMSE is not obtained only
for the brightest molecules.

For ranking purpose, we developed a single summary statistic for overall evaluation of software
performance, which we term the *efficiency (E)*, encapsulating both the software's ability to find
molecules, measured by the Jaccard index, and the software's ability to precisely localize molecules.

$$E = 100 - \sqrt{(100 - JAC)^2 + \alpha^2 \cdot RMSE^2}$$

The trade-off between these two metrics is controlled by a parameter α. In a retrospective analysis,
we chose $\alpha = 1$ nm^{-1} for the lateral efficiency E_{lat}, $\alpha = 0.5$ nm^{-1} for the axial efficiency E_{ax}, based on the
linear regression slope between the localization errors and Jaccard index (Fig 14A). Using this
definition, an average software performance has an efficiency in the range 25-75, ground-truth has
the maximum efficiency of 100. Overall 3D efficiency was calculated as the average of lateral and axial
efficiencies.

Performance of 3D software

Complete rankings for each imaging modality and spot density are presented (Fig 2, S13), together with summary information on all competition software (Table S1, *Supplementary Note 1*). As these data are continuously updated on the competition website, this resource provides microscopists with an easy quick reference for the current state of the art, including current best-in-class performers for each category.

After assembling an overall summary of best performers for each competition category, we investigated the performance of software within each imaging modality.

Astigmatic localization microscopy

Astigmatic localization microscopy is probably the most popular imaging 3D SMLM modality, reflected by the highest number of software submissions in the 3D competition (Fig 2). For astigmatism, we observed a large spread of software performance, even for the most straightforward high SNR, low spot density (LD) conditions (Fig 3A-B, Table S5). The best-in-class software (SMAP-2018) has significantly better localization error and Jaccard index performance than average (lateral RMSE 26 nm best vs 38 nm average, axial RMSE 29 nm best vs 66 nm average, Jaccard index 85 % best vs 74 % average). Clearly, the quality of the image reconstruction depends strongly on choice of 3D software.

To investigate the reasons for software variation, we inspected plots of software performance as a function of axial position in the low density, high SNR dataset for best-in-class and representative middle-range software (Fig S6A). We observed that the key cause of the spread in software performance is variation in software performance away from the focal plane. Near the focal plane, most software packages perform well. However, the axial and lateral RMSE away from the plane of focus is significantly higher for the best in class software, and the Jaccard index is also slightly improved (Fig 6A). This is also visibly apparent in the super-resolved images (Fig 4, top panel). We observed that best-in-class software had a Z-range (the FWHM range of axially resolved software recall, *Methods*) of 1170 nm, greater than two-thirds of the simulated range. Outside this range, the recall and Jaccard index dropped sharply, probably due the large increase in PSF size and decrease in effective SNR at significant defocus (Fig S9).

When we examined results for the low SNR, low density dataset (Fig 2B, 3B), we found an expected 2-fold degradation in best-in-class RMSE (lateral RMSE 39 nm, axial RMSE 60 nm), due to the decrease in image SNR. However, the best-in-class software (SMolPhot) Jaccard index was effectively constant between the low and high SNR datasets (86 % vs 85 %), although the Z-range did drop at lower SNR (930 nm vs 1120 nm). The best astigmatism software packages were thus remarkably good at finding spots at low SNR, even away from the plane of focus.

We analyzed how close software performance was to theoretical limits by calculating the Cramer-Rao Lower Bound (CRLB) as a function of axial position for each dataset and comparing it to the best-in-class software results (Fig S7, Fig S8). Close to the focus, best-in-class software was close to CRLB performance, but significant deviations for the CRLB limit occurred > 200 nm. This could be due to the difficulty in actually detecting the spots away from focus.

When we examined astigmatic software performance for the challenging high spot density datasets (Fig 2B, 3), performance was reduced. For the high SNR high spot density dataset (best software, SMolPhot), localization error increased and Jaccard index decreased significantly compared to the low density condition (lateral RMSE best HD 51 nm vs best LD 27 nm, axial RMSE best HD 66 nm vs best LD 29 nm, Jaccard index best HD 66 % vs best LD 85 %). Inspection of the super-resolved images (Fig 4) nevertheless shows acceptable results for the HD dataset, particularly in the lateral dimension. In many circumstances, the performance reduction at 10x higher spot density should be acceptable for 10x faster, potentially live-cell-compatible, imaging speed. We also observed a large spread of

software performance for the high density datasets, probably because a significant fraction of the
software packages were primarily designed for low density conditions.

We observed poor performance for the most challenging low SNR high spot density astigmatism
dataset (Fig 2, 3, S3, best software SMolPhot). Best-in-class localization precision and Jaccard index
decreased significantly (lateral RMSE 76 nm, axial RMSE 101 nm, Jaccard index 58 %). These data
suggest that low SNR high density 3D astigmatic localization microscopy entails a significant reduction
in image resolution.

Double helix point spread function localization microscopy

We next analyzed the performance of the double helix software (Fig S13). For the software in the high
SNR low spot density condition, double helix software showed more uniform performance than
astigmatism. Best-in-class software (SMAP-2018) showed only a limited improvement compared with
average software (Fig 3B, lateral RMSE, 27 nm best vs 37 nm average; axial RMSE 21 nm best vs 34 nm
average; Jaccard index 77 % best vs 73 % average). In general software localization performance was
close to the CRLB (Fig S7, S8). We observed that performance of the software away from the focal
plane is relatively uniform (Fig 4, S6A), and best-in-class Z-range at high SNR was large at 1180 nm (Fig
S6). Double helix imaging may show less software-to-software variation and large Z-range at low spot
density than astigmatic imaging because the PSF shape and intensity are fairly constant as a function
of Z – compared to astigmatic imaging, where spot size, shape and intensity vary greatly as a function
of Z (Fig S9).

Double helix software performance decreased significantly for the low spot density low SNR condition
(best software SMAP-2018), particularly in terms of best-in-class Jaccard index (66 % low SNR vs 77 %
high SNR, Figure 3B, S3, S13A). DH Jaccard index was also significantly worse than astigmatism results
at either high or low SNR (85 % high SNR, 86 % low SNR). This indicates that it was quite hard to
successfully find localizations in the low SNR DH dataset, likely because the large size of the DH PSF
spreads emitted photons over a large area, lowering effective image SNR. DH PSF designs with reduced
Z-range but more compact PSF would likely be less sensitive to this issue[22].

Double helix software performed poorly on the high spot density datasets at high SNR (best software
CSpline), especially in terms of the Jaccard index (Fig 3B, S13A, best lateral RMSE 67 nm, best axial
RMSE 69 nm, best Jaccard index 46 %). The poor performance at high spot density is again probably
because the large DH PSF size increases spot density and decreases SNR (Fig S9). DHPSF performance
at high spot density and low SNR was also not reliable (Fig. 3B, S13A, best software SMAP-2018).

Biplane localization microscopy

Best-in-class biplane software (SMAP-2018), at low spot density and for both high and low SNR,
delivered the best performance in any modality (high SNR: lateral RMSE 12.3 nm, axial RMSE 21.7 nm,
Jaccard 87 %), despite a slightly decreased image SNR for the biplane simulations (*Methods*). We
observed a significant spread in software performance in terms of lateral RMSE and Jaccard index,
with the best-in-class software significantly outperforming the other competitors (Fig S13B, 2D). At
low spot density, best-in-class biplane software (SMAP-2018) showed good performance as a function
of Z, with high Jaccard index over almost the entire Z-range of the simulations, and with a Z-range of
1200 nm at high SNR (Fig S6A, C, Table S5). The axial RMSE was relatively uniform as a function of Z
and close to the CRLB limit (Fig S7). As axial and lateral RMSE are both averaged over the entire Z-
range, the strong biplane results arise from good performance across a large Z-range (Fig S6).

At high spot density and high SNR, best-in-class biplane software (SMAP-2018) showed acceptable
super-resolved performance (Fig 3B, 4, S13B, best lateral RMSE 43 nm, best axial RMSE 49 nm, best
Jaccard index 61 %). Uniquely among the 3D modalities, best-in-class biplane software also gave
acceptable performance at high spot density and low SNR (Fig 3B, 4, S13B, best lateral RMSE 55 nm,
best axial RMSE 72 nm, best Jaccard index 61 %, best software SMAP-2018).

Performance of 2D software

Alongside the 3D challenge, we ran a second edition of the 2D localization microscopy software challenge[7] to assess how the latest 2D software performed on more challenging, more realistic datasets, and to provide an assessment of how the field had progressed since the last challenge. We used the new simulation software, including an experimentally derived PSF and a realistic blinking model, and also simulated a very high spot density condition (5 molecules/ μm^2). We created a more spatially extended test structure, "pseudo-endoplasmic reticulum" (pseudo-ER), composed of 150 nm diameter hollow tubes, to avoid artefacts due to 1D simulated structures[27]. We generated two different imaging conditions with overall similar SNR but different brightness properties; one with low fluorophore brightness and low autofluorescence (the low SNR condition for the 3D challenge, designed to simulate fluorescent protein based SMLM, Fig S4) and one with high fluorophore brightness and high autofluorescence (to simulate affinity-dye-based live cell SMLM, Fig S5). We used lateral RMSE, Jaccard index and overall lateral efficiency to rank the 2D software (Fig 2, S2, Table S1).

For the pseudo-ER dataset, at low density, best-in-class software (ADCG) performed well (Fig. S4, S5), with a Jaccard index of 90 % and lateral RMSE of 31 nm, substantially better than the class average (Jaccard index 72 %, lateral RMSE 36 nm). Low density results for the dimmer fluorophore microtubules dataset were similar to the brighter pseudo-ER dataset (Fig S2, best software SMolPhot). For the very high density 2D dataset, which had 25x higher spot density than the LD dataset, best-in-class software (ADCG) showed excellent performance, with Jaccard index of 75% and lateral RMSE of 45.5 nm (Fig S2). Best-in-class performance (ADCG) on the dimmer fluorophore data at high spot density was also strong (Fig S2, best Jaccard index 70 %, best lateral RMSE 51 nm).

Algorithms

We identified several classes of algorithm participant software (Table S1):

1) *Non-iterative* software tends to regroup the pixels in the local neighborhood of the candidates, like interpolation, center of mass (QuickPALM[28]) or template matching (WTM[29]). These (often older) algorithms are fast but tend to achieve poor performance (Table S1).

2) S*ingle emitter fitting* software is usually built on a multi-step strategy of detection, spot localization, and optional spot rejection. The detection step finds bright spots in noisy images on the pixel grid. The selection of candidates is usually performed by local maximum search after a denoising filter. Others rely on more complex algorithms like the wavelet transform (*e.g.,* WaveTracer[30]). We did not observe software ranking to depend significantly on the choice of optimization scheme, least-square, weighted least-square or maximum-likelihood estimator (Table S1).

3) *Multi-emitter fitting* software groups clusters of overlapping spots, and simultaneously fits multiple model PSFs to the data. Typically, fitted spots are added to the cluster until a stopping condition is met[4,5]. This leads to improved localization performance at high spot density, at the cost of reduced speed. This class of software (*e.g.,* 3D-DAOSTORM[14], CSpline[14], PeakFit, ThunderSTORM[31]) was amongst the top performers in each 2D and 3D competition category (Table S1).

As expected, single- and multiple-emitter fitting methods both performed well on low density data (Table S1); apparently at the densities studied, exclusion of occasionally overlapping spots by single-emitter software is sufficient for strong performance; explicit multi-emitter fitting is not required. For the 2D challenge, multi-emitter fitting showed a clear advantage over single emitter fitting at high density (Table S1). Surprisingly however, well-tuned single-emitter fitting algorithms (SMolPhot, SMAP-2018) outperformed multi-emitter algorithms for the 3D high density conditions.

4) *Compressed sensing algorithms.* One subset of these algorithms utilize deconvolution with sparsity constraints to reconstruct super-resolved images[32–34]. Although deconvolution approaches can give good results, they are limited by the necessary use of a sub-pixel grid; increased localization precision requires smaller grid resolution, which must be balanced against increased computational time.

Recent approaches address this issue by localizing the point sources in a grid-less manner using an
alternating descent conditional gradient scheme under some sparsity constraint (ADCG[35], SMfit,
SOLAR_STORM, TVSTORM[36]). This software class consistently gave the overall best performance for
2D high-density (ADCG[35] 1st, FALCON[34] 2nd, SMfit 3rd).

5) *Other approaches.* Of the alternative algorithmic approaches used (Table S1), the annihilating filter-
based method LEAP[37] gave good performance for biplane imaging (the only condition for which it was
entered).

Post-hoc temporal grouping

Because molecule on-time is stochastically distributed across multiple frames, a common post-
processing approach to improve localization precision is to group molecules detected multiple times
in adjacent frames, and average their position[38]. Temporal grouping was used by the top performers
(including SMolPhot[39], MIATool[40] and SMAP-2018[20]), and is visibly apparent as a more punctate super-
resolved image (Fig 4).

Choice of PSF model

Most software used a variant of Gaussian PSF model. A few participants designed more accurate PSF
models (Table S1). Either diffraction theory was used (MIATool[40], LEAP[37]) or spline fitting of an
analytical function to the experimental PSF was adopted (CSpline[41], SMAP-2018[20]). Although simple
Gaussian model PSFs were sufficient to obtain best-in-class performance for the 2D and astigmatic
modalities (ADCG[35], PeakFit, SMolPhot), top results for the more optically complex biplane and double
helix modalities were exclusively PSF-modelling algorithms (SMAP-2018, CSpline, MIATool, LEAP).

Multi-algorithm packages

Several software packages take a Swiss army knife approach of integrating multiple optional
localization algorithms into one program, to be flexible enough to suit various experimental
conditions[20,31]. SMAP and ThunderSTORM achieved strong across-the-board performance supporting
this rationale.

DISCUSSION

We performed the first broad evaluation of software for 3D single molecule localization microscopy,
to assess the state of the field and to allow non-specialists to determine the optimal software for their
experiments.

In order to provide a realistic assessment of 3D software performance we tested software on
simulations incorporating experimentally acquired microscope point spread functions. Our
experimental-PSF-derived simulation approach is readily adaptable to novel engineered 3D SMLM
PSFs[42] or to the PSF of individual microscopes. For instance, it would be possible to combine our
derived-PSF approach with the SMLM sample simulation tool SuReSim[43] in order to generate ultra-
realistic synthetic data, which could then be personalized to each experimentalists sample and
microscope, to easily determine the blocker factors to maximal resolution, for a given experiment.

The strongest conclusion we draw from the 3D localization microscopy challenge is that choice of
localization software greatly affects the quality of final super-resolution data, even at "easy" high SNR,
low spot density conditions. Biplane performance was particularly dependent on software choice, with
only one software (SMAP-2018[20]) achieving near-Cramer-Rao lower bound performance. Double helix
SMLM showed less sensitivity to choice of software than biplane, with astigmatic SMLM intermediate
between the two. The best software in each modality performed close to the Cramer-Rao lower
bounds over a wide focal range and successfully detected most molecules, even at low signal to noise.
Average software in all three modalities was significantly worse, with the obtained axial resolution
being particularly sensitive to software choice.

The second major conclusion of the 3D challenge is that localization software that explicitly includes
the experimental PSF in the fitting model gives a significant performance increase for 3D SMLM. For
the more optically complex biplane and double helix modalities in particular, the best results were
exclusively from software using PSF modelling approaches (SMAP, CSpline, MIATool). This result also
highlights the need for experimental PSF modelling not only in SMLM software, but also emphasizes
the high degree of experimental realism required of SMLM simulations. The clear performance
advantage of experimental PSF modelling software in the 3D software challenge would have been
entirely unobservable had it been run with a simple analytic PSF.

Of the different algorithm classes, well-tuned single-emitter and multi-emitter fitting algorithms (each
capable of dealing well with occasional molecule overlap) gave good results for low density 3D SMLM.
We also found that several software packages for astigmatic or biplane imaging gave adequate
performance for the challenging case of high molecule densities, as long as the image SNR was high.
Current software packages gave poor performance when molecule density was high and image SNR
was low. These results suggest that, at least with current algorithms, high density 3D SMLM
performance is mediocre at high SNR, and poor at low SNR. Surprisingly, multi-emitter fitting did not
show significant improvement over well-tuned single emitter fitting for 3D high-density; this may
indicate that significant potential for improvement remains in this category.

The second 2D localization microscopy challenge provided the opportunity to reassess the state of the
field. The performance of best-in-class 2D software over a range of conditions, at both high and low
spot density, is excellent. The performance of the best-in-class software at high spot density (ADCG[35])
was only moderately decreased compared with the low spot density results, with nearly identical
molecule detection performance, and a 30 % increase in localization error. Interestingly, the top three
performers in the 2D high density condition were all compressed sensing algorithms (ADCG[35],
FALCON[34], SMfit). In low density 2D conditions, the best single-emitter, multi-emitter and compressed
sensing algorithms all gave comparable, excellent, performance. We speculate that performance in
this category may now be near optimal levels.

Rapid improvements in sCMOS camera technology mean that these cameras are rapidly becoming a
major platform for single molecule localization microscopy[6]. Therefore, a key future goal for SMLM
software assessment should be to include sCMOS-specific localization microscopy software.
Furthermore, there remain two important classes of super-resolution microscopy for which software
performance is crucial, but no broad software assessment has yet been performed: fluorescence-
fluctuation-based super-resolution microscopies (*e.g.*, 3B[44], SOFI[45], SSRF[46]) and structured illumination
microscopy[47].

The results of this competition clearly demonstrate the formidable algorithmic performance of the
best 2D and 3D localization microscopy software. However, a key outstanding challenge that often
hinders adoption of new algorithms is that only a small subset of algorithms are packaged in, or
compatible with fast, well-maintained, user-friendly software packages, which include all stages of the
SMLM data analysis pipeline – analysis, visualization and quantification. One solution would be for the
SMLM software community to collectively adopt both a standard data format and a single software
platform for future software development, such as FIJI/ ImageJ[48]. Any new algorithm released in this
environment could be immediately and widely adopted by users, and easily integrated into existing
packages for SMLM analysis, visualization and quantification.

Both the 3D and 2D localization challenges remain open and continuously updated on the competition
website. This continuously evolving analysis of state of the art super-resolution software performance
provides a valuable resource to super-resolution microscopists, helping to ensure they use software
that gets the best out of hard-won data. It also provides SMLM software developers with a robust
means of benchmarking new algorithms against current state of the art.

ACKNOWLEDGEMENTS

Authors acknowledge the following funding sources: a Newcastle University Research Fellowship and a Wellcome Trust & Royal Society Sir Henry Dale Fellowship grant number 206670/Z/17/Z to SH; an European Research Council (ERC) under the European Union's Horizon 2020 research and innovation programme, Grant Agreement no. 692726 to DS, TAP, MU; UK BBSRC grants BB/M022374/1, BB/P027431/1, BB/R000697/1 grant and MRC grants MC-UU-12018/2, MR/K015826/1 to RH; European Research Council (ERC) grant CoG-724489, CellStructure to JR; National Institutes of Health grant 1R15GM128166-01 to GMH; and NSF SBIR grants 1353638, 1534745 to Double Helix LLC. We thank R. Piestun at University of Colorado for providing DH-PSF phase mask designs to Double Helix LLC. We thank all the localization microscopy challenge participants for their contribution: H. Babcock, F. Hauser, S. Watanabe, N. Boyd, J. Min, K. Jin, H. Rouault, E. Soubies, A. von Diezmann, C. Bayas, W.E. Moerner, J. Min, J.C. Ye, T. Vomhof, J. Reichel, H. Pan, Z. Huang, Y. Wang, R Belmurugan, A.V. Abraham, R. J. Ober, A. Herbert, K. Martens, J. Hohlbein, L. Li, R. Henriques, G. Tamas, J. Sinko, M. Kirchgessner, F. Gruell, Y. Li, J. Ries, H. Iloma, M. Pars, A. Loot, Y. Jung, N. Fakhri, A. Archetti, M. Ovesny, GM Hagen, P Krizek, J. Huang, A. Kechkar, J.B. Sibarita. We thank the SMLMS 2016 organizers (S. Manley and A. Radenovic, EPFL) for hosting a localization microscopy challenge special session. We also thank Double Helix LLC and Molecular Devices LLC for sponsoring the SMLMS 2016 special session. The sponsors had no input or influence on the research.

AUTHOR CONTRIBUTIONS

DS and SH conceived and coordinated the study. DS, SH, TAP, AAr, HB, SC, AW, GMH, RH, TL, TP, JBS designed the study. SH, AAg, RH, JBS collected experimental PSFs. DS, TAP, SH, TL wrote simulation code. BR shared unpublished software. DS generated simulated datasets. AH, JR, JC, RV provided feedback and quality control on simulations and analysis methods. TAP carried out the assessment of software performance. TAP, DS, SH analysed and interpreted the results. DS, HB, RO, BR, GMH, JBS, JR, RH, MU, SH directed research. SH, DS, TAP wrote the manuscript with feedback from all authors.

REFERENCES

1. Betzig, E. *et al.* Imaging Intracellular Fluorescent Proteins at Nanometer Resolution. *Science* **313,** 1642–1645 (2006).
2. Hess, S. T., Girirajan, T. P. K. & Mason, M. D. Ultra-High Resolution Imaging by Fluorescence Photoactivation Localization Microscopy. *Biophys. J.* **91,** 4258–4272 (2006).
3. Rust, M. J., Bates, M. & Zhuang, X. Sub-diffraction-limit imaging by stochastic optical reconstruction microscopy (STORM). *Nat Methods* **3,** 793–795 (2006).
4. Holden, S. J., Uphoff, S. & Kapanidis, A. N. DAOSTORM: an algorithm for high- density super-resolution microscopy. *Nat Meth* **8,** 279–280 (2011).
5. Huang, F., Schwartz, S. L., Byars, J. M. & Lidke, K. A. Simultaneous multiple-emitter fitting for single molecule super-resolution imaging. *Biomed. Opt. Express* **2,** 1377–1393 (2011).
6. Huang, F. *et al.* Video-rate nanoscopy using sCMOS camera-specific single-molecule localization algorithms. *Nat. Methods* **10,** 653–658 (2013).
7. Sage, D. *et al.* Quantitative evaluation of software packages for single-molecule localization microscopy. *Nat. Methods* **12,** 717–724 (2015).
8. Huang, B., Jones, S. A., Brandenburg, B. & Zhuang, X. Whole-cell 3D STORM reveals interactions between cellular structures with nanometer-scale resolution. *Nat Meth* **5,** 1047–1052 (2008).
9. Shtengel, G. *et al.* Interferometric fluorescent super-resolution microscopy resolves 3D cellular ultrastructure. *Proc. Natl. Acad. Sci.* **106,** 3125–3130 (2009).
10. Huang, B., Wang, W., Bates, M. & Zhuang, X. Three-Dimensional Super-Resolution Imaging by Stochastic Optical Reconstruction Microscopy. *Science* **319,** 810–813 (2008).
11. Pavani, S. R. P. *et al.* Three-dimensional, single-molecule fluorescence imaging beyond the diffraction limit by using a double-helix point spread function. *Proc. Natl. Acad. Sci.* **106,** 2995–2999 (2009).

12. Juette, M. F. *et al.* Three-dimensional sub–100 nm resolution fluorescence microscopy of thick
samples. *Nat. Methods* **5,** 527–529 (2008).
13. Shtengel, G. *et al.* Interferometric fluorescent super-resolution microscopy resolves 3D
cellular ultrastructure. *Proc. Natl. Acad. Sci.* **106,** 3125–3130 (2009).
14. Babcock, H., Sigal, Y. M. & Zhuang, X. A high-density 3D localization algorithm for stochastic
optical reconstruction microscopy. *Opt. Nanoscopy* **1,** 1–10 (2012).
15. Ovesný, M., Křížek, P., Švindrych, Z. & Hagen, G. M. High density 3D localization microscopy
using sparse support recovery. *Opt. Express* **22,** 31263–31276 (2014).
16. Min, J. *et al.* 3D high-density localization microscopy using hybrid astigmatic/ biplane imaging
and sparse image reconstruction. *Biomed. Opt. Express* **5,** 3935–3948 (2014).
17. Zhang, S., Chen, D. & Niu, H. 3D localization of high particle density images using sparse
recovery. *Appl. Opt.* **54,** 7859–7864 (2015).
18. Annibale, P., Vanni, S., Scarselli, M., Rothlisberger, U. & Radenovic, A. Quantitative Photo
Activated Localization Microscopy: Unraveling the Effects of Photoblinking. *PLOS ONE* **6,** e22678
(2011).
19. Collaboration through competition. *Nat. Methods* **11,** 695 (2014).
20. Li, Y. *et al.* Real-time 3D single-molecule localization using experimental point spread
functions. *Nat. Methods* (2018). doi:10.1038/nmeth.4661
21. Izeddin, I. *et al.* PSF shaping using adaptive optics for three-dimensional single-molecule
super-resolution imaging and tracking. *Opt. Express* **20,** 4957–4967 (2012).
22. Grover, G., DeLuca, K., Quirin, S., DeLuca, J. & Piestun, R. Super-resolution photon-efficient
imaging by nanometric double-helix point spread function localization of emitters (SPINDLE). *Opt.*
Express **20,** 26681–26695 (2012).
23. Carlini, L. & Manley, S. Live Intracellular Super-Resolution Imaging Using Site-Specific Stains.
ACS Chem. Biol. **8,** 2643–2648 (2013).
24. Shim, S.-H. *et al.* Super-resolution fluorescence imaging of organelles in live cells with
photoswitchable membrane probes. *Proc. Natl. Acad. Sci.* **109,** 13978–13983 (2012).
25. Single-Molecule Localization Microscopy • Software Benchmarking. Available at:
http://bigwww.epfl.ch/smlm/challenge2016/index.html?p=results. (Accessed: 15th June 2018)
26. Baddeley, D. & Bewersdorf, J. Biological Insight from Super-Resolution Microscopy: What We
Can Learn from Localization-Based Images. *Annu. Rev. Biochem.* **87,** 965–989 (2018).
27. Fox-Roberts, P. *et al.* Local dimensionality determines imaging speed in localization
microscopy. *Nat. Commun.* **8,** 13558 (2017).
28. Henriques, R. *et al.* QuickPALM: 3D real-time photoactivation nanoscopy image processing in
ImageJ. *Nat Meth* **7,** 339–340 (2010).
29. Takeshima, T., Takahashi, T., Yamashita, J., Okada, Y. & Watanabe, S. A multi-emitter fitting
algorithm for potential live cell super-resolution imaging over a wide range of molecular densities.
J. Microsc. **0,**
30. Kechkar, A., Nair, D., Heilemann, M., Choquet, D. & Sibarita, J.-B. Real-Time Analysis and
Visualization for Single-Molecule Based Super-Resolution Microscopy. *PLOS ONE* **8,** e62918 (2013).
31. Ovesný, M., Křížek, P., Borkovec, J., Švindrych, Z. & Hagen, G. M. ThunderSTORM: a
comprehensive ImageJ plug-in for PALM and STORM data analysis and super-resolution imaging.
Bioinformatics **30,** 2389–2390 (2014).
32. Soubies, E., Blanc-Féraud, L. & Aubert, G. A Continuous Exact ℓ_0 Penalty (CEL0) for Least
Squares Regularized Problem. *SIAM J. Imaging Sci.* **8,** 1607–1639 (2015).
33. Babcock, H. P., Moffitt, J. R., Cao, Y. & Zhuang, X. Fast compressed sensing analysis for super-
resolution imaging using L1-homotopy. *Opt. Express* **21,** 28583–28596 (2013).
34. Min, J. *et al.* FALCON: fast and unbiased reconstruction of high-density super-resolution
microscopy data. *Sci. Rep.* **4,** 4577 (2014).
35. Boyd, N., Schiebinger, G. & Recht, B. The Alternating Descent Conditional Gradient Method
for Sparse Inverse Problems. *SIAM J. Optim.* **27,** 616–639 (2017).

36.   Huang, J., Sun, M. & Chi, Y. Super-resolution image reconstruction for high-density 3D single-
molecule microscopy. in *2016 IEEE 13th International Symposium on Biomedical Imaging (ISBI)*
241–244 (2016). doi:10.1109/ISBI.2016.7493254
37.   Pan, H., Simeoni, M., Hurley, P., Blu, T. & Vetterli, M. LEAP: Looking beyond pixels with
continuous-space EstimAtion of Point sources. *Astron. Astrophys.* **608,** A136 (2017).
38.   Durisic, N., Laparra-Cuervo, L., Sandoval-Álvarez, Á., Borbely, J. S. & Lakadamyali, M. Single-
molecule evaluation of fluorescent protein photoactivation efficiency using an in vivo
nanotemplate. *Nat. Methods* **11,** 156–162 (2014).
39.   Martti Pars, A. L. SMolPhot Software. Available at: https://bitbucket.org/ardiloot/smolphot-
software/wiki/Home.
40.   Chao, J., Ward, E. S. & Ober, R. J. A software framework for the analysis of complex microscopy
image data. *IEEE Trans. Inf. Technol. Biomed. Publ. IEEE Eng. Med. Biol. Soc.* **14,** 1075–1087 (2010).
41.   Babcock, H. P. & Zhuang, X. Analyzing Single Molecule Localization Microscopy Data Using
Cubic Splines. *Sci. Rep.* **7,** 552 (2017).
42.   Shechtman, Y., Weiss, L. E., Backer, A. S., Sahl, S. J. & Moerner, W. E. Precise Three-
Dimensional Scan-Free Multiple-Particle Tracking over Large Axial Ranges with Tetrapod Point
Spread Functions. *Nano Lett.* **15,** 4194–4199 (2015).
43.   Venkataramani, V., Herrmannsdörfer, F., Heilemann, M. & Kuner, T. SuReSim: simulating
localization microscopy experiments from ground truth models. *Nat. Methods* **13,** 319–321 (2016).
44.   Cox, S. *et al.* Bayesian localization microscopy reveals nanoscale podosome dynamics. *Nat.*
Methods **9,** 195–200 (2012).
45.   Dertinger, T., Colyer, R., Iyer, G., Weiss, S. & Enderlein, J. Fast, background-free, 3D super-
resolution optical fluctuation imaging (SOFI). *Proc. Natl. Acad. Sci.* **106,** 22287–22292 (2009).
46.   Gustafsson, N. *et al.* Fast live-cell conventional fluorophore nanoscopy with ImageJ through
super-resolution radial fluctuations. *Nat. Commun.* **7,** (2016).
47.   Gustafsson, M. G. L. Surpassing the lateral resolution limit by a factor of two using structured
illumination microscopy. SHORT COMMUNICATION. *J. Microsc.* **198,** 82–87 (2000).
48.   Schindelin, J. *et al.* Fiji: an open-source platform for biological-image analysis. *Nat. Methods*
9, 676–682 (2012).
49.   Hanser B. M., Gustafsson M. G. L., Agard D. A. & Sedat J. W. Phase-retrieved pupil functions in
wide-field fluorescence microscopy. *J. Microsc.* **216,** 32–48 (2004).
50.   Hirsch, M., Wareham, R. J., Martin-Fernandez, M. L., Hobson, M. P. & Rolfe, D. J. A Stochastic
Model for Electron Multiplication Charge-Coupled Devices – From Theory to Practice. *PLOS ONE* **8,**
e53671 (2013).
51.   Basden, A. G., Haniff, C. A. & Mackay, C. D. Photon counting strategies with low-light-level
CCDs. *Mon. Not. R. Astron. Soc.* **345,** 985–991 (2003).
52.   Carlini, L., Holden, S. J., Douglass, K. M. & Manley, S. Correction of a Depth-Dependent Lateral
Distortion in 3D Super-Resolution Imaging. *PLoS ONE* **10,** e0142949 (2015).
552

Figure 1*: Summary of SMLM challenge simulations*. **A**. 3D rendering of microtubules and endoplasmic
reticulum samples in a 6.4 μm x 6.4 μm x 1.5 μm volume. **B**. *Key simulation steps.* The structure is
constructed from 3D tubes continuously defined by three B-spline functions in the volume of interest.
Membranes of the tubes are densely populated with possible positions. Fluorophores follow a 4-state
photophysics model. Activations of a given frame are convolved with the experimental PSF and shot
& camera noise is added. **C.** Summary of all 16 challenge datasets, calibration data and experimental
PSFs. Each dataset is characterized by its structure (endoplasmic reticulum (ER) or microtubules (MT)),
by it modality (2D, AS, DH, BP), its density (LD or HD) and by its SNR determined by the level of noise
N1, N2, and N3. Left column: orthogonal projections of the experimentally-derived PSF. Eight
categories were proposed for the challenge containing two datasets each, 2D-LD and 2D-HD, grey; AS-
LD and AS-HD, red, DH-LD and DH-HD, green; BP-LD and BP-HD, blue.

Figure 2: *Leaderboards for each competition category.* Ranking is based on the efficiency of software based on fraction of successfully detected molecules (Jaccard index) and precision of localization (RMSE, root mean square error, lateral & axial). The contribution of the high SNR dataset is plotted in orange and the contribution of low SNR dataset to the efficiency is plotted in blue.

573
Figure 3: *Comparison of 3D software performance*. Gold stars indicate top performers for each
dataset. Dashed lines in top, middle panels indicate overall efficiency (higher is better). ***A.***
Performance of all astigmatic SMLM software (for other modality results see Supporting Material). ***B.***

Average (colored marker with *s.d.* error bars) and best-in-class (colored marker with gold star)
software performance for all competition modalities. *AS, astigmatism; DH, double helix; BP, biplane.*

Figure 4*: Super-resolved images of 3D competition datasets for best-in-class (top) and representative*
average (bottom) software in each modality, for high SNR datasets. Box indicates zoomed region (left)
or region of line profile (middle). Red, ground truth; green, software results. GT, ground truth; AS,
astigmatism; DH, double helix; BP, biplane. Panel label key*: Software_name Ranking° (Efficiency).*

METHODS

1. CHALLENGE ORGANIZATION
We first ran the 3D SMLM software challenge as a time limited competition, with a results session
hosted as a special session of the 6[th] Annual Single Molecule Localization Microscopy Symposium in
August 2016. The competition has now been converted to a permanent software challenge accepting
new submissions. Special mention to the software SMAP and 3D-WTM that participate to our eight
categories (*density* x *modality*). The current list of participants is at:

http://bigwww.epfl.ch/smlm/challenge2016/index.html?p=participants

All datasets, methods, participations, and results of the challenge 2016 made available at
http://bigwww.epfl.ch/smlm/challenge2016/. Software for simulation and analysis is hosted on the
competition GitHub repository: https://github.com/SMLM-Challenge/Challenge2016/

2. LOCALIZATION MICROSCOPY SIMULATIONS

21. Structure
The synthetic datasets were designed to be similar to images derived from cellular structures in real
experimental conditions. We defined mathematical models for cellular structures that imitate
cytoskeletal filaments such as microtubules and larger tubular structures such as the endoplasmic
reticulum or mitochondria (Fig S1A). These structures have a tubular shape in the 3D space. Pseudo-
microtubules are defined with their central axis elongating in a 3D space having an average outer
diameter of 25 nm with an inner, hollow tube of 15 nm diameter. Pseudo-endoplasmic reticulum is
defined as having a diameter of approximately 150nm.

The underlying sample structure is formalized in a continuous space which allows rendering of digital
images at any scale, from very high resolution (up to 1 nm/pixel) to low resolution (camera resolution:
100 nm/ pixel). The continuous-domain 3D curve is represented by means of a polynomial spline. The
sample is imaged in a $6.4 \times 6.4 \ \mu m^2$ field of view, and the center lines of the microtubules have limited
variation along the z (vertical) axis, *i.e.*, less than 1.5 μm. The fluorescent markers are uniform
randomly distributed over the structure according to the required density. The photon emission rate
of each fluorophore is controlled by a photo-activation model (see below).

The exact locations of all fluorophores are stored at high precision floating-point numbers expressed
in nanometers. This ground-truth file is useful for conducting objective evaluations without human
bias.

2.2. Photophysics activation model
Given a list of source locations from the structure simulator, fluorophore blinking was simulated by a
4-states Markov chain model. The states are ON, OFF, BLEACH, DARK and the transition are Poisson
distributed (Fig S1C), except for the OFF to ON transitions which follow a uniform random distribution,
to reflect that in typical experimental conditions, constant imaging density is maintained by tuning the
photoactivation rate during the experiment. All switching is calculated at sub-frame resolution and
then total fluorophore on-time was integrated over each frame.

Due to two decay paths, the actual mean lifetime of the state ON is

$$\overline{T}_{LIFETIME} = \frac{1}{\frac{1}{T_{ON}} + \frac{1}{T_{BLEACH}}}$$

Switching rates were chosen to approximate photoactivatable fluorescent proteins T_{ON} = 3 frame, T_{DARK}
= 2.5 frames, and T_{BLEACH}=1.5 frames.

Fractional fluorophore ON-times per frame (between 0 and 1) were then multiplied by the mean flux
of photon emission. The flux of photons expressed in photons/seconds was given by the relation

$$\mathbf{F} = \frac{\phi \cdot P \cdot \sigma}{e}$$

Φ is the quantum yield of the dye, P is power of the laser in W/cm^2, e = h c / λ is the energy of one
photon, σ = 1000 ln(10) ε / N$_A$ is the absorption cross section in cm^2 and ε is the molar extinction
coefficient (EC) or absorptivity in cm^2/mol which is a characteristic of a given fluorophore. The laser
power was Gaussian distributed over the field of view. At the end of this process a list of XY positions,
on-frames and (noise-free) intensities for all activated fluorophores was obtained.

2.3. Experimental Point-Spread Function

Model PSFs, stored as high resolution look up tables, were derived from experimentally measured
PSFs. Although the algorithmic approach is distinct, this concept of accurately modelling the
experimental PSF based on calibration data bears relation to the PSF phase retrieval approach
previously employed by Hanser and coworkers[49].

Images of fluorescent beads were recorded for each modality (Table S4). Signal to noise ratio of
recorded PSFs was maximized in all cases by maximizing exposure time and averaging over several
frames to increase dynamic range.

To acquire experimental PSFs, we took 100 nm Tetraspek beads (Invitrogen) adsorbed to #1.5 (170 µm
thick) coverglass, imaged in water. The excitation wavelength was between 640 nm and 647 nm, and
a Cy5 emission filter was used. Exact data acquisition parameters for each modality are listed in Table
S4.

2.4 Simulation PSF construction

For each modality, 3-6 beads were selected within a small (< 32 µm) region, to minimize PSF variation
due to spherical aberration. Images for each selected bead were interpolated in XY to a pixel size of
10 nm. Beads were then coaligned by cross-correlation on the in-focus frame. Coaligned beads were
averaged in XY to minimize pixel quantization artefacts and to increase SNR. Where necessary, Z-stacks
were interpolated to a Z-step size of 10 nm. A central Z-range of 1.5 µm was selected that represents
151 optical planes with a Z-step of 10 nm. The Z-range covers -750 nm to +750 nm. The plane of best
focus was chosen as the simulation 0 nm plane. Each model PSF was normalized such that the total
intensity of the PSF in the in-focus frame within a diameter of 3 FWHM from the PSF center was equal
to 1.

For the DH PSF, the transmission of the combined phase mask system was measured as 96 %, which
was approximated as 100 % brightness relative to the 2D and astigmatic PSFs.

In biplane super-resolution microscopy, emitted fluorescence is split into two simultaneously imaged
channels, with a small (500-1000 nm) defocus introduced between the two channels[12]. As the small
defocus should introduce minimal additional aberration into an optical system, we semi-synthetically
constructed a realistic biplane PSF from the experimental 2D PSF. The two defocused PSFs were
constructed by duplicating the 2D PSF and offsetting it by -250 nm and 250 nm for each Z-plane.

This yielded five high SNR model PSFs with an isotropic voxel size of 10x10x10 nm^3. These normalized
PSFs are provided on the competition website: http://bigwww.epfl.ch/smlm/challenge2016/psf

The ground truth XY=0 was defined as the image centre of mass of the in-focus frame of the model
PSF, and Z=0 was defined as the in-focus frame. Accounts for shifts in the fitted XY centre of the model
PSF by localization software due to systematic offsets and Z-dependent variation of the model PSF
centre of mass are dealt with below (wobble correction).

2.4. Noise model

A constant mean autofluorescent background was added to the noise-free simulated images, and these images were then fed through the noise model representing Poisson distributed fluorescence emission recorded on a high quantum efficiency back-illuminated EMCCD[50,51].

The proposed noise model assumed as main contributions to the stochastic noise:

- σ_S , the shot noise produced by the fluorescence background and signal and the spurious charge. Shot noise can be derived from the second moment of the Poisson distribution
- σ_R, the read noise of EMCCD camera, which is described by second moment of the Gaussian distribution
- σ_{EM}, the electron multiplication noise introduced by the gain process, which is described by the second moment of the Gamma distribution[51].

We assumed as camera parameters the ones specified for the Photometrics Evolve Delta 512 EMCCD camera:

- QE = 0.9, Evolve quantum efficiency at 700 nm absorption wavelength.
- σ_R= 74.4 electrons, manufacturer measured root mean square noise for Evolve 512 camera
- c = 0.002 electrons, manufacturer quoted spurious charge (clock induced charge only, dark counts negligible)
- EM$_{gain}$ = 300
- e$_{adu}$ = 45 electron per analog to digital unit (ADU), analog to digital conversion factor
- G = 0.9*300/45 = 6, total system gain
- BL = 100 ADU

The final simulated photon electrons will thus be given by:

$$n_{ie} \ = \ \mathcal{P}(QE \cdot n_{photIn} + c)$$

$$n_{oe} \ = \ \Gamma\big(n_{ie}, EM_{gain}\big) + \mathcal{G}(0, \sigma_R)$$

which leads to the final pixel counts:

$$ADU_{out} \ = min\left(\frac{n_{oe} - n_{oe} \, mod \, e_{ADU}}{e_{per_{adu}}} + BL, 65535\right)$$

2.5. Depth-dependent lateral distortion: Wobble

As the PSF models are experimentally derived, the 3D estimated localizations exhibit a depth-dependent lateral distortion, here called *wobble*. This optical distortion is due to a combination of a systematic offset (arbitrary definition of PSF center) and optical aberrations[52]. In order to compare estimated and true localizations, we correct this effect during the assessment (Section 3.1).

2.5 Comparison of software results between different modalities.

The intensities of the PSF in each imaging modality were normalized to facilitate comparison of results between different modalities. Software results between 2D, 3D AS and 3D DH modalities are expected to be directly comparable.

For the biplane model PSF, as the emitted fluorescence is split into two channels, the intensity in each of the two simulated biplane channels was additionally reduced by 50 %. We note that the fluorescence background was not reduced by 50 % as intended, leading to artificially high background for the biplane simulation (*i.e.*, the background in each biplane channel is the same as in the single channel of the other modalities). However, due to the low background level in the 3D simulations, the effect on image SNR and thus localization error is small (see Fig S7), less than 5nm near the plane of

712 focus. Therefore, as long as the small drop in image SNR is taken into account, approximate
713 comparisons of the biplane data to the other modalities can still be made.

714 3. SOFTWARE ASSESSMENT

715 3.1 Protocol

Each localization file submitted by the participants was manually checked for erroneous systematic
errors in the definition of the dataset coordinate system, such as offsets, XY axis flips or clear scaling
errors. Datasets were then programmatically standardized into a consistent output format. All
modifications are publicly available. If required, the modifications consisted of columns reordering,
reversing axes, XY axis swap, and shifting the lateral positions by a half camera pixel.

The assessment pipeline includes three main parts: localization processing, the pairing between true
and estimated localization and the metrics calculations. The first one depends on the assessment
settings. There are two switchable properties: photon thresholding and wobble correction. Their
combinations yield four different assessment settings. Up to 64 assessment runs per software were
possible (*i.e.,* 4 modalities, 4 datasets per modality). For any setting, we excluded the fluorophores
within a lateral distance of 450 nm from the border. This value corresponds to the radius of the largest
PSF (*i.e.,* Double Helix). The activations too close from the border are more difficult to localise and
could bias the results.

The pairing between true and estimated localizations was performed frame by frame. The procedure
matches two sets of localizations. We deployed the presorted nearest-neighbor search for its
efficiency. The results are effectively similar to the computationally intensive Hungarian algorithm[7].

732 *Photon thresholding*

A photon threshold was required primarily due to the use of a realistic fluorophore blinking model.
Since a fluorophore could activate/ bleach at any point in a simulated frame, this led to many frames
containing very dim, undetectable localizations, eg. where a molecule had been active for one or more
frames previously, and then bleached during the first 5 % of a frame. These fractional localizations
should also be present but practically undetectable in an experimental dataset.

In order to focus the software analysis on the localizations where the molecule was active for the
majority of a frame, which we decided was most consistent with experimental expectations, we
implemented a photon threshold means where we kept the 75% brightest ground truth fluorophore
activations. Because this was performed *after* the pairing step, observed localizations that were paired
to discarded ground truth activations were also removed from the metric calculations.

743 *Wobble correction*

The centroid of experimental point spread functions shifts laterally by as much as 50 nm, as a function
of axial position[10,52]. This is most often ignored by localization software, and instead corrected post-
hoc by reference to a calibration curve[37]. Since our simulated PSF is experimentally derived, it was
necessary to correct for these artefactual shifts between the observed localizations and ground truth,
as part of the assessment process. This correction was performed using calibration data uploaded by
competitors, similar to the correction typically performed on experimental data[52].

Three scenarios were proposed to the participants: no correction was applied during the assessment;
the correction was based on a file provided by the participant itself or the correction was calculated
by ourselves. The latter nevertheless requires the participant to localize a stack of beads we provided.
Since the true positions of the beads are known, the difference between the estimated and true
positions could be calculated and averaged. It thus yields the values for wobble correction.

In certain specific cases (identified on the competition website), at the request of authors, we did not
apply this correction, for example because the software explicitly considered the whole 3D PSF during

fitting and was thus immune to this lateral shift artefact. For accurate results, application of lateral shift correction is critical for analysis of localization microscopy simulations using experimentally derived PSFs, as can be seen by comparison of typical software results with and without wobble correction (Fig S11).

3.2 Metrics
The metrics are split into two categories: localization based and image based metrics.

The former directly relies on the localizations positions and notably includes the Recall, the Precision, the Jaccard Index, the RMSE (axial and lateral) and the consolidated Z-range. For the calculation of average software performance (Fig 3B) outlier software with an efficiency less than *eff=-30* were excluded from the measurement.

The image based metrics are computed from a rendered image and includes the Signal-to-Noise Ratio (SNR) and the Fourier Ring / Shell Correlation (FRC/FSC). To render the image, we added the contribution of each localized molecule at the corresponding pixels. A contribution takes the form of a 3D additive Gaussian with a Full-Width Half Maximum (FWHM) of 20 nm. A complete list of all computed metrics is shown in the Supplementary Note 2.

We also calculated localization based metric results as a function of axial position. We proceeded by considering a subset of activations lying within an interval of axial positions (*i.e.,* from the true localizations). Then, most of the metrics (*e.g.,* Recall) are locally computed. This yields a curve providing information on the depth performance of each software / modality.

In order to summarize software axial performance, we analyzed how the recall varied as a function of Z. A typical recall versus axial position curve (Fig S9) will drop at positions far from the focal plane, *i.e.,* where software can no longer detect spots to defocus. We first smoothed the curve using a sliding window. Then we computed the software Z-range, defined as the full width half maximal Recall of the smoothed curve (Fig S12). This quantity is visually intuitive and useful for discussion of the recall performance if considered alongside a plot of recall vs axial position. However, because FHWM recall depends on the maximal recall, ranking based on this procedure would promote a software which poorly performed everywhere (*i.e.,* flat curve), whereas a software which performed well in the focal plane but less well outside would obtain a worse FWHM recall. This observation leads us to produce a so-called consolidated Z-range, by multiplying the Z-range value by the maximal Recall, which should provide a robust metric that avoids the previous case scenario.

Principal component analysis. In order to analyse the relationship between analysis metrics we computed the covariance matrix between each metric and the principal component analysis (PCA) on the metrics (Fig S14B). Each metric was standardized before applying the covariance and the PCA. For convenience, we took the additive inverse of the metrics for which lower values are best (i.e., FP, FN, RMSE, FRC, FSC).

Summary statistics and detailed results for each software are available on the competition website (http://bigwww.epfl.ch/smlm/challenge2016/index.html?p=results), which also includes a tool for side-by-side comparison of the results of multiple software packages

3.3 Baseline Localization Software
We developed a minimalist Java tool software that performs localizations of bright emitters on the 4 modalities of the challenge 2016: 2D, Astigmatism, Double-Helix, and Biplane. This SMLM_BaselineLocalization software is only designed to establish the performance baseline for the SMLM challenge. It has intentionally limited lines of code and relies only on few threshold parameters to localize particles. It has basic calibration tool that has to run on a z-stack of beads to find the linear f(x) relation between the axial position Z and the shape of the bead.

- Astigmatism: $Z = f(W_X - W_Y)$, where W_X and W_Y are respectively an estimation of the size in X and Y.
- Double-Helix: $Z = f(\theta)$, where θ is the angle formed the pairing of two close points.
- Biplane: $Z = f(W_{left} - W_{right})$, where W_{left} and W_{right} are respectively an estimation of the size of the spots in left and the right plane.

The Java code is available: https://github.com/SMLM-Challenge/Challenge2016

4 DATA AVAILABILITY

4.1 Data availability statement

Simulated competition datasets are available at http://bigwww.epfl.ch/smlm/challenge2016/, together with the parameters used to generate the data. The ground truth list of simulated molecule positions for each competition dataset remains secret in order to allow the software challenge to remain continuously open to new submissions. However, ground truth data is available for the simulated training datasets.

4.2 Code availability statement

All software is available at https://github.com/SMLM-Challenge/Challenge2016