1 **Next Generation-Targeted Amplicon Sequencing (NG-TAS): An optimised**

2 **protocol and computational pipeline for cost-effective profiling of circulating**

3 **tumour DNA.**

5 **Authors**

Meiling Gao[1,6]   mg390@cam.ac.uk

Maurizio Callari[1,6]   Maurizio.Callari@cruk.cam.ac.uk

Emma Beddowes[1,2]   Emma.Beddowes@cruk.cam.ac.uk

Stephen-John Sammut[1]   Stephen-John.Sammut@cruk.cam.ac.uk

Marta Grzelak[1]   Marta.Grzelak@cruk.cam.ac.uk

Heather Biggs[2]   heather.biggs@addenbrookes.nhs.uk

Linda Jones[2]   linda.jones@addenbrookes.nhs.uk

Abdelhamid Boumertit[2]   ab2421@medschl.cam.ac.uk

Sabine C Linn[3]   s.linn@nki.nl

Javier Cortes[4,5]   jacortes@vhio.net

Mafalda Oliveira[5]   moliveira@vhio.net

Richard Baird[2]   rdb39@medschl.cam.ac.uk

Suet-Feung Chin[1]*   Suet-Feung.Chin@cruk.cam.ac.uk

Carlos Caldas[1,2]*   Carlos.Caldas@cruk.cam.ac.uk

Institutional Addresses

[1]Department of Oncology and Cancer Research UK Cambridge Institute, Li Ka Shing

Centre, University of Cambridge, Cambridge CB2 0RE, UK

[2]Breast Cancer Programme, Cancer Research UK Cambridge Cancer Centre,

Cambridge University Hospitals NHS Foundation Trust, Cambridge CB2 2QQ, UK

[3]Netherland Cancer Institute, 1006 BE, Amsterdam, Netherland

[4]Ramon y Cajal University Hospital, 28034 Madrid, Spain

[5]Vall d'Hebron Institute of Oncology, 08035 Barcelona, Spain

[6]Co-first author

*Co-corresponding authors

Abstract

Circulating tumour DNA (ctDNA) detection and monitoring has enormous potential clinical utility in oncology. We describe here a fast, flexible and cost-effective method to profile multiple genes simultaneously in low input cell-free DNA (cfDNA): Next Generation-Targeted Amplicon Sequencing (NG-TAS). We designed a panel of 377 amplicons spanning 20 cancer genes and tested the NG-TAS pipeline using cell-free DNA from two hapmap lymphoblastoid cell lines. NG-TAS consistently detected mutations in cfDNA when mutation allele fraction was >1%. We applied NG-TAS to a clinical cohort of metastatic breast cancer patients, demonstrating its potential in monitoring the disease. The computational pipeline is available at: https://github.com/cclab-brca/NGTAS_pipeline.

Keywords

NG-TAS, ctDNA, liquid biopsy, mutation, multiplexing, deep sequencing, computational pipeline, cancer, heterogeneous

1 BACKGROUND

Cell free DNA (cfDNA) in plasma was first analysed in cancer patients nearly 50
years ago [1]. A fraction of cfDNA was shown to carry mutations found in the matched
tumour, and designated circulating tumour DNA (ctDNA) [1–3]. The utility of ctDNA as
a non-invasive diagnostic, prognostic or predictive biomarker in human cancer is now
well documented [4–8].

The amount of cfDNA in plasma is usually low and the ctDNA fraction is typically
only 1-30%, hence low mutant allele frequencies have to be detected. Human cancers
are genetically heterogeneous and mutations occur infrequently at recurrent hotspots.
Therefore, in most clinical scenarios (e.g. early diagnosis or monitoring of tumour
evolution), high sensitivity and the simultaneous investigation of multiple gene targets
are desirable features of any ctDNA detection and quantitation method.

There are a range of methods for detecting mutations in ctDNA, with the target
varying from a single nucleotide variant (SNV) to the whole genome. A widely used
method to detect mutations in ctDNA is digital polymerase chain reaction (dPCR)
performed in microfluidic devices or water-in-oil droplet emulsions [9,10]. Whilst dPCR
is able to detect rare mutations with extremely high sensitivity, it is restricted by the
number of targets that can be examined in a single reaction [11] .

Several sequencing based approaches have been developed to incorporate
multiple genomic loci, enabling *de novo* mutation identification in ctDNA. Previously,
we described Targeted Amplicon Sequencing (TAm-Seq), which utilised 48 primer
pairs to identify mutations in hotspots or selected regions of 6 key driver genes [12].
While TAm-Seq is useful, it is limited to a small number of targets. Capture based
sequencing methods can cover a larger number of genes (or the whole exome), but
are costly at the sequencing coverage (>300) required to detect allele frequencies
~1%.

There several ready to use commercial kits for ctDNA sequencing, which can cover
up to hundreds of mutation hotspots and many genes. These include Invision[TM]
(Inivata), Oncomine[TM] cfDNA assay (Thermo Fisher Scientific), Guardant360[TM]
(Guardant Health), and PlasmaSELECT™ (Personal Genome Diagnostics). These
products are expensive and test custom gene panels. Disturbingly, a recent study
comparing the performance of two of these commercial products (Guardant360[TM] and
PlasmaSELECT[TM]) in a cohort of plasma samples from prostate cancer patients,
revealed poor agreement [13].

Recently unique molecular barcodes have been developed to tag each cfDNA
template molecule before PCR amplification in order to reduce the error rate and allow
robust detection of rare mutant alleles in ctDNA [14].

In summary, using current ctDNA profiling methodology, the detection of mutations in a good number of cancer genes with sufficient sensitivity and in a cost-effective way poses significant challenges. Here we describe a new method for the profiling of ctDNA, designated Next Generation-Targeted Amplicon Sequencing (NG-TAS), with several unique features: i) optimised for low input ctDNA; ii) high level of multiplexing, enabling the analyses of multiple gene targets; iii) a bespoke computational pipeline for data analysis; and iv) very competitive costing. NG-TAS is designed to be flexible in terms of choice of gene targets and regions of interest; thus, it can be tailored to various cancer types and clinical contexts.

METHODS

Patient samples and blood processing

Patients were recruited from 3 different centres including Cambridge University Hospital, Netherland Cancer Institute (NKI) and Vall d'Hebron Institute of Oncology (VHIO). Metastatic breast cancer patients with hormone receptor positive tumours were recruited as a part of a clinical trial (patient number = 30, plasma samples number = 366). Blood samples were collected in EDTA tubes and processed within one hour to prevent lymphocyte lysis and fragmentation. Samples were centrifuged at 820g for 10min at room temperature to separate the plasma from the peripheral blood cells. The plasma was further centrifuged at 1400g for 10min to remove remaining cells and cell debris. The plasma was stored at -80°C until DNA extraction. This study was approved by the regulatory and ethic committees at each site and the reference number is NCT02285179 (https://clinicaltrials.gov/ct2/show/NCT02285179). All human samples used were collected after informed consent, and the study was fully compliant with the Helsinki Declaration

DNA extraction from plasma and buffy coat

Plasma DNA was extracted from between 2-4 ml of plasma with the QiaSymphony according to the manufacturer's instruction using Qiagen circulating DNA extraction kit. DNA was isolated from the buffy coat samples using DNeasy Blood & Tissue Kits (Qiagen).

Generation of cfDNA from NA12878 and NA11840

As previously reported [15] two lymphoblastoid cell lines, NA12878 and NA11840 from Human Genome Diversity Project (HGDP)-CEPH collection were obtained from the Coriell Cell Repository. A catalogue of highly accurate whole genome variant calls and homozygous reference calls has been derived for sample NA12878 by integrating independent sequencing data and the results of multiple pipelines (http://www.illumina.com/platinumgenomes). NA11840 cell line was chosen from a set of 17 available CEPH cell lines in our laboratory as it shared the least number of SNPs with NA12878, to generate the maximum number of virtual somatic SNVs.

The cell lines were grown as suspension in RPMI 1640-Glutamax (Invitrogen) supplemented with 10% foetal calf serum, 5% penicillin and streptomycin at 37°C and 5% CO_2. The media that the cell lines were grown in were collected when cells were passaged. The media were centrifuged at 1500rpm for 10 minutes at 4°C to remove cells and cellular debris. The clarified media were stored at -20°C until required. Cell-free DNA was extracted from the thawed media using the Qiagen circulating DNA

extraction kit (Qiagen) according to the manufacturer's instructions and quantified using Qubit High Sensitivity DNA quantification kit (Life Technologies). DNA from both cell lines was diluted to obtain 50 ng/µl stock concentrations. To generate the serial dilutions of one cell line with the other, we mixed by volume to obtain the percentage (volume/volume) as presented in Additional file 1: Table S1 (n=12).

Platinum variant calls for sample NA12878 (the virtual 'tumour') and confident regions (high confidence homozygous reference regions plus platinum calls) [16] were downloaded from http://www.illumina.com/platinumgenomes. Genotype data for sample NA11840 (the virtual 'normal') was obtained from the 1000 Genomes website. Platinum calls were intersected with our NG-TAS panel target regions and variants shared with the NA11840 sample were excluded. Five platinum calls were covered theoretically by our NG-TAS panel; however, one was targeted by one of the amplicons showing no coverage, therefore 4 SNVs were considered as identifiable 'somatic variants'.

NGS library construction

NGS libraries were prepared from 3-5 ng of cfDNA using the ThruPLEX® Plasma-seq kit (Rubicon Genomics, USA) as described in the manufacturer's instructions. NGS library was quantified using qPCR KAPA Library Quantification kit (KAPA Biosystem), while the fragment size and the NGS library yield were measured with 2200 TapeStation instrument (Agilent).

Digital PCR

BioMark system from Fluidigm has been used for dPCR, and the analyses have been performed as previously described [17]. As described in manufacturer's instructions, DNA samples were mixed with 2X TaqMan® Gene Expression Master Mix (Life Technology, 4369016), 20X GE Sample Loading Reagent (Fluidigm, 85000746) and 20X gene-specific assays. The reaction mix were loaded on the qdPCR 37K^{TM} IFC (Fluidigm, 100-6152). For *KRAS* (*G13D*) and *AKT1* (*E17K*) mutant and wild type PrimePCR^{TM} ddPCR^{TM} Mutation Assays were obtained from Bio-Rad (dHsaCP2000013 and dHsaCP2000014, dHsaCP2000032 and dHsaCP2000031 respectively). The *PIK3CA* and *ESR1* probes and primers were previously described [7,18], and the primer and probes used are listed in Additional file 1: Table S2.

NG-TAS protocol

Primer design for NG-TAS

Primers were designed with NCBI Primer-BLAST tool with Tm range of 59-61°C. The universal primer sequences (CS1 and CS2) were added at the 5' end of the designed primers. All primer pairs were tested alone and in multiplexed PCR reactions using 10 ng of TaqMan® Control Human Genomic DNA (Thermo Fisher Scientific) in 10 µl reaction volumes. The coverage and performance of primers were analysed using 2200 TapeStation instrument (Agilent) and Hi-seq 4000. The primers were grouped together as 7-8plex, and primers in each group were chosen to target different genes in order to minimise non-specific amplification and cross-reactivity.

Access Array™ microfluidic system

The 377 pairs of optimised primers were divided into 48 wells, with each well containing 7-8 pairs of primers for multiplexed PCR. Primers were diluted to the final concentration of 1 µM to make 20X primer solution. 4µl of the 20X primer solution from the 48 wells were added to the primer inlets of the Access Array™ IFC (Fluidigm). For the sample inlets, pre-sample master mix consisted of 2X Master Mix (Qiagen, 206143), 5X Q-solution, 20X Access Array™ Loading Reagent (Fluidigm), and DNA sample were added. The loaded IFC then moved to FC1™ Cycler (Fluidigm) for thermal cycles: 95°C for 15min, 30 cycles of 94°C for 30sec, 59°C for 90sec, 72°C for 90sec, and final extension step 60°C for 30min. The reaction products were harvested using Post-PCR IFC controller as described in manufacturer's instructions.

The harvested product was diluted (1:10) with water for further barcoding PCR. Barcoding PCR reaction master mix contains 2X Master Mix (Qiagen), diluted harvested product from Access Array™, and Access Array™ Barcode Library for Illumina® Sequencers single direction for barcoding primers (Fluidigm, 100-4876). The thermal cycle for barcoding is: 95°C for 10min, 15 cycles of 95°C for 15sec, 60°C for 30sec, 72°C for 1min, and final extension step of 72°C for 3min. The PCR reaction was performed using T100™ Thermal Cycler (Bio-Rad).

Quantification and clean-up of barcode Access Array™ harvest

After barcoding PCR, all samples were analysed using 2200 TapeStation (Agilent) to measure the concentration and size of the products (average 260bp). The PCR products were pooled and cleaned with AMPure XP beads (Beckman Coulter, A63880) following the manufacturer's instruction. Briefly, the samples were mixed with the magnetic beads to the ratio of 180:100 in volume. The beads were washed twice with 80% ethanol and dried by incubating at 30°C for 10min. Then the beads were eluted with water and the cleaned PCR product was run on the E-Gel® 2% agarose gel (Thermo Fisher Scientific, G501802) for further size selection and extraction. The band

1 between 200-300bp was cut out and DNA was isolated from the gel using the QIAquick
2 Gel Extraction kit (Qiagen, 28704), and 10-20nM of the eluents was submitted for
3 paired-end Hi-seq 4000 for sequencing.
4

Analysis of NG-TAS data
<u>Quality control, alignment and bam files annotation</u>
For each sequencing lane, quality control of raw data was performed using FastQC
(http://www.bioinformatics.babraham.ac.uk/projects/fastqc/). Up to 384 samples were
multiplexed in a single sequencing lane and demultiplexing was performed using in-
house software.
Alignment, read trimming (at 80bp) and base quality recalibration was performed in a
single step using Novoalign (v 3.08). However, to facilitate a broad use of the pipeline,
a version using BWA-MEM is also available. Alignment and bam metrics were
computed using Picard Tools (v 2.17). To remove potential off-target PCR products,
only reads mapped in proper pair and with insert size > 60bp were retained. After this
filtering, bam files were locally realigned using the Genome Analysis Toolkit (GATK, v
3.6). Reads were then assigned to the amplicon they belonged to using a custom java
script, in order to enable a per amplicon coverage and mutation calling analysis.
Coverage was computed for each amplicon in each sample using a custom java/R
script. One amplicon (SF3B1_D0069_001) showed an extremely high rate of
mismatches and indels in all the analysed samples, therefore we excluded it from
downstream analyses.
23

<u>Mutation calling</u>
Mutation calling was run separately for each amplicon in the panel. The core mutation
calling was performed for each pair of plasma and normal samples (or NA12878 an
NA11849 from the dilution series) using Mutect2 (included in GATK 3.6). The
minPruning parameter was set at 5 to reduce computational time with no significant
impact on the results. Besides the set of mutations passing all internal filters, we
included those failing the following internal filters or a combination of them:
"alt_allele_in_normal", "clustered_events", "homologous_mapping_event",
"multi_event_alt_allele_in_normal". On this set of candidate mutations, we applied the
following filtering criteria: coverage in normal and plasma >100x, alternative allele in
normal <1% and plasma/normal VAF ratio >5. The core mutation calling was repeated
for the three replicates generated for each pair and only mutations called in at least
two replicates were retained. For this set of mutations, we run HaplotypeCaller
(included in GATK 3.6) to compute the average VAF across the three replicates and

filter out mutations with an average VAF<1% and an average plasma/normal ratio <5

(Figure 4A). An extra filter was introduced for FFPE samples, where C>T and G>A

transitions with VAF<15% were filtered out because likely to be consequence of

cytosine deamination caused by fixation.

In calling somatic mutations from set of longitudinal samples from the same patient,

we first repeated the above procedure for all samples. Then, HaplotypeCaller was run

again to estimate in all samples the coverage and VAF of each mutation called in at

least one of them. This was followed by a variant annotation step using Annovar.

Finally, results obtained for all amplicons were merged to generate a single vcf file. A

final filter was applied at group level, that is, keeping only mutations that at least in one

sample had VAF higher than 5% (Additional file 1: Fig. S1).

RESULTS

Optimising targeted deep sequencing in cfDNA by NG-TAS

We designed 377 pairs of primers covering all exons or hotspots of 20 genes

commonly mutated in breast cancer (Table 1). To identify the genes or hotspots of

interest, we primarily looked at the comprehensive study carried out in our lab (Pereira

et al. Nat Comm 2016). Other genes genes (e.g. *ESR1*) were included because

reported as frequently mutated in metastasis [19]. Since the average cfDNA fragment

size is 160-170bp, NG-TAS primers were designed to generate amplicons of 69-157bp

(Additional file 2).

In a preliminary optimization step, individual primer pairs were tested in conventional

single and multiplexed (7-8plex) PCR reactions. The NG-TAS experimental workflow

(Figure 1A), starts with a multiplexed PCR step (7-8 primer pairs) performed using

Access Array™, a microfluidic system from Fluidigm. Each multiplexed reaction

contained primers targeting different genes to minimise the generation of unwanted

PCR products. The multiplexed PCR products were assessed using the Bioanalyser

and 2200 TapeStation instrument (Agilent Genomics; Additional file 1: Fig. S2).

Multiplexed PCR products were then pooled and barcoded with 384 unique barcodes

in a second PCR reaction. Barcoded products were pooled, and size selected to

remove primer dimers before submission for NGS paired-end 150bp sequencing.

Raw sequencing data were aligned and processed as described in Figure 1B and

Methods. Specific filters were applied to exclude reads from primer dimers or other

PCR artefacts. Since the amplicons are partially overlapping, each read was assigned

to its respective amplicon, to enable a per-amplicon analysis for coverage estimation

and mutation calling.

To optimise NG-TAS we used cfDNA isolated from the culture media of the Platinum

Genome Hapmap NA12878 cell line. The size profile of cfDNA isolated from the tissue

culture media was similar to that of plasma cfDNA (Additional file 1: Fig. S3). We tested

a range of input cfDNA amounts with NG-TAS (0.016 ng to 50 ng) in 4 replicates for

each input. For each cfDNA input we tested: i) a pre-amplification step and ii) the use

of the Qiagen Q solution. To assess the data generated the percentage of aligned

sequencing reads was computed (Figure 2A). In the TAM-Seq protocol addition of a

pre-amplification step reduced the probability of nonspecific amplification and biased

coverage [12]. However, using NG-TAS the pre-amplification step reduced the

percentage of aligned reads in all cfDNA input samples tested. Hence, we eliminated

pre-amplification from the NG-TAS protocol. Adding Q solution systematically

increased the percentage of aligned reads, with the largest improvement observed with

0.4 and 2 ng input samples (Figure 2A). Thus, we incorporated the Q solution in all

subsequent NG-TAS experiments.

We then used the optimised NG-TAS protocol in triplicate experiments for each input

NA12878 cfDNA (2 ng, 5 ng, and 10 ng). With 10ng of input cfDNA NG-TAS generated

a median read depth of 3064x, and only 22/377 amplicons (5.8%) had coverage less

than 100x (Figure 2B). In fact, high amplicon coverage was observed irrespective of

amount of input cfDNA (Additional file 1: Fig. S4A and S4B). The coverage heatmap

of individual amplicons showed similar patterns with 10 ng and 5 ng cfDNA input.

Strong consistency was observed within each triplicate (Figure 2C). However, with 2ng

cfDNA input we observed stochastic reduction in coverage for some of the amplicons.

This is probably due to a reduction in template availability, with the number of

amplifiable copies approaching zero for some of the amplicons.

Using these data, the background noise was estimated by computing the average

frequency for non-reference bases in each position, and for 99% of the targeted

genomic positions background noise was ≤0.3% (Figure 2D).

Sensitivity and specificity of mutation detection in control cfDNA

To establish an analysis pipeline and assess the performance of NG-TAS, we

generated a benchmark dilution series, similar to what we have previously described

[15], using cfDNA collected from the tissue culture media from two lymphoblastoid cell

lines from the HapMap/1000 Genome Project, NA12878 (the Platinum Genome

sample) and NA11840, to mimic a tumour-normal (or plasma-normal) pair. The dilution

series mixed cfDNA from NA12878 with an increasing amount of cfDNA from NA11840

(from 0 to 99.8% by volume, n=12, Additional file 1: Table S1). This cfDNA dilution

series was used to investigate the sensitivity in detecting mutations at high and low

allele frequency (50% - 0.1%). The 377-amplicon panel encompassed four

heterozygous single nucleotide polymorphisms (SNPs) present only in NA12878.

These SNPs were used as "somatic" mutations for the purpose of this analysis.

Using NG-TAS the cfDNA dilution series was tested in triplicate, varying the input

cfDNA from 5 ng to 50 ng. Since in clinical plasma samples the amount of ctDNA is

frequently a limiting factor, we also tested the ThruPlex plasma-seq kit (requiring as

little as 3 ng of cfDNA input) to generate a whole genome cfDNA library (termed NGS

cfDNA library). An aliquot of this NGS cfDNA library was then used as input for NG-

TAS.

These NG-TAS experiments showed a strong linear relationship between the

observed and expected variant allele frequencies (VAF) for the four "somatic"

mutations (Table 2, Figure 3). As the input cfDNA reduced from 50 ng to 5 ng the R

squared values decreased from 0.968 to 0.885. With 10 ng input cfDNA, VAFs as low as 1% could be consistently detected. Lowering the input cfDNA generated more variable results (i.e. VAF deviating from the expected values and higher standard deviations), in particular at low AF. This is probably caused by stochastic amplification of the alternative allele. NG-TAS performed using NGS cfDNA library as input performed better than 5 ng of cfDNA input (R^2=0.964, Table 2, Figure 3).

The NG-TAS analysis pipeline was developed and optimised using this dilution series data and later applied to data from clinical plasma samples. As illustrated in Figure 4A and in the Methods section, mutation calling was performed using MuTect2, processing each amplicon individually. To limit the number of false positives (FPs) caused by PCR errors, we only called mutations observed in at least two out of three replicates. With the reported settings and using 10 ng of input cfDNA from the dilution series, all four SNVs were called when the expected VAF was 5% or higher, and 3 of 4 SNVs when the expected VAF was 1% (Figure 4B). No FPs with VAF higher than 3% were called with 50 ng and 10 ng input cfDNA from the dilution series. NG-TAS of both the 5 ng cfDNA input and NGS cfDNA library input generated seven FPs above 3% in the dilution series (Figure 4C). Template scarcity and extra PCR cycles during library preparation could explain this increase in FPs.

Therefore, for NG-TAS in plasma samples we recommend the use of 10 ng cfDNA per replicate as input, and a threshold of 5% VAF for *de novo* mutation calling. In plasma samples with less cfDNA the use of NGS cfDNA library as input for NG-TAS enables ctDNA profiling in samples with as little as 3 ng of cfDNA. However, this approach is more suitable for tracking in plasma ctDNA mutations previously identified in the tumour, rather than for *de novo* plasma ctDNA mutation calling.

Testing NG-TAS performance in cancer patient samples

We applied NGTAS to a clinical cohort of 30 metastatic breast cancer patients from which we have collected 360 plasma samples (for 31 of these NGS cfDNA library samples were used) and buffy coats. This cohort is part of a clinical trial which will be comprehensively reported in a separate manuscript (Baird et al, in preparation).

To estimate the FP rate in blood samples, we used pairs of DNA extracted from the buffy coats collected at two different time points from four patients. Any mutation identified by NG-TAS in any of the eight possible buffy coat DNA pairs tested was considered a FP. Figure 5A shows that in these samples NG-TAS identified no FP with VAF greater than 5% (a result similar to NG-TAS performed using the cell line cfDNA dilution series, Figure 4C).

In 24 of the cases in our cohort, at least one tissue sample was also available and analysed. Sixteen of these cases had tissues from the primary tumour while in the remaining 8 cases, tissue samples were obtained from metastasis biopsies collected during the trial. Overall, we found at least one mutation in 21/24 patients (87.5%, Figure 5B). Forty-four mutations were detected in the tissue samples and 60 in at least one plasma sample; of these, 23 were observed in both tissue and plasma. The agreement was higher for the 8 cases where a metastasis biopsy was sequenced: 7 mutations detected in the tissue, 11 detected in plasma, 7 in common (100% of tissue mutations detected in plasma). In the 16 cases where a primary tumour was tested, 33 mutations were detected in the tissue, 41 in plasma, 19 in common (58% of tissue mutations detected in plasma, Figure 5B and Additional file 1: Fig. S5). The discordance seen in this cohort is probably due to the time gap between the primary tumour tissue sample and plasma, the latter obtained when the patients had metastatic disease. In addition, most of the tissue samples were formalin-fixed paraffin-embedded (FFPE) hence, we detected an increase of C>T/G>A SNVs not usually found in ctDNA samples (Additional file 1: Fig. S5).

We used dPCR to validate a subset of the mutations identified in seven patients in which NG-TAS was performed either directly on cfDNA (n=4) or using post-NGS library products (n=3). In the four direct NG-TAS samples, four hotspot mutations *PIK3CA* (*H1047R* and *E545K*), *KRAS* (*G13D*), *ESR1* (*D538G*) and *AKT1* (*E17K*)) were all validated by dPCR. A good concordance between VAFs estimated by NG-TAS and dPCR was found (R^2=0.64, Figure 5C). In the three patients where post-NGS library products was used as input, two *PIK3CA* hotspots (*H1047R and E545K*) were also validated by dPCR and a high concordance between the VAFs estimated by NG-TAS and dPCR was observed (R^2=0.80, Figure 5D).

Monitoring response in breast cancer patients using NG-TAS

We report the example of two patients from the above clinical trial to demonstrate the use of NG-TAS for metastatic breast cancer disease monitoring. Patients had clinical monitoring performed as per the trial protocol using RECIST (Response Evaluation Criteria in Solid Tumour), version 1.1.

The first patient had RECIST partial response in the first 28 weeks, and progression on day 197. NG-TAS identified mutations in *GATA3* (*F431fs*), *PIK3CA* (*E542K*), *CDKN1B (N124fs)* and *PTEN* (*137-139del*) (Figure 6A). *PTEN* mutation VAFs in ctDNA showed parallel dynamics to RECIST: initial drop, followed by continuous rise from day 85, preceding RECIST progression by over 100 days. The VAFs of the other mutations showed a parallel rise starting later.

The second patient had stable disease by RECIST during the 60 days of available follow up. Due to limited amount of cfDNA extracted in this case, NG-TAS was performed using NGS cfDNA libraries. NG-TAS detected *PIK3CA* (*H1047R*), *MAP3K1* (*E303fs*) and *TP53* (*R141H* and *P46fs*) mutations, and their VAFs showed stable values, then a slight reduction between days 20-56, followed by a slightly rise by the time monitoring was discontinued (Figure 6B).

These two examples demonstrate the use of NG-TAS in plasma cfDNA samples to monitor tumour burden in metastatic breast cancer patients.

Comparison of NG-TAS with other approaches

We finally compared NG-TAS to other existing technologies such as digital PCR, TAm-Seq and Oncomine™ Breast cfDNA Assay (Table 3). NG-TAS can be performed in 7 hours using the Fluidigm system as detailed in the methods session. Up to 384 samples can be processed at the same time. Lower limits of detection can be reached using Digital PCR or Oncomine technology, however this is limited to one target for the first and a set of pre-defined hotspots for the latter. Importantly, the cost of NG-TAS, estimated at 30 GBP per sample is significantly lower than any commercial solution, making it cost-effective for use in the clinics.

DISCUSSION

The genes frequently mutated in different human cancers have been characterized by large-scale sequencing studies such as The Cancer Genome Atlas [20,21]. These pan-cancer studies have revealed that most human tumours have at least 1-10 driver mutations, allowing the design of custom gene panels that could be used for generic cancer detection. But the challenge remaining is there are very few recurrent or hotspot mutations in tumours such as breast cancer, with mutations spread along the protein coding region, as observed in *TP53*, *GATA3* and *MAP3K1*. Therefore, it would be desirable to cover most exons of these genes simultaneously in a ctDNA mutation detection panel.

The detection of specific mutations in ctDNA is achievable by dPCR, now considered the gold standard to detect mutations with low VAFs. However, dPCR is constrained by the number of mutations that can be detected in a single reaction [11]. Thus, its high sensitivity and specificity is at the expense of the number of mutations that can be detected concurrently. At the other end of the spectrum, whole genome sequencing or whole exome sequencing suffer from reduced sensitivity at the current achievable level of sequencing depth [22].

We report here a new approach, NG-TAS, an optimised targeted amplicon sequencing pipeline that provides clinically relevant sensitivity in mutation calling across a targeted, but relatively broad and customizable panel of genes. The current version of NG-TAS covers all exons or hotspots of 20 breast cancer-associated genes in a total of 377 amplicons, has a lower detection limit of 1% VAF, and requires only three aliquots of 10 ng cfDNA input. The single step multiplexed PCR amplification makes it a less time consuming method and more cost effective than other assays, such as the commercially available Oncomine assay (Table 3). NG-TAS is flexible and custom designed primers can be adjusted to the needs of the end user, depending on the cancer type and the clinical context.

Importantly, we developed a bespoke NG-TAS computational pipeline for data analysis, with all the relevant open-source code available at GitHub (https://github.com/cclab-brca/NGTAS_pipeline). All sequencing data are also made available at: https://figshare.com/articles/NGTAS_NA12878/7387370 and https://www.ebi.ac.uk (EGAS00001003392). These will be instrumental to test and further develop the computational pipeline, as required by regulatory agencies.

The custom design of primers for NG-TAS is potentially challenging. Building a customised panel of primers manually, using the tool mentioned above is time-consuming and, in some cases difficult due to genomic sequence context (e.g. high GC and repetitive regions). The multiplex PCR requires a fixed annealing temperature,

but more complex PCR cycle design can circumvent this. Nevertheless, we were able to design primers that yielded in 94% of amplicons over 100x coverage (Figure 2B). We provide all primer sequences (Additional file 2) and an open source optimised primer library will be growing with an NG-TAS user community.

When using NG-TAS for accurate estimation of VAF, as required to do serial tumour burden monitoring, our data suggests that at least 10 ng of input cfDNA per replicate is required. NG-TAS has poor performance with cfDNA input below 5 ng (per replicate), with amplicon coverage reduced in a stochastic manner, probably due to the limited availability of template. A suitable alternative protocol for these cases is to generate an NGS cfDNA library, requiring only 3 ng of cfDNA, and use the library material as input for NG-TAS.

We applied NG-TAS to a cohort of 30 patients for which both tissue and serial plasma samples were available. The percentage of mutations identified in tissue and detected in ctDNA was 100% when the tissue was from a synchronous metastasis biopsy and 58% when the tissue was from the primary tumour. Such agreement is higher than what recently reported by Chae et al [23]. In their cohort of 45 patients, 60% of tissue samples were from primary tumours and 58% of the tissues were acquired more than 90 days before ctDNA testing. The FoundationOne panel was used for tissue analysis and the Guardant360 assay for ctDNA. They detected only 25.6% of tissue mutations in plasma when evaluating the common regions between the two targeted approaches.

A future development of NG-TAS will be the use of molecular barcoding, since this has been shown to improve sensitivity and specificity of amplicon-based deep sequencing [24]. This will have cost implications, potentially limiting one of the main advantages of the current NG-TAS protocol. The extra costs would be the result of generation of barcoded primers. For example, if 96 distinct barcodes are used, the primer cost will increase around 100 times. However, costs will be significantly diluted when considering laboratories processing a large number of samples, keeping the overall cost of NG-TAS within a very reasonable range.

Conclusions

We have described here the workflow for a highly multiplexed cfDNA deep sequencing method named NG-TAS. NG-TAS assesses the mutational status of several genes simultaneously, with high sensitivity (allowing quantification of AF) and competitive costs, and offers flexibility in the choice of target genes. We have also shown proof of principle that the monitoring of ctDNA using NG-TAS in metastatic breast cancer can allow detection of cancer progression earlier than conventional RECIST measurements.

1 **List of abbreviations**

Next Generation Targeted Amplicon Sequencing	NG-TAS
Cell Free DNA	cfDNA
Circulating Tumour DNA	ctDNA
Digital PCR	dPCR
Targeted Amplicon Sequencing	TAm-Seq
Next Generation Sequencing	NGS
Formalin Fixed Paraffin Embedded	FFPE
Variant Allele frequency	VAF
Single Nucleotide Variant	SNV
False Positive	FP
Uracil DNA Glycosylase	UDG
Response Evaluation Criteria in Solid Tumour	RECIST
Computed Tomography	CT

Declarations

Ethics approval and consent to participate

This study was approved by the East of England - Cambridge East Research Ethics

Committee (REC reference: 14/EE/1045). All human samples used were collected

after informed consent, and the study was fully compliant with the Helsinki

Declaration.

Consent for publication

Not applicable

Availability of data and material

The computational pipeline is available through GitHub (https://github.com/cclab-

brca/NGTAS_pipeline). All sequencing data are available at:

https://figshare.com/articles/NGTAS_NA12878/7387370 and https://www.ebi.ac.uk

(EGAS00001003392).

Competing interests

The authors declare that they have no competing interests.

Funding

This research was supported with funding from Cancer Research UK. MG has been

supported by a Genentech research grant (CLL-010907) awarded to the Caldas

Laboratory. MC has received funding from the European Union's Horizon 2020

research and innovation program under the Marie Skłodowska-Curie grant

agreement no. 660060.

Authors' contributions

MG, MC, EB, SFC and CC conceived the study; MG and EB designed primers and

generated the dilution series; MG performed NG-TAS in Fluidigm Access Array™

system; RB, KB, SCL, JR, JC and MO conducted the clinical trial; HB, LJ and AB

collected the clinical samples; MC developed the computational approach and

performed the analyses; SJS contributed in the computational pipeline development;

MC and MG performed data analysis; MG, MC, SFC and CC drafted the manuscript,

all authors revised and approved the final manuscript.

Acknowledgements

We are grateful to Cancer Research UK and the University of Cambridge for the

support. We thank the Cancer Molecular Diagnostics Lab and Cancer Research UK

Cambridge Institute Core Facilities including Genomics and Bio-repository that

supported this work. We thank Dr PA Edwards for the scientific advice and editing for

this manuscript.

Additional files

Additional file 1: Supplementary figures and tables, PDF 722 KB

Additional file 2: List and description of the 377 primers used, XLS 214 KB

Table 1. List of genes and regions covered in the panel

Gene	Target region	Hotspot position	No of ampicons
AKT1	Hotspot	E17 AA23 – 59 AA65 - 94	4
BRAF	Hotspot	V600	1
Her2	Hotspot	S310 AA428 - 438 AA746 – 797 AA832 – 986	14
HRAS	Hotspot	AA3 – 35 (G12 and G13) AA49 – 77 (Q61 and A66)	3
IDH2	Hotspot	AA132 - 162	1
KRAS	Hotspot	G12	1
SF3B1	Hotspot	K700	1
ESR1	Part of exons	Exon 8 - 10 (LBD)	10
SMAD4	Part of exons	Exon 8 – 12	10
CDH1	All exons		46
CDKN1B	All exons		9
FOXA1	All exons		18
GATA3	All exons		23
MAP2K4	All exons		22
MAP3K1	All exons		75
PIK3CA	All exons		59
PIK3R1	All exons		11
PTEN	All exons		24
RUNX1	All exons		24
TP53	All exons		21

1 **Table 2. Linear regression analysis for different cfDNA input**
2

Input DNA	R^2	Estimated coefficient	2.5% CI	97.5% CI
50ng	0.968	1.075	1.018	1.133
10ng	0.940	1.005	0.930	1.080
5ng	0.885	0.932	0.832	1.032
Library	0.964	1.123	1.059	1.187

3
4

Table 3: Comparison of different approaches for ctDNA detection

	NG-TAS	Digital PCR	TAm-Seq	Oncomine™ Breast cfDNA Assay
Approx. cost per sample (GBP)	3x10	2-3	10	200
Time (96 reactions)	7h	3h x 2	2 days	2 days
Number of amplicons	377	1	48	n/a
Amplicon size	<160bp	n/a	<200bp	<170bp
Number of genes	20	1 (hotspot)	6	10
Choice of targets	Flexible	Limited	Flexible	Limited
Median depth	3064	770	650 (avg)	n/a
Limit of detection	1-2%	>0.1%	1-2%	0.6-0.1%
Library material as input	Yes	Not tested	Not tested	Not tested
Ideal input	3x10ng	2-5ng	50ng	20ng
Multiplex	8plex	No	No	n/a
Number of samples per sequence run	384	n/a	96	12
Platform	Fluidigm Access Array	Fluidigm qdPCR	Fluidigm Access Array	Ion Chef System
Sequencing	Illumina HiSeq 4000	n/a	Illumina GAIIx	Ion S5/S5XL™

Figure Legends

Figure 1. NG-TAS workflow and alignment pipeline

(A) NG-TAS workflow. Primers were designed and multiplexed for direct amplification

in cfDNA obtained from plasma using Fluidigm Access Array™. The PCR products

were harvested and barcoded in a subsequent PCR reaction, the samples were pooled

and size selected for sequencing on an Illumina Hi-Seq 4000. (B) Schematic

representation of the computational pipeline for reads alignment, filtering and

annotation.

Figure 2. Optimising targeted deep sequencing by NG-TAS

(A) Percentage of aligned reads were compared in different samples where a variable

amount of input control genomic DNA was used (range 50 to 0.016ng). The effect of

pre-amplification and Q solutions are shown, red = No Q solution and no pre-

amplification step, green = With Q solution and no pre-amplification, blue = No Q

solution and with pre-amplification. (B) Density plot showing the log10 coverage values

for all primers in the 10 ng NA12878 cfDNA sample. The dotted line indicates 100x

coverage; median value for the distribution is 3064x. (C) Coverage heatmap of

individual primers for different amount of input NA12878 cfDNA. For each amount of

input DNA the analysis was performed in triplicate. (D) Distribution of all non-reference

base frequencies across all target regions in the NA12878 dilution series in (C); the

smaller plot on the right is a magnification of the main plot between 0 and 0.01.

Figure 3. Detection of SNVs in NA12878 cfDNA dilution series

(A) Expected versus observed VAF for 4 SNVs in the NA12878-NA11840 dilution

series starting from 50 ng input DNA (left) and zoom-in for expected VAF <5% (right).

(B) Expected versus observed VAF for 4 SNVs in the NA12878-NA11840 dilution

series starting from 10 ng input DNA (left) and zoom-in for expected VAF <5% (right).

(C) Expected versus observed VAF for 4 SNVs in the NA12878-NA11840 dilution

series starting from 5 ng input DNA (left) and zoom-in for expected VAF <5% (right).

(D) Expected versus observed VAF for 4 SNVs in the NA12878-NA11840 dilution

series starting from post-NGS library input DNA (left) and zoom-in for expected VAF

<5% (right).

Figure 4. Mutation calling in NA12878 cfDNA dilution series

(A) Schematic overview of the computational pipeline to identify somatic mutations in

NG-TAS data. (B) *De novo* mutation calling in the NA12878 dilution series was

evaluated for different amounts of input cfDNA. 4 SNVs can potentially be called using

our panel of 377 amplicons. (C) VAF for all FP calls in the NA12878 dilution series. The red dashed line represents 5% VAF.

Figure 5. Validation of NG-TAS performance in clinical plasma samples

(A) The specificity of NG-TAS in clinical samples was estimated using 4 pairs of buffy coats from the same patients (A, B, C and D). The mutation calling pipeline was applied using one buffy coat as normal and the other as 'tumour' and vice versa. All mutations called in this setting can be considered FPs. The red line indicates 5% VAF. (B) Oncoprint summary plot of genes mutated in 24 cases for which both tissue and plasma samples were tested. The vertical black line separates cases for which the primary tumour was analysed from cases for which a metastasis biopsy was analysed. (C-D) Comparison of VAF obtained by NG-TAS and dPCR. (C) In this comparison, four different hotspot mutations including *AKT1* (*E17K*), *ESR1* (*D538G*), *KRAS* (*G13D*) and *PIK3CA* (*H1047R*) identified in multiple plasma samples from 4 distinct patients were analysed (R^2 = 0.64). (D) Two *PIK3CA* hotspots (*H1047R* and *E545K*) were detected by NG-TAS using NGS library as input material in plasma samples from two distinct patients. The same mutations were detected using dPCR and a good correlation was found (R^2 = 0.80).

Figure 6. Monitoring response in metastatic breast cancer patients using NG-TAS

(A) Example of patient monitoring during treatment using direct NG-TAS in ctDNA. There are four mutations detected in more than one sample: *GATA3* (*F431fs*), *PIK3CA* (*E542K*), *CDKN1B (N124fs)* and *PTEN* (*137-139del*). The mutations called more than once in the longitudinal samples are shown including tumour and plasma samples. The arrow indicates the time of the disease considered as RECIST progressive disease. T indicates tumour samples, and SP indicates screening plasma sample which was collected prior to the treatment. (B) Example of patient monitoring during treatment using NGS library material for NG-TAS. This patient had a stable disease during the whole treatment period. There are three mutations detected, including *MAP3K1* (*E303 frame shift*), hotspot mutations *PIK3CA (E545K)* and *TP53* (*R141H* and *P46fs*). T indicates tumour samples.

REFERENCES

[1]   Leon SA, Shapiro B, Sklaroff DM, Leon SA, Shapiro B, Sklaroff DM, et al. Free DNA in the Serum of Cancer Patients and the Effect of Therapy Free DNA in the Serum of Cancer Patients and the Effect of Therapy. Cancer Res 1977;37:646–50.

[2]   Vasioukhin V, Anker P, Maurice P, Lyautey J, Lederrey C, Stroun M. Point mutations of the N-ras gene in the blood plasma DNA of patients with myelodysplastic syndrome or acute myelogenous leukaemia. Br J Haematol 1994;86:774–9. doi:10.1111/j.1365-2141.1994.tb04828.x.

[3]   Sorenson GD, Pribish DM, Valone FH, Memoli V a, Bzik DJ, Yao SL. Soluble Normal and Mutated Dna-Sequences From Single-Copy Genes in Human Blood. Cancer Epidemiol Biomarkers Prev 1994;3:67–71.

[4]   Murtaza M, Dawson S-J, Tsui DWY, Gale D, Forshew T, Piskorz AM, et al. Non-invasive analysis of acquired resistance to cancer therapy by sequencing of plasma DNA. Nature 2013;497:108–12. doi:10.1038/nature12065.

[5]   Diehl F, Schmidt K, Choti M a, Romans K, Goodman S, Li M, et al. Circulating mutant DNA to assess tumor dynamics. Nat Med 2008;14:985–90. doi:10.1038/nm.1789.

[6]   Schwarzenbach H, Hoon DS. B, Pantel K. Cell-free nucleic acids as biomarkers in cancer patients. Nat Rev Cancer 2011;11:426–37. doi:10.1038/nrc3066.

[7]   Dawson S-J, Tsui DWY, Murtaza M, Biggs H, Rueda OM, Chin S-F, et al. Analysis of circulating tumor DNA to monitor metastatic breast cancer. N Engl J Med 2013;368:1199–209.

[8]   Wan JCM, Massie C, Garcia-corbacho J, James D, Caldas C, Pacey S, et al. Liquid biopsies come of age: clinical applications of circulating tumour DNA. Nat Rev Cancer 2016. doi:10.1038/nrc.2017.7.

[9]   Vogelstein B, Kinzler KW. Digital PCR. Proc Natl Acad Sci U S A 1999;96:9236–41. doi:10.1073/pnas.96.16.9236.

[10]   Dressman D, Yan H, Traverso G, Kinzler KW, Vogelstein B. Transforming single DNA molecules into fluorescent magnetic particles for detection and enumeration of genetic variations. Proc Natl Acad Sci U S A 2003;100:8817–22. doi:10.1073/pnas.1133470100.

[11]   Hughesman CB, Lu XJD, Liu KYP, Zhu Y, Poh CF, Haynes C. Robust protocol for using multiplexed droplet digital PCR to quantify somatic copy number alterations in clinical tissue specimens. PLoS One 2016;11:1–22. doi:10.1371/journal.pone.0161274.

[12]   Forshew T, Murtaza M, Parkinson C, Gale D, Tsui DWY, Kaper F, et al. Noninvasive Identification and Monitoring of Cancer Mutations by Targeted Deep Sequencing of Plasma DNA. Sci Transl Med 2012;4:136ra68--136ra68. doi:10.1126/scitranslmed.3003726.

[13]   Torga G, Pienta KJ. Patient-Paired Sample Congruence Between 2 Commercial Liquid Biopsy Tests. JAMA Oncol 2017. doi:10.1001/jamaoncol.2017.4027.

[14]   Kinde I, Wu J, Papadopoulos N, Kinzler KW, Vogelstein B. Detection and quantification of rare mutations with massively parallel sequencing. Proc Natl Acad Sci 2011;108:9530–5. doi:10.1073/pnas.1105422108.

[15]   Callari M, Sammut S-J, De Mattos-Arruda L, Bruna A, Rueda OM, Chin S-F, et al. Intersect-then-combine approach: improving the performance of somatic variant calling in whole exome sequencing data using multiple aligners and callers. Genome Med 2017;9:35. doi:10.1186/s13073-017-0425-1.

[16]   Eberle MA, Fritzilas E, Krusche P, Källberg M, Moore BL, Bekritsky MA. A reference dataset of 5.4 million human variants validated by genetic inheritance from sequencing a three-generation 17-member pedigree. Genome Res 2017;27. doi:10.1101/gr.210500.116.

[17] Yung TKF, Chan KCA, Mok TSK, Tong J, To K-F, Lo YMD. Single-molecule detection of epidermal growth factor receptor mutations in plasma by microfluidics digital PCR in non-small cell lung cancer patients. Clin Cancer Res 2009;15:2076–84. doi:10.1158/1078-0432.CCR-08-2622.

[18] Schiavon G, Hrebien S, Garcia-Murillas I, Cutts RJ, Pearson a., Tarazona N, et al. Analysis of ESR1 mutation in circulating tumor DNA demonstrates evolution during therapy for metastatic breast cancer. Sci Transl Med 2015;7:313ra182--313ra182.

[19] Toy W, Shen Y, Won H, Green B, Sakr RA, Will M, et al. ESR1 ligand-binding domain mutations in hormone-resistant breast cancer. Nat Genet 2013;45:1439–45. doi:10.1038/ng.2822.

[20] Cerami E, Gao J, Dogrusoz U, Gross BE, Sumer SO, Arman B, et al. In Focus The cBio Cancer Genomics Portal : An Open Platform for Exploring Multidimensional Cancer Genomics Data 2012. doi:10.1158/2159-8290.CD-12-0095.

[21] Gao J, Aksoy BA, Dogrusoz U, Dresdner G, Gross B, Sumer SO, et al. Integrative Analysis of Complex Cancer Genomics and Clinical Profiles Using the cBioPortal Complementary Data Sources and Analysis Options 2013;6:1–20.

[22] Sims D, Sudbery I, Ilott NE, Heger A, Ponting CP. Sequencing depth and coverage: key considerations in genomic analyses. Nat Rev Genet 2014;15:121–32. doi:10.1038/nrg3642.

[23] Chae YK, Davis AA, Jain S, Santa-Maria C, Flaum L, Beaubier N, et al. Concordance of Genomic Alterations by Next-Generation Sequencing in Tumor Tissue versus Circulating Tumor DNA in Breast Cancer. Mol Cancer Ther 2017;16:1412–20. doi:10.1158/1535-7163.MCT-17-0061.

[24] Kou R, Lam H, Duan H, Ye L, Jongkam N, Chen W. Benefits and Challenges with Applying Unique Molecular Identifiers in Next Generation Sequencing to Detect Low Frequency Mutations 2016:1–15. doi:10.5061/dryad.n6068.Funding.

Figure 1

Figure 2

Figure 3

Figure 4

A

B

C

Figure 5

Figure 6

A Direct NGTAS

B Library - NGTAS