1 **Next Generation-Targeted Amplicon Sequencing (NG-TAS): An optimised**

2 **protocol and computational pipeline for cost-effective profiling of circulating**

3 **tumour DNA.**

5 **Authors**

Meiling Gao[1,6]   mg390@cam.ac.uk

Maurizio Callari[1,6]   Maurizio.Callari@cruk.cam.ac.uk

Emma Beddowes[1,2]   Emma.Beddowes@cruk.cam.ac.uk

Stephen-John Sammut[1]   Stephen-John.Sammut@cruk.cam.ac.uk

Marta Grzelak[1]   Marta.Grzelak@cruk.cam.ac.uk

Heather Biggs[2]   heather.biggs@addenbrookes.nhs.uk

Linda Jones[2]   linda.jones@addenbrookes.nhs.uk

Abdelhamid Boumertit[2]   ab2421@medschl.cam.ac.uk

Sabine C Linn[3]   s.linn@nki.nl

Javier Cortes[4,5]   jacortes@vhio.net

Richard Baird[2]   rdb39@medschl.cam.ac.uk

Suet-Feung Chin[1]*   Suet-Feung.Chin@cruk.cam.ac.uk

Carlos Caldas[1,2]*   Carlos.Caldas@cruk.cam.ac.uk

Institutional Addresses

[1]Department of Oncology and Cancer Research UK Cambridge Institute, Li Ka Shing

Centre, University of Cambridge, Cambridge CB2 0RE, UK

[2]Breast Cancer Programme, Cancer Research UK Cambridge Cancer Centre,

Cambridge University Hospitals NHS Foundation Trust, Cambridge CB2 2QQ, UK

[3]Netherland Cancer Institute, 1006 BE, Amsterdam, Netherland

[4]Ramon y Cajal University Hospital, 28034 Madrid, Spain

[5]Vall d'Hebron Institute of Oncology, 08035 Barcelona, Spain

[6]Co-first author

*Co-corresponding authors

Abstract

Circulating tumour DNA (ctDNA) detection and monitoring has enormous potential

clinical utility in oncology. We describe here a fast, flexible and cost-effective method

to profile multiple genes simultaneously in low input ctDNA: Next Generation-Targeted

Amplicon Sequencing (NG-TAS). We designed a panel of 377 amplicons spanning 20

cancer genes and tested the NG-TAS pipeline using cell-free DNA from two hapmap

lymphoblastoid cell lines. NG-TAS robustly detected mutations in cell-free DNA

(cfDNA) when mutation allele fraction was >1%. NG-TAS of plasma cfDNA is therefore

adequate to monitor metastatic cancer, as we show in a pilot study.

Keywords

NG-TAS, ctDNA, liquid biopsy, mutation, multiplexing, deep sequencing,

computational pipeline, cancer, heterogeneous

BACKGROUND

Cell free DNA (cfDNA) in plasma was first analysed in cancer patients nearly 50 years ago [1]. A fraction of cfDNA was shown to carry mutations found in the matched tumour, and designated circulating tumour DNA (ctDNA) [1–3]. The utility of ctDNA as a non-invasive diagnostic, prognostic or predictive biomarker in human cancer is now well documented [4–8].

The amount of cfDNA in plasma is usually low and the ctDNA fraction is typically only 1-30%, hence low mutant allele frequencies have to be robustly detected. Human cancers are genetically heterogeneous and mutations occur infrequently at recurrent hotspots. Therefore, in most clinical scenarios (e.g. early diagnosis or monitoring of tumour evolution), high sensitivity and the simultaneous investigation of multiple gene targets are desirable features of any ctDNA detection and quantitation method.

There are a range of methods for detecting mutations in ctDNA, with the target varying from a single nucleotide variant (SNV) to the whole genome. A widely used method to detect mutations in ctDNA is digital polymerase chain reaction (dPCR) performed in microfluidic devices or water-in-oil droplet emulsions [9,10]. Whilst dPCR is able to detect rare mutations with extremely high sensitivity, it is restricted by the number of targets that can be examined in a single reaction [11] .

Several sequencing based approaches have been developed to incorporate multiple genomic loci, enabling *de novo* mutation identification in ctDNA. Previously, we described Targeted Amplicon Sequencing (TAm-Seq), which utilised 48 primer pairs to identify mutations in hotspots or selected regions of 6 key driver genes [12]. While TAm-Seq is useful, it is limited to a small number of targets. Capture based sequencing methods can cover a larger number of genes (or the whole exome), but are costly at the sequencing coverage (>300) required to detect allele frequencies ~1%.

There several ready to use commercial kits for ctDNA sequencing, which can cover up to hundreds of mutation hotspots and many genes. These include Invision™ (Inivata), Oncomine™ cfDNA assay (Thermo Fisher Scientific), Guardant360™ (Guardant Health), and PlasmaSELECT™ (Personal Genome Diagnostics). These products are expensive and test custom gene panels. Disturbingly, a recent study comparing the performance of two of these commercial products (Guardant360™ and PlasmaSELECT™) in a cohort of plasma samples from prostate cancer patients, revealed poor agreement [13].

Recently unique molecular barcodes have been developed to tag each cfDNA template molecule before PCR amplification in order to reduce the error rate and allow robust detection of rare mutant alleles in ctDNA [14].

In summary, using current ctDNA profiling methodology, the detection of mutations

in a good number of cancer genes with sufficient sensitivity and in a cost-effective way

poses significant challenges. Here we describe a new method for the profiling of

ctDNA, designated Next Generation-Targeted Amplicon Sequencing (NG-TAS), with

several unique features: i) optimised for low input ctDNA; ii) high level of multiplexing,

enabling the analyses of multiple gene targets; iii) a bespoke computational pipeline

for data analysis; and iv) very competitive costing. NG-TAS is designed to be flexible

in terms of choice of gene targets and regions of interest; thus, it can be tailored to

various cancer types and clinical contexts.

RESULTS

Optimising targeted deep sequencing in cfDNA by NG-TAS

We designed 377 pairs of primers covering all exons or hotspots of 20 genes commonly mutated in breast cancer (Table 1). Since the average cfDNA fragment size is 160-170bp, NG-TAS primers were designed to generate amplicons of 69-157bp (Table S1).

In a preliminary optimization step, individual primer pairs were tested in conventional single and multiplexed (7-8plex) PCR reactions. The NG-TAS experimental workflow (Figure 1A), starts with a multiplexed PCR step (7-8 primer pairs) performed using Access Array™, a microfluidic system from Fluidigm. Each multiplexed reaction contained primers targeting different genes to minimise the generation of unwanted PCR products. The multiplexed PCR products were assessed using the Bioanalyser and 2200 TapeStation instrument (Agilent Genomics; Figure S1). Multiplexed PCR products were then pooled and barcoded with 384 unique barcodes in a second PCR reaction. Barcoded products were pooled, and size selected to remove primer dimers before submission for NGS paired-end 150bp sequencing.

Raw sequencing data were aligned and processed as described in Figure 1B and Methods. Specific filters were applied to exclude reads from primer dimers or other PCR artefacts. Since the amplicons are partially overlapping, each read was assigned to its respective amplicon, to enable a per-amplicon analysis for coverage estimation and mutation calling.

To optimise NG-TAS we used cfDNA isolated from the culture media of the Platinum Genome Hapmap NA12878 cell line. The size profile of cfDNA isolated from the tissue culture media was similar to that of plasma cfDNA (Figure S2). We tested a range of input cfDNA amounts with NG-TAS (0.016 ng to 50 ng) in 4 replicates for each input. For each cfDNA input we tested: i) a pre-amplification step and ii) the use of the Qiagen Q solution. To assess the data generated the percentage of aligned sequencing reads was computed (Figure 2A). In the TAM-Seq protocol addition of a pre-amplification step reduced the probability of nonspecific amplification and biased coverage [12]. However, using NG-TAS the pre-amplification step reduced the percentage of aligned reads in all cfDNA input samples tested. Hence, we eliminated pre-amplification from the NG-TAS protocol. Adding Q solution systematically increased the percentage of aligned reads, with the largest improvement observed with 0.4 and 2 ng input samples (Figure 2A). Thus, we incorporated the Q solution in all subsequent NG-TAS experiments.

We then used the optimised NG-TAS protocol in triplicate experiments for each input NA12878 cfDNA (2 ng, 5 ng, and 10 ng). With 10ng of input cfDNA NG-TAS generated

a median read depth of 3064x, and only 22/377 amplicons (5.8%) had coverage less

than 100x (Figure 2B). In fact, high amplicon coverage was observed irrespective of

amount of input cfDNA (Figure S3A and S3B). The coverage heatmap of individual

amplicons showed similar patterns with 10 ng and 5 ng cfDNA input.   Strong

consistency was observed within each triplicate (Figure 2C). However, with 2ng cfDNA

input we observed stochastic reduction in coverage for some of the amplicons. This is

probably due to a reduction in template availability, with the number of amplifiable

copies approaching zero for some of the amplicons.

Using these data, the background noise was estimated by computing the average

frequency for non-reference bases in each position, and for 99% of the targeted

genomic positions background noise was ≤0.3% (Figure 2D).

Sensitivity and specificity of mutation detection in control cfDNA

To establish an analysis pipeline and assess the performance of NG-TAS, we

generated a benchmark dilution series, similar to what we have previously described

[15], using cfDNA collected from the tissue culture media from two lymphoblastoid cell

lines from the HapMap/1000 Genome Project, NA12878 (the Platinum Genome

sample) and NA11840, to mimic a tumour-normal (or plasma-normal) pair. The dilution

series mixed cfDNA from NA12878 with an increasing amount of cfDNA from NA11840

(from 0 to 99.8% by volume, n=12, Table S2). This cfDNA dilution series was used to

investigate the sensitivity in detecting mutations at high and low allele frequency (50%

- 0.1%). The 377-amplicon panel encompassed four heterozygous single nucleotide

polymorphisms (SNPs) present only in NA12878. These SNPs were used as "somatic"

mutations for the purpose of this analysis.

Using NG-TAS the cfDNA dilution series was tested in triplicate, varying the input

cfDNA from 5 ng to 50 ng. Since in clinical plasma samples the amount of ctDNA is

frequently a limiting factor, we also tested the ThruPlex plasma-seq kit (requiring as

little as 3 ng of cfDNA input) to generate a whole genome cfDNA library (termed NGS

cfDNA library). An aliquot of this NGS cfDNA library was then used as input for NG-

TAS.

These NG-TAS experiments showed a strong linear relationship between the

observed and expected variant allele frequencies (VAF) for the four "somatic"

mutations (Table 2, Figure 3). As the input cfDNA reduced from 50 ng to 5 ng the R

squared values decreased from 0.968 to 0.885. With 10 ng input cfDNA, VAFs as low

as 1% could be robustly detected. Lowering the input cfDNA generated more variable

results, in particular at low AF. This is probably caused by stochastic amplification of

the alternative allele. NG-TAS performed using NGS cfDNA library as input performed
better than 5 ng of cfDNA input (R^2=0.964, Table 2, Figure 3).

The NG-TAS analysis pipeline was developed and optimised using this dilution
series data and later applied to data from clinical plasma samples. As illustrated in
Figure 4A and in the Methods section, mutation calling was performed using MuTect2,
processing each amplicon individually. To limit the number of false positives (FPs)
caused by PCR errors, we only called mutations observed in at least two out of three
replicates. With the reported settings and using 10 ng of input cfDNA from the dilution
series, all four SNVs were called when the expected VAF was 5% or higher, and 3 of
4 SNVs when the expected VAF was 1% (Figure 4B). No FPs with VAF higher than
3% were called with 50 ng and 10 ng input cfDNA from the dilution series. NG-TAS of
both the 5 ng cfDNA input and NGS cfDNA library input generated seven FPs above
3% in the dilution series (Figure 4C). Template scarcity and extra PCR cycles during
library preparation could explain this increase in FPs.

Therefore, for NG-TAS in plasma samples we recommend the use of 10 ng cfDNA
per replicate as input, and a threshold of 5% VAF for *de novo* mutation calling. In
plasma samples with less cfDNA the use of NGS cfDNA library as input for NG-TAS
enables ctDNA profiling in samples with as little as 3 ng of cfDNA. However, this
approach is more suitable for tracking in plasma ctDNA mutations previously identified
in the tumour, rather than for *de novo* plasma ctDNA mutation calling.

Testing NG-TAS performance in cancer patient samples
We applied NGTAS to a clinical cohort of 30 metastatic breast cancer patients from
which we have collected 360 plasma samples (for 31 of these NGS cfDNA library
samples were used) and buffy coats. This cohort is part of a clinical trial which will be
comprehensively reported in a separate manuscript (Baird et al, in preparation).

To estimate the FP rate in blood samples, we used pairs of DNA extracted from the
buffy coats collected at two different time points from four patients. Any mutation
identified by NG-TAS in any of the eight possible buffy coat DNA pairs tested was
considered a FP. Figure 5A shows that in these samples NG-TAS identified no FP with
VAF greater than 5% (a result similar to NG-TAS performed using the cell line cfDNA
dilution series, Figure 4C).

In 24 of the cases in our cohort, at least one tissue sample was also available and
analysed. Sixteen of these cases had tissues from the primary tumour while in the
remaining 8 cases, tissue samples were obtained from metastasis biopsies collected
during the trial. Overall, we found at least one mutation in 21/24 patients (87.5%, Figure
5B). Forty-four mutations were detected in the tissue samples and 60 in at least one

plasma sample; of these, 23 were observed in both tissue and plasma. The agreement was higher for the 8 cases where a metastasis biopsy was sequenced (7 mutations detected in the tissue, 11 detected in plasma, 7 in common) than in the 16 cases where a primary tumour was tested (33 mutations detected in the tissue, 41 in plasma, 19 in common, Supplementary Figure 5). The discordance in this case is probably due to the time gap between the primary tumour tissue sample and plasma, the latter obtained when the patients had metastatic disease. In addition, most of the tissue samples were formalin-fixed paraffin-embedded (FFPE) hence, we detected an increase of C>T/G>A SNVs not usually found in ctDNA samples (Supplementary Figure 5).

We used dPCR to validate a subset of the mutations identified in seven patients in which NG-TAS was performed either directly on cfDNA (n=4) or using post-NGS library products (n=3). In the four direct NG-TAS samples, four hotspot mutations *PIK3CA* (*H1047R* and *E545K*), *KRAS* (*G13D*), *ESR1* (*D538G*) and *AKT1* (*E17K*)) were all validated by dPCR. A good concordance between VAFs estimated by NG-TAS and dPCR was found (R^2=0.64, Figure 5C). In the three patients where post-NGS library products was used as input, two *PIK3CA* hotspots (*H1047R and E545K*) were also validated by dPCR and a high concordance between the VAFs estimated by NG-TAS and dPCR was observed (R^2=0.80, Figure 5D).

Monitoring response in breast cancer patients using NG-TAS

We report the example of two patients from the above clinical trial to demonstrate the use of NG-TAS for metastatic breast cancer disease monitoring. Patients had clinical monitoring performed as per the trial protocol using RECIST (Response Evaluation Criteria in Solid Tumour), version 1.1.

The first patient had RECIST partial response in the first 28 weeks, and progression on day 197. NG-TAS identified mutations in *GATA3* (*F431fs*), *PIK3CA* (*E542K*), *CDKN1B (N124fs)* and *PTEN* (*137-139del*) (Figure 6A). *PTEN* mutation VAFs in ctDNA showed parallel dynamics to RECIST: initial drop, followed by continuous rise from day 85, preceding RECIST progression by over 100 days. The VAFs of the other mutations showed a parallel rise starting later.

The second patient had stable disease by RECIST during the 60 days of available follow up. Due to limited amount of cfDNA extracted in this case, NG-TAS was performed using NGS cfDNA libraries. NG-TAS detected *PIK3CA* (*H1047R*), *MAP3K1* (*E303fs*) and *TP53* (*R141H* and *P46fs*) mutations, and their VAFs showed stable values, then a slight reduction between days 20-56, followed by a slightly rise by the time monitoring was discontinued (Figure 6B).

1   These two examples demonstrate the use of NG-TAS in plasma cfDNA samples to
2   monitor tumour burden in metastatic breast cancer patients.
3
4

DISCUSSION

The genes frequently mutated in different human cancers have been characterized by large-scale sequencing studies such as The Cancer Genome Atlas [16,17]. These pan-cancer studies have revealed that most human tumours have at least 1-10 driver mutations, allowing the design of custom gene panels that could be used for generic cancer detection. But the challenge remaining is there are very few recurrent or hotspot mutations in tumours such as breast cancer, with mutations spread along the protein coding region, as observed in *TP53*, *GATA3* and *MAP3K1*. Therefore, it would be desirable to cover most exons of these genes simultaneously in a ctDNA mutation detection panel.

The detection of specific mutations in ctDNA is achievable by dPCR, now considered the gold standard to detect mutations with low VAFs. However, dPCR is constrained by the number of mutations that can be detected in a single reaction [11]. Thus, its high sensitivity and specificity is at the expense of the number of mutations that can be detected concurrently. At the other end of the spectrum, whole genome sequencing or whole exome sequencing suffer from reduced sensitivity at the current achievable level of sequencing depth [18].

We report here a new approach, NG-TAS, an optimised targeted amplicon sequencing pipeline that provides clinically relevant sensitivity in mutation calling across a targeted, but relatively broad and customizable panel of genes. The current version of NG-TAS covers all exons or hotspots of 20 breast cancer-associated genes in a total of 377 amplicons, has a lower detection limit of 1% VAF, and requires only three aliquots of 10 ng cfDNA input. The single step multiplexed PCR amplification makes it a less time consuming method and more cost effective than other assays, such as the commercially available Oncomine assay (Table 3). NG-TAS is flexible and custom designed primers can be adjusted to the needs of the end user, depending on the cancer type and the clinical context.

Importantly, we developed a bespoke NG-TAS computational pipeline for data analysis, with all the relevant open-source code available at GitHub (https://github.com/cclab-brca/NGTAS_pipeline), together with the benchmark dataset generated using control cfDNA obtained from the Platinum Genome cell line NA12878. These will be instrumental to test and further develop the computational pipeline, as required by regulatory agencies (https://precision.fda.gov/).

The custom design of primers for NG-TAS is potentially challenging. Building a customised panel of primers manually, using the tool mentioned above is time-consuming and, in some cases difficult due to genomic sequence context (e.g. high GC and repetitive regions). The multiplex PCR requires a fixed annealing temperature,

but more complex PCR cycle design can circumvent this. Nevertheless, we were able to design primers that yielded in 94% of amplicons over 100x coverage (Figure 2B). We provide all primer sequences (Table S1) and an open source optimised primer library will be growing with an NG-TAS user community.

When using NG-TAS for accurate estimation of VAF, as required to do serial tumour burden monitoring, our data suggests that at least 10 ng of input cfDNA per replicate is required. NG-TAS does not perform robustly with cfDNA input below 5 ng (per replicate), with amplicon coverage reduced in a stochastic manner, probably due to the limited availability of template. A suitable alternative protocol for these cases is to generate an NGS cfDNA library, requiring only 3 ng of cfDNA, and use the library material as input for NG-TAS.

A future development of NG-TAS will be the use of molecular barcoding, since this has been shown to improve sensitivity and specificity of amplicon-based deep sequencing [19]. This will have cost implications that would limit one of the main advantages of the current NG-TAS protocol.

Conclusions

We have described here the workflow for a highly multiplexed cfDNA deep sequencing method named NG-TAS. NG-TAS assesses the mutational status of several genes simultaneously, with high sensitivity (allowing quantification of AF) and competitive costs, and offers flexibility in the choice of target genes. We have also shown proof of principle that the monitoring of ctDNA using NG-TAS in metastatic breast cancer can allow detection of cancer progression earlier than conventional RECIST measurements.

METHODS

Patient samples and blood processing

Patients were recruited from 3 different centres including Cambridge University

Hospital, Netherland Cancer Institute (NKI) and Vall d'Hebron Institute of Oncology

(VHIO). Metastatic breast cancer patients with hormone receptor positive tumours

were recruited as a part of a clinical trial (patient number = 30, plasma samples number

= 366). Blood samples were collected in EDTA tubes and processed within one hour

to prevent lymphocyte lysis and fragmentation. Samples were centrifuged at 820g for

10min at room temperature to separate the plasma from the peripheral blood cells.

The plasma was further centrifuged at 1400g for 10min to remove remaining cells and

cell debris. The plasma was stored at -80°C until DNA extraction. This study was

approved by the regulatory and ethic committees at each site and the reference

number is NCT02285179 (https://clinicaltrials.gov/ct2/show/NCT02285179).

DNA extraction from plasma and buffy coat

Plasma DNA was extracted from between 2-4 ml of plasma with the QiaSymphony

according to the manufacturer's instruction using Qiagen circulating DNA extraction

kit. DNA was isolated from the buffy coat samples using DNeasy Blood & Tissue Kits

(Qiagen).

Generation of cfDNA from NA12878 and NA11840

As previously reported [15] two lymphoblastoid cell lines, NA12878 and NA11840 from

Human Genome Diversity Project (HGDP)-CEPH collection were obtained from the

Coriell Cell Repository. The cell lines were grown as suspension in RPMI 1640-

Glutamax (Invitrogen) supplemented with 10% foetal calf serum, 5% penicillin and

streptomycin at 37°C and 5% CO_2. The media that the cell lines were grown in were

collected when cells were passaged. The media were centrifuged at 1500rpm for 10

minutes at 4°C to remove cells and cellular debris. The clarified media were stored at

-20°C until required. Cell-free DNA was extracted from the thawed media using the

Qiagen circulating DNA extraction kit (Qiagen) according to the manufacturer's

instructions and quantified using Qubit High Sensitivity DNA quantification kit (Life

Technologies). DNA from both cell lines was diluted to obtain 50 ng/µl stock

concentrations. To generate the serial dilutions of one cell line with the other, we mixed

by volume to obtain the percentage (volume/volume) as presented in Table S2 (n=12).

Platinum variant calls for sample NA12878 (the virtual 'tumour') and confident regions

(high confidence homozygous reference regions plus platinum calls) [20] were

downloaded from http://www.illumina.com/platinumgenomes/. Genotype data for

sample NA11840 (the virtual 'normal') was obtained from the 1000 Genomes website. Platinum calls were intersected with our NG-TAS panel target regions and variants shared with the NA11840 sample were excluded. Five platinum calls were covered theoretically by our NG-TAS panel; however, one was targeted by one of the amplicons showing no coverage, therefore 4 SNVs were considered as identifiable 'somatic variants'.

NGS library construction

NGS libraries were prepared from 3-5 ng of cfDNA using the ThruPLEX® Plasma-seq kit (Rubicon Genomics, USA) as described in the manufacturer's instructions. NGS library was quantified using qPCR KAPA Library Quantification kit (KAPA Biosystem), while the fragment size and the NGS library yield were measured with 2200 TapeStation instrument (Agilent).

Digital PCR

BioMark system from Fluidigm has been used for dPCR, and the analyses have been performed as previously described [21]. As described in manufacturer's instructions, DNA samples were mixed with 2X TaqMan® Gene Expression Master Mix (Life Technology, 4369016), 20X GE Sample Loading Reagent (Fluidigm, 85000746) and 20X gene-specific assays. The reaction mix were loaded on the qdPCR 37K^{TM} IFC (Fluidigm, 100-6152). For *KRAS* (*G13D*) and *AKT1* (*E17K*) mutant and wild type PrimePCR^{TM} ddPCR^{TM} Mutation Assays were obtained from Bio-Rad (dHsaCP2000013 and dHsaCP2000014, dHsaCP2000032 and dHsaCP2000031 respectively). The *PIK3CA* and *ESR1* probes and primers were previously described [7,22], and the primer and probes used are listed in Table S3.

NG-TAS protocol

Primer design for NG-TAS

Primers were designed with NCBI Primer-BLAST tool with Tm range of 59-61°C. The universal primer sequences (CS1 and CS2) were added at the 5' end of the designed primers. All primer pairs were tested alone and in multiplexed PCR reactions using 10 ng of TaqMan® Control Human Genomic DNA (Thermo Fisher Scientific) in 10 μl reaction volumes. The coverage and performance of primers were analysed using 2200 TapeStation instrument (Agilent) and Hi-seq 4000. The primers were grouped together as 7-8plex, and primers in each group were chosen to target different genes in order to minimise non-specific amplification and cross-reactivity.

Access Array™ microfluidic system

The 377 pairs of optimised primers were divided into 48 wells, with each well containing

7-8 pairs of primers for multiplexed PCR. Primers were diluted to the final concentration

of 1 µM to make 20X primer solution. 4µl of the 20X primer solution from the 48 wells

were added to the primer inlets of the Access Array™ IFC (Fluidigm). For the sample

inlets, pre-sample master mix consisted of 2X Master Mix (Qiagen, 206143), 5X Q-

solution, 20X Access Array™ Loading Reagent (Fluidigm), and DNA sample were

added. The loaded IFC then moved to FC1™ Cycler (Fluidigm) for thermal cycles:

95°C for 15min, 30 cycles of 94°C for 30sec, 59°C for 90sec, 72°C for 90sec, and final

extension step 60°C for 30min. The reaction products were harvested using Post-PCR

IFC controller as described in manufacturer's instructions.

The harvested product was diluted (1:10) with water for further barcoding PCR.

Barcoding PCR reaction master mix contains 2X Master Mix (Qiagen), diluted

harvested product from Access Array™, and Access Array™ Barcode Library for

Illumina® Sequencers single direction for barcoding primers (Fluidigm, 100-4876). The

thermal cycle for barcoding is: 95°C for 10min, 15 cycles of 95°C for 15sec, 60°C for

30sec, 72°C for 1min, and final extension step of 72°C for 3min. The PCR reaction

was performed using T100™ Thermal Cycler (Bio-Rad).

Quantification and clean-up of barcode Access Array™ harvest

After barcoding PCR, all samples were analysed using 2200 TapeStation (Agilent) to

measure the concentration and size of the products (average 260bp). The PCR

products were pooled and cleaned with AMPure XP beads (Beckman Coulter, A63880)

following the manufacturer's instruction. Briefly, the samples were mixed with the

magnetic beads to the ratio of 180:100 in volume. The beads were washed twice with

80% ethanol and dried by incubating at 30°C for 10min. Then the beads were eluted

with water and the cleaned PCR product was run on the E-Gel® 2% agarose gel

(Thermo Fisher Scientific, G501802) for further size selection and extraction. The band

between 200-300bp was cut out and DNA was isolated from the gel using the QIAquick

Gel Extraction kit (Qiagen, 28704), and 10-20nM of the eluents was submitted for

paired-end Hi-seq 4000 for sequencing.

Analysis of NG-TAS data

Quality control, alignment and bam files annotation

For each sequencing lane, quality control of raw data was performed using FastQC

(http://www.bioinformatics.babraham.ac.uk/projects/fastqc/). Up to 384 samples were

multiplexed in a single sequencing lane and demultiplexing was performed using in-
house software.
Alignment, read trimming (at 80bp) and base quality recalibration was performed in a
single step using Novoalign (v 3.02). Alignment and bam metrics were computed using
Picard Tools (v 1.140). To remove potential off-target PCR products, only reads
mapped in proper pair and with insert size > 60bp were retained. After this filtering,
bam files were locally realigned using the Genome Analysis Toolkit (GATK, v 3.5).
Reads were then assigned to the amplicon they belonged to using a custom java script,
in order to enable a per amplicon coverage and mutation calling analysis. Coverage
was computed for each amplicon in each sample using a custom java/R script. One
amplicon (SF3B1_D0069_001) showed an extremely high rate of mismatches and
indels in all the analysed samples, therefore we excluded it from downstream analyses.
13
<u>Mutation calling</u>
Mutation calling was run separately for each amplicon in the panel. The core mutation
calling was performed for each pair of plasma and normal samples (or NA12878 an
NA11849 from the dilution series) using Mutect2 (included in GATK 3.5). The
minPruning parameter was set at 5 to reduce computational time with no significant
impact on the results. Besides the set of mutations passing all internal filters, we
included those failing the following internal filters or a combination of them:
"alt_allele_in_normal",   "clustered_events",   "homologous_mapping_event",
"multi_event_alt_allele_in_normal". On this set of candidate mutations, we applied the
following filtering criteria: coverage in normal and plasma >100x, alternative allele in
normal <1% and plasma/normal VAF ratio >5. The core mutation calling was repeated
for the three replicates generated for each pair and only mutations called in at least
two replicates were retained. For this set of mutations, we run HaplotypeCaller
(included in GATK 3.5) to compute the average VAF across the three replicates and
filter out mutations with an average VAF<1% and an average plasma/normal ratio <5
(Figure 4A). An extra filter was introduced for FFPE samples, where C>T and G>A
transitions with VAF<15% were filtered out because likely to be consequence of
cytosine deamination caused by fixation.
In calling somatic mutations from set of longitudinal samples from the same patient,
we first repeated the above procedure for all samples. Then, HaplotypeCaller was run
again to estimate in all samples the coverage and VAF of each mutation called in at
least one of them. This was followed by a variant annotation step using Annovar.
Finally, results obtained for all amplicons were merged to generate a single vcf file. A

final filter was applied at group level, that is, keeping only mutations that at least in one

sample had VAF higher than 5% (Figure S4).

1 **List of abbreviations**

Next Generation Targeted Amplicon Sequencing	NG-TAS
Cell Free DNA	cfDNA
Circulating Tumour DNA	ctDNA
Digital PCR	dPCR
Targeted Amplicon Sequencing	TAm-Seq
Next Generation Sequencing	NGS
Formalin Fixed Paraffin Embedded	FFPE
Variant Allele frequency	VAF
Single Nucleotide Variant	SNV
False Positive	FP
Uracil DNA Glycosylase	UDG
Response Evaluation Criteria in Solid Tumour	RECIST
Computed Tomography	CT

Declarations

Ethics approval and consent to participate

This study was approved by the East of England - Cambridge East Research Ethics

Committee (REC reference: 14/EE/1045).

Consent for publication

Not applicable

Availability of data and material

The computational pipeline is available through GitHub (https://github.com/cclab-

brca/NGTAS_pipeline) and NA12878 dilution series data are available through the

precisionFDA portal (https://precision.fda.gov/).

Competing interests

All authors have no competing interests.

Funding

This research was supported with funding from Cancer Research UK. MG has been

supported by a Genentech research grant (CLL-010907) awarded to the Caldas

Laboratory. MC has received funding from the European Union's Horizon 2020

research and innovation program under the Marie Skłodowska-Curie grant agreement no. 660060.

Authors' contributions

MG, MC, EB, SFC and CC conceived the study; MG and EB designed primers and generated the dilution series; MG performed NG-TAS in Fluidigm Access Array™ system; RB, KB, SCL, JR, and JC conducted the clinical trial; HB, LJ and AB collected the clinical samples; MC developed the computational approach and performed the analyses; SJS contributed in the computational pipeline development; MC and MG performed data analysis; MG, MC, SFC and CC drafted the manuscript, all authors revised and approved the final manuscript.

Acknowledgements

We are grateful to Cancer Research UK and the University of Cambridge for the support. We thank the Cancer Molecular Diagnostics Lab and Cancer Research UK Cambridge Institute Core Facilities including Genomics and Bio-repository that supported this work. We thank Dr PA Edwards for the scientific advice and editing for this manuscript.

Table 1. List of genes and regions covered in the panel

Gene	Target region	Hotspot position
AKT1	Hotspot	E17 AA23 – 59 AA65 - 94
BRAF	Hotspot	V600
Her2	Hotspot	S310 AA428 - 438 AA746 – 797 AA832 – 986
HRAS	Hotspot	AA3 – 35 (G12 and G13) AA49 – 77 (Q61 and A66)
IDH2	Hotspot	AA132 - 162
KRAS	Hotspot	G12
SF3B1	Hotspot	K700
ESR1	Part of exons	Exon 8 - 10 (LBD)
SMAD4	Part of exons	Exon 8 – 12
CDH1	All exons	
CDKN1B	All exons	
FOXA1	All exons	
GATA3	All exons	
MAP2K4	All exons	
MAP3K1	All exons	
PIK3CA	All exons	
PIK3R1	All exons	
PTEN	All exons	
RUNX1	All exons	
TP53	All exons	

1   **Table 2. Linear regression analysis for different cfDNA input**
2

Input DNA	R^2	Estimated coefficient	2.5% CI	97.5% CI
50ng	0.968	1.075	1.018	1.133
10ng	0.940	1.005	0.930	1.080
5ng	0.885	0.932	0.832	1.032
Library	0.964	1.123	1.059	1.187

1 **Table 3: Comparison of different amplicon based sequencing**
2 **approaches for cfDNA**
3

	NG-TAS	Digital PCR	TAm-Seq	Oncomine[TM] Breast cfDNA Assay
Approx. cost per sample (GBP)	3x10	2-3	10	200
Time (96 reactions)	7h	3h x 2	2 days	2 days
Number of amplicons	377	1	48	n/a
Amplicon size	<160bp	n/a	<200bp	<170bp
Number of genes	20	1 (hotspot)	6	10
Choice of targets	Flexible	Limited	Flexible	Limited
Median depth	3064	770	650 (avg)	n/a
Limit of detection	1-2%	>0.1%	1-2%	0.6-0.1%
Library material as input	Yes	Not tested	Not tested	Not tested
Ideal input	3x10ng	2-5ng	50ng	20ng
Multiplex	8plex	No	No	n/a
Number of samples per sequence run	384	n/a	96	12
Platform	Fluidigm Access Array	Fluidigm qdPCR	Fluidigm Access Array	Ion Chef System
Sequencing	Illumina HiSeq 4000	n/a	Illumina GAIIx	Ion S5/S5XL[TM]

Figure Legends

Figure 1. NG-TAS workflow and alignment pipeline

(A) NG-TAS workflow. Primers were designed and multiplexed for direct amplification

in cfDNA obtained from plasma using Fluidigm Access Array™. The PCR products

were harvested and barcoded in a subsequent PCR reaction, the samples were pooled

and size selected for sequencing on an Illumina Hi-Seq 4000. (B) Schematic

representation of the computational pipeline for reads alignment, filtering and

annotation.

Figure 2. Optimising targeted deep sequencing by NG-TAS

(A) Percentage of aligned reads were compared in different samples where a variable

amount of input control genomic DNA was used (range 50 to 0.016ng). The effect of

pre-amplification and Q solutions are shown, red = No Q solution and no pre-

amplification step, green = With Q solution and no pre-amplification, blue = No Q

solution and with pre-amplification. (B) Density plot showing the log10 coverage values

for all primers in the 10 ng NA12878 cfDNA sample. The dotted line indicates 100x

coverage; median value for the distribution is 3064x. (C) Coverage heatmap of

individual primers for different amount of input NA12878 cfDNA. For each amount of

input DNA the analysis was performed in triplicate. (D) Distribution of all non-reference

base frequencies across all target regions in the NA12878 dilution series in (C).

Figure 3. Detection of SNVs in NA12878 cfDNA dilution series

(A) Expected versus observed VAF for 4 SNVs in the NA12878-NA11840 dilution

series starting from 50 ng input DNA (left) and zoom-in for expected VAF <5% (right).

(B) Expected versus observed VAF for 4 SNVs in the NA12878-NA11840 dilution

series starting from 10 ng input DNA (left) and zoom-in for expected VAF <5% (right).

(C) Expected versus observed VAF for 4 SNVs in the NA12878-NA11840 dilution

series starting from 5 ng input DNA (left) and zoom-in for expected VAF <5% (right).

(D) Expected versus observed VAF for 4 SNVs in the NA12878-NA11840 dilution

series starting from post-NGS library input DNA (left) and zoom-in for expected VAF

<5% (right).

Figure 4. Mutation calling in NA12878 cfDNA dilution series

(A) Schematic overview of the computational pipeline to identify somatic mutations in

NG-TAS data. (B) *De novo* mutation calling in the NA12878 dilution series was

evaluated for different amounts of input cfDNA. 4 SNVs can potentially be called using

our panel of 377 amplicons. (C) VAF for all FP calls in the NA12878 dilution series.
The red dashed line represents 5% VAF.
3
Figure 5. Validation of NG-TAS performance in clinical plasma samples
(A) The specificity of NG-TAS in clinical samples was estimated using 4 pairs of buffy
coats from the same patients (A, B, C and D). The mutation calling pipeline was applied
using one buffy coat as normal and the other as 'tumour' and vice versa. All mutations
called in this setting can be considered FPs. The red line indicates 5% VAF. (B)
Oncoprint summary plot of genes mutated in 24 cases for which both tissue and
plasma samples were tested. The vertical black line separates cases for which the
primary tumour was analysed from cases for which a metastasis biopsy was analysed.
(B-C) Comparison of VAF obtained by NG-TAS and dPCR. (C) In this comparison, four
different hotspot mutations including *AKT1* (*E17K*), *ESR1* (*D538G*), *KRAS* (*G13D*) and
PIK3CA (*H1047R*) identified in multiple plasma samples from 4 distinct patients were
analysed (R^2 = 0.64). (D) Two *PIK3CA* hotspots (*H1047R* and *E545K*) were detected
by NG-TAS using NGS library as input material in plasma samples from two distinct
patients. The same mutations were detected using dPCR and a good correlation was
found (R^2 = 0.80).
19
Figure 6. Monitoring response in metastatic breast cancer patients using NG-
TAS
(A) Example of patient monitoring during treatment using direct NG-TAS in ctDNA.
There are four mutations detected in more than one sample: *GATA3* (*F431fs*), *PIK3CA*
(*E542K*), *CDKN1B (N124fs)* and *PTEN* (*137-139del*). The mutations called more than
once in the longitudinal samples are shown including tumour and plasma samples.
The arrow indicates the time of the disease considered as RECIST progressive
disease. T indicates tumour samples, and SP indicates screening plasma sample
which was collected prior to the treatment. (B) Example of patient monitoring during
treatment using NGS library material for NG-TAS. This patient had a stable disease
during the whole treatment period. There are three mutations detected, including
MAP3K1 (E303 frame shift), hotspot mutations *PIK3CA (E545K)* and *TP53 (R141H*
and *P46fs*). T indicates tumour samples.
33
34

Supplementary Figures

Figure S1 - Representative image of the Bioanalyser gel plot. The 8plex PCR products

were analysed using Bioanalyser for primer efficiency and quality control.

Figure S2 - Fragment size distribution according to the Bioanalyser results for cfDNA

extracted from the media where NA12878 cells were grown (main peak at around 160-

170bp).

Figure S3 – (A) Percentage of amplicons having more than 100x coverage for 2, 5

and 10 ng of input cfDNA from NA12878 sample. (B) Percentage of reads on target for

2, 5 and 10 ng of input cfDNA from NA12878 sample.

12
13   **Figure S4** – Schematic overview of the computational pipeline to identify somatic

14   mutations in NG-TAS data from longitudinal samples.

16   **Figure S5** – Detailed representation of mutations identified in tumour or plasma

17   samples of 21 metastatic breast cancer cases. The colour gradient indicates the VAF

18   as indicated; PT = primary tumour, M = metastasis biopsy, V1…n = plasma.

REFERENCES

[1] Leon SA, Shapiro B, Sklaroff DM, Leon SA, Shapiro B, Sklaroff DM, et al. Free DNA in the Serum of Cancer Patients and the Effect of Therapy Free DNA in the Serum of Cancer Patients and the Effect of Therapy. Cancer Res 1977;37:646–50.

[2] Vasioukhin V, Anker P, Maurice P, Lyautey J, Lederrey C, Stroun M. Point mutations of the N-ras gene in the blood plasma DNA of patients with myelodysplastic syndrome or acute myelogenous leukaemia. Br J Haematol 1994;86:774–9. doi:10.1111/j.1365-2141.1994.tb04828.x.

[3] Sorenson GD, Pribish DM, Valone FH, Memoli V a, Bzik DJ, Yao SL. Soluble Normal and Mutated Dna-Sequences From Single-Copy Genes in Human Blood. Cancer Epidemiol Biomarkers Prev 1994;3:67–71.

[4] Murtaza M, Dawson S-J, Tsui DWY, Gale D, Forshew T, Piskorz AM, et al. Non-invasive analysis of acquired resistance to cancer therapy by sequencing of plasma DNA. Nature 2013;497:108–12. doi:10.1038/nature12065.

[5] Diehl F, Schmidt K, Choti M a, Romans K, Goodman S, Li M, et al. Circulating mutant DNA to assess tumor dynamics. Nat Med 2008;14:985–90. doi:10.1038/nm.1789.

[6] Schwarzenbach H, Hoon DS. B, Pantel K. Cell-free nucleic acids as biomarkers in cancer patients. Nat Rev Cancer 2011;11:426–37. doi:10.1038/nrc3066.

[7] Dawson S-J, Tsui DWY, Murtaza M, Biggs H, Rueda OM, Chin S-F, et al. Analysis of circulating tumor DNA to monitor metastatic breast cancer. N Engl J Med 2013;368:1199–209.

[8] Wan JCM, Massie C, Garcia-corbacho J, James D, Caldas C, Pacey S, et al. Liquid biopsies come of age: clinical applications of circulating tumour DNA. Nat Rev Cancer 2016. doi:10.1038/nrc.2017.7.

[9] Vogelstein B, Kinzler KW. Digital PCR. Proc Natl Acad Sci U S A 1999;96:9236–41. doi:10.1073/pnas.96.16.9236.

[10] Dressman D, Yan H, Traverso G, Kinzler KW, Vogelstein B. Transforming single DNA molecules into fluorescent magnetic particles for detection and enumeration of genetic variations. Proc Natl Acad Sci U S A 2003;100:8817–22. doi:10.1073/pnas.1133470100.

[11] Hughesman CB, Lu XJD, Liu KYP, Zhu Y, Poh CF, Haynes C. Robust protocol for using multiplexed droplet digital PCR to quantify somatic copy number alterations in clinical tissue specimens. PLoS One 2016;11:1–22. doi:10.1371/journal.pone.0161274.

[12] Forshew T, Murtaza M, Parkinson C, Gale D, Tsui DWY, Kaper F, et al. Noninvasive Identification and Monitoring of Cancer Mutations by Targeted Deep Sequencing of Plasma DNA. Sci Transl Med 2012;4:136ra68-136ra68. doi:10.1126/scitranslmed.3003726.

[13] Torga G, Pienta KJ. Patient-Paired Sample Congruence Between 2 Commercial Liquid Biopsy Tests. JAMA Oncol 2017. doi:10.1001/jamaoncol.2017.4027.

[14] Kinde I, Wu J, Papadopoulos N, Kinzler KW, Vogelstein B. Detection and quantification of rare mutations with massively parallel sequencing. Proc Natl Acad Sci 2011;108:9530–5. doi:10.1073/pnas.1105422108.

[15] Callari M, Sammut S, Mattos-arruda L De, Bruna A, Rueda OM, Chin S-F, et al. Intersect-then-combine approach : improving the performance of somatic variant calling in whole exome sequencing data using multiple aligners and callers 2017:1–11. doi:10.1186/s13073-017-0425-1.

[16] Cerami E, Gao J, Dogrusoz U, Gross BE, Sumer SO, Arman B, et al. In Focus The cBio Cancer Genomics Portal : An Open Platform for Exploring Multidimensional Cancer Genomics Data 2012. doi:10.1158/2159-8290.CD-

12-0095.

[17] Gao J, Aksoy BA, Dogrusoz U, Dresdner G, Gross B, Sumer SO, et al. Integrative Analysis of Complex Cancer Genomics and Clinical Profiles Using the cBioPortal Complementary Data Sources and Analysis Options 2013;6:1–20.

[18] Sims D, Sudbery I, Ilott NE, Heger A, Ponting CP. Sequencing depth and coverage: key considerations in genomic analyses. Nat Rev Genet 2014;15:121–32. doi:10.1038/nrg3642.

[19] Kou R, Lam H, Duan H, Ye L, Jongkam N, Chen W. Benefits and Challenges with Applying Unique Molecular Identifiers in Next Generation Sequencing to Detect Low Frequency Mutations 2016:1–15. doi:10.5061/dryad.n6068.Funding.

[20] Eberle MA, Fritzilas E, Krusche P, Källberg M, Moore BL, Bekritsky MA. A reference dataset of 5.4 million human variants validated by genetic inheritance from sequencing a three-generation 17-member pedigree. Genome Res 2017;27. doi:10.1101/gr.210500.116.

[21] Yung TKF, Chan KCA, Mok TSK, Tong J, To K-F, Lo YMD. Single-molecule detection of epidermal growth factor receptor mutations in plasma by microfluidics digital PCR in non-small cell lung cancer patients. Clin Cancer Res 2009;15:2076–84. doi:10.1158/1078-0432.CCR-08-2622.

[22] Schiavon G, Hrebien S, Garcia-Murillas I, Cutts RJ, Pearson a., Tarazona N, et al. Analysis of ESR1 mutation in circulating tumor DNA demonstrates evolution during therapy for metastatic breast cancer. Sci Transl Med 2015;7:313ra182-313ra182.

Figure 1

A

Primers design

Cell-free DNA from plasma

Direct amplification
Multiplex Access Array

Barcoding PCR

Sample pooling and size selection

Illumina sequencing

B

FastQ file

Novoalign
 -l=30
 -t=300
 -n=80
 -k
 -o=FullNW

**Picard
(ColllectAlignmentS
ummaryMetrics)**

**Picard
(CalculateHsMetrics)**

Bam filtering
- reads mapped in proper pair
- insert size >60

GATK Local realignment

Bam annotation
- assign amplicon to each read
- tollerance=3

Amplistats
- Compute per amplicon, per base coverage

Annotated bam file

Figure 2

A

B

C

D

Figure 3

Figure 4

A

Per amplicon:

Per replicate:

- **Mutect2 + filtering**
  - *minPruning=5*
  - *min_N_cov=100*
  - *min_P_cov=100*
  - *max_N_VAF=1%*
  - *P/N_ratio>5*

Keep mutations called in 2/3 replicates

- **Haplotypecaller**
  (in the 3 replicates P and N)
- **Filtering**
  - *max_avg_N_VAF=1%*
  - *min_avg_P_VAF=1%*
  - *Avg_P/N_ratio>5*

Mutations called in all amplicons

Variant annotation (Annovar)

B

C

Figure 5

Figure 6

A Direct NGTAS

B Library - NGTAS