<div style="background:#1a3a6b;color:white;padding:4px;">

METHODOLOGY

</div>

A fuzzy-registration approach to track cell divisions in time-lapse fluorescence microscopy

Saoirse Amarteifio[1*], Todd Fallesen[2], Gunnar Pruessner[1] and Giovanni Sena[2]

*Correspondence:
s.amarteifio14@imperial.ac.uk
[1]Department of Mathematics,
Imperial College London,
Kensington Campus, London, UK
Full list of author information is
available at the end of the article

Abstract

Background: Particle-tracking in 3D is an indispensable computational tool to extract critical information on dynamical processes from raw time-lapse imaging. This is particularly true with *in vivo* time-lapse fluorescence imaging in cell and developmental biology, where complex dynamics are observed at high temporal resolution. Common tracking algorithms used with time-lapse data in fluorescence microscopy typically assume a continuous signal where background, landscape keypoints and independently moving objects of interest are permanently visible. Under these conditions, simple registration and identity management algorithms can track the objects of interest over time. In contrast, here we consider the case of *transient* signals and objects whose movements are constrained within a tissue, where standard algorithms fail to provide robust tracking.

Results: To optimize 3D tracking in these conditions, we propose the merging of registration and tracking tasks into a fuzzy registration algorithm to solve the identity management problem. We describe the design and application of such an algorithm, illustrated in the domain of plant biology and make it available as an open source software implementation. The algorithm is tested on mitotic events in 4D datasets obtained with light-sheet fluorescence microscopy on growing *Arabidopsis thaliana* roots expressing CYCB::GFP. We validate the method by comparing the algorithm performance against both surrogate data and manual tracking.

Conclusion: This method fills a gap in existing tracking techniques, following mitotic events in challenging data-sets using transient fluorescent markers in un-registered images.

Keywords: Plant development; Plant root; Arabidopsis; Light-sheet microscopy; CYCB::GFP; Transient fluorescence; 3D Tracking; Point-set registration

1 Background

It is now generally acknowledged that automated computer vision and tracking methods can match or exceed manual tracking of objects in large data-sets. In life science applications, where time-lapse microscopy is a powerful and popular tool for capturing dynamics at cellular and sub-cellular scales [1], these methods are of growing importance. Indeed, a wide range of microscopy methods [2] and computational approaches [3, 4, 5, 6] have been developed and adapted to particle tracking in biology. More specifically, developmental biology has benefit enormously from automated *in vivo* tracking methods, ranging from cell lineage tracing in both animals [7, 8, 9] and plants [10, 11, 12], to cell shape tracking [13, 14] and even tracing of growing organs as in the case of plant roots [11, 15, 16, 17, 18].

It should be noted that in all these cases the signal to track is pretty much continuous in time, which generally makes the task easier notwithstanding occlusions or objects leaving and entering the field of view. Instead, when the signal is not continuous but intermittent or transient, as in the case of cell divisions or other short-lived events, tracking becomes a much more challenging task [19, 20, 21]. Moreover, tracking is always significantly more difficult with a 3D data-set, like in confocal or light-sheet microscopy stacks, compared to more common 2D data.

From a computational point of view, tracking in time-lapse image data involves detection of objects in individual frames and solving the data association or identity management problem between frames. The goal of identity management is to link an object in one frame to what is found to be the same object in a later frame, so that linked objects share the same identity label, or *identifier*. The set of linked identifications of a given object through time is referred to as its *lineage*.

Tracking algorithms are greatly influenced by the nature of the data. For example, macroscopic objects may have complex morphologies allowing objects to be distinguished on high-level image features, which can enhance tracking [22]. Objects smaller than imaging resolution, on the other hand, will appear as featureless blobs, so that in these cases tracking relies more strongly on dynamical models alone [23]. For example, multiple hypothesis tracking [24, 25] takes an exhaustive, deterministic approach to consider possible lineage trees. A similar approach has been previously applied to tracking plant root cell nuclei in 3D [11]. Stochastic algorithms have also been proposed [26], specifically to deal with particularly noisy datasets. Whenever persistent features, or reference points or *keypoints*, can be identified in the background, a common strategy is to take advantage of this extra information and to link the same keypoint in two frames [27, 28, 29, 30]. In this case, keypoints appearing in more than one frame are defined as *inliers*, while those appearing in only one frame are treated as *outliers*. Given a pairs of *inliers*, an affine transformation (*i.e.* a generalised transformation including translation, rotation or scaling) that takes one set of points to the other can be determined and used as hypothesis for tracking the objects of interest in the foreground.

Unfortunately, in many time-lapse data sets in life sciences, and specifically in plant developmental biology where soft tissue rarely offers usable reference points, tracking transient events such as cell divisions in 3D is still challenging and imprecise.

Here, we offer a novel solution to the problem of tracking short-lived events distributed in 3D and in absence of background *keypoints*. To count transient fluorescent markers we effectively exchange the common approach of segmentation with morphological analysis [20] for a fuzzy registration-tracking problem. The word *fuzzy* indicates the parametric *random* sampling point constellation approach used. As a case-study, we generated and analysed a 4D collection of cell division events detected in a single *Arabidopsis* root meristem throughout several consecutive days. This was achieved with time-lapse 3D scanning of growing transgenic roots expressing the fluorescent reporter CYCB::GFP , through a previously described light-sheet microscope setup [16]. The mitotic events, or *objects* from a computational point of view, are short-lived, sparsely distributed in a relatively soft tissue lacking fluorescent reference keypoints, and generally difficult to identify due to low signal-to-noise

60 ratio. A very common set of circumstances in time-lapse imaging of plant tissues.
61 Due to its general strategy and features, we believe that the methodology proposed
62 might be readily applied to similar 4D datasets collected from other tissues.

2 Results

2.1 Tracking Algorithm

2.1.1 General strategy

The identification of inliers in the experimental dataset described above is the key
challenge that we have tried to address, and should be understood as sitting between
a single particle tracking problem and a registration one. Single particle tracking
solutions may or may not require pre-registration. In the case that they do require
pre-registration, keypoint registration solutions may not be robust against lack of
guaranteed inliers. In the case where single-particle tracking methods are robust in
the absence of pre-registration, they typically emphasise individual object motion
models, which may not be optimal under certain data conditions *i.e.* transient events
in rigid structures and low signal-to-noise ratio across extended images sequences.

We treated the position, appearance and disappearance of transient objects as
a random spatial process. The frame-to-frame displacement of objects can be de-
scribed as a sum of two random variables: (1) large-scale movements of the tissue
within the field of view, due for example to growth of the structure itself or the
inability of the microscope to focus on a fixed point of the root tip over time; (2)
small-scale fluctuations of the objects of interests within the tissue.

The main purpose of the algorithm is to find the large-scale movement, in spite
of the noisy, small-scale dynamics. The strategy adopted was to find point set
correspondences between the random process at time t and the random process
at time $t - \tau$ where τ is a lag variable. Since here we consider the plant tissue as
"rigid" (a good approximation for small lags), in practice we only consider $\tau = 1$.
Objects that appear in both frames are called *inliers* and those that do not appear
in both frames are called *outliers* and cannot be linked by a global frame-frame
transformation. For example, outliers can be either debris, *i.e.* any object which is
determined *not* to be an object of interest, or could be an object of interest that
has just "exited" or "entered" the new frame. Any object outside the bounding box
will also be an outlier.

There are two main steps in the execution of the algorithm: (i) candidate trans-
formations are generated from the data using one or more strategies and (ii) among
these candidates, the optimal transformation is chosen with respect to some objec-
tive. Inliers are then mapped to each other by the chosen global transformation and,
in the linkage phase, identifiers (object labels) are propagated between the inliers in
each frame. Finally, new identifiers are generated for the outliers within the region
of interest.

Algorithm 1 provides a terse, high-level overview of the method. The key algorithm
stages, which amount to finding the best global transformation from the data, are
discussed below.

2.1.2 Proposing transformations

103 In the understanding that some objects are outliers, we wish to find the best global
104 transformation that explains most of the point's correspondence. Consider sets of

---

Algorithm 1 Transient point cloud tracking algorithm

High-level description of algorithm using 1-indexing
*Standard geometry or data structure functions are *lower camel case**
Other FUNCTIONS are explained in the main text
*Key tensor *sets* in **bold font** shorthand are understood from function calls*

1: $results := \{\}$
2: **for** $i \leftarrow lag + 1, frameCount$ **do**
3:     $U := frames(i)$
4:     $V := frames(i - lag)$
5:     $\mathbf{uv} := cartesianPointProduct(U, V)$
6:     $\mathbf{tr1} := translationsFor(\mathbf{uv})$ ▷ For points, k=1
7:     $ranked := \text{RANK}(\mathbf{tr1}, U, V, epsilon)$
8:     $linkages := \text{APPLY}(ranked, U, V, epsilon)$
9:     $\mathbf{kuv} := \text{CONSTELLATIONS}(U, V, linkages, k)$ ▷ constellation pairs
10:     $\mathbf{trk} := transformsFor(\mathbf{kuv})$
11:     $other := \text{PRIORS}(linkages, U, V)$
12:     $transforms := concatenate(\mathbf{tr1}, \mathbf{trk}, other)$

13:     $ranked := \text{RANK}(transforms, U, V, epsilon)$
14:     $result := \text{APPLY}(ranked, U, V, epsilon)$
15:     $results := concatenate(results, result)$

16: $results \leftarrow \text{MINFRAMEDETECTIONFILTERING}(result, minFrames)$
17: $results \leftarrow \text{ROIBELIEFFILTERING}(result)$
18: $results \leftarrow \text{ANISOTROPICCATCHMENTFILTERING}(result)$

---

points $U(t), V(t - \tau)$ at times offset by a lag τ. A subset of these points corresponds to true objects of interest (*i.e.* excluding noise and debris) and a further subset of these exists in both frames (due to short-lived events). There are N points at time t and M points at time $t - \tau$. A number of permutations $\mathcal{P}(N, k)$ are sampled from U and a number of permutations $\mathcal{P}(M, k)$ are sampled from V. In our method $k \in \{1, 3\}$, thus sampling either single points or triangular *constellations* from each frame. Points take a natural order and the same points can only appear in one constellation.

Let n be the number of k-constellations $c_i(t)$ sampled from $U(t)$ and m be the number of k-constellations $c_j(t')$ sampled from $V(t - \tau)$. In general, for two tensors $\mathbf{M}_1$ and $\mathbf{M}_2$ related by $\mathbf{M}_1 = \mathbf{A}\mathbf{M}_2$, the affine transformation $\mathbf{A}$ can be uniquely determined through a least squares method (this is, in general, over-determined). Here, the transformation γ_{ij} is the affine transformation, taking the point set $c_i(t)$ to the point set $c_j(t')$: $c_j(t') = \sum \gamma_{ij} c_i(t)$

On lines 3-6 of Algorithm 1 the constellation are sampled in frames U and V and translations are computed. Translations are then generated for each pair of points and the translation taking a point in $v \in V$ to a point $u \in U$ is added to a list of candidate transformations. For comparison with other stages, we think of points as *k-constellations* with $k = 1$.

Constellation sampling (line 9) was preferably seeded with linkages found by translations so that likely outliers are excluded from consideration. If a given translation maps n points in U to n points in V, then these point-point pairings can be considered candidate "linkages". This not only reduces the candidate points to those likely inlier points but constellation congruences can be identified as well, so the same set of points in one frame can be paired to the same set of points in the other frame when proposing transformations. Constellation pairs $\mathcal{P}(N_L, 3), \mathcal{P}(M_L, 3)$ can then be sampled from the L inlier objects that appear to be in both frames. These

Fig 1: Evaluating transformations: Red crosses correspond to the set $V(t-\tau)$ while the blue dots correspond to objects at the current time $U(t)$. The blue circles with radius ε correspond to the transformation of the set $U(t)$ under the proposal transformation. The transformation score is the average capped distance to nearest neighbours.

can then be used to generate proposal affine transformations, to provide a more optimal fit than translations alone.

2.1.3 Ranking transformations

Transformation ranking was carried out using a modified least squares loss function. The least-squares-loss objective is generally described as the minimization of $\sum |\mathbf{y} - \hat{\mathbf{y}}|^2$, where $\mathbf{y}$ is a proposal vector and $\hat{\mathbf{y}}$ is the target vector. In the registration problem, the proposal vectors correspond to the points in $U(t)$ under transformations from the set Γ, which is the set of all proposed transformations. The target vectors are points in $V(t-\tau)$. The least-squares-loss objective was modified based on the following prior: (i) we expect a global rigid transformation to explain the movement of object centroids up to small fluctuations due to individual object movement which are deemed negligible. (ii) A distance $> \varepsilon$ between objects' centroids is assumed. The value of ε in the context of the sample data corresponds to a value slightly larger than the average blob's radius. This hyperparameter ε was determined from the data. When considering distances between proposal points and target points, only distances to the first nearest neighbour within a radial distance ε were considered. Each candidate transformation γ was applied to all points $u_i(t) \in U(t), i \in \{1, 2, ..., N\}$. Let $\tilde{u}_i = \gamma(u_i)$. We used a cost function

$$C_\gamma = \sum_{i=1}^{N} \min(\varepsilon, |\tilde{u}_i - nn_V(\tilde{u}_i)|^2) \tag{1}$$

where $nn_V(\tilde{U}_i)$ is the nearest neighbour position in $V(t-\tau)$ to the projected point $\tilde{U}_i$. If the projected point had no nearest neighbour within a ball of radius ε, the capped distance $\varepsilon + 1$ was attributed. This is illustrated in Fig. 1.

The cost function will rank transformations by how well they explain the movement of the majority of points. If there are small differences in scores, we may prefer transformations that have the smallest displacement. For this reason, we can *regularise* this cost function to add a penalty to larger displacements.

2.2 Output: Life Matrix

The main output of the algorithm is a binary matrix called the Life Matrix, where each column is a time-frame, each row contains the lifetime of a single detected object, and each element is equal to 1 if an object is detected and to 0 otherwise. Debris exhibit movements that are not correlated with that of the objects of interest and therefore would appear in the Life Matrix as events with a life-span of a single frame. We discarded such single-frame events.

The algorithm was applied on time-lapse images of cell division events in an *Arabidopsis* root, collected over a few days through light-sheet microscopy on a transgenic line expressing the fluorescent reporter CYCB::GFP, following previous methods [16]. The correspondent Life Matrix is shown in Fig. 2.

Fig 2: Life matrix: Rows (y-axis) correspond to detected individuals with time on the x-axis. The life matrix shows a generally consistent trend in the appearance and persistence of objects. At the beginning of the sequence, objects persist for an abnormally long time for biological/experimental reasons.

The number of objects observed as a function of time (Fig. 3, top) and the average duration of events ("mean age", in number of frames) as a function of time (Fig. 3, bottom) are alternative representations of the data shown in the life matrix.

2.3 Validation against surrogate data

We used computationally generated (surrogate) data to validate the tracking strategy against a known spatial transformation. Surrogate data must be representative of the biological images intended to be tracked, and were generated respecting a minimum object separation ε. A number of parameters were introduced (Table 1) to determine how data are generated and the tracker was tested for different points in the parameter space. The parameter n determines how many objects are randomly and uniformly distributed and added to the set of points U. A random

Fig 3: The population time series reflect the information shown in the life matrix. Top panel, number of events as a function of time; bottom panel, average event's duration (in numbers of frames) as a function of time. The high spike in the mean age is a result of one object surviving for a long time while others disappear. After this point we observe more typical fluctuations in object activity.

Euclidean transformation (roto-translation) $\widehat{\gamma}$ is generated using the parameters r and s to control the magnitudes of rotation and translation, respectively. A set V is produced by applying $\widehat{\gamma}$ to U. The parameter u determines the strength of (small) random fluctuations to individual object locations in V, and p determines how many "noise" points are added to both U and V to represent debris or particles that exist in one but not the other frame. These added points also respect the minimum object separation ε. In the biological context this corresponds to cells that are no longer fluorescing or just beginning to fluoresce at time t.

Param.	Description
ε	Minimum object separation
n	Number of points to generate in the point-set U
p	Number of "noise" points to add to the point-set U and the point-set V
r	Generate random (global) rotations in the range $[-\pi/r,\,\pi/r]$
s	Generate random (global) translations in the range $[0, s]$
u	Random uniform shift of points in V on each axis in the range $[0, u]$

Table 1: *Surrogate data parameters:* A random Euclidean transformation determined by parameters (r, s, u) is applied to a set of points U to produce a new set of points V. To simulate debris or particle births and deaths, p points are added at random locations to both U and V.

The tracking objective is to find the transformation $\widehat{\gamma}$ from U to V. The tracking algorithm's ability to do this will depend on the amount of noise points and the extent of the magnitude of angle or displacement of the Euclidean transformation. To test the tracker's ability to find $\widehat{\gamma}$ we establish a classification task. Under the correct global transformation, correct identifiers (with respect to ground truth) should be assigned to inliers and the identifier value -1 is assigned to outliers. The *loss*, which we aim to minimise, was averaged over samples and corresponds to the number of objects in V that do not have the correct identifier. These results illustrate that for small rotations (approx 5 degrees) or for small translations the

tracker is robust and slowly degrades for larger rotations (approx 15 degrees), as shown in Fig 4.

Fig 4: *Surrogate data results:* the heatmap plot gives scores for different values of the parameters s and r with the loss given in the colour scale. The best score (loss) is 0 and appears black.

2.4 Validation against manual tracking

The automated detection by the algorithm was verified by manually marking the position of all GFP-expressing cells in a subset of 50 time-points by an experienced researcher without prior knowledge of the results of the algorithm output. Using the built-in FIJI plugin *CellCounter* the researcher manually identified 400 CYCB::GFP events which were present in at least 2 consecutive time-points, that is, using the same bias against single time-point events as the algorithm. In the same subset, the algorithm automatically detected 385 events. Given the 15-minute temporal resolution of our data-set and the longer persistence of the CYCB::GFP reporter, we reasoned that these should be considered either spurious events or plain artefacts. The rate of false negatives, *i.e.* events detected by the researcher but not by the algorithm was 16.5 %, while the rate of false positives *i.e.* events detected by the algorithm but not by the researcher was 12.8 %. It should be noted that, given the fact that the goal is to extract statistical information from a sample of existing events, it is generally preferable to minimize the false positives at the expense of the false negatives. An analogous setup based on light-sheet microscopy had been previously used with *Arabidopsis* roots constitutively expressing the nuclear-localised, fluorescent fusion protein H2B::YFP [11]. In that study, the time-lapse imagine was performed with a temporal resolution comparable to the data presented here, but cell division events were identified as branching points on nuclei trajectories through a completely independent and competing approach. As a result, the nuclei trajectory approach was validated against manual tracking with 26 % false negatives and 13 % false positives [11], indicating that the method presented here is preferable when detecting cell divisions at high temporal resolution.

2.5 Validation with lag test

A second, independent, validation of our method was performed with a "lag test". We counted the number of objects where identifiers are in agreement between dif-

Fig 5: Dis/agreement between lags: Lag-1 identifiers for objects of interest (excluding single-frame detections) are compared with identifiers assigned by $\lambda_\tau : \tau \in \{2, 3\}$ transformations for objects that are older than τ. We counted 734 objects in a sample sequence. 42 $\approx 6\%$ showed lag1-2 disagreements and excluding these, 29 $\approx 4\%$ showed lag1-3 disagreements in specific frames.

ferent lags τ. For example we may expect the following equivalence

$$(t-3) \xleftarrow{\lambda_1} (t-2) \xleftarrow{\lambda_1} (t-1) \xleftarrow{\lambda_1} t \equiv (t-3) \xleftarrow{\lambda_3} t \tag{2}$$

223 where, with an abuse of notation, λ_τ corresponds to propagation of object identifiers
224 using highest-scoring transformations at lag τ. Figure 5 shows presents the results
225 of this test.

2.6 Post-processing

Data filtering can be applied in a post-processing stage (Fig. 6 and Fig. 7). The
ellipse shows the region of interest (ROI) containing the most likely mitotic events
occurring in the root tip. The time colouring Fig. 6 illustrates when and where
events occur and indicates (roughly) if plotted points correspond to the same ob-
ject. In Fig. 7, large angle differences between object displacement vectors and a
global transformation vector appear in lighter colours (*e.g.* red) and are indicative
of possible inlier mis-classification. Objects outside of (or far from) the ellipse and
objects that show large angle differences are candidates for removal. In the case
of angle differences, actual displacements and ε values can be taken into account
when filtering objects. For example, in the final result output from the tracker, we
could filter objects out of our final analysis under the following conditions: we find
that (i) the local displacement of an object away from the location predicted by the
global transformation is above a certain threshold and (ii) the object is moving in
a direction that is sufficiently different to the rest of the object population.

Fluctuations of the ROI area across frames are an indicator for how well the image
processing step can identify the region of mitotic activity in the sample. When the
signal-to-noise ratio is good, this region can be accurately detected and its relatively
small area varies smoothly between frames. On the other hand, if there is too much
noise or few (biological) events, the identified ROI could become relatively large
and vary non-smoothly from frame to frame. In the later case, debris may lie within
an excessively large ROI and be picked up by the tracker. In Fig. 6, this debris
appears for example in the top part of the images, particularly towards the later
stages of the experiment due to drift in microscope calibration or other factors. The
root grows in the positive x direction (left to right) following the gravity vector,

Time-lapse with time(frames) in colour bar

Fig 6: Time-lapse with colouring used to distinguish time. The ellipse represents the (Gaussian) statistical region of interest over many frames.

so precipitating debris appear moving in the same direction of root growth, while static marks or defects in the imaging chamber appear as moving in the opposite direction.

3 Methods

3.1 Data acquisition

To visualize mitotic events, we used the established *Arabidopsis* transgenic line expressing a fusion between the cyclin protein CYCB1;1 and the fluorescent protein GFP, driven by the endogenous CYCB1;1 promoter [31]. The fluorescent reporter CYCB1;1::GFP accumulates in a cell transitioning between G2 and M phases of the cell cycle, and is quickly degraded after entering mitosis [31] and for this reasons it is widely adopted as a reliable live marker for mitotic events. *Arabidopsis* seeds were sterilised and stratified before sowing and germination on solid MS plates according to standard protocols [32]. Plants were imaged 5 days post-germination. A single primary root was grown and imaged on a custom-made light-sheet microscope setup, as previously described [16]. In essence, the root was hydroponically grown in a perfusion 5ml cuvette maintained under constant light and temperature, with its liquid medium fully exchanged every 2 minutes. A full 3D scan in fluorescence of the root tip generates a stack of 60 optical sections 4μm apart, captured every 15 minutes for up to 7 days. In our study, we accurately tracked and counted such division events in a temporal series of 3D stacks taken from a single root, as described.

3.2 Preprocessing

3.2.1 Noise analysis

Noise and histogram levels were used on normalised grey-scale images taking pixel values in $[0, 1]$. The noise shown in the top panel of Fig. 8b was estimated using

Time-lapse with global angle displacement (degrees) in colour bar

Fig 7: Time-lapse plotswith colouring are used to distinguish angle differences between the object displacement vector and the global transformation vector. The ellipse represents the (Gaussian) statistical region of interest over many frames.

a wavelet-based estimator of the Gaussian noise standard deviation[33, 34]. The middle panel in Fig. 8b describes the signal. This is a simple quasi-signal metric defined as a ratio between the 99^{th} and 95^{th} histogram percentile boundary values, which are determined empirically for our data to separate background from signal. If this value approaches 0 we find there is low signal/information. For values of this measure above a calibrated threshold, we find the data are more concentrated outside the 99^{th} percentile of the data, corresponding to bright singularities in the image. A noise range was used to guide downstream strategies in the image processing pipeline. An empirically determined "normal" noise band $[0.01, 0.045]$ was used in the sample parameters discussed here. If the noise is very low *e.g.* below 0.01, no denoising is required otherwise wavelet denoising will be applied. Excessive noise, exceeding 0.045, suggests light saturation and/or low signal-to-noise ratio typically due to a low number of biological events. In such instances additional thresholding is applied to the data before wavelet denoising. If the noise exceeds 0.1, the signal-to-noise ratio is so low that the frame is marked as degenerate. Fig. 8a contrasts different noise levels in sample images and Fig. 8b plots image properties over time for a given experiment.

3.2.2 Region of interest

Bright blobs are scattered across the full image and may correspond to mitotic events within the tissue or debris floating beyond it. For tracking accuracy and tracking speed, it is advisable to select a *region of interest* (ROI).

A 2D projection of the data was obtained by summing the 3D tensor along the z-axis. An adaptive threshold based on image histogram percentile ranges was used to construct a narrow-band filter for the 2D data. This range can vary erratically

between frames as shown in the bottom time series in Fig. 8b. A thin slice of the image data range (shaded band) was selected and an aggressive "Gaussian smoothing" (averaging neighbouring pixel values using a Gaussian kernel) with an empirically-determined sigma value ($\sigma = 8$) was used to find a mesh-like connected component corresponding to the region of activity in the root tip. The 2D mask of this largest component was extended to a 3D mask by projecting the 2D region back into the z-plane. The largest connected component in the thresholded image was identified as the root tip. These stages are illustrated in Fig. 9. The image processing pipeline continues to process data only within the ROI.

For cases where the region of activity is well-isolated (smaller region of interest), the processing is more efficient as the amount of volumetric data is reduced and debris beyond the root are filtered from the downstream pipeline. For our sample data, an ROI with area $\approx 10^5$ pixels corresponds to a well-isolated root tip. Noise and light saturation can affect this part of the process making the ROI area a proxy for image quality. While the actual area will depend on the data, it should vary smoothly and be relatively small unless root activity genuinely extends to the entire field of view.

3.2.3 Object detection

We performed a "pre-segmentation" procedure to detect blob centroids. Segmentation plays a central role in many image processing pipelines and typically involves (i) thresholding and identifying background, (ii) using distance/gradient transformations with peak detection to identify markers and (iii) routines such as watershed [35] to segment blob labels. For our data, we have found it appropriate not to carry out the final segmentation. Instead, we carried out the pre-segmentation steps from thresholding to peak detection in detecting object centroids. Given the variability in the data over a large frame sequence, we have found a simple "annealing thresholding" to be effective. This simply increases a threshold iteratively so as to remove large connected components beyond a maximum perimeter length (empirically determined from data to be in the range 1000-2000 pixels) thus removing the background from the image. Being iterative, this is not as efficient as simpler linear filters but it is simple and robust to noise and variability. In extreme cases where it was not possible to find a threshold level in this manner, we treated the 99^{th} percentile as image background and removed it.

Having applied the threshold filter, we identified centroids by (i) performing a difference of Gaussians to emphasise blob-like objects, (ii) applying a maximum filter and (iii) returning the coordinates of the local maxima (peaks) in the image. To allow for a fully automated routine that can cope with arbitrary datasets (in the scope of our light-sheet microscopy datasets) the emphasis in the centroid detection stage has been to avoid spurious centroid detections at the risk of under-sampling, while optimising for objects to be identified for at least two frames somewhere during the peak of their light intensity arc.

3.3 Notes on algorithm assumptions

We consider the role of certain model assumptions in coping with variable and noisy data.

3.3.1 Catchment region

Objects in the point cloud were expected to be separated by a minimum distance ε such that under global transformation single objects are matched. The parameter ε is of central importance at all stages of the image processing and tracking pipeline and can be chosen based on the data. For sparse point-cloud data there will typically be only one match within the ball of radius ε. An ε value can be chosen such that it is possible to find multiple objects matched to one object. The interpretation in this case is that all matched points in the catchment region correspond to the same object and that multiple points are the result of *e.g.* image processing anomalies. Merging lineages can be useful to avoid generating superfluous new identifiers. This adds robustness in the event of noisy data or image processing anomalies. Alternatives to using a ball of radius ε might consider anisotropic catchment regions that factor in the direction of the global transformation. The parameter ε value may result in the tracker merging objects that the object detection stage discriminates between if those objects are not sufficiently separated. In the case of accurately counting mitotic events, this reduces the introduction of false-positives *i.e.* the erroneous generation of new identifiers

3.3.2 Consensus translation

Object position updates were treated as small Gaussian perturbations of a global transformation. If the position of an object is given as

$$\mathbf{p}_i(t) = \gamma\big(\mathbf{p_i}(\mathbf{t} - \tau) + \xi_i$$

where $\xi_i \sim \mathcal{N}(0, \sigma_0)$ is the random fluctuation for one object, taking the "consensus" effectively filters out the noise to reveal the global movement. For example, the top m translations could be averaged to produce a new translation.

3.3.3 Minimum allowed frame detections

Transient events were expected to persist for a minimum number of frames greater than or equal to 1. For our data the minimum number of frames is 2. Objects that have been identified for less than 2 frames, termed *single-frame detections*, were excluded from the tracking result.

3.3.4 Gaussian distribution of ROI likelihood

In each frame, the objects of interest were expected to be found in the center of the field of view. Objects may be detected near the original frame boundaries depending on the size of the ROI. Objects can be treated as inliers/outliers based on standard outlier detection methods under the assumption of normality (for example within 1 or 2 standard deviations from the mean coordinate).

3.3.5 Further filtering of outliers

Object location and angle differences between object displacements and the global transformation may be used in a post-processing stage to remove misclassified outliers. In the PRIORS routine in Algorithm listing 1 additional transformations beyond the transformations generated from the data can be chosen based on modelling assumptions. Additionally, the filter routines on lines 16-18 can filter data

from the final result rather than during the tracking process based on modelling
assumptions.

3.4 Software implementation

The fully automated Python software has been made available on GitHub at the
link [36] and can be evaluated against test data available at the link [37]. The code
is easily installed over a Python scientific library distribution such as *Anaconda*
[38]. It includes interactive notebooks and can be run from a terminal window to
process a folder containing 3D image sequence data in fully automated fashion.

CPU usage was dominated by the image processing stage. For the light-sheet
microscopy experiments used in our evaluation, 3D images were approximately
$90,000$KB on disk with dimensions $(1392, 1040, 60)$, in (x, y, z) order. The software
was developed and tested on a personal laptop computer with 16GB RAM and
a 2.9-GHz Intel Core i7-3520M CPU. While times vary between different pipeline
modes, on average individual frame processing took around 40 seconds. More than
half of the time spent on image processing is spent on de-noising and smoothing
images. A negligible fraction of CPU time corresponds to the tracking stage of the
pipeline. The tracking stage alone processed 450 data frames in about 1.5 minutes.
The current implementation contains significant "meta analysis" overhead.

4 Discussion

In this work we address the problem of tracking transient events in structured 3D
point clouds and in absence of background keypoints, a typical scenario experienced
in extended *in vivo* time-lapse imaging of short-lived events in whole plant tissues.
We present a new tracking method and validate it by tracking cell divisions in
Arabidopsis root meristems, over the course of several days. The novelty of the
method proposed is in the way it deals at the same time with the transient nature of
the events to be tracked, and the lack of background reference points, or keypoints.
The validation presented against manual tracking with experimental data shows the
performance of the method compared to previously published on similar datasets.

Due to the transient nature of mitotic events and to the variation in image quality
in experiments lasting up to one week, we found a lack of guaranteed, permanent
features in image sequences. To circumvent the lack of fiducial markers, we in-
troduced the concept of fuzzy markers defined through sub-sampling the set of de-
tected events. In essence, we exchanged a morphological analysis problem for a fuzzy
registration-tracking problem. The algorithm then generates and evaluates candi-
date affine transformations between these fuzzy markers. This approach is distinct
from existing single-particle 3D tracking methods, which typically put greater em-
phasis on the existence of inlier objects. Furthermore, our method is characterised
by a rigid-structure prior, since the algorithm models frame-frame object movement
as a superposition of (a) long-range affine transformations due to global movement
of the tissue and (b) short-range transformations due to local fluctuations of the
nuclei within the tissue. The validation presented against both manual tracking and
surrogate data indicates the validity of the method in tracking transient events in
a 3D space.

It should be noted that the described algorithm has not been optimised for large
point clouds. The point clouds we worked with contain less than 100 objects in

each frame and often less than 20. As the frames are large 3D volumetric images, image processing is the performance bottleneck. Consequently, we have focused on the analysis and flexibility of the tracking algorithm instead of its computational optimisation. The transformation sampling and evaluation stages have been implemented efficiently within a vector-programming paradigm. Focus on performance improvements should emphasise efficiency of the *transformation proposal sampling.* This should, in general, be considered *NP-hard* as proposals are generated by finding congruences in large point cloud data. Also, for point clouds with high objects counts, proposal sampling could be applied on a suitable subregion of the data to avoid excessive evaluations.

5 Conclusion

In extended time-lapse microscopy, image processing and object detection relies heavily on checking and responding to variability in image properties. This goes beyond adaptive thresholding and may result in taking different routes through an image processing pipeline. Isolation of regions of interest within noisy point cloud data is also important.

Our approach was motivated by the case of transient florescent markers used in biology, where sub-cellular fluctuations are superposed with large-scale movements due to tissue growth. We present a clear case-study based on data on mitotic divisions in growing *Arabidopsis* roots, which could be immediately applied to similar datasets from other tissues. Given the vast variability in image quality and morphological characteristics (*e.g.* tissue types, density of objects to be tracked, landmark reference points, etc) in time-lapse microscopy datasets in life sciences and plant biology in particular, we believe that the novel approach proposed here will be a useful addition to the existing toolbox for single-particle tracking in 3D. Moreover, we expect that our approach will be applicable to any other situations where tracking transient events embedded in rigid structures is required.

6 Declarations

Author's contributions

SA wrote the manuscript and the code, TF and GS designed microscopy experiments and contributed to the writing of the manuscript, all authors reviewed the manuscript.

Availability of data and materials

The Python code [36] can be run against sample data hosted in a publicly accessible data repository [37]. The free-to-use code, which has an MIT license, can be run on any operating system running Python. The README at the project homepage provides instructions to run the code with the sample data.

Competing interests

The authors declare that they have no competing interests.

Ethics approval and consent to participate

Not applicable

Consent for publication

Not applicable

Funding

Not applicable

Acknowledgements

Not applicable

Author details
¹Department of Mathematics, Imperial College London, Kensington Campus, London, UK. ²Department of Life Sciences, Imperial College London, Kensington Campus, London, UK.

References
1. Meijering, E., Dzyubachyk, O., Smal, I., van Cappellen, W.A.: Tracking in cell and developmental biology. Seminars in Cell & Developmental Biology **20**(8), 894–902 (2009)
2. Ma, Y., Wang, X., Liu, H., Wei, L., Xiao, L.: Recent advances in optical microscopic methods for single-particle tracking in biological samples. Analytical and bioanalytical chemistry **411**(19), 4445–4463 (2019)
3. Ulman, V., Maška, M., Magnusson, K.E., Ronneberger, O., Haubold, C., Harder, N., Matula, P., Matula, P., Svoboda, D., Radojevic, M., *et al.*: An objective comparison of cell-tracking algorithms. Nature methods **14**(12), 1141 (2017)
4. Manzo, C., Garcia-Parajo, M.F.: A review of progress in single particle tracking: from methods to biophysical insights. Reports on progress in physics. Physical Society (Great Britain) **78**(12), 124601 (2015)
5. Chenouard, N., Smal, I., De Chaumont, F., Maška, M., Sbalzarini, I.F., Gong, Y., Cardinale, J., Carthel, C., Coraluppi, S., Winter, M., *et al.*: Objective comparison of particle tracking methods. Nature methods **11**(3), 281 (2014)
6. Kanade, T., Yin, Z., Bise, R., Huh, S., Eom, S., Sandbothe, M.F., Chen, M.: Cell image analysis: Algorithms, system and applications. In: Applications of Computer Vision (WACV), 2011 IEEE Workshop On, pp. 374–381 (2011). IEEE
7. Schnabel, R., Hutter, H., Moerman, D., Schnabel, H.: Assessing normal embryogenesis in Caenorhabditis elegans using a 4D microscope: variability of development and regional specification. Developmental biology **184**(2), 234–265 (1997)
8. Bao, Z., Murray, J.I., Boyle, T., Ooi, S.L., Sandel, M.J., Waterston, R.H.: Automated cell lineage tracing in Caenorhabditis elegans. Proc Natl Acad Sci U S A **103**(8), 2707–2712 (2006)
9. Keller, P.J., Schmidt, A.D., Wittbrodt, J., Stelzer, E.H.K.: Reconstruction of Zebrafish Early Embryonic Development by Scanned Light Sheet Microscopy. Science **322**(5904), 1065–1069 (2008)
10. Fick, R.H.J., Fedorov, D., Roeder, A.H.K., Manjunath, B.S.: Simultaneous cell tracking and image alignment in 3D CLSM imagery of growing Arabidopsis thaliana sepals. Biomedical Imaging (ISBI), 2013 IEEE 10th International Symposium on, 914–917 (2013)
11. Sena, G., Frentz, Z., Birnbaum, K.D., Leibler, S.: Quantitation of cellular dynamics in growing arabidopsis roots with light sheet microscopy. PLoS One **6**(6), 21303 (2011)
12. Gor, V., Elowitz, M., Bacarian, T., Mjolsness, E.: Tracking cell signals in fluorescent images, 142–142 (2005)
13. Liu, M., Yadav, R.K., Roy-Chowdhury, A., Reddy, G.V.: Automated tracking of stem cell lineages of Arabidopsis shoot apex using local graph matching. The Plant Journal **62**(1), 135–147 (2010)
14. Sethuraman, V., French, A., Wells, D., Kenobi, K., Pridmore, T.: Tissue-level segmentation and tracking of cells in growing plant roots. Machine Vision and Applications **23**(4), 639–658 (2011)
15. Campilho, A., Garcia, B., Toorn, H.V.D., Wijk, H.V., Campilho, A., Scheres, B.: Time-lapse analysis of stem-cell divisions in the Arabidopsis thaliana root meristem. The Plant journal : for cell and molecular biology **48**(4), 619–627 (2006)
16. Baesso, P., Randall, R.S., Sena, G.: Light sheet fluorescence microscopy optimized for long-term imaging of arabidopsis root development. In: Root Development, pp. 145–163. Springer, ??? (2018)
17. von Wangenheim, D., Hauschild, R., Fendrych, M., Barone, V., Benkova, E., Friml, J.: Live tracking of moving samples in confocal microscopy for vertically grown roots. eLife **6** (2017)
18. Rahni, R., Birnbaum, K.D.: Week-long imaging of cell divisions in the Arabidopsis root meristem. Plant Methods **15**(1) (2019)
19. Woell, D., Koelbl, C., Stempfle, B., Karrenbauer, A.: A novel method for automatic single molecule tracking of blinking molecules at low intensities. Physical Chemistry Chemical Physics **15**(17), 6196–6205 (2013)
20. Liu, A.-A., Lu, Y., Chen, M., Su, Y.-T.: Mitosis detection in phase contrast microscopy image sequences of stem cell populations: A critical review. IEEE Transactions on Big Data **3**(4), 443–457 (2017)
21. de Luis Balaguer, M.A., Ramos-Pezzotti, M., Rahhal, M.B., Melvin, C.E., Johannes, E., Horn, T.J., Sozzani, R.: Multi-sample Arabidopsis Growth and Imaging Chamber (MAGIC) for long term imaging in the ZEISS Lightsheet Z.1. Developmental biology **419**(1), 19–25 (2016)
22. Okuma, K., Taleghani, A., De Freitas, N., Little, J.J., Lowe, D.G.: A boosted particle filter: Multitarget detection and tracking. In: European Conference on Computer Vision, pp. 28–39 (2004). Springer
23. Godinez, W.J., Rohr, K.: Tracking multiple particles in fluorescence time-lapse microscopy images via probabilistic data association. IEEE Trans. Med. Imaging **34**(2), 415–432 (2015)
24. Reid, e.a. Donald: An algorithm for tracking multiple targets. IEEE transactions on Automatic Control **24**(6), 843–854 (1979)
25. Cox, I.J., Hingorani, S.L.: An efficient implementation of reid's multiple hypothesis tracking algorithm and its evaluation for the purpose of visual tracking. IEEE Transactions on pattern analysis and machine intelligence **18**(2), 138–150 (1996)
26. Schulz, D., Burgard, W., Fox, D., Cremers, A.B.: Tracking multiple moving targets with a mobile robot using particle filters and statistical data association. In: Robotics and Automation, 2001. Proceedings 2001 ICRA. IEEE International Conference On, vol. 2, pp. 1665–1670 (2001). IEEE
27. Besl, P.J., McKay, N.D.: A Method for Registration of 3-D Shapes. IEEE Transactions on Pattern Analysis and Machine Intelligence **14**(2), 239–256 (1992)
28. Fitzgibbon, A.W.: Robust registration of 2D and 3D point sets. In: Image and Vision Computing, pp. 1145–1153 (2003). University of Oxford, Oxford, United Kingdom
29. Jian, B., Vemuri, B.C.: Robust Point Set Registration Using Gaussian Mixture Models. IEEE Transactions on Pattern Analysis and Machine Intelligence **33**(8), 1633–1645 (2011)
30. Zhou, Z., Tu, J., Geng, C., Hu, J., Tong, B., Ji, J., Dai, Y.: Accurate and Robust Non-rigid Point Set

Registration using Student's-t Mixture Model with Prior Probability Modeling. Scientific reports **8**(1), 1–17
(2018)
31. Reddy, G.V., Heisler, M.G., Ehrhardt, D.W., Meyerowitz, E.M.: Real-time lineage analysis reveals oriented cell
divisions associated with morphogenesis at the shoot apex of arabidopsis thaliana. Development **131**(17),
4225–4237 (2004)
32. Kral N, S.G. Hanna Ougolnikova A: Externally imposed electric field enhances plant root tip regeneration.
Regeneration **3**(3), 156–167 (2016)
33. Donoho, D.L., Johnstone, J.M.: Ideal spatial adaptation by wavelet shrinkage. biometrika **81**(3), 425–455
(1994)
34. van der Walt, S., Schönberger, J.L., Nunez-Iglesias, J., Boulogne, F., Warner, J.D., Yager, N., Gouillart, E.,
Yu, T., the scikit-image contributors: scikit-image: image processing in Python. PeerJ **2**, 453 (2014)
35. Beucher, S.: Watershed, hierarchical segmentation and waterfall algorithm. In: Mathematical Morphology and
Its Applications to Image Processing, pp. 69–76. Springer, ??? (1994)
36. Amarteifio, S., Todd, F., Preussner, G., Sena, G.: Source code.
https://github.com/GiovanniSena/LSFM_CYCB_analysis_v2 (2019)
37. Todd, F., Sena, G.: Data. https://imperialcollegelondon.box.com/v/Amarteifio2019 (2019)
38. Anaconda. https://anaconda.org/anaconda/python (2018)

(a)

(b)

Fig 8: Variation in frame properties: The panels in (a) show four typical frames: A, "good" image; B, noisy image; C, noisy image with a low number of objects and some noise. D, over-saturated image to be marked as degenerate. In (b), quantitative analysis over 100 consecutive sample frames from one imaging session. The **signal** and **noise** levels are plotted with threshold values. The **narrow-band** plot shows the region between the 95^{th} and 99^{th} percentile ranges of the image histogram. Notably, these values fluctuate dramatically.

Fig 9: Isolating the Region of Interest (ROI): A narrow-band filter (B) and aggressive Gaussian smoothing (C) is used to find the largest connected component in the narrow band of the data corresponding to the region of activity in the root tip (D).