# Network-based instantaneous recording and video-rate reconstruction of 4D biological dynamics

*Zhaoqiang Wang[1, 3+], Hao Zhang[1+], Lanxin Zhu[1+], Guo Li[1], Yi Li[1,2], Yicong Yang[1], Yichen Ding[4], Mei Zhen[5], Shangbang Gao[2*], Tzung K. Hsiai[3,4*] and Peng Fei[1*]*

[1] *School of Optical and Electronic Information-Wuhan National Laboratory for Optoelectronics, Huazhong University of Science and Technology, Wuhan, 430074, China.*

[2] *College of Life Science and Technology, Huazhong University of Science and Technology, Wuhan, 430074, China.*

[3] *Department of Bioengineering, University of California, Los Angeles, Los Angeles, 90095, U.S.A.*

[4] *School of Medicine, University of California, Los Angeles, Los Angeles, 90095, U.S.A*

[5] *Lunenfeld-Tanenbaum Research Institute, Mount Sinai Hospital, University of Toronto, Toronto, M5G 1X5, Canada.*

[+] *These authors contribute equally to this work*

[*] *Correspondence: feipeng@hust.edu.cn (P.F.); sgao@hust.edu.cn (S.G.); thsiai@mednet.ucla.edu (T.K.H.)*

**Artefacts and non-uniform resolution together with slow reconstruction speed compromise the full capability of light-field microscopy for *in toto* observation of fast biological dynamics. Here, we demonstrate that combining a view-channel-depth (VCD) neural network with light-field microscopy can mitigate these limitations, yielding artefact-free three-dimensional image sequences with uniform spatial resolution and near three-orders-higher video-rate reconstruction throughput. We imaged neuronal activities across moving *C. elegans* and blood flow in a beating zebrafish heart at single-cell resolution with volume rates up to 200 Hz.**

A recurring challenge in biology is the quest to extract ever more spatiotemporal information from targets, as many millisecond-length transient cellular processes occur in three-dimensional (3D) tissues and across long time scales. Several imaging techniques, including epifluorescence and plane illumination methods, can image live samples in 3D at high spatial resolution[1-5]. However, they need to record a number of two-dimensional (2D) images to create a 3D volume, with the temporal resolution being compromised by the extended acquisition time of the camera.

Recently, light-field microscopy (LFM) has become the technique of choice for instantaneous volumetric imaging[6-13]. It permits the acquisition of transient 3D signals through post-processing the light-field information recorded by a single 2D camera snapshot. As LFM provides high-speed volumetric imaging limited only by the camera frame rate, it has delivered promising results in various applications, such as the recording of neuronal activities[7-10] and visualization of cardiac dynamics in model organisms[11]. Despite these advances, its generally low-and-nonuniform spatial resolution, together with the presence of reconstruction artefacts, still prevent its more widespread application in capturing millisecond time-scale biological processes at single-cell resolution. Although these problems can be mitigated through optimizing the way light-field is recorded[9,12], the extra system complexity could impede the wide dissemination of the LFM technique. Furthermore, current LFMs rely heavily on a computationally-demanding iterative recovery process that intrinsically limits the overall throughput of LFM reconstruction, compromising its potential for long time-scale applications.

Here, we propose a novel LFM strategy based on a View-Channel-Depth neural network processing of the light-field data, termed VCD-LFM. By light-field projection using a wave optics model[13], we generated synthetic light-field images from high-resolution 3D images experimentally obtained

beforehand, and readily paired them as input and ground-truth data, for network training. The VCD network (VCD-Net) was designed to extract multiple views from these 2D light-fields, and transform them back to 3D depth images, which would be compared with the high-resolution ground truth images to guide the optimization of the network. Through iteratively minimizing the loss of spatial resolution by incorporating abundant structural information from training data, this deep-learning VCD-Net could be gradually strengthened until it's capable of deducing 3D high-fidelity signals at uniform resolution across the imaging depth. In addition to higher resolution and minimum artefact, once the VCD-Net is properly trained, it can deduce an image stack from a light-field measurement at a millisecond time scale, showing ultra-high throughput suited for time-lapsed video processing. We demonstrated the ability of the VCD-LFM method via imaging the motor neuron activities of L4-stage *C. elegan*s rapidly moving inside a 300 × 300 × 50 μm microfluidic chamber, at an acquisition rate of 100 Hz and processing-throughput of 13 volumes per second. This allowed us to extract the 4D spatiotemporal patterns of neuronal calcium signaling and track correlated worm behaviors at single-cell resolution, notably better than conventional light-field deconvolution microscopy (LFDM). Furthermore, we *in toto* imaged blood flow in the beating heart of zebrafish larvae, enabling the velocity tracking of single blood cells and ejection fraction analysis of heartbeat across a 250 × 250 × 150 μm volume at 200 Hz.

Our VCD-LFM involves the training of a VCD convolutional neural network and its inferences based on the images obtained from LFM (**Fig. 1a, Methods, Supplementary Fig. 1, 2**). To create the data for network training, we first obtained a number of high-resolution (HR) 3D images of stationary samples using synthetic or experimental methods (**Fig. 1a, Methods**). Referring to the wave optic model of LFM[9], we projected these HR 3D images into 2D light-field images, which were then used as the input for the network training (**Fig. 1b. step 1, Supplementary Note 1**). Each synthetic light-field image was first re-arranged into different views, from which features were extracted and incorporated into multiple channels in each convolutional layer. The final output channels were then assigned to a number of planes representing different depths to generate an image stack. Using cascaded convolution and deconvolution layers (U-Net architecture, **Supplementary Fig. 3 and Supplementary Table 1**) to repetitively extract features, this VCD procedure generated intermediate 3D reconstructions (**Fig. 1b. step 2**). The pixel-wise mean-square-errors (MSE) were counted as the loss function to indicate how different these outputs were to the ground truth images. By iteratively minimizing the loss function (**Fig. 1b. step 3**), the network was gradually optimized until it could transform the synthetic light-fields into 3D images similar enough to the ground truth images (**Supplementary Note 2**). Following training on gigavoxels of prior data, the network was capable of instantly implementing VCD transformation of sequential light-field measurements that recorded dynamic processes, and inferring a sequence of 3D volumes at a rate up to 13 volumes s$^{-1}$ (**Fig. 1b. step 4**).

We built a custom microscope to enable *in situ* light-field recording and 3D wide-field imaging (**Methods, Supplementary Fig. 1**). To demonstrate the capability of VCD-LFM, we reconstructed sub-diffraction beads captured using a 40×/0.8w objective, and quantified the resolution improvement resulting from the network by comparing the results with those from conventional light-field deconvolution (LFD, **Fig. 1c-f**). As verified by 3D wide-field imaging of the same volume, the fluorescence of individual beads was correctly localized throughout the volume (**Fig. 1e, f, Supplementary Fig. 5**). VCD-LFM with network using wide-field 3D image as HR reference, yielded averaged x,y~1.1 ± 0.1 and z~3.0 ± 0.2 μm resolutions (n = 2527 beads) which were also uniform across 60-μm imaging depth (**Fig. 1g**, **Supplementary Fig. 6**), showing significant improvements as compared to x,y~2.6 ± 0.5 and z~5.0 ± 0.9 μm of the LFDM results (x,y~4.0 and z~5.7 μm at the outer edge of the

axial field of view). We also note that the performance of VCD-Net is highly training data-oriented, and
less coupled with the optical limit of the LFM system, indicating that its reconstruction quality can be
further improved to isotropic resolution (~1.0 ± 0.2 μm) by including higher-resolution data in the
training (**Fig. 1f,g, Supplementary Fig. 7**). Furthermore, VCD-LFM substantially removes the mosaic-
like artefacts near the focal plane that are common in LFDM (**Fig. 1d, Supplementary Fig. 5, 6**), and
perform well even when the signals are extremely weak with high background noises (**Supplementary
Fig. 8**).

We show VCD-LFM to be suitable for capturing dynamic processes in live animals by demonstrating
the imaging of neuronal activities of moving *C. elegans* and the beating heart of a zebrafish larva. A
microfluidic chip was used to permit *C. elegans* (L4-stage) to rapidly move inside a micro-chamber (300
× 300× 50 μm, **Fig. 2a**). We used a 40×/0.8w objective to image the activities of motor neurons labeled
with GCaMP (*Pacr-2(s)*::GCaMP6(f)) at a 100-Hz acquisition rate, yielding 6000 light-fields in a 1-
minute observation. The VCD-Net provided high-quality quantitatively-accurate reconstructions to
visualize the neuron calcium signaling at single-cell resolution during fast body movement (**Fig. 2b,
Supplementary Fig. 9-11**). In contrast, LFD suffered from ambiguous cellular resolution and
deteriorated image quality around the focal plane, a known limitation of LFD (**Fig. 2c**). Furthermore, the
non-iterative VCD reconstruction could sequentially recover 3D images at a volume rate of 13.5 Hz,
around 900-times faster than the iterative LFD method (**Fig. 2d**). Our VCD-LFM therefore shows
significant advantages for visualizing sustained biological dynamics, which is computationally
challenging to conventional methods (**Supplementary Table 3, Video 1**). By applying an automatic
segmentation of the worm body contours based on the location and morphology of neuron signals
(**Methods, Supplementary Note 3**), we calculated the worm's velocity and curvatures related to its
locomotion and behavior, thereby allowing classification of the worm motion into forward, backward, or
irregular crawling (**Fig. 2e, f**). Meanwhile, we identified the specific A- and B- motor neurons that have
been associated with motor-program selection, and mapped their calcium activities over time (**Fig. 2g,
Methods, Supplementary Video 2**). The patterns of transient $Ca^{2+}$ signaling were found to be relevant
to the switches from forward to backward crawling of the worm, consistent with previous findings[14, 15]
(**Fig. 2g-h**).

We also captured the cardiac hemodynamics in beating zebrafish heart. To reduce the background from
body tissue, a laser-rod was generated to selectively illuminate the heart region[20], and the light-field video
was recorded at 200 Hz volume rate using a 20×/0.5w objective (**Fig. 3a, Methods, Supplementary Fig.
2**). The flowing red blood cells (RBCs, line *Tg*(gata1a:dsRed)) and the beating cardiomyocytes nuclei
(line *Tg*(myl7:nls-gfp)) were both four-dimensionally reconstructed by VCD-Net with resolution,
structure similarity and processing throughput notably better than conventional LFD approaches (**Fig.
3b, c, f, g, Supplementary Fig. 12, 13, Video 3, 4**). Beyond the recovery of spare signals, a known
limitation of light-field microscopy, we further demonstrated our VCD-LFM can reconstruct 3D beating
heart expressing highly-dense trabecular myocardium (line *Tg*(cmlc2:gfp)). It is noted that the
reconstruction of such dense signals with acceptable resolution remains far unmet in earlier LFDM
demonstrations (**Fig. 3h, i, Supplementary Fig. 14, Video 5**). The fast VCD-LFM visualization with
sufficient quality then permitted the quantitative investigation of transient cardiac hemodynamics. We
tracked 19 individual RBCs throughout the entire cardiac cycle of 415 ms, during which the blood flow
was pumped in-and-out of the ventricle at a high speed of over 3000 μm s$^{-1}$ (**Fig. 3d, e**). With segmenting
the continuous heart boundary, we successfully quantified the volume change of myocardium during the
diastole and systole, and calculated the ejection fraction of heartbeat (**Fig. 3j, k**). Furthermore, through

a combination of VCD-LFM with SPIM, we realized the 3D visualization and quantitative analysis of
blood flow inside the static zebrafish vessels (line *Tg*(fli1:gfp; gata1a:dsRed), **Methods, Supplementary**
**15, Video 6, 7**).

To record the neural activities of acting *C. elegans* at single-cell resolution, previous studies captured
the worm in 2D or 3D at a frame/volume rate up to 50 Hz[7, 16,17]. An additional infrared channel is often
required to provide bright-field images of worm behaviors as a reference. Our VCD-LFM, capable of
100 Hz recording and 13.5 Hz reconstruction of fast 3D processes combined with network-based all-
fluorescence analysis, offers a new pipeline for the study of neural activities and relevant worm behaviors
in whole moving worms, without motion blur or computation traffic. For cardiovascular imaging, recent
scanning-based approaches for volumetrically imaging the beating heart and blood flow in zebrafish
larvae[18, 19] have required complicated optics, ultra-fast camera, and high-intensity excitation. In contrast,
our method based on a relatively simple system and easy-spread computation procedure offers a
compelling solution for investigating the dynamic properties and functions of the cardiovascular system
with single-cell resolution and low phototoxicity. Therefore, VCD-LFM could be a valuable tool for
studying dynamics on fast timescales, potentially benefiting a variety of applications such as neural
signaling-regulated phenotype studies and medically relevant dysfunctions of the heart and blood
transport system in model organisms.

In summary, we introduce a VCD-LFM approach and demonstrate its ability to image transient
biological dynamics with improved spatial resolution, minimal reconstruction artefacts, and increased
reconstruction throughput. These advances are crucial for sustained volumetric imaging of blood flow in
zebrafish larva and neural activities in behaving *C. elegans* at high resolution. While VCD-LFM greatly
improves reconstruction quality from one originally coupled with the optics system to one can be
optimized by prior data training, it still requires the pre-acquisition of training images from the specific
type of specimens. We expect this will improve with the unceasing development of deep learning,
allowing more versatile architectures and flexible training pipelines. Besides the combination with basic
LFM setup, we note that the VCD-Net is also compatible with modified LFM modalities, such as dual-
objective setup[12] or scanning light-sheet illumination[20]. Higher-quality light-field acquisitions will also
contribute to better VCD-Net reconstruction, especially under dense labelling conditions. Finally, we
expect that VCD-LFM could potentially bring new insights for computational imaging techniques, to
allow us to further push the spatiotemporal resolution limits in observations of biological dynamics.

161
**Figure 1. VCD-LFM and its performance.** (**a**) HR 3D imaging of steady samples by a confocal
microscope (upper row) and instantaneous recording of dynamic samples by a light-field microscope
(bottom row). (**b**) The VCD-Net reconstruction pipeline, containing: 1. Forward light-field projection
from the HR image stacks; 2. VCD transformation of synthetic light-field inputs into intermediate 3D
image stacks; 3. Network training via iteratively minimizing the difference between VCD inferences and
confocal ground truths; 4. Inference of 3D images from the recorded light-field images by well-trained
VCD-Net. (**c**)-(**f**) Maximum intensity projections (MIPs) of the same fluorescent beads and their
resolution (FWHM) by wide-field microscopy, LFDM and VCD-LFM trained with anisotropic and
isotropic HR data, respectively. The images were acquired using a 40×/0.8 NA water objective. The
results demonstrate a notable improvement in lateral and axial resolution for VCD-LFM and its data-
oriented reconstruction decoupling the achievable resolution with the optical limit of system. Inset: line
profiles across the beads at 14 µm off the focal plane. Scale bars, 10 µm. (**g**) Average axial (dashed lines)
and lateral (solid lines) FWHM of the beads across the volumes reconstructed by LFDM (n = 2,039
beads), anisotropic (n = 2,527 beads) and isotropic VCD-LFM (n = 2,731 beads) respectively. VCD-
LFM improves lateral/axial resolution across the entire 60-µm overlapping volume, with Aniso-VCD
achieving uniform resolution of ~1.1 ± 0.1 µm and ~3.0 ± 0.2 µm in x/y and z, respectively, and Iso-
VCD achieving near isotropic resolution of ~1.0 ± 0.2 µm.

**Figure 2. Whole-animal $Ca^{2+}$ imaging of moving *C. elegans* using VCD-LFM.** (**a**) Configuration

combining light field with microfluidic technique for imaging motor neuron activities of L4-stage *C.*

*elegans* (strain ZM9128 *hpIs595*[*Pacr-2(s)*::GCaMP6(f)::wCherry]) acting in a microfluidic chip (300 ×

300 × 50 μm, top panel) at 100 Hz recording rate. (**b**),(**c**) MIPs of one instantaneous volume reconstructed

by VCD and LFD, respectively. The magnified views of selected region show that VCD restores single-

cell resolution and removes artefacts. Scale bars, 10 μm. (**d**) The reconstruction throughput of VCD-

LFM and LFDM for processing light-field video. (**e**) Curvature kymograms along the body of moving

worm based on deep learning (U-Net) segmentation of reconstructed worm body contours. An ethogram

describing the worm behavior over the time (lower panel) is obtained by analyzing the curvature change.

(**f**) Selected volumes with time-coded traces (left and right, 300 ms; middle, 1000 ms) in accordance to

the ethogram visualizing the forward (left), backward (middle) and irregular (right) crawling tendency

of the worm. Scale bars, 20 μm. (**g**) Neural activities map of the acting worm (right) by quantifying the

$Ca^{2+}$ signal fluctuation of identified motor-neurons (left). Each row shows the fluctuation of an

individual neuron with color indicating percent fluorescence changes ($\Delta F/F_0$). (**h**) The velocity plot of

the worm during 1-minute observation.

**Figure 3. Imaging of various cardiac hemodynamics in beating zebrafish heart using VCD-LFM.**
(**a**) Schematic of selective volume illumination based light-field imaging and VCD reconstruction process. High-contrast light-field sequences of flowing RBCs (beating heart) are recorded under a rod-like beam illumination setup that selectively excites the fluorescence signals within the volumes of interest. (**b**),(**c**) MIPs in *x-y* (left) and *y-z* (right) planes of one instantaneous volume of flowing RBCs (sparse RFP signals) by VCD-LFM and LFDM, respectively. The dashlines indicate the heart. (**d**) Tracks of 19 single RBCs throughout the cardiac cycle. A static heart is outlined for reference. (**e**) Velocity map of two temporally adjacent volumes of RBCs during systole. (**f**),(**g**) MIPs in *x-y* (top) and *x-z* (bottom) planes of one instantaneous volume of beating myocytes (mid-density GFP signals) by VCD-LFM and LFDM, respectively. (**h**),(**i**) 3D visualization of beating myocardium in a transient moment by VCD-LFM and LFDM, respectively. The myocardium was densely labelled by GFP, with showing continuous trabecular structures. Arrows indicate the inlet and outlet of cardiac pumping. A: atrium; V: ventricle. (**j**) Volume change of myocardium during the diastole and systole in one cardiac cycle. (**k**) Volume change ratio of the ventricle during the diastole and systole in one cardiac cycle. The rate is calculated by ($V_{time}$ - *ESV*) / *EDV*, where $V_{time}$ is the time-varying volume of the ventricle during heartbeat, *ESV* and *EDV* represent the volumes at the end of systole and diastole, respectively. The ejection fraction of heartbeat (EF), a cardiological index given by (*EDV* - *ESV*) / *EDV*, also shown as the peak in the curve, is calculated to be ~71%. These quantifications are depended on the 3D segmentation of a series of clearly resolved myocardium, as one of them shown in (**h**). 20 of total 120 time points (1 of each 6) are selected for the analysis during an ~400 ms cardiac cycle. Scale bars, 50 µm.

**References**

1. Keller, P.J., Schmidt, A.D., Wittbrodt, J. & Stelzer, E.H.K. *Science* **322**, 1065-1069 (2008).
2. Planchon, T.A. et al. *Nature methods* **8**, 417 (2011).
3. Conchello, J.-A. & Lichtman, J.W. *Nature methods* **2**, 920 (2005).
4. Huisken, J., Swoger, J., Del Bene, F., Wittbrodt, J. & Stelzer, E.H.K. *Science* **305**, 1007-1009 (2004).
5. Yang, W. & Yuste, R. *Nature methods* **14**, 349 (2017).
6. Levoy, M., Ng, R., Adams, A., Footer, M. & Horowitz, M. *ACM Transactions on Graphics (TOG)* **25**, 924-934 (2006).
7. Prevedel, R. et al. *Nature methods* **11**, 727-730 (2014).
8. Pégard, N.C. et al. *Optica* **3**, 517 (2016).
9. Cong, L. et al. *eLife* **6**, e28158 (2017).
10. Nobauer, T. et al. *Nature methods* **14**, 811-818 (2017).
11. Wagner, N. et al. *Nature methods* **16**, 497-500 (2019).
12. Cohen, N. et al. *Opt. Express* **22**, 24817-24839 (2014).
13. Broxton, M. et al. *Optics express* **21**, 25418-25439 (2013).
14. Kawano, T. et al. *Neuron* **72**, 572-586 (2011).
15. Wen, Q., Gao, S. & Zhen, M. *Philosophical Transactions of the Royal Society B: Biological Sciences* **373**, 20170370 (2018).
16. Nguyen, J.P. et al. *Proceedings of the National Academy of Sciences* **113**, E1074-E1081 (2016).
17. Leifer, A.M., Fang-Yen, C., Gershow, M., Alkema, M.J. & Samuel, A.D. *Nature methods* **8**, 147 (2011).
18. Mickoleit, M. et al. *Nature methods* **11**, 919 (2014).
19. Voleti, V. et al. *Nature methods* **16**, 1054-1062 (2019).
20. Truong, T.V. et al. *bioRxiv*, 403303 (2018).

### Acknowledgements

We thank Dr. Jau-Nian Chen and Yuan Dong at UCLA for their help on zebrafish experiments. We thank Yanyi Huang at Peking University and Jianbin Wang at Tsinghua University for their discussions and comments on the manuscript. This work was supported by the following grants: National Key R&D program of China (2017YFA0700500 to P.F.), National Natural Science Foundation of China (21874052 to P.F., 31671052 to S.G.), National Science Foundation of Hubei (2018CFA039 to S.G.), National Institute of Health (NIH, 5R01HL129727-04, K99HL148493 to T.K.H.), American Heart Association (AHA 18CDA34110338 to T.K.H.), and the Junior Thousand Talents Program of China (P.F. and S.G.). Confocal experiment was done at the Advanced Light Microscopy/Spectroscopy Laboratory and the Leica Microsystems Center of Excellence at the California Nano Systems Institute at UCLA with funding support from NIH Shared Instrumentation Grant (S10OD025017) and NSF Major Research Instrumentation grant (CHE-0722519).

### Author contributions

P.F., Z.W. and H.Z. conceived the idea. P.F., S.G. and T.K.H. oversaw the project. Z.W., L.Z., G.L., Y.Y. developed the optical setups. H.Z., Z.W. and L.Z. developed the programs. Z.W., L.Z., G.L., H.Z, Y.D. and Y.L. conducted the experiments, processed the images. Z.W., L.Z., H.Z, M.Z., S.G., T.K.H. and P.F. analyzed the data and wrote the paper.

### Competing interests

The authors declare no conflicts of interest in this article.

## Methods

**Epi-illumination LFM setup.** An epi-fluorescence light-field setup was built on an upright microscope (BX51, Olympus). The light-field/wide-field detection paths were appended to the camera port of host microscope, with using a flip mirror to switch between two detection modes. A motorized z-stage (Z812B, Thorlabs) together with a water chamber were directly mounted onto the microscope stage (x-y), to three-dimensionally control the samples inside the chamber. A water immersion objective (LUMPlanFLN 40×/0.8W, Olympus) was used to collect the epifluorescence signals from samples. For recording the light field, a microlens array (MLA, APO-Q-P150-F3.5 (633), OKO Optics) was placed at the native image plane to collect the light-field signals. A 1:1 relay system (AF 60 mm 2.8D, Nikon) was used to conjugate the back focal plane of MLA with the camera sensor plane (Flash 4.0 V2, Hamamatsu). See **Supplementary Fig. 1** for more details of Epi-LFM setup.

**Selective plane/volume illumination LFM setup.** We also developed a LFM setup based on selective volume illumination. Two pairs of beam reducers combined with an adjustable iris were used to generate a scalable rod-like beam (473 or 532-nm), which was finally projected onto the sample through a 4× illumination objective (Plan Fluor 4×/0.13, Nikon) placed perpendicular to the detection path. It confined the fluorescence excitation within the heart region of zebrafish embryo, reducing the excessive emission from out of the volume of interest that could smear the desired signals. This selective volume illumination mode provided light-field image with less background noise and increased contrast[19]. For observing the dynamic process of blood flowing through vessels, we also integrated a standard SPIM channel (473-nm, 4-$\mu m$-thickness laser-sheet) to implement the 3D imaging of steady vessels. The light-rod/light-sheet illumination paths were aligned, providing double excitation to the sample from its dual sides. A water immersion objective (Fluor 20×/0.5w, Nikon) were used to collect the dual-color fluorescence signals from the RBCs/cardiomyocytes. A dichroic mirror split the dual-channel GFP (vessels)/DsRed (RBCs) signals for wide-field and light-field detection, respectively. The light-field detection here followed the same design used in the epi-illumination LFM. See **Supplementary Fig. 2** for more details of this hybrid setup.

**VCD light-field reconstruction Network.** In a general convolutional neural network (CNN), a certain $N$th convolutional layer receives feature maps from the last *(N-1)*th layer, and generates new feature maps using different convolution kernels. The network finally produces a multi-channel output, in which each channel is a non-linear combination of the original input. It's reasonable to share this concept with the synthetic refocusing in the light-field imaging domain, where each synthetic focal plane of the reconstructed volume can be interpreted as a superposition of different views extracted from the light-fields. Through cascaded layers, our model is expected to gradually transform the original angular/view information from the light-field raw image into depth features, eventually forming the conventional 3D image stack, and reconstructing the scene. In our implementation, the customized VCD-Net is based on a modified U-Net architecture of 6 convolution and 6 deconvolution layers[24]. Each layer has three parameters: **n, f** and **s**, denoting the output channels number, the filter size of convolution kernel and the step size of the moving kernel, respectively, as specified in **Supplementary Fig. 3 and Supplementary Table 1.** The pixels from the input 2D light-field raw image (dimension: **a × b**, height × width) are first reformatted into a series of different views (dimension: **a/d × b/d × d²**, height × width × views) according to their relative positions to each lenslet. A subpixel up-scaling part further interpolates these views to dimension **a × b × d²** (height × width × views). Then in the first VCD layer, the initial

transformation from "view" to "channel" is done by convolutions with all these views using different kernels of each channel, generating an output with dimensional $a \times b \times n$, where $n$ is the channel numbers. The following convolution and deconvolution layers keep combining old channels from previous layer and generating new ones to fully excavate the hidden features from input image. Finally the last layer outputs a 3D image with the channel number $n$ equal to the desired number of synthetic focal planes $c$, thereby finishing the transformation from "channel" to "depth" (dimension: $a \times b \times c$, height $\times$ width $\times$ depth).

For VCD-Net training, HR 3D reference images were from confocal microscopy of static samples or synthetic data. Then a light-field projection based on wave optics model[9] was applied to these HR references to generate corresponding synthetic 2D light-field images as inputs (**Supplementary Note 1**). The VCD-Net was gradually trained via iteratively minimizing the difference between its intermediate outputs and the HR references. With setting appropriate loss functions, such as the MSE (Mean Squared Error) of the pixel intensity, the VCD-Net obtained optimized kernel parameters for each layer and efficiently converged to a well-trained status, at which the network was capable of transforming the synthetic light-field inputs back to 3D images accurately enough as compared to the HR references.   At inference stage, unlike the previous LFD method, the trained VCD-Net could directly infer a sequence of 3D images from the input experimental light-field video that contained a number of light-field frames recording the dynamic biological processes. The time consumption depended on the dataset size and computation resources. As a reference point, the VCD-Net converged after training on 4580 pairs of blood cells image patches (size: $176 \times 176 \times (51)$ pixels) with 110 epochs, which cost ~4 hours on a single GPU. Then the trained network spent ~30 seconds to reconstruct 450 consecutive volumes (size: $341 \times 341 \times 51$ pixels) from acquired light-field videos. This 4D reconstruction throughput was compared to ~11.8 hours (~42467 s) by LFD (8 iterations) on the same workstation using the identical light-field PSF to reconstruct the volumes. The computation was performed on a workstation equipped with Intel(R) Core (TM) i9-7900X CPU @3.3GHz, 128G RAM, and Nvidia GeForce RTX 2080 Ti graphic cards. For more details in VCD-Net training and inference, see **Supplementary Note 2** and **Supplementary Software**.

**PSF measurements.** We used both light-field ($40\times/0.8$w objective) and 3D wide-field ($40\times/0.8$w objective plus 0.5-μm z-step) modes to image the same volume of sub-diffraction fluorescent beads (0.5 μm Lumisphere, BaseLine) distributed in a piece of hydrogel (0.7% low melting agarose solution, BBI Life Sciences). To demonstrate the tunable performance of VCD-Net with applying different training data, we trained anisotropic and isotropic VCD-Nets using synthetic anisotropic and isotropic beads, respectively. These synthetic beads with random 3D distributions were generated using a 3D Gaussian kernel with controllable FWHMs (anisotropic: $1 \times 1 \times 3$ μm; isotropic: $1 \times 1 \times 1$ μm). Then 3D image stacks with 60-μm depth were recovered from the recorded light-field image using LFD, anisotropic and isotropic VCD-Nets. The reconstruction yielded 61 z planes with voxel size $0.34 \times 0.34 \times 1$ μm for anisotropic VCD-Net and LFD (8 iterations), and 177 z planes, with voxel size $0.34 \times 0.34 \times 0.34$ μm for isotropic VCD-Net. The reconstructed beads were then detected and fitted with 1D Gaussian function in each dimension to determine the full width at half maximum (FWHM) as spatial resolution metrics using a custom-written MATLAB script. The PSFs of LFD and VCD-Nets at certain depth were then compared via plotting the line profiles of the same resolved beads. The achieved lateral and axial resolutions at certain depth by each method were indicated through calculating the averaged FWHM values of the resolved beads, e.g., z = -14 μm. The axial performance of anisotropic and isotropic VCD-

353 Net *versus* LFD at different depths were further analyzed via measuring the FWHMs of resolved beads
354 across a 60-μm depth, with their variation indicating the non-uniformity of light-field recovery.
355
***C. elegans* strain.** The strain ZM9128 *hpIs595*[*Pacr-2(s)*::GCaMP6(f)::wCherry], with calcium
indicators tagged to the A-and B- class motor neurons, was used to detect the neural activities in the
acting worm (**Fig. 2**). The strain QW1217 *hpIs467*[*Prab-3*::NLS::RFP], with nuclei labelled in all the
neuron cells, was used to demonstrate the imaging performance of VCD-LFM (**Supplementary Fig. 8**).
All *C. elegans* were cultured on standard Nematode Growth Medium (NGM) plates seeded with OP50
and maintained at 22 °C incubators until L4 stage.
362
***C. elegans* imaging.** To obtain HR 3D data for network training, the L4-stage anesthetized worms
(QW1217 *hpIs467*, ZM9128 *hpIs595*, by 2.5 mM levamisole in M9 buffer, Sigma-Aldrich) were first
imaged using a 40×/0.95 objective on a confocal microscope (FV3000, Olympus). To demonstrate the
performance of VCD-LFM, anesthetized worms (QW1217 *hpIs467*) embedded in agarose were *insitu*
imaged by both light-field and wide-field detection modes in our Epi-illumination LFM. The acquired
3D wide-field image stacks (1-μm step size) were deconvolved in Amira software (Thermo Scientific),
for comparison with those light-field reconstructions by LFD and VCD-Net. The 3D reconstructions
encompassed 31 z planes, with voxel size of $0.34 \times 0.34 \times 1$ μm. In the behavior studies of *C. elegans*,
the awake L4-stage worms (ZM9128 *hpIs595*) were loaded into a microfluidic chamber (~300 × 300 ×
50 μm), which allows the worms to act within the FOV of a 40× objective. The GFP calcium signals of
acting worm were then recorded for one minutes at 100 Hz frame rate (2-ms exposure time). A 2× pixel
binning was applied during the image acquisition of camera, for obtaining light-field images with better
signal-to-noise ratio.
376
**Quantitative Analysis of neural activities and behavior of acting *C. elegans*.** We performed semi-
automated tracking of the movement and intensity fluctuation of each individual neuron using TrackMate
Fiji Plugin[23]. Neurons were segmented automatically in each volume with applying a circular ROI
through the Difference of Gaussian (DoG) detector and then tracked using Kalman filter. In a few cases
when this automatic tracking failed due to fast movement of the neuron, the missing detections and
tracking mistakes were manually corrected. For each neuron in each volume, all the pixels within the
ROI were averaged to produce a single fluorescence value F. The F values in the time series was re-
normalized using an exponential-decay function, to remove the effect of photobleaching. To extract
$\Delta F/F_0$, we calculated $\Delta F/F_0 = (F(t)-F_0)/F_0$ with $F(t)$ being the fluorescence intensity at certain time point,
and $F_0$ being the mean value of $F(t)$. Worm behavior analysis was then implemented based on the same
fluorescence images. We successfully inferred the 2D worm outlines throughout the whole time period
using a U-Net based image segmentation[24], by which the center lines could be further extracted to
calculate the changing curvatures and motion velocities[25]. More details are given in **Supplementary
Note 3**. The body curvature map was shown in Fig. 2e, indicating the time-varying worm postures
during the 1-minute acquisition. The velocity curves shown in Fig. 2h were denoised by a sliding
average filter.
393
394 **Fish husbandry and lines.** Transgenic zebrafish lines *Tg*(cmlc2:gfp; gata1a:dsRed) (**Fig. 3**,
395 **Supplementary Fig. 12, 14, Video 3, 5**), *Tg*(fli1:gfp; gata1a:dsRed) (**Supplementary Fig. 15, Video 6**),
396 *Tg*(myl7: nls-gfp) (**Supplementary Fig. 13, Video 4**) were used in our experiments. Embryonic stage

fishes were maintained until 3-4 dpf in standard E3 medium, which was supplemented with extra PTU (Sigma Aldrich, MO) to inhibit melanogenesis. Then, the larvae were anesthetized with tricaine (3-amino benzoic acidethylester, Sigma Aldrich, MO) and immobilized in 1% low-melting-point agarose inside FEP (Fluorinated Ethylene Propylene) tube for further imaging. In the imaging of cardiac blood flow, the embryos were injected with gata2a morpholino oligonucleotide at single-cell stage, to slow down hematopoiesis, and thereby reduce the density of RBCs. All the experiments were performed in compliance with UCLA IACUC protocols.

**Zebrafish imaging.** The fish samples were precisely moved by a custom stage (xyz + rotation) so that their signals of interest could be positioned within the optimal imaging region, which was determined by the domain of selective volume illumination and the FOV of detection objective (LUMPlanFLN 20×/0.5w, Olympus). During cardiac imaging, the middle of heart was moved to the focal plane so that the volumetric illumination (473 or 532-nm) could selectively excite the RBCs/myocytes nuclei/myocardium. The light-field movies were recorded at 200 Hz frame rate, with $768 \times 768$ frame size that corresponds to a lateral FOV of ~$250 \times 250$ µm. The movies covered 4-5 cardiac cycles with containing 450 frames (5 ms exposure for each frame). For imaging the dynamic process of blood flow inside vessels, the cameras of dsRed and GFP channels recorded light-field video of rapidly flowing RBCs and SPIM image stack of static tail vessels, respectively. The light-field movies lasted for ~5.3 s with containing 527 frames (10 ms exposure for each frame). The SPIM stack contained 50 planes that covered a 100-µm z depth of fish tail (step size 2 µm).

To obtain HR 3D images of RBCs, myocyte nuclei and myocardium for VCD-Net training, deeply anesthetized fish larvae with immobilized hearts were embedded in 1% low-melting agarose for sustained confocal imaging (SP8-STED/FLIM/FCS, Leica) using a 20×/0.75 Objective (HC PL APO CS2). Then we trained three VCD-Nets based on these HR images of static blood cells (16 fishes), nuclei (23 fishes), and myocardium (11 fishes), and their paired LFPs. Besides the confocal imaging of immobilized heart, we also combined SPIM with retrospective gating method[18, 21, 22], to obtain the light-sheet based HR 3D image of periodically beating myocardium for network training. A movie stack of light-sheet images of the dense myocardium was first acquired, containing 60-70 movies at different planes (z step = 2 µm) that captured the complete myocardial structures. Each movie was then temporally aligned using the retrospective gating, to reconstruct a beating 3D heart. When the training using these prior data was finished, the network inferred 3D images from empirical light-fields with output voxel size $0.68 \times 0.68 \times 2$ µm (blood cells, myocardium) or $0.326 \times 0.326 \times 3$ µm (myocyte nucleus). The output image stacks contain 51 planes with depth spanning from -50 to 50 µm or -75 to 75 µm.

**Velocity map of RBCs and displacement analysis of cardiomyocytes in zebrafish heart.** Tracking of the flowing RBCs was applied using Imaris software (Bitplane). **Fig. 3d** shows the trajectories of 19 RBCs throughout one cardiac cycle (415 ms). **Fig. 3e** demonstrates the velocity map that has been extracted at one specific time point during systole, which contains 31 vectors from all trackable RBCs in that frame The analysis was written in MATLAB and the vector field was visualized by Mayavi[28].

**Data availability.** The datasets generated and analyzed in this study are available from the corresponding authors upon request.

**Code availability.** Custom codes for VCD-Net and quantitative analyses implemented in current study

are either available at https://github.com/aaronzq/vcdlfm or from the corresponding authors upon request.

## References

21.  Lee, J. et al. *The Journal of clinical investigation* **126**, 1679-1690 (2016).
22.  Liebling, M., Forouhar, A.S., Gharib, M., Fraser, S.E. & Dickinson, M.E. *Journal of biomedical optics* **10**, 054001 (2005).
23.  Tinevez, J.Y. et al. *Methods* **115**, 80-90 (2017).
24.  Falk, T. et al. *Nature methods* **16**, 67 (2019).
25.  Restif, C. et al. *PLoS computational biology* **10**, e1003702 (2014).
26.  Vercauteren, T., Pennec, X., Perchant, A. & Ayache, N. *NeuroImage* **45**, S61-S72 (2009).
27.  Thirion, J.-P. *Medical image analysis* **2**, 243-260 (1998).
28.  Ramachandran, P. & Varoquaux, G. *Computing in Science & Engineering* **13**, 40-51 (2011).